154154

To: C. Elly

From: S. Kim

Re: Contract Data Review from Mead CompuChem
(EPA Case # 725)

contractual

① Poor Surrogate Spike Recovery for Acids (or 4%)

② D6-phenol should not be used as a surrogate spike compound.

③ Poor duplicate test results for acid compounds, 2-Fluorophenol, d6-phenol and pentafluorophenol which were not detected at all in one of duplicate run.

sediments give poor reprod. ok

④ Sigma unit (standard deviation) is not consistant in the Internal standard area control chart.

is too!

⑤ They should include a GC/MS data of 2,3,7,8-tetra chlorodibenzodioxine standard or that of the surrogate spike.