From: Harry Roels
To: Lobdell, Danelle
Cc: Rosemarie Bowler; Colledge, Michelle; Cheryl.beseler@colostate.edu; Hudgens, Edward; Bollweg, George
Subject: Re: FW: normal ranges for samples from November 2011
Date: Tuesday, July 30, 2013 12:52:28 PM

Thank you Danelle,

Is it possible to send me also a brief description of the laboratory methods used for these serum analyses and the references going with these methods?

Thank you in advance.

Harry

****

Op 30/07/2013 19:16, Lobdell, Danelle schreef:

From: Richards, Judy
Sent: Tuesday, July 30, 2013 1:14 PM
To: Lobdell, Danelle
Subject: RE: normal ranges for samples from November 2011

These are the normal ranges for the kits I use.
Ferritin:
Male: 7 to 253 ng/ml
Female: 2-110 ng/ml
ALT:
10-35 U/l
AST:
Up to 40 U/l
GGT:
Male: <50 U/l
Female: < 30 U/l
Judy

From: Lobdell, Danelle

Sent: Tuesday, July 30, 2013 12:05 PM
To: Richards, Judy
Subject: normal ranges for samples from November 2011
Hi Judy,
It has been a long time. We are currently analyzing the biomarkers for an upcoming paper and I am looking through my emails with you and the data that you sent and do not see normal values for ferritin, ALT, AST, and GGT. Samples were ran on November 16, 2011.
Thanks,
Danelle

Danelle T. Lobdell, Ph.D., M.S.
Epidemiologist
National Health and Environmental Effects Research Laboratory
Environmental Public Health Division
Mail:
USEPA
MD 58A
Research Triangle Park, NC 27711
Package Delivery:
USEPA Human Studies Facility
104 Mason Farm Rd, Room 52
Chapel Hill, NC 27514-4512
Phone: 919-843-4434 Fax: 919-966-7584

--
Prof. emer. Harry Roels
J. Pauwelslaan 3
3271 Averbode
Belgium

Tel.: +32-13-556955
Email: roelsharry@telenet.be