Analytical Data Package

Prepared by:

Pace Analytical Services

Pace Project No.: 40151601

Table Of Contents

Project Overview

Final Report with COC/SCUR . . . 1

SDG Narrative . . . 30

Organic

GC-ECD PCBs

Surrogate Recovery Summary (Form 2) . . . 31

Laboratory Control Spike/Laboratory Control Spike Duplicate Summary (Form 3) . . . 34

Matrix Spike/Matrix Spike Duplicate Summary (Form 3) . . . 37

Duplicate Summary (Form 3) . . . 40

Method Blank Association Summary (Form 4) . . . 45

Analytical Results with Raw Data (Form 1) . . . 48

Initial Calibration Summary (Form 6) . . . 102

Raw Data for Initial Calibration Standards . . . 109

Continuing Calibration Summary (Form 7) . . . 249

Raw Data for Continuing Calibration Standards . . . 261

Raw Data QC . . . 323

Preparation Logs Raw Data . . . 366

Instrument Run Log . . . 372

Standards Traceability . . . 392

Certificate of Analysis . . . 457

InOrganic

Gravimetric

Analytical Results (Form 1-IN) . . . 484

Blanks (Form 3-IN) . . . 490

Duplicates (Form 6-IN) . . . 493

Preparation Logs Raw Data . . . 498

Biota Prep Logs

Biota Prep Log . . . 503

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

May 03, 2018

Jim Ryan
Anchor QEA
290 Elwood Davis Road
Suite 340
Liverpool, NY 13088

RE: Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Dear Jim Ryan:
Enclosed are the analytical results for sample(s) received by the laboratory on June 14, 2017. The results relate only to the samples included in this report. Results reported herein conform to the most current, applicable TNI/NELAC standards and the laboratory's Quality Assurance Manual, where applicable, unless otherwise noted in the body of the report.

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Tod Noltemeyer

Tod Noltemeyer
tod.noltemeyer@pacelabs.com
(920)469-2436
Project Manager

Enclosures

cc: David Blye, Environmental Standards Inc.
Mark Meyers, Anchor QEA, LLC
Meg Michell, Environmental Standards Inc.

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

CERTIFICATIONS

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Green Bay Certification IDs

1241 Bellevue Street, Green Bay, WI 54302
Florida/NELAP Certification #: E87948
Illinois Certification #: 200050
Kentucky UST Certification #: 82
Louisiana Certification #: 04168
Minnesota Certification #: 055-999-334
New York Certification #: 12064
North Dakota Certification #: R-150
Virginia VELAP ID: 460263
South Carolina Certification #: 83006001
Texas Certification #: T104704529-14-1
Wisconsin Certification #: 405132750
Wisconsin DATCP Certification #: 105-444
USDA Soil Permit #: P330-16-00157
Federal Fish & Wildlife Permit #: LE51774A-0

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

SAMPLE SUMMARY

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Lab ID	Sample ID	Matrix	Date Collected	Date Received
40151601001	**TD2-170612-01-YB-01**	Tissue	06/12/17 13:38	06/14/17 09:50
40151601002	**TD2-170612-01-SMB-01**	Tissue	06/12/17 13:18	06/14/17 09:50
40151601003	**TD2-170612-01-BB-04**	Tissue	06/12/17 13:40	06/14/17 09:50
40151601004	**TD2-170612-01-BB-03**	Tissue	06/12/17 13:36	06/14/17 09:50
40151601005	**TD2-170612-01-BB-02**	Tissue	06/12/17 13:32	06/14/17 09:50
40151601006	**TD2-170612-01-BB-01**	Tissue	06/12/17 13:31	06/14/17 09:50

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

SAMPLE ANALYTE COUNT

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Lab ID	Sample ID	Method	Analysts	Analytes Reported
40151601001	**TD2-170612-01-YB-01**	EPA 8082A Mod	BLM	10
		Pace Gender Typing	CWN	1
		Pace Lipid	ETB	1
40151601002	**TD2-170612-01-SMB-01**	EPA 8082A Mod	BLM	10
		Pace Gender Typing	CWN	1
		Pace Lipid	AMC	1
40151601003	**TD2-170612-01-BB-04**	EPA 8082A Mod	BLM	10
		Pace Gender Typing	CWN	1
		Pace Lipid	AMC	1
40151601004	**TD2-170612-01-BB-03**	EPA 8082A Mod	BLM	10
		Pace Gender Typing	CWN	1
		Pace Lipid	AMC	1
40151601005	**TD2-170612-01-BB-02**	EPA 8082A Mod	BLM	10
		Pace Gender Typing	CWN	1
		Pace Lipid	AMC	1
40151601006	**TD2-170612-01-BB-01**	EPA 8082A Mod	BLM	10
		Pace Gender Typing	CWN	1
		Pace Lipid	AMC	1

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

PROJECT NARRATIVE

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Date: May 03, 2018

Samples may have ice meltwater in the ziplock bags.

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

PROJECT NARRATIVE

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Method: **EPA 8082A Mod**
Description: 8082A Modified GCT PCB
Client: General Electric
Date: May 03, 2018

General Information:
6 samples were analyzed for EPA 8082A Mod. All samples were received in acceptable condition with any exceptions noted below or on the chain-of custody and/or the sample condition upon receipt form (SCUR) attached at the end of this report.

Hold Time:
The samples were analyzed within the method required hold times with any exceptions noted below.

Sample Preparation:
The samples were prepared in accordance with EPA 3541 with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):
All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:
All criteria were within method requirements with any exceptions noted below.

Surrogates:
All surrogates were within QC limits with any exceptions noted below.

QC Batch: 282598

S4: Surrogate recovery not evaluated against control limits due to sample dilution.
- DUP (Lab ID: 1655665)
  - Decachlorobiphenyl (S)
  - Tetrachloro-m-xylene (S)

Method Blank:
All analytes were below the report limit in the method blank, where applicable, with any exceptions noted below.

Laboratory Control Spike:
All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:
All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

QC Batch: 282598

A matrix spike and/or matrix spike duplicate (MS/MSD) were performed on the following sample(s): 40150295019

M1: Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.
- MS (Lab ID: 1655663)
  - PCB-1242 (Aroclor 1242)

QC Batch: 283116

A matrix spike and/or matrix spike duplicate (MS/MSD) were performed on the following sample(s): 40151552010

M1: Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.
- MS (Lab ID: 1658256)

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

PROJECT NARRATIVE

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Method: **EPA 8082A Mod**
Description: 8082A Modified GCT PCB
Client: General Electric
Date: May 03, 2018

QC Batch: 283116

A matrix spike and/or matrix spike duplicate (MS/MSD) were performed on the following sample(s): 40151552010

M1: Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

- PCB-1242 (Aroclor 1242)

Duplicate Sample:
All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

Batch Comments:

The set was re-extracted due to extraction error.

- QC Batch: 283147

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

PROJECT NARRATIVE

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Method: **Pace Gender Typing**
Description: Fish Gender Typing
Client: General Electric
Date: May 03, 2018

General Information:
6 samples were analyzed for Pace Gender Typing. All samples were received in acceptable condition with any exceptions noted below or on the chain-of custody and/or the sample condition upon receipt form (SCUR) attached at the end of this report.

Hold Time:
The samples were analyzed within the method required hold times with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):
All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:
All criteria were within method requirements with any exceptions noted below.

Internal Standards:
All internal standards were within QC limits with any exceptions noted below.

Surrogates:
All surrogates were within QC limits with any exceptions noted below.

Method Blank:
All analytes were below the report limit in the method blank, where applicable, with any exceptions noted below.

Laboratory Control Spike:
All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:
All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Additional Comments:

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

PROJECT NARRATIVE

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Method: **Pace Lipid**
Description: Lipid
Client: General Electric
Date: May 03, 2018

General Information:
6 samples were analyzed for Pace Lipid. All samples were received in acceptable condition with any exceptions noted below or on the chain-of custody and/or the sample condition upon receipt form (SCUR) attached at the end of this report.

Hold Time:
The samples were analyzed within the method required hold times with any exceptions noted below.

Initial Calibrations (including MS Tune as applicable):
All criteria were within method requirements with any exceptions noted below.

Continuing Calibration:
All criteria were within method requirements with any exceptions noted below.

Internal Standards:
All internal standards were within QC limits with any exceptions noted below.

Surrogates:
All surrogates were within QC limits with any exceptions noted below.

Method Blank:
All analytes were below the report limit in the method blank, where applicable, with any exceptions noted below.

Laboratory Control Spike:
All laboratory control spike compounds were within QC limits with any exceptions noted below.

Matrix Spikes:
All percent recoveries and relative percent differences (RPDs) were within acceptance criteria with any exceptions noted below.

Duplicate Sample:
All duplicate sample results were within method acceptance criteria with any exceptions noted below.

Additional Comments:

This data package has been reviewed for quality and completeness and is approved for release.

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

ANALYTICAL RESULTS

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Sample: TD2-170612-01-YB-01 **Lab ID: 40151601001** Collected: 06/12/17 13:38 Received: 06/14/17 09:50 Matrix: Tissue

Results reported on a "wet-weight" basis

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082A Modified GCT PCB	Analytical Method: EPA 8082A Mod Preparation Method: EPA 3541								
PCB-1016 (Aroclor 1016)	**<11.4**	ug/kg	49.9	11.4	1	03/28/18 08:09	03/30/18 18:25	12674-11-2	
PCB-1221 (Aroclor 1221)	**<11.4**	ug/kg	49.9	11.4	1	03/28/18 08:09	03/30/18 18:25	11104-28-2	
PCB-1232 (Aroclor 1232)	**<11.4**	ug/kg	49.9	11.4	1	03/28/18 08:09	03/30/18 18:25	11141-16-5	
PCB-1242 (Aroclor 1242)	**<11.4**	ug/kg	49.9	11.4	1	03/28/18 08:09	03/30/18 18:25	53469-21-9	
PCB-1248 (Aroclor 1248)	**80.1**	ug/kg	49.9	11.4	1	03/28/18 08:09	03/30/18 18:25	12672-29-6	
PCB-1254 (Aroclor 1254)	**40.2J**	ug/kg	49.9	11.4	1	03/28/18 08:09	03/30/18 18:25	11097-69-1	
PCB-1260 (Aroclor 1260)	**15.5J**	ug/kg	49.9	11.4	1	03/28/18 08:09	03/30/18 18:25	11096-82-5	
PCB, Total	**136**	ug/kg	49.9	11.4	1	03/28/18 08:09	03/30/18 18:25	1336-36-3	
Surrogates									
Tetrachloro-m-xylene (S)	89	%	60-140		1	03/28/18 08:09	03/30/18 18:25	877-09-8	
Decachlorobiphenyl (S)	90	%	60-140		1	03/28/18 08:09	03/30/18 18:25	2051-24-3	
Fish Gender Typing	Analytical Method: Pace Gender Typing								
Gender	**Female**				1		03/06/18 11:11		
Lipid	Analytical Method: Pace Lipid								
Lipid	**0.47**	%	0.030	0.030	1		03/29/18 07:08		

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

ANALYTICAL RESULTS

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Sample: TD2-170612-01-SMB-01 **Lab ID: 40151601002** Collected: 06/12/17 13:18 Received: 06/14/17 09:50 Matrix: Tissue

Results reported on a "wet-weight" basis

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082A Modified GCT PCB	Analytical Method: EPA 8082A Mod Preparation Method: EPA 3541								
PCB-1016 (Aroclor 1016)	**<11.4**	ug/kg	50.0	11.4	1	03/12/18 07:42	03/13/18 16:56	12674-11-2	
PCB-1221 (Aroclor 1221)	**<11.4**	ug/kg	50.0	11.4	1	03/12/18 07:42	03/13/18 16:56	11104-28-2	
PCB-1232 (Aroclor 1232)	**<11.4**	ug/kg	50.0	11.4	1	03/12/18 07:42	03/13/18 16:56	11141-16-5	
PCB-1242 (Aroclor 1242)	**<11.4**	ug/kg	50.0	11.4	1	03/12/18 07:42	03/13/18 16:56	53469-21-9	
PCB-1248 (Aroclor 1248)	**110**	ug/kg	50.0	11.4	1	03/12/18 07:42	03/13/18 16:56	12672-29-6	
PCB-1254 (Aroclor 1254)	**79.4**	ug/kg	50.0	11.4	1	03/12/18 07:42	03/13/18 16:56	11097-69-1	
PCB-1260 (Aroclor 1260)	**56.1**	ug/kg	50.0	11.4	1	03/12/18 07:42	03/13/18 16:56	11096-82-5	
PCB, Total	**245**	ug/kg	50.0	11.4	1	03/12/18 07:42	03/13/18 16:56	1336-36-3	
Surrogates									
Tetrachloro-m-xylene (S)	84	%	60-140		1	03/12/18 07:42	03/13/18 16:56	877-09-8	
Decachlorobiphenyl (S)	87	%	60-140		1	03/12/18 07:42	03/13/18 16:56	2051-24-3	
Fish Gender Typing	Analytical Method: Pace Gender Typing								
Gender	**Female**				1		03/06/18 11:11		
Lipid	Analytical Method: Pace Lipid								
Lipid	**0.37**	%	0.030	0.030	1		03/13/18 08:59		

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

ANALYTICAL RESULTS

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Sample: TD2-170612-01-BB-04 **Lab ID: 40151601003** Collected: 06/12/17 13:40 Received: 06/14/17 09:50 Matrix: Tissue

Results reported on a "wet-weight" basis

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082A Modified GCT PCB	Analytical Method: EPA 8082A Mod Preparation Method: EPA 3541								
PCB-1016 (Aroclor 1016)	**<11.4**	ug/kg	49.8	11.4	1	03/13/18 07:38	03/14/18 17:41	12674-11-2	
PCB-1221 (Aroclor 1221)	**<11.4**	ug/kg	49.8	11.4	1	03/13/18 07:38	03/14/18 17:41	11104-28-2	
PCB-1232 (Aroclor 1232)	**<11.4**	ug/kg	49.8	11.4	1	03/13/18 07:38	03/14/18 17:41	11141-16-5	
PCB-1242 (Aroclor 1242)	**<11.4**	ug/kg	49.8	11.4	1	03/13/18 07:38	03/14/18 17:41	53469-21-9	
PCB-1248 (Aroclor 1248)	**268**	ug/kg	49.8	11.4	1	03/13/18 07:38	03/14/18 17:41	12672-29-6	
PCB-1254 (Aroclor 1254)	**127**	ug/kg	49.8	11.4	1	03/13/18 07:38	03/14/18 17:41	11097-69-1	
PCB-1260 (Aroclor 1260)	**20.0J**	ug/kg	49.8	11.4	1	03/13/18 07:38	03/14/18 17:41	11096-82-5	
PCB, Total	**415**	ug/kg	49.8	11.4	1	03/13/18 07:38	03/14/18 17:41	1336-36-3	
Surrogates									
Tetrachloro-m-xylene (S)	99	%	60-140		1	03/13/18 07:38	03/14/18 17:41	877-09-8	
Decachlorobiphenyl (S)	100	%	60-140		1	03/13/18 07:38	03/14/18 17:41	2051-24-3	
Fish Gender Typing	Analytical Method: Pace Gender Typing								
Gender	**Female**				1		03/06/18 11:11		
Lipid	Analytical Method: Pace Lipid								
Lipid	**0.90**	%	0.030	0.030	1		03/14/18 06:49		

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

ANALYTICAL RESULTS

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Sample: TD2-170612-01-BB-03 **Lab ID: 40151601004** Collected: 06/12/17 13:36 Received: 06/14/17 09:50 Matrix: Tissue

Results reported on a "wet-weight" basis

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082A Modified GCT PCB	Analytical Method: EPA 8082A Mod Preparation Method: EPA 3541								
PCB-1016 (Aroclor 1016)	**<11.4**	ug/kg	50.0	11.4	1	03/13/18 07:38	03/14/18 17:57	12674-11-2	
PCB-1221 (Aroclor 1221)	**<11.4**	ug/kg	50.0	11.4	1	03/13/18 07:38	03/14/18 17:57	11104-28-2	
PCB-1232 (Aroclor 1232)	**<11.4**	ug/kg	50.0	11.4	1	03/13/18 07:38	03/14/18 17:57	11141-16-5	
PCB-1242 (Aroclor 1242)	**<11.4**	ug/kg	50.0	11.4	1	03/13/18 07:38	03/14/18 17:57	53469-21-9	
PCB-1248 (Aroclor 1248)	**96.8**	ug/kg	50.0	11.4	1	03/13/18 07:38	03/14/18 17:57	12672-29-6	
PCB-1254 (Aroclor 1254)	**49.3J**	ug/kg	50.0	11.4	1	03/13/18 07:38	03/14/18 17:57	11097-69-1	
PCB-1260 (Aroclor 1260)	**<11.4**	ug/kg	50.0	11.4	1	03/13/18 07:38	03/14/18 17:57	11096-82-5	
PCB, Total	**146**	ug/kg	50.0	11.4	1	03/13/18 07:38	03/14/18 17:57	1336-36-3	
Surrogates									
Tetrachloro-m-xylene (S)	77	%	60-140		1	03/13/18 07:38	03/14/18 17:57	877-09-8	
Decachlorobiphenyl (S)	83	%	60-140		1	03/13/18 07:38	03/14/18 17:57	2051-24-3	
Fish Gender Typing	Analytical Method: Pace Gender Typing								
Gender	**Male**				1		03/06/18 11:11		
Lipid	Analytical Method: Pace Lipid								
Lipid	**0.36**	%	0.030	0.030	1		03/14/18 06:49		

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

ANALYTICAL RESULTS

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Sample: TD2-170612-01-BB-02 **Lab ID: 40151601005** Collected: 06/12/17 13:32 Received: 06/14/17 09:50 Matrix: Tissue

Results reported on a "wet-weight" basis

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082A Modified GCT PCB	Analytical Method: EPA 8082A Mod Preparation Method: EPA 3541								
PCB-1016 (Aroclor 1016)	**<11.4**	ug/kg	49.9	11.4	1	03/13/18 07:38	03/14/18 18:13	12674-11-2	
PCB-1221 (Aroclor 1221)	**<11.4**	ug/kg	49.9	11.4	1	03/13/18 07:38	03/14/18 18:13	11104-28-2	
PCB-1232 (Aroclor 1232)	**<11.4**	ug/kg	49.9	11.4	1	03/13/18 07:38	03/14/18 18:13	11141-16-5	
PCB-1242 (Aroclor 1242)	**<11.4**	ug/kg	49.9	11.4	1	03/13/18 07:38	03/14/18 18:13	53469-21-9	
PCB-1248 (Aroclor 1248)	**217**	ug/kg	49.9	11.4	1	03/13/18 07:38	03/14/18 18:13	12672-29-6	
PCB-1254 (Aroclor 1254)	**130**	ug/kg	49.9	11.4	1	03/13/18 07:38	03/14/18 18:13	11097-69-1	
PCB-1260 (Aroclor 1260)	**29.6J**	ug/kg	49.9	11.4	1	03/13/18 07:38	03/14/18 18:13	11096-82-5	
PCB, Total	**377**	ug/kg	49.9	11.4	1	03/13/18 07:38	03/14/18 18:13	1336-36-3	
Surrogates									
Tetrachloro-m-xylene (S)	95	%	60-140		1	03/13/18 07:38	03/14/18 18:13	877-09-8	
Decachlorobiphenyl (S)	95	%	60-140		1	03/13/18 07:38	03/14/18 18:13	2051-24-3	
Fish Gender Typing	Analytical Method: Pace Gender Typing								
Gender	**Female**				1		03/06/18 11:11		
Lipid	Analytical Method: Pace Lipid								
Lipid	**0.62**	%	0.030	0.030	1		03/14/18 06:49		

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

ANALYTICAL RESULTS

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Sample: TD2-170612-01-BB-01 **Lab ID: 40151601006** Collected: 06/12/17 13:31 Received: 06/14/17 09:50 Matrix: Tissue

Results reported on a "wet-weight" basis

Parameters	Results	Units	PQL	MDL	DF	Prepared	Analyzed	CAS No.	Qual
8082A Modified GCT PCB	Analytical Method: EPA 8082A Mod Preparation Method: EPA 3541								
PCB-1016 (Aroclor 1016)	**<11.4**	ug/kg	50.1	11.4	1	03/13/18 07:38	03/14/18 18:29	12674-11-2	
PCB-1221 (Aroclor 1221)	**<11.4**	ug/kg	50.1	11.4	1	03/13/18 07:38	03/14/18 18:29	11104-28-2	
PCB-1232 (Aroclor 1232)	**<11.4**	ug/kg	50.1	11.4	1	03/13/18 07:38	03/14/18 18:29	11141-16-5	
PCB-1242 (Aroclor 1242)	**<11.4**	ug/kg	50.1	11.4	1	03/13/18 07:38	03/14/18 18:29	53469-21-9	
PCB-1248 (Aroclor 1248)	**239**	ug/kg	50.1	11.4	1	03/13/18 07:38	03/14/18 18:29	12672-29-6	
PCB-1254 (Aroclor 1254)	**108**	ug/kg	50.1	11.4	1	03/13/18 07:38	03/14/18 18:29	11097-69-1	
PCB-1260 (Aroclor 1260)	**22.1J**	ug/kg	50.1	11.4	1	03/13/18 07:38	03/14/18 18:29	11096-82-5	
PCB, Total	**369**	ug/kg	50.1	11.4	1	03/13/18 07:38	03/14/18 18:29	1336-36-3	
Surrogates									
Tetrachloro-m-xylene (S)	97	%	60-140		1	03/13/18 07:38	03/14/18 18:29	877-09-8	
Decachlorobiphenyl (S)	95	%	60-140		1	03/13/18 07:38	03/14/18 18:29	2051-24-3	
Fish Gender Typing	Analytical Method: Pace Gender Typing								
Gender	**Male**				1		03/06/18 11:11		
Lipid	Analytical Method: Pace Lipid								
Lipid	**1.4**	%	0.030	0.030	1		03/14/18 06:49		

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

QUALITY CONTROL DATA

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

QC Batch: 282598 | Analysis Method: EPA 8082A Mod
QC Batch Method: EPA 3541 | Analysis Description: 8082 GCS PCB
Associated Lab Samples: 40151601002

METHOD BLANK: 1655661 | Matrix: Tissue
Associated Lab Samples: 40151601002

Parameter	Units	Blank Result	Reporting Limit	MDL	Analyzed	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg	<11.4	50.0	11.4	03/13/18 14:50	
PCB-1221 (Aroclor 1221)	ug/kg	<11.4	50.0	11.4	03/13/18 14:50	
PCB-1232 (Aroclor 1232)	ug/kg	<11.4	50.0	11.4	03/13/18 14:50	
PCB-1242 (Aroclor 1242)	ug/kg	<11.4	50.0	11.4	03/13/18 14:50	
PCB-1248 (Aroclor 1248)	ug/kg	<11.4	50.0	11.4	03/13/18 14:50	
PCB-1254 (Aroclor 1254)	ug/kg	<11.4	50.0	11.4	03/13/18 14:50	
PCB-1260 (Aroclor 1260)	ug/kg	<11.4	50.0	11.4	03/13/18 14:50	
Decachlorobiphenyl (S)	%	83	60-140		03/13/18 14:50	
Tetrachloro-m-xylene (S)	%	73	60-140		03/13/18 14:50	

LABORATORY CONTROL SAMPLE: 1655662

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg		<11.4			
PCB-1221 (Aroclor 1221)	ug/kg		<11.4			
PCB-1232 (Aroclor 1232)	ug/kg		<11.4			
PCB-1242 (Aroclor 1242)	ug/kg	250	212	85	70-130	
PCB-1248 (Aroclor 1248)	ug/kg		<11.4			
PCB-1254 (Aroclor 1254)	ug/kg		<11.4			
PCB-1260 (Aroclor 1260)	ug/kg		<11.4			
Decachlorobiphenyl (S)	%			82	60-140	
Tetrachloro-m-xylene (S)	%			69	60-140	

MATRIX SPIKE SAMPLE: 1655663

Parameter	Units	40150295019 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg	<56.9		<56.9			
PCB-1221 (Aroclor 1221)	ug/kg	<56.9		<56.9			
PCB-1232 (Aroclor 1232)	ug/kg	<56.9		<56.9			
PCB-1242 (Aroclor 1242)	ug/kg	<56.9	250	519	208	70-130	M1
PCB-1248 (Aroclor 1248)	ug/kg	1040		1220			
PCB-1254 (Aroclor 1254)	ug/kg	1000		1060			
PCB-1260 (Aroclor 1260)	ug/kg	510		517			
Decachlorobiphenyl (S)	%				103	60-140	
Tetrachloro-m-xylene (S)	%				80	60-140	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

QUALITY CONTROL DATA

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

SAMPLE DUPLICATE: 1655664

Parameter	Units	40150295019 Result	Dup Result	RPD	Max RPD	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg	<56.9	<56.9		40	
PCB-1221 (Aroclor 1221)	ug/kg	<56.9	<56.9		40	
PCB-1232 (Aroclor 1232)	ug/kg	<56.9	<56.9		40	
PCB-1242 (Aroclor 1242)	ug/kg	<56.9	<56.9		40	
PCB-1248 (Aroclor 1248)	ug/kg	1040	786	28	40	
PCB-1254 (Aroclor 1254)	ug/kg	1000	746	29	40	
PCB-1260 (Aroclor 1260)	ug/kg	510	375	31	40	
Decachlorobiphenyl (S)	%	115	100	14		
Tetrachloro-m-xylene (S)	%	91	83	9		

SAMPLE DUPLICATE: 1655665

Parameter	Units	40151804007 Result	Dup Result	RPD	Max RPD	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg	<170	<171		40	
PCB-1221 (Aroclor 1221)	ug/kg	<170	<171		40	
PCB-1232 (Aroclor 1232)	ug/kg	<170	<171		40	
PCB-1242 (Aroclor 1242)	ug/kg	<170	<171		40	
PCB-1248 (Aroclor 1248)	ug/kg	2530	2760	9	40	
PCB-1254 (Aroclor 1254)	ug/kg	2420	2660	9	40	
PCB-1260 (Aroclor 1260)	ug/kg	693J	792		40	
Decachlorobiphenyl (S)	%	0	0			S4
Tetrachloro-m-xylene (S)	%	0	0			S4

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

QUALITY CONTROL DATA

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

QC Batch: 283116 | Analysis Method: EPA 8082A Mod
QC Batch Method: EPA 3541 | Analysis Description: 8082 GCS PCB
Associated Lab Samples: 40151601003, 40151601004, 40151601005, 40151601006

METHOD BLANK: 1658254 | Matrix: Tissue
Associated Lab Samples: 40151601003, 40151601004, 40151601005, 40151601006

Parameter	Units	Blank Result	Reporting Limit	MDL	Analyzed	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg	<11.4	50.0	11.4	03/14/18 13:44	
PCB-1221 (Aroclor 1221)	ug/kg	<11.4	50.0	11.4	03/14/18 13:44	
PCB-1232 (Aroclor 1232)	ug/kg	<11.4	50.0	11.4	03/14/18 13:44	
PCB-1242 (Aroclor 1242)	ug/kg	<11.4	50.0	11.4	03/14/18 13:44	
PCB-1248 (Aroclor 1248)	ug/kg	<11.4	50.0	11.4	03/14/18 13:44	
PCB-1254 (Aroclor 1254)	ug/kg	<11.4	50.0	11.4	03/14/18 13:44	
PCB-1260 (Aroclor 1260)	ug/kg	<11.4	50.0	11.4	03/14/18 13:44	
Decachlorobiphenyl (S)	%	83	60-140		03/14/18 13:44	
Tetrachloro-m-xylene (S)	%	76	60-140		03/14/18 13:44	

LABORATORY CONTROL SAMPLE: 1658255

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg		<11.4			
PCB-1221 (Aroclor 1221)	ug/kg		<11.4			
PCB-1232 (Aroclor 1232)	ug/kg		<11.4			
PCB-1242 (Aroclor 1242)	ug/kg	250	232	93	70-130	
PCB-1248 (Aroclor 1248)	ug/kg		<11.4			
PCB-1254 (Aroclor 1254)	ug/kg		<11.4			
PCB-1260 (Aroclor 1260)	ug/kg		<11.4			
Decachlorobiphenyl (S)	%			92	60-140	
Tetrachloro-m-xylene (S)	%			85	60-140	

MATRIX SPIKE SAMPLE: 1658256

Parameter	Units	40151552010 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg	<34.3		<34.2			
PCB-1221 (Aroclor 1221)	ug/kg	<34.3		<34.2			
PCB-1232 (Aroclor 1232)	ug/kg	<34.3		<34.2			
PCB-1242 (Aroclor 1242)	ug/kg	<34.3	250	354	142	70-130	M1
PCB-1248 (Aroclor 1248)	ug/kg	795		966			
PCB-1254 (Aroclor 1254)	ug/kg	589		630			
PCB-1260 (Aroclor 1260)	ug/kg	284		295			
Decachlorobiphenyl (S)	%				95	60-140	
Tetrachloro-m-xylene (S)	%				89	60-140	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

QUALITY CONTROL DATA

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

SAMPLE DUPLICATE: 1658257

Parameter	Units	40151552010 Result	Dup Result	RPD	Max RPD	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg	<34.3	<34.2		40	
PCB-1221 (Aroclor 1221)	ug/kg	<34.3	<34.2		40	
PCB-1232 (Aroclor 1232)	ug/kg	<34.3	<34.2		40	
PCB-1242 (Aroclor 1242)	ug/kg	<34.3	<34.2		40	
PCB-1248 (Aroclor 1248)	ug/kg	795	853	7	40	
PCB-1254 (Aroclor 1254)	ug/kg	589	628	6	40	
PCB-1260 (Aroclor 1260)	ug/kg	284	299	5	40	
Decachlorobiphenyl (S)	%	101	101	0		
Tetrachloro-m-xylene (S)	%	93	92	2		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

QUALITY CONTROL DATA

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

QC Batch: 284454 | Analysis Method: EPA 8082A Mod
QC Batch Method: EPA 3541 | Analysis Description: 8082 GCS PCB
Associated Lab Samples: 40151601001

METHOD BLANK: 1664793 | Matrix: Tissue
Associated Lab Samples: 40151601001

Parameter	Units	Blank Result	Reporting Limit	MDL	Analyzed	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg	<11.4	50.0	11.4	03/30/18 17:22	
PCB-1221 (Aroclor 1221)	ug/kg	<11.4	50.0	11.4	03/30/18 17:22	
PCB-1232 (Aroclor 1232)	ug/kg	<11.4	50.0	11.4	03/30/18 17:22	
PCB-1242 (Aroclor 1242)	ug/kg	<11.4	50.0	11.4	03/30/18 17:22	
PCB-1248 (Aroclor 1248)	ug/kg	<11.4	50.0	11.4	03/30/18 17:22	
PCB-1254 (Aroclor 1254)	ug/kg	<11.4	50.0	11.4	03/30/18 17:22	
PCB-1260 (Aroclor 1260)	ug/kg	<11.4	50.0	11.4	03/30/18 17:22	
Decachlorobiphenyl (S)	%	89	60-140		03/30/18 17:22	
Tetrachloro-m-xylene (S)	%	76	60-140		03/30/18 17:22	

LABORATORY CONTROL SAMPLE: 1664794

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg		<11.4			
PCB-1221 (Aroclor 1221)	ug/kg		<11.4			
PCB-1232 (Aroclor 1232)	ug/kg		<11.4			
PCB-1242 (Aroclor 1242)	ug/kg	250	236	95	70-130	
PCB-1248 (Aroclor 1248)	ug/kg		<11.4			
PCB-1254 (Aroclor 1254)	ug/kg		<11.4			
PCB-1260 (Aroclor 1260)	ug/kg		<11.4			
Decachlorobiphenyl (S)	%			92	60-140	
Tetrachloro-m-xylene (S)	%			81	60-140	

MATRIX SPIKE SAMPLE: 1664795

Parameter	Units	40151613001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg	<11.4		<11.4			
PCB-1221 (Aroclor 1221)	ug/kg	<11.4		<11.4			
PCB-1232 (Aroclor 1232)	ug/kg	<11.4		<11.4			
PCB-1242 (Aroclor 1242)	ug/kg	<11.4	251	238	95	70-130	
PCB-1248 (Aroclor 1248)	ug/kg	11.8J		<11.4			
PCB-1254 (Aroclor 1254)	ug/kg	16.9J		43.4J			
PCB-1260 (Aroclor 1260)	ug/kg	39.1J		35.7J			
Decachlorobiphenyl (S)	%				82	60-140	
Tetrachloro-m-xylene (S)	%				79	60-140	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

QUALITY CONTROL DATA

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

SAMPLE DUPLICATE: 1664796

Parameter	Units	40151613001 Result	Dup Result	RPD	Max RPD	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg	<11.4	<11.4		40	
PCB-1221 (Aroclor 1221)	ug/kg	<11.4	<11.4		40	
PCB-1232 (Aroclor 1232)	ug/kg	<11.4	<11.4		40	
PCB-1242 (Aroclor 1242)	ug/kg	<11.4	<11.4		40	
PCB-1248 (Aroclor 1248)	ug/kg	11.8J	11.4J		40	
PCB-1254 (Aroclor 1254)	ug/kg	16.9J	16.6J		40	
PCB-1260 (Aroclor 1260)	ug/kg	39.1J	36.6J		40	
Decachlorobiphenyl (S)	%	85	90	5		
Tetrachloro-m-xylene (S)	%	82	81	1		

SAMPLE DUPLICATE: 1664809

Parameter	Units	40157825001 Result	Dup Result	RPD	Max RPD	Qualifiers
PCB-1016 (Aroclor 1016)	ug/kg	<11.4	<11.4		40	
PCB-1221 (Aroclor 1221)	ug/kg	<11.4	<11.4		40	
PCB-1232 (Aroclor 1232)	ug/kg	<11.4	<11.4		40	
PCB-1242 (Aroclor 1242)	ug/kg	<11.4	<11.4		40	
PCB-1248 (Aroclor 1248)	ug/kg	272	344	23	40	
PCB-1254 (Aroclor 1254)	ug/kg	114	147	25	40	
PCB-1260 (Aroclor 1260)	ug/kg	20.1J	26.4J		40	
Decachlorobiphenyl (S)	%	90	93	2		
Tetrachloro-m-xylene (S)	%	86	85	1		

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

QUALITY CONTROL DATA

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

QC Batch: 282623 | Analysis Method: Pace Lipid
QC Batch Method: Pace Lipid | Analysis Description: LIPID
Associated Lab Samples: 40151601002

METHOD BLANK: 1655713 | Matrix: Tissue
Associated Lab Samples: 40151601002

Parameter	Units	Blank Result	Reporting Limit	MDL	Analyzed	Qualifiers
Lipid	%	<0.030	0.030	0.030	03/13/18 08:59	

MATRIX SPIKE SAMPLE: 1682460

Parameter	Units	40150295019 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lipid	%	3.0		3.0			

SAMPLE DUPLICATE: 1655716

Parameter	Units	40150295019 Result	Dup Result	RPD	Max RPD	Qualifiers
Lipid	%	3.0	2.1	36	40	

SAMPLE DUPLICATE: 1655717

Parameter	Units	40151804007 Result	Dup Result	RPD	Max RPD	Qualifiers
Lipid	%	0.39	0.41	3	40	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

QUALITY CONTROL DATA

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

QC Batch: 283118 | Analysis Method: Pace Lipid
QC Batch Method: Pace Lipid | Analysis Description: LIPID
Associated Lab Samples: 40151601003, 40151601004, 40151601005, 40151601006

METHOD BLANK: 1658258 | Matrix: Tissue
Associated Lab Samples: 40151601003, 40151601004, 40151601005, 40151601006

Parameter	Units	Blank Result	Reporting Limit	MDL	Analyzed	Qualifiers
Lipid	%	<0.030	0.030	0.030	03/14/18 06:49	

MATRIX SPIKE SAMPLE: 1682499

Parameter	Units	40151552010 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lipid	%	1.5		1.6			

SAMPLE DUPLICATE: 1658259

Parameter	Units	40151552010 Result	Dup Result	RPD	Max RPD	Qualifiers
Lipid	%	1.5	1.5	5	40	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

QUALITY CONTROL DATA

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

QC Batch: 284607 | Analysis Method: Pace Lipid
QC Batch Method: Pace Lipid | Analysis Description: LIPID
Associated Lab Samples: 40151601001

METHOD BLANK: 1665662 | Matrix: Tissue
Associated Lab Samples: 40151601001

Parameter	Units	Blank Result	Reporting Limit	MDL	Analyzed	Qualifiers
Lipid	%	<0.030	0.030	0.030	03/29/18 07:08	

MATRIX SPIKE SAMPLE: 1682554

Parameter	Units	40151613001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lipid	%	0.30		0.27			

SAMPLE DUPLICATE: 1665665

Parameter	Units	40151613001 Result	Dup Result	RPD	Max RPD	Qualifiers
Lipid	%	0.30	0.28	7	40	

SAMPLE DUPLICATE: 1665666

Parameter	Units	40157825001 Result	Dup Result	RPD	Max RPD	Qualifiers
Lipid	%	0.39	0.44	13	40	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

QUALIFIERS

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

WORKORDER QUALIFIERS

WO: 40151601

[1] Samples may have ice meltwater in the ziplock bags.

BATCH QUALIFIERS

Batch: 283147

[1] The set was re-extracted due to extraction error.

ANALYTE QUALIFIERS

M1 Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

S4 Surrogate recovery not evaluated against control limits due to sample dilution.

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace Analytical Services, LLC
1241 Bellevue Street - Suite 9
Green Bay, WI 54302
(920)469-2436

QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: HUDSON RIVER REMEDIAL ACTION
Pace Project No.: 40151601

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
40151601001	**TD2-170612-01-YB-01**	EPA 3541	284454	EPA 8082A Mod	284630
40151601002	**TD2-170612-01-SMB-01**	EPA 3541	282598	EPA 8082A Mod	283147
40151601003	**TD2-170612-01-BB-04**	EPA 3541	283116	EPA 8082A Mod	283254
40151601004	**TD2-170612-01-BB-03**	EPA 3541	283116	EPA 8082A Mod	283254
40151601005	**TD2-170612-01-BB-02**	EPA 3541	283116	EPA 8082A Mod	283254
40151601006	**TD2-170612-01-BB-01**	EPA 3541	283116	EPA 8082A Mod	283254
40151601001	**TD2-170612-01-YB-01**	Pace Gender Typing	282519		
40151601002	**TD2-170612-01-SMB-01**	Pace Gender Typing	282519		
40151601003	**TD2-170612-01-BB-04**	Pace Gender Typing	282519		
40151601004	**TD2-170612-01-BB-03**	Pace Gender Typing	282519		
40151601005	**TD2-170612-01-BB-02**	Pace Gender Typing	282519		
40151601006	**TD2-170612-01-BB-01**	Pace Gender Typing	282519		
40151601001	**TD2-170612-01-YB-01**	Pace Lipid	284607		
40151601002	**TD2-170612-01-SMB-01**	Pace Lipid	282623		
40151601003	**TD2-170612-01-BB-04**	Pace Lipid	283118		
40151601004	**TD2-170612-01-BB-03**	Pace Lipid	283118		
40151601005	**TD2-170612-01-BB-02**	Pace Lipid	283118		
40151601006	**TD2-170612-01-BB-01**	Pace Lipid	283118		

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

RMN

40151601

ANCHOR QEA

305 West Grand Avenue Montvale, NJ 07645 Ph: 201-930-9890

ENVIRONMENTAL SAMPLE CHAIN OF CUSTODY

COC ID: COC170612-T11-03
Sample Custodian: JER
Lab: Pace - Green Bay

Client: General Electric Company
Project: Hudson River Remedial Action Monitoring Program

COC Sample Number	Sampling Event ID	Field Sample ID	Sample Date	Sample Time	MS/LD	# CONTAINERS	Media*	Prep**	Composite	EPA Split	Total PCBs NE013_09	Total PCBs SW846 8082A	Percent Lipid (S-GB-O-068)	Fillet Weight (S-GB-L-009)	
001	TD5-170612-01	TD5-170612-01-YB-01	6/12/2017	15:28	☐	1	F	FLRC	☐	☐	☐	☑	☑	☑	
002	TD5-170612-01	TD5-170612-01-BB-09	6/12/2017	15:33	☐	1	F	FLRC	☐	☐	☐	☑	☑	☑	
003	TD5-170612-01	TD5-170612-01-BB-08	6/12/2017	15:31	☐	1	F	FLRC	☐	☐	☐	☑	☑	☑	
004	TD5-170612-01	TD5-170612-01-BB-07	6/12/2017	15:26	☐	1	F	FLRC	☐	☐	☐	☑	☑	☑	
005	TD5-170612-01	TD5-170612-01-BB-06	6/12/2017	15:24	☐	1	F	FLRC	☐	☐	☐	☑	☑	☑	
006	TD5-170612-01	TD5-170612-01-BB-05	6/12/2017	15:22	☐	1	F	FLRC	☐	☐	☐	☑	☑	☑	
007	TD5-170612-01	TD5-170612-01-BB-04	6/12/2017	15:20	☐	1	F	FLRC	☐	☐	☐	☑	☑	☑	
008	TD5-170612-01	TD5-170612-01-BB-03	6/12/2017	15:17	☐	1	F	FLRC	☐	☐	☐	☑	☑	☑	
009	TD5-170612-01	TD5-170612-01-BB-02	6/12/2017	15:14	☐	1	F	FLRC	☐	☐	☐	☑	☑	☑	
010	TD2-170612-01	TD2-170612-01-YB-01	6/12/2017	13:38	☐	1	F	FLRC	☐	☐	☐	☑	☑	☑	

001

1-ziploc A

Comments:

Relinquished by:	Received by:	Relinquished by:	Received by:	Relinquished by:	Received by:
Signature	Signature	Signature	Signature	Signature	Signature
Print Name	Print Name Alec Hlen	Print Name	Print Name	Print Name	Print Name
Company Feeler	Company Pace	Company	Company	Company	Company
Date/Time 6/14/17 0950	Date/Time 6/14/17 0950	Date/Time	Date/Time	Date/Time	Date/Time

Date Printed: 6/12/2017 * F= FISH **F=Fillet, WH=Whole Body, O=Offal ***COC TYPE:*** *TPCB-LIPD-PEST-MERCURY* ***Page 1 of 2***

40151601

ANCHOR QEA

305 West Grand Avenue Montvale, NJ 07645 Ph: 201-930-9890

ENVIRONMENTAL SAMPLE CHAIN OF CUSTODY

COC ID: COC170612-T11-03
Sample Custodian: JER
Lab: Pace - Green Bay

Client: General Electric Company
Project: Hudson River Remedial Action Monitoring Program

COC Sample Number	Sampling Event ID	Field Sample ID	Sample Date	Sample Time	MS/LD	# CONTAINERS	Media*	Prep**	Composite	EPA Split	Total PCBs NE013_09	Total PCBs SW846 8082A	Percent Lipid (S-GB-O-068)	Fillet Weight (S-GB-L-009)	
011	TD2-170612-01	TD2-170612-01-SMB-01	6/12/2017	13:18	☐	1	F	FLRC	☐	☐	☐	✓	✓	✓	
012	TD2-170612-01	TD2-170612-01-BB-04	6/12/2017	13:40	☐	1	F	FLRC	☐	☐	☐	✓	✓	✓	
013	TD2-170612-01	TD2-170612-01-BB-03	6/12/2017	13:36	☐	1	F	FLRC	☐	☐	☐	✓	✓	✓	
014	TD2-170612-01	TD2-170612-01-BB-02	6/12/2017	13:32	☐	1	F	FLRC	☐	☐	☐	✓	✓	✓	
015	TD2-170612-01	TD2-170612-01-BB-01	6/12/2017	13:31	☐	1	F	FLRC	☐	☐	☐	✓	✓	✓	
016	TD1-170612-01	TD1-170612-01-YB-01	6/12/2017	12:37	☐	1	F	FLRC	☐	☐	☐	✓	✓	✓	
017	TD1-170612-01	TD1-170612-01-SMB-05	6/12/2017	12:52	☐	1	F	FLRC	☐	☐	☐	✓	✓	✓	
018	TD1-170612-01	TD1-170612-01-SMB-04	6/12/2017	12:50	☐	1	F	FLRC	☐	☐	☐	✓	✓	✓	
019	TD1-170612-01	TD1-170612-01-SMB-01	6/12/2017	12:40	☐	1	F	FLRC	☐	☐	☐	✓	✓	✓	
020	TD1-170612-01	TD1-170612-01-BB-01	6/12/2017	12:33	✓	1	F	FLRC	☐	☐	☐	✓	✓	✓	

002
003
004
005
006

1-Ziploc A

JAO 6/14/17

Comments:

Relinquished by:	Received by:	Relinquished by:	Received by:	Relinquished by:	Received by:
Signature	Signature	Signature	Signature	Signature	Signature
Print Name	Print Name Alec Ft	Print Name	Print Name	Print Name	Print Name
Company Fedex	Company Pace	Company	Company	Company	Company
Date/Time 6/14/17 0950	Date/Time 6/14/17 0950	Date/Time	Date/Time	Date/Time	Date/Time

Date Printed: 6/12/2017 * F= FISH **F=Fillet, WH=Whole Body, O=Offal ***COC TYPE:*** *TPCB-LIPD-PEST-MERCURY* ***Page 2 of 2***

Pace Analytical™

Sample Condition Upon Receipt

Pace Analytical Services, LLC. - Green Bay WI
1241 Bellevue Street, Suite 9
Green Bay, WI 54302

Project #: **WO#: 40151601**

40151601

Client Name: Anchor

Courier: ☑ Fed Ex ☐ UPS ☐ Client ☐ Pace Other:

Tracking #: 779380651289

Custody Seal on Cooler/Box Present: ☑ yes ☐ no Seals intact: ☑ yes ☐ no

Custody Seal on Samples Present: ☐ yes ☑ no Seals intact: ☐ yes ☐ no

Packing Material: ☐ Bubble Wrap ☐ Bubble Bags ☑ None ☐ Other

Thermometer Used 88-70 **Type of Ice:** (Wet) Blue Dry None ☐ Samples on ice, cooling process has begun

Cooler Temperature Uncorr: 5 /Corr: 5 **Biological Tissue is Frozen:** ☐ yes ☑ no

Temp Blank Present: ☑ yes ☐ no

Temp should be above freezing to 6°C.
Biota Samples may be received at ≤ 0°C.

Person examining contents:
Date: 6/14/17
Initials: AR

		Comments:
Chain of Custody Present:	☑Yes ☐No ☐N/A	1.
Chain of Custody Filled Out:	☑Yes ☐No ☐N/A	2.
Chain of Custody Relinquished:	☐Yes ☑No ☐N/A	3.
Sampler Name & Signature on COC:	☐Yes ☑No ☐N/A	4.
Samples Arrived within Hold Time:	☑Yes ☐No ☐N/A	5.
- VOA Samples frozen upon receipt	☐Yes ☐No	Date/Time:
Short Hold Time Analysis (<72hr):	☐Yes ☑No ☐N/A	6.
Rush Turn Around Time Requested:	☐Yes ☑No ☐N/A	7.
Sufficient Volume:	☑Yes ☐No ☐N/A	8.
Correct Containers Used:	☑Yes ☐No ☐N/A	9.
-Pace Containers Used:	☐Yes ☑No ☐N/A	
-Pace IR Containers Used:	☐Yes ☐No ☑N/A	
Containers Intact:	☑Yes ☐No ☐N/A	10.
Filtered volume received for Dissolved tests	☐Yes ☐No ☑N/A	11.
Sample Labels match COC:	☑Yes ☐No ☐N/A	12.
-Includes date/time/ID/Analysis Matrix: B		
All containers needing preservation have been checked. (Non-Compliance noted in 13.)	☐Yes ☐No ☑N/A	13. ☐ HNO3 ☐ H2SO4 ☐ NaOH ☐ NaOH +ZnAct
All containers needing preservation are found to be in compliance with EPA recommendation. (HNO3, H2SO4 ≤2; NaOH+ZnAct ≥9, NaOH ≥12)	☐Yes ☐No ☑N/A	
exceptions: VOA, coliform, TOC, TOX, TOH, O&G, WIDROW, Phenolics, OTHER:	☐Yes ☑No	Initial when completed / Lab Std #ID of preservative / Date/Time:
Headspace in VOA Vials (>6mm):	☐Yes ☐No ☑N/A	14.
Trip Blank Present:	☐Yes ☐No ☑N/A	15.
Trip Blank Custody Seals Present	☐Yes ☐No ☑N/A	
Pace Trip Blank Lot # (if purchased):		

Client Notification/ Resolution: If checked, see attached form for additional comments ☐

Person Contacted: Date/Time:

Comments/ Resolution: All samples have ice melt water in ziplock AR 6/14/17

Project Manager Review: AL for TN **Date:** 6/14/17

F-GB-C-031-Rev.04 (12Dec2016) SCUR.xls
Pace Analytical Services LLC. - Green Bay WI

CASE NARRATIVE - PCB ANALYSIS

Lab Report Number (SDG): 40151601
Client: GENERAL ELECTRIC COMPANY
Project Name: HUDSON RIVER REMEDIAL ACTION MONITORING PROGRAM
Project Number: N/A

1. RECEIPT

Samples were received on ice at 5°C. Following homogenization, tissue samples were kept frozen at less than -10°C until time of preparation for analysis.

2. HOLDING TIMES

A. Sample Preparation: All method and QAPP specified holding times were met.
B. Sample Analysis: All method and QAPP specified holding times were met.

3. METHOD

Preparation: SW-846 3541
Analysis: SW-846 8082A Modified

4. PREPARATION

Sample preparation proceeded normally. All samples went through florisil and acid clean-up procedures prior to analysis.

5. ANALYSIS

A. Calibration:
1. **Initial verification:** All method acceptance criteria were met.
2. **Continuing verification:** All method acceptance criteria were met.

B. Blanks:
1. **Method:** All acceptance criteria were met for the associated method blanks.

C. Surrogates: All QAPP specified acceptance criteria were met.

D. Spikes:
1. **Lab Control Spike (LCS):** The control spikes were fortified with Aroclor 1242 and met the QAPP specified accuracy criteria.
2. **Matrix Spike (MS):** Batch MSs were analyzed with this SDG in order to cover the QAPP QC requirement. The QAPP specified accuracy criteria were met with the following exception. The accuracy criteria were not met for MS 1655663 and MS 1658256; the "M1" data qualifier was applied to the MS on the final report.

E. Samples: Sample analyses proceeded normally. ZB-1ms is the quantitation column.

F. Sample Duplicate: Batch duplicates were analyzed with this SDG in order to cover the QAPP QC requirement. The QAPP specified precision criteria were met. Duplicate results are not evaluated below the reporting limit.

G. Dilutions: None required for this SDG.

H. Reanalysis: None required for this SDG.

I. Comments: Due to rounding differences in the software programs used, the values found on the quantitation reports may not match the values found on the sample Form 1s.

I certify that this data package is in compliance with the terms and conditions agreed to by **Pace Analytical Services, LLC**, and by the client, both technically and for completeness, except for the conditions detailed above. The Laboratory Manager or his designee, as verified by the following signature, has authorized release of the data contained in this completed data package:

Signed: Leigh A Begalske Date: 05/08/18

Name: Leigh A. Begalske Position: Quality Assurance Auditor

PCB - FORM II SVOA-1
TISSUE SEMI-VOLATILE SURROGATE RECOVERY

Lab Name: Pace Analytical - Green Bay SDG No.: 40151601 Contract: HUDSON RIVER REMEDIAL

Instrument ID: 40GCSJ

LAB SAMPLE ID	SAMPLE NAME	DCBP	TCMX
1655661	1655661BLANK	83	73
1655662	1655662LCS	82	69
40151601002	TD2-170612-01-SMB-01	87	84

	QC LIMITS
(DCBP) = Decachlorobiphenyl (S)	(60-140)
(TCMX) = Tetrachloro-m-xylene (S)	(60-140)

* Values outside of QC Limits

PCB - FORM II SVOA-2
TISSUE SEMI-VOLATILE SURROGATE RECOVERY

Lab Name: Pace Analytical - Green Bay SDG No.: 40151601 Contract: HUDSON RIVER REMEDIAL

Instrument ID: 40GCSJ

LAB SAMPLE ID	SAMPLE NAME	DCBP	TCMX
1658254	1658254BLANK	83	76
1658255	1658255LCS	92	85
40151601003	TD2-170612-01-BB-04	100	99
40151601004	TD2-170612-01-BB-03	83	77
40151601005	TD2-170612-01-BB-02	95	95
40151601006	TD2-170612-01-BB-01	95	97

	QC LIMITS
(DCBP) = Decachlorobiphenyl (S)	(60-140)
(TCMX) = Tetrachloro-m-xylene (S)	(60-140)

* Values outside of QC Limits

PCB - FORM II SVOA-3
TISSUE SEMI-VOLATILE SURROGATE RECOVERY

Lab Name: Pace Analytical - Green Bay SDG No.: 40151601 Contract: HUDSON RIVER REMEDIAL

Instrument ID: 40GCSJ

LAB SAMPLE ID	SAMPLE NAME	DCBP	TCMX
1664793	1664793BLANK	89	76
1664794	1664794LCS	92	81
40151601001	TD2-170612-01-YB-01	90	89

	QC LIMITS
(DCBP) = Decachlorobiphenyl (S)	(60-140)
(TCMX) = Tetrachloro-m-xylene (S)	(60-140)

* Values outside of QC Limits

PCB - FORM III SVOA-1

TISSUE LABORATORY CONTROL SAMPLE RECOVERY

Lab Name: Pace Analytical - Green Bay

Date Extracted: 03/12/2018

Instrument: 40GCSJ

Lab File ID: 031318.B\031318007.D

Lab Sample ID: 1655662LCS

Date Analyzed (1): 03/13/2018

LCS Lot No: 182336

SDG No.: 40151601

COMPOUND	AMOUNT ADDED (ug/kg)	LCS CONCENTRATION (ug/kg)	LCS %REC	QC LIMITS REC.
PCB-1242 (Aroclor 1242)	250	212	85	70-130

Spike Recovery: 0 out of 1 outside limits.

PCB - FORM III SVOA-1

TISSUE LABORATORY CONTROL SAMPLE RECOVERY

Lab Name: Pace Analytical - Green Bay

Date Extracted: 03/13/2018

Instrument: 40GCSJ

Lab File ID: 031418.B\031418007.D

Lab Sample ID: 1658255LCS

Date Analyzed (1): 03/14/2018

LCS Lot No: 182336

SDG No.: 40151601

COMPOUND	AMOUNT ADDED (ug/kg)	LCS CONCENTRATION (ug/kg)	LCS %REC	QC LIMITS REC.
PCB-1242 (Aroclor 1242)	250	232	93	70-130

Spike Recovery: 0 out of 1 outside limits.

PCB - FORM III SVOA-1

TISSUE LABORATORY CONTROL SAMPLE RECOVERY

Lab Name: Pace Analytical - Green Bay

Date Extracted: 03/28/2018

Instrument: 40GCSJ

Lab File ID: 033018.B\033018011.D

Lab Sample ID: 1664794LCS

Date Analyzed (1): 03/30/2018

LCS Lot No: 182336

SDG No.: 40151601

COMPOUND	AMOUNT ADDED (ug/kg)	LCS CONCENTRATION (ug/kg)	LCS %REC	QC LIMITS REC.
PCB-1242 (Aroclor 1242)	250	236	95	70-130

Spike Recovery: 0 out of 1 outside limits.

PCB - FORM III SVOA-1

TISSUE SEMI-VOLATILE MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Pace Analytical - Green Bay

Date Extracted: 03/12/2018

Instrument: 40GCSJ

Parent Sample ID: 40150295019

Matrix Spike - Sample No: 1655663MS

Date Analyzed (1): 03/13/2018

Lab File ID: 031318.B\031318008.D

SDG No.: 40151601

COMPOUND	SPIKE ADDED (ug/kg)	SAMPLE CONCENTRATION (ug/kg)	MS CONCENTRATION (ug/kg)	MS %REC	QC LIMITS REC.
PCB-1242 (Aroclor 1242)	250	<56.9	519	208	70-130

Spike Recovery: 1 out of 1 outside limits.

PCB - FORM III SVOA-1

TISSUE SEMI-VOLATILE MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Pace Analytical - Green Bay

Date Extracted: 03/13/2018

Instrument: 40GCSJ

Parent Sample ID: 40151552010

Matrix Spike - Sample No: 1658256MS

Date Analyzed (1): 03/14/2018

Lab File ID: 031418.B\031418008.D

SDG No.: 40151601

COMPOUND	SPIKE ADDED (ug/kg)	SAMPLE CONCENTRATION (ug/kg)	MS CONCENTRATION (ug/kg)	MS %REC	QC LIMITS REC.
PCB-1242 (Aroclor 1242)	250	<34.3	354	142	70-130

Spike Recovery: 1 out of 1 outside limits.

PCB - FORM III SVOA-1

TISSUE SEMI-VOLATILE MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Pace Analytical - Green Bay

Date Extracted: 03/28/2018

Instrument: 40GCSJ

Parent Sample ID: 40151613001

Matrix Spike - Sample No: 1664795MS

Date Analyzed (1): 03/30/2018

Lab File ID: 033018.B\033018012.D

SDG No.: 40151601

COMPOUND	SPIKE ADDED (ug/kg)	SAMPLE CONCENTRATION (ug/kg)	MS CONCENTRATION (ug/kg)	MS %REC	QC LIMITS REC.
PCB-1242 (Aroclor 1242)	251	<11.4	238	95	70-130

Spike Recovery: 0 out of 1 outside limits.

PCB - FORM III SVOA-1

TISSUE SEMI-VOLATILE SAMPLE/DUPLICATE RECOVERY

Lab Name: Pace Analytical - Green Bay

Date Extracted: 03/12/2018

Instrument 40GCSJ

Lab Sample ID: 40150295019

Duplicate Sample No: 40150295019DUP

Date Analyzed: 03/13/2018

Lab File ID: 031318.B\031318010.D

SDG No.: 40151601

COMPOUND	SAMPLE CONCENTRATION (ug/kg)	DUPLICATE CONCENTRATION (ug/kg)	RPD	RPD LIMITS
PCB-1016 (Aroclor 1016)	<56.9	<56.9		0-40
PCB-1221 (Aroclor 1221)	<56.9	<56.9		0-40
PCB-1232 (Aroclor 1232)	<56.9	<56.9		0-40
PCB-1242 (Aroclor 1242)	<56.9	<56.9		0-40
PCB-1248 (Aroclor 1248)	1040	786	28	0-40
PCB-1254 (Aroclor 1254)	1000	746	29	0-40
PCB-1260 (Aroclor 1260)	510	375	31	0-40

RPD: 0 out of 3 outside limits.

PCB - FORM III SVOA-2

TISSUE SEMI-VOLATILE SAMPLE/DUPLICATE RECOVERY

Lab Name: Pace Analytical - Green Bay

Date Extracted: 03/12/2018

Instrument 40GCSJ

Lab Sample ID: 40151804007

Duplicate Sample No: 40151804007DUP

Date Analyzed: 03/13/2018

Lab File ID: 031318.B\031318031.D

SDG No.: 40151601

COMPOUND	SAMPLE CONCENTRATION (ug/kg)	DUPLICATE CONCENTRATION (ug/kg)	RPD	RPD LIMITS
PCB-1016 (Aroclor 1016)	<170	<171		0-40
PCB-1221 (Aroclor 1221)	<170	<171		0-40
PCB-1232 (Aroclor 1232)	<170	<171		0-40
PCB-1242 (Aroclor 1242)	<170	<171		0-40
PCB-1248 (Aroclor 1248)	2530	2760	9	0-40
PCB-1254 (Aroclor 1254)	2420	2660	9	0-40
PCB-1260 (Aroclor 1260)	693J	792		0-40

RPD: 0 out of 2 outside limits.

PCB - FORM III SVOA-1

TISSUE SEMI-VOLATILE SAMPLE/DUPLICATE RECOVERY

Lab Name: Pace Analytical - Green Bay

Date Extracted: 03/13/2018

Instrument 40GCSJ

Lab Sample ID: 40151552010

Duplicate Sample No: 40151552010DUP

Date Analyzed: 03/14/2018

Lab File ID: 031418.B\031418020.D

SDG No.: 40151601

COMPOUND	SAMPLE CONCENTRATION (ug/kg)	DUPLICATE CONCENTRATION (ug/kg)	RPD	RPD LIMITS
PCB-1016 (Aroclor 1016)	<34.3	<34.2		0-40
PCB-1221 (Aroclor 1221)	<34.3	<34.2		0-40
PCB-1232 (Aroclor 1232)	<34.3	<34.2		0-40
PCB-1242 (Aroclor 1242)	<34.3	<34.2		0-40
PCB-1248 (Aroclor 1248)	795	853	7	0-40
PCB-1254 (Aroclor 1254)	589	628	6	0-40
PCB-1260 (Aroclor 1260)	284	299	5	0-40

RPD: 0 out of 3 outside limits.

PCB - FORM III SVOA-1

TISSUE SEMI-VOLATILE SAMPLE/DUPLICATE RECOVERY

Lab Name: Pace Analytical - Green Bay

Date Extracted: 03/28/2018

Instrument 40GCSJ

Lab Sample ID: 40151613001

Duplicate Sample No: 40151613001DUP

Date Analyzed: 03/30/2018

Lab File ID: 033018.B\033018018.D

SDG No.: 40151601

COMPOUND	SAMPLE CONCENTRATION (ug/kg)	DUPLICATE CONCENTRATION (ug/kg)	RPD	RPD LIMITS
PCB-1016 (Aroclor 1016)	<11.4	<11.4		0-40
PCB-1221 (Aroclor 1221)	<11.4	<11.4		0-40
PCB-1232 (Aroclor 1232)	<11.4	<11.4		0-40
PCB-1242 (Aroclor 1242)	<11.4	<11.4		0-40
PCB-1248 (Aroclor 1248)	11.8J	11.4J		0-40
PCB-1254 (Aroclor 1254)	16.9J	16.6J		0-40
PCB-1260 (Aroclor 1260)	39.1J	36.6J		0-40

RPD: ____ out of __0__ outside limits.

PCB - FORM III SVOA-2

TISSUE SEMI-VOLATILE SAMPLE/DUPLICATE RECOVERY

Lab Name: Pace Analytical - Green Bay

Date Extracted: 03/28/2018

Instrument 40GCSJ

Lab Sample ID: 40157825001

Duplicate Sample No: 40157825001DUP

Date Analyzed: 03/30/2018

Lab File ID: 033018.B\033018034.D

SDG No.: 40151601

COMPOUND	SAMPLE CONCENTRATION (ug/kg)	DUPLICATE CONCENTRATION (ug/kg)	RPD	RPD LIMITS
PCB-1016 (Aroclor 1016)	<11.4	<11.4		0-40
PCB-1221 (Aroclor 1221)	<11.4	<11.4		0-40
PCB-1232 (Aroclor 1232)	<11.4	<11.4		0-40
PCB-1242 (Aroclor 1242)	<11.4	<11.4		0-40
PCB-1248 (Aroclor 1248)	272	344	23	0-40
PCB-1254 (Aroclor 1254)	114	147	25	0-40
PCB-1260 (Aroclor 1260)	20.1J	26.4J		0-40

RPD: 0 out of 2 outside limits.

SAMPLE NO.

1655661BLANK

PCB - FORM IV SVOA-1
SEMI-VOLATILE METHOD BLANK SUMMARY

Lab Name: Pace Analytical - Green Bay SDG No.: 40151601 Contract: HUDSON RIVER REMEDIAL

Instrument ID: 40GCSJ Matrix: Tissue Lab Sample ID: 1655661

Lab File ID: 031318.B\031318006.D Date Analyzed: 03/13/2018 Time:14:50

SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	ANALYZED
1655662LCS	1655662	031318.B\031318007.D	03/13/2018 15:05
1655664DUP	1655664	031318.B\031318010.D	03/13/2018 15:53
TD2-170612-01-SMB-01	40151601002	031318.B\031318014.D	03/13/2018 16:56
1655665DUP	1655665	031318.B\031318031.D	03/13/2018 21:25

SAMPLE NO.

1658254BLANK

PCB - FORM IV SVOA-1
SEMI-VOLATILE METHOD BLANK SUMMARY

Lab Name: Pace Analytical - Green Bay SDG No.: 40151601 Contract: HUDSON RIVER REMEDIAL

Instrument ID: 40GCSJ Matrix: Tissue Lab Sample ID: 1658254

Lab File ID: 031418.B\031418006.D Date Analyzed: 03/14/2018 Time:13:44

SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	ANALYZED
1658255LCS	1658255	031418.B\031418007.D	03/14/2018 14:00
1658257DUP	1658257	031418.B\031418020.D	03/14/2018 17:26
TD2-170612-01-BB-04	40151601003	031418.B\031418021.D	03/14/2018 17:41
TD2-170612-01-BB-03	40151601004	031418.B\031418022.D	03/14/2018 17:57
TD2-170612-01-BB-02	40151601005	031418.B\031418023.D	03/14/2018 18:13
TD2-170612-01-BB-01	40151601006	031418.B\031418024.D	03/14/2018 18:29

SAMPLE NO.

1664793BLANK

PCB - FORM IV SVOA-1
SEMI-VOLATILE METHOD BLANK SUMMARY

Lab Name: Pace Analytical - Green Bay SDG No.: 40151601 Contract: HUDSON RIVER REMEDIAL

Instrument ID: 40GCSJ Matrix: Tissue Lab Sample ID: 1664793

Lab File ID: 033018.B\033018010.D Date Analyzed: 03/30/2018 Time:17:22

SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	ANALYZED
1664794LCS	1664794	033018.B\033018011.D	03/30/2018 17:38
TD2-170612-01-YB-01	40151601001	033018.B\033018014.D	03/30/2018 18:25
1664796DUP	1664796	033018.B\033018018.D	03/30/2018 19:29
1664809DUP	1664809	033018.B\033018034.D	03/30/2018 23:43

SAMPLE NO.

TD2-170612-01-YB-01

PCB - FORM I SVOA-1

SEMI-VOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay
Contract: HUDSON RIVER REMEDIAL ACTION
Date Received: 06/14/2017 09:50
Matrix: Tissue SDG No.: 40151601
Date Extracted: 03/28/2018 08:09
Lab Sample ID: 40151601001
Date Analyzed: 03/30/2018 18:25
Lab File ID: 033018.B\033018014.D
Initial wt/vol: 10.0134 Final wt/vol: 5 mL Dilution: 1
Instrument: 40GCSJ Percent Moisture:

CAS NO.	COMPOUND	CONCENTRATION UNITS: ug/kg	Q
12674-11-2	PCB-1016 (Aroclor 1016)	<11.4	U
11104-28-2	PCB-1221 (Aroclor 1221)	<11.4	U
11141-16-5	PCB-1232 (Aroclor 1232)	<11.4	U
53469-21-9	PCB-1242 (Aroclor 1242)	<11.4	U
12672-29-6	PCB-1248 (Aroclor 1248)	80.1	
11097-69-1	PCB-1254 (Aroclor 1254)	40.2	J
11096-82-5	PCB-1260 (Aroclor 1260)	15.5	J
1336-36-3	PCB, Total	136	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018014.D
Lab Smp Id: 40151601001 Client Smp ID: TD2-170612-01-YB-01
Inj Date : 30-MAR-2018 18:25
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 40151601001,
Misc Info : 17591,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\033018.b\pcba030618.m
Meth Date : 05-Apr-2018 14:48 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 14
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.013	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.186	28545	20373	0.714	0.07		
1.426	11670	9755	0.836	0.03		
2.014	18310	13648	0.745	0.05		
2.196	18881	18636	0.987	0.05		
2.410	20334	14580	0.717	0.05	1 Aroclor 1221	
2.485	11024210	10660629	0.967	30.94	$ 2 Tetrachloro-m-xylene	
2.789	20251	18738	0.925	0.05	1 Aroclor 1221	
2.852	34291	21974	0.641	0.09	1 Aroclor 1221	
2.991	18696	18523	0.991	0.05		
3.022	25291	21997	0.870	0.06		
3.055	27863	23309	0.837	0.07		
3.318	26068	24478	0.939	0.07		
3.345	18887	16962	0.898	0.05		
3.395	10779	8394	0.779	0.02		
3.433	14845	12794	0.862	0.04		
3.519	71782	63720	0.888	0.19		
3.648	16353	13998	0.856	0.04		
3.731	144768	120577	0.833	0.40		
3.826	634670	372575	0.587	1.75		
3.939	38172	32435	0.850	0.10		
3.993	47052	43864	0.932	0.13		
4.056	36916	26809	0.726	0.10		
4.215	838936	768172	0.916	2.32		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.261	611777	586209	0.958	1.69	6 Aroclor 1248
4.292	1060040	929337	0.877	2.93	
4.417	128459	111673	0.869	0.35	
4.454	105461	81904	0.777	0.29	
4.546	416593	372760	0.895	1.15	6 Aroclor 1248
4.588	91092	57294	0.629	0.25	
4.681	44207	30902	0.699	0.12	
4.737	56044	39592	0.706	0.15	
4.770	126597	108896	0.860	0.35	
4.813	439289	393557	0.896	1.21	6 Aroclor 1248
4.849	164712	161004	0.977	0.45	
4.879	826783	551620	0.667	2.28	6 Aroclor 1248
4.958	97126	69456	0.715	0.26	
5.037	214486	176860	0.825	0.59	6 Aroclor 1248
5.103	284503	241515	0.849	0.78	
5.161	501919	424701	0.846	1.38	7 Aroclor 1254
5.212	560002	473456	0.845	1.54	
5.283	73445	61854	0.842	0.20	
5.308	78519	57271	0.729	0.21	
5.363	99200	59210	0.597	0.27	
5.406	263485	221215	0.840	0.72	7 Aroclor 1254
5.447	327646	261204	0.797	0.90	
5.511	407085	310724	0.763	1.12	7 Aroclor 1254
5.576	87667	39796	0.454	0.24	
5.674	162927	147803	0.907	0.45	
5.711	106035	77553	0.731	0.29	
5.741	162075	112010	0.691	0.44	
5.791	836978	630078	0.753	2.31	
5.888	75744	46309	0.611	0.20	
5.932	88563	64314	0.726	0.24	
5.993	198581	182538	0.919	0.54	
6.024	311281	274658	0.882	0.86	7 Aroclor 1254
6.055	656116	547482	0.834	1.81	
6.121	22962	14634	0.637	0.06	
6.168	137007	83167	0.607	0.37	
6.226	110654	60336	0.545	0.30	
6.300	841406	664189	0.789	2.32	7 Aroclor 1254
6.333	92047	79573	0.864	0.25	
6.375	37057	22186	0.599	0.10	
6.434	108799	70843	0.651	0.30	
6.509	247653	213160	0.861	0.68	
6.535	164166	145025	0.883	0.45	
6.560	84325	66523	0.789	0.23	
6.604	29091	23312	0.801	0.08	
6.664	18691	16097	0.861	0.05	
6.717	67415	51924	0.770	0.18	
6.757	36952	29769	0.806	0.10	
6.800	115524	95614	0.828	0.31	
6.838	59819	37941	0.634	0.16	
6.941	63225	55163	0.872	0.17	
7.004	267241	245771	0.920	0.73	8 Aroclor 1260
7.035	35712	32792	0.918	0.09	
7.075	16853	12102	0.718	0.04	
7.126	22816	16719	0.733	0.06	
7.250	126923	106470	0.839	0.35	8 Aroclor 1260
7.279	63278	44829	0.708	0.17	
7.390	81322	79272	0.975	0.22	

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
7.441	112339	94154	0.838	0.31	8 Aroclor 1260
7.578	16240	14053	0.865	0.04	
7.691	26069	27642	1.060	0.07	
7.728	53763	42778	0.796	0.14	
7.770	9803	10457	1.067	0.02	
7.799	45424	24132	0.531	0.12	
7.898	86462	70524	0.816	0.23	8 Aroclor 1260
7.966	19217	12213	0.636	0.05	
8.050	25474	21184	0.832	0.07	
8.273	133327	109598	0.822	0.36	
8.380	22859	17104	0.748	0.06	
8.569	9855938	8189922	0.831	27.26	$ 12 Decachlorobiphenyl
8.943	25589	14824	0.579	0.07	
9.165	14093	11139	0.790	0.03	
9.464	16063	14521	0.904	0.04	
9.961	27566	18911	0.686	0.07	
	36153171	31212262		100.000	

Total unknown % area = 27

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018014.D
Lab Smp Id: 40151601001 Client Smp ID: TD2-170612-01-YB-01
Inj Date : 30-MAR-2018 18:25
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 40151601001,
Misc Info : 17591,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\033018.b\pcba030618.m
Meth Date : 05-Apr-2018 14:48 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 14
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.013	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE

1 Aroclor 1221 CAS #: 11104-28-2

Compound Not Detected.

$ 2 Tetrachloro-m-xylene CAS #: 877-09-8

RT	EXP RT	DLT RT	RESPONSE	ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
2.484	2.484	0.000	11024209	0.08941	45			

---

3 Aroclor 1242 CAS #: 53469-21-9

Operator disabled compound identification.

---

4 Aroclor 1232 CAS #: 1114-16-5

Operator disabled compound identification.

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
5 Aroclor 1016					CAS #: 12674-11-2			
Operator disabled compound identification.								
6 Aroclor 1248					CAS #: 12672-29-6			
4.260	4.261	-0.001	611776	0.26111	130	0.00- 20.00	100.00(M)	BA
4.545	4.546	-0.001	416592	0.10038	50	0.00- 20.00	68.10	BA
4.812	4.812	0.000	439289	0.21639	110	0.00- 20.00	71.81	BA
4.878	4.879	-0.001	826783	0.16413	82	0.00- 20.00	135.14	BA
5.037	5.038	-0.001	214486	0.05977	30	0.00- 20.00	35.06	BA
Average of Peak Concentrations =					80			
7 Aroclor 1254					CAS #: 11097-69-1			
5.161	5.162	-0.001	501919	0.09587	48	0.00- 20.00	100.00(M)	CO
5.405	5.408	-0.003	263485	0.06126	30	0.00- 20.00	52.50	CO
5.510	5.510	0.000	407085	0.05242	26	0.00- 20.00	81.11	CO
6.023	6.027	-0.004	311280	0.06392	32	0.00- 20.00	62.02	CO
6.299	6.299	0.000	841406	0.12931	64	0.00- 20.00	167.64	CO
Average of Peak Concentrations =					40			
8 Aroclor 1260					CAS #: 11096-82-5			
6.300	6.300	0.000	0	0.00	0.00	0.00- 20.00	0.00(M)	CO
7.003	7.006	-0.003	267241	0.02669	13	0.00- 20.00	28.66	CO
7.249	7.251	-0.002	126922	0.02655	13	0.00- 20.00	13.61	CO
7.440	7.443	-0.003	112338	0.03854	19	0.00- 20.00	12.05	CO
7.897	7.901	-0.004	86462	0.03277	16	0.00- 20.00	9.27	CO
Average of Peak Concentrations =					16			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.569	8.571	-0.002	9855937	0.08996	45			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.
CO: Indicates that the analyst had to split two co-eluting peaks apart that were not (or could not be) separated by the computer system.

Data File: \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018014.D
Date : 30-MAR-2018 18:25
Client ID: TD2-170612-01-YB-01
Sample Info: 40151601001,
Volume Injected (uL): 1.0
Column phase: ZB-1ms

Instrument: 40GCSJ.i
Operator: BLM
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\033018.b/033018014.D (ManInt: BIOTA.sub Orig: BIOTA.sub)
Injection Date: 30-MAR-2018 18:25 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 40151601001





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	26068	0	
Aroclor 1016	26068	0	
Aroclor 1016	26068	0	
Aroclor 1016	26068	0	
Aroclor 1016	26068	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1232	34290	0	
Aroclor 1232	34290	0	
Aroclor 1232	34290	0	
Aroclor 1232	34290	0	

Aroclor 1232	34290	0	
Aroclor 1242	26068	0	
Aroclor 1242	26068	0	
Aroclor 1242	26068	0	
Aroclor 1242	26068	0	
Aroclor 1242	26068	0	
Aroclor 1248	598571	611776	BA
Aroclor 1248	598571	611776	BA
Aroclor 1248	598571	611776	BA
Aroclor 1248	598571	611776	BA
Aroclor 1248	598571	611776	BA
Aroclor 1254	466673	501919	CO
Aroclor 1254	466673	501919	CO
Aroclor 1254	466673	501919	CO
Aroclor 1254	466673	501919	CO
Aroclor 1254	466673	501919	CO
Aroclor 1260	910603	0	CO
Aroclor 1260	910603	0	CO
Aroclor 1260	910603	0	CO
Aroclor 1260	910603	0	CO
Aroclor 1260	910603	0	CO
Tetrachloro-m-xylene	11024209	11024209	
Decachlorobiphenyl	9855937	9855937	

==========

SAMPLE NO.

TD2-170612-01-SMB-01

PCB - FORM I SVOA-1

SEMI-VOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay
Contract: HUDSON RIVER REMEDIAL ACTION
Date Received: 06/14/2017 09:50
Matrix: Tissue SDG No.: 40151601
Date Extracted: 03/12/2018 07:42
Lab Sample ID: 40151601002
Date Analyzed: 03/13/2018 16:56
Lab File ID: 031318.B\031318014.D
Initial wt/vol: 9.9928 g Final wt/vol: 5 mL Dilution: 1
Instrument: 40GCSJ Percent Moisture:

CAS NO.	COMPOUND	CONCENTRATION UNITS: ug/kg	Q
12674-11-2	PCB-1016 (Aroclor 1016)	<11.4	U
11104-28-2	PCB-1221 (Aroclor 1221)	<11.4	U
11141-16-5	PCB-1232 (Aroclor 1232)	<11.4	U
53469-21-9	PCB-1242 (Aroclor 1242)	<11.4	U
12672-29-6	PCB-1248 (Aroclor 1248)	110	
11097-69-1	PCB-1254 (Aroclor 1254)	79.4	
11096-82-5	PCB-1260 (Aroclor 1260)	56.1	
1336-36-3	PCB, Total	245	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318014.D
Lab Smp Id: 40151601002 Client Smp ID: TD2-170612-01-SMB-0
Inj Date : 13-MAR-2018 16:56
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 40151601002,
Misc Info : 17521,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031318.b\pcba030618.m
Meth Date : 27-Mar-2018 10:38 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 14
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	9.993	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.027	15599	11791	0.756	0.03		
1.538	14524	9920	0.683	0.03		
2.409	19356	14291	0.738	0.04		
2.485	10403219	10022861	0.963	23.10	$ 2 Tetrachloro-m-xylene	
2.786	28603	19091	0.667	0.06		
2.856	25674	16034	0.625	0.05		
2.972	31078	15274	0.491	0.06		
3.021	17103	18201	1.064	0.03		
3.056	37588	28628	0.762	0.08		
3.317	46226	43252	0.936	0.10		
3.344	38926	38357	0.985	0.08		
3.403	12773	12029	0.942	0.02		
3.435	34183	30754	0.900	0.07		
3.519	72227	61979	0.858	0.15		
3.649	11284	8521	0.755	0.02		
3.691	10749	9793	0.911	0.02		
3.731	48988	50591	1.033	0.10		
3.755	33456	31564	0.943	0.07		
3.826	473453	307628	0.650	1.02		
3.931	174759	121307	0.694	0.37		
3.993	158509	148368	0.936	0.34		
4.058	73025	57000	0.781	0.15		
4.138	25497	18646	0.731	0.05		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.214	495672	470013	0.948	1.07	
4.240	93318	124513	1.334	0.20	
4.260	502614	465660	0.926	1.08	6 Aroclor 1248
4.292	546728	377128	0.690	1.18	
4.417	407097	378258	0.929	0.88	
4.451	317894	277025	0.871	0.68	
4.545	647824	611265	0.944	1.40	6 Aroclor 1248
4.591	37350	36064	0.966	0.08	
4.616	141184	109473	0.775	0.30	
4.652	25364	20787	0.820	0.05	
4.679	47800	35607	0.745	0.10	
4.729	81838	51227	0.626	0.17	
4.768	150133	128785	0.858	0.32	
4.812	547568	528813	0.966	1.18	6 Aroclor 1248
4.847	513659	507516	0.988	1.11	
4.878	1244651	823655	0.662	2.69	6 Aroclor 1248
4.957	209594	166483	0.794	0.45	
5.036	742983	684350	0.921	1.61	6 Aroclor 1248
5.100	395092	312725	0.792	0.85	
5.161	782359	699049	0.894	1.69	7 Aroclor 1254
5.211	648035	575533	0.888	1.40	
5.283	77801	62640	0.805	0.16	
5.308	127121	108562	0.854	0.27	
5.362	380683	293147	0.770	0.82	
5.405	597015	508278	0.851	1.29	7 Aroclor 1254
5.446	525544	434573	0.827	1.13	
5.490	71263	66794	0.937	0.15	
5.508	953850	856258	0.898	2.06	7 Aroclor 1254
5.578	47535	42338	0.891	0.10	
5.603	210591	178851	0.849	0.45	
5.674	227455	209220	0.920	0.49	
5.710	162820	120438	0.740	0.35	
5.741	246021	175008	0.711	0.53	
5.789	1435106	1125826	0.784	3.11	
5.838	41261	37509	0.909	0.08	
5.886	160835	116357	0.723	0.34	
5.928	135148	93439	0.691	0.29	
5.992	281986	260182	0.923	0.61	
6.024	790505	687042	0.869	1.71	7 Aroclor 1254
6.055	1012998	895467	0.884	2.19	
6.167	306097	227397	0.743	0.66	
6.227	239231	128330	0.536	0.51	
6.299	1434257	1216536	0.848	3.10	7 Aroclor 1254
6.332	199281	177926	0.893	0.43	
6.373	103439	81371	0.787	0.22	
6.431	224539	181934	0.810	0.48	
6.509	656769	602008	0.917	1.42	
6.535	292194	268475	0.919	0.63	
6.561	207289	184990	0.892	0.44	
6.604	70266	55897	0.796	0.15	
6.660	61396	46405	0.756	0.13	
6.717	226086	198804	0.879	0.49	
6.756	228204	192521	0.844	0.49	
6.800	304241	270596	0.889	0.65	
6.838	160741	122649	0.763	0.34	
6.911	59644	50395	0.845	0.12	
6.941	168456	152549	0.906	0.36	

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
7.004	774668	693370	0.895	1.67	8 Aroclor 1260
7.033	73426	68834	0.937	0.15	
7.077	23569	19135	0.812	0.05	
7.123	27843	21386	0.768	0.06	
7.250	347459	302573	0.871	0.75	8 Aroclor 1260
7.282	131311	108803	0.829	0.28	
7.337	12857	11422	0.888	0.02	
7.388	464605	393832	0.848	1.00	
7.442	475034	423156	0.891	1.02	8 Aroclor 1260
7.550	20733	20071	0.968	0.04	
7.619	12395	8296	0.669	0.02	
7.691	127266	115347	0.906	0.27	
7.726	131486	111227	0.846	0.28	
7.800	90254	57876	0.641	0.19	
7.899	357822	323587	0.904	0.77	8 Aroclor 1260
7.964	41353	26556	0.642	0.08	
8.049	64993	56375	0.867	0.14	
8.274	427741	390531	0.913	0.92	
8.373	15595	12924	0.829	0.03	
8.570	9539461	8347877	0.875	20.68	$ 12 Decachlorobiphenyl
8.724	15682	14608	0.932	0.03	
8.943	19747	12364	0.626	0.04	
9.052	16280	13105	0.805	0.03	
9.168	48970	43267	0.884	0.10	
9.469	24122	15167	0.629	0.05	
9.960	27151	17435	0.642	0.05	
	46125048	40539645		100.000	

Total unknown % area = 34

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318014.D
Lab Smp Id: 40151601002 Client Smp ID: TD2-170612-01-SMB-0
Inj Date : 13-MAR-2018 16:56
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 40151601002,
Misc Info : 17521,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031318.b\pcba030618.m
Meth Date : 27-Mar-2018 10:38 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 14
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	9.993	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221					CAS #: 11104-28-2			
Operator disabled compound identification.								
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.484	2.484	0.000	10403219	0.08438	42			
3 Aroclor 1242					CAS #: 53469-21-9			
Operator disabled compound identification.								
4 Aroclor 1232					CAS #: 1114-16-5			
Operator disabled compound identification.								

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
5 Aroclor 1016					CAS #: 12674-11-2			
Operator disabled compound identification.								
6 Aroclor 1248					CAS #: 12672-29-6			
4.260	4.261	-0.001	502614	0.21452	110	0.00- 20.00	100.00(M)	BA
4.545	4.546	-0.001	647824	0.15609	78	0.00- 20.00	128.89	BA
4.812	4.812	0.000	547568	0.26973	130	0.00- 20.00	108.94	BA
4.878	4.879	-0.001	1244650	0.24708	120	0.00- 20.00	247.64	BA
5.036	5.038	-0.002	742982	0.20704	100	0.00- 20.00	147.82	BA
Average of Peak Concentrations =					110			
7 Aroclor 1254					CAS #: 11097-69-1			
5.161	5.162	-0.001	782359	0.14944	75	0.00- 20.00	100.00(M)	BA
5.405	5.408	-0.003	597015	0.13881	69	0.00- 20.00	76.31	BA
5.508	5.510	-0.002	953849	0.12282	61	0.00- 20.00	121.92	BA
6.024	6.027	-0.003	790504	0.16233	81	0.00- 20.00	101.04	BA
6.298	6.299	-0.001	1434257	0.22041	110	0.00- 20.00	183.32	BA
Average of Peak Concentrations =					79			
8 Aroclor 1260					CAS #: 11096-82-5			
6.300	6.300	0.000	0	0.00	0.00	0.00- 20.00	0.00(M)	BA
7.003	7.006	-0.003	774668	0.07736	39	0.00- 20.00	54.01	BA
7.249	7.251	-0.002	347459	0.07269	36	0.00- 20.00	24.23	BA
7.441	7.443	-0.002	475034	0.16296	82	0.00- 20.00	33.12	BA
7.899	7.901	-0.002	357822	0.13560	68	0.00- 20.00	24.95	BA
Average of Peak Concentrations =					56			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.570	8.571	-0.001	9539461	0.08707	44			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318014.D

Date : 13-MAR-2018 16:56

Client ID: TD2-170612-01-SMB-0

Instrument: 40GCSJ.i

Sample Info: 40151601002,

Volume Injected (uL): 1.0

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031318.b/031318014.D (ManInt: BIOTA.sub Orig: BIOTA.sub)
Injection Date: 13-MAR-2018 16:56 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 40151601002





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	46225	0	
Aroclor 1016	46225	0	
Aroclor 1016	46225	0	
Aroclor 1016	46225	0	
Aroclor 1016	46225	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1232	25673	0	
Aroclor 1232	25673	0	
Aroclor 1232	25673	0	
Aroclor 1232	25673	0	

Aroclor 1232	25673	0	
Aroclor 1242	46225	0	
Aroclor 1242	46225	0	
Aroclor 1242	46225	0	
Aroclor 1242	46225	0	
Aroclor 1242	46225	0	
Aroclor 1248	493934	502614	BA
Aroclor 1248	493934	502614	BA
Aroclor 1248	493934	502614	BA
Aroclor 1248	493934	502614	BA
Aroclor 1248	493934	502614	BA
Aroclor 1254	752542	782359	BA
Aroclor 1254	752542	782359	BA
Aroclor 1254	752542	782359	BA
Aroclor 1254	752542	782359	BA
Aroclor 1254	752542	782359	BA
Aroclor 1260	1402795	0	BA
Aroclor 1260	1402795	0	BA
Aroclor 1260	1402795	0	BA
Aroclor 1260	1402795	0	BA
Aroclor 1260	1402795	0	BA
Tetrachloro-m-xylene	10403219	10403219	
Decachlorobiphenyl	9539461	9539461	

==========

SAMPLE NO.

TD2-170612-01-BB-04

PCB - FORM I SVOA-1

SEMI-VOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay
Contract: HUDSON RIVER REMEDIAL ACTION
Date Received: 06/14/2017 09:50
Matrix: Tissue SDG No.: 40151601
Date Extracted: 03/13/2018 07:38
Lab Sample ID: 40151601003
Date Analyzed: 03/14/2018 17:41
Lab File ID: 031418.B\031418021.D
Initial wt/vol: 10.0319 Final wt/vol: 5 mL Dilution: 1
Instrument: 40GCSJ Percent Moisture:

CAS NO.	COMPOUND	CONCENTRATION UNITS: ug/kg	Q
12674-11-2	PCB-1016 (Aroclor 1016)	<11.4	U
11104-28-2	PCB-1221 (Aroclor 1221)	<11.4	U
11141-16-5	PCB-1232 (Aroclor 1232)	<11.4	U
53469-21-9	PCB-1242 (Aroclor 1242)	<11.4	U
12672-29-6	PCB-1248 (Aroclor 1248)	268	
11097-69-1	PCB-1254 (Aroclor 1254)	127	
11096-82-5	PCB-1260 (Aroclor 1260)	20.0	J
1336-36-3	PCB, Total	415	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418021.D
Lab Smp Id: 40151601003 Client Smp ID: TD2-170612-01-BB-04
Inj Date : 14-MAR-2018 17:41
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 40151601003,
Misc Info : 17523,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 21
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.032	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.026	10783	8081	0.749	0.01		
1.154	12066	10121	0.839	0.01		
1.536	19803	12490	0.631	0.03		
1.954	15311	9619	0.628	0.02		
2.203	10516	8188	0.779	0.01		
2.410	36780	30868	0.839	0.05		
2.485	12266663	12086196	0.985	20.20	$ 2 Tetrachloro-m-xylene	
2.789	23183	23382	1.009	0.03		
2.854	39015	24878	0.638	0.06		
2.922	10473	9149	0.874	0.01		
2.975	80497	54762	0.680	0.12		
3.018	50011	48042	0.961	0.08		
3.055	26057	24477	0.939	0.04		
3.318	63451	58632	0.924	0.10		
3.344	42275	39170	0.927	0.06		
3.404	31774	24087	0.758	0.05		
3.434	79090	68040	0.860	0.12		
3.519	174756	152839	0.875	0.27		
3.648	27303	22522	0.825	0.04		
3.695	20522	20374	0.993	0.03		
3.732	161039	136128	0.845	0.25		
3.827	1656681	1238586	0.748	2.65		
3.934	84514	49439	0.585	0.13		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
3.995	292560	278872	0.953	0.46	
4.057	114250	90547	0.793	0.18	
4.135	31101	25372	0.816	0.04	
4.215	796305	727720	0.914	1.27	
4.261	1339080	1279603	0.956	2.14	6 Aroclor 1248
4.292	1677274	1406384	0.838	2.68	
4.348	19498	20678	1.061	0.03	
4.418	979408	931207	0.951	1.56	
4.453	525504	491272	0.935	0.84	
4.546	1574345	1472096	0.935	2.52	6 Aroclor 1248
4.588	72275	67048	0.928	0.11	
4.618	171248	153956	0.899	0.27	
4.681	83440	65759	0.788	0.13	
4.737	102285	64469	0.630	0.16	
4.768	282530	265682	0.940	0.45	
4.813	1459959	1431964	0.981	2.33	6 Aroclor 1248
4.848	745777	729564	0.978	1.19	
4.879	3020663	2254938	0.747	4.83	6 Aroclor 1248
4.958	333629	280901	0.842	0.53	
5.037	1502469	1408816	0.938	2.40	6 Aroclor 1248
5.101	757807	633644	0.836	1.21	
5.162	1201768	1079174	0.898	1.92	7 Aroclor 1254
5.213	1118813	1013045	0.905	1.79	
5.285	129362	123321	0.953	0.20	
5.310	289268	227057	0.785	0.46	
5.363	544431	489252	0.899	0.87	
5.407	1102589	977261	0.886	1.76	7 Aroclor 1254
5.447	1109757	927283	0.836	1.77	
5.486	72577	90870	1.252	0.11	
5.510	1920323	1688429	0.879	3.07	7 Aroclor 1254
5.580	50596	44048	0.871	0.08	
5.604	359693	302535	0.841	0.57	
5.675	246298	209130	0.849	0.39	
5.712	242742	184077	0.758	0.38	
5.741	380606	271912	0.714	0.60	
5.791	2291332	1806135	0.788	3.66	
5.888	299103	180499	0.603	0.47	
5.932	109205	83612	0.766	0.17	
5.995	297208	283576	0.954	0.47	
6.025	1408198	1258836	0.894	2.25	7 Aroclor 1254
6.057	918317	772274	0.841	1.47	
6.123	42507	24020	0.565	0.06	
6.169	335152	230299	0.687	0.53	
6.226	133005	132145	0.994	0.21	
6.242	191532	149568	0.781	0.30	
6.300	1660000	1386615	0.835	2.65	7 Aroclor 1254
6.335	221790	195640	0.882	0.35	
6.375	125932	86302	0.685	0.20	
6.431	150310	95987	0.639	0.24	
6.487	19723	21319	1.081	0.03	
6.510	436617	400029	0.916	0.69	
6.535	355034	325493	0.917	0.56	
6.561	124265	108622	0.874	0.19	
6.606	80402	52536	0.653	0.12	
6.662	50691	37900	0.748	0.08	
6.719	180485	144706	0.802	0.28	
6.758	178860	146599	0.820	0.28	

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
6.801	256953	231920	0.903	0.41	
6.841	99120	68566	0.692	0.15	
6.912	20871	18360	0.880	0.03	
6.942	200852	184468	0.918	0.32	
7.005	370290	328721	0.888	0.59	8 Aroclor 1260
7.034	28657	33164	1.157	0.04	
7.080	22063	12623	0.572	0.03	
7.122	31467	23022	0.732	0.05	
7.252	204486	177921	0.870	0.32	8 Aroclor 1260
7.281	97207	69812	0.718	0.15	
7.339	22386	17438	0.779	0.03	
7.390	115051	103538	0.900	0.18	
7.441	140755	111957	0.795	0.22	8 Aroclor 1260
7.583	13482	13584	1.008	0.02	
7.627	28796	14587	0.507	0.04	
7.695	36172	34001	0.940	0.05	
7.728	59097	52194	0.883	0.09	
7.812	74854	51544	0.689	0.11	
7.901	85784	66978	0.781	0.13	8 Aroclor 1260
8.050	139099	126961	0.913	0.22	
8.277	133341	116170	0.871	0.21	
8.380	21683	17142	0.791	0.03	
8.450	10377	9297	0.896	0.01	
8.572	10925078	9708929	0.889	17.49	$ 12 Decachlorobiphenyl
8.728	22598	19338	0.856	0.03	
9.057	15432	13458	0.872	0.02	
9.170	52315	46244	0.884	0.08	
9.473	10224	7704	0.754	0.01	
9.960	12225	8420	0.689	0.01	
	62453179	55478729		100.000	

Total unknown % area = 35

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418021.D
Lab Smp Id: 40151601003 Client Smp ID: TD2-170612-01-BB-04
Inj Date : 14-MAR-2018 17:41
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 40151601003,
Misc Info : 17523,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 21
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.032	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221					CAS #: 11104-28-2			
Operator disabled compound identification.								
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.484	2.484	0.000	12266663	0.09949	50			
3 Aroclor 1242					CAS #: 53469-21-9			
Operator disabled compound identification.								
4 Aroclor 1232					CAS #: 1114-16-5			
Operator disabled compound identification.								

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
5 Aroclor 1016					CAS #: 12674-11-2			
Operator disabled compound identification.								
6 Aroclor 1248					CAS #: 12672-29-6			
4.261	4.261	0.000	1339079	0.57153	280	0.00- 20.00	100.00(M)	BA
4.546	4.546	0.000	1574344	0.37933	190	0.00- 20.00	117.57	BA
4.813	4.812	0.001	1459958	0.71916	360	0.00- 20.00	109.03	BA
4.879	4.879	0.000	3020662	0.59964	300	0.00- 20.00	225.58	BA
5.037	5.038	-0.001	1502469	0.41868	210	0.00- 20.00	112.20	BA
Average of Peak Concentrations =					270			
7 Aroclor 1254					CAS #: 11097-69-1			
5.162	5.162	0.000	1201767	0.22955	110	0.00- 20.00	100.00(M)	CO
5.407	5.408	-0.001	1102589	0.25635	130	0.00- 20.00	91.75	CO
5.509	5.510	-0.001	1920322	0.24727	120	0.00- 20.00	159.79	CO
6.024	6.027	-0.003	1408197	0.28918	140	0.00- 20.00	117.18	CO
6.300	6.299	0.001	1660000	0.25511	130	0.00- 20.00	138.13	CO
Average of Peak Concentrations =					130			
8 Aroclor 1260					CAS #: 11096-82-5			
6.300	6.300	0.000	0	0.00	0.00	0.00- 20.00	0.00(M)	CO
7.005	7.006	-0.001	370290	0.03698	18	0.00- 20.00	24.01	CO
7.252	7.251	0.001	204485	0.04278	21	0.00- 20.00	12.32	CO
7.441	7.443	-0.002	140755	0.04828	24	0.00- 20.00	8.48	CO
7.900	7.901	-0.001	85784	0.03251	16	0.00- 20.00	5.17	CO
Average of Peak Concentrations =					20			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.571	8.571	0.000	10925078	0.09972	50			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.
CO: Indicates that the analyst had to split two co-eluting peaks apart that were not (or could not be) separated by the computer system.

Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418021.D

Date : 14-MAR-2018 17:41

Client ID: TD2-170612-01-BB-04

Instrument: 40GCSJ.i

Sample Info: 40151601003,

Volume Injected (uL): 1.0

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b/031418021.D (ManInt: BIOTA.sub Orig: BIOTA.sub)
Injection Date: 14-MAR-2018 17:41 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 40151601003





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	63451	0	
Aroclor 1016	63451	0	
Aroclor 1016	63451	0	
Aroclor 1016	63451	0	
Aroclor 1016	63451	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1232	39015	0	
Aroclor 1232	39015	0	
Aroclor 1232	39015	0	
Aroclor 1232	39015	0	

Aroclor 1232	39015	0	
Aroclor 1242	63451	0	
Aroclor 1242	63451	0	
Aroclor 1242	63451	0	
Aroclor 1242	63451	0	
Aroclor 1242	63451	0	
Aroclor 1248	1325700	1339079	BA
Aroclor 1248	1325700	1339079	BA
Aroclor 1248	1325700	1339079	BA
Aroclor 1248	1325700	1339079	BA
Aroclor 1248	1325700	1339079	BA
Aroclor 1254	1169671	1201767	CO
Aroclor 1254	1169671	1201767	CO
Aroclor 1254	1169671	1201767	CO
Aroclor 1254	1169671	1201767	CO
Aroclor 1254	1169671	1201767	CO
Aroclor 1260	1629435	0	CO
Aroclor 1260	1629435	0	CO
Aroclor 1260	1629435	0	CO
Aroclor 1260	1629435	0	CO
Aroclor 1260	1629435	0	CO
Tetrachloro-m-xylene	12266663	12266663	
Decachlorobiphenyl	10925078	10925078	

==========

SAMPLE NO.

TD2-170612-01-BB-03

PCB - FORM I SVOA-1

SEMI-VOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay
Contract: HUDSON RIVER REMEDIAL ACTION
Date Received: 06/14/2017 09:50
Matrix: Tissue SDG No.: 40151601
Date Extracted: 03/13/2018 07:38
Lab Sample ID: 40151601004
Date Analyzed: 03/14/2018 17:57
Lab File ID: 031418.B\031418022.D
Initial wt/vol: 9.9968 g Final wt/vol: 5 mL Dilution: 1
Instrument: 40GCSJ Percent Moisture:

CAS NO.	COMPOUND	CONCENTRATION UNITS: ug/kg	Q
12674-11-2	PCB-1016 (Aroclor 1016)	<11.4	U
11104-28-2	PCB-1221 (Aroclor 1221)	<11.4	U
11141-16-5	PCB-1232 (Aroclor 1232)	<11.4	U
53469-21-9	PCB-1242 (Aroclor 1242)	<11.4	U
12672-29-6	PCB-1248 (Aroclor 1248)	96.8	
11097-69-1	PCB-1254 (Aroclor 1254)	49.3	J
11096-82-5	PCB-1260 (Aroclor 1260)	<11.4	U
1336-36-3	PCB, Total	146	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418022.D
Lab Smp Id: 40151601004 Client Smp ID: TD2-170612-01-BB-03
Inj Date : 14-MAR-2018 17:57
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 40151601004,
Misc Info : 17523,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 22
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	9.997	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.024	12766	7710	0.604	0.03		
1.154	9667	9059	0.937	0.02		
1.186	23855	12590	0.528	0.06		
1.535	12924	10231	0.792	0.03		
2.200	17375	11009	0.634	0.05		
2.407	13976	12151	0.869	0.04	1 Aroclor 1221	
2.485	9553063	9196641	0.963	28.44	$ 2 Tetrachloro-m-xylene	
2.786	21375	16197	0.758	0.06	1 Aroclor 1221	
2.860	35938	20517	0.571	0.10	1 Aroclor 1221	
2.972	29553	25252	0.854	0.08		
3.019	24839	22834	0.919	0.07		
3.316	18876	17481	0.926	0.05		
3.397	11323	8209	0.725	0.03		
3.433	20166	16878	0.837	0.05		
3.520	34915	30920	0.886	0.10		
3.642	11641	8605	0.739	0.03		
3.732	68355	51407	0.752	0.20		
3.827	626728	449156	0.717	1.83		
3.927	27466	15508	0.565	0.08		
3.995	78192	70814	0.906	0.22		
4.057	65888	45133	0.685	0.19		
4.135	18113	13808	0.762	0.05		
4.215	363386	340689	0.938	1.06		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.240	42528	69693	1.639	0.12	
4.262	449945	406859	0.904	1.31	6 Aroclor 1248
4.292	577894	491955	0.851	1.69	
4.417	375619	344497	0.917	1.10	
4.452	164198	142454	0.868	0.48	
4.545	514749	484363	0.941	1.50	6 Aroclor 1248
4.588	33818	27657	0.818	0.09	
4.616	52530	44436	0.846	0.15	
4.681	39902	28416	0.712	0.11	
4.735	48305	27541	0.570	0.14	
4.768	117370	105787	0.901	0.34	
4.813	562651	534936	0.951	1.64	6 Aroclor 1248
4.848	367155	351676	0.958	1.07	
4.878	1197597	941284	0.786	3.50	6 Aroclor 1248
4.957	127053	97605	0.768	0.37	
5.037	489043	463494	0.948	1.43	6 Aroclor 1248
5.102	313755	251667	0.802	0.91	
5.162	431215	363363	0.843	1.26	7 Aroclor 1254
5.213	430580	363803	0.845	1.26	
5.283	62496	54404	0.871	0.18	
5.309	117550	91311	0.777	0.34	
5.364	227015	184067	0.811	0.66	
5.407	450499	390978	0.868	1.31	7 Aroclor 1254
5.448	422435	346369	0.820	1.23	
5.487	48891	49155	1.005	0.14	
5.510	713478	622065	0.872	2.09	7 Aroclor 1254
5.604	145302	115174	0.793	0.42	
5.675	130834	110477	0.844	0.38	
5.713	105218	78064	0.742	0.30	
5.742	148483	106265	0.716	0.43	
5.791	853436	698226	0.818	2.50	
5.888	126911	76998	0.607	0.37	
5.929	60527	37989	0.628	0.17	
5.995	117965	108016	0.916	0.34	
6.025	552064	478393	0.867	1.61	7 Aroclor 1254
6.056	332795	275744	0.829	0.97	
6.122	27378	15630	0.571	0.08	
6.170	160877	101031	0.628	0.47	
6.229	117765	63885	0.542	0.34	
6.301	653492	564319	0.864	1.91	7 Aroclor 1254
6.334	97032	85764	0.884	0.28	
6.373	60009	40005	0.667	0.17	
6.434	78983	44629	0.565	0.23	
6.510	175164	154265	0.881	0.51	
6.537	154685	134308	0.868	0.45	
6.562	39795	45936	1.154	0.11	
6.609	29398	18607	0.633	0.08	
6.660	21021	19938	0.948	0.06	
6.718	76302	62549	0.820	0.22	
6.759	88885	72341	0.814	0.26	
6.803	118438	100690	0.850	0.34	
6.841	47938	31028	0.647	0.14	
6.944	88584	76995	0.869	0.25	
7.004	163707	152595	0.932	0.47	8 Aroclor 1260
7.036	14096	16346	1.160	0.04	
7.123	22860	16170	0.707	0.06	
7.215	22047	17836	0.809	0.06	

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
7.252	112160	89610	0.799	0.32	8 Aroclor 1260
7.282	57370	38372	0.669	0.16	
7.337	17923	13005	0.726	0.05	
7.388	64070	50590	0.790	0.18	
7.444	71502	54838	0.767	0.20	8 Aroclor 1260
7.693	23039	22404	0.972	0.06	
7.727	46920	38355	0.817	0.13	
7.813	45301	28236	0.623	0.13	
7.901	49426	40776	0.825	0.14	8 Aroclor 1260
8.049	70897	66181	0.933	0.20	
8.274	102732	85800	0.835	0.30	
8.376	10994	8404	0.764	0.03	
8.572	9082978	7832106	0.862	26.61	$ 12 Decachlorobiphenyl
8.723	10458	9517	0.910	0.03	
9.168	30616	25580	0.836	0.08	
9.478	26457	14836	0.561	0.07	
	34133485	30031457		100.000	

Total unknown % area = 26

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418022.D
Lab Smp Id: 40151601004 Client Smp ID: TD2-170612-01-BB-03
Inj Date : 14-MAR-2018 17:57
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 40151601004,
Misc Info : 17523,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 22
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	9.997	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221						CAS #: 11104-28-2		
Compound Not Detected.								
$ 2 Tetrachloro-m-xylene						CAS #: 877-09-8		
2.484	2.484	0.000	9553063	0.07748	39			
3 Aroclor 1242						CAS #: 53469-21-9		
Operator disabled compound identification.								
4 Aroclor 1232						CAS #: 1114-16-5		
Operator disabled compound identification.								

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
5 Aroclor 1016					CAS #: 12674-11-2			
Operator disabled compound identification.								
6 Aroclor 1248					CAS #: 12672-29-6			
4.261	4.261	0.000	449944	0.19204	96	0.00- 20.00	100.00(M)	BA
4.545	4.546	-0.001	514748	0.12403	62	0.00- 20.00	114.40	BA
4.812	4.812	0.000	562651	0.27716	140	0.00- 20.00	125.05	BA
4.877	4.879	-0.002	1197596	0.23774	120	0.00- 20.00	266.17	BA
5.037	5.038	-0.001	489042	0.13628	68	0.00- 20.00	108.69	BA
Average of Peak Concentrations =					97			
7 Aroclor 1254					CAS #: 11097-69-1			
5.162	5.162	0.000	431214	0.08237	41	0.00- 20.00	100.00(M)	CO
5.406	5.408	-0.002	450499	0.10474	52	0.00- 20.00	104.47	CO
5.509	5.510	-0.001	713478	0.09187	46	0.00- 20.00	165.46	CO
6.024	6.027	-0.003	552063	0.11337	57	0.00- 20.00	128.03	CO
6.300	6.299	0.001	653492	0.10043	50	0.00- 20.00	151.55	CO
Average of Peak Concentrations =					49			
8 Aroclor 1260					CAS #: 11096-82-5			
Qualifier signal(s) failed ratio test.								CO
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.571	8.571	0.000	9082978	0.08290	41			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.
CO: Indicates that the analyst had to split two co-eluting peaks apart that were not (or could not be) separated by the computer system.

Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418022.D
Date : 14-MAR-2018 17:57
Client ID: TD2-170612-01-BB-03
Sample Info: 40151601004,
Volume Injected (uL): 1.0
Column phase: ZB-1ms

Instrument: 40GCSJ.i
Operator: BLM
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b/031418022.D (ManInt: BIOTA.sub Orig: BIOTA.sub)
Injection Date: 14-MAR-2018 17:57 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 40151601004





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	18876	0	
Aroclor 1016	18876	0	
Aroclor 1016	18876	0	
Aroclor 1016	18876	0	
Aroclor 1016	18876	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1232	35937	0	
Aroclor 1232	35937	0	
Aroclor 1232	35937	0	
Aroclor 1232	35937	0	

Aroclor 1232	35937	0	
Aroclor 1242	18876	0	
Aroclor 1242	18876	0	
Aroclor 1242	18876	0	
Aroclor 1242	18876	0	
Aroclor 1242	18876	0	
Aroclor 1248	435357	449944	BA
Aroclor 1248	435357	449944	BA
Aroclor 1248	435357	449944	BA
Aroclor 1248	435357	449944	BA
Aroclor 1248	435357	449944	BA
Aroclor 1254	403423	431214	CO
Aroclor 1254	403423	431214	CO
Aroclor 1254	403423	431214	CO
Aroclor 1254	403423	431214	CO
Aroclor 1254	403423	431214	CO
Aroclor 1260	627862	0	CO
Aroclor 1260	627862	0	CO
Aroclor 1260	627862	0	CO
Aroclor 1260	627862	0	CO
Aroclor 1260	627862	0	CO
Tetrachloro-m-xylene	9553063	9553063	
Decachlorobiphenyl	9082978	9082978	

==========

SAMPLE NO.

TD2-170612-01-BB-02

PCB - FORM I SVOA-1

SEMI-VOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay

Contract: HUDSON RIVER REMEDIAL ACTION

Date Received: 06/14/2017 09:50

Matrix: Tissue SDG No.: 40151601

Date Extracted: 03/13/2018 07:38

Lab Sample ID: 40151601005

Date Analyzed: 03/14/2018 18:13

Lab File ID: 031418.B\031418023.D

Initial wt/vol: 10.0156 Final wt/vol: 5 mL Dilution: 1

Instrument: 40GCSJ Percent Moisture:

CAS NO.	COMPOUND	CONCENTRATION UNITS: ug/kg	Q
12674-11-2	PCB-1016 (Aroclor 1016)	<11.4	U
11104-28-2	PCB-1221 (Aroclor 1221)	<11.4	U
11141-16-5	PCB-1232 (Aroclor 1232)	<11.4	U
53469-21-9	PCB-1242 (Aroclor 1242)	<11.4	U
12672-29-6	PCB-1248 (Aroclor 1248)	217	
11097-69-1	PCB-1254 (Aroclor 1254)	130	
11096-82-5	PCB-1260 (Aroclor 1260)	29.6	J
1336-36-3	PCB, Total	377	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418023.D
Lab Smp Id: 40151601005 Client Smp ID: TD2-170612-01-BB-02
Inj Date : 14-MAR-2018 18:13
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 40151601005,
Misc Info : 17523,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 23
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.016	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.040	12973	12370	0.954	0.02		
1.157	19400	10955	0.565	0.03		
1.540	17606	12366	0.702	0.02		
2.410	19505	18191	0.933	0.03	1 Aroclor 1221	
2.485	11720350	11426664	0.975	20.36	$ 2 Tetrachloro-m-xylene	
2.784	23123	19631	0.849	0.03	1 Aroclor 1221	
2.859	31894	20800	0.652	0.05	1 Aroclor 1221	
2.974	58590	29703	0.507	0.09		
3.021	40769	43983	1.079	0.06		
3.320	23985	22329	0.931	0.04		
3.343	18386	15701	0.854	0.03		
3.405	42728	34519	0.808	0.07		
3.433	19834	14907	0.752	0.03		
3.519	63277	56741	0.897	0.10		
3.643	12930	10228	0.791	0.02		
3.694	38223	34882	0.913	0.06		
3.732	75780	81270	1.072	0.12		
3.754	18953	20486	1.081	0.03		
3.827	1181727	854377	0.723	1.99		
3.938	54825	40243	0.734	0.09		
3.994	103515	95319	0.921	0.17		
4.056	63327	47289	0.747	0.10		
4.214	1132644	1074028	0.948	1.91		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.262	1008787	1005365	0.997	1.70	6 Aroclor 1248
4.293	1208563	1024006	0.847	2.04	
4.418	615004	572041	0.930	1.04	
4.453	300353	263384	0.877	0.50	
4.546	997355	947690	0.950	1.68	6 Aroclor 1248
4.588	50299	46774	0.930	0.08	
4.618	50491	41864	0.829	0.08	
4.680	53044	40201	0.758	0.08	
4.737	73176	51181	0.699	0.12	
4.768	216632	209822	0.969	0.36	
4.813	1327810	1264360	0.952	2.24	6 Aroclor 1248
4.848	920359	895255	0.973	1.55	
4.878	2872430	2134824	0.743	4.85	6 Aroclor 1248
4.957	253687	220194	0.868	0.42	
5.038	994174	941211	0.947	1.68	6 Aroclor 1248
5.101	527901	458396	0.868	0.89	
5.163	1297444	1157822	0.892	2.19	7 Aroclor 1254
5.212	1361782	1216646	0.893	2.30	
5.284	124635	112707	0.904	0.21	
5.310	189103	152406	0.806	0.31	
5.364	311325	258044	0.829	0.52	
5.407	960755	814951	0.848	1.62	7 Aroclor 1254
5.447	1130577	973139	0.861	1.91	
5.487	81487	97216	1.193	0.13	
5.510	1634446	1406260	0.860	2.76	7 Aroclor 1254
5.579	68926	46707	0.678	0.11	
5.605	124477	96390	0.774	0.21	
5.676	292060	264292	0.905	0.49	
5.712	241432	184754	0.765	0.40	
5.742	362950	260387	0.717	0.61	
5.791	2601228	2063660	0.793	4.40	
5.888	257902	172508	0.669	0.43	
5.933	112438	89734	0.798	0.19	
5.994	344020	318824	0.927	0.58	
6.024	1440554	1272890	0.884	2.43	7 Aroclor 1254
6.057	1466523	1267690	0.864	2.48	
6.124	33428	22761	0.681	0.05	
6.169	374255	261683	0.699	0.63	
6.227	174758	163383	0.935	0.29	
6.243	173675	153647	0.885	0.29	
6.300	2125030	1725140	0.812	3.59	7 Aroclor 1254
6.334	277563	243939	0.879	0.46	
6.375	105670	79535	0.753	0.17	
6.433	172372	123122	0.714	0.29	
6.509	558383	534573	0.957	0.94	
6.536	458070	410208	0.896	0.77	
6.560	163256	150101	0.919	0.27	
6.606	89810	71550	0.797	0.15	
6.659	52388	39589	0.756	0.08	
6.718	192131	163879	0.853	0.32	
6.758	171902	146644	0.853	0.29	
6.802	321380	286078	0.890	0.54	
6.842	114994	92491	0.804	0.19	
6.912	21441	19849	0.926	0.03	
6.943	205257	187354	0.913	0.34	
7.006	557997	523547	0.938	0.94	8 Aroclor 1260
7.037	49785	49321	0.991	0.08	

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
7.075	17365	14997	0.864	0.02	
7.122	43938	21419	0.487	0.07	
7.252	283555	257549	0.908	0.47	8 Aroclor 1260
7.283	116677	93117	0.798	0.19	
7.339	18614	15205	0.817	0.03	
7.389	185166	164251	0.887	0.31	
7.443	197108	161979	0.822	0.33	8 Aroclor 1260
7.554	15396	15071	0.979	0.02	
7.691	58036	52436	0.904	0.09	
7.727	75350	62554	0.830	0.12	
7.766	17490	12617	0.721	0.02	
7.813	74854	45356	0.606	0.12	
7.899	144443	120990	0.838	0.24	8 Aroclor 1260
7.958	22544	13611	0.604	0.03	
8.049	84060	79344	0.944	0.14	
8.137	9917	8103	0.817	0.01	
8.275	166663	147372	0.884	0.28	
8.374	19070	14112	0.740	0.03	
8.572	10369273	9074285	0.875	17.54	$ 12 Decachlorobiphenyl
8.726	22477	16375	0.729	0.03	
8.940	10612	11247	1.060	0.01	
9.053	13243	12319	0.930	0.02	
9.170	46812	39648	0.847	0.07	
9.466	20159	14776	0.733	0.03	
9.961	18396	16000	0.870	0.03	
	59113236	52273704		100.000	

Total unknown % area = 35

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418023.D
Lab Smp Id: 40151601005 Client Smp ID: TD2-170612-01-BB-02
Inj Date : 14-MAR-2018 18:13
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 40151601005,
Misc Info : 17523,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 23
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.016	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE

1 Aroclor 1221 CAS #: 11104-28-2

Compound Not Detected.

$ 2 Tetrachloro-m-xylene CAS #: 877-09-8

RT	EXP RT	DLT RT	RESPONSE	ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
2.484	2.484	0.000	11720350	0.09506	47			

---

3 Aroclor 1242 CAS #: 53469-21-9

Operator disabled compound identification.

---

4 Aroclor 1232 CAS #: 1114-16-5

Operator disabled compound identification.

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
5 Aroclor 1016					CAS #: 12674-11-2			
Operator disabled compound identification.								
6 Aroclor 1248					CAS #: 12672-29-6			
4.261	4.261	0.000	1008787	0.43056	210	0.00- 20.00	100.00(M)	BA
4.546	4.546	0.000	997355	0.24031	120	0.00- 20.00	98.87	BA
4.813	4.812	0.001	1327809	0.65407	330	0.00- 20.00	131.62	BA
4.878	4.879	-0.001	2872430	0.57022	280	0.00- 20.00	284.74	BA
5.038	5.038	0.000	994173	0.27704	140	0.00- 20.00	98.55	BA
			Average of Peak Concentrations =		220			
7 Aroclor 1254					CAS #: 11097-69-1			
5.162	5.162	0.000	1297443	0.24783	120	0.00- 20.00	100.00(M)	CO
5.407	5.408	-0.001	960754	0.22338	110	0.00- 20.00	74.05	CO
5.510	5.510	0.000	1634445	0.21046	100	0.00- 20.00	125.97	CO
6.023	6.027	-0.004	1440554	0.29582	150	0.00- 20.00	111.03	CO
6.299	6.299	0.000	2125029	0.32657	160	0.00- 20.00	163.79	CO
			Average of Peak Concentrations =		130			
8 Aroclor 1260					CAS #: 11096-82-5			
6.300	6.300	0.000	0	0.00	0.00	0.00- 20.00	0.00(M)	BA
7.005	7.006	-0.001	557997	0.05572	28	0.00- 20.00	26.26	BA
7.251	7.251	0.000	283554	0.05932	30	0.00- 20.00	13.34	BA
7.443	7.443	0.000	197107	0.06762	34	0.00- 20.00	9.28	BA
7.899	7.901	-0.002	144442	0.05474	27	0.00- 20.00	6.80	BA
			Average of Peak Concentrations =		30			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.571	8.571	0.000	10369272	0.09464	47			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.
CO: Indicates that the analyst had to split two co-eluting peaks apart that were not (or could not be) separated by the computer system.

Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418023.D
Date : 14-MAR-2018 18:13
Client ID: TD2-170612-01-BB-02
Sample Info: 40151601005,
Volume Injected (uL): 1.0
Column phase: ZB-1ms

Instrument: 40GCSJ.i
Operator: BLM
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b/031418023.D (ManInt: BIOTA.sub Orig: BIOTA.sub)
Injection Date: 14-MAR-2018 18:13 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 40151601005





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	23984	0	
Aroclor 1016	23984	0	
Aroclor 1016	23984	0	
Aroclor 1016	23984	0	
Aroclor 1016	23984	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1232	31893	0	
Aroclor 1232	31893	0	
Aroclor 1232	31893	0	
Aroclor 1232	31893	0	

Aroclor 1232	31893	0	
Aroclor 1242	23984	0	
Aroclor 1242	23984	0	
Aroclor 1242	23984	0	
Aroclor 1242	23984	0	
Aroclor 1242	23984	0	
Aroclor 1248	995501	1008787	BA
Aroclor 1248	995501	1008787	BA
Aroclor 1248	995501	1008787	BA
Aroclor 1248	995501	1008787	BA
Aroclor 1248	995501	1008787	BA
Aroclor 1254	1274614	1297443	CO
Aroclor 1254	1274614	1297443	CO
Aroclor 1254	1274614	1297443	CO
Aroclor 1254	1274614	1297443	CO
Aroclor 1254	1274614	1297443	CO
Aroclor 1260	2094907	0	BA
Aroclor 1260	2094907	0	BA
Aroclor 1260	2094907	0	BA
Aroclor 1260	2094907	0	BA
Aroclor 1260	2094907	0	BA
Tetrachloro-m-xylene	11720350	11720350	
Decachlorobiphenyl	10369272	10369272	

==========

SAMPLE NO.

TD2-170612-01-BB-01

PCB - FORM I SVOA-1

SEMI-VOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay
Contract: HUDSON RIVER REMEDIAL ACTION
Date Received: 06/14/2017 09:50
Matrix: Tissue SDG No.: 40151601
Date Extracted: 03/13/2018 07:38
Lab Sample ID: 40151601006
Date Analyzed: 03/14/2018 18:29
Lab File ID: 031418.B\031418024.D
Initial wt/vol: 9.9883 g Final wt/vol: 5 mL Dilution: 1
Instrument: 40GCSJ Percent Moisture:

CAS NO.	COMPOUND	CONCENTRATION UNITS: ug/kg	Q
12674-11-2	PCB-1016 (Aroclor 1016)	<11.4	U
11104-28-2	PCB-1221 (Aroclor 1221)	<11.4	U
11141-16-5	PCB-1232 (Aroclor 1232)	<11.4	U
53469-21-9	PCB-1242 (Aroclor 1242)	<11.4	U
12672-29-6	PCB-1248 (Aroclor 1248)	239	
11097-69-1	PCB-1254 (Aroclor 1254)	108	
11096-82-5	PCB-1260 (Aroclor 1260)	22.1	J
1336-36-3	PCB, Total	369	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418024.D
Lab Smp Id: 40151601006 Client Smp ID: TD2-170612-01-BB-01
Inj Date : 14-MAR-2018 18:29
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 40151601006,
Misc Info : 17523,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 24
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	9.988	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.033	18328	12282	0.670	0.03		
1.157	19236	14149	0.736	0.03		
1.427	13624	10746	0.789	0.02		
1.537	15951	11439	0.717	0.02		
2.411	62557	52656	0.842	0.10		
2.485	11911442	11668741	0.980	20.39	$ 2 Tetrachloro-m-xylene	
2.788	30967	27047	0.873	0.05		
2.852	64852	47546	0.733	0.10		
2.918	15966	10704	0.670	0.02		
2.973	90041	52084	0.578	0.15		
3.019	63480	64330	1.013	0.10		
3.054	35499	31088	0.876	0.05		
3.317	199459	193471	0.970	0.33		
3.344	80508	79990	0.994	0.13		
3.402	12024	14114	1.174	0.02		
3.433	163324	156159	0.956	0.27		
3.519	288219	259572	0.901	0.48		
3.645	41576	36979	0.889	0.06		
3.732	439564	377895	0.860	0.73		
3.827	1928053	1176601	0.610	3.21		
3.938	104879	59276	0.565	0.17		
3.994	245897	238503	0.970	0.40		
4.058	157489	145332	0.923	0.26		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.137	55751	44162	0.792	0.09	
4.215	1261963	1151213	0.912	2.10	
4.261	1638592	1583026	0.966	2.73	6 Aroclor 1248
4.292	2021009	1713590	0.848	3.36	
4.343	23742	21206	0.893	0.03	
4.417	837640	795913	0.950	1.39	
4.452	503209	474486	0.943	0.83	
4.545	1368851	1253668	0.916	2.28	6 Aroclor 1248
4.589	100964	92164	0.913	0.16	
4.618	203415	183234	0.901	0.33	
4.680	83936	68011	0.810	0.13	
4.734	108174	71521	0.661	0.18	
4.768	241001	231660	0.961	0.40	
4.813	1165016	1123280	0.964	1.94	6 Aroclor 1248
4.847	720470	697147	0.968	1.20	
4.878	2517284	1697471	0.674	4.19	6 Aroclor 1248
4.956	322107	267498	0.830	0.53	
5.037	1019889	972795	0.954	1.70	6 Aroclor 1248
5.100	762509	647879	0.850	1.27	
5.161	955715	829005	0.867	1.59	7 Aroclor 1254
5.212	999205	874397	0.875	1.66	
5.283	128008	118264	0.924	0.21	
5.308	227017	190257	0.838	0.37	
5.363	430455	374570	0.870	0.71	
5.407	846948	760009	0.897	1.41	7 Aroclor 1254
5.448	866663	734436	0.847	1.44	
5.484	95126	103826	1.091	0.15	
5.510	1591137	1381675	0.868	2.65	7 Aroclor 1254
5.575	61083	51963	0.851	0.10	
5.603	275024	233753	0.850	0.45	
5.675	301034	270422	0.898	0.50	
5.712	227461	182855	0.804	0.37	
5.742	349622	250537	0.717	0.58	
5.790	2022726	1585491	0.784	3.37	
5.887	242384	149864	0.618	0.40	
5.933	140800	103051	0.732	0.23	
5.993	322822	285490	0.884	0.53	
6.024	1153133	1030454	0.894	1.92	7 Aroclor 1254
6.055	867234	723354	0.834	1.44	
6.126	34660	23766	0.686	0.05	
6.169	289897	168244	0.580	0.48	
6.226	121412	123570	1.018	0.20	
6.241	175372	137522	0.784	0.29	
6.300	1677125	1403412	0.837	2.79	7 Aroclor 1254
6.333	191557	163004	0.851	0.31	
6.371	107484	72913	0.678	0.17	
6.433	187470	125943	0.672	0.31	
6.510	454043	396418	0.873	0.75	
6.535	364796	320761	0.879	0.60	
6.560	120214	101552	0.845	0.20	
6.605	92681	56946	0.614	0.15	
6.660	47620	33387	0.701	0.07	
6.717	172711	127594	0.739	0.28	
6.759	167257	138853	0.830	0.27	
6.801	251389	214658	0.854	0.41	
6.842	110829	76828	0.693	0.18	
6.911	25973	20244	0.779	0.04	

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
6.941	152377	137558	0.903	0.25	
7.005	353697	315714	0.893	0.58	8 Aroclor 1260
7.034	49319	46416	0.941	0.08	
7.077	21388	12880	0.602	0.03	
7.124	45882	21251	0.463	0.07	
7.194	41662	18333	0.440	0.06	
7.252	218891	183974	0.840	0.36	8 Aroclor 1260
7.283	102915	74546	0.724	0.17	
7.337	31588	20804	0.659	0.05	
7.390	136255	118000	0.866	0.22	
7.443	160998	124347	0.772	0.26	8 Aroclor 1260
7.553	13278	13788	1.038	0.02	
7.576	10663	9222	0.865	0.01	
7.692	37747	35111	0.930	0.06	
7.728	63740	55664	0.873	0.10	
7.810	48893	29331	0.600	0.08	
7.900	105873	84884	0.802	0.17	8 Aroclor 1260
8.049	73058	62822	0.860	0.12	
8.135	28994	21386	0.738	0.04	
8.274	131471	120933	0.920	0.21	
8.382	24479	14754	0.603	0.04	
8.447	13752	11820	0.860	0.02	
8.483	18591	11630	0.626	0.03	
8.572	10358743	9073089	0.876	17.26	$ 12 Decachlorobiphenyl
8.668	10443	8381	0.803	0.01	
8.723	33456	15032	0.449	0.05	
8.946	21038	10769	0.512	0.03	
9.172	23765	21041	0.885	0.03	
	59987496	52443396		100.000	

Total unknown % area = 38

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418024.D
Lab Smp Id: 40151601006 Client Smp ID: TD2-170612-01-BB-01
Inj Date : 14-MAR-2018 18:29
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 40151601006,
Misc Info : 17523,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 24
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	9.988	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221					CAS #: 11104-28-2			
Operator disabled compound identification.								
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.484	2.484	0.000	11911441	0.09661	48			
3 Aroclor 1242					CAS #: 53469-21-9			
Operator disabled compound identification.								
4 Aroclor 1232					CAS #: 1114-16-5			
Operator disabled compound identification.								

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
5 Aroclor 1016					CAS #: 12674-11-2			
Operator disabled compound identification.								
6 Aroclor 1248					CAS #: 12672-29-6			
4.260	4.261	-0.001	1638592	0.69936	350	0.00- 20.00	100.00(M)	BA
4.545	4.546	-0.001	1368850	0.32982	160	0.00- 20.00	83.54	BA
4.812	4.812	0.000	1165016	0.57388	290	0.00- 20.00	71.10	BA
4.878	4.879	-0.001	2517283	0.49971	250	0.00- 20.00	153.62	BA
5.036	5.038	-0.002	1019889	0.28421	140	0.00- 20.00	62.24	BA
			Average of Peak Concentrations =		240			
7 Aroclor 1254					CAS #: 11097-69-1			
5.160	5.162	-0.002	955714	0.18255	91	0.00- 20.00	100.00(M)	BA
5.406	5.408	-0.002	846947	0.19692	98	0.00- 20.00	88.62	BA
5.509	5.510	-0.001	1591136	0.20488	100	0.00- 20.00	166.49	BA
6.024	6.027	-0.003	1153132	0.23680	120	0.00- 20.00	120.66	BA
6.299	6.299	0.000	1677125	0.25774	130	0.00- 20.00	175.48	BA
			Average of Peak Concentrations =		110			
8 Aroclor 1260					CAS #: 11096-82-5			
6.300	6.300	0.000	0	0.00	0.00	0.00- 20.00	0.00(M)	BA
7.004	7.006	-0.002	353696	0.03532	18	0.00- 20.00	21.09	BA
7.251	7.251	0.000	218891	0.04579	23	0.00- 20.00	13.05	BA
7.442	7.443	-0.001	160997	0.05523	28	0.00- 20.00	9.60	BA
7.899	7.901	-0.002	105873	0.04012	20	0.00- 20.00	6.31	BA
			Average of Peak Concentrations =		22			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.571	8.571	0.000	10358742	0.09455	47			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418024.D
Date : 14-MAR-2018 18:29
Client ID: TD2-170612-01-BB-01
Instrument: 40GCSJ.i
Sample Info: 40151601006,
Volume Injected (uL): 1.0
Operator: BLM
Column phase: ZB-1ms
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b/031418024.D (ManInt: BIOTA.sub Orig: BIOTA.sub)
Injection Date: 14-MAR-2018 18:29 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 40151601006





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	199458	0	
Aroclor 1016	199458	0	
Aroclor 1016	199458	0	
Aroclor 1016	199458	0	
Aroclor 1016	199458	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1232	64851	0	
Aroclor 1232	64851	0	
Aroclor 1232	64851	0	
Aroclor 1232	64851	0	

Aroclor 1232	64851	0	
Aroclor 1242	199458	0	
Aroclor 1242	199458	0	
Aroclor 1242	199458	0	
Aroclor 1242	199458	0	
Aroclor 1242	199458	0	
Aroclor 1248	1623235	1638592	BA
Aroclor 1248	1623235	1638592	BA
Aroclor 1248	1623235	1638592	BA
Aroclor 1248	1623235	1638592	BA
Aroclor 1248	1623235	1638592	BA
Aroclor 1254	928520	955714	BA
Aroclor 1254	928520	955714	BA
Aroclor 1254	928520	955714	BA
Aroclor 1254	928520	955714	BA
Aroclor 1254	928520	955714	BA
Aroclor 1260	1645098	0	BA
Aroclor 1260	1645098	0	BA
Aroclor 1260	1645098	0	BA
Aroclor 1260	1645098	0	BA
Aroclor 1260	1645098	0	BA
Tetrachloro-m-xylene	11911441	11911441	
Decachlorobiphenyl	10358742	10358742	

==========

PCB - FORM VI SVOA-1

PCB INITIAL CALIBRATION (MULTIPOINT)

Lab Name: Pace Analytical - Green Bay Instrument ID: 40GCSJ GC Column: Col 1 SDG No.: 40151601

Calibration Date(s): 03/06/2018 03/06/2018 Calibration Time(s): 12:07 19:39

LAB FILE ID

CAL1A =	030618.B\030618058.D	CAL1B =	030618.B\030618022.D	CAL1C =	030618.B\030618046.D
CAL1D =	030618.B\030618034.D	CAL1F =	030618.B\030618010.D	CAL2A =	030618.B\030618060.D
CAL2B =	030618.B\030618024.D	CAL2C =	030618.B\030618048.D	CAL2D =	030618.B\030618036.D
CAL2F =	030618.B\030618012.D	CAL3A =	030618.B\030618062.D	CAL3B =	030618.B\030618026.D
CAL3C =	030618.B\030618050.D	CAL3D =	030618.B\030618038.D	CAL3F =	030618.B\030618014.D
CAL4A =	030618.B\030618064.D	CAL4B =	030618.B\030618028.D	CAL4C =	030618.B\030618052.D
CAL4D =	030618.B\030618040.D	CAL4F =	030618.B\030618016.D	CAL5A =	030618.B\030618066.D
CAL5B =	030618.B\030618030.D	CAL5C =	030618.B\030618054.D	CAL5D =	030618.B\030618042.D
CAL5F =	030618.B\030618018.D				

COMPOUND	PEAK	CURVE TYPE	CAL1	CAL2	CAL3	CAL4	CAL5
Aroclor 1016	1	Averaged	3346700.00	3270300.00	3026998.00	2916970.00	2843185.00
Aroclor 1016	2	Averaged	3361310.00	3225075.00	3064774.00	2953430.00	2890423.00
Aroclor 1016	3	Averaged	9619090.00	9294970.00	8629258.00	8432580.00	8333033.00
Aroclor 1016	4	Averaged	4072990.00	3972810.00	3750126.00	3581221.25	3523970.00
Aroclor 1016	5	Averaged	2912980.00	2694260.00	2551278.00	2433447.50	2424781.00
Aroclor 1221	1	Averaged	1168840.00	1110755.00	980478.000	981937.500	964273.000
Aroclor 1221	2	Averaged	612180.000	587450.000	526622.000	511427.500	501535.000
Aroclor 1221	3	Averaged	1505110.00	1446635.00	1316246.00	1333917.50	1314775.00
Aroclor 1221	4	Averaged	1181440.00	1150255.00	1038526.00	1067413.75	1091213.00
Aroclor 1221	5	Averaged	3492190.00	3470535.00	3167038.00	3088411.25	3058814.00
Aroclor 1242	1	Averaged	2772560.00	2688000.00	2505132.00	2371062.50	2329239.00
Aroclor 1242	2	Averaged	2739500.00	2600755.00	2506348.00	2392166.25	2361398.00
Aroclor 1242	3	Averaged	7695890.00	7473100.00	6971326.00	6735328.75	6706645.00
Aroclor 1242	4	Averaged	3287540.00	3169245.00	3017986.00	2875101.25	2857801.00
Aroclor 1242	5	Averaged	2247420.00	2077595.00	2062930.00	1928282.50	1922507.00
Aroclor 1248	1	Averaged	2522640.00	2478140.00	2346396.00	2184111.25	2183582.00
Aroclor 1248	2	Averaged	4441400.00	4466680.00	4104774.00	3847133.75	3891542.00
Aroclor 1248	3	Averaged	2175950.00	2100970.00	2034380.00	1911980.00	1927121.00
Aroclor 1248	4	Averaged	5384910.00	5302865.00	5054976.00	4695606.25	4748871.00
Aroclor 1248	5	Averaged	3757360.00	3770200.00	3588002.00	3386248.75	3440920.00
Aroclor 1254	1	Averaged	5783370.00	5652575.00	5120584.00	4858101.25	4761607.00
Aroclor 1254	2	Averaged	4616750.00	4581855.00	4225042.00	4064631.25	4016994.00
Aroclor 1254	3	Averaged	8435240.00	8286960.00	7623568.00	7277021.25	7207530.00
Aroclor 1254	4	Averaged	5140880.00	5151350.00	4826428.00	4638268.75	4591491.00
Aroclor 1254	5	Averaged	7015020.00	6981970.00	6368818.00	6122333.75	6047377.00
Aroclor 1260	1	Averaged	9569990.00	9323545.00	8657528.00	8388552.50	8372147.00
Aroclor 1260	2	Averaged	10746610.0	10429350.0	9816406.00	9526972.50	9552528.00
Aroclor 1260	3	Averaged	5052370.00	4986225.00	4721308.00	4577403.75	4561709.00
Aroclor 1260	4	Averaged	3131020.00	3037915.00	2887698.00	2771190.00	2747730.00

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

PCB - FORM VI SVOA-2

PCB INITIAL CALIBRATION (MULTIPOINT)

Lab Name: Pace Analytical - Green Bay Instrument ID: 40GCSJ GC Column: Col 1 SDG No.: 40151601

Calibration Date(s): 03/06/2018 03/06/2018 Calibration Time(s): 12:07 19:39

LAB FILE ID

CAL1A =	030618.B\030618058.D	CAL1B =	030618.B\030618022.D	CAL1C =	030618.B\030618046.D
CAL1D =	030618.B\030618034.D	CAL1F =	030618.B\030618010.D	CAL2A =	030618.B\030618060.D
CAL2B =	030618.B\030618024.D	CAL2C =	030618.B\030618048.D	CAL2D =	030618.B\030618036.D
CAL2F =	030618.B\030618012.D	CAL3A =	030618.B\030618062.D	CAL3B =	030618.B\030618026.D
CAL3C =	030618.B\030618050.D	CAL3D =	030618.B\030618038.D	CAL3F =	030618.B\030618014.D
CAL4A =	030618.B\030618064.D	CAL4B =	030618.B\030618028.D	CAL4C =	030618.B\030618052.D
CAL4D =	030618.B\030618040.D	CAL4F =	030618.B\030618016.D	CAL5A =	030618.B\030618066.D
CAL5B =	030618.B\030618030.D	CAL5C =	030618.B\030618054.D	CAL5D =	030618.B\030618042.D
CAL5F =	030618.B\030618018.D				

COMPOUND	PEAK	CURVE TYPE	CAL1	CAL2	CAL3	CAL4	CAL5
Aroclor 1260	5	Averaged	2853590.00	2718235.00	2625044.00	2511683.75	2485127.00
Decachlorobiphenyl (S)		Averaged	125223400.	114228350.	106255020.	101602250.	100490933.
Tetrachloro-m-xylene		Averaged	126550100.	124974300.	121144740.	121809550.	121996840.

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

PCB - FORM VI SVOA-3

PCB INITIAL CALIBRATION (MULTIPOINT)

Lab Name: Pace Analytical - Green Bay Instrument ID: 40GCSJ GC Column: Col 1 SDG No.: 40151601

Calibration Date(s): 03/06/2018 03/06/2018 Calibration Time(s): 12:07 19:39

LAB FILE ID

CAL1A =	030618.B\030618058.D	CAL1B =	030618.B\030618022.D	CAL1C =	030618.B\030618046.D
CAL1D =	030618.B\030618034.D	CAL1F =	030618.B\030618010.D	CAL2A =	030618.B\030618060.D
CAL2B =	030618.B\030618024.D	CAL2C =	030618.B\030618048.D	CAL2D =	030618.B\030618036.D
CAL2F =	030618.B\030618012.D	CAL3A =	030618.B\030618062.D	CAL3B =	030618.B\030618026.D
CAL3C =	030618.B\030618050.D	CAL3D =	030618.B\030618038.D	CAL3F =	030618.B\030618014.D
CAL4A =	030618.B\030618064.D	CAL4B =	030618.B\030618028.D	CAL4C =	030618.B\030618052.D
CAL4D =	030618.B\030618040.D	CAL4F =	030618.B\030618016.D	CAL5A =	030618.B\030618066.D
CAL5B =	030618.B\030618030.D	CAL5C =	030618.B\030618054.D	CAL5D =	030618.B\030618042.D
CAL5F =	030618.B\030618018.D				

COMPOUND	PEAK	CURVE TYPE	%RSD	R2	A1	A2	A3
Aroclor 1016	1	Averaged	7.12			3080830.60	
Aroclor 1016	2	Averaged	6.26			3099002.40	
Aroclor 1016	3	Averaged	6.38			8861786.20	
Aroclor 1016	4	Averaged	6.32			3780223.45	
Aroclor 1016	5	Averaged	7.86			2603349.30	
Aroclor 1221	1	Averaged	8.88			1041256.70	
Aroclor 1221	2	Averaged	8.95			547842.90	
Aroclor 1221	3	Averaged	6.31			1383336.70	
Aroclor 1221	4	Averaged	5.33			1105769.55	
Aroclor 1221	5	Averaged	6.45			3255397.65	
Aroclor 1242	1	Averaged	7.64			2533198.70	
Aroclor 1242	2	Averaged	6.15			2520033.45	
Aroclor 1242	3	Averaged	6.27			7116457.95	
Aroclor 1242	4	Averaged	6.12			3041534.65	
Aroclor 1242	5	Averaged	6.50			2047746.90	
Aroclor 1248	1	Averaged	6.78			2342973.85	
Aroclor 1248	2	Averaged	7.08			4150305.95	
Aroclor 1248	3	Averaged	5.55			2030080.20	
Aroclor 1248	4	Averaged	6.21			5037445.65	
Aroclor 1248	5	Averaged	4.91			3588546.15	
Aroclor 1254	1	Averaged	8.82			5235247.45	
Aroclor 1254	2	Averaged	6.58			4301054.45	
Aroclor 1254	3	Averaged	7.31			7766063.85	
Aroclor 1254	4	Averaged	5.48			4869683.55	
Aroclor 1254	5	Averaged	7.13			6507103.75	
Aroclor 1260	1	Averaged	6.23			8862352.50	
Aroclor 1260	2	Averaged	5.46			10014373.30	
Aroclor 1260	3	Averaged	4.78			4779803.15	
Aroclor 1260	4	Averaged	5.71			2915110.60	

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

PCB - FORM VI SVOA-4

PCB INITIAL CALIBRATION (MULTIPOINT)

Lab Name: Pace Analytical - Green Bay Instrument ID: 40GCSJ GC Column: Col 1 SDG No.: 40151601

Calibration Date(s): 03/06/2018 03/06/2018 Calibration Time(s): 12:07 19:39

LAB FILE ID

CAL1A =	030618.B\030618058.D	CAL1B =	030618.B\030618022.D	CAL1C =	030618.B\030618046.D
CAL1D =	030618.B\030618034.D	CAL1F =	030618.B\030618010.D	CAL2A =	030618.B\030618060.D
CAL2B =	030618.B\030618024.D	CAL2C =	030618.B\030618048.D	CAL2D =	030618.B\030618036.D
CAL2F =	030618.B\030618012.D	CAL3A =	030618.B\030618062.D	CAL3B =	030618.B\030618026.D
CAL3C =	030618.B\030618050.D	CAL3D =	030618.B\030618038.D	CAL3F =	030618.B\030618014.D
CAL4A =	030618.B\030618064.D	CAL4B =	030618.B\030618028.D	CAL4C =	030618.B\030618052.D
CAL4D =	030618.B\030618040.D	CAL4F =	030618.B\030618016.D	CAL5A =	030618.B\030618066.D
CAL5B =	030618.B\030618030.D	CAL5C =	030618.B\030618054.D	CAL5D =	030618.B\030618042.D
CAL5F =	030618.B\030618018.D				

COMPOUND	PEAK	CURVE TYPE	%RSD	R2	A1	A2	A3
Aroclor 1260	5	Averaged	5.76			2638735.95	
Decachlorobiphenyl (S)		Averaged	9.39			109559990.6	
Tetrachloro-m-xylene		Averaged	1.89			123295106.0	

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

PCB - FORM VI SVOA-1

PCB INITIAL CALIBRATION RETENTION TIME (MULTIPOINT)

Lab Name: Pace Analytical - Green Bay Instrument ID: 40GCSJ GC Column: Col 1 SDG No.: 40151601

Calibration Date(s): 03/06/2018 03/06/2018 Calibration Time(s): 12:07 19:39

LAB FILE ID

CAL1A =	030618.B\030618058.D	CAL1B =	030618.B\030618022.D	CAL1C =	030618.B\030618046.D
CAL1D =	030618.B\030618034.D	CAL1F =	030618.B\030618010.D	CAL2A =	030618.B\030618060.D
CAL2B =	030618.B\030618024.D	CAL2C =	030618.B\030618048.D	CAL2D =	030618.B\030618036.D
CAL2F =	030618.B\030618012.D	CAL3A =	030618.B\030618062.D	CAL3B =	030618.B\030618026.D
CAL3C =	030618.B\030618050.D	CAL3D =	030618.B\030618038.D	CAL3F =	030618.B\030618014.D
CAL4A =	030618.B\030618064.D	CAL4B =	030618.B\030618028.D	CAL4C =	030618.B\030618052.D
CAL4D =	030618.B\030618040.D	CAL4F =	030618.B\030618016.D	CAL5A =	030618.B\030618066.D
CAL5B =	030618.B\030618030.D	CAL5C =	030618.B\030618054.D	CAL5D =	030618.B\030618042.D
CAL5F =	030618.B\030618018.D				

COMPOUND	PEAK	CAL1	CAL2	CAL3	CAL4	CAL5	RT	RT WINDOW	
								FROM	TO
Aroclor 1016	1	3.318	3.317	3.317	3.317	3.318	3.318	3.284	3.352
Aroclor 1016	2	3.513	3.513	3.512	3.512	3.512	3.512	3.478	3.546
Aroclor 1016	3	3.826	3.826	3.826	3.826	3.826	3.826	3.792	3.86
Aroclor 1016	4	3.928	3.929	3.928	3.929	3.928	3.929	3.895	3.963
Aroclor 1016	5	3.995	3.994	3.994	3.995	3.994	3.994	3.96	4.028
Aroclor 1221	1	1.803	1.804	1.804	1.804	1.806	1.804	1.77	1.838
Aroclor 1221	2	2.406	2.407	2.407	2.407	2.407	2.407	2.373	2.441
Aroclor 1221	3	2.687	2.688	2.688	2.688	2.688	2.688	2.654	2.722
Aroclor 1221	4	2.789	2.789	2.79	2.789	2.789	2.789	2.755	2.823
Aroclor 1221	5	2.851	2.851	2.851	2.851	2.851	2.851	2.817	2.885
Aroclor 1242	1	3.318	3.318	3.317	3.318	3.318	3.318	3.284	3.352
Aroclor 1242	2	3.513	3.513	3.513	3.513	3.513	3.513	3.479	3.547
Aroclor 1242	3	3.826	3.826	3.827	3.827	3.827	3.827	3.793	3.861
Aroclor 1242	4	3.928	3.929	3.929	3.93	3.929	3.929	3.895	3.963
Aroclor 1242	5	3.995	3.994	3.995	3.995	3.994	3.995	3.961	4.029
Aroclor 1248	1	4.261	4.261	4.261	4.261	4.261	4.261	4.227	4.295
Aroclor 1248	2	4.546	4.545	4.545	4.545	4.545	4.546	4.512	4.58
Aroclor 1248	3	4.814	4.813	4.813	4.813	4.813	4.812	4.778	4.846
Aroclor 1248	4	4.879	4.878	4.879	4.878	4.879	4.879	4.845	4.913
Aroclor 1248	5	5.037	5.036	5.037	5.038	5.037	5.038	5.004	5.072
Aroclor 1254	1	5.163	5.162	5.163	5.162	5.163	5.162	5.128	5.196
Aroclor 1254	2	5.408	5.407	5.407	5.407	5.407	5.408	5.374	5.442
Aroclor 1254	3	5.51	5.51	5.51	5.509	5.51	5.51	5.476	5.544
Aroclor 1254	4	6.026	6.026	6.026	6.027	6.027	6.027	5.993	6.061
Aroclor 1254	5	6.299	6.299	6.299	6.299	6.299	6.299	6.265	6.333
Aroclor 1260	1	6.3	6.3	6.299	6.3	6.299	6.3	6.266	6.334
Aroclor 1260	2	7.006	7.006	7.005	7.005	7.005	7.006	6.972	7.04
Aroclor 1260	3	7.252	7.252	7.251	7.252	7.251	7.251	7.217	7.285

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

PCB - FORM VI SVOA-2

PCB INITIAL CALIBRATION RETENTION TIME (MULTIPOINT)

Lab Name: Pace Analytical - Green Bay Instrument ID: 40GCSJ GC Column: Col 1 SDG No.: 40151601

Calibration Date(s): 03/06/2018 03/06/2018 Calibration Time(s): 12:07 19:39

LAB FILE ID

CAL1A =	030618.B\030618058.D	CAL1B =	030618.B\030618022.D	CAL1C =	030618.B\030618046.D
CAL1D =	030618.B\030618034.D	CAL1F =	030618.B\030618010.D	CAL2A =	030618.B\030618060.D
CAL2B =	030618.B\030618024.D	CAL2C =	030618.B\030618048.D	CAL2D =	030618.B\030618036.D
CAL2F =	030618.B\030618012.D	CAL3A =	030618.B\030618062.D	CAL3B =	030618.B\030618026.D
CAL3C =	030618.B\030618050.D	CAL3D =	030618.B\030618038.D	CAL3F =	030618.B\030618014.D
CAL4A =	030618.B\030618064.D	CAL4B =	030618.B\030618028.D	CAL4C =	030618.B\030618052.D
CAL4D =	030618.B\030618040.D	CAL4F =	030618.B\030618016.D	CAL5A =	030618.B\030618066.D
CAL5B =	030618.B\030618030.D	CAL5C =	030618.B\030618054.D	CAL5D =	030618.B\030618042.D
CAL5F =	030618.B\030618018.D				

COMPOUND	PEAK	CAL1	CAL2	CAL3	CAL4	CAL5	RT	RT WINDOW	
								FROM	TO
Aroclor 1260	4	7.444	7.443	7.443	7.443	7.443	7.443	7.409	7.477
Aroclor 1260	5	7.902	7.9	7.901	7.901	7.9	7.901	7.867	7.935
Decachlorobiphenyl (S)		8.572	8.572	8.571	8.571	8.571	8.571	8.537	8.605
Tetrachloro-m-xylene		2.485	2.484	2.484	2.484	2.485	2.484	2.45	2.518

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

PCB - FORM VI SVOA-1

PCB INITIAL CALIBRATION DATA (SINGLE POINT)

Lab Name: Pace Analytical - Green Bay Instrument ID: 40GCSJ GC Column: Col 1 SDG No.: 40151601

Calibration Date(s): 03/06/2018 Calibration Time(s): 11:50

LAB FILE ID

CAL3S = 030618.B\030618008.D

COMPOUND	AMOUNT	PEAK	RT	RT WINDOW		RESPONSE FACTOR
				FROM	TO	
Aroclor 1232	.5	1	2.851	2.817	2.885	2500024.00
Aroclor 1232	.5	2	3.513	3.479	3.547	1339782.00
Aroclor 1232	.5	3	3.827	3.793	3.861	3749726.00
Aroclor 1232	.5	4	3.929	3.895	3.963	1606492.00
Aroclor 1232	.5	5	3.995	3.961	4.029	1091876.00

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618008.D
Lab Smp Id: 1232-CAL3S Client Smp ID: 1232-CAL3S
Inj Date : 06-MAR-2018 11:50
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1232-cal3s, 185539:1
Misc Info : 17498,NSAR1232
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 12:39 Cal File: 030618014.D
Als bottle: 5 Calibration Sample, Level: 3
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1232.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.183	64094	46184	0.721	0.24		
1.429	37762	31672	0.839	0.14		
1.535	16372	12525	0.765	0.06		
1.688	12215	10878	0.891	0.04		
1.804	291016	252430	0.867	1.10		
2.011	33236	24341	0.732	0.12		
2.198	82778	71506	0.864	0.31		
2.407	191146	142081	0.743	0.72		
2.485	5780151	5532286	0.957	22.35		
2.688	419741	347465	0.828	1.59		
2.789	391882	367897	0.939	1.48		
2.851	1250012	1144522	0.916	4.74	4 Aroclor 1232	BA
3.028	21736	20469	0.942	0.08		
3.055	85135	77802	0.914	0.32		
3.280	70307	52213	0.743	0.26		
3.319	723460	656407	0.907	2.74		
3.344	391395	363151	0.928	1.48		
3.434	120283	103424	0.860	0.45		
3.514	669892	480513	0.717	2.54	4 Aroclor 1232	BA
3.611	20898	19814	0.948	0.07		
3.698	27644	23229	0.840	0.10		
3.733	173570	179973	1.037	0.65		
3.757	122922	116095	0.944	0.46		
3.827	1874863	1087503	0.580	7.12	4 Aroclor 1232	BA
3.929	803247	685069	0.853	3.05	4 Aroclor 1232	BA
3.995	545938	536580	0.983	2.07	4 Aroclor 1232	BA
4.059	141928	130174	0.917	0.53		
4.137	62125	51170	0.824	0.23		
4.216	393609	374181	0.951	1.49		
4.261	373167	361789	0.970	1.41		
4.294	162815	169097	1.039	0.61		
4.310	204739	187151	0.914	0.77		
4.371	21919	15566	0.710	0.08		
4.418	536617	502916	0.937	2.03		
4.452	418871	298669	0.713	1.59		
4.546	683890	623608	0.912	2.59		
4.618	163410	151749	0.929	0.62		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.681	14650	12034	0.821	0.05	
4.730	43779	30964	0.707	0.16	
4.769	38954	36822	0.945	0.14	
4.814	304752	292334	0.959	1.15	
4.848	526734	522616	0.992	2.00	
4.880	691689	534310	0.772	2.62	
4.958	81081	59242	0.731	0.30	
5.038	573825	543275	0.947	2.17	
5.099	111947	85321	0.762	0.42	
5.163	131671	111235	0.845	0.50	
5.213	95848	80980	0.845	0.36	
5.307	24699	21694	0.878	0.09	
5.363	86500	72230	0.835	0.32	
5.408	138303	112479	0.813	0.52	
5.447	83276	72672	0.873	0.31	
5.492	30778	35064	1.139	0.11	
5.511	193150	166298	0.861	0.73	
5.605	68006	59493	0.875	0.25	
5.676	16334	15605	0.955	0.06	
5.708	52967	20705	0.391	0.20	
5.793	143283	113467	0.792	0.54	
5.888	27037	15011	0.555	0.10	
6.026	118842	104296	0.878	0.45	
6.056	34574	29464	0.852	0.13	
6.171	24413	14466	0.593	0.09	
6.302	43529	35888	0.824	0.16	
6.447	25328	16811	0.664	0.09	
6.511	16971	13602	0.801	0.06	
7.005	21863	21930	1.003	0.08	
7.129	17051	15791	0.926	0.06	
7.254	11045	10898	0.987	0.04	
7.728	22228	20096	0.904	0.08	
7.779	25468	15517	0.609	0.09	
8.277	43793	37397	0.854	0.16	
8.379	30735	21976	0.715	0.11	
8.573	4915144	4232662	0.861	18.67	
8.944	38971	23332	0.599	0.14	
9.470	37385	22307	0.597	0.14	
9.964	32743	21879	0.668	0.12	
	26324135	22924262		100.000	

Total unknown % area = 80

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618008.D
Lab Smp Id: 1232-CAL3S Client Smp ID: 1232-CAL3S
Inj Date : 06-MAR-2018 11:50
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1232-cal3s, 185539:1
Misc Info : 17498,NSAR1232
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 12:39 Cal File: 030618014.D
Als bottle: 5 Calibration Sample, Level: 3
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1232.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
4 Aroclor 1232					CAS #: 1114-16-5			
2.851	2.851	0.000	1250012	0.50000	0.50	0.00- 20.00	100.00(M)	BA
3.513	3.513	0.000	669891	0.50000	0.50	0.00- 20.00	53.59	BA
3.827	3.827	0.000	1874863	0.50000	0.50	0.00- 20.00	149.99	BA
3.929	3.929	0.000	803246	0.50000	0.50	0.00- 20.00	64.26	BA
3.995	3.995	0.000	545938	0.50000	0.50	0.00- 20.00	43.67	BA
Average of Peak Amounts =					0.50000			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618008.D

Date : 06-MAR-2018 11:50

Client ID: 1232-CAL3S

Instrument: 40GCSJ.i

Sample Info: 1232-CAL3S, 185539:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618008.D (ManInt: NSAR1232.sub Orig: NSAR1232.sub)
Injection Date: 06-MAR-2018 11:50 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1232-CAL3S





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1232	1234986	1250012	BA
Aroclor 1232	1234986	1250012	BA
Aroclor 1232	1234986	1250012	BA
Aroclor 1232	1234986	1250012	BA
Aroclor 1232	1234986	1250012	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618010.D
Lab Smp Id: 1221-CAL1F Client Smp ID: 1221-CAL1F
Inj Date : 06-MAR-2018 12:07
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1221-cal1f, 188138:1
Misc Info : 17498,NSAR1221
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 13:44 Cal File: 030618022.D
Als bottle: 6 Calibration Sample, Level: 1
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1221.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.180	13612	11288	0.829	0.38		
1.804	116885	98810	0.845	3.31	1 Aroclor 1221	BA
2.197	36039	30499	0.846	1.02		
2.407	61218	44733	0.731	1.73	1 Aroclor 1221	BA
2.485	1172586	1134895	0.968	33.37		
2.688	150512	119908	0.797	4.27	1 Aroclor 1221	BA
2.790	118144	107550	0.910	3.35	1 Aroclor 1221	BA
2.852	349220	309600	0.887	9.90	1 Aroclor 1221	BA
3.279	28274	21896	0.774	0.80		
3.325	85123	48119	0.565	2.41		
3.380	10358	5756	0.556	0.29		
3.436	15926	10893	0.684	0.45		
3.513	38645	25267	0.654	1.09		
3.733	9954	10584	1.063	0.28		
3.817	38159	24082	0.631	1.08		
3.928	37088	28113	0.758	1.05		
3.996	22534	20384	0.905	0.63		
4.260	9635	9376	0.973	0.27		
4.546	10833	12359	1.141	0.30		
4.881	15313	10912	0.713	0.43		
5.040	11422	9410	0.824	0.32		
8.278	12362	11919	0.964	0.35		
8.380	10282	8381	0.815	0.29		
8.572	1127356	960701	0.852	31.98		
8.936	10000	8119	0.812	0.28		
9.466	13059	9537	0.730	0.37		
	3524537	3093091		100.000		

Total unknown % area = 77

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618010.D
Lab Smp Id: 1221-CAL1F Client Smp ID: 1221-CAL1F
Inj Date : 06-MAR-2018 12:07
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1221-cal1f, 188138:1
Misc Info : 17498,NSAR1221
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 13:44 Cal File: 030618022.D
Als bottle: 6 Calibration Sample, Level: 1
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1221.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221					CAS #: 11104-28-2			
1.803	1.804	-0.001	116884	0.10000	0.11	0.00- 20.00	100.00(M)	BA
2.406	2.407	-0.001	61218	0.10000	0.11	0.00- 20.00	52.38	BA
2.687	2.688	-0.001	150511	0.10000	0.11	0.00- 20.00	128.77	BA
2.789	2.789	0.000	118144	0.10000	0.11	0.00- 20.00	101.08	BA
2.851	2.851	0.000	349219	0.10000	0.11	0.00- 20.00	298.77	BA
Average of Peak Amounts =					0.11000			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618010.D
Date : 06-MAR-2018 12:07
Client ID: 1221-CAL1F
Sample Info: 1221-CAL1F, 188138:1

Instrument: 40GCSJ.i
Operator: BLM
Column phase: ZB-1ms
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618010.D (ManInt: NSAR1221.sub Orig: NSAR1221.sub)
Injection Date: 06-MAR-2018 12:07 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1221-CAL1F





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1221	111828	116884	BA
Aroclor 1221	111828	116884	BA
Aroclor 1221	111828	116884	BA
Aroclor 1221	111828	116884	BA
Aroclor 1221	111828	116884	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618012.D
Lab Smp Id: 1221-CAL2F Client Smp ID: 1221-CAL2F
Inj Date : 06-MAR-2018 12:23
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1221-cal2f, 188139:1
Misc Info : 17498,NSAR1221
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 12:23 Cal File: 030618012.D
Als bottle: 7 Calibration Sample, Level: 2
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1221.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.185	21110	15992	0.758	0.29		
1.432	19890	14202	0.714	0.27		
1.804	222151	196763	0.886	3.10	1 Aroclor 1221	BA
2.006	14074	10640	0.756	0.19		
2.198	66696	59685	0.895	0.93		
2.408	117491	84755	0.721	1.64	1 Aroclor 1221	BA
2.485	2389575	2280739	0.954	33.51		
2.688	289327	237824	0.822	4.03	1 Aroclor 1221	BA
2.790	230052	213818	0.929	3.21	1 Aroclor 1221	BA
2.852	694108	629495	0.907	9.69	1 Aroclor 1221	BA
3.055	24441	15851	0.649	0.34		
3.283	52341	38105	0.728	0.73		
3.322	162359	92875	0.572	2.26		
3.433	17768	16060	0.904	0.24		
3.514	68265	44434	0.651	0.95		
3.610	18983	11562	0.609	0.26		
3.733	17041	17575	1.031	0.23		
3.755	20072	17438	0.869	0.28		
3.827	142441	84122	0.591	1.98		
3.930	69577	57831	0.831	0.97		
3.996	47705	40519	0.849	0.66		
4.216	23635	18011	0.762	0.33		
4.261	14899	13886	0.932	0.20		
4.418	12729	14384	1.130	0.17		
4.547	21625	20493	0.948	0.30		
4.618	11602	9168	0.790	0.16		
4.848	13653	14682	1.075	0.19		
4.880	20209	15214	0.753	0.28		
5.040	16835	15081	0.896	0.23		
7.133	16511	8656	0.524	0.23		
7.728	11846	11710	0.989	0.16		
8.278	23007	17531	0.762	0.32		
8.378	9408	9279	0.986	0.13		
8.573	2192346	1884280	0.859	30.61		
8.940	11831	13025	1.101	0.16		
9.473	27497	16611	0.604	0.38		
9.964	28645	19011	0.664	0.39		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
	7161743	6281307		100.000	

Total unknown % area = 78

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618012.D
Lab Smp Id: 1221-CAL2F Client Smp ID: 1221-CAL2F
Inj Date : 06-MAR-2018 12:23
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1221-cal2f, 188139:1
Misc Info : 17498,NSAR1221
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 12:23 Cal File: 030618012.D
Als bottle: 7 Calibration Sample, Level: 2
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1221.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221					CAS #: 11104-28-2			
1.804	1.804	0.000	222151	0.20000	0.21	0.00- 20.00	100.00(M)	BA
2.407	2.407	0.000	117490	0.20000	0.21	0.00- 20.00	52.89	BA
2.688	2.688	0.000	289327	0.20000	0.21	0.00- 20.00	130.24	BA
2.789	2.789	0.000	230051	0.20000	0.21	0.00- 20.00	103.56	BA
2.851	2.851	0.000	694107	0.20000	0.21	0.00- 20.00	312.45	BA
Average of Peak Amounts =					0.21000			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618012.D

Date : 06-MAR-2018 12:23

Client ID: 1221-CAL2F

Instrument: 40GCSJ.i

Sample Info: 1221-CAL2F, 188139:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618012.D (ManInt: NSAR1221.sub Orig: NSAR1221.sub)
Injection Date: 06-MAR-2018 12:23 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1221-CAL2F





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1221	219680	222151	BA
Aroclor 1221	219680	222151	BA
Aroclor 1221	219680	222151	BA
Aroclor 1221	219680	222151	BA
Aroclor 1221	219680	222151	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618014.D
Lab Smp Id: 1221-CAL3F Client Smp ID: 1221-CAL3F
Inj Date : 06-MAR-2018 12:39
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1221-cal3f, 185538:1
Misc Info : 17498,NSAR1221
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 12:39 Cal File: 030618014.D
Als bottle: 8 Calibration Sample, Level: 3
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1221.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.184	46886	34819	0.743	0.27		
1.429	37682	34000	0.902	0.22		
1.540	18475	13916	0.753	0.11		
1.686	12215	9839	0.805	0.07		
1.805	490239	441873	0.901	2.92	1 Aroclor 1221	BA
2.010	34778	21164	0.609	0.20		
2.197	148893	131934	0.886	0.88		
2.408	263311	192077	0.729	1.56	1 Aroclor 1221	BA
2.485	5899001	5597678	0.949	35.41		
2.688	658124	555463	0.844	3.92	1 Aroclor 1221	BA
2.790	519264	501204	0.965	3.09	1 Aroclor 1221	BA
2.851	1583519	1422292	0.898	9.43	1 Aroclor 1221	BA
3.030	16777	16577	0.988	0.09		
3.055	39004	30894	0.792	0.23		
3.213	9910	8832	0.891	0.05		
3.281	111839	84209	0.753	0.66		
3.325	365423	215871	0.591	2.17		
3.437	35737	30648	0.858	0.21		
3.514	140459	101998	0.726	0.83		
3.608	20183	17591	0.872	0.12		
3.732	32639	35941	1.101	0.19		
3.757	24903	28786	1.156	0.14		
3.827	309924	194764	0.628	1.84		
3.930	154650	131637	0.851	0.92		
3.996	105091	94784	0.902	0.62		
4.058	20977	14811	0.706	0.12		
4.215	34584	31722	0.917	0.20		
4.262	26916	30192	1.122	0.16		
4.292	16244	12494	0.769	0.09		
4.374	10462	9206	0.880	0.06		
4.418	41223	41705	1.012	0.24		
4.452	49711	28399	0.571	0.29		
4.545	54290	46656	0.859	0.32		
4.616	14003	11847	0.846	0.08		
4.812	21709	20229	0.932	0.12		
4.848	35650	33592	0.942	0.21		
4.882	55103	37630	0.683	0.32		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.039	36880	30922	0.838	0.21	
5.164	10614	9077	0.855	0.06	
5.703	10136	9806	0.967	0.06	
5.794	11063	9888	0.894	0.06	
6.450	15885	11521	0.725	0.09	
7.126	18110	12799	0.707	0.10	
7.730	23581	20558	0.872	0.14	
7.777	18011	13331	0.740	0.10	
7.801	14860	14996	1.009	0.08	
8.276	48701	37600	0.772	0.29	
8.381	26483	18054	0.682	0.15	
8.573	5020450	4347562	0.866	29.92	
8.945	30595	17805	0.582	0.18	
9.086	11785	9130	0.775	0.07	
9.963	22483	17745	0.789	0.13	
	16779431	14848068		100.000	

Total unknown % area = 79

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618014.D
Lab Smp Id: 1221-CAL3F Client Smp ID: 1221-CAL3F
Inj Date : 06-MAR-2018 12:39
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1221-cal3f, 185538:1
Misc Info : 17498,NSAR1221
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 12:39 Cal File: 030618014.D
Als bottle: 8 Calibration Sample, Level: 3
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1221.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221					CAS #: 11104-28-2			
1.804	1.804	0.000	490239	0.50000	0.47	0.00- 20.00	100.00(M)	BA
2.407	2.407	0.000	263311	0.50000	0.48	0.00- 20.00	53.71	BA
2.688	2.688	0.000	658123	0.50000	0.48	0.00- 20.00	134.25	BA
2.790	2.789	0.001	519263	0.50000	0.47	0.00- 20.00	105.92	BA
2.851	2.851	0.000	1583519	0.50000	0.49	0.00- 20.00	323.01	BA
Average of Peak Amounts =					0.47800			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618014.D

Date : 06-MAR-2018 12:39

Client ID: 1221-CAL3F

Instrument: 40GCSJ.i

Sample Info: 1221-CAL3F, 185538:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



40151601

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618014.D (ManInt: NSAR1221.sub Orig: NSAR1221.sub)
Injection Date: 06-MAR-2018 12:39 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1221-CAL3F





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1221	483261	490239	BA
Aroclor 1221	483261	490239	BA
Aroclor 1221	483261	490239	BA
Aroclor 1221	483261	490239	BA
Aroclor 1221	483261	490239	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618016.D
Lab Smp Id: 1221-CAL4F Client Smp ID: 1221-CAL4F
Inj Date : 06-MAR-2018 12:55
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1221-cal4f, 188140:1
Misc Info : 17498,NSAR1221
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 12:55 Cal File: 030618016.D
Als bottle: 9 Calibration Sample, Level: 4
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1221.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.181	15077	12970	0.860	0.04		
1.430	62013	58411	0.942	0.20		
1.540	29523	27434	0.929	0.09		
1.682	10134	8214	0.811	0.03		
1.804	785551	700381	0.892	2.59	1 Aroclor 1221	BA
2.162	17608	16089	0.914	0.05		
2.197	232212	214361	0.923	0.76		
2.408	409142	300406	0.734	1.35	1 Aroclor 1221	BA
2.485	11502194	11125302	0.967	38.28		
2.689	1067134	893347	0.837	3.52	1 Aroclor 1221	BA
2.789	853931	817692	0.958	2.82	1 Aroclor 1221	BA
2.851	2470730	2197877	0.890	8.16	1 Aroclor 1221	BA
3.053	58135	49830	0.857	0.19		
3.280	178171	134368	0.754	0.58		
3.325	581244	339314	0.584	1.92		
3.436	58621	50800	0.867	0.19		
3.513	227025	162531	0.716	0.75		
3.611	30500	29323	0.961	0.10		
3.696	15508	16983	1.095	0.05		
3.732	59581	61761	1.037	0.19		
3.757	63343	56081	0.885	0.20		
3.827	512825	311364	0.607	1.69		
3.929	244641	209892	0.858	0.80		
3.995	150703	143674	0.953	0.49		
4.059	21801	19537	0.896	0.07		
4.134	11088	10051	0.906	0.03		
4.215	55304	52296	0.946	0.18		
4.260	51932	50173	0.966	0.17		
4.291	25327	22345	0.882	0.08		
4.311	24902	25307	1.016	0.08		
4.418	49634	53105	1.070	0.16		
4.443	38631	29989	0.776	0.12		
4.547	83853	73357	0.875	0.27		
4.617	26087	19117	0.733	0.08		
4.814	30317	32424	1.069	0.10		
4.849	51073	52237	1.023	0.16		
4.880	70316	49193	0.700	0.23		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.038	59957	50669	0.845	0.19	
5.103	12689	10469	0.825	0.04	
5.163	17456	13811	0.791	0.05	
5.216	17786	12808	0.720	0.05	
5.360	16391	9844	0.601	0.05	
5.406	16004	13419	0.838	0.05	
5.445	10313	9432	0.915	0.03	
5.512	20459	14813	0.724	0.06	
5.793	13847	11803	0.852	0.04	
7.123	27333	14055	0.514	0.09	
7.694	9264	9447	1.020	0.03	
7.729	46203	36902	0.799	0.15	
7.800	22162	20669	0.933	0.07	
8.276	84580	73604	0.870	0.27	
8.381	14600	14602	1.000	0.04	
8.573	9619023	8259708	0.859	31.79	
8.941	10654	8547	0.802	0.03	
9.090	21504	11989	0.558	0.07	
9.466	18951	12488	0.659	0.06	
9.960	21699	14589	0.672	0.07	
	30256687	27051204		100.000	

Total unknown % area = 82

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618016.D
Lab Smp Id: 1221-CAL4F Client Smp ID: 1221-CAL4F
Inj Date : 06-MAR-2018 12:55
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1221-cal4f, 188140:1
Misc Info : 17498,NSAR1221
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 12:55 Cal File: 030618016.D
Als bottle: 9 Calibration Sample, Level: 4
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1221.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221					CAS #: 11104-28-2			
1.804	1.804	0.000	785550	0.80000	0.75	0.00- 20.00	100.00(M)	BA
2.407	2.407	0.000	409142	0.80000	0.75	0.00- 20.00	52.08	BA
2.688	2.688	0.000	1067134	0.80000	0.77	0.00- 20.00	135.85	BA
2.789	2.789	0.000	853931	0.80000	0.77	0.00- 20.00	108.70	BA
2.851	2.851	0.000	2470729	0.80000	0.76	0.00- 20.00	314.52	BA
Average of Peak Amounts =					0.76000			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618016.D

Date : 06-MAR-2018 12:55

Client ID: 1221-CAL4F

Instrument: 40GCSJ.i

Sample Info: 1221-CAL4F, 188140:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618016.D (ManInt: NSAR1221.sub Orig: NSAR1221.sub)
Injection Date: 06-MAR-2018 12:55 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1221-CAL4F





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1221	780911	785550	BA
Aroclor 1221	780911	785550	BA
Aroclor 1221	780911	785550	BA
Aroclor 1221	780911	785550	BA
Aroclor 1221	780911	785550	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618018.D
Lab Smp Id: 1221-CAL5F Client Smp ID: 1221-CAL5F
Inj Date : 06-MAR-2018 13:11
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1221-cal5f, 188141:1
Misc Info : 17498,NSAR1221
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 13:11 Cal File: 030618018.D
Als bottle: 10 Calibration Sample, Level: 5
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1221.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.429	91782	83949	0.915	0.21		
1.539	45153	34415	0.762	0.10		
1.806	964273	851004	0.883	2.23	1 Aroclor 1221	BA
2.014	11838	7838	0.662	0.02		
2.160	16492	15827	0.960	0.03		
2.198	286599	255763	0.892	0.66		
2.408	501535	366636	0.731	1.16	1 Aroclor 1221	BA
2.486	17378255	16694368	0.961	40.52		
2.688	1314776	1093955	0.832	3.04	1 Aroclor 1221	BA
2.790	1091213	1024965	0.939	2.52	1 Aroclor 1221	BA
2.851	3058815	2705599	0.885	7.08	1 Aroclor 1221	BA
3.053	74407	64323	0.864	0.17		
3.217	18486	13382	0.724	0.04		
3.281	228692	172494	0.754	0.52		
3.325	747509	424295	0.568	1.73		
3.436	77500	63495	0.819	0.17		
3.513	280490	203651	0.726	0.64		
3.612	49578	44845	0.905	0.11		
3.701	23500	20479	0.871	0.05		
3.734	76655	76322	0.996	0.17		
3.757	81462	74126	0.910	0.18		
3.827	635271	385105	0.606	1.47		
3.929	312825	252728	0.808	0.72		
3.995	186590	175263	0.939	0.43		
4.062	31750	23556	0.742	0.07		
4.138	17974	12098	0.673	0.04		
4.217	72031	63997	0.888	0.16		
4.261	66789	61014	0.914	0.15		
4.293	54107	32297	0.597	0.12		
4.418	84432	78861	0.934	0.19		
4.451	94426	56177	0.595	0.21		
4.546	109930	92350	0.840	0.25		
4.618	33492	25022	0.747	0.07		
4.815	42406	39475	0.931	0.09		
4.849	68687	65653	0.956	0.15		
4.881	95718	65369	0.683	0.22		
4.957	18933	12531	0.662	0.04		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.037	71300	62002	0.870	0.16	
5.101	11578	12947	1.118	0.02	
5.162	15068	13787	0.915	0.03	
5.208	16758	12735	0.760	0.03	
5.363	13212	11132	0.843	0.03	
5.408	17007	15696	0.923	0.03	
5.445	14075	10385	0.738	0.03	
5.512	25865	18240	0.705	0.05	
5.792	16968	13288	0.783	0.03	
6.028	13891	11907	0.857	0.03	
7.125	30918	19205	0.621	0.07	
7.692	13631	9457	0.694	0.03	
7.729	65348	56318	0.862	0.15	
7.802	42671	34037	0.798	0.09	
8.277	126824	104866	0.827	0.29	
8.379	18877	14713	0.779	0.04	
8.573	14268732	12292631	0.862	33.03	
8.940	12885	8913	0.692	0.02	
9.089	24129	17399	0.721	0.05	
9.468	12972	8897	0.686	0.03	
9.969	15122	10700	0.708	0.03	
	43192203	38492482		100.000	

Total unknown % area = 84

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618018.D
Lab Smp Id: 1221-CAL5F Client Smp ID: 1221-CAL5F
Inj Date : 06-MAR-2018 13:11
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1221-cal5f, 188141:1
Misc Info : 17498,NSAR1221
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 13:11 Cal File: 030618018.D
Als bottle: 10 Calibration Sample, Level: 5
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1221.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221					CAS #: 11104-28-2			
1.806	1.804	0.002	964273	1.00000	0.93	0.00- 20.00	100.00(M)	BA
2.407	2.407	0.000	501535	1.00000	0.92	0.00- 20.00	52.01	BA
2.688	2.688	0.000	1314775	1.00000	0.95	0.00- 20.00	136.35	BA
2.789	2.789	0.000	1091213	1.00000	0.99	0.00- 20.00	113.16	BA
2.851	2.851	0.000	3058814	1.00000	0.94	0.00- 20.00	317.21	BA
Average of Peak Amounts =					0.94600			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618018.D

Date : 06-MAR-2018 13:11

Client ID: 1221-CAL5F

Instrument: 40GCSJ.i

Sample Info: 1221-CAL5F, 188141:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618018.D (ManInt: NSAR1221.sub Orig: NSAR1221.sub)
Injection Date: 06-MAR-2018 13:11 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1221-CAL5F





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1221	958762	964273	BA
Aroclor 1221	958762	964273	BA
Aroclor 1221	958762	964273	BA
Aroclor 1221	958762	964273	BA
Aroclor 1221	958762	964273	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618022.D
Lab Smp Id: 1242-CAL1B Client Smp ID: 1242-CAL1B
Inj Date : 06-MAR-2018 13:44
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-cal1b, 188132:1
Misc Info : 17498,NSAR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 13:44 Cal File: 030618022.D
Als bottle: 12 Calibration Sample, Level: 1
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1242.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.183	48353	40712	0.842	0.55		
1.685	15584	12605	0.809	0.18		
2.012	33033	24593	0.745	0.38		
2.403	26701	21656	0.811	0.30		
2.485	1268339	1220330	0.962	14.93		
2.688	45903	34473	0.751	0.53		
2.791	63166	56222	0.890	0.73		
2.851	204981	187245	0.913	2.37		
3.026	22317	19810	0.888	0.25		
3.055	37376	31867	0.853	0.43		
3.318	277256	257375	0.928	3.21	3 Aroclor 1242	BA
3.345	154841	150344	0.971	1.79		
3.435	49043	39609	0.808	0.56		
3.514	273951	193620	0.707	3.17	3 Aroclor 1242	BA
3.733	69544	68947	0.991	0.80		
3.757	45882	42206	0.920	0.53		
3.827	769590	449825	0.584	8.91	3 Aroclor 1242	BA
3.929	328754	270344	0.822	3.80	3 Aroclor 1242	BA
3.996	224743	217453	0.968	2.60	3 Aroclor 1242	BA
4.059	72772	62645	0.861	0.84		
4.135	32124	25139	0.783	0.37		
4.215	184258	174093	0.945	2.13		
4.262	175232	166947	0.953	2.02		
4.295	76283	77062	1.010	0.88		
4.310	94862	88633	0.934	1.09		
4.417	242650	230875	0.951	2.80		
4.453	185260	138108	0.745	2.14		
4.547	310173	288905	0.931	3.59		
4.617	78937	69004	0.874	0.91		
4.726	21360	16230	0.760	0.24		
4.770	22579	18643	0.826	0.26		
4.813	144508	140479	0.972	1.67		
4.848	244505	231221	0.946	2.83		
4.880	333816	245294	0.735	3.86		
4.958	38640	32017	0.829	0.44		
5.039	262014	232176	0.886	3.03		
5.099	55213	47400	0.858	0.63		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.164	71096	60506	0.851	0.82	
5.213	54601	49311	0.903	0.63	
5.310	15720	13003	0.827	0.18	
5.363	43486	39659	0.912	0.50	
5.407	75906	62166	0.819	0.87	
5.447	47603	40647	0.854	0.55	
5.511	110554	91386	0.827	1.28	
5.605	38987	31752	0.814	0.45	
5.702	21923	17168	0.783	0.25	
5.793	90266	69831	0.774	1.04	
5.888	18665	11476	0.615	0.21	
6.027	83046	57198	0.689	0.96	
6.301	23663	20357	0.860	0.27	
6.450	25108	16371	0.652	0.29	
7.135	16747	14415	0.861	0.19	
7.722	10233	8765	0.857	0.11	
7.774	20981	14974	0.714	0.24	
8.276	10004	11485	1.148	0.11	
8.381	26018	18613	0.715	0.30	
8.572	1221508	1021745	0.836	14.14	
8.943	31681	20256	0.639	0.36	
9.468	21143	16317	0.772	0.24	
9.964	23088	17512	0.758	0.26	
	8636569	7349020		100.000	

Total unknown % area = 78

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618022.D
Lab Smp Id: 1242-CAL1B Client Smp ID: 1242-CAL1B
Inj Date : 06-MAR-2018 13:44
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-cal1b, 188132:1
Misc Info : 17498,NSAR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 13:44 Cal File: 030618022.D
Als bottle: 12 Calibration Sample, Level: 1
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1242.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
3 Aroclor 1242					CAS #: 53469-21-9			
3.318	3.318	0.000	277256	0.10000	0.11	0.00- 20.00	100.00(M)	BA
3.513	3.513	0.000	273950	0.10000	0.11	0.00- 20.00	98.81	BA
3.826	3.827	-0.001	769589	0.10000	0.11	0.00- 20.00	277.57	BA
3.928	3.929	-0.001	328754	0.10000	0.11	0.00- 20.00	118.57	BA
3.995	3.995	0.000	224742	0.10000	0.11	0.00- 20.00	81.06	BA
Average of Peak Amounts =					0.11000			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618022.D

Date : 06-MAR-2018 13:44

Client ID: 1242-CAL1B

Instrument: 40GCSJ.i

Sample Info: 1242-CAL1B, 188132:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618022.D (ManInt: NSAR1242.sub Orig: NSAR1242.sub)
Injection Date: 06-MAR-2018 13:44 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1242-CAL1B





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1242	273881	277256	BA
Aroclor 1242	273881	277256	BA
Aroclor 1242	273881	277256	BA
Aroclor 1242	273881	277256	BA
Aroclor 1242	273881	277256	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618024.D
Lab Smp Id: 1242-CAL2B Client Smp ID: 1242-CAL2B
Inj Date : 06-MAR-2018 14:00
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-cal2b, 185543:1
Misc Info : 17498,NSAR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 14:00 Cal File: 030618024.D
Als bottle: 13 Calibration Sample, Level: 2
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1242.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.184	14220	11208	0.788	0.08		
1.430	13376	14873	1.112	0.08		
1.537	12556	8533	0.680	0.07		
1.684	10310	8364	0.811	0.06		
1.803	9301	8955	0.963	0.05		
2.404	53409	39369	0.737	0.33		
2.484	2424898	2340795	0.965	15.45		
2.685	71092	59059	0.831	0.44		
2.789	112696	106870	0.948	0.70		
2.851	383763	360025	0.938	2.38		
3.054	60192	54572	0.907	0.37		
3.319	537600	506479	0.942	3.34	3 Aroclor 1242	BA
3.344	300840	286923	0.954	1.87		
3.434	87992	78217	0.889	0.54		
3.514	520152	378008	0.727	3.23	3 Aroclor 1242	BA
3.607	13378	11603	0.867	0.08		
3.696	19908	16213	0.814	0.12		
3.732	129346	131457	1.016	0.80		
3.756	89938	85088	0.946	0.56		
3.826	1494620	886299	0.593	9.30	3 Aroclor 1242	BA
3.929	633849	547272	0.863	3.94	3 Aroclor 1242	BA
3.994	415519	392907	0.946	2.58	3 Aroclor 1242	BA
4.059	126568	116783	0.923	0.78		
4.137	53777	45505	0.846	0.33		
4.216	349073	338096	0.969	2.17		
4.261	336841	321968	0.956	2.09		
4.293	139012	148834	1.071	0.86		
4.310	185086	169258	0.914	1.15		
4.372	13133	10098	0.769	0.08		
4.418	470013	436944	0.930	2.92		
4.452	357597	265664	0.743	2.22		
4.546	610105	562073	0.921	3.79		
4.618	143247	130258	0.909	0.89		
4.727	34627	28696	0.829	0.21		
4.768	38705	34109	0.881	0.24		
4.814	277192	264316	0.954	1.72		
4.849	476923	458153	0.961	2.97		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.879	613481	454847	0.741	3.82	
4.960	74104	57100	0.771	0.46	
5.038	507891	458844	0.903	3.16	
5.100	100349	88837	0.885	0.62	
5.164	134684	119177	0.885	0.83	
5.213	98854	90627	0.917	0.61	
5.283	12070	11231	0.930	0.07	
5.310	24719	19674	0.796	0.15	
5.363	85050	75533	0.888	0.52	
5.408	146635	119693	0.816	0.91	
5.447	90264	80310	0.890	0.56	
5.510	230506	174139	0.755	1.43	
5.605	75587	61685	0.816	0.47	
5.716	16948	14359	0.847	0.10	
5.741	23073	20242	0.877	0.14	
5.793	169665	128009	0.754	1.05	
5.887	31657	16739	0.529	0.19	
6.025	153536	114654	0.747	0.95	
6.171	21729	13262	0.610	0.13	
6.302	46462	36756	0.791	0.28	
6.534	18985	15865	0.836	0.11	
6.803	15791	11632	0.737	0.09	
7.725	15675	13882	0.886	0.09	
8.279	20164	17879	0.887	0.12	
8.380	27215	14895	0.547	0.16	
8.573	2205636	1873304	0.849	13.73	
8.944	29668	17498	0.590	0.18	
9.464	14497	11336	0.782	0.09	
9.964	31291	22255	0.711	0.19	
	16057042	13818108		100.000	

Total unknown % area = 78

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618024.D
Lab Smp Id: 1242-CAL2B Client Smp ID: 1242-CAL2B
Inj Date : 06-MAR-2018 14:00
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-cal2b, 185543:1
Misc Info : 17498,NSAR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 14:00 Cal File: 030618024.D
Als bottle: 13 Calibration Sample, Level: 2
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1242.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
3 Aroclor 1242					CAS #: 53469-21-9			
3.318	3.318	0.000	537600	0.20000	0.21	0.00- 20.00	100.00(M)	BA
3.513	3.513	0.000	520151	0.20000	0.21	0.00- 20.00	96.75	BA
3.826	3.827	-0.001	1494620	0.20000	0.21	0.00- 20.00	278.02	BA
3.929	3.929	0.000	633849	0.20000	0.21	0.00- 20.00	117.90	BA
3.994	3.995	-0.001	415519	0.20000	0.20	0.00- 20.00	77.29	BA
Average of Peak Amounts =					0.20800			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618024.D
Date : 06-MAR-2018 14:00
Client ID: 1242-CAL2B
Sample Info: 1242-CAL2B, 185543:1

Instrument: 40GCSJ.i
Operator: BLM
Column phase: ZB-1ms
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618024.D (ManInt: NSAR1242.sub Orig: NSAR1242.sub)
Injection Date: 06-MAR-2018 14:00 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1242-CAL2B





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1242	529987	537600	BA
Aroclor 1242	529987	537600	BA
Aroclor 1242	529987	537600	BA
Aroclor 1242	529987	537600	BA
Aroclor 1242	529987	537600	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618026.D
Lab Smp Id: 1242-CAL3B Client Smp ID: 1242-CAL3B
Inj Date : 06-MAR-2018 14:16
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-cal3b, 185544:1
Misc Info : 17498,NSAR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 14:16 Cal File: 030618026.D
Als bottle: 14 Calibration Sample, Level: 3
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1242.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.182	62004	50847	0.820	0.16		
1.431	37460	32055	0.856	0.09		
1.541	14316	12367	0.864	0.03		
1.687	15486	12743	0.823	0.04		
1.807	18746	14208	0.758	0.04		
2.011	43433	33553	0.773	0.11		
2.406	117538	91845	0.781	0.30		
2.485	5892252	5641423	0.957	15.80		
2.688	168721	139553	0.827	0.44		
2.789	263881	249457	0.945	0.68		
2.851	912957	864348	0.947	2.38		
3.032	23465	23649	1.008	0.06		
3.054	147736	128585	0.870	0.38		
3.278	22664	16866	0.744	0.05		
3.318	1252566	1174122	0.937	3.27	3 Aroclor 1242	BA
3.344	702584	667750	0.950	1.83		
3.435	195742	182887	0.934	0.51		
3.514	1253174	902840	0.720	3.27	3 Aroclor 1242	BA
3.610	20808	19442	0.934	0.05		
3.698	36264	35190	0.970	0.09		
3.733	310548	311302	1.002	0.81		
3.756	210733	195877	0.930	0.55		
3.827	3485663	2059745	0.591	9.11	3 Aroclor 1242	BA
3.929	1508994	1248466	0.827	3.94	3 Aroclor 1242	BA
3.995	1031466	974966	0.945	2.69	3 Aroclor 1242	BA
4.059	297827	269246	0.904	0.77		
4.137	124808	105679	0.847	0.32		
4.215	835282	795778	0.953	2.18		
4.262	800812	762657	0.952	2.09		
4.294	335573	342292	1.020	0.87		
4.310	418203	390877	0.935	1.09		
4.366	27246	20902	0.767	0.07		
4.418	1095729	1053734	0.962	2.86		
4.452	872734	643030	0.737	2.28		
4.546	1429253	1334739	0.934	3.73		
4.617	329513	299311	0.908	0.86		
4.679	25919	20313	0.784	0.06		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.728	75071	61492	0.819	0.19	
4.770	78410	75801	0.967	0.20	
4.814	666858	654940	0.982	1.74	
4.848	1139990	1092076	0.958	2.97	
4.879	1497562	1126464	0.752	3.91	
4.959	169941	133168	0.784	0.44	
5.038	1220947	1147297	0.940	3.19	
5.099	262922	203724	0.775	0.68	
5.163	307382	272531	0.887	0.80	
5.212	234571	204654	0.872	0.61	
5.286	17325	18279	1.055	0.04	
5.312	55833	44258	0.793	0.14	
5.365	200636	172920	0.862	0.52	
5.408	328555	284252	0.865	0.85	
5.447	222890	182949	0.821	0.58	
5.510	538895	423244	0.785	1.40	
5.605	171765	141640	0.825	0.44	
5.717	55818	31435	0.563	0.14	
5.742	42015	41544	0.989	0.10	
5.793	399108	317536	0.796	1.04	
5.891	59724	37531	0.628	0.15	
5.988	12622	14402	1.141	0.03	
6.025	317230	272827	0.860	0.82	
6.056	62897	52017	0.827	0.16	
6.173	35886	23021	0.641	0.09	
6.228	35394	17536	0.495	0.09	
6.300	107967	90981	0.843	0.28	
6.338	25191	21116	0.838	0.06	
6.373	22038	15834	0.718	0.05	
6.444	13051	11257	0.863	0.03	
6.540	47204	32635	0.691	0.12	
6.803	23591	20139	0.854	0.06	
7.005	15270	10968	0.718	0.03	
7.133	35322	22100	0.626	0.09	
7.250	13054	10560	0.809	0.03	
7.727	20164	18616	0.923	0.05	
7.782	26691	22023	0.825	0.06	
7.800	17830	16488	0.925	0.04	
8.277	43520	38062	0.875	0.11	
8.377	40259	27272	0.677	0.10	
8.573	5121279	4417666	0.863	13.38	
8.941	41704	28161	0.675	0.10	
9.469	46040	28501	0.619	0.12	
9.966	42365	27909	0.659	0.11	
	38254887	33030470		100.000	

Total unknown % area = 78

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618026.D
Lab Smp Id: 1242-CAL3B Client Smp ID: 1242-CAL3B
Inj Date : 06-MAR-2018 14:16
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-cal3b, 185544:1
Misc Info : 17498,NSAR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 14:16 Cal File: 030618026.D
Als bottle: 14 Calibration Sample, Level: 3
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1242.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
3 Aroclor 1242					CAS #: 53469-21-9			
3.317	3.318	-0.001	1252566	0.50000	0.49	0.00- 20.00	100.00(M)	BA
3.513	3.513	0.000	1253174	0.50000	0.50	0.00- 20.00	100.05	BA
3.827	3.827	0.000	3485663	0.50000	0.49	0.00- 20.00	278.28	BA
3.929	3.929	0.000	1508993	0.50000	0.50	0.00- 20.00	120.47	BA
3.995	3.995	0.000	1031465	0.50000	0.50	0.00- 20.00	82.35	BA
Average of Peak Amounts =					0.49600			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618026.D

Date : 06-MAR-2018 14:16

Client ID: 1242-CAL3B

Instrument: 40GCSJ.i

Sample Info: 1242-CAL3B, 185544:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618026.D (ManInt: NSAR1242.sub Orig: NSAR1242.sub)
Injection Date: 06-MAR-2018 14:16 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1242-CAL3B





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1242	1246512	1252566	BA
Aroclor 1242	1246512	1252566	BA
Aroclor 1242	1246512	1252566	BA
Aroclor 1242	1246512	1252566	BA
Aroclor 1242	1246512	1252566	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618028.D
Lab Smp Id: 1242-CAL4B Client Smp ID: 1242-CAL4B
Inj Date : 06-MAR-2018 14:32
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-cal4b, 185545:1
Misc Info : 17498,NSAR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 14:32 Cal File: 030618028.D
Als bottle: 15 Calibration Sample, Level: 4
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1242.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.185	10080	10187	1.011	0.01		
1.429	71364	59862	0.839	0.11		
1.538	29625	26596	0.898	0.04		
1.688	11915	10660	0.895	0.01		
1.804	23856	18863	0.791	0.03		
2.007	10897	7828	0.718	0.01		
2.404	181872	142910	0.786	0.29		
2.485	11627123	11366999	0.978	18.98		
2.688	264051	219269	0.830	0.42		
2.789	434211	416960	0.960	0.69		
2.851	1404465	1315051	0.936	2.24		
3.055	207612	200354	0.965	0.33		
3.277	30044	24512	0.816	0.04		
3.318	1896850	1767208	0.932	3.03	3 Aroclor 1242	BA
3.345	1097148	1042916	0.951	1.75		
3.435	308786	278730	0.903	0.49		
3.514	1913733	1396800	0.730	3.05	3 Aroclor 1242	BA
3.611	36879	32861	0.891	0.05		
3.650	32275	21129	0.655	0.05		
3.698	52994	51432	0.971	0.08		
3.732	483352	488922	1.012	0.77		
3.757	311671	302713	0.971	0.49		
3.827	5388263	3191450	0.592	8.61	3 Aroclor 1242	BA
3.930	2300081	1902191	0.827	3.67	3 Aroclor 1242	BA
3.995	1542626	1468087	0.952	2.46	3 Aroclor 1242	BA
4.058	448445	418306	0.933	0.71		
4.137	178413	160377	0.899	0.28		
4.216	1273682	1194446	0.938	2.03		
4.262	1227232	1169872	0.953	1.96		
4.294	525419	547237	1.042	0.83		
4.311	636195	609901	0.959	1.01		
4.372	36108	28635	0.793	0.05		
4.417	1692750	1588767	0.939	2.70		
4.453	1331839	978159	0.734	2.12		
4.546	2184137	1978129	0.906	3.49		
4.617	508881	469073	0.922	0.81		
4.680	35054	29148	0.832	0.05		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.726	105197	91550	0.870	0.16	
4.769	122661	117412	0.957	0.19	
4.814	1044933	1018344	0.975	1.66	
4.849	1747424	1640600	0.939	2.79	
4.879	2259358	1670288	0.739	3.61	
4.958	257618	206246	0.801	0.41	
5.037	1890389	1718139	0.909	3.02	
5.100	401551	318982	0.794	0.64	
5.163	472012	441234	0.935	0.75	
5.213	355660	314958	0.886	0.56	
5.286	29890	27564	0.922	0.04	
5.311	83705	68170	0.814	0.13	
5.364	301164	261411	0.868	0.48	
5.407	509309	451496	0.886	0.81	
5.445	334594	277269	0.829	0.53	
5.509	841921	651295	0.774	1.34	
5.605	255376	218910	0.857	0.40	
5.678	22556	17148	0.760	0.03	
5.717	46249	39239	0.848	0.07	
5.741	70982	58982	0.831	0.11	
5.793	601893	487520	0.810	0.96	
5.888	90739	56064	0.618	0.14	
5.996	18848	17967	0.953	0.03	
6.025	494570	434696	0.879	0.79	
6.055	94751	75807	0.800	0.15	
6.167	43225	29320	0.678	0.06	
6.227	36179	22391	0.619	0.05	
6.298	159553	135640	0.850	0.25	
6.334	38685	28936	0.748	0.06	
6.377	33304	20685	0.621	0.05	
6.537	60223	50647	0.841	0.09	
6.804	34153	30829	0.903	0.05	
7.006	17283	12447	0.720	0.02	
7.126	26579	14899	0.561	0.04	
7.251	12264	12902	1.052	0.01	
7.726	42507	35890	0.844	0.06	
7.802	24391	23419	0.960	0.03	
8.277	86483	69397	0.802	0.13	
8.377	21031	14261	0.678	0.03	
8.573	9665476	8353148	0.864	15.44	
8.947	18281	11560	0.632	0.02	
9.088	11569	10817	0.935	0.01	
9.469	15262	13186	0.864	0.02	
9.963	30595	19946	0.652	0.04	
	62580323	54528151		100.000	

Total unknown % area = 79

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618028.D
Lab Smp Id: 1242-CAL4B Client Smp ID: 1242-CAL4B
Inj Date : 06-MAR-2018 14:32
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-cal4b, 185545:1
Misc Info : 17498,NSAR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 14:32 Cal File: 030618028.D
Als bottle: 15 Calibration Sample, Level: 4
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1242.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
3 Aroclor 1242					CAS #: 53469-21-9			
3.318	3.318	0.000	1896850	0.80000	0.75	0.00- 20.00	100.00(M)	BA
3.513	3.513	0.000	1913733	0.80000	0.76	0.00- 20.00	100.89	BA
3.827	3.827	0.000	5388263	0.80000	0.76	0.00- 20.00	284.06	BA
3.930	3.929	0.001	2300081	0.80000	0.76	0.00- 20.00	121.26	BA
3.995	3.995	0.000	1542626	0.80000	0.75	0.00- 20.00	81.33	BA
Average of Peak Amounts =					0.75600			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618028.D

Date : 06-MAR-2018 14:32

Client ID: 1242-CAL4B

Instrument: 40GCSJ.i

Sample Info: 1242-CAL4B, 185545:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618028.D (ManInt: NSAR1242.sub Orig: NSAR1242.sub)
Injection Date: 06-MAR-2018 14:32 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1242-CAL4B





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1242	1890576	1896850	BA
Aroclor 1242	1890576	1896850	BA
Aroclor 1242	1890576	1896850	BA
Aroclor 1242	1890576	1896850	BA
Aroclor 1242	1890576	1896850	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618030.D
Lab Smp Id: 1242-CAL5B Client Smp ID: 1242-CAL5B
Inj Date : 06-MAR-2018 14:48
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-cal5b, 185546:1
Misc Info : 17498,NSAR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 14:48 Cal File: 030618030.D
Als bottle: 16 Calibration Sample, Level: 5
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1242.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.182	13142	9158	0.697	0.01		
1.429	100388	88167	0.878	0.12		
1.538	43436	37040	0.853	0.05		
1.802	27170	23817	0.877	0.03		
2.200	13193	7765	0.589	0.01		
2.404	223179	177740	0.796	0.26		
2.485	17637176	17309158	0.981	21.53		
2.687	332628	271512	0.816	0.39		
2.788	576258	556266	0.965	0.69		
2.850	1731132	1639872	0.947	2.07		
3.055	271375	257132	0.948	0.32		
3.199	9596	8183	0.853	0.01		
3.278	40145	29795	0.742	0.04		
3.318	2329240	2178022	0.935	2.79	3 Aroclor 1242	BA
3.344	1365925	1274431	0.933	1.63		
3.434	376796	349333	0.927	0.45		
3.514	2361399	1699901	0.720	2.82	3 Aroclor 1242	BA
3.611	53162	49737	0.936	0.06		
3.648	38548	26960	0.699	0.04		
3.697	65654	64327	0.980	0.07		
3.732	603958	618832	1.025	0.72		
3.756	391796	382124	0.975	0.46		
3.828	6706645	3996192	0.596	8.03	3 Aroclor 1242	BA
3.929	2857802	2355014	0.824	3.42	3 Aroclor 1242	BA
3.994	1922508	1833827	0.954	2.30	3 Aroclor 1242	BA
4.059	555199	530529	0.956	0.66		
4.136	222505	199818	0.898	0.26		
4.215	1575639	1501683	0.953	1.88		
4.261	1524934	1467406	0.962	1.82		
4.293	649073	667869	1.029	0.77		
4.311	810435	768258	0.948	0.97		
4.371	43433	37094	0.854	0.05		
4.417	2107983	1937856	0.919	2.52		
4.452	1647097	1208682	0.734	1.97		
4.546	2757523	2530813	0.918	3.30		
4.617	639793	601712	0.940	0.76		
4.681	44151	33585	0.761	0.05		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.728	132122	115497	0.874	0.15	
4.768	152881	144940	0.948	0.18	
4.814	1304656	1285158	0.985	1.56	
4.848	2176719	2065105	0.949	2.60	
4.878	2810648	2043331	0.727	3.36	
4.958	313624	254506	0.811	0.37	
5.038	2364110	2159153	0.913	2.83	
5.099	500540	389011	0.777	0.59	
5.163	588453	534323	0.908	0.70	
5.213	446583	410279	0.919	0.53	
5.283	33027	31901	0.966	0.03	
5.308	101588	84023	0.827	0.12	
5.363	379663	323299	0.852	0.45	
5.407	645531	557984	0.864	0.77	
5.447	411135	369925	0.900	0.49	
5.510	1059127	856260	0.808	1.26	
5.604	309697	267220	0.863	0.37	
5.676	26819	20072	0.748	0.03	
5.717	59263	49321	0.832	0.07	
5.742	85889	77700	0.905	0.10	
5.763	25900	36740	1.419	0.03	
5.792	730478	608087	0.832	0.87	
5.888	108449	69349	0.639	0.12	
5.993	23548	21446	0.911	0.02	
6.024	621770	555127	0.893	0.74	
6.056	116744	93958	0.805	0.13	
6.169	44881	36731	0.818	0.05	
6.227	46866	28002	0.597	0.05	
6.299	198407	171669	0.865	0.23	
6.334	48913	37408	0.765	0.05	
6.372	34798	26791	0.770	0.04	
6.535	86352	62639	0.725	0.10	
6.608	17596	12479	0.709	0.02	
6.719	9324	8755	0.939	0.01	
6.803	41284	37893	0.918	0.04	
7.004	14125	13843	0.980	0.01	
7.122	30632	19166	0.626	0.03	
7.251	21667	14526	0.670	0.02	
7.694	10065	10273	1.021	0.01	
7.729	57227	51664	0.903	0.06	
7.800	40976	33029	0.806	0.04	
8.276	125210	109793	0.877	0.14	
8.377	17875	15033	0.841	0.02	
8.572	14386052	12725920	0.885	17.23	
9.089	27631	16553	0.599	0.03	
9.466	10052	7309	0.727	0.01	
9.961	14595	11352	0.778	0.01	
	83483509	73604153		100.000	

Total unknown % area = 81

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618030.D
Lab Smp Id: 1242-CAL5B Client Smp ID: 1242-CAL5B
Inj Date : 06-MAR-2018 14:48
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-cal5b, 185546:1
Misc Info : 17498,NSAR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 14:48 Cal File: 030618030.D
Als bottle: 16 Calibration Sample, Level: 5
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1242.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
3 Aroclor 1242					CAS #: 53469-21-9			
3.318	3.318	0.000	2329239	1.00000	0.92	0.00- 20.00	100.00(M)	BA
3.513	3.513	0.000	2361398	1.00000	0.94	0.00- 20.00	101.38	BA
3.827	3.827	0.000	6706645	1.00000	0.94	0.00- 20.00	287.93	BA
3.929	3.929	0.000	2857801	1.00000	0.94	0.00- 20.00	122.69	BA
3.994	3.995	-0.001	1922507	1.00000	0.94	0.00- 20.00	82.54	BA
Average of Peak Amounts =					0.93600			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618030.D

Date : 06-MAR-2018 14:48

Client ID: 1242-CAL5B

Instrument: 40GCSJ.i

Sample Info: 1242-CAL5B, 185546:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618030.D (ManInt: NSAR1242.sub Orig: NSAR1242.sub)
Injection Date: 06-MAR-2018 14:48 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1242-CAL5B





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1242	2319478	2329239	BA
Aroclor 1242	2319478	2329239	BA
Aroclor 1242	2319478	2329239	BA
Aroclor 1242	2319478	2329239	BA
Aroclor 1242	2319478	2329239	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618034.D
Lab Smp Id: 1248-CAL1D Client Smp ID: 1248-CAL1D
Inj Date : 06-MAR-2018 15:21
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-cal1d, 188134:1
Misc Info : 17498,NSAR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 15:21 Cal File: 030618034.D
Als bottle: 18 Calibration Sample, Level: 1
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1248.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.181	14305	12931	0.904	0.15		
2.484	1247615	1209145	0.969	13.39		
2.788	24449	24176	0.989	0.25		
2.851	42332	37681	0.890	0.44		
3.054	17847	15627	0.876	0.18		
3.319	154829	149314	0.964	1.63		
3.345	69818	62303	0.892	0.73		
3.434	21759	16384	0.753	0.22		
3.514	129337	91632	0.708	1.36		
3.732	41192	29456	0.715	0.43		
3.815	249582	260696	1.045	2.62		
3.827	243092	265599	1.093	2.56		
3.930	196862	149009	0.757	2.07		
3.994	130100	120464	0.926	1.37		
4.060	81123	71377	0.880	0.85		
4.135	37994	31819	0.837	0.40		
4.215	323817	309946	0.957	3.41		
4.261	252265	240506	0.953	2.65	6 Aroclor 1248	BA
4.293	108819	107853	0.991	1.14		
4.310	114993	111028	0.966	1.21		
4.418	386700	369661	0.956	4.07		
4.452	202370	173038	0.855	2.13		
4.546	444141	405258	0.912	4.67	6 Aroclor 1248	BA
4.618	103930	97669	0.940	1.09		
4.723	15775	14602	0.926	0.16		
4.768	26846	24690	0.920	0.28		
4.814	217596	209298	0.962	2.29	6 Aroclor 1248	BA
4.848	417341	413186	0.990	4.39		
4.879	538491	358710	0.666	5.67	6 Aroclor 1248	BA
4.956	69249	57131	0.825	0.72		
5.038	375736	339065	0.902	3.95	6 Aroclor 1248	BA
5.100	148135	110594	0.747	1.56		
5.163	166243	148041	0.891	1.75		
5.212	121348	109604	0.903	1.27		
5.313	32609	23414	0.718	0.34		
5.363	107678	94673	0.879	1.13		
5.408	174507	147282	0.844	1.83		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.447	107960	95389	0.884	1.13	
5.510	292671	230307	0.787	3.08	
5.606	86504	77757	0.899	0.91	
5.677	14791	13016	0.880	0.15	
5.715	25257	18019	0.713	0.26	
5.741	25694	22207	0.864	0.27	
5.793	207991	150871	0.725	2.19	
5.891	26284	18285	0.696	0.27	
6.024	187756	132978	0.708	1.97	
6.173	26666	15778	0.592	0.28	
6.301	55473	43247	0.780	0.58	
6.372	11356	9835	0.866	0.11	
6.509	11520	12779	1.109	0.12	
6.535	21081	15879	0.753	0.22	
6.720	14478	10691	0.738	0.15	
6.801	13412	10673	0.796	0.14	
7.001	12573	13352	1.062	0.13	
7.250	10465	7692	0.735	0.11	
7.384	10819	8597	0.795	0.11	
7.445	14986	9827	0.656	0.15	
8.276	19065	13591	0.713	0.20	
8.376	10132	9028	0.891	0.10	
8.572	1203814	1072809	0.891	12.67	
8.944	15824	11638	0.735	0.16	
9.963	17710	12713	0.718	0.18	
	9495108	8449820		100.000	

Total unknown % area = 81

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618034.D
Lab Smp Id: 1248-CAL1D Client Smp ID: 1248-CAL1D
Inj Date : 06-MAR-2018 15:21
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-cal1d, 188134:1
Misc Info : 17498,NSAR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 15:21 Cal File: 030618034.D
Als bottle: 18 Calibration Sample, Level: 1
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1248.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
6 Aroclor 1248					CAS #: 12672-29-6			
4.261	4.261	0.000	252264	0.10000	0.11	0.00- 20.00	100.00(M)	BA
4.546	4.546	0.000	444140	0.10000	0.11	0.00- 20.00	176.06	BA
4.814	4.812	0.002	217595	0.10000	0.11	0.00- 20.00	86.26	BA
4.879	4.879	0.000	538491	0.10000	0.11	0.00- 20.00	213.46	BA
5.037	5.038	-0.001	375736	0.10000	0.10	0.00- 20.00	148.95	BA
Average of Peak Amounts =					0.10800			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618034.D

Date : 06-MAR-2018 15:21

Client ID: 1248-CAL1D

Instrument: 40GCSJ.i

Sample Info: 1248-CAL1D, 188134:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618034.D (ManInt: NSAR1248.sub Orig: NSAR1248.sub)
Injection Date: 06-MAR-2018 15:21 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1248-CAL1D





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1248	249611	252264	BA
Aroclor 1248	249611	252264	BA
Aroclor 1248	249611	252264	BA
Aroclor 1248	249611	252264	BA
Aroclor 1248	249611	252264	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618036.D
Lab Smp Id: 1248-CAL2D Client Smp ID: 1248-CAL2D
Inj Date : 06-MAR-2018 15:37
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-cal2d, 185548:1
Misc Info : 17498,NSAR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 15:37 Cal File: 030618036.D
Als bottle: 19 Calibration Sample, Level: 2
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1248.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.183	14828	12598	0.850	0.08		
1.430	17405	15532	0.892	0.09		
1.539	10978	9178	0.836	0.05		
2.484	2410492	2310272	0.958	13.49		
2.687	17007	12868	0.757	0.09		
2.788	48103	47022	0.978	0.26		
2.850	76500	74948	0.980	0.41		
3.054	25960	22998	0.886	0.14		
3.317	298832	283610	0.949	1.62		
3.344	124976	117615	0.941	0.68		
3.435	29128	26666	0.915	0.15		
3.515	253342	181031	0.715	1.37		
3.731	81061	56758	0.700	0.44		
3.815	462892	504630	1.090	2.51		
3.826	513591	533097	1.038	2.79		
3.930	374207	289234	0.773	2.03		
3.994	244564	239466	0.979	1.33		
4.058	144578	133216	0.921	0.78		
4.138	61375	54340	0.885	0.33		
4.215	629384	607757	0.966	3.42		
4.262	495628	473751	0.956	2.69	6 Aroclor 1248	BA
4.292	207266	214773	1.036	1.12		
4.310	218472	204865	0.938	1.18		
4.418	768532	734950	0.956	4.18		
4.453	378840	319389	0.843	2.06		
4.546	893337	815571	0.913	4.86	6 Aroclor 1248	BA
4.619	201356	185248	0.920	1.09		
4.680	15861	10971	0.692	0.08		
4.727	29910	23588	0.789	0.16		
4.768	49430	45109	0.913	0.26		
4.813	420194	408578	0.972	2.28	6 Aroclor 1248	BA
4.849	831772	785146	0.944	4.52		
4.879	1060573	728567	0.687	5.77	6 Aroclor 1248	BA
4.957	136900	106494	0.778	0.74		
5.037	754040	696546	0.924	4.10	6 Aroclor 1248	BA
5.100	272713	215866	0.792	1.48		
5.163	325949	288755	0.886	1.77		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.213	228553	210074	0.919	1.24	
5.284	17412	16457	0.945	0.09	
5.311	55636	44138	0.793	0.30	
5.365	206530	172716	0.836	1.12	
5.407	337971	295228	0.874	1.83	
5.446	206711	185231	0.896	1.12	
5.511	560479	467877	0.835	3.05	
5.606	165540	145657	0.880	0.90	
5.677	22681	19784	0.872	0.12	
5.717	37903	28891	0.762	0.20	
5.740	60086	37215	0.619	0.32	
5.793	371069	303747	0.819	2.02	
5.889	58211	32981	0.567	0.31	
5.992	13175	13785	1.046	0.07	
6.025	314967	267323	0.849	1.71	
6.057	68435	55542	0.812	0.37	
6.170	38384	24308	0.633	0.20	
6.299	99670	82914	0.832	0.54	
6.332	23115	17884	0.774	0.12	
6.374	20547	14764	0.719	0.11	
6.512	20781	18606	0.895	0.11	
6.536	45734	28211	0.617	0.24	
6.605	13055	7483	0.573	0.07	
6.719	20846	19042	0.913	0.11	
6.807	17328	17245	0.995	0.09	
7.005	39579	28763	0.727	0.21	
7.128	13661	9160	0.671	0.07	
7.254	16230	12319	0.759	0.08	
7.391	15959	13660	0.856	0.08	
7.448	19126	12843	0.671	0.10	
7.727	14015	12218	0.872	0.07	
7.904	9772	9677	0.990	0.05	
8.273	24829	20353	0.820	0.13	
8.382	9612	9180	0.955	0.05	
8.572	2222844	1814975	0.817	12.10	
8.942	19336	12310	0.637	0.10	
9.470	19278	15347	0.796	0.10	
9.966	19855	14911	0.751	0.10	
	18368893	16303822		100.000	

Total unknown % area = 80

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618036.D
Lab Smp Id: 1248-CAL2D Client Smp ID: 1248-CAL2D
Inj Date : 06-MAR-2018 15:37
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-cal2d, 185548:1
Misc Info : 17498,NSAR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 15:37 Cal File: 030618036.D
Als bottle: 19 Calibration Sample, Level: 2
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1248.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
6 Aroclor 1248					CAS #: 12672-29-6			
4.261	4.261	0.000	495628	0.20000	0.21	0.00- 20.00	100.00(M)	BA
4.545	4.546	-0.001	893336	0.20000	0.22	0.00- 20.00	180.24	BA
4.813	4.812	0.001	420194	0.20000	0.21	0.00- 20.00	84.78	BA
4.878	4.879	-0.001	1060573	0.20000	0.21	0.00- 20.00	213.99	BA
5.036	5.038	-0.002	754040	0.20000	0.21	0.00- 20.00	152.14	BA
Average of Peak Amounts =					0.21200			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618036.D

Date : 06-MAR-2018 15:37

Client ID: 1248-CAL2D

Instrument: 40GCSJ.i

Sample Info: 1248-CAL2D, 185548:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618036.D (ManInt: NSAR1248.sub Orig: NSAR1248.sub)
Injection Date: 06-MAR-2018 15:37 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1248-CAL2D





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1248	491137	495628	BA
Aroclor 1248	491137	495628	BA
Aroclor 1248	491137	495628	BA
Aroclor 1248	491137	495628	BA
Aroclor 1248	491137	495628	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618038.D
Lab Smp Id: 1248-CAL3D Client Smp ID: 1248-CAL3D
Inj Date : 06-MAR-2018 15:53
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-cal3d, 185549:1
Misc Info : 17498,NSAR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 15:53 Cal File: 030618038.D
Als bottle: 20 Calibration Sample, Level: 3
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1248.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.183	52291	42958	0.822	0.11		
1.429	43894	31713	0.722	0.09		
1.534	22011	15352	0.697	0.05		
1.684	14849	11921	0.803	0.03		
2.012	36627	27328	0.746	0.08		
2.407	26546	19456	0.733	0.06		
2.484	6017126	5772700	0.959	14.16		
2.690	32471	25048	0.771	0.07		
2.787	113251	108295	0.956	0.25		
2.852	188762	171319	0.908	0.42		
3.029	16871	18416	1.092	0.03		
3.055	67710	56713	0.838	0.15		
3.317	703226	686995	0.977	1.60		
3.344	286830	276759	0.965	0.65		
3.435	67890	58817	0.866	0.15		
3.516	596412	432178	0.725	1.35		
3.611	19167	18609	0.971	0.04		
3.647	15129	9511	0.629	0.03		
3.732	135042	128551	0.952	0.30		
3.756	64325	57019	0.886	0.14		
3.813	1062140	1201939	1.132	2.41		
3.826	1285788	1271741	0.989	2.92		
3.930	900477	692801	0.769	2.04		
3.994	605988	570061	0.941	1.37		
4.058	351201	319673	0.910	0.79		
4.137	149029	131722	0.884	0.33		
4.215	1475927	1383740	0.938	3.35		
4.261	1173198	1126600	0.960	2.66	6 Aroclor 1248	BA
4.293	494927	502286	1.015	1.12		
4.310	515717	488015	0.946	1.17		
4.417	1799324	1700045	0.945	4.09		
4.452	888654	777134	0.875	2.02		
4.546	2052388	1889189	0.920	4.66	6 Aroclor 1248	BA
4.588	34383	33944	0.987	0.07		
4.618	476094	438433	0.921	1.08		
4.681	29129	21872	0.751	0.06		
4.727	64061	51195	0.799	0.14		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.769	110931	104719	0.944	0.25	
4.813	1017191	989120	0.972	2.31	6 Aroclor 1248
4.848	1949790	1821792	0.934	4.43	
4.879	2527488	1652488	0.654	5.75	6 Aroclor 1248
4.957	303686	249082	0.820	0.69	
5.038	1794002	1655561	0.923	4.08	6 Aroclor 1248
5.100	643410	513336	0.798	1.46	
5.162	765077	698728	0.913	1.74	
5.213	547175	504652	0.922	1.24	
5.284	41401	37162	0.898	0.09	
5.308	120103	105343	0.877	0.27	
5.363	470348	418426	0.890	1.07	
5.407	811478	707819	0.872	1.84	
5.445	498126	442022	0.887	1.13	
5.510	1346051	1105489	0.821	3.06	
5.605	392604	337924	0.861	0.89	
5.674	48073	38822	0.808	0.10	
5.717	86755	62682	0.723	0.19	
5.740	101326	88143	0.870	0.23	
5.765	28818	38498	1.336	0.06	
5.792	911257	762608	0.837	2.07	
5.889	94463	79681	0.844	0.21	
5.913	50133	31292	0.624	0.11	
5.994	30626	28514	0.931	0.06	
6.025	762038	662252	0.869	1.73	
6.055	160892	129589	0.805	0.36	
6.171	81256	53475	0.658	0.18	
6.228	54811	30043	0.548	0.12	
6.300	235601	196601	0.834	0.53	
6.333	46391	39165	0.844	0.10	
6.376	41268	29678	0.719	0.09	
6.447	37982	15949	0.420	0.08	
6.509	37155	37405	1.007	0.08	
6.537	96382	65323	0.678	0.21	
6.604	13761	11378	0.827	0.03	
6.719	40218	33312	0.828	0.09	
6.759	20105	17496	0.870	0.04	
6.803	43393	37801	0.871	0.09	
6.843	24388	14465	0.593	0.05	
6.941	11214	8137	0.726	0.02	
7.004	73324	59305	0.809	0.16	
7.126	31431	17422	0.554	0.07	
7.250	27971	25551	0.913	0.06	
7.391	30533	26580	0.871	0.06	
7.444	27775	25655	0.924	0.06	
7.693	17906	15318	0.855	0.04	
7.728	28913	24590	0.850	0.06	
7.779	18403	15964	0.867	0.04	
7.802	17401	15137	0.870	0.03	
7.901	30922	23185	0.750	0.07	
8.276	52635	47411	0.901	0.11	
8.380	33122	21479	0.648	0.07	
8.571	5194473	4505821	0.867	11.81	
8.943	27069	16672	0.616	0.06	
9.470	33315	19626	0.589	0.07	
9.960	28297	18719	0.662	0.06	
	43949512	39304455		100.000	

Total unknown % area = 80

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618038.D
Lab Smp Id: 1248-CAL3D Client Smp ID: 1248-CAL3D
Inj Date : 06-MAR-2018 15:53
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-cal3d, 185549:1
Misc Info : 17498,NSAR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 15:53 Cal File: 030618038.D
Als bottle: 20 Calibration Sample, Level: 3
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1248.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
6 Aroclor 1248					CAS #: 12672-29-6			
4.261	4.261	0.000	1173198	0.50000	0.50	0.00- 20.00	100.00(M)	BA
4.545	4.546	-0.001	2052387	0.50000	0.49	0.00- 20.00	174.94	BA
4.813	4.812	0.001	1017190	0.50000	0.50	0.00- 20.00	86.70	BA
4.879	4.879	0.000	2527488	0.50000	0.50	0.00- 20.00	215.44	BA
5.037	5.038	-0.001	1794001	0.50000	0.50	0.00- 20.00	152.92	BA
Average of Peak Amounts =					0.49800			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618038.D

Date : 06-MAR-2018 15:53

Client ID: 1248-CAL3D

Instrument: 40GCSJ.i

Sample Info: 1248-CAL3D, 185549:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618038.D (ManInt: NSAR1248.sub Orig: NSAR1248.sub)
Injection Date: 06-MAR-2018 15:53 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1248-CAL3D





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1248	1161123	1173198	BA
Aroclor 1248	1161123	1173198	BA
Aroclor 1248	1161123	1173198	BA
Aroclor 1248	1161123	1173198	BA
Aroclor 1248	1161123	1173198	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618040.D
Lab Smp Id: 1248-CAL4D Client Smp ID: 1248-CAL4D
Inj Date : 06-MAR-2018 16:09
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-cal4d, 185550:1
Misc Info : 17498,NSAR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 16:09 Cal File: 030618040.D
Als bottle: 21 Calibration Sample, Level: 4
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1248.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.184	23783	14012	0.589	0.03		
1.430	68772	56288	0.818	0.09		
1.538	27818	21841	0.785	0.03		
2.406	41368	29218	0.706	0.05		
2.485	11662105	11112795	0.953	17.14		
2.688	47215	38573	0.817	0.06		
2.788	203552	191980	0.943	0.29		
2.851	287413	258349	0.899	0.41		
3.055	100338	85256	0.850	0.14		
3.318	1056102	1019716	0.966	1.50		
3.345	441526	421992	0.956	0.63		
3.434	97081	90204	0.929	0.13		
3.515	910659	658598	0.723	1.30		
3.611	32844	31062	0.946	0.04		
3.651	12163	10814	0.889	0.01		
3.695	12227	11029	0.902	0.01		
3.732	195496	195587	1.000	0.27		
3.756	83420	81444	0.976	0.11		
3.814	1567662	1791160	1.143	2.24		
3.827	1917579	1885095	0.983	2.74		
3.931	1340766	1031743	0.770	1.91		
3.994	893318	859435	0.962	1.27		
4.059	507945	474678	0.935	0.72		
4.137	209816	195120	0.930	0.29		
4.216	2174966	1996009	0.918	3.10		
4.262	1747290	1634508	0.935	2.49	6 Aroclor 1248	BA
4.293	719186	757380	1.053	1.02		
4.311	795770	743692	0.935	1.13		
4.418	2683345	2471775	0.921	3.83		
4.453	1336671	1172433	0.877	1.91		
4.546	3077708	2813367	0.914	4.39	6 Aroclor 1248	BA
4.592	50782	48315	0.951	0.07		
4.618	721531	680792	0.944	1.03		
4.682	39778	31109	0.782	0.05		
4.727	90436	74601	0.825	0.12		
4.768	164069	159220	0.970	0.23		
4.814	1529584	1458936	0.954	2.18	6 Aroclor 1248	BA

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.849	2912545	2807178	0.964	4.16	
4.879	3756486	2507650	0.668	5.36	6 Aroclor 1248
4.959	446081	367801	0.825	0.63	
5.038	2709000	2506705	0.925	3.87	6 Aroclor 1248
5.100	972027	775936	0.798	1.38	
5.163	1142709	1049261	0.918	1.63	
5.213	827376	755691	0.913	1.18	
5.285	59301	53069	0.895	0.08	
5.309	180372	153614	0.852	0.25	
5.364	725905	634309	0.874	1.03	
5.408	1217313	1074540	0.883	1.73	
5.447	755270	658840	0.872	1.07	
5.510	2000999	1684613	0.842	2.85	
5.605	577093	528233	0.915	0.82	
5.676	67052	59659	0.890	0.09	
5.718	120890	90293	0.747	0.17	
5.742	139371	130806	0.939	0.19	
5.763	48200	56761	1.178	0.06	
5.793	1366535	1115283	0.816	1.95	
5.888	190518	116597	0.612	0.27	
5.995	43830	40905	0.933	0.06	
6.025	1167064	1029378	0.882	1.66	
6.056	233192	189145	0.811	0.33	
6.171	117974	81314	0.689	0.16	
6.229	73058	40661	0.557	0.10	
6.299	353507	303097	0.857	0.50	
6.334	67864	56427	0.831	0.09	
6.373	56825	43449	0.765	0.08	
6.434	31012	16905	0.545	0.04	
6.512	66221	56258	0.850	0.09	
6.538	139332	94112	0.675	0.19	
6.604	20526	17194	0.838	0.02	
6.665	17050	12889	0.756	0.02	
6.720	59223	48657	0.822	0.08	
6.761	28886	25179	0.872	0.04	
6.804	64363	56227	0.874	0.09	
6.841	26723	19155	0.717	0.03	
7.005	107576	96122	0.894	0.15	
7.124	36317	20799	0.573	0.05	
7.253	47298	40406	0.854	0.06	
7.284	14796	11885	0.803	0.02	
7.392	50134	42799	0.854	0.07	
7.444	51593	44890	0.870	0.07	
7.690	21646	19744	0.912	0.03	
7.730	50158	42338	0.844	0.07	
7.804	37842	29484	0.779	0.05	
7.903	37446	35295	0.943	0.05	
8.276	107716	86008	0.798	0.15	
8.382	16440	13810	0.840	0.02	
8.573	9713869	8374648	0.862	13.88	
9.468	13816	9740	0.705	0.01	
9.968	11075	9075	0.819	0.01	
	69971497	62742960		100.000	

Total unknown % area = 82

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618040.D
Lab Smp Id: 1248-CAL4D Client Smp ID: 1248-CAL4D
Inj Date : 06-MAR-2018 16:09
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-cal4d, 185550:1
Misc Info : 17498,NSAR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 16:09 Cal File: 030618040.D
Als bottle: 21 Calibration Sample, Level: 4
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1248.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
6 Aroclor 1248					CAS #: 12672-29-6			
4.261	4.261	0.000	1747289	0.80000	0.74	0.00- 20.00	100.00(M)	BA
4.545	4.546	-0.001	3077707	0.80000	0.74	0.00- 20.00	176.14	BA
4.813	4.812	0.001	1529584	0.80000	0.75	0.00- 20.00	87.54	BA
4.878	4.879	-0.001	3756485	0.80000	0.74	0.00- 20.00	214.99	BA
5.038	5.038	0.000	2708999	0.80000	0.75	0.00- 20.00	155.04	BA
Average of Peak Amounts =					0.74400			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618040.D

Date : 06-MAR-2018 16:09

Client ID: 1248-CAL4D

Instrument: 40GCSJ.i

Sample Info: 1248-CAL4D, 185550:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618040.D (ManInt: NSAR1248.sub Orig: NSAR1248.sub)
Injection Date: 06-MAR-2018 16:09 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1248-CAL4D





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1248	1739344	1747289	BA
Aroclor 1248	1739344	1747289	BA
Aroclor 1248	1739344	1747289	BA
Aroclor 1248	1739344	1747289	BA
Aroclor 1248	1739344	1747289	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618042.D
Lab Smp Id: 1248-CAL5D Client Smp ID: 1248-CAL5D
Inj Date : 06-MAR-2018 16:25
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-cal5d, 185551:1
Misc Info : 17498,NSAR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 16:25 Cal File: 030618042.D
Als bottle: 22 Calibration Sample, Level: 5
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1248.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.184	15397	13334	0.866	0.01		
1.430	101369	87214	0.860	0.10		
1.537	43989	34591	0.786	0.04		
2.406	47903	33430	0.698	0.05		
2.485	17721229	17125552	0.966	19.37		
2.688	56165	47205	0.840	0.05		
2.788	288750	277495	0.961	0.30		
2.851	348759	325290	0.933	0.37		
3.054	123922	115841	0.935	0.13		
3.273	9293	8724	0.939	0.00		
3.318	1341983	1263092	0.941	1.43		
3.344	567538	536022	0.944	0.60		
3.434	126309	115528	0.915	0.13		
3.516	1155877	827361	0.716	1.23		
3.611	47908	47362	0.989	0.05		
3.647	10831	12550	1.159	0.01		
3.696	12947	14220	1.098	0.01		
3.732	246189	245791	0.998	0.26		
3.758	102423	103998	1.015	0.10		
3.814	1931729	2268598	1.174	2.06		
3.826	2479017	2389735	0.964	2.64		
3.931	1695811	1294913	0.764	1.80		
3.995	1133597	1088944	0.961	1.20		
4.059	642924	621662	0.967	0.68		
4.136	260443	245788	0.944	0.27		
4.216	2732694	2586455	0.946	2.91		
4.261	2183583	2056861	0.942	2.32	6 Aroclor 1248	BA
4.293	940491	949013	1.009	1.00		
4.310	984371	940127	0.955	1.05		
4.373	17082	14222	0.833	0.01		
4.418	3371096	3162549	0.938	3.59		
4.453	1703074	1454953	0.854	1.81		
4.546	3891543	3587538	0.922	4.15	6 Aroclor 1248	BA
4.591	55507	60505	1.090	0.05		
4.618	922889	867798	0.940	0.98		
4.679	46691	38089	0.816	0.04		
4.728	116729	92875	0.796	0.12		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.769	206094	200087	0.971	0.21	
4.814	1927122	1857150	0.964	2.05	6 Aroclor 1248
4.849	3694572	3560273	0.964	3.94	
4.880	4748871	3261898	0.687	5.06	6 Aroclor 1248
4.958	573809	487686	0.850	0.61	
5.038	3440920	3224544	0.937	3.67	6 Aroclor 1248
5.099	1231738	975474	0.792	1.31	
5.163	1442598	1311329	0.909	1.53	
5.213	1046539	961794	0.919	1.11	
5.285	77016	69190	0.898	0.08	
5.309	226198	194152	0.858	0.24	
5.364	920750	826849	0.898	0.98	
5.407	1532611	1323050	0.863	1.63	
5.446	983454	847993	0.862	1.04	
5.511	2532627	2102151	0.830	2.70	
5.605	736667	659983	0.896	0.78	
5.676	82056	71573	0.872	0.08	
5.717	150366	114784	0.763	0.16	
5.741	184149	167330	0.909	0.19	
5.762	57712	71463	1.238	0.06	
5.793	1735942	1359776	0.783	1.85	
5.890	231734	149127	0.644	0.24	
5.994	49770	49696	0.999	0.05	
6.025	1490983	1293880	0.868	1.59	
6.056	314721	242832	0.772	0.33	
6.169	146062	104225	0.714	0.15	
6.227	101525	50840	0.501	0.10	
6.299	449260	381467	0.849	0.47	
6.335	84948	72590	0.855	0.09	
6.375	71691	55450	0.773	0.07	
6.435	33526	20718	0.618	0.03	
6.510	77044	72111	0.936	0.08	
6.537	172087	123276	0.716	0.18	
6.609	25723	18897	0.735	0.02	
6.660	18434	15697	0.852	0.01	
6.721	70509	61602	0.874	0.07	
6.760	38763	32658	0.843	0.04	
6.802	83703	70972	0.848	0.08	
6.842	38188	23754	0.622	0.04	
6.914	13236	11502	0.869	0.01	
6.943	16695	13481	0.807	0.01	
7.006	152792	119576	0.783	0.16	
7.122	35477	27446	0.774	0.03	
7.252	56007	44086	0.787	0.05	
7.388	61340	50925	0.830	0.06	
7.446	59037	54928	0.930	0.06	
7.692	28579	27869	0.975	0.03	
7.730	76894	62730	0.816	0.08	
7.802	60957	40295	0.661	0.06	
7.901	57972	43159	0.744	0.06	
8.276	150141	122351	0.815	0.16	
8.379	21548	15836	0.735	0.02	
8.573	14394097	12289354	0.854	15.35	
8.946	18562	11531	0.621	0.01	
9.086	15319	13297	0.868	0.01	
9.963	9351	8698	0.930	0.00	
	93736539	84400610		100.000	

Total unknown % area = 83

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618042.D
Lab Smp Id: 1248-CAL5D Client Smp ID: 1248-CAL5D
Inj Date : 06-MAR-2018 16:25
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-cal5d, 185551:1
Misc Info : 17498,NSAR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 16:25 Cal File: 030618042.D
Als bottle: 22 Calibration Sample, Level: 5
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1248.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
6 Aroclor 1248					CAS #: 12672-29-6			
4.261	4.261	0.000	2183582	1.00000	0.93	0.00- 20.00	100.00(M)	BA
4.545	4.546	-0.001	3891542	1.00000	0.94	0.00- 20.00	178.22	BA
4.813	4.812	0.001	1927121	1.00000	0.95	0.00- 20.00	88.26	BA
4.879	4.879	0.000	4748871	1.00000	0.94	0.00- 20.00	217.48	BA
5.037	5.038	-0.001	3440920	1.00000	0.96	0.00- 20.00	157.58	BA
Average of Peak Amounts =					0.94400			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618042.D

Date : 06-MAR-2018 16:25

Client ID: 1248-CAL5D

Instrument: 40GCSJ.i

Sample Info: 1248-CAL5D, 185551:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618042.D (ManInt: NSAR1248.sub Orig: NSAR1248.sub)
Injection Date: 06-MAR-2018 16:25 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1248-CAL5D





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1248	2175285	2183582	BA
Aroclor 1248	2175285	2183582	BA
Aroclor 1248	2175285	2183582	BA
Aroclor 1248	2175285	2183582	BA
Aroclor 1248	2175285	2183582	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618046.D
Lab Smp Id: 1254-CAL1C Client Smp ID: 1254-CAL1C
Inj Date : 06-MAR-2018 16:57
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-cal1c, 188135:1
Misc Info : 17498,NSAR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 16:57 Cal File: 030618046.D
Als bottle: 24 Calibration Sample, Level: 1
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1254.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.183	48987	40093	0.818	0.41		
1.426	12583	10350	0.823	0.10		
1.686	14588	13145	0.901	0.12		
2.011	39185	27090	0.691	0.33		
2.484	1264026	1229666	0.973	11.03		
2.787	25737	23306	0.906	0.21		
3.028	26930	17363	0.645	0.22		
3.814	29490	14918	0.506	0.25		
4.000	27012	19985	0.740	0.22		
4.046	9554	8540	0.894	0.08		
4.215	267608	259554	0.970	2.27		
4.260	70496	69050	0.979	0.59		
4.293	14185	11720	0.826	0.12		
4.418	167997	155970	0.928	1.42		
4.452	20339	17534	0.862	0.17		
4.547	73516	70272	0.956	0.62		
4.615	15012	14316	0.954	0.12		
4.813	63092	62377	0.989	0.53		
4.848	232331	222423	0.957	1.97		
4.895	547413	435444	0.795	4.64		
4.958	89049	77209	0.867	0.75		
5.038	53508	47334	0.885	0.45		
5.101	310238	269708	0.869	2.63		
5.163	578338	533266	0.922	4.91	7 Aroclor 1254	BA
5.213	259231	232363	0.896	2.20		
5.286	13290	15715	1.182	0.11		
5.310	48017	42464	0.884	0.40		
5.363	222324	207379	0.933	1.88		
5.408	461676	411346	0.891	3.91	7 Aroclor 1254	BA
5.446	153810	143804	0.935	1.30		
5.478	60811	58639	0.964	0.51		
5.511	843524	762911	0.904	7.16	7 Aroclor 1254	BA
5.605	121504	105369	0.867	1.03		
5.676	76082	73395	0.965	0.64		
5.712	129494	82303	0.636	1.09		
5.740	72702	60950	0.838	0.61		
5.791	864719	646413	0.748	7.34		

Data File: ...wintarget\data2\chem\40GCSJ.i\030618.b\030618046.D Page 2
Report Date: 22-Mar-2018 13:34

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.885	76602	58309	0.761	0.65	
5.925	46537	28076	0.603	0.39	
5.994	69953	65994	0.943	0.59	
6.027	514088	428155	0.833	4.36	7 Aroclor 1254
6.056	378012	325072	0.860	3.20	
6.169	154351	125775	0.815	1.31	
6.228	120109	63109	0.525	1.01	
6.300	701502	599827	0.855	5.95	7 Aroclor 1254
6.335	126230	111073	0.880	1.07	
6.375	74344	60647	0.816	0.63	
6.443	34233	19876	0.581	0.29	
6.512	29729	27649	0.930	0.25	
6.536	235571	178655	0.758	2.00	
6.607	36815	29017	0.788	0.31	
6.719	47979	41305	0.861	0.40	
6.759	31561	26431	0.837	0.26	
6.803	129897	110305	0.849	1.10	
6.844	33654	23973	0.712	0.28	
6.948	16896	13072	0.774	0.14	
7.005	90785	79102	0.871	0.77	
7.138	27852	16095	0.578	0.23	
7.253	92553	68397	0.739	0.78	
7.776	24601	16237	0.660	0.20	
7.900	9733	7403	0.761	0.08	
8.278	11455	11514	1.005	0.09	
8.380	27407	19837	0.724	0.23	
8.573	1220451	1043440	0.855	10.36	
8.945	34590	22887	0.662	0.29	
9.468	24800	20749	0.837	0.21	
9.962	27633	19326	0.699	0.23	
	11778321	10154991		100.000	

Total unknown % area = 74

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618046.D
Lab Smp Id: 1254-CAL1C Client Smp ID: 1254-CAL1C
Inj Date : 06-MAR-2018 16:57
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-cal1c, 188135:1
Misc Info : 17498,NSAR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 16:57 Cal File: 030618046.D
Als bottle: 24 Calibration Sample, Level: 1
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1254.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
7 Aroclor 1254					CAS #: 11097-69-1			
5.163	5.162	0.001	578337	0.10000	0.11	0.00- 20.00	100.00(M)	BA
5.408	5.408	0.000	461675	0.10000	0.11	0.00- 20.00	79.83	BA
5.510	5.510	0.000	843524	0.10000	0.11	0.00- 20.00	145.85	BA
6.026	6.027	-0.001	514088	0.10000	0.10	0.00- 20.00	88.89	BA
6.299	6.299	0.000	701502	0.10000	0.11	0.00- 20.00	121.30	BA
Average of Peak Amounts =					0.10800			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618046.D

Date : 06-MAR-2018 16:57

Client ID: 1254-CAL1C

Instrument: 40GCSJ.i

Sample Info: 1254-CAL1C, 188135:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618046.D (ManInt: NSAR1254.sub Orig: NSAR1254.sub)
Injection Date: 06-MAR-2018 16:57 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1254-CAL1C





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1254	575308	578337	BA
Aroclor 1254	575308	578337	BA
Aroclor 1254	575308	578337	BA
Aroclor 1254	575308	578337	BA
Aroclor 1254	575308	578337	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618048.D
Lab Smp Id: 1254-CAL2C Client Smp ID: 1254-CAL2C
Inj Date : 06-MAR-2018 17:13
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-cal2c, 185553:1
Misc Info : 17498,NSAR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 17:13 Cal File: 030618048.D
Als bottle: 25 Calibration Sample, Level: 2
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1254.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.180	14785	11446	0.774	0.06		
1.428	22932	16453	0.717	0.10		
2.008	13252	7547	0.570	0.05		
2.484	2443152	2335532	0.956	11.18		
2.785	38923	36879	0.947	0.17		
3.317	15688	15090	0.962	0.06		
3.816	33314	23152	0.695	0.14		
3.936	25391	18861	0.743	0.11		
3.991	19776	14851	0.751	0.08		
4.057	14294	10125	0.708	0.06		
4.215	518990	504810	0.973	2.30		
4.262	130759	134614	1.029	0.57		
4.417	318089	306882	0.965	1.40		
4.453	36792	31401	0.853	0.16		
4.546	146634	140842	0.961	0.64		
4.594	9667	9066	0.938	0.04		
4.619	30811	24715	0.802	0.13		
4.679	14300	9754	0.682	0.06		
4.812	124031	118773	0.958	0.54		
4.848	455414	441589	0.970	2.01		
4.894	1053205	858019	0.815	4.66		
4.957	173883	154163	0.887	0.77		
5.038	101462	94180	0.928	0.44		
5.100	607083	520373	0.857	2.69		
5.163	1130516	1047365	0.926	5.01	7 Aroclor 1254	BA
5.213	500908	455276	0.909	2.22		
5.257	17625	16702	0.948	0.07		
5.285	29337	27985	0.954	0.13		
5.309	96003	86176	0.898	0.42		
5.363	431219	406625	0.943	1.91		
5.407	916372	831698	0.908	4.06	7 Aroclor 1254	BA
5.447	306543	274508	0.895	1.35		
5.478	113703	114036	1.003	0.50		
5.510	1657392	1480187	0.893	7.34	7 Aroclor 1254	BA
5.579	19853	16482	0.830	0.08		
5.605	231410	203474	0.879	1.02		
5.675	151855	139702	0.920	0.67		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.713	234337	153331	0.654	1.03	
5.740	130944	119042	0.909	0.58	
5.790	1692901	1288520	0.761	7.50	
5.884	152797	115170	0.754	0.67	
5.927	77749	50774	0.653	0.34	
5.994	139616	133887	0.959	0.61	
6.026	1030270	853221	0.828	4.56	7 Aroclor 1254
6.056	727544	649744	0.893	3.22	
6.086	209395	142010	0.678	0.92	
6.169	306146	246695	0.806	1.35	
6.228	123149	122142	0.992	0.54	
6.241	143819	124934	0.869	0.63	
6.299	1396395	1154592	0.827	6.18	7 Aroclor 1254
6.334	246981	213346	0.864	1.09	
6.375	147488	117497	0.797	0.65	
6.438	46079	24168	0.524	0.20	
6.512	58184	55972	0.962	0.25	
6.537	454237	352793	0.777	2.01	
6.605	73055	56160	0.769	0.32	
6.660	18073	13693	0.758	0.08	
6.720	93177	76403	0.820	0.41	
6.760	56477	50956	0.902	0.25	
6.802	250222	217408	0.869	1.10	
6.845	76101	47284	0.621	0.33	
6.945	29987	22146	0.739	0.13	
7.005	187120	153304	0.819	0.82	
7.252	146914	128748	0.876	0.65	
7.283	43303	26667	0.616	0.19	
7.557	16777	10505	0.626	0.07	
7.903	15055	9793	0.650	0.06	
8.278	21586	17365	0.804	0.09	
8.572	2251771	1853841	0.823	9.97	
	22563007	19541444		100.000	

Total unknown % area = 73

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618048.D
Lab Smp Id: 1254-CAL2C Client Smp ID: 1254-CAL2C
Inj Date : 06-MAR-2018 17:13
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-cal2c, 185553:1
Misc Info : 17498,NSAR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 17:13 Cal File: 030618048.D
Als bottle: 25 Calibration Sample, Level: 2
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1254.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
7 Aroclor 1254					CAS #: 11097-69-1			
5.162	5.162	0.000	1130515	0.20000	0.22	0.00- 20.00	100.00(M)	BA
5.407	5.408	-0.001	916371	0.20000	0.21	0.00- 20.00	81.06	BA
5.510	5.510	0.000	1657392	0.20000	0.21	0.00- 20.00	146.61	BA
6.026	6.027	-0.001	1030270	0.20000	0.21	0.00- 20.00	91.13	BA
6.299	6.299	0.000	1396394	0.20000	0.21	0.00- 20.00	123.52	BA
Average of Peak Amounts =					0.21200			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618048.D
Date : 06-MAR-2018 17:13
Client ID: 1254-CAL2C
Sample Info: 1254-CAL2C, 185553:1

Instrument: 40GCSJ.i

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618048.D (ManInt: NSAR1254.sub Orig: NSAR1254.sub)
Injection Date: 06-MAR-2018 17:13 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1254-CAL2C





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1254	1126550	1130515	BA
Aroclor 1254	1126550	1130515	BA
Aroclor 1254	1126550	1130515	BA
Aroclor 1254	1126550	1130515	BA
Aroclor 1254	1126550	1130515	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618050.D
Lab Smp Id: 1254-CAL3C Client Smp ID: 1254-CAL3C
Inj Date : 06-MAR-2018 17:30
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-cal3c, 185554:1
Misc Info : 17498,NSAR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 17:30 Cal File: 030618050.D
Als bottle: 26 Calibration Sample, Level: 3
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1254.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.182	51337	40044	0.780	0.09		
1.429	39957	34317	0.859	0.07		
1.537	17914	13995	0.781	0.03		
1.685	15629	15398	0.985	0.02		
2.011	29599	20853	0.705	0.05		
2.485	6057898	5851038	0.966	12.02		
2.686	9516	7676	0.807	0.01		
2.787	95008	87358	0.919	0.18		
2.851	23819	20877	0.877	0.04		
3.052	13713	14987	1.093	0.02		
3.316	31189	30919	0.991	0.05		
3.344	21096	17017	0.807	0.04		
3.515	29295	21880	0.747	0.05		
3.610	20788	17635	0.848	0.03		
3.728	13013	9960	0.765	0.02		
3.815	75390	52210	0.693	0.14		
3.933	62381	42408	0.680	0.11		
3.996	40445	30149	0.745	0.07		
4.057	24255	18114	0.747	0.04		
4.215	1217908	1150954	0.945	2.32		
4.262	330320	319433	0.967	0.63		
4.292	64830	48630	0.750	0.12		
4.418	748177	726093	0.970	1.42		
4.452	73920	66424	0.899	0.14		
4.545	335088	319422	0.953	0.63		
4.592	16526	15457	0.935	0.03		
4.619	59193	52479	0.887	0.11		
4.685	17734	16983	0.958	0.03		
4.770	20117	19880	0.988	0.03		
4.813	284745	275889	0.969	0.54		
4.849	1084246	1045959	0.965	2.06		
4.895	2405586	1902763	0.791	4.58		
4.957	390371	342893	0.878	0.74		
5.037	238296	218476	0.917	0.45		
5.100	1405616	1219340	0.867	2.68		
5.164	2560293	2304428	0.900	4.88	7 Aroclor 1254	BA
5.213	1173729	1059080	0.902	2.23		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.285	60363	60152	0.996	0.11	
5.309	214942	193512	0.900	0.41	
5.364	1020588	935288	0.916	1.94	
5.407	2112521	1886528	0.893	4.03	7 Aroclor 1254
5.447	733518	649777	0.886	1.39	
5.479	265192	257135	0.970	0.50	
5.511	3811784	3385999	0.888	7.27	7 Aroclor 1254
5.579	32531	33382	1.026	0.06	
5.605	530818	484241	0.912	1.01	
5.676	348702	323329	0.927	0.66	
5.714	533476	346229	0.649	1.01	
5.742	324759	270812	0.834	0.61	
5.791	3843360	2976827	0.775	7.33	
5.886	336244	269243	0.801	0.64	
5.927	179445	112871	0.629	0.34	
5.994	317667	298687	0.940	0.60	
6.026	2413214	1944030	0.806	4.60	7 Aroclor 1254
6.056	1748224	1477442	0.845	3.33	
6.170	686623	578812	0.843	1.31	
6.229	279467	280903	1.005	0.53	
6.242	317134	279276	0.881	0.60	
6.300	3184410	2766550	0.869	6.07	7 Aroclor 1254
6.334	582664	515473	0.885	1.11	
6.374	330007	275263	0.834	0.62	
6.438	89310	51831	0.580	0.17	
6.487	18679	18109	0.970	0.03	
6.511	122714	122347	0.997	0.23	
6.537	1081766	870038	0.804	2.06	
6.607	157551	128075	0.813	0.30	
6.662	30905	25338	0.820	0.05	
6.718	210754	178569	0.847	0.40	
6.761	129400	110612	0.855	0.24	
6.802	599534	539615	0.900	1.14	
6.844	171759	108076	0.629	0.32	
6.945	57172	47423	0.829	0.10	
7.006	434841	362564	0.834	0.82	
7.077	18727	15733	0.840	0.03	
7.130	24196	17055	0.705	0.04	
7.252	342578	310115	0.905	0.65	
7.282	72487	60104	0.829	0.13	
7.390	19578	16122	0.823	0.03	
7.445	35210	22193	0.630	0.06	
7.553	26390	18609	0.705	0.05	
7.694	16316	14388	0.882	0.03	
7.729	31518	23387	0.742	0.06	
7.778	19445	16730	0.860	0.03	
7.798	20263	16018	0.790	0.03	
7.904	22590	18864	0.835	0.04	
8.276	52953	39504	0.746	0.10	
8.382	28017	19746	0.705	0.05	
8.572	5213644	4428280	0.849	9.94	
8.943	21557	16555	0.768	0.04	
9.466	15645	11531	0.737	0.02	
9.966	18425	14389	0.781	0.03	
	52410519	45695099		100.000	

Total unknown % area = 73

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618050.D
Lab Smp Id: 1254-CAL3C Client Smp ID: 1254-CAL3C
Inj Date : 06-MAR-2018 17:30
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-cal3c, 185554:1
Misc Info : 17498,NSAR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 17:30 Cal File: 030618050.D
Als bottle: 26 Calibration Sample, Level: 3
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1254.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
7 Aroclor 1254					CAS #: 11097-69-1			
5.163	5.162	0.001	2560292	0.50000	0.49	0.00- 20.00	100.00(M)	BA
5.407	5.408	-0.001	2112521	0.50000	0.49	0.00- 20.00	82.51	BA
5.510	5.510	0.000	3811784	0.50000	0.49	0.00- 20.00	148.88	BA
6.026	6.027	-0.001	2413214	0.50000	0.50	0.00- 20.00	94.26	BA
6.299	6.299	0.000	3184409	0.50000	0.49	0.00- 20.00	124.38	BA
Average of Peak Amounts =					0.49200			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618050.D

Date : 06-MAR-2018 17:30

Client ID: 1254-CAL3C

Instrument: 40GCSJ.i

Sample Info: 1254-CAL3C, 185554:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618050.D (ManInt: NSAR1254.sub Orig: NSAR1254.sub)
Injection Date: 06-MAR-2018 17:30 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1254-CAL3C





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1254	2551487	2560292	BA
Aroclor 1254	2551487	2560292	BA
Aroclor 1254	2551487	2560292	BA
Aroclor 1254	2551487	2560292	BA
Aroclor 1254	2551487	2560292	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618052.D
Lab Smp Id: 1254-CAL4C Client Smp ID: 1254-CAL4C
Inj Date : 06-MAR-2018 17:46
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-cal4c, 185555:1
Misc Info : 17498,NSAR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 17:46 Cal File: 030618052.D
Als bottle: 27 Calibration Sample, Level: 4
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1254.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.181	20338	17643	0.867	0.02		
1.428	65922	60336	0.915	0.07		
1.538	31776	24460	0.770	0.03		
2.485	11777090	11571932	0.983	14.42		
2.787	164077	161759	0.986	0.19		
2.852	35610	28522	0.801	0.04		
3.052	34848	31099	0.892	0.04		
3.318	49721	48315	0.972	0.05		
3.345	27672	27798	1.005	0.03		
3.434	12631	8647	0.685	0.01		
3.512	47159	33510	0.711	0.05		
3.609	31888	32717	1.026	0.03		
3.732	12093	11883	0.983	0.01		
3.757	10828	11345	1.048	0.01		
3.814	107287	73196	0.682	0.12		
3.934	94015	65238	0.694	0.11		
3.996	37751	35743	0.947	0.04		
4.059	25966	21513	0.829	0.03		
4.136	10670	12088	1.133	0.01		
4.215	1826293	1728963	0.947	2.16		
4.262	495863	484615	0.977	0.58		
4.293	82438	62200	0.755	0.09		
4.418	1148007	1108250	0.965	1.35		
4.454	124829	105675	0.847	0.14		
4.545	520797	506348	0.972	0.61		
4.592	22494	21617	0.961	0.02		
4.617	91512	86993	0.951	0.10		
4.682	29611	23905	0.807	0.03		
4.738	13639	11326	0.830	0.01		
4.768	31653	29681	0.938	0.03		
4.813	436264	431646	0.989	0.51		
4.848	1661944	1593892	0.959	1.96		
4.895	3639982	2884202	0.792	4.30		
4.958	608805	546410	0.898	0.72		
5.038	362095	340028	0.939	0.42		
5.100	2123049	1825949	0.860	2.51		
5.163	3886482	3507162	0.902	4.59	7 Aroclor 1254	BA

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.213	1776836	1628217	0.916	2.10	
5.284	96560	90819	0.941	0.11	
5.309	325641	297506	0.914	0.38	
5.364	1556302	1425397	0.916	1.84	
5.407	3251705	3006981	0.925	3.84	7 Aroclor 1254
5.446	1124584	1040705	0.925	1.33	
5.479	401904	378797	0.943	0.47	
5.510	5821617	5399486	0.927	6.88	7 Aroclor 1254
5.575	58238	53397	0.917	0.06	
5.604	831473	749483	0.901	0.98	
5.676	530127	502418	0.948	0.62	
5.713	822568	545715	0.663	0.97	
5.742	512983	417833	0.815	0.60	
5.791	5870686	4701193	0.801	6.94	
5.885	522465	412152	0.789	0.61	
5.927	275865	173515	0.629	0.32	
5.994	487921	462752	0.948	0.57	
6.027	3710616	3100778	0.836	4.38	7 Aroclor 1254
6.057	2650507	2221374	0.838	3.13	
6.169	1066850	928060	0.870	1.26	
6.228	444612	436676	0.982	0.52	
6.243	480121	416542	0.868	0.56	
6.300	4897868	4117516	0.841	5.79	7 Aroclor 1254
6.334	897566	789045	0.879	1.06	
6.374	507098	430542	0.849	0.59	
6.441	124903	73312	0.587	0.14	
6.487	27937	25771	0.922	0.03	
6.510	183907	185808	1.010	0.21	
6.536	1677948	1278798	0.762	1.98	
6.605	242982	201019	0.827	0.28	
6.661	41322	38186	0.924	0.04	
6.719	331298	275607	0.832	0.39	
6.758	194741	175185	0.900	0.23	
6.802	936736	798625	0.853	1.10	
6.844	260030	171830	0.661	0.30	
6.907	13593	12018	0.884	0.01	
6.942	85729	72281	0.843	0.10	
7.006	675731	578164	0.856	0.79	
7.077	32184	26109	0.811	0.03	
7.122	28643	17643	0.616	0.03	
7.253	534471	485017	0.907	0.63	
7.282	121561	99258	0.817	0.14	
7.331	10539	7109	0.675	0.01	
7.390	27785	21939	0.790	0.03	
7.442	42859	32217	0.752	0.05	
7.555	39938	25488	0.638	0.04	
7.691	21961	21683	0.987	0.02	
7.727	55370	42642	0.770	0.06	
7.799	40475	27058	0.669	0.04	
7.900	38857	28824	0.742	0.04	
8.033	10655	8635	0.810	0.01	
8.277	84696	72245	0.853	0.10	
8.379	22615	16331	0.722	0.02	
8.571	9861374	8409353	0.853	11.66	
8.942	30130	19515	0.648	0.03	
9.085	14868	10676	0.718	0.01	
9.470	49128	26719	0.544	0.05	

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
9.961	49432	33887	0.686	0.05	
	84540213	74622457		100.000	

Total unknown % area = 74

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618052.D
Lab Smp Id: 1254-CAL4C Client Smp ID: 1254-CAL4C
Inj Date : 06-MAR-2018 17:46
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-cal4c, 185555:1
Misc Info : 17498,NSAR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 17:46 Cal File: 030618052.D
Als bottle: 27 Calibration Sample, Level: 4
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1254.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
7 Aroclor 1254					CAS #: 11097-69-1			
5.162	5.162	0.000	3886481	0.80000	0.74	0.00- 20.00	100.00(M)	BA
5.407	5.408	-0.001	3251705	0.80000	0.76	0.00- 20.00	83.67	BA
5.509	5.510	-0.001	5821617	0.80000	0.75	0.00- 20.00	149.79	BA
6.027	6.027	0.000	3710615	0.80000	0.76	0.00- 20.00	95.47	BA
6.299	6.299	0.000	4897867	0.80000	0.75	0.00- 20.00	126.02	BA
Average of Peak Amounts =					0.75200			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618052.D

Date : 06-MAR-2018 17:46

Client ID: 1254-CAL4C

Instrument: 40GCSJ.i

Sample Info: 1254-CAL4C, 185555:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618052.D (ManInt: NSAR1254.sub Orig: NSAR1254.sub)
Injection Date: 06-MAR-2018 17:46 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1254-CAL4C





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1254	3881378	3886481	BA
Aroclor 1254	3881378	3886481	BA
Aroclor 1254	3881378	3886481	BA
Aroclor 1254	3881378	3886481	BA
Aroclor 1254	3881378	3886481	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618054.D
Lab Smp Id: 1254-CAL5C Client Smp ID: 1254-CAL5C
Inj Date : 06-MAR-2018 18:02
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-cal5c, 185556:1
Misc Info : 17498,NSAR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 18:02 Cal File: 030618054.D
Als bottle: 28 Calibration Sample, Level: 5
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1254.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.181	15243	11962	0.785	0.01		
1.428	96747	89041	0.920	0.08		
1.540	40825	35144	0.861	0.03		
1.682	9327	8894	0.954	0.00		
2.008	10288	9685	0.941	0.00		
2.412	10350	7112	0.687	0.00		
2.485	17725886	17260144	0.974	16.66		
2.787	243100	236365	0.972	0.22		
2.851	36673	32993	0.900	0.03		
3.053	49195	43436	0.883	0.04		
3.319	55642	53004	0.953	0.05		
3.344	37001	30263	0.818	0.03		
3.514	61454	41534	0.676	0.05		
3.611	48883	46459	0.950	0.04		
3.734	9700	13664	1.409	0.00		
3.755	14530	9896	0.681	0.01		
3.814	160653	100699	0.627	0.14		
3.935	113785	77754	0.683	0.10		
3.993	48360	45579	0.942	0.04		
4.058	31144	26153	0.840	0.02		
4.136	14642	14082	0.962	0.01		
4.215	2246411	2105306	0.937	2.04		
4.262	624542	621927	0.996	0.56		
4.292	114086	82473	0.723	0.10		
4.418	1403386	1353229	0.964	1.27		
4.453	133440	127351	0.954	0.12		
4.546	635904	613877	0.965	0.57		
4.594	11699	17033	1.456	0.01		
4.617	92187	94751	1.028	0.08		
4.683	32935	28854	0.876	0.03		
4.735	16261	13245	0.815	0.01		
4.770	37428	36501	0.975	0.03		
4.814	542039	547513	1.010	0.49		
4.848	2045079	1950354	0.954	1.86		
4.896	4464777	3447952	0.772	4.07		
4.957	749670	674930	0.900	0.68		
5.037	447740	427307	0.954	0.40		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.100	2603383	2206600	0.848	2.37	
5.163	4761607	4381606	0.920	4.34	7 Aroclor 1254
5.212	2175024	1931040	0.888	1.98	
5.284	116466	109643	0.941	0.10	
5.309	404504	368472	0.911	0.36	
5.364	1900530	1737928	0.914	1.73	
5.407	4016995	3641293	0.906	3.66	7 Aroclor 1254
5.446	1380053	1254397	0.909	1.25	
5.477	506726	475165	0.938	0.46	
5.511	7207531	6391511	0.887	6.57	7 Aroclor 1254
5.577	67153	63150	0.940	0.06	
5.605	1023947	924404	0.903	0.93	
5.676	661316	643905	0.974	0.60	
5.714	1030048	680082	0.660	0.93	
5.740	635529	535004	0.842	0.57	
5.792	7230180	5601138	0.775	6.59	
5.886	645954	520462	0.806	0.58	
5.927	344827	212793	0.617	0.31	
5.995	613062	588236	0.960	0.55	
6.027	4591491	3986858	0.868	4.18	7 Aroclor 1254
6.056	3223664	2733295	0.848	2.93	
6.169	1328008	1109796	0.836	1.21	
6.228	1042954	573707	0.550	0.95	
6.300	6047377	5045174	0.834	5.51	7 Aroclor 1254
6.334	1117787	991343	0.887	1.01	
6.375	640740	547643	0.855	0.58	
6.437	148401	91961	0.620	0.13	
6.511	228443	228993	1.002	0.20	
6.537	2072914	1543722	0.745	1.89	
6.605	305742	240734	0.787	0.27	
6.662	54744	45287	0.827	0.04	
6.719	410032	340851	0.831	0.37	
6.759	242534	215295	0.888	0.22	
6.803	1169372	1018756	0.871	1.06	
6.844	311718	214607	0.688	0.28	
6.945	110710	91840	0.830	0.10	
7.006	832801	710785	0.853	0.75	
7.076	39055	35215	0.902	0.03	
7.124	42412	24716	0.583	0.03	
7.253	667072	625157	0.937	0.60	
7.283	149721	120610	0.806	0.13	
7.392	31016	25016	0.807	0.02	
7.444	53853	38361	0.712	0.04	
7.554	43717	29353	0.671	0.03	
7.693	32182	28611	0.889	0.02	
7.731	75133	56693	0.755	0.06	
7.801	54585	36993	0.678	0.04	
7.902	45857	33249	0.725	0.04	
8.278	123330	103442	0.839	0.11	
8.379	26036	20581	0.790	0.02	
8.572	14526385	12123277	0.835	13.24	
8.943	32605	19326	0.593	0.02	
9.087	19995	15238	0.762	0.01	
9.471	36056	22732	0.630	0.03	
9.962	37005	22458	0.607	0.03	
	109671291	95714970		100.000	

Total unknown % area = 76

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618054.D
Lab Smp Id: 1254-CAL5C Client Smp ID: 1254-CAL5C
Inj Date : 06-MAR-2018 18:02
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-cal5c, 185556:1
Misc Info : 17498,NSAR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 18:02 Cal File: 030618054.D
Als bottle: 28 Calibration Sample, Level: 5
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: NSAR1254.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
7 Aroclor 1254					CAS #: 11097-69-1			
5.163	5.162	0.001	4761607	1.00000	0.91	0.00- 20.00	100.00(M)	BA
5.407	5.408	-0.001	4016994	1.00000	0.93	0.00- 20.00	84.36	BA
5.510	5.510	0.000	7207530	1.00000	0.93	0.00- 20.00	151.37	BA
6.027	6.027	0.000	4591491	1.00000	0.94	0.00- 20.00	96.43	BA
6.299	6.299	0.000	6047377	1.00000	0.93	0.00- 20.00	127.00	BA
Average of Peak Amounts =					0.92800			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618054.D

Date : 06-MAR-2018 18:02

Client ID: 1254-CAL5C

Instrument: 40GCSJ.i

Sample Info: 1254-CAL5C, 185556:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618054.D (ManInt: NSAR1254.sub Orig: NSAR1254.sub)
Injection Date: 06-MAR-2018 18:02 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1254-CAL5C





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1254	4578985	4761607	BA
Aroclor 1254	4578985	4761607	BA
Aroclor 1254	4578985	4761607	BA
Aroclor 1254	4578985	4761607	BA
Aroclor 1254	4578985	4761607	BA

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618058.D
Lab Smp Id: 1660-CAL1A Client Smp ID: 1660-CAL1A
Inj Date : 06-MAR-2018 18:34
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1660-cal1a, 188136:1
Misc Info : 17498,AR1660
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 18:34 Cal File: 030618058.D
Als bottle: 30 Calibration Sample, Level: 1
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1660.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.184	75769	59956	0.791	0.39		
1.428	18267	12499	0.684	0.09		
1.686	19257	17034	0.885	0.10		
2.010	41793	33612	0.804	0.22		
2.405	34035	25626	0.753	0.17		
2.485	1265501	1240700	0.980	7.14	$ 2 Tetrachloro-m-xylene	
2.687	45905	38796	0.845	0.24		
2.791	69017	64076	0.928	0.36		
2.850	244595	230348	0.942	1.29		
3.030	26125	23948	0.917	0.13		
3.053	41274	35596	0.862	0.21		
3.319	334670	316544	0.946	1.76	5 Aroclor 1016	BA
3.345	198504	179987	0.907	1.04		
3.434	61488	51551	0.838	0.32		
3.514	336132	236249	0.703	1.77	5 Aroclor 1016	BA
3.732	87942	87307	0.993	0.46		
3.756	56569	52611	0.930	0.29		
3.827	961910	568396	0.591	5.07	5 Aroclor 1016	BA
3.929	407299	336226	0.826	2.14	5 Aroclor 1016	BA
3.995	291299	270962	0.930	1.53	5 Aroclor 1016	BA
4.057	91678	78986	0.862	0.48		
4.137	38165	30755	0.806	0.20		
4.215	245118	232129	0.947	1.29		
4.262	215370	204306	0.949	1.13		
4.294	97069	97854	1.008	0.51		
4.311	121075	108897	0.899	0.63		
4.418	296094	277838	0.938	1.56		
4.452	228098	168917	0.741	1.20		
4.546	376428	345686	0.918	1.98		
4.618	89285	79390	0.889	0.47		
4.725	18650	14732	0.790	0.09		
4.768	15175	13986	0.922	0.08		
4.814	70800	67479	0.953	0.37		
4.847	108064	102590	0.949	0.57		
4.895	365187	264463	0.724	1.92		
4.959	28108	18371	0.654	0.14		
5.037	24776	19872	0.802	0.13		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.100	75151	62358	0.830	0.39	
5.163	285696	262413	0.919	1.50	
5.214	17586	14232	0.809	0.09	
5.362	21031	19306	0.918	0.11	
5.407	87223	72021	0.826	0.46	
5.479	114689	109517	0.955	0.60	
5.511	195439	178257	0.912	1.03	
5.676	287789	270234	0.939	1.51	
5.710	226208	125913	0.557	1.19	
5.787	739489	648842	0.877	3.90	
5.884	50381	44276	0.879	0.26	
5.931	25620	20410	0.797	0.13	
5.994	121209	109219	0.901	0.63	
6.030	272564	261055	0.958	1.43	
6.056	810958	700847	0.864	4.27	
6.170	497456	309560	0.622	2.62	
6.257	116495	69428	0.596	0.61	
6.300	956999	807332	0.844	5.04	8 Aroclor 1260
6.335	137669	122786	0.892	0.72	
6.372	38735	30140	0.778	0.20	
6.433	122734	90057	0.734	0.64	
6.491	23519	28668	1.219	0.12	
6.511	474348	427351	0.901	2.50	
6.534	82021	92139	1.123	0.43	
6.562	294717	250391	0.850	1.55	
6.608	31351	24800	0.791	0.16	
6.660	93675	84477	0.902	0.49	
6.719	499348	457647	0.916	2.63	
6.760	283474	256610	0.905	1.49	
6.805	227703	192558	0.846	1.20	
6.837	47833	39344	0.823	0.25	
6.867	20777	22616	1.089	0.10	
6.913	37481	33795	0.902	0.19	
6.945	99494	87238	0.877	0.52	
7.006	1074661	952659	0.886	5.66	8 Aroclor 1260
7.034	85801	79294	0.924	0.45	
7.076	39830	29520	0.741	0.21	
7.124	76029	48693	0.640	0.40	
7.253	505238	443533	0.878	2.66	8 Aroclor 1260
7.283	161808	121906	0.753	0.85	
7.368	14051	20223	1.439	0.07	
7.390	249151	219715	0.882	1.31	
7.444	313103	259788	0.830	1.65	8 Aroclor 1260
7.555	28224	19645	0.696	0.14	
7.692	126890	112403	0.886	0.66	
7.727	24298	19768	0.814	0.12	
7.780	47376	23329	0.492	0.24	
7.902	285359	230087	0.806	1.50	8 Aroclor 1260
7.967	28076	18551	0.661	0.14	
8.275	73573	62778	0.853	0.38	
8.382	33580	24066	0.717	0.17	
8.573	1252235	1038307	0.829	6.60	$ 12 Decachlorobiphenyl
8.943	22477	18768	0.835	0.11	
9.467	21684	17832	0.822	0.11	
9.963	27944	21020	0.752	0.14	
	18954741	16115997		100.000	

Total unknown % area = 57

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618058.D
Lab Smp Id: 1660-CAL1A Client Smp ID: 1660-CAL1A
Inj Date : 06-MAR-2018 18:34
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1660-cal1a, 188136:1
Misc Info : 17498,AR1660
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 18:34 Cal File: 030618058.D
Als bottle: 30 Calibration Sample, Level: 1
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1660.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.485	2.484	0.001	1265501	0.01000	0.010			
5 Aroclor 1016					CAS #: 12674-11-2			
3.318	3.318	0.000	334670	0.10000	0.11	0.00- 20.00	100.00(M)	BA
3.513	3.512	0.001	336131	0.10000	0.11	0.00- 20.00	100.44	BA
3.826	3.826	0.000	961909	0.10000	0.11	0.00- 20.00	287.42	BA
3.928	3.929	-0.001	407299	0.10000	0.11	0.00- 20.00	121.70	BA
3.995	3.994	0.001	291298	0.10000	0.11	0.00- 20.00	87.04	BA
Average of Peak Amounts =					0.11000			
8 Aroclor 1260					CAS #: 11096-82-5			
6.300	6.300	0.000	956999	0.10000	0.11	0.00- 20.00	100.00(M)	CO
7.006	7.006	0.000	1074661	0.10000	0.11	0.00- 20.00	112.29	CO
7.252	7.251	0.001	505237	0.10000	0.10	0.00- 20.00	52.79	CO
7.444	7.443	0.001	313102	0.10000	0.11	0.00- 20.00	32.72	CO
7.902	7.901	0.001	285359	0.10000	0.11	0.00- 20.00	29.82	CO
Average of Peak Amounts =					0.10800			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.572	8.571	0.001	1252234	0.01000	0.011			

QC Flag Legend

M - Compound response manually integrated.

QC Flag Legend

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.
CO: Indicates that the analyst had to split two co-eluting peaks apart that were not (or could not be) separated by the computer system.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618058.D

Date : 06-MAR-2018 18:34

Client ID: 1660-CAL1A

Instrument: 40GCSJ.i

Sample Info: 1660-CAL1A, 188136:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618058.D (ManInt: AR1660.sub Orig: AR1660.sub)
Injection Date: 06-MAR-2018 18:34 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1660-CAL1A





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	331398	334670	BA
Aroclor 1016	331398	334670	BA
Aroclor 1016	331398	334670	BA
Aroclor 1016	331398	334670	BA
Aroclor 1016	331398	334670	BA
Aroclor 1260	948946	956999	CO
Aroclor 1260	948946	956999	CO
Aroclor 1260	948946	956999	CO
Aroclor 1260	948946	956999	CO
Aroclor 1260	948946	956999	CO
Tetrachloro-m-xylene	1265501	1265501	
Decachlorobiphenyl	1252234	1252234	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618060.D
Lab Smp Id: 1660-CAL2A Client Smp ID: 1660-CAL2A
Inj Date : 06-MAR-2018 18:51
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1660-cal2a, 185558:1
Misc Info : 17498,AR1660
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 18:51 Cal File: 030618060.D
Als bottle: 31 Calibration Sample, Level: 2
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1660.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.185	55176	41178	0.746	0.15		
1.429	21015	18585	0.884	0.05		
1.802	12282	8623	0.702	0.03		
2.011	32498	25901	0.797	0.09		
2.404	67139	48192	0.718	0.18		
2.484	2499487	2410686	0.964	7.40	$ 2 Tetrachloro-m-xylene	
2.687	94590	70872	0.749	0.26		
2.789	132356	122708	0.927	0.36		
2.851	469229	448920	0.957	1.30		
3.028	20462	19526	0.954	0.05		
3.054	77836	65563	0.842	0.21		
3.280	13769	10123	0.735	0.03		
3.318	654061	624883	0.955	1.81	5 Aroclor 1016	BA
3.345	373017	342479	0.918	1.03		
3.436	104520	94613	0.905	0.28		
3.513	645015	465357	0.721	1.78	5 Aroclor 1016	BA
3.614	11902	11206	0.942	0.03		
3.697	20307	21558	1.062	0.05		
3.732	166970	164519	0.985	0.46		
3.756	105123	106344	1.012	0.29		
3.827	1858995	1113513	0.599	5.15	5 Aroclor 1016	BA
3.929	794562	664351	0.836	2.20	5 Aroclor 1016	BA
3.995	538852	516167	0.958	1.49	5 Aroclor 1016	BA
4.059	153652	141937	0.924	0.42		
4.137	64387	55721	0.865	0.17		
4.215	463064	450944	0.974	1.28		
4.261	411934	395299	0.960	1.14		
4.293	176764	178584	1.010	0.48		
4.310	214978	205952	0.958	0.59		
4.368	18687	13075	0.700	0.05		
4.417	569211	550417	0.967	1.57		
4.452	428711	317474	0.741	1.18		
4.546	747149	680932	0.911	2.07		
4.617	172002	154821	0.900	0.47		
4.675	18051	11895	0.659	0.05		
4.728	31765	25328	0.797	0.08		
4.768	26719	24282	0.909	0.07		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.814	134761	128509	0.954	0.37	
4.849	207904	199310	0.959	0.57	
4.895	688062	504752	0.734	1.90	
4.955	49261	36191	0.735	0.13	
5.037	41204	35409	0.859	0.11	
5.101	147468	124667	0.845	0.40	
5.163	553608	515848	0.932	1.53	
5.211	28275	23336	0.825	0.07	
5.364	42478	33276	0.783	0.11	
5.406	154225	140226	0.909	0.42	
5.478	222337	212066	0.954	0.61	
5.510	375280	336790	0.897	1.03	
5.676	557330	520345	0.934	1.54	
5.710	420039	227627	0.542	1.16	
5.786	1434098	1216006	0.848	3.97	
5.881	99936	84741	0.848	0.27	
5.931	52909	37411	0.707	0.14	
5.994	236171	220443	0.933	0.65	
6.029	525050	521852	0.994	1.45	
6.055	1579576	1350262	0.855	4.37	
6.170	972431	621038	0.639	2.69	
6.256	228441	130747	0.572	0.63	
6.301	1864709	1458728	0.782	5.16	8 Aroclor 1260
6.335	262275	234816	0.895	0.72	
6.375	82478	55139	0.669	0.22	
6.433	206268	167261	0.811	0.57	
6.487	40545	49125	1.212	0.11	
6.510	932649	849322	0.911	2.58	
6.536	166259	180771	1.087	0.46	
6.561	564443	492928	0.873	1.56	
6.605	64137	48977	0.764	0.17	
6.661	181400	157714	0.869	0.50	
6.719	981159	893423	0.911	2.71	
6.759	544707	485293	0.891	1.50	
6.803	436009	386944	0.887	1.20	
6.839	89543	74514	0.832	0.24	
6.868	45779	37984	0.830	0.12	
6.913	64712	61547	0.951	0.17	
6.943	192241	174196	0.906	0.53	
7.006	2085870	1795515	0.861	5.78	8 Aroclor 1260
7.034	145420	142383	0.979	0.40	
7.075	69040	53643	0.777	0.19	
7.122	116391	79463	0.683	0.32	
7.252	997245	885990	0.888	2.76	8 Aroclor 1260
7.284	293602	234376	0.798	0.81	
7.390	506113	423868	0.837	1.40	
7.444	607583	530361	0.873	1.68	8 Aroclor 1260
7.553	47904	35161	0.734	0.13	
7.692	248075	211779	0.854	0.68	
7.729	41616	33711	0.810	0.11	
7.777	25315	19290	0.762	0.07	
7.800	38231	26898	0.704	0.10	
7.900	543647	451004	0.830	1.50	8 Aroclor 1260
7.966	45855	31122	0.679	0.12	
8.275	140866	122403	0.869	0.39	
8.382	29898	21612	0.723	0.08	
8.572	2284568	1922634	0.842	6.33	$ 12 Decachlorobiphenyl

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
8.945	25225	17796	0.705	0.06	
9.471	34351	19751	0.575	0.09	
9.964	23512	15629	0.665	0.06	
	36086723	30726451		100.000	

Total unknown % area = 57

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618060.D
Lab Smp Id: 1660-CAL2A Client Smp ID: 1660-CAL2A
Inj Date : 06-MAR-2018 18:51
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1660-cal2a, 185558:1
Misc Info : 17498,AR1660
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 18:51 Cal File: 030618060.D
Als bottle: 31 Calibration Sample, Level: 2
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1660.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.484	2.484	0.000	2499486	0.02000	0.020			
5 Aroclor 1016					CAS #: 12674-11-2			
3.317	3.318	-0.001	654060	0.20000	0.21	0.00- 20.00	100.00(M)	BA
3.513	3.512	0.001	645015	0.20000	0.21	0.00- 20.00	98.62	BA
3.826	3.826	0.000	1858994	0.20000	0.21	0.00- 20.00	284.22	BA
3.929	3.929	0.000	794562	0.20000	0.21	0.00- 20.00	121.48	BA
3.994	3.994	0.000	538852	0.20000	0.21	0.00- 20.00	82.39	BA
Average of Peak Amounts =					0.21000			
8 Aroclor 1260					CAS #: 11096-82-5			
6.300	6.300	0.000	1864709	0.20000	0.21	0.00- 20.00	100.00(M)	CO
7.006	7.006	0.000	2085870	0.20000	0.21	0.00- 20.00	111.86	CO
7.252	7.251	0.001	997245	0.20000	0.21	0.00- 20.00	53.48	CO
7.443	7.443	0.000	607583	0.20000	0.21	0.00- 20.00	32.58	CO
7.900	7.901	-0.001	543647	0.20000	0.21	0.00- 20.00	29.15	CO
Average of Peak Amounts =					0.21000			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.572	8.571	0.001	2284567	0.02000	0.021			

QC Flag Legend

M - Compound response manually integrated.

QC Flag Legend

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.
CO: Indicates that the analyst had to split two co-eluting peaks apart that were not (or could not be) separated by the computer system.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618060.D

Date : 06-MAR-2018 18:51

Client ID: 1660-CAL2A

Instrument: 40GCSJ.i

Sample Info: 1660-CAL2A, 185558:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618060.D (ManInt: AR1660.sub Orig: AR1660.sub)
Injection Date: 06-MAR-2018 18:51 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1660-CAL2A





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	649890	654060	BA
Aroclor 1016	649890	654060	BA
Aroclor 1016	649890	654060	BA
Aroclor 1016	649890	654060	BA
Aroclor 1016	649890	654060	BA
Aroclor 1260	1856530	1864709	CO
Aroclor 1260	1856530	1864709	CO
Aroclor 1260	1856530	1864709	CO
Aroclor 1260	1856530	1864709	CO
Aroclor 1260	1856530	1864709	CO
Tetrachloro-m-xylene	2499486	2499486	
Decachlorobiphenyl	2284567	2284567	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618062.D
Lab Smp Id: 1660-CAL3A Client Smp ID: 1660-CAL3A
Inj Date : 06-MAR-2018 19:07
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1660-cal3a, 185559:1
Misc Info : 17498,AR1660
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 17:30 Cal File: 030618050.D
Als bottle: 32 Calibration Sample, Level: 3
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1660.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.182	73481	56448	0.768	0.08		
1.429	38001	32608	0.858	0.04		
1.537	20915	15456	0.739	0.02		
1.685	19161	19298	1.007	0.02		
1.803	18038	14682	0.814	0.02		
2.010	53413	35717	0.669	0.06		
2.404	140743	111047	0.789	0.16		
2.484	6057237	5892667	0.973	7.63	$ 2 Tetrachloro-m-xylene	
2.687	201241	167032	0.830	0.23		
2.789	304363	285330	0.937	0.35		
2.851	1109963	1029461	0.927	1.31		
3.031	24750	24475	0.989	0.02		
3.055	178487	159145	0.892	0.21		
3.278	26481	21063	0.795	0.03		
3.318	1513499	1414991	0.935	1.78	5 Aroclor 1016	BA
3.345	880540	822356	0.934	1.04		
3.434	247038	220109	0.891	0.29		
3.513	1532387	1090208	0.711	1.81	5 Aroclor 1016	BA
3.611	27202	25734	0.946	0.03		
3.647	29738	19736	0.664	0.03		
3.697	43175	42217	0.978	0.05		
3.732	381498	383355	1.005	0.45		
3.756	273430	243142	0.889	0.32		
3.826	4314630	2564491	0.594	5.10	5 Aroclor 1016	BA
3.929	1875063	1563785	0.834	2.21	5 Aroclor 1016	BA
3.994	1275640	1205112	0.945	1.50	5 Aroclor 1016	BA
4.059	356271	321194	0.902	0.42		
4.137	142724	126589	0.887	0.16		
4.215	1091767	1043921	0.956	1.29		
4.261	983477	936778	0.953	1.16		
4.293	419201	426725	1.018	0.49		
4.310	507212	485823	0.958	0.59		
4.372	34974	24662	0.705	0.04		
4.418	1378494	1285521	0.933	1.63		
4.452	991992	777670	0.784	1.17		
4.545	1720081	1585302	0.922	2.03		
4.591	26254	25937	0.988	0.03		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.618	390639	352630	0.903	0.46	
4.680	45442	27438	0.604	0.05	
4.728	65437	52451	0.802	0.07	
4.768	60652	53450	0.881	0.07	
4.813	315747	306151	0.970	0.37	
4.848	490062	474651	0.969	0.57	
4.895	1609352	1151281	0.715	1.90	
4.957	109800	84138	0.766	0.12	
5.038	90170	81334	0.902	0.10	
5.101	330531	287340	0.869	0.39	
5.162	1311328	1213988	0.926	1.55	
5.213	62406	49170	0.788	0.07	
5.308	18942	13697	0.723	0.02	
5.364	102693	89831	0.875	0.12	
5.407	358555	326663	0.911	0.42	
5.444	17182	16887	0.983	0.02	
5.478	520000	499018	0.960	0.61	
5.510	901921	830532	0.921	1.06	
5.574	21617	18709	0.865	0.02	
5.607	22890	17618	0.770	0.02	
5.675	1323137	1193469	0.902	1.56	
5.710	966061	548770	0.568	1.14	
5.786	3269149	2832194	0.866	3.86	
5.883	222136	199869	0.900	0.26	
5.929	105245	84776	0.806	0.12	
5.994	562299	517914	0.921	0.66	
6.030	1251900	1210482	0.967	1.48	
6.056	3581989	3121113	0.871	4.23	
6.170	2328780	1484198	0.637	2.75	
6.257	533348	316626	0.594	0.63	
6.300	4328765	3602666	0.832	5.11	8 Aroclor 1260
6.332	635873	554609	0.872	0.75	
6.374	166590	132594	0.796	0.19	
6.433	464363	404511	0.871	0.54	
6.487	99285	119042	1.199	0.11	
6.509	2148304	1882394	0.876	2.54	
6.535	394424	422591	1.071	0.46	
6.561	1341525	1196072	0.892	1.58	
6.604	141409	115102	0.814	0.16	
6.661	417039	383638	0.920	0.49	
6.718	2289697	1970989	0.861	2.70	
6.759	1295745	1116495	0.862	1.53	
6.804	1058424	934717	0.883	1.25	
6.838	199169	171583	0.861	0.23	
6.866	100126	89213	0.891	0.11	
6.912	156733	140038	0.893	0.18	
6.944	453234	417139	0.920	0.53	
7.005	4908203	4291027	0.874	5.80	8 Aroclor 1260
7.035	325498	318921	0.980	0.38	
7.075	167281	127630	0.763	0.19	
7.122	238259	189781	0.797	0.28	
7.252	2360655	2040658	0.864	2.79	8 Aroclor 1260
7.282	691639	556699	0.805	0.81	
7.390	1210833	979598	0.809	1.43	
7.444	1443850	1242584	0.861	1.70	8 Aroclor 1260
7.553	121232	82650	0.682	0.14	
7.691	577449	516541	0.895	0.68	

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
7.728	106585	81318	0.763	0.12	
7.777	24225	28669	1.183	0.02	
7.800	83574	56954	0.681	0.09	
7.901	1312523	1129026	0.860	1.55	8 Aroclor 1260
7.965	102499	75394	0.736	0.12	
8.275	325997	297080	0.911	0.38	
8.377	45903	31430	0.685	0.05	
8.572	5312752	4519291	0.851	6.28	$ 12 Decachlorobiphenyl
8.739	26600	11588	0.436	0.03	
8.943	52533	33251	0.633	0.06	
8.993	10265	8473	0.825	0.01	
9.015	10050	7503	0.747	0.01	
9.470	52157	34850	0.668	0.06	
9.961	54045	36108	0.668	0.06	
	84557281	72306592		100.000	

Total unknown % area = 57

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618062.D
Lab Smp Id: 1660-CAL3A Client Smp ID: 1660-CAL3A
Inj Date : 06-MAR-2018 19:07
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1660-cal3a, 185559:1
Misc Info : 17498,AR1660
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 17:30 Cal File: 030618050.D
Als bottle: 32 Calibration Sample, Level: 3
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1660.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.484	2.484	0.000	6057237	0.05000	0.049			
5 Aroclor 1016					CAS #: 12674-11-2			
3.317	3.318	-0.001	1513499	0.50000	0.49	0.00- 20.00	100.00(M)	BA
3.512	3.512	0.000	1532387	0.50000	0.49	0.00- 20.00	101.25	BA
3.826	3.826	0.000	4314629	0.50000	0.49	0.00- 20.00	285.08	BA
3.928	3.929	-0.001	1875063	0.50000	0.50	0.00- 20.00	123.89	BA
3.994	3.994	0.000	1275639	0.50000	0.49	0.00- 20.00	84.28	BA
Average of Peak Amounts =					0.49200			
8 Aroclor 1260					CAS #: 11096-82-5			
6.299	6.300	-0.001	4328764	0.50000	0.49	0.00- 20.00	100.00(M)	CO
7.005	7.006	-0.001	4908203	0.50000	0.49	0.00- 20.00	113.39	CO
7.251	7.251	0.000	2360654	0.50000	0.49	0.00- 20.00	54.53	CO
7.443	7.443	0.000	1443849	0.50000	0.50	0.00- 20.00	33.35	CO
7.901	7.901	0.000	1312522	0.50000	0.50	0.00- 20.00	30.32	CO
Average of Peak Amounts =					0.49400			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.571	8.571	0.000	5312751	0.05000	0.048			

QC Flag Legend

M - Compound response manually integrated.

QC Flag Legend

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.
CO: Indicates that the analyst had to split two co-eluting peaks apart that were not (or could not be) separated by the computer system.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618062.D
Date : 06-MAR-2018 19:07
Client ID: 1660-CAL3A
Sample Info: 1660-CAL3A, 185559:1

Instrument: 40GCSJ.i
Operator: BLM
Column phase: ZB-1ms
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618062.D (ManInt: AR1660.sub Orig: AR1660.sub)
Injection Date: 06-MAR-2018 19:07 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1660-CAL3A





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	1506634	1513499	BA
Aroclor 1016	1506634	1513499	BA
Aroclor 1016	1506634	1513499	BA
Aroclor 1016	1506634	1513499	BA
Aroclor 1016	1506634	1513499	BA
Aroclor 1260	4317043	4328764	CO
Aroclor 1260	4317043	4328764	CO
Aroclor 1260	4317043	4328764	CO
Aroclor 1260	4317043	4328764	CO
Aroclor 1260	4317043	4328764	CO
Tetrachloro-m-xylene	6057237	6057237	
Decachlorobiphenyl	5312751	5312751	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618064.D
Lab Smp Id: 1660-CAL4A Client Smp ID: 1660-CAL4A
Inj Date : 06-MAR-2018 19:23
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1660-cal4a, 185560:1
Misc Info : 17498,AR1660
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:23 Cal File: 030618064.D
Als bottle: 33 Calibration Sample, Level: 4
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1660.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.183	51048	39743	0.779	0.03		
1.428	67658	59771	0.883	0.05		
1.537	35520	27064	0.762	0.02		
1.684	11345	12242	1.079	0.00		
1.804	26637	21421	0.804	0.01		
2.010	30322	24464	0.807	0.02		
2.405	217693	168016	0.772	0.16		
2.485	12180956	11703129	0.961	9.57	$ 2 Tetrachloro-m-xylene	
2.686	324056	263363	0.813	0.24		
2.788	503381	488978	0.971	0.37		
2.851	1715701	1591093	0.927	1.27		
3.033	12533	16182	1.291	0.00		
3.055	262518	238547	0.909	0.19		
3.277	37151	29842	0.803	0.02		
3.318	2333577	2156053	0.924	1.73	5 Aroclor 1016	BA
3.344	1347877	1252078	0.929	1.00		
3.435	374972	344552	0.919	0.27		
3.513	2362745	1662196	0.704	1.75	5 Aroclor 1016	BA
3.610	37684	33923	0.900	0.02		
3.644	36588	23793	0.650	0.02		
3.697	64716	64546	0.997	0.04		
3.733	614707	615053	1.001	0.45		
3.757	393294	383146	0.974	0.29		
3.827	6746065	3969633	0.588	5.01	5 Aroclor 1016	BA
3.929	2864977	2413896	0.843	2.12	5 Aroclor 1016	BA
3.995	1946759	1862950	0.957	1.44	5 Aroclor 1016	BA
4.059	556550	510919	0.918	0.41		
4.136	217971	201589	0.925	0.16		
4.215	1671721	1557184	0.931	1.24		
4.262	1509426	1424890	0.944	1.12		
4.293	644274	673004	1.045	0.47		
4.310	801855	741139	0.924	0.59		
4.372	46455	38102	0.820	0.03		
4.418	2086121	1916658	0.919	1.55		
4.453	1561367	1168414	0.748	1.16		
4.546	2691796	2429747	0.903	2.00		
4.618	614879	582697	0.948	0.45		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.681	40268	29964	0.744	0.02	
4.726	97409	80777	0.829	0.07	
4.769	84631	79528	0.940	0.06	
4.813	496125	477191	0.962	0.36	
4.848	769915	758419	0.985	0.57	
4.895	2431191	1773420	0.729	1.80	
4.957	158097	128688	0.814	0.11	
5.037	141886	125690	0.886	0.10	
5.100	509908	441468	0.866	0.37	
5.163	1980314	1775935	0.897	1.47	
5.213	95707	77428	0.809	0.07	
5.309	25285	19192	0.759	0.01	
5.364	136686	126726	0.927	0.10	
5.407	557655	538024	0.965	0.41	
5.447	22003	24599	1.118	0.01	
5.478	803193	761709	0.948	0.59	
5.511	1404548	1270014	0.904	1.04	
5.606	31361	23033	0.734	0.02	
5.676	2014749	1859939	0.923	1.49	
5.711	1494302	813162	0.544	1.11	
5.786	4990155	4308283	0.863	3.70	
5.884	340816	299117	0.878	0.25	
5.929	164472	137221	0.834	0.12	
5.994	852561	813094	0.954	0.63	
6.031	1912065	1795299	0.939	1.42	
6.057	5492951	4849840	0.883	4.08	
6.169	3549907	2212097	0.623	2.63	
6.255	812378	476866	0.587	0.60	
6.300	6710843	5584923	0.832	4.98	8 Aroclor 1260
6.334	997961	870632	0.872	0.74	
6.374	242400	190548	0.786	0.18	
6.432	705567	608791	0.863	0.52	
6.487	129043	188452	1.460	0.09	
6.511	3293621	2899546	0.880	2.44	
6.536	647596	661803	1.022	0.48	
6.562	2057795	1823223	0.886	1.52	
6.605	225392	179290	0.795	0.16	
6.661	658391	628771	0.955	0.48	
6.719	3522957	3237666	0.919	2.61	
6.760	1974466	1743981	0.883	1.46	
6.804	1651358	1397674	0.846	1.22	
6.839	305859	257806	0.843	0.22	
6.868	154872	136206	0.879	0.11	
6.913	224919	208551	0.927	0.16	
6.944	715915	663774	0.927	0.53	
7.005	7621579	7030639	0.922	5.66	8 Aroclor 1260
7.036	534386	507176	0.949	0.39	
7.076	240666	196245	0.815	0.17	
7.122	345690	285294	0.825	0.25	
7.222	14224	14841	1.043	0.01	
7.252	3661924	3266511	0.892	2.72	8 Aroclor 1260
7.283	1069136	892362	0.835	0.79	
7.390	1852151	1502868	0.811	1.37	
7.443	2216952	1888648	0.852	1.64	8 Aroclor 1260
7.552	171302	132932	0.776	0.12	
7.692	911316	814697	0.894	0.67	
7.728	156368	125258	0.801	0.11	

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
7.777	11035	17805	1.614	0.00	
7.802	122291	99621	0.815	0.09	
7.901	2009348	1700247	0.846	1.49	8 Aroclor 1260
7.965	142693	115147	0.807	0.10	
8.076	14691	9375	0.638	0.01	
8.277	552327	476813	0.863	0.41	
8.383	35077	22743	0.648	0.02	
8.572	10160226	8786744	0.865	7.55	$ 12 Decachlorobiphenyl
8.945	25819	16899	0.655	0.01	
9.091	15971	14108	0.883	0.01	
9.470	23386	15627	0.668	0.01	
9.964	26995	17470	0.647	0.02	
	134567937	116018447		100.000	

Total unknown % area = 54

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618064.D
Lab Smp Id: 1660-CAL4A Client Smp ID: 1660-CAL4A
Inj Date : 06-MAR-2018 19:23
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1660-cal4a, 185560:1
Misc Info : 17498,AR1660
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:23 Cal File: 030618064.D
Als bottle: 33 Calibration Sample, Level: 4
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1660.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.484	2.484	0.000	12180955	0.10000	0.099			
5 Aroclor 1016					CAS #: 12674-11-2			
3.317	3.318	-0.001	2333576	0.80000	0.76	0.00- 20.00	100.00(M)	BA
3.512	3.512	0.000	2362744	0.80000	0.76	0.00- 20.00	101.25	BA
3.826	3.826	0.000	6746064	0.80000	0.76	0.00- 20.00	289.09	BA
3.929	3.929	0.000	2864977	0.80000	0.76	0.00- 20.00	122.77	BA
3.995	3.994	0.001	1946758	0.80000	0.75	0.00- 20.00	83.42	BA
Average of Peak Amounts =					0.75800			
8 Aroclor 1260					CAS #: 11096-82-5			
6.300	6.300	0.000	6710842	0.80000	0.76	0.00- 20.00	100.00(M)	BA
7.005	7.006	-0.001	7621578	0.80000	0.76	0.00- 20.00	113.57	BA
7.252	7.251	0.001	3661923	0.80000	0.77	0.00- 20.00	54.57	BA
7.443	7.443	0.000	2216952	0.80000	0.76	0.00- 20.00	33.04	BA
7.901	7.901	0.000	2009347	0.80000	0.76	0.00- 20.00	29.94	BA
Average of Peak Amounts =					0.76200			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.571	8.571	0.000	10160225	0.10000	0.093			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618064.D

Date : 06-MAR-2018 19:23

Client ID: 1660-CAL4A

Instrument: 40GCSJ.i

Sample Info: 1660-CAL4A, 185560:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618064.D (ManInt: AR1660.sub Orig: AR1660.sub)
Injection Date: 06-MAR-2018 19:23 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1660-CAL4A





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	2324400	2333576	BA
Aroclor 1016	2324400	2333576	BA
Aroclor 1016	2324400	2333576	BA
Aroclor 1016	2324400	2333576	BA
Aroclor 1016	2324400	2333576	BA
Aroclor 1260	6681873	6710842	BA
Aroclor 1260	6681873	6710842	BA
Aroclor 1260	6681873	6710842	BA
Aroclor 1260	6681873	6710842	BA
Aroclor 1260	6681873	6710842	BA
Tetrachloro-m-xylene	12180955	12180955	
Decachlorobiphenyl	10160225	10160225	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618066.D
Lab Smp Id: 1660-CAL5A Client Smp ID: 1660-CAL5A
Inj Date : 06-MAR-2018 19:39
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1660-cal5a, 185561:1
Misc Info : 17498,AR1660
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 34 Calibration Sample, Level: 5
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1660.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.187	97781	71878	0.735	0.05		
1.429	107893	87519	0.811	0.06		
1.538	44327	37507	0.846	0.02		
1.689	14954	14333	0.958	0.00		
1.805	29740	28315	0.952	0.01		
2.010	60528	44274	0.731	0.03		
2.195	10025	8478	0.846	0.00		
2.405	262681	207562	0.790	0.15		
2.485	18299527	17728359	0.969	11.17	$ 2 Tetrachloro-m-xylene	
2.688	390097	317641	0.814	0.22		
2.789	657209	632240	0.962	0.38		
2.851	2084131	1917446	0.920	1.20		
3.030	26436	29009	1.097	0.01		
3.056	320575	297217	0.927	0.18		
3.278	50946	36697	0.720	0.02		
3.318	2843186	2604947	0.916	1.64	5 Aroclor 1016	BA
3.345	1662911	1538320	0.925	0.96		
3.435	468814	430601	0.918	0.27		
3.513	2890424	2043297	0.707	1.67	5 Aroclor 1016	BA
3.609	62043	51821	0.835	0.03		
3.645	34136	30033	0.880	0.01		
3.697	81037	81549	1.006	0.04		
3.732	758995	759602	1.001	0.43		
3.756	489036	474710	0.971	0.28		
3.827	8333033	5039841	0.605	4.83	5 Aroclor 1016	BA
3.929	3523970	2978332	0.845	2.04	5 Aroclor 1016	BA
3.994	2424782	2289525	0.944	1.40	5 Aroclor 1016	BA
4.059	690218	639344	0.926	0.40		
4.137	268754	245577	0.914	0.15		
4.215	2047181	1927649	0.942	1.18		
4.262	1851518	1768522	0.955	1.07		
4.293	812332	825594	1.016	0.47		
4.310	980066	916238	0.935	0.56		
4.372	57766	44143	0.764	0.03		
4.418	2565892	2366938	0.922	1.48		
4.453	1929565	1448450	0.751	1.11		
4.546	3331677	3063380	0.919	1.93		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.617	763681	725726	0.950	0.44	
4.681	43434	35484	0.817	0.02	
4.728	114565	101125	0.883	0.06	
4.768	108255	99422	0.918	0.06	
4.814	619087	601641	0.972	0.35	
4.849	956824	929295	0.971	0.55	
4.895	3006449	2129145	0.708	1.74	
4.958	194324	157336	0.810	0.11	
5.038	174178	158253	0.909	0.10	
5.101	636910	553004	0.868	0.36	
5.163	2438926	2194509	0.900	1.41	
5.212	111384	92369	0.829	0.06	
5.308	27827	26460	0.951	0.01	
5.364	165309	157324	0.952	0.09	
5.408	704023	646364	0.918	0.40	
5.443	29112	28546	0.981	0.01	
5.479	1000519	928454	0.928	0.57	
5.510	1729074	1534039	0.887	1.00	
5.576	19550	16048	0.821	0.01	
5.604	34848	27975	0.803	0.02	
5.675	2490301	2308914	0.927	1.44	
5.711	1866212	995046	0.533	1.08	
5.787	6185492	5328228	0.861	3.58	
5.883	417579	397243	0.951	0.24	
5.930	202848	167345	0.825	0.11	
5.995	1064293	993125	0.933	0.61	
6.031	2357002	2281362	0.968	1.36	
6.056	6802716	5913555	0.869	3.94	
6.169	4398269	2916854	0.663	2.54	
6.256	1013848	595870	0.588	0.58	
6.299	8372147	6873251	0.821	4.85	8 Aroclor 1260
6.334	1228651	1117268	0.909	0.71	
6.374	310797	237866	0.765	0.18	
6.433	875594	776637	0.887	0.50	
6.486	186696	226162	1.211	0.10	
6.510	4055406	3619288	0.892	2.35	
6.536	804268	870936	1.083	0.46	
6.562	2530676	2228116	0.880	1.46	
6.606	275537	224114	0.813	0.15	
6.660	821602	744134	0.906	0.47	
6.718	4363168	3923038	0.899	2.52	
6.759	2436592	2160749	0.887	1.41	
6.804	2046255	1732909	0.847	1.18	
6.840	379385	321813	0.848	0.21	
6.866	193686	163537	0.844	0.11	
6.913	279672	253923	0.908	0.16	
6.943	893745	829547	0.928	0.51	
7.006	9552528	8232909	0.862	5.53	8 Aroclor 1260
7.036	634159	627991	0.990	0.36	
7.075	290251	237040	0.817	0.16	
7.122	440812	358276	0.813	0.25	
7.222	14569	17676	1.213	0.00	
7.252	4561710	4033993	0.884	2.64	8 Aroclor 1260
7.283	1317583	1086644	0.825	0.76	
7.390	2273253	1907971	0.839	1.31	
7.444	2747730	2350698	0.856	1.59	8 Aroclor 1260
7.554	205956	170355	0.827	0.11	

Data File: ...wintarget\data2\chem\40GCSJ.i\030618.b\030618066.D Page 3
Report Date: 22-Mar-2018 13:35

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
7.692	1143671	1008027	0.881	0.66	
7.728	196307	160265	0.816	0.11	
7.778	26925	26862	0.998	0.01	
7.801	147398	123041	0.835	0.08	
7.901	2485127	2093291	0.842	1.44	8 Aroclor 1260
7.966	182440	138184	0.757	0.10	
8.071	16818	11151	0.663	0.00	
8.276	710759	618559	0.870	0.41	
8.379	47324	33393	0.706	0.02	
8.572	15073641	13074115	0.867	8.73	$ 12 Decachlorobiphenyl
8.944	40002	27870	0.697	0.02	
9.088	20422	15432	0.756	0.01	
9.469	45604	27683	0.607	0.02	
9.964	44216	30829	0.697	0.02	
	172514109	148782492		100.000	

Total unknown % area = 52

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618066.D
Lab Smp Id: 1660-CAL5A Client Smp ID: 1660-CAL5A
Inj Date : 06-MAR-2018 19:39
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1660-cal5a, 185561:1
Misc Info : 17498,AR1660
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 34 Calibration Sample, Level: 5
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1660.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.485	2.484	0.001	18299526	0.15000	0.15			
5 Aroclor 1016					CAS #: 12674-11-2			
3.318	3.318	0.000	2843185	1.00000	0.92	0.00- 20.00	100.00(M)	BA
3.512	3.512	0.000	2890423	1.00000	0.93	0.00- 20.00	101.66	BA
3.826	3.826	0.000	8333033	1.00000	0.94	0.00- 20.00	293.09	BA
3.928	3.929	-0.001	3523970	1.00000	0.93	0.00- 20.00	123.94	BA
3.994	3.994	0.000	2424781	1.00000	0.93	0.00- 20.00	85.28	BA
Average of Peak Amounts =					0.93000			
8 Aroclor 1260					CAS #: 11096-82-5			
6.299	6.300	-0.001	8372147	1.00000	0.94	0.00- 20.00	100.00(M)	BA
7.005	7.006	-0.001	9552528	1.00000	0.95	0.00- 20.00	114.10	BA
7.251	7.251	0.000	4561709	1.00000	0.95	0.00- 20.00	54.49	BA
7.443	7.443	0.000	2747730	1.00000	0.94	0.00- 20.00	32.82	BA
7.900	7.901	-0.001	2485127	1.00000	0.94	0.00- 20.00	29.68	BA
Average of Peak Amounts =					0.94400			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.571	8.571	0.000	15073640	0.15000	0.14			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618066.D

Date : 06-MAR-2018 19:39

Client ID: 1660-CAL5A

Instrument: 40GCSJ.i

Sample Info: 1660-CAL5A, 185561:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618066.D (ManInt: AR1660.sub Orig: AR1660.sub)
Injection Date: 06-MAR-2018 19:39 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1660-CAL5A





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	2831739	2843185	BA
Aroclor 1016	2831739	2843185	BA
Aroclor 1016	2831739	2843185	BA
Aroclor 1016	2831739	2843185	BA
Aroclor 1016	2831739	2843185	BA
Aroclor 1260	8339409	8372147	BA
Aroclor 1260	8339409	8372147	BA
Aroclor 1260	8339409	8372147	BA
Aroclor 1260	8339409	8372147	BA
Aroclor 1260	8339409	8372147	BA
Tetrachloro-m-xylene	18299526	18299526	
Decachlorobiphenyl	15073640	15073640	

PCB Low Level Calibration % Difference Check

Lab Name: Pace Analytical - Green Bay
Instrument ID: 40GCSJ
Calibration Date: 3/6/2018

Aroclor	True Value	Abundance Concentration	% Difference	Acceptance Criteria %	Pass/Fail
AR1221-CAL1	.1	0.109383	9.383	20	Pass
AR1242-CAL1	.1	0.108828	8.828	20	Pass
AR1248-CAL1	.1	0.106694	6.694	20	Pass
AR1254-CAL1	.1	0.10796	7.96	20	Pass
AR1016-CAL1	.1	0.109056	9.056	20	Pass
AR1260-CAL1	.1	0.10731	7.31	20	Pass

SAMPLE NO.

10757904ICV

PCB - FORM VII SVOA-1
PCB CALIBRATION VERIFICATION SUMMARY

Lab Name: Pace Analytical - Green Bay
Calibration Date: 03/06/2018 Time: 20:11

Instrument ID: 40GCSJ GC Column: Col 1
Init. Calib. Date(s): 03/06/2018 03/06/2018

Lab File ID: 030618.B\030618070.D
Init. Calib. Time(s): 11:02 19:39

SDG No.: 40151601

COMPOUND	PEAK	CURVE	RT	RT WINDOW		$\overline{RRF}$ or Amount	RRF or Amount	%D
				FROM	TO			
Aroclor 1221	1	Averaged	1.804	1.77	1.838	1041256.70	1074308.00	3.1742
Aroclor 1221	2	Averaged	2.407	2.373	2.441	547842.900	558458.000	1.9376
Aroclor 1221	3	Averaged	2.688	2.654	2.722	1383336.70	1446642.00	4.5763
Aroclor 1221	4	Averaged	2.789	2.755	2.823	1105769.55	1114194.00	0.7619
Aroclor 1221	5	Averaged	2.851	2.817	2.885	3255397.65	3446964.00	5.8846
Decachlorobiphenyl (S)		Averaged	8.571	8.537	8.605	109559990.	106438520.	-2.8491
Tetrachloro-m-xylene		Averaged	2.484	2.45	2.518	123295106.	119469080.	-3.1031

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

SAMPLE NO.

10757932ICV

PCB - FORM VII SVOA-1
PCB CALIBRATION VERIFICATION SUMMARY

Lab Name: Pace Analytical - Green Bay
Calibration Date: 03/06/2018 Time: 20:27

Instrument ID: 40GCSJ GC Column: Col 1
Init. Calib. Date(s): 03/06/2018 03/06/2018

Lab File ID: 030618.B\030618072.D
Init. Calib. Time(s): 11:02 19:39

SDG No.: 40151601

COMPOUND	PEAK	CURVE	RT	RT WINDOW		$\overline{RRF}$ or Amount	RRF or Amount	%D
				FROM	TO			
Aroclor 1242	1	Averaged	3.317	3.284	3.352	2533198.70	2574948.00	1.6481
Aroclor 1242	2	Averaged	3.513	3.479	3.547	2520033.45	2587724.00	2.6861
Aroclor 1242	3	Averaged	3.825	3.793	3.861	7116457.95	7162442.00	0.6462
Aroclor 1242	4	Averaged	3.928	3.895	3.963	3041534.65	3109342.00	2.2294
Aroclor 1242	5	Averaged	3.993	3.961	4.029	2047746.90	2082504.00	1.6973
Decachlorobiphenyl (S)		Averaged	8.571	8.537	8.605	109559990.	100680260.	-8.1049
Tetrachloro-m-xylene		Averaged	2.484	2.45	2.518	123295106.	117472720.	-4.7223

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

SAMPLE NO.

10757923ICV

PCB - FORM VII SVOA-1
PCB CALIBRATION VERIFICATION SUMMARY

Lab Name: Pace Analytical - Green Bay
Calibration Date: 03/06/2018 Time: 20:44

Instrument ID: 40GCSJ GC Column: Col 1
Init. Calib. Date(s): 03/06/2018 03/06/2018

Lab File ID: 030618.B\030618074.D
Init. Calib. Time(s): 11:02 19:39

SDG No.: 40151601

COMPOUND	PEAK	CURVE	RT	RT WINDOW		$\overline{RRF}$ or Amount	RRF or Amount	%D
				FROM	TO			
Aroclor 1248	1	Averaged	4.261	4.227	4.295	2342973.85	2416006.00	3.1171
Aroclor 1248	2	Averaged	4.545	4.512	4.58	4150305.95	4272006.00	2.9323
Aroclor 1248	3	Averaged	4.813	4.778	4.846	2030080.20	2145142.00	5.6678
Aroclor 1248	4	Averaged	4.879	4.845	4.913	5037445.65	5257174.00	4.3619
Aroclor 1248	5	Averaged	5.037	5.004	5.072	3588546.15	3781714.00	5.3829
Decachlorobiphenyl (S)		Averaged	8.57	8.537	8.605	109559990.	103468660.	-5.5598
Tetrachloro-m-xylene		Averaged	2.485	2.45	2.518	123295106.	117995220.	-4.2985

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

SAMPLE NO.

10757930ICV

PCB - FORM VII SVOA-1
PCB CALIBRATION VERIFICATION SUMMARY

Lab Name: Pace Analytical - Green Bay
Calibration Date: 03/06/2018 Time: 21:00

Instrument ID: 40GCSJ GC Column: Col 1
Init. Calib. Date(s): 03/06/2018 03/06/2018

Lab File ID: 030618.B\030618076.D
Init. Calib. Time(s): 11:02 19:39

SDG No.: 40151601

COMPOUND	PEAK	CURVE	RT	RT WINDOW		$\overline{RRF}$ or Amount	RRF or Amount	%D
				FROM	TO			
Aroclor 1254	1	Averaged	5.162	5.128	5.196	5235247.45	4908596.00	-6.2395
Aroclor 1254	2	Averaged	5.406	5.374	5.442	4301054.45	4023996.00	-6.4416
Aroclor 1254	3	Averaged	5.51	5.476	5.544	7766063.85	7114120.00	-8.3948
Aroclor 1254	4	Averaged	6.026	5.993	6.061	4869683.55	4559312.00	-6.3735
Aroclor 1254	5	Averaged	6.299	6.265	6.333	6507103.75	6351522.00	-2.3910
Decachlorobiphenyl (S)		Averaged	8.571	8.537	8.605	109559990.	103892160.	-5.1733
Tetrachloro-m-xylene		Averaged	2.485	2.45	2.518	123295106.	119128800.	-3.3791

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

SAMPLE NO.

10757911ICV

PCB - FORM VII SVOA-1
PCB CALIBRATION VERIFICATION SUMMARY

Lab Name: Pace Analytical - Green Bay
Calibration Date: 03/06/2018 Time: 21:16

Instrument ID: 40GCSJ GC Column: Col 1
Init. Calib. Date(s): 03/06/2018 03/06/2018

Lab File ID: 030618.B\030618078.D
Init. Calib. Time(s): 11:02 19:39

SDG No.: 40151601

COMPOUND	PEAK	CURVE	RT	RT WINDOW		$\overline{RRF}$ or Amount	RRF or Amount	%D
				FROM	TO			
Aroclor 1016	1	Averaged	3.317	3.284	3.352	3080830.60	3151828.00	2.3045
Aroclor 1016	2	Averaged	3.512	3.478	3.546	3099002.40	3167044.00	2.1956
Aroclor 1016	3	Averaged	3.826	3.792	3.86	8861786.20	9017370.00	1.7557
Aroclor 1016	4	Averaged	3.928	3.895	3.963	3780223.45	3853260.00	1.9321
Aroclor 1016	5	Averaged	3.994	3.96	4.028	2603349.30	2572474.00	-1.1860
Aroclor 1260	1	Averaged	6.3	6.266	6.334	8862352.50	8433522.00	-4.8388
Aroclor 1260	2	Averaged	7.005	6.972	7.04	10014373.3	10338502.0	3.2366
Aroclor 1260	3	Averaged	7.251	7.217	7.285	4779803.15	4863954.00	1.7606
Aroclor 1260	4	Averaged	7.442	7.409	7.477	2915110.60	2985660.00	2.4201
Aroclor 1260	5	Averaged	7.901	7.867	7.935	2638735.95	2816390.00	6.7325
Decachlorobiphenyl (S)		Averaged	8.571	8.537	8.605	109559990.	104964120.	-4.1948
Tetrachloro-m-xylene		Averaged	2.484	2.45	2.518	123295106.	118641800.	-3.7741

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

SAMPLE NO.

10780908CCV

PCB - FORM VII SVOA-1
PCB CALIBRATION VERIFICATION SUMMARY

Lab Name: Pace Analytical - Green Bay
Calibration Date: 03/13/2018 Time: 14:18

Instrument ID: 40GCSJ GC Column: Col 1
Init. Calib. Date(s): 03/06/2018 03/06/2018

Lab File ID: 031318.B\031318004.D
Init. Calib. Time(s): 11:02 19:39

SDG No.: 40151601

COMPOUND	PEAK	CURVE	RT	RT WINDOW		$\overline{RRF}$ or Amount	RRF or Amount	%D
				FROM	TO			
Aroclor 1242	1	Averaged	3.317	3.284	3.352	2533198.70	2333598.00	-7.8794
Aroclor 1242	2	Averaged	3.513	3.479	3.547	2520033.45	2303554.00	-8.5903
Aroclor 1242	3	Averaged	3.826	3.793	3.861	7116457.95	6211628.00	-12.7146
Aroclor 1242	4	Averaged	3.928	3.895	3.963	3041534.65	2669418.00	-12.2345
Aroclor 1242	5	Averaged	3.993	3.961	4.029	2047746.90	1775018.00	-13.3185
Decachlorobiphenyl (S)		Averaged	8.571	8.537	8.605	109559990.	92920240.0	-15.1878
Tetrachloro-m-xylene		Averaged	2.484	2.45	2.518	123295106.	111046180.	-9.9346

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

SAMPLE NO.

10780906CCV

PCB - FORM VII SVOA-1
PCB CALIBRATION VERIFICATION SUMMARY

Lab Name: Pace Analytical - Green Bay
Calibration Date: 03/13/2018 Time: 17:28

Instrument ID: 40GCSJ GC Column: Col 1
Init. Calib. Date(s): 03/06/2018 03/06/2018

Lab File ID: 031318.B\031318016.D
Init. Calib. Time(s): 11:02 19:39

SDG No.: 40151601

COMPOUND	PEAK	CURVE	RT	RT WINDOW		$\overline{RRF}$ or Amount	RRF or Amount	%D
				FROM	TO			
Aroclor 1248	1	Averaged	4.26	4.227	4.295	2342973.85	2229724.00	-4.8336
Aroclor 1248	2	Averaged	4.545	4.512	4.58	4150305.95	3900338.00	-6.0229
Aroclor 1248	3	Averaged	4.812	4.778	4.846	2030080.20	1923484.00	-5.2508
Aroclor 1248	4	Averaged	4.877	4.845	4.913	5037445.65	4751220.00	-5.6820
Aroclor 1248	5	Averaged	5.036	5.004	5.072	3588546.15	3372468.00	-6.0213
Decachlorobiphenyl (S)		Averaged	8.57	8.537	8.605	109559990.	98260660.0	-10.3134
Tetrachloro-m-xylene		Averaged	2.484	2.45	2.518	123295106.	114350280.	-7.2548

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

SAMPLE NO.

10842796CCV

PCB - FORM VII SVOA-1
PCB CALIBRATION VERIFICATION SUMMARY

Lab Name: Pace Analytical - Green Bay
Calibration Date: 03/14/2018 Time: 13:12

Instrument ID: 40GCSJ GC Column: Col 1
Init. Calib. Date(s): 03/06/2018 03/06/2018

Lab File ID: 031418.B\031418004.D
Init. Calib. Time(s): 11:02 19:39

SDG No.: 40151601

COMPOUND	PEAK	CURVE	RT	RT WINDOW		$\overline{RRF}$ or Amount	RRF or Amount	%D
				FROM	TO			
Aroclor 1242	1	Averaged	3.317	3.284	3.352	2533198.70	2510940.00	-0.8787
Aroclor 1242	2	Averaged	3.513	3.479	3.547	2520033.45	2473072.00	-1.8635
Aroclor 1242	3	Averaged	3.826	3.793	3.861	7116457.95	6921176.00	-2.7441
Aroclor 1242	4	Averaged	3.929	3.895	3.963	3041534.65	2975824.00	-2.1604
Aroclor 1242	5	Averaged	3.993	3.961	4.029	2047746.90	1971886.00	-3.7046
Decachlorobiphenyl (S)		Averaged	8.572	8.537	8.605	109559990.	101941740.	-6.9535
Tetrachloro-m-xylene		Averaged	2.484	2.45	2.518	123295106.	117766700.	-4.4839

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

SAMPLE NO.

10842797CCV

PCB - FORM VII SVOA-1
PCB CALIBRATION VERIFICATION SUMMARY

Lab Name: Pace Analytical - Green Bay
Calibration Date: 03/14/2018 Time: 16:22

Instrument ID: 40GCSJ GC Column: Col 1
Init. Calib. Date(s): 03/06/2018 03/06/2018

Lab File ID: 031418.B\031418016.D
Init. Calib. Time(s): 11:02 19:39

SDG No.: 40151601

COMPOUND	PEAK	CURVE	RT	RT WINDOW		$\overline{RRF}$ or Amount	RRF or Amount	%D
				FROM	TO			
Aroclor 1248	1	Averaged	4.261	4.227	4.295	2342973.85	2400320.00	2.4476
Aroclor 1248	2	Averaged	4.545	4.512	4.58	4150305.95	4188910.00	0.9301
Aroclor 1248	3	Averaged	4.813	4.778	4.846	2030080.20	2112654.00	4.0675
Aroclor 1248	4	Averaged	4.878	4.845	4.913	5037445.65	5172652.00	2.6840
Aroclor 1248	5	Averaged	5.037	5.004	5.072	3588546.15	3700542.00	3.1209
Decachlorobiphenyl (S)		Averaged	8.572	8.537	8.605	109559990.	107145860.	-2.2035
Tetrachloro-m-xylene		Averaged	2.484	2.45	2.518	123295106.	120659120.	-2.1379

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

SAMPLE NO.

10842799CCV

PCB - FORM VII SVOA-1
PCB CALIBRATION VERIFICATION SUMMARY

Lab Name: Pace Analytical - Green Bay
Calibration Date: 03/14/2018 Time: 18:45

Instrument ID: 40GCSJ GC Column: Col 1
Init. Calib. Date(s): 03/06/2018 03/06/2018

Lab File ID: 031418.B\031418025.D
Init. Calib. Time(s): 11:02 19:39

SDG No.: 40151601

COMPOUND	PEAK	CURVE	RT	RT WINDOW		$\overline{RRF}$ or Amount	RRF or Amount	%D
				FROM	TO			
Aroclor 1254	1	Averaged	5.162	5.128	5.196	5235247.45	5204698.00	-0.5835
Aroclor 1254	2	Averaged	5.406	5.374	5.442	4301054.45	4348410.00	1.1010
Aroclor 1254	3	Averaged	5.51	5.476	5.544	7766063.85	7768508.00	0.0315
Aroclor 1254	4	Averaged	6.026	5.993	6.061	4869683.55	5096960.00	4.6672
Aroclor 1254	5	Averaged	6.299	6.265	6.333	6507103.75	6645918.00	2.1333
Decachlorobiphenyl (S)		Averaged	8.572	8.537	8.605	109559990.	108046740.	-1.3812
Tetrachloro-m-xylene		Averaged	2.484	2.45	2.518	123295106.	120437160.	-2.3180

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

SAMPLE NO.

10850971CCV

PCB - FORM VII SVOA-1
PCB CALIBRATION VERIFICATION SUMMARY

Lab Name: Pace Analytical - Green Bay
Calibration Date: 03/30/2018 Time: 15:47

Instrument ID: 40GCSJ GC Column: Col 1
Init. Calib. Date(s): 03/06/2018 03/06/2018

Lab File ID: 033018.B\033018004.D
Init. Calib. Time(s): 11:02 19:39

SDG No.: 40151601

COMPOUND	PEAK	CURVE	RT	RT WINDOW		$\overline{RRF}$ or Amount	RRF or Amount	%D
				FROM	TO			
Aroclor 1232	1	Averaged	2.851	2.817	2.885	2500024.00	2538698.00	1.5469
Aroclor 1232	2	Averaged	3.513	3.479	3.547	1339782.00	1338104.00	-0.1252
Aroclor 1232	3	Averaged	3.826	3.793	3.861	3749726.00	3730710.00	-0.5071
Aroclor 1232	4	Averaged	3.929	3.895	3.963	1606492.00	1594924.00	-0.7201
Aroclor 1232	5	Averaged	3.994	3.961	4.029	1091876.00	1054644.00	-3.4099
Decachlorobiphenyl (S)		Averaged	8.571	8.537	8.605	109559990.	102155020.	-6.7588
Tetrachloro-m-xylene		Averaged	2.485	2.45	2.518	123295106.	115728840.	-6.1367

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

SAMPLE NO.

10850970CCV

PCB - FORM VII SVOA-1
PCB CALIBRATION VERIFICATION SUMMARY

Lab Name: Pace Analytical - Green Bay
Calibration Date: 03/30/2018 Time: 18:57

Instrument ID: 40GCSJ GC Column: Col 1
Init. Calib. Date(s): 03/06/2018 03/06/2018

Lab File ID: 033018.B\033018016.D
Init. Calib. Time(s): 11:02 19:39

SDG No.: 40151601

COMPOUND	PEAK	CURVE	RT	RT WINDOW		$\overline{RRF}$ or Amount	RRF or Amount	%D
				FROM	TO			
Aroclor 1242	1	Averaged	3.318	3.284	3.352	2533198.70	2590076.00	2.2453
Aroclor 1242	2	Averaged	3.512	3.479	3.547	2520033.45	2573384.00	2.1171
Aroclor 1242	3	Averaged	3.825	3.793	3.861	7116457.95	7354464.00	3.3444
Aroclor 1242	4	Averaged	3.928	3.895	3.963	3041534.65	3157966.00	3.8280
Aroclor 1242	5	Averaged	3.994	3.961	4.029	2047746.90	2096922.00	2.4014
Decachlorobiphenyl (S)		Averaged	8.57	8.537	8.605	109559990.	108080680.	-1.3502
Tetrachloro-m-xylene		Averaged	2.484	2.45	2.518	123295106.	120300920.	-2.4285

The values for compounds reported as total are based on a summation of the components within the laboratory information management system.

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618070.D
Lab Smp Id: 1221-ICV Client Smp ID: 1221-ICV
Inj Date : 06-MAR-2018 20:11
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1221-icv, 188142:1
Misc Info : 17498,AR1221
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 36 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1221.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.430	40840	33323	0.816	0.23		
1.537	17254	12915	0.749	0.09		
1.688	12270	8463	0.690	0.07		
1.804	537154	469858	0.875	3.07	1 Aroclor 1221	BA
2.158	11282	10928	0.969	0.06		
2.198	158378	141258	0.892	0.90		
2.407	279229	204848	0.734	1.60	1 Aroclor 1221	BA
2.485	5973454	5698500	0.954	34.47	$ 2 Tetrachloro-m-xylene	BA
2.688	723322	608616	0.841	4.14	1 Aroclor 1221	BA
2.789	557097	532699	0.956	3.19	1 Aroclor 1221	BA
2.851	1723482	1520837	0.882	9.87	1 Aroclor 1221	BA
3.053	42842	31274	0.730	0.24		
3.281	125722	92702	0.737	0.72		
3.324	409120	233739	0.571	2.34		
3.433	41957	33921	0.808	0.24		
3.514	152248	110173	0.724	0.87		
3.609	20491	17352	0.847	0.11		
3.693	13172	11957	0.908	0.07		
3.731	39776	40682	1.023	0.22		
3.755	38316	38664	1.009	0.21		
3.827	333479	205587	0.616	1.91		
3.930	166253	139975	0.842	0.95		
3.994	101078	94563	0.936	0.57		
4.058	16842	14008	0.832	0.09		
4.214	43249	38654	0.894	0.24		
4.260	30501	31686	1.039	0.17		
4.291	19839	13076	0.659	0.11		
4.418	34788	38094	1.095	0.19		
4.447	27970	19186	0.686	0.16		
4.544	57594	51561	0.895	0.33		
4.616	16740	13977	0.835	0.09		
4.815	23044	20925	0.908	0.13		
4.847	36853	34664	0.941	0.21		
4.879	53269	36786	0.691	0.30		
5.037	34294	32438	0.946	0.19		
5.103	10234	8648	0.845	0.05		
5.408	11500	8513	0.740	0.06		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.512	18851	12253	0.650	0.10	
5.783	9303	7027	0.755	0.05	
6.347	10003	6690	0.669	0.05	
6.867	9564	7223	0.755	0.05	
7.729	30210	25344	0.839	0.17	
7.802	20371	17570	0.862	0.11	
8.278	45102	38714	0.858	0.25	
8.572	5321926	4509273	0.847	30.49	$ 12 Decachlorobiphenyl
8.942	13753	10341	0.752	0.07	
9.470	11610	10452	0.900	0.06	
9.960	25620	14968	0.584	0.14	
	17451246	15314905		100.000	

Total unknown % area = 13

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618070.D
Lab Smp Id: 1221-ICV Client Smp ID: 1221-ICV
Inj Date : 06-MAR-2018 20:11
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1221-icv, 188142:1
Misc Info : 17498,AR1221
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 36 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1221.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO		REVIEW CODE
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8				
2.484	2.484	0.000	5973454	0.05000	0.048			(M)	BA
1 Aroclor 1221					CAS #: 11104-28-2				
1.804	1.804	0.000	537154	0.50000	0.52	0.00- 20.00	100.00	(M)	BA
2.407	2.407	0.000	279229	0.50000	0.51	0.00- 20.00	51.98		BA
2.688	2.688	0.000	723321	0.50000	0.52	0.00- 20.00	134.66		BA
2.789	2.789	0.000	557097	0.50000	0.50	0.00- 20.00	103.71		BA
2.851	2.851	0.000	1723482	0.50000	0.53	0.00- 20.00	320.85		BA
Average of Peak Amounts =					0.51600				
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3				
8.571	8.571	0.000	5321926	0.05000	0.048				

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618070.D
Date : 06-MAR-2018 20:11
Client ID: 1221-ICV
Sample Info: 1221-ICV, 188142:1

Instrument: 40GCSJ.i
Operator: BLM
Column phase: ZB-1ms
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618070.D (ManInt: AR1221.sub Orig: AR1221.sub)
Injection Date: 06-MAR-2018 20:11 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1221-ICV





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1221	531813	537154	BA
Aroclor 1221	531813	537154	BA
Aroclor 1221	531813	537154	BA
Aroclor 1221	531813	537154	BA
Aroclor 1221	531813	537154	BA
Tetrachloro-m-xylene	5952668	5973454	BA
Decachlorobiphenyl	5321926	5321926	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618072.D
Lab Smp Id: 1242-ICV Client Smp ID: 1242-ICV
Inj Date : 06-MAR-2018 20:27
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-icv, 185562:1
Misc Info : 17498,AR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 37 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1242.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.182	14288	12968	0.908	0.03		
1.428	37134	32784	0.883	0.09		
1.539	20685	13744	0.664	0.05		
1.806	19059	17804	0.934	0.05		
2.404	134346	103552	0.771	0.35		
2.485	5873636	5676099	0.966	16.03	$ 2 Tetrachloro-m-xylene	
2.686	193190	158466	0.820	0.51		
2.788	283870	272236	0.959	0.75		
2.850	999209	952371	0.953	2.66		
3.054	145787	130280	0.894	0.38		
3.279	24447	19522	0.799	0.06		
3.317	1287474	1200238	0.932	3.43	3 Aroclor 1242	BA
3.344	730459	693075	0.949	1.94		
3.435	207883	189325	0.911	0.55		
3.513	1293863	944289	0.730	3.44	3 Aroclor 1242	BA
3.612	22860	18377	0.804	0.06		
3.647	23439	16349	0.698	0.06		
3.696	36476	35506	0.973	0.09		
3.731	322838	319621	0.990	0.86		
3.756	213073	205263	0.963	0.56		
3.826	3581222	2097110	0.586	9.54	3 Aroclor 1242	BA
3.928	1554671	1293077	0.832	4.14	3 Aroclor 1242	BA
3.994	1041253	986439	0.947	2.77	3 Aroclor 1242	BA
4.058	279247	251769	0.902	0.74		
4.136	113510	101678	0.896	0.30		
4.215	824770	788034	0.955	2.19		
4.261	781405	765245	0.979	2.08		
4.292	312524	328346	1.051	0.83		
4.311	404517	367665	0.909	1.07		
4.372	27372	22538	0.823	0.07		
4.417	1118646	1040424	0.930	2.98		
4.452	854457	630460	0.738	2.27		
4.545	1421692	1283823	0.903	3.78		
4.618	334338	302865	0.906	0.89		
4.679	28007	18667	0.667	0.07		
4.728	68438	58015	0.848	0.18		
4.768	73081	69415	0.950	0.19		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.812	640662	618887	0.966	1.70	
4.848	1131783	1062264	0.939	3.01	
4.878	1453285	1083159	0.745	3.87	
4.957	156083	126825	0.813	0.41	
5.037	1224571	1114748	0.910	3.26	
5.099	245239	185939	0.758	0.65	
5.162	276625	246249	0.890	0.73	
5.212	201147	175293	0.871	0.53	
5.285	18730	16061	0.858	0.04	
5.312	50903	39235	0.771	0.13	
5.362	178111	159411	0.895	0.47	
5.407	296369	257265	0.868	0.78	
5.446	185486	163132	0.879	0.49	
5.509	494170	376849	0.763	1.31	
5.604	157705	129475	0.821	0.42	
5.670	15998	12262	0.766	0.04	
5.716	27499	21293	0.774	0.07	
5.741	38080	32982	0.866	0.10	
5.792	324066	249256	0.769	0.86	
5.891	46598	28590	0.614	0.12	
6.025	300272	225833	0.752	0.80	
6.173	22376	16996	0.760	0.05	
6.299	61219	54211	0.886	0.16	
6.335	9981	10468	1.049	0.02	
6.367	9764	8232	0.843	0.02	
6.801	13109	12142	0.926	0.03	
7.008	20895	16561	0.793	0.05	
7.124	16181	11660	0.721	0.04	
7.253	9479	8109	0.855	0.02	
7.392	10765	9021	0.838	0.02	
7.727	30633	26862	0.877	0.08	
7.802	22807	20254	0.888	0.06	
8.275	44979	37205	0.827	0.11	
8.381	12768	10411	0.815	0.03	
8.572	5034014	4309550	0.856	13.41	$ 12 Decachlorobiphenyl
9.466	9953	6796	0.683	0.02	
9.962	21579	14184	0.657	0.05	
	37517051	32315109		100.000	

Total unknown % area = 47

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618072.D
Lab Smp Id: 1242-ICV Client Smp ID: 1242-ICV
Inj Date : 06-MAR-2018 20:27
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-icv, 185562:1
Misc Info : 17498,AR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 37 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1242.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene						CAS #: 877-09-8		
2.484	2.484	0.000	5873636	0.05000	0.048			
3 Aroclor 1242						CAS #: 53469-21-9		
3.317	3.318	-0.001	1287474	0.50000	0.51	0.00- 20.00	100.00(M)	BA
3.513	3.513	0.000	1293862	0.50000	0.51	0.00- 20.00	100.50	BA
3.825	3.827	-0.002	3581221	0.50000	0.50	0.00- 20.00	278.16	BA
3.928	3.929	-0.001	1554671	0.50000	0.51	0.00- 20.00	120.75	BA
3.993	3.995	-0.002	1041252	0.50000	0.51	0.00- 20.00	80.88	BA
Average of Peak Amounts =					0.50800			
$ 12 Decachlorobiphenyl						CAS #: 2051-24-3		
8.571	8.571	0.000	5034013	0.05000	0.046			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618072.D

Date : 06-MAR-2018 20:27

Client ID: 1242-ICV

Instrument: 40GCSJ.i

Sample Info: 1242-ICV, 185562:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618072.D (ManInt: AR1242.sub Orig: AR1242.sub)
Injection Date: 06-MAR-2018 20:27 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1242-ICV





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1242	1280662	1287474	BA
Aroclor 1242	1280662	1287474	BA
Aroclor 1242	1280662	1287474	BA
Aroclor 1242	1280662	1287474	BA
Aroclor 1242	1280662	1287474	BA
Tetrachloro-m-xylene	5873636	5873636	
Decachlorobiphenyl	5034013	5034013	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618074.D
Lab Smp Id: 1248-ICV Client Smp ID: 1248-ICV
Inj Date : 06-MAR-2018 20:44
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-icv, 185563:1
Misc Info : 17498,AR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 38 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1248.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.181	17909	14420	0.805	0.04		
1.427	41300	30754	0.745	0.09		
1.534	17534	12165	0.694	0.04		
1.683	10383	10019	0.965	0.02		
2.405	14088	10428	0.740	0.03		
2.485	5899761	5544150	0.940	14.05	$ 2 Tetrachloro-m-xylene	
2.689	16668	12362	0.742	0.03		
2.788	95501	90846	0.951	0.22		
2.852	103096	93433	0.906	0.23		
3.055	53842	46986	0.873	0.12		
3.318	652092	633367	0.971	1.50		
3.344	255476	242530	0.949	0.58		
3.433	54381	45369	0.834	0.12		
3.515	528681	378059	0.715	1.22		
3.612	17192	16665	0.969	0.03		
3.733	139942	107922	0.771	0.32		
3.815	998138	1180553	1.183	2.30		
3.825	1248615	1194584	0.957	2.88		
3.931	835250	638277	0.764	1.92		
3.995	550006	533435	0.970	1.26		
4.059	343387	326192	0.950	0.79		
4.137	145276	132173	0.910	0.33		
4.215	1486323	1403883	0.945	3.43		
4.262	1208004	1159133	0.960	2.78	6 Aroclor 1248	BA
4.293	511269	526568	1.030	1.17		
4.311	543428	513963	0.946	1.25		
4.418	1823450	1718104	0.942	4.20		
4.452	908943	795006	0.875	2.09		
4.546	2136003	1914178	0.896	4.92	6 Aroclor 1248	BA
4.588	34235	33628	0.982	0.07		
4.618	485337	462281	0.952	1.12		
4.680	28053	23489	0.837	0.06		
4.728	63611	50641	0.796	0.14		
4.769	114596	110060	0.960	0.26		
4.813	1072572	1047618	0.977	2.47	6 Aroclor 1248	BA
4.848	2019974	1872521	0.927	4.66		
4.880	2628588	1729939	0.658	6.06	6 Aroclor 1248	BA

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.958	300549	254865	0.848	0.69	
5.037	1890857	1718796	0.909	4.36	6 Aroclor 1248
5.099	658439	515140	0.782	1.51	
5.163	778429	704282	0.905	1.79	
5.212	561933	511880	0.911	1.29	
5.284	37417	36848	0.985	0.08	
5.309	124762	104101	0.834	0.28	
5.364	496549	440266	0.887	1.14	
5.408	826216	719936	0.871	1.90	
5.446	528194	462528	0.876	1.21	
5.510	1373479	1108021	0.807	3.16	
5.604	398982	348600	0.874	0.92	
5.675	36019	29954	0.832	0.08	
5.717	78189	59388	0.760	0.18	
5.739	98549	89525	0.908	0.22	
5.793	935796	748637	0.800	2.15	
5.889	98768	81563	0.826	0.22	
5.996	31650	25516	0.806	0.07	
6.024	912880	698617	0.765	2.10	
6.169	70680	48903	0.692	0.16	
6.228	45926	27843	0.606	0.10	
6.301	222697	182730	0.821	0.51	
6.335	46487	36946	0.795	0.10	
6.372	40036	31572	0.789	0.09	
6.512	15047	14089	0.936	0.03	
6.536	93754	63840	0.681	0.21	
6.607	16189	11424	0.706	0.03	
6.719	21159	16941	0.801	0.04	
6.759	10928	9914	0.907	0.02	
6.804	35912	32894	0.916	0.08	
7.005	37431	32956	0.880	0.08	
7.126	11973	10422	0.870	0.02	
7.253	25115	21545	0.858	0.05	
7.683	12978	7477	0.576	0.02	
7.730	34860	28466	0.817	0.08	
7.799	36493	21426	0.587	0.08	
7.900	14089	11492	0.816	0.03	
8.274	45874	41456	0.904	0.10	
8.380	17315	11279	0.651	0.03	
8.571	5173433	4424436	0.855	11.93	$ 12 Decachlorobiphenyl
8.940	12157	9669	0.795	0.02	
9.960	18949	14032	0.741	0.04	
	43330043	38425916		100.000	

Total unknown % area = 53

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618074.D
Lab Smp Id: 1248-ICV Client Smp ID: 1248-ICV
Inj Date : 06-MAR-2018 20:44
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-icv, 185563:1
Misc Info : 17498,AR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 38 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1248.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.485	2.484	0.001	5899761	0.05000	0.048			
6 Aroclor 1248					CAS #: 12672-29-6			
4.261	4.261	0.000	1208003	0.50000	0.52	0.00- 20.00	100.00(M)	BA
4.545	4.546	-0.001	2136003	0.50000	0.51	0.00- 20.00	176.82	BA
4.813	4.812	0.001	1072571	0.50000	0.53	0.00- 20.00	88.79	BA
4.879	4.879	0.000	2628587	0.50000	0.52	0.00- 20.00	217.60	BA
5.037	5.038	-0.001	1890857	0.50000	0.53	0.00- 20.00	156.53	BA
Average of Peak Amounts =					0.52200			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.570	8.571	-0.001	5173433	0.05000	0.047			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618074.D
Date : 06-MAR-2018 20:44
Client ID: 1248-ICV
Sample Info: 1248-ICV, 185563:1

Instrument: 40GCSJ.i
Operator: BLM
Column phase: ZB-1ms
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618074.D (ManInt: AR1248.sub Orig: AR1248.sub)
Injection Date: 06-MAR-2018 20:44 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1248-ICV





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1248	1202018	1208003	BA
Aroclor 1248	1202018	1208003	BA
Aroclor 1248	1202018	1208003	BA
Aroclor 1248	1202018	1208003	BA
Aroclor 1248	1202018	1208003	BA
Tetrachloro-m-xylene	5899761	5899761	
Decachlorobiphenyl	5173433	5173433	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618076.D
Lab Smp Id: 1254-ICV Client Smp ID: 1254-ICV
Inj Date : 06-MAR-2018 21:00
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-icv, 185564:1
Misc Info : 17498,AR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 39 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1254.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.181	11241	11145	0.991	0.02		
1.429	37023	29359	0.793	0.07		
1.536	18870	14547	0.771	0.03		
2.485	5956441	5624399	0.944	12.17	$ 2 Tetrachloro-m-xylene	
2.787	83932	83545	0.995	0.16		
2.847	23487	12751	0.543	0.04		
3.053	22053	16758	0.760	0.04		
3.320	24424	22137	0.906	0.04		
3.343	16533	14772	0.894	0.03		
3.511	21956	15136	0.689	0.04		
3.610	23047	18723	0.812	0.04		
3.813	73837	44648	0.605	0.14		
3.936	54806	35146	0.641	0.10		
3.991	21322	19067	0.894	0.04		
4.057	16084	14470	0.900	0.03		
4.216	1104737	1042404	0.944	2.18		
4.261	294439	287244	0.976	0.58		
4.293	55355	38898	0.703	0.10		
4.417	645490	619185	0.959	1.27		
4.453	65034	55154	0.848	0.12		
4.546	297966	280995	0.943	0.58		
4.590	13297	14947	1.124	0.02		
4.615	52626	43507	0.827	0.10		
4.682	14760	14823	1.004	0.02		
4.769	19864	18032	0.908	0.03		
4.813	255256	243907	0.956	0.50		
4.848	954635	917189	0.961	1.88		
4.895	2240553	1806193	0.806	4.42		
4.957	378746	331044	0.874	0.74		
5.039	197483	178932	0.906	0.39		
5.100	1331660	1130632	0.849	2.63		
5.163	2454299	2245562	0.915	4.84	7 Aroclor 1254	BA
5.212	1112261	987906	0.888	2.19		
5.285	62054	59649	0.961	0.12		
5.310	206656	186551	0.903	0.40		
5.363	957753	888665	0.928	1.89		
5.407	2011999	1826444	0.908	3.97	7 Aroclor 1254	BA

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.445	683522	603976	0.884	1.35	
5.478	263671	258349	0.980	0.52	
5.510	3557061	3172502	0.892	7.02	7 Aroclor 1254
5.577	41056	36384	0.886	0.08	
5.605	496084	429134	0.865	0.98	
5.676	353874	332146	0.939	0.69	
5.711	557893	345779	0.620	1.10	
5.741	316074	271838	0.860	0.62	
5.790	3702485	2959912	0.799	7.31	
5.885	349557	271841	0.778	0.69	
5.928	191368	119957	0.627	0.37	
5.994	332278	306209	0.922	0.65	
6.027	2279656	1881319	0.825	4.50	7 Aroclor 1254
6.055	1742805	1415445	0.812	3.44	
6.169	710083	597581	0.842	1.40	
6.228	301903	298212	0.988	0.59	
6.242	317472	282700	0.890	0.62	
6.300	3175762	2557382	0.805	6.27	7 Aroclor 1254
6.335	613369	538823	0.878	1.21	
6.374	338945	280771	0.828	0.66	
6.441	98349	54990	0.559	0.19	
6.487	20933	19173	0.916	0.04	
6.510	122159	127890	1.047	0.24	
6.537	1093807	836036	0.764	2.16	
6.606	171691	137034	0.798	0.33	
6.662	39774	29158	0.733	0.07	
6.720	225492	189217	0.839	0.44	
6.760	134652	118490	0.880	0.26	
6.802	627925	538577	0.858	1.24	
6.843	186255	113960	0.612	0.36	
6.943	65536	50712	0.774	0.12	
7.006	455898	370505	0.813	0.90	
7.074	28987	20562	0.709	0.05	
7.121	17906	16606	0.927	0.03	
7.253	362457	317889	0.877	0.71	
7.281	86561	71028	0.821	0.17	
7.390	15134	15178	1.003	0.02	
7.445	25464	19955	0.784	0.05	
7.555	25575	19764	0.773	0.05	
7.692	17214	14547	0.845	0.03	
7.731	30335	26152	0.862	0.05	
7.802	25782	20334	0.789	0.05	
7.898	15405	15400	1.000	0.03	
8.277	51144	41782	0.817	0.10	
8.571	5194608	4290400	0.826	10.26	$ 12 Decachlorobiphenyl
9.963	10137	8532	0.842	0.02	
	50606079	43638597		100.000	

Total unknown % area = 51

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618076.D
Lab Smp Id: 1254-ICV Client Smp ID: 1254-ICV
Inj Date : 06-MAR-2018 21:00
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-icv, 185564:1
Misc Info : 17498,AR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 39 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1254.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.485	2.484	0.001	5956440	0.05000	0.048			
7 Aroclor 1254					CAS #: 11097-69-1			
5.162	5.162	0.000	2454298	0.50000	0.47	0.00- 20.00	100.00(M)	BA
5.406	5.408	-0.002	2011998	0.50000	0.47	0.00- 20.00	81.98	BA
5.510	5.510	0.000	3557060	0.50000	0.46	0.00- 20.00	144.93	BA
6.026	6.027	-0.001	2279656	0.50000	0.47	0.00- 20.00	92.88	BA
6.299	6.299	0.000	3175761	0.50000	0.49	0.00- 20.00	129.40	BA
Average of Peak Amounts =					0.47200			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.571	8.571	0.000	5194608	0.05000	0.047			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618076.D

Date : 06-MAR-2018 21:00

Client ID: 1254-ICV

Instrument: 40GCSJ.i

Sample Info: 1254-ICV, 185564:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618076.D (ManInt: AR1254.sub Orig: AR1254.sub)
Injection Date: 06-MAR-2018 21:00 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1254-ICV





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1254	2444116	2454298	BA
Aroclor 1254	2444116	2454298	BA
Aroclor 1254	2444116	2454298	BA
Aroclor 1254	2444116	2454298	BA
Aroclor 1254	2444116	2454298	BA
Tetrachloro-m-xylene	5956440	5956440	
Decachlorobiphenyl	5194608	5194608	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618078.D
Lab Smp Id: 1660-ICV Client Smp ID: 1660-ICV
Inj Date : 06-MAR-2018 21:16
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1660-icv, 185565:1
Misc Info : 17498,AR1660
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 40 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1660.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.182	21221	17184	0.810	0.02		
1.431	40838	31713	0.777	0.04		
1.536	17776	15618	0.879	0.02		
1.804	13311	13960	1.049	0.01		
2.011	10268	7617	0.742	0.01		
2.405	140776	109473	0.778	0.16		
2.484	5932090	5726918	0.965	7.56	$ 2 Tetrachloro-m-xylene	
2.687	201068	164753	0.819	0.24		
2.788	295756	279701	0.946	0.35		
2.851	1113041	1029534	0.925	1.33		
3.055	169967	158668	0.934	0.20		
3.277	24278	19456	0.801	0.02		
3.317	1575915	1463198	0.928	1.88	5 Aroclor 1016	BA
3.344	915512	854579	0.933	1.09		
3.435	250094	229311	0.917	0.29		
3.513	1583523	1159648	0.732	1.89	5 Aroclor 1016	BA
3.610	21286	19905	0.935	0.02		
3.645	25464	17411	0.684	0.03		
3.696	44251	42008	0.949	0.05		
3.732	398780	398305	0.999	0.47		
3.756	256467	255771	0.997	0.30		
3.827	4508685	2620847	0.581	5.39	5 Aroclor 1016	BA
3.929	1926630	1599827	0.830	2.30	5 Aroclor 1016	BA
3.994	1286237	1237455	0.962	1.54	5 Aroclor 1016	BA
4.058	377180	342508	0.908	0.45		
4.136	150296	140468	0.935	0.17		
4.214	1102455	1039229	0.943	1.32		
4.261	1029907	980757	0.952	1.23		
4.293	425562	439845	1.034	0.50		
4.310	540061	519083	0.961	0.64		
4.370	26189	19343	0.739	0.03		
4.417	1387264	1301927	0.938	1.66		
4.453	895338	753232	0.841	1.07		
4.545	1576424	1430306	0.907	1.88		
4.591	17433	23996	1.376	0.02		
4.618	301465	279443	0.927	0.36		
4.680	19837	14137	0.713	0.02		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.728	37013	29354	0.793	0.04	
4.768	30417	27693	0.910	0.03	
4.814	156037	149062	0.955	0.18	
4.848	239300	233617	0.976	0.28	
4.895	1221737	985481	0.807	1.46	
4.958	58840	41257	0.701	0.07	
5.037	85057	75952	0.893	0.10	
5.100	246503	215312	0.873	0.29	
5.163	1183240	1075758	0.909	1.41	
5.212	41947	29637	0.707	0.05	
5.310	13373	10637	0.795	0.01	
5.363	57749	52832	0.915	0.06	
5.407	268815	254856	0.948	0.32	
5.446	13974	15175	1.086	0.01	
5.478	527498	492186	0.933	0.63	
5.510	745390	669137	0.898	0.89	
5.576	10212	10260	1.005	0.01	
5.603	17026	12475	0.733	0.02	
5.676	1384577	1280481	0.925	1.65	
5.710	943302	526346	0.558	1.12	
5.786	3233298	2804467	0.867	3.87	
5.882	208249	189092	0.908	0.24	
5.930	96479	76874	0.797	0.11	
5.993	545401	501421	0.919	0.65	
6.029	1181491	1139915	0.965	1.41	
6.055	3616031	3141367	0.869	4.33	
6.169	2379439	1526966	0.642	2.84	
6.256	501646	311664	0.621	0.60	
6.301	4216761	3413991	0.810	5.04	8 Aroclor 1260
6.334	589361	516529	0.876	0.70	
6.373	133856	105313	0.787	0.16	
6.433	474923	416467	0.877	0.56	
6.486	96367	124001	1.287	0.11	
6.510	2278472	2020249	0.887	2.72	
6.535	319186	362341	1.135	0.38	
6.561	1371229	1239953	0.904	1.64	
6.606	126514	100966	0.798	0.15	
6.661	435030	401966	0.924	0.52	
6.718	2436948	2139504	0.878	2.91	
6.760	1379678	1221074	0.885	1.65	
6.803	1012423	886702	0.876	1.21	
6.839	204549	179535	0.878	0.24	
6.867	101951	86190	0.845	0.12	
6.915	150840	138650	0.919	0.18	
6.944	471585	431542	0.915	0.56	
7.006	5169251	4474270	0.866	6.18	8 Aroclor 1260
7.035	349824	352477	1.008	0.41	
7.076	161438	129444	0.802	0.19	
7.122	226639	191125	0.843	0.27	
7.252	2431978	2174432	0.894	2.91	8 Aroclor 1260
7.282	689708	552481	0.801	0.82	
7.391	1295375	1038752	0.802	1.55	
7.443	1492830	1281759	0.859	1.78	8 Aroclor 1260
7.554	105447	79274	0.752	0.12	
7.692	611618	562958	0.920	0.73	
7.728	153573	121839	0.793	0.18	
7.801	84844	63857	0.753	0.10	

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
7.902	1408195	1200861	0.853	1.68	8 Aroclor 1260
7.965	99066	72839	0.735	0.11	
8.277	469900	412226	0.877	0.56	
8.379	12566	10774	0.857	0.01	
8.571	5248206	4347540	0.828	6.28	$ 12 Decachlorobiphenyl
8.945	16720	11361	0.679	0.02	
9.471	16512	10298	0.624	0.01	
	83510046	71509948		100.000	

Total unknown % area = 56

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618078.D
Lab Smp Id: 1660-ICV Client Smp ID: 1660-ICV
Inj Date : 06-MAR-2018 21:16
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1660-icv, 185565:1
Misc Info : 17498,AR1660
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\030618.b\pcba030618.m
Meth Date : 22-Mar-2018 13:34 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 40 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1660.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.484	2.484	0.000	5932090	0.05000	0.048			
5 Aroclor 1016					CAS #: 12674-11-2			
3.317	3.318	-0.001	1575914	0.50000	0.51	0.00- 20.00	100.00(M)	BA
3.512	3.512	0.000	1583522	0.50000	0.51	0.00- 20.00	100.48	BA
3.826	3.826	0.000	4508685	0.50000	0.51	0.00- 20.00	286.10	BA
3.928	3.929	-0.001	1926630	0.50000	0.51	0.00- 20.00	122.25	BA
3.994	3.994	0.000	1286237	0.50000	0.49	0.00- 20.00	81.62	BA
Average of Peak Amounts =					0.50600			
8 Aroclor 1260					CAS #: 11096-82-5			
6.300	6.300	0.000	4216761	0.50000	0.48	0.00- 20.00	100.00(M)	CO
7.005	7.006	-0.001	5169251	0.50000	0.52	0.00- 20.00	122.59	CO
7.251	7.251	0.000	2431977	0.50000	0.51	0.00- 20.00	57.67	CO
7.442	7.443	-0.001	1492830	0.50000	0.51	0.00- 20.00	35.40	CO
7.901	7.901	0.000	1408195	0.50000	0.53	0.00- 20.00	33.40	CO
Average of Peak Amounts =					0.51000			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.571	8.571	0.000	5248206	0.05000	0.048			

QC Flag Legend

M - Compound response manually integrated.

QC Flag Legend

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.
CO: Indicates that the analyst had to split two co-eluting peaks apart that were not (or could not be) separated by the computer system.

Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b\030618078.D

Date : 06-MAR-2018 21:16

Client ID: 1660-ICV

Instrument: 40GCSJ.i

Sample Info: 1660-ICV, 185565:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\030618.b/030618078.D (ManInt: AR1660.sub Orig: AR1660.sub)
Injection Date: 06-MAR-2018 21:16 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1660-ICV





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	1566175	1575914	BA
Aroclor 1016	1566175	1575914	BA
Aroclor 1016	1566175	1575914	BA
Aroclor 1016	1566175	1575914	BA
Aroclor 1016	1566175	1575914	BA
Aroclor 1260	4198281	4216761	CO
Aroclor 1260	4198281	4216761	CO
Aroclor 1260	4198281	4216761	CO
Aroclor 1260	4198281	4216761	CO
Aroclor 1260	4198281	4216761	CO
Tetrachloro-m-xylene	5932090	5932090	
Decachlorobiphenyl	5248206	5248206	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318004.D
Lab Smp Id: 1242-CCV Client Smp ID: 1242-CCV
Inj Date : 13-MAR-2018 14:18
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-ccv, 185544:1
Misc Info : 17498,AR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031318.b\pcba030618.m
Meth Date : 27-Mar-2018 10:38 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 4 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1242.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.182	27729	26319	0.949	0.08		
1.428	33453	29727	0.889	0.09		
1.538	20488	14619	0.714	0.05		
1.801	11778	11760	0.998	0.03		
2.008	19921	15710	0.789	0.05		
2.405	108223	87198	0.806	0.31		
2.484	5552310	5342617	0.962	16.54	$ 2 Tetrachloro-m-xylene	
2.686	159923	131103	0.820	0.46		
2.788	237799	230665	0.970	0.69		
2.851	817185	785446	0.961	2.37		
3.025	11065	12008	1.085	0.03		
3.054	132892	121653	0.915	0.38		
3.276	18728	15746	0.841	0.05		
3.318	1166799	1106855	0.949	3.39	3 Aroclor 1242	BA
3.344	654278	620865	0.949	1.90		
3.434	184801	173408	0.938	0.53		
3.514	1151778	833455	0.724	3.35	3 Aroclor 1242	BA
3.610	15692	16614	1.059	0.04		
3.649	20511	14959	0.729	0.05		
3.698	31944	32138	1.006	0.09		
3.731	284983	280826	0.985	0.82		
3.755	180299	179544	0.996	0.52		
3.826	3105814	1818355	0.585	9.03	3 Aroclor 1242	BA
3.929	1334709	1124161	0.842	3.88	3 Aroclor 1242	BA
3.994	887509	843140	0.950	2.58	3 Aroclor 1242	BA
4.058	278578	253577	0.910	0.81		
4.137	109261	101496	0.929	0.31		
4.215	774002	740714	0.957	2.25		
4.261	744466	719464	0.966	2.16		
4.292	303869	318652	1.049	0.88		
4.309	381691	354979	0.930	1.11		
4.373	20006	17736	0.887	0.05		
4.418	1024314	997672	0.974	2.97		
4.452	739738	581880	0.787	2.15		
4.546	1301660	1210386	0.930	3.78		
4.589	16687	17909	1.073	0.04		
4.617	302738	279821	0.924	0.88		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.680	21300	17459	0.820	0.06	
4.727	63095	52304	0.829	0.18	
4.767	71898	67959	0.945	0.20	
4.812	572644	562711	0.983	1.66	
4.848	997508	968292	0.971	2.90	
4.878	1270628	964793	0.759	3.69	
4.958	151973	119495	0.786	0.44	
5.037	1026402	978327	0.953	2.98	
5.098	245668	194112	0.790	0.71	
5.163	278059	248516	0.894	0.80	
5.212	210820	189885	0.901	0.61	
5.284	19653	18711	0.952	0.05	
5.309	49055	39540	0.806	0.14	
5.363	170284	157223	0.923	0.49	
5.407	306212	258824	0.845	0.89	
5.445	187785	168850	0.899	0.54	
5.510	472598	368310	0.779	1.37	
5.604	140444	124847	0.889	0.40	
5.718	35346	26125	0.739	0.10	
5.741	38304	33772	0.882	0.11	
5.792	329358	269604	0.819	0.95	
5.889	51373	30362	0.591	0.14	
6.024	261152	224985	0.862	0.75	
6.056	52368	45070	0.861	0.15	
6.170	25509	17193	0.674	0.07	
6.298	90655	75572	0.834	0.26	
6.335	21148	16461	0.778	0.06	
6.375	12751	11969	0.939	0.03	
6.450	10705	7776	0.726	0.03	
6.537	35492	26941	0.759	0.10	
6.605	11787	7404	0.628	0.03	
6.801	20073	18126	0.903	0.05	
7.126	19729	13103	0.664	0.05	
7.687	9475	7483	0.790	0.02	
7.727	27675	22834	0.825	0.08	
7.802	30004	17582	0.586	0.08	
8.274	45852	37803	0.824	0.13	
8.380	23787	14608	0.614	0.06	
8.571	4646012	3951323	0.850	13.51	$ 12 Decachlorobiphenyl
8.943	19091	10692	0.560	0.05	
9.833	31547	19171	0.608	0.09	
9.851	23300	25620	1.100	0.06	
9.870	30135	28982	0.962	0.08	
9.884	53193	29344	0.552	0.15	
	34379470	29953240		100.000	

Total unknown % area = 48

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318004.D
Lab Smp Id: 1242-CCV Client Smp ID: 1242-CCV
Inj Date : 13-MAR-2018 14:18
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-ccv, 185544:1
Misc Info : 17498,AR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031318.b\pcba030618.m
Meth Date : 27-Mar-2018 10:38 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 4 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1242.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.484	2.484	0.000	5552309	0.05000	0.045			
3 Aroclor 1242					CAS #: 53469-21-9			
3.317	3.318	-0.001	1166799	0.50000	0.46	0.00- 20.00	100.00(M)	BA
3.513	3.513	0.000	1151777	0.50000	0.46	0.00- 20.00	98.71	BA
3.826	3.827	-0.001	3105814	0.50000	0.44	0.00- 20.00	266.18	BA
3.928	3.929	-0.001	1334709	0.50000	0.44	0.00- 20.00	114.39	BA
3.993	3.995	-0.002	887509	0.50000	0.43	0.00- 20.00	76.06	BA
Average of Peak Amounts =					0.44600			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.571	8.571	0.000	4646012	0.05000	0.042			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318004.D
Date : 13-MAR-2018 14:18
Client ID: 1242-CCV
Sample Info: 1242-CCV, 185544:1

Instrument: 40GCSJ.i

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031318.b/031318004.D (ManInt: AR1242.sub Orig: AR1242.sub)
Injection Date: 13-MAR-2018 14:18 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1242-CCV





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1242	1161730	1166799	BA
Aroclor 1242	1161730	1166799	BA
Aroclor 1242	1161730	1166799	BA
Aroclor 1242	1161730	1166799	BA
Aroclor 1242	1161730	1166799	BA
Tetrachloro-m-xylene	5552309	5552309	
Decachlorobiphenyl	4646012	4646012	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318016.D
Lab Smp Id: 1248-CCV Client Smp ID: 1248-CCV
Inj Date : 13-MAR-2018 17:28
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-ccv, 185549:1
Misc Info : 17498,AR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031318.b\pcba030618.m
Meth Date : 27-Mar-2018 10:38 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 16 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1248.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.183	34547	26137	0.757	0.08		
1.430	38532	28426	0.738	0.09		
1.538	17792	14244	0.801	0.04		
1.687	26790	16710	0.624	0.06		
2.012	16736	14881	0.889	0.04		
2.402	25266	16834	0.666	0.06		
2.484	5717515	5404411	0.945	14.26	$ 2 Tetrachloro-m-xylene	
2.688	31911	24623	0.772	0.07		
2.787	108403	101621	0.937	0.26		
2.850	181579	166250	0.916	0.43		
3.029	11468	12149	1.059	0.02		
3.053	56715	51363	0.906	0.13		
3.317	666400	648700	0.973	1.61		
3.343	281689	262357	0.931	0.68		
3.434	63551	57232	0.901	0.15		
3.515	565654	405242	0.716	1.36		
3.610	19845	19312	0.973	0.04		
3.646	11099	6999	0.631	0.02		
3.731	124142	124676	1.004	0.30		
3.754	61967	53224	0.859	0.15		
3.814	994841	1154620	1.161	2.40		
3.825	1233767	1194510	0.968	2.98		
3.930	845064	650154	0.769	2.04		
3.994	559045	538233	0.963	1.35		
4.057	330303	301258	0.912	0.79		
4.136	140355	120359	0.858	0.33		
4.215	1409463	1316932	0.934	3.41		
4.261	1114862	1076234	0.965	2.69	6 Aroclor 1248	BA
4.292	463155	471683	1.018	1.12		
4.309	485188	464077	0.956	1.17		
4.416	1717291	1589511	0.926	4.15		
4.452	836688	731802	0.875	2.02		
4.546	1950169	1837348	0.942	4.72	6 Aroclor 1248	BA
4.591	28359	30347	1.070	0.06		
4.617	447635	419819	0.938	1.08		
4.678	26595	19613	0.737	0.06		
4.728	60694	48703	0.802	0.14		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.768	104450	99117	0.949	0.25	
4.812	961742	956981	0.995	2.32	6 Aroclor 1248
4.847	1854263	1753514	0.946	4.48	
4.878	2375611	1575651	0.663	5.75	6 Aroclor 1248
4.957	277211	232416	0.838	0.67	
5.036	1686235	1578524	0.936	4.08	6 Aroclor 1248
5.098	601218	504142	0.839	1.45	
5.162	718836	651581	0.906	1.74	
5.212	514473	466688	0.907	1.24	
5.283	38226	35590	0.931	0.09	
5.308	114377	97670	0.854	0.27	
5.362	454967	386178	0.849	1.10	
5.406	755181	677323	0.897	1.82	
5.445	463914	420122	0.906	1.12	
5.509	1264136	1067561	0.844	3.06	
5.604	358061	318483	0.889	0.86	
5.674	40456	36539	0.903	0.09	
5.715	81181	61479	0.757	0.19	
5.741	96448	82759	0.858	0.23	
5.763	28052	36234	1.292	0.06	
5.791	850989	716872	0.842	2.06	
5.886	86937	76656	0.882	0.21	
5.910	56059	30482	0.544	0.13	
5.995	26545	26566	1.001	0.06	
6.023	715923	664644	0.928	1.73	
6.055	148446	120732	0.813	0.35	
6.167	76183	49935	0.655	0.18	
6.227	49695	27459	0.553	0.12	
6.298	222047	190922	0.860	0.53	
6.333	42944	36638	0.853	0.10	
6.373	36944	28876	0.782	0.08	
6.508	37862	37090	0.980	0.09	
6.535	87189	61825	0.709	0.21	
6.603	15343	11595	0.756	0.03	
6.718	37237	33179	0.891	0.09	
6.756	19515	17280	0.885	0.04	
6.800	39294	34828	0.886	0.09	
6.841	12660	12553	0.992	0.03	
7.004	64752	61162	0.945	0.15	
7.126	17005	12165	0.715	0.04	
7.252	28814	24913	0.865	0.06	
7.390	29822	24928	0.836	0.07	
7.442	31654	27393	0.865	0.07	
7.694	15473	13759	0.889	0.03	
7.729	24068	21868	0.909	0.05	
7.898	26492	24740	0.934	0.06	
8.273	51775	43579	0.842	0.12	
8.380	19276	13488	0.700	0.04	
8.570	4913033	4227369	0.860	11.89	$ 12 Decachlorobiphenyl
8.941	17085	13032	0.763	0.04	
9.463	11934	9303	0.780	0.02	
	41307112	37154977		100.000	

Total unknown % area = 54

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318016.D
Lab Smp Id: 1248-CCV Client Smp ID: 1248-CCV
Inj Date : 13-MAR-2018 17:28
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-ccv, 185549:1
Misc Info : 17498,AR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031318.b\pcba030618.m
Meth Date : 27-Mar-2018 10:38 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 16 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1248.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.484	2.484	0.000	5717514	0.05000	0.046			
6 Aroclor 1248					CAS #: 12672-29-6			
4.260	4.261	-0.001	1114862	0.50000	0.48	0.00- 20.00	100.00(M)	BA
4.545	4.546	-0.001	1950169	0.50000	0.47	0.00- 20.00	174.92	BA
4.812	4.812	0.000	961742	0.50000	0.47	0.00- 20.00	86.27	BA
4.877	4.879	-0.002	2375610	0.50000	0.47	0.00- 20.00	213.09	BA
5.036	5.038	-0.002	1686234	0.50000	0.47	0.00- 20.00	151.25	BA
Average of Peak Amounts =					0.47200			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.570	8.571	-0.001	4913033	0.05000	0.045			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318016.D
Date : 13-MAR-2018 17:28
Client ID: 1248-CCV
Sample Info: 1248-CCV, 185549:1

Instrument: 40GCSJ.i
Operator: BLM
Column phase: ZB-1ms
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031318.b/031318016.D (ManInt: AR1248.sub Orig: AR1248.sub)
Injection Date: 13-MAR-2018 17:28 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1248-CCV





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1248	1104511	1114862	BA
Aroclor 1248	1104511	1114862	BA
Aroclor 1248	1104511	1114862	BA
Aroclor 1248	1104511	1114862	BA
Aroclor 1248	1104511	1114862	BA
Tetrachloro-m-xylene	5717514	5717514	
Decachlorobiphenyl	4913033	4913033	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418004.D
Lab Smp Id: 1242-CCV Client Smp ID: 1242-CCV
Inj Date : 14-MAR-2018 13:12
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-ccv, 185544:1
Misc Info : 17498,AR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 4 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1242.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.182	22076	17935	0.812	0.05		
1.430	39238	30892	0.787	0.10		
1.541	19164	13983	0.730	0.05		
1.685	16120	10882	0.675	0.04		
1.806	16509	13973	0.846	0.04		
2.009	17580	12745	0.725	0.04		
2.405	115981	91487	0.789	0.30		
2.485	5888335	5613246	0.953	15.88	$ 2 Tetrachloro-m-xylene	
2.687	171095	140396	0.821	0.44		
2.789	260946	244164	0.936	0.68		
2.851	909776	851980	0.936	2.38		
3.055	137084	129521	0.945	0.36		
3.279	22074	16599	0.752	0.05		
3.318	1255470	1168373	0.931	3.29	3 Aroclor 1242	BA
3.344	702067	669480	0.954	1.84		
3.435	197936	180141	0.910	0.51		
3.514	1236536	900390	0.728	3.24	3 Aroclor 1242	BA
3.609	18910	16358	0.865	0.04		
3.695	38171	33394	0.875	0.10		
3.733	310197	314001	1.012	0.81		
3.756	205378	191849	0.934	0.53		
3.827	3460588	2056369	0.594	9.09	3 Aroclor 1242	BA
3.930	1487913	1260705	0.847	3.90	3 Aroclor 1242	BA
3.994	985943	960466	0.974	2.58	3 Aroclor 1242	BA
4.059	296137	272644	0.921	0.77		
4.137	117009	105726	0.904	0.30		
4.215	835166	816313	0.977	2.19		
4.261	804392	773226	0.961	2.11		
4.293	310468	338069	1.089	0.81		
4.310	430859	390008	0.905	1.13		
4.371	29754	19790	0.665	0.07		
4.418	1104858	1056191	0.956	2.90		
4.452	840769	631429	0.751	2.20		
4.546	1420053	1329669	0.936	3.73		
4.592	19452	19420	0.998	0.05		
4.617	326036	304299	0.933	0.85		
4.678	26145	19683	0.753	0.06		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.727	68612	60112	0.876	0.18	
4.769	81072	73992	0.913	0.21	
4.814	657026	631706	0.961	1.72	
4.848	1121735	1105088	0.985	2.94	
4.879	1445784	1073292	0.742	3.79	
4.960	162404	131137	0.807	0.42	
5.037	1182939	1100605	0.930	3.10	
5.100	264233	208210	0.788	0.69	
5.163	303561	268294	0.884	0.79	
5.212	233196	207188	0.888	0.61	
5.285	18911	19607	1.037	0.04	
5.311	53463	44553	0.833	0.14	
5.364	200175	170063	0.850	0.52	
5.407	325034	280038	0.862	0.85	
5.446	218342	180420	0.826	0.57	
5.511	529005	409685	0.774	1.38	
5.604	161355	137557	0.853	0.42	
5.718	37481	28404	0.758	0.09	
5.740	45007	38765	0.861	0.11	
5.793	359972	305266	0.848	0.94	
5.889	60748	35982	0.592	0.15	
6.024	349104	269609	0.772	0.91	
6.172	23208	22123	0.953	0.06	
6.299	101788	88029	0.865	0.26	
6.331	25182	19687	0.782	0.06	
6.375	17173	16197	0.943	0.04	
6.536	34760	31749	0.913	0.09	
6.608	14704	8524	0.580	0.03	
6.801	23516	21659	0.921	0.06	
7.247	13141	9241	0.703	0.03	
7.730	25515	23370	0.916	0.06	
7.801	16274	14505	0.891	0.04	
8.277	49081	39467	0.804	0.12	
8.375	13242	10800	0.816	0.03	
8.572	5097087	4226833	0.829	13.38	$ 12 Decachlorobiphenyl
8.943	15637	12998	0.831	0.04	
9.469	14777	10007	0.677	0.03	
9.877	607033	91945	0.151	1.59	
	38067466	32442503		100.000	

Total unknown % area = 49

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418004.D
Lab Smp Id: 1242-CCV Client Smp ID: 1242-CCV
Inj Date : 14-MAR-2018 13:12
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-ccv, 185544:1
Misc Info : 17498,AR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 4 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1242.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.484	2.484	0.000	5888335	0.05000	0.048			
3 Aroclor 1242					CAS #: 53469-21-9			
3.317	3.318	-0.001	1255470	0.50000	0.50	0.00- 20.00	100.00(M)	BA
3.513	3.513	0.000	1236536	0.50000	0.49	0.00- 20.00	98.49	BA
3.826	3.827	-0.001	3460588	0.50000	0.49	0.00- 20.00	275.64	BA
3.929	3.929	0.000	1487912	0.50000	0.49	0.00- 20.00	118.51	BA
3.993	3.995	-0.002	985943	0.50000	0.48	0.00- 20.00	78.53	BA
Average of Peak Amounts =					0.49000			
$ 12 Decachlorobiphenyl					CAS #: 2051-24-3			
8.572	8.571	0.001	5097087	0.05000	0.046			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418004.D
Date : 14-MAR-2018 13:12
Client ID: 1242-CCV
Sample Info: 1242-CCV, 185544:1

Instrument: 40GCSJ.i
Operator: BLM
Column phase: ZB-1ms
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b/031418004.D (ManInt: AR1242.sub Orig: AR1242.sub)
Injection Date: 14-MAR-2018 13:12 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1242-CCV





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1242	1248314	1255470	BA
Aroclor 1242	1248314	1255470	BA
Aroclor 1242	1248314	1255470	BA
Aroclor 1242	1248314	1255470	BA
Aroclor 1242	1248314	1255470	BA
Tetrachloro-m-xylene	5888335	5888335	
Decachlorobiphenyl	5097087	5097087	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418016.D
Lab Smp Id: 1248-CCV Client Smp ID: 1248-CCV
Inj Date : 14-MAR-2018 16:22
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-ccv, 185549:1
Misc Info : 17498,AR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 16 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1248.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.183	18627	13256	0.712	0.04		
1.428	36768	32675	0.889	0.08		
1.534	14066	14342	1.020	0.03		
1.686	13130	13208	1.006	0.02		
2.011	10133	9076	0.896	0.02		
2.405	27213	17737	0.652	0.06		
2.484	6032956	5829574	0.966	13.98	$ 2 Tetrachloro-m-xylene	
2.687	34908	26724	0.766	0.07		
2.788	119717	111094	0.928	0.26		
2.850	195041	177689	0.911	0.43		
3.054	64449	55436	0.860	0.14		
3.317	718105	686764	0.956	1.61		
3.345	299217	285038	0.953	0.67		
3.433	72953	60617	0.831	0.16		
3.516	603279	437537	0.725	1.35		
3.611	17289	18206	1.053	0.03		
3.732	136608	137299	1.005	0.30		
3.757	57934	57809	0.998	0.13		
3.815	1057961	1244674	1.176	2.37		
3.826	1355215	1290034	0.952	3.04		
3.931	912472	714020	0.783	2.04		
3.994	602488	593646	0.985	1.35		
4.059	344628	324565	0.942	0.77		
4.135	146530	131406	0.897	0.32		
4.214	1499401	1400578	0.934	3.36		
4.262	1200161	1137794	0.948	2.69	6 Aroclor 1248	BA
4.292	493781	506105	1.025	1.10		
4.310	521623	495104	0.949	1.17		
4.417	1835807	1730766	0.943	4.12		
4.453	902046	773821	0.858	2.02		
4.545	2094456	1955584	0.934	4.70	6 Aroclor 1248	BA
4.590	29033	31907	1.099	0.06		
4.617	483093	462687	0.958	1.08		
4.680	26626	21385	0.803	0.05		
4.727	64575	50218	0.778	0.14		
4.768	110472	109009	0.987	0.24		
4.813	1056328	1019316	0.965	2.37	6 Aroclor 1248	BA

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.848	2008478	1926373	0.959	4.50	
4.879	2586326	1744443	0.674	5.80	6 Aroclor 1248
4.957	302271	250896	0.830	0.67	
5.037	1850271	1791747	0.968	4.15	6 Aroclor 1248
5.099	648459	513391	0.792	1.45	
5.163	779707	705528	0.905	1.75	
5.212	552822	504757	0.913	1.24	
5.285	38549	36530	0.948	0.08	
5.309	123207	108433	0.880	0.27	
5.363	497849	437511	0.879	1.11	
5.406	821310	728844	0.887	1.84	
5.446	510589	462807	0.906	1.14	
5.511	1374644	1178494	0.857	3.08	
5.604	389101	353439	0.908	0.87	
5.673	47947	39201	0.818	0.10	
5.717	81548	63494	0.779	0.18	
5.740	103264	90439	0.876	0.23	
5.764	29666	41638	1.404	0.06	
5.792	942978	767008	0.813	2.11	
5.890	127991	82458	0.644	0.28	
5.994	29893	29133	0.975	0.06	
6.024	799758	717924	0.898	1.79	
6.057	155771	128120	0.822	0.34	
6.169	86517	55456	0.641	0.19	
6.227	52978	30864	0.583	0.11	
6.300	242370	208493	0.860	0.54	
6.335	45347	39677	0.875	0.10	
6.374	35107	30664	0.873	0.07	
6.510	40730	38512	0.946	0.09	
6.536	98531	71340	0.724	0.22	
6.603	17977	11312	0.629	0.04	
6.662	10230	9472	0.926	0.02	
6.720	43466	37297	0.858	0.09	
6.759	22376	19324	0.864	0.05	
6.804	47588	41152	0.865	0.10	
6.840	20806	15662	0.753	0.04	
7.005	80748	65198	0.807	0.18	
7.120	10187	10010	0.983	0.02	
7.253	32716	26820	0.820	0.07	
7.281	9398	8624	0.918	0.02	
7.390	31404	28309	0.901	0.07	
7.446	39542	30627	0.775	0.08	
7.691	15327	14556	0.950	0.03	
7.728	25197	23694	0.940	0.05	
7.801	22218	13854	0.624	0.04	
7.903	29108	26720	0.918	0.06	
8.275	59263	49344	0.833	0.13	
8.377	15156	10204	0.673	0.03	
8.572	5357294	4462099	0.833	12.02	$ 12 Decachlorobiphenyl
8.935	9803	6990	0.713	0.02	
9.469	13670	9800	0.717	0.03	
9.961	11612	9138	0.787	0.02	
	44538155	40084520		100.000	

Total unknown % area = 54

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418016.D
Lab Smp Id: 1248-CCV Client Smp ID: 1248-CCV
Inj Date : 14-MAR-2018 16:22
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1248-ccv, 185549:1
Misc Info : 17498,AR1248
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 16 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1248.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene						CAS #: 877-09-8		
2.484	2.484	0.000	6032956	0.05000	0.049			
6 Aroclor 1248						CAS #: 12672-29-6		
4.261	4.261	0.000	1200160	0.50000	0.51	0.00- 20.00	100.00(M)	BA
4.545	4.546	-0.001	2094455	0.50000	0.50	0.00- 20.00	174.51	BA
4.813	4.812	0.001	1056327	0.50000	0.52	0.00- 20.00	88.02	BA
4.878	4.879	-0.001	2586326	0.50000	0.51	0.00- 20.00	215.50	BA
5.037	5.038	-0.001	1850271	0.50000	0.52	0.00- 20.00	154.17	BA
Average of Peak Amounts =					0.51200			
$ 12 Decachlorobiphenyl						CAS #: 2051-24-3		
8.572	8.571	0.001	5357293	0.05000	0.049			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418016.D
Date : 14-MAR-2018 16:22
Client ID: 1248-CCV
Sample Info: 1248-CCV, 185549:1

Instrument: 40GCSJ.i
Operator: BLM
Column phase: ZB-1ms
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b/031418016.D (ManInt: AR1248.sub Orig: AR1248.sub)
Injection Date: 14-MAR-2018 16:22 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1248-CCV





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1248	1190556	1200160	BA
Aroclor 1248	1190556	1200160	BA
Aroclor 1248	1190556	1200160	BA
Aroclor 1248	1190556	1200160	BA
Aroclor 1248	1190556	1200160	BA
Tetrachloro-m-xylene	6032956	6032956	
Decachlorobiphenyl	5357293	5357293	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418025.D
Lab Smp Id: 1254-CCV Client Smp ID: 1254-CCV
Inj Date : 14-MAR-2018 18:45
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-ccv, 185554:1
Misc Info : 17498,AR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 25 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1254.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.184	24572	20116	0.819	0.04		
1.429	36622	31232	0.853	0.06		
1.539	20957	13856	0.661	0.03		
1.688	17722	13538	0.764	0.03		
2.010	17777	12901	0.726	0.03		
2.485	6021859	5676111	0.943	11.72	$ 2 Tetrachloro-m-xylene	
2.786	86525	81827	0.946	0.16		
2.853	21069	18842	0.894	0.03		
3.032	12297	10731	0.873	0.02		
3.051	21824	21017	0.963	0.04		
3.317	34182	27244	0.797	0.06		
3.513	30637	22358	0.730	0.05		
3.611	19057	18113	0.950	0.03		
3.813	75556	52102	0.690	0.14		
3.933	61417	41905	0.682	0.11		
3.995	36753	29239	0.796	0.06		
4.057	26484	21425	0.809	0.04		
4.129	10042	9411	0.937	0.01		
4.215	1234432	1175472	0.952	2.31		
4.261	326766	322407	0.987	0.61		
4.293	62101	46515	0.749	0.11		
4.417	758074	740736	0.977	1.42		
4.454	72519	67618	0.932	0.13		
4.545	344351	327986	0.952	0.64		
4.591	14249	14184	0.995	0.02		
4.616	60501	55621	0.919	0.11		
4.682	22449	16649	0.742	0.04		
4.728	10906	9182	0.842	0.02		
4.768	22259	19942	0.896	0.04		
4.813	292766	286646	0.979	0.54		
4.848	1132250	1069440	0.945	2.12		
4.894	2407481	1904557	0.791	4.51		
4.957	384890	349458	0.908	0.72		
5.037	246240	230251	0.935	0.46		
5.099	1413141	1244721	0.881	2.65		
5.162	2602349	2407445	0.925	4.88	7 Aroclor 1254	BA
5.212	1183950	1095673	0.925	2.22		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
5.283	60819	58379	0.960	0.11	
5.309	213913	202347	0.946	0.40	
5.364	1036263	949891	0.917	1.94	
5.407	2174206	1971673	0.907	4.08	7 Aroclor 1254
5.446	737824	691000	0.937	1.38	
5.478	258469	254600	0.985	0.48	
5.510	3884254	3534104	0.910	7.28	7 Aroclor 1254
5.577	28886	30260	1.048	0.05	
5.604	539802	503344	0.932	1.01	
5.675	346396	327137	0.944	0.65	
5.711	533598	353736	0.663	1.00	
5.740	339950	285535	0.840	0.63	
5.791	3962745	3181602	0.803	7.43	
5.885	344321	278009	0.807	0.64	
5.926	174589	115122	0.659	0.32	
5.994	325318	317899	0.977	0.61	
6.026	2548480	2148892	0.843	4.78	7 Aroclor 1254
6.056	1723792	1505010	0.873	3.23	
6.169	707776	596569	0.843	1.32	
6.228	288011	298024	1.035	0.54	
6.241	321442	295463	0.919	0.60	
6.299	3322960	2873175	0.865	6.23	7 Aroclor 1254
6.334	572000	551168	0.964	1.07	
6.373	325427	277824	0.854	0.61	
6.436	80149	49444	0.617	0.15	
6.489	18645	16257	0.872	0.03	
6.509	117671	125936	1.070	0.22	
6.536	1131240	904878	0.800	2.12	
6.604	161818	138021	0.853	0.30	
6.662	27108	23701	0.874	0.05	
6.718	219047	185836	0.848	0.41	
6.758	130281	116871	0.897	0.24	
6.802	638961	602524	0.943	1.19	
6.844	166163	118835	0.715	0.31	
6.941	56974	51490	0.904	0.10	
7.005	419420	401028	0.956	0.78	
7.038	30589	26386	0.863	0.05	
7.076	19919	17593	0.883	0.03	
7.252	364881	329635	0.903	0.68	
7.282	70393	65760	0.934	0.13	
7.393	21162	15756	0.745	0.03	
7.443	27862	22845	0.820	0.05	
7.554	25351	17779	0.701	0.04	
7.692	14805	13185	0.891	0.02	
7.727	20847	20709	0.993	0.03	
7.778	9359	10676	1.141	0.01	
7.801	15639	14010	0.896	0.02	
7.899	27971	20098	0.719	0.05	
8.273	53919	43404	0.805	0.10	
8.572	5402337	4646474	0.860	10.13	$ 12 Decachlorobiphenyl
8.939	13359	10613	0.794	0.02	
9.472	37063	18505	0.499	0.06	
9.961	27608	21916	0.794	0.05	
	53288806	47157399		100.000	

Total unknown % area = 51

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418025.D
Lab Smp Id: 1254-CCV Client Smp ID: 1254-CCV
Inj Date : 14-MAR-2018 18:45
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1254-ccv, 185554:1
Misc Info : 17498,AR1254
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 25 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1254.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene						CAS #: 877-09-8		
2.484	2.484	0.000	6021858	0.05000	0.049			
7 Aroclor 1254						CAS #: 11097-69-1		
5.162	5.162	0.000	2602349	0.50000	0.50	0.00- 20.00	100.00(M)	BA
5.406	5.408	-0.002	2174205	0.50000	0.50	0.00- 20.00	83.55	BA
5.510	5.510	0.000	3884254	0.50000	0.50	0.00- 20.00	149.26	BA
6.026	6.027	-0.001	2548480	0.50000	0.52	0.00- 20.00	97.93	BA
6.299	6.299	0.000	3322959	0.50000	0.51	0.00- 20.00	127.69	BA
Average of Peak Amounts =					0.50600			
$ 12 Decachlorobiphenyl						CAS #: 2051-24-3		
8.572	8.571	0.001	5402337	0.05000	0.049			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418025.D
Date : 14-MAR-2018 18:45
Client ID: 1254-CCV
Sample Info: 1254-CCV, 185554:1

Instrument: 40GCSJ.i

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b/031418025.D (ManInt: AR1254.sub Orig: AR1254.sub)
Injection Date: 14-MAR-2018 18:45 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1254-CCV





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1254	2593104	2602349	BA
Aroclor 1254	2593104	2602349	BA
Aroclor 1254	2593104	2602349	BA
Aroclor 1254	2593104	2602349	BA
Aroclor 1254	2593104	2602349	BA
Tetrachloro-m-xylene	6021858	6021858	
Decachlorobiphenyl	5402337	5402337	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018004.D
Lab Smp Id: 1232-CCV Client Smp ID: 1232-CCV
Inj Date : 30-MAR-2018 15:47
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1232-ccv, 185539:1
Misc Info : 17498,AR1232
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\033018.b\pcba030618.m
Meth Date : 04-Apr-2018 13:12 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 4 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1232.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.184	28611	24110	0.843	0.10		
1.432	38050	29406	0.773	0.14		
1.537	16847	14036	0.833	0.06		
1.686	18795	16246	0.864	0.07		
1.806	293863	246585	0.839	1.11		
2.011	20417	15641	0.766	0.07		
2.198	82173	72036	0.877	0.31		
2.406	195034	143989	0.738	0.73		
2.485	5786443	5418296	0.936	22.29	$ 2 Tetrachloro-m-xylene	
2.688	423326	346437	0.818	1.60		
2.789	394101	358789	0.910	1.49		
2.851	1269350	1141797	0.900	4.80	4 Aroclor 1232	BA
3.029	11732	14025	1.195	0.04		
3.054	90724	73974	0.815	0.34		
3.281	66768	49387	0.740	0.25		
3.318	736732	655403	0.890	2.79		
3.344	402334	361823	0.899	1.52		
3.433	113342	103719	0.915	0.42		
3.513	669053	481887	0.720	2.53	4 Aroclor 1232	BA
3.609	20062	17884	0.891	0.07		
3.698	24610	20932	0.851	0.09		
3.732	176560	173194	0.981	0.66		
3.756	126992	114145	0.899	0.48		
3.826	1865355	1089801	0.584	7.06	4 Aroclor 1232	BA
3.930	797462	669657	0.840	3.02	4 Aroclor 1232	BA
3.995	527323	498762	0.946	1.99	4 Aroclor 1232	BA
4.059	153252	131901	0.861	0.58		
4.137	59676	53239	0.892	0.22		
4.215	394581	370771	0.940	1.49		
4.261	383584	358821	0.935	1.45		
4.293	168081	172238	1.025	0.63		
4.310	197644	184817	0.935	0.74		
4.417	533725	493147	0.924	2.02		
4.452	414687	311008	0.750	1.57		
4.547	676915	620446	0.917	2.56		
4.618	163928	152581	0.931	0.62		
4.677	16561	13343	0.806	0.06		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.727	33649	30922	0.919	0.12	
4.768	37020	33477	0.904	0.14	
4.813	297483	283159	0.952	1.12	
4.848	519297	483260	0.931	1.96	
4.879	679028	512647	0.755	2.57	
4.958	74870	56417	0.754	0.28	
5.037	551301	496890	0.901	2.08	
5.098	114768	86161	0.751	0.43	
5.163	127977	114468	0.894	0.48	
5.212	90559	78383	0.866	0.34	
5.308	25509	19884	0.780	0.09	
5.363	78766	67279	0.854	0.29	
5.407	138700	114980	0.829	0.52	
5.446	84273	70209	0.833	0.31	
5.510	220958	164243	0.743	0.83	
5.605	67711	57426	0.848	0.25	
5.672	13331	11647	0.874	0.05	
5.791	146328	110676	0.756	0.55	
5.884	18310	13481	0.736	0.06	
6.025	106645	95460	0.895	0.40	
6.056	33501	27285	0.814	0.12	
6.170	14328	11980	0.836	0.05	
6.299	39749	32967	0.829	0.15	
6.509	12262	11270	0.919	0.04	
6.719	16301	13619	0.835	0.06	
7.003	30494	20946	0.687	0.11	
7.124	27056	15787	0.583	0.10	
7.251	15684	11663	0.744	0.05	
7.438	9492	7819	0.824	0.03	
7.692	13467	9906	0.736	0.05	
7.726	30600	26257	0.858	0.11	
7.768	16249	11313	0.696	0.06	
7.799	27557	22489	0.816	0.10	
8.277	47916	40368	0.842	0.18	
8.378	21581	16221	0.752	0.08	
8.571	5107752	4316134	0.845	19.35	$ 12 Decachlorobiphenyl
8.938	10862	9594	0.883	0.04	
9.410	9812	13044	1.329	0.03	
9.426	12356	16262	1.316	0.04	
9.465	69759	22340	0.320	0.26	
9.854	13935	10507	0.754	0.05	
9.867	13247	9964	0.752	0.05	
9.957	13957	11464	0.821	0.05	
	26393094	22604541		100.000	

Total unknown % area = 39

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018004.D
Lab Smp Id: 1232-CCV Client Smp ID: 1232-CCV
Inj Date : 30-MAR-2018 15:47
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1232-ccv, 185539:1
Misc Info : 17498,AR1232
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\033018.b\pcba030618.m
Meth Date : 04-Apr-2018 13:12 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 4 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1232.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene						CAS #: 877-09-8		
2.485	2.484	0.001	5786442	0.05000	0.047			
4 Aroclor 1232						CAS #: 1114-16-5		
2.851	2.851	0.000	1269349	0.50000	0.51	0.00- 20.00	100.00(M)	BA
3.513	3.513	0.000	669052	0.50000	0.50	0.00- 20.00	52.71	BA
3.826	3.827	-0.001	1865355	0.50000	0.50	0.00- 20.00	146.95	BA
3.929	3.929	0.000	797462	0.50000	0.50	0.00- 20.00	62.82	BA
3.994	3.995	-0.001	527322	0.50000	0.48	0.00- 20.00	41.54	BA
Average of Peak Amounts =					0.49800			
$ 12 Decachlorobiphenyl						CAS #: 2051-24-3		
8.571	8.571	0.000	5107751	0.05000	0.047			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018004.D

Date : 30-MAR-2018 15:47

Client ID: 1232-CCV

Instrument: 40GCSJ.i

Sample Info: 1232-CCV, 185539:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\033018.b/033018004.D (ManInt: AR1232.sub Orig: AR1232.sub)
Injection Date: 30-MAR-2018 15:47 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1232-CCV





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1232	1259674	1269349	BA
Aroclor 1232	1259674	1269349	BA
Aroclor 1232	1259674	1269349	BA
Aroclor 1232	1259674	1269349	BA
Aroclor 1232	1259674	1269349	BA
Tetrachloro-m-xylene	5786442	5786442	
Decachlorobiphenyl	5107751	5107751	

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018016.D
Lab Smp Id: 1242-CCV Client Smp ID: 1242-CCV
Inj Date : 30-MAR-2018 18:57
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-ccv, 185544:1
Misc Info : 17498,AR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\033018.b\pcba030618.m
Meth Date : 04-Apr-2018 13:12 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 16 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1242.sub
Target Version: RC10A Sample Matrix: None

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.183	22405	18532	0.827	0.05		
1.431	40276	30743	0.763	0.10		
1.539	21364	14547	0.681	0.05		
1.688	39933	33591	0.841	0.10		
1.806	15485	14673	0.948	0.03		
2.405	121508	92266	0.759	0.30		
2.485	6015047	5620296	0.934	15.57	$ 2 Tetrachloro-m-xylene	
2.686	179100	142673	0.797	0.45		
2.788	272860	250720	0.919	0.68		
2.851	954082	877697	0.920	2.40		
3.054	143309	127113	0.887	0.36		
3.278	26506	18895	0.713	0.06		
3.318	1295039	1187612	0.917	3.27	3 Aroclor 1242	BA
3.344	747922	694661	0.929	1.88		
3.435	208397	185067	0.888	0.52		
3.512	1286692	915338	0.711	3.25	3 Aroclor 1242	BA
3.610	24609	19552	0.795	0.06		
3.651	23747	16018	0.675	0.05		
3.697	36677	35434	0.966	0.09		
3.731	326714	316997	0.970	0.82		
3.756	224700	206128	0.917	0.56		
3.826	3677233	2126529	0.578	9.28	3 Aroclor 1242	BA
3.929	1578983	1297420	0.822	3.98	3 Aroclor 1242	BA
3.994	1048461	998756	0.953	2.64	3 Aroclor 1242	BA
4.058	319715	283616	0.887	0.80		
4.137	129269	107709	0.833	0.32		
4.214	874601	807996	0.924	2.20		
4.260	836630	806365	0.964	2.11		
4.292	341738	359002	1.051	0.86		
4.310	446814	398295	0.891	1.12		
4.370	31927	21945	0.687	0.08		
4.416	1142103	1082234	0.948	2.88		
4.451	913690	659762	0.722	2.30		
4.545	1510362	1343728	0.890	3.81		
4.617	346342	309594	0.894	0.87		
4.680	28187	21496	0.763	0.07		
4.725	76119	62203	0.817	0.19		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.768	83108	78177	0.941	0.20	
4.812	699910	670567	0.958	1.76	
4.847	1202699	1133950	0.943	3.03	
4.878	1544124	1135518	0.735	3.90	
4.957	178912	133914	0.748	0.45	
5.037	1280421	1148948	0.897	3.23	
5.098	277704	209878	0.756	0.70	
5.161	321799	274805	0.854	0.81	
5.212	247007	219787	0.890	0.62	
5.284	20360	19125	0.939	0.05	
5.306	56120	46311	0.825	0.14	
5.362	209417	177043	0.845	0.52	
5.405	350926	280474	0.799	0.88	
5.444	220385	194602	0.883	0.55	
5.509	566724	432643	0.763	1.43	
5.603	171864	143328	0.834	0.43	
5.671	16234	12840	0.791	0.04	
5.714	36570	29147	0.797	0.09	
5.740	47521	42277	0.890	0.12	
5.760	14752	20251	1.373	0.03	
5.792	396864	315002	0.794	1.00	
5.888	62266	38258	0.614	0.15	
6.023	379459	274459	0.723	0.95	
6.168	37049	22222	0.600	0.09	
6.225	31667	17003	0.537	0.07	
6.298	108595	93635	0.862	0.27	
6.332	23600	21579	0.914	0.05	
6.372	18762	14987	0.799	0.04	
6.533	46511	33158	0.713	0.11	
6.799	24472	18645	0.762	0.06	
7.250	11396	10214	0.896	0.02	
7.726	29012	23418	0.807	0.07	
7.796	24839	18196	0.733	0.06	
8.274	54802	45908	0.838	0.13	
8.570	5404034	4589082	0.849	13.65	$ 12 Decachlorobiphenyl
8.934	12372	9481	0.766	0.03	
9.463	21668	13742	0.634	0.05	
9.959	26183	16834	0.643	0.06	
	39588648	33484611		100.000	

Total unknown % area = 48

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018016.D
Lab Smp Id: 1242-CCV Client Smp ID: 1242-CCV
Inj Date : 30-MAR-2018 18:57
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1242-ccv, 185544:1
Misc Info : 17498,AR1242
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\033018.b\pcba030618.m
Meth Date : 04-Apr-2018 13:12 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 16 Continuing Calibration Sample
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: AR1242.sub
Target Version: RC10A Sample Matrix: None

RT	EXP RT	DLT RT	RESPONSE	AMOUNTS CAL-AMT (ug/ml)	AMOUNTS ON-COL (ug/ml)	TARGET RANGE	RATIO	REVIEW CODE
$ 2 Tetrachloro-m-xylene						CAS #: 877-09-8		
2.484	2.484	0.000	6015046	0.05000	0.049			
3 Aroclor 1242						CAS #: 53469-21-9		
3.318	3.318	0.000	1295038	0.50000	0.51	0.00- 20.00	100.00(M)	BA
3.512	3.513	-0.001	1286692	0.50000	0.51	0.00- 20.00	99.36	BA
3.825	3.827	-0.002	3677232	0.50000	0.52	0.00- 20.00	283.95	BA
3.928	3.929	-0.001	1578983	0.50000	0.52	0.00- 20.00	121.93	BA
3.994	3.995	-0.001	1048461	0.50000	0.51	0.00- 20.00	80.96	BA
Average of Peak Amounts =					0.51400			
$ 12 Decachlorobiphenyl						CAS #: 2051-24-3		
8.570	8.571	-0.001	5404034	0.05000	0.049			

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018016.D

Date : 30-MAR-2018 18:57

Client ID: 1242-CCV

Instrument: 40GCSJ.i

Sample Info: 1242-CCV, 185544:1

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\033018.b/033018016.D (ManInt: AR1242.sub Orig: AR1242.sub)
Injection Date: 30-MAR-2018 18:57 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1242-CCV





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1242	1286633	1295038	BA
Aroclor 1242	1286633	1295038	BA
Aroclor 1242	1286633	1295038	BA
Aroclor 1242	1286633	1295038	BA
Aroclor 1242	1286633	1295038	BA
Tetrachloro-m-xylene	6015046	6015046	
Decachlorobiphenyl	5404034	5404034	

SAMPLE NO.

BLANK

PCB - FORM I SVOA-1

SEMI-VOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay

Contract: HUDSON RIVER REMEDIAL ACTION

Date Received:

Matrix: Tissue SDG No.: 40151601

Date Extracted: 03/12/2018 07:42

Lab Sample ID: 1655661

Date Analyzed: 03/13/2018 14:50

Lab File ID: 031318.B\031318006.D

Initial wt/vol: 10 g Final wt/vol: 5 mL Dilution: 1

Instrument: 40GCSJ Percent Moisture:

CAS NO.	COMPOUND	CONCENTRATION UNITS: ug/kg	Q
12674-11-2	PCB-1016 (Aroclor 1016)	<11.4	U
11104-28-2	PCB-1221 (Aroclor 1221)	<11.4	U
11141-16-5	PCB-1232 (Aroclor 1232)	<11.4	U
53469-21-9	PCB-1242 (Aroclor 1242)	<11.4	U
12672-29-6	PCB-1248 (Aroclor 1248)	<11.4	U
11097-69-1	PCB-1254 (Aroclor 1254)	<11.4	U
11096-82-5	PCB-1260 (Aroclor 1260)	<11.4	U

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318006.D
Lab Smp Id: 1655661 Client Smp ID: MB
Inj Date : 13-MAR-2018 14:50
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1655661,
Misc Info : 17521,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031318.b\pcba030618.m
Meth Date : 27-Mar-2018 10:38 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 6 QC Sample: BLANK
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.000	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.180	14983	13696	0.914	0.08		
1.428	13915	10443	0.750	0.07		
1.540	12689	9107	0.718	0.06		
2.011	11512	8184	0.711	0.06		
2.484	8963759	8758350	0.977	48.60	$ 2 Tetrachloro-m-xylene	
2.785	16109	15841	0.983	0.08		
2.861	20012	15272	0.763	0.10		
6.445	11005	7697	0.699	0.05		
7.123	20596	15757	0.765	0.11		
7.689	10803	11523	1.067	0.05		
7.727	51054	42336	0.829	0.27		
7.778	9569	9201	0.962	0.05		
7.800	41503	32244	0.777	0.22		
8.275	87389	71886	0.823	0.47		
8.378	19258	14526	0.754	0.10		
8.570	9095946	7799740	0.857	49.41	$ 12 Decachlorobiphenyl	
8.942	16917	10192	0.602	0.09		
9.463	10126	7844	0.775	0.05		
9.964	16090	12348	0.767	0.08		
	18443237	16866187		100.000		

Total unknown % area = 2.0

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318006.D
Lab Smp Id: 1655661 Client Smp ID: MB
Inj Date : 13-MAR-2018 14:50
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1655661,
Misc Info : 17521,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031318.b\pcba030618.m
Meth Date : 27-Mar-2018 10:38 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 6 QC Sample: BLANK
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.000	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221						CAS #: 11104-28-2		
Peaks not detected for Quant. or Qual. signal(s).								
$ 2 Tetrachloro-m-xylene						CAS #: 877-09-8		
2.484	2.484	0.000	8963758	0.07270	36			
3 Aroclor 1242						CAS #: 53469-21-9		
Peaks not detected for Quant. or Qual. signal(s).								
4 Aroclor 1232						CAS #: 1114-16-5		
Peaks not detected for Quant. or Qual. signal(s).								

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE

5 Aroclor 1016 CAS #: 12674-11-2

Peaks not detected for Quant. or Qual. signal(s).

---

6 Aroclor 1248 CAS #: 12672-29-6

Peaks not detected for Quant. or Qual. signal(s).

---

7 Aroclor 1254 CAS #: 11097-69-1

Peaks not detected for Quant. or Qual. signal(s).

---

8 Aroclor 1260 CAS #: 11096-82-5

Peaks not detected for Quant. or Qual. signal(s).

$ 12 Decachlorobiphenyl CAS #: 2051-24-3

8.570	8.571	-0.001	9095946	0.08302	42			

Data File: \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318006.D

Date : 13-MAR-2018 14:50

Client ID: MB

Instrument: 40GCSJ.i

Sample Info: 1655661,

Volume Injected (uL): 1.0

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031318.b/031318006.D (ManInt: BIOTA.sub Orig: BIOTA.sub)
Injection Date: 13-MAR-2018 14:50 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1655661

NO SIGNAL MANUAL INTEGRATIONS DONE FOR THIS DATA FILE

Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	0	0	
Aroclor 1016	0	0	
Aroclor 1016	0	0	
Aroclor 1016	0	0	
Aroclor 1016	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1232	0	0	
Aroclor 1232	0	0	
Aroclor 1232	0	0	
Aroclor 1232	0	0	
Aroclor 1232	0	0	
Aroclor 1242	0	0	
Aroclor 1242	0	0	
Aroclor 1242	0	0	
Aroclor 1242	0	0	
Aroclor 1242	0	0	
Aroclor 1248	0	0	
Aroclor 1248	0	0	
Aroclor 1248	0	0	
Aroclor 1248	0	0	
Aroclor 1248	0	0	
Aroclor 1254	0	0	
Aroclor 1254	0	0	
Aroclor 1254	0	0	
Aroclor 1254	0	0	
Aroclor 1254	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Tetrachloro-m-xylene	8963758	8963758	
Decachlorobiphenyl	9095946	9095946	

SAMPLE NO.

LCS

PCB - FORM I SVOA-1

SEMI-VOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay

Contract: HUDSON RIVER REMEDIAL ACTION

Date Received:

Matrix: Tissue SDG No.: 40151601

Date Extracted: 03/12/2018 07:42

Lab Sample ID: 1655662

Date Analyzed: 03/13/2018 15:05

Lab File ID: 031318.B\031318007.D

Initial wt/vol: 10 g Final wt/vol: 5 mL Dilution: 1

Instrument: 40GCSJ Percent Moisture:

CAS NO.	COMPOUND	CONCENTRATION UNITS: ug/kg	Q
12674-11-2	PCB-1016 (Aroclor 1016)	<11.4	U
11104-28-2	PCB-1221 (Aroclor 1221)	<11.4	U
11141-16-5	PCB-1232 (Aroclor 1232)	<11.4	U
53469-21-9	PCB-1242 (Aroclor 1242)	212	
12672-29-6	PCB-1248 (Aroclor 1248)	<11.4	U
11097-69-1	PCB-1254 (Aroclor 1254)	<11.4	U
11096-82-5	PCB-1260 (Aroclor 1260)	<11.4	U

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318007.D
Lab Smp Id: 1655662 Client Smp ID: MBLCS
Inj Date : 13-MAR-2018 15:05
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1655662,
Misc Info : 17521,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031318.b\pcba030618.m
Meth Date : 27-Mar-2018 10:38 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 7 QC Sample: LCS
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.000	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.185	19401	16917	0.872	0.04		
1.426	12365	11403	0.922	0.03		
1.535	14196	8107	0.571	0.03		
1.805	22348	12680	0.567	0.05		
2.406	108102	82068	0.759	0.27		
2.485	8493218	8075538	0.951	21.27	$ 2 Tetrachloro-m-xylene	
2.687	156314	128161	0.820	0.39		
2.790	183945	177565	0.965	0.46		
2.850	820415	762840	0.930	2.05		
3.054	113141	102677	0.908	0.28		
3.275	21779	14710	0.675	0.05		
3.317	1106729	1033392	0.934	2.77	3 Aroclor 1242	BA
3.343	620375	583365	0.940	1.55		
3.434	177645	159986	0.901	0.44		
3.513	1086406	784656	0.722	2.72	3 Aroclor 1242	BA
3.646	17193	12918	0.751	0.04		
3.694	30109	30818	1.024	0.07		
3.731	280259	275971	0.985	0.70		
3.755	174390	169690	0.973	0.43		
3.826	2990868	1735712	0.580	7.49	3 Aroclor 1242	BA
3.929	1287186	1091790	0.848	3.22	3 Aroclor 1242	BA
3.994	840859	818462	0.973	2.10	3 Aroclor 1242	BA
4.058	248978	228753	0.919	0.62		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.136	101087	89834	0.889	0.25	
4.215	734527	706049	0.961	1.84	
4.261	692196	669619	0.967	1.73	
4.293	284423	292508	1.028	0.71	
4.310	336024	326879	0.973	0.84	
4.369	27879	19257	0.691	0.06	
4.417	973317	924199	0.950	2.43	
4.451	692398	528492	0.763	1.73	
4.545	1197332	1099274	0.918	2.99	
4.617	280846	261679	0.932	0.70	
4.679	22164	17635	0.796	0.05	
4.727	59276	51858	0.875	0.14	
4.769	64675	61097	0.945	0.16	
4.812	528392	515934	0.976	1.32	
4.848	949085	915745	0.965	2.37	
4.878	1193294	906286	0.759	2.98	
4.957	131010	108965	0.832	0.32	
5.037	966431	919835	0.952	2.42	
5.100	217315	161359	0.743	0.54	
5.161	243995	218214	0.894	0.61	
5.213	181001	161325	0.891	0.45	
5.286	16035	14127	0.881	0.04	
5.309	44130	37416	0.848	0.11	
5.362	161681	132524	0.820	0.40	
5.407	259141	221182	0.854	0.64	
5.445	156892	142430	0.908	0.39	
5.509	412088	322351	0.782	1.03	
5.604	118456	108941	0.920	0.29	
5.673	14672	12533	0.854	0.03	
5.715	30106	21614	0.718	0.07	
5.738	40119	26533	0.661	0.10	
5.792	259953	214280	0.824	0.65	
5.890	39989	26208	0.655	0.10	
6.023	242894	192155	0.791	0.60	
6.298	51661	45860	0.888	0.12	
6.535	16238	13492	0.831	0.04	
6.714	13429	10744	0.800	0.03	
6.801	12202	11321	0.928	0.03	
7.003	22046	16804	0.762	0.05	
7.123	24227	13784	0.569	0.06	
7.393	12076	9301	0.770	0.03	
7.686	11849	10184	0.859	0.02	
7.727	44027	38858	0.883	0.11	
7.776	9870	9889	1.002	0.02	
7.800	32693	31016	0.949	0.08	
8.274	82909	74720	0.901	0.20	
8.375	11899	10576	0.889	0.02	
8.570	9013608	8120658	0.901	22.90	$ 12 Decachlorobiphenyl
8.942	20551	11149	0.543	0.05	
9.467	12006	9397	0.783	0.03	
9.963	21447	12323	0.575	0.05	
	39911776	35196592		100.000	

Total unknown % area = 38

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318007.D
Lab Smp Id: 1655662 Client Smp ID: MBLCS
Inj Date : 13-MAR-2018 15:05
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1655662,
Misc Info : 17521,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031318.b\pcba030618.m
Meth Date : 27-Mar-2018 10:38 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 7 QC Sample: LCS
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.000	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221					CAS #: 11104-28-2			
Operator disabled compound identification.								
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.484	2.484	0.000	8493218	0.06889	34			
3 Aroclor 1242					CAS #: 53469-21-9			
3.316	3.318	-0.002	1106729	0.43689	220	0.00- 20.00	100.00(M)	BA
3.513	3.513	0.000	1086405	0.43111	220	0.00- 20.00	98.16	BA
3.825	3.827	-0.002	2990867	0.42027	210	0.00- 20.00	270.24	BA
3.928	3.929	-0.001	1287185	0.42320	210	0.00- 20.00	116.31	BA
3.993	3.995	-0.002	840859	0.41063	200	0.00- 20.00	75.98	BA
Average of Peak Concentrations =					210			

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE

4 Aroclor 1232 CAS #: 1114-16-5

Operator disabled compound identification.

---

5 Aroclor 1016 CAS #: 12674-11-2

Operator disabled compound identification.

---

6 Aroclor 1248 CAS #: 12672-29-6

Operator disabled compound identification.

---

7 Aroclor 1254 CAS #: 11097-69-1

Operator disabled compound identification.

---

8 Aroclor 1260 CAS #: 11096-82-5

Peaks not detected for Quant. or Qual. signal(s).

$ 12 Decachlorobiphenyl CAS #: 2051-24-3

8.569	8.571	-0.002	9013607	0.08227	41	

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\031318.b\031318007.D

Date : 13-MAR-2018 15:05

Client ID: MBLCS

Instrument: 40GCSJ.i

Sample Info: 1655662,

Volume Injected (uL): 1.0

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031318.b/031318007.D (ManInt: BIOTA.sub Orig: BIOTA.sub)
Injection Date: 13-MAR-2018 15:05 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1655662





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	1101039	0	
Aroclor 1016	1101039	0	
Aroclor 1016	1101039	0	
Aroclor 1016	1101039	0	
Aroclor 1016	1101039	0	
Aroclor 1221	22347	0	
Aroclor 1221	22347	0	
Aroclor 1221	22347	0	
Aroclor 1221	22347	0	
Aroclor 1221	22347	0	
Aroclor 1232	820414	0	
Aroclor 1232	820414	0	
Aroclor 1232	820414	0	
Aroclor 1232	820414	0	

Aroclor 1232	820414	0	
Aroclor 1242	1101039	1106729	BA
Aroclor 1242	1101039	1106729	BA
Aroclor 1242	1101039	1106729	BA
Aroclor 1242	1101039	1106729	BA
Aroclor 1242	1101039	1106729	BA
Aroclor 1248	686690	0	
Aroclor 1248	686690	0	
Aroclor 1248	686690	0	
Aroclor 1248	686690	0	
Aroclor 1248	686690	0	
Aroclor 1254	235448	0	
Aroclor 1254	235448	0	
Aroclor 1254	235448	0	
Aroclor 1254	235448	0	
Aroclor 1254	235448	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Tetrachloro-m-xylene	8493218	8493218	
Decachlorobiphenyl	9013607	9013607	

==========

SAMPLE NO.

BLANK

PCB - FORM I SVOA-1

SEMI-VOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay

Contract: HUDSON RIVER REMEDIAL ACTION

Date Received:

Matrix: Tissue SDG No.: 40151601

Date Extracted: 03/13/2018 07:38

Lab Sample ID: 1658254

Date Analyzed: 03/14/2018 13:44

Lab File ID: 031418.B\031418006.D

Initial wt/vol: 10 g Final wt/vol: 5 mL Dilution: 1

Instrument: 40GCSJ Percent Moisture:

CAS NO.	COMPOUND	CONCENTRATION UNITS: ug/kg	Q
12674-11-2	PCB-1016 (Aroclor 1016)	<11.4	U
11104-28-2	PCB-1221 (Aroclor 1221)	<11.4	U
11141-16-5	PCB-1232 (Aroclor 1232)	<11.4	U
53469-21-9	PCB-1242 (Aroclor 1242)	<11.4	U
12672-29-6	PCB-1248 (Aroclor 1248)	<11.4	U
11097-69-1	PCB-1254 (Aroclor 1254)	<11.4	U
11096-82-5	PCB-1260 (Aroclor 1260)	<11.4	U

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418006.D
Lab Smp Id: 1658254 Client Smp ID: MB
Inj Date : 14-MAR-2018 13:44
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1658254,
Misc Info : 17523,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 6 QC Sample: BLANK
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.000	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.184	9308	9168	0.985	0.04		
1.426	13229	9984	0.755	0.06		
1.538	10968	9927	0.905	0.05		
2.485	9322527	9158044	0.982	49.40	$ 2 Tetrachloro-m-xylene	
2.787	19842	16132	0.813	0.10		
2.854	14749	10951	0.743	0.07		
4.049	11412	11989	1.051	0.06		
4.239	12234	11659	0.953	0.06		
4.420	10103	7444	0.737	0.05		
5.506	12536	9405	0.750	0.06		
5.792	14830	11542	0.778	0.07		
6.056	19989	13624	0.682	0.10		
6.300	14777	12555	0.850	0.07		
7.005	11590	9464	0.817	0.06		
7.125	20450	13890	0.679	0.10		
7.727	42265	36227	0.857	0.22		
7.803	31627	26512	0.838	0.16		
8.276	77820	71472	0.918	0.41		
8.382	22898	15893	0.694	0.12		
8.572	9139833	7852094	0.859	48.32	$ 12 Decachlorobiphenyl	
8.938	15536	12441	0.801	0.08		
9.469	31302	19141	0.611	0.16		
9.964	35242	23488	0.666	0.18		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
	18915068	17373046		100.000	

Total unknown % area = 2.3

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418006.D
Lab Smp Id: 1658254 Client Smp ID: MB
Inj Date : 14-MAR-2018 13:44
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1658254,
Misc Info : 17523,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 6 QC Sample: BLANK
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.000	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE

1 Aroclor 1221 CAS #: 11104-28-2

Peaks not detected for Quant. or Qual. signal(s).

$ 2 Tetrachloro-m-xylene CAS #: 877-09-8

2.484	2.484	0.000	9322527	0.07561	38			

---

3 Aroclor 1242 CAS #: 53469-21-9

Peaks not detected for Quant. or Qual. signal(s).

---

4 Aroclor 1232 CAS #: 1114-16-5

Peaks not detected for Quant. or Qual. signal(s).

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE

5 Aroclor 1016 CAS #: 12674-11-2

Peaks not detected for Quant. or Qual. signal(s).

---

6 Aroclor 1248 CAS #: 12672-29-6

Peaks not detected for Quant. or Qual. signal(s).

---

7 Aroclor 1254 CAS #: 11097-69-1

Peaks not detected for Quant. or Qual. signal(s).

---

8 Aroclor 1260 CAS #: 11096-82-5

Peaks not detected for Quant. or Qual. signal(s).

$ 12 Decachlorobiphenyl CAS #: 2051-24-3

8.572	8.571	0.001	9139833	0.08342		42		

Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418006.D

Date : 14-MAR-2018 13:44

Client ID: MB

Instrument: 40GCSJ.i

Sample Info: 1658254,

Volume Injected (uL): 1.0

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b/031418006.D (ManInt: BIOTA.sub Orig: BIOTA.sub)
Injection Date: 14-MAR-2018 13:44 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1658254

NO SIGNAL MANUAL INTEGRATIONS DONE FOR THIS DATA FILE

Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	0	0	
Aroclor 1016	0	0	
Aroclor 1016	0	0	
Aroclor 1016	0	0	
Aroclor 1016	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1232	0	0	
Aroclor 1232	0	0	
Aroclor 1232	0	0	
Aroclor 1232	0	0	
Aroclor 1232	0	0	
Aroclor 1242	0	0	
Aroclor 1242	0	0	
Aroclor 1242	0	0	
Aroclor 1242	0	0	
Aroclor 1242	0	0	
Aroclor 1248	0	0	
Aroclor 1248	0	0	
Aroclor 1248	0	0	
Aroclor 1248	0	0	
Aroclor 1248	0	0	
Aroclor 1254	0	0	
Aroclor 1254	0	0	
Aroclor 1254	0	0	
Aroclor 1254	0	0	
Aroclor 1254	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Tetrachloro-m-xylene	9322527	9322527	
Decachlorobiphenyl	9139833	9139833	

SAMPLE NO.

LCS

PCB - FORM I SVOA-1

SEMI-VOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay

Contract: HUDSON RIVER REMEDIAL ACTION

Date Received:

Matrix: Tissue SDG No.: 40151601

Date Extracted: 03/13/2018 07:38

Lab Sample ID: 1658255

Date Analyzed: 03/14/2018 14:00

Lab File ID: 031418.B\031418007.D

Initial wt/vol: 10 g Final wt/vol: 5 mL Dilution: 1

Instrument: 40GCSJ Percent Moisture:

CAS NO.	COMPOUND	CONCENTRATION UNITS: ug/kg	Q
12674-11-2	PCB-1016 (Aroclor 1016)	<11.4	U
11104-28-2	PCB-1221 (Aroclor 1221)	<11.4	U
11141-16-5	PCB-1232 (Aroclor 1232)	<11.4	U
53469-21-9	PCB-1242 (Aroclor 1242)	232	
12672-29-6	PCB-1248 (Aroclor 1248)	<11.4	U
11097-69-1	PCB-1254 (Aroclor 1254)	<11.4	U
11096-82-5	PCB-1260 (Aroclor 1260)	<11.4	U

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418007.D
Lab Smp Id: 1658255 Client Smp ID: MBLCS
Inj Date : 14-MAR-2018 14:00
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1658255,
Misc Info : 17523,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 7 QC Sample: LCS
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.000	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.767	9946	7433	0.747	0.02		
1.803	17693	14637	0.827	0.03		
2.404	111911	87457	0.781	0.24		
2.485	10488341	10397400	0.991	23.61	$ 2 Tetrachloro-m-xylene	
2.687	170402	142152	0.834	0.37		
2.790	209653	195891	0.934	0.46		
2.850	926104	857106	0.925	2.05		
3.055	115628	109822	0.950	0.25		
3.103	10592	6907	0.652	0.02		
3.277	23783	17688	0.744	0.05		
3.317	1181695	1113787	0.943	2.61	3 Aroclor 1242	BA
3.344	665653	631762	0.949	1.47		
3.434	186207	172651	0.927	0.41		
3.514	1157603	831922	0.719	2.56	3 Aroclor 1242	BA
3.697	34912	32671	0.936	0.07		
3.732	304043	304608	1.002	0.67		
3.756	196485	192148	0.978	0.43		
3.826	3337551	1956979	0.586	7.39	3 Aroclor 1242	BA
3.929	1423369	1184695	0.832	3.15	3 Aroclor 1242	BA
3.994	935105	919454	0.983	2.07	3 Aroclor 1242	BA
4.058	258971	236457	0.913	0.57		
4.136	103923	92474	0.890	0.23		
4.214	791632	760649	0.961	1.75		

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.261	744052	738740	0.993	1.64	
4.293	303425	314392	1.036	0.67	
4.310	371058	337081	0.908	0.82	
4.369	26503	20717	0.782	0.05	
4.418	1046039	987604	0.944	2.31	
4.452	778452	577300	0.742	1.72	
4.546	1293913	1203273	0.930	2.86	
4.590	19786	18103	0.915	0.04	
4.617	296490	272037	0.918	0.65	
4.678	23183	16757	0.723	0.05	
4.728	66128	56825	0.859	0.14	
4.768	73265	66800	0.912	0.16	
4.813	604724	580382	0.960	1.33	
4.848	1064518	1039421	0.976	2.35	
4.878	1349776	1024261	0.759	2.99	
4.958	145182	115220	0.794	0.32	
5.037	1106774	1036097	0.936	2.45	
5.099	229188	178256	0.778	0.50	
5.163	263878	238330	0.903	0.58	
5.211	199502	171553	0.860	0.44	
5.309	48069	37998	0.790	0.10	
5.363	171461	150475	0.878	0.37	
5.408	287689	240408	0.836	0.63	
5.446	171215	152690	0.892	0.37	
5.511	460004	350202	0.761	1.01	
5.605	139233	117902	0.847	0.30	
5.675	15012	13326	0.888	0.03	
5.717	31085	23808	0.766	0.06	
5.740	34830	32214	0.925	0.07	
5.793	295008	248821	0.843	0.65	
5.889	43717	30083	0.688	0.09	
6.026	284789	216969	0.762	0.63	
6.171	30416	17725	0.583	0.06	
6.300	65116	55728	0.856	0.14	
6.334	16324	13350	0.818	0.03	
6.379	22429	12929	0.576	0.04	
6.512	13440	13494	1.004	0.02	
6.536	22696	17123	0.754	0.05	
6.801	17773	13249	0.745	0.03	
7.007	18694	16572	0.886	0.04	
7.124	13289	12322	0.927	0.02	
7.691	14721	9396	0.638	0.03	
7.729	50729	41436	0.817	0.11	
7.799	44103	29773	0.675	0.09	
8.276	89292	77692	0.870	0.19	
8.378	14689	13560	0.923	0.03	
8.572	10029384	8558329	0.853	22.21	$ 12 Decachlorobiphenyl
8.943	11416	8249	0.723	0.02	
9.961	13880	12395	0.893	0.03	
	45137541	39798117		100.000	

Total unknown % area = 36

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418007.D
Lab Smp Id: 1658255 Client Smp ID: MBLCS
Inj Date : 14-MAR-2018 14:00
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1658255,
Misc Info : 17523,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\031418.b\pcba030618.m
Meth Date : 02-Apr-2018 11:46 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 7 QC Sample: LCS
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.000	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221					CAS #: 11104-28-2			
Operator disabled compound identification.								
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.485	2.484	0.001	10488341	0.08507	42			
3 Aroclor 1242					CAS #: 53469-21-9			
3.317	3.318	-0.001	1181695	0.46648	230	0.00- 20.00	100.00(M)	BA
3.514	3.513	0.001	1157603	0.45936	230	0.00- 20.00	97.96	BA
3.826	3.827	-0.001	3337551	0.46899	230	0.00- 20.00	282.44	BA
3.928	3.929	-0.001	1423368	0.46798	230	0.00- 20.00	120.45	BA
3.994	3.995	-0.001	935105	0.45665	230	0.00- 20.00	79.13	BA
Average of Peak Concentrations =					230			

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE

4 Aroclor 1232 CAS #: 1114-16-5

Operator disabled compound identification.

---

5 Aroclor 1016 CAS #: 12674-11-2

Operator disabled compound identification.

---

6 Aroclor 1248 CAS #: 12672-29-6

Operator disabled compound identification.

---

7 Aroclor 1254 CAS #: 11097-69-1

Operator disabled compound identification.

---

8 Aroclor 1260 CAS #: 11096-82-5

Peaks not detected for Quant. or Qual. signal(s).

$ 12 Decachlorobiphenyl CAS #: 2051-24-3

8.572	8.571	0.001	10029383	0.09154	46	

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b\031418007.D

Date : 14-MAR-2018 14:00

Client ID: MBLCS

Instrument: 40GCSJ.i

Sample Info: 1658255,

Volume Injected (uL): 1.0

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\031418.b/031418007.D (ManInt: BIOTA.sub Orig: BIOTA.sub)
Injection Date: 14-MAR-2018 14:00 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1658255





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	1177781	0	
Aroclor 1016	1177781	0	
Aroclor 1016	1177781	0	
Aroclor 1016	1177781	0	
Aroclor 1016	1177781	0	
Aroclor 1221	17693	0	
Aroclor 1221	17693	0	
Aroclor 1221	17693	0	
Aroclor 1221	17693	0	
Aroclor 1221	17693	0	
Aroclor 1232	918724	0	
Aroclor 1232	918724	0	
Aroclor 1232	918724	0	
Aroclor 1232	918724	0	

Aroclor 1232	918724	0	
Aroclor 1242	1177781	1181695	BA
Aroclor 1242	1177781	1181695	BA
Aroclor 1242	1177781	1181695	BA
Aroclor 1242	1177781	1181695	BA
Aroclor 1242	1177781	1181695	BA
Aroclor 1248	737221	0	
Aroclor 1248	737221	0	
Aroclor 1248	737221	0	
Aroclor 1248	737221	0	
Aroclor 1248	737221	0	
Aroclor 1254	250152	0	
Aroclor 1254	250152	0	
Aroclor 1254	250152	0	
Aroclor 1254	250152	0	
Aroclor 1254	250152	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Tetrachloro-m-xylene	10488341	10488341	
Decachlorobiphenyl	10029383	10029383	

==========

SAMPLE NO.

BLANK

PCB - FORM I SVOA-1

SEMI-VOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay

Contract: HUDSON RIVER REMEDIAL ACTION

Date Received:

Matrix: Tissue SDG No.: 40151601

Date Extracted: 03/28/2018 08:09

Lab Sample ID: 1664793

Date Analyzed: 03/30/2018 17:22

Lab File ID: 033018.B\033018010.D

Initial wt/vol: 10 g Final wt/vol: 5 mL Dilution: 1

Instrument: 40GCSJ Percent Moisture:

CAS NO.	COMPOUND	CONCENTRATION UNITS: ug/kg	Q
12674-11-2	PCB-1016 (Aroclor 1016)	<11.4	U
11104-28-2	PCB-1221 (Aroclor 1221)	<11.4	U
11141-16-5	PCB-1232 (Aroclor 1232)	<11.4	U
53469-21-9	PCB-1242 (Aroclor 1242)	<11.4	U
12672-29-6	PCB-1248 (Aroclor 1248)	<11.4	U
11097-69-1	PCB-1254 (Aroclor 1254)	<11.4	U
11096-82-5	PCB-1260 (Aroclor 1260)	<11.4	U

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018010.D
Lab Smp Id: 1664793 Client Smp ID: MB
Inj Date : 30-MAR-2018 17:22
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1664793,
Misc Info : 17591,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\033018.b\pcba030618.m
Meth Date : 05-Apr-2018 14:48 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 10 QC Sample: BLANK
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.000	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.184	14570	13384	0.919	0.07		
1.539	10323	9103	0.882	0.05		
1.688	11405	8619	0.756	0.05		
2.199	61443	30725	0.500	0.31		
2.485	9345991	8986785	0.962	48.06	$ 2 Tetrachloro-m-xylene	
2.788	19349	15309	0.791	0.09		
7.125	21633	13203	0.610	0.11		
7.725	44824	39078	0.872	0.23		
7.800	29631	28981	0.978	0.15		
8.274	87961	73547	0.836	0.45		
8.378	20602	13224	0.642	0.10		
8.570	9756823	8203778	0.841	50.24	$ 12 Decachlorobiphenyl	
8.938	10515	8884	0.845	0.05		
9.958	9403	7914	0.842	0.04		
	19444473	17452534		100.000		

Total unknown % area = 1.7

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018010.D
Lab Smp Id: 1664793 Client Smp ID: MB
Inj Date : 30-MAR-2018 17:22
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1664793,
Misc Info : 17591,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\033018.b\pcba030618.m
Meth Date : 05-Apr-2018 14:48 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 10 QC Sample: BLANK
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.000	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221						CAS #: 11104-28-2		
Peaks not detected for Quant. or Qual. signal(s).								
$ 2 Tetrachloro-m-xylene						CAS #: 877-09-8		
2.485	2.484	0.001	9345991	0.07580	38			
3 Aroclor 1242						CAS #: 53469-21-9		
Peaks not detected for Quant. or Qual. signal(s).								
4 Aroclor 1232						CAS #: 1114-16-5		
Peaks not detected for Quant. or Qual. signal(s).								

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE

5 Aroclor 1016 CAS #: 12674-11-2

Peaks not detected for Quant. or Qual. signal(s).

---

6 Aroclor 1248 CAS #: 12672-29-6

Peaks not detected for Quant. or Qual. signal(s).

---

7 Aroclor 1254 CAS #: 11097-69-1

Peaks not detected for Quant. or Qual. signal(s).

---

8 Aroclor 1260 CAS #: 11096-82-5

Peaks not detected for Quant. or Qual. signal(s).

$ 12 Decachlorobiphenyl CAS #: 2051-24-3

8.569	8.571	-0.002	9756823	0.08905		44		

Data File: \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018010.D
Date : 30-MAR-2018 17:22
Client ID: MB
Instrument: 40GCSJ.i
Sample Info: 1664793,
Volume Injected (uL): 1.0
Operator: BLM
Column phase: ZB-1ms
Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\033018.b/033018010.D (ManInt: BIOTA.sub Orig: BIOTA.sub)
Injection Date: 30-MAR-2018 17:22 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1664793

NO SIGNAL MANUAL INTEGRATIONS DONE FOR THIS DATA FILE

Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	0	0	
Aroclor 1016	0	0	
Aroclor 1016	0	0	
Aroclor 1016	0	0	
Aroclor 1016	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1232	0	0	
Aroclor 1232	0	0	
Aroclor 1232	0	0	
Aroclor 1232	0	0	
Aroclor 1232	0	0	
Aroclor 1242	0	0	
Aroclor 1242	0	0	
Aroclor 1242	0	0	
Aroclor 1242	0	0	
Aroclor 1242	0	0	
Aroclor 1248	0	0	
Aroclor 1248	0	0	
Aroclor 1248	0	0	
Aroclor 1248	0	0	
Aroclor 1248	0	0	
Aroclor 1254	0	0	
Aroclor 1254	0	0	
Aroclor 1254	0	0	
Aroclor 1254	0	0	
Aroclor 1254	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Aroclor 1260	0	0	
Tetrachloro-m-xylene	9345991	9345991	
Decachlorobiphenyl	9756823	9756823	

SAMPLE NO.

LCS

PCB - FORM I SVOA-1

SEMI-VOLATILE ORGANICS ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay

Contract: HUDSON RIVER REMEDIAL ACTION

Date Received:

Matrix: Tissue SDG No.: 40151601

Date Extracted: 03/28/2018 08:09

Lab Sample ID: 1664794

Date Analyzed: 03/30/2018 17:38

Lab File ID: 033018.B\033018011.D

Initial wt/vol: 10 g Final wt/vol: 5 mL Dilution: 1

Instrument: 40GCSJ Percent Moisture:

CAS NO.	COMPOUND	CONCENTRATION UNITS: ug/kg	Q
12674-11-2	PCB-1016 (Aroclor 1016)	<11.4	U
11104-28-2	PCB-1221 (Aroclor 1221)	<11.4	U
11141-16-5	PCB-1232 (Aroclor 1232)	<11.4	U
53469-21-9	PCB-1242 (Aroclor 1242)	236	
12672-29-6	PCB-1248 (Aroclor 1248)	<11.4	U
11097-69-1	PCB-1254 (Aroclor 1254)	<11.4	U
11096-82-5	PCB-1260 (Aroclor 1260)	<11.4	U

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018011.D
Lab Smp Id: 1664794 Client Smp ID: MBLCS
Inj Date : 30-MAR-2018 17:38
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1664794,
Misc Info : 17591,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\033018.b\pcba030618.m
Meth Date : 05-Apr-2018 14:48 40GCSJ.i Quant Type: AREA%
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 11 QC Sample: LCS
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.000	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS	REVIEW CODE
1.185	14371	12627	0.879	0.03		
1.423	10530	7685	0.730	0.02		
1.536	13414	7924	0.591	0.02		
1.686	11381	10053	0.883	0.02		
2.198	54145	32734	0.605	0.11		
2.405	112460	86945	0.773	0.24		
2.484	10031155	9542303	0.951	22.16	$ 2 Tetrachloro-m-xylene	
2.687	169882	137424	0.809	0.37		
2.789	206912	194056	0.938	0.45		
2.850	916005	843455	0.921	2.02		
3.054	109450	106780	0.976	0.24		
3.276	24277	17988	0.741	0.05		
3.318	1194942	1102529	0.923	2.63	3 Aroclor 1242	BA
3.344	673074	628279	0.933	1.48		
3.433	189004	166505	0.881	0.41		
3.513	1179824	838056	0.710	2.60	3 Aroclor 1242	BA
3.647	20606	14963	0.726	0.04		
3.696	33251	32704	0.984	0.07		
3.731	305888	299746	0.980	0.67		
3.756	201569	189205	0.939	0.44		
3.827	3390735	1931255	0.570	7.49	3 Aroclor 1242	BA
3.928	1455755	1202014	0.826	3.21	3 Aroclor 1242	BA
3.993	960656	896360	0.933	2.12	3 Aroclor 1242	BA

RT	AREA	HEIGHT	HT/AREA	% AREA	COMPOUNDS
4.057	264291	236277	0.894	0.58	
4.136	107261	95815	0.893	0.23	
4.214	794554	739961	0.931	1.75	
4.260	755016	715461	0.948	1.66	
4.292	310234	311821	1.005	0.68	
4.309	379143	345005	0.910	0.83	
4.370	23790	19392	0.815	0.05	
4.417	1056407	979658	0.927	2.33	
4.451	808986	599820	0.741	1.78	
4.545	1342685	1194659	0.890	2.96	
4.617	306653	277666	0.905	0.67	
4.681	21743	18158	0.835	0.04	
4.727	68527	54682	0.798	0.15	
4.767	71956	65521	0.911	0.15	
4.812	606613	573568	0.946	1.34	
4.847	1080460	1003008	0.928	2.38	
4.878	1382549	1012969	0.733	3.05	
4.957	157241	119163	0.758	0.34	
5.037	1141778	1008137	0.883	2.52	
5.099	236608	175642	0.742	0.52	
5.162	271430	237736	0.876	0.59	
5.212	196259	173485	0.884	0.43	
5.284	17550	15371	0.876	0.03	
5.308	48879	38335	0.784	0.10	
5.362	170525	144062	0.845	0.37	
5.406	294036	234927	0.799	0.64	
5.446	178715	155302	0.869	0.39	
5.509	476178	345620	0.726	1.05	
5.605	140916	119367	0.847	0.31	
5.674	15337	14250	0.929	0.03	
5.716	27794	21374	0.769	0.06	
5.741	51456	31217	0.607	0.11	
5.792	299958	231188	0.771	0.66	
5.888	48680	27142	0.558	0.10	
6.023	280916	206418	0.735	0.62	
6.167	30563	17934	0.587	0.06	
6.299	63446	50975	0.803	0.14	8 Aroclor 1260
6.334	16217	13312	0.821	0.03	
6.511	9835	10746	1.093	0.02	
6.535	25592	16656	0.651	0.05	
6.805	14085	11480	0.815	0.03	
7.005	20456	15271	0.747	0.04	8 Aroclor 1260
7.124	18063	14848	0.822	0.03	
7.245	11908	10334	0.868	0.02	8 Aroclor 1260
7.444	13524	9088	0.672	0.02	8 Aroclor 1260
7.727	48321	42048	0.870	0.10	
7.801	31610	31396	0.993	0.06	
7.899	9566	7815	0.817	0.02	8 Aroclor 1260
8.275	89863	76286	0.849	0.19	
8.378	17149	13847	0.807	0.03	
8.570	10109187	8088140	0.800	22.73	$ 12 Decachlorobiphenyl
9.462	11459	9498	0.829	0.02	
9.956	11260	11145	0.990	0.02	
	45266512	38294586		100.000	

Total unknown % area = 37

Pace Analytical Services, Inc.

SW846-8082

Data file : \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018011.D
Lab Smp Id: 1664794 Client Smp ID: MBLCS
Inj Date : 30-MAR-2018 17:38
Operator : BLM Inst ID: 40GCSJ.i
Smp Info : 1664794,
Misc Info : 17591,BIOTA
Comment :
Method : \\40wintarget\data2\chem\40GCSJ.i\033018.b\pcba030618.m
Meth Date : 05-Apr-2018 14:48 40GCSJ.i Quant Type: ESTD
Cal Date : 06-MAR-2018 19:39 Cal File: 030618066.D
Als bottle: 11 QC Sample: LCS
Dil Factor: 1.00000
Integrator: Falcon Compound Sublist: BIOTA.sub
Target Version: RC10A Sample Matrix: Biota

Concentration Formula: Amt * DF * Uf * Vt/(Vi * Ws * (100 - M)/100) * CpndV

Name	Value	Description
DF	1.000	Dilution Factor
Uf	1.000	Correction Factor
Vt	5000.000	Volume of final extract (uL)(1000 low, 2
Vi	1.000	Volume injected (uL)
Ws	10.000	Weight of sample extracted
M	0.00000	% Moisture
Cpnd Variable		Local Compound Variable

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE
1 Aroclor 1221					CAS #: 11104-28-2			
Operator disabled compound identification.								
$ 2 Tetrachloro-m-xylene					CAS #: 877-09-8			
2.484	2.484	0.000	10031154	0.08136	41			
3 Aroclor 1242					CAS #: 53469-21-9			
3.317	3.318	-0.001	1194941	0.47171	240	0.00- 20.00	100.00(M)	BA
3.513	3.513	0.000	1179824	0.46818	230	0.00- 20.00	98.73	BA
3.826	3.827	-0.001	3390734	0.47646	240	0.00- 20.00	283.76	BA
3.928	3.929	-0.001	1455755	0.47863	240	0.00- 20.00	121.83	BA
3.993	3.995	-0.002	960656	0.46913	230	0.00- 20.00	80.39	BA
Average of Peak Concentrations =					240			

RT	EXP RT	DLT RT	RESPONSE	CONCENTRATIONS ON-COL (ug/ml)	FINAL (ug/Kg)	TARGET RANGE	RATIO	REVIEW CODE

4 Aroclor 1232 CAS #: 1114-16-5

Operator disabled compound identification.

---

5 Aroclor 1016 CAS #: 12674-11-2

Operator disabled compound identification.

---

6 Aroclor 1248 CAS #: 12672-29-6

Operator disabled compound identification.

---

7 Aroclor 1254 CAS #: 11097-69-1

Operator disabled compound identification.

---

8 Aroclor 1260 CAS #: 11096-82-5

Qualifier signal(s) failed ratio test. BA

$ 12 Decachlorobiphenyl CAS #: 2051-24-3

8.570	8.571	-0.001	10109187	0.09227		46		

QC Flag Legend

M - Compound response manually integrated.

Review Codes Legend

:
BA: Indicates that the baseline had to be adjusted correctly by the analyst.

Data File: \\40wintarget\data2\chem\40GCSJ.i\033018.b\033018011.D

Date : 30-MAR-2018 17:38

Client ID: MBLCS

Instrument: 40GCSJ.i

Sample Info: 1664794,

Volume Injected (uL): 1.0

Operator: BLM

Column phase: ZB-1ms

Column diameter: 0.18



Data File: \\40wintarget\data2\chem\40GCSJ.i\033018.b/033018011.D (ManInt: BIOTA.sub Orig: BIOTA.sub)
Injection Date: 30-MAR-2018 17:38 Operator: BLM
Instrument: 40GCSJ.i
Lab Sample ID: 1664794





Manually Integrated Compounds

Compound	Area (before)	Area (after)	Code
Aroclor 1016	1186083	0	
Aroclor 1016	1186083	0	
Aroclor 1016	1186083	0	
Aroclor 1016	1186083	0	
Aroclor 1016	1186083	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1221	0	0	
Aroclor 1232	916004	0	
Aroclor 1232	916004	0	
Aroclor 1232	916004	0	
Aroclor 1232	916004	0	

Aroclor 1232	916004	0	
Aroclor 1242	1186083	1194941	BA
Aroclor 1242	1186083	1194941	BA
Aroclor 1242	1186083	1194941	BA
Aroclor 1242	1186083	1194941	BA
Aroclor 1242	1186083	1194941	BA
Aroclor 1248	753062	0	
Aroclor 1248	753062	0	
Aroclor 1248	753062	0	
Aroclor 1248	753062	0	
Aroclor 1248	753062	0	
Aroclor 1254	261395	0	
Aroclor 1254	261395	0	
Aroclor 1254	261395	0	
Aroclor 1254	261395	0	
Aroclor 1254	261395	0	
Aroclor 1260	0	0	BA
Aroclor 1260	0	0	BA
Aroclor 1260	0	0	BA
Aroclor 1260	0	0	BA
Aroclor 1260	0	0	BA
Tetrachloro-m-xylene	10031154	10031154	
Decachlorobiphenyl	10109187	10109187	

==========

Prep Log Report

Batch Information: OEXT 282598 PCB-B 2

Template Version: F-GB-O-113-Rev.03 (19Feb2018)

Prep Method	EPA 3541	Analysis Method	EPA 8082A Mod	Extracted By	AMC	Extracted Date/Time	03/12/2018 07:42:08:001
Spiked By	AMC	Spiked By Date/Time	03/12/2018 07:42:09:160	Witnessed By	ETB	Witnessed By Date/Time	03/12/2018 07:42:09:160
Instrument	40BALL	Conc. Temp #1	40	Conc. Temp #2	40	Methylene Chloride	None Added
Hexane	188917	Sodium Sulfate	188340	Florisil	186910	3620B Date/Initials	03/13/18 AMC
H2SO4 - Conc.	179360	Acid 3665A Date/Initials	03/13/18 AMC	Tuna	None Added	Reviewed By	BLM
Reviewed By Date	03/23/2018 13:52	Batch Notes	Hex/Ace 50:50-191186, Bacth was re-extracted under the same batch number.				

Sample Information:

QC Rule	Sample Type	Lab Sample ID	Initial Weight (g)	Final Volume (mL)	Volume Delivered (mL)	Sample Notes	8081-SS (mL)	AR1242-SPK (mL)
8082A TNYP	BLANK	**1655661**	10	5	3		189973 (.25)	
8082A TNYP	LCS	**1655662**	10	5	3		189973 (.25)	182336 (.125)
8082A TNYP	PS	**40150295019**	10.012	5	3		189973 (.25)	
8082A TNYP	MS	**1655663**	10.0094	5	3		189973 (.25)	182336 (.125)
8082A TNYP	DUP	**1655664**	10.0106	5	3		189973 (.25)	
8082A TNYP	PS	**40150295020**	9.9616	5	3		189973 (.25)	
8082A TNYP	PS	**40151597002**	10.0054	5	3		189973 (.25)	
8082A TNYP	PS	**40151597003**	9.9731	5	3		189973 (.25)	
8082A TNYP	PS	**40151601002**	9.9928	5	3		189973 (.25)	
8082A TNYP	PS	**40151608001**	9.9824	5	3		189973 (.25)	
8082A TNYP	PS	**40151608002**	10.0008	5	3		189973 (.25)	
8082A TNYP	PS	**40151608011**	10.0021	5	3		189973 (.25)	
8082A TNYP	PS	**40151800006**	10.0226	5	3		189973 (.25)	
8082A TNYP	PS	**40151800007**	10.0192	5	3		189973 (.25)	
8082A TNYP	PS	**40151800008**	9.98	5	3		189973 (.25)	

Prep Log Report

QC Rule	Sample Type	Lab Sample ID	Initial Weight (g)	Final Volume (mL)	Volume Delivered (mL)	Sample Notes	8081-SS (mL)	AR1242-SPK (mL)
8082A TNYP	PS	**40151800009**	10.0226	5	3		189973 (.25)	
8082A TNYP	PS	**40151800010**	9.9707	5	3		189973 (.25)	
8082A TNYP	PS	**40151804004**	10.0181	5	3		189973 (.25)	
8082A TNYP	PS	**40151804005**	9.9782	5	3		189973 (.25)	
8082A TNYP	PS	**40151804006**	9.9895	5	3		189973 (.25)	
8082A TNYP	RQS	**40151804007**	10.0333	5	3		189973 (.25)	
8082A TNYP	DUP	**1655665**	9.9824	5	3		189973 (.25)	
8082A TNYP	PS	**40151804008**	9.9931	5	3		189973 (.25)	

Standard Notes:

182336: Ar1242 Spike @20ug/mL 40GCS9. **189973: Pest/PCB Surrogate Spike 40GCSC.**

Prep Log Report

Batch Information: OEXT 283116 PCB-B

Template Version: F-GB-O-113-Rev.03 (19Feb2018)

Prep Method	EPA 3541	Analysis Method	EPA 8082A Mod	Extracted By	AMC	Extracted Date/Time	03/13/2018 07:38:05:819
Spiked By	AMC	Spiked By Date/Time	03/13/2018 07:38:06:514	Witnessed By	BLM	Witnessed By Date/Time	03/13/2018 07:38:06:514
Instrument	40BALL	Conc. Temp #1	40	Conc. Temp #2	40	Methylene Chloride	None Added
Hexane	188917	Sodium Sulfate	188340	Florisil	186910	3620B Date/Initials	3/14/18 AMC
H2SO4 - Conc.	179360	Acid 3665A Date/Initials	3/14/18 AMC	Tuna	None Added	Reviewed By	BLM
Reviewed By Date	03/23/2018 14:36	Batch Notes	Hex/Ace 50:50-191186				

Sample Information:

QC Rule	Sample Type	Lab Sample ID	Initial Weight (g)	Final Volume (mL)	Volume Delivered (mL)	Sample Notes	8081-SS (mL)	AR1242-SPK (mL)
8082A TNYP	BLANK	**1658254**	10	5	3		189973 (.25)	
8082A TNYP	LCS	**1658255**	10	5	3		189973 (.25)	182336 (.125)
8082A TNYP	PS	**40151549002**	9.986	5	3		189973 (.25)	
8082A TNYP	PS	**40151549003**	10.003	5	3		189973 (.25)	
8082A TNYP	PS	**40151549004**	10.0055	5	3		189973 (.25)	
8082A TNYP	PS	**40151549005**	10.0008	5	3		189973 (.25)	
8082A TNYP	PS	**40151549006**	10.0138	5	3		189973 (.25)	
8082A TNYP	PS	**40151549007**	10.0021	5	3		189973 (.25)	
8082A TNYP	PS	**40151549008**	10.0178	5	3		189973 (.25)	
8082A TNYP	PS	**40151549009**	10.0005	5	3		189973 (.25)	
8082A TNYP	RQS	**40151552010**	9.9677	5	3		189973 (.25)	
8082A TNYP	MS	**1658256**	9.9946	5	3		189973 (.25)	182336 (.125)
8082A TNYP	DUP	**1658257**	10.0089	5	3		189973 (.25)	
8082A TNYP	PS	**40151601003**	10.0319	5	3		189973 (.25)	
8082A TNYP	PS	**40151601004**	9.9968	5	3		189973 (.25)	
8082A TNYP	PS	**40151601005**	10.0156	5	3		189973 (.25)	

Prep Log Report

40151601

QC Rule	Sample Type	Lab Sample ID	Initial Weight (g)	Final Volume (mL)	Volume Delivered (mL)	Sample Notes	8081-SS (mL)	AR1242-SPK (mL)
8082A TNYP	PS	**40151601006**	9.9883	5	3		189973 (.25)	

Standard Notes:

182336: Ar1242 Spike @20ug/mL 40GCS9. **189973: Pest/PCB Surrogate Spike 40GCSC.**

Prep Log Report

Batch Information: OEXT 284454 PCB-B

Template Version: F-GB-O-113-Rev.05 (22Mar2018)

Prep Method	EPA 3541	Analysis Method	EPA 8082A Mod	Extracted By	AMC	Extracted Date/Time	03/28/2018 08:09:30:539
Spiked By	AMC	Spiked By Date/Time	03/28/2018 08:09:31:564	Witnessed By	ETB	Witnessed By Date/Time	03/28/2018 08:09:35:108
Instrument	40BALL	Conc. Temp #1	40	Conc. Temp #2	40	Methylene Chloride	None Added
Hexane/Acetone 1:1	192488	Hexane	188917	Sodium Sulfate	190953	Florisil	186910
3620B Date/Initials	03/29/18 ETB	H2SO4 - Conc.	179360	Acid 3665A Date/Initials	03/29/18 ETB	Tuna	None Added
Reviewed By	BLM	Reviewed By Date	04/06/2018 08:50	Batch Notes			

Sample Information:

QC Rule	Sample Type	Lab Sample ID	Initial Weight (g)	Final Volume (mL)	Volume Delivered (mL)	Sample Notes	8081-SS (mL)	AR1242-SPK (mL)
8082A TNYP	BLANK	**1664793**	10	5	3		189973 (.25)	
8082A TNYP	LCS	**1664794**	10	5	3		189973 (.25)	182336 (.125)
8082A TNYP	PS	**40151549001**	10.0051	5	3		189973 (.25)	
8082A TNYP	PS	**40151601001**	10.0134	5	3		189973 (.25)	
8082A TNYP	PS	**40151613001**	9.9785	5	3		189973 (.25)	
8082A TNYP	MS	**1664795**	9.9787	5	3		189973 (.25)	182336 (.125)
8082A TNYP	DUP	**1664796**	9.9864	5	3		189973 (.25)	
8082A TNYP	PS	**40151613002**	9.9976	5	3		189973 (.25)	
8082A TNYP	PS	**40151614002**	10.0031	5	3		189973 (.25)	
8082A TNYP	PS	**40151614003**	9.9906	5	3		189973 (.25)	
8082A TNYP	PS	**40151614004**	9.986	5	3		189973 (.25)	
8082A TNYP	PS	**40151613014**	10	5	3		189973 (.25)	
8082A TNYP	PS	**40151800005**	10.0085	5	3		189973 (.25)	
8082A TNYP	PS	**40151802001**	9.9815	5	3		189973 (.25)	
8082A TNYP	PS	**40151802002**	9.994	5	2.5	1*	189973 (.25)	
8082A TNYP	PS	**40151802003**	9.9604	5	3		189973 (.25)	

Prep Log Report

QC Rule	Sample Type	Lab Sample ID	Initial Weight (g)	Final Volume (mL)	Volume Delivered (mL)	Sample Notes	8081-SS (mL)	AR1242-SPK (mL)
8082A TNYP	PS	**40151802004**	10.0221	5	3		189973 (.25)	
8082A TNYP	PS	**40151802005**	9.9766	5	3		189973 (.25)	
8082A TNYP	PS	**40151805001**	10.0137	5	3		189973 (.25)	
8082A TNYP	PS	**40157825001**	9.9773	5	3		189973 (.25)	
8082A TNYP	DUP	**1664809**	10.019	5	3		189973 (.25)	

Sample Notes:

1*: An extra half a mL was added to the lipids tin so the delivered volumes is 0.5mL lower than usual. AMC 4/6/18

Standard Notes:

182336: Ar1242 Spike @20ug/mL 40GCS9. **189973: Pest/PCB Surrogate Spike 40GCSC.**

```
    Sequence Name: C:\msdchem\1\sequence\031318.s
          Comment:
         Operator:
        Data Path: C:\MSDCHEM\1\DATA\031318\
Instrument Control  Pre-Seq Cmd:
Data Analysis       Pre-Seq Cmd:

Instrument Control Post-Seq Cmd:
Data Analysis      Post-Seq Cmd:

    Method Sections To Run       Sequence Barcode Options
    (X) Full Method              (X) On Mismatch, Inject Anyway
    ( ) Reprocessing Only        ( ) On Mismatch, Don't Inject
                                 ( ) Barcode Disabled
   ---------------------------------------------------------------------
   Line                    Sample Name/Misc Info
 1) RearSamp        1  HEXANE
     Datafile          --------
     Method            SCREEN2014
 2) Sample          2  HEXANE
     Datafile          --------
     Method            SCREEN2014
 3) RearSamp        3  1658171MB
     Datafile          --------
     Method            SCREEN2014
 4) Sample          4  1658172LCS
     Datafile          --------
     Method            SCREEN2014
 5) RearSamp        5  1658173MS
     Datafile          --------
     Method            SCREEN2014
 6) Sample          6  1658174MSD
     Datafile          --------
     Method            SCREEN2014
 7) RearSamp        7  40165731-001
     Datafile          --------
     Method            SCREEN2014
 8) Sample          8  40165479-001
     Datafile          --------
     Method            SCREEN2014
 9) RearSamp        9  40165479-002
     Datafile          --------
     Method            SCREEN2014
10) Sample         10  40165479-003
     Datafile          --------
     Method            SCREEN2014
11) RearSamp       11  40165479-004
     Datafile          --------
     Method            SCREEN2014
12) Sample         12  40165650-001
     Datafile          --------
     Method            SCREEN2014
13) RearSamp       13  40165650-002
     Datafile          --------
     Method            SCREEN2014
14) Sample         14  40165650-003P
     Datafile          --------
     Method            SCREEN2014
15) RearSamp       15  40165650-004
     Datafile          --------
     Method            SCREEN2014
16) Sample         16  40165696-001
     Datafile          --------
     Method            SCREEN2014
17) RearSamp       17  1655661MB
     Datafile          --------
     Method            SCREEN2014
18) Sample         18  1655662LCS
     Datafile          --------
     Method            SCREEN2014
19) RearSamp       19  1655663MS x5
     Datafile          --------

   Last Modified: Tue Mar 13 12:00:34 2018                Page:  1
```

GE BIOTA

BLM 3/15/18

Continued on Page 10

Read and Understood By

Signed 3/15/18 Date Signed 3/15/18 Date

	Method		SCREEN2014	
20)	Sample	20	40150295-019	x5
	Datafile		--------	
	Method		SCREEN2014	
21)	RearSamp	21	1655664DUP	x5
	Datafile		--------	
	Method		SCREEN2014	
22)	Sample	22	40150295-020	—
	Datafile		--------	
	Method		SCREEN2014	
23)	RearSamp	23	40151597-002	—
	Datafile		--------	
	Method		SCREEN2014	
24)	Sample	24	40151597-003	—
	Datafile		--------	
	Method		SCREEN2014	
25)	RearSamp	25	40151601-002	—
	Datafile		--------	
	Method		SCREEN2014	
26)	Sample	26	40151608-001	x2
	Datafile		--------	
	Method		SCREEN2014	
27)	RearSamp	27	40151608-002	—
	Datafile		--------	
	Method		SCREEN2014	
28)	Sample	28	40151608-011	—
	Datafile		--------	
	Method		SCREEN2014	
29)	RearSamp	29	40151800-006	x3
	Datafile		--------	
	Method		SCREEN2014	
30)	Sample	30	40151800-007	x3
	Datafile		--------	
	Method		SCREEN2014	
31)	RearSamp	31	40151800-008	x3
	Datafile		--------	
	Method		SCREEN2014	
32)	Sample	32	40151800-009	x3
	Datafile		--------	
	Method		SCREEN2014	
33)	RearSamp	33	40151800-010	x3
	Datafile		--------	
	Method		SCREEN2014	
34)	Sample	34	40151804-004	x3
	Datafile		--------	
	Method		SCREEN2014	
35)	RearSamp	35	40151804-005	—
	Datafile		--------	
	Method		SCREEN2014	
36)	Sample	36	40151804-006	x3
	Datafile		--------	
	Method		SCREEN2014	
37)	RearSamp	37	40151804-007	x15
	Datafile		--------	
	Method		SCREEN2014	
38)	Sample	38	1655665DUP	x15
	Datafile		--------	
	Method		SCREEN2014	
39)	RearSamp	39	40151804-008	x3
	Datafile		--------	
	Method		SCREEN2014	
40)	Sample	40	HEXANE	
	Datafile		--------	
	Method		SCREEN2014	

Last Modified: Tue Mar 13 12:00:34 2018

Page: 2

BLM 3/15/18

Continued on Page —

Read and Understood By

Signed 3/15/18 Date

Signed 3/15/18 Date

```
   Sequence Name: C:\msdchem\1\sequence\031418.s
         Comment:
        Operator:
       Data Path: C:\MSDCHEM\1\DATA\031418\
Instrument Control  Pre-Seq Cmd:
Data Analysis       Pre-Seq Cmd:

Instrument Control Post-Seq Cmd:
Data Analysis      Post-Seq Cmd:

    Method Sections To Run       Sequence Barcode Options
    (X) Full Method              (X) On Mismatch, Inject Anyway
    ( ) Reprocessing Only        ( ) On Mismatch, Don't Inject
                                 ( ) Barcode Disabled
   ---------------------------------------------------------------------------
   Line                     Sample Name/Misc Info
 1) RearSamp        1  HEXANE
     Datafile          --------
     Method            SCREEN2014
 2) Sample          2  HEXANE
     Datafile          --------
     Method            SCREEN2014
 3) RearSamp        3  1658516MB
     Datafile          --------
     Method            SCREEN2014
 4) Sample          4  1658517LCS
     Datafile          --------
     Method            SCREEN2014
 5) RearSamp        5  1658518MS
     Datafile          --------
     Method            SCREEN2014
 6) Sample          6  1658519MSD
     Datafile          --------
     Method            SCREEN2014
 7) RearSamp        7  40165780-001
     Datafile          --------
     Method            SCREEN2014
 8) Sample          8  40165780-002
     Datafile          --------
     Method            SCREEN2014
 9) RearSamp        9  40165780-003
     Datafile          --------
     Method            SCREEN2014
10) Sample         10  40165627-002
     Datafile          --------
     Method            SCREEN2014
11) RearSamp       11  40165627-003
     Datafile          --------
     Method            SCREEN2014
12) Sample         12  40165627-004
     Datafile          --------
     Method            SCREEN2014
13) RearSamp       13  40165627-005P
     Datafile          --------
     Method            SCREEN2014
14) Sample         14  40165627-006
     Datafile          --------
     Method            SCREEN2014
15) RearSamp       15  1658254MB
     Datafile          --------
     Method            SCREEN2014
16) Sample         16  1658255LCS
     Datafile          --------
     Method            SCREEN2014
17) RearSamp       17  1658256MS x 3
     Datafile          --------
     Method            SCREEN2014
18) Sample         18  40151549-002
     Datafile          --------
     Method            SCREEN2014
19) RearSamp       19  40151549-003
     Datafile          --------

   Last Modified: Wed Mar 14 10:16:14 2018                Page:  1
```

GE BIOTA

BLM 3/15/18

Continued on Page 12

Read and Understood By

Signed 3/15/18 Date Signed 3/15/18 Date

```
      Method                SCREEN2014
20) Sample          20  40151549-004 —
      Datafile              --------
      Method                SCREEN2014
21) RearSamp        21  40151549-005 x5
      Datafile              --------
      Method                SCREEN2014
22) Sample          22  40151549-006 x2
      Datafile              --------
      Method                SCREEN2014
23) RearSamp        23  40151549-007 x3
      Datafile              --------
      Method                SCREEN2014
24) Sample          24  40151549-008 x2
      Datafile              --------
      Method                SCREEN2014
25) RearSamp        25  40151549-009 —
      Datafile              --------
      Method                SCREEN2014
26) Sample          26  40151552-010P x3
      Datafile              --------
      Method                SCREEN2014
27) RearSamp        27  1658257DUP x3
      Datafile              --------
      Method                SCREEN2014
28) Sample          28  40151601-003 —
      Datafile              --------
      Method                SCREEN2014
29) RearSamp        29  40151601-004 —
      Datafile              --------
      Method                SCREEN2014
30) Sample          30  40151601-005 —
      Datafile              --------
      Method                SCREEN2014
31) RearSamp        31  40151601-006 —
      Datafile              --------
      Method                SCREEN2014
32) Sample          32  HEXANE
      Datafile              --------
      Method                SCREEN2014
```

Last Modified: Wed Mar 14 10:16:14 2018 Page: 2

BLM 3/15/18

Continued on Page —

Read and Understood By

Signed 03/15/18 Date Signed 3/15/18 Date

```
   Sequence Name: C:\msdchem\1\sequence\033018.s
         Comment:
        Operator:
       Data Path: C:\MSDCHEM\1\DATA\033018\
Instrument Control  Pre-Seq Cmd:
Data Analysis       Pre-Seq Cmd:

Instrument Control Post-Seq Cmd:
Data Analysis      Post-Seq Cmd:

    Method Sections To Run       Sequence Barcode Options
    (X) Full Method              (X) On Mismatch, Inject Anyway
    ( ) Reprocessing Only        ( ) On Mismatch, Don't Inject
                                 ( ) Barcode Disabled
   ---------------------------------------------------------------------
   Line                     Sample Name/Misc Info
 1) RearSamp        1  HEXANE
     Datafile          --------
     Method            SCREEN2014
 2) Sample          2  HEXANE
     Datafile          --------
     Method            SCREEN2014
 3) RearSamp        3  1664793MB
     Datafile          --------
     Method            SCREEN2014
 4) Sample          4  1664794LCS
     Datafile          --------
     Method            SCREEN2014
 5) RearSamp        5  1664795MS —
     Datafile          --------
     Method            SCREEN2014
 6) Sample          6  40151549-001 x3
     Datafile          --------
     Method            SCREEN2014
 7) RearSamp        7  40151601-001 —
     Datafile          --------
     Method            SCREEN2014
 8) Sample          8  40151613-001 —
     Datafile          --------
     Method            SCREEN2014
 9) RearSamp        9  1664796DUP —
     Datafile          --------
     Method            SCREEN2014
10) Sample         10  40151613-002 —
     Datafile          --------
     Method            SCREEN2014
11) RearSamp       11  40151614-002 x2
     Datafile          --------
     Method            SCREEN2014
12) Sample         12  40151614-003 x3
     Datafile          --------
     Method            SCREEN2014
13) RearSamp       13  40151614-004 —
     Datafile          --------
     Method            SCREEN2014
14) Sample         14  40151613-014 —
     Datafile          --------
     Method            SCREEN2014
15) RearSamp       15  40151800-005 —
     Datafile          --------
     Method            SCREEN2014
16) Sample         16  40151802-001 —
     Datafile          --------
     Method            SCREEN2014
17) RearSamp       17  40151802-002 —
     Datafile          --------
     Method            SCREEN2014
18) Sample         18  40151802-003 —
     Datafile          --------
     Method            SCREEN2014
19) RearSamp       19  40151802-004 —
     Datafile          --------

   Last Modified: Fri Mar 30 07:49:18 2018                Page:  1
```

GEBIOTA

BLM 4/3/18

Continued on Page 32

Read and Understood By

Signed 4/3/18 Date Signed 4-3-18 Date

```
      Method              SCREEN2014
20) Sample          20  40151802-005
      Datafile            --------
      Method              SCREEN2014
21) RearSamp        21  40151805-001
      Datafile            --------
      Method              SCREEN2014
22) Sample          22  40157825-001
      Datafile            --------
      Method              SCREEN2014
23) RearSamp        23  1664809DUP
      Datafile            --------
      Method              SCREEN2014
24) Sample          24  HEXANE
      Datafile            --------
      Method              SCREEN2014
25) RearSamp        25  40166572-014
      Datafile            --------
      Method              SCREEN2014
26) Sample          26  40166575-010
      Datafile            --------
      Method              SCREEN2014
27) RearSamp        27  40166548-002 X1000
      Datafile            --------
      Method              SCREEN2014
28) Sample          28  1666113MS
      Datafile            --------
      Method              SCREEN2014
29) RearSamp        29  1666114DUP
      Datafile            --------
      Method              SCREEN2014
30) Sample          30  40166571-015
      Datafile            --------
      Method              SCREEN2014
31) RearSamp        31  40166568-002
      Datafile            --------
      Method              SCREEN2014
32) Sample          32  40166568-004
      Datafile            --------
      Method              SCREEN2014
33) RearSamp        33  40166568-005
      Datafile            --------
      Method              SCREEN2014
34) Sample          34  40166568-006
      Datafile            --------
      Method              SCREEN2014
35) RearSamp        35  40166568-007
      Datafile            --------
      Method              SCREEN2014
36) Sample          36  40166568-008
      Datafile            --------
      Method              SCREEN2014
37) RearSamp        37  40166568-009
      Datafile            --------
      Method              SCREEN2014
38) Sample          38  40166568-010
      Datafile            --------
      Method              SCREEN2014
39) RearSamp        39  40166568-011
      Datafile            --------
      Method              SCREEN2014
40) Sample          40  40166568-012
      Datafile            --------
      Method              SCREEN2014
41) RearSamp        41  40166568-013
      Datafile            --------
      Method              SCREEN2014
42) Sample          42  40166568-014
      Datafile            --------
      Method              SCREEN2014
43) RearSamp        43  40166568-016
      Datafile            --------

   Last Modified: Fri Mar 30 07:49:18 2018                Page:  2
```

BLM 4/3/18

Continued on Page 33

Read and Understood By

Signed: Bryan Mayl Date: 4/3/18 Signed: Dawn Nagy Date: 4-3-18

Sequence Name: C:\msdchem\1\sequence\033018.s

Line Type	Vial	DataFile	Method	Sample Name
44) Sample	44	40166568-017		
Datafile		--------		
Method			SCREEN2014	
45) RearSamp	45	40166568-018		
Datafile		--------		
Method			SCREEN2014	
46) Sample	46	40166568-019		
Datafile		--------		
Method			SCREEN2014	
47) RearSamp	47	40166568-020		
Datafile		--------		
Method			SCREEN2014	
48) Sample	48	HEXANE		
Datafile		--------		
Method			SCREEN2014	
49) RearSamp	49	40166548-002 X100		
Datafile		--------		
Method			SCREEN2014	
50) Sample	50	1665641MB		
Datafile		--------		
Method			SCREEN2014	
51) RearSamp	51	1665642LCS		
Datafile		--------		
Method			SCREEN2014	
52) Sample	52	1665643MS × 2		
Datafile		--------		
Method			SCREEN2014	
53) RearSamp	53	40157488-001 —		
Datafile		--------		
Method			SCREEN2014	
54) Sample	54	40157488-002 × 2		
Datafile		--------		
Method			SCREEN2014	
55) RearSamp	55	40157488-003 —		
Datafile		--------		
Method			SCREEN2014	
56) Sample	56	40157488-004 × 3		
Datafile		--------		
Method			SCREEN2014	
57) RearSamp	57	40157488-005P —		
Datafile		--------		
Method			SCREEN2014	
58) Sample	58	1665644DUP —		
Datafile		--------		
Method			SCREEN2014	
59) RearSamp	59	40157489-001 —		
Datafile		--------		
Method			SCREEN2014	
60) Sample	60	40157489-002 —		
Datafile		--------		
Method			SCREEN2014	
61) RearSamp	61	40157489-003 —		
Datafile		--------		
Method			SCREEN2014	
62) Sample	62	40157489-004 —		
Datafile		--------		
Method			SCREEN2014	
63) RearSamp	63	40157489-005 ×2		
Datafile		--------		
Method			SCREEN2014	
64) Sample	64	40157825-002 —		
Datafile		--------		
Method			SCREEN2014	
65) RearSamp	65	40157825-003 —		
Datafile		--------		
Method			SCREEN2014	
66) Sample	66	40157825-004 —		
Datafile		--------		

GE BIOTA

Last Modified: Fri Mar 30 09:59:08 2018

Page: 4

BLM 4/3/18

Continued on Page 34

Read and Understood By

Signed 4/3/18 Date

Signed 4-3-18 Date

```
      Method              SCREEN2014
67) RearSamp       67  40157825-005 —
      Datafile            --------
      Method              SCREEN2014
68) Sample         68  40157825-006 —
      Datafile            --------
      Method              SCREEN2014
69) RearSamp       69  1665645DUP —
      Datafile            --------
      Method              SCREEN2014
70) Sample         70  40157825-007 ×2
      Datafile            --------
      Method              SCREEN2014
71) RearSamp       71  40157825-008 ×3
      Datafile            --------
      Method              SCREEN2014
72) Sample         72  40157825-009 ×5
      Datafile            --------
      Method              SCREEN2014
73) RearSamp       73  40157825-010 —
      Datafile            --------
      Method              SCREEN2014
74) Sample         74  HEXANE
      Datafile            --------
      Method              SCREEN2014
```

Last Modified: Fri Mar 30 09:59:08 2018

Page: 5

BLM 4/3/18

Continued on Page —

Read and Understood By

Signed 4/3/18 Date

Signed 4-3-18 Date

```
   Sequence Name: C:\msdchem\1\sequence\030618.s
         Comment:
        Operator: BLM
       Data Path: C:\MSDCHEM\1\DATA\030618\
Instrument Control  Pre-Seq Cmd:
Data Analysis       Pre-Seq Cmd:

Instrument Control Post-Seq Cmd:
Data Analysis      Post-Seq Cmd:

    Method Sections To Run       Sequence Barcode Options
    (X) Full Method              (X) On Mismatch, Inject Anyway
    ( ) Reprocessing Only        ( ) On Mismatch, Don't Inject
                                 ( ) Barcode Disabled
   ---------------------------------------------------------------------
   Line                    Sample Name/Misc Info
 1)  Unlinked
 2) RearSamp        1  prime
     Datafile          030618001
     Method            GEBIOTA2017
 3) Sample          2  prime
     Datafile          030618002
     Method            GEBIOTA2017
 4) RearSamp        3  HEXANE
     Datafile          030618003
     Method            GEBIOTA2017
 5) Sample          4  HEXANE
     Datafile          030618004
     Method            GEBIOTA2017
 6) RearSamp        5  1232-CAL3S,185539:1,17499,NSA
     Datafile          030618005
     Method            GEBIOTA2017
 7) Sample          4  HEXANE
     Datafile          030618006
     Method            GEBIOTA2017
 8) RearSamp        6  1221-CAL1F,188138:1,17499,NSA
     Datafile          030618007
     Method            GEBIOTA2017
 9) Sample          5  1232-CAL3S,185539:1,17498,NSA
     Datafile          030618008
     Method            GEBIOTA2017
10) RearSamp        7  1221-CAL2F,188139:1,17499,NSA
     Datafile          030618009
     Method            GEBIOTA2017
11) Sample          6  1221-CAL1F,188138:1,17498,NSA
     Datafile          030618010
     Method            GEBIOTA2017
12) RearSamp        8  1221-CAL3F,185538:1,17499,NSA
     Datafile          030618011
     Method            GEBIOTA2017
13) Sample          7  1221-CAL2F,188139:1,17498,NSA
     Datafile          030618012
     Method            GEBIOTA2017
14) RearSamp        9  1221-CAL4F,188140:1,17499,NSA
     Datafile          030618013
     Method            GEBIOTA2017
15) Sample          8  1221-CAL3F,185538:1,17498,NSA
     Datafile          030618014
     Method            GEBIOTA2017
16) RearSamp       10  1221-CAL5F,188141:1,17499,NSA
     Datafile          030618015
     Method            GEBIOTA2017
17) Sample          9  1221-CAL4F,188140:1,17498,NSA
     Datafile          030618016
     Method            GEBIOTA2017
18) RearSamp       11  HEXANE
     Datafile          030618017
     Method            GEBIOTA2017
19) Sample         10  1221-CAL5F,188141:1,17498,NSA
     Datafile          030618018
     Method            GEBIOTA2017
20) RearSamp       12  1242-CAL1B,188132:1,17499,NSA

   Last Modified: Tue Mar 06 10:58:59 2018                Page:  1
```

BLM 3/7/18

Continued on Page 5

Read and Understood By

Bryan Mayl — Signed — 3/7/18 — Date

Signed — 3/7/18 — Date

```
      Datafile          030618019
      Method            GEBIOTA2017
21) Sample        11    HEXANE
      Datafile          030618020
      Method            GEBIOTA2017
22) RearSamp      13    1242-CAL2B,185543:1,17499,NSA
      Datafile          030618021
      Method            GEBIOTA2017
23) Sample        12    1242-CAL1B,188132:1,17498,NSA
      Datafile          030618022
      Method            GEBIOTA2017
24) RearSamp      14    1242-CAL3B,185544:1,17499,NSA
      Datafile          030618023
      Method            GEBIOTA2017
25) Sample        13    1242-CAL2B,185543:1,17498,NSA
      Datafile          030618024
      Method            GEBIOTA2017
26) RearSamp      15    1242-CAL4B,185545:1,17499,NSA
      Datafile          030618025
      Method            GEBIOTA2017
27) Sample        14    1242-CAL3B,185544:1,17498,NSA
      Datafile          030618026
      Method            GEBIOTA2017
28) RearSamp      16    1242-CAL5B,185546:1,17499,NSA
      Datafile          030618027
      Method            GEBIOTA2017
29) Sample        15    1242-CAL4B,185545:1,17498,NSA
      Datafile          030618028
      Method            GEBIOTA2017
30) RearSamp      17    HEXANE
      Datafile          030618029
      Method            GEBIOTA2017
31) Sample        16    1242-CAL5B,185546:1,17498,NSA
      Datafile          030618030
      Method            GEBIOTA2017
32) RearSamp      18    1248-CAL1D,188134:1,17499,NSA
      Datafile          030618031
      Method            GEBIOTA2017
33) Sample        17    HEXANE
      Datafile          030618032
      Method            GEBIOTA2017
34) RearSamp      19    1248-CAL2D,185548:1,17499,NSA
      Datafile          030618033
      Method            GEBIOTA2017
35) Sample        18    1248-CAL1D,188134:1,17498,NSA
      Datafile          030618034
      Method            GEBIOTA2017
36) RearSamp      20    1248-CAL3D,185549:1,17499,NSA
      Datafile          030618035
      Method            GEBIOTA2017
37) Sample        19    1248-CAL2D,185548:1,17498,NSA
      Datafile          030618036
      Method            GEBIOTA2017
38) RearSamp      21    1248-CAL4D,185550:1,17499,NSA
      Datafile          030618037
      Method            GEBIOTA2017
39) Sample        20    1248-CAL3D,185549:1,17498,NSA
      Datafile          030618038
      Method            GEBIOTA2017
40) RearSamp      22    1248-CAL5D,185551:1,17499,NSA
      Datafile          030618039
      Method            GEBIOTA2017
41) Sample        21    1248-CAL4D,185550:1,17498,NSA
      Datafile          030618040
      Method            GEBIOTA2017
42) RearSamp      23    HEXANE
      Datafile          030618041
      Method            GEBIOTA2017
43) Sample        22    1248-CAL5D,185551:1,17498,NSA
      Datafile          030618042
      Method            GEBIOTA2017

Last Modified: Tue Mar 06 10:58:59 2018          Page:  2
```

BLM 3/7/18

Continued on Page 6

Read and Understood By

Signed 3/7/18 Date Signed 3/5/18 Date

Sequence Name: C:\msdchem\1\sequence\030618.s

	Line Type	Vial	DataFile	Method	Sample Name
44)	RearSamp	24	1254-CAL1C,188135:1,17499,NSA		
	Datafile		030618043		
	Method			GEBIOTA2017	
45)	Sample	23	HEXANE		
	Datafile		030618044		
	Method			GEBIOTA2017	
46)	RearSamp	25	1254-CAL2C,185553:1,17499,NSA		
	Datafile		030618045		
	Method			GEBIOTA2017	
47)	Sample	24	1254-CAL1C,188135:1,17498,NSA		
	Datafile		030618046		
	Method			GEBIOTA2017	
48)	RearSamp	26	1254-CAL3C,185554:1,17499,NSA		
	Datafile		030618047		
	Method			GEBIOTA2017	
49)	Sample	25	1254-CAL2C,185553:1,17498,NSA		
	Datafile		030618048		
	Method			GEBIOTA2017	
50)	RearSamp	27	1254-CAL4C,185555:1,17499,NSA		
	Datafile		030618049		
	Method			GEBIOTA2017	
51)	Sample	26	1254-CAL3C,185554:1,17498,NSA		
	Datafile		030618050		
	Method			GEBIOTA2017	
52)	RearSamp	28	1254-CAL5C,185556:1,17499,NSA		
	Datafile		030618051		
	Method			GEBIOTA2017	
53)	Sample	27	1254-CAL4C,185555:1,17498,NSA		
	Datafile		030618052		
	Method			GEBIOTA2017	
54)	RearSamp	29	HEXANE		
	Datafile		030618053		
	Method			GEBIOTA2017	
55)	Sample	28	1254-CAL5C,185556:1,17498,NSA		
	Datafile		030618054		
	Method			GEBIOTA2017	
56)	RearSamp	30	1660-CAL1A,188136:1,17499,AR1		
	Datafile		030618055		
	Method			GEBIOTA2017	
57)	Sample	29	HEXANE		
	Datafile		030618056		
	Method			GEBIOTA2017	
58)	RearSamp	31	1660-CAL2A,185558:1,17499,AR1		
	Datafile		030618057		
	Method			GEBIOTA2017	
59)	Sample	30	1660-CAL1A,188136:1,17498,AR1		
	Datafile		030618058		
	Method			GEBIOTA2017	
60)	RearSamp	32	1660-CAL3A,185559:1,17499,AR1		
	Datafile		030618059		
	Method			GEBIOTA2017	
61)	Sample	31	1660-CAL2A,185558:1,17498,AR1		
	Datafile		030618060		
	Method			GEBIOTA2017	
62)	RearSamp	33	1660-CAL4A,185560:1,17499,AR1		
	Datafile		030618061		
	Method			GEBIOTA2017	
63)	Sample	32	1660-CAL3A,185559:1,17498,AR1		
	Datafile		030618062		
	Method			GEBIOTA2017	
64)	RearSamp	34	1660-CAL5A,185561:1,17499,AR1		
	Datafile		030618063		
	Method			GEBIOTA2017	
65)	Sample	33	1660-CAL4A,185560:1,17498,AR1		
	Datafile		030618064		
	Method			GEBIOTA2017	
66)	RearSamp	35	HEXANE		
	Datafile		030618065		

Last Modified: Tue Mar 06 10:58:59 2018

Page: 3

BLM 3/7/18

Continued on Page 7

Read and Understood By

Bryan Meyer — Signed — 3/7/18 Date

Signed — 3/7/18 Date

```
      Method                GEBIOTA2017
67) Sample          34  1660-CAL5A,185561:1,17498,AR1
      Datafile            030618066
      Method              GEBIOTA2017
68) RearSamp        36  1221-ICV,188142:1,17499,AR122
      Datafile            030618067
      Method              GEBIOTA2017
69) Sample          35  HEXANE
      Datafile            030618068
      Method              GEBIOTA2017
70) RearSamp        37  1242-ICV,185562:1,17499,AR124 ok
      Datafile            030618069
      Method              GEBIOTA2017
71) Sample          36  1221-ICV,188142:1,17498,AR122 ok
      Datafile            030618070
      Method              GEBIOTA2017
72) RearSamp        38  1248-ICV,185563:1,17499,AR124 ok
      Datafile            030618071
      Method              GEBIOTA2017
73) Sample          37  1242-ICV,185562:1,17498,AR124 ok
      Datafile            030618072
      Method              GEBIOTA2017
74) RearSamp        39  1254-ICV,185564:1,17499,AR125 ok
      Datafile            030618073
      Method              GEBIOTA2017
75) Sample          38  1248-ICV,185563:1,17498,AR124 ok
      Datafile            030618074
      Method              GEBIOTA2017
76) RearSamp        40  1660-ICV,185565:1,17499,AR166 ok
      Datafile            030618075
      Method              GEBIOTA2017
77) Sample          39  1254-ICV,185564:1,17498,AR125 ok
      Datafile            030618076
      Method              GEBIOTA2017
78) RearSamp        41  1221-CRDL,188138:1,17499,RL12 ok
      Datafile            030618077
      Method              GEBIOTA2017
79) Sample          40  1660-ICV,185565:1,17498,AR166 ok
      Datafile            030618078
      Method              GEBIOTA2017
80) RearSamp        42  1242-CRDL,188132:1,17499,RL12 ok
      Datafile            030618079
      Method              GEBIOTA2017
81) Sample          41  1221-CRDL,188138:1,17498,RL12 ok
      Datafile            030618080
      Method              GEBIOTA2017
82) RearSamp        43  1248-CRDL,188134:1,17499,RL12 ok
      Datafile            030618081
      Method              GEBIOTA2017
83) Sample          42  1242-CRDL,188132:1,17498,RL12 ok
      Datafile            030618082
      Method              GEBIOTA2017
84) RearSamp        44  1254-CRDL,188135:1,17499,RL12 ok
      Datafile            030618083
      Method              GEBIOTA2017
85) Sample          43  1248-CRDL,188134:1,17498,RL12 ok
      Datafile            030618084
      Method              GEBIOTA2017
86) RearSamp        45  1660-CRDL,188136:1,17499,RL16 ok
      Datafile            030618085
      Method              GEBIOTA2017
87) Sample          44  1254-CRDL,188135:1,17498,RL12 ok
      Datafile            030618086
      Method              GEBIOTA2017
88) RearSamp        46  HEXANE
      Datafile            030618087
      Method              GEBIOTA2017
89) Sample          45  1660-CRDL,188136:1,17498,RL16 ok
      Datafile            030618088
      Method              GEBIOTA2017

Last Modified: Tue Mar 06 10:58:59 2018          Page:  4
```

BLM 3/7/18

Continued on Page —

Read and Understood By

Signed: [signature] Date: 3/7/18

Signed: [signature] Date: 3/7/18

```
   Sequence Name: C:\msdchem\1\sequence\031318.s
         Comment:
        Operator: BLM
       Data Path: C:\MSDCHEM\1\DATA\031318\
Instrument Control  Pre-Seq Cmd:
Data Analysis       Pre-Seq Cmd:

Instrument Control Post-Seq Cmd:
Data Analysis      Post-Seq Cmd:

    Method Sections To Run       Sequence Barcode Options
    (X) Full Method              (X) On Mismatch, Inject Anyway
    ( ) Reprocessing Only        ( ) On Mismatch, Don't Inject
                                 ( ) Barcode Disabled
   ---------------------------------------------------------------------
   Line                     Sample Name/Misc Info
 1) Sample          1  prime
     Datafile          031318001
     Method            GEBIOTA2017
 2) Sample          2  HEXANE
     Datafile          031318002
     Method            GEBIOTA2017
 3) Sample          3  HEXANE
     Datafile          031318003
     Method            GEBIOTA2017
 4) Sample          4  1242-CCV,185544:1,17498,AR124 OK
     Datafile          031318004
     Method            GEBIOTA2017
 5) Sample          5  1242-CCV,185544:1,17498,AR124 OK
     Datafile          031318005
     Method            GEBIOTA2017
 6) Sample          6  1655661,,17521,BIOTA
     Datafile          031318006
     Method            GEBIOTA2017
 7) Sample          7  1655662,,17521,BIOTA
     Datafile          031318007
     Method            GEBIOTA2017
 8) Sample          8  1655663 x5,,17521,BIOTA
     Datafile          031318008
     Method            GEBIOTA2017
 9) Sample          9  40150295019 x5,,17521,BIOTA
     Datafile          031318009
     Method            GEBIOTA2017
10) Sample         10  1655664 x5,,17521,BIOTA
     Datafile          031318010
     Method            GEBIOTA2017
11) Sample         11  40150295020,,17521,BIOTA
     Datafile          031318011
     Method            GEBIOTA2017
12) Sample         12  40151597002,,17521,BIOTA RS
     Datafile          031318012
     Method            GEBIOTA2017
13) Sample         13  40151597003,,17521,BIOTA
     Datafile          031318013
     Method            GEBIOTA2017
14) Sample         14  40151601002,,17521,BIOTA
     Datafile          031318014
     Method            GEBIOTA2017
15) Sample         15  40151608001 x2,,17521,BIOTA
     Datafile          031318015
     Method            GEBIOTA2017
16) Sample         16  1248-CCV,185549:1,17498,AR124 OK
     Datafile          031318016
     Method            GEBIOTA2017
17) Sample         17  1248-CCV,185549:1,17498,AR124 OK
     Datafile          031318017
     Method            GEBIOTA2017
18) Sample         18  40151608002,,17521,BIOTA
     Datafile          031318018
     Method            GEBIOTA2017
19) Sample         19  40151608011,,17521,BIOTA RS
     Datafile          031318019

   Last Modified: Tue Mar 13 13:26:45 2018                Page:  1
```

BLM 3/14/18

Continued on Page 15

Read and Understood By

Signed 3/14/18 Date Signed 3/14/18 Date

```
      Method              GEBIOTA2017
20) Sample          20  40151800006 x3,,17521,BIOTA
      Datafile            031318020
      Method              GEBIOTA2017
21) Sample          21  40151800007 x3,,17521,BIOTA
      Datafile            031318021
      Method              GEBIOTA2017
22) Sample          22  40151800008 x3,,17521,BIOTA
      Datafile            031318022
      Method              GEBIOTA2017
23) Sample          23  40151800009 x3,,17521,BIOTA
      Datafile            031318023
      Method              GEBIOTA2017
24) Sample          24  40151800010 x3,,17521,BIOTA RS
      Datafile            031318024
      Method              GEBIOTA2017
25) Sample          25  40151804004 x3,,17521,BIOTA
      Datafile            031318025
      Method              GEBIOTA2017
26) Sample          26  40151804005,,17521,BIOTA RS
      Datafile            031318026
      Method              GEBIOTA2017
27) Sample          27  40151804006 x3,,17521,BIOTA
      Datafile            031318027
      Method              GEBIOTA2017
28) Sample          28  1254-CCV,185554:1,17498,AR125 OK
      Datafile            031318028
      Method              GEBIOTA2017
29) Sample          29  1254-CCV,185554:1,17498,AR125 OK
      Datafile            031318029
      Method              GEBIOTA2017
30) Sample          30  40151804007 x15,,17521,BIOTA
      Datafile            031318030
      Method              GEBIOTA2017
31) Sample          31  1655665 x15,,17521,BIOTA
      Datafile            031318031
      Method              GEBIOTA2017
32) Sample          32  40151804008 x3,,17521,BIOTA
      Datafile            031318032
      Method              GEBIOTA2017
33) Sample          33  1660-CCV,185559:1,17498,AR166 OK
      Datafile            031318033
      Method              GEBIOTA2017
34) Sample          34  1660-CCV,185559:1,17498,AR166 OK
      Datafile            031318034
      Method              GEBIOTA2017
35) Sample          35  1656929,,17518,BIOTA
      Datafile            031318035
      Method              GEBIOTA2017
36) Sample          36  1656930,,17518,BIOTA
      Datafile            031318036
      Method              GEBIOTA2017
37) Sample          37  1656931,,17518,BIOTA
      Datafile            031318037
      Method              GEBIOTA2017
38) Sample          38  1656932,,17518,BIOTA
      Datafile            031318038
      Method              GEBIOTA2017
39) Sample          39  1656933,,17518,BIOTA
      Datafile            031318039
      Method              GEBIOTA2017
40) Sample          40  1656934,,17518,BIOTA
      Datafile            031318040
      Method              GEBIOTA2017
41) Sample          41  1656935,,17518,BIOTA
      Datafile            031318041
      Method              GEBIOTA2017
42) Sample          42  1656936,,17518,BIOTA
      Datafile            031318042
      Method              GEBIOTA2017
43) Sample          43  1656937,,17518,BIOTA
      Datafile            031318043

   Last Modified: Tue Mar 13 13:26:45 2018                Page:  2
```

BLM 3/14/18

Continued on Page 16

Read and Understood By

Signed 3/14/18 Date Signed 3/14/18 Date

Sequence Name: C:\msdchem\1\sequence\031318.s

	Line Type	Vial	DataFile Method	Sample Name
44)	Sample	44	1656938,,17518,BIOTA	
	Datafile		031318044	
	Method		GEBIOTA2017	
45)	Sample	45	1221-CCV,185538:1,17498,AR122	ok
	Datafile		031318045	
	Method		GEBIOTA2017	
46)	Sample	46	1221-CCV,185538:1,17498,AR122	ok
	Datafile		031318046	
	Method		GEBIOTA2017	
47)	Sample	47	1656939,,17518,BIOTA	
	Datafile		031318047	
	Method		GEBIOTA2017	
48)	Sample	48	1656940,,17518,BIOTA	
	Datafile		031318048	
	Method		GEBIOTA2017	
49)	Sample	49	1656941,,17518,BIOTA	
	Datafile		031318049	
	Method		GEBIOTA2017	
50)	Sample	50	1656942,,17518,BIOTA	
	Datafile		031318050	
	Method		GEBIOTA2017	
51)	Sample	51	1656943,,17518,BIOTA	
	Datafile		031318051	
	Method		GEBIOTA2017	
52)	Sample	52	1656944,,17518,BIOTA	
	Datafile		031318052	
	Method		GEBIOTA2017	
53)	Sample	53	1656945,,17518,BIOTA	
	Datafile		031318053	
	Method		GEBIOTA2017	
54)	Sample	54	1232-CCV,185539:1,17498,AR123	ok
	Datafile		031318054	
	Method		GEBIOTA2017	
55)	Sample	55	1232-CCV,185539:1,17498,AR123	ok
	Datafile		031318055	
	Method		GEBIOTA2017	

Screener run on 40GCSI 03/13/18
BLK 3/23/18

Last Modified: Tue Mar 13 13:26:45 2018 Page: 4

BLK 3/14/18

Continued on Page —

Read and Understood By

Signed 3/14/18 Date Signed 3/14/18 Date

```
    Sequence Name: C:\msdchem\1\sequence\031418.s
          Comment:
         Operator: BLM
        Data Path: C:\MSDCHEM\1\DATA\031418\
Instrument Control  Pre-Seq Cmd:
Data Analysis       Pre-Seq Cmd:

Instrument Control Post-Seq Cmd:
Data Analysis      Post-Seq Cmd:

    Method Sections To Run       Sequence Barcode Options
    (X) Full Method              (X) On Mismatch, Inject Anyway
    ( ) Reprocessing Only        ( ) On Mismatch, Don't Inject
                                 ( ) Barcode Disabled
   ---------------------------------------------------------------------
   Line                    Sample Name/Misc Info
 1) Sample          1  prime
     Datafile          031418001
     Method            GEBIOTA2017
 2) Sample          2  HEXANE
     Datafile          031418002
     Method            GEBIOTA2017
 3) Sample          3  HEXANE
     Datafile          031418003
     Method            GEBIOTA2017
 4) Sample          4  1242-CCV,185544:1,17498,AR124 OK
     Datafile          031418004
     Method            GEBIOTA2017
 5) Sample          5  1242-CCV,185544:1,17498,AR124 OK
     Datafile          031418005
     Method            GEBIOTA2017
 6) Sample          6  1658254,,17523,BIOTA
     Datafile          031418006
     Method            GEBIOTA2017
 7) Sample          7  1658255,,17523,BIOTA
     Datafile          031418007
     Method            GEBIOTA2017
 8) Sample          8  1658256 x3,,17523,BIOTA
     Datafile          031418008
     Method            GEBIOTA2017
 9) Sample          9  40151549002,,17523,BIOTA
     Datafile          031418009
     Method            GEBIOTA2017
10) Sample         10  40151549003,,17523,BIOTA RS
     Datafile          031418010
     Method            GEBIOTA2017
11) Sample         11  40151549004,,17523,BIOTA RS
     Datafile          031418011
     Method            GEBIOTA2017
12) Sample         12  40151549005 x5,,17523,BIOTA
     Datafile          031418012
     Method            GEBIOTA2017
13) Sample         13  40151549006 x2,,17523,BIOTA
     Datafile          031418013
     Method            GEBIOTA2017
14) Sample         14  40151549007 x3,,17523,BIOTA
     Datafile          031418014
     Method            GEBIOTA2017
15) Sample         15  40151549008 x2,,17523,BIOTA
     Datafile          031418015
     Method            GEBIOTA2017
16) Sample         16  1248-CCV,185549:1,17498,AR124 OK
     Datafile          031418016
     Method            GEBIOTA2017
17) Sample         17  1248-CCV,185549:1,17498,AR124 OK
     Datafile          031418017
     Method            GEBIOTA2017
18) Sample         18  40151549009,,17523,BIOTA
     Datafile          031418018
     Method            GEBIOTA2017
19) Sample         19  40151552010 x3,,17523,BIOTA
     Datafile          031418019

   Last Modified: Wed Mar 14 12:18:30 2018                Page:  1
```

BLM 3/23/18

Continued on Page 18

Read and Understood By

Signed 3/23/18 Date Signed 3/26/18 Date

```
      Method              GEBIOTA2017
20) Sample          20  1658257 x3,,17523,BIOTA
      Datafile            031418020
      Method              GEBIOTA2017
21) Sample          21  40151601003,,17523,BIOTA
      Datafile            031418021
      Method              GEBIOTA2017
22) Sample          22  40151601004,,17523,BIOTA
      Datafile            031418022
      Method              GEBIOTA2017
23) Sample          23  40151601005,,17523,BIOTA
      Datafile            031418023
      Method              GEBIOTA2017
24) Sample          24  40151601006,,17523,BIOTA
      Datafile            031418024
      Method              GEBIOTA2017
25) Sample          25  1254-CCV,185554:1,17498,AR125 OK
      Datafile            031418025
      Method              GEBIOTA2017
26) Sample          26  1254-CCV,185554:1,17498,AR125 OK
      Datafile            031418026
      Method              GEBIOTA2017
```

Screen ran on 40GCSI 03/14/18

Last Modified: Wed Mar 14 12:18:30 2018 Page: 2

BLM 3/23/18

Continued on Page —

Read and Understood By

Signed 3/23/18 Date Signed 3/26/18 Date

```
   Sequence Name: C:\msdchem\1\sequence\033018.s
         Comment:
        Operator: BLM
       Data Path: C:\MSDCHEM\1\DATA\033018\
Instrument Control  Pre-Seq Cmd:
Data Analysis       Pre-Seq Cmd:

Instrument Control Post-Seq Cmd:
Data Analysis      Post-Seq Cmd:

    Method Sections To Run       Sequence Barcode Options
    (X) Full Method              (X) On Mismatch, Inject Anyway
    ( ) Reprocessing Only        ( ) On Mismatch, Don't Inject
                                 ( ) Barcode Disabled
  ---------------------------------------------------------------------
  Line                    Sample Name/Misc Info
 1) Sample          1  prime
     Datafile          033018001
     Method            GEBIOTA2017
 2) Sample          2  HEXANE
     Datafile          033018002
     Method            GEBIOTA2017
 3) Sample          3  HEXANE
     Datafile          033018003
     Method            GEBIOTA2017
 4) Sample          4  1232-CCV,185539:1,17498,AR123 ok
     Datafile          033018004
     Method            GEBIOTA2017
 5) Sample          5  1232-CCV,185539:1,17498,AR123 ok
     Datafile          033018005
     Method            GEBIOTA2017
 6) Sample          6  1662968 x40,,17575,BIOTA
     Datafile          033018006
     Method            GEBIOTA2017
 7) Sample          7  40151616010 x40,,17575,BIOTA
     Datafile          033018007
     Method            GEBIOTA2017
 8) Sample          8  40157828010 x10,,17584,BIOTA
     Datafile          033018008
     Method            GEBIOTA2017
 9) Sample          9  40157828019 x3,,17584,BIOTA
     Datafile          033018009
     Method            GEBIOTA2017
10) Sample         10  1664793,,17591,BIOTA
     Datafile          033018010
     Method            GEBIOTA2017
11) Sample         11  1664794,,17591,BIOTA
     Datafile          033018011
     Method            GEBIOTA2017
12) Sample         12  1664795,,17591,BIOTA
     Datafile          033018012
     Method            GEBIOTA2017
13) Sample         13  40151549001 x3,,17591,BIOTA
     Datafile          033018013
     Method            GEBIOTA2017
14) Sample         14  40151601001,,17591,BIOTA
     Datafile          033018014
     Method            GEBIOTA2017
15) Sample         15  40151613001,,17591,BIOTA
     Datafile          033018015
     Method            GEBIOTA2017
16) Sample         16  1242-CCV,185544:1,17498,AR124 ok
     Datafile          033018016
     Method            GEBIOTA2017
17) Sample         17  1242-CCV,185544:1,17498,AR124 ok
     Datafile          033018017
     Method            GEBIOTA2017
18) Sample         18  1664796,,17591,BIOTA
     Datafile          033018018
     Method            GEBIOTA2017
19) Sample         19  40151613002,,17591,BIOTA
     Datafile          033018019

   Last Modified: Fri Mar 30 14:55:13 2018                Page:  1
```

BLM 4/3/18

Continued on Page 37

Read and Understood By

Signed: Bryan Mayl — Date: 4/3/18

Signed: Dawn Haag — Date: 4-3-18

```
      Method              GEBIOTA2017
20) Sample          20  40151614002 x2,,17591,BIOTA
      Datafile            033018020
      Method              GEBIOTA2017
21) Sample          21  40151614003 x3,,17591,BIOTA
      Datafile            033018021
      Method              GEBIOTA2017
22) Sample          22  40151614004,,17591,BIOTA
      Datafile            033018022
      Method              GEBIOTA2017
23) Sample          23  40151613014,,17591,BIOTA
      Datafile            033018023
      Method              GEBIOTA2017
24) Sample          24  40151800005,,17591,BIOTA
      Datafile            033018024
      Method              GEBIOTA2017
25) Sample          25  40151802001,,17591,BIOTA
      Datafile            033018025
      Method              GEBIOTA2017
26) Sample          26  40151802002,,17591,BIOTA
      Datafile            033018026
      Method              GEBIOTA2017
27) Sample          27  40151802003,,17591,BIOTA
      Datafile            033018027
      Method              GEBIOTA2017
28) Sample          28  1248-CCV,185549:1,17498,AR124 ok
      Datafile            033018028
      Method              GEBIOTA2017
29) Sample          29  1248-CCV,185549:1,17498,AR124 ok
      Datafile            033018029
      Method              GEBIOTA2017
30) Sample          30  40151802004,,17591,BIOTA
      Datafile            033018030
      Method              GEBIOTA2017
31) Sample          31  40151802005,,17591,BIOTA
      Datafile            033018031
      Method              GEBIOTA2017
32) Sample          32  40151805001,,17591,BIOTA
      Datafile            033018032
      Method              GEBIOTA2017
33) Sample          33  40157825001,,17591,BIOTA
      Datafile            033018033
      Method              GEBIOTA2017
34) Sample          34  1664809,,17591,BIOTA
      Datafile            033018034
      Method              GEBIOTA2017
35) Sample          35  1665641,,17602,BIOTA
      Datafile            033018035
      Method              GEBIOTA2017
36) Sample          36  1665642,,17602,BIOTA
      Datafile            033018036
      Method              GEBIOTA2017
37) Sample          37  1665643 x2,,17602,BIOTA
      Datafile            033018037
      Method              GEBIOTA2017
38) Sample          38  40157488001,,17602,BIOTA
      Datafile            033018038
      Method              GEBIOTA2017
39) Sample          39  40157488002 x2,,17602,BIOTA
      Datafile            033018039
      Method              GEBIOTA2017
40) Sample          40  1254-CCV,185554:1,17498,AR125 ok
      Datafile            033018040
      Method              GEBIOTA2017
41) Sample          41  1254-CCV,185554:1,17498,AR125 ok
      Datafile            033018041
      Method              GEBIOTA2017
42) Sample          42  40157488003,,17602,BIOTA
      Datafile            033018042
      Method              GEBIOTA2017
43) Sample          43  40157488004 x3,,17602,BIOTA
      Datafile            033018043

   Last Modified: Fri Mar 30 14:55:13 2018                Page:  2
```

BLM 4/3/18

Continued on Page 38

Read and Understood By

Signed: [signature] Date: 4/3/18

Signed: [signature] Date: 4-3-18

Sequence Name: C:\msdchem\1\sequence\033018.s

	Line Type	Vial	DataFile	Method	Sample Name
44)	Sample	44	40157488005,,17602,BIOTA		
	Datafile		033018044		
	Method			GEBIOTA2017	
45)	Sample	45	1665644,,17602,BIOTA		
	Datafile		033018045		
	Method			GEBIOTA2017	
46)	Sample	46	40157489001,,17602,BIOTA		
	Datafile		033018046		
	Method			GEBIOTA2017	
47)	Sample	47	40157489002,,17602,BIOTA		
	Datafile		033018047		
	Method			GEBIOTA2017	
48)	Sample	48	40157489003,,17602,BIOTA		
	Datafile		033018048		
	Method			GEBIOTA2017	
49)	Sample	49	40157489004,,17602,BIOTA		
	Datafile		033018049		
	Method			GEBIOTA2017	
50)	Sample	50	40157489005 x2,,17602,BIOTA		
	Datafile		033018050		
	Method			GEBIOTA2017	
51)	Sample	51	40157825002,,17602,BIOTA		
	Datafile		033018051		
	Method			GEBIOTA2017	
52)	Sample	52	1660-CCV,185559:1,17498,AR166 ok		
	Datafile		033018052		
	Method			GEBIOTA2017	
53)	Sample	53	1660-CCV,185559:1,17498,AR166 ok		
	Datafile		033018053		
	Method			GEBIOTA2017	
54)	Sample	54	40157825003,,17602,BIOTA		
	Datafile		033018054		
	Method			GEBIOTA2017	
55)	Sample	55	40157825004,,17602,BIOTA		
	Datafile		033018055		
	Method			GEBIOTA2017	
56)	Sample	56	40157825005,,17602,BIOTA RS x2		
	Datafile		033018056		
	Method			GEBIOTA2017	
57)	Sample	57	40157825006,,17602,BIOTA		
	Datafile		033018057		
	Method			GEBIOTA2017	
58)	Sample	58	1665645,,17602,BIOTA		
	Datafile		033018058		
	Method			GEBIOTA2017	
59)	Sample	59	40157825007 x2,,17602,BIOTA		
	Datafile		033018059		
	Method			GEBIOTA2017	
60)	Sample	60	40157825008 x3,,17602,BIOTA		
	Datafile		033018060		
	Method			GEBIOTA2017	
61)	Sample	61	40157825009 x5,,17602,BIOTA		
	Datafile		033018061		
	Method			GEBIOTA2017	
62)	Sample	62	40157825010,,17602,BIOTA RS x2		
	Datafile		033018062		
	Method			GEBIOTA2017	
63)	Sample	63	1221-CCV,185538:1,17498,AR122 ok		
	Datafile		033018063		
	Method			GEBIOTA2017	
64)	Sample	64	1221-CCV,185538:1,17498,AR122 ok		
	Datafile		033018064		
	Method			GEBIOTA2017	

GCSV Batch 17591 HBN: 284630 AND 17602 HBN: 284746
Screened: 03/30/18 40GCSJ

Last Modified: Fri Mar 30 14:55:13 2018

Page: 4

BLM 4/3/18

Continued on Page —

Read and Understood By

Bryan Meyl — Signed — 4/3/18 — Date

Dawn Haag — Signed — 4-3-18 — Date

Standard Log Report

Standards Log Information for Standard #192488

WORKING STANDARD

Created By:	AMC	**Volume of Standard:**	4 L	**Lot ID:**	NA
Create Date:	03/28/2018	**Manufacturer:**	N/A	**Part ID:**	N/A
Expire Date:	09/28/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	RGNT498

Notes:

Hexane/Acetone 1:1 By AMC

Compound Name and Concentration for Standard #192488

Compound Name	Concentration	Compound Name	Concentration
Acetone	ug/mL	Hexanes	ug/mL

Composed of Information for Standard #192488

Composed of Standard Seq	Volume	Units
189481	2000	mL
188917	2000	mL

Standard Log Report

Standards Log Information for Standard #191186

WORKING STANDARD

Created By:	AMC	**Volume of Standard:**	4 L	**Lot ID:**	NA
Create Date:	03/12/2018	**Manufacturer:**	N/A	**Part ID:**	N/A
Expire Date:	09/12/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	RGNT498

Notes:

Hexane/Acetone 1:1 By AMC

Compound Name and Concentration for Standard #191186

Compound Name	Concentration	Compound Name	Concentration
Acetone	ug/mL	Hexanes	ug/mL

Composed of Information for Standard #191186

Composed of Standard Seq	Volume	Units
190951	2000	mL
188917	2000	mL

Standard Log Report

Standards Log Information for Standard #190953

STOCK STANDARD

Created By:	MRN	**Volume of Standard:**	45.5 kg	**Lot ID:**	405550
Create Date:	03/08/2018	**Manufacturer:**	JT Baker	**Part ID:**	3375-09
Expire Date:	05/22/2022	**Manufacturer Lot ID:**	0000177456	**Standard ID:**	RGNT301

Notes:

Sodium Sulfate SVOA

Compound Name and Concentration for Standard #190953

Compound Name	Concentration	Compound Name	Concentration
Sodium Sulfate			

The standards listed are from Pace's Green Bay Laboratory.

Standard Log Report

Standards Log Information for Standard #190951

STOCK STANDARD

Created By:	MRN	**Volume of Standard:**	4 L	**Lot ID:**	405758
Create Date:	03/08/2018	**Manufacturer:**	Burdick & Jackson	**Part ID:**	010-4
Expire Date:	03/21/2019	**Manufacturer Lot ID:**	DS923	**Standard ID:**	RGNT036

Notes:

Acetone - 4L container

Compound Name and Concentration for Standard #190951

Compound Name	Concentration	Compound Name	Concentration
Acetone	ug/mL		

The standards listed are from Pace's Green Bay Laboratory.

Standard Log Report

Standards Log Information for Standard #189973

WORKING STANDARD

Created By:	ASD	**Volume of Standard:**	1000 mL	**Lot ID:**	GCSV
Create Date:	02/26/2018	**Manufacturer:**	N/A	**Part ID:**	N/A
Expire Date:	08/26/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	8081-SS

Notes:

Pest/PCB Surrogate Spike 40GCSC.i\022618SurrVal.b\A226037 TCMX-116.01% DCB-98.69%

Compound Name and Concentration for Standard #189973

Compound Name	Concentration	Compound Name	Concentration
Acetone	ug/mL	Decachlorobiphenyl (S)	2 ug/mL
Tetrachloro-m-xylene (S)	2 ug/mL		

Composed of Information for Standard #189973

Composed of Standard Seq	Volume	Units
179366	999	mL
172034	1	mL

Standard Log Report

Standards Log Information for Standard #189481

STOCK STANDARD

Created By:	MRN	**Volume of Standard:**	4 L	**Lot ID:**	405758
Create Date:	02/19/2018	**Manufacturer:**	Burdick & Jackson	**Part ID:**	010-4
Expire Date:	03/21/2019	**Manufacturer Lot ID:**	DS923	**Standard ID:**	RGNT036

Notes:

Acetone - 4L container

Compound Name and Concentration for Standard #189481

Compound Name	Concentration	Compound Name	Concentration
Acetone	ug/mL		

Standard Log Report

Standards Log Information for Standard #188917

STOCK STANDARD

Created By:	ETB	**Volume of Standard:**	215 L	**Lot ID:**	405750
Create Date:	02/12/2018	**Manufacturer:**	J.T. Baker	**Part ID:**	9262-8P
Expire Date:	01/15/2019	**Manufacturer Lot ID:**	0000183983	**Standard ID:**	RGNT150

Notes:

Hexane SVOA TANK Changed ETB 02/12/18

Compound Name and Concentration for Standard #188917

Compound Name	Concentration	Compound Name	Concentration
Hexanes	ug/mL		

Standard Log Report

Standards Log Information for Standard #188340

STOCK STANDARD

Created By:	MRN	**Volume of Standard:**	45.5 kg	**Lot ID:**	405550
Create Date:	02/01/2018	**Manufacturer:**	JT Baker	**Part ID:**	3375-09
Expire Date:	05/22/2022	**Manufacturer Lot ID:**	0000177456	**Standard ID:**	RGNT301

Notes:

Sodium Sulfate SVOA

Compound Name and Concentration for Standard #188340

Compound Name	Concentration	Compound Name	Concentration
Sodium Sulfate			

The standards listed are from Pace's Green Bay Laboratory.

Standard Log Report

Standards Log Information for Standard #188142

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	01/30/2018	**Manufacturer:**	O2Si	**Part ID:**	N/A
Expire Date:	07/30/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1221-ICV

Notes:

1221 ICV Std @ 0.5 ug/mL

Compound Name and Concentration for Standard #188142

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.05 ug/mL	Hexanes	ug/mL
PCB-1221 (Aroclor 1221)	.5 ug/mL	Tetrachloro-m-xylene (S)	.05 ug/mL

Composed of Information for Standard #188142

Composed of Standard Seq	Volume	Units
188137	500	uL
188129	500	uL
184640	99	mL

Standard Log Report

Standards Log Information for Standard #188137

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	10 mL	**Lot ID:**	GCSV
Create Date:	01/30/2018	**Manufacturer:**	O2Si	**Part ID:**	N/A
Expire Date:	12/15/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1221-STK

Notes:

1221 ICV STK Std @ 100 ug/mL

Compound Name and Concentration for Standard #188137

Compound Name	Concentration	Compound Name	Concentration
Hexanes	ug/mL	PCB-1221 (Aroclor 1221)	100 ug/mL

Composed of Information for Standard #188137

Composed of Standard Seq	Volume	Units
184640	9	mL
179816	1000	uL

Standard Log Report

Standards Log Information for Standard #188136

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	01/30/2018	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	07/30/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1660-CAL1

Notes:

NY 1660 Cal 1 Std @ 0.1 ug/mL

Compound Name and Concentration for Standard #188136

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.01 ug/mL	Hexanes	ug/mL
PCB-1016 (Aroclor 1016)	.1 ug/mL	PCB-1260 (Aroclor 1260)	.1 ug/mL
Tetrachloro-m-xylene (S)	.01 ug/mL		

Composed of Information for Standard #188136

Composed of Standard Seq	Volume	Units
188129	100	uL
185533	100	uL
184640	99.8	mL

Standard Log Report

Standards Log Information for Standard #188135

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	01/30/2018	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	07/30/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1254-CAL1

Notes:

NY 1254 Cal 1 Std @ 0.1 ug/mL

Compound Name and Concentration for Standard #188135

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.01 ug/mL	Hexanes	ug/mL
PCB-1254 (Aroclor 1254)	.1 ug/mL	Tetrachloro-m-xylene (S)	.01 ug/mL

Composed of Information for Standard #188135

Composed of Standard Seq	Volume	Units
188129	100	uL
185530	100	uL
184640	99.8	mL

Standard Log Report

Standards Log Information for Standard #188134

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	01/30/2018	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	07/30/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1248-CAL1

Notes:

NY 1248 Cal 1 Std @ 0.1 ug/mL

Compound Name and Concentration for Standard #188134

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.01 ug/mL	Hexanes	ug/mL
PCB-1248 (Aroclor 1248)	.1 ug/mL	Tetrachloro-m-xylene (S)	.01 ug/mL

Composed of Information for Standard #188134

Composed of Standard Seq	Volume	Units
188129	100	uL
185529	100	uL
184640	99.8	mL

Standard Log Report

Standards Log Information for Standard #188132

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	01/30/2018	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	07/30/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1242-CAL1

Notes:

NY 1242 Cal 1 Std @ 0.1 ug/mL

Compound Name and Concentration for Standard #188132

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.01 ug/mL	Hexanes	ug/mL
PCB-1242 (Aroclor 1242)	.1 ug/mL	PCB-1242 (Aroclor 1242)	.1 ug/mL
PCB-1242 (Aroclor 1242)-6	.1 ug/mL	PCB-1242 (Aroclor 1242)-7	.1 ug/mL
PCB-1242 (Aroclor 1242)-8	.1 ug/mL	PCB-1242 (Aroclor 1242)-9	.1 ug/mL
Tetrachloro-m-xylene (S)	.01 ug/mL		

Composed of Information for Standard #188132

Composed of Standard Seq	Volume	Units
188129	100	uL
185528	100	uL
184640	99.8	mL

Standard Log Report

Standards Log Information for Standard #188129

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	20 mL	**Lot ID:**	GCSV
Create Date:	01/30/2018	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	12/15/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	PEST-SUR

Notes:

Pest/PCB Surr Std @ 10 ug/mL

Compound Name and Concentration for Standard #188129

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	10 ug/mL	Hexanes	ug/mL
Tetrachloro-m-xylene (S)	10 ug/mL		

Composed of Information for Standard #188129

Composed of Standard Seq	Volume	Units
185958	1	mL
184640	19	mL

Standard Log Report

Standards Log Information for Standard #186910

STOCK STANDARD

Created By:	AMC	**Volume of Standard:**	3 kg	**Lot ID:**	405706
Create Date:	01/12/2018	**Manufacturer:**	Fisher Scientific	**Part ID:**	F100-3
Expire Date:	01/11/2020	**Manufacturer Lot ID:**	174228	**Standard ID:**	RGNT142

Notes:

Florisil, 60-100 Mesh by AMC

Compound Name and Concentration for Standard #186910

Compound Name	Concentration	Compound Name	Concentration
Florisil			

The standards listed are from Pace's Green Bay Laboratory.

Standard Log Report

Standards Log Information for Standard #185958

INDIVIDUAL STANDARD

Created By:	BLM	**Volume of Standard:**	1 mL	**Lot ID:**	405693
Create Date:	01/02/2018	**Manufacturer:**	Restek	**Part ID:**	32000
Expire Date:	11/30/2023	**Manufacturer Lot ID:**	A0130424	**Standard ID:**	PEST-SUR

Notes:

Pest/PCB Surr @ 200 ug/mL

Compound Name and Concentration for Standard #185958

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	200 ug/mL	Tetrachloro-m-xylene (S)	200 ug/mL

Standard Log Report

Standards Log Information for Standard #185565

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	O2Si	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1660-ICV

Notes:

1660 ICV Std @ 0.5 ug/mL

Compound Name and Concentration for Standard #185565

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.05 ug/mL	Hexanes	ug/mL
PCB-1016 (Aroclor 1016)	.5 ug/mL	PCB-1260 (Aroclor 1260)	.5 ug/mL
Tetrachloro-m-xylene (S)	.05 ug/mL		

Composed of Information for Standard #185565

Composed of Standard Seq	Volume	Units
185537	500	uL
185524	500	uL
184640	99	mL

Standard Log Report

Standards Log Information for Standard #185564

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	O2Si	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1254-ICV

Notes:

1254 ICV Std @ 0.5 ug/mL

Compound Name and Concentration for Standard #185564

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.05 ug/mL	Hexanes	ug/mL
PCB-1254 (Aroclor 1254)	.5 ug/mL	Tetrachloro-m-xylene (S)	.05 ug/mL

Composed of Information for Standard #185564

Composed of Standard Seq	Volume	Units
185536	500	uL
185524	500	uL
184640	99	mL

Standard Log Report

Standards Log Information for Standard #185563

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	O2Si	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1248-ICV

Notes:

1248 ICV Std @ 0.5 ug/mL

Compound Name and Concentration for Standard #185563

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.05 ug/mL	Hexanes	ug/mL
PCB-1248 (Aroclor 1248)	.5 ug/mL	Tetrachloro-m-xylene (S)	.05 ug/mL

Composed of Information for Standard #185563

Composed of Standard Seq	Volume	Units
185535	500	uL
185524	500	uL
184640	99	mL

Standard Log Report

Standards Log Information for Standard #185562

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	O2Si	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1242-ICV

Notes:

1242 ICV Std @ 0.5 ug/mL

Compound Name and Concentration for Standard #185562

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.05 ug/mL	Hexanes	ug/mL
PCB-1242 (Aroclor 1242)	.5 ug/mL	Tetrachloro-m-xylene (S)	.05 ug/mL

Composed of Information for Standard #185562

Composed of Standard Seq	Volume	Units
185534	500	uL
185524	500	uL
184640	99	mL

Standard Log Report

Standards Log Information for Standard #185561

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1660-CAL5

Notes:

1660 Cal 5 Std @ 1.0 ug/mL

Compound Name and Concentration for Standard #185561

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.15 ug/mL	Hexanes	ug/mL
PCB-1016 (Aroclor 1016)	1 ug/mL	PCB-1260 (Aroclor 1260)	1 ug/mL
Tetrachloro-m-xylene (S)	.15 ug/mL		

Composed of Information for Standard #185561

Composed of Standard Seq	Volume	Units
185533	1	mL
185524	1500	uL
184640	97.5	mL

Standard Log Report

Standards Log Information for Standard #185560

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1660-CAL4

Notes:

1660 Cal 4 Std @ 0.8 ug/mL

Compound Name and Concentration for Standard #185560

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.1 ug/mL	Hexanes	ug/mL
PCB-1016 (Aroclor 1016)	.8 ug/mL	PCB-1260 (Aroclor 1260)	.8 ug/mL
Tetrachloro-m-xylene (S)	.1 ug/mL		

Composed of Information for Standard #185560

Composed of Standard Seq	Volume	Units
185533	800	uL
185524	1	mL
184640	98.2	mL

Standard Log Report

Standards Log Information for Standard #185559

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	200 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1660-CAL3

Notes:

1660 Cal 3 Std @ 0.5 ug/mL

Compound Name and Concentration for Standard #185559

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.05 ug/mL	Hexanes	ug/mL
PCB-1016 (Aroclor 1016)	.5 ug/mL	PCB-1260 (Aroclor 1260)	.5 ug/mL
Tetrachloro-m-xylene (S)	.05 ug/mL		

Composed of Information for Standard #185559

Composed of Standard Seq	Volume	Units
185533	1	mL
185524	1	mL
184640	198	mL

Standard Log Report

Standards Log Information for Standard #185558

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1660-CAL2

Notes:

1660 Cal 2 Std @ 0.2 ug/mL

Compound Name and Concentration for Standard #185558

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.02 ug/mL	Hexanes	ug/mL
PCB-1016 (Aroclor 1016)	.2 ug/mL	PCB-1260 (Aroclor 1260)	.2 ug/mL
Tetrachloro-m-xylene (S)	.02 ug/mL		

Composed of Information for Standard #185558

Composed of Standard Seq	Volume	Units
185533	200	uL
185524	200	uL
184640	99.6	mL

Standard Log Report

Standards Log Information for Standard #185556

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1254-CAL5

Notes:

1254 Cal 5 Std @ 1.0 ug/mL

Compound Name and Concentration for Standard #185556

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.15 ug/mL	Hexanes	ug/mL
PCB-1254 (Aroclor 1254)	1 ug/mL	Tetrachloro-m-xylene (S)	.15 ug/mL

Composed of Information for Standard #185556

Composed of Standard Seq	Volume	Units
185530	1000	uL
185524	1500	uL
184640	97.5	mL

Standard Log Report

Standards Log Information for Standard #185555

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1254-CAL4

Notes:

1254 Cal 4 Std @ 0.8 ug/mL

Compound Name and Concentration for Standard #185555

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.1 ug/mL	Hexanes	ug/mL
PCB-1254 (Aroclor 1254)	.8 ug/mL	Tetrachloro-m-xylene (S)	.1 ug/mL

Composed of Information for Standard #185555

Composed of Standard Seq	Volume	Units
185530	800	uL
185524	1000	uL
184640	98.2	mL

Standard Log Report

Standards Log Information for Standard #185554

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	200 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1254-CAL3

Notes:

1254 Cal 3 Std @ 0.5 ug/mL

Compound Name and Concentration for Standard #185554

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.05 ug/mL	Hexanes	ug/mL
PCB-1254 (Aroclor 1254)	.5 ug/mL	Tetrachloro-m-xylene (S)	.05 ug/mL

Composed of Information for Standard #185554

Composed of Standard Seq	Volume	Units
185530	1000	uL
185524	1000	uL
184640	198	mL

Standard Log Report

Standards Log Information for Standard #185553

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1254-CAL2

Notes:

1254 Cal 2 Std @ 0.2 ug/mL

Compound Name and Concentration for Standard #185553

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.02 ug/mL	Hexanes	ug/mL
PCB-1254 (Aroclor 1254)	.2 ug/mL	Tetrachloro-m-xylene (S)	.02 ug/mL

Composed of Information for Standard #185553

Composed of Standard Seq	Volume	Units
185530	200	uL
185524	200	uL
184640	99.6	mL

Standard Log Report

Standards Log Information for Standard #185551

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1248-CAL5

Notes:

1248 Cal 5 Std @ 1.0 ug/mL

Compound Name and Concentration for Standard #185551

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.15 ug/mL	Hexanes	ug/mL
PCB-1248 (Aroclor 1248)	1 ug/mL	Tetrachloro-m-xylene (S)	.15 ug/mL

Composed of Information for Standard #185551

Composed of Standard Seq	Volume	Units
185529	1	mL
185524	1500	uL
184640	97.5	mL

Standard Log Report

Standards Log Information for Standard #185550

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1248-CAL4

Notes:

1248 Cal 4 Std @ 0.8 ug/mL

Compound Name and Concentration for Standard #185550

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.1 ug/mL	Hexanes	ug/mL
PCB-1248 (Aroclor 1248)	.8 ug/mL	Tetrachloro-m-xylene (S)	.1 ug/mL

Composed of Information for Standard #185550

Composed of Standard Seq	Volume	Units
185529	800	uL
185524	1	mL
184640	98.2	mL

Standard Log Report

Standards Log Information for Standard #185549

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1248-CAL3

Notes:

1248 Cal 3 Std @ 0.5 ug/mL

Compound Name and Concentration for Standard #185549

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.05 ug/mL	Hexanes	ug/mL
PCB-1248 (Aroclor 1248)	.5 ug/mL	Tetrachloro-m-xylene (S)	.05 ug/mL

Composed of Information for Standard #185549

Composed of Standard Seq	Volume	Units
185529	500	uL
185524	500	uL
184640	99	mL

Standard Log Report

Standards Log Information for Standard #185548

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1248-CAL2

Notes:

1248 Cal 2 Std @ 0.2 ug/mL

Compound Name and Concentration for Standard #185548

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.02 ug/mL	Hexanes	ug/mL
PCB-1248 (Aroclor 1248)	.2 ug/mL	Tetrachloro-m-xylene (S)	.02 ug/mL

Composed of Information for Standard #185548

Composed of Standard Seq	Volume	Units
185529	200	uL
185524	200	uL
184640	99.6	mL

Standard Log Report

Standards Log Information for Standard #185546

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1242-CAL5

Notes:

1242 Cal 5 Std @ 1.0 ug/mL

Compound Name and Concentration for Standard #185546

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.15 ug/mL	Hexanes	ug/mL
PCB-1242 (Aroclor 1242)	1 ug/mL	PCB-1242 (Aroclor 1242)	1 ug/mL
PCB-1242 (Aroclor 1242)-6	1 ug/mL	PCB-1242 (Aroclor 1242)-7	1 ug/mL
PCB-1242 (Aroclor 1242)-8	1 ug/mL	PCB-1242 (Aroclor 1242)-9	1 ug/mL
Tetrachloro-m-xylene (S)	.15 ug/mL		

Composed of Information for Standard #185546

Composed of Standard Seq	Volume	Units
185528	1000	uL
185524	1500	uL
184640	97.5	mL

Standard Log Report

Standards Log Information for Standard #185545

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1242-CAL4

Notes:

1242 Cal 4 Std @ 0.8 ug/mL

Compound Name and Concentration for Standard #185545

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.1 ug/mL	Hexanes	ug/mL
PCB-1242 (Aroclor 1242)	.8 ug/mL	PCB-1242 (Aroclor 1242)	.8 ug/mL
PCB-1242 (Aroclor 1242)-6	.8 ug/mL	PCB-1242 (Aroclor 1242)-7	.8 ug/mL
PCB-1242 (Aroclor 1242)-8	.8 ug/mL	PCB-1242 (Aroclor 1242)-9	.8 ug/mL
Tetrachloro-m-xylene (S)	.1 ug/mL		

Composed of Information for Standard #185545

Composed of Standard Seq	Volume	Units
185528	800	uL
185524	1000	uL
184640	98.2	mL

Standard Log Report

Standards Log Information for Standard #185544

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	200 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1242-CAL3

Notes:

1242 Cal 3 Std @ 0.5 ug/mL

Compound Name and Concentration for Standard #185544

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.05 ug/mL	Hexanes	ug/mL
PCB-1242 (Aroclor 1242)	.5 ug/mL	PCB-1242 (Aroclor 1242)	.5 ug/mL
PCB-1242 (Aroclor 1242)-6	.5 ug/mL	PCB-1242 (Aroclor 1242)-7	.5 ug/mL
PCB-1242 (Aroclor 1242)-8	.5 ug/mL	PCB-1242 (Aroclor 1242)-9	.5 ug/mL
Tetrachloro-m-xylene (S)	.05 ug/mL		

Composed of Information for Standard #185544

Composed of Standard Seq	Volume	Units
185528	1000	uL
185524	1000	uL
184640	198	mL

Standard Log Report

Standards Log Information for Standard #185543

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1242-CAL2

Notes:

1242 Cal 2 Std @ 0.2 ug/mL

Compound Name and Concentration for Standard #185543

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.02 ug/mL	Hexanes	ug/mL
PCB-1242 (Aroclor 1242)	.2 ug/mL	PCB-1242 (Aroclor 1242)	.2 ug/mL
PCB-1242 (Aroclor 1242)-6	.2 ug/mL	PCB-1242 (Aroclor 1242)-7	.2 ug/mL
PCB-1242 (Aroclor 1242)-8	.2 ug/mL	PCB-1242 (Aroclor 1242)-9	.2 ug/mL
Tetrachloro-m-xylene (S)	.02 ug/mL		

Composed of Information for Standard #185543

Composed of Standard Seq	Volume	Units
185528	200	uL
185524	200	uL
184640	99.6	mL

Standard Log Report

Standards Log Information for Standard #185539

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	100 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	06/27/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1232-CAL3

Notes:

1232 Cal 3 Std @ 0.5 ug/mL

Compound Name and Concentration for Standard #185539

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	.05 ug/mL	Hexanes	ug/mL
PCB-1232 (Aroclor 1232)	.5 ug/mL	Tetrachloro-m-xylene (S)	.05 ug/mL

Composed of Information for Standard #185539

Composed of Standard Seq	Volume	Units
185527	500	uL
185524	500	uL
184640	99	mL

Standard Log Report

Standards Log Information for Standard #185537

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	10 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	O2Si	**Part ID:**	N/A
Expire Date:	08/09/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1660-STK

Notes:

1660 ICV STK Std @ 100 ug/mL

Compound Name and Concentration for Standard #185537

Compound Name	Concentration	Compound Name	Concentration
Hexanes	ug/mL	PCB-1016 (Aroclor 1016)	100 ug/mL
PCB-1260 (Aroclor 1260)	100 ug/mL		

Composed of Information for Standard #185537

Composed of Standard Seq	Volume	Units
184640	9	mL
175600	1	mL

The standards listed are from Pace's Green Bay Laboratory.

Standard Log Report

Standards Log Information for Standard #185536

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	10 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	O2Si	**Part ID:**	N/A
Expire Date:	12/15/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1254-STK

Notes:

1254 ICV STK Std @ 100 ug/mL

Compound Name and Concentration for Standard #185536

Compound Name	Concentration	Compound Name	Concentration
Hexanes	ug/mL	PCB-1254 (Aroclor 1254)	100 ug/mL

Composed of Information for Standard #185536

Composed of Standard Seq	Volume	Units
184640	9	mL
173678	1000	uL

Standard Log Report

Standards Log Information for Standard #185535

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	10 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	O2Si	**Part ID:**	N/A
Expire Date:	12/15/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1248-STK

Notes:

1248 ICV STK Std @ 100 ug/mL

Compound Name and Concentration for Standard #185535

Compound Name	Concentration	Compound Name	Concentration
Hexanes	ug/mL	PCB-1248 (Aroclor 1248)	100 ug/mL

Composed of Information for Standard #185535

Composed of Standard Seq	Volume	Units
184640	9	mL
173676	1000	uL

Standard Log Report

Standards Log Information for Standard #185534

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	10 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	O2Si	**Part ID:**	N/A
Expire Date:	12/15/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1242-STK

Notes:

1242 ICV STK Std @ 100 ug/mL

Compound Name and Concentration for Standard #185534

Compound Name	Concentration	Compound Name	Concentration
Hexanes	ug/mL	PCB-1242 (Aroclor 1242)	100 ug/mL

Composed of Information for Standard #185534

Composed of Standard Seq	Volume	Units
184640	9	mL
173674	1000	uL

Standard Log Report

Standards Log Information for Standard #185533

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	10 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	12/15/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1660-STK

Notes:

1660 STK Std @ 100 ug/mL

Compound Name and Concentration for Standard #185533

Compound Name	Concentration	Compound Name	Concentration
Hexanes	ug/mL	PCB-1016 (Aroclor 1016)	100 ug/mL
PCB-1260 (Aroclor 1260)	100 ug/mL		

Composed of Information for Standard #185533

Composed of Standard Seq	Volume	Units
184640	9	mL
173188	1	mL

The standards listed are from Pace's Green Bay Laboratory.

Standard Log Report

Standards Log Information for Standard #185530

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	10 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	12/15/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1254-STK

Notes:

1254 STK Std @ 100 ug/mL

Compound Name and Concentration for Standard #185530

Compound Name	Concentration	Compound Name	Concentration
Hexanes	ug/mL	PCB-1254 (Aroclor 1254)	100 ug/mL

Composed of Information for Standard #185530

Composed of Standard Seq	Volume	Units
184640	9	mL
173182	1000	uL

Standard Log Report

Standards Log Information for Standard #185529

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	10 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	12/15/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1248-STK

Notes:

1248 STK Std @ 100 ug/mL

Compound Name and Concentration for Standard #185529

Compound Name	Concentration	Compound Name	Concentration
Hexanes	ug/mL	PCB-1248 (Aroclor 1248)	100 ug/mL

Composed of Information for Standard #185529

Composed of Standard Seq	Volume	Units
184640	9	mL
173180	1000	uL

Standard Log Report

Standards Log Information for Standard #185528

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	10 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	12/15/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1242-STK

Notes:

1242 STK Std @ 100 ug/mL

Compound Name and Concentration for Standard #185528

Compound Name	Concentration	Compound Name	Concentration
Hexanes	ug/mL	PCB-1242 (Aroclor 1242)	100 ug/mL
PCB-1242 (Aroclor 1242)	100 ug/mL	PCB-1242 (Aroclor 1242)-6	100 ug/mL
PCB-1242 (Aroclor 1242)-7	100 ug/mL	PCB-1242 (Aroclor 1242)-8	100 ug/mL
PCB-1242 (Aroclor 1242)-9	100 ug/mL		

Composed of Information for Standard #185528

Composed of Standard Seq	Volume	Units
184640	9	mL
173178	1000	uL

Standard Log Report

Standards Log Information for Standard #185527

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	10 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	12/15/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	1232-STK

Notes:

1232 STK Std @ 100 ug/mL

Compound Name and Concentration for Standard #185527

Compound Name	Concentration	Compound Name	Concentration
Hexanes	ug/mL	PCB-1232 (Aroclor 1232)	100 ug/mL

Composed of Information for Standard #185527

Composed of Standard Seq	Volume	Units
184640	9	mL
173176	1000	uL

Standard Log Report

Standards Log Information for Standard #185524

WORKING STANDARD

Created By:	BDS	**Volume of Standard:**	20 mL	**Lot ID:**	GCSV
Create Date:	12/27/2017	**Manufacturer:**	Restek	**Part ID:**	N/A
Expire Date:	12/15/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	PEST-SUR

Notes:

Pest/PCB Surr Std @ 10 ug/mL

Compound Name and Concentration for Standard #185524

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	10 ug/mL	Hexanes	ug/mL
Tetrachloro-m-xylene (S)	10 ug/mL		

Composed of Information for Standard #185524

Composed of Standard Seq	Volume	Units
184640	19	mL
178398	1	mL

Standard Log Report

Standards Log Information for Standard #184640

STOCK STANDARD

Created By:	AMC	**Volume of Standard:**	215 L	**Lot ID:**	405661
Create Date:	12/12/2017	**Manufacturer:**	J.T. Baker	**Part ID:**	9262-8P
Expire Date:	12/15/2018	**Manufacturer Lot ID:**	0000183983	**Standard ID:**	RGNT150

Notes:

Hexane SVOA TANK Changed AMC 12/12/17

Compound Name and Concentration for Standard #184640

Compound Name	Concentration	Compound Name	Concentration
Hexanes	ug/mL		

Standard Log Report

Standards Log Information for Standard #182336

WORKING STANDARD

Created By:	ASD	**Volume of Standard:**	200 mL	**Lot ID:**	GCSV
Create Date:	11/09/2017	**Manufacturer:**	N/A	**Part ID:**	N/A
Expire Date:	03/30/2018	**Manufacturer Lot ID:**	N/A	**Standard ID:**	AR1242-SPK

Notes:

Ar1242 Spike @20ug/mL 40GCS9.i\111017spk.b\A110047 96% Recovery

Compound Name and Concentration for Standard #182336

Compound Name	Concentration	Compound Name	Concentration
Acetone	ug/mL	PCB-1242 (Aroclor 1242)	20 ug/mL

Composed of Information for Standard #182336

Composed of Standard Seq	Volume	Units
175728	196	mL
164832	4000	uL

Standard Log Report

Standards Log Information for Standard #179816

INDIVIDUAL STANDARD

Created By:	MRN	**Volume of Standard:**	1 mL	**Lot ID:**	405540
Create Date:	10/09/2017	**Manufacturer:**	o2si	**Part ID:**	030273-11
Expire Date:	09/01/2021	**Manufacturer Lot ID:**	316120	**Standard ID:**	1221-STK

Notes:

1221 STK Std @ 1000 ug/mL

Compound Name and Concentration for Standard #179816

Compound Name	Concentration	Compound Name	Concentration
PCB-1221 (Aroclor 1221)	1000 ug/mL		

Standard Log Report

Standards Log Information for Standard #179366

STOCK STANDARD

Created By:	MRN	**Volume of Standard:**	4 L	**Lot ID:**	405532
Create Date:	10/02/2017	**Manufacturer:**	Burdick & Jackson	**Part ID:**	010-4
Expire Date:	01/08/2019	**Manufacturer Lot ID:**	DS490	**Standard ID:**	RGNT036

Notes:

Acetone - 4L container

Compound Name and Concentration for Standard #179366

Compound Name	Concentration	Compound Name	Concentration
Acetone	ug/mL		

The standards listed are from Pace's Green Bay Laboratory.

Standard Log Report

Standards Log Information for Standard #178398

INDIVIDUAL STANDARD

Created By:	DMH	**Volume of Standard:**	1 mL	**Lot ID:**	405520
Create Date:	09/18/2017	**Manufacturer:**	Restek	**Part ID:**	32000
Expire Date:	06/30/2023	**Manufacturer Lot ID:**	A0125833	**Standard ID:**	PEST-SUR

Notes:

Pest/PCB Surr @ 200 ug/mL

Compound Name and Concentration for Standard #178398

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	200 ug/mL	Tetrachloro-m-xylene (S)	200 ug/mL

Standard Log Report

Standards Log Information for Standard #175728

STOCK STANDARD

Created By:	MRN	**Volume of Standard:**	4 L	**Lot ID:**	405434
Create Date:	08/10/2017	**Manufacturer:**	Burdick & Jackson	**Part ID:**	010-4
Expire Date:	11/10/2018	**Manufacturer Lot ID:**	DS178	**Standard ID:**	RGNT036

Notes:

Acetone - 4L container -

Compound Name and Concentration for Standard #175728

Compound Name	Concentration	Compound Name	Concentration
Acetone	ug/mL		

The standards listed are from Pace's Green Bay Laboratory.

Standard Log Report

Standards Log Information for Standard #175600

INDIVIDUAL STANDARD

Created By:	BDS	**Volume of Standard:**	1 mL	**Lot ID:**	405431
Create Date:	08/09/2017	**Manufacturer:**	O2SI	**Part ID:**	130011-03
Expire Date:	08/09/2018	**Manufacturer Lot ID:**	311587	**Standard ID:**	1660-STK

Notes:

1016/1260 STK Std @ 1000 ug/mL

Compound Name and Concentration for Standard #175600

Compound Name	Concentration	Compound Name	Concentration
PCB-1016 (Aroclor 1016)	1000 ug/mL	PCB-1260 (Aroclor 1260)	1000 ug/mL

Standard Log Report

Standards Log Information for Standard #173678

INDIVIDUAL STANDARD

Created By:	BDS	**Volume of Standard:**	1 mL	**Lot ID:**	405386
Create Date:	07/17/2017	**Manufacturer:**	O2Si	**Part ID:**	030277-06
Expire Date:	03/21/2020	**Manufacturer Lot ID:**	302980	**Standard ID:**	1254-STK

Notes:

1254 ICV STK Std @ 1000 ug/mL

Compound Name and Concentration for Standard #173678

Compound Name	Concentration	Compound Name	Concentration
PCB-1254 (Aroclor 1254)	1000 ug/mL		

Standard Log Report

Standards Log Information for Standard #173676

INDIVIDUAL STANDARD

Created By:	BDS	**Volume of Standard:**	1 mL	**Lot ID:**	405385
Create Date:	07/17/2017	**Manufacturer:**	O2Si	**Part ID:**	030276-15
Expire Date:	01/14/2020	**Manufacturer Lot ID:**	296389	**Standard ID:**	1248-STK

Notes:

1248 ICV STK Std @ 1000 ug/mL

Compound Name and Concentration for Standard #173676

Compound Name	Concentration	Compound Name	Concentration
PCB-1248 (Aroclor 1248)	1000 ug/mL		

The standards listed are from Pace's Green Bay Laboratory.

Standard Log Report

Standards Log Information for Standard #173674

INDIVIDUAL STANDARD

Created By:	BDS	**Volume of Standard:**	1 mL	**Lot ID:**	405384
Create Date:	07/17/2017	**Manufacturer:**	O2Si	**Part ID:**	030275-03
Expire Date:	03/20/2027	**Manufacturer Lot ID:**	302979	**Standard ID:**	1242-STK

Notes:

1242 ICV STK Std @ 1000 ug/mL

Compound Name and Concentration for Standard #173674

Compound Name	Concentration	Compound Name	Concentration
PCB-1242 (Aroclor 1242)	1000 ug/mL		

Standard Log Report

Standards Log Information for Standard #173188

INDIVIDUAL STANDARD

Created By:	BDS	**Volume of Standard:**	1 mL	**Lot ID:**	405375
Create Date:	07/07/2017	**Manufacturer:**	Restek	**Part ID:**	32039
Expire Date:	04/30/2023	**Manufacturer Lot ID:**	A0124547	**Standard ID:**	1660-STK

Notes:

Ar1660 STK Std @ 1000 ug/mL

Compound Name and Concentration for Standard #173188

Compound Name	Concentration	Compound Name	Concentration
PCB-1016 (Aroclor 1016)	1000 ug/mL	PCB-1260 (Aroclor 1260)	1000 ug/mL

Standard Log Report

Standards Log Information for Standard #173182

INDIVIDUAL STANDARD

Created By:	BDS	**Volume of Standard:**	1 mL	**Lot ID:**	405372
Create Date:	07/07/2017	**Manufacturer:**	Restek	**Part ID:**	32011
Expire Date:	05/31/2023	**Manufacturer Lot ID:**	A0125350	**Standard ID:**	1254-STK

Notes:

1254 STK Std @ 1000 ug/mL L

Compound Name and Concentration for Standard #173182

Compound Name	Concentration	Compound Name	Concentration
PCB-1254 (Aroclor 1254)	1000 ug/mL		

The standards listed are from Pace's Green Bay Laboratory.

Standard Log Report

Standards Log Information for Standard #173180

INDIVIDUAL STANDARD

Created By:	BDS	**Volume of Standard:**	1 mL	**Lot ID:**	405371
Create Date:	07/07/2017	**Manufacturer:**	Restek	**Part ID:**	32010
Expire Date:	12/31/2022	**Manufacturer Lot ID:**	A0121842	**Standard ID:**	1248-STK

Notes:

1248 STK Std @ 1000 ug/mL

Compound Name and Concentration for Standard #173180

Compound Name	Concentration	Compound Name	Concentration
PCB-1248 (Aroclor 1248)	1000 ug/mL		

Standard Log Report

Standards Log Information for Standard #173178

INDIVIDUAL STANDARD

Created By:	BDS	**Volume of Standard:**	1 mL	**Lot ID:**	405370
Create Date:	07/07/2017	**Manufacturer:**	Restek	**Part ID:**	32009
Expire Date:	04/30/2023	**Manufacturer Lot ID:**	A0124535	**Standard ID:**	1242-STK

Notes:

1242 STK Std @ 1000 ug/mL

Compound Name and Concentration for Standard #173178

Compound Name	Concentration	Compound Name	Concentration
PCB-1242 (Aroclor 1242)	1000 ug/mL	PCB-1242 (Aroclor 1242)	1000 ug/mL
PCB-1242 (Aroclor 1242)-6	1000 ug/mL	PCB-1242 (Aroclor 1242)-7	1000 ug/mL
PCB-1242 (Aroclor 1242)-8	1000 ug/mL	PCB-1242 (Aroclor 1242)-9	1000 ug/mL

Standard Log Report

Standards Log Information for Standard #173176

INDIVIDUAL STANDARD

Created By:	BDS	**Volume of Standard:**	1 mL	**Lot ID:**	405369
Create Date:	07/07/2017	**Manufacturer:**	Restek	**Part ID:**	32008
Expire Date:	04/30/2023	**Manufacturer Lot ID:**	A0124534	**Standard ID:**	1232-STK

Notes:

1232 STK Std @ 1000 ug/mL

Compound Name and Concentration for Standard #173176

Compound Name	Concentration	Compound Name	Concentration
PCB-1232 (Aroclor 1232)	1000 ug/mL		

The standards listed are from Pace's Green Bay Laboratory.

Standard Log Report

Standards Log Information for Standard #172034

INDIVIDUAL STANDARD

Created By:	MRN	**Volume of Standard:**	10 mL	**Lot ID:**	405341
Create Date:	06/23/2017	**Manufacturer:**	O2SI	**Part ID:**	130023-15
Expire Date:	11/06/2018	**Manufacturer Lot ID:**	311236	**Standard ID:**	PEST-SUR

Notes:

Pest/PCB Surr @ 2000 ug/mL - opened 11/6/17 MRN

Compound Name and Concentration for Standard #172034

Compound Name	Concentration	Compound Name	Concentration
Decachlorobiphenyl (S)	2000 ug/mL	Tetrachloro-m-xylene (S)	2000 ug/mL

Standard Log Report

Standards Log Information for Standard #164832

INDIVIDUAL STANDARD

Created By:	MRN	**Volume of Standard:**	25 mL	**Lot ID:**	405214
Create Date:	03/27/2017	**Manufacturer:**	O2SI	**Part ID:**	030275-07
Expire Date:	03/30/2018	**Manufacturer Lot ID:**	302694	**Standard ID:**	1242-STK

Notes:

1242 STK Std @ 1000 ug/mL

Compound Name and Concentration for Standard #164832

Compound Name	Concentration	Compound Name	Concentration
PCB-1242 (Aroclor 1242)	1000 ug/mL		

The standards listed are from Pace's Green Bay Laboratory.

405214
REC. 03/27/17 BY MRM
Exp. __/__/__ By __

2030 Savage Road
Charleston, South Carolina 29407
Phone: 866.272.0932
Fax: 866.509.5146
www.o2si.com

o2si smart solutions®

Quality System Audited & Registered by NSF-ISR to ISO 9001:2008

ACCREDITED

ISO 17025 Accredited Chemical Testing Lab Cert. No. 3031.01

Date Received: _____

Certificate of Analysis

Rev 0 **Page 1**

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
030275-07	302694	≤ Ambient	Tol.:Hex. 1:1	15-Jan-2021	Aroclor 1242 Solution, 1,000 mg/L, 25 ml

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration	
Aroclor 1242	53469-21-9	100	275.3.5.2P	1002 +/- 5.75	mg/L

Certified By: _____

Dreema Murray

Manufacture Date 17-Mar-2017

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetriclly.

405341
REC. 06/23/17 BY MRN
Exp. --/--/-- By --

2030 Savage Road
Charleston, South Carolina 29407
Phone: 866.272.0932
Fax: 866.509.5146
www.o2si.com

o2si smart solutions®

Quality System Audited & Registered by NSF-ISR to ISO 9001:2008

ACCREDITED

ISO 17025 Accredited Chemical Testing Lab Cert. No. 3031.01

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
130023-15	311236	≤ 6 °C	Tol.:Hex. 1:1	8-Jun-2019	Pesticide Surrogate Solution, 2,000 mg/L, 10 mL

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration	
decachlorobiphenyl (BZ# 209)	2051-24-3	99.1	301.115.5P	2022 +/- 4.56	mg/L
2,4,5,6-tetrachloro-m-xylene	877-09-8	98.6	300.115.3P	1996 +/- 14.54	mg/L

Certified By: Kaylie Gillenkirk

Kaylie Gillenkirk

Manufacture Date 8-Jun-2017

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ISO Guide 34 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008 **Lot No.:** A0124534

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2023 **Storage:** 25°C nominal

Handling: This product contains PCBs.

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Aroclor 1232	1,000.0 µg/mL	+/- 5.9397	µg/mL	Gravimetric
	CAS # 11141-16-5 (Lot W-107-05)		+/- 31.7072	µg/mL	Unstressed
	Purity ----%		+/- 41.4130	µg/mL	Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

405369
REC.07/06/17 BY BDS
Exp. __/__/__ By --
Open __/__/__ By --

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Cathleen Soltis - Mix Technician

Date Mixed: 27-Jan-2017 **Balance:** B442140311

Justine Albertson - Operations Tech.ARM QC

Date Passed: 31-Jan-2017

Manufactured under Restek's ISO 9001:2008
Registered Quality System
Certificate #FM 80397

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ISO Guide 34 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32009 **Lot No.:** A0124535

Description : Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2023 **Storage:** 25°C nominal

Handling: This product contains PCBs.

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Aroclor 1242	1,002.0 µg/mL	+/- 5.9516 µg/mL	Gravimetric
	CAS # 53469-21-9 (Lot 01141-A)		+/- 31.7706 µg/mL	Unstressed
	Purity ----%		+/- 41.4958 µg/mL	Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

405370
REC.07/06/17 BY BDS
Exp. __/__/__ By __
Open __/__/__ By __

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Cathleen Soltis - Mix Technician

Date Mixed: 27-Jan-2017 **Balance:** B442140311

Justine Albertson - Operations Tech-ARM QC

Date Passed: 31-Jan-2017

Manufactured under Restek's ISO 9001:2008
Registered Quality System
Certificate #FM 80397

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO Guide 34 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32010 **Lot No.:** A0121842

Description : Aroclor® 1248 Standard

Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2022 **Storage:** 25°C nominal

Handling: This product contains PCBs.

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Aroclor 1248	996.8 µg/mL	+/- 5.8498 µg/mL	Gravimetric
	CAS # 12672-29-6 (Lot 07)		+/- 31.5925 µg/mL	Unstressed
	Purity ----%		+/- 41.2704 µg/mL	Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

405371
REC.07/06/17 BY BDS
Exp. __/__/__ By --

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Dawn Brownson - Mix Technician

Date Mixed: 29-Sep-2016 **Balance:** 1128360905

Jennifer L. Pollino - QC Analyst

Date Passed: 03-Oct-2016

Manufactured under Restek's ISO 9001:2008
Registered Quality System
Certificate #FM 80397

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis

ACCREDITED
ISO Guide 34 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0125350

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2023 **Storage:** 25°C nominal

Handling: This product contains PCBs.

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Aroclor 1254		1,005.6 µg/mL	+/- 5.9014	µg/mL	Gravimetric
	CAS # 11097-69-1	(Lot 124-191-B)		+/- 31.8715	µg/mL	Unstressed
	Purity ----%			+/- 41.6347	µg/mL	Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

405372
REC.07/06/17 BY BDS
Exp. __/__/__ By __

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Dawn Brownson - Mix Technician **Date Mixed:** 24-Feb-2017 **Balance:** 1128360905

Jennifer Pollino - Operations Tech-ARM QC **Date Passed:** 13-Mar-2017

Manufactured under Restek's ISO 9001:2008
Registered Quality System
Certificate #FM 80397

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ISO Guide 34 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0124547

Description : Aroclor® 1016/1260 Mix

Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2023 **Storage:** 25°C nominal

Handling: This product contains PCBs.

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Aroclor 1016		1,002.0 µg/mL	+/- 5.8392	µg/mL	Gravimetric
	CAS # 12674-11-2	(Lot W-125-04)		+/- 31.7498	µg/mL	Unstressed
	Purity ----%			+/- 41.4799	µg/mL	Stressed
2	Aroclor 1260		1,002.2 µg/mL	+/- 5.8403	µg/mL	Gravimetric
	CAS # 11096-82-5	(Lot W-129-06)		+/- 31.7561	µg/mL	Unstressed
	Purity ----%			+/- 41.4881	µg/mL	Stressed

Solvent: Hexane
CAS # 110-54-3
Purity 99%

405375
REC. 07/06/17 BY BDS
Exp. --/--/-- By --
Open / / By

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Cathleen Soltis - Mix Technician

Date Mixed: 30-Jan-2017 **Balance:** B442140311

Jennifer Pollino - Operations Tech-ARM QC

Date Passed: 01-Feb-2017

Manufactured under Restek's ISO 9001:2008
Registered Quality System
Certificate #FM 80397

o2si smart solutions®

2030 Savage Road
Charleston, South Carolina 29407
Phone: 866.272.0932
Fax: 866.509.5146
www.o2si.com

Quality System Audited & Registered by NSF-ISR to ISO 9001:2008

ACCREDITED

ISO 17025 Accredited Chemical Testing Lab
Cert. No. 3031.01

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
030275-03	302979	Ambient or Less	Hexane	20-Mar-2027	Aroclor 1242 Solution, 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration	
Aroclor 1242	53469-21-9	100	275.3.5.2P	1010 +/- 7.38	mg/L

405384
REC.07/13/17 BY BDS
Exp. __/__/__ By --
Open / / By

Certified By: _____

Katrina Emelianoff

Manufacture Date 22-Mar-2017

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

o2si smart solutions®

2030 Savage Road
Charleston, South Carolina 29407
Phone: 866.272.0932
Fax: 866.509.5146
www.o2si.com

Quality System Audited & Registered by NSF-ISR to ISO 9001:2008

ACCREDITED

ISO 17025 Accredited Chemical Testing Lab
Cert. No. 3031.01

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
030276-15	296389	≤ Ambient	Hexane	14-Jan-2020	Aroclor 1248 Solution, 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration	
Aroclor 1248	12672-29-6	100	276.3.1P	1008 +/- 4.47	mg/L

405385
REC. 07/13/17 BY BDS
Exp. --/--/-- By --

Certified By: _____
Katrina Emelianoff
Manufacture Date 9-Jan-2017

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

o2si smart solutions®

2030 Savage Road
Charleston, South Carolina 29407
Phone: 866.272.0932
Fax: 866.509.5146
www.o2si.com

Quality System Audited & Registered by NSF-ISR to ISO 9001:2008

ACCREDITED

ISO 17025 Accredited Chemical Testing Lab Cert. No. 3031.01

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
030277-06	302980	≤ Ambient	Hexane	21-Mar-2020	Aroclor 1254 Solution, 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration	
Aroclor 1254	11097-69-1	100	277.6.4P	1004 +/- 4.32	mg/L

405386
REC. 07/13/17 BY BDS
Exp. __/__/__ By __
Open __/__/__ By

Certified By: _____

Katrina Emelianoff

Manufacture Date 22-Mar-2017

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

o2si smart solutions®

2030 Savage Road
Charleston, South Carolina 29407
Phone: 866.272.0932
Fax: 866.509.5146
www.o2si.com

Quality System Audited & Registered by NSF-ISR to ISO 9001:2008

ACCREDITED

ISO 17025 Accredited Chemical Testing Lab
Cert. No. 3031.01

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
130011-03	311587	≤ Ambient	Hexane	28-Jul-2020	Aroclor 1016 + 1260 Solution, 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration	
Aroclor 1016	12674-11-2	100	272.5.1.6P	1006 +/- 3.01	mg/L
Aroclor 1260	11096-82-5	100	278.3.10P	1016 +/- 3.02	mg/L

405431
REC. 08/07/17 BY BDS
Exp. --/--/-- By --

Certified By: _____
Christina Elder

Manufacture Date 24-Jul-2017

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetriclly.

405434
REC 08/10/17 BY MRN
Exp. --/--/-- By --
Open --/--/-- By --

Honeywell

CERTIFICATE OF ANALYSIS

Honeywell Burdick & Jackson®
1953 South Harvey Street
Muskegon, MI 49442
Phone: (800) 368-0050
Fax: (231) 728-8226
www.lab-honeywell.com

Brand:	Research Chemicals - B&J
Product:	010
Lot No.:	DS178
Production Date:	12-May-2017
Best Before:	10-Nov-2018

Acetone, B&J Brand™, >99.9%
for HPLC, GC, pesticide residue analysis and spectrophotometry

Parameter	Specification Min.	Max.	Result	Units
Water by Karl Fischer Titration		0.50	0.29	%
UV Cutoff		330	328	nm
Refractive Index (20°C)	1.3583	1.3589	1.3586	
Residue		1	<0.5	mg/L
GC Analysis (excluding water)	99.9		99.95	%
Electron Capture GC		10	<10	ng/L
UV Absorbance @ 340 nm		0.060	0.0527	AU
UV Absorbance @ 350 nm		0.010	0.0048	AU
UV Absorbance @ 375 nm		0.005	0.0013	AU
UV Absorbance @ 400 nm		0.005	<0.0001	AU

Honeywell
Quality Control Approval

Janna Dickinson

Muskegon 5/12/2017 LIMS Sample No.: AI06733

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis

ACCREDITED
ISO Guide 34 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0125833

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2023 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use.

DMH 9-18-17

405520
REC.09/18/17 BY DMH
Exp. __/__/__ By --
Open __/__/__ By

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2,4,5,6-Tetrachloro-m-xylene		201.5 µg/mL	+/- 1.1742	µg/mL	Gravimetric
	CAS # 877-09-8	(Lot 0052481)		+/- 6.3844	µg/mL	Unstressed
	Purity 98%			+/- 8.3410	µg/mL	Stressed
2	Decachlorobiphenyl (BZ# 209)		201.9 µg/mL	+/- 1.1766	µg/mL	Gravimetric
	CAS # 2051-24-3	(Lot ER071509-01)		+/- 6.3975	µg/mL	Unstressed
	Purity 99%			+/- 8.3581	µg/mL	Stressed

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Matt Fragassi - Mix Technician

Date Mixed: 15-Mar-2017 **Balance:** B345965662

Diane Shaffer - Operations Tech-ARM QC

Date Passed: 17-Mar-2017

Manufactured under Restek's ISO 9001:2008 Registered Quality System Certificate #FM 80397

405532
REC. 09/29/17 BY MRN
Exp. --/--/-- By --

Honeywell

CERTIFICATE OF ANALYSIS

Honeywell Burdick & Jackson®
1953 South Harvey Street
Muskegon, MI 49442
Phone: (800) 368-0050
Fax: (231) 728-8226
www.lab-honeywell.com

Brand:	Research Chemicals - B&J
Product:	010
Lot No.:	DS490
Production Date:	10-Jul-2017
Best Before:	8-Jan-2019

Acetone, B&J Brand™, >99.9%

for HPLC, GC, pesticide residue analysis and spectrophotometry

Parameter	Specification Min.	Max.	Result	Units
Water by Karl Fischer Titration		0.50	0.31	%
UV Cutoff		330	328	nm
Refractive Index (20°C)	1.3583	1.3589	1.3586	
Residue		1	<0.5	mg/L
GC Analysis (excluding water)	99.9		99.97	%
Electron Capture GC		10	<10	ng/L
UV Absorbance @ 340 nm		0.060	0.0481	AU
UV Absorbance @ 350 nm		0.010	0.0014	AU
UV Absorbance @ 375 nm		0.005	<0.0001	AU
UV Absorbance @ 400 nm		0.005	<0.0001	AU

Honeywell
Quality Control Approval

Janna Dickinson

Muskegon 7/11/2017 LIMS Sample No.: AI07427

405540
REC. 10/09/17 BY MRI
Exp. --/--/-- By --
Open / / By

o2si smart solutions®

Quality System Audited & Registered by NSF-ISR to ISO 9001:2008

2030 Savage Road
Charleston, South Carolina 29407
Phone: 866.272.0932
Fax: 866.509.5146
www.o2si.com

ACCREDITED

ISO 17025 Accredited Chemical Testing Lab
Cert. No. 3031.01

Date Received: _____

Certificate of Analysis

Rev 0 **Page 1 of 1**

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
030273-11	316120	≤ Ambient	Hexane	1-Sep-2021	Aroclor 1221 Solution, 1,000 mg/L, 1 ml

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration	
Aroclor 1221	11104-28-2	100	273.3.2P	1018 +/- 3.01	mg/L

Certified By: _____
Kara Catron
Manufacture Date 24-Jul-2017

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

Sodium Sulfate, Anhydrous, Granular (12-60 Mesh)
ULTRA RESI-ANALYZED®, ACS Reagent
Tested for Hexane Extractable Impurities
Suitable for Use in Pesticide Residue Analysis

405550
REC. 10/16/17 BY MRN
Exp. --/--/-- By --
Gen. / / By

J.T.Baker® | avantor™

Material No.: 3375-09
Batch No.: 0000177456
Manufactured Date: 2017/05/23
Retest Date: 2022/05/22

Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
ACS - Assay (Na_2SO_4)	>= 99.0 %	99.2
ACS - Insoluble Matter	<= 0.01 %	< 0.01
ACS - Loss on Ignition	<= 0.5 %	0.1
ACS - pH of 5% Solution at 25°C	5.2 - 9.2	6.0
ACS - Chloride (Cl)	<= 0.001 %	< 0.001
ACS - Phosphate (PO_4)	<= 0.001 %	< 0.001
ACS - Calcium (Ca)	<= 0.01 %	< 0.01
ACS - Iron (Fe)	<= 0.001 %	< 0.001
ACS - Magnesium (Mg)	<= 0.005 %	< 0.005
ACS - Potassium (K)	<= 0.01 %	< 0.01
ACS - Extraction-Concentration Suitability *	Passes Test	PT
ACS - Heavy Metals (as Pb)	<= 5 ppm	< 5
Nitrogen Compounds (as N)	<= 5 ppm	< 5

For Laboratory, Research or Manufacturing Use
* Result obtained on a sample dried at 400°C for one hour (ACS method)

Country of Origin: MX
Packaging Site: Paris Mfg Ctr & DC

ISO
Phillipsburg, NJ 9001:2008, 14001:2004, FSSC 22000
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 13485:2012
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008
Panoli, India 9001:2008

James T Ethier
Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor Performance Materials, LLC.
3477 Corporate Parkway. Center Valley, PA 18034. U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610

Hexanes (95% n-hexane)
ULTRA RESI-ANALYZED
For Organic Residue Analysis

405661
REC. 10/26/17 BY AMC
Exp. --/--/-- By --

J.T.Baker | avantor

Material No.: 9262-8P
Batch No.: 0000183983
Manufactured Date: 2017/09/15
Expiration Date: 2018/12/15

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	<= 5	1
Assay (Total Saturated C_6 Isomers) (by GC, corrected for water)	>= 99.5 %	99.8
Assay (as n-Hexane) (by GC, corrected for water)	>= 95 %	98
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.5
Substances Darkened by H_2SO_4	Passes Test	PT
Water (by KF, coulometric)	<= 0.05 %	0.01

For Laboratory, Research or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC

ISO
Phillipsburg, NJ 9001:2008, 14001:2004, FSSC 22000
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 13485:2012
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008
Panoli, India 9001:2008

James Ethier
Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor Performance Materials, LLC.
3477 Corporate Parkway. Center Valley, PA 18034. U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610

RESTEK® CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis

ACCREDITED
ISO Guide 34 Accredited
Reference Material Producer
Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0130424

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2023 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use.

405693
REC.01/02/18 BY BLM
Exp: --/--/-- By --
Open / / By

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	2,4,5,6-Tetrachloro-m-xylene		201.2 µg/mL	+/- 1.1725	µg/mL	Gravimetric
	CAS # 877-09-8	(Lot 0052481)		+/- 6.3751	µg/mL	Unstressed
	Purity 98%			+/- 8.3288	µg/mL	Stressed
2	Decachlorobiphenyl (BZ# 209)		201.5 µg/mL	+/- 1.1742	µg/mL	Gravimetric
	CAS # 2051-24-3	(Lot ER071509-01)		+/- 6.3848	µg/mL	Unstressed
	Purity 99%			+/- 8.3415	µg/mL	Stressed

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 31-Aug-2017 **Balance:** B345965662

Jennifer Pollino - Operations Tech-ARM QC

Date Passed: 07-Sep-2017

Manufactured under Restek's ISO 9001:2008
Registered Quality System
Certificate #FM 80397

Hexanes (95% n-hexane)
ULTRA RESI-ANALYZED
For Organic Residue Analysis

405750
REC. 02/12/18 BY ETB
Exp. --/--/-- By --

J.T.Baker | avantor

Material No.: 9262-8P
Batch No.: 0000187681
Manufactured Date: 2017/10/16
Expiration Date: 2019/01/15

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	5
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	<= 5	< 1
Assay (Total Saturated C_6 Isomers) (by GC, corrected for water)	>= 99.5 %	99.9
Assay (as n-Hexane) (by GC, corrected for water)	>= 95 %	99
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.2
Substances Darkened by H_2SO_4	Passes Test	PT
Water (by KF, coulometric)	<= 0.05 %	0.01

For Laboratory, Research or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC

ISO
Phillipsburg, NJ 9001:2008, 14001:2004, FSSC 22000
Paris, KY 9001:2008
Mexico City, Mexico 9001:2008
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003
Gliwice, Poland 9001:2008, 13485:2012
Selangor, Malaysia 9001:2008
Dehradun, India, 9001:2008, 14001:2004, 13485:2003
Mumbai, India, 9001:2008
Panoli, India 9001:2008

James T Ethier
Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor Performance Materials, LLC.
3477 Corporate Parkway. Center Valley, PA 18034. U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610

o2si smart solutions®

2030 Savage Road
Charleston, South Carolina 29407
Phone: 866.272.0932
Fax: 866.509.5146
www.o2si.com

Quality System Audited & Registered by NSF-ISR to ISO 9001:2008

ISO 17025 Accredited Chemical Testing Lab Cert. No. 3031.01

Date Received: _____

Certificate of Analysis

Rev 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
020986-02 -02-SS	297165	≤ -10 °C	P/T Methanol	16-Jan-2020	Dichlorofluoromethane (Freon 21) Gas Solution, Second Source, 2000 mg/L, 2 x 0.6 ml

Compound	CAS No.	Purity (%)	Neat Material Lot No.	Concentration
dichlorofluoromethane (Freon 21)	75-43-4	99	986.158.2.2.1S	2008 +/- 20.85 mg/L

405766
REC. 02/26/18 BY SMT
Exp. --/--/-- By --
Open / / By

4388-13

189977 – 189980

Certified By: Madeline Luby

Madeline Luby
Manufacture Date 16-Jan-2017

Follow all storage requirements, keep tightly closed when not in use, and use good laboratory practices when handling.
This Reference Material was manufactured, produced, and/or certified under a quality management system that is accredited to ISO 17034 and ISO/IEC 17025.

Second Source standards are prepared according to NELAC

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.
The stated uncertainty is the expanded uncertainty with a coverage factor of two to give a 95% confidence level.

SAMPLE NO.

TD2-170612-01-YB-01

FORM I INORGANIC-1

INORGANIC ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract: HUDSON RIVER REMEDIAL

Lab Sample ID: 40151601001 Percent Moisture:

CAS No.	Analyte	Concentration	Q	Units	DF	Analysis Date/Time
	Lipid	0.47		%	1	03/29/2018 07:08

SAMPLE NO.

TD2-170612-01-SMB-01

FORM I INORGANIC-1

INORGANIC ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract: HUDSON RIVER REMEDIAL

Lab Sample ID: 40151601002 Percent Moisture:

CAS No.	Analyte	Concentration	Q	Units	DF	Analysis Date/Time
	Lipid	0.37		%	1	03/13/2018 08:59

SAMPLE NO.

TD2-170612-01-BB-04

FORM I INORGANIC-1

INORGANIC ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract: HUDSON RIVER REMEDIAL

Lab Sample ID: 40151601003 Percent Moisture:

CAS No.	Analyte	Concentration	Q	Units	DF	Analysis Date/Time
	Lipid	0.90		%	1	03/14/2018 06:49

SAMPLE NO.

TD2-170612-01-BB-03

FORM I INORGANIC-1

INORGANIC ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract: HUDSON RIVER REMEDIAL

Lab Sample ID: 40151601004 Percent Moisture:

CAS No.	Analyte	Concentration	Q	Units	DF	Analysis Date/Time
	Lipid	0.36		%	1	03/14/2018 06:49

SAMPLE NO.

TD2-170612-01-BB-02

FORM I INORGANIC-1

INORGANIC ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract: HUDSON RIVER REMEDIAL

Lab Sample ID: 40151601005 Percent Moisture:

CAS No.	Analyte	Concentration	Q	Units	DF	Analysis Date/Time
	Lipid	0.62		%	1	03/14/2018 06:49

SAMPLE NO.

TD2-170612-01-BB-01

FORM I INORGANIC-1

INORGANIC ANALYSIS DATA SHEET

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract: HUDSON RIVER REMEDIAL

Lab Sample ID: 40151601006 Percent Moisture:

CAS No.	Analyte	Concentration	Q	Units	DF	Analysis Date/Time
	Lipid	1.4		%	1	03/14/2018 06:49

FORM III INORGANIC-1

BLANKS

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract : HUDSON RIVER REMEDIAL ACTION

Method Blank Matrix: Tissue Instrument ID: 40BALL

Method Blank Concentration Units: %

Analyte	Initial Calibration Blank		Continuing Calibration Blank						Method Blank	
		C		C		C		C	1655713	C
Lipid									<0.030	U

FORM III INORGANIC-1

BLANKS

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract : HUDSON RIVER REMEDIAL ACTION

Method Blank Matrix: Tissue Instrument ID: 40BALL

Method Blank Concentration Units: %

Analyte	Initial Calibration Blank		Continuing Calibration Blank						Method Blank	
		C		C		C		C	1658258	C
Lipid									<0.030	U

FORM III INORGANIC-1

BLANKS

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract : HUDSON RIVER REMEDIAL ACTION

Method Blank Matrix: Tissue Instrument ID: 40BALL

Method Blank Concentration Units: %

Analyte	Initial Calibration Blank		Continuing Calibration Blank						Method Blank	
		C		C		C		C	1665662	C
Lipid									<0.030	U

SAMPLE NO.

1655716DUP

FORM VI INORGANIC-1
DUPLICATES

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract: HUDSON RIVER REMEDIAL

Matrix: Tissue Concentration Units: %

Percent Moisture: Basis: Wet

Analyte	Control Limit	Sample	Duplicate	RPD
Lipid	40	3.0	2.1	36

SAMPLE NO.

1655717DUP

FORM VI INORGANIC-2
DUPLICATES

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract: HUDSON RIVER REMEDIAL

Matrix: Tissue Concentration Units: %

Percent Moisture: Basis: Wet

Analyte	Control Limit	Sample	Duplicate	RPD
Lipid	40	0.39	0.41	3

SAMPLE NO.

1658259DUP

FORM VI INORGANIC-1
DUPLICATES

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract: HUDSON RIVER REMEDIAL

Matrix: Tissue Concentration Units: %

Percent Moisture: Basis: Wet

Analyte	Control Limit	Sample	Duplicate	RPD
Lipid	40	1.5	1.5	5

SAMPLE NO.

1665665DUP

FORM VI INORGANIC-1
DUPLICATES

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract: HUDSON RIVER REMEDIAL

Matrix: Tissue Concentration Units: %

Percent Moisture: Basis: Wet

Analyte	Control Limit	Sample	Duplicate	RPD
Lipid	40	0.30	0.28	7

SAMPLE NO.

1665666DUP

FORM VI INORGANIC-2
DUPLICATES

Lab Name: Pace Analytical - Green Bay SDG No. : 40151601 Contract: HUDSON RIVER REMEDIAL

Matrix: Tissue Concentration Units: %

Percent Moisture: Basis: Wet

Analyte	Control Limit	Sample	Duplicate	RPD
Lipid	40	0.39	0.44	13

Prep Log Report

Batch Information: OEXT 282623 LIPIDS **Template Version: F-GB-O-145-Rev.00 (17Feb2015) LIPID**

Analysis Method	Pace Lipid	Analyzed By	AMC	Instrument	40BALL	Batch Notes	
Reviewed By	BLM	Reviewed By Date	05/03/2018 12:46				

Sample Information:

QC Rule	Sample Type	Lab Sample ID	Select	Lipid % Posted	Run Date/Time	ID	Sample Weight (g)	Sample Volume (mL)	Aliquot (mL)	Dish Weight (g)	Dry Weight 1 (g)	Dry Wt Use 1	Sample Notes
LIPID	BLANK	**1655713**	Y	-0.02500	03/13/2018 08:59:48		10	5	1	0.9256	0.9251	M	
LIPID	BLANK	**1655714**	Y	-0.005000	03/13/2018 08:59:48		10	5	1	0.9253	0.9252	M	
LIPID	PS	**40150295019**	Y	2.996	03/13/2018 08:59:48		10.012	5	1	0.9258	0.9858	M	
LIPID	DUP	**1655716**	Y	2.088	03/13/2018 08:59:48		10.0106	5	1	0.9241	0.9659	M	
LIPID	PS	**40150295020**	Y	0.7228	03/13/2018 08:59:48		9.9616	5	1	0.9251	0.9395	M	
LIPID	PS	**40151597002**	Y	0.6496	03/13/2018 08:59:48		10.0054	5	1	0.9272	0.9402	M	
LIPID	PS	**40151597003**	Y	0.3158	03/13/2018 08:59:48		9.9731	5	1	0.9222	0.9285	M	
LIPID	PS	**40151601002**	Y	0.3703	03/13/2018 08:59:48		9.9928	5	1	0.9241	0.9315	M	
LIPID	PS	**40151608001**	Y	0.7363	03/13/2018 08:59:48		9.9824	5	1	0.9213	0.936	M	
LIPID	PS	**40151608002**	Y	0.6200	03/13/2018 08:59:48		10.0008	5	1	0.9218	0.9342	M	
LIPID	PS	**40151608011**	Y	0.4249	03/13/2018 08:59:48		10.0021	5	1	0.9218	0.9303	M	
LIPID	PS	**40151800006**	Y	0.5787	03/13/2018 08:59:48		10.0226	5	1	0.9249	0.9365	M	
LIPID	PS	**40151800007**	Y	0.4591	03/13/2018 08:59:48		10.0192	5	1	0.9286	0.9378	M	
LIPID	PS	**40151800008**	Y	0.3457	03/13/2018 08:59:48		9.98	5	1	0.9229	0.9298	M	
LIPID	PS	**40151800009**	Y	0.3592	03/13/2018 08:59:48		10.0226	5	1	0.926	0.9332	M	
LIPID	PS	**40151800010**	Y	0.4864	03/13/2018 08:59:48		9.9707	5	1	0.9251	0.9348	M	
LIPID	PS	**40151804004**	Y	0.3344	03/13/2018 08:59:48		10.0181	5	1	0.9272	0.9339	M	
LIPID	PS	**40151804005**	Y	0.2505	03/13/2018 08:59:48		9.9782	5	1	0.9275	0.9325	M	
LIPID	PS	**40151804006**	Y	0.4104	03/13/2018 08:59:48		9.9895	5	1	0.9304	0.9386	M	

Prep Log Report

40151601

QC Rule	Sample Type	Lab Sample ID	Select	Lipid % Posted	Run Date/Time	ID	Sample Weight (g)	Sample Volume (mL)	Aliquot (mL)	Dish Weight (g)	Dry Weight 1 (g)	Dry Wt Use 1	Sample Notes
LIPID	RQS	**40151804007**	Y	0.3937	03/13/2018 08:59:48		10.0333	5	1	0.9225	0.9304	M	
LIPID	DUP	**1655717**	Y	0.4057	03/13/2018 08:59:48		9.9824	5	1	0.9236	0.9317	M	
LIPID	PS	**40151804008**	Y	0.2902	03/13/2018 08:59:48		9.9931	5	1	0.9254	0.9312	M	
LIPID	MS	**1682460**	Y	2.962	03/13/2018 08:59:48		10.0094	5	1	.9242	0.9835	M	

Prep Log Report

Batch Information: OEXT 283118 LIPIDS **Template Version: F-GB-O-145-Rev.00 (17Feb2015) LIPID**

Analysis Method	Pace Lipid	Analyzed By	AMC	Instrument	40BALL	Batch Notes	
Reviewed By	BLM	Reviewed By Date	05/03/2018 12:23				

Sample Information:

QC Rule	Sample Type	Lab Sample ID	Select	Lipid % Posted	Run Date/Time	ID	Sample Weight (g)	Sample Volume (mL)	Aliquot (mL)	Dish Weight (g)	Dry Weight 1 (g)	Dry Wt Use 1	Sample Notes
LIPID	BLANK	**1658693**	Y	-0.01500	03/14/2018 06:49:16		10	5	1	0.9256	0.9253	M	
LIPID	BLANK	**1658258**	Y	-0.02500	03/14/2018 06:49:16		10	5	1	0.9284	0.9279	M	
LIPID	PS	**40151549002**	Y	0.7210	03/14/2018 06:49:16		9.986	5	1	0.9245	0.9389	M	
LIPID	PS	**40151549003**	Y	1.380	03/14/2018 06:49:16		10.003	5	1	0.9214	0.949	M	
LIPID	PS	**40151549004**	Y	1.174	03/14/2018 06:49:16		10.0055	5	1	0.925	0.9485	M	
LIPID	PS	**40151549005**	Y	1.205	03/14/2018 06:49:16		10.0008	5	1	0.9265	0.9506	M	
LIPID	PS	**40151549006**	Y	0.6691	03/14/2018 06:49:16		10.0138	5	1	0.9244	0.9378	M	
LIPID	PS	**40151549007**	Y	0.7099	03/14/2018 06:49:16		10.0021	5	1	0.9212	0.9354	M	
LIPID	PS	**40151549008**	Y	0.9483	03/14/2018 06:49:16		10.0178	5	1	0.9285	0.9475	M	
LIPID	PS	**40151549009**	Y	0.5400	03/14/2018 06:49:16		10.0005	5	1	0.9247	0.9355	M	
LIPID	RQS	**40151552010**	Y	1.455	03/14/2018 06:49:16		9.9677	5	1	0.9245	0.9535	M	
LIPID	DUP	**1658259**	Y	1.529	03/14/2018 06:49:16		10.0089	5	1	0.9211	0.9517	M	
LIPID	PS	**40151601003**	Y	0.8971	03/14/2018 06:49:16		10.0319	5	1	0.924	0.942	M	
LIPID	PS	**40151601004**	Y	0.3551	03/14/2018 06:49:16		9.9968	5	1	0.923	0.9301	M	
LIPID	PS	**40151601005**	Y	0.6190	03/14/2018 06:49:16		10.0156	5	1	0.9209	0.9333	M	
LIPID	PS	**40151601006**	Y	1.412	03/14/2018 06:49:16		9.9883	5	1	0.923	0.9512	M	
LIPID	MS	**1682499**	Y	1.561	03/14/2018 06:49:16		9.9946	5	1	.9183	0.9495	M	

Prep Log Report

Batch Information: OEXT 284607 LIPIDS **Template Version: F-GB-O-145-Rev.01 (22Mar2018) LIPID**

Analysis Method	Pace Lipid	Analyzed By	ETB	Instrument	40BALL	Batch Notes	
Reviewed By	BLM	Reviewed By Date	05/02/2018 14:45				

Sample Information:

QC Rule	Sample Type	Lab Sample ID	Select	Lipid % Posted	Run Date/Time	ID	Sample Weight (g)	Sample Volume (mL)	Aliquot (mL)	Dish Weight (g)	Dry Weight 1 (g)	Dry Wt Use 1	Sample Notes
LIPID	BLANK	**1665662**	Y	0.000	03/29/2018 07:08:13		10	5	1	0.9302	0.9302	M	1664793
LIPID	BLANK	**1665663**	Y	-0.01000	03/29/2018 07:08:26		10	5	1	0.929	0.9288	M	1664794
LIPID	PS	**40151549001**	Y	0.8646	03/29/2018 07:08:40		10.0051	5	1	0.9306	0.9479	M	
LIPID	PS	**40151601001**	Y	0.4694	03/29/2018 07:08:51		10.0134	5	1	0.9305	0.9399	M	
LIPID	PS	**40151613001**	Y	0.2956	03/29/2018 07:09:02		9.9785	5	1	0.9291	0.935	M	
LIPID	DUP	**1665665**	Y	0.2754	03/29/2018 07:09:22		9.9864	5	1	0.928	0.9335	M	1664796
LIPID	PS	**40151613002**	Y	0.3351	03/29/2018 07:09:35		9.9976	5	1	0.9266	0.9333	M	
LIPID	PS	**40151614002**	Y	0.5648	03/29/2018 07:09:44		10.0031	5	1	0.9267	0.938	M	
LIPID	PS	**40151614003**	Y	0.2853	03/29/2018 07:09:58		9.9906	5	1	0.9255	0.9312	M	
LIPID	PS	**40151614004**	Y	0.2303	03/29/2018 07:10:08		9.986	5	1	0.9277	0.9323	M	
LIPID	PS	**40151613014**	Y	0.3100	03/29/2018 07:10:21		10	5	1	0.9278	0.934	M	
LIPID	PS	**40151800005**	Y	0.5196	03/29/2018 07:10:32		10.0085	5	1	0.926	0.9364	M	
LIPID	PS	**40151802001**	Y	0.2955	03/29/2018 07:10:38		9.9815	5	1	0.9251	0.931	M	
LIPID	PS	**40151802002**	Y	0.2035	03/29/2018 07:10:45		9.994	5	1.5	0.9294	0.9355	M	1*
LIPID	PS	**40151802003**	Y	0.3614	03/29/2018 07:10:53		9.9604	5	1	0.9296	0.9368	M	
LIPID	PS	**40151802004**	Y	0.3492	03/29/2018 07:11:00		10.0221	5	1	0.9287	0.9357	M	
LIPID	PS	**40151802005**	Y	0.3408	03/29/2018 07:11:07		9.9766	5	1	0.928	0.9348	M	
LIPID	PS	**40151805001**	Y	1.029	03/29/2018 07:11:17		10.0137	5	1	0.9291	0.9497	M	
LIPID	PS	**40157825001**	Y	0.3859	03/29/2018 07:11:23		9.9773	5	1	0.9255	0.9332	M	

Prep Log Report

QC Rule	Sample Type	Lab Sample ID	Select	Lipid % Posted	Run Date/Time	ID	Sample Weight (g)	Sample Volume (mL)	Aliquot (mL)	Dish Weight (g)	Dry Weight 1 (g)	Dry Wt Use 1	Sample Notes
LIPID	DUP	**1665666**	Y	0.4392	03/29/2018 07:11:30		10.019	5	1	0.924	0.9328	M	1664809
LIPID	MS	**1682554**	Y	0.2706	03/29/2018 07:11:30		9.9787	5	1	.9302	0.9356	M	

Sample Notes:

1*: 0.5mL of samples was added on accident to the tin so the Lipids are out of 1.5 instead of the usual 1.0mL. AMC 4/6/18

40151601

Client Name: GE
Project #: 40151601
Sample Type: Fish
Date/Time Removed to Thaw: 03-06-18
Date/Time Prepped: 03-06-18
Prepped by: CWN + JL + RM
Location Thawed: 40FRG21 / Room Temperature (circle one)

Pace Analytical

Biota Homogenization Log (NYDOH Method)

Logbook: 4413

Sample ID	A) Scales Removed B) Skin Removed (Note A or B Below)	Whole Body/Fillet (WB/F)	Picture of Fish/Weight Taken? (Y/N)	Total Fish Weight (g)	Internal Organs Intact (Y/N)	Gender (Optional) M/F/U[1]	Fillet Weight (g)	Left/Both Fillet[2]?	Picture of Fillet/Weight Taken? (Y/N)	Container Code[3]	Number of Containers Used	Picture of Jars/Weight Taken? (Y/N)	Weight of homogenized Sample (g)	Processing Type (C/G/R/RM)[4]	Sample Comments
1	B	F	Y	261.9	Y	F	38.0	Left / Both	Y	9AG	1	Y	36.2	C	
2	A	F	Y	619.4	Y	F	106.9	Left / Both	Y	9AG	1	Y	102.2	C	
3	B	F	Y	365.0	Y	F	69.1	Left / Both	Y	9AG	1	Y	67.7	C	
4	B	F	Y	328.1	Y	M	77.7	Left / Both	Y	9AG	1	Y	75.5	C	
5	B	F	Y	975.0	Y	F	91.6	Left / Both	Y	9AG	1	Y	90.8	C	
6	B	F	Y	301.5	Y	M	70.6	Left / Both	Y	9AG	1	Y	68.3	C	
								Left / Both							
								Left / Both							
								Left / Both							
								Left / Both							
								Left / Both							
								Left / Both							
								Left / Both							
						CWN	03-06-18	Left / Both							
								Left / Both							
								Left / Both							
								Left / Both							
								Left / Both							
								Left / Both							
								Left / Both							

503 of 503

(1) M = Male, F = Female, U = Undetermined
(2) If left fillet weight is under 40g, the right fillet in included in the sample weight.
(3) Note Container Type / Lot # using Codes Below:

Container Code	Container Size	Lot#
2CG	2 oz Clear Glass	
4AG	4 oz Amber Glass	
9AG	9 oz Amber Glass	F-7-212-04AB
32AG	32 oz Amber Glass	

(4) Processing Type Code Descriptions:

Code	Description of Code
C	Chopped
G	Grind
R	Robot coupe
RM	Robot coupe multiple batches

Reviewed By

Date 3/16/18