Lancaster Laboratories
2425 New Holland Pike, PO Box 12425, Lancaster, PA 17605-2425 • 717-656-2300 Fax: 717-656-2681 • www.lancasterlabs.com

Type I Data Package
for
Stantec

SDG# ASX22

Project: Amtrak Wilmington
Water and Soil Samples
Collected on 10/28/09-10/29/09

GROUP	SAMPLE NUMBERS
1168860	5822062-5822080

PA Cert. # 36-00037
NY Cert. # 10670
NJ Cert. # PA011
NC Cert. # 521
TX Cert. # T104704194-08A-TX

Prepared by _____

Reviewed by _____

Date 11/21/09

Lancaster Laboratories
2425 New Holland Pike, PO Box 12425, Lancaster, PA 17605-2425 • 717-656-2300 Fax: 717-656-2681 • www.lancasterlabs.com

Table of Contents for SDG# ASX22

1. Sample Reference List 1

2. Analysis Request, Field Chain-of-Custody Record 2

3. Methodology Summary/Reference 5

4. Analysis Reports 6

5. Polychlorinated Biphenyls(PCBs) Data 28

   a. Case Narrative-Conformance/Nonconform. Summary 29

   b. QC Summary 33

   c. Sample Data 44

   d. Standards Data 197

   e. Raw QC Data 820

   f. Extraction/Distillation/Digestion Logs 901

6. Moisture Data 909

Lancaster Laboratories
2425 New Holland Pike, PO Box 12425, Lancaster, PA 17605-2425 • 717-656-2300 Fax: 717-656-2681 • www.lancasterlabs.com

Sample Reference List for SDG Number ASX22 with a Data Package Type of I

11289 - Stantec

Project: Amtrak Wilmington

Lab Sample Number	Lab Sample Code	Client Sample Description
5822062	EBS28	EB-Sed10282009 Grab Water Sample
5822063	CARS1	Car_Shop-1 Grab Water Sample
5822064	CARS2	Car_Shop-2 Unspiked Grab Water Sample
5822065	CARS2	Car_Shop-2MS Matrix Spike Grab Water Sample
5822066	CARS2	Car_Shop-2MSD Matrix Spike Dup Grab Water Sample
5822067	DPW29	Dup-10292009 Grab Water Sample
5822068	STEA1	Steambay-1 Grab Water Sample
5822069	STEA2	Steambay-2 Grab Soil Sample
5822070	STEA3	Steambay-3 Unspiked Grab Soil Sample
5822071	STEA3	Steambay-3MS Matrix Spike Grab Soil Sample
5822072	STEA3	Steambay-3MSD Matrix Spike Dup Grab Soil Sample
5822073	DPS29	Dup-Sed10292009 Grab Soil Sample
5822074	ELEC3	Electric_Shop-3 Grab Soil Sample
5822075	WHEE1	Wheel_Shop-1 Grab Soil Sample
5822076	WHEE2	Wheel_Shop-2 Grab Soil Sample
5822077	WHEE3	Wheel_Shop-3 Grab Soil Sample
5822078	BB--1	B&B-1 Grab Water Sample
5822079	EBW29	EB-Aq10292009 Grab Water Sample
5822080	EBS29	EB-Sed10292009 Grab Water Sample

Analysis Request/ Environmental Services Chain of Custody

Lancaster Laboratories

For Lancaster Laboratories use only

Acct. # 11289 Group# 1168860 Sample # 5822062-80

COC # 190103

Please print. Instructions on reverse side correspond with circled numbers.

(1)
Client: Stantec Acct. #:
Project Name/#: Amtrak-Wilmington PWSID #:
Project Manager: S. Baggett P.O.#:
Sampler: P. Miller/M. Schaeffer Quote #:
Name of state where samples were collected: Delaware

(2) (3) (4) (5) Analyses Requested — Preservation Codes

For Lab Use Only
FSC:
SCR#:

Preservation Codes
H=HCl T=Thiosulfate
N=HNO_3 B=NaOH
S=H_2SO_4 O=Other

(6)

Sample Identification	Date Collected	Time Collected	[illegible]	[illegible]	[illegible]	[illegible]	[illegible]	RB 8082								Remarks
EB-Sed10282009	10-28-09	1430	X			X	2									
CarShop-1	10-29-09	0900				X	2									
CarShop-2		0915				X	2									
CarShop-2MS		0915				X	2									
CarShop-2MSD		0915				X	2									
Dup-10292009		0915				X	2									
Steambay-1		0945				X	2									
Steambay-2		0950			X		1									
Steambay-3		1000			X		1									
Steambay-3MS		1000			X		1									

(7) **Turnaround Time Requested** (TAT) (please circle): Normal Rush
(Rush TAT is subject to Lancaster Laboratories approval and surcharge.)
Date results are needed:
Rush results requested by (please circle): Phone Fax E-mail
Phone #: Fax #:
E-mail address:

(8) **Data Package Options** (please circle if required)
Type I (validation/NJ Reg) TX TRRP-13
Type II (Tier II) MA MCP CT RCP
Type III (Reduced NJ)
Type IV (CLP SOW)
Type VI (Raw Data Only)

SDG Complete? Yes No

Site-specific QC (MS/MSD/Dup)? Yes No
(If yes, indicate QC sample and submit triplicate volume.)
Internal COC Required? Yes / No

(9)

Relinquished by:	Date	Time	Received by:	Date	Time
[signature]	10-30-09	1100	Jim Lydey	10/30	1100
Jim Lydey	10/30	1655			
			Karen Hawthorne	10/30/09	16:55

Lancaster Laboratories, Inc., 2425 New Holland Pike, Lancaster, PA 17601 (717) 656-2300 Fax: (717) 656-6766
Copies: White and yellow should accompany samples to Lancaster Laboratories. The pink copy should be retained by the client.

2102.03

Analysis Request/ Environmental Services Chain of Custody

Lancaster Laboratories

For Lancaster Laboratories use only

Acct. # 11289 Group# 1168860 Sample # 5822062-80

COC # 190104

Please print. Instructions on reverse side correspond with circled numbers.

(1)
Client: Stantec Acct. #:
Project Name/#: Amtrak Wilmington PWSID #:
Project Manager: S. Baggett P.O.#:
Sampler: P. Miller M. Schaeffer Quote #:
Name of state where samples were collected: Delaware

(2) (3) (4) (5) Analyses Requested — Preservation Codes

For Lab Use Only
FSC:
SCR#:

Preservation Codes
H=HCl T=Thiosulfate
N=HNO_3 B=NaOH
S=H_2SO_4 O=Other

(6)

Sample Identification	Date Collected	Time Collected	Matrix						PCBs 8082								Remarks
Steam bay-3 MSD	10-29-09	1000	X		X			1									
Dup-Sed 10292009		1000			X			1									
Electric Shop 3		1030			X			1									
Wheel Shop-1		1045			X			1									
Wheel Shop-2		1050			X			1									
Wheel Shop 3		1100			X			1									
B&B-1		1130				X		2									
EB-Aq 10292009		1230				X		2									
EB-Sed 10292009		1230				X		2									

(7) **Turnaround Time Requested** (TAT) (please circle): Normal Rush
(Rush TAT is subject to Lancaster Laboratories approval and surcharge.)
Date results are needed:
Rush results requested by (please circle): Phone Fax E-mail
Phone #: Fax #:
E-mail address:

(8) **Data Package Options** (please circle if required)
Type I (validation/NJ Reg) TX TRRP-13
Type II (Tier II) MA MCP CT RCP
Type III (Reduced NJ)
Type IV (CLP SOW)
Type VI (Raw Data Only)

SDG Complete? Yes No

Site-specific QC (MS/MSD/Dup)? Yes No
(If yes, indicate QC sample and submit triplicate volume.)
Internal COC Required? Yes / No

(9)

Relinquished by:	Date	Time	Received by:	Date	Time
[signature]	10-30-09	11	Jim Lyody	10/30	1100
Jim Lyody	10/30	1655			
			Kevin Hartlaw	10/30/09	16:55

Lancaster Laboratories, Inc., 2425 New Holland Pike, Lancaster, PA 17601 (717) 656-2300 Fax: (717) 656-6766
Copies: White and yellow should accompany samples to Lancaster Laboratories. The pink copy should be retained by the client.

2102.03

Lancaster Laboratories

Environmental Sample Administration
Receipt Documentation Log

Client/Project: Stantec

Date of Receipt: 10/30/09

Time of Receipt: 16:55

Source Code: 01

Unpacker Emp. No.: 2114

Shipping Container Sealed: YES (NO)

Custody Seal Present *: YES (NO)

* Custody seal was intact unless otherwise noted in the discrepancy section

Package: (Chilled) Not Chilled

Temperature of Shipping Containers							
Cooler #	Thermometer ID	Temperature (°C)	Temp Bottle (TB) or Surface Temp (ST)	Wet Ice (WI) or Dry Ice (DI) or Ice Packs (IP)	Ice Present? Y/N	Loose (L) Bagged Ice (B) or NA	Comments
1	0429975	4.2	TB	WI	Y	B	
2	↓	2.1	↓	↓	↓	↓	
3	↓	1.3	↓	↓	↓	↓	
4	↓	2.3	↓	↓	↓	↓	
5	↓	1.4	↓	↓	↓	↓	
6							

Number of Trip Blanks received NOT listed on chain of custody: 0

Paperwork Discrepancy/Unpacking Problems:

10/30/09 ~~Also~~ rcvd 2 jars lanco 8.

Sample Administration Internal Chain of Custody			
Name	Date	Time	Reason for Transfer
Katie Hartlove	10/30/09	17:45	Unpacking to storage
Tammy Helsel	10/30/09	1823	Place in Storage or (Entry)
			Entry
			Entry

2425 New Holland Pike, PO Box 12425, Lancaster, PA 17605-2425 · 717-656-2300 Fax: 717-656-2681 · www.lancasterlabs.com

01216 PCBs in Solids
00173 PCBs in Water

The sample is solvent extracted and exchanged to hexane. Sulfuric acid and florisil cleanups are performed to reduce potential matrix interferences. The extract is analyzed using gas chromatography on two capillary columns (primary and confirmation) with electron capture detectors.

Reference: Test Methods for Evaluating Solid Waste, SW-846 Method 8082, December 1996.

00817 Water Sample Pest. Extraction

The sample is solvent extracted and exchanged to hexane. Florisil cleanup is performed to reduce potential matrix interferences. The extract is analyzed using gas chromatography on two capillary columns (primary and confirmation) with electron capture detectors.

Reference: Test Methods for Evaluating Solid Wastes, SW-846 Method 3510C, December 1996

00819 Solid Sample Pesticide Extract

Pesticides and/or PCBs are extracted from the sample using sonic probe with 1:1 methylene chloride and acetone. The extract is then concentrated and exchanged to hexane. Sulfuric acid, copper and florisil cleanups are performed to minimize matrix interferences.

Reference: Test Methods for Evaluating Solid Wastes, SW-846 Method 3550B, December 1996.

00111 Moisture
00118 Moisture
00121 Moisture Duplicate

A well-mixed sample is placed in a tared container and dried to a constant weight in an oven at 103-105C. The increase in weight is the total solids.

Reference: Standard Methods for the Examination of Water and Wastewater, 20th Edition, 1998, Method 2540 G

Lancaster Laboratories

ANALYTICAL RESULTS

Prepared for:

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

610-840-2538

Prepared by:

Lancaster Laboratories
2425 New Holland Pike
Lancaster, PA 17605-2425

November 11, 2009

Project: Amtrak Wilmington

Samples arrived at the laboratory on Friday, October 30, 2009. The PO# for this group is 1102. The group number for this submittal is 1168860.

Client Sample Description	Lancaster Labs (LLI) #
EB-Sed10282009 Grab Water Sample	5822062
Car_Shop-1 Grab Water Sample	5822063
Car_Shop-2 Unspiked Grab Water Sample	5822064
Car_Shop-2MS Matrix Spike Grab Water Sample	5822065
Car_Shop-2MSD Matrix Spike Dup Grab Water Sample	5822066
Dup-10292009 Grab Water Sample	5822067
Steambay-1 Grab Water Sample	5822068
Steambay-2 Grab Soil Sample	5822069
Steambay-3 Unspiked Grab Soil Sample	5822070
Steambay-3MS Matrix Spike Grab Soil Sample	5822071
Steambay-3MSD Matrix Spike Dup Grab Soil Sample	5822072
Dup-Sed10292009 Grab Soil Sample	5822073
Electric_Shop-3 Grab Soil Sample	5822074
Wheel_Shop-1 Grab Soil Sample	5822075
Wheel_Shop-2 Grab Soil Sample	5822076
Wheel_Shop-3 Grab Soil Sample	5822077
B&B-1 Grab Water Sample	5822078
EB-Aq10292009 Grab Water Sample	5822079
EB-Sed10292009 Grab Water Sample	5822080

The specific methodologies used in obtaining the enclosed analytical results are indicated on the Laboratory Sample Analysis Record.

ASXZZ [illegible]

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

ELECTRONIC COPY TO	Stantec	Attn: Steve Baggett
ELECTRONIC COPY TO	Stantec	Attn: Paul Miller
1 COPY TO	Data Package Group	

Questions? Contact your Client Services Representative
Loran A Carter at (717) 656-2300

Respectfully Submitted,

Sarah Snyder
Specialist

ASXZZ. 6687

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Lancaster Laboratories
2425 New Holland Pike • Lancaster, PA 17601

Explanation of Symbols and Abbreviations

The following defines common symbols and abbreviations used in reporting technical data:

RL	Reporting Limit	**BMQL**	Below Minimum Quantitation Level
N.D.	none detected	**MPN**	Most Probable Number
TNTC	Too Numerous To Count	**CP Units**	cobalt-chloroplatinate units
IU	International Units	**NTU**	nephelometric turbidity units
umhos/cm	micromhos/cm		
C	degrees Celsius	**F**	degrees Fahrenheit
meq	milliequivalents	**lb.**	pound(s)
g	gram(s)	**kg**	kilogram(s)
ug	microgram(s)	**mg**	milligram(s)
ml	milliliter(s)	**l**	liter(s)
m3	cubic meter(s)	**ul**	microliter(s)

< less than - The number following the sign is the limit of quantitation, the smallest amount of analyte which can be reliably determined using this specific test.

> greater than

J estimated value – The result is ≥ the Method Detection Limit (MDL) and < the Limit of Quantitation (LOQ).

ppm parts per million - One ppm is equivalent to one milligram per kilogram (mg/kg), or one gram per million grams. For aqueous liquids, ppm is usually taken to be equivalent to milligrams per liter (mg/l), because one liter of water has a weight very close to a kilogram. For gases or vapors, one ppm is equivalent to one microliter of gas per liter of gas.

ppb parts per billion

Dry weight basis Results printed under this heading have been adjusted for moisture content. This increases the analyte weight concentration to approximate the value present in a similar sample without moisture. All other results are reported on an as-received basis.

U.S. EPA CLP Data Qualifiers:

Organic Qualifiers

A	TIC is a possible aldol-condensation product
B	Analyte was also detected in the blank
C	Pesticide result confirmed by GC/MS
D	Compound quantitated on a diluted sample
E	Concentration exceeds the calibration range of the instrument
N	Presumptive evidence of a compound (TICs only)
P	Concentration difference between primary and confirmation columns >25%
U	Compound was not detected
X,Y,Z	Defined in case narrative

Inorganic Qualifiers

B	Value is <CRDL, but ≥IDL
E	Estimated due to interference
M	Duplicate injection precision not met
N	Spike sample not within control limits
S	Method of standard additions (MSA) used for calculation
U	Compound was not detected
W	Post digestion spike out of control limits
*****	Duplicate analysis not within control limits
+	Correlation coefficient for MSA <0.995

Analytical test results for methods listed on the laboratories' accreditation scope meet all requirements of NELAC unless otherwise noted under the individual analysis.

Measurement uncertainty values, as applicable, are available upon request.

Tests results relate only to the sample tested. Clients should be aware that a critical step in a chemical or microbiological analysis is the collection of the sample. Unless the sample analyzed is truly representative of the bulk of material involved, the test results will be meaningless. If you have questions regarding the proper techniques of collecting samples, please contact us. We cannot be held responsible for sample integrity, however, unless sampling has been performed by a member of our staff. This report shall not be reproduced except in full, without the written approval of the laboratory.

WARRANTY AND LIMITS OF LIABILITY - In accepting analytical work, we warrant the accuracy of test results for the sample as submitted. THE FOREGOING EXPRESS WARRANTY IS EXCLUSIVE AND IS GIVEN IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED. WE DISCLAIM ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING A WARRANTY OF FITNESS FOR PARTICULAR PURPOSE AND WARRANTY OF MERCHANTABILITY. IN NO EVENT SHALL LANCASTER LABORATORIES BE LIABLE FOR INDIRECT, SPECIAL, CONSEQUENTIAL, OR INCIDENTAL DAMAGES INCLUDING, BUT NOT LIMITED TO, DAMAGES FOR LOSS OF PROFIT OR GOODWILL REGARDLESS OF (A) THE NEGLIGENCE (EITHER SOLE OR CONCURRENT) OF LANCASTER LABORATORIES AND (B) WHETHER LANCASTER LABORATORIES HAS BEEN INFORMED OF THE POSSIBILITY OF SUCH DAMAGES. We accept no legal responsibility for the purposes for which the client uses the test results. No purchase order or other order for work shall be accepted by Lancaster Laboratories which includes any conditions that vary from the Standard Terms and Conditions of Lancaster Laboratories and we hereby object to any conflicting terms contained in any acceptance or order submitted by client.

Analysis Report

Lancaster Laboratories

Sample Description: EB-Sed10282009 Grab Water Sample
Amtrak Wilmington

LLI Sample # WW 5822062
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/28/2009 14:30 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

EBS28 SDG#: ASX22-01EB

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		ug/l	ug/l	
00173	PCB-1016	12674-11-2	< 0.47	0.47	1
00173	PCB-1221	11104-28-2	< 0.47	0.47	1
00173	PCB-1232	11141-16-5	< 0.47	0.47	1
00173	PCB-1242	53469-21-9	< 0.47	0.47	1
00173	PCB-1248	12672-29-6	< 0.47	0.47	1
00173	PCB-1254	11097-69-1	< 0.47	0.47	1
00173	PCB-1260	11096-82-5	< 0.47	0.47	1

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
00173	PCBs in Water	SW-846 8082	1	093060018A	11/06/2009 00:33	Jamie L Brillhart	1
00817	Water Sample Pest. Extraction	SW-846 3510C	1	093060018A	11/03/2009 04:00	Roman Kuropatkin	1

ASX22 0089

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Car_Shop-1 Grab Water Sample
Amtrak Wilmington

LLI Sample # WW 5822063
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 09:00 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

CARS1 SDG#: ASX22-02

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		**ug/l**	**ug/l**	
00173	PCB-1016	12674-11-2	< 62	62	100
00173	PCB-1221	11104-28-2	< 62	62	100
00173	PCB-1232	11141-16-5	< 62	62	100
00173	PCB-1242	53469-21-9	< 62	62	100
00173	PCB-1248	12672-29-6	200	62	100
00173	PCB-1254	11097-69-1	400	62	100
00173	PCB-1260	11096-82-5	420	62	100

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
00173	PCBs in Water	SW-846 8082	1	093060018A	11/06/2009 00:46	Jamie L Brillhart	100
00817	Water Sample Pest. Extraction	SW-846 3510C	1	093060018A	11/03/2009 04:00	Roman Kuropatkin	1

ASX22. 0018

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Car_Shop-2 Unspiked Grab Water Sample
Amtrak Wilmington

LLI Sample # WW 5822064
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 09:15 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

CARS2 SDG#: ASX22-03BKG

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		ug/l	ug/l	
00173	PCB-1016	12674-11-2	< 11	11	20
00173	PCB-1221	11104-28-2	< 11	11	20
00173	PCB-1232	11141-16-5	< 11	11	20
00173	PCB-1242	53469-21-9	< 11	11	20
00173	PCB-1248	12672-29-6	17	11	20
00173	PCB-1254	11097-69-1	110	11	20
00173	PCB-1260	11096-82-5	93	11	20

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
00173	PCBs in Water	SW-846 8082	1	093060018A	11/06/2009 00:59	Jamie L Brillhart	20
00817	Water Sample Pest. Extraction	SW-846 3510C	1	093060018A	11/03/2009 04:00	Roman Kuropatkin	1

ASXZZ 0811

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Car_Shop-2MS Matrix Spike Grab Water Sample
Amtrak Wilmington

LLI Sample # WW 5822065
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 09:15 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

CARS2 SDG#: ASX22-03MS

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		ug/l	ug/l	
00173	PCB-1016	12674-11-2	13	9.8	20
00173	PCB-1221	11104-28-2	< 9.8	9.8	20
00173	PCB-1232	11141-16-5	< 9.8	9.8	20
00173	PCB-1242	53469-21-9	< 9.8	9.8	20
00173	PCB-1248	12672-29-6	20	9.8	20
00173	PCB-1254	11097-69-1	56	9.8	20
00173	PCB-1260	11096-82-5	58	9.8	20

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
00173	PCBs in Water	SW-846 8082	1	093060018A	11/06/2009 01:12	Jamie L Brillhart	20
00817	Water Sample Pest. Extraction	SW-846 3510C	1	093060018A	11/03/2009 04:00	Roman Kuropatkin	1

ASX22 0012

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Car_Shop-2MSD Matrix Spike Dup Grab Water Sample
Amtrak Wilmington

LLI Sample # WW 5822066
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 09:15 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

CARS2 SDG#: ASX22-03MSD

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		ug/l	ug/l	
00173	PCB-1016	12674-11-2	13	9.6	20
00173	PCB-1221	11104-28-2	< 9.6	9.6	20
00173	PCB-1232	11141-16-5	< 9.6	9.6	20
00173	PCB-1242	53469-21-9	< 9.6	9.6	20
00173	PCB-1248	12672-29-6	19	9.6	20
00173	PCB-1254	11097-69-1	52	9.6	20
00173	PCB-1260	11096-82-5	53	9.6	20

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
00173	PCBs in Water	SW-846 8082	1	093060018A	11/06/2009 01:25	Jamie L Brillhart	20
00817	Water Sample Pest. Extraction	SW-846 3510C	1	093060018A	11/03/2009 04:00	Roman Kuropatkin	1

ASXZZ 6013

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Dup-10292009 Grab Water Sample
Amtrak Wilmington

LLI Sample # WW 5822067
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 09:15 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

DPW29 SDG#: ASX22-04FD

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		ug/l	ug/l	
00173	PCB-1016	12674-11-2	< 9.8	9.8	20
00173	PCB-1221	11104-28-2	< 9.8	9.8	20
00173	PCB-1232	11141-16-5	< 9.8	9.8	20
00173	PCB-1242	53469-21-9	< 9.8	9.8	20
00173	PCB-1248	12672-29-6	< 9.8	9.8	20
00173	PCB-1254	11097-69-1	49	9.8	20
00173	PCB-1260	11096-82-5	54	9.8	20

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
00173	PCBs in Water	SW-846 8082	1	093060018A	11/06/2009 01:38	Jamie L Brillhart	20
00817	Water Sample Pest. Extraction	SW-846 3510C	1	093060018A	11/03/2009 04:00	Roman Kuropatkin	1

ASX22 0014

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Steambay-1 Grab Water Sample
Amtrak Wilmington

LLI Sample # WW 5822068
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 09:45 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

STEA1 SDG#: ASX22-05

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		**ug/l**	**ug/l**	
00173	PCB-1016	12674-11-2	< 0.48	0.48	1
00173	PCB-1221	11104-28-2	< 0.48	0.48	1
00173	PCB-1232	11141-16-5	< 0.48	0.48	1
00173	PCB-1242	53469-21-9	< 0.48	0.48	1
00173	PCB-1248	12672-29-6	< 0.48	0.48	1
00173	PCB-1254	11097-69-1	1.9	0.48	1
00173	PCB-1260	11096-82-5	3.0	0.48	1

The surrogate data is outside the QC limits due to unresolvable matrix problems evident in the sample chromatogram.

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
00173	PCBs in Water	SW-846 8082	1	093060018A	11/06/2009 01:51	Jamie L Brillhart	1
00817	Water Sample Pest. Extraction	SW-846 3510C	1	093060018A	11/03/2009 04:00	Roman Kuropatkin	1

ASX22. 0015

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Steambay-2 Grab Soil Sample
Amtrak Wilmington

LLI Sample # SW 5822069
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 09:50 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

STEA2 SDG#: ASX22-06

CAT No.	Analysis Name	CAS Number	Dry Result	Dry Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		**ug/kg**	**ug/kg**	
01216	PCB-1016	12674-11-2	< 870	870	50
01216	PCB-1221	11104-28-2	< 870	870	50
01216	PCB-1232	11141-16-5	< 870	870	50
01216	PCB-1242	53469-21-9	< 870	870	50
01216	PCB-1248	12672-29-6	< 870	870	50
01216	PCB-1254	11097-69-1	3,000	870	50
01216	PCB-1260	11096-82-5	8,700	870	50
Wet Chemistry	**SM20 2540 G**		**%**	**%**	
00111	Moisture	n.a.	2.8	0.50	1

"Moisture" represents the loss in weight of the sample after oven drying at 103 - 105 degrees Celsius. The moisture result reported above is on an as-received basis.

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
01216	PCBs in Solids	SW-846 8082	1	093070028A	11/07/2009 11:00	Andrea J Covey	50
00819	Solid Sample Pesticide Extract	SW-846 3550B	1	093070028A	11/04/2009 09:30	Kerrie A Freeburn	1
00111	Moisture	SM20 2540 G	1	09307820007B	11/03/2009 16:40	Scott W Freisher	1

ASX22 0016

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Steambay-3 Unspiked Grab Soil Sample
Amtrak Wilmington

LLI Sample # SW 5822070
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 10:00 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

STEA3 SDG#: ASX22-07BKG

CAT No.	Analysis Name	CAS Number	Dry Result	Dry Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		**ug/kg**	**ug/kg**	
01216	PCB-1016	12674-11-2	< 26,000	26,000	1000
01216	PCB-1221	11104-28-2	< 26,000	26,000	1000
01216	PCB-1232	11141-16-5	< 26,000	26,000	1000
01216	PCB-1242	53469-21-9	< 26,000	26,000	1000
01216	PCB-1248	12672-29-6	< 26,000	26,000	1000
01216	PCB-1254	11097-69-1	120,000	26,000	1000
01216	PCB-1260	11096-82-5	110,000	26,000	1000
Wet Chemistry	**SM20 2540 G**		**%**	**%**	
00111	Moisture	n.a.	35.1	0.50	1

"Moisture" represents the loss in weight of the sample after oven drying at 103 - 105 degrees Celsius. The moisture result reported above is on an as-received basis.

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
01216	PCBs in Solids	SW-846 8082	1	093070028A	11/07/2009 11:24	Andrea J Covey	1000
00819	Solid Sample Pesticide Extract	SW-846 3550B	1	093070028A	11/04/2009 09:30	Kerrie A Freeburn	1
00111	Moisture	SM20 2540 G	2	09308820003A	11/04/2009 19:46	Scott W Freisher	1

ASX22 0017

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Steambay-3MS Matrix Spike Grab Soil Sample
Amtrak Wilmington

LLI Sample # SW 5822071
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 10:00 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

STEA3 SDG#: ASX22-07MS

CAT No.	Analysis Name	CAS Number	Dry Result	Dry Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		ug/kg	ug/kg	
01216	PCB-1016	12674-11-2	< 26,000	26,000	1000
01216	PCB-1221	11104-28-2	< 26,000	26,000	1000
01216	PCB-1232	11141-16-5	< 26,000	26,000	1000
01216	PCB-1242	53469-21-9	< 26,000	26,000	1000
01216	PCB-1248	12672-29-6	< 26,000	26,000	1000
01216	PCB-1254	11097-69-1	120,000	26,000	1000
01216	PCB-1260	11096-82-5	120,000	26,000	1000
Wet Chemistry	**SM20 2540 G**		%	%	
00118	Moisture	n.a.	35.1	0.50	1

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
01216	PCBs in Solids	SW-846 8082	1	093070028A	11/07/2009 11:48	Andrea J Covey	1000
00819	Solid Sample Pesticide Extract	SW-846 3550B	1	093070028A	11/04/2009 09:30	Kerrie A Freeburn	1
00118	Moisture	SM20 2540 G	2	09308820003A	11/04/2009 19:46	Scott W Freisher	1

ASX22 0018

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Steambay-3MSD Matrix Spike Dup Grab Soil Sample
Amtrak Wilmington

LLI Sample # SW 5822072
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 10:00 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

STEA3 SDG#: ASX22-07MSD

CAT No.	Analysis Name	CAS Number	Dry Result	Dry Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		**ug/kg**	**ug/kg**	
01216	PCB-1016	12674-11-2	< 26,000	26,000	1000
01216	PCB-1221	11104-28-2	< 26,000	26,000	1000
01216	PCB-1232	11141-16-5	< 26,000	26,000	1000
01216	PCB-1242	53469-21-9	< 26,000	26,000	1000
01216	PCB-1248	12672-29-6	< 26,000	26,000	1000
01216	PCB-1254	11097-69-1	120,000	26,000	1000
01216	PCB-1260	11096-82-5	100,000	26,000	1000

The sample was injected numerous times. Each time the responses for various analytes in the calibration check standard injected after the sample were outside the acceptance criteria. Therefore, this effect is attributed to the sample matrix and the data is reported.

Wet Chemistry	**SM20 2540 G**		**%**	**%**	
00118	Moisture	n.a.	35.1	0.50	1
00121	Moisture Duplicate	n.a.	32.2	0.50	1

The duplicate moisture value is provided to assess the precision of the moisture test. For comparability purposes, the initial moisture determination is the value used to perform dry weight calculations.

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
01216	PCBs in Solids	SW-846 8082	1	093070028A	11/09/2009 19:48	Andrea J Covey	1000
00819	Solid Sample Pesticide Extract	SW-846 3550B	1	093070028A	11/04/2009 09:30	Kerrie A Freeburn	1
00118	Moisture	SM20 2540 G	2	09308820003A	11/04/2009 19:46	Scott W Freisher	1
00121	Moisture Duplicate	SM20 2540 G	2	09308820003A	11/04/2009 19:46	Scott W Freisher	1

ASX22 0019

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Dup-Sed10292009 Grab Soil Sample
Amtrak Wilmington

LLI Sample # SW 5822073
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 10:00 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

DPS29 SDG#: ASX22-08FD

CAT No.	Analysis Name	CAS Number	Dry Result	Dry Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		**ug/kg**	**ug/kg**	
01216	PCB-1016	12674-11-2	< 24,000	24,000	1000
01216	PCB-1221	11104-28-2	< 24,000	24,000	1000
01216	PCB-1232	11141-16-5	< 24,000	24,000	1000
01216	PCB-1242	53469-21-9	< 24,000	24,000	1000
01216	PCB-1248	12672-29-6	< 24,000	24,000	1000
01216	PCB-1254	11097-69-1	190,000	24,000	1000
01216	PCB-1260	11096-82-5	170,000	24,000	1000

The sample was injected numerous times. Each time the responses for various analytes in the calibration check standard injected after the sample were outside the acceptance criteria. Therefore, this effect is attributed to the sample matrix and the data is reported.

Wet Chemistry	**SM20 2540 G**		**%**	**%**	
00111	Moisture	n.a.	28.3	0.50	1

"Moisture" represents the loss in weight of the sample after oven drying at 103 - 105 degrees Celsius. The moisture result reported above is on an as-received basis.

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
01216	PCBs in Solids	SW-846 8082	1	093070028A	11/09/2009 20:00	Andrea J Covey	1000
00819	Solid Sample Pesticide Extract	SW-846 3550B	1	093070028A	11/04/2009 09:30	Kerrie A Freeburn	1
00111	Moisture	SM20 2540 G	1	09307820007B	11/03/2009 16:40	Scott W Freisher	1

ASX22 0028

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Electric_Shop-3 Grab Soil Sample
Amtrak Wilmington

LLI Sample # SW 5822074
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 10:30 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

ELEC3 SDG#: ASX22-09

CAT No.	Analysis Name	CAS Number	Dry Result	Dry Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		**ug/kg**	**ug/kg**	
01216	PCB-1016	12674-11-2	< 1,800	1,800	100
01216	PCB-1221	11104-28-2	< 1,800	1,800	100
01216	PCB-1232	11141-16-5	< 1,800	1,800	100
01216	PCB-1242	53469-21-9	< 1,800	1,800	100
01216	PCB-1248	12672-29-6	< 1,800	1,800	100
01216	PCB-1254	11097-69-1	2,700	1,800	100
01216	PCB-1260	11096-82-5	12,000	1,800	100

The sample was injected numerous times. Each time the responses for various analytes in the calibration check standard injected after the sample were outside the acceptance criteria. Therefore, this effect is attributed to the sample matrix and the data is reported.

Wet Chemistry	**SM20 2540 G**		**%**	**%**	
00111	Moisture	n.a.	3.3	0.50	1

"Moisture" represents the loss in weight of the sample after oven drying at 103 - 105 degrees Celsius. The moisture result reported above is on an as-received basis.

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
01216	PCBs in Solids	SW-846 8082	1	093070028A	11/09/2009 20:12	Andrea J Covey	100
00819	Solid Sample Pesticide Extract	SW-846 3550B	1	093070028A	11/04/2009 09:30	Kerrie A Freeburn	1
00111	Moisture	SM20 2540 G	1	09307820007B	11/03/2009 16:40	Scott W Freisher	1

ASX22 0021

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Wheel_Shop-1 Grab Soil Sample
Amtrak Wilmington

LLI Sample # SW 5822075
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 10:45 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

WHEE1 SDG#: ASX22-10

CAT No.	Analysis Name	CAS Number	Dry Result	Dry Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		**ug/kg**	**ug/kg**	
01216	PCB-1016	12674-11-2	< 470	470	20
01216	PCB-1221	11104-28-2	< 470	470	20
01216	PCB-1232	11141-16-5	< 470	470	20
01216	PCB-1242	53469-21-9	< 470	470	20
01216	PCB-1248	12672-29-6	< 470	470	20
01216	PCB-1254	11097-69-1	< 470	470	20
01216	PCB-1260	11096-82-5	680	470	20

The sample was injected numerous times. Each time the responses for various analytes in the calibration check standard injected after the sample were outside the acceptance criteria. Therefore, this effect is attributed to the sample matrix and the data is reported.

Wet Chemistry	**SM20 2540 G**		**%**	**%**	
00111	Moisture	n.a.	28.3	0.50	1

"Moisture" represents the loss in weight of the sample after oven drying at 103 - 105 degrees Celsius. The moisture result reported above is on an as-received basis.

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
01216	PCBs in Solids	SW-846 8082	1	093070028A	11/09/2009 20:24	Andrea J Covey	20
00819	Solid Sample Pesticide Extract	SW-846 3550B	1	093070028A	11/04/2009 09:30	Kerrie A Freeburn	1
00111	Moisture	SM20 2540 G	1	09307820007B	11/03/2009 16:40	Scott W Freisher	1

ASX22 0822

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Wheel_Shop-2 Grab Soil Sample
Amtrak Wilmington

LLI Sample # SW 5822076
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 10:50 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

WHEE2 SDG#: ASX22-11

CAT No.	Analysis Name	CAS Number	Dry Result	Dry Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		**ug/kg**	**ug/kg**	
01216	PCB-1016	12674-11-2	< 50	50	1
01216	PCB-1221	11104-28-2	< 50	50	1
01216	PCB-1232	11141-16-5	< 50	50	1
01216	PCB-1242	53469-21-9	< 50	50	1
01216	PCB-1248	12672-29-6	< 50	50	1
01216	PCB-1254	11097-69-1	290	50	1
01216	PCB-1260	11096-82-5	270	50	1

The sample was injected numerous times. Each time the responses for various analytes in the calibration check standard injected after the sample were outside the acceptance criteria. Therefore, this effect is attributed to the sample matrix and the data is reported.

Due to the nature of the sample extract matrix, the extract could only be concentrated to a final volume of 25 ml instead of the usual volume of 10 ml. The reporting limits were raised accordingly.
The surrogate data is outside the QC limits due to unresolvable matrix problems evident in the sample chromatogram.

Wet Chemistry	**SM20 2540 G**		**%**	**%**	
00111	Moisture	n.a.	14.6	0.50	1

"Moisture" represents the loss in weight of the sample after oven drying at 103 - 105 degrees Celsius. The moisture result reported above is on an as-received basis.

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
01216	PCBs in Solids	SW-846 8082	1	093070028A	11/09/2009 20:36	Andrea J Covey	1
00819	Solid Sample Pesticide Extract	SW-846 3550B	1	093070028A	11/04/2009 09:30	Kerrie A Freeburn	1
00111	Moisture	SM20 2540 G	1	09307820007B	11/03/2009 16:40	Scott W Freisher	1

ASX22 0023

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: Wheel_Shop-3 Grab Soil Sample
Amtrak Wilmington

LLI Sample # SW 5822077
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 11:00 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

WHEE3 SDG#: ASX22-12

CAT No.	Analysis Name	CAS Number	Dry Result	Dry Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		ug/kg	ug/kg	
01216	PCB-1016	12674-11-2	< 1,800	1,800	100
01216	PCB-1221	11104-28-2	< 1,800	1,800	100
01216	PCB-1232	11141-16-5	< 1,800	1,800	100
01216	PCB-1242	53469-21-9	< 1,800	1,800	100
01216	PCB-1248	12672-29-6	< 1,800	1,800	100
01216	PCB-1254	11097-69-1	< 1,800	1,800	100
01216	PCB-1260	11096-82-5	5,200	1,800	100

The sample was injected numerous times. Each time the responses for various analytes in the calibration check standard injected after the sample were outside the acceptance criteria. Therefore, this effect is attributed to the sample matrix and the data is reported.

Due to interfering peaks on the chromatogram, the values reported represent the lowest reporting limits attainable.
Despite numerous cleanup methods, our usual reporting limits were not attained.

Wet Chemistry	**SM20 2540 G**		%	%	
00111	Moisture	n.a.	4.8	0.50	1

"Moisture" represents the loss in weight of the sample after oven drying at 103 - 105 degrees Celsius. The moisture result reported above is on an as-received basis.

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
01216	PCBs in Solids	SW-846 8082	1	093070028A	11/09/2009 20:48	Andrea J Covey	100
00819	Solid Sample Pesticide Extract	SW-846 3550B	1	093070028A	11/04/2009 09:30	Kerrie A Freeburn	1
00111	Moisture	SM20 2540 G	1	09307820007B	11/03/2009 16:40	Scott W Freisher	1

ASX22 0024

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: B&B-1 Grab Water Sample
Amtrak Wilmington

LLI Sample # WW 5822078
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 11:30 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

BB--1 SDG#: ASX22-13

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		ug/l	ug/l	
00173	PCB-1016	12674-11-2	< 0.49	0.49	1
00173	PCB-1221	11104-28-2	< 0.49	0.49	1
00173	PCB-1232	11141-16-5	< 0.49	0.49	1
00173	PCB-1242	53469-21-9	< 0.49	0.49	1
00173	PCB-1248	12672-29-6	< 0.49	0.49	1
00173	PCB-1254	11097-69-1	1.3	0.49	1
00173	PCB-1260	11096-82-5	1.8	0.49	1

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
00173	PCBs in Water	SW-846 8082	1	093060018A	11/06/2009 02:04	Jamie L Brillhart	1
00817	Water Sample Pest. Extraction	SW-846 3510C	1	093060018A	11/03/2009 04:00	Roman Kuropatkin	1

ASX22 0025

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: EB-Aq10292009 Grab Water Sample
Amtrak Wilmington

LLI Sample # WW 5822079
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 12:30 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

EBW29 SDG#: ASX22-14EB

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		ug/l	ug/l	
00173	PCB-1016	12674-11-2	< 0.47	0.47	1
00173	PCB-1221	11104-28-2	< 0.47	0.47	1
00173	PCB-1232	11141-16-5	< 0.47	0.47	1
00173	PCB-1242	53469-21-9	< 0.47	0.47	1
00173	PCB-1248	12672-29-6	< 0.47	0.47	1
00173	PCB-1254	11097-69-1	< 0.47	0.47	1
00173	PCB-1260	11096-82-5	< 0.47	0.47	1

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
00173	PCBs in Water	SW-846 8082	1	093060018A	11/06/2009 02:43	Jamie L Brillhart	1
00817	Water Sample Pest. Extraction	SW-846 3510C	1	093060018A	11/03/2009 04:00	Roman Kuropatkin	1

ASX22 0026

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Analysis Report

Lancaster Laboratories

Sample Description: EB-Sed10292009 Grab Water Sample
Amtrak Wilmington

LLI Sample # WW 5822080
LLI Group # 1168860
DE

Project Name: Amtrak Wilmington

Collected: 10/29/2009 12:30 by PM

Account Number: 11289

Submitted: 10/30/2009 16:55
Reported: 11/11/2009 at 14:55
Discard: 11/19/2009

Stantec
1060 Andrew Drive
Suite 140
West Chester PA 19380

EBS29 SDG#: ASX22-15EB*

CAT No.	Analysis Name	CAS Number	As Received Result	As Received Limit of Quantitation	Dilution Factor
Pesticides/PCBs	**SW-846 8082**		**ug/l**	**ug/l**	
00173	PCB-1016	12674-11-2	< 0.48	0.48	1
00173	PCB-1221	11104-28-2	< 0.48	0.48	1
00173	PCB-1232	11141-16-5	< 0.48	0.48	1
00173	PCB-1242	53469-21-9	< 0.48	0.48	1
00173	PCB-1248	12672-29-6	< 0.48	0.48	1
00173	PCB-1254	11097-69-1	< 0.48	0.48	1
00173	PCB-1260	11096-82-5	< 0.48	0.48	1

General Sample Comments

All QC is compliant unless otherwise noted. Please refer to the Quality Control Summary for overall QC performance data and associated samples.

Laboratory Sample Analysis Record

CAT No.	Analysis Name	Method	Trial#	Batch#	Analysis Date and Time	Analyst	Dilution Factor
00173	PCBs in Water	SW-846 8082	1	093060018A	11/06/2009 02:56	Jamie L Brillhart	1
00817	Water Sample Pest. Extraction	SW-846 3510C	1	093060018A	11/03/2009 04:00	Roman Kuropatkin	1

ASX22 0027

Lancaster Laboratories, Inc.
2425 New Holland Pike
PO Box 12425
Lancaster, PA 17605-2425
717-656-2300 Fax: 717-656-2681

Polychlorinated Biphenyls (PCBs) Data

ASXZZ 8028

Case Narrative Conformance/Nonconformance Summary

Lancaster Laboratories

CLIENT: Stantec
SDG: ASX22

LANCASTER LABORATORIES

PCBs

LLI SAMPLE #	SAMPLE CODE	MATRIX WATER	MATRIX SOLID	COMMENT
BLANKA 11/3/09	PBLK65	X		Method Blank
LCSA	LCSBG	X		Lab Control Spike
5822062	EBS28	X		
5822063	CARS1	X		DF100
5822064	CARS2	X		Unspiked DF20
5822065MS	CARS2MS	X		Matrix Spike DF20
5822066MSD	CARS2MSD	X		Matrix Spike Dup DF20
5822067	DPW29	X		DF20
5822068	STEA1	X		
5822078	BB--1	X		
5822079	EBW29	X		
5822080	EBS29	X		
BLANKA 11/4/09	PBLK8D		X	Method Blank
LCSA	LCSDS		X	Lab Control Spike
5822069	STEA2		X	DF50
5822070	STEA3		X	Unspiked DF1000
5822071MS	STEA3MS		X	Matrix Spike DF1000
5822072MSD	STEA3MSD		X	Matrix Spike Dup DF1000
5822073	DPS29		X	DF1000
5822074	ELEC3		X	DF100
5822075	WHEE1		X	DF20
5822076	WHEE2		X	
5822077	WHEE3		X	DF100

A. Sample Preparation:

Due to the nature of the sample matrix, the extract for 5822076 could only be concentrated to a final volume of 24 ml. instead of the usual 10 ml. volume. The reporting limits were raised accordingly.

No other problems were encountered with the preparation of the samples.

B. Analysis:

All compounds in the Initial Calibration Verification standards are within the 20% criteria.

No problems were encountered during analysis.

Samples 5822072, 5822073, 5822074, 5822075, 2855076, and 5822077 were injected numerous times. Each time the responses for various analytes in the continuing calibration standard injected after the samples were outside the 15% criteria. Therefore, this effect is attributed to the sample matrix and the data is reported.

Lancaster Laboratories, Inc. • 2425 New Holland Pike, PO Box 12425, Lancaster, PA 17605-2425 • 717-656-2300 Fax: 717-656-2681

The average %D and various compounds in continuing calibration standards 3P7307.63R on Column B and 3C6310.57R on Column A and Column B are outside the 15% criteria.

All other continuing calibration data are within method specifications since if any individual target analytes have a %D value outside the 15% criteria, then the average %D value of the standard is within the 15% criteria.

C. Quality Control:

Matrix QC may not be included if site-specific QC were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD was performed, unless otherwise specified in the method or by the client.

For preparation/method blank results >LOQ, corrective action is not required if the sample result is >10 times the blank concentration, unless otherwise specified in the method or by the client.

Surrogate recoveries that are outside the QC window are confirmed unless attributed to a dilution or otherwise noted.

The DCB surrogate recoveries for 5822068 and 5822076 are outside QC limits due to unresolvable matrix interference.

All other QC data are within specifications.

D. Data Interpretation:

Due to interfering peaks on the chromatogram for 5822077 and despite numerous cleanup methods, the values reported represent the lowest limits attainable.

No further interpretation is needed.

Data codes:

Detected analytes are reported using the higher of the two column results in accordance with the method. Data with analytes showing disparity >40% are further evaluated for reporting based on chromatographic appearance. The analytes will be coded with a comment number on the data sheet. The comments are explained below:

- 1569 - A disparity of >40% between the primary and confirmatory analysis occurred. Due to suspected interference, the lower result was reported.
- 1570 - A disparity of >40% between the primary and confirmatory analysis occurred. Since no chromatographic anomalies were apparent, the higher result was reported.

Data that indicates that manual integration was required would include the following codes:
1 = missed peak and 2 = improper baseline. The peaks that have been manually changed are indicated with an "M" on the raw data.

Lancaster Laboratories, Inc. • 2425 New Holland Pike, PO Box 12425, Lancaster, PA 17605-2425 • 717-656-2300 Fax: 717-656-2681

Lancaster Laboratories

Narrative reviewed and approved by:

Dana Kauffman, Manager Data Deliverables **Date** 11/21/05

ASX22 0832

QC Summary

2E
WATER SURROGATE RECOVERY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.: ASX22

GC Column (1): RTXCLPI ID: .32 GC Column (2): RTXCLPII ID: .32

Batchnumber: 093060018A

SAMPLE	SAMPLE CODE NO.	TCX 1 % REC #	TCX 2 % REC #	DCB 1 % REC #	DCB 2 % REC #	TOT OUT
5822062	EBS28	109	106	83	96	0
5822063	CARS1	-138 4D	97	-142 8D	174 D	0
5822064	CARS2	-161D	103	-226D	98	0
5822065 MS	CARS2MS	-159D	98	-234D	87	0
5822066 MSD	CARS2MSD	-167D	93	-235D	80	0
5822067	DPW29	-167D	94	-237D	79	0
5822068	STEA1	34 *	41 *	-1.3*	43	~~2~~ 3
5822078	BB--1	77	80	59	75	0
5822079	EBW29	99	99	83	94	0
5822080	EBS29	104	103	85	97	0
BLANKA	PBLK65	106	105	69	83	0
LCSA	LCSBG	110	109	52	66	0

@ MSK 16/
11/20/09

		ADVISORY QC LIMITS	NOMINAL CONCENTRATION	
TCX	= Tetrachloro-m-xylene	(55 - 132)	0.307	ug/l
DCB	= Decachlorobiphenyl	(36 - 153)	0.305	ug/l

ASX22 0034

Column to be used to flag recovery values
* Values outside of QC Limits
D Surrogate diluted out

2F
SOIL SURROGATE RECOVERY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.: ASX22

GC Column (1): RTX CLP ID: .32 GC Column (2): RTX CLPII ID: .32

Batchnumber: 093070028A

SAMPLE	SAMPLE CODE NO.	TCX 1 % REC #	TCX 2 % REC #	DCB 1 % REC #	DCB 2 % REC #	TOT OUT
5822069	STEA2	103	86	196 D	3621 D	0
5822070	STEA3	209 D	191 D	714 D	1315 D	0
5822071 MS	STEA3MS	193 D	239 D	OD	1880 D	0
5822072 MSD	STEA3MSD	239 D	188 D	458 D	1046 D	0
5822073	DPS29	OD	237 D	OD	2192 D	0
5822074	ELEC3	69	106	225 D	2028 D	0
5822075	WHEE1	106	96	155 D	111	0
5822076	WHEE2	76	54	0 *	OK	~~0~~ 2
5822077	WHEE3	85	OD	OD	307 D	0
BLANKA	PBLK8D	98	88	106	91	0
LCSA	LCSDS	108	104	123	106	0

MSK 5/
11/20/09

		ADVISORY QC LIMITS	NOMINAL CONCENTRATION	
TCX	= Tetrachloro-m-xylene	(53 - 139)	10.2	ug/kg
DCB	= Decachlorobiphenyl	(53 - 133)	10.2	ug/kg

ASX22 8835

Column to be used to flag recovery values
* Values outside of QC Limits
D Surrogate diluted out

FORM II - 2

3E
Water Matrix Spike/Matrix Spike Duplicate Recovery

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Matrix Spike - Sample Code No.: CARS2

Compound	Spike Added (ug/l)	Sample Concen (ug/l)	MS Concen (ug/l)	MSD Concen (ug/l)	MS % Rec #	MSD % Rec #	MS-MSD % REC Limits	% RPD #	% RPD Lim
Aroclor-1016	4.9	0	13	13	265*	271*	(76 - 116)	0	30
Aroclor-1260	4.9	93	58	53	-714*	-833*	(58 - 133)	9	30

Column to be used to flag recovery and RPD values with an asterisk

* Values outside of QC limits

RPD: 0 out of 2 outside limits

Spike Recovery: 4 out of 4 outside limits

Comments: Results calculated on as-received basis.

Sample No.: 5822064 Batch: 093060018A

ASX22 0036

3F

Soil Matrix Spike/Matrix Spike Duplicate Recovery

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Matrix Spike - Sample Code No.: STEA3

Compound	Spike Added (ug/kg)	Sample Concen (ug/kg)	MS Concen (ug/kg)	MSD Concen (ug/kg)	MS % Rec #	MSD % Rec #	MS-MSD % REC Limits	% RPD #	% RPD Lim
Aroclor-1016	166.67	0	0	0	0*	0*	(45 - 130)		50
Aroclor-1260	166.67	68912.82	74642.18	66203.78	3440*	-1630*	(39 - 149)	12	50

Column to be used to flag recovery and RPD values with an asterisk

* Values outside of QC limits

RPD: 0 out of 2 outside limits

Spike Recovery: 4 out of 4 outside limits

Comments: Results calculated on as-received basis.

Sample No.: 5822070 Batch: 093070028A

ASX22. 0037

3E

Water Lab Control Spike/Lab Control Spike Duplicate Recovery

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Laboratory Control Spike - Sample Code No.: LCSBG

Compound	Spike Added (ug/l)	LCS Concen (ug/l)	LCSD Concen (ug/l)	LCS % Rec #	LCSD % Rec #	LCS-LCSD % REC Limits	% RPD #	% RPD Lim
Aroclor-1016	5.0	5.1		102		(69 - 114)		30
Aroclor-1260	5.0	5.1		102		(75 - 111)		30

Column to be used to flag recovery and RPD values with an asterisk

* Values outside of QC limits

RPD: 0 out of 2 outside limits

Spike Recovery: 0 out of 2 outside limits

Comments: Results calculated on as-received basis.

Sample No.: LCSA Batch: 093060018A

ASX22 8838

3F

Soil Lab Control Spike/Lab Control Spike Duplicate Recovery

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Laboratory Control Spike - Sample Code No.: LCSDS

Compound	Spike Added (ug/kg)	LCS Concen (ug/kg)	LCSD Concen (ug/kg)	LCS % Rec #	LCSD % Rec #	LCS-LCSD % REC Limits	% RPD #	% RPD Lim
Aroclor-1016	166.67	154.06		92		(68 - 116)		50
Aroclor-1260	166.67	183.15		110		(72 - 120)		50

Column to be used to flag recovery and RPD values with an asterisk

* Values outside of QC limits

RPD: 0 out of 2 outside limits

Spike Recovery: 0 out of 2 outside limits

Comments: Results calculated on as-received basis.

Sample No.: LCSA Batch: 093070028A

ASX22 0039

4C

METHOD BLANK SUMMARY

SAMPLE CODE NO.

PBLK65

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.: ASX22

Lab Sample ID BLANKA Batch 093060018A Lab File ID: 3P7307.42R 3P7307B.42R

Matrix: (soil/water) WATER Extraction: (SepF/Cont/Sonc) SEPF

Sulfur Cleanup: (Y/N) N Date Extracted: 11/3/2009

Date Analyzed (1): 11/6/2009 Date Analyzed (2): 11/6/2009

Time Analyzed (1): 00:07:24 Time Analyzed (2): 00:07:24

Instrument ID (1): H5088A Instrument ID (2): H5088B

GC Column: RTXCLPI ID: 0.32 (mm) GC Column: RTXCLPII ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS, AND MSD

	SAMPLE CODE NO.	LAB SAMPLEID	DATE ANALYZED 1	DATE ANALYZED 2
01	EBS28	5822062	11/6/2009	11/6/2009
02	CARS1	5822063	11/6/2009	11/6/2009
03	CARS2	5822064	11/6/2009	11/6/2009
04	CARS2MS	5822065MS	11/6/2009	11/6/2009
05	CARS2MSD	5822066MSD	11/6/2009	11/6/2009
06	DPW29	5822067	11/6/2009	11/6/2009
07	STEA1	5822068	11/6/2009	11/6/2009
08	BB--1	5822078	11/6/2009	11/6/2009
09	EBW29	5822079	11/6/2009	11/6/2009
10	EBS29	5822080	11/6/2009	11/6/2009
11	PBLK65	BLANKA	11/6/2009	11/6/2009
12	LCSBG	LCSA	11/6/2009	11/6/2009

ASX22. 0040

COMMENTS:

FORM IV

1D

ORGANICS ANALYSIS DATA SHEET

SAMPLE CODE NO.

PBLK65

Lab Name: Lancaster Laboratories **Contract:** **Batchnumber:** 093060018A

Lab Code: **Case No.:** **SAS No.:** **SDG No.:**

Matrix: (soil/water) WATER **Lab Sample ID:** BLANKA

Sample wt/vol: 1000 **(g/ml)** ml **Lab File ID:** 3P7307.42R

% Moisture: **Decanted: (Y/N)** **Date Received:**

Extraction: (SepF/Cont/Sonc) SEPF **Date Extracted:** 11/3/2009

Concentrated Extract Volume: 10000 **(uL)** **Date Analyzed:** 11/6/2009

Injection Volume: 1 **(uL)** **Dilution Factor:** 1

GPC Cleanup: (Y/N) N **pH:** **Sulfur Cleanup: (Y/N)** N

CAS NO.	COMPOUND	CONCENTRATION UNITS (UG/L or UG/KG) ug/l	Q
12674-11-2	Aroclor-1016	0.10	U
11104-28-2	Aroclor-1221	0.10	U
11141-16-5	Aroclor-1232	0.10	U
53469-21-9	Aroclor-1242	0.10	U
12672-29-6	Aroclor-1248	0.10	U
11097-69-1	Aroclor-1254	0.10	U
11096-82-5	Aroclor-1260	0.10	U

ASX22 0841

FORM I

4C

METHOD BLANK SUMMARY

SAMPLE CODE NO.

PBLK8D

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.: ASX22

Lab Sample ID BLANKA Batch 093070028A Lab File ID: 2C6310.03R 2C6310B.03R

Matrix: (soil/water) SOIL Extraction: (SepF/Cont/Sonc) SONC

Sulfur Cleanup: (Y/N) Y Date Extracted: 11/4/2009

Date Analyzed (1): 11/7/2009 Date Analyzed (2): 11/7/2009

Time Analyzed (1): 10:24:40 Time Analyzed (2): 10:24:40

Instrument ID (1): S9147A Instrument ID (2): S9147B

GC Column: RTX CLP ID: 0.32 (mm) GC Column: RTX CLPII ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS, AND MSD

	SAMPLE CODE NO.	LAB SAMPLEID	DATE ANALYZED 1	DATE ANALYZED 2
01	STEA2	5822069	11/7/2009	11/7/2009
02	STEA3	5822070	11/7/2009	11/7/2009
03	STEA3MS	5822071MS	11/7/2009	11/7/2009
04	STEA3MSD	5822072MSD	11/9/2009	11/9/2009
05	DPS29	5822073	11/9/2009	11/9/2009
06	ELEC3	5822074	11/9/2009	11/9/2009
07	WHEE1	5822075	11/9/2009	11/9/2009
08	WHEE2	5822076	11/9/2009	11/9/2009
09	WHEE3	5822077	11/9/2009	11/9/2009
10	PBLK8D	BLANKA	11/7/2009	11/7/2009
11	LCSDS	LCSA	11/7/2009	11/7/2009

ASX22 0842

COMMENTS:

FORM IV

1D

ORGANICS ANALYSIS DATA SHEET

SAMPLE CODE NO.

PBLK8D

Lab Name: Lancaster Laboratories **Contract:** **Batchnumber:** 093070028A

Lab Code: **Case No.:** **SAS No.:** **SDG No.:**

Matrix: (soil/water) SOIL **Lab Sample ID:** BLANKA

Sample wt/vol: 30 **(g/ml)** g **Lab File ID:** 2C6310.03R

% Moisture: **Decanted: (Y/N)** **Date Received:**

Extraction: (SepF/Cont/Sonc) SONC **Date Extracted:** 11/4/2009

Concentrated Extract Volume: 10000 **(uL)** **Date Analyzed:** 11/7/2009

Injection Volume: 1 **(uL)** **Dilution Factor:** 1

GPC Cleanup: (Y/N) N **pH:** **Sulfur Cleanup: (Y/N)** Y

CAS NO.	COMPOUND	CONCENTRATION UNITS (UG/L or UG/KG) ug/kg	Q
12674-11-2	Aroclor-1016	3.3	U
11104-28-2	Aroclor-1221	4.0	U
11141-16-5	Aroclor-1232	5.3	U
53469-21-9	Aroclor-1242	3.3	U
12672-29-6	Aroclor-1248	3.3	U
11097-69-1	Aroclor-1254	3.3	U
11096-82-5	Aroclor-1260	3.3	U

ASX22: 0043

FORM I

Sample Data

Analysis LOQ/MDL Report

Analysis: 00173 Name: PCBs in Water

Description: Default Values

Compound	Units	LOQ	MDL
Aroclor-1260	ug/l	0.5	0.1
Aroclor-1254	ug/l	0.5	0.1
Aroclor-1221	ug/l	0.5	0.1
Aroclor-1232	ug/l	0.5	0.1
Aroclor-1248	ug/l	0.5	0.1
Aroclor-1016	ug/l	0.5	0.1
Aroclor-1242	ug/l	0.5	0.1

ASX22: 0045

Analysis LOQ/MDL Report

Analysis: 01216 Name: PCBs in Solids

Description: Default Values

Compound	Units	LOQ	MDL
Aroclor-1260	ug/kg	17	3.3
Aroclor-1254	ug/kg	17	3.3
Aroclor-1221	ug/kg	17	3.3
Aroclor-1232	ug/kg	17	3.3
Aroclor-1248	ug/kg	17	3.3
Aroclor-1016	ug/kg	17	3.3
Aroclor-1242	ug/kg	17	3.3

ASX22 0046

Multiple Component Data Summary

Sample Name: 5822062 AF EBS28 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1060 ml Total Volume: 10 ml Analyst: 2131 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on	Nov 06, 2009 00:33:23
Instrument	H5088A
Result file	3P7307.44R
Calibration file	1P7307
Method file	PESTD

%SSR(TCX) 108.6% Conc: 0.314709
%SSR(DCB) 82.8% Conc: 0.238114

Analysis Report (B)

Injected on	Nov 06, 2009 00:33:23
Instrument	H5088B
Result file	3P7307B.44R
Calibration file	1P7307B
Method file	PESTDB

%SSR(TCX) 105.9% Conc: 0.306814
%SSR(DCB) 95.6% Conc: 0.274888

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<0.4717	<0.0943			4	15	
Aroclor-1221			<0.4717	<0.0943			3	5	
Aroclor-1232			<0.4717	<0.0943			4	10	
Aroclor-1242			<0.4717	<0.0943			4	30	
Aroclor-1248			<0.4717	<0.0943			4	30	
Aroclor-1254			<0.4717	<0.0943			4	40	
Aroclor-1260			<0.4717	<0.0943			4	40	

Units: ug/l

%Difference = High - Low divided by the Average times 100

Reviewed By [signature] Date: 11-6-09

Verified By: [signature] Date: 11/9/09
[signature] 11/9/09

ASX22: 0847

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822062** AF EBS28 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1060 ml Total Volume: 10 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 00:33:23
Instrument : CP07--H5088A
Result file : 3P7307.44R
Calibration file : 1P7307.CAL
Method file : PESTD.MET
%SSR(TCX) : 108.6% Conc.: 0.314709
%SSR(DCB) : 82.8% Conc.: 0.238114

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					5	76.94	
3.35	3.39	3.41	43.732914	0.003945			1
3.83	3.88	3.89	90.883118	0.006761			2
4.15	4.20	4.21	173.918182	0.013575			4
4.24	4.26	4.30	79.859413	0.006582			5
4.62	4.68	4.68	219.093781	0.025338			6
Height Summation:			607.487408				
Amount Avg CF:			0.01124	Linear:			
Aroclor-1221					3	168.93	
3.20	3.22	3.26	44.365826	0.006737			1
3.28	3.29	3.34	2659.868408	0.572575			2
3.35	3.39	3.41	43.732914	0.002838			3
Height Summation:			2747.967148				
Amount Avg CF:			0.19405	Linear:			
Aroclor-1232					5	88.94	
3.35	3.39	3.41	43.732914	0.003361			1
3.83	3.88	3.89	90.883118	0.015195			2
4.15	4.20	4.21	173.918182	0.028364			4
4.23	4.26	4.29	79.859413	0.014035			5
4.62	4.68	4.68	219.093781	0.058346			6
Height Summation:			607.487408				
Amount Avg CF:			0.02386	Linear:			
Aroclor-1242					5	44.86	
3.35	3.39	3.41	43.732914	0.004614			1
3.83	3.88	3.89	90.883118	0.008274			2
4.09	4.09	4.15	247.867477	0.009944			3
4.15	4.20	4.21	173.918182	0.015872			4
4.23	4.26	4.29	79.859413	0.007674			5
Height Summation:			636.261104				
Amount Avg CF:			0.009276	Linear:			
Aroclor-1254					2	36.98	
5.34	5.38	5.40	333.700226	0.013287			4
5.98	6.01	6.04	133.551514	0.007779			6
Height Summation:			467.251740				
Amount Avg CF:			0.010533	Linear:			
Aroclor-1260					1		
5.98	6.01	6.04	133.551514	0.007256			2
Height Summation:			133.551514				
Amount Avg CF:			0.007256	Linear:			
Aroclor-1268					2	116.65	
6.87	6.88	6.93	645.00824	0.016852			4
7.11	7.15	7.17	1327.650391	0.175601			6
Height Summation:			1972.658631				
Amount Avg CF:			0.096227	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 00:33:23
Instrument : CP07--H5088B
Result file : 3P7307B.44R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET
%SSR(TCX) : 105.9% Conc.: 0.306814
%SSR(DCB) : 95.6% Conc.: 0.274888

A Uncl

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					1		
3.14	3.17	3.20	55.602303	0.004448			3
Height Summation:			55.602303				
Amount Avg CF:			0.004448	Linear:			
Aroclor-1221					1		
2.61	2.62	2.67	1805.081299	0.600143			2
Height Summation:			1805.081299				
Amount Avg CF:			0.600143	Linear:			
Aroclor-1232					1		
3.14	3.17	3.20	55.602303	0.009508			3
Height Summation:			55.602303				
Amount Avg CF:			0.009508	Linear:			
Aroclor-1242					1		
3.14	3.17	3.20	55.602303	0.005479			3
Height Summation:			55.602303				
Amount Avg CF:			0.005479	Linear:			
Aroclor-1248					3	149.11	
3.14	3.17	3.20	55.602303	0.012445			1
4.26	4.31	4.32	337.296173	0.030985			5
4.55	4.60	4.61	2786.897461	0.422542			6
Height Summation:			3179.795937				
Amount Avg CF:			0.155324	Linear:			
Aroclor-1254					3	136.87	
4.55	4.60	4.61	2786.897461	0.126387			2
4.85	4.87	4.91	98.555	0.010567			5
5.11	5.14	5.17	191.576157	0.009986			6
Height Summation:			3077.028618				
Amount Avg CF:			0.04898	Linear:			
Aroclor-1260					2	48.76	
4.84	4.87	4.90	98.555	0.004683			2
5.60	5.63	5.66	327.204865	0.009611			5
Height Summation:			425.759865				
Amount Avg CF:			0.007147	Linear:			
Aroclor-1262					2	25.88	
4.84	4.87	4.90	98.555	0.005129			1
5.60	5.63	5.66	327.204865	0.007427			3
Height Summation:			425.759865				
Amount Avg CF:			0.006278	Linear:			

Lmr

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			0.4717	0.0943		** 86.59	4	15	
Aroclor-1221			0.4717	0.0943		** 102.27	3	5	
Aroclor-1232			0.4717	0.0943		** 86.02	4	10	
Aroclor-1242			0.4717	0.0943		** 51.47	4	30	ASX22. 0040

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:10:27

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822062** AF EBS28 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1060 ml Total Volume: 10 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 00:33:23
Instrument : CP07--H5088A
Result file : 3P7307.44R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

Analysis Report (B)

Injected on : NOV 06, 2009 00:33:23
Instrument : CP07--H5088B
Result file : 3P7307B.44R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1248			0.4717	0.0943			4	30	
Aroclor-1254			0.4717	0.0943		** 129.21	4	40	
Aroclor-1260			0.4717	0.0943		1.51	4	40	
Aroclor-1262			0	0			4	40	
Aroclor-1268			0	0			4	40	

Units: ug/L

ASX22 5049

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:10:27

Sample Name: **5822062AF ABEBS28 T 093060018A 00173**
Acquired from CP07--H5088A via port 3 on 11/6/09 12:42:31am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.44R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.584		0.0000	0.000	765.6	0.150	BV	0.060	212.80	0.090
2	1.826		0.0000	0.000	8895.5	1.747	VV	0.199	746.87	0.317
3	2.078		0.0000	0.000	14295.2	2.807	VV	0.130	1830.68	0.776
4	2.177		0.0000	0.000	4479.3	0.880	VV	0.065	1148.98	0.487
5	2.213		0.0000	0.000	2919.2	0.573	VV	0.044	1104.91	0.469
6	2.290		0.0000	0.000	2778.3	0.546	VV	0.038	1231.00	0.522
7	2.569		0.0000	0.000	1176.2	0.231	VB	0.079	248.50	0.105
8	2.850		0.0000	0.000	2977.0	0.585	BB	0.113	437.18	0.185
9	3.001	TCX	333.5919	56.928	250455.7	49.187	BB	0.027	155670.80	66.024
10	3.221		0.0000	0.000	53.0	0.010	BB	0.020	44.37	0.019
11	3.289		0.0000	0.000	4391.8	0.862	BB	0.028	2659.87	1.128
12	3.395		0.0000	0.000	74.0	0.015	BB	0.028	43.73	0.019
13	3.474		0.0000	0.000	720.4	0.141	BB	0.030	404.96	0.172
14	3.720		0.0000	0.000	275.6	0.054	BB	0.075	61.29	0.026
15	3.765		0.0000	0.000	97.7	0.019	BB	0.022	75.48	0.032
16	3.800		0.0000	0.000	126.6	0.025	BB	0.022	97.77	0.041
17	3.878		0.0000	0.000	243.6	0.048	BB	0.045	90.88	0.039
18	3.940		0.0000	0.000	667.1	0.131	BB	0.025	441.24	0.187
19	4.025		0.0000	0.000	164.0	0.032	BB	0.037	74.42	0.032
20	4.094		0.0000	0.000	703.7	0.138	BB	0.047	247.87	0.105
21	4.201		0.0000	0.000	539.9	0.106	BB	0.052	173.92	0.074
22	4.264		0.0000	0.000	117.1	0.023	BB	0.024	79.86	0.034
23	4.462		0.0000	0.000	1020.0	0.200	BB	0.069	245.60	0.104
24	4.679		0.0000	0.000	685.3	0.135	BB	0.052	219.09	0.093
25	4.804		0.0000	0.000	5894.4	1.158	BB	0.165	596.92	0.253
26	4.964		0.0000	0.000	4161.7	0.817	BB	0.167	415.77	0.176
27	5.175		0.0000	0.000	215.0	0.042	BB	0.016	224.03	0.095
28	5.380		0.0000	0.000	593.7	0.117	BB	0.030	333.70	0.142
29	5.475		0.0000	0.000	151.1	0.030	BB	0.023	110.49	0.047
30	5.910		0.0000	0.000	2043.9	0.401	BB	0.388	87.81	0.037
31	5.966		0.0000	0.000	81.0	0.016	BB	0.027	49.54	0.021
32	6.012		0.0000	0.000	240.4	0.047	BB	0.030	133.55	0.057
33	6.420		0.0000	0.000	37014.6	7.269	BB	0.265	2324.63	0.986
34	6.877		0.0000	0.000	5083.4	0.998	BB	0.131	645.01	0.274
35	6.976		0.0000	0.000	102.2	0.020	BB	0.039	43.73	0.019
36	7.155		0.0000	0.000	14680.8	2.883	BB	0.184	1327.65	0.563
37	7.584		0.0000	0.000	4638.1	0.911	BB	0.140	551.14	0.234
38	7.731		0.0000	0.000	1331.7	0.262	BB	0.035	640.57	0.272
39	7.901		0.0000	0.000	142.8	0.028	BB	0.044	53.56	0.023
40	7.976	DCB	252.4003	43.072	134198.6	26.355	BB	0.037	60648.52	25.723

Total Area = 509195.1, Total Amount = 585.992, Total Height = 235778.7, Sample Units = ug/l

ASX22 0058

Sample Name: **5822062AF ABEBS28 T 093060018A 00173**
Acquired from CP07--H5088B via port 4 on 11/6/09 12:42:31am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.44R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.590		0.0000	0.000	5557.5	0.910	BV	0.079	1173.33	0.464
2	1.710		0.0000	0.000	1550.3	0.254	VV	0.043	598.49	0.237
3	1.767		0.0000	0.000	2837.7	0.465	VB	0.052	909.19	0.359
4	2.300		0.0000	0.000	93.7	0.015	BB	0.045	34.91	0.014
5	2.343	TCX	325.2232	52.744	322478.1	52.826	BB	0.038	141044.00	55.745
6	2.624		0.0000	0.000	3815.4	0.625	BB	0.035	1805.08	0.713
7	2.776		0.0000	0.000	711.6	0.117	BB	0.041	290.04	0.115
8	3.080		0.0000	0.000	262.0	0.043	BB	0.037	118.88	0.047
9	3.170		0.0000	0.000	98.0	0.016	BB	0.029	55.60	0.022
10	3.227		0.0000	0.000	64.6	0.011	BB	0.029	37.70	0.015
11	3.269		0.0000	0.000	3895.1	0.638	BB	0.198	328.32	0.130
12	4.307		0.0000	0.000	1834.2	0.300	BB	0.091	337.30	0.133
13	4.599		0.0000	0.000	5410.2	0.886	BB	0.032	2786.90	1.101
14	4.827		0.0000	0.000	533.1	0.087	BB	0.033	269.40	0.106
15	4.870		0.0000	0.000	201.0	0.033	BB	0.034	98.56	0.039
16	5.142		0.0000	0.000	3976.5	0.651	BB	0.346	191.58	0.076
17	5.627		0.0000	0.000	1105.9	0.181	BB	0.056	327.20	0.129
18	5.800		0.0000	0.000	3540.8	0.580	BB	0.170	347.08	0.137
19	6.567		0.0000	0.000	537.8	0.088	BB	0.374	24.00	0.009
20	6.622		0.0000	0.000	3248.5	0.532	BB	0.046	1179.01	0.466
21	6.800	DCB	291.3818	47.256	228940.0	37.503	BB	0.038	100591.30	39.757
22	7.343		0.0000	0.000	390.9	0.064	BB	0.046	142.22	0.056
23	7.713		0.0000	0.000	19369.8	3.173	BB	0.988	326.67	0.129

Total Area = 610452.9, Total Amount = 616.605, Total Height = 253016.8, Sample Units = ug/l

ASX22 0051

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\3P7307.44R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 12:33:22 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.44R



Instrument ID: CP07--H5088B Injected On: 11/6/2009 12:33:22 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1060
Analyst: 2280

Dilution Factor: 10

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.001	155671	.31471	TCX	2.343	141044	.30681	TCX
7.976	60649	.23811	DCB	6.8	100591	.27489	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.44A
Area File: C:\CPWIN\DATA1\3P7307B.44A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 12:42:38 AM
File Reported On: 11/6/2009 at 12:42:50 AM

10X REPLOTS

ASX22 0053

C:\CPWIN\DATA1\3P7307.44R



C:\CPWIN\DATA1\3P7307B.44R



ASX22 0054

Multiple Component Data Summary

Sample Name: 5822063 AF DF100 CARS1 Sample ID: AC **Batchnumber:** 093060018A
Sample Amount: 805 ml Total Volume: 1000 ml Analyst: 2131 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on Nov 06, 2009 00:46:24
Instrument H5088A
Result file 3P7307.45R
Calibration file 1P7307
Method file PESTD

%SSR(TCX) * 0.0% Conc: 0
%SSR(DCB) * 0.0% Conc: 0

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1248						**42.38**	
4.10	4.13	4.16	32855.28906:	272.213265			1
4.33	4.36	4.39	19348.95703	230.342846			2
4.50	4.56	4.56	13423.65527:	106.694368			3
Height summation:			**65627.901367**				
Concentration		CF:	**203.083493**	L:			
Aroclor-1254						**28.70**	
5.03	5.05	5.09	77241.57812!	528.687701			2
5.27	5.30	5.33	31957.70312!	300.49784			3
5.34	5.37	5.40	72613.73437!	380.710428			4
Height summation:			**181813.015625**				
Concentration		CF:	**403.298656**	L:			
Aroclor-1260						**9.54**	
6.25	6.28	6.31	45315.42578	374.62419			4
6.55	6.58	6.61	80327.90625	311.2365			5
6.80	6.83	6.86	39321.19140	331.677424			6
Height summation:			**164964.523437**				
Concentration		CF:	**339.179371**	L:			

Analysis Report (B)

Injected on Nov 06, 2009 00:46:24
Instrument H5088B
Result file 3P7307B.45R
Calibration file 1P7307B
Method file PESTDB

%SSR(TCX) 96.7% Conc: 0.368686
%SSR(DCB) * 173.9% Conc: 0.658936

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1248						**18.50**	
3.14	3.17	3.20	7250.137207	213.680618			1
3.37	3.42	3.43	15025.07617:	146.767068			2
3.89	3.92	3.95	9203.025391	182.558584			4
Height summation:			**31478.23877**				
Concentration		CF:	**181.00209**	L:			
Aroclor-1254						**20.43**	
4.55	4.59	4.61	107205.3593:	640.189938			2
4.66	4.69	4.72	84789.11718	893.174136			3
4.76	4.79	4.82	103309.25	963.458993			4
Height summation:			**295303.726563**				
Concentration		CF:	**832.274356**	L:			
Aroclor-1260						**11.08**	
5.36	5.39	5.42	48538.09765	373.093861			4
5.60	5.62	5.66	120492.1484:	466.056945			5
5.89	5.92	5.95	74945.19531:	429.867666			6
Height summation:			**243975.441407**				
Concentration		CF:	**423.006157**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<62.1118	<12.4224			4	15	
Aroclor-1221			<62.1118	<12.4224			3	5	
Aroclor-1232			<62.1118	<12.4224			4	10	
Aroclor-1242			<62.1118	<12.4224			4	30	
Aroclor-1248	A	203.083493	62.1118	12.4224		11.50	4	30	
Aroclor-1254	A	403.298656	62.1118	12.4224		69.44	4	40	1269
Aroclor-1260	B	423.006157	62.1118	12.4224		22.00	4	40	

Units: ug/l

%Difference = High - Low divided by the Average times 100

Reviewed By: [signature] Date: 11-6-09

Verified By: [signature] Date: 11/9/09

ASX22 0855

11/6/2009

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822063** AF DF100 CARS1 Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 805 ml Total Volume: 1000 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 00:46:24
Instrument : CP07--H5088A
Result file : 3P7307.45R
Calibration file : 1P7307.CAL
Method file : PESTD.MET
%SSR(TCX) : *-1383.7% Conc.: -5.27801
%SSR(DCB) : *-1427.7% Conc.: -5.408259

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	64.21	
3.35	3.38	3.41	1079.554077	12.822567			1
3.83	3.86	3.89	8239.131836	80.708459			2
4.10	4.13	4.16	32855.28906:	150.122093			3
4.15	4.19	4.21	17569.44335!	180.576938			4
4.24	4.27	4.30	5085.767578	55.195233			5
4.62	4.64	4.68	11926.12109·	181.612715			6
Height Summation:			**76755.307007**				
Amount Avg CF:			**110.173001**	Linear:			
Aroclor-1221					3	19.36	
3.20	3.22	3.26	311.871429	6.236103			1
3.28	3.31	3.34	271.858398	7.705928			2
3.35	3.38	3.41	1079.554077	9.226108			3
Height Summation:			**1663.283904**				
Amount Avg CF:			**7.722713**	Linear:			
Aroclor-1232					6	67.34	
3.35	3.38	3.41	1079.554077	10.924695			1
3.83	3.86	3.89	8239.131836	181.390277			2
E 4.10	4.13	4.16	32855.28906:	313.640388			3
E 4.15	4.19	4.21	17569.44335!	377.308246			4
4.23	4.27	4.29	5085.767578	117.692319			5
E 4.62	4.64	4.68	11926.12109·	418.209097			6
Height Summation:			**76755.307007**				
Amount Avg CF:			**236.527504**	Linear:			
Aroclor-1242					6	63.24	
3.35	3.38	3.41	1079.554077	14.996488			1
3.83	3.86	3.89	8239.131836	98.77385			2
4.09	4.13	4.15	32855.28906:	173.562219			3
4.15	4.19	4.21	17569.44335!	211.128372			4
4.23	4.27	4.29	5085.767578	64.351655			5
4.62	4.64	4.68	11926.12109·	207.247076			6
Height Summation:			**76755.307007**				
Amount Avg CF:			**128.343277**	Linear:			
Aroclor-1248					6	122.94	
4.10	4.13	4.16	32855.28906:	272.213265			1
4.33	4.36	4.39	19348.95703·	230.342846			2
+ 4.50	4.53	4.56	12129.23730!	96.40603			3
4.50	4.56	4.56	13423.65527:	106.694368			3
4.62	4.64	4.68	11926.12109·	143.499746			4
5.02	5.05	5.08	77241.57812!	767.404522			5
E 5.61	5.64	5.67	92736.15625	1937.60313			6
Height Summation:			**247531.756836**				
Amount Avg CF:			**576.29298**	Linear:			
Aroclor-1254					6	39.31	
4.83	4.86	4.89	35839.01562!	370.844543			1
+ 4.83	4.88	4.89	15058.89843!	155.82209			1
5.03	5.05	5.09	77241.57812!	528.687701			2
5.27	5.30	5.33	31957.70312!	300.49784			3
5.34	5.37	5.40	72613.73437!	380.710428			4
5.85	5.87	5.91	75271.86718!	764.617869			5
5.98	6.01	6.04	98985.85156:	759.161335			6
Height Summation:			**391909.750001**				
Amount Avg CF:			**517.419953**	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 00:46:24
Instrument : CP07--H5088B
Result file : 3P7307B.45R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

A linear

%SSR(TCX) 96.7% Conc.: 0.368686
%SSR(DCB) *173.9% Conc.: 0.658936

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	137.81	
2.69	2.71	2.75	377.621521	6.082128			1
+ 2.69	2.74	2.75	325.934204	5.249631			1
2.98	3.01	3.04	6779.455078	49.193052			2
3.14	3.17	3.20	7250.137207	76.374458			3
* 3.42	3.42	3.48	15025.07617:	545.371668			5
* 3.37	3.42	3.43	15025.07617:	106.316137			4
+ 3.42	3.47	3.48	10017.06054:	363.593566			5
3.77	3.82	3.83	6649.952148	84.455308			6
Height Summation:			**51107.318298**				
Amount Avg CF:			**144.632125**	Linear:			
Aroclor-1221					2	72.67	
2.61	2.64	2.67	312.412476	13.677182			2
2.69	2.71	2.75	377.621521	4.39207			3
+ 2.69	2.74	2.75	325.934204	3.790901			3
Height Summation:			**690.033997**				
Amount Avg CF:			**9.034626**	Linear:			
Aroclor-1232					6	135.81	
2.69	2.71	2.75	377.621521	5.314621			1
+ 2.69	2.74	2.75	325.934204	4.587177			1
2.98	3.01	3.04	6779.455078	108.505873			2
3.14	3.17	3.20	7250.137207	163.247809			3
E* 3.42	3.42	3.48	15025.07617:	1210.42094			5
E* 3.38	3.42	3.44	15025.07617:	258.066933			4
E+ 3.42	3.47	3.48	10017.06054:	806.974933			5
3.77	3.82	3.83	6649.952148	207.766854			6
Height Summation:			**51107.318298**				
Amount Avg CF:			**325.553838**	Linear:			
Aroclor-1242					6	129.72	
2.68	2.71	2.74	377.621521	7.326748			1
2.98	3.01	3.04	6779.455078	60.379146			2
3.14	3.17	3.20	7250.137207	94.071252			3
E* 3.42	3.42	3.48	15025.07617:	598.707006			5
* 3.37	3.42	3.43	15025.07617:	137.558985			4
E+ 3.42	3.47	3.48	10017.06054:	399.151675			5
3.77	3.82	3.83	6649.952148	101.840721			6
Height Summation:			**51107.318298**				
Amount Avg CF:			**166.64731**	Linear:			
Aroclor-1248					6	137.68	
3.14	3.17	3.20	7250.137207	213.680618			1
3.37	3.42	3.43	15025.07617:	146.767068			2
3.77	3.82	3.83	6649.952148	61.747093			3
3.89	3.92	3.95	9203.025391	182.558584			4
4.26	4.29	4.32	61774.69140(	747.230189			5
E 4.55	4.59	4.61	107205.3593:	2140.30297			6
Height Summation:			**207108.241699**				
Amount Avg CF:			**582.047754**	Linear:			

ASX22: 0056

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:11:45

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822063** AF DF100 CARS1 Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 805 ml Total Volume: 1000 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 00:46:24
Instrument : CP07--H5088A
Result file : 3P7307.45R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	39.76	
5.74	5.77	5.80	87331.07812	531.377989			1
5.98	6.01	6.04	98985.85156	708.138917			2
6.05	6.08	6.11	29698.49023	268.526115			3
6.25	6.28	6.31	45315.42578	374.62419			4
6.55	6.58	6.61	80327.90625	311.2365			5
6.80	6.83	6.86	39321.19140	331.677424			6
Height Summation:			380979.943359				
Amount Avg CF:			420.930189	Linear:			
Aroclor-1262					6	50.50	
5.49	5.52	5.55	52472.77734	391.778125			1
5.73	5.77	5.79	87331.07812	625.932163			2
6.05	6.08	6.11	29698.49023	160.847942			3
6.25	6.28	6.31	45315.42578	315.457202			4
6.55	6.58	6.61	80327.90625	243.538669			5
6.79	6.83	6.85	39321.19140	231.581335			6
Height Summation:			334466.869140				
Amount Avg CF:			328.189239	Linear:			
Aroclor-1268					5	103.05	
6.05	6.08	6.11	29698.49023	406.512875			1
6.26	6.28	6.32	45315.42578	486.128828			2
6.79	6.83	6.85	39321.19140	116.842795			3
6.87	6.88	6.93	14664.14843	50.447929			4
7.03	7.06	7.09	790.157959	2.802369			5
Height Summation:			129789.413818				
Amount Avg CF:			212.546959	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 00:46:24
Instrument : CP07--H5088B
Result file : 3P7307B.45R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					6	50.64	
4.08	4.11	4.14	28567.22070	190.133252			1
4.55	4.59	4.61	107205.3593	640.189938			2
4.66	4.69	4.72	84789.11718	893.174136			3
4.76	4.79	4.82	103309.25	963.456993			4
E 4.85	4.87	4.91	105830.8984	1494.204151			5
5.11	5.14	5.17	126914.1171	871.075168			6
Height Summation:			556615.962892				
Amount Avg CF:			842.039273	Linear:			
Aroclor-1260					6	21.46	
4.66	4.69	4.72	84789.11718	561.075955			1
4.84	4.87	4.90	105830.8984	662.117913			2
4.96	4.98	5.02	52102.10156	440.399304			3
5.36	5.39	5.42	48538.09765	373.093861			4
5.60	5.62	5.66	120492.1484	466.056945			5
5.89	5.92	5.95	74945.19531	429.867666			6
Height Summation:			486697.558596				
Amount Avg CF:			488.768607	Linear:			
Aroclor-1262					6	57.11	
4.84	4.87	4.90	105830.8984	725.277676			1
5.36	5.39	5.42	48538.09765	267.015687			2
5.60	5.62	5.66	120492.1484	360.155669			3
5.84	5.87	5.90	18599.59570	153.955299			4
5.89	5.92	5.95	74945.19531	353.637588			5
6.35	6.38	6.41	32601.83593	236.55093			6
Height Summation:			401007.771486				
Amount Avg CF:			349.432142	Linear:			
Aroclor-1268					6	99.94	
5.35	5.39	5.41	48538.09765	443.787885			1
5.84	5.87	5.90	18599.59570	53.08349			2
5.90	5.92	5.96	74945.19531	243.890458			3
6.06	6.09	6.12	1194.103882	3.560076			4
6.14	6.17	6.20	3000.849121	38.190222			5
6.36	6.38	6.42	32601.83593	233.52074			6
Height Summation:			178879.677613				
Amount Avg CF:			169.338812	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			62.1118	12.4224		27.05	4	15	
Aroclor-1221			62.1118	12.4224		15.66	3	5	
Aroclor-1232			62.1118	12.4224	E	31.68	4	10	
Aroclor-1242			62.1118	12.4224		25.97	4	30	
Aroclor-1248			62.1118	12.4224	E	0.99	4	30	
Aroclor-1254			62.1118	12.4224		** 47.76	4	40	
Aroclor-1260			62.1118	12.4224		14.91	4	40	
Aroclor-1262			0	0		6.27	4	40	
Aroclor-1268			0	0		22.63	4	40	

Units: ug/L

ASX22 0057

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:11:45

Sample Name: **5822063AF DF100 ACCARS1 T 093060018A 00173**
Acquired from CP07--H5088A via port 3 on 11/6/09 12:55:32am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.45R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.509		0.0000	0.000	413.9	0.018	BB	0.019	359.69	0.024
2	1.664		0.0000	0.000	1517.6	0.066	BV	0.059	430.01	0.029
3	1.875		0.0000	0.000	7557.0	0.329	VV	0.123	1025.48	0.070
4	2.060		0.0000	0.000	16450.7	0.716	VV	0.056	4866.05	0.330
5	2.174		0.0000	0.000	4061.3	0.177	VV	0.066	1032.84	0.070
6	2.276		0.0000	0.000	1060.4	0.046	VB	0.052	339.82	0.023
7	2.573		0.0000	0.000	1336.5	0.058	BB	0.048	463.13	0.031
8	2.626		0.0000	0.000	5752.4	0.250	BB	0.022	4321.22	0.293
9	2.737		0.0000	0.000	141.1	0.006	BB	0.035	67.77	0.005
10	2.848		0.0000	0.000	1534.1	0.067	BB	0.054	470.28	0.032
11	3.000	TCX	-4248.7980	49.391	3359.3	0.146	BB	0.026	2113.62	0.144
12	3.067		0.0000	0.000	1144.7	0.050	BB	0.055	344.28	0.023
13	3.169		0.0000	0.000	709.6	0.031	BB	0.077	152.74	0.010
14	3.225		0.0000	0.000	583.8	0.025	BB	0.031	311.87	0.021
15	3.280		0.0000	0.000	59.6	0.003	BB	0.015	65.31	0.004
16	3.309		0.0000	0.000	297.8	0.013	BB	0.018	271.86	0.018
17	3.381		0.0000	0.000	1717.4	0.075	BB	0.027	1079.55	0.073
18	3.452		0.0000	0.000	1296.5	0.056	BB	0.025	868.50	0.059
19	3.514		0.0000	0.000	1995.3	0.087	BB	0.026	1266.75	0.086
20	3.597		0.0000	0.000	509.1	0.022	BB	0.036	236.73	0.016
21	3.715		0.0000	0.000	13992.8	0.609	BB	0.037	6296.25	0.427
22	3.772		0.0000	0.000	9593.7	0.417	BB	0.026	6049.49	0.411
23	3.827		0.0000	0.000	109.7	0.005	BB	0.015	121.90	0.008
24	3.864		0.0000	0.000	13004.2	0.566	BB	0.026	8239.13	0.559
25	3.962		0.0000	0.000	1839.5	0.080	BB	0.024	1274.23	0.087
26	4.046		0.0000	0.000	28569.5	1.243	BB	0.030	15628.07	1.061
27	4.132		0.0000	0.000	58726.5	2.555	BB	0.030	32855.29	2.231
28	4.186		0.0000	0.000	25616.9	1.114	BB	0.024	17569.44	1.193
29	4.265		0.0000	0.000	7546.2	0.328	BB	0.025	5085.77	0.345
30	4.315		0.0000	0.000	6381.5	0.278	BB	0.021	5124.13	0.348
31	4.359		0.0000	0.000	22323.5	0.971	BB	0.019	19348.96	1.314
32	4.392		0.0000	0.000	28335.8	1.233	BB	0.017	27239.39	1.849
33	4.421		0.0000	0.000	43926.0	1.911	BB	0.020	37217.00	2.527
34	4.526		0.0000	0.000	14166.4	0.616	BB	0.019	12129.24	0.824
35	4.556		0.0000	0.000	14327.8	0.623	BB	0.018	13423.66	0.911
36	4.598		0.0000	0.000	36833.6	1.602	BB	0.021	28795.25	1.955
37	4.644		0.0000	0.000	16206.2	0.705	BB	0.023	11926.12	0.810
38	4.700		0.0000	0.000	5266.9	0.229	BB	0.020	4410.02	0.299
39	4.731		0.0000	0.000	1158.4	0.050	BB	0.014	1333.18	0.091
40	4.807		0.0000	0.000	14358.2	0.625	BB	0.028	8605.96	0.584
41	4.856		0.0000	0.000	41010.7	1.784	BB	0.019	35839.02	2.433
42	4.885		0.0000	0.000	11870.7	0.516	BB	0.013	15058.90	1.022
43	4.913		0.0000	0.000	79555.6	3.461	BB	0.020	65277.36	4.432
44	5.001		0.0000	0.000	21029.7	0.915	BB	0.020	17300.58	1.175
45	5.054		0.0000	0.000	120010.6	5.221	BB	0.026	77241.58	5.244
46	5.097		0.0000	0.000	61280.5	2.666	BB	0.021	49091.65	3.333
47	5.149		0.0000	0.000	2738.8	0.119	BB	0.015	3054.00	0.207
48	5.175		0.0000	0.000	5371.1	0.234	BB	0.016	5445.89	0.370
49	5.231		0.0000	0.000	37575.7	1.635	BB	0.025	24974.93	1.696
50	5.299		0.0000	0.000	47228.1	2.055	BB	0.025	31957.70	2.170
51	5.332		0.0000	0.000	6549.9	0.285	BB	0.013	8594.87	0.584
52	5.373		0.0000	0.000	100984.3	4.393	BB	0.023	72613.73	4.930
53	5.428		0.0000	0.000	29549.3	1.286	BB	0.026	18827.59	1.278
54	5.474		0.0000	0.000	10596.1	0.461	BB	0.020	8864.75	0.602

ASX22: 0050

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.520		0.0000	0.000	82617.1	3.594	BB	0.026	52472.78	3.563
56	5.588		0.0000	0.000	11658.5	0.507	BB	0.019	10040.30	0.682
57	5.637		0.0000	0.000	145350.9	6.323	BB	0.026	92736.16	6.296
58	5.702		0.0000	0.000	22716.0	0.988	BB	0.024	15719.02	1.067
59	5.766		0.0000	0.000	184720.6	8.036	BB	0.035	87331.08	5.929
60	5.875		0.0000	0.000	114123.0	4.965	BB	0.025	75271.87	5.111
61	5.919		0.0000	0.000	1460.7	0.064	BB	0.011	2193.00	0.149
62	5.964		0.0000	0.000	8038.7	0.350	BB	0.023	5705.25	0.387
63	6.011		0.0000	0.000	172474.9	7.504	BB	0.029	98985.85	6.721
64	6.082		0.0000	0.000	38631.3	1.681	BB	0.022	29698.49	2.016
65	6.134		0.0000	0.000	27728.0	1.206	BB	0.024	19024.48	1.292
66	6.276		0.0000	0.000	111435.7	4.848	BB	0.041	45315.43	3.077
67	6.351		0.0000	0.000	24446.3	1.064	BB	0.025	16456.42	1.117
68	6.400		0.0000	0.000	10134.4	0.441	BB	0.022	7854.27	0.533
69	6.444		0.0000	0.000	980.8	0.043	BB	0.019	874.38	0.059
70	6.512		0.0000	0.000	35271.4	1.534	BB	0.027	22141.47	1.503
71	6.578		0.0000	0.000	133240.7	5.797	BB	0.028	80327.91	5.454
72	6.730		0.0000	0.000	741.4	0.032	BB	0.026	483.81	0.033
73	6.833		0.0000	0.000	66540.7	2.895	BB	0.028	39321.19	2.670
74	6.884		0.0000	0.000	34761.5	1.512	BB	0.040	14664.15	0.996
75	7.057		0.0000	0.000	1341.3	0.058	BB	0.028	790.16	0.054
76	7.181		0.0000	0.000	27061.0	1.177	BB	0.044	10322.89	0.701
77	7.416		0.0000	0.000	36512.4	1.588	BB	0.030	20331.29	1.380
78	7.489		0.0000	0.000	2603.2	0.113	BB	0.042	1035.53	0.070
79	7.584		0.0000	0.000	890.8	0.039	BB	0.098	151.97	0.010
80	7.730		0.0000	0.000	10264.3	0.447	BB	0.033	5174.58	0.351
81	7.820		0.0000	0.000	1288.7	0.056	BB	0.029	731.83	0.050
82	7.978	DCB	-4353.6490	50.609	1940.6	0.084	BB	0.045	713.80	0.048
83	8.148		0.0000	0.000	476.4	0.021	BB	0.070	112.88	0.008

Total Area = 2298584.0, Total Amount = -8602.447, Total Height = 1472885.0, Sample Units = ug/l

ASX22 0059

Sample Name: **5822063AF DF100 ACCARS1 T 093060018A 00173**
Acquired from CP07--H5088B via port 4 on 11/6/09 12:55:32am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.45R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.573		0.0000	0.000	802.1	0.026	BV	0.032	424.04	0.029
2	1.617		0.0000	0.000	12548.4	0.406	VV	0.053	3956.24	0.271
3	1.884		0.0000	0.000	217.5	0.007	VB	0.058	62.93	0.004
4	1.961		0.0000	0.000	110.0	0.004	BB	0.075	24.32	0.002
5	2.043		0.0000	0.000	8360.2	0.270	BB	0.036	3881.48	0.266
6	2.160		0.0000	0.000	862.3	0.028	BB	0.033	439.78	0.030
7	2.300		0.0000	0.000	206.1	0.007	BB	0.111	31.07	0.002
8	2.340	TCX	296.7925	35.878	2170.9	0.070	BB	0.028	1287.14	0.088
9	2.391		0.0000	0.000	3251.2	0.105	BB	0.059	924.50	0.063
10	2.610		0.0000	0.000	267.0	0.009	BB	0.262	16.98	0.001
11	2.644		0.0000	0.000	640.3	0.021	BB	0.034	312.41	0.021
12	2.714		0.0000	0.000	558.9	0.018	BB	0.025	377.62	0.026
13	2.744		0.0000	0.000	495.7	0.016	BB	0.025	325.93	0.022
14	2.860		0.0000	0.000	3520.0	0.114	BB	0.038	1532.45	0.105
15	3.011		0.0000	0.000	14375.0	0.465	BB	0.035	6779.46	0.465
16	3.072		0.0000	0.000	8058.9	0.261	BB	0.031	4311.59	0.296
17	3.168		0.0000	0.000	17139.5	0.554	BB	0.039	7250.14	0.497
18	3.319		0.0000	0.000	31509.3	1.019	BB	0.046	11516.42	0.790
19	3.422		0.0000	0.000	37622.2	1.217	BB	0.042	15025.08	1.030
20	3.465		0.0000	0.000	15154.4	0.490	BB	0.025	10017.06	0.687
21	3.556		0.0000	0.000	7589.9	0.245	BB	0.029	4436.83	0.304
22	3.617		0.0000	0.000	9166.7	0.296	BB	0.037	4092.42	0.281
23	3.664		0.0000	0.000	165267.5	5.344	BB	0.044	62435.78	4.282
24	3.820		0.0000	0.000	21267.1	0.688	BB	0.053	6649.95	0.456
25	3.853		0.0000	0.000	39139.0	1.266	BB	0.029	22155.74	1.519
26	3.921		0.0000	0.000	14581.2	0.471	BB	0.026	9203.03	0.631
27	3.971		0.0000	0.000	4898.3	0.158	BB	0.026	3110.92	0.213
28	4.038		0.0000	0.000	8671.2	0.280	BB	0.031	4592.42	0.315
29	4.110		0.0000	0.000	42429.1	1.372	BB	0.025	28567.22	1.959
30	4.145		0.0000	0.000	67102.4	2.170	BB	0.029	38909.27	2.668
31	4.196		0.0000	0.000	6345.8	0.205	BB	0.017	6256.56	0.429
32	4.248		0.0000	0.000	63594.3	2.056	BB	0.021	51096.82	3.504
33	4.286		0.0000	0.000	101367.7	3.278	BB	0.027	61774.69	4.237
34	4.384		0.0000	0.000	3347.6	0.108	BB	0.019	2873.59	0.197
35	4.429		0.0000	0.000	46968.1	1.519	BB	0.032	24455.01	1.677
36	4.502		0.0000	0.000	37695.0	1.219	BB	0.023	27225.13	1.867
37	4.536		0.0000	0.000	13161.2	0.426	BB	0.017	13073.83	0.897
38	4.590		0.0000	0.000	276957.0	8.956	BB	0.043	107205.40	7.352
39	4.691		0.0000	0.000	149004.9	4.818	BB	0.029	84789.12	5.815
40	4.744		0.0000	0.000	2981.9	0.096	BB	0.015	3313.02	0.227
41	4.789		0.0000	0.000	235200.8	7.605	BB	0.038	103309.30	7.085
42	4.871		0.0000	0.000	213530.3	6.905	BB	0.034	105830.90	7.258
43	4.983		0.0000	0.000	133869.9	4.329	BB	0.043	52102.10	3.573
44	5.044		0.0000	0.000	89018.9	2.878	BB	0.027	55474.34	3.804
45	5.093		0.0000	0.000	10227.6	0.331	BB	0.017	9827.45	0.674
46	5.140		0.0000	0.000	332732.4	10.759	BB	0.044	126914.10	8.704
47	5.217		0.0000	0.000	12372.1	0.400	BB	0.020	10265.49	0.704
48	5.335		0.0000	0.000	6883.9	0.223	BB	0.022	5109.97	0.350
49	5.391		0.0000	0.000	141950.5	4.590	BB	0.049	48538.10	3.329
50	5.471		0.0000	0.000	21953.3	0.710	BB	0.022	16278.49	1.116
51	5.509		0.0000	0.000	6601.7	0.213	BB	0.019	5745.16	0.394
52	5.560		0.0000	0.000	4079.0	0.132	BB	0.018	3774.43	0.259
53	5.625		0.0000	0.000	293966.2	9.506	BB	0.041	120492.20	8.264
54	5.794		0.0000	0.000	817.5	0.026	BB	0.022	624.26	0.043

ASX22 0060

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.867		0.0000	0.000	28144.4	0.910	BB	0.025	18599.60	1.276
56	5.921		0.0000	0.000	166828.4	5.395	BB	0.037	74945.20	5.140
57	6.087		0.0000	0.000	2337.2	0.076	BB	0.033	1194.10	0.082
58	6.172		0.0000	0.000	5481.1	0.177	BB	0.030	3000.85	0.206
59	6.237		0.0000	0.000	29302.5	0.948	BB	0.036	13493.20	0.925
60	6.382		0.0000	0.000	78491.8	2.538	BB	0.040	32601.84	2.236
61	6.620		0.0000	0.000	20583.2	0.666	BB	0.039	8899.44	0.610
62	6.811	DCB	530.4431	64.122	4824.5	0.156	BB	0.044	1831.20	0.126
63	7.054		0.0000	0.000	245.1	0.008	BB	0.030	136.19	0.009
64	7.294		0.0000	0.000	30.5	0.001	BB	0.014	37.31	0.003
65	7.371		0.0000	0.000	286.8	0.009	BB	0.115	41.48	0.003
66	7.454		0.0000	0.000	8588.9	0.278	BB	0.038	3756.43	0.258
67	7.567		0.0000	0.000	4797.1	0.155	BB	0.136	588.23	0.040

Total Area = 3092552.0, Total Amount = 827.236, Total Height = 1458125.0, Sample Units = ug/l

ASX22 BB61

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\3P7307.45R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 12:46:23 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.45R



ASX22 0062

Instrument ID: CP07--H5088B Injected On: 11/6/2009 12:46:23 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 805
Analyst: 2280

Dilution Factor: 1000

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.	2114	-5.27801	TCX	2.34	1287	.36869	TCX
7.978	714	-5.40826	DCB	6.811	1831	.65894	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.45A
Area File: C:\CPWIN\DATA1\3P7307B.45A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 12:55:40 AM
File Reported On: 11/6/2009 at 12:55:51 AM

10X REPLOTS

ASX22 0063

C:\CPWIN\DATA1\3P7307.45R



C:\CPWIN\DATA1\3P7307B.45R



ASX22 0064

Multiple Component Data Summary

Sample Name: 5822064 RI AF DF20 CARS2 Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 893 ml Total Volume: 200 ml Analyst: 2131 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on Nov 06, 2009 00:59:28
Instrument H5088A
Result file 3P7307.46R
Calibration file 1P7307
Method file PESTD

%SSR(TCX) * 0.0% Conc: 0
%SSR(DCB) * 0.0% Conc: 0

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1248						**25.95**	
4.10	4.13	4.16	18039.703125	26.946837			1
4.33	4.36	4.39	16150.955078	34.664887			2
4.50	4.52	4.56	14291.775390	20.480064			3
Height summation:			**48482.433594**				
Concentration		CF:	**27.363929**	L:			
Aroclor-1254						**13.94**	
5.03	5.05	5.09	83651.382813	103.227548			2
5.27	5.30	5.33	46001.054688	77.984448			3
5.34	5.37	5.40	99825.296875	94.360673			4
Height summation:			**229477.734376**				
Concentration		CF:	**91.857556**	L:			
Aroclor-1260						**12.37**	
6.25	6.27	6.31	56920.84375	84.838992			4
6.55	6.58	6.61	95158.210938	66.472921			5
6.80	6.83	6.86	48318.148438	73.480811			6
Height summation:			**200397.203126**				
Concentration		CF:	**74.930908**	L:			

Analysis Report (B)

Injected on Nov 06, 2009 00:59:28
Instrument H5088B
Result file 3P7307B.46R
Calibration file 1P7307B
Method file PESTDB

%SSR(TCX) 102.8% Conc: 0.353444
%SSR(DCB) 97.6% Conc: 0.333327

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1248						**25.61**	
3.14	3.17	3.20	4200.269531	22.318782			1
3.37	3.42	3.43	8088.845703	14.245343			2
3.89	3.92	3.95	4244.121094	15.178674			4
Height summation:			**16533.236328**				
Concentration		CF:	**17.2476**	L:			
Aroclor-1254						**61.58**	
4.08	4.11	4.14	38020.855469	45.623268			1
4.55	4.59	4.61	98256.90625	105.786403			2
4.66	4.69	4.72	95992.390625	182.308639			3
Height summation:			**232270.152344**				
Concentration		CF:	**111.239437**	L:			
Aroclor-1260						**9.65**	
5.36	5.41	5.42	60672.230469	84.08136			4
5.60	5.62	5.66	146317	102.035064			5
5.89	5.92	5.95	89846.695313	92.91108			6
Height summation:			**296835.925782**				
Concentration		CF:	**93.009168**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<11.1982	<2.2396			4	15	
Aroclor-1221			<11.1982	<2.2396			3	5	
Aroclor-1232			<11.1982	<2.2396			4	10	
Aroclor-1242			<11.1982	<2.2396			4	30	
Aroclor-1248	B	17.2476	11.1982	2.2396		45.35	4	30	1529
Aroclor-1254	B	111.239437	11.1982	2.2396		19.09	4	40	
Aroclor-1260	B	93.009168	11.1982	2.2396		21.53	4	40	

Units: ug/l

%Difference = High - Low divided by the Average times 100

Reviewed By: [signature] Date: 11-6-09
Verified By: [signature] Date: 11/9/09

ASX22 9865

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822064** RI AF DF20 CARS2 Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 893 ml Total Volume: 200 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173 00516 06030 07394 10227

Analysis Report (A)

Injected on : NOV 06, 2009 00:59:28
Instrument : CP07--H5088A
Result file : 3P7307.46R
Calibration file : 1P7307.CAL
Method file : PESTD.MET
%SSR(TCX) : *-161% Conc.: -0.553478
%SSR(DCB) : *-225.7% Conc.: -0.770806

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					5	56.92	
3.83	3.86	3.89	4950.303223	8.742665			2
4.10	4.13	4.16	18039.70312	14.860832			3
4.15	4.19	4.21	8941.068359	16.567925			4
4.24	4.26	4.30	3736.44165	7.311015			5
4.62	4.65	4.68	10689.51171	29.348055			6
Height Summation:			46357.028076				
Amount Avg CF:			15.366098	Linear:			
Aroclor-1221					2	134.98	
3.20	3.22	3.26	231.64827	0.835105			1
3.28	3.28	3.34	7013.565918	35.842269			2
Height Summation:			7245.214188				
Amount Avg CF:			18.338687	Linear:			
Aroclor-1232					5	60.84	
3.83	3.86	3.89	4950.303223	19.648924			2
4.10	4.13	4.16	18039.70312	31.047776			3
4.15	4.19	4.21	8941.068359	34.618012			4
4.23	4.26	4.29	3736.44165	15.589215			5
E 4.62	4.65	4.68	10689.51171	67.581301			6
Height Summation:			46357.028076				
Amount Avg CF:			33.697046	Linear:			
Aroclor-1242					5	54.99	
3.83	3.86	3.89	4950.303223	10.699581			2
4.09	4.13	4.15	18039.70312	17.181208			3
4.15	4.19	4.21	8941.068359	19.371017			4
4.23	4.26	4.29	3736.44165	8.523851			5
4.62	4.65	4.68	10689.51171	33.490489			6
Height Summation:			46357.028076				
Amount Avg CF:			17.853229	Linear:			
Aroclor-1248					6	146.89	
4.10	4.13	4.16	18039.70312	26.946837			1
4.33	4.36	4.39	16150.95507	34.664887			2
4.50	4.52	4.56	14291.77539	20.480964			3
+ 4.50	4.55	4.56	4687.961914	6.717833			3
4.62	4.65	4.68	10689.51171	23.189116			4
5.02	5.05	5.08	83651.38281	149.837583			5
E 5.61	5.64	5.67	123724.7421	466.065237			6
Height Summation:			266548.070314				
Amount Avg CF:			120.197287	Linear:			
Aroclor-1254					6	47.98	
4.83	4.86	4.89	36312.74609	67.743765			1
+ 4.83	4.88	4.89	17890.44335	33.375774			1
5.03	5.05	5.09	83651.38281	103.227548			2
5.27	5.30	5.33	46001.05468	77.984448			3
5.34	5.37	5.40	99825.29687	94.360673			4
E 5.85	5.87	5.91	114846.1640	210.330699			5
5.98	6.01	6.04	130691.9843	180.710891			6
Height Summation:			511328.628908				
Amount Avg CF:			122.393004	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 00:59:28
Instrument : CP07--H5088B
Result file : 3P7307B.46R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET
%SSR(TCX) : 102.8% Conc.: 0.353444
%SSR(DCB) : 97.6% Conc.: 0.333327

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	72.17	
2.69	2.74	2.75	10616.82226	30.829592			1
2.98	3.01	3.04	4539.249512	5.938373			2
3.14	3.17	3.20	4200.269531	7.977255			3
3.37	3.42	3.43	8088.845703	10.319139			4
3.42	3.46	3.48	4264.824219	27.909461			5
3.77	3.80	3.83	3828.759766	8.766797			6
+ 3.77	3.82	3.83	2321.437744	5.315448			6
Height Summation:			35538.770997				
Amount Avg CF:			15.290103	Linear:			
Aroclor-1221					1		
2.69	2.74	2.75	10616.82226	22.262888			3
Height Summation:			10616.822266				
Amount Avg CF:			22.262888	Linear:			
Aroclor-1232					6	63.66	
2.69	2.74	2.75	10616.82226	26.939192			1
2.98	3.01	3.04	4539.249512	13.09836			2
3.14	3.17	3.20	4200.269531	17.051112			3
3.38	3.42	3.44	8088.845703	25.048207			4
E 3.42	3.46	3.48	4264.824219	61.943437			5
3.77	3.80	3.83	3828.759766	21.567027			6
+ 3.77	3.82	3.83	2321.437744	13.07643			6
Height Summation:			35538.770997				
Amount Avg CF:			27.607889	Linear:			
Aroclor-1242					6	90.23	
2.68	2.74	2.74	10616.82226	37.138426			1
2.98	3.01	3.04	4539.249512	7.28871			2
3.14	3.17	3.20	4200.269531	9.825672			3
E* 3.42	3.42	3.48	8088.845703	58.111037			5
* 3.37	3.42	3.43	8088.845703	13.351598			4
+ 3.42	3.46	3.48	4264.824219	30.638903			5
3.77	3.80	3.83	3828.759766	10.571472			6
+ 3.77	3.82	3.83	2321.437744	6.409651			6
Height Summation:			39362.792481				
Amount Avg CF:			22.714486	Linear:			
Aroclor-1248					6	149.94	
3.14	3.17	3.20	4200.269531	22.316782			1
3.37	3.42	3.43	8088.845703	14.245343			2
3.77	3.80	3.83	3828.759766	6.409594			3
+ 3.77	3.82	3.83	2321.437744	3.886239			3
3.89	3.92	3.95	4244.121094	15.178674			4
4.26	4.29	4.32	63868.52343	139.285247			5
E 4.55	4.59	4.61	98256.90625	353.668402			6
Height Summation:			182487.425782				
Amount Avg CF:			91.851007	Linear:			

ASX22 8866

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:14:23

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822064** RI AF DF20 CARS2 Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 893 ml Total Volume: 200 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173 00516 06030 07394 10227

Analysis Report (A)

Injected on : NOV 06, 2009 00:59:28
Instrument : CP07–H5088A
Result file : 3P7307.46R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	45.82	
5.74	5.76	5.80	102275.1718	112.196517			1
5.98	6.01	6.04	130691.9843	168.565506			2
6.05	6.08	6.11	31092.14453	50.684743			3
6.25	6.27	6.31	56920.84375	84.838992			4
6.55	6.58	6.61	95158.21093	66.472921			5
6.80	6.83	6.86	48318.14843	73.480811			6
Height Summation:			464456.503907				
Amount Avg CF:			92.706582	Linear:			
Aroclor-1262					6	51.34	
5.49	5.52	5.55	52307.76562	70.412006			1
5.73	5.76	5.79	102275.1718	132.160929			2
6.05	6.08	6.11	31092.14453	30.360312			3
6.25	6.27	6.31	56920.84375	71.439783			4
6.55	6.58	6.61	95158.21093	52.014229			5
6.79	6.83	6.85	48318.14843	51.305224			6
Height Summation:			386072.285157				
Amount Avg CF:			67.948747	Linear:			
Aroclor-1268					5	103.74	
6.05	6.08	6.11	31092.14453	76.72997			1
6.26	6.27	6.32	56920.84375	110.090808			2
6.79	6.83	6.85	48318.14843	25.885703			3
6.87	6.89	6.93	18368.26953	11.392765			4
7.03	7.06	7.09	1091.349609	0.69783			5
Height Summation:			155790.755859				
Amount Avg CF:			44.959415	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 00:59:28
Instrument : CP07–H5088B
Result file : 3P7307B.46R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					6	52.77	
4.08	4.11	4.14	38020.85546	45.628268			1
+ 4.08	4.14	4.14	20765.37695	24.917492			1
4.55	4.59	4.61	98256.90625	105.786403			2
4.66	4.69	4.72	95992.39062	182.308639			3
E 4.76	4.79	4.82	144433.3437	242.84877			4
E 4.85	4.87	4.91	123981.3906	315.593737			5
E 5.11	5.14	5.17	164338.2656	203.356855			6
Height Summation:			665023.152344				
Amount Avg CF:			182.586279	Linear:			
Aroclor-1260					6	18.76	
4.66	4.69	4.72	95992.39062	114.523014			1
4.84	4.87	4.90	123981.3906	139.8472			2
4.96	4.98	5.02	64210.82421	97.852977			3
5.36	5.41	5.42	60672.23046	84.08136			4
5.60	5.62	5.66	146317	102.035064			5
5.89	5.92	5.95	89846.69531	92.91108			6
Height Summation:			581020.531251				
Amount Avg CF:			105.208449	Linear:			
Aroclor-1262					6	57.55	
4.84	4.87	4.90	123981.3906	153.187295			1
5.36	5.41	5.42	60672.23046	60.175319			2
5.60	5.62	5.66	146317	78.849821			3
5.84	5.87	5.90	19132.91992	28.552675			4
5.89	5.92	5.95	89846.69531	76.434803			5
6.35	6.38	6.41	37642.26171	49.241665			6
Height Summation:			477592.498048				
Amount Avg CF:			74.40693	Linear:			
Aroclor-1268					6	102.08	
5.35	5.41	5.41	60672.23046	100.013141			1
5.84	5.87	5.90	19132.91992	9.844907			2
5.90	5.92	5.96	89846.69531	52.71419			3
6.06	6.09	6.12	1213.047852	0.652033			4
6.14	6.17	6.20	4424.620117	10.15216			5
6.36	6.38	6.42	37642.26171	48.610885			6
Height Summation:			212931.775392				
Amount Avg CF:			36.997886	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			11.1982	2.2396		0.50	4	15	
Aroclor-1221			11.1982	2.2396		19.33	3	5	
Aroclor-1232			11.1982	2.2396	E	19.87	4	10	
			11.1982	4.4793	E				
Aroclor-1242			11.1982	2.2396		23.97	4	30	
Aroclor-1248			11.1982	2.2396	E	26.74	4	30	
Aroclor-1254			11.1982	2.2396	E	39.47	4	40	
Aroclor-1260			11.1982	2.2396		12.63	4	40	
Aroclor-1262			11.1982	2.2396		9.07	4	40	
			11.1982	4.4793					

ASX22 0067

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:14:23

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822064** RI AF DF20 CARS2 Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 893 ml Total Volume: 200 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173 00516 06030 07394 10227

Analysis Report (A)

Injected on	: NOV 06, 2009 00:59:28
Instrument	: CP07--H5088A
Result file	: 3P7307.46R
Calibration file	: 1P7307.CAL
Method file	: PESTD.MET

Analysis Report (B)

Injected on	: NOV 06, 2009 00:59:28
Instrument	: CP07--H5088B
Result file	: 3P7307B.46R
Calibration file	: 1P7307B.CAL
Method file	: PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1268			11.1982	2.2396		19.43	4	40	

Units: ug/L

ASX22 8068

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:14:23

Sample Name: **5822064 RI AF DF20 ACCARS2 T 093060018A 00173**
Acquired from CP07--H5088A via port 3 on 11/6/09 01:08:36am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.46R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.511		0.0000	0.000	8100.7	0.283	BB	0.022	6200.69	0.351
2	1.659		0.0000	0.000	55699.6	1.944	BV	0.036	25579.48	1.449
3	1.768		0.0000	0.000	12164.4	0.425	VV	0.069	2942.46	0.167
4	1.879		0.0000	0.000	24033.4	0.839	VV	0.075	5358.82	0.304
5	2.029		0.0000	0.000	11274.8	0.394	VV	0.042	4465.47	0.253
6	2.062		0.0000	0.000	71892.7	2.509	VV	0.033	36858.07	2.088
7	2.177		0.0000	0.000	10091.8	0.352	VV	0.045	3704.23	0.210
8	2.278		0.0000	0.000	22410.1	0.782	VV	0.034	10963.08	0.621
9	2.447		0.0000	0.000	1196.8	0.042	VV	0.045	438.41	0.025
10	2.480		0.0000	0.000	838.9	0.029	VV	0.034	414.84	0.023
11	2.529		0.0000	0.000	2088.1	0.073	VV	0.038	920.60	0.052
12	2.574		0.0000	0.000	1175.3	0.041	VB	0.025	775.03	0.044
13	2.627		0.0000	0.000	11964.7	0.418	BB	0.022	9243.43	0.524
14	2.745		0.0000	0.000	10366.2	0.362	BB	0.028	6261.67	0.355
15	2.813		0.0000	0.000	80.9	0.003	BB	0.021	63.13	0.004
16	2.855		0.0000	0.000	874.5	0.031	BB	0.030	491.25	0.028
17	2.952		0.0000	0.000	191.3	0.007	BB	0.020	158.00	0.009
18	3.001	TCX	-494.2558	41.795	13763.6	0.480	BB	0.024	9371.40	0.531
19	3.053		0.0000	0.000	16456.6	0.574	BB	0.028	9809.80	0.556
20	3.183		0.0000	0.000	705.2	0.025	BB	0.023	505.00	0.029
21	3.219		0.0000	0.000	187.1	0.007	BB	0.013	231.65	0.013
22	3.281		0.0000	0.000	17270.4	0.603	BB	0.041	7013.57	0.397
23	3.344		0.0000	0.000	10628.1	0.371	BB	0.027	6484.87	0.367
24	3.457		0.0000	0.000	3435.2	0.120	BB	0.027	2114.54	0.120
25	3.512		0.0000	0.000	2603.1	0.091	BB	0.023	1893.27	0.107
26	3.596		0.0000	0.000	220.4	0.008	BB	0.023	160.59	0.009
27	3.639		0.0000	0.000	1041.0	0.036	BB	0.022	781.66	0.044
28	3.694		0.0000	0.000	961.4	0.034	BB	0.020	814.70	0.046
29	3.714		0.0000	0.000	1683.2	0.059	BB	0.018	1592.49	0.090
30	3.773		0.0000	0.000	4561.4	0.159	BB	0.029	2629.03	0.149
31	3.864		0.0000	0.000	8172.9	0.285	BB	0.028	4950.30	0.280
32	3.927		0.0000	0.000	346.7	0.012	BB	0.027	216.12	0.012
33	3.966		0.0000	0.000	967.1	0.034	BB	0.030	540.83	0.031
34	4.046		0.0000	0.000	15173.5	0.530	BB	0.032	7788.14	0.441
35	4.131		0.0000	0.000	33911.1	1.184	BB	0.031	18039.70	1.022
36	4.186		0.0000	0.000	12394.4	0.433	BB	0.023	8941.07	0.506
37	4.265		0.0000	0.000	5935.6	0.207	BB	0.026	3736.44	0.212
38	4.315		0.0000	0.000	2611.3	0.091	BB	0.021	2108.96	0.119
39	4.359		0.0000	0.000	18786.9	0.656	BB	0.019	16150.96	0.915
40	4.393		0.0000	0.000	14620.7	0.510	BB	0.017	14421.38	0.817
41	4.424		0.0000	0.000	30180.9	1.053	BB	0.023	21635.36	1.225
42	4.525		0.0000	0.000	17094.5	0.597	BB	0.020	14291.78	0.809
43	4.555		0.0000	0.000	4578.3	0.160	BB	0.016	4687.96	0.266
44	4.598		0.0000	0.000	18348.4	0.640	BB	0.021	14392.16	0.815
45	4.646		0.0000	0.000	14905.3	0.520	BB	0.023	10689.51	0.605
46	4.699		0.0000	0.000	4145.5	0.145	BB	0.020	3470.04	0.197
47	4.732		0.0000	0.000	3551.7	0.124	BB	0.026	2279.25	0.129
48	4.811		0.0000	0.000	11928.4	0.416	BB	0.031	6314.30	0.358
49	4.856		0.0000	0.000	43072.8	1.503	BB	0.020	36312.75	2.057
50	4.883		0.0000	0.000	14333.2	0.500	BB	0.013	17890.44	1.013
51	4.912		0.0000	0.000	47757.9	1.667	BB	0.017	47151.21	2.671
52	5.001		0.0000	0.000	20199.0	0.705	BB	0.021	16106.51	0.912
53	5.054		0.0000	0.000	128247.8	4.476	BB	0.026	83651.38	4.738
54	5.096		0.0000	0.000	53734.7	1.876	BB	0.020	45129.92	2.556

ASX22. 8869

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.148		0.0000	0.000	1208.9	0.042	BB	0.012	1615.31	0.091
56	5.175		0.0000	0.000	8057.4	0.281	BB	0.018	7645.08	0.433
57	5.230		0.0000	0.000	50652.3	1.768	BB	0.025	33388.37	1.891
58	5.298		0.0000	0.000	63725.0	2.224	BB	0.023	46001.05	2.605
59	5.330		0.0000	0.000	9143.0	0.319	BB	0.013	11299.52	0.640
60	5.373		0.0000	0.000	141753.5	4.948	BB	0.024	99825.30	5.654
61	5.428		0.0000	0.000	27054.6	0.944	BB	0.026	17669.29	1.001
62	5.473		0.0000	0.000	17851.8	0.623	BB	0.020	15002.02	0.850
63	5.519		0.0000	0.000	84411.9	2.946	BB	0.027	52307.77	2.963
64	5.587		0.0000	0.000	13663.0	0.477	BB	0.019	11691.14	0.662
65	5.637		0.0000	0.000	195173.6	6.812	BB	0.026	123724.70	7.007
66	5.701		0.0000	0.000	30371.6	1.060	BB	0.024	21391.31	1.212
67	5.765		0.0000	0.000	215562.7	7.524	BB	0.035	102275.20	5.793
68	5.874		0.0000	0.000	173123.6	6.043	BB	0.025	114846.20	6.505
69	5.919		0.0000	0.000	2001.2	0.070	BB	0.011	3091.45	0.175
70	5.963		0.0000	0.000	10013.8	0.350	BB	0.023	7211.42	0.408
71	6.011		0.0000	0.000	222292.0	7.759	BB	0.028	130692.00	7.402
72	6.082		0.0000	0.000	41399.8	1.445	BB	0.022	31092.14	1.761
73	6.134		0.0000	0.000	32001.5	1.117	BB	0.025	21165.71	1.199
74	6.274		0.0000	0.000	143428.3	5.006	BB	0.042	56920.84	3.224
75	6.351		0.0000	0.000	19065.3	0.665	BB	0.021	15131.19	0.857
76	6.395		0.0000	0.000	65286.9	2.279	BB	0.033	32623.53	1.848
77	6.511		0.0000	0.000	49746.5	1.736	BB	0.026	31995.60	1.812
78	6.577		0.0000	0.000	155152.0	5.415	BB	0.027	95158.21	5.390
79	6.729		0.0000	0.000	509.3	0.018	BB	0.023	369.56	0.021
80	6.832		0.0000	0.000	80537.8	2.811	BB	0.028	48318.15	2.737
81	6.886		0.0000	0.000	41004.4	1.431	BB	0.037	18368.27	1.040
82	6.986		0.0000	0.000	420.2	0.015	BB	0.027	262.23	0.015
83	7.058		0.0000	0.000	1947.8	0.068	BB	0.030	1091.35	0.062
84	7.181		0.0000	0.000	29448.0	1.028	BB	0.042	11771.14	0.667
85	7.414		0.0000	0.000	44546.2	1.555	BB	0.031	23733.90	1.344
86	7.488		0.0000	0.000	2905.0	0.101	BB	0.040	1210.80	0.069
87	7.584		0.0000	0.000	956.5	0.033	BB	0.087	182.25	0.010
88	7.730		0.0000	0.000	12629.0	0.441	BB	0.034	6121.22	0.347
89	7.820		0.0000	0.000	857.7	0.030	BB	0.027	534.77	0.030
90	7.977	DCB	-688.3293	58.205	5768.4	0.201	BB	0.038	2560.84	0.145
91	8.148		0.0000	0.000	301.5	0.011	BB	0.036	139.53	0.008
92	8.241		0.0000	0.000	576.1	0.020	BB	0.040	238.62	0.014
93	8.373		0.0000	0.000	24460.2	0.854	BB	0.055	7430.79	0.421
94	8.610		0.0000	0.000	990.9	0.035	BB	0.064	256.54	0.015
95	9.093		0.0000	0.000	2033.1	0.071	BB	0.294	115.17	0.007

Total Area = 2864991.0, Total Amount = -1182.585, Total Height = 1765607.0, Sample Units = ug/l

ASX22 18070

Sample Name: **5822064 RI AF DF20 ACCARS2 T 093060018A 00173**
Acquired from CP07--H5088B via port 4 on 11/6/09 01:08:36am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.46R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.568		0.0000	0.000	2494.2	0.067	BV	0.028	1510.20	0.084
2	1.620		0.0000	0.000	85347.1	2.283	VB	0.044	32539.03	1.814
3	1.817		0.0000	0.000	18189.7	0.487	BB	0.037	8160.49	0.455
4	2.052		0.0000	0.000	30237.7	0.809	BB	0.037	13473.86	0.751
5	2.186		0.0000	0.000	4975.2	0.133	BB	0.048	1713.42	0.096
6	2.249		0.0000	0.000	269.2	0.007	BB	0.020	220.24	0.012
7	2.283		0.0000	0.000	361.9	0.010	BB	0.020	307.10	0.017
8	2.342	TCX	315.6255	51.465	12643.9	0.338	BB	0.031	6844.08	0.382
9	2.433		0.0000	0.000	18731.0	0.501	BB	0.042	7412.50	0.413
10	2.493		0.0000	0.000	10774.6	0.288	BB	0.045	3958.82	0.221
11	2.598		0.0000	0.000	3314.6	0.089	BB	0.039	1402.69	0.078
12	2.743		0.0000	0.000	21503.3	0.575	BB	0.034	10616.82	0.592
13	2.861		0.0000	0.000	4344.2	0.116	BB	0.039	1837.92	0.102
14	3.009		0.0000	0.000	10096.3	0.270	BB	0.037	4539.25	0.253
15	3.071		0.0000	0.000	3253.6	0.087	BB	0.032	1718.12	0.096
16	3.168		0.0000	0.000	13264.7	0.355	BB	0.053	4200.27	0.234
17	3.319		0.0000	0.000	16392.0	0.439	BB	0.048	5706.85	0.318
18	3.420		0.0000	0.000	23035.1	0.616	BB	0.047	8088.85	0.451
19	3.464		0.0000	0.000	6821.8	0.183	BB	0.027	4264.82	0.238
20	3.559		0.0000	0.000	5224.4	0.140	BB	0.028	3110.84	0.173
21	3.617		0.0000	0.000	9203.7	0.246	BB	0.027	5579.63	0.311
22	3.663		0.0000	0.000	79175.1	2.118	BB	0.044	30083.83	1.677
23	3.795		0.0000	0.000	7631.6	0.204	BB	0.033	3828.76	0.213
24	3.817		0.0000	0.000	1992.0	0.053	BB	0.014	2321.44	0.129
25	3.861		0.0000	0.000	31476.9	0.842	BB	0.041	12641.54	0.705
26	3.922		0.0000	0.000	5972.7	0.160	BB	0.023	4244.12	0.237
27	3.973		0.0000	0.000	3814.7	0.102	BB	0.026	2480.68	0.138
28	4.042		0.0000	0.000	4549.9	0.122	BB	0.042	1788.09	0.100
29	4.111		0.0000	0.000	58359.5	1.561	BB	0.026	38020.86	2.120
30	4.139		0.0000	0.000	31758.6	0.850	BB	0.025	20765.38	1.158
31	4.197		0.0000	0.000	7625.9	0.204	BB	0.019	6741.49	0.376
32	4.248		0.0000	0.000	65326.0	1.748	BB	0.020	54049.43	3.013
33	4.285		0.0000	0.000	108275.8	2.897	BB	0.028	63868.52	3.561
34	4.382		0.0000	0.000	4922.9	0.132	BB	0.019	4303.45	0.240
35	4.430		0.0000	0.000	53815.2	1.440	BB	0.031	28556.82	1.592
36	4.502		0.0000	0.000	52305.1	1.399	BB	0.023	37643.18	2.099
37	4.535		0.0000	0.000	17120.9	0.458	BB	0.016	17524.18	0.977
38	4.590		0.0000	0.000	204460.6	5.470	BB	0.035	98256.91	5.478
39	4.630		0.0000	0.000	12125.4	0.324	BB	0.018	11374.84	0.634
40	4.690		0.0000	0.000	165425.6	4.426	BB	0.029	95992.39	5.352
41	4.743		0.0000	0.000	3429.3	0.092	BB	0.014	4025.87	0.224
42	4.788		0.0000	0.000	322773.1	8.636	BB	0.037	144433.30	8.053
43	4.870		0.0000	0.000	246713.8	6.601	BB	0.033	123981.40	6.912
44	4.983		0.0000	0.000	163398.2	4.372	BB	0.042	64210.82	3.580
45	5.043		0.0000	0.000	153543.4	4.108	BB	0.027	94850.91	5.288
46	5.091		0.0000	0.000	6401.2	0.171	BB	0.016	6774.01	0.378
47	5.140		0.0000	0.000	423282.5	11.325	BB	0.043	164338.30	9.163
48	5.216		0.0000	0.000	16669.4	0.446	BB	0.020	13697.90	0.764
49	5.334		0.0000	0.000	7183.0	0.192	BB	0.022	5502.56	0.307
50	5.406		0.0000	0.000	180704.6	4.835	BB	0.050	60672.23	3.383
51	5.471		0.0000	0.000	22853.8	0.611	BB	0.022	17420.98	0.971
52	5.509		0.0000	0.000	9405.2	0.252	BB	0.020	7861.61	0.438
53	5.559		0.0000	0.000	4352.8	0.116	BB	0.018	4033.12	0.225
54	5.624		0.0000	0.000	365977.1	9.791	BB	0.042	146317.00	8.158

ASX22 0071

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.798		0.0000	0.000	143294.0	3.834	BB	0.031	77671.61	4.331
56	5.866		0.0000	0.000	26845.7	0.718	BB	0.023	19132.92	1.067
57	5.920		0.0000	0.000	195872.3	5.240	BB	0.036	89846.70	5.009
58	6.086		0.0000	0.000	2324.0	0.062	BB	0.032	1213.05	0.068
59	6.172		0.0000	0.000	8117.0	0.217	BB	0.031	4424.62	0.247
60	6.237		0.0000	0.000	34466.0	0.922	BB	0.037	15449.06	0.861
61	6.381		0.0000	0.000	87540.3	2.342	BB	0.039	37642.26	2.099
62	6.523		0.0000	0.000	415.2	0.011	BB	0.026	262.59	0.015
63	6.620		0.0000	0.000	23418.1	0.627	BB	0.038	10218.91	0.570
64	6.804	DCB	297.6609	48.535	15912.7	0.426	BB	0.052	5137.95	0.286
65	7.006		0.0000	0.000	1511.4	0.040	BB	0.060	422.93	0.024
66	7.348		0.0000	0.000	18250.7	0.488	BB	0.053	5791.92	0.323
67	7.452		0.0000	0.000	9810.1	0.262	BB	0.036	4518.53	0.252
68	7.713		0.0000	0.000	640.0	0.017	BB	0.049	218.59	0.012
69	8.193		0.0000	0.000	14965.3	0.400	BB	0.391	638.67	0.036
70	8.530		0.0000	0.000	7021.0	0.188	BB	0.128	917.11	0.051
71	8.670		0.0000	0.000	4032.5	0.108	BB	0.252	266.98	0.015

Total Area = 3737736.0, Total Amount = 613.286, Total Height = 1793586.0, Sample Units = ug/l

ASX22 0872

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\3P7307.46R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 12:59:27 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.46R



Instrument ID: CP07--H5088B Injected On: 11/6/2009 12:59:27 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 893
Analyst: 2280

Dilution Factor: 200

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.001	9371	-.55348	TCX	2.342	6844	.35344	TCX
7.977	2561	-.77081	DCB	6.804	5138	.33333	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.46A
Area File: C:\CPWIN\DATA1\3P7307B.46A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 1:08:44 AM
File Reported On: 11/6/2009 at 1:08:57 AM

10X REPLOTS

ASX22 8874

C:\CPWIN\DATA1\3P7307.46R



C:\CPWIN\DATA1\3P7307B.46R



ASX22 0075

Multiple Component Data Summary

Sample Name: 5822067 RI AF DF20 DPW29 Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 1018 ml Total Volume: 200 ml Analyst: 2131 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on Nov 06, 2009 01:38:29
Instrument H5088A
Result file 3P7307.49R
Calibration file 1P7307
Method file PESTD

%SSR(TCX) * 0.0% Conc: 0
%SSR(DCB) * 0.0% Conc: 0

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**20.89**	
4.83	4.86	4.89	23992.33984	39.263275			1
5.03	5.06	5.09	55120.23046	59.667411			2
5.27	5.30	5.33	32308.68164	48.046652			3
Height summation:			**111421.251954**				
Concentration		CF:	**48.992446**	L:			
Aroclor-1260						**14.78**	
6.25	6.27	6.31	38500.63671	50.337977			4
6.55	6.58	6.61	62122.47656	38.0672			5
6.80	6.83	6.86	30841.88476	41.144165			6
Height summation:			**131464.998048**				
Concentration		CF:	**43.183114**	L:			

Analysis Report (B)

Injected on Nov 06, 2009 01:38:29
Instrument H5088B
Result file 3P7307B.49R
Calibration file 1P7307B
Method file PESTDB

%SSR(TCX) 94.3% Conc: 0.284544
%SSR(DCB) 79.3% Conc: 0.237523

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**25.24**	
4.08	4.11	4.14	62853.66406	66.160501			1
4.55	4.59	4.61	74271.66406	70.144496			2
4.66	4.69	4.72	61719.75390	102.824903			3
Height summation:			**198845.082032**				
Concentration		CF:	**79.709967**	L:			
Aroclor-1260						**7.83**	
5.36	5.41	5.42	41594.83593	50.565332			4
5.60	5.62	5.66	95873.25781	58.648353			5
5.89	5.92	5.95	57875.21875	52.500293			6
Height summation:			**195343.312501**				
Concentration		CF:	**53.904659**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<9.8232	<1.9646			4	15	
Aroclor-1221			<9.8232	<1.9646			3	5	
Aroclor-1232			<9.8232	<1.9646			4	10	
Aroclor-1242			<9.8232	<1.9646			4	30	
Aroclor-1248			<9.8232	<1.9646			4	30	
Aroclor-1254	A	48.992446	9.8232	1.9646		47.73	4	40	1269
Aroclor-1260	B	53.904659	9.8232	1.9646		22.09	4	40	

Units: ug/l

%Difference = High - Low divided by the Average times 100

Reviewed By: [signature] Date: 11-6-09

Verified By: [signature] Date: 11/9/09

ASX22 0076

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822067** RI AF DF20 DPW29 Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 1018 ml Total Volume: 200 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 01:38:29
Instrument : CP07–H5088A
Result file : 3P7307.49R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

%SSR(TCX) : *-167% Conc.: -0.503589
%SSR(DCB) : *-237% Conc.: -0.710005

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					5	70.15	
3.83	3.86	3.89	2436.875244	3.775279			2
4.10	4.13	4.16	8757.089844	6.328157			3
4.15	4.19	4.21	3710.564697	6.031459			4
4.24	4.27	4.30	2123.583496	3.644958			5
4.62	4.65	4.68	6534.521973	15.737616			6
Height Summation:			23562.635254				
Amount Avg CF:			7.103494	Linear:			
Aroclor-1221					2	125.07	
3.20	3.22	3.26	148.93956	0.471005			1
3.28	3.28	3.34	1712.028076	7.674873			2
Height Summation:			1860.967636				
Amount Avg CF:			4.072939	Linear:			
Aroclor-1232					5	75.03	
3.83	3.86	3.89	2436.875244	8.484846			2
4.10	4.13	4.16	8757.089844	13.221009			3
4.15	4.19	4.21	3710.564697	12.602491			4
4.23	4.27	4.29	2123.583496	7.772112			5
4.62	4.65	4.68	6534.521973	36.239832			6
Height Summation:			23562.635254				
Amount Avg CF:			15.664058	Linear:			
Aroclor-1242					5	68.05	
3.83	3.86	3.89	2436.875244	4.620319			2
4.09	4.13	4.15	8757.089844	7.316238			3
4.15	4.19	4.21	3710.564697	7.051909			4
4.23	4.27	4.29	2123.583496	4.249625			5
4.62	4.65	4.68	6534.521973	17.958957			6
Height Summation:			23562.635254				
Amount Avg CF:			8.23941	Linear:			
Aroclor-1248					6	152.52	
4.10	4.13	4.16	8757.089844	11.474715			1
4.33	4.36	4.39	10276.41113:	19.348029			2
4.50	4.52	4.56	13063.04687!	16.420761			3
4.62	4.65	4.68	6534.521973	12.434944			4
5.02	5.06	5.08	55120.23046!	86.608863			5
E 5.61	5.64	5.67	89107.32812!	294.447083			6
Height Summation:			182858.628419				
Amount Avg CF:			73.455733	Linear:			
Aroclor-1254					6	50.54	
4.83	4.86	4.89	23992.33984	39.263275			1
+ 4.83	4.88	4.89	15716.83496	25.720477			1
5.03	5.06	5.09	55120.23046!	59.667411			2
5.27	5.30	5.33	32308.68164	48.046652			3
5.34	5.37	5.40	70810	58.71495			4
5.85	5.87	5.91	83642.30468!	134.374185			5
5.98	6.01	6.04	89122.20312!	108.09982			6
Height Summation:			354995.759767				
Amount Avg CF:			74.694382	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 01:38:29
Instrument : CP07–H5088B
Result file : 3P7307B.49R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET A linear

%SSR(TCX) : 94.3% Conc.: 0.284544
%SSR(DCB) : 79.3% Conc.: 0.237523

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	85.72	
2.69	2.74	2.75	8436.113281	21.489153			1
2.98	3.01	3.04	2511.842285	2.882567			2
3.14	3.17	3.20	1542.849609	2.570417			3
3.37	3.41	3.43	4774.445313	5.34298			4
3.42	3.46	3.48	1167.285645	6.700869			5
3.77	3.79	3.83	8815.224609	17.705978			6
Height Summation:			27247.760742				
Amount Avg CF:			9.448661	Linear:			
Aroclor-1221					1		
2.69	2.74	2.75	8436.113281	15.517903			3
Height Summation:			8436.113281				
Amount Avg CF:			15.517903	Linear:			
Aroclor-1232					6	82.11	
2.69	2.74	2.75	8436.113281	18.777428			1
2.98	3.01	3.04	2511.842285	6.358123			2
3.14	3.17	3.20	1542.849609	5.494179			3
3.38	3.41	3.44	4774.445313	12.969305			4
3.42	3.46	3.48	1167.285645	14.872192			5
E 3.77	3.79	3.83	8815.224609	43.55813			6
Height Summation:			27247.760742				
Amount Avg CF:			17.004893	Linear:			
Aroclor-1242					6	85.73	
2.68	2.74	2.74	8436.113281	25.886601			1
2.98	3.01	3.04	2511.842285	3.538039			2
3.14	3.17	3.20	1542.849609	3.166011			3
3.37	3.41	3.43	4774.445313	6.913107			4
3.42	3.46	3.48	1167.285645	7.356189			5
3.77	3.79	3.83	8815.224609	21.350814			6
Height Summation:			27247.760742				
Amount Avg CF:			11.36846	Linear:			
Aroclor-1248					6	157.37	
3.14	3.17	3.20	1542.849609	7.191518			1
3.37	3.41	3.43	4774.445313	7.375865			2
3.77	3.79	3.83	8815.224609	12.945221			3
3.89	3.92	3.95	2782.03833	8.727962			4
4.26	4.29	4.32	39018.59375	74.643765			5
E 4.55	4.59	4.61	74271.66406:	234.509267			6
Height Summation:			131204.815674				
Amount Avg CF:			57.5656	Linear:			
Aroclor-1254					6	39.24	
4.08	4.11	4.14	62853.66406:	86.160501			1
4.55	4.59	4.61	74271.66406:	70.144496			2
4.66	4.69	4.72	61719.75390!	102.824903			3
4.76	4.79	4.82	100756.6406:	148.609237			4
E 4.85	4.87	4.91	81801.66406:	182.657577			5
5.11	5.14	5.17	115485.7656:	125.358058			6
Height Summation:			496889.152345				
Amount Avg CF:			115.959129	Linear:			

ASX22. 0077

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:20:43

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822067** RI AF DF20 DPW29 Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 1018 ml Total Volume: 200 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 01:38:29
Instrument : CP07--H5088A
Result file : 3P7307.49R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	48.53	
5.74	5.77	5.80	69363.27343	66.748639			1
5.98	6.01	6.04	89122.20312	100.834547			2
6.05	6.08	6.11	19636.875	28.080357			3
6.25	6.27	6.31	38500.63671	50.337977			4
6.55	6.58	6.61	62122.47656	38.0672			5
6.80	6.83	6.86	30841.88476	41.144165			6
Height Summation:			**309587.349611**				
Amount Avg CF:			**54.202148**	Linear:			
Aroclor-1262					6	53.57	
5.49	5.52	5.55	35924.08984	42.419936			1
5.73	5.77	5.79	69363.27343	78.62599			2
6.05	6.08	6.11	19636.875	16.820218			3
6.25	6.27	6.31	38500.63671	42.387753			4
6.55	6.58	6.61	62122.47656	29.787108			5
6.79	6.83	6.85	30841.88476	28.727372			6
Height Summation:			**256389.236330**				
Amount Avg CF:			**39.79473**	Linear:			
Aroclor-1268					5	106.62	
6.05	6.08	6.11	19636.875	42.509932			1
6.26	6.27	6.32	38500.63671	65.320774			2
6.79	6.83	6.85	30841.88476	14.494201			3
6.87	6.88	6.93	10932.98046	5.948441			4
7.03	7.06	7.09	656.097473	0.368008			5
Height Summation:			**100568.474427**				
Amount Avg CF:			**25.728271**	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 01:38:29
Instrument : CP07--H5088B
Result file : 3P7307B.49R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	18.37	
4.66	4.69	4.72	61719.75390	64.592758			1
4.84	4.87	4.90	81801.66406	80.93998			2
4.96	4.98	5.02	42038.76953	56.197821			3
5.36	5.41	5.42	41594.83593	58.585332			4
5.60	5.62	5.66	95873.25781	58.648353			5
5.89	5.92	5.95	57875.21875	52.500293			6
Height Summation:			**380903.500001**				
Amount Avg CF:			**60.57409**	Linear:			
Aroclor-1262					6	58.09	
4.84	4.87	4.90	81801.66406	88.660885			1
5.36	5.41	5.42	41594.83593	36.18858			2
5.60	5.62	5.66	95873.25781	45.321794			3
5.84	5.87	5.90	12496.50390	16.359038			4
5.89	5.92	5.95	57875.21875	43.190215			5
6.35	6.38	6.41	23745.37109	27.248316			6
Height Summation:			**313386.851564**				
Amount Avg CF:			**42.828138**	Linear:			
Aroclor-1268					6	105.88	
5.35	5.41	5.41	41594.83593	60.146478			1
5.84	5.87	5.90	12496.50390	5.640565			2
5.90	5.92	5.96	57875.21875	29.786656			3
6.06	6.08	6.12	785.62085	0.370432			4
6.14	6.17	6.20	2597.943359	5.228966			5
6.36	6.38	6.42	23745.37109	26.899268			6
Height Summation:			**139095.493897**				
Amount Avg CF:			**21.345394**	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			9.8232	1.9646		28.34	4	15	
Aroclor-1221			9.8232	1.9646		** 116.84	3	5	
Aroclor-1232			9.8232	1.9646		8.21	4	10	
Aroclor-1242			9.8232	1.9646		31.92	4	30	
Aroclor-1248			9.8232	1.9646	E	24.26	4	30	
Aroclor-1254			9.8232	1.9646		** 43.29	4	40	
Aroclor-1260			9.8232	1.9646		11.10	4	40	
Aroclor-1262			0	0		7.34	4	40	
Aroclor-1268			0	0		18.62	4	40	

Units: ug/L

ASX22 0078

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:20:43

Sample Name: **5822067 RI AF DF20 ACDPW29 T 093060018A 00173**
Acquired from CP07--H5088A via port 3 on 11/6/09 01:47:37am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.49R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.511		0.0000	0.000	9962.9	0.515	BB	0.022	7581.93	0.640
2	1.659		0.0000	0.000	54352.9	2.812	BV	0.032	27978.36	2.361
3	1.763		0.0000	0.000	8164.4	0.422	VV	0.066	2056.15	0.173
4	1.879		0.0000	0.000	16821.6	0.870	VV	0.061	4585.07	0.387
5	1.969		0.0000	0.000	6677.7	0.345	VV	0.063	1762.82	0.149
6	2.062		0.0000	0.000	73673.1	3.811	VV	0.033	37008.30	3.122
7	2.178		0.0000	0.000	6772.2	0.350	VV	0.053	2124.81	0.179
8	2.278		0.0000	0.000	13494.0	0.698	VB	0.028	7895.81	0.666
9	2.481		0.0000	0.000	685.1	0.035	BB	0.039	296.00	0.025
10	2.528		0.0000	0.000	623.1	0.032	BV	0.025	419.27	0.035
11	2.574		0.0000	0.000	1311.9	0.068	VB	0.028	794.41	0.067
12	2.627		0.0000	0.000	10011.5	0.518	BB	0.021	7814.86	0.659
13	2.699		0.0000	0.000	93.9	0.005	BB	0.018	84.67	0.007
14	2.745		0.0000	0.000	10486.4	0.542	BB	0.027	6471.61	0.546
15	2.854		0.0000	0.000	1602.0	0.083	BB	0.036	748.95	0.063
16	3.001	TCX	-512.6532	41.496	13469.3	0.697	BB	0.025	8995.81	0.759
17	3.053		0.0000	0.000	5631.0	0.291	BB	0.031	3029.49	0.256
18	3.183		0.0000	0.000	630.3	0.033	BB	0.023	453.58	0.038
19	3.223		0.0000	0.000	141.3	0.007	BB	0.016	148.94	0.013
20	3.268		0.0000	0.000	2146.4	0.111	BB	0.023	1589.58	0.134
21	3.282		0.0000	0.000	1259.7	0.065	BB	0.012	1712.03	0.144
22	3.344		0.0000	0.000	8230.6	0.426	BB	0.027	5050.44	0.426
23	3.457		0.0000	0.000	2784.9	0.144	BB	0.027	1696.78	0.143
24	3.513		0.0000	0.000	1380.6	0.071	BB	0.023	1007.74	0.085
25	3.598		0.0000	0.000	372.6	0.019	BB	0.027	227.13	0.019
26	3.640		0.0000	0.000	124.4	0.006	BB	0.018	114.87	0.010
27	3.695		0.0000	0.000	822.6	0.043	BB	0.019	724.14	0.061
28	3.781		0.0000	0.000	2720.7	0.141	BB	0.030	1530.78	0.129
29	3.864		0.0000	0.000	4022.9	0.208	BB	0.028	2436.88	0.206
30	3.924		0.0000	0.000	268.0	0.014	BB	0.029	152.48	0.013
31	3.974		0.0000	0.000	163.5	0.008	BB	0.027	101.28	0.009
32	4.045		0.0000	0.000	7225.2	0.374	BB	0.034	3586.07	0.303
33	4.129		0.0000	0.000	18109.6	0.937	BB	0.034	8757.09	0.739
34	4.186		0.0000	0.000	5341.9	0.276	BB	0.024	3710.56	0.313
35	4.265		0.0000	0.000	3642.4	0.188	BB	0.029	2123.58	0.179
36	4.316		0.0000	0.000	684.5	0.035	BB	0.019	599.54	0.051
37	4.359		0.0000	0.000	12400.8	0.642	BB	0.020	10276.41	0.867
38	4.393		0.0000	0.000	5399.9	0.279	BB	0.016	5596.47	0.472
39	4.424		0.0000	0.000	8761.5	0.453	BB	0.022	6707.28	0.566
40	4.525		0.0000	0.000	25211.6	1.304	BB	0.032	13063.05	1.102
41	4.599		0.0000	0.000	5577.7	0.289	BB	0.021	4456.56	0.376
42	4.647		0.0000	0.000	9486.6	0.491	BB	0.024	6534.52	0.551
43	4.698		0.0000	0.000	2242.3	0.116	BB	0.020	1836.03	0.155
44	4.813		0.0000	0.000	5974.4	0.309	BB	0.038	2629.41	0.222
45	4.856		0.0000	0.000	28075.8	1.452	BB	0.020	23992.34	2.024
46	4.884		0.0000	0.000	13695.7	0.709	BB	0.015	15716.83	1.326
47	4.912		0.0000	0.000	21077.7	1.090	BB	0.016	22583.38	1.905
48	5.002		0.0000	0.000	12153.5	0.629	BB	0.023	8763.05	0.739
49	5.055		0.0000	0.000	86183.7	4.459	BB	0.026	55120.23	4.651
50	5.097		0.0000	0.000	31437.2	1.626	BB	0.020	26300.87	2.219
51	5.148		0.0000	0.000	223.4	0.012	BB	0.009	395.03	0.033
52	5.176		0.0000	0.000	8850.7	0.458	BB	0.019	7593.07	0.641
53	5.231		0.0000	0.000	37597.1	1.945	BB	0.025	24584.13	2.074
54	5.299		0.0000	0.000	46102.9	2.385	BB	0.024	32308.68	2.726

ASX22 0879

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.332		0.0000	0.000	8364.3	0.433	BB	0.013	10412.41	0.879
56	5.373		0.0000	0.000	100671.9	5.208	BB	0.024	70810.00	5.974
57	5.428		0.0000	0.000	16513.2	0.854	BB	0.025	10997.97	0.928
58	5.474		0.0000	0.000	14546.7	0.753	BB	0.020	12062.47	1.018
59	5.520		0.0000	0.000	58814.1	3.043	BB	0.027	35924.09	3.031
60	5.587		0.0000	0.000	8962.5	0.464	BB	0.019	7677.74	0.648
61	5.637		0.0000	0.000	139183.9	7.200	BB	0.026	89107.33	7.518
62	5.701		0.0000	0.000	34474.6	1.783	BB	0.024	23830.95	2.011
63	5.767		0.0000	0.000	143429.6	7.420	BB	0.034	69363.27	5.852
64	5.875		0.0000	0.000	126338.6	6.536	BB	0.025	83642.30	7.057
65	5.919		0.0000	0.000	1379.4	0.071	BB	0.011	2096.86	0.177
66	5.963		0.0000	0.000	6543.9	0.339	BB	0.023	4841.17	0.408
67	6.011		0.0000	0.000	157727.7	8.160	BB	0.029	89122.20	7.519
68	6.082		0.0000	0.000	25144.3	1.301	BB	0.021	19636.88	1.657
69	6.134		0.0000	0.000	20682.7	1.070	BB	0.025	13791.54	1.164
70	6.274		0.0000	0.000	96830.1	5.009	BB	0.042	38500.64	3.248
71	6.352		0.0000	0.000	13702.7	0.709	BB	0.022	10374.60	0.875
72	6.397		0.0000	0.000	18017.4	0.932	BB	0.029	10479.99	0.884
73	6.511		0.0000	0.000	34783.7	1.799	BB	0.026	22124.23	1.867
74	6.578		0.0000	0.000	102063.4	5.280	BB	0.027	62122.48	5.241
75	6.833		0.0000	0.000	51732.2	2.676	BB	0.028	30841.88	2.602
76	6.884		0.0000	0.000	25313.7	1.310	BB	0.039	10932.98	0.922
77	6.982		0.0000	0.000	264.5	0.014	BB	0.027	162.24	0.014
78	7.059		0.0000	0.000	1192.2	0.062	BB	0.030	656.10	0.055
79	7.181		0.0000	0.000	20205.6	1.045	BB	0.045	7445.14	0.628
80	7.415		0.0000	0.000	27324.3	1.414	BB	0.030	14997.89	1.265
81	7.513		0.0000	0.000	2580.9	0.134	BB	0.048	893.50	0.075
82	7.582		0.0000	0.000	507.3	0.026	BB	0.068	124.40	0.010
83	7.730		0.0000	0.000	8084.5	0.418	BB	0.034	3914.83	0.330
84	7.820		0.0000	0.000	688.8	0.036	BB	0.027	422.94	0.036
85	7.977	DCB	-722.7852	58.504	5464.7	0.283	BB	0.041	2211.92	0.187
86	8.149		0.0000	0.000	157.7	0.008	BB	0.032	80.93	0.007
87	8.242		0.0000	0.000	212.7	0.011	BB	0.037	95.44	0.008
88	8.310		0.0000	0.000	72.3	0.004	BB	0.026	46.10	0.004
89	8.374		0.0000	0.000	3621.0	0.187	BB	0.040	1519.99	0.128
90	9.105		0.0000	0.000	3048.5	0.158	BB	0.364	139.57	0.012

Total Area = 1933003.0, Total Amount = -1235.438, Total Height = 1185234.0, Sample Units = ug/l

ASX22 0000

Sample Name: **5822067 RI AF DF20 ACDPW29 T 093060018A 00173**
Acquired from CP07--H5088B via port 4 on 11/6/09 01:47:37am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.49R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.572		0.0000	0.000	1087.8	0.041	BV	0.026	685.69	0.057
2	1.620		0.0000	0.000	84631.2	3.202	VV	0.043	32753.04	2.720
3	1.816		0.0000	0.000	13248.0	0.501	VB	0.035	6355.83	0.528
4	1.957		0.0000	0.000	154.2	0.006	BB	0.032	81.48	0.007
5	2.053		0.0000	0.000	28067.8	1.062	BB	0.037	12717.62	1.056
6	2.167		0.0000	0.000	425.6	0.016	BB	0.043	166.35	0.014
7	2.187		0.0000	0.000	1024.6	0.039	BB	0.024	707.42	0.059
8	2.250		0.0000	0.000	343.7	0.013	BB	0.020	281.22	0.023
9	2.283		0.0000	0.000	437.0	0.017	BB	0.021	350.01	0.029
10	2.341	TCX	289.6657	54.503	11218.0	0.424	BB	0.030	6281.17	0.522
11	2.432		0.0000	0.000	19300.2	0.730	BB	0.045	7204.54	0.598
12	2.487		0.0000	0.000	188.9	0.007	BB	0.013	234.91	0.020
13	2.520		0.0000	0.000	442.2	0.017	BB	0.023	322.19	0.027
14	2.597		0.0000	0.000	3133.6	0.119	BB	0.040	1290.71	0.107
15	2.743		0.0000	0.000	16883.4	0.639	BB	0.033	8436.11	0.700
16	2.862		0.0000	0.000	2562.6	0.097	BB	0.039	1094.35	0.091
17	3.009		0.0000	0.000	5518.6	0.209	BB	0.037	2511.84	0.209
18	3.073		0.0000	0.000	1307.1	0.049	BB	0.032	688.69	0.057
19	3.167		0.0000	0.000	3912.7	0.148	BB	0.042	1542.85	0.128
20	3.206		0.0000	0.000	625.3	0.024	BB	0.023	448.02	0.037
21	3.268		0.0000	0.000	782.7	0.030	BB	0.019	700.25	0.058
22	3.318		0.0000	0.000	5162.7	0.195	BB	0.041	2115.04	0.176
23	3.406		0.0000	0.000	12658.6	0.479	BB	0.044	4774.45	0.396
24	3.462		0.0000	0.000	1543.8	0.058	BB	0.022	1167.29	0.097
25	3.561		0.0000	0.000	3762.3	0.142	BB	0.025	2485.55	0.206
26	3.616		0.0000	0.000	6849.6	0.259	BB	0.026	4454.89	0.370
27	3.644		0.0000	0.000	896.4	0.034	BB	0.017	901.76	0.075
28	3.661		0.0000	0.000	5068.3	0.192	BB	0.027	3174.98	0.264
29	3.795		0.0000	0.000	19570.7	0.740	BB	0.037	8815.22	0.732
30	3.856		0.0000	0.000	9142.7	0.346	BB	0.044	3446.46	0.286
31	3.924		0.0000	0.000	4283.1	0.162	BB	0.026	2782.04	0.231
32	3.972		0.0000	0.000	2144.0	0.081	BB	0.027	1321.68	0.110
33	4.046		0.0000	0.000	1111.3	0.042	BB	0.079	234.48	0.019
34	4.111		0.0000	0.000	221654.6	8.385	BB	0.059	62853.66	5.219
35	4.198		0.0000	0.000	4310.3	0.163	BB	0.019	3702.66	0.307
36	4.249		0.0000	0.000	47175.6	1.785	BB	0.023	34227.25	2.842
37	4.286		0.0000	0.000	67775.8	2.564	BB	0.029	39018.59	3.240
38	4.382		0.0000	0.000	2929.5	0.111	BB	0.020	2428.84	0.202
39	4.430		0.0000	0.000	40809.7	1.544	BB	0.031	21728.72	1.804
40	4.502		0.0000	0.000	38028.6	1.439	BB	0.023	27252.24	2.263
41	4.536		0.0000	0.000	11661.3	0.441	BB	0.016	11779.89	0.978
42	4.592		0.0000	0.000	155062.0	5.866	BB	0.035	74271.66	6.167
43	4.634		0.0000	0.000	23138.2	0.875	BB	0.020	19218.77	1.596
44	4.691		0.0000	0.000	106032.7	4.011	BB	0.029	61719.75	5.125
45	4.743		0.0000	0.000	1651.3	0.062	BB	0.012	2202.49	0.183
46	4.789		0.0000	0.000	222881.7	8.432	BB	0.037	100756.60	8.366
47	4.871		0.0000	0.000	167774.2	6.347	BB	0.034	81801.66	6.792
48	4.984		0.0000	0.000	109062.5	4.126	BB	0.043	42038.77	3.491
49	5.044		0.0000	0.000	116780.3	4.418	BB	0.027	71002.05	5.895
50	5.090		0.0000	0.000	4116.9	0.156	BB	0.015	4500.55	0.374
51	5.140		0.0000	0.000	295904.3	11.194	BB	0.043	115485.80	9.589
52	5.216		0.0000	0.000	8750.7	0.331	BB	0.020	7321.40	0.608
53	5.334		0.0000	0.000	4418.4	0.167	BB	0.022	3332.80	0.277
54	5.407		0.0000	0.000	123192.3	4.660	BB	0.049	41594.84	3.454

4SX22 5091

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.471		0.0000	0.000	14278.5	0.540	BB	0.022	10910.16	0.906
56	5.509		0.0000	0.000	6147.0	0.233	BB	0.020	5047.58	0.419
57	5.559		0.0000	0.000	2662.7	0.101	BB	0.018	2504.55	0.208
58	5.624		0.0000	0.000	244462.9	9.248	BB	0.042	95873.26	7.961
59	5.797		0.0000	0.000	32311.3	1.222	BB	0.030	18206.50	1.512
60	5.866		0.0000	0.000	18538.3	0.701	BB	0.025	12496.50	1.038
61	5.920		0.0000	0.000	129407.7	4.896	BB	0.037	57875.22	4.806
62	6.085		0.0000	0.000	1566.5	0.059	BB	0.033	785.62	0.065
63	6.172		0.0000	0.000	4839.9	0.183	BB	0.031	2597.94	0.216
64	6.237		0.0000	0.000	22390.5	0.847	BB	0.039	9554.09	0.793
65	6.382		0.0000	0.000	57465.3	2.174	BB	0.040	23745.37	1.972
66	6.620		0.0000	0.000	14707.9	0.556	BB	0.038	6373.34	0.529
67	6.803	DCB	241.7984	45.497	11135.7	0.421	BB	0.044	4173.70	0.347
68	7.006		0.0000	0.000	692.7	0.026	BB	0.063	183.94	0.015
69	7.349		0.0000	0.000	3928.9	0.149	BB	0.053	1229.58	0.102
70	7.453		0.0000	0.000	6919.7	0.262	BB	0.038	3053.63	0.254
71	7.580		0.0000	0.000	264.2	0.010	BB	0.059	74.60	0.006
72	7.720		0.0000	0.000	23743.6	0.898	BB	2.096	188.77	0.016
73	8.517		0.0000	0.000	3764.1	0.142	BB	0.119	528.88	0.044
74	8.643		0.0000	0.000	3503.1	0.133	BB	0.320	182.39	0.015

Total Area = 2643330.0, Total Amount = 531.464, Total Height = 1204353.0, Sample Units = ug/l

ASX22: 0002:

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\3P7307.49R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 1:38:28 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.49R



Instrument ID: CP07--H5088B Injected On: 11/6/2009 1:38:28 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1018
Analyst: 2280

Dilution Factor: 200

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.001	8996	-.50359	TCX	2.341	6281	.28454	TCX
7.977	2212	-.71001	DCB	6.803	4174	.23752	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.49A
Area File: C:\CPWIN\DATA1\3P7307B.49A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 1:47:44 AM
File Reported On: 11/6/2009 at 1:47:57 AM

10X REPLOTS

ASX22 0084

C:\CPWIN\DATA1\3P7307.49R



C:\CPWIN\DATA1\3P7307B.49R



ASX22: 0005

Multiple Component Data Summary

Sample Name: 5822068 AF STEA1 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1041 ml Total Volume: 10 ml Analyst: 2131 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on Nov 06, 2009 01:51:31
Instrument H5088A
Result file 3P7307.50R
Calibration file 1P7307
Method file PESTD

%SSR(TCX) * 34.0% Conc: 0.100431
%SSR(DCB) * 0.0% Conc: 0

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**26.07**	
4.83	4.86	4.89	16932.36132	1.354873			1
5.03	5.06	5.09	43451.73828	2.299855			2
5.27	5.30	5.33	29149.58593	2.119548			3
Height summation:			**89533.685547**				
Concentration		CF:	**1.924759**	L:			
Aroclor-1260						**10.91**	
6.25	6.28	6.31	46111.3125	2.94783			4
6.55	6.58	6.61	79749.65625	2.389451			5
6.80	6.83	6.86	39150.72265	2.553725			6
Height summation:			**165011.691406**				
Concentration		CF:	**2.630335**	L:			

Analysis Report (B)

Injected on Nov 06, 2009 01:51:31
Instrument H5088B
Result file 3P7307B.50R
Calibration file 1P7307B
Method file PESTDB

%SSR(TCX) * 41.5% Conc: 0.12228
%SSR(DCB) 43.2% Conc: 0.126565

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**68.31**	
4.08	4.11	4.14	27800.72070	1.430841			1
4.66	4.69	4.72	69386.44531	5.652178			3
4.76	4.79	4.82	117376.9765	8.464905			4
Height summation:			**214564.142579**				
Concentration		CF:	**5.182641**	L:			
Aroclor-1260						**16.41**	
5.36	5.41	5.42	50035.85156	2.974143			4
5.60	5.63	5.66	116536.9765	3.485693			5
5.89	5.92	5.95	56482.58203	2.505248			6
Height summation:			**223055.410157**				
Concentration		CF:	**2.988361**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<0.4803	<0.0961			4	15	
Aroclor-1221			<0.4803	<0.0961			3	5	
Aroclor-1232			<0.4803	<0.0961			4	10	
Aroclor-1242			<0.4803	<0.0961			4	30	
Aroclor-1248			<0.4803	<0.0961			4	30	
Aroclor-1254	A	1.924759	0.4803	0.0961		91.68	4	40	1269
Aroclor-1260	B	2.988361	0.4803	0.0961		12.74	4	40	

Units: ug/l

%Difference = High - Low divided by the Average times 100

Reviewed By: [signature] Date: 11-6-09
Verified By: [signature] Date: 11/10/09

ASX22 8886

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822068** AF STEA1 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1041 ml Total Volume: 10 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 01:51:31
Instrument : CP07--H5088A
Result file : 3P7307.50R
Calibration file : 1P7307.CAL
Method file : PESTD.MET
%SSR(TCX) : *34% Conc.: 0.100431
%SSR(DCB) : *-1.3% Conc.: -0.003761

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					4	95.12	
3.35	3.38	3.41	5613.971191	0.515639			1
4.10	4.10	4.16	82365.34375	2.910241			3
4.15	4.19	4.21	1338.724121	0.1064			4
4.24	4.28	4.30	22219.98632	1.864807			5
Height Summation:			111538.025390				
Amount Avg CF:			1.349272	Linear:			
Aroclor-1221					3	87.73	
E 3.20	3.23	3.26	15284.93652	2.36345			1
3.28	3.29	3.34	3796.705566	0.832213			2
3.35	3.38	3.41	5613.971191	0.371013			3
Height Summation:			24695.613280				
Amount Avg CF:			1.188892	Linear:			
Aroclor-1232					4	106.22	
3.35	3.38	3.41	5613.971191	0.439319			1
E 4.10	4.10	4.16	82365.34375	6.080179			3
4.15	4.19	4.21	1338.724121	0.222318			4
E 4.23	4.28	4.29	22219.98632	3.976313			5
Height Summation:			111538.025390				
Amount Avg CF:			2.679532	Linear:			
Aroclor-1242					4	94.76	
3.35	3.38	3.41	5613.971191	0.60306			1
E 4.09	4.10	4.15	82365.34375	3.364647			3
4.15	4.19	4.21	1338.724121	0.124401			4
E 4.23	4.28	4.29	22219.98632	2.174163			5
Height Summation:			111538.025390				
Amount Avg CF:			1.566568	Linear:			
Aroclor-1248					5	125.33	
4.10	4.10	4.16	82365.34375	5.27708			1
4.33	4.36	4.39	4890.974121	0.450254			2
4.50	4.53	4.56	3879.619873	0.238454			3
5.02	5.06	5.08	43451.73828	3.338301			5
E 5.61	5.64	5.67	94594.10156	15.283575			6
Height Summation:			229181.777588				
Amount Avg CF:			4.917533	Linear:			
Aroclor-1254					6	65.40	
4.83	4.86	4.89	16932.36132	1.354878			1
5.03	5.06	5.09	43451.73828	2.299853			2
5.27	5.30	5.33	29149.58593	2.119548			3
5.34	5.37	5.40	69739.70312	2.827491			4
5.85	5.88	5.91	86906.30468	6.826659			5
5.98	6.01	6.04	110877.1718	6.575795			6
Height Summation:			357056.865235				
Amount Avg CF:			3.66737	Linear:			
Aroclor-1260					6	48.97	
5.74	5.77	5.80	85457.39062	4.020958			1
5.98	6.01	6.04	110877.1718	6.133843			2
6.05	6.08	6.11	23173.55859	1.62028			3
6.25	6.28	6.31	46111.3125	2.94783			4
6.55	6.58	6.61	79749.65625	2.389451			5
6.80	6.83	6.86	39150.72265	2.553725			6
Height Summation:			384519.812500				
Amount Avg CF:			3.277681	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 01:51:31
Instrument : CP07--H5088B
Result file : 3P7307B.50R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET
%SSR(TCX) : *41.5% Conc.: 0.12228
%SSR(DCB) : 43.2% Conc.: 0.126565

A linear matrix or hazards

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					4	84.98	
2.69	2.73	2.75	38441.59765	4.787903			1
2.98	3.01	3.04	1916.857666	0.107558			2
3.42	3.46	3.48	10883.61816	3.054881			5
3.77	3.80	3.83	15508.82812	1.523115			6
Height Summation:			66750.901611				
Amount Avg CF:			2.368364	Linear:			
Aroclor-1221					2	99.37	
2.53	2.59	2.59	2759.169189	0.603827			1
E 2.69	2.73	2.75	38441.59765	3.457476			3
Height Summation:			41200.766845				
Amount Avg CF:			2.030652	Linear:			
Aroclor-1232					4	71.98	
E 2.69	2.73	2.75	38441.59765	4.183716			1
2.98	3.01	3.04	1916.857666	0.237243			2
E 3.42	3.46	3.48	10883.61816	6.780133			5
E 3.77	3.80	3.83	15508.82812	3.746984			6
Height Summation:			66750.901611				
Amount Avg CF:			3.737019	Linear:			
Aroclor-1242					4	86.26	
E 2.68	2.73	2.74	38441.59765	5.767679			1
2.98	3.01	3.04	1916.857666	0.132016			2
E 3.42	3.46	3.48	10883.61816	3.353637			5
3.77	3.80	3.83	15508.82812	1.836653			6
Height Summation:			66750.901611				
Amount Avg CF:			2.772496	Linear:			
Aroclor-1248					4	186.59	
3.77	3.80	3.83	15508.82812	1.113582			3
3.89	3.94	3.95	9125.625	1.399843			4
4.26	4.29	4.32	37587.25390	3.515844			5
E 4.55	4.60	4.61	735386.1875	113.532284			6
Height Summation:			797607.894531				
Amount Avg CF:			29.890388	Linear:			
Aroclor-1254					6	103.83	
4.08	4.11	4.14	27800.72070	1.430841			1
E 4.55	4.60	4.61	735386.1875	33.958849			2
4.66	4.69	4.72	69386.44531	5.652178			3
4.76	4.79	4.82	117376.9765	8.464905			4
E 4.85	4.87	4.91	97705.47656	10.667472			5
5.11	5.14	5.17	127923.7890	6.789568			6
Height Summation:			1175579.595705				
Amount Avg CF:			11.160636	Linear:			
Aroclor-1260					6	22.17	
4.66	4.69	4.72	69386.44531	3.550597			1
4.84	4.87	4.90	97705.47656	4.727014			2
4.96	4.99	5.02	48101.49609	3.144091			3
5.36	5.41	5.42	50035.85156	2.974143			4
5.60	5.63	5.66	116536.9765	3.485693			5
5.89	5.92	5.95	56482.58203	2.505248			6
Height Summation:			438248.828127				
Amount Avg CF:			3.397798	Linear:			

ASX22: 8887

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:22:11

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822068** AF STEA1 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1041 ml Total Volume: 10 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 01:51:31
Instrument : CP07--H5088A
Result file : 3P7307.50R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1262					6	53.91	
5.49	5.52	5.55	41246.19140	2.381415			1
5.73	5.77	5.79	85457.39062	4.736453			2
6.05	6.08	6.11	23173.55859	0.970552			3
6.25	6.28	6.31	46111.3125	2.482259			4
6.55	6.58	6.61	79749.65625	1.869715			5
6.79	6.83	6.85	39150.72265	1.783043			6
Height Summation:			**314888.832031**				
Amount Avg CF:			**2.370573**	Linear:			
Aroclor-1268					5	104.44	
6.05	6.08	6.11	23173.55859	2.452889			1
6.26	6.28	6.32	46111.3125	3.825233			2
6.79	6.83	6.85	39150.72265	0.899622			3
6.87	6.89	6.93	14430.31835	0.383891			4
7.03	7.06	7.09	1238.179688	0.033958			5
Height Summation:			**124104.091797**				
Amount Avg CF:			**1.519119**	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 01:51:31
Instrument : CP07--H5088B
Result file : 3P7307B.50R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1262					6	47.95	
4.84	4.87	4.90	97705.47656	5.177927			1
5.36	5.41	5.42	50035.85156	2.128533			2
5.60	5.63	5.66	116536.9765	2.693645			3
5.84	5.86	5.90	38556.82031	2.467956			4
5.89	5.92	5.95	56482.58203	2.060983			5
6.35	6.38	6.41	27593.21679	1.548211			6
Height Summation:			**386910.923830**				
Amount Avg CF:			**2.679543**	Linear:			
Aroclor-1268					6	98.46	
5.35	5.41	5.41	50035.85156	3.537685			1
5.84	5.86	5.90	38556.82031	0.850947			2
5.90	5.92	5.96	56482.58203	1.421382			3
6.06	6.11	6.12	1281.70752	0.02955			4
6.14	6.18	6.20	2973.898682	0.292671			5
6.36	6.38	6.42	27593.21679	1.528378			6
Height Summation:			**176924.076906**				
Amount Avg CF:			**1.276769**	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			0.4803	0.0961		** 54.82	4	15	
Aroclor-1221			0.4803	0.0961	E	** 52.29	3	5	
Aroclor-1232			0.4803	0.0961	E	32.96	4	10	
Aroclor-1242			0.4803	0.0961	E	** 55.58	4	30	
Aroclor-1248			0.4803	0.0961	E	** 143.49	4	30	
Aroclor-1254			0.4803	0.0961		** 101.07	4	40	
Aroclor-1260			0.4803	0.0961		3.60	4	40	
Aroclor-1262			0	0		12.24	4	40	
Aroclor-1268			0	0		17.34	4	40	

Units: ug/L

ASX22 0000

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:22:11

Sample Name: **5822068AF ABSTEA1 T 093060018A 00173**
Acquired from CP07--H5088A via port 3 on 11/6/09 02:00:39am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.50R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.618		0.0000	0.000	14440.9	0.443	BV	0.031	7851.35	0.377
2	1.657		0.0000	0.000	18644.7	0.572	VB	0.030	10507.31	0.505
3	1.768		0.0000	0.000	22527.6	0.692	BV	0.045	8261.46	0.397
4	1.845		0.0000	0.000	8275.3	0.254	VV	0.039	3535.38	0.170
5	1.895		0.0000	0.000	11545.7	0.355	VV	0.041	4711.30	0.226
6	1.960		0.0000	0.000	4406.7	0.135	VV	0.041	1800.32	0.087
7	2.001		0.0000	0.000	2302.1	0.071	VV	0.035	1100.09	0.053
8	2.062		0.0000	0.000	15702.8	0.482	VV	0.029	8902.43	0.428
9	2.120		0.0000	0.000	1889.4	0.058	VV	0.028	1127.56	0.054
10	2.176		0.0000	0.000	7347.0	0.226	VV	0.040	3037.66	0.146
11	2.271		0.0000	0.000	20223.9	0.621	VV	0.038	8938.73	0.430
12	2.312		0.0000	0.000	10440.8	0.321	VB	0.036	4839.24	0.233
13	2.394		0.0000	0.000	1274.3	0.039	BB	0.027	786.29	0.038
14	2.578		0.0000	0.000	103836.0	3.188	BV	0.026	67435.55	3.241
15	2.628		0.0000	0.000	4858.4	0.149	VB	0.022	3696.01	0.178
16	2.701		0.0000	0.000	726.0	0.022	BB	0.023	530.21	0.025
17	2.761		0.0000	0.000	2206.8	0.068	BB	0.035	1039.51	0.050
18	2.854		0.0000	0.000	222919.9	6.845	BB	0.033	112963.60	5.429
19	2.965		0.0000	0.000	671.7	0.021	BB	0.015	754.50	0.036
20	3.002	TCX	104.5487	103.890	97795.6	3.003	BB	0.026	62150.17	2.987
21	3.095		0.0000	0.000	2135.8	0.066	BB	0.024	1488.70	0.072
22	3.146		0.0000	0.000	127.9	0.004	BB	0.015	140.68	0.007
23	3.181		0.0000	0.000	446.8	0.014	BB	0.021	360.45	0.017
24	3.227		0.0000	0.000	21215.9	0.651	BB	0.023	15284.94	0.735
25	3.289		0.0000	0.000	6000.4	0.184	BB	0.026	3796.71	0.182
26	3.349		0.0000	0.000	3368.9	0.103	BB	0.022	2608.32	0.125
27	3.382		0.0000	0.000	7092.8	0.218	BB	0.021	5613.97	0.270
28	3.448		0.0000	0.000	4254.3	0.131	BB	0.023	3140.67	0.151
29	3.474		0.0000	0.000	1189.0	0.037	BB	0.020	1003.10	0.048
30	3.519		0.0000	0.000	1882.9	0.058	BB	0.019	1635.11	0.079
31	3.552		0.0000	0.000	383.8	0.012	BB	0.015	440.43	0.021
32	3.599		0.0000	0.000	10363.8	0.318	BB	0.027	6481.83	0.312
33	3.787		0.0000	0.000	335323.0	10.296	BB	0.024	237385.50	11.410
34	3.960		0.0000	0.000	5757.2	0.177	BB	0.066	1443.61	0.069
35	4.043		0.0000	0.000	4853.3	0.149	BB	0.028	2891.86	0.139
36	4.098		0.0000	0.000	124992.5	3.838	BB	0.025	82365.34	3.959
37	4.188		0.0000	0.000	1625.8	0.050	BB	0.020	1338.72	0.064
38	4.218		0.0000	0.000	182.0	0.006	BB	0.012	244.31	0.012
39	4.282		0.0000	0.000	34882.4	1.071	BB	0.026	22219.99	1.068
40	4.363		0.0000	0.000	6867.2	0.211	BB	0.023	4890.97	0.235
41	4.395		0.0000	0.000	3355.1	0.103	BB	0.016	3419.80	0.164
42	4.426		0.0000	0.000	5010.7	0.154	BB	0.020	4168.47	0.200
43	4.477		0.0000	0.000	13682.9	0.420	BB	0.024	9481.11	0.456
44	4.530		0.0000	0.000	5425.5	0.167	BB	0.023	3879.62	0.186
45	4.709		0.0000	0.000	16430.7	0.504	BB	0.047	5817.58	0.280
46	4.749		0.0000	0.000	558.5	0.017	BB	0.015	635.76	0.031
47	4.817		0.0000	0.000	4062.8	0.125	BB	0.028	2376.68	0.114
48	4.858		0.0000	0.000	18851.5	0.579	BB	0.019	16932.36	0.814
49	4.887		0.0000	0.000	8675.6	0.266	BB	0.015	9533.34	0.458
50	4.914		0.0000	0.000	20217.5	0.621	BB	0.020	17266.43	0.830
51	5.002		0.0000	0.000	5579.1	0.171	BB	0.018	5157.55	0.248
52	5.056		0.0000	0.000	65511.6	2.011	BB	0.025	43451.74	2.088
53	5.098		0.0000	0.000	22697.6	0.697	BB	0.021	18357.88	0.882
54	5.179		0.0000	0.000	276736.0	8.497	BB	0.020	227010.00	10.911

ASX22 0009

822068AF ABSTEA1 T 093060018A 00173

K#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.231		0.0000	0.000	19601.6	0.602	BB	0.021	15743.04	0.757
56	5.299		0.0000	0.000	41032.7	1.260	BB	0.023	29149.59	1.401
57	5.333		0.0000	0.000	6342.3	0.195	BB	0.013	8189.07	0.394
58	5.375		0.0000	0.000	97145.0	2.983	BB	0.023	69739.70	3.352
59	5.430		0.0000	0.000	19126.4	0.587	BB	0.025	12670.35	0.609
60	5.475		0.0000	0.000	15727.2	0.483	BB	0.019	13577.41	0.653
61	5.521		0.0000	0.000	63760.5	1.958	BB	0.026	41246.19	1.982
62	5.586		0.0000	0.000	10231.7	0.314	BB	0.021	8283.66	0.398
63	5.639		0.0000	0.000	146878.4	4.510	BB	0.026	94594.10	4.546
64	5.702		0.0000	0.000	24008.0	0.737	BB	0.025	16196.70	0.778
65	5.768		0.0000	0.000	171283.6	5.259	BB	0.033	85457.39	4.107
66	5.876		0.0000	0.000	119602.6	3.672	BB	0.023	86906.30	4.177
67	5.919		0.0000	0.000	36341.8	1.116	BB	0.018	32801.71	1.577
68	5.964		0.0000	0.000	3110.9	0.096	BB	0.017	3091.67	0.149
69	6.013		0.0000	0.000	192580.6	5.913	BB	0.029	110877.20	5.329
70	6.084		0.0000	0.000	29825.2	0.916	BB	0.021	23173.56	1.114
71	6.136		0.0000	0.000	24240.1	0.744	BB	0.024	16782.47	0.807
72	6.206		0.0000	0.000	224.5	0.007	BB	0.018	211.50	0.010
73	6.277		0.0000	0.000	115505.9	3.547	BB	0.042	46111.31	2.216
74	6.353		0.0000	0.000	21165.8	0.650	BB	0.023	15036.71	0.723
75	6.401		0.0000	0.000	10872.8	0.334	BB	0.023	8025.00	0.386
76	6.513		0.0000	0.000	41872.9	1.286	BB	0.027	26092.54	1.254
77	6.579		0.0000	0.000	134859.5	4.141	BB	0.028	79749.66	3.833
78	6.731		0.0000	0.000	20348.1	0.625	BB	0.023	14499.16	0.697
79	6.779		0.0000	0.000	17784.9	0.546	BB	0.021	14320.27	0.688
80	6.834		0.0000	0.000	63321.5	1.944	BB	0.027	39150.72	1.882
81	6.886		0.0000	0.000	32372.6	0.994	BB	0.037	14430.32	0.694
82	6.979		0.0000	0.000	5800.9	0.178	BB	0.032	3054.99	0.147
83	7.061		0.0000	0.000	2237.7	0.069	BB	0.030	1238.18	0.060
84	7.182		0.0000	0.000	25895.5	0.795	BB	0.048	9074.03	0.436
85	7.418		0.0000	0.000	32560.3	1.000	BB	0.029	18469.81	0.888
86	7.513		0.0000	0.000	10708.9	0.329	BB	0.036	4900.76	0.236
87	7.733		0.0000	0.000	8133.8	0.250	BB	0.032	4242.40	0.204
88	7.823		0.0000	0.000	46453.3	1.426	BB	0.033	23792.38	1.144
89	7.933		0.0000	0.000	171.8	0.005	BB	0.021	136.41	0.007
90	7.980	DCB	-3.9148	-3.890	19297.0	0.592	BB	0.037	8738.21	0.420
91	8.146		0.0000	0.000	264.0	0.008	BB	0.024	180.24	0.009
92	8.208		0.0000	0.000	935.3	0.029	BB	0.039	402.24	0.019
93	8.375		0.0000	0.000	2530.8	0.078	BB	0.098	429.65	0.021
94	8.570		0.0000	0.000	3197.9	0.098	BB	0.043	1225.90	0.059
95	8.670		0.0000	0.000	3865.3	0.119	BB	0.037	1720.47	0.083
96	8.870		0.0000	0.000	100.4	0.003	BB	0.037	44.80	0.002
97	9.088		0.0000	0.000	3788.4	0.116	BB	0.105	604.18	0.029

Total Area = 3256885.0, Total Amount = 100.634, Total Height = 2080594.0, Sample Units = ug/l

ASX22: 0090

Sample Name: **5822068AF ABSTEA1 T 093060018A 00173**
Acquired from CP07--H5088B via port 4 on 11/6/09 02:00:39am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.50R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.571		0.0000	0.000	482.3	0.008	BB	0.021	380.79	0.012
2	1.620		0.0000	0.000	22035.6	0.343	BV	0.042	8742.56	0.270
3	1.748		0.0000	0.000	18458.1	0.288	VV	0.037	8400.83	0.260
4	1.813		0.0000	0.000	10741.5	0.167	VB	0.042	4287.39	0.132
5	1.909		0.0000	0.000	29837.0	0.465	BB	0.043	11566.19	0.357
6	2.039		0.0000	0.000	25666.2	0.400	BB	0.047	9128.17	0.282
7	2.117		0.0000	0.000	1322.7	0.021	BB	0.023	967.40	0.030
8	2.166		0.0000	0.000	194058.5	3.025	BB	0.032	99769.63	3.083
9	2.343	TCX	127.2936	49.139	88854.0	1.385	BB	0.027	55205.16	1.706
10	2.397		0.0000	0.000	325012.3	5.065	BB	0.038	142484.70	4.403
11	2.590		0.0000	0.000	7878.1	0.123	BB	0.048	2759.17	0.085
12	2.679		0.0000	0.000	1773.8	0.028	BB	0.026	1131.77	0.035
13	2.734		0.0000	0.000	86300.8	1.345	BB	0.037	38441.60	1.188
14	2.811		0.0000	0.000	1044.7	0.016	BB	0.020	864.65	0.027
15	2.867		0.0000	0.000	11984.8	0.187	BB	0.025	7883.79	0.244
16	2.902		0.0000	0.000	12309.0	0.192	BB	0.023	8778.76	0.271
17	3.011		0.0000	0.000	2935.1	0.046	BB	0.026	1916.86	0.059
18	3.271		0.0000	0.000	809940.1	12.623	BB	0.026	525191.40	16.229
19	3.329		0.0000	0.000	25806.5	0.402	BB	0.021	20462.44	0.632
20	3.458		0.0000	0.000	29075.0	0.453	BB	0.045	10883.62	0.336
21	3.561		0.0000	0.000	462284.8	7.205	BB	0.031	245217.80	7.578
22	3.651		0.0000	0.000	45837.9	0.714	BB	0.032	24060.42	0.743
23	3.803		0.0000	0.000	36201.3	0.564	BB	0.039	15508.83	0.479
24	3.945		0.0000	0.000	16630.2	0.259	BB	0.030	9125.63	0.282
25	3.985		0.0000	0.000	259.5	0.004	BB	0.012	375.40	0.012
26	4.112		0.0000	0.000	47848.1	0.746	BB	0.029	27800.72	0.859
27	4.141		0.0000	0.000	11186.7	0.174	BB	0.024	7670.91	0.237
28	4.207		0.0000	0.000	6036.0	0.094	BB	0.024	4189.63	0.129
29	4.251		0.0000	0.000	49494.2	0.771	BB	0.023	36292.56	1.121
30	4.289		0.0000	0.000	66957.4	1.044	BB	0.030	37587.25	1.161
31	4.380		0.0000	0.000	6111.8	0.095	BB	0.033	3102.36	0.096
32	4.436		0.0000	0.000	52866.7	0.824	BB	0.035	25374.52	0.784
33	4.501		0.0000	0.000	58459.8	0.911	BB	0.025	39082.73	1.208
34	4.537		0.0000	0.000	11471.9	0.179	BB	0.017	11549.80	0.357
35	4.595		0.0000	0.000	1564761.0	24.388	BB	0.035	735386.20	22.724
36	4.694		0.0000	0.000	134008.2	2.089	BB	0.032	69386.45	2.144
37	4.790		0.0000	0.000	253726.7	3.954	BB	0.036	117377.00	3.627
38	4.873		0.0000	0.000	199022.2	3.102	BB	0.034	97705.48	3.019
39	4.987		0.0000	0.000	124644.5	1.943	BB	0.043	48101.50	1.486
40	5.045		0.0000	0.000	95061.1	1.482	BB	0.023	69240.53	2.140
41	5.088		0.0000	0.000	89574.5	1.396	BB	0.024	63435.03	1.960
42	5.141		0.0000	0.000	314227.9	4.897	BB	0.041	127923.80	3.953
43	5.219		0.0000	0.000	9906.7	0.154	BB	0.019	8670.05	0.268
44	5.336		0.0000	0.000	5193.2	0.081	BB	0.022	3931.38	0.121
45	5.409		0.0000	0.000	144651.3	2.254	BB	0.048	50035.85	1.546
46	5.474		0.0000	0.000	26909.9	0.419	BB	0.024	19074.21	0.589
47	5.510		0.0000	0.000	4284.1	0.067	BB	0.021	3426.08	0.106
48	5.561		0.0000	0.000	2772.1	0.043	BB	0.017	2768.09	0.086
49	5.627		0.0000	0.000	286166.2	4.460	BB	0.041	116537.00	3.601
50	5.797		0.0000	0.000	7589.5	0.118	BB	0.027	4724.65	0.146
51	5.859		0.0000	0.000	74266.8	1.157	BB	0.032	38556.82	1.191
52	5.922		0.0000	0.000	116764.2	1.820	BB	0.034	56482.58	1.745
53	6.108		0.0000	0.000	2715.2	0.042	BB	0.035	1281.71	0.040
54	6.176		0.0000	0.000	5495.4	0.086	BB	0.031	2973.90	0.092

ASX22 0891

822068AF ABSTEA1 T 093060018A 00173

K#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.241		0.0000	0.000	27689.4	0.432	BB	0.039	11849.20	0.366
56	6.331		0.0000	0.000	112.9	0.002	BB	0.015	128.27	0.004
57	6.385		0.0000	0.000	61419.9	0.957	BB	0.037	27593.22	0.853
58	6.505		0.0000	0.000	1238.4	0.019	BB	0.030	680.20	0.021
59	6.617		0.0000	0.000	43370.2	0.676	BB	0.063	11447.70	0.354
60	6.819	DCB	131.7543	50.861	111999.2	1.746	BB	0.041	45484.43	1.406
61	6.952		0.0000	0.000	10156.7	0.158	BB	0.084	2019.14	0.062
62	7.344		0.0000	0.000	5630.2	0.088	BB	0.048	1938.50	0.060
63	7.401		0.0000	0.000	3349.2	0.052	BB	0.024	2297.21	0.071
64	7.458		0.0000	0.000	92082.8	1.435	BB	0.036	42524.55	1.314
65	7.743		0.0000	0.000	2636.9	0.041	BB	0.069	633.27	0.020
66	7.821		0.0000	0.000	8290.1	0.129	BB	0.034	4009.13	0.124
67	8.147		0.0000	0.000	2050.3	0.032	BB	0.059	574.63	0.018
68	8.280		0.0000	0.000	5028.8	0.078	BB	0.086	974.45	0.030
69	8.663		0.0000	0.000	1422.8	0.022	BB	0.047	502.09	0.016
70	8.993		0.0000	0.000	835.8	0.013	BB	0.052	270.42	0.008

Total Area = 6416216.0, Total Amount = 259.048, Total Height = 3236131.0, Sample Units = ug/l

ASX22 0892

822068AF ABSTEA1 T 093060018A 00173

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\3P7307.50R

C:\CPWIN\DATA1\3P7307.50R



Instrument ID: CP07-H5088A Injected On: 11/6/2009 1:51:30 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.50R

C:\CPWIN\DATA1\3P7307B.50R



ASX22 0093

Instrument ID: CP07--H5088B Injected On: 11/6/2009 1:51:30 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

)ven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

)etector A Parameters:
Threshold: 2 Width: 0.01
:alibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1041
Analyst: 2280

Dilution Factor: 10

RT A	Height A	Amount A	Compound A
3.002	62150	.10043	TCX
7.98	8738	-.00376	DCB

RT B	Height B	Amount B	Compound B
2.343	55205	.12228	TCX
6.819	45484	.12657	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.50A
Area File: C:\CPWIN\DATA1\3P7307B.50A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 2:00:46 AM
File Reported On: 11/6/2009 at 2:00:59 AM

10X REPLOTS

ASX22 0094

C:\CPWIN\DATA1\3P7307.50R



C:\CPWIN\DATA1\3P7307B.50R



ASX22 0095

Multiple Component Data Summary

Sample Name: 5822069 ACF DF50 STEA2 Sample ID:AC **Batchnumber: 093070028A**
Sample Amount:30 g Total Volume: 500 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on	Nov 07, 2009 11:00:36
Instrument	S9147A
Result file	2C6310.06R
Calibration file	1C6310
Method file	PESTD

%SSR(TCX) 102.9% Conc: 10.52769
%SSR(DCB) * 196.4% Conc: 19.96719

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**39.54**	
4.68	4.71	4.74	6147.285645	2853.606902			2
4.85	4.87	4.91	12143.140625	4721.246579			3
5.13	5.15	5.19	14707.223633	6586.283449			4
Height summation:			**32997.649903**				
Concentration		CF:	**4720.378977**	L:			
Aroclor-1260						**8.06**	
5.89	5.92	5.95	10906.928711	8029.885114			4
6.12	6.15	6.18	21277.726563	8126.008819			5
6.33	6.36	6.39	11863.74707	9257.501256			6
Height summation:			**44048.402344**				
Concentration		CF:	**8471.13173**	L:			

Analysis Report (B)

Injected on	Nov 07, 2009 11:00:36
Instrument	S9147B
Result file	2C6310B.06R
Calibration file	1C6310B
Method file	PESTDB

%SSR(TCX) 85.5% Conc: 8.751721
%SSR(DCB) * 3620.9% Conc: 368.0679

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**32.42**	
4.07	4.10	4.13	24835.605469	2018.086492			1
4.18	4.21	4.24	29600.085938	2930.620333			2
4.45	4.48	4.51	46524.765625	3938.03242			3
Height summation:			**100960.457032**				
Concentration		CF:	**2962.246415**	L:			
Aroclor-1260						**47.76**	
4.92	4.95	4.98	79132.695313	8921.526982			4
5.34	5.37	5.40	73739.15625	4492.883095			5
5.60	5.63	5.66	42003.144531	3875.692639			6
Height summation:			**194874.996094**				
Concentration		CF:	**5763.367572**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<850	<165			4	40	
Aroclor-1221			<850	<200			3	5	
Aroclor-1232			<850	<265			4	10	
Aroclor-1242			<850	<165			4	30	
Aroclor-1248			<850	<165			4	40	
Aroclor-1254	B	2962.246415	850	165		45.77	4	40	1569
Aroclor-1260	A	8471.13173	850	165	~~E~~ Baji 1157 11-10-09	38.05	4	40	

Units: ug/kg

%Difference = High - Low divided by the Average times 100

Reviewed By: Andrea Carey Date: 11-10-09

Verified By: [signature] Date: 11/11/09

ASX22 0896

11/10/2009

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822069** ACF DF50 STEA2 Sample ID: AC **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 500 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 07, 2009 11:00:36
Instrument : CP06--S9147A
Result file : 2C6310.06R
Calibration file : 1C6310.CAL
Method file : PESTD.MET
%SSR(TCX) : 102.9% Conc.: 10.527692
%SSR(DCB) : *196.4% Conc.: 19.96719 diluted out

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					5	138.50	
3.88	3.91	3.94	103.898201	45.447396			2
4.08	4.11	4.14	276.215759	89.923514			3
4.13	4.16	4.19	160.262299	93.122437			4
4.19	4.22	4.25	134.649597	87.512087			5
4.40	4.43	4.46	821.582458	717.928012			6
Height Summation:			1496.608314				
Amount Avg CF:			206.786689	Linear:			
Aroclor-1221					1		
3.43	3.46	3.49	245.350372	214.09282			1
Height Summation:			245.350372				
Amount Avg CF:			214.09282	Linear:			
Aroclor-1232					5	141.07	
3.88	3.91	3.94	103.898201	102.131329			2
4.08	4.11	4.14	276.215759	207.790385			3
4.13	4.16	4.19	160.262299	212.493104			4
4.19	4.22	4.25	134.649597	203.644279			5
4.40	4.43	4.46	821.582458	1723.479039			6
Height Summation:			1496.608314				
Amount Avg CF:			489.907627	Linear:			
Aroclor-1242					5	135.98	
3.88	3.91	3.94	103.898201	58.610143			2
4.08	4.11	4.14	276.215759	117.618702			3
4.13	4.16	4.19	160.262299	123.85988			4
4.19	4.22	4.25	134.649597	113.141414			5
4.40	4.43	4.46	821.582458	899.378717			6
Height Summation:			1496.608314				
Amount Avg CF:			262.521771	Linear:			
Aroclor-1248					6	150.49	
3.88	3.91	3.94	103.898201	119.49904			1
4.19	4.22	4.25	134.649597	142.314665			2
4.26	4.29	4.32	1146.116699	673.62362			3
4.40	4.43	4.46	821.582458	488.541657			4
4.50	4.53	4.56	614.499512	483.6766			5
4.70	4.71	4.76	6147.285645	3927.064058			6
+ 4.70	4.76	4.76	3287.404785	2100.08936			6
Height Summation:			8968.032112				
Amount Avg CF:			972.453273	Linear:			
Aroclor-1254					6	82.84	
4.26	4.29	4.32	1146.116699	780.499509			1
4.68	4.71	4.74	6147.285645	2853.606902 /			2
4.85	4.87	4.91	12143.140625	4721.246579 /			3
5.13	5.15	5.19	14707.223633	6586.283449 /			4
5.35	5.38	5.41	12568.688477	11486.940332			5
E 5.66	5.69	5.72	25475.525391	16887.626773			6
Height Summation:			72187.980470				
Amount Avg CF:			7219.367257	Linear:			

Analysis Report (B)

Injected on : NOV 07, 2009 11:00:36
Instrument : CP06--S9147B
Result file : 2C6310B.06R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET
%SSR(TCX) : 85.5% Conc.: 8.751721
%SSR(DCB) : *3620.9% Conc.: 368.067963

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	164.66	
+ 3.07	3.07	3.13	39.983147	7.701353			1
3.07	3.12	3.13	9135.780273	1759.688201			1
3.40	3.43	3.46	379.872406	53.916659			2
3.56	3.59	3.62	901.910461	71.908853			3
* 3.59	3.61	3.65	420.777588	61.71848			4
+* 3.56	3.61	3.62	420.777588	33.548379			3
3.67	3.70	3.73	447.479218	68.672522			5
3.84	3.87	3.90	2760.157715	456.986412			6
Height Summation:			14045.977661				
Amount Avg CF:			412.148521	Linear:			
Aroclor-1221					3	156.93	
2.95	2.98	3.01	200.185989	61.580531			1
+ 2.95	3.01	3.01	83.585693	25.712346			1
* 3.02	3.07	3.08	39.983147	22.860576			2
+* 3.07	3.07	3.13	39.983147	5.509142			3
3.07	3.12	3.13	9135.780273	1258.788067			3
Height Summation:			9375.949409				
Amount Avg CF:			447.743058	Linear:			
Aroclor-1232					6	117.45	
+ 3.07	3.07	3.13	39.983147	7.03941			1
3.07	3.12	3.13	9135.780273	1608.440338			1
3.40	3.43	3.46	379.872406	119.909219			2
3.56	3.59	3.62	901.910461	161.658773			3
* 3.60	3.61	3.66	420.777588	139.132225			4
+* 3.56	3.61	3.62	420.777588	75.420334			3
3.67	3.70	3.73	447.479218	161.918953			5
3.84	3.87	3.90	2760.157715	1118.468966			6
Height Summation:			14045.977661				
Amount Avg CF:			551.588079	Linear:			
Aroclor-1242					6	162.31	
+ 3.07	3.07	3.13	39.983147	9.878236			1
3.07	3.12	3.13	9135.780273	2257.085748			1
3.40	3.43	3.46	379.872406	70.487717			2
3.56	3.59	3.62	901.910461	94.453744			3
* 3.59	3.61	3.65	420.777588	81.20141			4
+* 3.56	3.61	3.62	420.777588	44.066479			3
3.67	3.70	3.73	447.479218	91.352119			5
3.84	3.87	3.90	2760.157715	621.951309			6
Height Summation:			14045.977661				
Amount Avg CF:			536.088675	Linear:			
Aroclor-1248					6	161.86	
3.28	3.31	3.34	387.463165	104.661503			1
3.56	3.59	3.62	901.910461	141.634248			2
+ 3.56	3.61	3.62	420.777588	66.078086			2
3.71	3.74	3.77	3565.405273	456.678178			3
+ 3.71	3.76	3.77	1945.350464	249.171984			3
3.84	3.87	3.90	2760.157715	329.229851			4
4.21	4.24	4.27	20754.886719	4234.95826			5
+ 4.45	4.45	4.51	12345.128906	2979.395101			6
4.45	4.48	4.51	46524.765625	11228.368683			6
Height Summation:			74894.588958				
Amount Avg CF:			2749.255121	Linear:			

ASX22 0097

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/9/09 05:36:02

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822069** ACF DF50 STEA2 Sample ID: AC **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 500 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 07, 2009 11:00:36
Instrument : CP06–S9147A
Result file : 2C6310.06R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	31.49	
5.25	5.28	5.31	12273.158203	6573.999907			1
5.45	5.48	5.51	17567.402344	10311.39642			2
E 5.66	5.69	5.72	25475.525391	15175.384747			3
5.89	5.92	5.95	10906.928711	8029.885114			4
6.12	6.15	6.18	21277.726563	8126.008819			5
6.33	6.36	6.39	11863.74707	9257.501256			6
Height Summation:			**99364.488282**				
Amount Avg CF:			**9579.029377**	Linear:			
Aroclor-1262					6	37.27	
5.25	5.28	5.31	12273.158203	7224.026216			1
5.45	5.48	5.51	17567.402344	11324.572957			2
5.73	5.75	5.79	7275.892578	3414.853885			3
5.89	5.92	5.95	10906.928711	6635.305085			4
6.12	6.15	6.18	21277.726563	6315.199699			5
6.32	6.36	6.38	11863.74707	6234.640129			6
Height Summation:			**81164.855469**				
Amount Avg CF:			**6858.099662**	Linear:			
Aroclor-1268					6	114.40	
5.90	5.92	5.96	10906.928711	9161.482975			1
6.32	6.36	6.38	11863.74707	3266.478964			2
6.38	6.40	6.44	3420.008057	1177.854209			3
6.51	6.53	6.57	323.398499	103.854283			4
6.74	6.77	6.80	4714.747559	4364.149264			5
6.96	6.99	7.02	720.028809	98.162794			6
Height Summation:			**31948.858705**				
Amount Avg CF:			**3028.663748**	Linear:			

Analysis Report (B)

Injected on : NOV 07, 2009 11:00:36
Instrument : CP06–S9147B
Result file : 2C6310B.06R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					7	68.85	
+ 4.07	4.08	4.13	4421.435547	359.276095			1
4.07	4.10	4.13	24835.605469	2018.086492			1
4.07	4.13	4.13	8054.189453	654.465661			1
4.18	4.21	4.24	29600.085938	2930.620333			2
+ 4.45	4.45	4.51	12345.128906	1044.93848			3
4.45	4.48	4.51	46524.765625	3938.03242			3
4.62	4.64	4.68	43233.152344	6432.758154			4
+ 4.62	4.67	4.68	14238.881836	2118.635313			4
4.79	4.82	4.85	44191.457031	8967.895431			5
4.92	4.95	4.98	79132.695313	9012.819033			6
Height Summation:			**275571.951173**				
Amount Avg CF:			**4850.668218**	Linear:			
Aroclor-1260					6	34.54	
4.55	4.57	4.61	44164.210938	4491.009577			1
4.68	4.71	4.74	61923.230469	5894.021884			2
4.79	4.82	4.85	44191.457031	4599.660291			3
4.92	4.95	4.98	79132.695313	8921.526982			4
+ 5.34	5.34	5.40	11174.800781	680.87399			5
5.34	5.37	5.40	73739.15625	4492.883095			5
+ 5.60	5.60	5.66	14481.734375	1336.251177			6
5.60	5.63	5.66	42003.144531	3875.692639			6
Height Summation:			**345153.894532**				
Amount Avg CF:			**5379.132411**	Linear:			
Aroclor-1262					6	52.11	
4.68	4.71	4.74	61923.230469	6625.50662			1
4.95	4.98	5.01	26744.076172	2047.458596			2
5.15	5.18	5.21	28556.134766	2321.854434			3
+ 5.15	5.20	5.21	19711.871094	1602.741256			3
+ 5.34	5.34	5.40	11174.800781	528.558203			4
5.34	5.37	5.40	73739.15625	3487.796934			4
+ 5.60	5.60	5.66	14481.734375	1011.35306			5
5.60	5.63	5.66	42003.144531	2933.350913			5
5.97	6.00	6.03	17808.884766	2326.407854			6
Height Summation:			**250774.626954**				
Amount Avg CF:			**3290.395892**	Linear:			
Aroclor-1268					7	112.41	
5.15	5.18	5.21	28556.134766	4008.871076			1
+ 5.15	5.20	5.21	19711.871094	2767.263515			1
+ 5.55	5.56	5.61	405.154449	16.711182			2
5.55	5.58	5.61	3358.341553	138.51966			2
+* 5.60	5.60	5.66	14481.734375	709.211434			3
* 5.55	5.60	5.61	14481.734375	597.320102			2
5.60	5.63	5.66	42003.144531	2057.012621			3
5.75	5.78	5.81	657.316528	31.239753			4
5.97	6.00	6.03	17808.884766	2302.738836			5
6.17	6.18	6.23	11600.320313	216.612323			6
Height Summation:			**118465.876832**				
Amount Avg CF:			**1336.04491**	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			850	165		** 66.36	4	40	
Aroclor-1221			850	165		** 70.61	3	5	
Aroclor-1232			850	165		11.84	4	10	
Aroclor-1242			850	165		** 68.51	4	30	~~ASX22 8898~~

*Peak found within more than one window
+Duplicate Peak In window - not included in average

Printed on: 11/9/09 05:36:02

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822069** ACF DF50 STEA2 Sample ID: AC **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 500 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on	: NOV 07, 2009 11:00:36
Instrument	: CP06–S9147A
Result file	: 2C6310.06R
Calibration file	: 1C6310.CAL
Method file	: PESTD.MET

Analysis Report (B)

Injected on	: NOV 07, 2009 11:00:36
Instrument	: CP06–S9147B
Result file	: 2C6310B.06R
Calibration file	: 1C6310B.CAL
Method file	: PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1248			850	165		** 95.48	4	40	
Aroclor-1254			850	165	E	39.25	4	40	
Aroclor-1260			850	165	E	** 56.16	4	40	
Aroclor-1262			0	0		** 70.31	4	40	
Aroclor-1268			0	0		** 77.56	4	40	

Units: _____

ASX22 0099

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/9/09 05:36:02

Lancaster Laboratories

Sample Name: **5822069ACF DF50 ACSTEA2 T 093070028A 01216**
Acquired from CP06--S9147A via port 1 on 11/7/09 11:08:06am by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\2C6310.06R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.027		0.0000	0.000	777.5	0.218	BB	0.021	622.46	0.231
2	3.124		0.0000	0.000	445.8	0.125	BB	0.066	113.28	0.042
3	3.216		0.0000	0.000	59.7	0.017	BB	0.019	51.36	0.019
4	3.249		0.0000	0.000	301.6	0.084	BB	0.026	195.40	0.073
5	3.311	TCX	315.8307	34.523	1292.4	0.362	BB	0.017	1235.59	0.459
6	3.414		0.0000	0.000	70.3	0.020	BB	0.013	89.40	0.033
7	3.464		0.0000	0.000	230.8	0.065	BB	0.016	245.35	0.091
8	3.618		0.0000	0.000	4147.4	1.161	BB	0.017	3981.82	1.479
9	3.733		0.0000	0.000	107.3	0.030	BB	0.018	99.49	0.037
10	3.786		0.0000	0.000	216.1	0.060	BB	0.027	135.28	0.050
11	3.912		0.0000	0.000	115.8	0.032	BB	0.019	103.90	0.039
12	4.073		0.0000	0.000	529.0	0.148	BB	0.015	602.29	0.224
13	4.113		0.0000	0.000	381.5	0.107	BB	0.023	276.22	0.103
14	4.160		0.0000	0.000	218.0	0.061	BB	0.023	160.26	0.060
15	4.219		0.0000	0.000	122.1	0.034	BB	0.015	134.65	0.050
16	4.259		0.0000	0.000	375.9	0.105	BB	0.017	363.19	0.135
17	4.291		0.0000	0.000	1077.6	0.302	BB	0.016	1146.12	0.426
18	4.322		0.0000	0.000	343.1	0.096	BB	0.014	414.78	0.154
19	4.347		0.0000	0.000	130.7	0.037	BB	0.011	191.51	0.071
20	4.432		0.0000	0.000	846.2	0.237	BB	0.017	821.58	0.305
21	4.494		0.0000	0.000	160.0	0.045	BB	0.014	195.95	0.073
22	4.532		0.0000	0.000	672.6	0.188	BB	0.018	614.50	0.228
23	4.577		0.0000	0.000	178.2	0.050	BB	0.016	184.32	0.068
24	4.668		0.0000	0.000	247.8	0.069	BV	0.019	223.17	0.083
25	4.708		0.0000	0.000	11616.6	3.251	VV	0.031	6147.29	2.284
26	4.756		0.0000	0.000	4840.7	1.355	VV	0.025	3287.40	1.221
27	4.828		0.0000	0.000	2386.3	0.668	VV	0.020	1944.98	0.723
28	4.875		0.0000	0.000	16062.4	4.495	VV	0.022	12143.14	4.511
29	4.914		0.0000	0.000	9145.4	2.559	VV	0.024	6305.52	2.342
30	4.958		0.0000	0.000	865.5	0.242	VV	0.021	697.25	0.259
31	4.983		0.0000	0.000	1030.8	0.288	VV	0.020	839.17	0.312
32	5.032		0.0000	0.000	4743.4	1.327	VB	0.020	3917.78	1.455
33	5.088		0.0000	0.000	9504.9	2.660	BV	0.023	6783.91	2.520
34	5.121		0.0000	0.000	5147.6	1.440	VV	0.021	4115.36	1.529
35	5.154		0.0000	0.000	17562.8	4.915	VB	0.020	14707.22	5.464
36	5.202		0.0000	0.000	5146.8	1.440	BB	0.022	3907.01	1.451
37	5.243		0.0000	0.000	2283.1	0.639	BB	0.015	2572.73	0.956
38	5.283		0.0000	0.000	14810.8	4.144	BB	0.020	12273.16	4.559
39	5.332		0.0000	0.000	1090.6	0.305	BB	0.016	1144.41	0.425
40	5.375		0.0000	0.000	15333.8	4.291	BB	0.020	12568.69	4.669
41	5.431		0.0000	0.000	2372.8	0.664	BB	0.017	2335.73	0.868
42	5.484		0.0000	0.000	27875.7	7.800	BB	0.026	17567.40	6.526
43	5.538		0.0000	0.000	1436.7	0.402	BB	0.014	1697.49	0.631
44	5.577		0.0000	0.000	16641.0	4.657	BB	0.020	13675.56	5.080
45	5.620		0.0000	0.000	524.4	0.147	BB	0.011	822.75	0.306
46	5.654		0.0000	0.000	1394.9	0.390	BB	0.019	1212.83	0.451
47	5.693		0.0000	0.000	34399.0	9.626	BB	0.023	25475.53	9.464
48	5.754		0.0000	0.000	7321.5	2.049	BB	0.017	7275.89	2.703
49	5.797		0.0000	0.000	5662.2	1.584	BB	0.019	5055.88	1.878
50	5.918		0.0000	0.000	21840.8	6.112	BB	0.033	10906.93	4.052
51	5.977		0.0000	0.000	4249.3	1.189	BB	0.018	3866.75	1.436
52	6.015		0.0000	0.000	2285.0	0.639	BB	0.017	2235.02	0.830
53	6.058		0.0000	0.000	297.0	0.083	BB	0.014	345.00	0.128
54	6.097		0.0000	0.000	5357.8	1.499	BB	0.020	4364.32	1.621
55	6.152		0.0000	0.000	28650.0	8.017	BB	0.022	21277.73	7.905
56	6.267		0.0000	0.000	10320.0	2.888	BB	0.023	7546.13	2.803
57	6.357		0.0000	0.000	13561.8	3.795	BB	0.019	11863.75	4.407
58	6.403		0.0000	0.000	5816.0	1.627	BB	0.028	3420.01	1.271
59	6.530		0.0000	0.000	530.3	0.148	BB	0.027	323.40	0.120
60	6.597		0.0000	0.000	146.4	0.041	BB	0.016	150.60	0.056
61	6.625		0.0000	0.000	2319.0	0.649	BB	0.019	2056.64	0.764
62	6.733		0.0000	0.000	77.2	0.022	BB	0.018	71.27	0.026

ASX22 0100

Printed on 11/7/2009 11:39:52 AM

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.774		0.0000	0.000	5886.8	1.647	BB	0.021	4714.75	1.752
64	6.834		0.0000	0.000	4296.1	1.202	BB	0.024	2960.71	1.100
65	6.988		0.0000	0.000	782.2	0.219	BB	0.018	720.03	0.267
66	7.030		0.0000	0.000	17837.4	4.991	BB	0.027	11082.61	4.117
67	7.159	DCB	599.0157	65.477	530.9	0.149	BB	0.023	380.10	0.141
68	7.403		0.0000	0.000	106.3	0.030	BB	0.032	55.53	0.021
69	7.489		0.0000	0.000	227.3	0.064	BB	0.023	162.51	0.060

Total Area = 357364.5, Total Amount = 914.846, Total Height = 269179.8, Sample Units = PPB

ASXZZ 0101

Lancaster Laboratories

Sample Name: **5822069ACF DF50 ACSTEA2 T 093070028A 01216**
Acquired from CP06--S9147B via port 2 on 11/7/09 11:08:06am by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\2C6310B.06R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.550		0.0000	0.000	1693.8	0.173	BV	0.021	1359.81	0.143
2	2.608		0.0000	0.000	451.2	0.046	VB	0.038	196.19	0.021
3	2.807	TCX	262.5516	2.323	2306.7	0.235	BB	0.016	2402.56	0.252
4	2.876		0.0000	0.000	66.1	0.007	BB	0.014	79.60	0.008
5	2.906		0.0000	0.000	429.7	0.044	BB	0.016	439.92	0.046
6	2.980		0.0000	0.000	129.3	0.013	BB	0.011	200.19	0.021
7	3.008		0.0000	0.000	175.4	0.018	BB	0.035	83.59	0.009
8	3.073		0.0000	0.000	42.0	0.004	BB	0.018	39.98	0.004
9	3.121		0.0000	0.000	8211.2	0.838	BB	0.015	9135.78	0.958
10	3.251		0.0000	0.000	87.4	0.009	BB	0.015	98.89	0.010
11	3.308		0.0000	0.000	456.1	0.047	BB	0.020	387.46	0.041
12	3.432		0.0000	0.000	498.7	0.051	BV	0.022	379.87	0.040
13	3.530		0.0000	0.000	4316.7	0.440	VV	0.017	4308.65	0.452
14	3.588		0.0000	0.000	1046.8	0.107	VV	0.019	901.91	0.095
15	3.611		0.0000	0.000	775.6	0.079	VV	0.031	420.78	0.044
16	3.700		0.0000	0.000	373.2	0.038	VV	0.014	447.48	0.047
17	3.736		0.0000	0.000	2990.8	0.305	VV	0.014	3565.41	0.374
18	3.760		0.0000	0.000	1545.3	0.158	VV	0.013	1945.35	0.204
19	3.779		0.0000	0.000	712.0	0.073	VV	0.013	948.12	0.099
20	3.812		0.0000	0.000	247.4	0.025	VB	0.014	290.44	0.030
21	3.870		0.0000	0.000	2564.0	0.262	BB	0.015	2760.16	0.289
22	3.918		0.0000	0.000	493.6	0.050	BB	0.013	648.43	0.068
23	3.965		0.0000	0.000	1473.0	0.150	BB	0.014	1716.60	0.180
24	4.010		0.0000	0.000	495.7	0.051	BB	0.014	585.17	0.061
25	4.052		0.0000	0.000	528.1	0.054	BV	0.016	542.54	0.057
26	4.079		0.0000	0.000	3941.0	0.402	VV	0.015	4421.44	0.463
27	4.103		0.0000	0.000	20997.9	2.142	VV	0.014	24835.61	2.603
28	4.128		0.0000	0.000	7079.7	0.722	VV	0.015	8054.19	0.844
29	4.150		0.0000	0.000	3208.9	0.327	VV	0.015	3470.73	0.364
30	4.169		0.0000	0.000	4563.8	0.466	VV	0.014	5486.76	0.575
31	4.206		0.0000	0.000	26206.7	2.673	VV	0.015	29600.09	3.103
32	4.240		0.0000	0.000	29312.5	2.990	VV	0.024	20754.89	2.176
33	4.322		0.0000	0.000	2572.3	0.262	VV	0.018	2396.69	0.251
34	4.358		0.0000	0.000	11742.1	1.198	VB	0.016	12417.33	1.302
35	4.417		0.0000	0.000	19119.6	1.950	BB	0.015	21935.91	2.300
36	4.451		0.0000	0.000	12056.0	1.230	BV	0.016	12345.13	1.294
37	4.479		0.0000	0.000	51219.7	5.225	VB	0.018	46524.77	4.877
38	4.521		0.0000	0.000	6355.5	0.648	BB	0.016	6530.61	0.685
39	4.575		0.0000	0.000	43753.8	4.463	BV	0.017	44164.21	4.630
40	4.610		0.0000	0.000	6258.6	0.638	VV	0.017	6065.71	0.636
41	4.643		0.0000	0.000	41588.8	4.242	VV	0.016	43233.15	4.532
42	4.670		0.0000	0.000	13334.2	1.360	VV	0.016	14238.88	1.493
43	4.713		0.0000	0.000	64919.2	6.622	VB	0.017	61923.23	6.491
44	4.820		0.0000	0.000	46476.4	4.741	BB	0.018	44191.46	4.633
45	4.871		0.0000	0.000	34150.1	3.484	BV	0.017	32752.56	3.433
46	4.913		0.0000	0.000	12617.2	1.287	VB	0.016	13107.41	1.374
47	4.952		0.0000	0.000	91455.3	9.329	BV	0.019	79132.70	8.295
48	4.984		0.0000	0.000	27232.5	2.778	VB	0.017	26744.08	2.804
49	5.024		0.0000	0.000	12660.2	1.291	BB	0.014	14871.95	1.559
50	5.088		0.0000	0.000	871.8	0.089	BB	0.018	815.32	0.085
51	5.127		0.0000	0.000	8781.8	0.896	BV	0.026	5648.81	0.592
52	5.178		0.0000	0.000	27826.9	2.839	VV	0.016	28556.13	2.994
53	5.202		0.0000	0.000	19412.2	1.980	VV	0.016	19711.87	2.066
54	5.251		0.0000	0.000	15359.1	1.567	VV	0.017	15257.58	1.599
55	5.287		0.0000	0.000	8221.6	0.839	VV	0.016	8461.62	0.887
56	5.313		0.0000	0.000	6405.5	0.653	VV	0.017	6309.67	0.661
57	5.345		0.0000	0.000	9382.8	0.957	VV	0.014	11174.80	1.171
58	5.367		0.0000	0.000	81416.6	8.305	VB	0.018	73739.16	7.730
59	5.444		0.0000	0.000	2996.4	0.306	BB	0.017	3000.95	0.315
60	5.557		0.0000	0.000	548.4	0.056	BB	0.023	405.15	0.042
61	5.582		0.0000	0.000	3931.2	0.401	BB	0.020	3358.34	0.352
62	5.600		0.0000	0.000	9891.1	1.009	BB	0.011	14481.73	1.518

ASX22 0182

Printed on 11/7/2009 11:39:59 AM

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.626		0.0000	0.000	39375.8	4.017	BB	0.016	42003.14	4.403
64	5.703		0.0000	0.000	396.3	0.040	BB	0.013	509.01	0.053
65	5.725		0.0000	0.000	250.1	0.026	BB	0.009	443.74	0.047
66	5.781		0.0000	0.000	599.4	0.061	BB	0.015	657.32	0.069
67	5.820		0.0000	0.000	1263.0	0.129	BB	0.015	1391.37	0.146
68	5.849		0.0000	0.000	382.5	0.039	BB	0.013	482.55	0.051
69	5.897		0.0000	0.000	7702.4	0.786	BB	0.017	7740.57	0.811
70	6.000		0.0000	0.000	16889.2	1.723	BB	0.016	17808.88	1.867
71	6.029		0.0000	0.000	331.0	0.034	BB	0.010	559.21	0.059
72	6.120		0.0000	0.000	219.7	0.022	BB	0.016	229.85	0.024
73	6.181		0.0000	0.000	16403.5	1.673	BB	0.024	11600.32	1.216
74	6.253		0.0000	0.000	156.1	0.016	BB	0.017	151.68	0.016
75	6.355	DCB	11042.0400	97.677	60706.7	6.192	BB	0.019	52708.28	5.525
76	6.508		0.0000	0.000	206.4	0.021	BB	0.018	192.80	0.020
77	6.690		0.0000	0.000	378.1	0.039	BB	0.023	276.40	0.029
78	6.748		0.0000	0.000	1921.2	0.196	BB	0.015	2133.73	0.224
79	6.782		0.0000	0.000	4554.5	0.465	BB	0.017	4504.89	0.472
80	6.882		0.0000	0.000	136.4	0.014	BB	0.032	70.27	0.007
81	6.990		0.0000	0.000	150.9	0.015	BB	0.017	150.81	0.016
82	7.043		0.0000	0.000	7127.8	0.727	BB	0.022	5410.88	0.567
83	7.244		0.0000	0.000	776.9	0.079	BB	0.021	606.76	0.064
84	7.296		0.0000	0.000	142.9	0.015	BB	0.025	95.20	0.010
85	7.342		0.0000	0.000	284.8	0.029	BB	0.024	199.64	0.021

Total Area = 980332.6, Total Amount = 11304.59, Total Height = 953932.6, Sample Units = PPB

ASX22 0103

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\2C6310.06R

5822069ACF DF50 ACSTEA2 T 093070028A 01216

C:\CPWIN\DATA1\2C6310.06R



Instrument ID: CP06--S9147A Injected On: 11/7/2009 11:00:36 AM Column ID: RTX CLP

File: C:\CPWIN\DATA1\2C6310B.06R

5822069ACF DF50 ACSTEA2 T 093070028A 01216

C:\CPWIN\DATA1\2C6310B.06R



ASX22 0194

Instrument ID: CP06--S9147B Injected On: 11/7/2009 11:00:36 AM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 30 Dilution Factor: 500
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	1236	10.52769	TCX	2.807	2403	8.75172	TCX
7.159	380	19.96719	DCB	6.355	52708	368.068	DCB

Files:
Area File: C:\CPWIN\DATA1\2C6310.06A
Area File: C:\CPWIN\DATA1\2C6310B.06A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/7/2009 11:39:52 AM
File Reported On: 11/7/2009 at 11:40:02 AM

REPLOTS

ASX22 8105

5822069ACF DF50 ACSTEA2 T 093070028A 01216
C:\CPWIN\DATA1\2C6310.06R



5822069ACF DF50 ACSTEA2 T 093070028A 01216
C:\CPWIN\DATA1\2C6310B.06R



ASX22 0186

Multiple Component Data Summary

Sample Name: 5822070 ACF DF1000 STEA3 Sample ID:AD **Batchnumber: 093070028A**
Sample Amount:30 g Total Volume: 10000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on Nov 07, 2009 11:24:36
Instrument S9147A
Result file 2C6310.08R
Calibration file 1C6310
Method file PESTD

%SSR(TCX) * 209.4% Conc: 21.42911:
%SSR(DCB) * 714.3% Conc: 72.61128:

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**29.26**	
4.68	4.71	4.74	5549.405762	51521.34941			2
4.85	4.88	4.91	10376.939453	80690.970156			3
5.13	5.15	5.19	10608.741211	95017.49399			4
Height summation:			**26535.086426**				
Concentration		CF:	**75743.271185**	L:			
Aroclor-1260						**8.82**	
5.89	5.92	5.95	5096.042969	75036.044812			4
6.12	6.15	6.18	8232.876953	62883.06274			5
6.33	6.36	6.39	4409.696777	68819.359031			6
Height summation:			**17738.616699**				
Concentration		CF:	**68912.822194**	L:			

Analysis Report (B)

Injected on Nov 07, 2009 11:24:36
Instrument S9147B
Result file 2C6310B.08R
Calibration file 1C6310B
Method file PESTDB

%SSR(TCX) * 191.3% Conc: 19.57619:
%SSR(DCB) * 1314.7% Conc: 133.6405:

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**18.48**	
4.07	4.10	4.13	33338.808594	54180.760252			1
4.18	4.21	4.24	33871.328125	67070.077513			2
4.45	4.48	4.51	46562.558594	78824.627202			3
Height summation:			**113772.695313**				
Concentration		CF:	**66691.821656**	L:			
Aroclor-1260						**17.83**	
4.79	4.82	4.85	32646.623047	67960.364178			3
5.34	5.37	5.40	41465.144531	50528.933707			5
5.60	5.62	5.66	27412.640625	50588.102708			6
Height summation:			**101524.408203**				
Concentration		CF:	**56359.133531**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<17000	<3300			4	40	
Aroclor-1221			<17000	<4000			3	5	
Aroclor-1232			<17000	<5300			4	10	
Aroclor-1242			<17000	<3300			4	30	
Aroclor-1248			<17000	<3300			4	40	
Aroclor-1254	A	75743.271185	17000	3300		12.71	4	40	
Aroclor-1260	A	68912.822194	17000	3300		20.04	4	40	

Units: ug/kg

%Difference = High - Low divided by the Average times 100

Reviewed By: Andrea Cavey Date: 11/10/09
Verified By: [illegible] Date: 11/11/09

ASX22 0107

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822070** ACF DF1000 STEA3 Sample ID: AD **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10000 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on : NOV 07, 2009 11:24:36
Instrument : CP06--S9147A
Result file : 2C6310.08R
Calibration file : 1C6310.CAL
Method file : PESTD.MET diluted out

%SSR(TCX) : *209.4% Conc.: 21.429113
%SSR(DCB) : *714.3% Conc.: 72.611282

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	176.58	
3.54	3.56	3.60	146.55217	1516.594538			1
3.88	3.91	3.94	90.821983	794.551318			2
4.08	4.11	4.14	285.147766	1856.627527			3
4.13	4.16	4.19	99.150192	1152.249479			4
4.19	4.22	4.25	99.644638	1295.230047			5
4.40	4.43	4.46	1244.970581	21757.992642			6
Height Summation:			**1966.287330**				
Amount Avg CF:			**4728.874259**	Linear:			
Aroclor-1221					1		
3.53	3.56	3.59	146.55217	1087.020991			3
Height Summation:			**146.552170**				
Amount Avg CF:			**1087.020991**	Linear:			
Aroclor-1232					6	186.19	
3.54	3.56	3.60	146.55217	1395.53559			1
3.88	3.91	3.94	90.821983	1785.549651			2
4.08	4.11	4.14	285.147766	4290.194328			3
4.13	4.16	4.19	99.150192	2629.281146			4
4.19	4.22	4.25	99.644638	3014.054386			5
4.40	4.43	4.46	1244.970581	52232.875226			6
Height Summation:			**1966.287330**				
Amount Avg CF:			**10891.248388**	Linear:			
Aroclor-1242					6	174.81	
3.54	3.56	3.60	146.55217	1910.096709			1
3.88	3.91	3.94	90.821983	1024.674034			2
4.08	4.11	4.14	285.147766	2428.442906			3
4.13	4.16	4.19	99.150192	1532.578901			4
4.19	4.22	4.25	99.644638	1674.559079			5
4.40	4.43	4.46	1244.970581	27257.155577			6
Height Summation:			**1966.287330**				
Amount Avg CF:			**5971.251201**	Linear:			
Aroclor-1248					7	108.71	
3.88	3.91	3.94	90.821983	2089.187231			1
4.19	4.22	4.25	99.644638	2106.340246			2
4.26	4.29	4.32	1604.02002	18855.074232			3
4.40	4.43	4.46	1244.970581	14806.060769			4
4.50	4.53	4.56	747.559631	11768.181863			5
4.70	4.71	4.76	5549.405762	70902.42156			6
4.70	4.73	4.76	4801.156738	61342.358734			6
+ 4.70	4.76	4.76	3483.45874	44506.68606			6
Height Summation:			**14137.579353**				
Amount Avg CF:			**25981.374948**	Linear:			
Aroclor-1254					6	61.10	
4.26	4.29	4.32	1604.02002	21846.585764			1
+ 4.26	4.32	4.32	500.390137	6815.261597			1
4.68	4.71	4.74	5549.405762	51521.34941 /			2
+ 4.68	4.73	4.74	4801.156738	44574.515629			2
4.85	4.88	4.91	10376.939453	80690.970156 /			3
5.13	5.15	5.19	10608.741211	95017.49399 /			4
5.35	5.37	5.41	9170.236328	167619.648977			5
5.66	5.69	5.72	12391.71875	164288.443997			6
Height Summation:			**49701.061524**				
Amount Avg CF:			**96830.748716**	Linear:			

Analysis Report (B)

Injected on : NOV 07, 2009 11:24:36
Instrument : CP06--S9147B
Result file : 2C6310B.08R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

%SSR(TCX) : *191.3% Conc.: 19.576193
%SSR(DCB) : *1314.7% Conc.: 133.640533

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					5	152.96	
3.07	3.12	3.13	294.860687	1135.89175			1
3.40	3.43	3.46	371.773407	1055.34277			2
3.56	3.59	3.62	1085.537231	1730.986402			3
3.67	3.70	3.73	489.783386	1503.294857			5
3.84	3.87	3.90	4831.681641	15999.178932			6
Height Summation:			**7073.636352**				
Amount Avg CF:			**4284.938942**	Linear:			
Aroclor-1221					1		
3.07	3.12	3.13	294.860687	812.557008			3
Height Summation:			**294.860687**				
Amount Avg CF:			**812.557008**	Linear:			
Aroclor-1232					5	163.47	
3.07	3.12	3.13	294.860687	1038.260135			1
3.40	3.43	3.46	371.773407	2347.054337			2
3.56	3.59	3.62	1085.537231	3891.4421			3
3.67	3.70	3.73	489.783386	3544.531669			5
3.84	3.87	3.90	4831.681641	39157.805665			6
Height Summation:			**7073.636352**				
Amount Avg CF:			**9995.818781**	Linear:			
Aroclor-1242					5	154.94	
3.07	3.12	3.13	294.860687	1456.965545			1
3.40	3.43	3.46	371.773407	1379.69794			2
3.56	3.59	3.62	1085.537231	2273.685907			3
3.67	3.70	3.73	489.783386	1999.768847			5
3.84	3.87	3.90	4831.681641	21774.630528			6
Height Summation:			**7073.636352**				
Amount Avg CF:			**5776.949753**	Linear:			
Aroclor-1248					6	150.24	
3.28	3.31	3.34	265.457611	1434.107556			1
3.56	3.59	3.62	1085.537231	3409.412706			2
3.71	3.73	3.77	6365.876953	16307.582822			3
+ 3.71	3.76	3.77	3014.244629	7721.645312			3
3.84	3.87	3.90	4831.681641	11526.398052			4
4.21	4.24	4.27	23799.279297	97123.107254			5
+ 4.45	4.45	4.51	13480.265625	65067.019834			6
4.45	4.48	4.51	46562.558594	224749.794093			6
Height Summation:			**82910.391327**				
Amount Avg CF:			**59091.733747**	Linear:			
Aroclor-1254					7	52.57	
+ 4.07	4.08	4.13	6378.438965	10365.957481			1
4.07	4.10	4.13	33338.808594	54180.760252 /			1
4.07	4.13	4.13	11555.474609	18779.447311			1
4.18	4.21	4.24	33871.328125	67070.077513 /			2
4.45	4.48	4.51	46562.558594	78824.627202 /			3
4.62	4.64	4.68	47043.910156	139995.387904			4
+ 4.62	4.67	4.68	12620.246094	37555.897066			4
4.79	4.82	4.85	32646.623047	132501.402456			5
4.92	4.95	4.98	58738.828125	133801.18698			6
+ 4.92	4.98	4.98	17296.84375	39400.483439			6
Height Summation:			**263757.531250**				
Amount Avg CF:			**89307.55566**	Linear:			

ASX22 0188

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/9/09 05:38:45

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822070** ACF DF1000 STEA3 Sample ID: AD **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10000 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on : NOV 07, 2009 11:24:36
Instrument : CP06--S9147A
Result file : 2C6310.08R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	37.47	
5.25	5.28	5.31	6235.838379	66803.344742			1
5.45	5.48	5.51	8259.695313	96962.534372			2
5.66	5.69	5.72	12391.71875	147631.184698			3
5.89	5.92	5.95	5096.042969	75036.044812			4
6.12	6.15	6.18	8232.876953	62883.06274			5
6.33	6.36	6.39	4409.696777	68819.359031			6
Height Summation:			**44625.869141**				
Amount Avg CF:			**86355.921733**	Linear:			
Aroclor-1262					6	45.82	
5.25	5.28	5.31	6235.838379	73408.7497			1
5.45	5.48	5.51	8259.695313	106489.872942			2
5.73	5.75	5.79	2728.170898	25608.693062			3
5.89	5.92	5.95	5096.042969	62004.255688			4
6.12	6.15	6.18	8232.876953	48870.128959			5
6.32	6.36	6.38	4409.696777	46347.705021			6
Height Summation:			**34962.321289**				
Amount Avg CF:			**60454.900895**	Linear:			
Aroclor-1268					5	106.83	
5.90	5.92	5.96	5096.042969	85610.371418			1
6.32	6.36	6.38	4409.696777	24282.685183			2
6.38	6.40	6.44	1297.956177	8940.34821			3
6.74	6.77	6.80	1943.483032	35979.232979			5
6.96	6.99	7.02	384.976898	1049.691558			6
Height Summation:			**13132.155853**				
Amount Avg CF:			**31172.46587**	Linear:			

Analysis Report (B)

Injected on : NOV 07, 2009 11:24:36
Instrument : CP06--S9147B
Result file : 2C6310B.08R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	39.60	
4.55	4.57	4.61	36557.730469	74350.300468			1
4.68	4.71	4.74	44548.9375	84805.78631			2
4.79	4.82	4.85	32646.623047	67960.364178			3
4.92	4.95	4.98	58738.828125	132445.896835			4
+ 5.34	5.34	5.40	10320.878906	12576.900718			5
5.34	5.37	5.40	41465.144531	50528.933707			5
5.60	5.62	5.66	27412.640625	50588.102708			6
Height Summation:			**241369.904297**				
Amount Avg CF:			**76779.897368**	Linear:			
Aroclor-1262					6	63.67	
4.68	4.71	4.74	44548.9375	95330.711302			1
4.95	4.98	5.01	17296.84375	26484.049176			2
5.15	5.18	5.21	17278.445313	28097.664615			3
+ 5.15	5.20	5.21	16434.158203	26724.711458			3
+ 5.34	5.34	5.40	10320.878906	9763.369054			4
5.34	5.37	5.40	41465.144531	39225.293951			4
5.60	5.62	5.66	27412.640625	38288.035485			5
5.97	6.00	6.03	9575.873047	25018.283359			6
Height Summation:			**157577.884766**				
Amount Avg CF:			**42074.006315**	Linear:			
Aroclor-1268					6	104.95	
5.15	5.18	5.21	17278.445313	48512.909899			1
+ 5.15	5.20	5.21	16434.158203	46142.39428			1
+ 5.55	5.56	5.61	234.245178	193.23563			2
5.55	5.58	5.61	8173.791504	6742.797301			2
5.60	5.62	5.66	27412.640625	26849.488709			3
5.75	5.78	5.81	400.545105	380.727695			4
5.97	6.00	6.03	9575.873047	24763.745786			5
6.17	6.20	6.23	2229.477051	832.618738			6
Height Summation:			**65070.772645**				
Amount Avg CF:			**18013.714688**	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			17000	3300		9.85	4	40	
Aroclor-1221			17000	3300		28.90	3	5	
Aroclor-1232			17000	3300		8.57	4	10	
Aroclor-1242			17000	3300		3.31	4	30	
Aroclor-1248			17000	3300		** 77.84	4	40	
Aroclor-1254			17000	3300		8.08	4	40	
Aroclor-1260			17000	3300		11.74	4	40	
Aroclor-1262			17000	3300		35.86	4	40	
Aroclor-1268			17000	3300		** 53.51	4	40	

Units: _____

ASX22 0109

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/9/09 05:38:45

Lancaster Laboratories

Sample Name: **5822070ACF DF1000 ADSTEA3 T 093070028A 01216**
Acquired from CP06--S9147A via port 1 on 11/7/09 11:32:06am by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\2C6310.08R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.027		0.0000	0.000	335.1	0.169	BB	0.021	263.19	0.182
2	3.310 TCX		642.8734	22.787	170.2	0.086	BB	0.023	125.75	0.087
3	3.564		0.0000	0.000	131.2	0.066	BB	0.015	146.55	0.102
4	3.618		0.0000	0.000	186.6	0.094	BB	0.028	109.29	0.076
5	3.787		0.0000	0.000	54.1	0.027	BB	0.013	67.80	0.047
6	3.913		0.0000	0.000	98.5	0.050	BB	0.018	90.82	0.063
7	4.073		0.0000	0.000	75.0	0.038	BB	0.015	84.16	0.058
8	4.112		0.0000	0.000	417.9	0.211	BB	0.024	285.15	0.198
9	4.157		0.0000	0.000	101.7	0.051	BB	0.017	99.15	0.069
10	4.219		0.0000	0.000	110.2	0.056	BB	0.018	99.64	0.069
11	4.291		0.0000	0.000	1546.7	0.779	BB	0.016	1604.02	1.112
12	4.321		0.0000	0.000	408.6	0.206	BB	0.014	500.39	0.347
13	4.431		0.0000	0.000	1653.3	0.833	BB	0.022	1244.97	0.863
14	4.493		0.0000	0.000	174.6	0.088	BB	0.015	197.77	0.137
15	4.531		0.0000	0.000	818.3	0.412	BB	0.018	747.56	0.518
16	4.578		0.0000	0.000	237.1	0.119	BB	0.018	222.21	0.154
17	4.667		0.0000	0.000	189.6	0.096	BB	0.017	186.93	0.130
18	4.708		0.0000	0.000	6135.9	3.092	BV	0.018	5549.41	3.846
19	4.727		0.0000	0.000	5694.5	2.869	VV	0.020	4801.16	3.328
20	4.755		0.0000	0.000	5784.0	2.914	VV	0.028	3483.46	2.414
21	4.828		0.0000	0.000	2248.5	1.133	VV	0.023	1611.26	1.117
22	4.875		0.0000	0.000	14012.0	7.060	VV	0.023	10376.94	7.193
23	4.913		0.0000	0.000	9493.0	4.783	VV	0.029	5423.83	3.759
24	4.983		0.0000	0.000	941.2	0.474	VV	0.023	687.04	0.476
25	5.031		0.0000	0.000	3677.7	1.853	VB	0.020	3063.44	2.123
26	5.087		0.0000	0.000	7737.8	3.899	BV	0.023	5570.99	3.861
27	5.121		0.0000	0.000	4073.3	2.052	VV	0.021	3232.73	2.241
28	5.153		0.0000	0.000	13555.5	6.830	VV	0.021	10608.74	7.353
29	5.202		0.0000	0.000	2934.3	1.479	VB	0.024	2070.79	1.435
30	5.243		0.0000	0.000	1483.9	0.748	BB	0.015	1647.44	1.142
31	5.282		0.0000	0.000	7824.3	3.943	BB	0.021	6235.84	4.322
32	5.332		0.0000	0.000	645.9	0.325	BB	0.014	747.17	0.518
33	5.375		0.0000	0.000	11867.0	5.980	BB	0.022	9170.24	6.356
34	5.429		0.0000	0.000	846.9	0.427	BB	0.015	922.42	0.639
35	5.483		0.0000	0.000	14265.6	7.188	BB	0.029	8259.70	5.725
36	5.537		0.0000	0.000	537.4	0.271	BB	0.012	741.12	0.514
37	5.577		0.0000	0.000	12800.4	6.450	BB	0.023	9120.85	6.322
38	5.652		0.0000	0.000	623.7	0.314	BB	0.017	602.90	0.418
39	5.692		0.0000	0.000	18366.0	9.254	BB	0.025	12391.72	8.589
40	5.753		0.0000	0.000	2639.0	1.330	BB	0.016	2728.17	1.891
41	5.796		0.0000	0.000	2330.0	1.174	BB	0.019	2021.37	1.401
42	5.917		0.0000	0.000	10549.9	5.316	BB	0.035	5096.04	3.532
43	5.975		0.0000	0.000	1573.1	0.793	BB	0.017	1501.34	1.041
44	6.014		0.0000	0.000	899.4	0.453	BB	0.017	859.96	0.596
45	6.095		0.0000	0.000	2940.7	1.482	BB	0.021	2363.62	1.638
46	6.150		0.0000	0.000	12008.4	6.051	BB	0.024	8232.88	5.706
47	6.356		0.0000	0.000	5464.9	2.754	BB	0.021	4409.70	3.056
48	6.400		0.0000	0.000	2329.6	1.174	BB	0.030	1297.96	0.900
49	6.598		0.0000	0.000	60.7	0.031	BB	0.013	76.59	0.053
50	6.625		0.0000	0.000	1023.7	0.516	BB	0.020	853.22	0.591
51	6.774		0.0000	0.000	3691.4	1.860	BB	0.032	1943.48	1.347
52	6.988		0.0000	0.000	517.5	0.261	BB	0.022	384.98	0.267
53	7.157 DCB		2178.3380	77.213	112.3	0.057	BB	0.027	69.11	0.048
54	7.376		0.0000	0.000	60.1	0.030	BB	0.025	40.13	0.028

Total Area = 198458.4, Total Amount = 2821.212, Total Height = 144273.1, Sample Units = PPB

ASX22 0118

Lancaster Laboratories

Sample Name: **5822070ACF DF1000 ADSTEA3 T 093070028A 01216**
Acquired from CP06--S9147B via port 2 on 11/7/09 11:32:06am by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\2C6310B.08R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.551		0.0000	0.000	568.4	0.073	BB	0.017	573.19	0.079
2	2.611		0.0000	0.000	374.0	0.048	BB	0.086	72.08	0.010
3	2.807	TCX	587.2858	12.777	280.6	0.036	BB	0.017	268.71	0.037
4	2.847		0.0000	0.000	67.7	0.009	BB	0.018	63.77	0.009
5	3.014		0.0000	0.000	172.2	0.022	BB	0.040	72.47	0.010
6	3.121		0.0000	0.000	188.3	0.024	BB	0.011	294.86	0.041
7	3.141		0.0000	0.000	268.2	0.035	BB	0.011	394.14	0.054
8	3.308		0.0000	0.000	476.7	0.062	BV	0.030	265.46	0.037
9	3.388		0.0000	0.000	210.0	0.027	VV	0.031	114.72	0.016
10	3.431		0.0000	0.000	370.5	0.048	VV	0.017	371.77	0.051
11	3.530		0.0000	0.000	1185.8	0.153	VV	0.021	947.70	0.130
12	3.587		0.0000	0.000	2099.0	0.271	VB	0.032	1085.54	0.149
13	3.701		0.0000	0.000	384.0	0.050	BB	0.013	489.78	0.067
14	3.735		0.0000	0.000	5207.5	0.673	BV	0.014	6365.88	0.875
15	3.760		0.0000	0.000	3279.9	0.424	VB	0.018	3014.24	0.414
16	3.869		0.0000	0.000	4589.0	0.593	BB	0.016	4831.68	0.664
17	3.917		0.0000	0.000	604.2	0.078	BB	0.012	842.62	0.116
18	3.965		0.0000	0.000	2589.6	0.334	BV	0.015	2859.70	0.393
19	4.010		0.0000	0.000	824.8	0.107	VB	0.014	970.00	0.133
20	4.052		0.0000	0.000	689.5	0.089	BV	0.017	689.28	0.095
21	4.079		0.0000	0.000	5271.4	0.681	VV	0.014	6378.44	0.877
22	4.103		0.0000	0.000	29621.4	3.826	VV	0.015	33338.81	4.584
23	4.128		0.0000	0.000	10746.0	1.388	VV	0.015	11555.47	1.589
24	4.150		0.0000	0.000	4140.2	0.535	VV	0.014	4875.07	0.670
25	4.168		0.0000	0.000	6548.1	0.846	VV	0.016	6745.49	0.928
26	4.206		0.0000	0.000	31205.8	4.031	VV	0.015	33871.33	4.658
27	4.239		0.0000	0.000	38201.6	4.934	VV	0.027	23799.28	3.273
28	4.322		0.0000	0.000	2904.3	0.375	VV	0.017	2807.32	0.386
29	4.357		0.0000	0.000	13549.0	1.750	VB	0.016	13949.21	1.918
30	4.416		0.0000	0.000	24071.1	3.109	BV	0.015	26154.99	3.596
31	4.451		0.0000	0.000	13207.9	1.706	VV	0.016	13480.27	1.854
32	4.479		0.0000	0.000	50153.4	6.478	VB	0.018	46562.56	6.403
33	4.520		0.0000	0.000	4912.4	0.634	BB	0.016	5262.74	0.724
34	4.574		0.0000	0.000	37265.4	4.813	BV	0.017	36557.73	5.027
35	4.610		0.0000	0.000	6269.0	0.810	VV	0.017	6119.46	0.841
36	4.643		0.0000	0.000	47710.5	6.162	VV	0.017	47043.91	6.469
37	4.668		0.0000	0.000	13093.9	1.691	VV	0.017	12620.25	1.735
38	4.713		0.0000	0.000	52319.0	6.758	VB	0.020	44548.94	6.126
39	4.819		0.0000	0.000	35566.5	4.594	BB	0.018	32646.62	4.489
40	4.870		0.0000	0.000	38742.7	5.004	BV	0.018	36259.11	4.986
41	4.913		0.0000	0.000	9876.8	1.276	VB	0.017	9959.28	1.369
42	4.952		0.0000	0.000	69778.4	9.013	BV	0.020	58738.83	8.077
43	4.983		0.0000	0.000	18409.9	2.378	VB	0.018	17296.84	2.378
44	5.024		0.0000	0.000	7325.5	0.946	BB	0.014	8962.34	1.232
45	5.085		0.0000	0.000	534.2	0.069	BB	0.016	540.55	0.074
46	5.128		0.0000	0.000	5532.2	0.715	BB	0.027	3366.91	0.463
47	5.178		0.0000	0.000	17324.8	2.238	BV	0.017	17278.45	2.376
48	5.202		0.0000	0.000	16546.2	2.137	VV	0.017	16434.16	2.260
49	5.251		0.0000	0.000	9089.1	1.174	VV	0.017	8951.90	1.231
50	5.286		0.0000	0.000	4832.6	0.624	VV	0.016	5053.13	0.695
51	5.312		0.0000	0.000	3100.4	0.400	VV	0.016	3192.79	0.439
52	5.344		0.0000	0.000	8918.3	1.152	VV	0.014	10320.88	1.419
53	5.367		0.0000	0.000	48909.2	6.317	VB	0.020	41465.14	5.702
54	5.442		0.0000	0.000	1182.9	0.153	BB	0.014	1401.27	0.193
55	5.557		0.0000	0.000	180.2	0.023	BB	0.013	234.25	0.032
56	5.581		0.0000	0.000	7550.9	0.975	BB	0.015	8173.79	1.124
57	5.625		0.0000	0.000	33031.4	4.266	BB	0.020	27412.64	3.769
58	5.702		0.0000	0.000	145.6	0.019	BB	0.013	185.69	0.026
59	5.780		0.0000	0.000	372.5	0.048	BB	0.016	400.55	0.055
60	5.819		0.0000	0.000	646.4	0.083	BB	0.015	730.47	0.100
61	5.849		0.0000	0.000	238.4	0.031	BB	0.013	296.84	0.041
62	5.896		0.0000	0.000	4606.1	0.595	BB	0.017	4427.67	0.609

ASX22 0111

Lancaster Laboratories

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.999		0.0000	0.000	11255.9	1.454	BB	0.020	9575.87	1.317
64	6.198		0.0000	0.000	2559.1	0.331	BB	0.019	2229.48	0.307
65	6.355	DCB	4009.2160	87.223	1306.0	0.169	BB	0.023	956.88	0.132
66	6.772		0.0000	0.000	430.3	0.056	BB	0.026	281.02	0.039
67	7.039		0.0000	0.000	471.6	0.061	BB	0.060	131.43	0.018
68	7.187		0.0000	0.000	53.7	0.007	BB	0.023	38.51	0.005
69	7.364		0.0000	0.000	72.9	0.009	BB	0.033	37.20	0.005

Total Area = 774210.8, Total Amount = 4596.502, Total Height = 727239.3, Sample Units = PPB

ASX22: 0112

Lancaster Laboratories

LANCASTER LABORATORIES
5822070ACF DF1000 ADSTEA3 T 093070028A 01216
C:\CPWIN\DATA1\2C6310.08R

File: C:\CPWIN\DATA1\2C6310.08R



Instrument ID: CP06–S9147A Injected On: 11/7/2009 11:24:36 AM Column ID: RTX CLP

File: C:\CPWIN\DATA1\2C6310B.08R

5822070ACF DF1000 ADSTEA3 T 093070028A 01216
C:\CPWIN\DATA1\2C6310B.08R



ASX22 8113

Instrument ID: CP06–S9147B Injected On: 11/7/2009 11:24:36 AM Column ID: RTX CLPII

Printed on 11/7/2009 11:40:45 AM

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 30 Dilution Factor: 10000
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.31	126	21.42911	TCX	2.807	269	19.57619	TCX
7.157	69	72.61128	DCB	6.355	957	133.6405	DCB

Files:
Area File: C:\CPWIN\DATA1\2C6310.08A
Area File: C:\CPWIN\DATA1\2C6310B.08A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/7/2009 11:40:36 AM
File Reported On: 11/7/2009 at 11:40:44 AM

REPLOTS

ASX22 8114

5822070ACF DF1000 ADSTEA3 T 093070028A 01216
C:\CPWIN\DATA1\2C6310.08R



5822070ACF DF1000 ADSTEA3 T 093070028A 01216
C:\CPWIN\DATA1\2C6310B.08R



ASX22 0115

Multiple Component Data Summary

Sample Name: 5822073 RI ACF DF100 DPS29 Sample ID:AD **Batchnumber: 093070028A**
Sample Amount:30 g Total Volume: 10000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on Nov 09, 2009 20:00:42
Instrument S9147A
Result file 3C6310.48R
Calibration file 1C6310
Method file PESTD

%SSR(TCX) * 0.0% Conc: 0
%SSR(DCB) * 0.0% Conc: 0

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**30.23**	
4.68	4.71	4.74	10431.082031	96843.418032			2
4.85	4.88	4.91	17575.232422	136664.819263			3
5.13	5.15	5.19	20116.767578	180176.404008			4
Height summation:			**48123.082031**				
Concentration		CF:	**137894.880434**	L:			
Aroclor-1260						**8.42**	
5.89	5.92	5.95	8781.307617	129299.261381			4
6.12	6.15	6.18	14585.040039	111401.153337			5
6.33	6.36	6.39	8297.264648	129490.181221			6
Height summation:			**31663.612304**				
Concentration		CF:	**123396.865313**	L:			

Analysis Report (B)

Injected on Nov 09, 2009 20:00:42
Instrument S9147B
Result file 3C6310B.48R
Calibration file 1C6310B
Method file PESTDB

%SSR(TCX) * 236.7% Conc: 24.22862
%SSR(DCB) * 2191.8% Conc: 222.7945

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**18.77**	
4.07	4.10	4.13	53028.074219	86178.885721			1
4.18	4.21	4.24	53337.109375	105615.110423			2
4.45	4.48	4.51	74392.8125	125937.789693			3
Height summation:			**180757.996094**				
Concentration		CF:	**105910.595279**	L:			
Aroclor-1260						**14.21**	
4.79	4.82	4.85	48073.519531	100074.482128			3
5.34	5.37	5.40	65643.320313	79992.172171			5
5.60	5.63	5.66	42243.820313	77958.003025			6
Height summation:			**155960.660157**				
Concentration		CF:	**86008.219108**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<17000	<3300			4	40	
Aroclor-1221			<17000	<4000			3	5	
Aroclor-1232			<17000	<5300			4	10	
Aroclor-1242			<17000	<3300			4	30	
Aroclor-1248			<17000	<3300			4	40	
Aroclor-1254	A	137894.880434	17000	3300		26.24	4	40	
Aroclor-1260	A	123396.865313	17000	3300		35.71	4	40	

Units: ug/kg

%Difference = High - Low divided by the Average times 100

Reviewed By: Andrea Carey Date: 11-10-09

Verified By: [illegible] Date: 11/11/09

ASX22 0116

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822073** RI ACF DF100 DPS29 Sample ID: AD **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:00:42
Instrument : CP06--S9147A
Result file : 3C6310.48R
Calibration file : 1C6310.CAL
Method file : PESTD.MET
%SSR(TCX) : *0% Conc.: 0
%SSR(DCB) : *0% Conc.: 0

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	156.42	
3.54	3.56	3.60	313.535736	3244.623294			1
3.88	3.91	3.94	211.885193	1853.666413			2
4.08	4.11	4.14	573.374268	3733.301033			3
4.13	4.16	4.19	224.108276	2604.419005			4
4.19	4.22	4.25	232.853958	3026.750351			5
4.40	4.43	4.46	1913.796265	33446.866688			6
Height Summation:			**3469.553696**				
Amount Avg CF:			**7984.937797**	Linear:			
Aroclor-1221					1		
3.53	3.56	3.59	313.535736	2325.587717			3
Height Summation:			**313.535736**				
Amount Avg CF:			**2325.587717**	Linear:			
Aroclor-1232					6	167.79	
3.54	3.56	3.60	313.535736	2985.62811			1
3.88	3.91	3.94	211.885193	4165.638317			2
4.08	4.11	4.14	573.374268	8626.709817			3
4.13	4.16	4.19	224.108276	5942.940228			4
4.19	4.22	4.25	232.853958	7043.37441			5
E 4.40	4.43	4.46	1913.796265	80293.529054			6
Height Summation:			**3469.553696**				
Amount Avg CF:			**18176.303323**	Linear:			
Aroclor-1242					6	154.33	
3.54	3.56	3.60	313.535736	4086.487273			1
3.88	3.91	3.94	211.885193	2390.536391			2
4.08	4.11	4.14	573.374268	4883.105672			3
4.13	4.16	4.19	224.108276	3464.074132			4
4.19	4.22	4.25	232.853958	3913.18306			5
4.40	4.43	4.46	1913.796265	41900.301368			6
Height Summation:			**3469.553696**				
Amount Avg CF:			**10106.281316**	Linear:			
Aroclor-1248					6	133.55	
3.88	3.91	3.94	211.885193	4874.016456			1
4.19	4.22	4.25	232.853958	4922.188217			2
4.26	4.29	4.32	2665.429932	31331.827908			3
4.40	4.43	4.46	1913.796265	22760.20352			4
4.50	4.53	4.56	1347.223511	21208.169395			5
4.70	4.71	4.76	10431.082031	133273.54445			6
Height Summation:			**16802.270890**				
Amount Avg CF:			**36394.991658**	Linear:			
Aroclor-1254					6	64.07	
4.26	4.29	4.32	2665.429932	36302.87832			1
4.68	4.71	4.74	10431.082031	96843.418032 /			2
4.85	4.88	4.91	17575.232422	136664.819263 /			3
5.13	5.15	5.19	20116.767578	180176.404008 /			4
5.35	5.37	5.41	17709.193359	323700.357147			5
5.66	5.69	5.72	23367.275391	309801.52083			6
Height Summation:			**91864.980713**				
Amount Avg CF:			**180581.566267**	Linear:			

Analysis Report (B)

Injected on : NOV 09, 2009 20:00:42
Instrument : CP06--S9147B
Result file : 3C6310B.48R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET
%SSR(TCX) : *236.7% Conc.: 24.228626
%SSR(DCB) : *2191.8% Conc.: 222.79454

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	134.02	
+ 3.07	3.10	3.13	155.279984	598.185043			1
3.07	3.12	3.13	452.331238	1742.515513			1
3.40	3.43	3.46	914.853943	2596.970294			2
3.56	3.59	3.62	2560.754395	4083.352381			3
3.59	3.63	3.65	1395.931763	4095.027358			4
3.67	3.70	3.73	1747.632568	5364.01831			5
+ 3.84	3.84	3.90	1283.869995	4251.287088			6
3.84	3.87	3.90	8784.054688	29086.697581			6
Height Summation:			**15855.558595**				
Amount Avg CF:			**7828.096906**	Linear:			
Aroclor-1221					2	34.91	
3.02	3.07	3.08	180.465347	2063.640332			2
+ 3.07	3.10	3.13	155.279984	427.910009			3
3.07	3.12	3.13	452.331238	1246.503632			3
Height Summation:			**632.796585**				
Amount Avg CF:			**1655.071982**	Linear:			
Aroclor-1232					6	143.38	
+ 3.07	3.10	3.13	155.279984	546.770133			1
3.07	3.12	3.13	452.331238	1592.743668			1
3.40	3.43	3.46	914.853943	5775.593074			2
3.56	3.59	3.62	2560.754395	9179.81178			3
3.60	3.63	3.66	1395.931763	9231.437113			4
3.67	3.70	3.73	1747.632568	12647.507367			5
+ 3.84	3.84	3.90	1283.869995	10404.976052			6
E 3.84	3.87	3.90	8784.054688	71189.356415			6
Height Summation:			**15855.558595**				
Amount Avg CF:			**18269.408236**	Linear:			
Aroclor-1242					6	136.38	
+ 3.07	3.10	3.13	155.279984	767.269414			1
3.07	3.12	3.13	452.331238	2235.058988			1
3.40	3.43	3.46	914.853943	3395.138213			2
3.56	3.59	3.62	2560.754395	5363.566548			3
3.59	3.63	3.65	1395.931763	5387.721735			4
3.67	3.70	3.73	1747.632568	7135.524122			5
+ 3.84	3.84	3.90	1283.869995	5785.934766			6
3.84	3.87	3.90	8784.054688	39586.537272			6
Height Summation:			**15855.558595**				
Amount Avg CF:			**10517.257813**	Linear:			
Aroclor-1248					6	147.19	
3.28	3.31	3.34	808.96698	4370.361258			1
3.56	3.59	3.62	2560.754395	8042.716843			2
3.71	3.74	3.77	10703.68457	27419.823523			3
+ 3.71	3.76	3.77	5580.794434	14296.422648			3
+ 3.84	3.84	3.90	1283.869995	3062.783873			4
3.84	3.87	3.90	8784.054688	20955.128745			4
4.21	4.24	4.27	37679.492188	153767.234519			5
+ 4.45	4.45	4.51	20115.222656	97092.863592			6
E 4.45	4.48	4.51	74392.8125	359081.84164			6
Height Summation:			**134929.765321**				
Amount Avg CF:			**95606.184421**	Linear:			

ASX22 5117

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:49:41

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822073** RI ACF DF100 DPS29 Sample ID: AD **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:00:42
Instrument : CP06–S9147A
Result file : 3C6310.48R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	40.19	
5.25	5.28	5.31	11239.061523	120401.918053			1
5.45	5.48	5.51	15375.237305	180493.579869			2
5.66	5.69	5.72	23367.275391	278390.642875			3
5.89	5.92	5.95	8781.307617	129299.261381 /			4
6.12	6.15	6.18	14585.040039	111401.153337 /			5
6.33	6.36	6.39	8297.264648	129490.181221 /			6
Height Summation:			**81645.186523**				
Amount Avg CF:			**158246.122789**	Linear:			
Aroclor-1262					6	47.71	
5.25	5.28	5.31	11239.061523	132307.061867			1
5.45	5.48	5.51	15375.237305	198228.506624			2
5.73	5.76	5.79	4774.506836	44817.163095			3
5.89	5.92	5.95	8781.307617	106843.377512			4
6.12	6.15	6.18	14585.040039	86576.392633			5
6.32	6.36	6.38	8297.264648	87207.623072			6
Height Summation:			**63052.417968**				
Amount Avg CF:			**109330.020801**	Linear:			
Aroclor-1268					6	122.48	
5.90	5.92	5.96	8781.307617	147520.53921			1
6.32	6.36	6.38	8297.264648	45690.185861			2
6.38	6.40	6.44	2400.456787	16534.39455			3
6.51	6.54	6.57	251.74733	1616.893002			4
6.74	6.78	6.80	3070.797119	56848.926982			5
6.96	6.99	7.02	696.403687	1898.8388			6
Height Summation:			**23497.977188**				
Amount Avg CF:			**45018.296401**	Linear:			

Analysis Report (B)

Injected on : NOV 09, 2009 20:00:42
Instrument : CP06–S9147B
Result file : 3C6310B.48R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					7	52.25	
+ 4.07	4.08	4.13	11913.235352	19360.86428			1
4.07	4.10	4.13	53028.074219	86178.885721 /			1
4.07	4.13	4.13	21569.298828	35053.472451			1
4.18	4.21	4.24	53337.109375	105615.110423 /			2
+ 4.45	4.45	4.51	20115.222656	34052.573029			3
4.45	4.48	4.51	74392.8125	125937.789693 /			3
4.62	4.64	4.68	80671.289063	240065.257494			4
+ 4.62	4.67	4.68	18285.023438	54413.396772			4
4.79	4.82	4.85	48073.519531	195113.863682			5
4.92	4.95	4.98	92826.664063	211449.874495			6
Height Summation:			**423898.767579**				
Amount Avg CF:			**142773.464851**	Linear:			
Aroclor-1260					6	41.37	
4.55	4.58	4.61	54297.464844	110428.978332			1
4.68	4.71	4.74	68296.234375	130012.435385			2
4.79	4.82	4.85	48073.519531	100074.482128 /			3
4.92	4.95	4.98	92826.664063	209308.070393			4
+ 5.34	5.35	5.40	17230.828125	20997.282944			5
5.34	5.37	5.40	65643.320313	79992.172171 /			5
5.60	5.63	5.66	42243.820313	77958.003025 /			6
Height Summation:			**371381.023439**				
Amount Avg CF:			**117962.356906**	Linear:			
Aroclor-1262					6	63.94	
4.68	4.71	4.74	68296.234375	146147.786402			1
4.95	4.99	5.01	24629.736328	37711.80208			2
5.15	5.18	5.21	26519.982422	43125.961752			3
+ 5.15	5.20	5.21	26309.537109	42783.742199			3
+ 5.34	5.35	5.40	17230.828125	16300.058902			4
5.34	5.37	5.40	65643.320313	62097.420962			4
5.60	5.63	5.66	42243.820313	59003.177158			5
5.97	6.00	6.03	15003.958008	39199.900741			6
Height Summation:			**242337.051759**				
Amount Avg CF:			**64547.674849**	Linear:			
Aroclor-1268					6	106.38	
5.15	5.18	5.21	26519.982422	74460.490771			1
+ 5.15	5.20	5.21	26309.537109	73869.620799			1
5.55	5.58	5.61	10649.459961	8785.047899			2
5.60	5.63	5.66	42243.820313	41375.983877			3
5.75	5.78	5.81	606.668091	576.65252			4
5.97	6.00	6.03	15003.958008	38801.07851			5
6.17	6.20	6.23	3285.913086	1227.154505			6
Height Summation:			**98309.801881**				
Amount Avg CF:			**27537.73468**	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			17000	3300		1.98	4	40	
Aroclor-1221			17000	3300		33.69	3	5	
Aroclor-1232			17000	3300	E	0.51	4	10	
Aroclor-1242			17000	3300		3.99	4	30	
Aroclor-1248			17000	3300		** 89.71	4	40	
Aroclor-1254			17000	3300		23.38	4	40	
Aroclor-1260			17000	3300		29.17	4	40	ASX22. 0116

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:49:41

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822073** RI ACF DF100 DPS29 Sample ID: AD **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:00:42
Instrument : CP06--S9147A
Result file : 3C6310.48R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Analysis Report (B)

Injected on : NOV 09, 2009 20:00:42
Instrument : CP06--S9147B
Result file : 3C6310B.48R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1262			0	0		** 51.51	4	40	
Aroclor-1268			0	0		** 48.19	4	40	

Units: _____

ASX22 0119

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:49:41

Sample Name: **5822073 RI ACF DF100ADDPS29 T 093070028A 01216**
Acquired from CP06--S9147A via port 1 on 11/9/09 08:08:12pm by 1151
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\3C6310.48R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.027		0.0000	0.000	507.0	0.146	BB	0.016	520.34	0.204
2	3.332		0.0000	0.000	1613.2	0.465	BB	0.017	1542.45	0.605
3	3.563		0.0000	0.000	329.4	0.095	BB	0.018	313.54	0.123
4	3.618		0.0000	0.000	232.3	0.067	BB	0.018	214.27	0.084
5	3.747		0.0000	0.000	174.1	0.050	BB	0.018	157.01	0.062
6	3.789		0.0000	0.000	174.5	0.050	BB	0.019	149.27	0.059
7	3.914		0.0000	0.000	426.0	0.123	BB	0.034	211.89	0.083
8	4.035		0.0000	0.000	107.3	0.031	BB	0.020	88.53	0.035
9	4.076		0.0000	0.000	130.6	0.038	BB	0.014	152.96	0.060
10	4.113		0.0000	0.000	865.9	0.249	BB	0.025	573.37	0.225
11	4.159		0.0000	0.000	246.7	0.071	BB	0.018	224.11	0.088
12	4.221		0.0000	0.000	250.6	0.072	BB	0.018	232.85	0.091
13	4.292		0.0000	0.000	2646.1	0.762	BB	0.017	2665.43	1.045
14	4.324		0.0000	0.000	289.9	0.083	BB	0.013	375.11	0.147
15	4.341		0.0000	0.000	945.0	0.272	BB	0.016	1002.17	0.393
16	4.433		0.0000	0.000	1855.6	0.534	BB	0.016	1913.80	0.750
17	4.494		0.0000	0.000	313.9	0.090	BB	0.014	366.88	0.144
18	4.533		0.0000	0.000	1518.1	0.437	BB	0.019	1347.22	0.528
19	4.578		0.0000	0.000	658.8	0.190	BB	0.018	616.28	0.242
20	4.669		0.0000	0.000	1639.1	0.472	BV	0.043	630.13	0.247
21	4.709		0.0000	0.000	18764.3	5.403	VV	0.030	10431.08	4.089
22	4.759		0.0000	0.000	6539.2	1.883	VV	0.029	3784.39	1.483
23	4.829		0.0000	0.000	4154.2	1.196	VV	0.024	2835.83	1.112
24	4.876		0.0000	0.000	23698.8	6.824	VV	0.022	17575.23	6.889
25	4.914		0.0000	0.000	15656.3	4.508	VV	0.028	9339.73	3.661
26	4.986		0.0000	0.000	1467.0	0.422	VV	0.021	1186.66	0.465
27	5.032		0.0000	0.000	6966.2	2.006	VB	0.020	5707.23	2.237
28	5.089		0.0000	0.000	14278.4	4.111	BV	0.025	9654.70	3.784
29	5.122		0.0000	0.000	6855.5	1.974	VV	0.019	5921.22	2.321
30	5.155		0.0000	0.000	26181.1	7.539	VV	0.022	20116.77	7.885
31	5.203		0.0000	0.000	4833.7	1.392	VB	0.022	3600.33	1.411
32	5.243		0.0000	0.000	1901.0	0.547	BB	0.014	2213.87	0.868
33	5.283		0.0000	0.000	13912.6	4.006	BB	0.021	11239.06	4.405
34	5.333		0.0000	0.000	1612.2	0.464	BB	0.016	1634.71	0.641
35	5.374		0.0000	0.000	19023.1	5.478	BB	0.018	17709.19	6.941
36	5.432		0.0000	0.000	1950.5	0.562	BB	0.017	1861.79	0.730
37	5.484		0.0000	0.000	25488.7	7.339	BB	0.028	15375.24	6.026
38	5.538		0.0000	0.000	1080.8	0.311	BB	0.014	1297.26	0.508
39	5.579		0.0000	0.000	21897.7	6.305	BB	0.021	17463.13	6.845
40	5.620		0.0000	0.000	543.3	0.156	BB	0.011	789.15	0.309
41	5.653		0.0000	0.000	1266.9	0.365	BB	0.018	1175.40	0.461
42	5.694		0.0000	0.000	33364.3	9.607	BB	0.024	23367.28	9.159
43	5.755		0.0000	0.000	4557.7	1.312	BB	0.016	4774.51	1.871
44	5.798		0.0000	0.000	4157.6	1.197	BB	0.019	3690.55	1.447
45	5.919		0.0000	0.000	18830.5	5.422	BB	0.036	8781.31	3.442
46	5.978		0.0000	0.000	1343.8	0.387	BB	0.020	1136.73	0.446
47	6.014		0.0000	0.000	3253.1	0.937	BB	0.024	2283.55	0.895
48	6.098		0.0000	0.000	5571.9	1.604	BB	0.020	4659.79	1.826
49	6.152		0.0000	0.000	19778.3	5.695	BB	0.023	14585.04	5.717
50	6.264		0.0000	0.000	206.5	0.059	BB	0.022	157.98	0.062
51	6.358		0.0000	0.000	9938.7	2.862	BB	0.020	8297.26	3.252
52	6.404		0.0000	0.000	3976.3	1.145	BB	0.028	2400.46	0.941
53	6.542		0.0000	0.000	380.9	0.110	BB	0.025	251.75	0.099
54	6.599		0.0000	0.000	211.4	0.061	BB	0.021	170.18	0.067
55	6.626		0.0000	0.000	1584.4	0.456	BB	0.018	1454.34	0.570
56	6.731		0.0000	0.000	239.9	0.069	BB	0.019	208.69	0.082
57	6.776		0.0000	0.000	4068.1	1.171	BB	0.022	3070.80	1.204
58	6.826		0.0000	0.000	194.6	0.056	BB	0.017	189.42	0.074
59	6.990		0.0000	0.000	871.4	0.251	BB	0.021	696.40	0.273
60	7.029		0.0000	0.000	173.9	0.050	BB	0.018	162.93	0.064
61	7.296		0.0000	0.000	628.6	0.181	BB	0.060	174.55	0.068
62	7.370		0.0000	0.000	830.8	0.239	BB	0.043	322.14	0.126

ASX22 8128

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	7.445		0.0000	0.000	131.9	0.038	BB	0.025	86.75	0.034

Total Area = 347289.6, Total Amount = 0.0, Total Height = 255132.0, Sample Units = PPB

ASX22 8121

Sample Name: **5822073 RI ACF DF100ADDPS29 T 093070028A 01216**
Acquired from CP06--S9147B via port 2 on 11/9/09 08:08:12pm by 1151
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\3C6310B.48R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.553		0.0000	0.000	875.5	0.070	BB	0.015	967.33	0.083
2	2.612		0.0000	0.000	896.4	0.072	BB	0.076	196.28	0.017
3	2.809 TCX		726.8588	9.808	324.7	0.026	BB	0.016	332.57	0.029
4	2.899		0.0000	0.000	111.4	0.009	BB	0.022	83.11	0.007
5	2.929		0.0000	0.000	3768.5	0.303	BB	0.016	3981.60	0.342
6	3.068		0.0000	0.000	194.9	0.016	BB	0.018	180.47	0.016
7	3.100		0.0000	0.000	117.5	0.009	BB	0.013	155.28	0.013
8	3.122		0.0000	0.000	256.3	0.021	BB	0.009	452.33	0.039
9	3.142		0.0000	0.000	471.4	0.038	BB	0.010	761.12	0.065
10	3.213		0.0000	0.000	316.3	0.025	BV	0.016	321.27	0.028
11	3.307		0.0000	0.000	2542.2	0.205	VV	0.052	808.97	0.070
12	3.432		0.0000	0.000	918.2	0.074	VV	0.017	914.85	0.079
13	3.466		0.0000	0.000	1049.3	0.084	VV	0.031	566.09	0.049
14	3.497		0.0000	0.000	708.9	0.057	VV	0.018	638.92	0.055
15	3.531		0.0000	0.000	3768.0	0.303	VV	0.028	2235.72	0.192
16	3.589		0.0000	0.000	3932.6	0.317	VV	0.026	2560.75	0.220
17	3.625		0.0000	0.000	3248.9	0.261	VV	0.039	1395.93	0.120
18	3.702		0.0000	0.000	2952.2	0.238	VV	0.028	1747.63	0.150
19	3.736		0.0000	0.000	9452.8	0.761	VV	0.015	10703.68	0.920
20	3.761		0.0000	0.000	6227.4	0.501	VV	0.019	5580.79	0.480
21	3.813		0.0000	0.000	2349.4	0.189	VV	0.033	1171.28	0.101
22	3.844		0.0000	0.000	1726.1	0.139	VV	0.022	1283.87	0.110
23	3.871		0.0000	0.000	10084.2	0.812	VB	0.019	8784.05	0.755
24	3.931		0.0000	0.000	7850.7	0.632	BV	0.020	6541.10	0.562
25	3.966		0.0000	0.000	6162.3	0.496	VV	0.019	5303.06	0.456
26	4.010		0.0000	0.000	3662.9	0.295	VV	0.027	2299.21	0.198
27	4.054		0.0000	0.000	2500.2	0.201	VV	0.022	1908.72	0.164
28	4.080		0.0000	0.000	10767.4	0.867	VV	0.015	11913.24	1.024
29	4.104		0.0000	0.000	47152.5	3.795	VV	0.015	53028.07	4.558
30	4.129		0.0000	0.000	29345.4	2.362	VV	0.023	21569.30	1.854
31	4.170		0.0000	0.000	9541.7	0.768	VV	0.016	10248.49	0.881
32	4.207		0.0000	0.000	46708.4	3.759	VV	0.015	53337.11	4.585
33	4.241		0.0000	0.000	53938.2	4.341	VV	0.024	37679.49	3.239
34	4.323		0.0000	0.000	4981.4	0.401	VV	0.019	4273.02	0.367
35	4.359		0.0000	0.000	20923.8	1.684	VB	0.016	21855.49	1.879
36	4.417		0.0000	0.000	18617.5	1.498	BB	0.015	20579.01	1.769
37	4.452		0.0000	0.000	18713.6	1.506	BV	0.016	20115.22	1.729
38	4.480		0.0000	0.000	77242.2	6.217	VB	0.017	74392.81	6.395
39	4.522		0.0000	0.000	7591.3	0.611	BB	0.016	7681.73	0.660
40	4.576		0.0000	0.000	55329.8	4.453	BV	0.017	54297.46	4.667
41	4.611		0.0000	0.000	10527.3	0.847	VV	0.017	10220.22	0.879
42	4.645		0.0000	0.000	78700.3	6.334	VV	0.016	80671.29	6.935
43	4.670		0.0000	0.000	24503.6	1.972	VV	0.022	18285.02	1.572
44	4.714		0.0000	0.000	78248.9	6.298	VB	0.019	68296.23	5.871
45	4.820		0.0000	0.000	48582.5	3.910	BB	0.017	48073.52	4.132
46	4.872		0.0000	0.000	64603.3	5.199	BV	0.017	62444.06	5.368
47	4.914		0.0000	0.000	14817.2	1.193	VB	0.017	14771.56	1.270
48	4.953		0.0000	0.000	108318.7	8.718	BV	0.019	92826.66	7.980
49	4.985		0.0000	0.000	23784.2	1.914	VB	0.016	24629.74	2.117
50	5.025		0.0000	0.000	10277.9	0.827	BB	0.013	12889.59	1.108
51	5.088		0.0000	0.000	1030.0	0.083	BB	0.019	905.38	0.078
52	5.139		0.0000	0.000	8753.7	0.705	BB	0.027	5407.28	0.465
53	5.179		0.0000	0.000	25225.5	2.030	BV	0.016	26519.98	2.280
54	5.203		0.0000	0.000	25970.6	2.090	VV	0.016	26309.54	2.262
55	5.252		0.0000	0.000	13335.0	1.073	VV	0.016	13529.20	1.163
56	5.288		0.0000	0.000	7144.1	0.575	VV	0.015	7878.03	0.677
57	5.314		0.0000	0.000	5135.9	0.413	VV	0.017	5146.55	0.442
58	5.346		0.0000	0.000	16507.0	1.329	VV	0.016	17230.83	1.481
59	5.368		0.0000	0.000	72630.3	5.845	VB	0.018	65643.32	5.643
60	5.443		0.0000	0.000	2168.4	0.175	BB	0.016	2280.12	0.196
61	5.540		0.0000	0.000	23812.5	1.916	BB	0.019	20817.43	1.790
62	5.581		0.0000	0.000	9091.7	0.732	BB	0.014	10649.46	0.915

ASX22 8122

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.626		0.0000	0.000	48315.9	3.889	BB	0.019	42243.82	3.631
64	5.704		0.0000	0.000	276.1	0.022	BB	0.013	348.76	0.030
65	5.782		0.0000	0.000	589.3	0.047	BB	0.016	606.67	0.052
66	5.821		0.0000	0.000	2540.8	0.204	BB	0.032	1342.02	0.115
67	5.897		0.0000	0.000	8721.0	0.702	BB	0.021	6896.38	0.593
68	6.000		0.0000	0.000	16831.4	1.355	BB	0.019	15003.96	1.290
69	6.135		0.0000	0.000	148.2	0.012	BB	0.017	147.92	0.013
70	6.199		0.0000	0.000	4215.8	0.339	BB	0.021	3285.91	0.282
71	6.327		0.0000	0.000	289.0	0.023	BB	0.019	250.30	0.022
72	6.355	DCB	6683.8360	90.192	1938.7	0.156	BB	0.020	1595.24	0.137
73	6.473		0.0000	0.000	497.0	0.040	BB	0.020	416.65	0.036
74	6.648		0.0000	0.000	862.9	0.069	BB	0.020	710.98	0.061
75	6.698		0.0000	0.000	78.8	0.006	BB	0.016	84.33	0.007
76	6.750		0.0000	0.000	5536.3	0.446	BB	0.051	1791.68	0.154
77	6.993		0.0000	0.000	157.6	0.013	BB	0.022	117.47	0.010
78	7.469		0.0000	0.000	1012.1	0.081	BB	0.086	196.26	0.017

Total Area = 1242502.0, Total Amount = 7410.695, Total Height = 1163310.0, Sample Units = PPB

ASX22 0123

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3C6310.48R
5822073 RI ACF DF100ADDPS29 T 093070028A 01216
——— C:\CPWIN\DATA1\3C6310.48R



Instrument ID: CP06--S9147A Injected On: 11/9/2009 8:00:42 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\3C6310B.48R
5822073 RI ACF DF100ADDPS29 T 093070028A 01216
——— C:\CPWIN\DATA1\3C6310B.48R



Instrument ID: CP06--S9147B Injected On: 11/9/2009 8:00:42 PM Column ID: RTX CLPII

ASX22 0124

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 30 Dilution Factor: 10000
Analyst: 1151

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
.	0	.	TCX	2.809	333	24.22863	TCX
.	0	.	DCB	6.355	1595	222.7945	DCB

Files:
Area File: C:\CPWIN\DATA1\3C6310.48A
Area File: C:\CPWIN\DATA1\3C6310B.48A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/9/2009 8:08:18 PM
File Reported On: 11/9/2009 at 8:08:27 PM

REPLOTS

ASX22 0125

5822073 RI ACF DF100ADDPS29 T 093070028A 01216
C:\CPWIN\DATA1\3C6310.48R



5822073 RI ACF DF100ADDPS29 T 093070028A 01216
C:\CPWIN\DATA1\3C6310B.48R



ASX22 0126

Multiple Component Data Summary

Sample Name: 5822074 RI ACF DF100 ELEC3 Sample ID:AC **Batchnumber: 093070028A**
Sample Amount:30 g Total Volume: 1000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on	Nov 09, 2009 20:12:41
Instrument	S9147A
Result file	3C6310.49R
Calibration file	1C6310
Method file	PESTD

%SSR(TCX) 69.4% Conc: 7.100471
%SSR(DCB) * 225.1% Conc: 22.87846

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**18.64**	
4.68	4.71	4.74	5820.625	5403.936138			2
4.85	4.88	4.91	4835.666504	3760.209099			3
5.13	5.15	5.19	4810.625488	4308.650473			4
Height summation:			**15466.916992**				
Concentration		CF:	**4490.93257**	L:			
Aroclor-1260						**9.93**	
5.89	5.92	5.95	11398.618164	16783.752189			4
6.12	6.15	6.18	25699.683594	19629.527139			5
6.33	6.36	6.39	13027.736328	20331.567204			6
Height summation:			**50126.038086**				
Concentration		CF:	**18914.948844**	L:			

Analysis Report (B)

Injected on	Nov 09, 2009 20:12:41
Instrument	S9147B
Result file	3C6310B.49R
Calibration file	1C6310B
Method file	PESTDB

%SSR(TCX) 105.9% Conc: 10.83376
%SSR(DCB) * 2028.3% Conc: 206.1802

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**5.52**	
4.07	4.12	4.13	17176.072266	2791.379455			1
4.18	4.21	4.24	13532.663086	2679.660977			2
4.45	4.49	4.51	14771.770508	2500.677236			3
Height summation:			**45480.50586**				
Concentration		CF:	**2657.239223**	L:			
Aroclor-1260						**11.94**	
4.92	4.95	4.98	60248.804688	13585.063279			4
5.34	5.37	5.40	95460.757813	11632.734813			5
5.60	5.63	5.66	58481.707031	10792.388235			6
Height summation:			**214191.269532**				
Concentration		CF:	**12003.395442**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<1700	<330			4	40	
Aroclor-1221			<1700	<400			3	5	
Aroclor-1232			<1700	<530			4	10	
Aroclor-1242			<1700	<330			4	30	
Aroclor-1248			<1700	<330			4	40	
Aroclor-1254	B	2657.239223	1700	330		51.31	4	40	1569
Aroclor-1260	B	12003.395442	1700	330		44.71	4	40	1569

Units: ug/kg

%Difference = High - Low divided by the Average times 100

Reviewed By: Andrea Carey Date: 11/1009

Verified By: [illegible] Date: 11/11/09

ASX22 8127

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822074** RI ACF DF100 ELEC3 Sample ID: AC **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 1000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:12:41
Instrument : CP06--S9147A
Result file : 3C6310.49R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

%SSR(TCX) : 69.4% Conc.: 7.100471
%SSR(DCB) : *225.1% Conc.: 22.87846

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					5	70.61	
3.54	3.56	3.60	204.333832	211.454783			1
4.08	4.12	4.14	238.410477	155.231605			3
4.13	4.15	4.19	671.945801	780.885225			4
4.19	4.22	4.25	236.272812	307.11903			5
4.40	4.43	4.46	173.932343	303.976551			6
Height Summation:			1524.895265				
Amount Avg CF:			351.733439	Linear:			
Aroclor-1221					2	86.69	
+ 3.43	3.45	3.49	90.792801	158.45166			1
3.43	3.47	3.49	361.937805	631.65411			1
3.53	3.56	3.59	204.333832	151.560475			3
Height Summation:			566.271637				
Amount Avg CF:			391.607293	Linear:			
Aroclor-1232					5	81.76	
3.54	3.56	3.60	204.333832	194.575853			1
4.08	4.12	4.14	238.410477	358.700785			3
4.13	4.15	4.19	671.945801	1781.876958			4
4.19	4.22	4.25	236.272812	714.678802			5
4.40	4.43	4.46	173.932343	729.735024			6
Height Summation:			1524.895265				
Amount Avg CF:			755.913484	Linear:			
Aroclor-1242					5	73.25	
3.54	3.56	3.60	204.333832	266.319755			1
4.08	4.12	4.14	238.410477	203.040774			3
4.13	4.15	4.19	671.945801	1038.636372			4
4.19	4.22	4.25	236.272812	397.063796			5
4.40	4.43	4.46	173.932343	380.804254			6
Height Summation:			1524.895265				
Amount Avg CF:			457.17299	Linear:			
Aroclor-1248					5	176.08	
4.19	4.22	4.25	236.272812	499.445773			2
4.26	4.29	4.32	362.85199	426.528417			3
4.40	4.43	4.46	173.932343	206.852506			4
4.50	4.53	4.56	250.779373	394.780182			5
4.70	4.71	4.76	5820.625	7436.76756			6
Height Summation:			6844.461518				
Amount Avg CF:			1792.874888	Linear:			
Aroclor-1254					6	113.54	
4.26	4.29	4.32	362.85199	494.200635			1
4.68	4.71	4.74	5820.625	5403.938138 /			2
4.85	4.88	4.91	4835.666504	3760.209099 /			3
5.13	5.15	5.19	4810.625488	4308.650473 /			4
5.35	5.37	5.41	3770.419434	6891.822189			5
5.66	5.69	5.72	18539.40625	24579.400702			6
Height Summation:			38139.594666				
Amount Avg CF:			7573.036873	Linear:			

Analysis Report (B)

Injected on : NOV 09, 2009 20:12:41
Instrument : CP06--S9147B
Result file : 3C6310B.49R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

%SSR(TCX) : 105.9% Conc.: 10.83376
%SSR(DCB) : *2028.3% Conc.: 206.180237

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	65.02	
3.07	3.12	3.13	1072.903564	413.314613			1
3.40	3.41	3.46	485.041351	137.687331			2
+ 3.40	3.43	3.46	404.46463	114.814243			2
3.40	3.45	3.46	692.876831	196.685008			2
+ 3.56	3.58	3.62	1257.36731	200.498486			3
* 3.59	3.61	3.65	1301.0896	381.680369			4
* 3.56	3.61	3.62	1301.0896	207.470397			3
3.84	3.87	3.90	2274.609619	753.192967			6
Height Summation:			7127.610565				
Amount Avg CF:			348.338448	Linear:			
Aroclor-1221					4	63.77	
2.95	2.96	3.01	910.146729	559.95246			1
+ 3.02	3.02	3.08	340.115814	388.926031			2
3.02	3.04	3.08	1330.21582	1521.11586			2
3.02	3.07	3.08	853.451233	975.930512			2
3.07	3.12	3.13	1072.903564	295.66346			3
Height Summation:			4166.717346				
Amount Avg CF:			838.165573	Linear:			
Aroclor-1232					6	81.86	
3.07	3.12	3.13	1072.903564	377.789596			1
3.40	3.41	3.46	485.041351	306.212974			2
+ 3.40	3.43	3.46	404.46463	255.343832			2
3.40	3.45	3.46	692.876831	437.422242			2
+ 3.56	3.58	3.62	1257.36731	450.741987			3
* 3.60	3.61	3.66	1301.0896	860.423635			4
* 3.56	3.61	3.62	1301.0896	466.415587			3
3.84	3.87	3.90	2274.609619	1843.431088			6
Height Summation:			7127.610565				
Amount Avg CF:			715.28252	Linear:			
Aroclor-1242					6	67.23	
3.07	3.12	3.13	1072.903564	530.143079			1
3.40	3.41	3.46	485.041351	180.004955			2
+ 3.40	3.43	3.46	404.46463	150.101919			2
3.40	3.45	3.46	692.876831	257.135319			2
+ 3.56	3.58	3.62	1257.36731	263.358847			3
* 3.59	3.61	3.65	1301.0896	502.167004			4
* 3.56	3.61	3.62	1301.0896	272.516594			3
3.84	3.87	3.90	2274.609619	1025.083764			6
Height Summation:			7127.610565				
Amount Avg CF:			461.175119	Linear:			
Aroclor-1248					6	136.67	
3.28	3.30	3.34	727.232971	392.880165			1
+ 3.56	3.58	3.62	1257.36731	394.908987			2
3.56	3.61	3.62	1301.0896	408.641112			2
3.71	3.74	3.77	2580.738281	661.112421			3
+ 3.71	3.76	3.77	1981.286865	507.549861			3
3.84	3.87	3.90	2274.609619	542.627967			4
4.21	4.24	4.27	6354.910645	2593.392264			5
+ 4.45	4.46	4.51	3710.751221	1791.118439			6
4.45	4.49	4.51	14771.770508	7130.090099			6
Height Summation:			28010.351624				
Amount Avg CF:			1954.790671	Linear:			

ASX22 0128

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:50:41

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822074** RI ACF DF100 ELEC3 Sample ID: AC **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 1000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:12:41
Instrument : CP06--S9147A
Result file : 3C6310.49R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	18.23	
5.25	5.28	5.31	11752.773438	12590.521562			1
5.45	5.48	5.51	16888.888672	19826.269448			2
5.66	5.69	5.72	18539.40625	22087.287192			3
5.89	5.92	5.95	11398.618164	16783.752189 /			4
6.12	6.15	6.18	25699.683594	19629.527139 /			5
6.33	6.36	6.39	13027.736328	20331.567204 /			6
Height Summation:			97307.106446				
Amount Avg CF:			18541.487456	Linear:			
Aroclor-1262					6	29.81	
5.25	5.28	5.31	11752.773438	13835.451645			1
5.45	5.48	5.51	16888.888672	21774.357778			2
5.73	5.76	5.79	8886.287109	8341.346915			3
5.89	5.92	5.95	11398.618164	13868.855491			4
6.12	6.15	6.18	25699.683594	15255.260811			5
6.32	6.36	6.38	13027.736328	13692.680267			6
Height Summation:			87653.987305				
Amount Avg CF:			14461.325485	Linear:			
Aroclor-1268					6	111.25	
5.90	5.92	5.96	11398.618164	19148.973833			1
6.32	6.36	6.38	13027.736328	7173.926823			2
6.38	6.41	6.44	3937.249512	2711.98537			3
6.51	6.53	6.57	327.534607	210.365057			4
6.74	6.78	6.80	5251.809082	9722.547581			5
6.96	6.99	7.02	833.655701	227.307497			6
Height Summation:			34776.603394				
Amount Avg CF:			6532.517694	Linear:			

Analysis Report (B)

Injected on : NOV 09, 2009 20:12:41
Instrument : CP06--S9147B
Result file : 3C6310B.49R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					7	97.43	
+ 4.07	4.08	4.13	2347.812012	381.556046			1
4.07	4.10	4.13	9996.032227	1624.511038			1
4.07	4.12	4.13	17176.072266	2791.379455 /			1
+ 4.18	4.18	4.24	3159.199707	625.566758			2
4.18	4.21	4.24	13532.663086	2679.660977 /			2
+ 4.45	4.46	4.51	3710.751221	628.184083			3
4.45	4.49	4.51	14771.770508	2500.677236 /			3
4.62	4.65	4.68	11259.913086	3350.77568			4
+ 4.62	4.67	4.68	9777.251953	2909.558698			4
4.79	4.82	4.85	37386.382813	15173.845541			5
4.92	4.95	4.98	60248.804688	13724.076286			6
Height Summation:			164371.638674				
Amount Avg CF:			5977.846602	Linear:			
Aroclor-1260					6	23.75	
4.55	4.58	4.61	34803.546875	7078.268083			1
4.68	4.71	4.74	57109.203125	10871.619276			2
4.79	4.82	4.85	37386.382813	7782.710597			3
4.92	4.95	4.98	60248.804688	13585.063279 /			4
+ 5.34	5.35	5.40	6664.306152	812.10445			5
5.34	5.37	5.40	95460.757813	11632.734813 /			5
5.60	5.63	5.66	58481.707031	10792.388235 /			6
Height Summation:			343490.402345				
Amount Avg CF:			10290.464047	Linear:			
Aroclor-1262					6	34.40	
4.68	4.71	4.74	57109.203125	12220.854775			1
4.95	4.99	5.01	32974.292969	5048.856365			2
5.15	5.18	5.21	36932	6005.765743			3
+ 5.34	5.35	5.40	6664.306152	630.431585			4
5.34	5.37	5.40	95460.757813	9030.41899			4
5.60	5.63	5.66	58481.707031	8168.310761			5
5.97	6.00	6.03	22744.671875	5942.357873			6
Height Summation:			303702.632813				
Amount Avg CF:			7736.094085	Linear:			
Aroclor-1268					6	107.64	
5.15	5.18	5.21	36932	10369.444449			1
5.55	5.58	5.61	12446.205078	1026.723498			2
5.60	5.63	5.66	58481.707031	5728.028737			3
5.75	5.78	5.81	796.330627	75.693129			4
5.97	6.00	6.03	22744.671875	5881.899954			5
+ 6.17	6.18	6.23	1503.636719	56.154698			6
6.17	6.20	6.23	2599.59082	97.084113			6
Height Summation:			134000.505431				
Amount Avg CF:			3863.145647	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			1700	330		0.97	4	40	
Aroclor-1221			1700	330		** 72.62	3	5	
Aroclor-1232			1700	330		5.52	4	10	
Aroclor-1242			1700	330		0.87	4	30	
Aroclor-1248			1700	330		8.64	4	40	
Aroclor-1254			1700	330		23.54	4	40	
Aroclor-1260			1700	330		** 57.24	4	40	

ASX22 6129

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:50:41

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822074** RI ACF DF100 ELEC3 Sample ID: AC **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 1000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on	: NOV 09, 2009 20:12:41
Instrument	: CP06–S9147A
Result file	: 3C6310.49R
Calibration file	: 1C6310.CAL
Method file	: PESTD.MET

Analysis Report (B)

Injected on	: NOV 09, 2009 20:12:41
Instrument	: CP06–S9147B
Result file	: 3C6310B.49R
Calibration file	: 1C6310B.CAL
Method file	: PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1262			0	0		** 60.59	4	40	
Aroclor-1268			0	0		** 51.36	4	40	

Units:

ASX22 0130

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:50:41

Sample Name: **5822074 RI ACF DF100ACELEC3 T 093070028A 01216**
Acquired from CP06--S9147A via port 1 on 11/9/09 08:20:11pm by 1151
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\3C6310.49R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.029		0.0000	0.000	2520.1	0.894	BB	0.015	2743.52	1.323
2	3.057		0.0000	0.000	105.6	0.037	BB	0.012	146.34	0.071
3	3.111		0.0000	0.000	55.0	0.020	BB	0.011	86.48	0.042
4	3.170		0.0000	0.000	236.3	0.084	BB	0.024	161.79	0.078
5	3.219		0.0000	0.000	340.8	0.121	BB	0.022	259.39	0.125
6	3.276		0.0000	0.000	527.8	0.187	BB	0.025	354.21	0.171
7	3.313	TCX	213.0141	23.685	304.4	0.108	BB	0.012	416.68	0.201
8	3.333		0.0000	0.000	1844.4	0.655	BB	0.013	2350.64	1.134
9	3.374		0.0000	0.000	141.4	0.050	BB	0.015	154.14	0.074
10	3.447		0.0000	0.000	55.6	0.020	BB	0.010	90.79	0.044
11	3.472		0.0000	0.000	730.6	0.259	BB	0.034	361.94	0.175
12	3.563		0.0000	0.000	294.2	0.104	BB	0.024	204.33	0.099
13	3.620		0.0000	0.000	233.0	0.083	BB	0.015	252.12	0.122
14	3.653		0.0000	0.000	156.7	0.056	BB	0.015	174.71	0.084
15	3.748		0.0000	0.000	206.0	0.073	BB	0.016	215.28	0.104
16	3.840		0.0000	0.000	73.0	0.026	BB	0.018	66.09	0.032
17	3.859		0.0000	0.000	83.0	0.029	BB	0.010	136.39	0.066
18	4.023		0.0000	0.000	55.4	0.020	BB	0.027	34.54	0.017
19	4.074		0.0000	0.000	1071.4	0.380	BB	0.017	1025.90	0.495
20	4.117		0.0000	0.000	230.8	0.082	BB	0.016	238.41	0.115
21	4.152		0.0000	0.000	1251.1	0.444	BB	0.031	671.95	0.324
22	4.215		0.0000	0.000	266.3	0.095	BB	0.019	236.27	0.114
23	4.292		0.0000	0.000	582.2	0.207	BB	0.027	362.85	0.175
24	4.341		0.0000	0.000	2143.2	0.761	BB	0.018	2025.98	0.977
25	4.376		0.0000	0.000	541.2	0.192	BB	0.014	665.37	0.321
26	4.434		0.0000	0.000	257.9	0.092	BB	0.025	173.93	0.084
27	4.530		0.0000	0.000	560.3	0.199	BB	0.037	250.78	0.121
28	4.576		0.0000	0.000	452.7	0.161	BB	0.020	369.08	0.178
29	4.672		0.0000	0.000	2397.3	0.851	BV	0.065	610.39	0.294
30	4.709		0.0000	0.000	8069.6	2.864	VB	0.023	5820.63	2.808
31	4.766		0.0000	0.000	1829.5	0.649	BV	0.044	700.06	0.338
32	4.830		0.0000	0.000	2396.2	0.850	VV	0.026	1530.43	0.738
33	4.877		0.0000	0.000	6453.6	2.290	VV	0.022	4835.67	2.333
34	4.909		0.0000	0.000	5559.4	1.973	VV	0.040	2303.29	1.111
35	5.033		0.0000	0.000	962.4	0.342	VB	0.021	771.81	0.372
36	5.087		0.0000	0.000	11376.4	4.037	BV	0.024	7751.98	3.739
37	5.155		0.0000	0.000	8053.9	2.858	VB	0.028	4810.63	2.320
38	5.204		0.0000	0.000	8629.7	3.063	BB	0.030	4756.64	2.294
39	5.284		0.0000	0.000	15443.2	5.481	BB	0.022	11752.77	5.669
40	5.374		0.0000	0.000	4087.0	1.450	BB	0.018	3770.42	1.819
41	5.433		0.0000	0.000	2623.5	0.931	BB	0.021	2121.15	1.023
42	5.484		0.0000	0.000	25828.7	9.167	BB	0.025	16888.89	8.146
43	5.539		0.0000	0.000	1905.4	0.676	BB	0.014	2228.85	1.075
44	5.579		0.0000	0.000	11553.8	4.100	BB	0.022	8696.60	4.195
45	5.657		0.0000	0.000	1159.9	0.412	BB	0.019	1037.69	0.501
46	5.694		0.0000	0.000	25047.2	8.889	BB	0.023	18539.41	8.943
47	5.756		0.0000	0.000	9620.0	3.414	BB	0.018	8886.29	4.286
48	5.799		0.0000	0.000	6622.5	2.350	BB	0.019	5767.03	2.782
49	5.920		0.0000	0.000	18463.6	6.553	BB	0.027	11398.62	5.498
50	5.979		0.0000	0.000	3846.0	1.365	BB	0.018	3513.83	1.695
51	6.015		0.0000	0.000	3489.0	1.238	BB	0.019	3072.08	1.482
52	6.098		0.0000	0.000	4078.5	1.447	BB	0.020	3338.12	1.610
53	6.153		0.0000	0.000	33905.6	12.033	BB	0.022	25699.68	12.396
54	6.268		0.0000	0.000	3136.5	1.113	BB	0.023	2303.99	1.111
55	6.359		0.0000	0.000	15472.3	5.491	BB	0.020	13027.74	6.284
56	6.405		0.0000	0.000	6299.7	2.236	BB	0.027	3937.25	1.899
57	6.532		0.0000	0.000	586.2	0.208	BB	0.030	327.53	0.158
58	6.597		0.0000	0.000	224.2	0.080	BB	0.015	248.49	0.120
59	6.627		0.0000	0.000	2635.7	0.935	BB	0.019	2333.52	1.126
60	6.734		0.0000	0.000	249.6	0.089	BB	0.021	200.53	0.097
61	6.776		0.0000	0.000	6597.1	2.341	BB	0.021	5251.81	2.533
62	6.833		0.0000	0.000	1056.5	0.375	BB	0.023	762.89	0.368

ASX22 0131

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.991		0.0000	0.000	943.5	0.335	BB	0.019	833.66	0.402
64	7.031		0.0000	0.000	3925.7	1.393	BB	0.027	2451.49	1.182
65	7.162	DCB	686.3538	76.315	365.7	0.130	BB	0.028	217.76	0.105
66	7.285		0.0000	0.000	483.3	0.172	BB	0.054	148.27	0.072
67	7.368		0.0000	0.000	1044.4	0.371	BB	0.044	393.31	0.190
68	7.487		0.0000	0.000	28.8	0.010	BB	0.011	44.85	0.022

Total Area = 281771.6, Total Amount = 899.368, Total Height = 207315.9, Sample Units = PPB

ASX22 0132

Sample Name: **5822074 RI ACF DF100ACELEC3 T 093070028A 01216**
Acquired from CP06--S9147B via port 2 on 11/9/09 08:20:11pm by 1151
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\3C6310B.49R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.553		0.0000	0.000	6215.0	0.711	BV	0.016	6595.98	0.820
2	2.606		0.0000	0.000	1341.0	0.153	VV	0.030	755.00	0.094
3	2.642		0.0000	0.000	324.1	0.037	VV	0.017	317.40	0.039
4	2.667		0.0000	0.000	521.5	0.060	VV	0.024	356.23	0.044
5	2.731		0.0000	0.000	1656.8	0.189	VB	0.023	1182.46	0.147
6	2.790		0.0000	0.000	158.7	0.018	BB	0.011	240.57	0.030
7	2.809	TCX	325.0128	4.992	1127.0	0.129	BB	0.013	1487.07	0.185
8	2.859		0.0000	0.000	180.2	0.021	BB	0.019	158.30	0.020
9	2.876		0.0000	0.000	990.8	0.113	BB	0.013	1302.83	0.162
10	2.899		0.0000	0.000	718.2	0.082	BB	0.010	1234.24	0.153
11	2.928		0.0000	0.000	5311.1	0.607	BB	0.014	6211.93	0.772
12	2.965		0.0000	0.000	661.9	0.076	BB	0.012	910.15	0.113
13	3.017		0.0000	0.000	399.5	0.046	BB	0.020	340.12	0.042
14	3.040		0.0000	0.000	1106.1	0.126	BB	0.014	1330.22	0.165
15	3.067		0.0000	0.000	575.8	0.066	BB	0.011	853.45	0.106
16	3.121		0.0000	0.000	1394.3	0.159	BB	0.022	1072.90	0.133
17	3.171		0.0000	0.000	531.7	0.061	BB	0.024	374.23	0.047
18	3.236		0.0000	0.000	415.2	0.047	BV	0.040	171.77	0.021
19	3.301		0.0000	0.000	946.5	0.108	VV	0.022	727.23	0.090
20	3.338		0.0000	0.000	1285.2	0.147	VV	0.040	539.39	0.067
21	3.412		0.0000	0.000	518.6	0.059	VV	0.018	485.04	0.060
22	3.434		0.0000	0.000	272.3	0.031	VV	0.011	404.46	0.050
23	3.450		0.0000	0.000	1064.0	0.122	VV	0.026	692.88	0.086
24	3.511		0.0000	0.000	1591.2	0.182	VV	0.021	1246.76	0.155
25	3.531		0.0000	0.000	8114.0	0.928	VV	0.019	7258.11	0.902
26	3.578		0.0000	0.000	1325.1	0.151	VV	0.018	1257.37	0.156
27	3.613		0.0000	0.000	2552.5	0.292	VV	0.033	1301.09	0.162
28	3.657		0.0000	0.000	8072.7	0.923	VV	0.059	2289.52	0.285
29	3.737		0.0000	0.000	4152.2	0.475	VV	0.027	2580.74	0.321
30	3.761		0.0000	0.000	2801.5	0.320	VB	0.024	1981.29	0.246
31	3.813		0.0000	0.000	6835.8	0.781	BV	0.028	4058.32	0.505
32	3.871		0.0000	0.000	4779.8	0.546	VB	0.035	2274.61	0.283
33	3.931		0.0000	0.000	9966.7	1.139	BV	0.019	8814.12	1.096
34	3.967		0.0000	0.000	3597.2	0.411	VV	0.027	2257.49	0.281
35	4.006		0.0000	0.000	3390.4	0.388	VV	0.026	2139.97	0.266
36	4.049		0.0000	0.000	4864.5	0.556	VV	0.041	1964.22	0.244
37	4.080		0.0000	0.000	3609.0	0.413	VV	0.026	2347.81	0.292
38	4.104		0.0000	0.000	8364.1	0.956	VV	0.014	9996.03	1.243
39	4.122		0.0000	0.000	19318.6	2.208	VV	0.019	17176.07	2.135
40	4.184		0.0000	0.000	5781.2	0.661	VV	0.030	3159.20	0.393
41	4.207		0.0000	0.000	14049.9	1.606	VV	0.017	13532.66	1.683
42	4.242		0.0000	0.000	9806.6	1.121	VB	0.026	6354.91	0.790
43	4.323		0.0000	0.000	2279.8	0.261	BV	0.034	1131.04	0.141
44	4.359		0.0000	0.000	3806.5	0.435	VV	0.019	3367.62	0.419
45	4.388		0.0000	0.000	20211.2	2.311	VV	0.015	22356.29	2.780
46	4.417		0.0000	0.000	1871.8	0.214	VB	0.013	2430.29	0.302
47	4.463		0.0000	0.000	4000.1	0.457	BV	0.018	3710.75	0.461
48	4.493		0.0000	0.000	21913.8	2.505	VB	0.025	14771.77	1.837
49	4.523		0.0000	0.000	5108.8	0.584	BB	0.015	5549.19	0.690
50	4.575		0.0000	0.000	32270.9	3.689	BV	0.015	34803.55	4.327
51	4.612		0.0000	0.000	1092.0	0.125	VV	0.018	1012.76	0.126
52	4.645		0.0000	0.000	10746.5	1.229	VV	0.016	11259.91	1.400
53	4.671		0.0000	0.000	15509.7	1.773	VV	0.026	9777.25	1.216
54	4.715		0.0000	0.000	54534.1	6.234	VV	0.016	57109.20	7.100
55	4.771		0.0000	0.000	7018.2	0.802	VB	0.016	7182.68	0.893
56	4.821		0.0000	0.000	38251.7	4.373	BV	0.017	37386.38	4.648
57	4.873		0.0000	0.000	10933.5	1.250	VB	0.018	10089.26	1.254
58	4.916		0.0000	0.000	10216.1	1.168	BV	0.016	10901.67	1.355
59	4.952		0.0000	0.000	67607.8	7.729	VV	0.019	60248.80	7.491
60	4.986		0.0000	0.000	31973.0	3.655	VB	0.016	32974.29	4.100
61	5.026		0.0000	0.000	18269.0	2.089	BB	0.015	20112.59	2.501
62	5.128		0.0000	0.000	8613.3	0.985	BV	0.022	6668.93	0.829

ASX22 0133

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.179		0.0000	0.000	43319.0	4.952	VV	0.020	36932.00	4.592
64	5.252		0.0000	0.000	20507.3	2.344	VV	0.017	19966.78	2.482
65	5.288		0.0000	0.000	11372.1	1.300	VV	0.017	10985.52	1.366
66	5.314		0.0000	0.000	8184.8	0.936	VV	0.017	7952.93	0.989
67	5.348		0.0000	0.000	5336.3	0.610	VV	0.013	6664.31	0.829
68	5.368		0.0000	0.000	101534.9	11.607	VB	0.018	95460.76	11.869
69	5.444		0.0000	0.000	2938.6	0.336	BB	0.016	3047.10	0.379
70	5.540		0.0000	0.000	20356.5	2.327	BB	0.019	17754.29	2.207
71	5.582		0.0000	0.000	12568.0	1.437	BB	0.017	12446.21	1.547
72	5.626		0.0000	0.000	60480.1	6.914	BB	0.017	58481.71	7.271
73	5.703		0.0000	0.000	451.1	0.052	BB	0.025	301.71	0.038
74	5.780		0.0000	0.000	730.0	0.083	BB	0.015	796.33	0.099
75	5.820		0.0000	0.000	1259.9	0.144	BB	0.015	1413.59	0.176
76	5.847		0.0000	0.000	232.1	0.027	BB	0.012	312.72	0.039
77	5.897		0.0000	0.000	11731.4	1.341	BB	0.020	9810.89	1.220
78	6.000		0.0000	0.000	24167.2	2.763	BB	0.018	22744.67	2.828
79	6.080		0.0000	0.000	980.5	0.112	BB	0.028	579.61	0.072
80	6.181		0.0000	0.000	1199.9	0.137	BB	0.013	1503.64	0.187
81	6.199		0.0000	0.000	1625.2	0.186	BB	0.010	2599.59	0.323
82	6.355	DCB	6185.4070	95.008	18829.2	2.153	BB	0.021	14762.77	1.835
83	6.471		0.0000	0.000	491.6	0.056	BB	0.024	334.45	0.042
84	6.647		0.0000	0.000	351.1	0.040	BB	0.020	286.15	0.036
85	6.693		0.0000	0.000	161.5	0.018	BB	0.021	130.12	0.016
86	6.748		0.0000	0.000	2184.1	0.250	BB	0.019	1954.63	0.243
87	6.782		0.0000	0.000	934.8	0.107	BB	0.014	1100.84	0.137
88	6.824		0.0000	0.000	600.4	0.069	BB	0.017	594.44	0.074
89	6.901		0.0000	0.000	374.7	0.043	BB	0.018	346.26	0.043
90	6.992		0.0000	0.000	209.9	0.024	BB	0.016	216.06	0.027
91	7.020		0.0000	0.000	81.8	0.009	BB	0.012	113.46	0.014
92	7.043		0.0000	0.000	1516.7	0.173	BB	0.017	1493.41	0.186
93	7.174		0.0000	0.000	140.6	0.016	BB	0.026	89.90	0.011
94	7.247		0.0000	0.000	141.1	0.016	BB	0.017	136.17	0.017
95	7.301		0.0000	0.000	1743.6	0.199	BB	0.151	192.35	0.024
96	7.467		0.0000	0.000	1085.8	0.124	BB	0.078	232.75	0.029

Total Area = 874737.7, Total Amount = 6510.42, Total Height = 804316.3, Sample Units = PPB

ASX22 0134

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3C6310.49R
5822074 RI ACF DF100ACELEC3 T 093070028A 01216
C:\CPWIN\DATA1\3C6310.49R



Instrument ID: CP06--S9147A Injected On: 11/9/2009 8:12:41 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\3C6310B.49R
5822074 RI ACF DF100ACELEC3 T 093070028A 01216
C:\CPWIN\DATA1\3C6310B.49R



ASX22 8135

Instrument ID: CP06--S9147B Injected On: 11/9/2009 8:12:41 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min; hold 1 Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 30 Dilution Factor: 1000
Analyst: 1151

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.313	417	7.10047	TCX	2.809	1487	10.83376	TCX
7.162	218	22.87846	DCB	6.355	14763	206.1802	DCB

Files:
Area File: C:\CPWIN\DATA1\3C6310.49A
Area File: C:\CPWIN\DATA1\3C6310B.49A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/9/2009 8:20:16 PM
File Reported On: 11/9/2009 at 8:20:26 PM

REPLOTS

ASX22 8136

5822074 RI ACF DF100ACELEC3 T 093070028A 01216
——— C:\CPWIN\DATA1\3C6310.49R



5822074 RI ACF DF100ACELEC3 T 093070028A 01216
——— C:\CPWIN\DATA1\3C6310B.49R



ASX22 0137

Multiple Component Data Summary

Sample Name: 5822075 RI ACF DF20 WHEE1 Sample ID:AC **Batchnumber: 093070028A**
Sample Amount:30 g Total Volume: 200 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on Nov 09, 2009 20:24:40
Instrument S9147A
Result file 3C6310.50R
Calibration file 1C6310
Method file PESTD

%SSR(TCX) 106.0% Conc: 10.84641:
%SSR(DCB) * 155.0% Conc: 15.76014:

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**31.53**	
4.68	4.71	4.74	889.26947	165.121691			2
4.85	4.88	4.91	1717.205811	267.05948			3
5.13	5.15	5.19	1791.854492	320.97592			4
Height summation:			**4398.329773**				
Concentration		CF:	**251.052364**	L:			
Aroclor-1260						**8.61**	
5.89	5.92	5.95	2584.630615	761.141379			4
6.12	6.15	6.18	4885.296387	746.282013			5
6.33	6.36	6.39	2791.394287	871.270635			6
Height summation:			**10261.321289**				
Concentration		CF:	**792.898009**	L:			

Analysis Report (B)

Injected on Nov 09, 2009 20:24:40
Instrument S9147B
Result file 3C6310B.50R
Calibration file 1C6310B
Method file PESTDB

%SSR(TCX) 96.0% Conc: 9.824996
%SSR(DCB) 110.8% Conc: 11.26436

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**27.40**	
4.07	4.10	4.13	3224.644775	104.811007			1
4.18	4.21	4.24	3987.851563	157.930337			2
4.45	4.48	4.51	5472.142578	185.273152			3
Height summation:			**12684.638916**				
Concentration		CF:	**149.338165**	L:			
Aroclor-1260						**26.10**	
4.92	4.95	4.98	13962.87793	629.677489			4
5.34	5.37	5.40	17513.46875	426.834109			5
5.60	5.63	5.66	10765.908203	397.354547			6
Height summation:			**42242.254883**				
Concentration		CF:	**484.622048**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<340	<66			4	40	
Aroclor-1221			<340	<80			3	5	
Aroclor-1232			<340	<106			4	10	
Aroclor-1242			<340	<66			4	30	
Aroclor-1248			<340	<66			4	40	
Aroclor-1254	B	149.338165	<340	66	J	50.81	4	40	1569
Aroclor-1260	B	484.622048	340	66		48.26	4	40	1569

Units: ug/kg

%Difference = High - Low divided by the Average times 100

Reviewed By: Andrea Carey Date: 11.1009
Verified By: [illegible] Date: 11/11/09

ASX22 0138

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822075** RI ACF DF20 WHEE1 Sample ID: AC **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 200 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:24:40
Instrument : CP06--S9147A
Result file : 3C6310.50R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

%SSR(TCX) : 106% Conc.: 10.846412
%SSR(DCB) : *155% Conc.: 15.760148

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					3	68.74	
4.08	4.12	4.14	190.56929	24.81634			3
4.13	4.16	4.19	116.982529	27.189672			4
4.19	4.21	4.25	297.191254	77.260764			5
Height Summation:			604.743073				
Amount Avg CF:			43.088925	Linear:			
Aroclor-1232					3	69.57	
4.08	4.12	4.14	190.56929	57.344253			3
4.13	4.16	4.19	116.982529	62.04324			4
4.19	4.21	4.25	297.191254	179.789022			5
Height Summation:			604.743073				
Amount Avg CF:			99.725505	Linear:			
Aroclor-1242					3	67.48	
4.08	4.12	4.14	190.56929	32.459426			3
4.13	4.16	4.19	116.982529	36.164318			4
4.19	4.21	4.25	297.191254	99.887826			5
Height Summation:			604.743073				
Amount Avg CF:			56.170523	Linear:			
Aroclor-1248					3	56.03	
4.19	4.21	4.25	297.191254	125.64367			2
4.26	4.29	4.32	302.7883	71.184846			3
4.70	4.71	4.76	889.26947	227.236434			6
Height Summation:			1489.249024				
Amount Avg CF:			141.354983	Linear:			
Aroclor-1254					6	93.93	
4.26	4.29	4.32	302.7883	82.478903 /			1
4.68	4.71	4.74	889.26947	165.121691 /			2
4.85	4.88	4.91	1717.205811	267.05948 /			3
5.13	5.15	5.19	1791.854492	320.97592			4
5.35	5.38	5.41	1479.534302	540.878144			5
5.66	5.69	5.72	4436.745117	1176.440438			6
Height Summation:			10617.397492				
Amount Avg CF:			425.492429	Linear:			
Aroclor-1260					6	20.21	
5.25	5.28	5.31	2630.580322	563.618085			1
5.45	5.49	5.51	3517.672607	825.895964			2
5.66	5.69	5.72	4436.745117	1057.160756			3
5.89	5.92	5.95	2584.630615	761.141379 /			4
6.12	6.15	6.18	4885.296387	746.282013 /			5
6.33	6.36	6.39	2791.394287	871.270635 /			6
Height Summation:			20846.319335				
Amount Avg CF:			804.228139	Linear:			
Aroclor-1262					6	28.77	
5.25	5.28	5.31	2630.580322	619.347715			1
5.45	5.49	5.51	3517.672607	907.046797			2
5.73	5.76	5.79	1888.041504	354.451955			3
5.89	5.92	5.95	2584.630615	628.951123			4
6.12	6.15	6.18	4885.296387	579.979674			5
6.32	6.36	6.38	2791.394287	586.773765			6
Height Summation:			18297.615722				
Amount Avg CF:			612.758505	Linear:			

Analysis Report (B)

Injected on : NOV 09, 2009 20:24:40
Instrument : CP06--S9147B
Result file : 3C6310B.50R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

%SSR(TCX) : 96% Conc.: 9.824996
%SSR(DCB) : 110.8% Conc.: 11.264367

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	44.65	
3.07	3.12	3.13	702.343445	54.112749			1
3.40	3.45	3.46	340.712738	19.343434			2
+ 3.56	3.58	3.62	270.31369	8.620788			3
* 3.59	3.61	3.65	728.516479	42.742704			4
* 3.56	3.61	3.62	728.516479	23.233696			3
3.67	3.70	3.73	382.266174	23.465834			5
3.84	3.87	3.90	366.408356	24.265808			6
Height Summation:			3248.763671				
Amount Avg CF:			31.194038	Linear:			
Aroclor-1221					4	72.73	
2.95	2.97	3.01	658.039612	80.96956			1
+ 2.95	3.00	3.01	156.513809	19.258497			1
3.02	3.02	3.08	882.945801	201.931573			2
+ 3.02	3.04	3.08	222.438583	50.872174			2
3.02	3.07	3.08	1196.942017	273.743171			2
3.07	3.12	3.13	702.343445	38.709405			3
Height Summation:			3440.270875				
Amount Avg CF:			148.838427	Linear:			
Aroclor-1232					6	32.00	
3.07	3.12	3.13	702.343445	49.461677			1
3.40	3.45	3.46	340.712738	43.019285			2
+ 3.56	3.58	3.62	270.31369	19.380451			3
* 3.60	3.61	3.66	728.516479	96.355055			4
* 3.56	3.61	3.62	728.516479	52.231828			3
3.67	3.70	3.73	382.266174	55.328727			5
3.84	3.87	3.90	366.408356	59.390284			6
Height Summation:			3248.763671				
Amount Avg CF:			59.297809	Linear:			
Aroclor-1242					6	43.07	
3.07	3.12	3.13	702.343445	69.408385			1
3.40	3.45	3.46	340.712738	25.288558			2
+ 3.56	3.58	3.62	270.31369	11.323581			3
* 3.59	3.61	3.65	728.516479	56.235472			4
* 3.56	3.61	3.62	728.516479	30.517933			3
3.67	3.70	3.73	382.266174	31.215595			5
3.84	3.87	3.90	366.408356	33.025382			6
Height Summation:			3248.763671				
Amount Avg CF:			40.948554	Linear:			
Aroclor-1248					6	147.30	
3.28	3.30	3.34	270.284607	29.203698			1
+ 3.56	3.58	3.62	270.31369	16.979812			2
3.56	3.61	3.62	728.516479	45.761919			2
3.71	3.74	3.77	530.244629	27.166746			3
+ 3.71	3.76	3.77	470.143158	24.087486			3
3.84	3.87	3.90	366.408356	17.481982			4
4.21	4.24	4.27	2050.504639	167.35917			5
+ 4.45	4.45	4.51	1289.889893	124.521717			6
4.45	4.48	4.51	5472.142578	528.262602			6
Height Summation:			9418.101288				
Amount Avg CF:			135.872686	Linear:			

ASX22 S139

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:51:35

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822075** RI ACF DF20 WHEE1 Sample ID: AC **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 200 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:24:40
Instrument : CP06--S9147A
Result file : 3C6310.50R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1268					6	113.50	
5.90	5.92	5.96	2584.630615	868.403929			1
6.32	6.36	6.38	2791.394287	307.424987			2
6.38	6.40	6.44	767.002014	105.662505			3
6.51	6.53	6.57	165.576584	21.268914			4
6.74	6.78	6.80	1123.706055	416.057988			5
6.96	6.99	7.02	192.619583	10.504067			6
Height Summation:			7624.929138				
Amount Avg CF:			288.220398	Linear:			

Analysis Report (B)

Injected on : NOV 09, 2009 20:24:40
Instrument : CP06--S9147B
Result file : 3C6310B.50R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					6	75.49	
+ 4.07	4.08	4.13	464.687103	15.103779			1
4.07	4.10	4.13	3224.644775	104.811007 /			1
4.18	4.21	4.24	3987.851563	157.930337 /			2
+ 4.45	4.45	4.51	1289.889893	43.672467			3
4.45	4.48	4.51	5472.142578	185.273152 /			3
4.62	4.65	4.68	5082.720215	302.507757			4
+ 4.62	4.67	4.68	2542.706543	151.334014			4
4.79	4.82	4.85	9050.78125	734.680097			5
4.92	4.95	4.98	13962.87793	636.120842			6
Height Summation:			40781.018311				
Amount Avg CF:			353.553865	Linear:			
Aroclor-1260					6	23.07	
4.55	4.58	4.61	8486.277344	345.184049			1
+ 4.68	4.69	4.74	1685.973999	64.190241			2
4.68	4.71	4.74	12038.143555	458.329328			2
4.79	4.82	4.85	9050.78125	376.819611			3
4.92	4.95	4.98	13962.87793	629.677489 /			4
+ 5.34	5.35	5.40	1937.338867	47.216364			5
5.34	5.37	5.40	17513.46875	426.834109 /			5
5.60	5.63	5.66	10765.908203	397.354547 /			6
Height Summation:			71817.457032				
Amount Avg CF:			439.033189	Linear:			
Aroclor-1262					6	34.91	
+ 4.68	4.69	4.74	1685.973999	72.156648			1
4.68	4.71	4.74	12038.143555	515.210846			1
4.95	4.99	5.01	6553.525391	200.688509			2
5.15	5.18	5.21	7202.063477	234.235385			3
+ 5.15	5.20	5.21	3058.689453	99.478893			3
+ 5.34	5.35	5.40	1937.338867	36.653767			4
5.34	5.37	5.40	17513.46875	331.348639			4
5.60	5.63	5.66	10765.908203	300.741166			5
5.97	6.00	6.03	5826.541992	304.452822			6
Height Summation:			59899.651368				
Amount Avg CF:			314.446228	Linear:			
Aroclor-1268					6	105.45	
5.15	5.18	5.21	7202.063477	404.426498			1
+ 5.15	5.20	5.21	3058.689453	171.758423			1
5.55	5.58	5.61	2650.787598	43.734229			2
5.60	5.63	5.66	10765.908203	210.894773			3
5.75	5.78	5.81	180.319626	3.427962			4
5.97	6.00	6.03	5826.541992	301.3553			5
+ 6.17	6.18	6.23	122.871239	0.917748			6
6.17	6.20	6.23	534.6203	3.993177			6
Height Summation:			27160.241196				
Amount Avg CF:			161.305323	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			340	66		32.03	4	40	
Aroclor-1221			340	66			3	5	
Aroclor-1232			340	66		** 50.84	4	10	
Aroclor-1242			340	66		31.35	4	30	
Aroclor-1248			340	66		3.96	4	40	
Aroclor-1254			340	66		18.47	4	40	~~ASX22 8148~~

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:51:35

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822075** RI ACF DF20 WHEE1 Sample ID: AC **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 200 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:24:40
Instrument : CP06–S9147A
Result file : 3C6310.50R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Analysis Report (B)

Injected on : NOV 09, 2009 20:24:40
Instrument : CP06–S9147B
Result file : 3C6310B.50R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1260			340	66		** 58.75	4	40	
Aroclor-1262			0	0		** 64.35	4	40	
Aroclor-1268			0	0		** 56.47	4	40	

Units: _____

ASX22 0141

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:51:35

Sample Name: **5822075 RI ACF DF20 ACWHEE1 T 093070028A 01216**
Acquired from CP06--S9147A via port 1 on 11/9/09 08:32:10pm by 1151
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\3C6310.50R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.008		0.0000	0.000	40.1	0.056	BB	0.008	85.35	0.162
2	3.029		0.0000	0.000	326.2	0.457	BB	0.012	455.93	0.863
3	3.055		0.0000	0.000	282.6	0.396	BB	0.012	391.96	0.742
4	3.100		0.0000	0.000	204.3	0.286	BB	0.021	165.70	0.314
5	3.215		0.0000	0.000	142.1	0.199	BB	0.033	71.54	0.135
6	3.313	TCX	325.3924	40.766	3198.2	4.481	BB	0.017	3182.50	6.027
7	3.620		0.0000	0.000	387.7	0.543	BB	0.030	214.93	0.407
8	3.780		0.0000	0.000	458.2	0.642	BB	0.024	317.16	0.601
9	3.842		0.0000	0.000	149.9	0.210	BB	0.026	96.24	0.182
10	3.951		0.0000	0.000	181.6	0.254	BB	0.020	149.66	0.283
11	4.044		0.0000	0.000	89.7	0.126	BB	0.015	102.11	0.193
12	4.075		0.0000	0.000	1241.3	1.739	BB	0.017	1237.08	2.343
13	4.119		0.0000	0.000	165.7	0.232	BB	0.014	190.57	0.361
14	4.162		0.0000	0.000	170.5	0.239	BB	0.024	116.98	0.222
15	4.215		0.0000	0.000	377.3	0.529	BB	0.021	297.19	0.563
16	4.294		0.0000	0.000	428.7	0.601	BB	0.024	302.79	0.573
17	4.327		0.0000	0.000	169.6	0.238	BB	0.026	109.66	0.208
18	4.378		0.0000	0.000	534.0	0.748	BB	0.027	327.42	0.620
19	4.622		0.0000	0.000	322.5	0.452	BV	0.024	223.61	0.423
20	4.709		0.0000	0.000	1636.1	2.292	VV	0.031	889.27	1.684
21	4.762		0.0000	0.000	1138.5	1.595	VV	0.035	534.85	1.013
22	4.831		0.0000	0.000	546.2	0.765	VV	0.025	359.49	0.681
23	4.877		0.0000	0.000	2334.9	3.271	VV	0.023	1717.21	3.252
24	4.915		0.0000	0.000	1748.7	2.450	VB	0.039	739.87	1.401
25	5.034		0.0000	0.000	540.7	0.758	BB	0.020	442.25	0.838
26	5.089		0.0000	0.000	2430.5	3.405	BV	0.031	1294.97	2.452
27	5.154		0.0000	0.000	2517.4	3.527	VB	0.023	1791.85	3.393
28	5.205		0.0000	0.000	1240.4	1.738	BB	0.021	972.69	1.842
29	5.244		0.0000	0.000	303.5	0.425	BB	0.015	346.21	0.656
30	5.285		0.0000	0.000	3326.9	4.661	BB	0.021	2630.58	4.982
31	5.377		0.0000	0.000	1770.1	2.480	BB	0.020	1479.53	2.802
32	5.433		0.0000	0.000	615.2	0.862	BB	0.019	544.16	1.031
33	5.486		0.0000	0.000	5557.7	7.787	BB	0.026	3517.67	6.662
34	5.540		0.0000	0.000	458.6	0.643	BB	0.015	519.25	0.983
35	5.580		0.0000	0.000	3038.8	4.258	BB	0.023	2226.09	4.216
36	5.658		0.0000	0.000	253.0	0.354	BB	0.017	242.41	0.459
37	5.695		0.0000	0.000	5870.4	8.225	BB	0.022	4436.75	8.402
38	5.757		0.0000	0.000	2030.4	2.845	BB	0.018	1888.04	3.576
39	5.800		0.0000	0.000	1468.0	2.057	BB	0.019	1289.55	2.442
40	5.921		0.0000	0.000	4463.4	6.253	BB	0.029	2584.63	4.895
41	5.980		0.0000	0.000	846.0	1.185	BB	0.018	795.10	1.506
42	6.015		0.0000	0.000	792.3	1.110	BB	0.019	679.83	1.287
43	6.057		0.0000	0.000	63.8	0.089	BB	0.012	86.31	0.163
44	6.098		0.0000	0.000	1213.3	1.700	BB	0.026	776.35	1.470
45	6.154		0.0000	0.000	6174.3	8.651	BB	0.021	4885.30	9.252
46	6.267		0.0000	0.000	421.0	0.590	BB	0.024	289.36	0.548
47	6.359		0.0000	0.000	3328.1	4.663	BB	0.020	2791.39	5.286
48	6.405		0.0000	0.000	1374.4	1.926	BB	0.030	767.00	1.453
49	6.533		0.0000	0.000	314.0	0.440	BB	0.032	165.58	0.314
50	6.601		0.0000	0.000	57.9	0.081	BB	0.014	70.70	0.134
51	6.628		0.0000	0.000	632.2	0.886	BB	0.020	515.77	0.977
52	6.776		0.0000	0.000	1438.5	2.015	BB	0.021	1123.71	2.128
53	6.832		0.0000	0.000	197.2	0.276	BB	0.026	127.48	0.241
54	6.991		0.0000	0.000	196.6	0.275	BB	0.017	192.62	0.365
55	7.031		0.0000	0.000	423.3	0.593	BB	0.023	301.12	0.570
56	7.162	DCB	472.8044	59.234	1741.8	2.440	BB	0.039	750.03	1.420

Total Area = 71374.6, Total Amount = 798.197, Total Height = 52803.36, Sample Units = PPB

ASX22 8142

Sample Name: **5822075 RI ACF DF20 ACWHEE1 T 093070028A 01216**
Acquired from CP06--S9147B via port 2 on 11/9/09 08:32:10pm by 1151
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\3C6310B.50R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.508		0.0000	0.000	225.0	0.106	BV	0.012	318.28	0.160
2	2.529		0.0000	0.000	241.1	0.113	VV	0.013	314.70	0.158
3	2.553		0.0000	0.000	1101.9	0.517	VB	0.016	1175.94	0.590
4	2.604		0.0000	0.000	872.4	0.409	BV	0.015	999.77	0.502
5	2.631		0.0000	0.000	235.9	0.111	VB	0.018	212.56	0.107
6	2.731		0.0000	0.000	889.0	0.417	BB	0.021	696.01	0.349
7	2.761		0.0000	0.000	107.3	0.050	BB	0.010	180.91	0.091
8	2.810	TCX	294.7499	46.587	5230.9	2.455	BB	0.013	6743.00	3.384
9	2.857		0.0000	0.000	540.1	0.254	BB	0.013	704.19	0.353
10	2.900		0.0000	0.000	1234.5	0.579	BB	0.016	1257.83	0.631
11	2.925		0.0000	0.000	1738.1	0.816	BB	0.015	1995.65	1.002
12	2.965		0.0000	0.000	320.2	0.150	BB	0.008	658.04	0.330
13	2.997		0.0000	0.000	99.4	0.047	BB	0.011	156.51	0.079
14	3.018		0.0000	0.000	505.4	0.237	BB	0.010	882.95	0.443
15	3.039		0.0000	0.000	193.9	0.091	BB	0.015	222.44	0.112
16	3.066		0.0000	0.000	676.0	0.317	BB	0.009	1196.94	0.601
17	3.121		0.0000	0.000	719.0	0.337	BB	0.017	702.34	0.353
18	3.166		0.0000	0.000	152.7	0.072	BB	0.014	183.42	0.092
19	3.200		0.0000	0.000	128.5	0.060	BB	0.012	180.59	0.091
20	3.261		0.0000	0.000	491.8	0.231	BV	0.039	211.00	0.106
21	3.301		0.0000	0.000	417.7	0.196	VV	0.026	270.28	0.136
22	3.341		0.0000	0.000	431.9	0.203	VB	0.018	391.76	0.197
23	3.449		0.0000	0.000	321.8	0.151	BV	0.016	340.71	0.171
24	3.487		0.0000	0.000	305.2	0.143	VV	0.013	393.67	0.198
25	3.510		0.0000	0.000	504.6	0.237	VV	0.014	602.39	0.302
26	3.531		0.0000	0.000	6353.3	2.982	VV	0.014	7531.48	3.780
27	3.584		0.0000	0.000	251.5	0.118	VV	0.016	270.31	0.136
28	3.613		0.0000	0.000	859.3	0.403	VB	0.020	728.52	0.366
29	3.657		0.0000	0.000	1119.2	0.525	BV	0.027	690.68	0.347
30	3.703		0.0000	0.000	352.1	0.165	VV	0.015	382.27	0.192
31	3.738		0.0000	0.000	518.8	0.243	VV	0.016	530.24	0.266
32	3.762		0.0000	0.000	511.0	0.240	VB	0.018	470.14	0.236
33	3.813		0.0000	0.000	893.4	0.419	BB	0.014	1082.94	0.544
34	3.871		0.0000	0.000	364.4	0.171	BB	0.017	366.41	0.184
35	3.932		0.0000	0.000	909.1	0.427	BV	0.030	506.01	0.254
36	4.007		0.0000	0.000	827.5	0.388	VV	0.056	245.61	0.123
37	4.081		0.0000	0.000	567.5	0.266	VV	0.020	464.69	0.233
38	4.105		0.0000	0.000	4070.5	1.910	VV	0.021	3224.64	1.618
39	4.171		0.0000	0.000	1099.8	0.516	VV	0.025	718.91	0.361
40	4.207		0.0000	0.000	3680.9	1.728	VV	0.015	3987.85	2.002
41	4.241		0.0000	0.000	2879.4	1.351	VV	0.023	2050.50	1.029
42	4.307		0.0000	0.000	340.4	0.160	VB	0.025	228.11	0.114
43	4.359		0.0000	0.000	1123.9	0.527	BV	0.014	1316.66	0.661
44	4.389		0.0000	0.000	1465.3	0.688	VV	0.014	1714.42	0.860
45	4.418		0.0000	0.000	2199.0	1.032	VB	0.014	2559.37	1.285
46	4.454		0.0000	0.000	1584.8	0.744	BV	0.020	1289.89	0.647
47	4.481		0.0000	0.000	7034.3	3.301	VB	0.021	5472.14	2.747
48	4.522		0.0000	0.000	1289.0	0.605	BB	0.016	1382.99	0.694
49	4.576		0.0000	0.000	8059.3	3.782	BV	0.016	8486.28	4.259
50	4.612		0.0000	0.000	596.0	0.280	VV	0.016	605.33	0.304
51	4.646		0.0000	0.000	4837.1	2.270	VV	0.016	5082.72	2.551
52	4.671		0.0000	0.000	2447.9	1.149	VV	0.016	2542.71	1.276
53	4.692		0.0000	0.000	1460.0	0.685	VV	0.014	1685.97	0.846
54	4.715		0.0000	0.000	12033.4	5.648	VV	0.017	12038.14	6.042
55	4.773		0.0000	0.000	201.8	0.095	VB	0.017	197.33	0.099
56	4.822		0.0000	0.000	9686.7	4.546	BB	0.018	9050.78	4.543
57	4.873		0.0000	0.000	4119.6	1.933	BB	0.017	3953.60	1.984
58	4.916		0.0000	0.000	2706.6	1.270	BV	0.017	2728.86	1.370
59	4.954		0.0000	0.000	15779.4	7.406	VV	0.019	13962.88	7.008
60	4.986		0.0000	0.000	6718.8	3.153	VB	0.017	6553.53	3.289
61	5.026		0.0000	0.000	3147.7	1.477	BB	0.014	3695.50	1.855
62	5.129		0.0000	0.000	2174.3	1.020	BB	0.022	1650.49	0.828

ASX22 0143

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.180		0.0000	0.000	7037.5	3.303	BV	0.016	7202.06	3.615
64	5.203		0.0000	0.000	2856.9	1.341	VV	0.016	3058.69	1.535
65	5.253		0.0000	0.000	4303.8	2.020	VV	0.018	3973.05	1.994
66	5.287		0.0000	0.000	3414.1	1.602	VV	0.024	2324.00	1.166
67	5.315		0.0000	0.000	1542.2	0.724	VV	0.016	1560.37	0.783
68	5.347		0.0000	0.000	1438.9	0.675	VV	0.012	1937.34	0.972
69	5.369		0.0000	0.000	18750.5	8.800	VB	0.018	17513.47	8.790
70	5.445		0.0000	0.000	489.0	0.230	BB	0.015	532.72	0.267
71	5.541		0.0000	0.000	2619.9	1.230	BB	0.018	2369.30	1.189
72	5.582		0.0000	0.000	2608.0	1.224	BB	0.016	2650.79	1.330
73	5.626		0.0000	0.000	12100.4	5.679	BB	0.019	10765.91	5.404
74	5.781		0.0000	0.000	201.1	0.094	BB	0.019	180.32	0.091
75	5.822		0.0000	0.000	338.2	0.159	BB	0.016	347.68	0.175
76	5.898		0.0000	0.000	1633.4	0.767	BB	0.015	1770.30	0.889
77	6.000		0.0000	0.000	13408.0	6.293	BB	0.038	5826.54	2.924
78	6.081		0.0000	0.000	792.1	0.372	BB	0.023	580.07	0.291
79	6.183		0.0000	0.000	117.5	0.055	BB	0.016	122.87	0.062
80	6.200		0.0000	0.000	429.0	0.201	BB	0.013	534.62	0.268
81	6.359	DCB	337.9310	53.413	4727.4	2.219	BB	0.020	4032.72	2.024
82	6.604		0.0000	0.000	99.4	0.047	BB	0.058	28.34	0.014
83	6.749		0.0000	0.000	120.7	0.057	BB	0.017	116.01	0.058
84	6.783		0.0000	0.000	391.6	0.184	BB	0.022	301.96	0.152
85	6.824		0.0000	0.000	231.9	0.109	BB	0.017	232.26	0.117
86	6.901		0.0000	0.000	200.8	0.094	BB	0.017	192.47	0.097
87	6.996		0.0000	0.000	120.0	0.056	BB	0.019	103.26	0.052
88	7.044		0.0000	0.000	752.6	0.353	BB	0.032	389.66	0.196
89	7.296		0.0000	0.000	399.9	0.188	BB	0.045	149.22	0.075

Total Area = 213072.0, Total Amount = 632.681, Total Height = 199236.8, Sample Units = PPB

ASX22 0144

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3C6310.50R

5822075 RI ACF DF20 ACWHEE1 T 093070028A 01216

C:\CPWIN\DATA1\3C6310.50R



Instrument ID: CP06--S9147A Injected On: 11/9/2009 8:24:40 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\3C6310B.50R

5822075 RI ACF DF20 ACWHEE1 T 093070028A 01216

C:\CPWIN\DATA1\3C6310B.50R



Instrument ID: CP06--S9147B Injected On: 11/9/2009 8:24:40 PM Column ID: RTX CLPII

ASX22 0145

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Valid Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 30 Dilution Factor: 200
Analyst: 1151

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.313	3182	10.84641	TCX	2.81	6743	9.825	TCX
7.162	750	15.76015	DCB	6.359	4033	11.26437	DCB

Files:
Area File: C:\CPWIN\DATA1\3C6310.50A
Area File: C:\CPWIN\DATA1\3C6310B.50A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/9/2009 8:32:16 PM
File Reported On: 11/9/2009 at 8:32:25 PM

REPLOTS

ASX22 0146

5822075 RI ACF DF20 ACWHEE1 T 093070028A 01216
——— C:\CPWIN\DATA1\3C6310.50R



5822075 RI ACF DF20 ACWHEE1 T 093070028A 01216
——— C:\CPWIN\DATA1\3C6310B.50R



ASX22 0147

Multiple Component Data Summary

Sample Name: 5822076 RI ACF WHEE2 Sample ID:AB **Batchnumber: 093070028A**
Sample Amount:30 g Total Volume: 25 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on Nov 09, 2009 20:36:40
Instrument S9147A
Result file 3C6310.51R
Calibration file 1C6310
Method file PESTD

%SSR(TCX) 76.3% Conc: 7.808618
%SSR(DCB) * 0.0% Conc: 0

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**25.75**	
4.68	4.72	4.74	9404.490234	218.281041			2
4.85	4.89	4.91	16325.822266	317.373548			3
5.13	5.14	5.19	8932.963867	200.020866			4
Height summation:			**34663.276367**				
Concentration		CF:	**245.225152**	L:			
Aroclor-1260						**3.15**	
5.89	5.94	5.95	13509.357422	497.292093			4
6.12	6.17	6.18	25014.779297	477.659858			5
6.33	6.38	6.39	13031.361328	508.430613			6
Height summation:			**51555.498047**				
Concentration		CF:	**494.460855**	L:			

Analysis Report (B)

Injected on Nov 09, 2009 20:36:40
Instrument S9147B
Result file 3C6310B.51R
Calibration file 1C6310B
Method file PESTDB

%SSR(TCX) 54.4% Conc: 5.569533
%SSR(DCB) * 0.0% Conc: 0

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**5.52**	
4.07	4.12	4.13	43084.476563	175.047474			1
4.18	4.22	4.24	34316.238281	169.877658			2
4.45	4.50	4.51	37118.152344	157.091052			3
Height summation:			**114518.867188**				
Concentration		CF:	**167.338728**	L:			
Aroclor-1260						**37.15**	
4.79	4.84	4.85	38288.480469	199.262499			3
4.92	4.98	4.98	59040.386719	332.814648			4
5.60	5.63	5.66	36784.894531	169.709847			6
Height summation:			**134113.761719**				
Concentration		CF:	**233.928998**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<42.5	<8.25			4	40	
Aroclor-1221			<42.5	<10			3	5	
Aroclor-1232			<42.5	<13.25			4	10	
Aroclor-1242			<42.5	<8.25			4	30	
Aroclor-1248			<42.5	<8.25			4	40	
Aroclor-1254	A	245.225152	42.5	8.25		37.76	4	40	
Aroclor-1260	B	233.928998	42.5	8.25		71.54	4	40	1569

Units: ug/kg

%Difference = High - Low divided by the Average times 100

Reviewed By: Andrea Carey Date: 11-10-09
Verified By: [illegible] Date: 11/11/09

ASX22 8148

Lancaster Laboratories: Multiple Component Peak Data Report

Sample Name: **5822076** RI ACF WHEE2 Sample ID: AB **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 25 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:36:40
Instrument : CP06--S9147A
Result file : 3C6310.51R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

%SSR(TCX) : 76.3% Conc.: 7.808618
%SSR(DCB) : *0% Conc.: 0

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	172.94	
3.54	3.57	3.60	4168.10498	107.834044			1
E 3.88	3.91	3.94	38227.886719	836.086617			2
+ 4.08	4.09	4.14	1444.696655	23.516435			3
4.08	4.12	4.14	1777.008667	28.925733			3
4.13	4.15	4.19	1221.828125	35.497935			4
+ 4.19	4.19	4.25	535.026062	17.38633			5
4.19	4.23	4.25	840.798157	27.322771			5
4.40	4.44	4.46	1738.987549	75.979463			6
Height Summation:			47974.614197				
Amount Avg CF:			185.274427	Linear:			
Aroclor-1221					3	77.48	
+ 3.43	3.45	3.49	581.364929	25.364962			1
3.43	3.47	3.49	737.792542	32.189901			1
3.49	3.52	3.55	333.441528	15.032077			2
3.53	3.57	3.59	4168.10498	77.290183			3
Height Summation:			5239.339050				
Amount Avg CF:			41.504054	Linear:			
Aroclor-1232					6	184.18	
3.54	3.57	3.60	4168.10498	99.22642			1
E 3.88	3.91	3.94	38227.886719	1878.889547			2
+ 4.08	4.09	4.14	1444.696655	54.340505			3
4.08	4.12	4.14	1777.008667	66.840016			3
4.13	4.15	4.19	1221.828125	81.001599			4
+ 4.19	4.19	4.25	535.026062	40.458716			5
4.19	4.23	4.25	840.798157	63.581228			5
E 4.40	4.44	4.46	1738.987549	182.398526			6
Height Summation:			47974.614197				
Amount Avg CF:			395.322889	Linear:			
Aroclor-1242					6	173.49	
3.54	3.57	3.60	4168.10498	135.813131			1
E 3.88	3.91	3.94	38227.886719	1078.239034			2
+ 4.08	4.09	4.14	1444.696655	30.759169			3
4.08	4.12	4.14	1777.008667	37.834455			3
4.13	4.15	4.19	1221.828125	47.214936			4
+ 4.19	4.19	4.25	535.026062	22.478198			5
4.19	4.23	4.25	840.798157	35.324685			5
4.40	4.44	4.46	1738.987549	95.182679			6
Height Summation:			47974.614197				
Amount Avg CF:			238.268153	Linear:			
Aroclor-1248					6	185.15	
E 3.88	3.91	3.94	38227.886719	2198.399831			1
+ 4.19	4.19	4.25	535.026062	28.274149			2
4.19	4.23	4.25	840.798157	44.433073			2
4.26	4.30	4.32	2163.501221	63.579281			3
4.40	4.44	4.46	1738.987549	51.70314			4
4.50	4.53	4.56	2926.458984	115.171754			5
4.70	4.72	4.76	9404.490234	300.393033			6
Height Summation:			55302.122864				
Amount Avg CF:			462.280019	Linear:			

Analysis Report (B)

Injected on : NOV 09, 2009 20:36:40
Instrument : CP06--S9147B
Result file : 3C6310B.51R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

%SSR(TCX) : 54.4% Conc.: 5.569533
%SSR(DCB) : *0% Conc.: 0

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	164.18	
+ 3.07	3.07	3.13	2756.095703	26.543267			1
3.07	3.12	3.13	13427.104492	129.313078			1
+ 3.40	3.42	3.46	2340.677734	16.611041			2
3.40	3.44	3.46	88166.75	625.691217			2
3.56	3.57	3.62	3943.615234	15.72112			3
* 3.59	3.60	3.65	3248.962646	23.827438			4
+* 3.56	3.60	3.62	3248.962646	12.951905			3
3.67	3.71	3.73	5740.854004	44.051087			5
3.84	3.88	3.90	4128.541504	34.177166			6
Height Summation:			118655.827880				
Amount Avg CF:			145.463518	Linear:			
Aroclor-1221					2	11.32	
+ 3.02	3.02	3.08	2097.980957	59.976585			2
+* 3.07	3.07	3.13	2756.095703	18.987652			3
* 3.02	3.07	3.08	2756.095703	78.790615			2
3.07	3.12	3.13	13427.104492	92.503751			3
Height Summation:			16183.200195				
Amount Avg CF:			85.647183	Linear:			
Aroclor-1232					6	180.28	
+ 3.07	3.07	3.13	2756.095703	24.261833			1
3.07	3.12	3.13	13427.104492	118.198423			1
+ 3.40	3.42	3.46	2340.677734	36.942515			2
E 3.40	3.44	3.46	88166.75	1391.520676			2
3.56	3.57	3.62	3943.615234	35.342755			3
* 3.60	3.60	3.66	3248.962646	53.714292			4
+* 3.56	3.60	3.62	3248.962646	29.117265			3
3.67	3.71	3.73	5740.854004	103.865502			5
3.84	3.88	3.90	4128.541504	83.648219			6
Height Summation:			118655.827880				
Amount Avg CF:			297.714978	Linear:			
Aroclor-1242					6	164.06	
+ 3.07	3.07	3.13	2756.095703	34.046048			1
3.07	3.12	3.13	13427.104492	165.865013			1
+ 3.40	3.42	3.46	2340.677734	21.716375			2
E 3.40	3.44	3.46	88166.75	817.994786			2
3.56	3.57	3.62	3943.615234	20.650011			3
* 3.59	3.60	3.65	3248.962646	31.349145			4
+* 3.56	3.60	3.62	3248.962646	17.012591			3
3.67	3.71	3.73	5740.854004	58.599277			5
3.84	3.88	3.90	4128.541504	46.514585			6
Height Summation:			118655.827880				
Amount Avg CF:			190.162136	Linear:			

ASX22 0149

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:52:42

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822076** RI ACF WHEE2 Sample ID: AB **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 25 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:36:40
Instrument : CP06--S9147A
Result file : 3C6310.51R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					6	77.46	
4.26	4.30	4.32	2163.501221	73.666654			1
4.68	4.72	4.74	9404.490234	218.281041			2
4.85	4.89	4.91	16325.822266	317.373548			3
+ 5.13	5.14	5.19	8932.963867	200.020866			4
5.13	5.17	5.19	19941.650391	446.519905			4
+ 5.35	5.35	5.41	1253.914429	57.299695			5
5.35	5.40	5.41	16428.373047	750.721689			5
+ 5.66	5.68	5.72	855.414978	28.352563			6
E 5.66	5.71	5.72	32579.462891	1079.839967			6
Height Summation:			96843.300050				
Amount Avg CF:			481.067134	Linear:			
Aroclor-1260					6	32.42	
+ 5.25	5.26	5.31	3291.110596	88.142597			1
5.25	5.30	5.31	18694.90625	500.687394			1
5.45	5.51	5.51	25265.140625	741.483788			2
+ 5.66	5.68	5.72	855.414978	25.477888			3
E 5.66	5.71	5.72	32579.462891	970.354638			3
5.89	5.94	5.95	13509.357422	497.292093			4
6.12	6.17	6.18	25014.779297	477.659858			5
6.33	6.38	6.39	13031.361328	508.430613			6
Height Summation:			128095.007813				
Amount Avg CF:			615.984731	Linear:			
Aroclor-1262					6	44.25	
+ 5.25	5.26	5.31	3291.110596	96.857992			1
5.25	5.30	5.31	18694.90625	550.194541			1
E 5.45	5.51	5.51	25265.140625	814.340455			2
5.73	5.78	5.79	10488.611328	246.135266			3
5.89	5.94	5.95	13509.357422	410.925524			4
6.12	6.17	6.18	25014.779297	371.217588			5
6.32	6.38	6.38	13031.361328	342.412257			6
Height Summation:			106004.156250				
Amount Avg CF:			455.870939	Linear:			
Aroclor-1268					5	112.36	
5.90	5.94	5.96	13509.357422	567.372132			1
6.32	6.38	6.38	13031.361328	179.398075			2
6.38	6.42	6.44	3427.201904	59.016589			3
+ 6.74	6.74	6.80	64.472473	2.983908			5
6.74	6.80	6.80	3795.679688	175.671258			5
6.96	7.02	7.02	224.98999	1.533664			6
Height Summation:			33988.590332				
Amount Avg CF:			196.598344	Linear:			

Analysis Report (B)

Injected on : NOV 09, 2009 20:36:40
Instrument : CP06--S9147B
Result file : 3C6310B.51R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1248					6	117.86	
3.28	3.31	3.34	5095.561523	68.820623			1
3.56	3.57	3.62	3943.615234	30.96488			2
+ 3.56	3.60	3.62	3248.962646	25.510536			2
3.71	3.71	3.77	5740.854004	36.766125			3
+ 3.71	3.75	3.77	5700.927246	36.510423			3
3.84	3.88	3.90	4128.541504	24.62249			4
4.21	4.22	4.27	34316.238281	350.105107			5
+ 4.21	4.26	4.27	20372.628906	207.847998			5
+ 4.45	4.47	4.51	14524.655273	175.270291			6
4.45	4.50	4.51	37118.152344	447.908005			6
Height Summation:			90342.962890				
Amount Avg CF:			159.864538	Linear:			
Aroclor-1254					6	39.11	
+ 4.07	4.10	4.13	8402.566406	34.1387			1
4.07	4.12	4.13	43084.476563	175.047474			1
+ 4.18	4.19	4.24	5058.486328	25.041317			2
4.18	4.22	4.24	34316.238281	169.877658			2
+ 4.45	4.47	4.51	14524.655273	61.471092			3
4.45	4.50	4.51	37118.152344	157.091052			3
+ 4.62	4.63	4.68	2607.360352	19.397751			4
4.62	4.67	4.68	39360.363281	292.825857			4
4.79	4.84	4.85	38288.480469	388.499398			5
+ 4.92	4.94	4.98	6990.227051	39.807599			6
4.92	4.98	4.98	59040.386719	336.220268			6
Height Summation:			251208.097657				
Amount Avg CF:			253.260285	Linear:			
Aroclor-1260					6	70.26	
4.55	4.59	4.61	49514.972656	251.75613			1
4.68	4.70	4.74	11622.043945	55.310891			2
4.79	4.84	4.85	38288.480469	199.262499			3
+ 4.92	4.94	4.98	6990.227051	39.404382			4
4.92	4.98	4.98	59040.386719	332.814648			4
5.34	5.35	5.40	4447.907227	13.550418			5
5.60	5.63	5.66	36784.894531	169.709847			6
Height Summation:			199698.685547				
Amount Avg CF:			170.400739	Linear:			
Aroclor-1262					5	74.64	
4.68	4.70	4.74	11622.043945	62.175317			1
4.95	4.98	5.01	59040.386719	225.999108			2
+ 5.15	5.16	5.21	5990.119141	24.352359			3
5.15	5.21	5.21	27696.335938	112.597277			3
5.34	5.35	5.40	4447.907227	10.519104			4
5.60	5.63	5.66	36784.894531	128.44634			5
Height Summation:			139591.568360				
Amount Avg CF:			107.947429	Linear:			
Aroclor-1268					4	109.63	
5.15	5.16	5.21	5990.119141	42.046334			1
5.60	5.63	5.66	36784.894531	90.073009			3
+ 5.60	5.66	5.66	19307.017578	47.275959			3
5.75	5.75	5.81	904.843811	2.150189			4
6.17	6.21	6.23	11551.21582	10.78477			6
Height Summation:			55231.073303				
Amount Avg CF:			36.263576	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			42.5	8.25	E	24.07	4	40	ASX22 0150

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:52:42

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822076** RI ACF WHEE2 Sample ID: AB **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 25 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on	: NOV 09, 2009 20:36:40
Instrument	: CP06–S9147A
Result file	: 3C6310.51R
Calibration file	: 1C6310.CAL
Method file	: PESTD.MET

Analysis Report (B)

Injected on	: NOV 09, 2009 20:36:40
Instrument	: CP06–S9147B
Result file	: 3C6310B.51R
Calibration file	: 1C6310B.CAL
Method file	: PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1221			42.5	8.25		** 69.43	3	5	
Aroclor-1232			42.5	8.25	E	28.17	4	10	
Aroclor-1242			42.5	8.25	E	22.46	4	30	
Aroclor-1248			42.5	8.25	E	** 97.22	4	40	
Aroclor-1254			42.5	8.25	E	** 62.05	4	40	
Aroclor-1260			42.5	8.25	E	** 113.32	4	40	
Aroclor-1262			0	0	E	** 123.42	4	40	
Aroclor-1268			0	0		** 137.71	4	40	

Units: _____

ASX22 8151

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:52:42

Sample Name: **5822076 RI ACF ABWHEE2 T 093070028A 01216**
Acquired from CP06--S9147A via port 1 on 11/9/09 08:44:10pm by 1151
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\3C6310.51R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.033		0.0000	0.000	7891.4	1.220	BB	0.018	7113.14	1.455
2	3.090		0.0000	0.000	1283.0	0.198	BB	0.019	1124.69	0.230
3	3.118		0.0000	0.000	259.1	0.040	BB	0.009	462.68	0.095
4	3.149		0.0000	0.000	1300.0	0.201	BB	0.023	929.27	0.190
5	3.218		0.0000	0.000	411.0	0.064	BB	0.016	435.74	0.089
6	3.250		0.0000	0.000	142.8	0.022	BB	0.018	135.70	0.028
7	3.278		0.0000	0.000	30.6	0.005	BB	0.006	83.77	0.017
8	3.316 TCX		234.2585	100.000	16772.0	2.593	BB	0.015	18329.32	3.748
9	3.367		0.0000	0.000	8766.0	1.355	BB	0.017	8761.03	1.792
10	3.449		0.0000	0.000	478.4	0.074	BB	0.014	581.36	0.119
11	3.467		0.0000	0.000	669.7	0.104	BB	0.015	737.79	0.151
12	3.518		0.0000	0.000	306.3	0.047	BB	0.015	333.44	0.068
13	3.567		0.0000	0.000	4750.5	0.734	BB	0.019	4168.10	0.852
14	3.624		0.0000	0.000	1179.3	0.182	BB	0.015	1331.51	0.272
15	3.641		0.0000	0.000	3673.5	0.568	BB	0.012	4941.89	1.011
16	3.690		0.0000	0.000	42.5	0.007	BB	0.009	78.10	0.016
17	3.721		0.0000	0.000	176.6	0.027	BB	0.014	212.75	0.044
18	3.742		0.0000	0.000	173.9	0.027	BB	0.014	214.37	0.044
19	3.783		0.0000	0.000	2672.3	0.413	BB	0.015	3017.69	0.617
20	3.799		0.0000	0.000	2304.9	0.356	BB	0.010	4039.14	0.826
21	3.864		0.0000	0.000	500.4	0.077	BB	0.017	484.98	0.099
22	3.911		0.0000	0.000	37247.2	5.759	BB	0.016	38227.89	7.818
23	3.953		0.0000	0.000	1783.7	0.276	BB	0.024	1234.48	0.252
24	4.010		0.0000	0.000	274.8	0.042	BB	0.014	327.70	0.067
25	4.050		0.0000	0.000	1391.8	0.215	BB	0.035	665.83	0.136
26	4.090		0.0000	0.000	1566.4	0.242	BB	0.018	1444.70	0.295
27	4.121		0.0000	0.000	1452.3	0.225	BB	0.014	1777.01	0.363
28	4.150		0.0000	0.000	1611.2	0.249	BB	0.022	1221.83	0.250
29	4.192		0.0000	0.000	471.0	0.073	BB	0.015	535.03	0.109
30	4.228		0.0000	0.000	969.4	0.150	BB	0.019	840.80	0.172
31	4.257		0.0000	0.000	1447.3	0.224	BB	0.013	1812.63	0.371
32	4.299		0.0000	0.000	2510.8	0.388	BB	0.019	2163.50	0.442
33	4.330		0.0000	0.000	472.1	0.073	BB	0.015	539.83	0.110
34	4.356		0.0000	0.000	113.8	0.018	BB	0.008	250.93	0.051
35	4.385		0.0000	0.000	2340.2	0.362	BB	0.019	2047.38	0.419
36	4.442		0.0000	0.000	1860.6	0.288	BB	0.018	1738.99	0.356
37	4.535		0.0000	0.000	6766.7	1.046	BB	0.039	2926.46	0.598
38	4.587		0.0000	0.000	165.1	0.026	BB	0.013	211.78	0.043
39	4.626		0.0000	0.000	7315.7	1.131	BV	0.019	6360.80	1.301
40	4.719		0.0000	0.000	21039.0	3.253	VV	0.037	9404.49	1.923
41	4.770		0.0000	0.000	14452.7	2.234	VV	0.034	7054.37	1.443
42	4.842		0.0000	0.000	2397.2	0.371	VB	0.019	2158.66	0.441
43	4.890		0.0000	0.000	24120.1	3.729	BV	0.025	16325.82	3.339
44	4.926		0.0000	0.000	11808.3	1.826	VV	0.027	7241.58	1.481
45	4.978		0.0000	0.000	3852.5	0.596	VB	0.023	2812.40	0.575
46	5.049		0.0000	0.000	4890.6	0.756	BV	0.022	3689.55	0.755
47	5.102		0.0000	0.000	31645.9	4.893	VV	0.030	17459.08	3.570
48	5.136		0.0000	0.000	11021.6	1.704	VV	0.021	8932.96	1.827
49	5.169		0.0000	0.000	25244.8	3.903	VB	0.021	19941.65	4.078
50	5.220		0.0000	0.000	6915.2	1.069	BB	0.022	5178.87	1.059
51	5.257		0.0000	0.000	3048.3	0.471	BB	0.015	3291.11	0.673
52	5.300		0.0000	0.000	22276.5	3.444	BB	0.020	18694.91	3.823
53	5.353		0.0000	0.000	1391.5	0.215	BB	0.018	1253.91	0.256
54	5.395		0.0000	0.000	22386.3	3.461	BB	0.023	16428.37	3.360
55	5.453		0.0000	0.000	3003.9	0.464	BB	0.018	2757.83	0.564
56	5.507		0.0000	0.000	39465.5	6.102	BB	0.026	25265.14	5.167
57	5.556		0.0000	0.000	3483.6	0.539	BB	0.015	3755.79	0.768
58	5.598		0.0000	0.000	27686.0	4.280	BB	0.022	21170.96	4.330
59	5.675		0.0000	0.000	973.0	0.150	BB	0.019	855.41	0.175
60	5.713		0.0000	0.000	42762.1	6.611	BB	0.022	32579.46	6.663
61	5.777		0.0000	0.000	11428.0	1.767	BB	0.018	10488.61	2.145
62	5.820		0.0000	0.000	7412.5	1.146	BB	0.020	6230.54	1.274

ASX22 0152

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.938		0.0000	0.000	25983.5	4.017	BB	0.032	13509.36	2.763
64	5.997		0.0000	0.000	6293.6	0.973	BB	0.019	5604.44	1.146
65	6.035		0.0000	0.000	3068.6	0.474	BB	0.020	2527.16	0.517
66	6.076		0.0000	0.000	212.8	0.033	BB	0.016	228.51	0.047
67	6.120		0.0000	0.000	7100.1	1.098	BB	0.022	5354.62	1.095
68	6.175		0.0000	0.000	32049.8	4.955	BB	0.021	25014.78	5.116
69	6.288		0.0000	0.000	27686.0	4.280	BB	0.024	19182.12	3.923
70	6.378		0.0000	0.000	15799.0	2.443	BB	0.020	13031.36	2.665
71	6.421		0.0000	0.000	6532.0	1.010	BB	0.032	3427.20	0.701
72	6.646		0.0000	0.000	2342.2	0.362	BB	0.023	1707.12	0.349
73	6.714		0.0000	0.000	146.2	0.023	BB	0.017	147.12	0.030
74	6.744		0.0000	0.000	61.8	0.010	BB	0.016	64.47	0.013
75	6.798		0.0000	0.000	4937.6	0.763	BB	0.022	3795.68	0.776
76	6.858		0.0000	0.000	7146.8	1.105	BB	0.024	4948.39	1.012
77	6.922		0.0000	0.000	612.8	0.095	BB	0.021	483.70	0.099
78	7.015		0.0000	0.000	280.3	0.043	BB	0.021	224.99	0.046
79	7.056		0.0000	0.000	34702.0	5.365	BB	0.026	22408.88	4.583
80	7.182		0.0000	0.000	5000.1	0.773	BB	0.037	2256.31	0.461
81	7.444		0.0000	0.000	611.3	0.095	BB	0.040	255.45	0.052

Total Area = 646815.7, Total Amount = 234.259, Total Height = 488987.2, Sample Units = PPB

ASXZZ 8153

Sample Name: **5822076 RI ACF ABWHEE2 T 093070028A 01216**
Acquired from CP06--S9147B via port 2 on 11/9/09 08:44:10pm by 1151
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\3C6310B.51R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.521		0.0000	0.000	1930.8	0.133	BV	0.028	1129.82	0.100
2	2.558		0.0000	0.000	16960.3	1.164	VB	0.021	13368.67	1.180
3	2.608		0.0000	0.000	2334.3	0.160	BV	0.017	2280.56	0.201
4	2.635		0.0000	0.000	921.2	0.063	VV	0.025	613.70	0.054
5	2.670		0.0000	0.000	1249.1	0.086	VV	0.026	804.14	0.071
6	2.726		0.0000	0.000	1106.4	0.076	VB	0.025	742.37	0.066
7	2.768		0.0000	0.000	1190.3	0.082	BB	0.020	982.38	0.087
8	2.815 TCX		167.0860	100.000	29469.4	2.023	BB	0.016	30579.45	2.698
9	2.862		0.0000	0.000	10492.7	0.720	BB	0.013	13069.62	1.153
10	2.910		0.0000	0.000	520.6	0.036	BB	0.017	522.64	0.046
11	2.923		0.0000	0.000	3245.9	0.223	BB	0.017	3093.57	0.273
12	3.021		0.0000	0.000	2109.2	0.145	BB	0.017	2097.98	0.185
13	3.072		0.0000	0.000	3132.2	0.215	BB	0.019	2756.10	0.243
14	3.123		0.0000	0.000	13432.0	0.922	BB	0.017	13427.10	1.185
15	3.145		0.0000	0.000	4050.6	0.278	BB	0.009	7629.40	0.673
16	3.223		0.0000	0.000	1125.4	0.077	BB	0.015	1247.70	0.110
17	3.269		0.0000	0.000	12079.4	0.829	BV	0.014	14491.83	1.279
18	3.315		0.0000	0.000	9921.4	0.681	VV	0.032	5095.56	0.450
19	3.346		0.0000	0.000	7035.7	0.483	VB	0.026	4558.59	0.402
20	3.417		0.0000	0.000	3774.7	0.259	BV	0.027	2340.68	0.207
21	3.442		0.0000	0.000	71537.1	4.911	VV	0.014	88166.75	7.780
22	3.492		0.0000	0.000	4674.2	0.321	VV	0.021	3716.46	0.328
23	3.516		0.0000	0.000	6578.6	0.452	VV	0.016	6934.35	0.612
24	3.544		0.0000	0.000	4940.5	0.339	VV	0.020	4057.99	0.358
25	3.566		0.0000	0.000	3828.6	0.263	VV	0.016	3943.62	0.348
26	3.599		0.0000	0.000	8499.9	0.583	VV	0.044	3248.96	0.287
27	3.662		0.0000	0.000	12492.6	0.858	VV	0.031	6760.22	0.597
28	3.710		0.0000	0.000	5973.4	0.410	VB	0.017	5740.85	0.507
29	3.747		0.0000	0.000	5999.2	0.412	BV	0.018	5700.93	0.503
30	3.770		0.0000	0.000	5432.7	0.373	VB	0.022	4177.78	0.369
31	3.826		0.0000	0.000	6065.8	0.416	BB	0.023	4315.47	0.381
32	3.881		0.0000	0.000	3519.5	0.242	BB	0.014	4128.54	0.364
33	3.950		0.0000	0.000	18290.2	1.256	BV	0.030	10038.54	0.886
34	3.973		0.0000	0.000	16816.8	1.154	VV	0.028	9839.14	0.868
35	4.020		0.0000	0.000	5398.7	0.371	VV	0.024	3701.76	0.327
36	4.096		0.0000	0.000	21171.0	1.453	VV	0.042	8402.57	0.741
37	4.119		0.0000	0.000	59261.1	4.068	VV	0.023	43084.48	3.802
38	4.187		0.0000	0.000	6045.0	0.415	VV	0.020	5058.49	0.446
39	4.225		0.0000	0.000	43527.6	2.988	VV	0.021	34316.24	3.028
40	4.259		0.0000	0.000	36616.1	2.514	VB	0.030	20372.63	1.798
41	4.342		0.0000	0.000	1295.7	0.089	BV	0.018	1208.15	0.107
42	4.377		0.0000	0.000	18897.9	1.297	VV	0.021	14704.83	1.298
43	4.405		0.0000	0.000	28607.4	1.964	VV	0.019	24905.42	2.198
44	4.433		0.0000	0.000	35791.8	2.457	VV	0.022	27533.61	2.430
45	4.473		0.0000	0.000	16507.9	1.133	VV	0.019	14524.66	1.282
46	4.498		0.0000	0.000	49085.2	3.370	VB	0.022	37118.15	3.275
47	4.544		0.0000	0.000	8235.1	0.565	BB	0.020	7007.30	0.618
48	4.595		0.0000	0.000	54113.2	3.715	BB	0.018	49514.97	4.369
49	4.634		0.0000	0.000	2641.9	0.181	BV	0.017	2607.36	0.230
50	4.669		0.0000	0.000	47890.9	3.288	VV	0.020	39360.36	3.473
51	4.698		0.0000	0.000	16241.6	1.115	VV	0.023	11622.04	1.026
52	4.744		0.0000	0.000	87997.8	6.041	VV	0.025	58512.89	5.163
53	4.844		0.0000	0.000	61665.8	4.233	VV	0.027	38288.48	3.379
54	4.895		0.0000	0.000	36082.3	2.477	VB	0.020	30065.60	2.653
55	4.942		0.0000	0.000	7336.9	0.504	BB	0.017	6990.23	0.617
56	4.981		0.0000	0.000	70998.2	4.874	BB	0.020	59040.39	5.210
57	5.018		0.0000	0.000	15939.5	1.094	BB	0.017	15898.50	1.403
58	5.059		0.0000	0.000	14034.1	0.963	BB	0.019	12490.20	1.102
59	5.158		0.0000	0.000	14723.8	1.011	BV	0.041	5990.12	0.529
60	5.208		0.0000	0.000	50822.4	3.489	VV	0.031	27696.34	2.444
61	5.279		0.0000	0.000	19880.5	1.365	VV	0.023	14704.34	1.298
62	5.316		0.0000	0.000	8142.1	0.559	VV	0.020	6714.29	0.592

ASX22 0154

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.348		0.0000	0.000	5424.4	0.372	VV	0.020	4447.91	0.392
64	5.404		0.0000	0.000	86083.5	5.909	VB	0.025	57350.61	5.061
65	5.472		0.0000	0.000	2803.1	0.192	BB	0.017	2807.02	0.248
66	5.523		0.0000	0.000	1112.4	0.076	BB	0.027	688.44	0.061
67	5.629		0.0000	0.000	40904.8	2.808	BB	0.019	36784.89	3.246
68	5.658		0.0000	0.000	21164.2	1.453	BB	0.018	19307.02	1.704
69	5.753		0.0000	0.000	774.6	0.053	BB	0.014	904.84	0.080
70	5.814		0.0000	0.000	491.7	0.034	BB	0.017	480.50	0.042
71	5.860		0.0000	0.000	615.8	0.042	BB	0.020	504.24	0.044
72	5.894		0.0000	0.000	87.9	0.006	BB	0.020	72.47	0.006
73	5.928		0.0000	0.000	5947.3	0.408	BB	0.024	4213.32	0.372
74	6.036		0.0000	0.000	14636.5	1.005	BB	0.028	8739.47	0.771
75	6.147		0.0000	0.000	1651.5	0.113	BB	0.022	1247.75	0.110
76	6.214		0.0000	0.000	18296.8	1.256	BB	0.026	11551.22	1.019
77	6.281		0.0000	0.000	241.7	0.017	BB	0.021	194.47	0.017
78	6.337		0.0000	0.000	220.1	0.015	BB	0.018	208.94	0.018
79	6.387		0.0000	0.000	88527.5	6.077	BB	0.024	61350.04	5.414
80	6.511		0.0000	0.000	96.1	0.007	BB	0.013	125.86	0.011
81	6.542		0.0000	0.000	6982.7	0.479	BB	0.019	6135.42	0.541
82	6.658		0.0000	0.000	235.7	0.016	BB	0.023	170.77	0.015
83	6.725		0.0000	0.000	604.6	0.042	BB	0.026	388.48	0.034
84	6.776		0.0000	0.000	2721.2	0.187	BB	0.019	2398.95	0.212
85	6.817		0.0000	0.000	4644.9	0.319	BB	0.022	3596.68	0.317
86	6.942		0.0000	0.000	1197.6	0.082	BB	0.029	696.93	0.061
87	7.022		0.0000	0.000	145.9	0.010	BB	0.021	114.64	0.010
88	7.083		0.0000	0.000	9408.4	0.646	BB	0.032	4901.23	0.432
89	7.285		0.0000	0.000	823.1	0.057	BB	0.026	519.52	0.046
90	7.380		0.0000	0.000	1215.6	0.083	BB	0.067	301.08	0.027

Total Area = 1456742.0, Total Amount = 167.086, Total Height = 1133239.0, Sample Units = PPB

ASX22: 8:55

LANCASTER LABORATORIES
File: C:\CPWIN\DATA1\3C6310.51R
5822076 RI ACF ABWHEE2 T 093070028A 01216
C:\CPWIN\DATA1\3C6310.51R



Instrument ID: CP06--S9147A Injected On: 11/9/2009 8:36:40 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\3C6310B.51R
5822076 RI ACF ABWHEE2 T 093070028A 01216
C:\CPWIN\DATA1\3C6310B.51R



ASX22 0156

Instrument ID: CP06--S9147B Injected On: 11/9/2009 8:36:40 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 30 Dilution Factor: 25
Analyst: 1151

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.316	18329	7.80862	TCX	2.815	30579	5.56953	TCX

Files:
Area File: C:\CPWIN\DATA1\3C6310.51A
Area File: C:\CPWIN\DATA1\3C6310B.51A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/9/2009 8:44:16 PM
File Reported On: 11/9/2009 at 8:44:25 PM

REPLOTS

ASX22 0157

5822076 RI ACF ABWHEE2 T 093070028A 01216
C:\CPWIN\DATA1\3C6310.51R



5822076 RI ACF ABWHEE2 T 093070028A 01216
C:\CPWIN\DATA1\3C6310B.51R



ASXZ2 0158

Multiple Component Data Summary

Sample Name: 5822077 RI ACF DF100 WHEE3 Sample ID:AC **Batchnumber: 093070028A**
Sample Amount:30 g Total Volume: 1000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on Nov 09, 2009 20:48:36
Instrument S9147A
Result file 3C6310.52R
Calibration file 1C6310
Method file PESTD

%SSR(TCX) 84.8% Conc: 8.676577
%SSR(DCB) * 0.0% Conc: 0

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260						**31.19**	
5.45	5.46	5.51	2930.27832	3439.923648			2
5.66	5.70	5.72	6117.302246	7287.96865			3
5.89	5.92	5.95	2479.633545	3651.105278			4
6.12	6.16	6.18	6939.099609	5300.113661			5
6.33	6.36	6.39	3298.336914	5147.506592			6

Height summation: **21764.650634**
Concentration CF: **4965.323566** L:

Analysis Report (B)

Injected on Nov 09, 2009 20:48:36
Instrument S9147B
Result file 3C6310B.52R
Calibration file 1C6310B
Method file PESTDB

%SSR(TCX) * 0.0% Conc: 0
%SSR(DCB) * 307.2% Conc: 31.22949:

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260						**67.01**	
4.55	4.60	4.61	37000.019531	7524.981814			1
4.68	4.72	4.74	40804.054688	7767.682321			2
4.79	4.82	4.85	11219.433594	2335.545676			3
4.92	4.96	4.98	14061.806641	3170.694158			4
5.34	5.37	5.40	19457.546875	2371.073602			5
5.60	5.63	5.66	8758.964844	1616.405436			6

Height summation: **131301.826173**
Concentration CF: **4131.063835** L:

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<1700	<330			4	40	
Aroclor-1221			<1700	<400			3	5	
Aroclor-1232			<1700	<530			4	10	
Aroclor-1242			<1700	<330			4	30	
Aroclor-1248			<1700	<330			4	40	
Aroclor-1254			<1700	<1700			4	40	W2C - interferent
Aroclor-1260	A	4965.323566	1700	330		18.34	4	40	

Units: ug/kg

%Difference = High - Low divided by the Average times 100

Reviewed By: Andrea Carey Date: 11-10-09
Verified By: [illegible] Date: 11/11/09

ASX22 0159

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822077** RI ACF DF100 WHEE3 Sample ID: AC **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 1000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:48:36
Instrument : CP06--S9147A
Result file : 3C6310.52R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

%SSR(TCX) : 84.8% Conc.: 8.676577
%SSR(DCB) : *0% Conc.: 0

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	153.05	
3.54	3.55	3.60	1887.845459	1953.636109			1
3.88	3.89	3.94	4444.454102	3888.207182			2
+ 4.08	4.09	4.14	17632.416016	11480.654191			3
E 4.08	4.11	4.14	71298.414063	46423.158089			3
4.13	4.17	4.19	2049.137207	2381.354219			4
E 4.19	4.19	4.25	76347.625	99240.400614			5
+ 4.19	4.23	4.25	1946.022705	2529.53609			5
4.40	4.42	4.46	1437.729126	2512.677827			6
Height Summation:			157465.204957				
Amount Avg CF:			26066.57234	Linear:			
Aroclor-1221					3	31.39	
3.43	3.45	3.49	424.314423	740.513827			1
3.49	3.49	3.55	740.793213	1335.845664			2
3.53	3.55	3.59	1887.845459	1400.271072			3
Height Summation:			3052.953095				
Amount Avg CF:			1158.876854	Linear:			
Aroclor-1232					6	155.11	
3.54	3.55	3.60	1887.845459	1797.691243			1
E 3.88	3.89	3.94	4444.454102	8737.745212			2
E+ 4.08	4.09	4.14	17632.416016	26528.873867			3
E 4.08	4.11	4.14	71298.414063	107272.119255			3
4.13	4.17	4.19	2049.137207	5433.935845			4
E 4.19	4.19	4.25	76347.625	230936.554749			5
+ 4.19	4.23	4.25	1946.022705	5886.336071			5
4.40	4.42	4.46	1437.729126	6032.008081			6
Height Summation:			157465.204957				
Amount Avg CF:			60035.009064	Linear:			
Aroclor-1242					6	152.83	
3.54	3.55	3.60	1887.845459	2460.534974			1
3.88	3.89	3.94	4444.454102	5014.333054			2
E+ 4.08	4.09	4.14	17632.416016	15016.535527			3
E 4.08	4.11	4.14	71298.414063	60720.843181			3
4.13	4.17	4.19	2049.137207	3167.381107			4
E 4.19	4.19	4.25	76347.625	128304.554239			5
+ 4.19	4.23	4.25	1946.022705	3270.351575			5
4.40	4.42	4.46	1437.729126	3147.73755			6
Height Summation:			157465.204957				
Amount Avg CF:			33802.564018	Linear:			
Aroclor-1248					6	217.19	
3.88	3.89	3.94	4444.454102	10223.622579			1
E 4.19	4.19	4.25	76347.625	161387.585347			2
+ 4.19	4.23	4.25	1946.022705	4113.604128			2
+ 4.26	4.26	4.32	510.942535	600.607181			3
4.26	4.30	4.32	1289.932495	1516.301084			3
4.40	4.42	4.46	1437.729126	1709.848018			4
4.50	4.53	4.56	1211.627563	1907.360018			5
4.70	4.71	4.76	1311.498657	1675.64663			6
Height Summation:			86042.866943				
Amount Avg CF:			29736.727279	Linear:			

Analysis Report (B)

Injected on : NOV 09, 2009 20:48:36
Instrument : CP06--S9147B
Result file : 3C6310B.52R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

%SSR(TCX) : *0% Conc.: 0
%SSR(DCB) : *307.2% Conc.: 31.229492

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					7	211.72	
3.07	3.12	3.13	828.221252	319.055652			1
E 3.40	3.42	3.46	145035.5625	41170.839373			2
E+ 3.40	3.44	3.46	114519.57031	32508.350043			2
+* 3.59	3.61	3.65	7133.649414	2092.6875			4
* 3.56	3.61	3.62	7133.649414	1137.524332			3
3.59	3.63	3.65	9370.262695	2748.807865			4
3.67	3.72	3.73	6774.140625	2079.19073			5
+ 3.84	3.85	3.90	77.736572	25.740962			6
3.84	3.87	3.90	3707.667725	1227.722432			6
3.84	3.89	3.90	3170.60083	1049.883121			6
Height Summation:			176020.105041				
Amount Avg CF:			7104.717644	Linear:			
Aroclor-1221					3	122.71	
2.95	2.95	3.01	2583.462891	1589.432073			1
+ 2.95	3.00	3.01	1079.57605	664.190999			1
E 3.02	3.05	3.08	6259.057129	7157.298032			2
3.07	3.12	3.13	828.221252	228.235574			3
Height Summation:			9670.741272				
Amount Avg CF:			2991.655226	Linear:			
Aroclor-1232					7	210.66	
3.07	3.12	3.13	828.221252	291.632336			1
E 3.40	3.42	3.46	145035.5625	91562.855114			2
E+ 3.40	3.44	3.46	114519.57031	72297.708531			2
+* 3.60	3.61	3.66	7133.649414	4717.554088			4
* 3.56	3.61	3.62	7133.649414	2557.276053			3
3.60	3.63	3.66	9370.262695	6196.64894			4
3.67	3.72	3.73	6774.140625	4902.403116			5
+ 3.84	3.85	3.90	77.736572	63.000707			6
3.84	3.87	3.90	3707.667725	3004.836474			6
3.84	3.89	3.90	3170.60083	2569.576813			6
Height Summation:			176020.105041				
Amount Avg CF:			15869.318407	Linear:			
Aroclor-1242					7	210.97	
3.07	3.12	3.13	828.221252	409.240662			1
E 3.40	3.42	3.46	145035.5625	53824.524048			2
E+ 3.40	3.44	3.46	114519.57031	42499.654981			2
+* 3.59	3.61	3.65	7133.649414	2753.294897			4
* 3.56	3.61	3.62	7133.649414	1494.161386			3
3.59	3.63	3.65	9370.262695	3616.535516			4
3.67	3.72	3.73	6774.140625	2765.858495			5
+ 3.84	3.85	3.90	77.736572	35.033044			6
3.84	3.87	3.90	3707.667725	1670.910893			6
3.84	3.89	3.90	3170.60083	1428.874391			6
Height Summation:			176020.105041				
Amount Avg CF:			9315.729342	Linear:			

ASX22 0160

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:53:53

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822077** RI ACF DF100 WHEE3 Sample ID: AC **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 1000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:48:36
Instrument : CP06–S9147A
Result file : 3C6310.52R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					**4**	**111.03**	
+ 4.26	4.26	4.32	510.942535	695.898416			1
4.26	4.30	4.32	1289.932495	1756.874636			1
4.68	4.71	4.74	1311.498657	1217.6111			2
4.85	4.89	4.91	25009.148438	19447.087066			3
5.66	5.70	5.72	6117.302246	8110.271769			6
Height Summation:			**33727.881836**				
Amount Avg CF:			**7632.961143**	Linear:			
Aroclor-1260					**5**	**31.19**	
5.45	5.46	5.51	2930.27832	3439.923648			2
5.66	5.70	5.72	6117.302246	7287.96865			3
5.89	5.92	5.95	2479.633545	3651.105278			4
+ 6.12	6.13	6.18	187.318069	143.074334			5
6.12	6.16	6.18	6939.099609	5300.113661			5
6.33	6.36	6.39	3298.336914	5147.506592			6
Height Summation:			**21764.650634**				
Amount Avg CF:			**4965.323566**	Linear:			
Aroclor-1262					**5**	**33.97**	
5.45	5.46	5.51	2930.27832	3777.923448			2
5.73	5.76	5.79	1469.572388	1379.45274			3
5.89	5.92	5.95	2479.633545	3017.004238			4
+ 6.12	6.13	6.18	187.318069	111.191486			5
6.12	6.16	6.18	6939.099609	4119.030257			5
6.32	6.36	6.38	3298.336914	3466.686126			6
Height Summation:			**17116.920776**				
Amount Avg CF:			**3152.019362**	Linear:			
Aroclor-1268					**6**	**106.21**	
5.90	5.92	5.96	2479.633545	4165.63106			1
6.32	6.36	6.38	3298.336914	1816.280823			2
6.38	6.41	6.44	1290.844238	889.136103			3
6.51	6.51	6.57	264.704926	170.011552			4
6.74	6.78	6.80	803.750916	1487.9647			5
6.96	6.99	7.02	109.602325	29.884556			6
Height Summation:			**8246.872864**				
Amount Avg CF:			**1426.484799**	Linear:			

Analysis Report (B)

Injected on : NOV 09, 2009 20:48:36
Instrument : CP06–S9147B
Result file : 3C6310B.52R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1248					**6**	**146.48**	
+ 3.28	3.30	3.34	2067.664551	1117.034597			1
3.28	3.33	3.34	34780.230469	18789.663306			1
3.56	3.61	3.62	7133.649414	2240.508594			2
3.71	3.72	3.77	6774.140625	1735.34393			3
+ 3.84	3.85	3.90	77.736572	18.544737			4
3.84	3.87	3.90	3707.667725	884.49648			4
3.84	3.89	3.90	3170.60083	756.374487			4
E 4.45	4.46	4.51	79103.390625	38181.902567			6
+ 4.45	4.50	4.51	54863.480469	26481.697551			6
Height Summation:			**134669.679688**				
Amount Avg CF:			**10431.381561**	Linear:			
Aroclor-1254					**6**	**77.91**	
+ 4.07	4.09	4.13	3652.799805	593.636902			1
4.07	4.11	4.13	5741.506836	933.08435			1
4.18	4.21	4.24	92490.429688	18314.428845			2
4.45	4.46	4.51	79103.390625	13391.221326			3
+ 4.45	4.50	4.51	54863.480469	9287.705671			3
+ 4.62	4.63	4.68	38747.363281	11530.614982			4
4.62	4.65	4.68	43977.230469	13086.942429			4
4.79	4.82	4.85	11219.433594	4553.581802			5
4.92	4.96	4.98	14061.806641	3203.139184			6
Height Summation:			**246593.797853**				
Amount Avg CF:			**8913.732989**	Linear:			
Aroclor-1260					**6**	**125.14**	
4.55	4.56	4.61	130332.28125	26506.68996			1
+ 4.55	4.60	4.61	37000.019531	7524.981814 /			1
4.68	4.69	4.74	57312.320313	10910.285771			2
+ 4.68	4.72	4.74	40804.054688	7767.682321 /			2
4.79	4.82	4.85	11219.433594	2335.545676 /			3
4.92	4.96	4.98	14061.806641	3170.694158 /			4
5.34	5.37	5.40	19457.546875	2371.073602 /			5
5.60	5.63	5.66	8758.964844	1616.405436 /			6
Height Summation:			**241142.353517**				
Amount Avg CF:			**7818.449101**	Linear:			
Aroclor-1262					**7**	**147.35**	
4.68	4.69	4.74	57312.320313	12264.320023			1
+ 4.68	4.72	4.74	40804.054688	8731.699959			1
4.95	4.96	5.01	14061.806641	2153.072456			2
+ 4.95	4.99	5.01	6888.255859	1054.694773			2
+ 5.15	5.16	5.21	2295.750488	373.327728			3
5.15	5.18	5.21	7016.916992	1141.068983			3
5.15	5.21	5.21	3861.119629	627.883137			3
+ 5.34	5.34	5.40	2882.238281	272.654648			4
5.34	5.37	5.40	19457.546875	1840.649549			4
5.60	5.63	5.66	8758.964844	1223.390192			5
+ 5.97	5.97	6.03	162.392105	42.427168			6
5.97	6.00	6.03	2536.26416	662.633841			6
Height Summation:			**113004.939454**				
Amount Avg CF:			**2844.716883**	Linear:			

ASX22 8161

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:53:53

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822077** RI ACF DF100 WHEE3 Sample ID: AC **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 1000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on : NOV 09, 2009 20:48:36
Instrument : CP06–S9147A
Result file : 3C6310.52R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Analysis Report (B)

Injected on : NOV 09, 2009 20:48:36
Instrument : CP06–S9147B
Result file : 3C6310B.52R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1268					7	102.27	
+ 5.15	5.16	5.21	2295.750488	644.580774			1
5.15	5.18	5.21	7016.916992	1970.148677			1
5.15	5.21	5.21	3861.119629	1084.091452			1
5.55	5.59	5.61	1736.892334	143.281278			2
5.60	5.63	5.66	8758.964844	857.902494			3
5.75	5.79	5.81	195.868973	18.617814			4
+ 5.97	5.97	6.03	162.392105	41.995511			5
5.97	6.00	6.03	2536.26416	655.892164			5
6.17	6.18	6.23	1944.839111	72.631808			6

Height Summation: **26050.866043**
Amount Avg CF: **686.080812** Linear:

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			1700	330	E	** 114.33	4	40	
Aroclor-1221			1700	330		** 88.32	3	5	
Aroclor-1232			1700	330	E	** 116.37	4	10	
Aroclor-1242			1700	330	E	** 113.58	4	30	
Aroclor-1248			1700	330	E	** 96.12	4	40	
Aroclor-1254			1700	330		15.48	4	40	
Aroclor-1260			1700	330		** 44.64	4	40	
Aroclor-1262			0	0		10.25	4	40	
Aroclor-1268			0	0		** 70.10	4	40	

Units: _____

ASX22 0162

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:53:53

Sample Name: **5822077 RI ACF DF100ACWHEE3 T 093070028A 01216**
Acquired from CP06--S9147A via port 1 on 11/9/09 08:56:06pm by 1151
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\3C6310.52R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.031		0.0000	0.000	750.4	0.069	BB	0.020	625.26	0.094
2	3.080		0.0000	0.000	145.3	0.013	BB	0.023	104.57	0.016
3	3.143		0.0000	0.000	4377.7	0.401	BB	0.024	3005.10	0.454
4	3.234		0.0000	0.000	336.2	0.031	BB	0.018	311.78	0.047
5	3.296		0.0000	0.000	78.5	0.007	BB	0.012	112.74	0.017
6	3.315	TCX	260.2973	100.000	397.1	0.036	BB	0.013	509.17	0.077
7	3.446		0.0000	0.000	594.8	0.054	BB	0.023	424.31	0.064
8	3.493		0.0000	0.000	1497.3	0.137	BB	0.034	740.79	0.112
9	3.547		0.0000	0.000	1743.2	0.160	BB	0.015	1887.85	0.285
10	3.597		0.0000	0.000	1414.1	0.130	BB	0.027	869.18	0.131
11	3.643		0.0000	0.000	1761.9	0.161	BB	0.029	1030.08	0.155
12	3.753		0.0000	0.000	6495.7	0.595	BB	0.015	7205.13	1.087
13	3.787		0.0000	0.000	637.3	0.058	BB	0.018	581.44	0.088
14	3.816		0.0000	0.000	423.7	0.039	BB	0.012	612.83	0.092
15	3.856		0.0000	0.000	415.3	0.038	BB	0.020	352.92	0.053
16	3.889		0.0000	0.000	4154.1	0.381	BB	0.016	4444.45	0.671
17	3.951		0.0000	0.000	17722.6	1.624	BB	0.021	14357.06	2.167
18	3.981		0.0000	0.000	1234.7	0.113	BB	0.016	1297.78	0.196
19	4.037		0.0000	0.000	148751.9	13.629	BB	0.035	71706.74	10.822
20	4.090		0.0000	0.000	11451.9	1.049	BB	0.011	17632.42	2.661
21	4.115		0.0000	0.000	75491.8	6.917	BB	0.018	71298.41	10.760
22	4.167		0.0000	0.000	2146.6	0.197	BB	0.017	2049.14	0.309
23	4.192		0.0000	0.000	93034.3	8.524	BB	0.020	76347.63	11.522
24	4.229		0.0000	0.000	1421.4	0.130	BB	0.012	1946.02	0.294
25	4.263		0.0000	0.000	348.5	0.032	BB	0.011	510.94	0.077
26	4.303		0.0000	0.000	962.5	0.088	BB	0.012	1289.93	0.195
27	4.421		0.0000	0.000	2552.3	0.234	BB	0.030	1437.73	0.217
28	4.474		0.0000	0.000	2445.5	0.224	BB	0.018	2266.00	0.342
29	4.534		0.0000	0.000	2315.8	0.212	BB	0.032	1211.63	0.183
30	4.621		0.0000	0.000	798.8	0.073	BB	0.022	613.26	0.093
31	4.675		0.0000	0.000	1078.7	0.099	BV	0.023	776.25	0.117
32	4.715		0.0000	0.000	1981.4	0.182	VB	0.025	1311.50	0.198
33	4.764		0.0000	0.000	150.7	0.014	BB	0.014	183.25	0.028
34	4.793		0.0000	0.000	2467.1	0.226	BB	0.021	1932.66	0.292
35	4.892		0.0000	0.000	37012.6	3.391	BV	0.025	25009.15	3.774
36	4.965		0.0000	0.000	16655.8	1.526	VB	0.020	13647.45	2.060
37	5.051		0.0000	0.000	32105.1	2.942	BV	0.020	27032.23	4.080
38	5.075		0.0000	0.000	53048.9	4.860	VV	0.022	39676.38	5.988
39	5.123		0.0000	0.000	233047.4	21.352	VV	0.042	93402.17	14.096
40	5.185		0.0000	0.000	16830.2	1.542	VB	0.022	12963.74	1.956
41	5.241		0.0000	0.000	69029.3	6.325	BB	0.037	30808.03	4.649
42	5.346		0.0000	0.000	12796.0	1.172	BB	0.023	9209.62	1.390
43	5.408		0.0000	0.000	87547.6	8.021	BB	0.032	45619.40	6.885
44	5.464		0.0000	0.000	3245.9	0.297	BB	0.018	2930.28	0.442
45	5.547		0.0000	0.000	81084.8	7.429	BB	0.039	35086.29	5.295
46	5.626		0.0000	0.000	1860.4	0.170	BB	0.021	1458.65	0.220
47	5.701		0.0000	0.000	12520.3	1.147	BB	0.034	6117.30	0.923
48	5.760		0.0000	0.000	1278.2	0.117	BB	0.014	1469.57	0.222
49	5.801		0.0000	0.000	2488.8	0.228	BB	0.021	1966.22	0.297
50	5.923		0.0000	0.000	4602.5	0.422	BB	0.031	2479.63	0.374
51	5.983		0.0000	0.000	1462.0	0.134	BB	0.023	1064.73	0.161
52	6.021		0.0000	0.000	217.7	0.020	BB	0.013	289.96	0.044
53	6.059		0.0000	0.000	222.6	0.020	BB	0.018	201.62	0.030
54	6.100		0.0000	0.000	933.8	0.086	BB	0.018	877.52	0.132
55	6.125		0.0000	0.000	98.0	0.009	BB	0.009	187.32	0.028
56	6.158		0.0000	0.000	9732.7	0.892	BB	0.023	6939.10	1.047
57	6.292		0.0000	0.000	1443.1	0.132	BB	0.081	295.81	0.045
58	6.319		0.0000	0.000	1343.4	0.123	BB	0.016	1384.31	0.209
59	6.359		0.0000	0.000	4032.1	0.369	BB	0.020	3298.34	0.498
60	6.412		0.0000	0.000	2289.6	0.210	BB	0.030	1290.84	0.195
61	6.450		0.0000	0.000	344.7	0.032	BB	0.014	396.38	0.060
62	6.513		0.0000	0.000	372.6	0.034	BB	0.023	264.70	0.040

ASX22 0163

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.630		0.0000	0.000	1085.6	0.099	BB	0.038	474.27	0.072
64	6.780		0.0000	0.000	964.7	0.088	BB	0.020	803.75	0.121
65	6.843		0.0000	0.000	488.0	0.045	BB	0.027	299.42	0.045
66	6.895		0.0000	0.000	260.8	0.024	BB	0.024	178.51	0.027
67	6.948		0.0000	0.000	125.0	0.011	BB	0.020	102.38	0.015
68	6.994		0.0000	0.000	96.9	0.009	BB	0.015	109.60	0.017
69	7.038		0.0000	0.000	1601.7	0.147	BB	0.024	1124.96	0.170
70	7.107		0.0000	0.000	3459.1	0.317	BB	0.053	1085.38	0.164
71	7.209		0.0000	0.000	531.5	0.049	BB	0.024	363.44	0.055
72	7.269		0.0000	0.000	3604.1	0.330	BB	0.035	1717.20	0.259

Total Area = 1091442.0, Total Amount = 260.297, Total Height = 662611.5, Sample Units = PPB

ASX22 8164

Sample Name: **5822077 RI ACF DF100ACWHEE3 T 093070028A 01216**
Acquired from CP06--S9147B via port 2 on 11/9/09 08:56:06pm by 1151
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\3C6310B.52R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.525		0.0000	0.000	1628.4	0.053	BV	0.014	1939.70	0.082
2	2.575		0.0000	0.000	1638.2	0.053	VV	0.022	1244.39	0.053
3	2.600		0.0000	0.000	1290.6	0.042	VB	0.021	1033.21	0.044
4	2.647		0.0000	0.000	3982.2	0.129	BV	0.023	2856.72	0.121
5	2.708		0.0000	0.000	850.4	0.028	VV	0.021	666.26	0.028
6	2.734		0.0000	0.000	7581.6	0.246	VB	0.017	7547.20	0.320
7	2.777		0.0000	0.000	11467.4	0.372	BB	0.027	7124.22	0.302
8	2.857		0.0000	0.000	3682.4	0.119	BB	0.015	4006.09	0.170
9	2.901		0.0000	0.000	12456.3	0.404	BB	0.023	9176.82	0.389
10	2.926		0.0000	0.000	9013.9	0.292	BB	0.014	10916.14	0.463
11	2.954		0.0000	0.000	2794.4	0.091	BB	0.018	2583.46	0.110
12	2.996		0.0000	0.000	898.1	0.029	BB	0.014	1079.58	0.046
13	3.053		0.0000	0.000	22238.0	0.721	BB	0.059	6259.06	0.266
14	3.120		0.0000	0.000	688.1	0.022	BB	0.014	828.22	0.035
15	3.168		0.0000	0.000	12847.1	0.417	BB	0.013	16445.37	0.698
16	3.203		0.0000	0.000	861.6	0.028	BB	0.009	1552.29	0.066
17	3.228		0.0000	0.000	1300.6	0.042	BB	0.013	1703.04	0.072
18	3.270		0.0000	0.000	9220.9	0.299	BV	0.014	10916.35	0.463
19	3.304		0.0000	0.000	1624.2	0.053	VV	0.013	2067.66	0.088
20	3.335		0.0000	0.000	33810.7	1.097	VV	0.016	34780.23	1.476
21	3.363		0.0000	0.000	7201.7	0.234	VV	0.017	7199.27	0.306
22	3.420		0.0000	0.000	273192.0	8.862	VV	0.031	145035.60	6.155
23	3.440		0.0000	0.000	174095.8	5.648	VV	0.025	114519.60	4.860
24	3.493		0.0000	0.000	205358.9	6.662	VV	0.017	199361.30	8.460
25	3.550		0.0000	0.000	251248.5	8.151	VV	0.025	165666.60	7.030
26	3.605		0.0000	0.000	8612.9	0.279	VV	0.020	7133.65	0.303
27	3.627		0.0000	0.000	20886.9	0.678	VV	0.037	9370.26	0.398
28	3.721		0.0000	0.000	8639.1	0.280	VV	0.021	6774.14	0.287
29	3.770		0.0000	0.000	3432.8	0.111	VV	0.017	3427.69	0.145
30	3.821		0.0000	0.000	3690.8	0.120	VB	0.020	3006.91	0.128
31	3.853		0.0000	0.000	42.1	0.001	BB	0.009	77.74	0.003
32	3.871		0.0000	0.000	2982.4	0.097	BV	0.013	3707.67	0.157
33	3.890		0.0000	0.000	3103.3	0.101	VB	0.016	3170.60	0.135
34	3.936		0.0000	0.000	1913.1	0.062	BV	0.021	1549.71	0.066
35	3.958		0.0000	0.000	4801.0	0.156	VV	0.021	3891.48	0.165
36	3.998		0.0000	0.000	555.2	0.018	VB	0.011	815.52	0.035
37	4.025		0.0000	0.000	2277.6	0.074	BB	0.015	2462.10	0.104
38	4.066		0.0000	0.000	4334.8	0.141	BV	0.017	4261.28	0.181
39	4.087		0.0000	0.000	3446.9	0.112	VV	0.016	3652.80	0.155
40	4.110		0.0000	0.000	6954.5	0.226	VV	0.020	5741.51	0.244
41	4.147		0.0000	0.000	14170.8	0.460	VB	0.027	8739.73	0.371
42	4.212		0.0000	0.000	100904.0	3.273	BB	0.018	92490.43	3.925
43	4.283		0.0000	0.000	43302.6	1.405	BV	0.017	43041.18	1.827
44	4.312		0.0000	0.000	6516.3	0.211	VB	0.014	7547.64	0.320
45	4.368		0.0000	0.000	224047.1	7.268	BV	0.029	128400.20	5.449
46	4.389		0.0000	0.000	108252.4	3.512	VV	0.015	122212.00	5.186
47	4.412		0.0000	0.000	410061.0	13.302	VV	0.021	318201.20	13.503
48	4.463		0.0000	0.000	71783.1	2.329	VB	0.015	79103.39	3.357
49	4.505		0.0000	0.000	52750.9	1.711	BV	0.016	54863.48	2.328
50	4.538		0.0000	0.000	203026.2	6.586	VV	0.024	138949.50	5.897
51	4.562		0.0000	0.000	192889.9	6.257	VV	0.025	130332.30	5.531
52	4.598		0.0000	0.000	44824.4	1.454	VV	0.020	37000.02	1.570
53	4.629		0.0000	0.000	61368.0	1.991	VV	0.026	38747.36	1.644
54	4.654		0.0000	0.000	52025.7	1.688	VV	0.020	43977.23	1.866
55	4.691		0.0000	0.000	80224.6	2.602	VV	0.023	57312.32	2.432
56	4.720		0.0000	0.000	44617.7	1.447	VV	0.018	40804.05	1.732
57	4.751		0.0000	0.000	62924.7	2.041	VV	0.023	45226.89	1.919
58	4.823		0.0000	0.000	14722.4	0.478	VB	0.022	11219.43	0.476
59	4.876		0.0000	0.000	4184.9	0.136	BV	0.016	4450.86	0.189
60	4.912		0.0000	0.000	16461.1	0.534	VV	0.023	12125.30	0.515
61	4.957		0.0000	0.000	15898.0	0.516	VV	0.019	14061.81	0.597
62	4.990		0.0000	0.000	7439.6	0.241	VB	0.018	6888.26	0.292

ASX22 6165

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.031		0.0000	0.000	3235.3	0.105	BV	0.015	3581.33	0.152
64	5.053		0.0000	0.000	2155.1	0.070	VB	0.016	2268.36	0.096
65	5.098		0.0000	0.000	556.7	0.018	BV	0.016	573.41	0.024
66	5.133		0.0000	0.000	1939.9	0.063	VV	0.023	1385.71	0.059
67	5.163		0.0000	0.000	2503.9	0.081	VV	0.018	2295.75	0.097
68	5.184		0.0000	0.000	7331.9	0.238	VV	0.017	7016.92	0.298
69	5.206		0.0000	0.000	3311.4	0.107	VV	0.014	3861.12	0.164
70	5.229		0.0000	0.000	11520.0	0.374	VV	0.020	9603.77	0.408
71	5.257		0.0000	0.000	5084.1	0.165	VB	0.014	6031.62	0.256
72	5.292		0.0000	0.000	4089.9	0.133	BV	0.017	4128.68	0.175
73	5.315		0.0000	0.000	3666.5	0.119	VV	0.016	3749.40	0.159
74	5.337		0.0000	0.000	3059.9	0.099	VV	0.018	2882.24	0.122
75	5.373		0.0000	0.000	28051.9	0.910	VB	0.024	19457.55	0.826
76	5.428		0.0000	0.000	3061.3	0.099	BB	0.015	3477.39	0.148
77	5.498		0.0000	0.000	3702.5	0.120	BB	0.028	2202.95	0.093
78	5.543		0.0000	0.000	2016.3	0.065	BB	0.019	1812.12	0.077
79	5.586		0.0000	0.000	2015.4	0.065	BB	0.019	1736.89	0.074
80	5.630		0.0000	0.000	9027.0	0.293	BB	0.017	8758.96	0.372
81	5.791		0.0000	0.000	218.4	0.007	BB	0.019	195.87	0.008
82	5.828		0.0000	0.000	170.7	0.006	BB	0.016	182.25	0.008
83	5.854		0.0000	0.000	66.5	0.002	BB	0.014	77.22	0.003
84	5.902		0.0000	0.000	1266.9	0.041	BB	0.016	1348.92	0.057
85	5.974		0.0000	0.000	336.2	0.011	BB	0.035	162.39	0.007
86	6.005		0.0000	0.000	2291.7	0.074	BB	0.015	2536.26	0.108
87	6.085		0.0000	0.000	784.5	0.025	BB	0.036	366.22	0.016
88	6.116		0.0000	0.000	662.8	0.022	BB	0.021	520.26	0.022
89	6.180		0.0000	0.000	4070.9	0.132	BB	0.035	1944.84	0.083
90	6.244		0.0000	0.000	3468.8	0.113	BB	0.029	1966.80	0.083
91	6.311		0.0000	0.000	806.0	0.026	BB	0.018	733.11	0.031
92	6.359	DCB	936.8848	100.000	2262.8	0.073	BB	0.017	2236.07	0.095
93	6.391		0.0000	0.000	1189.0	0.039	BB	0.015	1327.53	0.056
94	6.495		0.0000	0.000	225.1	0.007	BB	0.026	142.36	0.006
95	6.620		0.0000	0.000	249.5	0.008	BB	0.039	106.74	0.005
96	6.675		0.0000	0.000	233.1	0.008	BB	0.028	138.89	0.006
97	6.788		0.0000	0.000	268.7	0.009	BB	0.024	188.03	0.008
98	6.838		0.0000	0.000	571.3	0.019	BB	0.025	377.47	0.016
99	6.909		0.0000	0.000	359.0	0.012	BB	0.019	307.62	0.013
100	6.944		0.0000	0.000	562.3	0.018	BB	0.018	519.59	0.022
101	6.984		0.0000	0.000	582.8	0.019	BB	0.021	468.43	0.020
102	7.009		0.0000	0.000	326.7	0.011	BB	0.016	350.27	0.015
103	7.054		0.0000	0.000	71.4	0.002	BB	0.015	78.72	0.003
104	7.085		0.0000	0.000	1658.7	0.054	BB	0.024	1166.58	0.050
105	7.124		0.0000	0.000	1278.6	0.041	BB	0.022	950.50	0.040
106	7.170		0.0000	0.000	1786.7	0.058	BB	0.033	909.63	0.039
107	7.234		0.0000	0.000	121.9	0.004	BB	0.019	106.10	0.005
108	7.286		0.0000	0.000	1349.4	0.044	BB	0.023	969.66	0.041

Total Area = 3082612.0, Total Amount = 936.885, Total Height = 2356448.0, Sample Units = PPB

ASX22 0166

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3C6310.52R

5822077 RI ACF DF100ACWHEE3 T 093070028A 01216

——— C:\CPWIN\DATA1\3C6310.52R



Instrument ID: CP06--S9147A Injected On: 11/9/2009 8:48:36 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\3C6310B.52R

5822077 RI ACF DF100ACWHEE3 T 093070028A 01216

——— C:\CPWIN\DATA1\3C6310B.52R



Instrument ID: CP06--S9147B Injected On: 11/9/2009 8:48:36 PM Column ID: RTX CLPII

ASX22 8167

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 30 Dilution Factor: 1000
Analyst: 1151

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.315	509	8.67658	TCX	.	0	.	TCX
.	0	.	DCB	6.359	2236	31.22949	DCB

Files:
Area File: C:\CPWIN\DATA1\3C6310.52A
Area File: C:\CPWIN\DATA1\3C6310B.52A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/9/2009 8:56:12 PM
File Reported On: 11/9/2009 at 8:56:20 PM

REPLOTS

ASX22 0168

5822077 RI ACF DF100ACWHEE3 T 093070028A 01216
——— C:\CPWIN\DATA1\3C6310.52R



5822077 RI ACF DF100ACWHEE3 T 093070028A 01216
——— C:\CPWIN\DATA1\3C6310B.52R



ASX22 8169

Multiple Component Data Summary

Sample Name: 5822078 AF BB--1 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1030 ml Total Volume: 10 ml Analyst: 2131 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on Nov 06, 2009 02:04:34
Instrument H5088A
Result file 3P7307.51R
Calibration file 1P7307
Method file PESTD

%SSR(TCX) 77.4% Conc: 0.230594
%SSR(DCB) 59.2% Conc: 0.175412

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						44.90	
4.83	4.85	4.89	10061.869147	0.813717			1
5.03	5.05	5.09	21221.203125	1.135211			2
5.27	5.30	5.33	26462.611321	1.94472			3
Height summation:			57745.683594				
Concentration		CF:	1.297883	L:			
Aroclor-1260						3.87	
6.25	6.28	6.31	22642.189455	1.462941			4
6.55	6.58	6.61	47588.984375	1.441084			5
6.80	6.83	6.86	23506.332037	1.549647			6
Height summation:			93737.505859				
Concentration		CF:	1.484557	L:			

Analysis Report (B)

Injected on Nov 06, 2009 02:04:34
Instrument H5088B
Result file 3P7307B.51R
Calibration file 1P7307B
Method file PESTDB

%SSR(TCX) 80.2% Conc: 0.238961
%SSR(DCB) 74.6% Conc: 0.22098

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						62.86	
4.08	4.11	4.14	11352.577148	0.590532			1
4.66	4.69	4.72	36347.703125	2.992483			3
4.76	4.79	4.82	39181.480469	2.855838			4
Height summation:			86881.760742				
Concentration		CF:	2.146284	L:			
Aroclor-1260						15.43	
5.36	5.39	5.42	25710.076172	1.544534			4
5.60	5.62	5.66	69813.226563	2.110457			5
5.89	5.92	5.95	42007.195312	1.8831			6
Height summation:			137530.498048				
Concentration		CF:	1.84603	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<0.4854	<0.0971			4	15	
Aroclor-1221			<0.4854	<0.0971			3	5	
Aroclor-1232			<0.4854	<0.0971			4	10	
Aroclor-1242			<0.4854	<0.0971			4	30	
Aroclor-1248			<0.4854	<0.0971			4	30	
Aroclor-1254	A	1.297883	0.4854	0.0971		49.27	4	40	1809
Aroclor-1260	B	1.84603	0.4854	0.0971		21.71	4	40	

Units: ug/l

%Difference = High - Low divided by the Average times 100

Reviewed By: [signature] Date: 11-6-09
Verified By: [signature] Date: 11/9/09

ASX22 0170

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822078** AF BB--1 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1030 ml Total Volume: 10 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 02:04:34
Instrument : CP07--H5088A
Result file : 3P7307.51R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

%SSR(TCX) : 77.4% Conc.: 0.230594
%SSR(DCB) : 59.2% Conc.: 0.175412

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					5	87.88	
3.83	3.86	3.89	225.352615	0.017253			2
4.10	4.13	4.16	1208.792114	0.043167			3
4.15	4.19	4.21	1994.921021	0.160247			4
4.24	4.28	4.30	802.549194	0.068073			5
4.62	4.65	4.68	2039.50354	0.242734			6
Height Summation:			6271.118484				
Amount Avg CF:			0.106295	Linear:			
Aroclor-1221					1		
3.28	3.28	3.34	1709.767212	0.378772			2
Height Summation:			1709.767212				
Amount Avg CF:			0.378772	Linear:			
Aroclor-1232					5	91.43	
3.83	3.86	3.89	225.352615	0.038775			2
4.10	4.13	4.16	1208.792114	0.090186			3
4.15	4.19	4.21	1994.921021	0.334829			4
4.23	4.28	4.29	802.549194	0.145152			5
4.62	4.65	4.68	2039.50354	0.558956			6
Height Summation:			6271.118484				
Amount Avg CF:			0.23358	Linear:			
Aroclor-1242					5	91.54	
3.83	3.86	3.89	225.352615	0.021115			2
4.09	4.09	4.15	12906.97851	0.532885			3
+ 4.09	4.13	4.15	1208.792114	0.049907			3
4.15	4.19	4.21	1994.921021	0.187358			4
4.23	4.28	4.29	802.549194	0.079366			5
4.62	4.65	4.68	2039.50354	0.276995			6
Height Summation:			17969.304886				
Amount Avg CF:			0.219544	Linear:			
Aroclor-1248					6	163.40	
4.10	4.13	4.16	1208.792114	0.078273			1
4.33	4.36	4.39	3109.576172	0.289319			2
4.50	4.53	4.56	1961.635254	0.121856			3
+ 4.50	4.56	4.56	1345.272583	0.083568			3
4.62	4.65	4.68	2039.50354	0.191794			4
5.02	5.05	5.08	21221.20312	1.64779			5
5.61	5.64	5.67	33295.62109	5.437027			6
Height Summation:			62836.331299				
Amount Avg CF:			1.294343	Linear:			
Aroclor-1254					6	48.83	
4.83	4.85	4.89	10061.86914	0.813717			1
+ 4.83	4.88	4.89	4204.963867	0.340061			1
5.03	5.05	5.09	21221.20312	1.352411			2
5.27	5.30	5.33	26462.61132	1.94472			3
5.34	5.37	5.40	26775.25195	1.097156			4
5.85	5.87	5.91	32386.63281	2.571202			5
5.98	6.01	6.04	47427.96875	2.842851			6
Height Summation:			164335.537110				
Amount Avg CF:			1.734143	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 02:04:34
Instrument : CP07--H5088B
Result file : 3P7307B.51R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Linear mA

%SSR(TCX) : 80.2% Conc.: 0.238961
%SSR(DCB) : 74.6% Conc.: 0.22098

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					5	93.35	
2.69	2.74	2.75	10190.08984	1.282731			1
2.98	3.01	3.04	1206.908203	0.068445			2
3.14	3.16	3.20	867.262451	0.071402			3
3.42	3.46	3.48	4692.800781	1.331272			5
3.77	3.80	3.83	5762.888672	0.572015			6
Height Summation:			22719.949951				
Amount Avg CF:			0.665173	Linear:			
Aroclor-1221					2	57.90	
2.61	2.62	2.67	1134.355835	0.388129			2
+ 2.61	2.66	2.67	272.891815	0.093372			2
2.69	2.74	2.75	10190.08984	0.926295			3
Height Summation:			11324.445679				
Amount Avg CF:			0.657212	Linear:			
Aroclor-1232					5	99.62	
2.69	2.74	2.75	10190.08984	1.120862			1
2.98	3.01	3.04	1206.908203	0.15097			2
3.14	3.16	3.20	867.262451	0.15262			3
E 3.42	3.46	3.48	4692.800781	2.95468			5
3.77	3.80	3.83	5762.888672	1.407203			6
Height Summation:			22719.949951				
Amount Avg CF:			1.157267	Linear:			
Aroclor-1242					5	91.88	
2.68	2.74	2.74	10190.08984	1.545223			1
2.98	3.01	3.04	1206.908203	0.084009			2
3.14	3.16	3.20	867.262451	0.087947			3
3.42	3.46	3.48	4692.800781	1.461465			5
3.77	3.80	3.83	5762.888672	0.689766			6
Height Summation:			22719.949951				
Amount Avg CF:			0.773682	Linear:			
Aroclor-1248					5	213.17	
3.14	3.16	3.20	867.262451	0.199769			1
3.77	3.80	3.83	5762.888672	0.418212			3
3.89	3.92	3.95	1862.494629	0.288752			4
4.26	4.29	4.32	17562.03710	1.660265			5
E 4.55	4.60	4.61	422320.9687	65.896153			6
Height Summation:			448375.651611				
Amount Avg CF:			13.69263	Linear:			
Aroclor-1254					6	119.46	
4.08	4.11	4.14	11352.57714	0.590532			1
+ 4.08	4.14	4.14	5131.420898	0.266923			1
E 4.55	4.60	4.61	422320.9687	19.710319			2
4.66	4.69	4.72	36347.70312	2.992483			3
4.76	4.79	4.82	39181.48046	2.855938			4
4.85	4.87	4.91	49712.05468	5.48552			5
5.11	5.14	5.17	63131.27734	3.386483			6
Height Summation:			622046.061524				
Amount Avg CF:			5.836863	Linear:			

ASX22 0171

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:23:38

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822078** AF BB--1 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1030 ml Total Volume: 10 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 02:04:34
Instrument : CP07--H5088A
Result file : 3P7307.51R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	33.45	
5.74	5.77	5.80	45702.44921	2.173366			1
5.98	6.01	6.04	47427.96875	2.651786			2
6.05	6.08	6.11	15182.57617	1.072893			3
6.25	6.28	6.31	22642.18945	1.462941			4
6.55	6.58	6.61	47588.98437	1.441084			5
6.80	6.83	6.86	23506.33203	1.549647			6
Height Summation:			**202050.500000**				
Amount Avg CF:			**1.725286**	Linear:			
Aroclor-1262					6	48.26	
5.49	5.52	5.55	24371.56640	1.422159			1
5.73	5.77	5.79	45702.44921	2.560098			2
6.05	6.08	6.11	15182.57617	0.642666			3
6.25	6.28	6.31	22642.18945	1.231889			4
6.55	6.58	6.61	47588.98437	1.12763			5
6.79	6.83	6.85	23506.33203	1.081983			6
Height Summation:			**178994.097656**				
Amount Avg CF:			**1.344404**	Linear:			
Aroclor-1268					5	97.24	
6.05	6.08	6.11	15182.57617	1.624217			1
6.26	6.28	6.32	22642.18945	1.898377			2
6.79	6.83	6.85	23506.33203	0.545907			3
6.87	6.89	6.93	8797.543945	0.236541			4
7.03	7.05	7.09	1026.644897	0.028457			5
Height Summation:			**71155.286498**				
Amount Avg CF:			**0.8667**	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 02:04:34
Instrument : CP07--H5088B
Result file : 3P7307B.51R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	17.66	
4.66	4.69	4.72	36347.70312	1.879824			1
4.84	4.87	4.90	49712.05468	2.430766			2
4.96	4.98	5.02	23714.37695	1.566613			3
5.36	5.39	5.42	25710.07617	1.544534			4
5.60	5.62	5.66	69813.22656	2.110457			5
5.89	5.92	5.95	42007.19531	1.8831			6
Height Summation:			**247304.632814**				
Amount Avg CF:			**1.902549**	Linear:			
Aroclor-1262					6	52.76	
4.84	4.87	4.90	49712.05468	2.662638			1
5.36	5.39	5.42	25710.07617	1.105391			2
5.60	5.62	5.66	69813.22656	1.630902			3
5.84	5.87	5.90	7578.708496	0.490281			4
5.89	5.92	5.95	42007.19531	1.549162			5
6.35	6.38	6.41	17524.22460	0.993757			6
Height Summation:			**212345.485841**				
Amount Avg CF:			**1.405355**	Linear:			
Aroclor-1268					6	100.98	
5.35	5.39	5.41	25710.07617	1.837193			1
5.84	5.87	5.90	7578.708496	0.169048			2
5.90	5.92	5.96	42007.19531	1.068399			3
6.06	6.09	6.12	552.477356	0.012873			4
6.14	6.17	6.20	1701.240479	0.169212			5
6.36	6.38	6.42	17524.22460	0.981027			6
Height Summation:			**95073.922425**				
Amount Avg CF:			**0.706292**	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			0.4854	0.0971		** 144.89	4	15	
Aroclor-1221			0.4854	0.0971		** 53.75	3	5	
Aroclor-1232			0.4854	0.0971		** 132.82	4	10	
Aroclor-1242			0.4854	0.0971		** 111.58	4	30	
Aroclor-1248			0.4854	0.0971		** 165.45	4	30	
Aroclor-1254			0.4854	0.0971		** 108.38	4	40	
Aroclor-1260			0.4854	0.0971		9.77	4	40	
Aroclor-1262			0	0		4.43	4	40	
Aroclor-1268			0	0		20.40	4	40	

Units: ug/l

ASX22 8172

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:23:38

Sample Name: **5822078AF ABBB--1 T 093060018A 00173**
Acquired from CP07--H5088A via port 3 on 11/6/09 02:13:42am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.51R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.545		0.0000	0.000	4339.7	0.294	BV	0.044	1643.29	0.192
2	1.605		0.0000	0.000	1900.4	0.129	VB	0.031	1007.31	0.117
3	1.654		0.0000	0.000	3827.1	0.259	BV	0.024	2611.17	0.305
4	1.712		0.0000	0.000	5298.2	0.359	VV	0.051	1739.36	0.203
5	1.762		0.0000	0.000	4918.2	0.333	VB	0.034	2419.01	0.282
6	1.839		0.0000	0.000	11180.5	0.758	BV	0.042	4465.45	0.521
7	1.891		0.0000	0.000	9202.7	0.624	VV	0.039	3910.52	0.456
8	1.986		0.0000	0.000	22741.9	1.542	VV	0.096	3966.05	0.463
9	2.072		0.0000	0.000	14644.0	0.993	VV	0.065	3740.63	0.436
10	2.179		0.0000	0.000	14179.7	0.961	VB	0.060	3907.91	0.456
11	2.277		0.0000	0.000	33352.6	2.261	BV	0.036	15615.58	1.821
12	2.393		0.0000	0.000	4067.4	0.276	VV	0.041	1657.89	0.193
13	2.432		0.0000	0.000	4015.8	0.272	VV	0.043	1570.37	0.183
14	2.489		0.0000	0.000	3888.3	0.264	VB	0.047	1372.70	0.160
15	2.571		0.0000	0.000	4442.2	0.301	BV	0.042	1778.28	0.207
16	2.624		0.0000	0.000	1153.5	0.078	VB	0.028	697.47	0.081
17	2.701		0.0000	0.000	730.6	0.050	BB	0.027	445.90	0.052
18	2.739		0.0000	0.000	959.2	0.065	BB	0.022	724.75	0.085
19	2.774		0.0000	0.000	303.3	0.021	BB	0.018	279.18	0.033
20	2.852		0.0000	0.000	27118.0	1.839	BB	0.032	13931.69	1.625
21	2.944		0.0000	0.000	555.4	0.038	BB	0.025	366.88	0.043
22	2.998	TCX	237.5113	56.796	182002.4	12.341	BB	0.026	116440.10	13.579
23	3.121		0.0000	0.000	695.1	0.047	BB	0.024	484.00	0.056
24	3.193		0.0000	0.000	1855.3	0.126	BB	0.044	698.32	0.081
25	3.284		0.0000	0.000	3233.4	0.219	BB	0.032	1709.77	0.199
26	3.344		0.0000	0.000	8980.7	0.609	BB	0.027	5615.94	0.655
27	3.448		0.0000	0.000	2458.7	0.167	BB	0.028	1472.99	0.172
28	3.491		0.0000	0.000	855.0	0.058	BB	0.032	446.42	0.052
29	3.619		0.0000	0.000	488.7	0.033	BB	0.032	257.03	0.030
30	3.654		0.0000	0.000	197.0	0.013	BB	0.020	161.30	0.019
31	3.714		0.0000	0.000	1971.0	0.134	BB	0.040	830.16	0.097
32	3.784		0.0000	0.000	3169.1	0.215	BB	0.028	1871.78	0.218
33	3.855		0.0000	0.000	577.4	0.039	BB	0.043	225.35	0.026
34	3.935		0.0000	0.000	1568.6	0.106	BB	0.036	735.86	0.086
35	4.001		0.0000	0.000	764.1	0.052	BB	0.029	439.65	0.051
36	4.047		0.0000	0.000	2120.2	0.144	BB	0.025	1432.69	0.167
37	4.095		0.0000	0.000	15711.9	1.065	BB	0.020	12906.98	1.505
38	4.129		0.0000	0.000	1036.4	0.070	BB	0.014	1208.79	0.141
39	4.186		0.0000	0.000	3076.4	0.209	BB	0.026	1994.92	0.233
40	4.276		0.0000	0.000	1810.4	0.123	BB	0.038	802.55	0.094
41	4.321		0.0000	0.000	378.1	0.026	BB	0.023	277.26	0.032
42	4.359		0.0000	0.000	3884.5	0.263	BB	0.021	3109.58	0.363
43	4.393		0.0000	0.000	3561.0	0.241	BB	0.018	3324.80	0.388
44	4.422		0.0000	0.000	4463.2	0.303	BB	0.019	3856.96	0.450
45	4.527		0.0000	0.000	2490.1	0.169	BB	0.021	1961.64	0.229
46	4.556		0.0000	0.000	1428.9	0.097	BB	0.018	1345.27	0.157
47	4.598		0.0000	0.000	3330.8	0.226	BB	0.021	2601.19	0.303
48	4.647		0.0000	0.000	3078.7	0.209	BB	0.025	2039.50	0.238
49	4.700		0.0000	0.000	514.5	0.035	BB	0.021	414.57	0.048
50	4.736		0.0000	0.000	965.1	0.065	BB	0.034	468.72	0.055
51	4.807		0.0000	0.000	977.7	0.066	BB	0.024	693.05	0.081
52	4.855		0.0000	0.000	11513.5	0.781	BB	0.019	10061.87	1.173
53	4.884		0.0000	0.000	3482.6	0.236	BB	0.014	4204.96	0.490
54	4.912		0.0000	0.000	12528.5	0.850	BB	0.019	11003.75	1.283

ASX22 0173

5822078AF ABBB--1 T 093060018A 00173

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.000		0.0000	0.000	5256.1	0.356	BB	0.021	4216.35	0.492
56	5.055		0.0000	0.000	31141.4	2.112	BB	0.024	21221.20	2.475
57	5.096		0.0000	0.000	13391.0	0.908	BB	0.020	11139.63	1.299
58	5.175		0.0000	0.000	73337.7	4.973	BB	0.021	59230.66	6.908
59	5.229		0.0000	0.000	9779.5	0.663	BB	0.023	7172.69	0.836
60	5.301		0.0000	0.000	51456.6	3.489	BB	0.032	26462.61	3.086
61	5.372		0.0000	0.000	38219.8	2.592	BB	0.024	26775.25	3.123
62	5.426		0.0000	0.000	10528.3	0.714	BB	0.025	6915.84	0.807
63	5.472		0.0000	0.000	4235.7	0.287	BB	0.018	3856.53	0.450
64	5.519		0.0000	0.000	38886.1	2.637	BB	0.027	24371.57	2.842
65	5.586		0.0000	0.000	3137.0	0.213	BB	0.019	2786.27	0.325
66	5.636		0.0000	0.000	53215.3	3.608	BB	0.027	33295.62	3.883
67	5.701		0.0000	0.000	8437.6	0.572	BB	0.024	5834.00	0.680
68	5.766		0.0000	0.000	95739.5	6.492	BB	0.035	45702.45	5.330
69	5.874		0.0000	0.000	52937.7	3.589	BB	0.027	32386.63	3.777
70	5.963		0.0000	0.000	3723.2	0.252	BB	0.024	2611.88	0.305
71	6.010		0.0000	0.000	83614.8	5.670	BB	0.029	47427.97	5.531
72	6.081		0.0000	0.000	20475.3	1.388	BB	0.022	15182.58	1.771
73	6.134		0.0000	0.000	15488.2	1.050	BB	0.025	10484.68	1.223
74	6.276		0.0000	0.000	54131.4	3.670	BB	0.040	22642.19	2.641
75	6.350		0.0000	0.000	12087.2	0.820	BB	0.024	8287.55	0.967
76	6.397		0.0000	0.000	6594.7	0.447	BB	0.023	4811.02	0.561
77	6.444		0.0000	0.000	422.2	0.029	BB	0.017	425.68	0.050
78	6.510		0.0000	0.000	16528.3	1.121	BB	0.027	10243.50	1.195
79	6.576		0.0000	0.000	78442.8	5.319	BB	0.027	47588.98	5.550
80	6.728		0.0000	0.000	5183.7	0.351	BB	0.026	3352.67	0.391
81	6.831		0.0000	0.000	39632.8	2.687	BB	0.028	23506.33	2.741
82	6.888		0.0000	0.000	19611.6	1.330	BB	0.037	8797.54	1.026
83	7.054		0.0000	0.000	2650.5	0.180	BB	0.043	1026.64	0.120
84	7.180		0.0000	0.000	14234.5	0.965	BB	0.041	5722.34	0.667
85	7.414		0.0000	0.000	23585.1	1.599	BB	0.032	12154.16	1.417
86	7.509		0.0000	0.000	4488.2	0.304	BB	0.052	1433.51	0.167
87	7.729		0.0000	0.000	6838.7	0.464	BB	0.034	3349.31	0.391
88	7.820		0.0000	0.000	11706.1	0.794	BB	0.031	6250.94	0.729
89	7.975	DCB	180.6748	43.204	98336.5	6.668	BB	0.036	46122.29	5.379
90	8.287		0.0000	0.000	2334.1	0.158	BB	0.165	236.45	0.028
91	8.559		0.0000	0.000	3385.8	0.230	BB	0.081	699.71	0.082
92	8.662		0.0000	0.000	865.4	0.059	BB	0.032	445.19	0.052
93	9.099		0.0000	0.000	2754.6	0.187	BB	0.128	357.44	0.042

Total Area = 1474802.0, Total Amount = 418.186, Total Height = 857477.0, Sample Units = ug/l

ASX22 8174

Sample Name: **5822078AF ABBB--1 T 093060018A 00173**
Acquired from CP07--H5088B via port 4 on 11/6/09 02:13:42am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.51R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.620		0.0000	0.000	8229.2	0.340	BB	0.051	2689.70	0.209
2	1.772		0.0000	0.000	9343.5	0.386	BV	0.055	2820.87	0.219
3	1.813		0.0000	0.000	20726.4	0.857	VB	0.034	10231.92	0.794
4	2.001		0.0000	0.000	334.0	0.014	BB	0.099	56.05	0.004
5	2.044		0.0000	0.000	3382.3	0.140	BB	0.038	1480.80	0.115
6	2.112		0.0000	0.000	555.9	0.023	BB	0.021	448.82	0.035
7	2.164		0.0000	0.000	1641.1	0.068	BB	0.038	725.13	0.056
8	2.261		0.0000	0.000	566.0	0.023	BB	0.125	75.35	0.006
9	2.339	TCX	246.1295	51.955	186938.0	7.730	BB	0.029	106742.40	8.282
10	2.393		0.0000	0.000	25101.4	1.038	BB	0.024	17095.20	1.326
11	2.620		0.0000	0.000	2230.3	0.092	BB	0.033	1134.36	0.088
12	2.663		0.0000	0.000	404.8	0.017	BB	0.025	272.89	0.021
13	2.741		0.0000	0.000	18926.7	0.783	BB	0.031	10190.09	0.791
14	2.853		0.0000	0.000	336.9	0.014	BB	0.026	213.48	0.017
15	2.901		0.0000	0.000	426.4	0.018	BB	0.027	265.89	0.021
16	3.008		0.0000	0.000	3262.1	0.135	BB	0.045	1206.91	0.094
17	3.085		0.0000	0.000	1154.9	0.048	BB	0.039	496.60	0.039
18	3.163		0.0000	0.000	2735.8	0.113	BB	0.053	867.26	0.067
19	3.221		0.0000	0.000	404.1	0.017	BB	0.024	286.21	0.022
20	3.269		0.0000	0.000	3037.2	0.126	BB	0.026	1935.27	0.150
21	3.325		0.0000	0.000	2082.6	0.086	BB	0.034	1007.92	0.078
22	3.459		0.0000	0.000	19236.4	0.795	BB	0.068	4692.80	0.364
23	3.561		0.0000	0.000	125488.0	5.189	BB	0.028	74282.83	5.763
24	3.663		0.0000	0.000	19924.7	0.824	BB	0.042	7848.16	0.609
25	3.798		0.0000	0.000	13344.9	0.552	BB	0.039	5762.89	0.447
26	3.853		0.0000	0.000	3817.6	0.158	BB	0.026	2401.52	0.186
27	3.923		0.0000	0.000	3287.7	0.136	BB	0.029	1862.49	0.145
28	3.986		0.0000	0.000	670.1	0.028	BB	0.047	238.96	0.019
29	4.036		0.0000	0.000	3866.9	0.160	BB	0.034	1881.78	0.146
30	4.110		0.0000	0.000	17881.3	0.739	BB	0.026	11352.58	0.881
31	4.140		0.0000	0.000	8008.3	0.331	BB	0.026	5131.42	0.398
32	4.201		0.0000	0.000	2885.8	0.119	BB	0.020	2465.69	0.191
33	4.248		0.0000	0.000	23876.1	0.987	BB	0.023	17481.80	1.356
34	4.286		0.0000	0.000	28280.5	1.169	BB	0.027	17562.04	1.363
35	4.387		0.0000	0.000	965.0	0.040	BB	0.025	655.81	0.051
36	4.431		0.0000	0.000	9332.3	0.386	BB	0.026	5967.07	0.463
37	4.496		0.0000	0.000	31358.8	1.297	BB	0.040	13049.62	1.012
38	4.536		0.0000	0.000	4561.2	0.189	BB	0.017	4478.78	0.347
39	4.598		0.0000	0.000	616522.2	25.495	BB	0.024	422321.00	32.766
40	4.690		0.0000	0.000	67357.7	2.785	BB	0.031	36347.70	2.820
41	4.788		0.0000	0.000	91449.3	3.782	BB	0.039	39181.48	3.040
42	4.870		0.0000	0.000	99685.5	4.122	BB	0.033	49712.05	3.857
43	4.982		0.0000	0.000	65523.5	2.710	BB	0.046	23714.38	1.840
44	5.042		0.0000	0.000	36104.0	1.493	BB	0.027	22485.82	1.745
45	5.093		0.0000	0.000	4882.8	0.202	BB	0.018	4540.29	0.352
46	5.139		0.0000	0.000	167554.7	6.929	BB	0.044	63131.28	4.898
47	5.217		0.0000	0.000	8732.9	0.361	BB	0.023	6442.24	0.500
48	5.333		0.0000	0.000	3680.7	0.152	BB	0.023	2624.76	0.204
49	5.389		0.0000	0.000	72043.3	2.979	BB	0.047	25710.08	1.995
50	5.471		0.0000	0.000	12527.6	0.518	BB	0.023	8896.02	0.690
51	5.507		0.0000	0.000	3631.0	0.150	BB	0.021	2910.79	0.226
52	5.558		0.0000	0.000	2336.7	0.097	BB	0.018	2164.13	0.168
53	5.623		0.0000	0.000	163313.7	6.753	BB	0.039	69813.23	5.416
54	5.794		0.0000	0.000	3323.6	0.137	BB	0.026	2116.40	0.164

ASX22 8175

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.868		0.0000	0.000	12548.4	0.519	BB	0.028	7578.71	0.588
56	5.918		0.0000	0.000	91929.9	3.802	BB	0.036	42007.20	3.259
57	6.092		0.0000	0.000	1320.0	0.055	BB	0.040	552.48	0.043
58	6.170		0.0000	0.000	2952.6	0.122	BB	0.029	1701.24	0.132
59	6.235		0.0000	0.000	14647.1	0.606	BB	0.034	7139.30	0.554
60	6.380		0.0000	0.000	37095.7	1.534	BB	0.035	17524.22	1.360
61	6.493		0.0000	0.000	806.3	0.033	BB	0.049	275.01	0.021
62	6.522		0.0000	0.000	829.1	0.034	BB	0.030	456.51	0.035
63	6.617		0.0000	0.000	13578.6	0.562	BB	0.041	5486.16	0.426
64	6.800	DCB	227.6089	48.045	186111.1	7.696	BB	0.039	78575.52	6.096
65	6.961		0.0000	0.000	686.2	0.028	BB	0.035	329.81	0.026
66	7.051		0.0000	0.000	2092.8	0.087	BB	0.038	919.23	0.071
67	7.336		0.0000	0.000	2611.4	0.108	BB	0.051	855.09	0.066
68	7.400		0.0000	0.000	1692.6	0.070	BB	0.029	986.05	0.077
69	7.450		0.0000	0.000	1647.7	0.068	BB	0.042	651.75	0.051
70	7.734		0.0000	0.000	2480.8	0.103	BB	0.047	870.58	0.068
71	7.816		0.0000	0.000	2344.9	0.097	BB	0.038	1042.05	0.081
72	8.133		0.0000	0.000	473.6	0.020	BB	0.052	152.16	0.012
73	8.266		0.0000	0.000	10279.9	0.425	BB	0.142	1207.58	0.094
74	8.834		0.0000	0.000	1898.0	0.078	BB	0.060	525.33	0.041
75	8.933		0.0000	0.000	2926.6	0.121	BB	0.081	599.65	0.047

Total Area = 2418229.0, Total Amount = 473.738, Total Height = 1288907.0, Sample Units = ug/l

ASX22 8176

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\3P7307.51R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 2:04:33 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.51R



Instrument ID: CP07–H5088B Injected On: 11/6/2009 2:04:33 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1030
Analyst: 2280

Dilution Factor: 10

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.998	116440	.23059	TCX	2.339	106742	.23896	TCX
7.975	46122	.17541	DCB	6.8	78576	.22098	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.51A
Area File: C:\CPWIN\DATA1\3P7307B.51A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 2:13:50 AM
File Reported On: 11/6/2009 at 2:14:02 AM

10X REPLOTS

ASX22 8178

C:\CPWIN\DATA1\3P7307.51R



C:\CPWIN\DATA1\3P7307B.51R



ASX22 0179

Multiple Component Data Summary

Sample Name: 5822079 AF EBW29 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1056 ml Total Volume: 10 ml Analyst: 2131 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on	Nov 06, 2009 02:43:40
Instrument	H5088A
Result file	3P7307.54R
Calibration file	1P7307
Method file	PESTD

%SSR(TCX)	99.4%	Conc:	0.289075
%SSR(DCB)	83.1%	Conc:	0.240011

Analysis Report (B)

Injected on	Nov 06, 2009 02:43:40
Instrument	H5088B
Result file	3P7307B.54R
Calibration file	1P7307B
Method file	PESTDB

%SSR(TCX)	99.2%	Conc:	0.288324
%SSR(DCB)	93.9%	Conc:	0.271119

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<0.4735	<0.0947			4	15	
Aroclor-1221			<0.4735	<0.0947			3	5	
Aroclor-1232			<0.4735	<0.0947			4	10	
Aroclor-1242			<0.4735	<0.0947			4	30	
Aroclor-1248			<0.4735	<0.0947			4	30	
Aroclor-1254			<0.4735	<0.0947			4	40	
Aroclor-1260			<0.4735	<0.0947			4	40	

Units: ug/l

%Difference = High - Low divided by the Average times 100

Reviewed By: [illegible] Date: 11-11-09

Verified By: [illegible] Date: 11/9/09

ASX22 0188

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822079** AF EBW29 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1056 ml Total Volume: 10 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 02:43:40
Instrument : CP07--H5088A
Result file : 3P7307.54R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

%SSR(TCX) : 99.4% Conc.: 0.289075
%SSR(DCB) : 83.1% Conc.: 0.240011

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					3	42.52	
3.83	3.86	3.89	366.034485	0.027333			2
4.15	4.20	4.21	228.609665	0.017911			4
4.24	4.26	4.30	137.427383	0.01137			5
Height Summation:			732.071533				
Amount Avg CF:			0.018871	Linear:			
Aroclor-1221					1		
3.28	3.29	3.34	2343.927979	0.506475			2
Height Summation:			2343.927979				
Amount Avg CF:			0.506475	Linear:			
Aroclor-1232					3	45.95	
3.83	3.86	3.89	366.034485	0.061431			2
4.15	4.20	4.21	228.609665	0.037425			4
4.23	4.26	4.29	137.427383	0.024244			5
Height Summation:			732.071533				
Amount Avg CF:			0.041033	Linear:			
Aroclor-1242					3	45.20	
3.83	3.86	3.89	366.034485	0.033451			2
4.15	4.20	4.21	228.609665	0.020942			4
4.23	4.26	4.29	137.427383	0.013256			5
Height Summation:			732.071533				
Amount Avg CF:			0.02255	Linear:			
Aroclor-1254					4	65.62	
5.27	5.32	5.33	50.380333	0.003611			3
5.34	5.38	5.40	333.933594	0.013347			4
5.85	5.89	5.91	104.013199	0.008054			5
5.98	6.03	6.04	59.06543	0.003453			6
Height Summation:			547.392556				
Amount Avg CF:			0.007116	Linear:			
Aroclor-1260					2	27.20	
5.98	6.03	6.04	59.06543	0.003221			2
6.80	6.83	6.86	73.941757	0.004755			6
Height Summation:			133.007187				
Amount Avg CF:			0.003988	Linear:			
Aroclor-1262					1		
6.79	6.83	6.85	73.941757	0.00332			6
Height Summation:			73.941757				
Amount Avg CF:			0.00332	Linear:			
Aroclor-1268					3	166.99	
6.79	6.83	6.85	73.941757	0.001675			3
6.87	6.90	6.93	78.644409	0.002062			4
7.11	7.15	7.17	1148.295776	0.152454			6
Height Summation:			1300.881942				
Amount Avg CF:			0.052064	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 02:43:40
Instrument : CP07--H5088B
Result file : 3P7307B.54R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET A linear

%SSR(TCX) : 99.2% Conc.: 0.288324
%SSR(DCB) : 93.9% Conc.: 0.271119

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					2	128.87	
2.98	3.02	3.04	32.087219	0.001775			2
3.14	3.15	3.20	475.828644	0.038211			3
Height Summation:			507.915863				
Amount Avg CF:			0.019993	Linear:			
Aroclor-1221					1		
2.61	2.62	2.67	1689.551147	0.56386			2
Height Summation:			1689.551147				
Amount Avg CF:			0.56386	Linear:			
Aroclor-1232					2	128.48	
2.98	3.02	3.04	32.087219	0.003915			2
3.14	3.15	3.20	475.828644	0.081674			3
Height Summation:			507.915863				
Amount Avg CF:			0.042795	Linear:			
Aroclor-1242					2	128.91	
2.98	3.02	3.04	32.087219	0.002178			2
3.14	3.15	3.20	475.828644	0.047064			3
Height Summation:			507.915863				
Amount Avg CF:			0.024621	Linear:			
Aroclor-1248					3	149.65	
3.14	3.15	3.20	475.828644	0.106906			1
4.26	4.31	4.32	189.937286	0.017514			5
4.55	4.60	4.61	8074.44873	1.228863			6
Height Summation:			8740.214660				
Amount Avg CF:			0.451094	Linear:			
Aroclor-1254					2	133.50	
4.08	4.12	4.14	208.635757	0.010585			1
4.55	4.60	4.61	8074.44873	0.367568			2
Height Summation:			8283.084487				
Amount Avg CF:			0.189077	Linear:			
Aroclor-1262					1		
6.35	6.40	6.41	43.827114	0.002424			6
Height Summation:			43.827114				
Amount Avg CF:			0.002424	Linear:			
Aroclor-1268					1		
6.36	6.40	6.42	43.827114	0.002393			6
Height Summation:			43.827114				
Amount Avg CF:			0.002393	Linear:			

Sample Clean

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Quatifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			0.4735	0.0947		5.77	4	15	
Aroclor-1221			0.4735	0.0947		10.72	3	5	
Aroclor-1232			0.4735	0.0947		4.20	4	10	

ASX22 0181

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:25:38

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822079** AF EBW29 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1056 ml Total Volume: 10 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 02:43:40
Instrument : CP07--H5088A
Result file : 3P7307.54R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

Analysis Report (B)

Injected on : NOV 06, 2009 02:43:40
Instrument : CP07--H5088B
Result file : 3P7307B.54R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1242			0.4735	0.0947		8.78	4	30	
Aroclor-1248			0.4735	0.0947			4	30	
Aroclor-1254			0.4735	0.0947		** 185.49	4	40	
Aroclor-1260			0.4735	0.0947			4	40	
Aroclor-1262			0	0		31.20	4	40	
Aroclor-1268			0	0		** 182.42	4	40	

Units: ug/L

ASX22 8582

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:25:38

Sample Name: **5822079AF ABEBW29 T 093060018A 00173**
Acquired from CP07--H5088A via port 3 on 11/6/09 02:52:48am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.54R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.075		0.0000	0.000	22846.0	4.918	BV	0.240	1587.63	0.720
2	2.176		0.0000	0.000	3042.6	0.655	VV	0.052	976.40	0.443
3	2.210		0.0000	0.000	2263.7	0.487	VV	0.042	889.58	0.403
4	2.287		0.0000	0.000	1464.1	0.315	VV	0.037	658.99	0.299
5	2.523		0.0000	0.000	502.4	0.108	VB	0.099	84.39	0.038
6	2.877		0.0000	0.000	1959.2	0.422	BB	0.198	165.18	0.075
7	2.999	TCX	305.2628	54.637	232118.5	49.968	BB	0.027	144103.70	65.322
8	3.288		0.0000	0.000	3722.4	0.801	BB	0.026	2343.93	1.062
9	3.472		0.0000	0.000	720.9	0.155	BB	0.030	394.74	0.179
10	3.678		0.0000	0.000	650.1	0.140	BB	0.071	152.00	0.069
11	3.799		0.0000	0.000	356.2	0.077	BB	0.056	105.38	0.048
12	3.856		0.0000	0.000	673.4	0.145	BB	0.031	366.03	0.166
13	3.939		0.0000	0.000	526.2	0.113	BB	0.025	350.64	0.159
14	4.197		0.0000	0.000	829.0	0.178	BB	0.060	228.61	0.104
15	4.265		0.0000	0.000	188.3	0.041	BB	0.023	137.43	0.062
16	4.440		0.0000	0.000	455.1	0.098	BB	0.079	96.46	0.044
17	4.805		0.0000	0.000	1232.9	0.265	BB	0.087	234.93	0.106
18	5.102		0.0000	0.000	831.2	0.179	BB	0.035	395.20	0.179
19	5.175		0.0000	0.000	1269.4	0.273	BB	0.020	1077.84	0.489
20	5.254		0.0000	0.000	269.4	0.058	BB	0.050	90.43	0.041
21	5.324		0.0000	0.000	130.4	0.028	BB	0.043	50.38	0.023
22	5.379		0.0000	0.000	2317.6	0.499	BB	0.116	333.93	0.151
23	5.807		0.0000	0.000	495.8	0.107	BB	0.115	71.87	0.033
24	5.891		0.0000	0.000	322.3	0.069	BB	0.052	104.01	0.047
25	6.025		0.0000	0.000	138.3	0.030	BB	0.039	59.07	0.027
26	6.422		0.0000	0.000	29408.1	6.331	BB	0.220	2228.49	1.010
27	6.830		0.0000	0.000	710.1	0.153	BB	0.160	73.94	0.034
28	6.899		0.0000	0.000	237.0	0.051	BB	0.050	78.64	0.036
29	6.969		0.0000	0.000	90.7	0.020	BB	0.033	45.35	0.021
30	7.154		0.0000	0.000	14359.6	3.091	BB	0.208	1148.30	0.521
31	7.592		0.0000	0.000	4181.7	0.900	BB	0.134	521.84	0.237
32	7.731		0.0000	0.000	1220.2	0.263	BB	0.034	590.09	0.267
33	7.976	DCB	253.4514	45.363	134999.6	29.061	BB	0.037	60861.38	27.588

Total Area = 464532.2, Total Amount = 558.714, Total Height = 220606.7, Sample Units = ug/l

ASX22 0183

Sample Name: **5822079AF ABEBW29 T 093060018A 00173**
Acquired from CP07--H5088B via port 4 on 11/6/09 02:52:48am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.54R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.587		0.0000	0.000	5050.2	0.811	BV	0.076	1113.38	0.450
2	1.706		0.0000	0.000	1271.2	0.204	VV	0.045	471.23	0.190
3	1.894		0.0000	0.000	1257.7	0.202	VB	0.642	32.62	0.013
4	1.964		0.0000	0.000	103.1	0.017	BB	0.054	31.92	0.013
5	2.034		0.0000	0.000	184.5	0.030	BB	0.086	35.73	0.014
6	2.301		0.0000	0.000	65.7	0.011	BB	0.051	21.36	0.009
7	2.340	TCX	304.4706	51.538	299899.6	48.187	BB	0.038	132044.00	53.346
8	2.621		0.0000	0.000	3545.8	0.570	BB	0.035	1689.55	0.683
9	2.775		0.0000	0.000	668.0	0.107	BB	0.045	247.31	0.100
10	3.020		0.0000	0.000	49.5	0.008	BB	0.026	32.09	0.013
11	3.097		0.0000	0.000	231.3	0.037	BB	0.056	68.73	0.028
12	3.153		0.0000	0.000	1523.4	0.245	BB	0.053	475.83	0.192
13	3.264		0.0000	0.000	2386.6	0.383	BB	0.150	265.89	0.107
14	4.124		0.0000	0.000	1523.3	0.245	BB	0.122	208.64	0.084
15	4.307		0.0000	0.000	680.8	0.109	BB	0.060	189.94	0.077
16	4.599		0.0000	0.000	13388.4	2.151	BB	0.028	8074.45	3.262
17	5.230		0.0000	0.000	31017.6	4.984	BB	0.522	991.01	0.400
18	5.797		0.0000	0.000	3314.1	0.532	BB	0.113	488.94	0.198
19	6.287		0.0000	0.000	358.5	0.058	BB	0.030	199.23	0.080
20	6.400		0.0000	0.000	70.7	0.011	BB	0.027	43.83	0.018
21	6.560		0.0000	0.000	249.8	0.040	BB	0.152	27.40	0.011
22	6.622		0.0000	0.000	3227.2	0.519	BB	0.049	1103.38	0.446
23	6.799	DCB	286.3017	48.462	222107.6	35.687	BB	0.037	98837.54	39.931
24	7.613		0.0000	0.000	29202.0	4.692	BB	0.837	581.67	0.235
25	8.987		0.0000	0.000	993.8	0.160	BB	0.067	247.26	0.100

Total Area = 622370.4, Total Amount = 590.772, Total Height = 247522.9, Sample Units = ug/l

ASX22 8184

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\3P7307.54R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 2:43:39 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.54R



Instrument ID: CP07--H5088B Injected On: 11/6/2009 2:43:39 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1056
Analyst: 2280

Dilution Factor: 10

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	144104	.28907	TCX	2.34	132044	.28832	TCX
7.976	60861	.24001	DCB	6.799	98838	.27112	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.54A
Area File: C:\CPWIN\DATA1\3P7307B.54A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 2:52:56 AM
File Reported On: 11/6/2009 at 2:53:08 AM

10X REPLOTS

ASX22 8186

C:\CPWIN\DATA1\3P7307.54R



C:\CPWIN\DATA1\3P7307B.54R



ASX22 0187

Multiple Component Data Summary

Sample Name: 5822080 RI AF EBS29 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1044 ml Total Volume: 10 ml Analyst: 2131 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on	Nov 06, 2009 02:56:41
Instrument	H5088A
Result file	3P7307.55R
Calibration file	1P7307
Method file	PESTD

%SSR(TCX)	103.6%	Conc:	0.304659
%SSR(DCB)	85.1%	Conc:	0.248709

Analysis Report (B)

Injected on	Nov 06, 2009 02:56:41
Instrument	H5088B
Result file	3P7307B.55R
Calibration file	1P7307B
Method file	PESTDB

%SSR(TCX)	102.6%	Conc:	0.301642
%SSR(DCB)	97.3%	Conc:	0.28422

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<0.4789	<0.0958			4	15	
Aroclor-1221			<0.4789	<0.0958			3	5	
Aroclor-1232			<0.4789	<0.0958			4	10	
Aroclor-1242			<0.4789	<0.0958			4	30	
Aroclor-1248			<0.4789	<0.0958			4	30	
Aroclor-1254			<0.4789	<0.0958			4	40	
Aroclor-1260			<0.4789	<0.0958			4	40	

Units: ug/l

%Difference = High - Low divided by the Average times 100

Reviewed By: [illegible] Date: 11-11-09

Verified By: [illegible] Date: 11/9/09

ASX22 0188

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822080** RI AF EBS29 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1044 ml Total Volume: 10 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 02:56:41
Instrument : CP07–H5088A
Result file : 3P7307.55R
Calibration file : 1P7307.CAL
Method file : PESTD.MET
%SSR(TCX) : 103.6% Conc.: 0.304659
%SSR(DCB) : 85.1% Conc.: 0.248709

BOOS sample clean

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					4	40.84	
3.35	3.37	3.41	138.02034	0.012641			1
3.83	3.86	3.89	525.782654	0.039714			2
4.24	4.26	4.30	441.628448	0.036957			5
4.62	4.67	4.68	254.108505	0.029837			6
Height Summation:			1359.539947				
Amount Avg CF:			0.029787	Linear:			
Aroclor-1221					3	158.94	
3.20	3.21	3.26	152.566025	0.023523			1
3.28	3.29	3.34	2560.599609	0.559654			2
3.35	3.37	3.41	138.02034	0.009095			3
Height Summation:			2851.185974				
Amount Avg CF:			0.197424	Linear:			
Aroclor-1232					4	56.71	
3.35	3.37	3.41	138.02034	0.01077			1
3.83	3.86	3.89	525.782654	0.089255			2
4.23	4.26	4.29	441.628448	0.078803			5
4.62	4.67	4.68	254.108505	0.068708			6
Height Summation:			1359.539947				
Amount Avg CF:			0.061884	Linear:			
Aroclor-1242					4	42.22	
3.35	3.37	3.41	138.02034	0.014784			1
3.83	3.86	3.89	525.782654	0.048603			2
4.23	4.26	4.29	441.628448	0.043088			5
4.62	4.67	4.68	254.108505	0.034049			6
Height Summation:			1359.539947				
Amount Avg CF:			0.035131	Linear:			
Aroclor-1248					1		
4.62	4.67	4.68	254.108505	0.023576			4
Height Summation:			254.108505				
Amount Avg CF:			0.023576	Linear:			
Aroclor-1254					4	141.69	
4.83	4.87	4.89	16195.25878!	1.292168			1
5.27	5.30	5.33	908.762695	0.065889			3
5.34	5.38	5.40	7315.569336	0.295746			4
5.85	5.90	5.91	298.169617	0.023354			5
Height Summation:			24717.760437				
Amount Avg CF:			0.419289	Linear:			
Aroclor-1260					2	124.12	
6.05	6.09	6.11	2074.114014	0.144604			3
6.80	6.83	6.86	144.876312	0.009423			6
Height Summation:			2218.990326				
Amount Avg CF:			0.077014	Linear:			
Aroclor-1262					2	121.45	
6.05	6.09	6.11	2074.114014	0.086618			3
6.79	6.83	6.85	144.876312	0.006579			6
Height Summation:			2218.990326				
Amount Avg CF:			0.046599	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 02:56:41
Instrument : CP07–H5088B
Result file : 3P7307B.55R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET
%SSR(TCX) : 102.6% Conc.: 0.301642
%SSR(DCB) : 97.3% Conc.: 0.28422

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					3	88.55	
2.69	2.71	2.75	99.409958	0.012346			1
2.98	3.01	3.04	149.792786	0.008381			2
3.14	3.15	3.20	523.327454	0.042508			3
Height Summation:			772.530198				
Amount Avg CF:			0.021078	Linear:			
Aroclor-1221					2	137.16	
2.61	2.62	2.67	1727.91748	0.583293			2
2.69	2.71	2.75	99.409958	0.008915			3
Height Summation:			1827.327438				
Amount Avg CF:			0.296104	Linear:			
Aroclor-1232					3	110.32	
2.69	2.71	2.75	99.409958	0.010788			1
2.98	3.01	3.04	149.792786	0.018486			2
3.14	3.15	3.20	523.327454	0.090859			3
Height Summation:			772.530198				
Amount Avg CF:			0.040044	Linear:			
Aroclor-1242					3	89.32	
2.68	2.71	2.74	99.409958	0.014872			1
2.98	3.01	3.04	149.792786	0.010287			2
3.14	3.15	3.20	523.327454	0.052358			3
Height Summation:			772.530198				
Amount Avg CF:			0.025839	Linear:			
Aroclor-1248					3	138.23	
3.14	3.15	3.20	523.327454	0.118929			1
4.26	4.30	4.32	748.217346	0.069786			5
4.55	4.60	4.61	7853.223145	1.208933			6
Height Summation:			9124.767945				
Amount Avg CF:			0.465883	Linear:			
Aroclor-1254					2	39.88	
4.08	4.13	4.14	12581.56152:	0.645684			1
4.55	4.60	4.61	7853.223145	0.361606			2
Height Summation:			20434.784668				
Amount Avg CF:			0.503645	Linear:			
Aroclor-1260					1		
4.96	5.00	5.02	2290.588379	0.149291			3
Height Summation:			2290.588379				
Amount Avg CF:			0.149291	Linear:			

no pattern

ASX22 0189

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:26:27

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822080** RI AF EBS29 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1044 ml Total Volume: 10 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on : NOV 06, 2009 02:56:41
Instrument : CP07--H5088A
Result file : 3P7307.55R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1268					**3**	**88.28**	
6.05	6.09	6.11	2074.114014	0.218911			1
6.79	6.83	6.85	144.876312	0.003319			3
7.11	7.15	7.17	1141.72583	0.153324			6

Height Summation: **3360.716156**
Amount Avg CF: **0.125185** Linear:

Analysis Report (B)

Injected on : NOV 06, 2009 02:56:41
Instrument : CP07--H5088B
Result file : 3P7307B.55R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			0.4789	0.0958		34.24	4	15	
Aroclor-1221			0.4789	0.0958		39.99	3	5	
Aroclor-1232			0.4789	0.0958		** 42.85	4	10	
Aroclor-1242			0.4789	0.0958		30.48	4	30	
Aroclor-1248			0.4789	0.0958		** 180.73	4	30	
Aroclor-1254			0.4789	0.0958		18.28	4	40	
Aroclor-1260			0.4789	0.0958		** 63.88	4	40	
Aroclor-1262			0	0			4	40	
Aroclor-1268			0	0			4	40	

Units: ug/L

ASX22 0198

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:26:27

Sample Name: **5822080 RI AF ABEBS29 T 093060018A 00173**
Acquired from CP07--H5088A via port 3 on 11/6/09 03:05:49am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.55R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.579		0.0000	0.000	4219.2	0.558	BB	0.047	1483.28	0.494
2	1.796		0.0000	0.000	22813.3	3.015	BV	0.190	2004.83	0.667
3	1.849		0.0000	0.000	7073.2	0.935	VV	0.061	1942.44	0.646
4	2.028		0.0000	0.000	22250.5	2.940	VV	0.101	3672.36	1.222
5	2.076		0.0000	0.000	6680.6	0.883	VV	0.042	2649.94	0.882
6	2.127		0.0000	0.000	4513.1	0.596	VV	0.039	1911.46	0.636
7	2.175		0.0000	0.000	10630.3	1.405	VV	0.064	2748.49	0.915
8	2.288		0.0000	0.000	4202.8	0.555	VV	0.049	1424.44	0.474
9	2.340		0.0000	0.000	3654.5	0.483	VV	0.062	983.14	0.327
10	2.523		0.0000	0.000	3090.7	0.408	VV	0.134	383.79	0.128
11	2.576		0.0000	0.000	564.3	0.075	VB	0.049	190.30	0.063
12	2.703		0.0000	0.000	96.5	0.013	BB	0.029	55.35	0.018
13	2.880		0.0000	0.000	1353.4	0.179	BB	0.230	98.25	0.033
14	2.912		0.0000	0.000	133.3	0.018	BB	0.024	91.52	0.030
15	2.999	TCX	318.0642	55.055	244811.6	32.350	BB	0.027	149330.70	49.690
16	3.114		0.0000	0.000	120.7	0.016	BB	0.017	118.14	0.039
17	3.214		0.0000	0.000	226.6	0.030	BB	0.025	152.57	0.051
18	3.288		0.0000	0.000	4407.4	0.582	BB	0.029	2560.60	0.852
19	3.366		0.0000	0.000	282.4	0.037	BB	0.034	138.02	0.046
20	3.473		0.0000	0.000	860.8	0.114	BB	0.033	429.40	0.143
21	3.510		0.0000	0.000	88.6	0.012	BB	0.016	94.04	0.031
22	3.562		0.0000	0.000	347.7	0.046	BB	0.035	163.74	0.054
23	3.639		0.0000	0.000	1165.7	0.154	BB	0.043	447.20	0.149
24	3.796		0.0000	0.000	606.3	0.080	BB	0.051	199.11	0.066
25	3.858		0.0000	0.000	1151.1	0.152	BB	0.036	525.78	0.175
26	3.938		0.0000	0.000	792.6	0.105	BB	0.026	506.64	0.169
27	4.032		0.0000	0.000	224.0	0.030	BB	0.036	104.43	0.035
28	4.085		0.0000	0.000	580.6	0.077	BB	0.039	249.16	0.083
29	4.264		0.0000	0.000	1655.2	0.219	BB	0.062	441.63	0.147
30	4.316		0.0000	0.000	617.6	0.082	BB	0.033	312.52	0.104
31	4.435		0.0000	0.000	11636.6	1.538	BB	0.030	6387.25	2.125
32	4.598		0.0000	0.000	149.8	0.020	BB	0.025	99.69	0.033
33	4.675		0.0000	0.000	713.1	0.094	BB	0.047	254.11	0.085
34	4.739		0.0000	0.000	2461.9	0.325	BB	0.022	1848.05	0.615
35	4.866		0.0000	0.000	39700.3	5.246	BB	0.041	16195.26	5.389
36	4.913		0.0000	0.000	3869.9	0.511	BB	0.020	3290.01	1.095
37	5.009		0.0000	0.000	10465.5	1.383	BB	0.104	1671.47	0.556
38	5.101		0.0000	0.000	1551.4	0.205	BB	0.039	664.18	0.221
39	5.176		0.0000	0.000	1332.4	0.176	BB	0.020	1129.75	0.376
40	5.235		0.0000	0.000	1771.0	0.234	BB	0.030	972.87	0.324
41	5.305		0.0000	0.000	3343.4	0.442	BB	0.061	908.76	0.302
42	5.381		0.0000	0.000	29537.6	3.903	BB	0.067	7315.57	2.434
43	5.582		0.0000	0.000	20734.6	2.740	BB	0.143	2417.82	0.805
44	5.822		0.0000	0.000	10853.7	1.434	BB	0.289	626.24	0.208
45	5.904		0.0000	0.000	1071.2	0.142	BB	0.060	298.17	0.099
46	6.092		0.0000	0.000	5350.5	0.707	BB	0.043	2074.11	0.690
47	6.392		0.0000	0.000	100024.0	13.218	BB	0.139	11974.45	3.985
48	6.728		0.0000	0.000	156.4	0.021	BB	0.029	88.59	0.029
49	6.835		0.0000	0.000	566.0	0.075	BB	0.065	144.88	0.048
50	7.153		0.0000	0.000	13118.3	1.733	BB	0.191	1141.73	0.380
51	7.379		0.0000	0.000	665.2	0.088	BB	0.027	411.80	0.137
52	7.591		0.0000	0.000	4147.0	0.548	BB	0.135	511.00	0.170
53	7.731		0.0000	0.000	1409.0	0.186	BB	0.037	640.03	0.213
54	7.975	DCB	259.6523	44.945	137217.5	18.132	BB	0.037	62117.22	20.670

ASX22 0191

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	8.246		0.0000	0.000	494.5	0.065	BB	0.043	193.36	0.064
56	8.317		0.0000	0.000	104.8	0.014	BB	0.028	63.50	0.021
57	8.374		0.0000	0.000	3897.9	0.515	BB	0.044	1478.74	0.492
58	8.610		0.0000	0.000	1195.8	0.158	BB	0.104	192.48	0.064

Total Area = 756753.8, Total Amount = 577.717, Total Height = 300524.3, Sample Units = ug/l

ASX22 0192

Sample Name: **5822080 RI AF ABEBS29 T 093060018A 00173**
Acquired from CP07--H5088B via port 4 on 11/6/09 03:05:49am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.55R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.605		0.0000	0.000	12974.5	1.514	BV	0.087	2497.67	0.789
2	1.767		0.0000	0.000	869.0	0.101	VV	0.036	404.39	0.128
3	1.809		0.0000	0.000	322.9	0.038	VV	0.027	196.66	0.062
4	1.893		0.0000	0.000	144.4	0.017	VB	0.042	56.69	0.018
5	1.942		0.0000	0.000	484.4	0.057	BB	0.051	158.36	0.050
6	2.060		0.0000	0.000	165.4	0.019	BB	0.077	35.60	0.011
7	2.301		0.0000	0.000	77.9	0.009	BB	0.045	28.71	0.009
8	2.341	TCX	314.9147	51.487	303109.0	35.375	BB	0.037	136573.40	43.116
9	2.494		0.0000	0.000	5587.2	0.652	BB	0.045	2046.88	0.646
10	2.622		0.0000	0.000	3429.3	0.400	BB	0.033	1727.92	0.546
11	2.713		0.0000	0.000	145.6	0.017	BB	0.024	99.41	0.031
12	2.776		0.0000	0.000	544.8	0.064	BB	0.039	231.89	0.073
13	2.866		0.0000	0.000	461.7	0.054	BB	0.034	223.37	0.071
14	3.007		0.0000	0.000	364.6	0.043	BB	0.041	149.79	0.047
15	3.153		0.0000	0.000	2111.4	0.246	BB	0.067	523.33	0.165
16	3.273		0.0000	0.000	517.6	0.060	BB	0.037	234.00	0.074
17	3.350		0.0000	0.000	2237.3	0.261	BB	0.201	185.22	0.058
18	3.660		0.0000	0.000	4660.6	0.544	BB	0.180	431.46	0.136
19	3.835		0.0000	0.000	2766.0	0.323	BB	0.434	106.29	0.034
20	3.880		0.0000	0.000	14262.9	1.665	BB	0.034	6976.21	2.202
21	4.051		0.0000	0.000	4858.2	0.567	BB	0.833	97.25	0.031
22	4.132		0.0000	0.000	33945.4	3.962	BB	0.045	12581.56	3.972
23	4.296		0.0000	0.000	10479.3	1.223	BB	0.233	748.22	0.236
24	4.540		0.0000	0.000	20873.7	2.436	BB	0.107	3258.99	1.029
25	4.600		0.0000	0.000	22933.7	2.676	BB	0.049	7853.22	2.479
26	4.826		0.0000	0.000	9773.0	1.141	BB	0.038	4315.14	1.362
27	4.997		0.0000	0.000	14970.5	1.747	BB	0.109	2290.59	0.723
28	5.247		0.0000	0.000	5564.6	0.649	BB	1.065	87.07	0.027
29	5.269		0.0000	0.000	1650.1	0.193	BB	0.251	109.44	0.035
30	5.710		0.0000	0.000	544.0	0.063	BB	0.040	225.10	0.071
31	5.798		0.0000	0.000	79325.4	9.258	BB	0.053	25100.84	7.924
32	6.304		0.0000	0.000	1805.7	0.211	BB	0.048	624.19	0.197
33	6.527		0.0000	0.000	418.3	0.049	BB	0.038	185.08	0.058
34	6.621		0.0000	0.000	2400.4	0.280	BB	0.040	996.15	0.314
35	6.799	DCB	296.7258	48.513	234141.3	27.326	BB	0.038	102436.20	32.339
36	7.250		0.0000	0.000	236.2	0.028	BB	0.033	119.42	0.038
37	7.353		0.0000	0.000	10144.2	1.184	BB	0.129	1308.98	0.413
38	7.726		0.0000	0.000	32371.9	3.778	BB	7.022	76.83	0.024
39	8.510		0.0000	0.000	6559.0	0.765	BB	0.113	963.15	0.304
40	8.713		0.0000	0.000	8624.2	1.006	BB	0.293	491.27	0.155

Total Area = 856855.7, Total Amount = 611.641, Total Height = 316755.9, Sample Units = ug/l

ASX22 P193

5822080 RI AF ABEBS29 T 093060018A 00173

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\3P7307.55R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 2:56:40 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.55R



ASX22 8194

Instrument ID: CP07--H5088B Injected On: 11/6/2009 2:56:40 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1044
Analyst: 2280

Dilution Factor: 10

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	149331	.30466	TCX	2.341	136573	.30164	TCX
7.975	62117	.24871	DCB	6.799	102436	.28422	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.55A
Area File: C:\CPWIN\DATA1\3P7307B.55A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 3:05:56 AM
File Reported On: 11/6/2009 at 3:06:09 AM

10X REPLOTS

ASX22 8195

C:\CPWIN\DATA1\3P7307.55R



C:\CPWIN\DATA1\3P7307B.55R



ASX22 8196

Standards Data

6D
INITIAL CALIBRATION - RETENTION TIME SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: H5088A
Calibration File: 1P7307
GC Column (1) : RTXCLPI
ID: 0.32 (mm)
Update File:
Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	RT OF STANDARDS LEVEL 1	LEVEL 2	LEVEL 3	LEVEL 4	LEVEL 5	MIDPOINT RT	RT WINDOW FROM	TO
Tetrachloro-m-xylene	3.00	3.00	3.00	3.00	3.00	3.00	2.97	3.03
Decachlorobiphenyl	7.98	7.98	7.98	7.98	7.98	7.98	7.95	8.01

ASX22 0198

FORM VI-1

6D
INITIAL CALIBRATION - RETENTION TIME SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: H5088B
Calibration File: 1P7307B
GC Column (2) : RTXCLPII
ID: 0.32 (mm)
Update File:
Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	RT OF STANDARDS					MIDPOINT	RT WINDOW	
	LEVEL 1	LEVEL 2	LEVEL 3	LEVEL 4	LEVEL 5	RT	FROM	TO
Tetrachloro-m-xylene	2.34	2.34	2.34	2.34	2.34	2.34	2.31	2.37
Decachlorobiphenyl	6.81	6.80	6.80	6.80	6.80	6.80	6.77	6.83

ASX22 0199

FORM VI-1

6E

INITIAL CALIBRATION - CALIBRATION FACTOR SUMMARY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088A Calibration File: 1P7307

GC Column (1): RTXCLPI ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	CALIBRATION FACTORS LEVEL 1	LEVEL 2	LEVEL 3	LEVEL 4	LEVEL 5	MEAN	%RSD	
Tetrachloro-m-xylene	7.04E+03	5.20E+03	4.71E+03	4.48E+03	4.23E+03	5.13E+03	21.9	Linear
Decachlorobiphenyl	3.38E+03	2.59E+03	2.36E+03	2.20E+03	2.10E+03	2.53E+03	20.3	Linear
						Average % RSD:	21.1	

[illegible] 11-4-09

ASX22 0288

FORM VI-2

Calibration File Name: **C:\CPWIN\DATA1\1P7307.CAL** Version = 8
External standard calibration
Standard injection volume = 1
No sample weight correction
Area reject threshold = 0
Reference peak area reject threshold = 0
Amount units = ug/l
2 components with 5 levels each

1 TCX

Retention time = 2.999 min., Search window = 0.030 min.
Low alarm amount = 0, High alarm amount = 0
Group number = 0, Component constant = 0
No retention time reference component
Single peak quantification by height

Level	Amount	Height	Height/Amt	Source	Date and time
1	4.094	28830.4	7042.111	1P7307.07A	11/3/2009 4:09:1
2	20.470	106490.3	5202.262	1P7307.08A	11/3/2009 4:22:1
3	40.940	192864.9	4710.917	1P7307.09A	11/3/2009 4:35:0
4	61.410	275302.2	4483.019	1P7307.10A	11/3/2009 4:48:0
5	81.880	346348.3	4229.95	1P7307.11A	11/3/2009 5:01:0



Calibration formula: Y = 4083.1 X + 19461.88
Fit type = Linear with equal weighting
Coefficient of determination = 0.9971, Average error = 6.20%
Average CF = 5133.6520 with RSD = 21.92%

2 DCB

Retention time = 7.978 min., Search window = 0.030 min.
Low alarm amount = 0, High alarm amount = 0
Group number = 0, Component constant = 0
No retention time reference component
Single peak quantification by height

Level	Amount	Height	Height/Amt	Source	Date and time
1	4.066	13762.4	3384.742	1P7307.07A	11/3/2009 4:09:1
2	20.330	52688.0	2591.637	1P7307.08A	11/3/2009 4:22:1
3	40.660	96000.0	2361.043	1P7307.09A	11/3/2009 4:35:0
4	60.990	134069.2	2198.216	1P7307.10A	11/3/2009 4:48:0
5	81.320	171104.4	2104.088	1P7307.11A	11/3/2009 5:01:0

ASX22 0201



Calibration formula: Y = 2025.254 X + 9531.046
Fit type = Linear with equal weighting
Coefficient of determination = 0.9970, Average error = 6.70%
Average CF = 2527.9450 with RSD = 20.31%

ASX22 8202

6E
INITIAL CALIBRATION - CALIBRATION FACTOR SUMMARY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088B Calibration File: 1P7307B

GC Column (2) : RTXCLPII ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	CALIBRATION FACTORS						%RSD
	LEVEL 1	LEVEL 2	LEVEL 3	LEVEL 4	LEVEL 5	MEAN	
Tetrachloro-m-xylene	4.98E+03	4.47E+03	4.28E+03	4.04E+03	3.91E+03	4.34E+03	9.7
Decachlorobiphenyl	4.46E+03	3.56E+03	3.29E+03	3.04E+03	2.91E+03	3.45E+03	17.8
						Average % RSD:	13.8

ASX22 0283

FORM VI-2

Calibration File Name: **C:\CPWIN\DATA1\1P7307B.CAL** Version = 8
External standard calibration
Standard injection volume = 1
No sample weight correction
Area reject threshold = 0
Reference peak area reject threshold = 0
Amount units = ug/l
2 components with 5 levels each

1 TCX

Retention time = 2.340 min., Search window = 0.030 min.
Low alarm amount = 0, High alarm amount = 0
Group number = 0, Component constant = 0
No retention time reference component
Single peak quantification by height

Level	Amount	Height	Height/Amt	Source	Date and time
1	4.094	20397.5	4982.282	1P7307B.07A	11/3/2009 4:09:3
2	20.470	91446.6	4467.346	1P7307B.08A	11/3/2009 4:22:2
3	40.940	175402.2	4284.373	1P7307B.09A	11/3/2009 4:35:2
4	61.410	248224.9	4042.093	1P7307B.10A	11/3/2009 4:48:2
5	81.880	319994.7	3908.094	1P7307B.11A	11/3/2009 5:01:2



Calibration formula: Y = 4336.837 X
Fit type = Avg CF with equal weighting, forced to origin
Coefficient of determination = 0.9724, Average error = 7.16%
Average CF = 4336.8370 with RSD = 9.69%

2 DCB

Retention time = 6.801 min., Search window = 0.030 min.
Low alarm amount = 0, High alarm amount = 0
Group number = 0, Component constant = 0
No retention time reference component
Single peak quantification by height

Level	Amount	Height	Height/Amt	Source	Date and time
1	4.066	18116.3	4455.556	1P7307B.07A	11/3/2009 4:09:3
2	20.330	72456.6	3564.024	1P7307B.08A	11/3/2009 4:22:2
3	40.660	133662.5	3287.322	1P7307B.09A	11/3/2009 4:35:2
4	60.990	185401.3	3039.864	1P7307B.10A	11/3/2009 4:48:2
5	81.320	236992.3	2914.317	1P7307B.11A	11/3/2009 5:01:2

ASX22 0204



Calibration formula: Y = 3452.217 X
Fit type = Avg CF with equal weighting, forced to origin
Coefficient of determination = 0.9139, Average error = 12.92%
Average CF = 3452.2170 with RSD = 17.77%

ASX22 0285

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088A Calibration File: 1P7307

GC Column (1): RTXCLPI ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1016	1	3.38	3.35	3.41	119	105	1	50	5954	12.0
					113		2	100	11253	
					107		3	200	21384	
					97		4	500	48586	
					87		5	1000	87228	
	2	3.86	3.83	3.89	148	127	1	50	7381	13.6
					138		2	100	13765	
					129		3	200	25774	
					116		4	500	58058	
					104		5	1000	103813	
	3	4.13	4.10	4.16	307	272	1	50	15354	12.7
					297		2	100	29721	
					281		3	200	56218	
					251		4	500	125257	
					223		5	1000	223467	
	4	4.18	4.15	4.21	134	121	1	50	6681	10.6
					131		2	100	13067	
					124		3	200	24854	
					113		4	500	56458	
					103		5	1000	102848	
	5	4.27	4.24	4.30	128	114	1	50	6395	10.1
					121		2	100	12068	
					117		3	200	23478	
					108		4	500	54005	
					98		5	1000	98327	
	6	4.65	4.62	4.68	91	82	1	50	4531	10.0
					87		2	100	8707	
					84		3	200	16701	
					76		4	500	38070	
					71		5	1000	70540	
Aroclor-1221	1	3.23	3.20	3.26	62	62	1	200	12425	
	2	3.31	3.28	3.34	44	44	1	200	8765	
	3	3.38	3.35	3.41	145	145	1	200	29071	
Aroclor-1232	1	3.38	3.35	3.41	123	123	1	200	24551	
	2	3.86	3.83	3.89	56	56	1	200	11285	
	3	4.13	4.10	4.16	130	130	1	200	26026	
	4	4.18	4.15	4.21	58	58	1	200	11569	
	5	4.26	4.23	4.29	54	54	1	200	10736	
	6	4.65	4.62	4.68	35	35	1	200	7085	

ASX22 0286

FORM VI-3

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088A Calibration File: 1P7307

GC Column (1) : RTXCLPI ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1242	1	3.38	3.35	3.41	89	89	1	200	17885	
	2	3.86	3.83	3.89	104	104	1	200	20724	
	3	4.12	4.09	4.15	235	235	1	200	47031	
	4	4.18	4.15	4.21	103	103	1	200	20675	
	5	4.26	4.23	4.29	98	98	1	200	19635	
	6	4.65	4.62	4.68	71	71	1	200	14297	
Aroclor-1248	1	4.13	4.10	4.16	170	150	1	50	8520	11.0
					162		2	100	16240	
					145		3	200	28940	
					144		4	500	71780	
					129		5	1000	128610	
	2	4.36	4.33	4.39	124	104	1	50	6197	15.1
					116		2	100	11642	
					100		3	200	19983	
					97		4	500	48509	
					84		5	1000	84452	
	3	4.53	4.50	4.56	182	156	1	50	9119	13.4
					172		2	100	17206	
					150		3	200	30093	
					147		4	500	73402	
					130		5	1000	129745	
	4	4.65	4.62	4.68	119	103	1	50	5956	12.6
					114		2	100	11409	
					98		3	200	19650	
					97		4	500	48604	
					88		5	1000	87537	
	5	5.05	5.02	5.08	114	125	1	50	5676	10.4
					112		2	100	11186	
					141		3	200	28146	
					136		4	500	67873	
					123		5	1000	123318	
	6	5.64	5.61	5.67	67	59	1	50	3336	9.4
					63		2	100	6341	
					57		3	200	11317	
					58		4	500	28874	
					53		5	1000	52812	

ASX22 0287

FORM VI-3

6F
INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:
Lab Code: Case No.: SAS No.: SDG No.:
Instrument: H5088A Calibration File: 1P7307
GC Column (1): RTXCLPI ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1254	1	4.86	4.83	4.89	152	120	1	50	7584	17.5
					131		2	100	13056	
					109		3	250	27272	
					109		4	500	54647	
					100		5	1000	99639	
	2	5.06	5.03	5.09	226	181	1	50	11310	17.5
					200		2	100	19952	
					174		3	250	43436	
					162		4	500	81030	
					146		5	1000	145933	
	3	5.30	5.27	5.33	170	132	1	50	8506	16.9
					131		2	100	13068	
					114		3	250	28418	
					128		4	500	63864	
					118		5	1000	118354	
	4	5.37	5.34	5.40	301	237	1	50	15037	18.6
					260		2	100	26043	
					226		3	250	56379	
					206		4	500	103120	
					192		5	1000	191746	
	5	5.88	5.85	5.91	146	122	1	50	7316	12.6
					128		2	100	12835	
					116		3	250	29107	
					112		4	500	55839	
					109		5	1000	108676	
	6	6.01	5.98	6.04	166	162	1	50	8294	6.0
					174		2	100	17384	
					166		3	250	41458	
					155		4	500	77285	
					150		5	1000	149744	

ASX22 0288

FORM VI-3

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088A Calibration File: 1P7307

GC Column (1) : RTXCLPI ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1260	1	5.77	5.74	5.80	231	204	1	50	11560	12.8
					223		2	100	22282	
					212		3	200	42420	
					186		4	500	93032	
					169		5	1000	168613	
	2	6.01	5.98	6.04	164	174	1	50	8186	7.3
					186		2	100	18615	
					188		3	200	37511	
					170		4	500	85243	
					160		5	1000	160307	
	3	6.08	6.05	6.11	156	137	1	50	7794	12.4
					149		2	100	14895	
					142		3	200	28386	
					126		4	500	62878	
					114		5	1000	114429	
	4	6.28	6.25	6.31	162	150	1	50	8099	9.3
					161		2	100	16096	
					157		3	200	31490	
					140		4	500	69901	
					131		5	1000	131127	
	5	6.58	6.55	6.61	364	321	1	50	18183	11.4
					342		2	100	34202	
					330		3	200	66033	
					294		4	500	147022	
					273		5	1000	273172	
	6	6.83	6.80	6.86	168	147	1	50	8415	12.6
					160		2	100	15985	
					152		3	200	30349	
					132		4	500	65767	
					125		5	1000	124922	

ASX22 0289

FORM VI-3

6F
INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:
Lab Code: Case No.: SAS No.: SDG No.:
Instrument: H5088A Calibration File: 1P7307
GC Column (1): RTXCLPI ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1262	1	5.52	5.49	5.55	201	166	1	50	10027	20.0
					194		2	100	19394	
					174		3	200	34767	
					136		4	500	68019	
					128		5	1000	127541	
	2	5.76	5.73	5.79	207	173	1	50	10342	19.3
					202		2	100	20201	
					180		3	200	36053	
					145		4	500	72584	
					132		5	1000	132311	
	3	6.08	6.05	6.11	277	229	1	50	13849	20.0
					269		2	100	26901	
					236		3	200	47282	
					188		4	500	93938	
					177		5	1000	176538	
	4	6.28	6.25	6.31	213	178	1	50	10667	19.2
					208		2	100	20750	
					185		3	200	37085	
					149		4	500	74482	
					137		5	1000	137007	
	5	6.58	6.55	6.61	484	410	1	50	24216	18.0
					472		2	100	47162	
					427		3	200	85374	
					344		4	500	171806	
					322		5	1000	322251	
	6	6.82	6.79	6.85	253	211	1	50	12636	18.8
					242		2	100	24242	
					220		3	200	43964	
					178		4	500	88887	
					162		5	1000	161889	

ASX22 0210

FORM VI-3

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088A Calibration File: 1P7307

GC Column (1) : RTXCLPI ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1268	1	6.08	6.05	6.11	97	91	1	50	4855	7.0
					96		2	100	9556	
					92		3	200	18324	
					88		4	500	44208	
					81		5	1000	81075	
	2	6.29	6.26	6.32	128	116	1	50	6377	10.8
					127		2	100	12658	
					118		3	200	23565	
					109		4	500	54690	
					98		5	1000	97662	
	3	6.82	6.79	6.85	491	418	1	50	24550	14.0
					460		2	100	45966	
					415		3	200	83035	
					376		4	500	188224	
					348		5	1000	347970	
	4	6.90	6.87	6.93	418	361	1	50	20923	13.6
					398		2	100	39823	
					363		3	200	72533	
					327		4	500	163421	
					299		5	1000	299262	
	5	7.06	7.03	7.09	407	350	1	50	20325	13.3
					385		2	100	38547	
					348		3	200	69596	
					317		4	500	158499	
					294		5	1000	294361	
	6	7.14	7.11	7.17	49	71	1	50	2452	17.7
					78		2	100	7781	
					79		3	200	15713	
					78		4	500	38873	
					73		5	1000	73472	

ASX22 0211

FORM VI-3

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088B Calibration File: 1P7307B

GC Column (2) : RTXCLPII ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1016	1	2.72	2.69	2.75	84	77	1	50	4210	7.8
					80		2	100	8042	
					79		3	200	15727	
					74		4	500	36793	
					69		5	1000	68793	
	2	3.01	2.98	3.04	193	171	1	50	9640	11.4
					184		2	100	18408	
					176		3	200	35207	
					159		4	500	79377	
					144		5	1000	144314	
	3	3.17	3.14	3.20	133	118	1	50	6630	10.3
					125		2	100	12510	
					120		3	200	23982	
					110		4	500	54990	
					102		5	1000	102030	
	4	3.40	3.37	3.43	183	176	1	50	9150	5.1
					179		2	100	17947	
					181		3	200	36230	
					173		4	500	86556	
					161		5	1000	161060	
	5	3.45	3.42	3.48	37	34	1	50	1825	5.3
					34		2	100	3443	
					35		3	200	7053	
					33		4	500	16534	
					32		5	1000	31856	
	6	3.80	3.77	3.83	106	98	1	50	5320	8.4
					103		2	100	10288	
					101		3	200	20206	
					92		4	500	46168	
					86		5	1000	86418	
Aroclor-1221	1	2.56	2.53	2.59	44	44	1	200	8779	
	2	2.64	2.61	2.67	28	28	1	200	5675	
	3	2.72	2.69	2.75	107	107	1	200	21361	
Aroclor-1232	1	2.72	2.69	2.75	88	88	1	200	17653	
	2	3.01	2.98	3.04	78	78	1	200	15523	
	3	3.17	3.14	3.20	55	55	1	200	11034	
	4	3.41	3.38	3.44	72	72	1	200	14465	
	5	3.45	3.42	3.48	15	15	1	200	3084	
	6	3.80	3.77	3.83	40	40	1	200	7952	

ASX22 0212

FORM VI-3

6F
INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:
Lab Code: Case No.: SAS No.: SDG No.:
Instrument: H5088B Calibration File: 1P7307B
GC Column (2) : RTXCLPII ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1242	1	2.71	2.68	2.74	64	64	1	200	12805	
	2	3.01	2.98	3.04	139	139	1	200	27896	
	3	3.17	3.14	3.20	96	96	1	200	19148	
	4	3.40	3.37	3.43	136	136	1	200	27137	
	5	3.45	3.42	3.48	31	31	1	200	6235	
	6	3.80	3.77	3.83	81	81	1	200	16223	
Aroclor-1248	1	3.17	3.14	3.20	46	42	1	50	2308	8.7
					46		2	100	4585	
					39		3	200	7860	
					41		4	500	20574	
					38		5	1000	38286	
	2	3.40	3.37	3.43	132	127	1	50	6624	4.8
					133		2	100	13319	
					120		3	200	24014	
					129		4	500	64311	
					121		5	1000	121499	
	3	3.80	3.77	3.83	154	134	1	50	7703	13.1
					149		2	100	14921	
					131		3	200	26224	
					113		4	500	56549	
					121		5	1000	121435	
	4	3.92	3.89	3.95	67	63	1	50	3347	6.6
					67		2	100	6662	
					59		3	200	11818	
					62		4	500	31211	
					58		5	1000	58042	
	5	4.29	4.26	4.32	128	103	1	50	6416	17.0
					113		2	100	11261	
					91		3	200	18112	
					95		4	500	47648	
					87		5	1000	86702	
	6	4.58	4.55	4.61	69	62	1	50	3431	7.9
					66		2	100	6585	
					58		3	200	11683	
					61		4	500	30597	
					57		5	1000	57032	

ASX22: 0213

FORM VI-3

6F
INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:
Lab Code: Case No.: SAS No.: SDG No.:
Instrument: H5088B Calibration File: 1P7307B
GC Column (2) : RTXCLPII ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1254	1	4.11	4.08	4.14	220	187	1	50	10998	12.2
					198		2	100	19778	
					182		3	250	45423	
					170		4	500	85113	
					164		5	1000	163562	
	2	4.58	4.55	4.61	244	208	1	50	12195	12.1
					222		2	100	22226	
					203		3	250	50647	
					188		4	500	94030	
					183		5	1000	183308	
	3	4.69	4.66	4.72	141	118	1	50	7055	13.4
					126		2	100	12553	
					115		3	250	28670	
					106		4	500	53222	
					102		5	1000	101874	
	4	4.79	4.76	4.82	155	133	1	50	7760	10.5
					137		2	100	13713	
					131		3	250	32770	
					122		4	500	60999	
					121		5	1000	120601	
	5	4.88	4.85	4.91	102	88	1	50	5083	10.3
					91		2	100	9077	
					87		3	250	21867	
					81		4	500	40655	
					79		5	1000	78715	
	6	5.14	5.11	5.17	214	181	1	50	10700	12.4
					192		2	100	19188	
					176		3	250	44051	
					164		4	500	82180	
					159		5	1000	158514	

ASX22 P214

FORM VI-3

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088B Calibration File: 1P7307B

GC Column (2): RTXCLPII ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1260	1	4.69	4.66	4.72	210	188	1	50	10504	11.0
					201		2	100	20144	
					194		3	200	38779	
					173		4	500	86726	
					160		5	1000	159759	
	2	4.87	4.84	4.90	215	199	1	50	10760	8.9
					211		2	100	21130	
					206		3	200	41273	
					186		4	500	92928	
					174		5	1000	174055	
	3	4.99	4.96	5.02	149	147	1	50	7464	6.3
					155		2	100	15547	
					154		3	200	30856	
					143		4	500	71323	
					133		5	1000	133147	
	4	5.39	5.36	5.42	176	162	1	50	8811	9.4
					172		2	100	17229	
					168		3	200	33592	
					151		4	500	75481	
					141		5	1000	140619	
	5	5.63	5.60	5.66	339	321	1	50	16964	7.5
					341		2	100	34126	
					334		3	200	66748	
					304		4	500	151766	
					288		5	1000	287997	
	6	5.92	5.89	5.95	236	217	1	50	11799	8.8
					230		2	100	23022	
					223		3	200	44624	
					203		4	500	101367	
					191		5	1000	190833	

ASX22 0215

FORM VI-3

6F
INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:
Lab Code: Case No.: SAS No.: SDG No.:
Instrument: H5088B Calibration File: 1P7307B
GC Column (2) : RTXCLPII ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1262	1	4.87	4.84	4.90	212	181	1	50	10604	16.6
					207		2	100	20737	
					186		3	200	37164	
					157		4	500	78319	
					144		5	1000	144425	
	2	5.39	5.36	5.42	265	226	1	50	13260	17.3
					260		2	100	25984	
					234		3	200	46703	
					191		4	500	95742	
					179		5	1000	179029	
	3	5.63	5.60	5.66	473	416	1	50	23646	14.5
					469		2	100	46929	
					431		3	200	86270	
					360		4	500	180114	
					344		5	1000	344198	
	4	5.87	5.84	5.90	174	150	1	50	8675	16.9
					174		2	100	17357	
					157		3	200	31307	
					124		4	500	62206	
					122		5	1000	122366	
	5	5.92	5.89	5.95	310	263	1	50	15492	16.3
					304		2	100	30362	
					260		3	200	52027	
					226		4	500	113119	
					216		5	1000	216481	
	6	6.38	6.35	6.41	203	171	1	50	10151	17.1
					193		2	100	19326	
					178		3	200	35621	
					145		4	500	72691	
					136		5	1000	136268	

ASX22 0216

FORM VI-3

6F
INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:
Lab Code: Case No.: SAS No.: SDG No.:
Instrument: H5088B Calibration File: 1P7307B
GC Column (2): RTXCLPII ID: 0.32 (mm) Date(s) Analyzed: 11/3/2009 11/3/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1268	1	5.38	5.35	5.41	153	136	1	50	7626	11.7
					149		2	100	14909	
					137		3	200	27485	
					126		4	500	63025	
					114		5	1000	114246	
	2	5.87	5.84	5.90	489	435	1	50	24427	10.0
					470		2	100	46987	
					430		3	200	86037	
					403		4	500	201376	
					385		5	1000	384950	
	3	5.93	5.90	5.96	429	382	1	50	21469	10.0
					411		2	100	41071	
					378		3	200	75635	
					352		4	500	175987	
					338		5	1000	338397	
	4	6.09	6.06	6.12	469	417	1	50	23465	10.9
					453		2	100	45327	
					416		3	200	83182	
					384		4	500	192182	
					360		5	1000	360481	
	5	6.17	6.14	6.20	104	98	1	50	5185	8.0
					104		2	100	10414	
					101		3	200	20250	
					93		4	500	46288	
					86		5	1000	86386	
	6	6.39	6.36	6.42	194	173	1	50	9705	9.5
					184		2	100	18393	
					174		3	200	34824	
					162		4	500	80889	
					153		5	1000	153215	

ASX22 8217

FORM VI-3

6D
INITIAL CALIBRATION - RETENTION TIME SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: S9147A
Calibration File: 1C6310
GC Column (1) : RTX CLP
ID: 0.32 (mm)
Update File:
Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	RT OF STANDARDS					MIDPOINT RT	RT WINDOW	
	LEVEL 1	LEVEL 2	LEVEL 3	LEVEL 4	LEVEL 5		FROM	TO
Tetrachloro-m-xylene	3.31	3.31	3.31	3.31	3.31	3.31	3.29	3.33
Decachlorobiphenyl	7.16	7.16	7.16	7.16	7.16	7.16	7.14	7.18

ASX22 0210

FORM VI-1

6D

INITIAL CALIBRATION - RETENTION TIME SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: S9147B
Calibration File: 1C6310B
GC Column (2) : RTX CLPII
ID: 0.32 (mm)
Update File:
Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	RT OF STANDARDS					MIDPOINT	RT WINDOW	
	LEVEL 1	LEVEL 2	LEVEL 3	LEVEL 4	LEVEL 5	RT	FROM	TO
Tetrachloro-m-xylene	2.81	2.81	2.81	2.81	2.81	2.81	2.79	2.83
Decachlorobiphenyl	6.36	6.36	6.36	6.36	6.36	6.36	6.34	6.38

ASX22 0219

FORM VI-1

6E

INITIAL CALIBRATION - CALIBRATION FACTOR SUMMARY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: S9147A Calibration File: 1C6310

GC Column (1): RTX CLP ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	CALIBRATION FACTORS						%RSD
	LEVEL 1	LEVEL 2	LEVEL 3	LEVEL 4	LEVEL 5	MEAN	
Tetrachloro-m-xylene	2.41E+03	1.96E+03	1.89E+03	1.84E+03	1.67E+03	1.96E+03	14.2
Decachlorobiphenyl	3.87E+02	3.17E+02	3.03E+02	2.98E+02	2.82E+02	3.17E+02	12.9

Average % RSD: 13.6

ASX22 8220

FORM VI-2

Calibration File Name: **C:\CPWIN\DATA1\1C6310.CAL** Version = 7
External standard calibration
Standard injection volume = 1
No sample weight correction
Area reject threshold = 0
Reference peak area reject threshold = 50
Amount units = PPB
2 components with 5 levels each

1 TCX

Retention time = 3.311 min., Search window = 0.020 min.
Low alarm amount = 0, High alarm amount = 0.1
Group number = 0, Component constant = 0
No retention time reference component
Single peak quantification by height

Level	Amount	Height	Height/Amt	Source	Date and time
1	4.094	9884.6	2414.42	1C6310.05A	11/6/2009 4:01:0
2	20.470	40131.8	1960.517	1C6310.06A	11/6/2009 4:13:0
3	40.940	77543.1	1894.067	1C6310.07A	11/6/2009 4:25:0
4	61.410	113177.8	1842.986	1C6310.08A	11/6/2009 4:37:0
5	81.880	136617.2	1668.505	1C6310.09A	11/6/2009 4:49:0



Calibration formula: Y = 1956.099 X
Fit type = Avg CF with equal weighting, forced to origin
Coefficient of determination = 0.9428, Average error = 9.46%
Average CF = 1956.0990 with RSD = 14.22%

2 DCB

Retention time = 7.160 min., Search window = 0.020 min.
Low alarm amount = 0, High alarm amount = 0.2
Group number = 0, Component constant = 0
No retention time reference component
Single peak quantification by height

Level	Amount	Height	Height/Amt	Source	Date and time
1	4.066	1572.0	386.6313	1C6310.05A	11/6/2009 4:01:0
2	20.330	6448.8	317.2078	1C6310.06A	11/6/2009 4:13:0
3	40.660	12307.6	302.6955	1C6310.07A	11/6/2009 4:25:0
4	60.990	18185.8	298.1759	1C6310.08A	11/6/2009 4:37:0
5	81.320	22902.5	281.6343	1C6310.09A	11/6/2009 4:49:0

ASX22 8221



Calibration formula: Y = 317.269 X
Fit type = Avg CF with equal weighting, forced to origin
Coefficient of determination = 0.9656, Average error = 8.74%
Average CF = 317.2690 with RSD = 12.86%

ASX22 8222

6E

INITIAL CALIBRATION - CALIBRATION FACTOR SUMMARY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: S9147B Calibration File: 1C6310B

GC Column (2) : RTX CLPII ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	CALIBRATION FACTORS						%RSD
	LEVEL 1	LEVEL 2	LEVEL 3	LEVEL 4	LEVEL 5	MEAN	
Tetrachloro-m-xylene	5.59E+03	4.62E+03	4.50E+03	4.43E+03	3.74E+03	4.58E+03	14.5
Decachlorobiphenyl	2.78E+03	2.37E+03	2.37E+03	2.32E+03	2.10E+03	2.39E+03	10.3

Average % RSD: 12.4

ASX22 0223

FORM VI-2

Calibration File Name: **C:\CPWIN\DATA1\1C6310B.CAL** Version = 7
External standard calibration
Standard injection volume = 1
No sample weight correction
Area reject threshold = 0
Reference peak area reject threshold = 50
Amount units = PPB
2 components with 5 levels each

1 TCX

Retention time = 2.808 min., Search window = 0.020 min.
Low alarm amount = 0, High alarm amount = 0.1
Group number = 0, Component constant = 0
No retention time reference component
Single peak quantification by height

Level	Amount	Height	Height/Amt	Source	Date and time
1	4.094	22877.9	5588.146	1C6310B.05A	11/6/2009 4:01:2
2	20.470	94601.8	4621.487	1C6310B.06A	11/6/2009 4:13:2
3	40.940	184135.8	4497.699	1C6310B.07A	11/6/2009 4:25:1
4	61.410	271897.9	4427.583	1C6310B.08A	11/6/2009 4:37:1
5	81.880	306404.7	3742.119	1C6310B.09A	11/6/2009 4:49:1



Calibration formula: Y = 4575.407 X
Fit type = Avg CF with equal weighting, forced to origin
Coefficient of determination = 0.9154, Average error = 9.26%
Average CF = 4575.4070 with RSD = 14.46%

2 DCB

Retention time = 6.359 min., Search window = 0.020 min.
Low alarm amount = 0, High alarm amount = 0.2
Group number = 0, Component constant = 0
No retention time reference component
Single peak quantification by height

Level	Amount	Height	Height/Amt	Source	Date and time
1	4.066	11302.6	2779.779	1C6310B.05A	11/6/2009 4:01:2
2	20.330	48138.1	2367.833	1C6310B.06A	11/6/2009 4:13:2
3	40.660	96410.5	2371.139	1C6310B.07A	11/6/2009 4:25:1
4	60.990	141265.9	2316.214	1C6310B.08A	11/6/2009 4:37:1
5	81.320	170656.8	2098.583	1C6310B.09A	11/6/2009 4:49:1

ASX22 0224



Calibration formula: Y = 2386.71 X
Fit type = Avg CF with equal weighting, forced to origin
Coefficient of determination = 0.9665, Average error = 6.59%
Average CF = 2386.7100 with RSD = 10.33%

ASX22 0225

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: S9147A
Calibration File: 1C6310
GC Column (1): RTX CLP
ID: 0.32 (mm)
Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1016	1	3.57	3.54	3.60	32	1	50	1937	15.3
						2	100	3467	
						3	200	6445	
						4	500	14811	
						5	1000	25797	
	2	3.91	3.88	3.94	38	1	50	2232	13.5
						2	100	4087	
						3	200	7734	
						4	500	17491	
						5	1000	31348	
	3	4.11	4.08	4.14	51	1	50	2895	9.3
						2	100	5307	
						3	200	10321	
						4	500	23767	
						5	1000	45864	
	4	4.16	4.13	4.19	29	1	50	1637	9.6
						2	100	2968	
						3	200	5737	
						4	500	13189	
						5	1000	25932	
	5	4.22	4.19	4.25	26	1	50	1473	11.2
						2	100	2691	
						3	200	5168	
						4	500	12081	
						5	1000	21848	
	6	4.43	4.40	4.46	19	1	50	1123	12.3
						2	100	2000	
						3	200	3753	
						4	500	8989	
						5	1000	16162	
Aroclor-1221	1	3.46	3.43	3.49	19	1	200	3820	
	2	3.52	3.49	3.55	18	1	200	3697	
	3	3.56	3.53	3.59	45	1	200	8988	
Aroclor-1232	1	3.57	3.54	3.60	35	1	200	7001	
	2	3.91	3.88	3.94	17	1	200	3391	
	3	4.11	4.08	4.14	22	1	200	4431	
	4	4.16	4.13	4.19	13	1	200	2514	
	5	4.22	4.19	4.25	11	1	200	2204	
	6	4.43	4.40	4.46	8	1	200	1589	

ASX22 8226

FORM VI-3

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: S9147A Calibration File: 1C6310

GC Column (1): RTX CLP ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1242	1	3.57	3.54	3.60	26	1	200	5115	
	2	3.91	3.88	3.94	30	1	200	5909	
	3	4.11	4.08	4.14	39	1	200	7828	
	4	4.16	4.13	4.19	22	1	200	4313	
	5	4.22	4.19	4.25	20	1	200	3967	
	6	4.43	4.40	4.46	15	1	200	3045	
Aroclor-1248	1	3.91	3.88	3.94	14	1	50	856	13.9
						2	100	1611	
						3	200	2709	
						4	500	6641	
						5	1000	12397	
	2	4.22	4.19	4.25	16	1	50	942	14.0
						2	100	1737	
						3	200	2887	
						4	500	7234	
						5	1000	13732	
	3	4.29	4.26	4.32	28	1	50	1647	13.4
						2	100	3178	
						3	200	5352	
						4	500	13096	
						5	1000	24113	
	4	4.43	4.40	4.46	28	1	50	1655	12.7
						2	100	3014	
						3	200	5403	
						4	500	12770	
						5	1000	24347	
	5	4.53	4.50	4.56	21	1	50	1235	11.8
						2	100	2295	
						3	200	3867	
						4	500	9865	
						5	1000	19158	
	6	4.73	4.70	4.76	26	1	50	1495	10.8
						2	100	2815	
						3	200	4709	
						4	500	12536	
						5	1000	23785	

ASX22 0227

FORM VI-3

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: S9147A Calibration File: 1C6310

GC Column (1): RTX CLP ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1254	1	4.29	4.26	4.32	24	1	50	1482	13.8
						2	100	2587	
						3	250	5684	
						4	500	11593	
						5	1000	20938	
	2	4.71	4.68	4.74	36	1	50	2170	13.8
						2	100	3832	
						3	250	8229	
						4	500	16815	
						5	1000	31252	
	3	4.88	4.85	4.91	43	1	50	2484	10.4
						2	100	4485	
						3	250	10065	
						4	500	20556	
						5	1000	38433	
	4	5.16	5.13	5.19	37	1	50	2198	11.2
						2	100	3805	
						3	250	8675	
						4	500	18020	
						5	1000	33334	
	5	5.38	5.35	5.41	18	1	50	988	6.0
						2	100	1782	
						3	250	4207	
						4	500	9378	
						5	1000	18017	
	6	5.69	5.66	5.72	25	1	50	1498	11.2
						2	100	2464	
						3	250	5737	
						4	500	12386	
						5	1000	23391	

ASX22 0228

FORM VI-3

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: S9147A Calibration File: 1C6310

GC Column (1): RTX CLP ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1260	1	5.28	5.25	5.31	31	1	50	1784	10.4
						2	100	3243	
						3	200	6198	
						4	500	14676	
						5	1000	27125	
	2	5.48	5.45	5.51	28	1	50	1567	7.9
						2	100	2977	
						3	200	5596	
						4	500	13688	
						5	1000	25508	
	3	5.69	5.66	5.72	28	1	50	1601	9.4
						2	100	2881	
						3	200	5509	
						4	500	13166	
						5	1000	25188	
	4	5.92	5.89	5.95	23	1	50	1284	9.1
						2	100	2295	
						3	200	4546	
						4	500	10925	
						5	1000	19981	
	5	6.15	6.12	6.18	44	1	50	2430	8.3
						2	100	4509	
						3	200	8859	
						4	500	20381	
						5	1000	39459	
	6	6.36	6.33	6.39	21	1	50	1206	9.9
						2	100	2258	
						3	200	4257	
						4	500	10047	
						5	1000	18715	

ASX22 8229

FORM VI-3

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: S9147A Calibration File: 1C6310

GC Column (1): RTX CLP ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1262	1	5.28	5.25	5.31	28	1	50	1696	15.8
						2	100	3166	
						3	200	5596	
						4	500	12181	
						5	1000	23656	
	2	5.48	5.45	5.51	26	1	50	1520	14.2
						2	100	2864	
						3	200	5133	
						4	500	11251	
						5	1000	22065	
	3	5.76	5.73	5.79	36	1	50	2140	15.2
						2	100	3911	
						3	200	6915	
						4	500	15222	
						5	1000	30626	
	4	5.92	5.89	5.95	27	1	50	1582	13.1
						2	100	3067	
						3	200	5265	
						4	500	11775	
						5	1000	24796	
	5	6.15	6.12	6.18	56	1	50	3244	12.3
						2	100	6102	
						3	200	11230	
						4	500	24604	
						5	1000	49516	
	6	6.35	6.32	6.38	32	1	50	1842	14.6
						2	100	3620	
						3	200	6244	
						4	500	13575	
						5	1000	27163	

ASX22 0238

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: S9147A Calibration File: 1C6310

GC Column (1): RTX CLP ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1268	1	5.93	5.90	5.96	20	1	50	1211	14.3
						2	100	2062	
						3	200	3916	
						4	500	8940	
						5	1000	16910	
	2	6.35	6.32	6.38	61	1	50	3629	11.2
						2	100	5793	
						3	200	11686	
						4	500	27956	
						5	1000	57812	
	3	6.41	6.38	6.44	48	1	50	2864	10.3
						2	100	4651	
						3	200	9373	
						4	500	22860	
						5	1000	45591	
	4	6.54	6.51	6.57	52	1	50	2903	6.7
						2	100	5074	
						3	200	10173	
						4	500	24676	
						5	1000	50480	
	5	6.77	6.74	6.80	18	1	50	1079	11.5
						2	100	1765	
						3	200	3526	
						4	500	8404	
						5	1000	16360	
	6	6.99	6.96	7.02	122	1	50	6997	8.5
						2	100	11618	
						3	200	22845	
						4	500	59276	
						5	1000	122357	

ASX22 0231

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: S9147B Calibration File: 1C6310B

GC Column (2) : RTX CLPII ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1016	1	3.10	3.07	3.13	102	87	1	50	5124	14.5
					92		2	100	9194	
					89		3	200	17700	
					81		4	500	40358	
					69		5	1000	69006	
	2	3.43	3.40	3.46	138	117	1	50	6882	13.7
					125		2	100	12475	
					119		3	200	23872	
					111		4	500	55436	
					95		5	1000	94507	
	3	3.59	3.56	3.62	238	209	1	50	11917	12.1
					224		2	100	22437	
					211		3	200	42214	
					200		4	500	99758	
					172		5	1000	171905	
	4	3.62	3.59	3.65	133	114	1	50	6654	13.9
					121		2	100	12102	
					117		3	200	23460	
					105		4	500	52435	
					92		5	1000	91871	
	5	3.70	3.67	3.73	125	109	1	50	6254	11.9
					114		2	100	11394	
					112		3	200	22304	
					102		4	500	50794	
					91		5	1000	90883	
	6	3.87	3.84	3.90	117	101	1	50	5873	11.9
					102		2	100	10159	
					104		3	200	20784	
					96		4	500	47761	
					85		5	1000	84834	
Aroclor-1221	1	2.98	2.95	3.01	54	54	1	200	10836	
	2	3.05	3.02	3.08	29	29	1	200	5830	
	3	3.10	3.07	3.13	121	121	1	200	24192	
Aroclor-1232	1	3.10	3.07	3.13	95	95	1	200	18933	
	2	3.43	3.40	3.46	53	53	1	200	10560	
	3	3.59	3.56	3.62	93	93	1	200	18597	
	4	3.63	3.60	3.66	50	50	1	200	10081	
	5	3.70	3.67	3.73	46	46	1	200	9212	
	6	3.87	3.84	3.90	41	41	1	200	8226	

ASX22 8232

FORM VI-3

6F
INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:
Lab Code: Case No.: SAS No.: SDG No.:
Instrument: S9147B Calibration File: 1C6310B
GC Column (2) : RTX CLPII ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1242	1	3.10	3.07	3.13	67	67	1	200	13492	
	2	3.43	3.40	3.46	90	90	1	200	17964	
	3	3.59	3.56	3.62	159	159	1	200	31829	
	4	3.62	3.59	3.65	86	86	1	200	17273	
	5	3.70	3.67	3.73	82	82	1	200	16328	
	6	3.87	3.84	3.90	74	74	1	200	14793	
Aroclor-1248	1	3.31	3.28	3.34	74	62	1	50	3678	14.4
					68		2	100	6808	
					58		3	200	11575	
					57		4	500	28575	
					52		5	1000	51840	
	2	3.59	3.56	3.62	125	106	1	50	6235	13.0
					116		2	100	11647	
					99		3	200	19876	
					98		4	500	49099	
					92		5	1000	91909	
	3	3.74	3.71	3.77	155	130	1	50	7744	14.5
					144		2	100	14410	
					125		3	200	24921	
					118		4	500	59002	
					109		5	1000	109016	
	4	3.87	3.84	3.90	166	140	1	50	8307	14.0
					153		2	100	15345	
					131		3	200	26273	
					130		4	500	64966	
					118		5	1000	117753	
	5	4.24	4.21	4.27	91	82	1	50	4545	9.3
					89		2	100	8888	
					76		3	200	15168	
					78		4	500	38939	
					75		5	1000	74906	
	6	4.48	4.45	4.51	78	69	1	50	3925	11.4
					77		2	100	7669	
					64		3	200	12740	
					64		4	500	32017	
					62		5	1000	62374	

ASX22 0233

FORM VI-3

6F
INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:
Lab Code: Case No.: SAS No.: SDG No.:
Instrument: S9147B Calibration File: 1C6310B
GC Column (2) : RTX CLPII ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1254	1	4.10	4.07	4.13	249	205	1	50	12470	13.6
					211		2	100	21128	
					197		3	250	49195	
					193		4	500	96676	
					175		5	1000	174728	
	2	4.21	4.18	4.24	194	168	1	50	9705	10.1
					176		2	100	17575	
					161		3	250	40223	
					160		4	500	80045	
					151		5	1000	150858	
	3	4.48	4.45	4.51	237	197	1	50	11874	12.3
					199		2	100	19850	
					182		3	250	45533	
					191		4	500	95349	
					176		5	1000	175708	
	4	4.65	4.62	4.68	129	112	1	50	6446	9.0
					111		2	100	11138	
					103		3	250	25753	
					111		4	500	55575	
					106		5	1000	105603	
	5	4.82	4.79	4.85	83	82	1	50	4169	1.9
					84		2	100	8353	
					83		3	250	20715	
					81		4	500	40303	
					80		5	1000	80269	
	6	4.95	4.92	4.98	177	146	1	50	8846	13.4
					150		2	100	14971	
					129		3	250	32339	
					147		4	500	73508	
					129		5	1000	128666	

ASX22 8234

FORM VI-3

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: S9147B Calibration File: 1C6310B

GC Column (2) : RTX CLPII ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1260	1	4.58	4.55	4.61	189	164	1	50	9429	12.0
					174		2	100	17420	
					164		3	200	32836	
					157		4	500	78276	
					136		5	1000	135981	
	2	4.71	4.68	4.74	197	175	1	50	9854	10.0
					184		2	100	18390	
					177		3	200	35336	
					168		4	500	83769	
					150		5	1000	150311	
	3	4.82	4.79	4.85	182	160	1	50	9115	10.5
					167		2	100	16680	
					161		3	200	32110	
					154		4	500	77208	
					137		5	1000	136563	
	4	4.95	4.92	4.98	164	148	1	50	8186	9.0
					155		2	100	15453	
					148		3	200	29660	
					145		4	500	72389	
					128		5	1000	127827	
	5	5.37	5.34	5.40	296	274	1	50	14779	7.9
					283		2	100	28329	
					280		3	200	55925	
					271		4	500	135537	
					238		5	1000	238134	
	6	5.63	5.60	5.66	195	181	1	50	9756	7.5
					188		2	100	18792	
					185		3	200	37032	
					175		4	500	87410	
					160		5	1000	160112	

ASX22 0235

FORM VI-3

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: S9147B Calibration File: 1C6310B

GC Column (2) : RTX CLPII ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1262	1	4.71	4.68	4.74	181	156	1	50	9054	14.5
					175		2	100	17488	
					157		3	200	31387	
					134		4	500	66902	
					132		5	1000	132150	
	2	4.98	4.95	5.01	251	218	1	50	12544	13.9
					244		2	100	24351	
					219		3	200	43875	
					193		4	500	96627	
					181		5	1000	181488	
	3	5.18	5.15	5.21	243	205	1	50	12133	15.3
					231		2	100	23139	
					200		3	200	39962	
					180		4	500	89938	
					171		5	1000	171168	
	4	5.37	5.34	5.40	405	352	1	50	20265	13.8
					394		2	100	39374	
					357		3	200	71364	
					309		4	500	154698	
					297		5	1000	296581	
	5	5.63	5.60	5.66	275	239	1	50	13756	13.8
					261		2	100	26127	
					248		3	200	49635	
					200		4	500	100196	
					208		5	1000	208307	
	6	6.00	5.97	6.03	147	128	1	50	7362	13.2
					143		2	100	14349	
					123		3	200	24505	
					110		4	500	55103	
					114		5	1000	114464	

ASX22 0236

FORM VI-3

6F

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: S9147B Calibration File: 1C6310B

GC Column (2) : RTX CLPII ID: 0.32 (mm) Date(s) Analyzed: 11/6/2009 11/6/2009

COMPOUND	PEAK	RT	RT WINDOW FROM	RT WINDOW TO	CALIBRATION FACTOR	AVERAGE CF	LEVEL	AMOUNT	PEAK HEIGHT	%RSD
Aroclor-1268	1	5.18	5.15	5.21	143	119	1	50	7129	12.7
					122		2	100	12161	
					117		3	200	23496	
					108		4	500	54202	
					104		5	1000	103529	
	2	5.58	5.55	5.61	481	404	1	50	24055	12.3
					411		2	100	41059	
					404		3	200	80847	
					377		4	500	188446	
					348		5	1000	347559	
	3	5.63	5.60	5.66	387	340	1	50	19373	8.1
					341		2	100	34080	
					332		3	200	66318	
					323		4	500	161474	
					319		5	1000	318826	
	4	5.78	5.75	5.81	397	351	1	50	19854	8.9
					360		2	100	35961	
					348		3	200	69687	
					335		4	500	167729	
					313		5	1000	312836	
	5	6.00	5.97	6.03	152	129	1	50	7616	10.5
					126		2	100	12572	
					126		3	200	25137	
					124		4	500	61861	
					117		5	1000	117035	
	6	6.20	6.17	6.23	1044	893	1	50	52203	11.0
					901		2	100	90057	
					894		3	200	178811	
					850		4	500	424808	
					774		5	1000	774480	

ASX22 0237

FORM VI-3

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: H5088A
Init. Calib Date(s): 11/03/09 11/03/09
GC Column (1) : RTXCLPI ID: .32 (mm)
Date Analyzed: 11/03/09
Lab File ID: 1P7307.39R
Time Analyzed: 22:55
Lab Standard ID: IC16XAA
Initial Calibration: 1P7307

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	3.00	2.97	3.03	44.36	40.94	8.3
Decachlorobiphenyl	7.98	7.95	8.01	46.08	40.66	13.3
Aroclor-1016	3.38	3.35	3.41	215.83	200.60	7.6
	3.86	3.83	3.89			
	4.12	4.10	4.16			
	4.18	4.15	4.21			
	4.26	4.24	4.30			
	4.65	4.62	4.68			
Aroclor-1260	5.77	5.74	5.80	219.71	200.20	9.7
	6.01	5.98	6.04			
	6.08	6.05	6.11			
	6.28	6.25	6.31			
	6.58	6.55	6.61			
	6.83	6.80	6.86			

Average of %D: 9.7

ASX22 0238

FORM VII

7E

CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088B Init. Calib Date(s): 11/03/09 11/03/09

GC Column (2) : RTXCLPII ID: .32 (mm) Date Analyzed: 11/03/09

Lab File ID: 1P7307B.39R Time Analyzed: 22:55

Lab Standard ID: IC16XAA Initial Calibration: 1P7307B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	2.34	2.31	2.37	41.55	40.94	1.5
Decachlorobiphenyl	6.80	6.77	6.83	40.56	40.66	-0.2
Aroclor-1016	2.72	2.69	2.75	213.36	200.60	6.4
	3.01	2.98	3.04			
	3.17	3.14	3.20			
	3.40	3.37	3.43			
	3.45	3.42	3.48			
	3.80	3.77	3.83			
Aroclor-1260	4.69	4.66	4.72	215.18	200.20	7.5
	4.87	4.84	4.90			
	4.99	4.96	5.02			
	5.39	5.36	5.42			
	5.63	5.60	5.66			
	5.92	5.89	5.95			
					Average of %D:	3.9

ASX22 0239

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: H5088A
Init. Calib Date(s): 11/03/09 11/03/09
GC Column (1) : RTXCLPI ID: .32 (mm)
Date Analyzed: 11/03/09
Lab File ID: 1P7307.40R
Time Analyzed: 23:08
Lab Standard ID: IC48XAA
Initial Calibration: 1P7307

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	3.00	2.97	3.03	45.47	40.94	11.1
Decachlorobiphenyl	7.97	7.95	8.01	46.56	40.66	14.5
Aroclor-1248	4.12	4.10	4.16	217.76	200.20	8.8
	4.36	4.33	4.39			
	4.53	4.50	4.56			
	4.65	4.62	4.68			
	5.05	5.02	5.08			
	5.64	5.61	5.67			
					Average of %D:	11.5

ASX22: 0240

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: H5088B
Init. Calib Date(s): 11/03/09 11/03/09
GC Column (2) : RTXCLPII ID: .32 (mm)
Date Analyzed: 11/03/09
Lab File ID: 1P7307B.40R
Time Analyzed: 23:08
Lab Standard ID: IC48XAA
Initial Calibration: 1P7307B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	2.34	2.31	2.37	41.37	40.94	1.0
Decachlorobiphenyl	6.80	6.77	6.83	40.80	40.66	0.3
Aroclor-1248	3.17	3.14	3.20	203.65	200.20	1.7
	3.40	3.37	3.43			
	3.80	3.77	3.83			
	3.92	3.89	3.95			
	4.29	4.26	4.32			
	4.58	4.55	4.61			

Average of %D: 1.0

ASX22 0241

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: H5088A
Init. Calib Date(s): 11/03/09 11/03/09
GC Column (1) : RTXCLPI ID: .32 (mm)
Date Analyzed: 11/03/09
Lab File ID: 1P7307.41R
Time Analyzed: 23:21
Lab Standard ID: IC54XAA
Initial Calibration: 1P7307

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	3.00	2.97	3.03	44.19	40.94	7.9
Decachlorobiphenyl	7.98	7.95	8.01	46.75	40.66	15.0
Aroclor-1254	4.86	4.83	4.89	256.79	250.00	2.7
	5.06	5.03	5.09			
	5.30	5.27	5.33			
	5.37	5.34	5.40			
	5.88	5.85	5.91			
	6.01	5.98	6.04			

Average of %D: 8.5

ASX22 0242

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: H5088B
Init. Calib Date(s): 11/03/09 11/03/09
GC Column (2) : RTXCLPII ID: .32 (mm)
Date Analyzed: 11/03/09
Lab File ID: 1P7307B.41R
Time Analyzed: 23:21
Lab Standard ID: IC54XAA
Initial Calibration: 1P7307B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	2.34	2.31	2.37	41.86	40.94	2.2
Decachlorobiphenyl	6.80	6.77	6.83	41.07	40.66	1.0
Aroclor-1254	4.11	4.08	4.14	256.37	250.00	2.5
	4.58	4.55	4.61			
	4.70	4.66	4.72			
	4.80	4.76	4.82			
	4.88	4.85	4.91			
	5.15	5.11	5.17			
					Average of %D:	1.9

ASX22 0243

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088A Init. Calib Date(s): 11/03/09 11/03/09

GC Column (1) : RTXCLPI ID: .32 (mm) Date Analyzed: 11/05/09

Lab File ID: 3P7307.40R Time Analyzed: 23:41

Lab Standard ID: AR164FN Initial Calibration: 1P7307

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	3.00	2.97	3.03	64.40	61.41	4.9
Decachlorobiphenyl	7.97	7.95	8.01	47.41	60.99	-22.3
Aroclor-1016	3.38	3.35	3.41	488.22	500.00	-2.4
	3.86	3.83	3.89			
	4.12	4.10	4.16			
	4.18	4.15	4.21			
	4.26	4.24	4.30			
	4.65	4.62	4.68			
Aroclor-1260	5.77	5.74	5.80	377.15	500.00	-24.6
	6.01	5.98	6.04			
	6.08	6.05	6.11			
	6.28	6.25	6.31			
	6.58	6.55	6.61			
	6.83	6.80	6.86			

Average of %D: 13.6

ASX22 0244

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088B Init. Calib Date(s): 11/03/09 11/03/09

GC Column (2) : RTXCLPII ID: .32 (mm) Date Analyzed: 11/05/09

Lab File ID: 3P7307B.40R Time Analyzed: 23:41

Lab Standard ID: AR164FN Initial Calibration: 1P7307B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	2.34	2.31	2.37	60.39	61.41	-1.7
Decachlorobiphenyl	6.80	6.77	6.83	50.76	60.99	-16.8
Aroclor-1016	2.71	2.69	2.75	523.92	500.00	4.8
	3.01	2.98	3.04			
	3.17	3.14	3.20			
	3.40	3.37	3.43			
	3.45	3.42	3.48			
	3.80	3.77	3.83			
Aroclor-1260	4.69	4.66	4.72	496.26	500.00	-0.7
	4.87	4.84	4.90			
	4.98	4.96	5.02			
	5.39	5.36	5.42			
	5.63	5.60	5.66			
	5.92	5.89	5.95			

Average of %D: 6.0

ASX22 18245

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088A Init. Calib Date(s): 11/03/09 11/03/09

GC Column (1) : RTXCLPI ID: .32 (mm) Date Analyzed: 11/06/09

Lab File ID: 3P7307.52R Time Analyzed: 2:17

Lab Standard ID: AR163OU Initial Calibration: 1P7307

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	3.00	2.97	3.03	40.91	40.94	-0.1
Decachlorobiphenyl	7.97	7.95	8.01	33.45	40.66	-17.7
Aroclor-1016	3.38	3.35	3.41	186.71	200.00	-6.6
	3.86	3.83	3.89			
	4.12	4.10	4.16			
	4.18	4.15	4.21			
	4.26	4.24	4.30			
	4.65	4.62	4.68			
Aroclor-1260	5.77	5.74	5.80	167.38	200.00	-16.3
	6.01	5.98	6.04			
	6.08	6.05	6.11			
	6.28	6.25	6.31			
	6.58	6.55	6.61			
	6.83	6.80	6.86			

Average of %D: 10.2

ASX22 8246

7E

CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088B Init. Calib Date(s): 11/03/09 11/03/09

GC Column (2) : RTXCLPII ID: .32 (mm) Date Analyzed: 11/06/09

Lab File ID: 3P7307B.52R Time Analyzed: 2:17

Lab Standard ID: AR163OU Initial Calibration: 1P7307B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	2.34	2.31	2.37	39.68	40.94	-3.1
Decachlorobiphenyl	6.80	6.77	6.83	36.36	40.66	-10.6
Aroclor-1016	2.71	2.69	2.75	214.87	200.00	7.4
	3.00	2.98	3.04			
	3.17	3.14	3.20			
	3.40	3.37	3.43			
	3.45	3.42	3.48			
	3.79	3.77	3.83			
Aroclor-1260	4.69	4.66	4.72	214.29	200.00	7.1
	4.87	4.84	4.90			
	4.98	4.96	5.02			
	5.39	5.36	5.42			
	5.63	5.60	5.66			
	5.92	5.89	5.95			
					Average of %D:	7.1

ASX22 8247

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088A Init. Calib Date(s): 11/03/09 11/03/09

GC Column (1) : RTXCLPI ID: .32 (mm) Date Analyzed: 11/06/09

Lab File ID: 3P7307.63R Time Analyzed: 4:41

Lab Standard ID: AR164FO Initial Calibration: 1P7307

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	3.00	2.97	3.03	73.39	61.41	19.5
Decachlorobiphenyl	7.97	7.95	8.01	65.25	60.99	7.0
Aroclor-1016	3.38	3.35	3.41	541.64	500.00	8.3
	3.86	3.83	3.89			
	4.12	4.10	4.16			
	4.18	4.15	4.21			
	4.26	4.24	4.30			
	4.65	4.62	4.68			
Aroclor-1260	5.76	5.74	5.80	542.29	500.00	8.5
	6.01	5.98	6.04			
	6.08	6.05	6.11			
	6.28	6.25	6.31			
	6.57	6.55	6.61			
	6.83	6.80	6.86			

Average of %D: 10.8

8SX22 0248

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: H5088B Init. Calib Date(s): 11/03/09 11/03/09

GC Column (2) : RTXCLPII ID: .32 (mm) Date Analyzed: 11/06/09

Lab File ID: 3P7307B.63R Time Analyzed: 4:41

Lab Standard ID: AR164FO Initial Calibration: 1P7307B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	2.34	2.31	2.37	71.32	61.41	16.1
Decachlorobiphenyl	6.80	6.77	6.83	67.69	60.99	11.0
Aroclor-1016	2.71	2.69	2.75	614.46	500.00	22.9
	3.00	2.98	3.04			
	3.17	3.14	3.20			
	3.40	3.37	3.43			
	3.45	3.42	3.48			
	3.79	3.77	3.83			
Aroclor-1260	4.69	4.66	4.72	628.33	500.00	25.7
	4.87	4.84	4.90			
	4.98	4.96	5.02			
	5.39	5.36	5.42			
	5.62	5.60	5.66			
	5.92	5.89	5.95			
					Average of %D:	18.9

ASX22 0249

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: S9147A
Init. Calib Date(s): 11/06/09 11/06/09
GC Column (1) : RTX CLP ID: .32 (mm)
Date Analyzed: 11/06/09
Lab File ID: 1C6310.37R
Time Analyzed: 22:16
Lab Standard ID: IC16XAA
Initial Calibration: 1C6310

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	3.31	3.29	3.33	40.34	40.94	-1.5
Decachlorobiphenyl	7.16	7.14	7.18	38.12	40.66	-6.2
Aroclor-1016	3.56	3.54	3.60	194.66	200.60	-3.0
	3.91	3.88	3.94			
	4.11	4.08	4.14			
	4.15	4.13	4.19			
	4.22	4.19	4.25			
	4.43	4.40	4.46			
Aroclor-1260	5.28	5.25	5.31	181.05	200.20	-9.6
	5.48	5.45	5.51			
	5.69	5.66	5.72			
	5.92	5.89	5.95			
	6.15	6.12	6.18			
	6.36	6.33	6.39			

Average of %D: 5.1

ASX22 0258

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: S9147B
Init. Calib Date(s): 11/06/09 11/06/09
GC Column (2) : RTX CLPII ID: .32 (mm)
Date Analyzed: 11/06/09
Lab File ID: 1C6310B.37R
Time Analyzed: 22:16
Lab Standard ID: IC16XAA
Initial Calibration: 1C6310B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	2.81	2.79	2.83	38.68	40.94	-5.5
Decachlorobiphenyl	6.36	6.34	6.38	36.65	40.66	-9.9
Aroclor-1016	3.10	3.07	3.13	195.83	200.60	-2.4
	3.43	3.40	3.46			
	3.59	3.56	3.62			
	3.62	3.59	3.65			
	3.70	3.67	3.73			
	3.87	3.84	3.90			
Aroclor-1260	4.57	4.55	4.61	186.46	200.20	-6.9
	4.71	4.68	4.74			
	4.82	4.79	4.85			
	4.95	4.92	4.98			
	5.37	5.34	5.40			
	5.63	5.60	5.66			

Average of %D: 6.2

ASX22 0251

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: S9147A
Init. Calib Date(s): 11/06/09 11/06/09
GC Column (1) : RTX CLP ID: .32 (mm)
Date Analyzed: 11/06/09
Time Analyzed: 22:28
Lab File ID: 1C6310.38R
Initial Calibration: 1C6310
Lab Standard ID: IC48XAA

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/l)	NOM AMOUNT (ug/l)	%D
Tetrachloro-m-xylene	3.31	3.29	3.33	42.77	40.94	4.5
Decachlorobiphenyl	7.16	7.14	7.18	41.29	40.66	1.6
Aroclor-1248	3.91	3.88	3.94	183.25	200.20	-8.5
	4.22	4.19	4.25			
	4.29	4.26	4.32			
	4.43	4.40	4.46			
	4.53	4.50	4.56			
	4.76	4.70	4.76			

Average of %D: 4.9

ASX22 0252

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories Contract:

Lab Code: Case No.: SAS No.: SDG No.:

Instrument: S9147B Init. Calib Date(s): 11/06/09 11/06/09

GC Column (2) : RTX CLPII ID: .32 (mm) Date Analyzed: 11/06/09

Lab File ID: 1C6310B.38R Time Analyzed: 22:28

Lab Standard ID: IC48XAA Initial Calibration: 1C6310B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/l)	NOM AMOUNT (ug/l)	%D
Tetrachloro-m-xylene	2.81	2.79	2.83	41.74	40.94	2.0
Decachlorobiphenyl	6.36	6.34	6.38	41.35	40.66	1.7
Aroclor-1248	3.31	3.28	3.34	190.78	200.20	-4.7
	3.58	3.56	3.62			
	3.74	3.71	3.77			
	3.87	3.84	3.90			
	4.24	4.21	4.27			
	4.48	4.45	4.51			

Average of %D: 2.8

ASX22 0253

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: S9147A
Init. Calib Date(s): 11/06/09 11/06/09
GC Column (1) : RTX CLP ID: .32 (mm)
Date Analyzed: 11/06/09
Time Analyzed: 22:40
Lab File ID: 1C6310.39R
Initial Calibration: 1C6310
Lab Standard ID: IC54XAA

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	3.31	3.29	3.33	43.42	40.94	6.1
Decachlorobiphenyl	7.16	7.14	7.18	41.40	40.66	1.8
Aroclor-1254	4.29	4.26	4.32	238.67	250.00	-4.5
	4.71	4.68	4.74			
	4.87	4.85	4.91			
	5.15	5.13	5.19			
	5.37	5.35	5.41			
	5.69	5.66	5.72			

Average of %D: 4.1

ASX22 8254

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: S9147B
Init. Calib Date(s): 11/06/09 11/06/09
GC Column (2) : RTX CLPII ID: .32 (mm)
Date Analyzed: 11/06/09
Lab File ID: 1C6310B.39R
Time Analyzed: 22:40
Lab Standard ID: IC54XAA
Initial Calibration: 1C6310B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	2.81	2.79	2.83	43.29	40.94	5.7
Decachlorobiphenyl	6.36	6.34	6.38	41.57	40.66	2.2
Aroclor-1254	4.10	4.07	4.13	233.87	250.00	-6.5
	4.20	4.18	4.24			
	4.48	4.45	4.51			
	4.64	4.62	4.68			
	4.82	4.79	4.85			
	4.95	4.92	4.98			

Average of %D: 4.8

ASX22 0255

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: S9147A
Init. Calib Date(s): 11/06/09 11/06/09
GC Column (1) : RTX CLP ID: .32 (mm)
Date Analyzed: 11/07/09
Lab File ID: 1C6310.93R
Time Analyzed: 9:28
Lab Standard ID: AR164FS
Initial Calibration: 1C6310

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/l)	NOM AMOUNT (ug/l)	%D
Tetrachloro-m-xylene	3.31	3.29	3.33	64.32	61.41	4.7
Decachlorobiphenyl	7.16	7.14	7.18	66.06	60.99	8.3
Aroclor-1016	3.56	3.54	3.60	493.68	500.00	-1.3
	3.91	3.88	3.94			
	4.11	4.08	4.14			
	4.15	4.13	4.19			
	4.22	4.19	4.25			
	4.43	4.40	4.46			
Aroclor-1260	5.28	5.25	5.31	521.76	500.00	4.4
	5.48	5.45	5.51			
	5.69	5.66	5.72			
	5.92	5.89	5.95			
	6.15	6.12	6.18			
	6.36	6.33	6.39			

Average of %D: 4.7

ASX22 0256

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: S9147B
Init. Calib Date(s): 11/06/09 11/06/09
GC Column (2) : RTX CLPII ID: .32 (mm)
Date Analyzed: 11/07/09
Lab File ID: 1C6310B.93R
Time Analyzed: 9:28
Lab Standard ID: AR164FS
Initial Calibration: 1C6310B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/l)	NOM AMOUNT (ug/l)	%D
Tetrachloro-m-xylene	2.81	2.79	2.83	65.28	61.41	6.3
Decachlorobiphenyl	6.36	6.34	6.38	58.34	60.99	-4.3
Aroclor-1016	3.10	3.07	3.13	493.93	500.00	-1.2
	3.43	3.40	3.46			
	3.59	3.56	3.62			
	3.62	3.59	3.65			
	3.70	3.67	3.73			
	3.87	3.84	3.90			
Aroclor-1260	4.57	4.55	4.61	518.40	500.00	3.7
	4.71	4.68	4.74			
	4.82	4.79	4.85			
	4.95	4.92	4.98			
	5.37	5.34	5.40			
	5.62	5.60	5.66			

Average of %D: 3.9

ASX22 8257

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories Contract:
Lab Code: Case No.: SAS No.: SDG No.:
Instrument: S9147A Init. Calib Date(s): 11/06/09 11/06/09
GC Column (1) : RTX CLP ID: .32 (mm) Date Analyzed: 11/07/09
Lab File ID: 2C6310.11R Time Analyzed: 12:00
Lab Standard ID: AR163PM Initial Calibration: 1C6310

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	3.31	3.29	3.33	38.52	40.94	-5.9
Decachlorobiphenyl	7.16	7.14	7.18	41.06	40.66	1.0
Aroclor-1016	3.56	3.54	3.60	188.50	200.00	-5.8
	3.91	3.88	3.94			
	4.11	4.08	4.14			
	4.15	4.13	4.19			
	4.22	4.19	4.25			
	4.43	4.40	4.46			
Aroclor-1260	5.28	5.25	5.31	198.04	200.00	-1.0
	5.48	5.45	5.51			
	5.69	5.66	5.72			
	5.92	5.89	5.95			
	6.15	6.12	6.18			
	6.36	6.33	6.39			

Average of %D: 3.4

ASXZZ 0258

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories Contract:
Lab Code: Case No.: SAS No.: SDG No.:
Instrument: S9147B Init. Calib Date(s): 11/06/09 11/06/09
GC Column (2) : RTX CLPII ID: .32 (mm) Date Analyzed: 11/07/09
Lab File ID: 2C6310B.11R Time Analyzed: 12:00
Lab Standard ID: AR163PM Initial Calibration: 1C6310B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	2.81	2.79	2.83	37.93	40.94	-7.3
Decachlorobiphenyl	6.36	6.34	6.38	34.91	40.66	-14.2
Aroclor-1016	3.10	3.07	3.13	197.86	200.00	-1.1
	3.43	3.40	3.46			
	3.59	3.56	3.62			
	3.62	3.59	3.65			
	3.70	3.67	3.73			
	3.87	3.84	3.90			
Aroclor-1260	4.57	4.55	4.61	198.26	200.00	-0.9
	4.71	4.68	4.74			
	4.82	4.79	4.85			
	4.95	4.92	4.98			
	5.37	5.34	5.40			
	5.63	5.60	5.66			

Average of %D: 5.9

ASX22 0259

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code: Case No.: SAS No.: SDG No.:
Instrument: S9147A
Init. Calib Date(s): 11/06/09 11/06/09
GC Column (1) : RTX CLP ID: .32 (mm)
Date Analyzed: 11/09/09
Lab File ID: 3C6310.45R
Time Analyzed: 19:24
Lab Standard ID: AR163QD
Initial Calibration: 1C6310

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	3.31	3.29	3.33	45.01	40.94	9.9
Decachlorobiphenyl	7.16	7.14	7.18	43.34	40.66	6.6
Aroclor-1016	3.57	3.54	3.60	201.49	200.00	0.7
	3.92	3.88	3.94			
	4.11	4.08	4.14			
	4.16	4.13	4.19			
	4.22	4.19	4.25			
	4.43	4.40	4.46			
Aroclor-1260	5.28	5.25	5.31	219.34	200.00	9.7
	5.49	5.45	5.51			
	5.69	5.66	5.72			
	5.92	5.89	5.95			
	6.15	6.12	6.18			
	6.36	6.33	6.39			

Average of %D: 6.7

ASX22. 0260

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories Contract:
Lab Code: Case No.: SAS No.: SDG No.:
Instrument: S9147B Init. Calib Date(s): 11/06/09 11/06/09
GC Column (2) : RTX CLPII ID: .32 (mm) Date Analyzed: 11/09/09
Lab File ID: 3C6310B.45R Time Analyzed: 19:24
Lab Standard ID: AR163QD Initial Calibration: 1C6310B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	2.81	2.79	2.83	45.38	40.94	10.9
Decachlorobiphenyl	6.36	6.34	6.38	39.04	40.66	-4.0
Aroclor-1016	3.10	3.07	3.13	208.93	200.00	4.5
	3.43	3.40	3.46			
	3.59	3.56	3.62			
	3.63	3.59	3.65			
	3.70	3.67	3.73			
	3.87	3.84	3.90			
Aroclor-1260	4.58	4.55	4.61	214.32	200.00	7.2
	4.72	4.68	4.74			
	4.82	4.79	4.85			
	4.95	4.92	4.98			
	5.37	5.34	5.40			
	5.63	5.60	5.66			

Average of %D: 6.7

ASX22 0261

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: S9147A
Init. Calib Date(s): 11/06/09 11/06/09
GC Column (1) : RTX CLP ID: .32 (mm)
Date Analyzed: 11/09/09
Time Analyzed: 21:48
Lab File ID: 3C6310.57R
Initial Calibration: 1C6310
Lab Standard ID: AR162OT

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	3.31	3.29	3.33	23.79	20.47	16.2
Decachlorobiphenyl	7.16	7.14	7.18	14.76	20.33	-27.4
Aroclor-1016	3.57	3.54	3.60	82.19	100.00	-17.8
	3.92	3.88	3.94			
	4.11	4.08	4.14			
	4.16	4.13	4.19			
	4.22	4.19	4.25			
	4.43	4.40	4.46			
Aroclor-1260	5.29	5.25	5.31	80.61	100.00	-19.4
	5.49	5.45	5.51			
	5.70	5.66	5.72			
	5.92	5.89	5.95			
	6.16	6.12	6.18			
	6.36	6.33	6.39			

Average of %D: 20.2

ASX22 8262

FORM VII

7E
CALIBRATION VERIFICATION SUMMARY

Lab Name: Lancaster Laboratories
Contract:
Lab Code:
Case No.:
SAS No.:
SDG No.:
Instrument: S9147B
Init. Calib Date(s): 11/06/09 11/06/09
GC Column (2) : RTX CLPII ID: .32 (mm)
Date Analyzed: 11/09/09
Lab File ID: 3C6310B.57R
Time Analyzed: 21:48
Lab Standard ID: AR162OT
Initial Calibration: 1C6310B

COMPOUND	RT	RT WINDOW FROM	TO	CALC AMOUNT (ug/kg)	NOM AMOUNT (ug/kg)	%D
Tetrachloro-m-xylene	2.81	2.79	2.83	22.99	20.47	12.3
Decachlorobiphenyl	6.36	6.34	6.38	8.48	20.33	-58.3
Aroclor-1016	3.10	3.07	3.13	101.26	100.00	1.3
	3.43	3.40	3.46			
	3.59	3.56	3.62			
	3.63	3.59	3.65			
	3.70	3.67	3.73			
	3.87	3.84	3.90			
Aroclor-1260	4.58	4.55	4.61	67.08	100.00	-32.9
	4.72	4.68	4.74			
	4.82	4.79	4.85			
	4.95	4.92	4.98			
	5.37	5.34	5.40			
	5.63	5.60	5.66			

Average of %D: 26.2

ASX22 0263

FORM VII

8D
ANALYTICAL SEQUENCE

Sequence: 1P7307 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTXCLPI ID: 0.32

Instrument: H5088A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
001		CONDITIONER	11/03/2009	14:28:03	1P7307		
002		CONDITIONER	11/03/2009	14:41:03	1P7307		
003		CONDITIONER	11/03/2009	14:53:58	1P7307		
004		CONDITIONER	11/03/2009	15:06:57	1P7307	3.00	7.98
005		CONDITIONER	11/03/2009	15:19:54	1P7307	3.00	7.98
006	PIBLKAA	IBLKX0924B	11/03/2009	15:45:59	1P7307	3.00	7.98
007	AR161AA	AR1610924A	11/03/2009	15:59:53	1P7307	3.00	7.98
008	AR162AA	AR1620924A	11/03/2009	16:12:52	1P7307	3.00	7.98
009	AR163AA	AR1630924A	11/03/2009	16:25:48	1P7307	3.00	7.98
010	AR164AA	AR1640924A	11/03/2009	16:38:47	1P7307	3.00	7.98
011	AR165AA	AR1650924A	11/03/2009	16:51:47	1P7307	3.00	7.98
012	AR481AA	AR4810924B	11/03/2009	17:04:47	1P7307	3.00	7.98
013	AR482AA	AR4820924B	11/03/2009	17:17:43	1P7307	3.00	7.98
014	AR483AA	AR4830924B	11/03/2009	17:30:43	1P7307	3.00	7.98
015	AR484AA	AR4840924B	11/03/2009	17:43:41	1P7307	3.00	7.98
016	AR485AA	AR4850924B	11/03/2009	17:56:41	1P7307	3.00	7.98
017	AR541AA	AR5410924B	11/03/2009	18:09:39	1P7307	3.00	7.98
018	AR542AA	AR5420924B	11/03/2009	18:22:38	1P7307	3.00	7.98
019	AR543AA	AR5430924B	11/03/2009	18:35:34	1P7307	3.00	7.98
020	AR544AA	AR5440924B	11/03/2009	18:48:34	1P7307	3.00	7.98
021	AR545AA	AR5450924B	11/03/2009	19:01:35	1P7307	3.00	7.97
022	AR621AA	AR6210924B	11/03/2009	19:14:35	1P7307	3.02	7.97
023	AR622AA	AR6220924B	11/03/2009	19:27:32	1P7307	2.99	7.97
024	AR623AA	AR6230924B	11/03/2009	19:40:31	1P7307	3.02	7.97
025	AR624AA	AR6240924B	11/03/2009	19:53:29	1P7307	3.02	7.97
026	AR625AA	AR6250924B	11/03/2009	20:06:31	1P7307	3.01	7.98
027	AR681AA	AR6810924C	11/03/2009	20:19:30	1P7307	2.98	7.98
028	AR682AA	AR6820924C	11/03/2009	20:32:31	1P7307		7.98
029	AR683AA	AR6830924C	11/03/2009	20:45:28	1P7307	2.98	7.98
030	AR684AA	AR6840924C	11/03/2009	20:58:28	1P7307		7.97
031	AR685AA	AR6850924C	11/03/2009	21:11:28	1P7307	2.99	7.97

	ICAL Dates		ICAL RT	QC Limits
1P7307	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	3.	(2.97 - 3.03 Minutes)
		DCB = Decachlorobiphenyl	7.98	(7.95 - 8.01 Minutes)

aSX22 0264

8D
ANALYTICAL SEQUENCE

Sequence: 1P7307 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTXCLPI ID: 0.32

Instrument: H5088A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
032	AR213AA	AR2130924B	11/03/2009	21:24:27	1P7307	3.00	7.98
033	AR32XAA	AR32X0924B	11/03/2009	21:37:25	1P7307	3.00	7.98
034	AR423AA	AR4230924B	11/03/2009	21:50:24	1P7307	3.00	7.97
035	AR16XAA	AR16X0924A	11/03/2009	22:03:25	1P7307	3.00	7.98
036	MD16XAA	MD16X0924A	11/03/2009	22:16:26	1P7307	3.00	7.98
037	MD62XAA	MD62X0924B	11/03/2009	22:29:25	1P7307	3.00	7.98
038	MD68XAA	MD68X0924C	11/03/2009	22:42:25	1P7307		7.98
039	IC16XAA	IC16X0924A	11/03/2009	22:55:21	1P7307	3.00	7.98
040	IC48XAA	IC48X0924B	11/03/2009	23:08:22	1P7307	3.00	7.97
041	IC54XAA	IC54X0924B	11/03/2009	23:21:23	1P7307	3.00	7.98
042	PBLK47	BLANKA	11/03/2009	23:34:24	1P7307	3.00	7.97
043	LCS9Q	LCSA	11/03/2009	23:47:23	1P7307	3.00	7.97
044	LCSDOD	LCSDA	11/04/2009	00:00:22	1P7307	3.00	7.98
045	21FE-	5812843	11/04/2009	00:13:23	1P7307	3.00	7.98
046	WW101	5815988	11/04/2009	00:26:25	1P7307	3.00	7.97
047	AR163OH	AR1630924A	11/04/2009	00:39:25	1P7307	3.00	7.98
048	PIBLKEY	IBLKX0924B	11/04/2009	00:52:27	1P7307	3.00	7.97
049	LOB-N	5813449	11/04/2009	01:05:25	1P7307		
050	LO--E	5813450	11/04/2009	01:18:27	1P7307		
051	A4VP1	5813451	11/04/2009	01:31:29	1P7307		
052	50VP1	5813453	11/04/2009	01:44:32	1P7307		
053	50VP1MS	5813454	11/04/2009	01:57:31	1P7307		
054	50VP1MSD	5813455	11/04/2009	02:10:31	1P7307		
055	70V10	5813458	11/04/2009	02:23:34	1P7307		
056	13V05	5813462	11/04/2009	02:36:39	1P7307		
057	AA1FD	5813470	11/04/2009	02:49:41	1P7307		7.98
058	CSF-6	5813158	11/04/2009	03:02:42	1P7307	3.00	7.98
059	CSF12	5813164	11/04/2009	03:15:41	1P7307		
060	AR162NS	AR1620924A	11/04/2009	03:28:44	1P7307	3.00	7.97
061	PIBLKEY	IBLKX0924B	11/04/2009	03:41:47	1P7307	3.00	7.98
062	PBLK2S	BLANKA	11/04/2009	03:54:53	1P7307	3.00	7.98

	ICAL Dates		ICAL RT	QC Limits
1P7307	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	3.	(2.97 - 3.03 Minutes)
		DCB = Decachlorobiphenyl	7.98	(7.95 - 8.01 Minutes)

ASX22 0265

8D
ANALYTICAL SEQUENCE

Sequence: 1P7307 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTXCLPI ID: 0.32

Instrument: H5088A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
063	LCS8E	LCSA	11/04/2009	04:07:54	1P7307	3.00	7.98
064	LCSDNK	LCSDA	11/04/2009	04:20:56	1P7307	3.00	7.97
065	PPRF4	5820183	11/04/2009	04:33:58	1P7307	3.00	7.97
066	PPR13	5820184	11/04/2009	04:47:02	1P7307	3.00	7.97
067	PPR05	5820185	11/04/2009	05:00:03	1P7307	3.00	7.97
068	PPR15	5820186	11/04/2009	05:13:05	1P7307	3.00	7.97
069	PPRF5	5820187	11/04/2009	05:26:03	1P7307	3.00	7.98
070	PPR17	5820188	11/04/2009	05:39:05	1P7307	3.00	7.97
071	PPR16	5820189	11/04/2009	05:52:09	1P7307	3.00	7.97
072	AR163OK	AR1630924A	11/04/2009	06:05:15	1P7307	3.00	7.98
073	PIBLKEY	IBLKX0924B	11/04/2009	06:18:16	1P7307	3.00	7.97
074	PBLK4B	BLANKA	11/04/2009	06:31:18	1P7307	3.00	7.97
075	LCS9U	LCSA	11/04/2009	06:44:19	1P7307	3.00	7.97
076	SMNES	5815455	11/04/2009	06:57:25	1P7307	3.01	7.98
077	G-7-8	5816565	11/04/2009	07:10:29	1P7307	3.00	7.97
078	G-7-8MS	5816565	11/04/2009	07:23:34	1P7307	3.00	7.98
079	G-7-8MSD	5816565	11/04/2009	07:36:38	1P7307	3.00	7.98
080	F-CAR	5819873	11/04/2009	07:49:42	1P7307	3.00	7.97
081	7-CAR	5819874	11/04/2009	08:02:48	1P7307	3.00	7.98
082	8-CAR	5819875	11/04/2009	08:15:54	1P7307	3.00	7.98
083	9-CAR	5819876	11/04/2009	08:28:58	1P7307	3.00	7.98
084	AR164FF	AR1640924A	11/04/2009	08:42:04	1P7307	3.00	7.98
085	PIBLKEY	IBLKX0924B	11/04/2009	08:55:11	1P7307	3.00	7.98

	ICAL Dates		ICAL RT	QC Limits
1P7307	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	3.	(2.97 - 3.03 Minutes)
		DCB = Decachlorobiphenyl	7.98	(7.95 - 8.01 Minutes)

ASX22 8266

FORM VIII PEST

8D
ANALYTICAL SEQUENCE

Sequence: 1P7307B Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTXCLPII ID: 0.32

Instrument: H5088B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
001		CONDITIONER	11/03/2009	14:28:03	1P7307B		
002		CONDITIONER	11/03/2009	14:41:03	1P7307B		
003		CONDITIONER	11/03/2009	14:53:58	1P7307B		
004		CONDITIONER	11/03/2009	15:06:57	1P7307B	2.34	6.80
005		CONDITIONER	11/03/2009	15:19:54	1P7307B	2.34	6.81
006	PIBLKAA	IBLKX0924B	11/03/2009	15:45:59	1P7307B	2.34	6.80
007	AR161AA	AR1610924A	11/03/2009	15:59:53	1P7307B	2.34	6.80
008	AR162AA	AR1620924A	11/03/2009	16:12:52	1P7307B	2.34	6.80
009	AR163AA	AR1630924A	11/03/2009	16:25:48	1P7307B	2.34	6.80
010	AR164AA	AR1640924A	11/03/2009	16:38:47	1P7307B	2.34	6.80
011	AR165AA	AR1650924A	11/03/2009	16:51:47	1P7307B	2.34	6.80
012	AR481AA	AR4810924B	11/03/2009	17:04:47	1P7307B	2.34	6.80
013	AR482AA	AR4820924B	11/03/2009	17:17:43	1P7307B	2.34	6.80
014	AR483AA	AR4830924B	11/03/2009	17:30:43	1P7307B	2.34	6.80
015	AR484AA	AR4840924B	11/03/2009	17:43:41	1P7307B	2.34	6.80
016	AR485AA	AR4850924B	11/03/2009	17:56:41	1P7307B	2.34	6.80
017	AR541AA	AR5410924B	11/03/2009	18:09:39	1P7307B	2.34	6.80
018	AR542AA	AR5420924B	11/03/2009	18:22:38	1P7307B	2.34	6.80
019	AR543AA	AR5430924B	11/03/2009	18:35:34	1P7307B	2.34	6.80
020	AR544AA	AR5440924B	11/03/2009	18:48:34	1P7307B	2.34	6.80
021	AR545AA	AR5450924B	11/03/2009	19:01:35	1P7307B	2.34	6.80
022	AR621AA	AR6210924B	11/03/2009	19:14:35	1P7307B		6.81
023	AR622AA	AR6220924B	11/03/2009	19:27:32	1P7307B		6.80
024	AR623AA	AR6230924B	11/03/2009	19:40:31	1P7307B		6.80
025	AR624AA	AR6240924B	11/03/2009	19:53:29	1P7307B		6.80
026	AR625AA	AR6250924B	11/03/2009	20:06:31	1P7307B		6.80
027	AR681AA	AR6810924C	11/03/2009	20:19:30	1P7307B		6.80
028	AR682AA	AR6820924C	11/03/2009	20:32:31	1P7307B		6.80
029	AR683AA	AR6830924C	11/03/2009	20:45:28	1P7307B		6.80
030	AR684AA	AR6840924C	11/03/2009	20:58:28	1P7307B		6.80
031	AR685AA	AR6850924C	11/03/2009	21:11:28	1P7307B		6.80
032	AR213AA	AR2130924B	11/03/2009	21:24:27	1P7307B	2.34	6.80
033	AR32XAA	AR32X0924B	11/03/2009	21:37:25	1P7307B	2.34	6.80

	ICAL Dates		ICAL RT	QC Limits
1P7307B	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	2.34	(2.31 - 2.37 Minutes)
		DCB = Decachlorobiphenyl	6.8	(6.77 - 6.83 Minutes)

ASX22 #267

FORM VIII PEST

8D
ANALYTICAL SEQUENCE

Sequence: 1P7307B | Lab Name: Lancaster laboratories | Contract:

Lab Code: | Case No.: | SAS No: | SDG No.:

GC Column: RTXCLPII | ID: 0.32

Instrument: H5088B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
034	AR423AA	AR4230924B	11/03/2009	21:50:24	1P7307B	2.34	6.80
035	AR16XAA	AR16X0924A	11/03/2009	22:03:25	1P7307B	2.34	6.80
036	MD16XAA	MD16X0924A	11/03/2009	22:16:26	1P7307B	2.34	6.80
037	MD62XAA	MD62X0924B	11/03/2009	22:29:25	1P7307B		6.80
038	MD68XAA	MD68X0924C	11/03/2009	22:42:25	1P7307B		6.80
039	IC16XAA	IC16X0924A	11/03/2009	22:55:21	1P7307B	2.34	6.80
040	IC48XAA	IC48X0924B	11/03/2009	23:08:22	1P7307B	2.34	6.80
041	IC54XAA	IC54X0924B	11/03/2009	23:21:23	1P7307B	2.34	6.80
042	PBLK47	BLANKA	11/03/2009	23:34:24	1P7307B	2.34	6.80
043	LCS9Q	LCSA	11/03/2009	23:47:23	1P7307B	2.34	6.80
044	LCSDOD	LCSDA	11/04/2009	00:00:22	1P7307B	2.34	6.80
045	21FE-	5812843	11/04/2009	00:13:23	1P7307B	2.34	6.80
046	WW101	5815988	11/04/2009	00:26:25	1P7307B	2.34	6.80
047	AR163OH	AR1630924A	11/04/2009	00:39:25	1P7307B	2.34	6.80
048	PIBLKEY	IBLKX0924B	11/04/2009	00:52:27	1P7307B	2.34	6.80
049	LOB-N	5813449	11/04/2009	01:05:25	1P7307B	2.34	6.80
050	LO--E	5813450	11/04/2009	01:18:27	1P7307B		6.81
051	A4VP1	5813451	11/04/2009	01:31:29	1P7307B	2.34	6.80
052	50VP1	5813453	11/04/2009	01:44:32	1P7307B	2.34	6.80
053	50VP1MS	5813454	11/04/2009	01:57:31	1P7307B	2.34	6.80
054	50VP1MSD	5813455	11/04/2009	02:10:31	1P7307B	2.34	6.80
055	70V10	5813458	11/04/2009	02:23:34	1P7307B	2.34	6.81
056	13V05	5813462	11/04/2009	02:36:39	1P7307B	2.34	6.80
057	AA1FD	5813470	11/04/2009	02:49:41	1P7307B	2.34	6.80
058	CSF-6	5813158	11/04/2009	03:02:42	1P7307B	2.34	6.80
059	CSF12	5813164	11/04/2009	03:15:41	1P7307B	2.34	6.80
060	AR162NS	AR1620924A	11/04/2009	03:28:44	1P7307B	2.34	6.80
061	PIBLKEY	IBLKX0924B	11/04/2009	03:41:47	1P7307B	2.34	6.80
062	PBLK2S	BLANKA	11/04/2009	03:54:53	1P7307B	2.34	6.80
063	LCS8E	LCSA	11/04/2009	04:07:54	1P7307B	2.34	6.80
064	LCSDNK	LCSDA	11/04/2009	04:20:56	1P7307B	2.34	6.80
065	PPRF4	5820183	11/04/2009	04:33:58	1P7307B	2.34	6.80
066	PPR13	5820184	11/04/2009	04:47:02	1P7307B	2.34	6.80

	ICAL Dates		ICAL RT	QC Limits
1P7307B	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	2.34	(2.31 - 2.37 Minutes)
		DCB = Decachlorobiphenyl	6.8	(6.77 - 6.83 Minutes)

ASX22 8268

FORM VIII PEST

8D
ANALYTICAL SEQUENCE

Sequence: 1P7307B Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTXCLPII ID: 0.32

Instrument: H5088B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
067	PPR05	5820185	11/04/2009	05:00:03	1P7307B	2.34	6.80
068	PPR15	5820186	11/04/2009	05:13:05	1P7307B	2.34	6.80
069	PPRF5	5820187	11/04/2009	05:26:03	1P7307B	2.34	6.80
070	PPR17	5820188	11/04/2009	05:39:05	1P7307B	2.34	6.80
071	PPR16	5820189	11/04/2009	05:52:09	1P7307B	2.34	6.80
072	AR163OK	AR1630924A	11/04/2009	06:05:15	1P7307B	2.34	6.80
073	PIBLKEY	IBLKX0924B	11/04/2009	06:18:16	1P7307B	2.34	6.80
074	PBLK4B	BLANKA	11/04/2009	06:31:18	1P7307B	2.34	6.80
075	LCS9U	LCSA	11/04/2009	06:44:19	1P7307B	2.34	6.80
076	SMNES	5815455	11/04/2009	06:57:25	1P7307B		6.83
077	G-7-8	5816565	11/04/2009	07:10:29	1P7307B	2.34	6.80
078	G-7-8MS	5816565	11/04/2009	07:23:34	1P7307B	2.34	6.80
079	G-7-8MSD	5816565	11/04/2009	07:36:38	1P7307B	2.34	6.80
080	F-CAR	5819873	11/04/2009	07:49:42	1P7307B	2.34	6.80
081	7-CAR	5819874	11/04/2009	08:02:48	1P7307B	2.34	6.80
082	8-CAR	5819875	11/04/2009	08:15:54	1P7307B	2.35	6.81
083	9-CAR	5819876	11/04/2009	08:28:58	1P7307B	2.34	6.80
084	AR164FF	AR1640924A	11/04/2009	08:42:04	1P7307B	2.34	6.80
085	PIBLKEY	IBLKX0924B	11/04/2009	08:55:11	1P7307B	2.34	6.80

	ICAL Dates		ICAL RT	QC Limits
1P7307B	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	2.34	(2.31 - 2.37 Minutes)
		DCB = Decachlorobiphenyl	6.8	(6.77 - 6.83 Minutes)

RSX22 0269

FORM VIII PEST

8D
ANALYTICAL SEQUENCE

Sequence: 3P7307 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTXCLPI ID: 0.32

Instrument: H5088A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
001		CONDITIONER	11/05/2009	15:14:18	1P7307	3.00	7.99
002		CONDITIONER	11/05/2009	15:27:30	1P7307	3.00	7.98
003		CONDITIONER	11/05/2009	15:40:24	1P7307	3.00	7.98
004		CONDITIONER	11/05/2009	15:53:21	1P7307	3.00	7.98
005		CONDITIONER	11/05/2009	16:06:17	1P7307	3.00	7.98
006	AR163OU	AR1630924A	11/05/2009	16:20:10	1P7307	3.00	7.97
007	PIBLKFH	IBLKX0924B	11/05/2009	16:33:07	1P7307	3.00	7.97
008	AA	MS0926824A	11/05/2009	16:46:04	1P7307	3.01	7.98
009	PBLK90	BLANKA	11/05/2009	16:59:00	1P7307	3.00	7.97
010	LCSE5	LCSA	11/05/2009	17:11:59	1P7307	3.00	7.97
011	BLWO3	5817880	11/05/2009	17:24:56	1P7307	3.00	7.97
012	BLWO3MS	5817880	11/05/2009	17:37:54	1P7307	3.00	7.97
013	BLWO3MSD	5817880	11/05/2009	17:50:49	1P7307	3.00	7.97
014	92146	5822168	11/05/2009	18:03:47	1P7307	3.00	7.97
015	89FS1	5822197	11/05/2009	18:16:46	1P7307	3.01	7.97
016	AR164FM	AR1640924A	11/05/2009	18:29:47	1P7307	3.00	7.97
017	PIBLKFH	IBLKX0924B	11/05/2009	18:42:44	1P7307	3.00	7.97
018	PBLK9I	BLANKA	11/05/2009	18:55:42	1P7307	3.00	7.97
019	LCSEQ	LCSA	11/05/2009	19:08:39	1P7307	3.00	7.97
020	LCSDRC	LCSDA	11/05/2009	19:21:37	1P7307	3.00	7.97
021	CSFEB	5813174	11/05/2009	19:34:36	1P7307	3.00	7.97
022	AA1EB	5813471	11/05/2009	19:47:35	1P7307	3.00	7.97
023	44001	5826632	11/05/2009	20:00:31	1P7307	3.01	
024	44002	5826633	11/05/2009	20:13:30	1P7307	3.01	
025	44004	5826635	11/05/2009	20:26:30	1P7307	3.01	
026	44005	5826636	11/05/2009	20:39:31	1P7307	3.01	7.99

	ICAL Dates		ICAL RT	QC Limits
1P7307	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	3.	(2.97 - 3.03 Minutes)
		DCB = Decachlorobiphenyl	7.98	(7.95 - 8.01 Minutes)

ASX22 8278

8D
ANALYTICAL SEQUENCE

Sequence: 3P7307 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTXCLPI ID: 0.32

Instrument: H5088A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
027	44006	5826637	11/05/2009	20:52:28	1P7307	3.00	7.98
028	AR163OU	AR1630924A	11/05/2009	21:05:27	1P7307	3.00	7.98
029	PIBLKFH	IBLKX0924B	11/05/2009	21:18:24	1P7307	3.00	7.98
030	PBLK4C	BLANKA	11/05/2009	21:31:23	1P7307	3.00	7.98
031	LCS9V	LCSA	11/05/2009	21:44:22	1P7307	3.00	7.98
032	D1305	5820657	11/05/2009	21:57:23	1P7307	3.00	7.98
033	D1310	5820658	11/05/2009	22:10:21	1P7307	3.00	7.98
034	D1315	5820659	11/05/2009	22:23:20	1P7307	3.00	7.98
035	D1320	5820660	11/05/2009	22:36:20	1P7307	3.00	7.98
036	DFDP2	5820661	11/05/2009	22:49:22	1P7307	3.00	7.98
037	D1425	5820662	11/05/2009	23:02:21	1P7307	3.00	7.97
038	D1430	5820663	11/05/2009	23:15:21	1P7307	3.00	7.98
039	D1435	5820664	11/05/2009	23:28:19	1P7307	3.00	7.97
040	AR164FN	AR1640924A	11/05/2009	23:41:19	1P7307	3.00	7.97
041	PIBLKFH	IBLKX0924B	11/05/2009	23:54:21	1P7307	3.00	7.98
042	PBLK65	BLANKA	11/06/2009	00:07:24	1P7307	3.00	7.98
043	LCSBG	LCSA	11/06/2009	00:20:23	1P7307	3.00	7.98
044	EBS28	5822062	11/06/2009	00:33:23	1P7307	3.00	7.98
045	CARS1	5822063	11/06/2009	00:46:24	1P7307		
046	CARS2	5822064	11/06/2009	00:59:28	1P7307		
047	CARS2MS	5822065	11/06/2009	01:12:28	1P7307		
048	CARS2MSD	5822066	11/06/2009	01:25:29	1P7307		
049	DPW29	5822067	11/06/2009	01:38:29	1P7307		
050	STEA1	5822068	11/06/2009	01:51:31	1P7307	3.00	
051	BB-1	5822078	11/06/2009	02:04:34	1P7307	3.00	7.97
052	AR163OU	AR1630924A	11/06/2009	02:17:39	1P7307	3.00	7.97

	ICAL Dates		ICAL RT	QC Limits
1P7307	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	3.	(2.97 - 3.03 Minutes)
		DCB = Decachlorobiphenyl	7.98	(7.95 - 8.01 Minutes)

ASX22 0271

8D
ANALYTICAL SEQUENCE

Sequence: 3P7307 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTXCLPI ID: 0.32

Instrument: H5088A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
053	PIBLKFH	IBLKX0924B	11/06/2009	02:30:39	1P7307	3.00	7.98
054	EBW29	5822079	11/06/2009	02:43:40	1P7307	3.00	7.98
055	EBS29	5822080	11/06/2009	02:56:41	1P7307	3.00	7.98
056	FE101	5822187	11/06/2009	03:09:46	1P7307	3.00	7.97
057	FE102	5822188	11/06/2009	03:22:49	1P7307	3.00	7.98
058	FE103	5822189	11/06/2009	03:35:52	1P7307	3.00	7.97
059	FE201	5822190	11/06/2009	03:48:53	1P7307	3.00	7.98
060	FE202	5822191	11/06/2009	04:01:56	1P7307	3.00	7.98
061	FEM4-	5822192	11/06/2009	04:15:01	1P7307	3.00	7.98
062	FEM5-	5822193	11/06/2009	04:28:05	1P7307	3.00	7.97
063	AR164FO	AR1640924A	11/06/2009	04:41:07	1P7307	3.00	7.97
064	PIBLKFH	IBLKX0924B	11/06/2009	04:54:09	1P7307	3.00	7.97
065	LOCO6	5822047	11/06/2009	05:07:11	1P7307		
066	LOCO7	5822048	11/06/2009	05:20:19	1P7307		
067	PBLK7O	BLANKA	11/06/2009	05:33:22	1P7307	3.00	7.97
068	LCSD2	LCSA	11/06/2009	05:46:26	1P7307	3.00	7.97
069	F0544	5823532	11/06/2009	05:59:28	1P7307	3.00	7.97
070	20-25	5821263	11/06/2009	06:12:32	1P7307	3.00	7.97
071	20-25MS	5821263	11/06/2009	06:25:39	1P7307	3.00	7.97
072	20-25MSD	5821263	11/06/2009	06:38:43	1P7307	3.00	7.97
073	20-75	5821264	11/06/2009	06:51:47	1P7307	3.00	7.97
074	10000	5821265	11/06/2009	07:04:51	1P7307	3.00	7.97
075	AR163OU	AR1630924A	11/06/2009	07:17:53	1P7307	3.00	7.97
076	PIBLKFH	IBLKX0924B	11/06/2009	07:31:01	1P7307	3.00	7.97
077	10017	5821266	11/06/2009	07:44:05	1P7307	3.00	7.97
078	10100	5821267	11/06/2009	07:57:10	1P7307	3.00	7.97

	ICAL Dates		ICAL RT	QC Limits
1P7307	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	3.	(2.97 - 3.03 Minutes)
		DCB = Decachlorobiphenyl	7.98	(7.95 - 8.01 Minutes)

ASX22 0272

8D
ANALYTICAL SEQUENCE

Sequence: 3P7307 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTXCLPI ID: 0.32

Instrument: H5088A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
079	10150	5821268	11/06/2009	08:10:15	1P7307	3.00	7.97
080	10200	5821269	11/06/2009	08:23:23	1P7307	3.00	7.97
081	10250	5821270	11/06/2009	08:36:28	1P7307	3.00	7.97
082	PBLK6A	BLANKA	11/06/2009	08:49:34	1P7307	3.00	7.97
083	LCSBM	LCSA	11/06/2009	09:02:38	1P7307	3.00	7.98
084	LCSDPH	LCSDA	11/06/2009	09:15:44	1P7307	3.00	7.98
085	36AR1	5822183	11/06/2009	09:28:48	1P7307	3.01	7.98
086	36AC1	5822184	11/06/2009	09:41:56	1P7307	3.01	7.98
087	AR164FP	AR1640924A	11/06/2009	09:54:59	1P7307	3.00	7.97
088	PIBLKFH	IBLKX0924B	11/06/2009	10:08:02	1P7307	3.00	7.97
089	36AC2	5822185	11/06/2009	10:21:05	1P7307	3.01	7.97
090	36AC3	5822186	11/06/2009	10:34:11	1P7307	3.01	7.97
091	AR162OJ	AR1620924A	11/06/2009	10:47:15	1P7307	3.00	7.97
092	PIBLKFH	IBLKX0924B	11/06/2009	11:00:19	1P7307	3.00	7.97

	ICAL Dates		ICAL RT	QC Limits
1P7307	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	3.	(2.97 - 3.03 Minutes)
		DCB = Decachlorobiphenyl	7.98	(7.95 - 8.01 Minutes)

ASX22 0273

8D
ANALYTICAL SEQUENCE

Sequence: 3P7307B | Lab Name: Lancaster laboratories | Contract:

Lab Code: | Case No.: | SAS No: | SDG No.:

GC Column: RTXCLPII | ID: 0.32

Instrument: H5088B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
001		CONDITIONER	11/05/2009	15:14:18	1P7307B	2.34	6.81
002		CONDITIONER	11/05/2009	15:27:30	1P7307B	2.34	6.80
003		CONDITIONER	11/05/2009	15:40:24	1P7307B	2.34	6.80
004		CONDITIONER	11/05/2009	15:53:21	1P7307B	2.34	6.80
005		CONDITIONER	11/05/2009	16:06:17	1P7307B	2.34	6.80
006	AR163OU	AR1630924A	11/05/2009	16:20:10	1P7307B	2.34	6.80
007	PIBLKFH	IBLKX0924B	11/05/2009	16:33:07	1P7307B	2.34	6.80
008	AA	MS0926824A	11/05/2009	16:46:04	1P7307B	2.35	6.80
009	PBLK90	BLANKA	11/05/2009	16:59:00	1P7307B	2.34	6.80
010	LCSE5	LCSA	11/05/2009	17:11:59	1P7307B	2.34	6.80
011	BLWO3	5817880	11/05/2009	17:24:56	1P7307B	2.34	6.80
012	BLWO3MS	5817880	11/05/2009	17:37:54	1P7307B	2.34	6.80
013	BLWO3MSD	5817880	11/05/2009	17:50:49	1P7307B	2.34	6.80
014	92146	5822168	11/05/2009	18:03:47	1P7307B	2.34	6.79
015	89FS1	5822197	11/05/2009	18:16:46	1P7307B	2.35	6.80
016	AR164FM	AR1640924A	11/05/2009	18:29:47	1P7307B	2.34	6.80
017	PIBLKFH	IBLKX0924B	11/05/2009	18:42:44	1P7307B	2.34	6.80
018	PBLK9I	BLANKA	11/05/2009	18:55:42	1P7307B	2.34	6.80
019	LCSEQ	LCSA	11/05/2009	19:08:39	1P7307B	2.34	6.80
020	LCSDRC	LCSDA	11/05/2009	19:21:37	1P7307B	2.34	6.80
021	CSFEB	5813174	11/05/2009	19:34:36	1P7307B	2.34	6.80
022	AA1EB	5813471	11/05/2009	19:47:35	1P7307B	2.34	6.80
023	44001	5826632	11/05/2009	20:00:31	1P7307B	2.35	
024	44002	5826633	11/05/2009	20:13:30	1P7307B	2.35	
025	44004	5826635	11/05/2009	20:26:30	1P7307B	2.35	
026	44005	5826636	11/05/2009	20:39:31	1P7307B	2.35	6.81
027	44006	5826637	11/05/2009	20:52:28	1P7307B	2.34	6.80
028	AR163OU	AR1630924A	11/05/2009	21:05:27	1P7307B	2.34	6.80
029	PIBLKFH	IBLKX0924B	11/05/2009	21:18:24	1P7307B	2.34	6.80
030	PBLK4C	BLANKA	11/05/2009	21:31:23	1P7307B	2.34	6.80
031	LCS9V	LCSA	11/05/2009	21:44:22	1P7307B	2.34	6.80

	ICAL Dates		ICAL RT	QC Limits
1P7307B	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	2.34	(2.31 - 2.37 Minutes)
		DCB = Decachlorobiphenyl	6.8	(6.77 - 6.83 Minutes)

ASX22 0274

8D
ANALYTICAL SEQUENCE

Sequence: 3P7307B Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTXCLPII ID: 0.32

Instrument: H5088B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
032	D1305	5820657	11/05/2009	21:57:23	1P7307B	2.34	6.80
033	D1310	5820658	11/05/2009	22:10:21	1P7307B	2.34	6.80
034	D1315	5820659	11/05/2009	22:23:20	1P7307B	2.34	6.80
035	D1320	5820660	11/05/2009	22:36:20	1P7307B	2.34	6.80
036	DFDP2	5820661	11/05/2009	22:49:22	1P7307B	2.34	6.80
037	D1425	5820662	11/05/2009	23:02:21	1P7307B	2.34	6.80
038	D1430	5820663	11/05/2009	23:15:21	1P7307B	2.34	6.80
039	D1435	5820664	11/05/2009	23:28:19	1P7307B	2.34	6.80
040	AR164FN	AR1640924A	11/05/2009	23:41:19	1P7307B	2.34	6.80
041	PIBLKFH	IBLKX0924B	11/05/2009	23:54:21	1P7307B	2.34	6.80
042	PBLK65	BLANKA	11/06/2009	00:07:24	1P7307B	2.34	6.80
043	LCSBG	LCSA	11/06/2009	00:20:23	1P7307B	2.34	6.80
044	EBS28	5822062	11/06/2009	00:33:23	1P7307B	2.34	6.80
045	CARS1	5822063	11/06/2009	00:46:24	1P7307B	2.34	6.81
046	CARS2	5822064	11/06/2009	00:59:28	1P7307B	2.34	6.80
047	CARS2MS	5822065	11/06/2009	01:12:28	1P7307B	2.34	6.80
048	CARS2MSD	5822066	11/06/2009	01:25:29	1P7307B	2.34	6.80
049	DPW29	5822067	11/06/2009	01:38:29	1P7307B	2.34	6.80
050	STEA1	5822068	11/06/2009	01:51:31	1P7307B	2.34	6.82
051	BB–1	5822078	11/06/2009	02:04:34	1P7307B	2.34	6.80
052	AR163OU	AR1630924A	11/06/2009	02:17:39	1P7307B	2.34	6.80
053	PIBLKFH	IBLKX0924B	11/06/2009	02:30:39	1P7307B	2.34	6.80
054	EBW29	5822079	11/06/2009	02:43:40	1P7307B	2.34	6.80
055	EBS29	5822080	11/06/2009	02:56:41	1P7307B	2.34	6.80
056	FE101	5822187	11/06/2009	03:09:46	1P7307B	2.34	6.80
057	FE102	5822188	11/06/2009	03:22:49	1P7307B	2.34	6.80
058	FE103	5822189	11/06/2009	03:35:52	1P7307B	2.34	6.80
059	FE201	5822190	11/06/2009	03:48:53	1P7307B	2.34	6.80
060	FE202	5822191	11/06/2009	04:01:56	1P7307B	2.35	6.80
061	FEM4-	5822192	11/06/2009	04:15:01	1P7307B	2.34	6.80
062	FEM5-	5822193	11/06/2009	04:28:05	1P7307B	2.34	6.80

	ICAL Dates		ICAL RT	QC Limits
1P7307B	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	2.34	(2.31 - 2.37 Minutes)
		DCB = Decachlorobiphenyl	6.8	(6.77 - 6.83 Minutes)

ASX22 0275

8D
ANALYTICAL SEQUENCE

Sequence: 3P7307B Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTXCLPII ID: 0.32

Instrument: H5088B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
063	AR164FO	AR1640924A	11/06/2009	04:41:07	1P7307B	2.34	6.80
064	PIBLKFH	IBLKX0924B	11/06/2009	04:54:09	1P7307B	2.34	6.80
065	LOCO6	5822047	11/06/2009	05:07:11	1P7307B	2.34	6.81
066	LOCO7	5822048	11/06/2009	05:20:19	1P7307B	2.34	6.81
067	PBLK7O	BLANKA	11/06/2009	05:33:22	1P7307B	2.34	6.80
068	LCSD2	LCSA	11/06/2009	05:46:26	1P7307B	2.34	6.80
069	F0544	5823532	11/06/2009	05:59:28	1P7307B	2.34	6.80
070	20-25	5821263	11/06/2009	06:12:32	1P7307B	2.34	6.80
071	20-25MS	5821263	11/06/2009	06:25:39	1P7307B	2.34	6.80
072	20-25MSD	5821263	11/06/2009	06:38:43	1P7307B	2.34	6.80
073	20-75	5821264	11/06/2009	06:51:47	1P7307B	2.34	6.80
074	10000	5821265	11/06/2009	07:04:51	1P7307B	2.34	6.80
075	AR163OU	AR1630924A	11/06/2009	07:17:53	1P7307B	2.34	6.80
076	PIBLKFH	IBLKX0924B	11/06/2009	07:31:01	1P7307B	2.34	6.80
077	10017	5821266	11/06/2009	07:44:05	1P7307B	2.34	6.80
078	10100	5821267	11/06/2009	07:57:10	1P7307B	2.34	6.80
079	10150	5821268	11/06/2009	08:10:15	1P7307B	2.34	6.80
080	10200	5821269	11/06/2009	08:23:23	1P7307B	2.34	6.80
081	10250	5821270	11/06/2009	08:36:28	1P7307B	2.34	6.79
082	PBLK6A	BLANKA	11/06/2009	08:49:34	1P7307B	2.34	6.80
083	LCSBM	LCSA	11/06/2009	09:02:38	1P7307B	2.34	6.80
084	LCSDPH	LCSDA	11/06/2009	09:15:44	1P7307B	2.34	6.80
085	36AR1	5822183	11/06/2009	09:28:48	1P7307B	2.35	6.80
086	36AC1	5822184	11/06/2009	09:41:56	1P7307B	2.35	6.80
087	AR164FP	AR1640924A	11/06/2009	09:54:59	1P7307B	2.34	6.80
088	PIBLKFH	IBLKX0924B	11/06/2009	10:08:02	1P7307B	2.34	6.80
089	36AC2	5822185	11/06/2009	10:21:05	1P7307B	2.35	6.80
090	36AC3	5822186	11/06/2009	10:34:11	1P7307B	2.35	6.80
091	AR162OJ	AR1620924A	11/06/2009	10:47:15	1P7307B	2.34	6.80
092	PIBLKFH	IBLKX0924B	11/06/2009	11:00:19	1P7307B	2.34	6.80

	ICAL Dates		ICAL RT	QC Limits
1P7307B	11/03/2009 - 11/03/2009	TCX = Tetrachloro-m-xylene	2.34	(2.31 - 2.37 Minutes)
		DCB = Decachlorobiphenyl	6.8	(6.77 - 6.83 Minutes)

ASX22 8276

8D
ANALYTICAL SEQUENCE

Sequence: 1C6310 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLP ID: 0.32

Instrument: S9147A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
001		CONDITIONER	11/06/2009	15:05:28	1C6310	3.31	7.16
002		CONDITIONER	11/06/2009	15:17:24	1C6310	3.31	7.16
003		CONDITIONER	11/06/2009	15:29:25	1C6310	3.31	7.16
004	PIBLKAA	IBLKX0924B	11/06/2009	15:41:27	1C6310	3.31	7.16
005	AR161AA	AR1610924A	11/06/2009	15:53:29	1C6310	3.31	7.16
006	AR162AA	AR1620924A	11/06/2009	16:05:29	1C6310	3.31	7.16
007	AR163AA	AR1630924A	11/06/2009	16:17:25	1C6310	3.31	7.16
008	AR164AA	AR1640924A	11/06/2009	16:29:24	1C6310	3.31	7.16
009	AR165AA	AR1650924A	11/06/2009	16:41:23	1C6310	3.31	7.16
010	AR481AA	AR4810924B	11/06/2009	16:53:24	1C6310	3.31	7.16
011	AR482AA	AR4820924B	11/06/2009	17:05:21	1C6310	3.31	7.16
012	AR483AA	AR4830924B	11/06/2009	17:17:14	1C6310	3.31	7.16
013	AR484AA	AR4840924B	11/06/2009	17:29:14	1C6310	3.31	7.16
014	AR485AA	AR4850924B	11/06/2009	17:41:15	1C6310	3.31	7.16
015	AR541AA	AR5410924B	11/06/2009	17:53:10	1C6310	3.31	7.16
016	AR542AA	AR5420924B	11/06/2009	18:05:05	1C6310	3.31	7.16
017	AR543AA	AR5430924B	11/06/2009	18:17:07	1C6310	3.31	7.16
018	AR544AA	AR5440924B	11/06/2009	18:29:06	1C6310	3.31	7.16
019	AR545AA	AR5450924B	11/06/2009	18:41:07	1C6310	3.31	7.16
020	AR621AA	AR6210924B	11/06/2009	18:53:06	1C6310	3.31	7.16
021	AR622AA	AR6220924B	11/06/2009	19:05:01	1C6310	3.33 *	7.16
022	AR623AA	AR6230924B	11/06/2009	19:17:03	1C6310	3.31	7.16
023	AR624AA	AR6240924B	11/06/2009	19:29:04	1C6310	3.30	7.16
024	AR625AA	AR6250924B	11/06/2009	19:41:03	1C6310	3.30	
025	AR681AA	AR6810924C	11/06/2009	19:53:03	1C6310	3.33	7.16
026	AR682AA	AR6820924C	11/06/2009	20:04:58	1C6310		7.16
027	AR683AA	AR6830924C	11/06/2009	20:16:55	1C6310	3.33	7.16
028	AR684AA	AR6840924C	11/06/2009	20:28:54	1C6310		7.16
029	AR685AA	AR6850924C	11/06/2009	20:40:54	1C6310	3.33	7.16
030	AR213AA	AR2130924B	11/06/2009	20:52:53	1C6310	3.31	7.16
031	AR32XAA	AR32X0924B	11/06/2009	21:04:50	1C6310	3.31	7.16

	ICAL Dates		ICAL RT	QC Limits
1C6310	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	3.31	(3.29 - 3.33 Minutes)
		DCB = Decachlorobiphenyl	7.16	(7.14 - 7.18 Minutes)

ASX22 0277

8D
ANALYTICAL SEQUENCE

Sequence: 1C6310 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLP ID: 0.32

Instrument: S9147A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
032	AR423AA	AR4230924B	11/06/2009	21:16:45	1C6310	3.31	7.16
033	AR16XAA	AR16X0924A	11/06/2009	21:28:52	1C6310	3.31	7.16
034	MD16XAA	MD16X0924A	11/06/2009	21:40:51	1C6310	3.31	7.16
035	MD62XAA	MD62X0924B	11/06/2009	21:52:51	1C6310	3.31	7.16
036	MD68XAA	MD68X0924C	11/06/2009	22:04:48	1C6310	3.33	7.16
037	IC16XAA	IC16X0924A	11/06/2009	22:16:45	1C6310	3.31	7.16
038	IC48XAA	IC48X0924B	11/06/2009	22:28:49	1C6310	3.31	7.16
039	IC54XAA	IC54X0924B	11/06/2009	22:40:48	1C6310	3.31	7.16
040	PBLK9I	BLANKA	11/06/2009	22:52:47	1C6310	3.31	7.16
041	LCSEQ	LCSA	11/06/2009	23:04:45	1C6310	3.31	7.16
042	LCSDRC	LCSDA	11/06/2009	23:16:41	1C6310	3.31	7.16
043	44001	5826632	11/06/2009	23:28:42	1C6310	3.31	7.16
044	44001	5826632	11/06/2009	23:40:39	1C6310	3.31	7.16
045	44002	5826633	11/06/2009	23:52:40	1C6310	3.31	7.16
046	44002	5826633	11/07/2009	00:04:38	1C6310	3.31	7.16
047	44004	5826635	11/07/2009	00:16:33	1C6310	3.31	7.16
048	44004	5826635	11/07/2009	00:28:32	1C6310	3.31	7.16
049	44005	5826636	11/07/2009	00:40:39	1C6310	3.31	7.16
050	AR163PB	AR1630924A	11/07/2009	00:52:39	1C6310	3.31	7.16
051	PIBLKFJ	IBLKX0924B	11/07/2009	01:04:36	1C6310	3.31	7.16
052	44005	5826636	11/07/2009	01:16:32	1C6310	3.31	7.16
053	44006	5826637	11/07/2009	01:28:32	1C6310	3.31	7.16
054	PBLK9K	BLANKA	11/07/2009	01:40:34	1C6310	3.31	7.16
055	LCSES	LCSA	11/07/2009	01:52:34	1C6310	3.31	7.16
056	PRFB1	5813035	11/07/2009	02:04:31	1C6310	3.31	7.16
057	PR-GN	5813036	11/07/2009	02:16:27	1C6310	3.31	7.16
058	PR-GS	5813037	11/07/2009	02:28:26	1C6310	3.31	7.16
059	PRBW6	5813038	11/07/2009	02:40:26	1C6310	3.31	7.16
060	PR-FD	5813039	11/07/2009	02:52:26	1C6310	3.31	7.16
061	PRFB2	5813040	11/07/2009	03:04:24	1C6310	3.31	7.16
062	AR162ON	AR1620924A	11/07/2009	03:16:19	1C6310	3.31	7.16

	ICAL Dates		ICAL RT	QC Limits
1C6310	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	3.31	(3.29 - 3.33 Minutes)
		DCB = Decachlorobiphenyl	7.16	(7.14 - 7.18 Minutes)

ASX22 0278

8D
ANALYTICAL SEQUENCE

Sequence: 1C6310 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLP ID: 0.32

Instrument: S9147A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
063	PIBLKFJ	IBLKX0924B	11/07/2009	03:28:20	1C6310	3.31	7.16
064	PRB-1	5813041	11/07/2009	03:40:20	1C6310	3.31	7.16
065	PRP22	5813042	11/07/2009	03:52:20	1C6310	3.31	7.16
066	AGWF3	5816066	11/07/2009	04:04:17	1C6310	3.31	7.16
067	AGWP1	5816067	11/07/2009	04:16:13	1C6310	3.31	7.16
068	AGW14	5816068	11/07/2009	04:28:13	1C6310	3.31	7.16
069	AGW22	5816069	11/07/2009	04:40:14	1C6310	3.31	7.16
070	AGW22MS	5816070	11/07/2009	04:52:21	1C6310	3.31	7.16
071	AGW22MSD	5816071	11/07/2009	05:04:19	1C6310	3.31	7.16
072	AGWEB	5816072	11/07/2009	05:16:15	1C6310	3.31	7.16
073	PPRF4	5820183	11/07/2009	05:28:16	1C6310	3.31	7.16
074	AR163PB	AR1630924A	11/07/2009	05:40:16	1C6310	3.31	7.16
075	PIBLKFJ	IBLKX0924B	11/07/2009	05:52:22	1C6310	3.31	7.16
076	PPR13	5820184	11/07/2009	06:04:19	1C6310	3.31	7.16
077	PPR05	5820185	11/07/2009	06:16:17	1C6310	3.31	7.16
078	PPR15	5820186	11/07/2009	06:28:17	1C6310	3.31	7.16
079	PPRF5	5820187	11/07/2009	06:40:20	1C6310	3.31	7.16
080	PPR17	5820188	11/07/2009	06:52:29	1C6310	3.31	7.16
081	PPR16	5820189	11/07/2009	07:04:26	1C6310	3.31	7.16
082	PBLK93	BLANKA	11/07/2009	07:16:23	1C6310	3.31	7.16
083	LCSE8	LCSA	11/07/2009	07:28:25	1C6310	3.31	7.16
084	10714	5823539	11/07/2009	07:40:27	1C6310	3.31	7.16
085	GTA44	5823540	11/07/2009	07:52:30	1C6310		7.17
086	AR162OO	AR1620924A	11/07/2009	08:04:25	1C6310	3.31	7.16
087	PIBLKFJ	IBLKX0924B	11/07/2009	08:16:23	1C6310	3.31	7.16
088	OLDC1	5825593	11/07/2009	08:28:24	1C6310	3.31	7.16
089	OLDC1MS	5825593	11/07/2009	08:40:25	1C6310	3.31	7.16
090	OLDC1MSD	5825593	11/07/2009	08:52:27	1C6310	3.31	7.16
091	OLDC2	5825594	11/07/2009	09:04:23	1C6310	3.31	7.16
092	OLDC3	5825595	11/07/2009	09:16:19	1C6310	3.31	7.16
093	AR164FS	AR1640924A	11/07/2009	09:28:21	1C6310	3.31	7.16

	ICAL Dates		ICAL RT	QC Limits
1C6310	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	3.31	(3.29 - 3.33 Minutes)
		DCB = Decachlorobiphenyl	7.16	(7.14 - 7.18 Minutes)

ASX22 0279

FORM VIII PEST

8D
ANALYTICAL SEQUENCE

Sequence: 1C6310 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLP ID: 0.32

Instrument: S9147A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
094	PIBLKFJ	IBLKX0924B	11/07/2009	09:40:22	1C6310	3.31	7.16

	ICAL Dates		ICAL RT	QC Limits
1C6310	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	3.31	(3.29 - 3.33 Minutes)
		DCB = Decachlorobiphenyl	7.16	(7.14 - 7.18 Minutes)

ASX22 0288

FORM VIII PEST

8D
ANALYTICAL SEQUENCE

Sequence: 1C6310B Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLPII ID: 0.32

Instrument: S9147B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
001		CONDITIONER	11/06/2009	15:05:28	1C6310B	2.81	6.36
002		CONDITIONER	11/06/2009	15:17:24	1C6310B	2.81	6.36
003		CONDITIONER	11/06/2009	15:29:25	1C6310B	2.81	6.36
004	PIBLKAA	IBLKX0924B	11/06/2009	15:41:27	1C6310B	2.81	6.36
005	AR161AA	AR1610924A	11/06/2009	15:53:29	1C6310B	2.81	6.36
006	AR162AA	AR1620924A	11/06/2009	16:05:29	1C6310B	2.81	6.36
007	AR163AA	AR1630924A	11/06/2009	16:17:25	1C6310B	2.81	6.36
008	AR164AA	AR1640924A	11/06/2009	16:29:24	1C6310B	2.81	6.36
009	AR165AA	AR1650924A	11/06/2009	16:41:23	1C6310B	2.81	6.36
010	AR481AA	AR4810924B	11/06/2009	16:53:24	1C6310B	2.81	6.36
011	AR482AA	AR4820924B	11/06/2009	17:05:21	1C6310B	2.81	6.36
012	AR483AA	AR4830924B	11/06/2009	17:17:14	1C6310B	2.81	6.36
013	AR484AA	AR4840924B	11/06/2009	17:29:14	1C6310B	2.81	6.36
014	AR485AA	AR4850924B	11/06/2009	17:41:15	1C6310B	2.81	6.36
015	AR541AA	AR5410924B	11/06/2009	17:53:10	1C6310B	2.81	6.36
016	AR542AA	AR5420924B	11/06/2009	18:05:05	1C6310B	2.81	6.36
017	AR543AA	AR5430924B	11/06/2009	18:17:07	1C6310B	2.81	6.36
018	AR544AA	AR5440924B	11/06/2009	18:29:06	1C6310B	2.81	6.36
019	AR545AA	AR5450924B	11/06/2009	18:41:07	1C6310B	2.81	6.36
020	AR621AA	AR6210924B	11/06/2009	18:53:06	1C6310B	2.81	6.36
021	AR622AA	AR6220924B	11/06/2009	19:05:01	1C6310B		6.36
022	AR623AA	AR6230924B	11/06/2009	19:17:03	1C6310B	2.79	6.36
023	AR624AA	AR6240924B	11/06/2009	19:29:04	1C6310B	2.80	6.36
024	AR625AA	AR6250924B	11/06/2009	19:41:03	1C6310B		6.36
025	AR681AA	AR6810924C	11/06/2009	19:53:03	1C6310B	2.81	6.36
026	AR682AA	AR6820924C	11/06/2009	20:04:58	1C6310B		6.36
027	AR683AA	AR6830924C	11/06/2009	20:16:55	1C6310B	2.81	6.36
028	AR684AA	AR6840924C	11/06/2009	20:28:54	1C6310B	2.81	6.36
029	AR685AA	AR6850924C	11/06/2009	20:40:54	1C6310B		6.36
030	AR213AA	AR2130924B	11/06/2009	20:52:53	1C6310B	2.81	6.36
031	AR32XAA	AR32X0924B	11/06/2009	21:04:50	1C6310B	2.81	6.36
032	AR423AA	AR4230924B	11/06/2009	21:16:45	1C6310B	2.81	6.36
033	AR16XAA	AR16X0924A	11/06/2009	21:28:52	1C6310B	2.81	6.36

	ICAL Dates		ICAL RT	QC Limits
1C6310B	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	2.81	(2.79 - 2.83 Minutes)
		DCB = Decachlorobiphenyl	6.36	(6.34 - 6.38 Minutes)

ASX22 8281

FORM VIII PEST

8D
ANALYTICAL SEQUENCE

Sequence: 1C6310B Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLPII ID: 0.32

Instrument: S9147B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
034	MD16XAA	MD16X0924A	11/06/2009	21:40:51	1C6310B	2.81	6.36
035	MD62XAA	MD62X0924B	11/06/2009	21:52:51	1C6310B	2.81	6.36
036	MD68XAA	MD68X0924C	11/06/2009	22:04:48	1C6310B		6.36
037	IC16XAA	IC16X0924A	11/06/2009	22:16:45	1C6310B	2.81	6.36
038	IC48XAA	IC48X0924B	11/06/2009	22:28:49	1C6310B	2.81	6.36
039	IC54XAA	IC54X0924B	11/06/2009	22:40:48	1C6310B	2.81	6.36
040	PBLK9I	BLANKA	11/06/2009	22:52:47	1C6310B	2.81	6.36
041	LCSEQ	LCSA	11/06/2009	23:04:45	1C6310B	2.81	6.36
042	LCSDRC	LCSDA	11/06/2009	23:16:41	1C6310B	2.81	6.36
043	44001	5826632	11/06/2009	23:28:42	1C6310B	2.81	6.36
044	44001	5826632	11/06/2009	23:40:39	1C6310B	2.81	6.36
045	44002	5826633	11/06/2009	23:52:40	1C6310B	2.81	6.36
046	44002	5826633	11/07/2009	00:04:38	1C6310B	2.81	6.36
047	44004	5826635	11/07/2009	00:16:33	1C6310B	2.81	6.36
048	44004	5826635	11/07/2009	00:28:32	1C6310B	2.81	6.36
049	44005	5826636	11/07/2009	00:40:39	1C6310B	2.81	6.36
050	AR163PB	AR1630924A	11/07/2009	00:52:39	1C6310B	2.81	6.36
051	PIBLKFJ	IBLKX0924B	11/07/2009	01:04:36	1C6310B	2.81	6.36
052	44005	5826636	11/07/2009	01:16:32	1C6310B	2.81	6.36
053	44006	5826637	11/07/2009	01:28:32	1C6310B	2.81	6.36
054	PBLK9K	BLANKA	11/07/2009	01:40:34	1C6310B	2.81	6.36
055	LCSES	LCSA	11/07/2009	01:52:34	1C6310B	2.81	6.36
056	PRFB1	5813035	11/07/2009	02:04:31	1C6310B	2.81	6.36
057	PR-GN	5813036	11/07/2009	02:16:27	1C6310B	2.81	6.36
058	PR-GS	5813037	11/07/2009	02:28:26	1C6310B	2.81	6.36
059	PRBW6	5813038	11/07/2009	02:40:26	1C6310B	2.81	6.36
060	PR-FD	5813039	11/07/2009	02:52:26	1C6310B	2.81	6.36
061	PRFB2	5813040	11/07/2009	03:04:24	1C6310B	2.81	6.36
062	AR162ON	AR1620924A	11/07/2009	03:16:19	1C6310B	2.81	6.36
063	PIBLKFJ	IBLKX0924B	11/07/2009	03:28:20	1C6310B	2.81	6.36
064	PRB-1	5813041	11/07/2009	03:40:20	1C6310B	2.81	6.36
065	PRP22	5813042	11/07/2009	03:52:20	1C6310B	2.81	6.36
066	AGWF3	5816066	11/07/2009	04:04:17	1C6310B	2.81	6.36

	ICAL Dates		ICAL RT	QC Limits
1C6310B	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	2.81	(2.79 - 2.83 Minutes)
		DCB = Decachlorobiphenyl	6.36	(6.34 - 6.38 Minutes)

ASX22 0282

FORM VIII PEST

8D
ANALYTICAL SEQUENCE

Sequence: 1C6310B Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLPII ID: 0.32

Instrument: S9147B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
067	AGWP1	5816067	11/07/2009	04:16:13	1C6310B	2.81	6.36
068	AGW14	5816068	11/07/2009	04:28:13	1C6310B	2.82	6.36
069	AGW22	5816069	11/07/2009	04:40:14	1C6310B	2.81	6.36
070	AGW22MS	5816070	11/07/2009	04:52:21	1C6310B	2.81	6.36
071	AGW22MSD	5816071	11/07/2009	05:04:19	1C6310B	2.81	6.36
072	AGWEB	5816072	11/07/2009	05:16:15	1C6310B	2.81	6.36
073	PPRF4	5820183	11/07/2009	05:28:16	1C6310B	2.81	6.36
074	AR163PB	AR1630924A	11/07/2009	05:40:16	1C6310B	2.81	6.36
075	PIBLKFJ	IBLKX0924B	11/07/2009	05:52:22	1C6310B	2.81	6.36
076	PPR13	5820184	11/07/2009	06:04:19	1C6310B	2.81	6.36
077	PPR05	5820185	11/07/2009	06:16:17	1C6310B	2.81	6.36
078	PPR15	5820186	11/07/2009	06:28:17	1C6310B	2.81	6.36
079	PPRF5	5820187	11/07/2009	06:40:20	1C6310B	2.81	6.36
080	PPR17	5820188	11/07/2009	06:52:29	1C6310B	2.81	6.36
081	PPR16	5820189	11/07/2009	07:04:26	1C6310B	2.81	6.36
082	PBLK93	BLANKA	11/07/2009	07:16:23	1C6310B	2.81	6.36
083	LCSE8	LCSA	11/07/2009	07:28:25	1C6310B	2.81	6.36
084	10714	5823539	11/07/2009	07:40:27	1C6310B	2.81	6.36
085	GTA44	5823540	11/07/2009	07:52:30	1C6310B	2.81	6.36
086	AR162OO	AR1620924A	11/07/2009	08:04:25	1C6310B	2.81	6.36
087	PIBLKFJ	IBLKX0924B	11/07/2009	08:16:23	1C6310B	2.81	6.36
088	OLDC1	5825593	11/07/2009	08:28:24	1C6310B	2.81	6.36
089	OLDC1MS	5825593	11/07/2009	08:40:25	1C6310B	2.81	6.36
090	OLDC1MSD	5825593	11/07/2009	08:52:27	1C6310B	2.81	6.36
091	OLDC2	5825594	11/07/2009	09:04:23	1C6310B	2.81	6.36
092	OLDC3	5825595	11/07/2009	09:16:19	1C6310B	2.81	6.36
093	AR164FS	AR1640924A	11/07/2009	09:28:21	1C6310B	2.81	6.36
094	PIBLKFJ	IBLKX0924B	11/07/2009	09:40:22	1C6310B	2.81	6.36

	ICAL Dates		ICAL RT	QC Limits
1C6310B	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	2.81	(2.79 - 2.83 Minutes)
		DCB = Decachlorobiphenyl	6.36	(6.34 - 6.38 Minutes)

ASX22 0283

FORM VIII PEST

8D
ANALYTICAL SEQUENCE

Sequence: 2C6310 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLP ID: 0.32

Instrument: S9147A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
001	FE103	5822189	11/07/2009	10:00:40	1C6310	3.31	7.16
002	FE202	5822191	11/07/2009	10:12:39	1C6310	3.31	7.16
003	PBLK8D	BLANKA	11/07/2009	10:24:40	1C6310	3.31	7.16
004	LCSDS	LCSA	11/07/2009	10:36:39	1C6310	3.31	7.16
005	WOO54	5814908	11/07/2009	10:48:35	1C6310	3.31	7.16
006	STEA2	5822069	11/07/2009	11:00:36	1C6310	3.31	7.16
007	STEA3	5822070	11/07/2009	11:12:38	1C6310	3.31	7.16
008	STEA3	5822070	11/07/2009	11:24:36	1C6310	3.31	7.16
009	STEA3MS	5822071	11/07/2009	11:36:37	1C6310	3.31	7.16
010	STEA3MS	5822071	11/07/2009	11:48:35	1C6310	3.31	
011	AR163PM	AR1630924A	11/07/2009	12:00:31	1C6310	3.31	7.16
012	PIBLKFU	IBLKX0924B	11/07/2009	12:12:30	1C6310	3.31	7.16
013	STEA3MSD	5822072	11/07/2009	12:24:31	1C6310	3.31	7.16
014	STEA3MSD	5822072	11/07/2009	12:36:29	1C6310		
015	DPS29	5822073	11/07/2009	12:48:27	1C6310	3.31	7.16
016	DPS29	5822073	11/07/2009	13:00:24	1C6310	3.31	7.16
017	ELEC3	5822074	11/07/2009	13:12:24	1C6310	3.31	7.16
018	ELEC3	5822074	11/07/2009	13:24:24	1C6310	3.31	
019	WHEE1	5822075	11/07/2009	13:36:25	1C6310	3.31	7.16
020	WHEE2	5822076	11/07/2009	13:48:22	1C6310	3.31	7.18
021	WHEE3	5822077	11/07/2009	14:00:19	1C6310	3.32	
022	WHEE3	5822077	11/07/2009	14:12:23	1C6310	3.32	
023	AR162OS	AR1620924A	11/07/2009	14:24:24	1C6310	3.31	7.16
024	PIBLKFV	IBLKX0924B	11/07/2009	14:36:25	1C6310	3.31	7.16
025	PBLK8C	BLANKA	11/07/2009	14:48:22	1C6310	3.31	7.16
026	LCSDR	LCSA	11/07/2009	15:00:22	1C6310	3.31	7.16
027	13181	5821274	11/07/2009	15:12:27	1C6310	3.31	7.16
028	13181MS	5821275	11/07/2009	15:24:27	1C6310	3.31	7.16
029	13181MSD	5821276	11/07/2009	15:36:27	1C6310	3.31	7.16
030	13182	5821278	11/07/2009	15:48:22	1C6310	3.31	7.16
031	13183	5821279	11/07/2009	16:00:23	1C6310	3.31	7.16

	ICAL Dates		ICAL RT	QC Limits
1C6310	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	3.31	(3.29 - 3.33 Minutes)
		DCB = Decachlorobiphenyl	7.16	(7.14 - 7.18 Minutes)

ASX22 0284

8D
ANALYTICAL SEQUENCE

Sequence: 2C6310 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLP ID: 0.32

Instrument: S9147A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
032	13184	5821280	11/07/2009	16:12:21	1C6310	3.31	7.16
033	13185	5821281	11/07/2009	16:24:22	1C6310	3.31	7.16
034	13186	5821282	11/07/2009	16:36:23	1C6310	3.31	7.16
035	AR163PO	AR1630924A	11/07/2009	16:48:20	1C6310	3.31	7.16
036	PIBLKFW	IBLKX0924B	11/07/2009	17:00:20	1C6310	3.31	7.16
037	13187	5821283	11/07/2009	17:12:20	1C6310	3.31	7.16
038	13188	5821284	11/07/2009	17:24:20	1C6310	3.31	7.16
039	13189	5821285	11/07/2009	17:36:21	1C6310	3.31	7.16
040	31810	5821286	11/07/2009	17:48:15	1C6310	3.31	7.16
041	31811	5821287	11/07/2009	18:00:17	1C6310	3.31	7.16
042	31812	5821288	11/07/2009	18:12:18	1C6310	3.31	7.16
043	31813	5821289	11/07/2009	18:24:26	1C6310	3.31	7.16
044	31814	5821290	11/07/2009	18:36:24	1C6310	3.31	7.16
045	31815	5821291	11/07/2009	18:48:21	1C6310	3.31	7.16
046	31816	5821292	11/07/2009	19:00:23	1C6310	3.31	7.16
047	AR164FU	AR1640924A	11/07/2009	19:12:20	1C6310	3.31	7.16
048	PIBLKFX	IBLKX0924B	11/07/2009	19:24:22	1C6310	3.31	7.16
049	31817	5821293	11/07/2009	19:36:23	1C6310	3.31	7.16
050	31818	5821294	11/07/2009	19:48:20	1C6310	3.31	7.16
051	31819	5821295	11/07/2009	20:00:22	1C6310	3.31	7.16
052	31820	5821296	11/07/2009	20:12:23	1C6310	3.31	7.16
053	PBLKB2	BLANKA	11/07/2009	20:24:23	1C6310	3.31	7.16
054	LCSG7	LCSA	11/07/2009	20:36:23	1C6310	3.31	7.16
055	BRIRB	5826082	11/07/2009	20:48:20	1C6310	3.31	7.16
056	BRIRBMS	5826082	11/07/2009	21:00:21	1C6310	3.31	7.16
057	BRIRBMSD	5826082	11/07/2009	21:12:23	1C6310	3.31	7.16
058	AR163PQ	AR1630924A	11/07/2009	21:24:23	1C6310	3.31	7.16
059	PIBLKFY	IBLKX0924B	11/07/2009	21:36:24	1C6310	3.31	7.16

	ICAL Dates		ICAL RT	QC Limits
1C6310	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	3.31	(3.29 - 3.33 Minutes)
		DCB = Decachlorobiphenyl	7.16	(7.14 - 7.18 Minutes)

ASX22 0285

8D
ANALYTICAL SEQUENCE

Sequence: 2C6310B | Lab Name: Lancaster laboratories | Contract:

Lab Code: | Case No.: | SAS No: | SDG No.:

GC Column: RTX CLPII | ID: 0.32

Instrument: S9147B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
001	FE103	5822189	11/07/2009	10:00:40	1C6310B	2.81	6.36
002	FE202	5822191	11/07/2009	10:12:39	1C6310B	2.81	6.36
003	PBLK8D	BLANKA	11/07/2009	10:24:40	1C6310B	2.81	6.36
004	LCSDS	LCSA	11/07/2009	10:36:39	1C6310B	2.81	6.36
005	WOO54	5814908	11/07/2009	10:48:35	1C6310B	2.81	6.36
006	STEA2	5822069	11/07/2009	11:00:36	1C6310B	2.81	6.35
007	STEA3	5822070	11/07/2009	11:12:38	1C6310B	2.81	6.35
008	STEA3	5822070	11/07/2009	11:24:36	1C6310B	2.81	6.36
009	STEA3MS	5822071	11/07/2009	11:36:37	1C6310B	2.81	6.36
010	STEA3MS	5822071	11/07/2009	11:48:35	1C6310B	2.81	6.35
011	AR163PM	AR1630924A	11/07/2009	12:00:31	1C6310B	2.81	6.36
012	PIBLKFU	IBLKX0924B	11/07/2009	12:12:30	1C6310B	2.81	6.36
013	STEA3MSD	5822072	11/07/2009	12:24:31	1C6310B	2.81	6.35
014	STEA3MSD	5822072	11/07/2009	12:36:29	1C6310B		6.36
015	DPS29	5822073	11/07/2009	12:48:27	1C6310B	2.81	6.35
016	DPS29	5822073	11/07/2009	13:00:24	1C6310B	2.81	6.35
017	ELEC3	5822074	11/07/2009	13:12:24	1C6310B	2.81	6.35
018	ELEC3	5822074	11/07/2009	13:24:24	1C6310B	2.81	6.36
019	WHEE1	5822075	11/07/2009	13:36:25	1C6310B	2.81	6.36
020	WHEE2	5822076	11/07/2009	13:48:22	1C6310B	2.81	
021	WHEE3	5822077	11/07/2009	14:00:19	1C6310B		6.36
022	WHEE3	5822077	11/07/2009	14:12:23	1C6310B	2.81	6.36
023	AR162OS	AR1620924A	11/07/2009	14:24:24	1C6310B	2.81	6.36
024	PIBLKFV	IBLKX0924B	11/07/2009	14:36:25	1C6310B	2.81	6.36
025	PBLK8C	BLANKA	11/07/2009	14:48:22	1C6310B	2.81	6.36
026	LCSDR	LCSA	11/07/2009	15:00:22	1C6310B	2.81	6.36
027	13181	5821274	11/07/2009	15:12:27	1C6310B	2.81	6.36
028	13181MS	5821275	11/07/2009	15:24:27	1C6310B	2.81	6.36
029	13181MSD	5821276	11/07/2009	15:36:27	1C6310B	2.81	6.36
030	13182	5821278	11/07/2009	15:48:22	1C6310B	2.81	6.36
031	13183	5821279	11/07/2009	16:00:23	1C6310B	2.81	6.36
032	13184	5821280	11/07/2009	16:12:21	1C6310B	2.81	6.36
033	13185	5821281	11/07/2009	16:24:22	1C6310B	2.81	6.36

	ICAL Dates		ICAL RT	QC Limits
1C6310B	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	2.81	(2.79 - 2.83 Minutes)
		DCB = Decachlorobiphenyl	6.36	(6.34 - 6.38 Minutes)

ASX22 0286

8D
ANALYTICAL SEQUENCE

Sequence: 2C6310B Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLPII ID: 0.32

Instrument: S9147B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
034	13186	5821282	11/07/2009	16:36:23	1C6310B	2.81	6.36
035	AR163PO	AR1630924A	11/07/2009	16:48:20	1C6310B	2.81	6.36
036	PIBLKFW	IBLKX0924B	11/07/2009	17:00:20	1C6310B	2.81	6.36
037	13187	5821283	11/07/2009	17:12:20	1C6310B	2.81	6.36
038	13188	5821284	11/07/2009	17:24:20	1C6310B	2.81	6.36
039	13189	5821285	11/07/2009	17:36:21	1C6310B	2.81	6.36
040	31810	5821286	11/07/2009	17:48:15	1C6310B	2.81	6.36
041	31811	5821287	11/07/2009	18:00:17	1C6310B	2.81	6.36
042	31812	5821288	11/07/2009	18:12:18	1C6310B	2.81	6.36
043	31813	5821289	11/07/2009	18:24:26	1C6310B	2.81	6.36
044	31814	5821290	11/07/2009	18:36:24	1C6310B	2.81	6.36
045	31815	5821291	11/07/2009	18:48:21	1C6310B	2.81	6.36
046	31816	5821292	11/07/2009	19:00:23	1C6310B	2.81	6.36
047	AR164FU	AR1640924A	11/07/2009	19:12:20	1C6310B	2.81	6.36
048	PIBLKFX	IBLKX0924B	11/07/2009	19:24:22	1C6310B	2.81	6.36
049	31817	5821293	11/07/2009	19:36:23	1C6310B	2.81	6.36
050	31818	5821294	11/07/2009	19:48:20	1C6310B	2.81	6.36
051	31819	5821295	11/07/2009	20:00:22	1C6310B	2.81	6.36
052	31820	5821296	11/07/2009	20:12:23	1C6310B	2.81	6.36
053	PBLKB2	BLANKA	11/07/2009	20:24:23	1C6310B	2.81	6.36
054	LCSG7	LCSA	11/07/2009	20:36:23	1C6310B	2.81	6.36
055	BRIRB	5826082	11/07/2009	20:48:20	1C6310B	2.81	6.36
056	BRIRBMS	5826082	11/07/2009	21:00:21	1C6310B	2.81	6.36
057	BRIRBMSD	5826082	11/07/2009	21:12:23	1C6310B	2.81	6.36
058	AR163PQ	AR1630924A	11/07/2009	21:24:23	1C6310B	2.81	6.36
059	PIBLKFY	IBLKX0924B	11/07/2009	21:36:24	1C6310B	2.81	6.36

	ICAL Dates		ICAL RT	QC Limits
1C6310B	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	2.81	(2.79 - 2.83 Minutes)
		DCB = Decachlorobiphenyl	6.36	(6.34 - 6.38 Minutes)

ASXZZ 0287

8D
ANALYTICAL SEQUENCE

Sequence: 3C6310 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLP ID: 0.32

Instrument: S9147A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
001		CONDITIONER	11/09/2009	10:19:08	1C6310	3.31	7.16
002		CONDITIONER	11/09/2009	10:31:05	1C6310	3.31	7.16
003		CONDITIONER	11/09/2009	10:43:05	1C6310	3.31	7.16
004	AR164FY	AR1640924A	11/09/2009	10:55:06	1C6310	3.31	7.16
005	PIBLKGD	IBLKX0924B	11/09/2009	11:07:05	1C6310	3.31	7.16
006	PBLK8C	BLANKA	11/09/2009	11:19:06	1C6310	3.31	7.16
007	LCSDR	LCSA	11/09/2009	11:30:59	1C6310	3.31	7.16
008	13181	5821274	11/09/2009	11:43:00	1C6310	3.31	7.16
009	13181MS	5821275	11/09/2009	11:54:59	1C6310	3.31	7.16
010	13181MSD	5821276	11/09/2009	12:06:59	1C6310	3.31	7.16
011	13182	5821278	11/09/2009	12:18:56	1C6310	3.31	7.16
012	13183	5821279	11/09/2009	12:30:51	1C6310	3.31	7.16
013	13184	5821280	11/09/2009	12:42:51	1C6310	3.31	7.16
014	13185	5821281	11/09/2009	12:54:51	1C6310	3.31	7.16
015	13186	5821282	11/09/2009	13:06:50	1C6310	3.31	7.16
016	13187	5821283	11/09/2009	13:18:48	1C6310	3.31	7.16
017	AR163PZ	AR1630924A	11/09/2009	13:30:49	1C6310	3.31	7.16
018	PIBLKGE	IBLKX0924B	11/09/2009	13:42:51	1C6310	3.31	7.16
019	PBLKBP	BLANKA	11/09/2009	14:13:21	1C6310	3.32	7.17
020	LCSGV	LCSA	11/09/2009	14:25:19	1C6310	3.31	7.16
021	CS-1B	5827102	11/09/2009	14:37:23	1C6310	3.31	7.16
022	CS-1BMS	5827102	11/09/2009	14:49:20	1C6310	3.31	7.16
023	CS-1BMSD	5827102	11/09/2009	15:01:16	1C6310	3.31	7.16
024	DS-1B	5827103	11/09/2009	15:13:17	1C6310	3.31	7.16
025	CS-2B	5827104	11/09/2009	15:25:15	1C6310	3.31	7.16
026	DS-2B	5827105	11/09/2009	15:37:22	1C6310	3.31	7.16
027	CS-3B	5827106	11/09/2009	15:49:18	1C6310	3.31	7.16
028	DS-3B	5827107	11/09/2009	16:01:18	1C6310	3.31	7.16
029	CS-4B	5827108	11/09/2009	16:13:17	1C6310	3.31	7.16
030	DS-4B	5827109	11/09/2009	16:25:16	1C6310	3.31	7.16
031	CS-5B	5827110	11/09/2009	16:37:15	1C6310	3.31	7.16

	ICAL Dates		ICAL RT	QC Limits
1C6310	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	3.31	(3.29 - 3.33 Minutes)
		DCB = Decachlorobiphenyl	7.16	(7.14 - 7.18 Minutes)

ASX22 0288

8D
ANALYTICAL SEQUENCE

Sequence: 3C6310 Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLP ID: 0.32

Instrument: S9147A

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
032	DS-5B	5827111	11/09/2009	16:49:12	1C6310	3.31	7.16
033	CS-6B	5827112	11/09/2009	17:01:07	1C6310	3.31	7.16
034	DS-6B	5827113	11/09/2009	17:13:08	1C6310	3.31	7.16
035	CS-8B	5827114	11/09/2009	17:25:04	1C6310	3.31	7.16
036	DS-8B	5827115	11/09/2009	17:37:05	1C6310	3.31	7.16
037	CS-9B	5827116	11/09/2009	17:49:06	1C6310	3.31	7.16
038	13182	5821278	11/09/2009	18:01:02	1C6310	3.31	7.16
039	PIBLKGG	IBLKX0924B	11/09/2009	18:13:00	1C6310	3.31	7.16
040	CS10B	5827117	11/09/2009	18:25:00	1C6310	3.31	7.16
041	CS10B	5827117	11/09/2009	18:36:59	1C6310	3.31	
042	DS-9B	5827118	11/09/2009	18:48:55	1C6310	3.31	7.16
043	DS-9B	5827118	11/09/2009	19:00:51	1C6310	3.31	
044	13182	5821278	11/09/2009	19:12:50	1C6310	3.31	7.16
045	AR163QD	AR1630924A	11/09/2009	19:24:50	1C6310	3.31	7.16
046	PIBLKGH	IBLKX0924B	11/09/2009	19:36:50	1C6310	3.31	7.16
047	STEA3MSD	5822072	11/09/2009	19:48:45	1C6310	3.31	7.15
048	DPS29	5822073	11/09/2009	20:00:42	1C6310		
049	ELEC3	5822074	11/09/2009	20:12:41	1C6310	3.31	7.16
050	WHEE1	5822075	11/09/2009	20:24:40	1C6310	3.31	7.16
051	WHEE2	5822076	11/09/2009	20:36:40	1C6310	3.32	
052	WHEE3	5822077	11/09/2009	20:48:36	1C6310	3.32	
053	WHEE3	5822077	11/09/2009	21:00:40	1C6310	3.31	
054	AA	CONDITIONER	11/09/2009	21:12:41	1C6310	3.31	7.16
055		CONDITIONER	11/09/2009	21:24:41	1C6310	3.31	7.16
056		CONDITIONER	11/09/2009	21:36:40	1C6310	3.31	7.16
057	AR162OT	AR1620924A	11/09/2009	21:48:35	1C6310	3.31	7.16
058	PIBLKGI	IBLKX0924B	11/09/2009	22:00:38	1C6310	3.31	7.16

	ICAL Dates		ICAL RT	QC Limits
1C6310	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	3.31	(3.29 - 3.33 Minutes)
		DCB = Decachlorobiphenyl	7.16	(7.14 - 7.18 Minutes)

ASX22 0289

8D
ANALYTICAL SEQUENCE

Sequence: 3C6310B | Lab Name: Lancaster laboratories | Contract:

Lab Code: | Case No.: | SAS No: | SDG No.:

GC Column: RTX CLPII | ID: 0.32

Instrument: S9147B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
001		CONDITIONER	11/09/2009	10:19:08	1C6310B	2.81	6.36
002		CONDITIONER	11/09/2009	10:31:05	1C6310B	2.81	6.36
003		CONDITIONER	11/09/2009	10:43:05	1C6310B	2.81	6.36
004	AR164FY	AR1640924A	11/09/2009	10:55:06	1C6310B	2.81	6.36
005	PIBLKGD	IBLKX0924B	11/09/2009	11:07:05	1C6310B	2.81	6.36
006	PBLK8C	BLANKA	11/09/2009	11:19:06	1C6310B	2.81	6.36
007	LCSDR	LCSA	11/09/2009	11:30:59	1C6310B	2.81	6.36
008	13181	5821274	11/09/2009	11:43:00	1C6310B	2.81	6.36
009	13181MS	5821275	11/09/2009	11:54:59	1C6310B	2.81	6.36
010	13181MSD	5821276	11/09/2009	12:06:59	1C6310B	2.81	6.36
011	13182	5821278	11/09/2009	12:18:56	1C6310B	2.81	6.36
012	13183	5821279	11/09/2009	12:30:51	1C6310B	2.81	6.36
013	13184	5821280	11/09/2009	12:42:51	1C6310B	2.81	6.36
014	13185	5821281	11/09/2009	12:54:51	1C6310B	2.81	6.36
015	13186	5821282	11/09/2009	13:06:50	1C6310B	2.81	6.36
016	13187	5821283	11/09/2009	13:18:48	1C6310B	2.81	6.36
017	AR163PZ	AR1630924A	11/09/2009	13:30:49	1C6310B	2.81	6.36
018	PIBLKGE	IBLKX0924B	11/09/2009	13:42:51	1C6310B	2.81	6.36
019	PBLKBP	BLANKA	11/09/2009	14:13:21	1C6310B	2.81	6.36
020	LCSGV	LCSA	11/09/2009	14:25:19	1C6310B	2.81	6.36
021	CS-1B	5827102	11/09/2009	14:37:23	1C6310B	2.81	6.36
022	CS-1BMS	5827102	11/09/2009	14:49:20	1C6310B	2.81	6.36
023	CS-1BMSD	5827102	11/09/2009	15:01:16	1C6310B	2.81	6.36
024	DS-1B	5827103	11/09/2009	15:13:17	1C6310B	2.81	6.36
025	CS-2B	5827104	11/09/2009	15:25:15	1C6310B	2.81	6.36
026	DS-2B	5827105	11/09/2009	15:37:22	1C6310B	2.81	6.36
027	CS-3B	5827106	11/09/2009	15:49:18	1C6310B	2.81	6.36
028	DS-3B	5827107	11/09/2009	16:01:18	1C6310B	2.81	6.36
029	CS-4B	5827108	11/09/2009	16:13:17	1C6310B	2.81	6.36
030	DS-4B	5827109	11/09/2009	16:25:16	1C6310B	2.81	6.36
031	CS-5B	5827110	11/09/2009	16:37:15	1C6310B	2.81	6.36
032	DS-5B	5827111	11/09/2009	16:49:12	1C6310B	2.81	6.36
033	CS-6B	5827112	11/09/2009	17:01:07	1C6310B	2.81	6.36

	ICAL Dates		ICAL RT	QC Limits
1C6310B	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	2.81	(2.79 - 2.83 Minutes)
		DCB = Decachlorobiphenyl	6.36	(6.34 - 6.38 Minutes)

ASX22 8290

FORM VIII PEST

8D
ANALYTICAL SEQUENCE

Sequence: 3C6310B Lab Name: Lancaster laboratories Contract:

Lab Code: Case No.: SAS No: SDG No.:

GC Column: RTX CLPII ID: 0.32

Instrument: S9147B

THIS ANALYTICAL SEQUENCE OF BLANKS, SAMPLES AND STANDARDS IS GIVEN BELOW:

	Sample Code No.	Lab Sample ID	Date Analyzed	Time Analyzed	Calibration File	TCX	DCB
034	DS-6B	5827113	11/09/2009	17:13:08	1C6310B	2.81	6.36
035	CS-8B	5827114	11/09/2009	17:25:04	1C6310B	2.81	6.36
036	DS-8B	5827115	11/09/2009	17:37:05	1C6310B	2.81	6.36
037	CS-9B	5827116	11/09/2009	17:49:06	1C6310B	2.81	6.36
038	13182	5821278	11/09/2009	18:01:02	1C6310B	2.81	6.36
039	PIBLKGG	IBLKX0924B	11/09/2009	18:13:00	1C6310B	2.81	6.36
040	CS10B	5827117	11/09/2009	18:25:00	1C6310B	2.81	6.36
041	CS10B	5827117	11/09/2009	18:36:59	1C6310B	2.81	6.36
042	DS-9B	5827118	11/09/2009	18:48:55	1C6310B	2.81	6.36
043	DS-9B	5827118	11/09/2009	19:00:51	1C6310B	2.81	6.36
044	13182	5821278	11/09/2009	19:12:50	1C6310B	2.81	6.36
045	AR163QD	AR1630924A	11/09/2009	19:24:50	1C6310B	2.81	6.36
046	PIBLKGH	IBLKX0924B	11/09/2009	19:36:50	1C6310B	2.81	6.36
047	STEA3MSD	5822072	11/09/2009	19:48:45	1C6310B	2.81	6.36
048	DPS29	5822073	11/09/2009	20:00:42	1C6310B	2.81	6.36
049	ELEC3	5822074	11/09/2009	20:12:41	1C6310B	2.81	6.35
050	WHEE1	5822075	11/09/2009	20:24:40	1C6310B	2.81	6.36
051	WHEE2	5822076	11/09/2009	20:36:40	1C6310B	2.81	
052	WHEE3	5822077	11/09/2009	20:48:36	1C6310B		6.36
053	WHEE3	5822077	11/09/2009	21:00:40	1C6310B	2.81	6.36
054	AA	CONDITIONER	11/09/2009	21:12:41	1C6310B	2.81	6.36
055		CONDITIONER	11/09/2009	21:24:41	1C6310B	2.81	6.36
056		CONDITIONER	11/09/2009	21:36:40	1C6310B	2.81	6.36
057	AR162OT	AR1620924A	11/09/2009	21:48:35	1C6310B	2.81	6.36
058	PIBLKGI	IBLKX0924B	11/09/2009	22:00:38	1C6310B	2.81	6.36

	ICAL Dates		ICAL RT	QC Limits
1C6310B	11/06/2009 - 11/06/2009	TCX = Tetrachloro-m-xylene	2.81	(2.79 - 2.83 Minutes)
		DCB = Decachlorobiphenyl	6.36	(6.34 - 6.38 Minutes)

HSX22 0291

FORM VIII PEST

Lancaster Laboratories

CHROM PERFECT SEQUENCE FILE

Sequence File: \\cp7\C-Drive\CPWIN\DATA1\1p7307.seq

Chromatography Directory: \\cp7\C-Drive\cpwin\DATA1

Method Directory: \\cp7\C-Drive\cpwin\DATA1

Number of Entries: 85

Samplename	Code	ID	FileName	Method	Samp Wt	DF	Int Std	Batch Number	Analysis
1 CONDITIONER	MISC	AA	1p7307.01R	PESTD.MET	1	1	1	0930699999	
2 CONDITIONER	MISC	AA	1p7307.02R	PESTD.MET	1	1	1	0930699999	
3 CONDITIONER	MISC	AA	1p7307.03R	PESTD.MET	1	1	1	0930699999	
4 CONDITIONER	MISC	AA	1p7307.04R	PESTD.MET	1	1	1	0930699999	
5 CONDITIONER	MISC	AA	1p7307.05R	PESTD.MET	1	1	1	0930699999	
6 IBLKX0924B	PIBLK	AA	1p7307.06R	PESTD.MET	1000	10	1	0930699999	
7 AR1610924A	ICAL	AA	1p7307.07R	PESTD.MET	1	1	1	0930699999	
8 AR1620924A	ICAL	AA	1p7307.08R	PESTD.MET	1	1	1	0930699999	
9 AR1630924A	ICAL	AA	1p7307.09R	PESTD.MET	1	1	1	0930699999	
10 AR1640924A	ICAL	AA	1p7307.10R	PESTD.MET	1	1	1	0930699999	
11 AR1650924A	ICAL	AA	1p7307.11R	PESTD.MET	1	1	1	0930699999	
12 AR4810924B	ICAL	AA	1p7307.12R	PESTD.MET	1	1	1	0930699999	
13 AR4820924B	ICAL	AA	1p7307.13R	PESTD.MET	1	1	1	0930699999	
14 AR4830924B	ICAL	AA	1p7307.14R	PESTD.MET	1	1	1	0930699999	
15 AR4840924B	ICAL	AA	1p7307.15R	PESTD.MET	1	1	1	0930699999	
16 AR4850924B	ICAL	AA	1p7307.16R	PESTD.MET	1	1	1	0930699999	
17 AR5410924B	ICAL	AA	1p7307.17R	PESTD.MET	1	1	1	0930699999	
18 AR5420924B	ICAL	AA	1p7307.18R	PESTD.MET	1	1	1	0930699999	
19 AR5430924B	ICAL	AA	1p7307.19R	PESTD.MET	1	1	1	0930699999	
20 AR5440924B	ICAL	AA	1p7307.20R	PESTD.MET	1	1	1	0930699999	
21 AR5450924B	ICAL	AA	1p7307.21R	PESTD.MET	1	1	1	0930699999	
22 AR6210924B	ICAL	AA	1p7307.22R	PESTD.MET	1	1	1	0930699999	
23 AR6220924B	ICAL	AA	1p7307.23R	PESTD.MET	1	1	1	0930699999	
24 AR6230924B	ICAL	AA	1p7307.24R	PESTD.MET	1	1	1	0930699999	
25 AR6240924B	ICAL	AA	1p7307.25R	PESTD.MET	1	1	1	0930699999	
26 AR6250924B	ICAL	AA	1p7307.26R	PESTD.MET	1	1	1	0930699999	
27 AR6810924C	ICAL	AA	1p7307.27R	PESTD.MET	1	1	1	0930699999	
28 AR6820924C	ICAL	AA	1p7307.28R	PESTD.MET	1	1	1	0930699999	
29 AR6830924C	ICAL	AA	1p7307.29R	PESTD.MET	1	1	1	0930699999	
30 AR6840924C	ICAL	AA	1p7307.30R	PESTD.MET	1	1	1	0930699999	
31 AR6850924C	ICAL	AA	1p7307.31R	PESTD.MET	1	1	1	0930699999	
32 AR2130924B	ICAL	AA	1p7307.32R	PESTD.MET	1	1	1	0930699999	
33 AR32X0924B	ICAL	AA	1p7307.33R	PESTD.MET	1	1	1	0930699999	
34 AR4230924B	ICAL	AA	1p7307.34R	PESTD.MET	1	1	1	0930699999	
35 AR16X0924A	ICAL	AA	1p7307.35R	PESTD.MET	1	1	1	0930699999	
36 MD16X0924A	ICAL	AA	1p7307.36R	PESTD.MET	1	1	1	0930699999	
37 MD62X0924B	ICAL	AA	1p7307.37R	PESTD.MET	1	1	1	0930699999	
38 MD68X0924C	ICAL	AA	1p7307.38R	PESTD.MET	1	1	1	0930699999	
39 IC16X0924A	CCAL	AA	1p7307.39R	PESTD.MET	1	1	1	0930699999	
40 IC48X0924B	CCAL	AA	1p7307.40R	PESTD.MET	1	1	1	0930699999	
41 IC54X0924B	CCAL	AA	1p7307.41R	PESTD.MET	1	1	1	0930699999	
42 BLANKA 11/2/09AF	BLK	AB	1p7307.42R	PESTD.MET	1000	10	1	093040007A	02034
43 LCSA 11/2/09AF	LCS	AB	1p7307.43R	PESTD.MET	1000	10	1	093040007A	02034
44 LCSDA 11/2/09AF	LCSD	AB	1p7307.44R	PESTD.MET	1000	10	1	093040007A	02034
45 5812843RAF	T	AB	1p7307.45R	PESTD.MET	1056	10	1	093040007A	02034
46 5815988RAF	T	AB	1p7307.46R	PESTD.MET	923	10	1	093040007A	02034
47 AR1630924A	CCAL	OH	1p7307.47R	PESTD.MET	1	1	1	0930699999	
48 IBLKX0924B	PIBLK	EY	1p7307.48R	PESTD.MET	1000	10	1	0930699999	
49 5813449ACF DF100	T	AD	1p7307.49R	PESTD.MET	30	1000	1	093000011A	01216
50 5813450ACF DF10000	T	AE	1p7307.50R	PESTD.MET	30	100000	1	093000011A	01216

A9142 8292

Lancaster Laboratories

CHROM PERFECT SEQUENCE FILE

Sequence File: \\cp7\C-Drive\CPWIN\DATA1\1p7307.seq

Chromatography Directory: \\cp7\C-Drive\cpwin\DATA1

Method Directory: \\cp7\C-Drive\cpwin\DATA1

Number of Entries: 85

Samplename	Code	ID	FileName	Method	Samp Wt	DF	Int Std	Batch Number	Analysis
51 5813451ACF DF50	T	AE	1p7307.51R	PESTD.MET	30	500	1	093000011A	01216
52 5813453ACF DF100	T	AD	1p7307.52R	PESTD.MET	30	1000	1	093000011A	01216
53 5813454MSACF DF100	MS	AD	1p7307.53R	PESTD.MET	30	1000	1	093000011A	01216
54 5813455MSDACF DF100	MSD	AD	1p7307.54R	PESTD.MET	30	1000	1	093000011A	01216
55 5813458ACF DF100	T	AD	1p7307.55R	PESTD.MET	30	1000	1	093000011A	01216
56 5813462ACF DF100	T	AD	1p7307.56R	PESTD.MET	30	1000	1	093000011A	01216
57 5813470ACF DF10	T	AD	1p7307.57R	PESTD.MET	30	100	1	093000011A	01216
58 5813158ACF DF10	T	AD	1p7307.58R	PESTD.MET	30	100	1	092990030A	01216
59 5813164ACF DF20	T	AE	1p7307.59R	PESTD.MET	30	200	1	092990030A	01216
60 AR1620924A	CCAL	NS	1p7307.60R	PESTD.MET	1	1	1	0930699999	
61 IBLKX0924B	PIBLK	EY	1p7307.61R	PESTD.MET	1000	10	1	0930699999	
62 BLANKA 10/30/09CAF	BLK	AC	1p7307.62R	PESTD.MET	1000	10	1	093030004A	02034
63 LCSA 10/30/09CAF	LCS	AC	1p7307.63R	PESTD.MET	1000	10	1	093030004A	02034
64 LCSDA 10/30/09CAF	LCSD	AC	1p7307.64R	PESTD.MET	1000	10	1	093030004A	02034
65 5820183CAF	T	AC	1p7307.65R	PESTD.MET	1054	10	1	093030004A	02034
66 5820184CAF	T	AC	1p7307.66R	PESTD.MET	1058	10	1	093030004A	02034
67 5820185CAF	T	AC	1p7307.67R	PESTD.MET	1050	10	1	093030004A	02034
68 5820186CAF	T	AC	1p7307.68R	PESTD.MET	1055	10	1	093030004A	02034
69 5820187CAF	T	AC	1p7307.69R	PESTD.MET	1055	10	1	093030004A	02034
70 5820188CAF	T	AC	1p7307.70R	PESTD.MET	1055	10	1	093030004A	02034
71 5820189CAF	T	AC	1p7307.71R	PESTD.MET	1053	10	1	093030004A	02034
72 AR1630924A	CCAL	OK	1p7307.72R	PESTD.MET	1	1	1	0930699999	
73 IBLKX0924B	PIBLK	EY	1p7307.73R	PESTD.MET	1000	10	1	0930699999	
74 BLANKA 11/2/09ACF	BLK	AB	1p7307.74R	PESTD.MET	30	10	1	093040011A	01216
75 LCSA 11/2/09ACF	LCS	AB	1p7307.75R	PESTD.MET	30	10	1	093040011A	01216
76 5815455RACF	T	AB	1p7307.76R	PESTD.MET	30	10	1	093040011A	01216
77 5816565RACF	T	AB	1p7307.77R	PESTD.MET	30	10	1	093040011A	01216
78 5816565RMSACF	MS	AB	1p7307.78R	PESTD.MET	30	10	1	093040011A	01216
79 5816565RMSDACF	MSD	AB	1p7307.79R	PESTD.MET	30	10	1	093040011A	01216
80 5819873ACF	T	AB	1p7307.80R	PESTD.MET	30	10	1	093040011A	01216
81 5819874ACF	T	AB	1p7307.81R	PESTD.MET	30	10	1	093040011A	01216
82 5819875ACF	T	AB	1p7307.82R	PESTD.MET	30	10	1	093040011A	01216
83 5819876ACF	T	AB	1p7307.83R	PESTD.MET	30	10	1	093040011A	01216
84 AR1640924A	CCAL	FF	1p7307.84R	PESTD.MET	1	1	1	0930699999	
85 IBLKX0924B	PIBLK	EY	1p7307.85R	PESTD.MET	1000	10	1	0930699999	

ASX22 0293

Set-up by: [illegible] Date: 11-4-09

11/4/2009

Lancaster Laboratories

CHROM PERFECT SEQUENCE FILE

Sequence File: \\cp7\C-Drive\cpwin\DATA1\3p7307.seq

Chromatography Directory: \\cp7\C-Drive\cpwin\DATA1

Method Directory: \\cp7\C-Drive\cpwin\DATA1

Number of Entries: 92

Samplename	Code	ID	FileName	Method	Samp Wt	DF	Int Std	Batch Number	Analysis
1 CONDITIONER	MISC	AA	3p7307.01R	PESTD.MET	1	1	1	0930899999	
2 CONDITIONER	MISC	AA	3p7307.02R	PESTD.MET	1	1	1	0930899999	
3 CONDITIONER	MISC	AA	3p7307.03R	PESTD.MET	1	1	1	0930899999	
4 CONDITIONER	MISC	AA	3p7307.04R	PESTD.MET	1	1	1	0930899999	
5 CONDITIONER	MISC	AA	3p7307.05R	PESTD.MET	1	1	1	0930899999	
6 AR1630924A	CCAL	OU	3p7307.06R	PESTD.MET	1	1	1	0930899999	
7 IBLKX0924B	PIBLK	FH	3p7307.07R	PESTD.MET	1000	10	1	0930899999	
8 MS0926824A	MISC	AA	3p7307.08R	PESTD.MET	1	1	1	0930899999	
9 BLANKA 11/4/09ACF	BLK	AB	3p7307.09R	PESTD.MET	30	10	1	093080003A	01216
10 LCSA 11/4/09ACF	LCS	AB	3p7307.10R	PESTD.MET	30	10	1	093080003A	01216
11 5817880ACF	T	AB	3p7307.11R	PESTD.MET	30	10	1	093080003A	01216
12 5817880MSACF	MS	AB	3p7307.12R	PESTD.MET	30	10	1	093080003A	01216
13 5817880MSDACF	MSD	AB	3p7307.13R	PESTD.MET	30	10	1	093080003A	01216
14 5822168ACF	T	AB	3p7307.14R	PESTD.MET	30	10	1	093080003A	01216
15 5822197ACF	T	AB	3p7307.15R	PESTD.MET	30	10	1	093080003A	01216
16 AR1640924A	CCAL	FM	3p7307.16R	PESTD.MET	1	1	1	0930899999	
17 IBLKX0924B	PIBLK	FH	3p7307.17R	PESTD.MET	1000	10	1	0930899999	
18 BLANKA 11/5/09AF	BLK	AB	3p7307.18R	PESTD.MET	1000	10	1	093080009A	00173
19 LCSA 11/5/09AF	LCS	AB	3p7307.19R	PESTD.MET	1000	10	1	093080009A	00173
20 LCSDA 11/5/09AF	LCSD	AB	3p7307.20R	PESTD.MET	1000	10	1	093080009A	00173
21 5813174RAF	T	AB	3p7307.21R	PESTD.MET	1016	10	1	093080009A	00173
22 5813471RAF	T	AB	3p7307.22R	PESTD.MET	993	10	1	093080009A	00173
23 5826632AF	T	AB	3p7307.23R	PESTD.MET	1013	10	1	093080009A	00173
24 5826633AF	T	AB	3p7307.24R	PESTD.MET	982	10	1	093080009A	00173
25 5826635AF	T	AB	3p7307.25R	PESTD.MET	992	10	1	093080009A	00173
26 5826636AF	T	AB	3p7307.26R	PESTD.MET	985	10	1	093080009A	00173
27 5826637AF	T	AB	3p7307.27R	PESTD.MET	973	10	1	093080009A	00173
28 AR1630924A	CCAL	OU	3p7307.28R	PESTD.MET	1	1	1	0930899999	
29 IBLKX0924B	PIBLK	FH	3p7307.29R	PESTD.MET	1000	10	1	0930899999	
30 BLANKA 11/2/09 RI ACF	BLK	AB	3p7307.30R	PESTD.MET	30	10	1	093040012A	01216
31 LCSA 11/2/09 RI ACF	LCS	AB	3p7307.31R	PESTD.MET	30	10	1	093040012A	01216
32 5820657 RI ACF	T	AB	3p7307.32R	PESTD.MET	30	10	1	093040012A	01216
33 5820658 RI ACF	T	AB	3p7307.33R	PESTD.MET	30	10	1	093040012A	01216
34 5820659 RI ACF	T	AB	3p7307.34R	PESTD.MET	30	10	1	093040012A	01216
35 5820660 RI ACF	T	AB	3p7307.35R	PESTD.MET	30	10	1	093040012A	01216
36 5820661 RI ACF	T	AB	3p7307.36R	PESTD.MET	30	10	1	093040012A	01216
37 5820662 RI ACF	T	AB	3p7307.37R	PESTD.MET	30	10	1	093040012A	01216
38 5820663 RI ACF	T	AB	3p7307.38R	PESTD.MET	30	10	1	093040012A	01216
39 5820664 RI ACF	T	AB	3p7307.39R	PESTD.MET	30	10	1	093040012A	01216
40 AR1640924A	CCAL	FN	3p7307.40R	PESTD.MET	1	1	1	0930899999	
41 IBLKX0924B	PIBLK	FH	3p7307.41R	PESTD.MET	1000	10	1	0930899999	
42 BLANKA 11/3/09 RI AF	BLK	AB	3p7307.42R	PESTD.MET	1000	10	1	093060018A	00173
43 LCSA 11/3/09 RI AF	LCS	AB	3p7307.43R	PESTD.MET	1000	10	1	093060018A	00173
44 5822062AF	T	AB	3p7307.44R	PESTD.MET	1060	10	1	093060018A	00173
45 5822063AF DF100	T	AC	3p7307.45R	PESTD.MET	805	1000	1	093060018A	00173
46 5822064 RI AF DF20	T	AC	3p7307.46R	PESTD.MET	893	200	1	093060018A	00173
47 5822065MS RI AF DF20	MS	AC	3p7307.47R	PESTD.MET	1020	200	1	093060018A	00173
48 5822066MSD RI AF DF20	MSD	AC	3p7307.48R	PESTD.MET	1041	200	1	093060018A	00173
49 5822067 RI AF DF20	T	AC	3p7307.49R	PESTD.MET	1018	200	1	093060018A	00173
50 5822068AF	T	AB	3p7307.50R	PESTD.MET	1041	10	1	093060018A	00173

RSX22 0294

Lancaster Laboratories

CHROM PERFECT SEQUENCE FILE

Sequence File: \\cp7\C-Drive\cpwin\DATA1\3p7307.seq

Chromatography Directory: \\cp7\C-Drive\cpwin\DATA1

Method Directory: \\cp7\C-Drive\cpwin\DATA1

Number of Entries: 92

Samplename	Code	ID	FileName	Method	Samp Wt	DF	Int Std	Batch Number	Analysis
51 5822078AF	T	AB	3p7307.51R	PESTD.MET	1030	10	1	093060018A	00173
52 AR1630924A	CCAL	OU	3p7307.52R	PESTD.MET	1	1	1	0930899999	
53 IBLKX0924B	PIBLK	FH	3p7307.53R	PESTD.MET	1000	10	1	0930899999	
54 5822079AF	T	AB	3p7307.54R	PESTD.MET	1056	10	1	093060018A	00173
55 5822080 RI AF	T	AB	3p7307.55R	PESTD.MET	1044	10	1	093060018A	00173
56 5822187 RI AF	T	AB	3p7307.56R	PESTD.MET	1046	10	1	093060018A	00173
57 5822188 RI AF	T	AB	3p7307.57R	PESTD.MET	1051	10	1	093060018A	00173
58 5822189 RI AF	T	AB	3p7307.58R	PESTD.MET	1057	10	1	093060018A	00173
59 5822190 RI AF	T	AB	3p7307.59R	PESTD.MET	1055	10	1	093060018A	00173
60 5822191 RI AF	T	AB	3p7307.60R	PESTD.MET	1046	10	1	093060018A	00173
61 5822192 RI AF	T	AB	3p7307.61R	PESTD.MET	1060	10	1	093060018A	00173
62 5822193 RI AF	T	AB	3p7307.62R	PESTD.MET	1059	10	1	093060018A	00173
63 AR1640924A	CCAL	FO	3p7307.63R	PESTD.MET	1	1	1	0930899999	
64 IBLKX0924B	PIBLK	FH	3p7307.64R	PESTD.MET	1000	10	1	0930899999	
65 5822047ACF DF1000	T	AD	3p7307.65R	PESTD.MET	30	10000	1	093070002A	01216
66 5822048ACF DF100	T	AD	3p7307.66R	PESTD.MET	30	1000	1	093070002A	01216
67 BLANKA 11/4/09 RI ACF	BLK	AB	3p7307.67R	PESTD.MET	30	10	1	093070012A	02033
68 LCSA 11/4/09 RI ACF	LCS	AB	3p7307.68R	PESTD.MET	30	10	1	093070012A	02033
69 5823532 RI ACF	T	AB	3p7307.69R	PESTD.MET	30	10	1	093070012A	02033
70 5821263 RI ACF	T	AB	3p7307.70R	PESTD.MET	30	10	1	093070012A	02033
71 5821263MS RI ACF	MS	AB	3p7307.71R	PESTD.MET	30	10	1	093070012A	02033
72 5821263MSD RI ACF	MSD	AB	3p7307.72R	PESTD.MET	30	10	1	093070012A	02033
73 5821264 RI ACF	T	AB	3p7307.73R	PESTD.MET	30	10	1	093070012A	02033
74 5821265 RI ACF	T	AB	3p7307.74R	PESTD.MET	30	10	1	093070012A	02033
75 AR1630924A	CCAL	OU	3p7307.75R	PESTD.MET	1	1	1	0930899999	
76 IBLKX0924B	PIBLK	FH	3p7307.76R	PESTD.MET	1000	10	1	0930899999	
77 5821266 RI ACF	T	AB	3p7307.77R	PESTD.MET	30	10	1	093070012A	02033
78 5821267 RI ACF	T	AB	3p7307.78R	PESTD.MET	30	10	1	093070012A	02033
79 5821268 RI ACF	T	AB	3p7307.79R	PESTD.MET	30	10	1	093070012A	02033
80 5821269 RI ACF	T	AB	3p7307.80R	PESTD.MET	30	10	1	093070012A	02033
81 5821270 RI ACF	T	AB	3p7307.81R	PESTD.MET	30	10	1	093070012A	02033
82 BLANKA 11/3/09AF	BLK	AB	3p7307.82R	PESTD.MET	1	50	1	093060023A	00174
83 LCSA 11/3/09AF	LCS	AB	3p7307.83R	PESTD.MET	1	50	1	093060023A	00174
84 LCSDA 11/3/09AF	LCSD	AB	3p7307.84R	PESTD.MET	1	50	1	093060023A	00174
85 5822183AF	T	AB	3p7307.85R	PESTD.MET	1	50	1	093060023A	00174
86 5822184AF	T	AB	3p7307.86R	PESTD.MET	1	50	1	093060023A	00174
87 AR1640924A	CCAL	FP	3p7307.87R	PESTD.MET	1	1	1	0930899999	
88 IBLKX0924B	PIBLK	FH	3p7307.88R	PESTD.MET	1000	10	1	0930899999	
89 5822185AF	T	AB	3p7307.89R	PESTD.MET	1	50	1	093060023A	00174
90 5822186AF	T	AB	3p7307.90R	PESTD.MET	1	50	1	093060023A	00174
91 AR1620924A	CCAL	OJ	3p7307.91R	PESTD.MET	1	1	1	0930899999	
92 IBLKX0924B	PIBLK	FH	3p7307.92R	PESTD.MET	1000	10	1	0930899999	

ASX22 6295

Set-up by: [illegible] Date: 11-6-09

11/5/2009

Sample Name: **IBLKX0924B AAPIBLKAA PIBLK0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 03:55:07pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.06R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.583		0.0000	0.000	2491.4	0.411	BV	0.049	840.01	0.390
2	1.929		0.0000	0.000	48117.4	7.944	VB	0.218	3679.53	1.707
3	2.071		0.0000	0.000	38216.9	6.310	BV	0.142	4494.25	2.085
4	2.172		0.0000	0.000	18085.2	2.986	VV	0.068	4429.95	2.055
5	2.312		0.0000	0.000	9096.3	1.502	VV	0.097	1561.12	0.724
6	2.367		0.0000	0.000	5249.4	0.867	VV	0.067	1304.43	0.605
7	2.527		0.0000	0.000	8016.1	1.323	VV	0.140	952.73	0.442
8	2.621		0.0000	0.000	1227.0	0.203	VB	0.083	246.89	0.115
9	2.782		0.0000	0.000	1219.7	0.201	BB	0.053	383.74	0.178
10	2.879		0.0000	0.000	6120.1	1.010	BB	0.062	1654.39	0.767
11	2.998	TCX	200.4876	47.417	184648.0	30.486	BB	0.028	108292.30	50.229
12	3.286		0.0000	0.000	2509.2	0.414	BB	0.026	1613.98	0.749
13	3.339		0.0000	0.000	214.5	0.035	BB	0.019	187.60	0.087
14	3.471		0.0000	0.000	1482.9	0.245	BB	0.068	365.55	0.170
15	3.774		0.0000	0.000	1428.8	0.236	BB	0.077	308.75	0.143
16	3.859		0.0000	0.000	279.6	0.046	BB	0.037	125.86	0.058
17	3.939		0.0000	0.000	379.2	0.063	BB	0.024	258.76	0.120
18	4.102		0.0000	0.000	669.7	0.111	BB	0.026	421.56	0.196
19	4.212		0.0000	0.000	289.6	0.048	BB	0.054	89.94	0.042
20	4.263		0.0000	0.000	253.1	0.042	BB	0.039	109.30	0.051
21	4.403		0.0000	0.000	186.8	0.031	BB	0.036	85.34	0.040
22	4.571		0.0000	0.000	572.9	0.095	BB	0.049	194.10	0.090
23	4.662		0.0000	0.000	401.1	0.066	BB	0.044	153.17	0.071
24	4.806		0.0000	0.000	3532.0	0.583	BB	0.065	902.01	0.418
25	4.928		0.0000	0.000	1416.2	0.234	BB	0.042	560.98	0.260
26	4.995		0.0000	0.000	1341.4	0.221	BB	0.036	626.52	0.291
27	5.105		0.0000	0.000	10880.0	1.796	BB	0.040	4547.11	2.109
28	5.200		0.0000	0.000	1403.1	0.232	BB	0.034	695.46	0.323
29	5.384		0.0000	0.000	1703.2	0.281	BB	0.050	571.13	0.265
30	5.475		0.0000	0.000	205.6	0.034	BB	0.030	112.88	0.052
31	5.522		0.0000	0.000	2853.9	0.471	BB	0.097	488.42	0.227
32	5.636		0.0000	0.000	685.3	0.113	BB	0.037	304.86	0.141
33	5.713		0.0000	0.000	827.6	0.137	BB	0.054	255.88	0.119
34	5.819		0.0000	0.000	1428.1	0.236	BB	0.052	454.12	0.211
35	5.904		0.0000	0.000	1351.7	0.223	BB	0.036	626.11	0.290
36	5.970		0.0000	0.000	3258.3	0.538	BB	0.037	1466.52	0.680
37	6.158		0.0000	0.000	9943.2	1.642	BB	0.275	603.37	0.280
38	6.341		0.0000	0.000	12478.8	2.060	BB	0.101	2057.33	0.954
39	6.471		0.0000	0.000	17331.8	2.862	BB	0.130	2228.32	1.034
40	6.733		0.0000	0.000	8736.6	1.442	BB	0.066	2207.66	1.024
41	6.906		0.0000	0.000	8302.9	1.371	BB	0.287	482.49	0.224
42	6.962		0.0000	0.000	118.4	0.020	BB	0.030	66.43	0.031
43	7.104		0.0000	0.000	10200.0	1.684	BB	0.083	2051.96	0.952
44	7.227		0.0000	0.000	5852.7	0.966	BB	0.086	1136.26	0.527
45	7.292		0.0000	0.000	766.7	0.127	BB	0.038	337.27	0.156
46	7.348		0.0000	0.000	8577.1	1.416	BB	0.101	1410.75	0.654
47	7.526		0.0000	0.000	20632.5	3.406	BB	0.123	2806.43	1.302
48	7.658		0.0000	0.000	4628.4	0.764	BB	0.120	641.80	0.298
49	7.734		0.0000	0.000	1361.3	0.225	BB	0.038	599.97	0.278
50	7.878		0.0000	0.000	1173.2	0.194	BB	0.050	389.89	0.181
51	7.980	DCB	222.3317	52.583	117592.8	19.415	BB	0.037	52318.52	24.267
52	8.126		0.0000	0.000	3376.2	0.557	BB	0.114	495.12	0.230
53	8.393		0.0000	0.000	722.1	0.119	BB	0.067	179.33	0.083
54	8.506		0.0000	0.000	759.4	0.125	BB	0.059	214.17	0.099

ASX22 5296

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	8.606		0.0000	0.000	824.2	0.136	BB	0.042	327.11	0.152
56	8.714		0.0000	0.000	6968.7	1.151	BB	0.097	1203.05	0.558
57	8.967		0.0000	0.000	3294.9	0.544	BB	0.116	473.03	0.219

Total Area = 605683.2, Total Amount = 422.819, Total Height = 215595.4, Sample Units = ug/l

ASXZZ 8297

Sample Name: **IBLKX0924B AAPIBLKAA PIBLK0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 03:55:07pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.06R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.527		0.0000	0.000	124.9	0.025	BV	0.017	121.31	0.066
2	1.572		0.0000	0.000	7470.7	1.506	VV	0.052	2392.31	1.303
3	1.635		0.0000	0.000	9156.4	1.846	VB	0.070	2188.74	1.192
4	1.883		0.0000	0.000	81.9	0.017	BB	0.024	56.40	0.031
5	2.023		0.0000	0.000	600.0	0.121	BB	0.056	178.32	0.097
6	2.113		0.0000	0.000	157.1	0.032	BB	0.021	125.61	0.068
7	2.261		0.0000	0.000	263.5	0.053	BB	0.020	222.59	0.121
8	2.338	TCX	202.1875	49.545	217832.6	43.907	BB	0.039	92691.25	50.494
9	2.618		0.0000	0.000	2589.5	0.522	BB	0.039	1096.96	0.598
10	2.770		0.0000	0.000	434.9	0.088	BB	0.045	159.97	0.087
11	3.143		0.0000	0.000	315.3	0.064	BB	0.038	137.40	0.075
12	3.304		0.0000	0.000	218.2	0.044	BB	0.100	36.50	0.020
13	3.352		0.0000	0.000	2254.0	0.454	BB	0.064	582.58	0.317
14	4.140		0.0000	0.000	1254.7	0.253	BB	0.106	197.23	0.107
15	4.303		0.0000	0.000	11619.6	2.342	BB	0.086	2255.19	1.229
16	4.826		0.0000	0.000	986.2	0.199	BB	0.029	574.45	0.313
17	4.943		0.0000	0.000	3047.2	0.614	BB	0.096	528.38	0.288
18	5.167		0.0000	0.000	3156.3	0.636	BB	0.080	658.81	0.359
19	5.383		0.0000	0.000	6576.9	1.326	BB	0.152	722.15	0.393
20	5.504		0.0000	0.000	23.1	0.005	BB	0.013	28.95	0.016
21	5.553		0.0000	0.000	597.8	0.120	BB	0.033	304.50	0.166
22	5.720		0.0000	0.000	1891.3	0.381	BB	0.128	246.09	0.134
23	5.804		0.0000	0.000	3121.5	0.629	BB	0.045	1160.76	0.632
24	6.027		0.0000	0.000	1397.0	0.282	BB	0.054	433.94	0.236
25	6.097		0.0000	0.000	694.3	0.140	BB	0.050	232.83	0.127
26	6.237		0.0000	0.000	6182.3	1.246	BB	0.093	1109.45	0.604
27	6.508		0.0000	0.000	4816.7	0.971	BB	1.583	50.72	0.028
28	6.524		0.0000	0.000	398.6	0.080	BB	0.033	203.95	0.111
29	6.590		0.0000	0.000	108.9	0.022	BB	0.029	61.66	0.034
30	6.627		0.0000	0.000	1781.5	0.359	BB	0.051	582.21	0.317
31	6.803	DCB	205.9033	50.455	180451.8	36.373	BB	0.041	73087.41	39.815
32	7.170		0.0000	0.000	6980.8	1.407	BB	0.817	142.33	0.078
33	7.637		0.0000	0.000	15016.3	3.027	BB	0.360	694.37	0.378
34	8.310		0.0000	0.000	4516.3	0.910	BB	0.248	303.72	0.165

Total Area = 496118.3, Total Amount = 408.091, Total Height = 183569.0, Sample Units = ug/l

4SX22 8298

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.06R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 3:45:58 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.06R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 3:45:58 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1000
Analyst: 2280

Dilution Factor: 10

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.998	108292	.20049	TCX	2.338	92691	.20219	TCX
7.98	52319	.22233	DCB	6.803	73087	.2059	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.06A
Area File: C:\CPWIN\DATA1\1P7307B.06A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 3:55:14 PM
File Reported On: 11/3/2009 at 3:55:27 PM

10X REPLOTS

ASX22 0388

C:\CPWIN\DATA1\1P7307.06R



C:\CPWIN\DATA1\1P7307B.06R



ASX22 8381

Sample Name: **AR1610924A AAAR161AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 04:09:01pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.07R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.515		0.0000	0.000	64.5	0.010	BB	0.019	56.97	0.021
2	1.585		0.0000	0.000	1590.6	0.251	BV	0.040	664.55	0.242
3	1.798		0.0000	0.000	14198.3	2.237	VV	0.160	1482.73	0.539
4	1.867		0.0000	0.000	7565.7	1.192	VV	0.070	1791.95	0.652
5	1.932		0.0000	0.000	6612.2	1.042	VB	0.052	2132.76	0.776
6	2.072		0.0000	0.000	18934.1	2.984	BV	0.121	2611.29	0.950
7	2.113		0.0000	0.000	6487.1	1.022	VV	0.046	2376.07	0.864
8	2.174		0.0000	0.000	10014.8	1.578	VV	0.046	3651.25	1.328
9	2.319		0.0000	0.000	3204.9	0.505	VV	0.077	690.10	0.251
10	2.372		0.0000	0.000	1386.1	0.218	VV	0.044	527.77	0.192
11	2.529		0.0000	0.000	4278.9	0.674	VV	0.081	884.49	0.322
12	2.598		0.0000	0.000	1438.8	0.227	VB	0.037	642.83	0.234
13	2.782		0.0000	0.000	1551.3	0.244	BB	0.062	413.88	0.151
14	2.824		0.0000	0.000	13250.0	2.088	BB	0.167	1324.98	0.482
15	3.000	TCX	0.8318	12.452	71965.9	11.340	BB	0.042	28830.40	10.485
16	3.227		0.0000	0.000	1413.5	0.223	BB	0.025	949.70	0.345
17	3.311		0.0000	0.000	2033.2	0.320	BB	0.029	1156.52	0.421
18	3.381		0.0000	0.000	9311.3	1.467	BB	0.026	5954.17	2.165
19	3.515		0.0000	0.000	1394.6	0.220	BB	0.030	765.55	0.278
20	3.694		0.0000	0.000	18439.6	2.906	BB	0.038	8040.99	2.924
21	3.772		0.0000	0.000	1072.7	0.169	BB	0.027	656.58	0.239
22	3.825		0.0000	0.000	265.1	0.042	BB	0.015	286.57	0.104
23	3.863		0.0000	0.000	11590.9	1.826	BB	0.026	7381.13	2.684
24	3.974		0.0000	0.000	193.3	0.030	BB	0.021	153.20	0.056
25	4.030		0.0000	0.000	5011.1	0.790	BB	0.038	2183.46	0.794
26	4.127		0.0000	0.000	32057.0	5.051	BB	0.035	15353.77	5.584
27	4.181		0.0000	0.000	9796.6	1.544	BB	0.024	6680.84	2.430
28	4.266		0.0000	0.000	10174.4	1.603	BB	0.027	6395.18	2.326
29	4.317		0.0000	0.000	288.5	0.045	BB	0.016	291.92	0.106
30	4.361		0.0000	0.000	4542.3	0.716	BB	0.020	3880.69	1.411
31	4.397		0.0000	0.000	7867.0	1.240	BB	0.032	4098.36	1.491
32	4.528		0.0000	0.000	12779.4	2.014	BB	0.036	5975.25	2.173
33	4.599		0.0000	0.000	1081.8	0.170	BB	0.017	1080.71	0.393
34	4.649		0.0000	0.000	7630.9	1.202	BB	0.028	4531.45	1.648
35	4.698		0.0000	0.000	1387.6	0.219	BB	0.020	1148.73	0.418
36	4.813		0.0000	0.000	1121.6	0.177	BB	0.041	451.34	0.164
37	4.857		0.0000	0.000	4218.3	0.665	BB	0.020	3551.16	1.292
38	4.887		0.0000	0.000	199.1	0.031	BB	0.010	328.46	0.119
39	4.916		0.0000	0.000	1335.7	0.210	BB	0.020	1098.55	0.400
40	5.001		0.0000	0.000	1900.3	0.299	BB	0.032	988.37	0.359
41	5.060		0.0000	0.000	4946.3	0.779	BB	0.025	3325.27	1.209
42	5.104		0.0000	0.000	1843.5	0.290	BB	0.036	841.84	0.306
43	5.202		0.0000	0.000	2113.4	0.333	BB	0.043	825.85	0.300
44	5.311		0.0000	0.000	3683.9	0.580	BB	0.029	2130.52	0.775
45	5.378		0.0000	0.000	2772.7	0.437	BB	0.025	1875.49	0.682
46	5.430		0.0000	0.000	11713.2	1.846	BB	0.037	5276.27	1.919
47	5.523		0.0000	0.000	18704.9	2.947	BB	0.029	10861.18	3.950
48	5.636		0.0000	0.000	2569.6	0.405	BB	0.037	1143.17	0.416
49	5.706		0.0000	0.000	2290.6	0.361	BB	0.026	1484.20	0.540
50	5.769		0.0000	0.000	20215.4	3.185	BB	0.029	11559.69	4.204
51	5.813		0.0000	0.000	1389.7	0.219	BB	0.017	1342.17	0.488
52	5.879		0.0000	0.000	8801.0	1.387	BB	0.031	4692.19	1.707
53	5.971		0.0000	0.000	1336.0	0.211	BB	0.025	903.02	0.328
54	6.014		0.0000	0.000	13944.4	2.197	BB	0.028	8186.33	2.977

ASX22 B302

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.086		0.0000	0.000	11332.0	1.786	BB	0.024	7794.12	2.835
56	6.137		0.0000	0.000	5822.2	0.917	BB	0.025	3831.22	1.393
57	6.283		0.0000	0.000	16217.0	2.555	BB	0.033	8098.89	2.946
58	6.354		0.0000	0.000	7328.4	1.155	BB	0.031	3981.06	1.448
59	6.452		0.0000	0.000	2034.1	0.321	BB	0.044	764.61	0.278
60	6.513		0.0000	0.000	2675.6	0.422	BB	0.026	1719.05	0.625
61	6.582		0.0000	0.000	33372.9	5.259	BB	0.031	18183.39	6.613
62	6.735		0.0000	0.000	11132.9	1.754	BB	0.071	2597.29	0.945
63	6.836		0.0000	0.000	16130.0	2.542	BB	0.032	8415.07	3.060
64	6.896		0.0000	0.000	8089.3	1.275	BB	0.037	3618.79	1.316
65	7.062		0.0000	0.000	232.8	0.037	BB	0.030	127.63	0.046
66	7.100		0.0000	0.000	2817.7	0.444	BB	0.048	973.26	0.354
67	7.188		0.0000	0.000	9735.2	1.534	BB	0.052	3116.52	1.133
68	7.347		0.0000	0.000	14112.0	2.224	BB	0.096	2459.74	0.895
69	7.529		0.0000	0.000	28080.9	4.425	BB	0.132	3556.94	1.294
70	7.663		0.0000	0.000	4222.6	0.665	BB	0.157	449.66	0.164
71	7.735		0.0000	0.000	3074.7	0.484	BB	0.037	1382.30	0.503
72	7.871		0.0000	0.000	1320.5	0.208	BB	0.057	387.37	0.141
73	7.982	DCB	5.8484	87.548	32641.1	5.143	BB	0.040	13762.36	5.005
74	8.134		0.0000	0.000	2068.2	0.326	BB	0.086	402.35	0.146
75	8.389		0.0000	0.000	3155.4	0.497	BB	0.079	663.88	0.241
76	8.517		0.0000	0.000	1770.6	0.279	BB	0.060	488.81	0.178
77	8.604		0.0000	0.000	1087.3	0.171	BB	0.044	407.66	0.148
78	8.717		0.0000	0.000	8267.6	1.303	BB	0.091	1519.42	0.553
79	8.964		0.0000	0.000	1917.1	0.302	BB	0.093	344.30	0.125

Total Area = 634611.9, Total Amount = 6.68, Total Height = 274957.6, Sample Units = ug/l

ASX22 0303

Sample Name: **AR1610924A AAAR161AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 04:09:01pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.07R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.535		0.0000	0.000	117.9	0.022	BV	0.017	116.95	0.054
2	1.569		0.0000	0.000	6638.2	1.219	VV	0.044	2539.03	1.180
3	1.636		0.0000	0.000	9653.7	1.773	VB	0.074	2178.97	1.013
4	2.064		0.0000	0.000	687.0	0.126	BB	0.037	307.93	0.143
5	2.339	TCX	4.4493	79.905	50080.8	9.199	BB	0.041	20397.46	9.482
6	2.561		0.0000	0.000	1274.9	0.234	BB	0.032	660.44	0.307
7	2.643		0.0000	0.000	1676.1	0.308	BB	0.035	790.80	0.368
8	2.714		0.0000	0.000	10805.5	1.985	BB	0.043	4210.37	1.957
9	2.860		0.0000	0.000	1440.3	0.265	BB	0.036	666.92	0.310
10	3.007		0.0000	0.000	28242.8	5.188	BB	0.049	9640.04	4.481
11	3.170		0.0000	0.000	16539.7	3.038	BB	0.042	6630.34	3.082
12	3.311		0.0000	0.000	3144.3	0.578	BB	0.045	1169.10	0.543
13	3.404		0.0000	0.000	19112.9	3.511	BB	0.035	9150.14	4.254
14	3.450		0.0000	0.000	3149.0	0.578	BB	0.029	1825.31	0.849
15	3.557		0.0000	0.000	7337.9	1.348	BB	0.032	3837.68	1.784
16	3.618		0.0000	0.000	1341.5	0.246	BB	0.023	961.95	0.447
17	3.645		0.0000	0.000	6689.9	1.229	BB	0.040	2778.48	1.292
18	3.798		0.0000	0.000	17829.5	3.275	BB	0.056	5319.69	2.473
19	3.922		0.0000	0.000	3725.2	0.684	BB	0.029	2141.82	0.996
20	3.972		0.0000	0.000	844.9	0.155	BB	0.023	603.88	0.281
21	4.109		0.0000	0.000	17123.4	3.145	BB	0.055	5187.03	2.411
22	4.254		0.0000	0.000	15862.9	2.914	BB	0.071	3702.91	1.721
23	4.513		0.0000	0.000	443.3	0.081	BB	0.073	100.89	0.047
24	4.543		0.0000	0.000	1584.5	0.291	BB	0.024	1089.86	0.507
25	4.598		0.0000	0.000	16439.7	3.020	BB	0.047	5797.63	2.695
26	4.695		0.0000	0.000	22980.2	4.221	BB	0.036	10504.16	4.883
27	4.823		0.0000	0.000	3127.5	0.574	BB	0.034	1528.60	0.711
28	4.875		0.0000	0.000	24735.3	4.544	BB	0.038	10760.01	5.002
29	4.985		0.0000	0.000	24331.9	4.470	BB	0.054	7464.02	3.470
30	5.101		0.0000	0.000	1658.4	0.305	BB	0.023	1201.58	0.559
31	5.144		0.0000	0.000	7660.2	1.407	BB	0.024	5221.76	2.427
32	5.178		0.0000	0.000	7795.1	1.432	BB	0.026	4911.25	2.283
33	5.339		0.0000	0.000	1541.9	0.283	BB	0.027	953.02	0.443
34	5.392		0.0000	0.000	19588.4	3.598	BB	0.037	8810.93	4.096
35	5.475		0.0000	0.000	4517.9	0.830	BB	0.025	2974.73	1.383
36	5.512		0.0000	0.000	773.7	0.142	BB	0.019	682.45	0.317
37	5.562		0.0000	0.000	1166.3	0.214	BB	0.019	1028.32	0.478
38	5.630		0.0000	0.000	44134.1	8.107	BB	0.043	16964.27	7.886
39	5.804		0.0000	0.000	1852.1	0.340	BB	0.030	1038.58	0.483
40	5.871		0.0000	0.000	5313.1	0.976	BB	0.026	3376.59	1.570
41	5.926		0.0000	0.000	26473.8	4.863	BB	0.037	11799.47	5.485
42	6.094		0.0000	0.000	810.2	0.149	BB	0.036	376.69	0.175
43	6.186		0.0000	0.000	409.4	0.075	BB	0.041	165.82	0.077
44	6.241		0.0000	0.000	18299.0	3.361	BB	0.054	5654.96	2.629
45	6.389		0.0000	0.000	22109.5	4.061	BB	0.058	6325.13	2.940
46	6.533		0.0000	0.000	1159.6	0.213	BB	0.050	387.48	0.180
47	6.626		0.0000	0.000	4351.1	0.799	BB	0.046	1565.24	0.728
48	6.805	DCB	1.1189	20.095	45797.1	8.412	BB	0.042	18116.29	8.421
49	7.017		0.0000	0.000	392.4	0.072	BB	0.042	154.37	0.072
50	7.637		0.0000	0.000	9716.4	1.785	BB	0.221	733.42	0.341
51	8.313		0.0000	0.000	950.5	0.175	BB	0.058	274.58	0.128
52	8.466		0.0000	0.000	163.2	0.030	BB	0.035	76.69	0.036
53	8.544		0.0000	0.000	569.6	0.105	BB	0.049	194.89	0.091
54	8.906		0.0000	0.000	235.3	0.043	BB	0.057	69.22	0.032

ASX22 8384

Total Area = 544399.0, Total Amount = 5.568, Total Height = 215120.1, Sample Units = ug/l

ASX22 8385

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.07R

C:\CPWIN\DATA1\1P7307.07R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 3:59:52 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.07R

C:\CPWIN\DATA1\1P7307B.07R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 3:59:52 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.	28830	.83182	TCX	2.339	20397	4.4493	TCX
7.982	13762	5.84842	DCB	6.805	18116	1.11893	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.07A
Area File: C:\CPWIN\DATA1\1P7307B.07A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 4:09:08 PM
File Reported On: 11/3/2009 at 4:09:23 PM

10X REPLOTS

ASX22 0387

C:\CPWIN\DATA1\1P7307.07R



C:\CPWIN\DATA1\1P7307B.07R



ASX22 0388

Sample Name: **AR1620924A AAAR162AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 04:22:00pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.08R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.615		0.0000	0.000	6849.5	0.574	BV	0.068	1674.09	0.281
2	1.661		0.0000	0.000	5235.1	0.439	VV	0.047	1875.22	0.315
3	1.796		0.0000	0.000	22288.4	1.869	VV	0.112	3324.95	0.558
4	1.868		0.0000	0.000	15313.9	1.284	VV	0.063	4044.67	0.678
5	1.927		0.0000	0.000	14920.2	1.251	VB	0.053	4705.19	0.789
6	2.033		0.0000	0.000	21847.6	1.832	BV	0.078	4660.46	0.782
7	2.074		0.0000	0.000	20263.7	1.700	VV	0.069	4863.28	0.816
8	2.173		0.0000	0.000	28469.4	2.388	VV	0.079	6000.71	1.007
9	2.402		0.0000	0.000	21543.2	1.807	VV	0.210	1712.22	0.287
10	2.446		0.0000	0.000	5789.1	0.486	VV	0.066	1456.58	0.244
11	2.523		0.0000	0.000	4155.7	0.349	VV	0.060	1153.38	0.193
12	2.588		0.0000	0.000	1388.3	0.116	VV	0.051	454.18	0.076
13	2.625		0.0000	0.000	442.8	0.037	VB	0.034	214.75	0.036
14	2.789		0.0000	0.000	2271.1	0.190	BB	0.064	587.22	0.099
15	2.881		0.0000	0.000	5487.6	0.460	BB	0.062	1478.28	0.248
16	2.998	TCX	19.9153	47.858	176744.0	14.824	BB	0.028	106490.30	17.863
17	3.226		0.0000	0.000	2936.7	0.246	BB	0.026	1910.88	0.321
18	3.309		0.0000	0.000	5848.0	0.490	BB	0.039	2478.18	0.416
19	3.381		0.0000	0.000	17711.3	1.485	BB	0.026	11252.64	1.888
20	3.471		0.0000	0.000	251.5	0.021	BB	0.019	215.31	0.036
21	3.514		0.0000	0.000	2132.8	0.179	BB	0.025	1417.58	0.238
22	3.583		0.0000	0.000	187.9	0.016	BB	0.038	82.65	0.014
23	3.694		0.0000	0.000	35589.2	2.985	BB	0.039	15271.24	2.562
24	3.771		0.0000	0.000	1994.7	0.167	BB	0.027	1237.31	0.208
25	3.824		0.0000	0.000	577.0	0.048	BB	0.016	599.12	0.101
26	3.862		0.0000	0.000	21565.9	1.809	BB	0.026	13765.32	2.309
27	3.939		0.0000	0.000	151.2	0.013	BB	0.017	147.25	0.025
28	3.970		0.0000	0.000	350.7	0.029	BB	0.022	267.10	0.045
29	4.030		0.0000	0.000	10154.8	0.852	BB	0.038	4440.68	0.745
30	4.126		0.0000	0.000	58728.6	4.926	BB	0.033	29721.02	4.986
31	4.180		0.0000	0.000	19018.5	1.595	BB	0.024	13067.38	2.192
32	4.265		0.0000	0.000	19395.2	1.627	BB	0.027	12068.33	2.024
33	4.315		0.0000	0.000	515.8	0.043	BB	0.016	550.41	0.092
34	4.360		0.0000	0.000	8908.1	0.747	BB	0.020	7530.33	1.263
35	4.396		0.0000	0.000	15319.0	1.285	BB	0.032	7908.67	1.327
36	4.528		0.0000	0.000	25010.0	2.098	BB	0.036	11727.67	1.967
37	4.599		0.0000	0.000	2291.1	0.192	BB	0.017	2231.23	0.374
38	4.649		0.0000	0.000	14568.5	1.222	BB	0.028	8707.42	1.461
39	4.698		0.0000	0.000	2825.7	0.237	BB	0.021	2214.48	0.371
40	4.762		0.0000	0.000	254.8	0.021	BB	0.018	234.81	0.039
41	4.815		0.0000	0.000	607.7	0.051	BB	0.024	416.72	0.070
42	4.857		0.0000	0.000	13322.1	1.117	BB	0.027	8329.68	1.397
43	4.915		0.0000	0.000	1849.8	0.155	BB	0.019	1600.52	0.268
44	5.002		0.0000	0.000	1933.7	0.162	BB	0.027	1188.57	0.199
45	5.059		0.0000	0.000	10230.7	0.858	BB	0.025	6939.75	1.164
46	5.105		0.0000	0.000	3751.3	0.315	BB	0.026	2374.79	0.398
47	5.202		0.0000	0.000	1530.4	0.128	BB	0.040	632.46	0.106
48	5.310		0.0000	0.000	7598.0	0.637	BB	0.030	4263.75	0.715
49	5.376		0.0000	0.000	5056.3	0.424	BB	0.024	3461.37	0.581
50	5.430		0.0000	0.000	23199.2	1.946	BB	0.037	10513.15	1.764
51	5.522		0.0000	0.000	33850.0	2.839	BB	0.027	20543.28	3.446
52	5.636		0.0000	0.000	4056.5	0.340	BB	0.037	1833.78	0.308
53	5.704		0.0000	0.000	4723.6	0.396	BB	0.025	3111.11	0.522
54	5.768		0.0000	0.000	38418.3	3.222	BB	0.029	22281.65	3.738

ASX22 0389

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.813		0.0000	0.000	2789.9	0.234	BB	0.017	2780.73	0.466
56	5.877		0.0000	0.000	17868.9	1.499	BB	0.030	10065.71	1.688
57	5.969		0.0000	0.000	1617.5	0.136	BB	0.023	1148.46	0.193
58	6.013		0.0000	0.000	32607.5	2.735	BB	0.029	18614.85	3.123
59	6.085		0.0000	0.000	21362.7	1.792	BB	0.024	14894.77	2.499
60	6.136		0.0000	0.000	10669.0	0.895	BB	0.024	7363.85	1.235
61	6.281		0.0000	0.000	32257.2	2.705	BB	0.033	16096.45	2.700
62	6.354		0.0000	0.000	12314.5	1.033	BB	0.027	7549.75	1.266
63	6.402		0.0000	0.000	812.3	0.068	BB	0.015	881.42	0.148
64	6.449		0.0000	0.000	1086.5	0.091	BB	0.027	675.56	0.113
65	6.512		0.0000	0.000	5672.6	0.476	BB	0.026	3591.67	0.602
66	6.580		0.0000	0.000	61479.5	5.156	BB	0.030	34202.23	5.737
67	6.727		0.0000	0.000	3530.2	0.296	BB	0.061	972.20	0.163
68	6.834		0.0000	0.000	28478.3	2.389	BB	0.030	15985.15	2.681
69	6.895		0.0000	0.000	15441.4	1.295	BB	0.036	7076.86	1.187
70	7.059		0.0000	0.000	607.8	0.051	BB	0.028	367.95	0.062
71	7.186		0.0000	0.000	11973.6	1.004	BB	0.042	4696.99	0.788
72	7.289		0.0000	0.000	549.4	0.046	BB	0.035	260.93	0.044
73	7.352		0.0000	0.000	156.9	0.013	BB	0.027	98.38	0.017
74	7.419		0.0000	0.000	11976.6	1.004	BB	0.029	6947.62	1.165
75	7.493		0.0000	0.000	4659.5	0.391	BB	0.078	993.31	0.167
76	7.654		0.0000	0.000	4169.6	0.350	BB	0.112	620.64	0.104
77	7.733		0.0000	0.000	6022.9	0.505	BB	0.035	2907.82	0.488
78	7.878		0.0000	0.000	921.3	0.077	BB	0.046	330.93	0.056
79	7.979	DCB	21.6982	52.142	116366.5	9.760	BB	0.037	52687.98	8.838
80	8.122		0.0000	0.000	3447.0	0.289	BB	0.113	509.22	0.085
81	8.316		0.0000	0.000	70.6	0.006	BB	0.023	50.69	0.009
82	8.393		0.0000	0.000	687.9	0.058	BB	0.056	205.54	0.034
83	8.598		0.0000	0.000	444.4	0.037	BB	0.042	175.76	0.029
84	8.708		0.0000	0.000	9946.4	0.834	BB	0.096	1733.13	0.291
85	8.964		0.0000	0.000	2895.0	0.243	BB	0.116	416.50	0.070

Total Area = 1192307.0, Total Amount = 41.614, Total Height = 596136.1, Sample Units = ug/l

ASX22 0310

Sample Name: **AR1620924A AAAR162AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 04:22:00pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.08R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.573		0.0000	0.000	8611.6	0.718	BV	0.051	2796.40	0.567
2	1.637		0.0000	0.000	11425.8	0.953	VV	0.072	2640.27	0.536
3	1.793		0.0000	0.000	513.6	0.043	VV	0.047	181.67	0.037
4	1.877		0.0000	0.000	268.8	0.022	VB	0.035	129.60	0.026
5	1.974		0.0000	0.000	219.4	0.018	BV	0.038	95.59	0.019
6	2.016		0.0000	0.000	923.1	0.077	VB	0.049	316.95	0.064
7	2.107		0.0000	0.000	198.4	0.017	BB	0.023	141.46	0.029
8	2.269		0.0000	0.000	395.9	0.033	BB	0.024	271.05	0.055
9	2.340	TCX	20.4848	49.852	213075.6	17.771	BB	0.039	91446.56	18.551
10	2.563		0.0000	0.000	2467.9	0.206	BB	0.032	1304.14	0.265
11	2.640		0.0000	0.000	4315.9	0.360	BB	0.040	1778.92	0.361
12	2.715		0.0000	0.000	20771.8	1.732	BB	0.043	8042.40	1.632
13	2.858		0.0000	0.000	2784.8	0.232	BB	0.036	1277.20	0.259
14	3.007		0.0000	0.000	53889.2	4.495	BB	0.049	18408.09	3.734
15	3.170		0.0000	0.000	30488.7	2.543	BB	0.041	12509.70	2.538
16	3.313		0.0000	0.000	7562.2	0.631	BB	0.046	2733.54	0.555
17	3.404		0.0000	0.000	36894.3	3.077	BB	0.034	17947.49	3.641
18	3.449		0.0000	0.000	5957.7	0.497	BB	0.029	3443.24	0.699
19	3.557		0.0000	0.000	13810.6	1.152	BB	0.031	7345.57	1.490
20	3.618		0.0000	0.000	2734.2	0.228	BB	0.025	1858.80	0.377
21	3.644		0.0000	0.000	13018.7	1.086	BB	0.040	5419.02	1.099
22	3.797		0.0000	0.000	32414.7	2.704	BB	0.053	10287.58	2.087
23	3.885		0.0000	0.000	140.7	0.012	BB	0.011	222.74	0.045
24	3.923		0.0000	0.000	7606.9	0.634	BB	0.029	4373.17	0.887
25	3.971		0.0000	0.000	1663.4	0.139	BB	0.026	1077.86	0.219
26	4.108		0.0000	0.000	32798.3	2.736	BB	0.054	10066.20	2.042
27	4.253		0.0000	0.000	25603.5	2.135	BB	0.062	6872.68	1.394
28	4.507		0.0000	0.000	658.8	0.055	BB	0.028	388.26	0.079
29	4.542		0.0000	0.000	3129.8	0.261	BB	0.024	2131.29	0.432
30	4.598		0.0000	0.000	31212.8	2.603	BB	0.050	10501.10	2.130
31	4.694		0.0000	0.000	43650.2	3.641	BB	0.036	20143.68	4.086
32	4.822		0.0000	0.000	5044.6	0.421	BB	0.034	2502.45	0.508
33	4.874		0.0000	0.000	47499.8	3.962	BB	0.037	21130.13	4.287
34	4.985		0.0000	0.000	50809.4	4.238	BB	0.054	15547.44	3.154
35	5.101		0.0000	0.000	3296.8	0.275	BB	0.023	2395.78	0.486
36	5.144		0.0000	0.000	14985.0	1.250	BB	0.025	10022.43	2.033
37	5.177		0.0000	0.000	15493.5	1.292	BB	0.030	8600.43	1.745
38	5.337		0.0000	0.000	3216.5	0.268	BB	0.026	2083.85	0.423
39	5.392		0.0000	0.000	37616.2	3.137	BB	0.036	17228.91	3.495
40	5.475		0.0000	0.000	9089.8	0.758	BB	0.026	5731.98	1.163
41	5.511		0.0000	0.000	1358.3	0.113	BB	0.021	1083.10	0.220
42	5.561		0.0000	0.000	2276.1	0.190	BB	0.023	1640.12	0.333
43	5.629		0.0000	0.000	90681.3	7.563	BB	0.044	34125.64	6.923
44	5.801		0.0000	0.000	1110.4	0.093	BB	0.023	793.37	0.161
45	5.870		0.0000	0.000	10830.0	0.903	BB	0.027	6711.66	1.362
46	5.925		0.0000	0.000	52171.9	4.351	BB	0.038	23022.37	4.670
47	6.093		0.0000	0.000	1230.9	0.103	BB	0.035	591.46	0.120
48	6.176		0.0000	0.000	1108.9	0.092	BB	0.031	589.72	0.120
49	6.240		0.0000	0.000	12482.1	1.041	BB	0.044	4725.65	0.959
50	6.385		0.0000	0.000	28747.0	2.398	BB	0.045	10605.40	2.151
51	6.624		0.0000	0.000	9045.4	0.754	BB	0.043	3483.75	0.707
52	6.802	DCB	20.6064	50.148	177411.6	14.797	BB	0.041	72456.62	14.699
53	7.510		0.0000	0.000	4288.9	0.358	BB	0.116	615.08	0.125
54	7.624		0.0000	0.000	7692.7	0.642	BB	0.221	580.33	0.118

ASX22 8311

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	8.203		0.0000	0.000	605.4	0.050	BB	0.060	168.88	0.034
56	8.323		0.0000	0.000	469.1	0.039	BB	0.042	185.08	0.038
57	8.910		0.0000	0.000	1217.3	0.102	BB	0.118	171.22	0.035

Total Area = 1198986.0, Total Amount = 41.091, Total Height = 492945.2, Sample Units = ug/l

ASX22 8312

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.08R

C:\CPWIN\DATA1\1P7307.08R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 4:12:51 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.08R

C:\CPWIN\DATA1\1P7307B.08R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 4:12:51 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.998	106490	19.91528	TCX	2.34	91447	20.48481	TCX
7.979	52688	21.69823	DCB	6.802	72457	20.60636	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.08A
Area File: C:\CPWIN\DATA1\1P7307B.08A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 4:22:08 PM
File Reported On: 11/3/2009 at 4:22:21 PM

10X REPLOTS

ASX22 0314

C:\CPWIN\DATA1\1P7307.08R



C:\CPWIN\DATA1\1P7307B.08R



ASX22 03.15

Sample Name: **AR1630924A AAAR163AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 04:34:56pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.09R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.579		0.0000	0.000	2364.5	0.114	BV	0.048	815.18	0.075
2	1.787		0.0000	0.000	30752.5	1.486	VV	0.146	3501.78	0.321
3	1.876		0.0000	0.000	24022.4	1.160	VV	0.087	4579.89	0.420
4	1.927		0.0000	0.000	11959.5	0.578	VB	0.038	5247.03	0.481
5	2.032		0.0000	0.000	23993.5	1.159	BV	0.077	5209.65	0.478
6	2.074		0.0000	0.000	22307.1	1.078	VV	0.063	5886.15	0.540
7	2.174		0.0000	0.000	33142.6	1.601	VV	0.096	5742.83	0.527
8	2.268		0.0000	0.000	8586.8	0.415	VV	0.048	2972.60	0.273
9	2.447		0.0000	0.000	17460.9	0.843	VV	0.183	1589.33	0.146
10	2.525		0.0000	0.000	3638.0	0.176	VV	0.047	1285.18	0.118
11	2.594		0.0000	0.000	1900.2	0.092	VB	0.063	500.54	0.046
12	2.778		0.0000	0.000	4029.5	0.195	BB	0.078	861.06	0.079
13	2.890		0.0000	0.000	5031.7	0.243	BB	0.063	1327.12	0.122
14	2.999	TCX	41.0852	51.177	312535.3	15.098	BB	0.027	192864.90	17.687
15	3.227		0.0000	0.000	5993.6	0.290	BB	0.026	3837.80	0.352
16	3.309		0.0000	0.000	11599.5	0.560	BB	0.039	4928.20	0.452
17	3.381		0.0000	0.000	33595.2	1.623	BB	0.026	21383.81	1.961
18	3.473		0.0000	0.000	462.1	0.022	BB	0.019	412.98	0.038
19	3.514		0.0000	0.000	4324.7	0.209	BB	0.025	2883.79	0.264
20	3.576		0.0000	0.000	229.4	0.011	BB	0.049	77.56	0.007
21	3.694		0.0000	0.000	68140.2	3.292	BB	0.039	28984.22	2.658
22	3.771		0.0000	0.000	3785.2	0.183	BB	0.027	2363.31	0.217
23	3.824		0.0000	0.000	1193.1	0.058	BB	0.016	1241.24	0.114
24	3.863		0.0000	0.000	40245.9	1.944	BB	0.026	25774.42	2.364
25	3.939		0.0000	0.000	318.5	0.015	BB	0.017	320.74	0.029
26	3.970		0.0000	0.000	659.4	0.032	BB	0.020	537.12	0.049
27	4.030		0.0000	0.000	19452.1	0.940	BB	0.038	8453.58	0.775
28	4.126		0.0000	0.000	111135.8	5.369	BB	0.033	56217.88	5.156
29	4.181		0.0000	0.000	36094.1	1.744	BB	0.024	24854.33	2.279
30	4.266		0.0000	0.000	37411.1	1.807	BB	0.027	23478.22	2.153
31	4.316		0.0000	0.000	1125.2	0.054	BB	0.016	1189.84	0.109
32	4.360		0.0000	0.000	16859.8	0.814	BB	0.020	14371.74	1.318
33	4.396		0.0000	0.000	28575.5	1.380	BB	0.032	14775.14	1.355
34	4.528		0.0000	0.000	48816.4	2.358	BB	0.036	22628.24	2.075
35	4.599		0.0000	0.000	4460.6	0.215	BB	0.017	4334.18	0.397
36	4.649		0.0000	0.000	27783.7	1.342	BB	0.028	16700.94	1.532
37	4.698		0.0000	0.000	5330.0	0.257	BB	0.021	4301.10	0.394
38	4.763		0.0000	0.000	717.4	0.035	BB	0.020	586.39	0.054
39	4.815		0.0000	0.000	866.6	0.042	BB	0.023	634.45	0.058
40	4.857		0.0000	0.000	28055.3	1.355	BB	0.028	16807.88	1.541
41	4.914		0.0000	0.000	2831.1	0.137	BB	0.017	2726.17	0.250
42	5.002		0.0000	0.000	2892.9	0.140	BB	0.025	1966.17	0.180
43	5.058		0.0000	0.000	21694.0	1.048	BB	0.025	14334.30	1.315
44	5.105		0.0000	0.000	2853.9	0.138	BB	0.022	2127.22	0.195
45	5.202		0.0000	0.000	470.2	0.023	BB	0.023	337.16	0.031
46	5.232		0.0000	0.000	283.7	0.014	BB	0.018	269.16	0.025
47	5.310		0.0000	0.000	15210.5	0.735	BB	0.030	8567.01	0.786
48	5.376		0.0000	0.000	9812.3	0.474	BB	0.024	6837.80	0.627
49	5.429		0.0000	0.000	45094.7	2.178	BB	0.037	20431.50	1.874
50	5.521		0.0000	0.000	62631.2	3.026	BB	0.027	38918.96	3.569
51	5.636		0.0000	0.000	7628.4	0.369	BB	0.037	3458.19	0.317
52	5.704		0.0000	0.000	9376.0	0.453	BB	0.025	6185.41	0.567
53	5.767		0.0000	0.000	72817.8	3.518	BB	0.029	42420.06	3.890
54	5.812		0.0000	0.000	5450.0	0.263	BB	0.017	5499.95	0.504

A9X2Z 0316

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.876		0.0000	0.000	35041.8	1.693	BB	0.029	20310.79	1.863
56	5.969		0.0000	0.000	2428.3	0.117	BB	0.023	1795.04	0.165
57	6.012		0.0000	0.000	66083.8	3.192	BB	0.029	37511.36	3.440
58	6.084		0.0000	0.000	40649.4	1.964	BB	0.024	28386.14	2.603
59	6.136		0.0000	0.000	21218.8	1.025	BB	0.024	14696.57	1.348
60	6.281		0.0000	0.000	64831.3	3.132	BB	0.034	31489.60	2.888
61	6.353		0.0000	0.000	22716.4	1.097	BB	0.027	14257.90	1.308
62	6.401		0.0000	0.000	3248.0	0.157	BB	0.019	2827.23	0.259
63	6.447		0.0000	0.000	1379.5	0.067	BB	0.022	1039.02	0.095
64	6.511		0.0000	0.000	12026.3	0.581	BB	0.027	7525.48	0.690
65	6.579		0.0000	0.000	117848.2	5.693	BB	0.030	66033.00	6.056
66	6.833		0.0000	0.000	53863.2	2.602	BB	0.030	30349.23	2.783
67	6.893		0.0000	0.000	29830.1	1.441	BB	0.037	13574.65	1.245
68	7.058		0.0000	0.000	6866.3	0.332	BB	0.080	1434.61	0.132
69	7.184		0.0000	0.000	18745.9	0.906	BB	0.037	8429.52	0.773
70	7.288		0.0000	0.000	715.0	0.035	BB	0.038	312.27	0.029
71	7.349		0.0000	0.000	138.2	0.007	BB	0.022	102.58	0.009
72	7.418		0.0000	0.000	27461.8	1.327	BB	0.030	15204.22	1.394
73	7.490		0.0000	0.000	5017.7	0.242	BB	0.064	1306.40	0.120
74	7.656		0.0000	0.000	4277.6	0.207	BB	0.120	596.41	0.055
75	7.732		0.0000	0.000	11911.4	0.575	BB	0.034	5885.54	0.540
76	7.875		0.0000	0.000	910.5	0.044	BB	0.047	324.22	0.030
77	7.978	DCB	39.1958	48.823	211997.3	10.241	BB	0.037	95999.99	8.804
78	8.112		0.0000	0.000	2905.5	0.140	BB	0.116	415.80	0.038
79	8.386		0.0000	0.000	856.4	0.041	BB	0.056	255.78	0.023
80	8.602		0.0000	0.000	371.8	0.018	BB	0.032	193.09	0.018
81	8.701		0.0000	0.000	6366.3	0.308	BB	0.086	1233.94	0.113
82	8.964		0.0000	0.000	2351.1	0.114	BB	0.098	398.39	0.037

Total Area = 2070060.0, Total Amount = 80.281, Total Height = 1090430.0, Sample Units = ug/l

ASX22 0317

Sample Name: **AR1630924A AAAR163AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 04:34:56pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.09R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.580		0.0000	0.000	8119.6	0.359	BV	0.050	2697.00	0.284
2	1.637		0.0000	0.000	9528.4	0.421	VV	0.060	2663.75	0.281
3	1.700		0.0000	0.000	5378.6	0.238	VV	0.082	1097.32	0.116
4	2.013		0.0000	0.000	2060.1	0.091	VB	0.072	477.97	0.050
5	2.118		0.0000	0.000	317.6	0.014	BB	0.029	183.24	0.019
6	2.265		0.0000	0.000	167.0	0.007	BB	0.024	115.34	0.012
7	2.340	TCX	39.7045	48.356	411679.3	18.200	BB	0.039	175402.20	18.474
8	2.563		0.0000	0.000	4990.5	0.221	BB	0.032	2591.57	0.273
9	2.641		0.0000	0.000	8702.2	0.385	BB	0.041	3550.59	0.374
10	2.715		0.0000	0.000	40625.4	1.796	BB	0.043	15727.15	1.656
11	2.860		0.0000	0.000	5668.1	0.251	BB	0.036	2628.87	0.277
12	3.007		0.0000	0.000	102839.5	4.546	BB	0.049	35206.81	3.708
13	3.171		0.0000	0.000	57076.1	2.523	BB	0.040	23982.21	2.526
14	3.312		0.0000	0.000	15734.7	0.696	BB	0.046	5745.51	0.605
15	3.404		0.0000	0.000	73098.6	3.232	BB	0.034	36229.63	3.816
16	3.450		0.0000	0.000	12023.6	0.532	BB	0.028	7053.33	0.743
17	3.557		0.0000	0.000	26887.1	1.189	BB	0.030	14705.09	1.549
18	3.619		0.0000	0.000	5069.9	0.224	BB	0.043	1943.51	0.205
19	3.644		0.0000	0.000	25111.9	1.110	BB	0.040	10353.72	1.090
20	3.798		0.0000	0.000	63048.8	2.787	BB	0.052	20205.74	2.128
21	3.886		0.0000	0.000	343.7	0.015	BB	0.010	545.58	0.057
22	3.923		0.0000	0.000	15423.4	0.682	BB	0.029	8904.47	0.938
23	3.971		0.0000	0.000	3683.7	0.163	BB	0.027	2293.08	0.242
24	4.109		0.0000	0.000	64567.7	2.854	BB	0.054	19775.57	2.083
25	4.253		0.0000	0.000	42700.0	1.888	BB	0.053	13371.97	1.408
26	4.430		0.0000	0.000	514.6	0.023	BB	0.029	291.58	0.031
27	4.505		0.0000	0.000	1285.7	0.057	BB	0.028	774.29	0.082
28	4.542		0.0000	0.000	6048.9	0.267	BB	0.024	4244.93	0.447
29	4.598		0.0000	0.000	61651.4	2.726	BB	0.049	20837.31	2.195
30	4.694		0.0000	0.000	82311.3	3.639	BB	0.035	38779.26	4.084
31	4.822		0.0000	0.000	9712.8	0.429	BB	0.034	4752.78	0.501
32	4.874		0.0000	0.000	91145.9	4.029	BB	0.037	41273.27	4.347
33	4.985		0.0000	0.000	100475.7	4.442	BB	0.054	30856.01	3.250
34	5.101		0.0000	0.000	6643.4	0.294	BB	0.023	4846.36	0.510
35	5.144		0.0000	0.000	28872.3	1.276	BB	0.025	19491.53	2.053
36	5.177		0.0000	0.000	20631.3	0.912	BB	0.021	16155.08	1.701
37	5.218		0.0000	0.000	7022.8	0.310	BB	0.024	4928.75	0.519
38	5.337		0.0000	0.000	6446.7	0.285	BB	0.026	4194.41	0.442
39	5.391		0.0000	0.000	72014.8	3.184	BB	0.036	33592.43	3.538
40	5.474		0.0000	0.000	17682.9	0.782	BB	0.025	11707.88	1.233
41	5.510		0.0000	0.000	2566.0	0.113	BB	0.019	2196.19	0.231
42	5.561		0.0000	0.000	4016.7	0.178	BB	0.022	2980.70	0.314
43	5.629		0.0000	0.000	172093.4	7.608	BB	0.043	66748.00	7.030
44	5.801		0.0000	0.000	787.2	0.035	BB	0.022	586.87	0.062
45	5.869		0.0000	0.000	21340.6	0.943	BB	0.027	13251.69	1.396
46	5.924		0.0000	0.000	101458.5	4.485	BB	0.038	44624.39	4.700
47	6.088		0.0000	0.000	2167.2	0.096	BB	0.036	1002.15	0.106
48	6.174		0.0000	0.000	2270.7	0.100	BB	0.031	1238.35	0.130
49	6.240		0.0000	0.000	22490.2	0.994	BB	0.041	9188.90	0.968
50	6.385		0.0000	0.000	62432.7	2.760	BB	0.049	21330.28	2.247
51	6.623		0.0000	0.000	17746.5	0.785	BB	0.041	7261.80	0.765
52	6.801	DCB	42.4039	51.644	321640.3	14.219	BB	0.040	133662.50	14.078
53	7.617		0.0000	0.000	9600.9	0.424	BB	0.255	628.00	0.066
54	8.057		0.0000	0.000	2898.2	0.128	BB	0.133	362.71	0.038

ASX22. 318

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	8.320		0.0000	0.000	1202.5	0.053	BB	0.087	229.14	0.024

Total Area = 2262018.0, Total Amount = 82.108, Total Height = 949468.7, Sample Units = ug/l

ASX22 8319

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.09R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 4:25:47 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.09R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 4:25:47 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	192865	41.08517	TCX	2.34	175402	39.70447	TCX
7.978	96000	39.1958	DCB	6.801	133662	42.40391	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.09A
Area File: C:\CPWIN\DATA1\1P7307B.09A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 4:35:04 PM
File Reported On: 11/3/2009 at 4:35:18 PM

10X REPLOTS

ASX22 0321

C:\CPWIN\DATA1\1P7307.09R



C:\CPWIN\DATA1\1P7307B.09R



ASX22 8322

Sample Name: **AR1640924A AAAR164AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 04:47:55pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.10R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.612		0.0000	0.000	4145.3	0.106	BV	0.064	1083.31	0.050
2	1.663		0.0000	0.000	3584.7	0.092	VV	0.056	1070.80	0.050
3	1.793		0.0000	0.000	8343.5	0.213	VV	0.087	1591.32	0.074
4	1.929		0.0000	0.000	14567.5	0.372	VB	0.112	2171.52	0.101
5	2.034		0.0000	0.000	9694.6	0.248	BV	0.075	2147.91	0.100
6	2.073		0.0000	0.000	11780.5	0.301	VV	0.054	3638.82	0.169
7	2.172		0.0000	0.000	9631.1	0.246	VV	0.058	2764.03	0.129
8	2.207		0.0000	0.000	5215.3	0.133	VV	0.036	2432.67	0.113
9	2.271		0.0000	0.000	3541.2	0.090	VV	0.056	1051.23	0.049
10	2.369		0.0000	0.000	3658.0	0.093	VV	0.073	830.67	0.039
11	2.459		0.0000	0.000	2716.0	0.069	VV	0.071	635.28	0.030
12	2.520		0.0000	0.000	4351.6	0.111	VV	0.054	1345.37	0.063
13	2.622		0.0000	0.000	545.2	0.014	VB	0.035	262.24	0.012
14	2.771		0.0000	0.000	1522.0	0.039	BB	0.060	426.28	0.020
15	2.823		0.0000	0.000	21853.3	0.558	BB	0.169	2157.72	0.100
16	2.997	TCX	61.1758	53.050	470232.7	12.009	BB	0.028	275302.20	12.818
17	3.147		0.0000	0.000	367.4	0.009	BB	0.042	147.25	0.007
18	3.226		0.0000	0.000	14935.8	0.381	BB	0.027	9364.91	0.436
19	3.309		0.0000	0.000	22496.1	0.574	BB	0.033	11305.36	0.526
20	3.379		0.0000	0.000	77736.6	1.985	BB	0.027	48585.89	2.262
21	3.471		0.0000	0.000	591.5	0.015	BB	0.018	554.07	0.026
22	3.513		0.0000	0.000	11219.6	0.287	BB	0.026	7187.69	0.335
23	3.692		0.0000	0.000	152646.2	3.898	BB	0.039	64570.71	3.006
24	3.770		0.0000	0.000	8677.2	0.222	BB	0.026	5476.83	0.255
25	3.823		0.0000	0.000	3036.4	0.078	BB	0.017	3037.27	0.141
26	3.861		0.0000	0.000	90053.0	2.300	BB	0.026	58058.17	2.703
27	3.937		0.0000	0.000	384.7	0.010	BB	0.015	440.00	0.020
28	3.969		0.0000	0.000	1915.4	0.049	BB	0.021	1527.52	0.071
29	4.029		0.0000	0.000	46189.7	1.180	BB	0.039	19848.34	0.924
30	4.124		0.0000	0.000	248686.7	6.351	BB	0.033	125256.70	5.832
31	4.179		0.0000	0.000	80507.6	2.056	BB	0.024	56457.63	2.629
32	4.264		0.0000	0.000	85325.1	2.179	BB	0.026	54005.36	2.514
33	4.314		0.0000	0.000	2865.1	0.073	BB	0.017	2869.55	0.134
34	4.358		0.0000	0.000	39235.9	1.002	BB	0.020	32502.37	1.513
35	4.395		0.0000	0.000	65229.2	1.666	BB	0.033	33060.70	1.539
36	4.526		0.0000	0.000	111392.5	2.845	BB	0.036	51143.04	2.381
37	4.598		0.0000	0.000	10282.4	0.263	BB	0.017	10276.34	0.478
38	4.648		0.0000	0.000	64017.1	1.635	BB	0.028	38070.38	1.772
39	4.697		0.0000	0.000	13314.7	0.340	BB	0.021	10687.58	0.498
40	4.762		0.0000	0.000	1865.5	0.048	BB	0.021	1508.61	0.070
41	4.814		0.0000	0.000	1839.9	0.047	BB	0.021	1443.38	0.067
42	4.856		0.0000	0.000	65135.1	1.663	BB	0.028	38390.38	1.787
43	4.912		0.0000	0.000	6847.8	0.175	BB	0.016	7068.18	0.329
44	5.002		0.0000	0.000	5744.1	0.147	BB	0.023	4184.68	0.195
45	5.057		0.0000	0.000	68092.0	1.739	BB	0.031	36856.18	1.716
46	5.202		0.0000	0.000	294.4	0.008	BB	0.021	238.55	0.011
47	5.231		0.0000	0.000	1138.6	0.029	BB	0.022	867.96	0.040
48	5.309		0.0000	0.000	36211.2	0.925	BB	0.030	20285.16	0.944
49	5.374		0.0000	0.000	22558.3	0.576	BB	0.023	16074.89	0.748
50	5.428		0.0000	0.000	102812.3	2.626	BB	0.037	46714.74	2.175
51	5.520		0.0000	0.000	136979.2	3.498	BB	0.027	85901.27	3.999
52	5.635		0.0000	0.000	18126.9	0.463	BB	0.038	8040.23	0.374
53	5.702		0.0000	0.000	22307.5	0.570	BB	0.025	14788.04	0.689
54	5.766		0.0000	0.000	159151.9	4.064	BB	0.029	93032.28	4.331

ASX22 0323

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.811		0.0000	0.000	11472.1	0.293	BB	0.017	10945.63	0.510
56	5.875		0.0000	0.000	81367.9	2.078	BB	0.029	47390.61	2.206
57	5.968		0.0000	0.000	5899.4	0.151	BB	0.021	4722.59	0.220
58	6.011		0.0000	0.000	150839.7	3.852	BB	0.029	85242.67	3.969
59	6.083		0.0000	0.000	88452.6	2.259	BB	0.023	62878.09	2.928
60	6.134		0.0000	0.000	48198.8	1.231	BB	0.024	33294.70	1.550
61	6.279		0.0000	0.000	147816.0	3.775	BB	0.035	69901.27	3.254
62	6.352		0.0000	0.000	49457.1	1.263	BB	0.026	31559.34	1.469
63	6.400		0.0000	0.000	10359.9	0.265	BB	0.020	8723.50	0.406
64	6.446		0.0000	0.000	2689.1	0.069	BB	0.020	2225.45	0.104
65	6.511		0.0000	0.000	28763.1	0.735	BB	0.027	18044.55	0.840
66	6.579		0.0000	0.000	278632.2	7.116	BB	0.032	147022.20	6.845
67	6.832		0.0000	0.000	116222.4	2.968	BB	0.029	65766.97	3.062
68	6.892		0.0000	0.000	66741.8	1.704	BB	0.037	30248.44	1.408
69	7.057		0.0000	0.000	5664.6	0.145	BB	0.042	2232.38	0.104
70	7.183		0.0000	0.000	38671.6	0.988	BB	0.036	18022.24	0.839
71	7.344		0.0000	0.000	93.4	0.002	BB	0.021	72.73	0.003
72	7.417		0.0000	0.000	72149.9	1.843	BB	0.031	38974.09	1.815
73	7.488		0.0000	0.000	5787.7	0.148	BB	0.041	2331.15	0.109
74	7.656		0.0000	0.000	3945.8	0.101	BB	0.127	516.38	0.024
75	7.732		0.0000	0.000	25983.0	0.664	BB	0.034	12761.29	0.594
76	7.874		0.0000	0.000	628.9	0.016	BB	0.040	261.80	0.012
77	7.978	DCB	54.1418	46.950	292130.5	7.460	BB	0.036	134069.20	6.242
78	8.103		0.0000	0.000	1436.3	0.037	BB	0.100	238.69	0.011
79	8.316		0.0000	0.000	249.2	0.006	BB	0.030	138.26	0.006
80	8.604		0.0000	0.000	255.4	0.007	BB	0.034	125.83	0.006
81	8.702		0.0000	0.000	4951.6	0.126	BB	0.075	1106.10	0.051
82	8.953		0.0000	0.000	1350.8	0.034	BB	0.081	279.62	0.013

Total Area = 3915805.0, Total Amount = 115.318, Total Height = 2147835.0, Sample Units = ug/l

ASX22 8324

Sample Name: **AR1640924A AAAR164AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 04:47:55pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.10R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.586		0.0000	0.000	7724.4	0.175	BV	0.047	2740.43	0.144
2	1.636		0.0000	0.000	8338.7	0.189	VV	0.052	2657.55	0.140
3	1.699		0.0000	0.000	3488.0	0.079	VB	0.053	1103.10	0.058
4	1.963		0.0000	0.000	1726.4	0.039	BB	0.059	487.94	0.026
5	2.104		0.0000	0.000	8161.6	0.185	BB	1.413	96.27	0.005
6	2.339	TCX	56.6115	48.279	600796.6	13.589	BB	0.040	248224.90	13.049
7	2.561		0.0000	0.000	12978.1	0.294	BB	0.033	6634.72	0.349
8	2.641		0.0000	0.000	17531.0	0.397	BB	0.037	7851.47	0.413
9	2.713		0.0000	0.000	93185.5	2.108	BB	0.042	36793.03	1.934
10	2.858		0.0000	0.000	14310.3	0.324	BB	0.036	6589.50	0.346
11	3.005		0.0000	0.000	229345.5	5.188	BB	0.048	79376.80	4.173
12	3.169		0.0000	0.000	131010.6	2.963	BB	0.040	54990.23	2.891
13	3.310		0.0000	0.000	39902.7	0.903	BB	0.045	14804.39	0.778
14	3.402		0.0000	0.000	168447.1	3.810	BB	0.032	86556.01	4.550
15	3.447		0.0000	0.000	29033.9	0.657	BB	0.029	16534.38	0.869
16	3.556		0.0000	0.000	64627.0	1.462	BB	0.031	35091.33	1.845
17	3.617		0.0000	0.000	12947.9	0.293	BB	0.027	7858.78	0.413
18	3.642		0.0000	0.000	53397.4	1.208	BB	0.038	23407.72	1.231
19	3.796		0.0000	0.000	143727.8	3.251	BB	0.052	46167.95	2.427
20	3.885		0.0000	0.000	805.5	0.018	BB	0.011	1218.08	0.064
21	3.922		0.0000	0.000	36117.7	0.817	BB	0.028	21121.57	1.110
22	3.970		0.0000	0.000	8340.6	0.189	BB	0.024	5701.63	0.300
23	4.108		0.0000	0.000	132079.9	2.987	BB	0.049	44506.64	2.340
24	4.196		0.0000	0.000	1153.7	0.026	BB	0.018	1059.79	0.056
25	4.251		0.0000	0.000	83601.9	1.891	BB	0.045	30688.16	1.613
26	4.432		0.0000	0.000	1411.6	0.032	BB	0.029	802.43	0.042
27	4.504		0.0000	0.000	2768.4	0.063	BB	0.023	2009.21	0.106
28	4.541		0.0000	0.000	13565.0	0.307	BB	0.022	10266.54	0.540
29	4.597		0.0000	0.000	141152.7	3.193	BB	0.050	47327.84	2.488
30	4.692		0.0000	0.000	179316.9	4.056	BB	0.034	86726.13	4.559
31	4.820		0.0000	0.000	23845.6	0.539	BB	0.034	11618.42	0.611
32	4.872		0.0000	0.000	200457.6	4.534	BB	0.036	92927.59	4.885
33	4.984		0.0000	0.000	231885.5	5.245	BB	0.054	71323.06	3.749
34	5.099		0.0000	0.000	15631.0	0.354	BB	0.023	11102.41	0.584
35	5.142		0.0000	0.000	62082.2	1.404	BB	0.022	46560.25	2.448
36	5.175		0.0000	0.000	47134.4	1.066	BB	0.020	38916.29	2.046
37	5.217		0.0000	0.000	16858.9	0.381	BB	0.021	13330.41	0.701
38	5.336		0.0000	0.000	15357.8	0.347	BB	0.025	10097.13	0.531
39	5.390		0.0000	0.000	158491.3	3.585	BB	0.035	75481.45	3.968
40	5.473		0.0000	0.000	38984.6	0.882	BB	0.025	26146.80	1.375
41	5.509		0.0000	0.000	7629.0	0.173	BB	0.019	6578.13	0.346
42	5.560		0.0000	0.000	8804.9	0.199	BB	0.021	6848.56	0.360
43	5.627		0.0000	0.000	382692.3	8.656	BB	0.042	151765.60	7.978
44	5.868		0.0000	0.000	47033.4	1.064	BB	0.025	30991.65	1.629
45	5.923		0.0000	0.000	231718.7	5.241	BB	0.038	101367.30	5.329
46	6.088		0.0000	0.000	4328.3	0.098	BB	0.033	2186.61	0.115
47	6.173		0.0000	0.000	5893.6	0.133	BB	0.031	3203.23	0.168
48	6.239		0.0000	0.000	49048.1	1.109	BB	0.039	21018.28	1.105
49	6.384		0.0000	0.000	135411.9	3.063	BB	0.047	48531.14	2.551
50	6.622		0.0000	0.000	39174.4	0.886	BB	0.040	16128.77	0.848
51	6.800	DCB	60.6479	51.721	443094.2	10.022	BB	0.040	185401.30	9.746
52	7.513		0.0000	0.000	1561.6	0.035	BB	0.053	488.12	0.026
53	7.633		0.0000	0.000	8496.8	0.192	BB	0.280	505.34	0.027
54	8.323		0.0000	0.000	1446.4	0.033	BB	0.133	181.43	0.010

ASX22 0325

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	8.553		0.0000	0.000	3044.3	0.069	BB	0.324	156.79	0.008

Total Area = 4421102.0, Total Amount = 117.259, Total Height = 1902250.0, Sample Units = ug/l

ASX22. 0326

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.10R

C:\CPWIN\DATA1\1P7307.10R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 4:38:46 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.10R

C:\CPWIN\DATA1\1P7307B.10R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 4:38:46 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.997	275302	61.17582	TCX	2.339	248225	56.61152	TCX
7.978	134069	54.14183	DCB	6.8	185401	60.64785	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.10A
Area File: C:\CPWIN\DATA1\1P7307B.10A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 4:48:02 PM
File Reported On: 11/3/2009 at 4:48:17 PM

10X REPLOTS

ASX22 0328

C:\CPWIN\DATA1\1P7307.10R



C:\CPWIN\DATA1\1P7307B.10R



ASX22 8329

Sample Name: **AR1650924A AAAR165AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 05:00:55pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.11R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.666		0.0000	0.000	12461.8	0.189	BV	0.116	1786.29	0.048
2	1.868		0.0000	0.000	30564.8	0.463	VV	0.139	3661.48	0.099
3	1.929		0.0000	0.000	14259.8	0.216	VB	0.055	4285.63	0.116
4	2.034		0.0000	0.000	18564.5	0.281	BV	0.074	4186.00	0.113
5	2.074		0.0000	0.000	23304.9	0.353	VV	0.064	6072.16	0.164
6	2.175		0.0000	0.000	14005.5	0.212	VV	0.054	4340.78	0.117
7	2.208		0.0000	0.000	9342.6	0.141	VV	0.038	4092.68	0.110
8	2.315		0.0000	0.000	11849.0	0.179	VV	0.096	2050.23	0.055
9	2.372		0.0000	0.000	7017.3	0.106	VV	0.067	1741.74	0.047
10	2.454		0.0000	0.000	4883.6	0.074	VV	0.065	1246.85	0.034
11	2.520		0.0000	0.000	7177.3	0.109	VV	0.055	2181.76	0.059
12	2.620		0.0000	0.000	735.8	0.011	VB	0.040	309.32	0.008
13	2.776		0.0000	0.000	1885.0	0.029	BB	0.066	475.15	0.013
14	2.888		0.0000	0.000	15656.9	0.237	BB	0.058	4499.65	0.121
15	2.998	TCX	78.8829	53.548	575562.7	8.710	BB	0.028	346348.30	9.339
16	3.155		0.0000	0.000	410.2	0.006	BB	0.043	159.28	0.004
17	3.227		0.0000	0.000	28628.4	0.433	BB	0.027	17918.70	0.483
18	3.310		0.0000	0.000	38255.3	0.579	BB	0.031	20667.69	0.557
19	3.380		0.0000	0.000	138566.0	2.097	BB	0.026	87228.45	2.352
20	3.472		0.0000	0.000	759.0	0.011	BB	0.018	706.87	0.019
21	3.514		0.0000	0.000	21554.9	0.326	BB	0.026	13823.01	0.373
22	3.693		0.0000	0.000	272228.5	4.120	BB	0.039	115802.30	3.122
23	3.771		0.0000	0.000	16739.9	0.253	BB	0.027	10448.78	0.282
24	3.823		0.0000	0.000	5789.2	0.088	BB	0.016	5859.41	0.158
25	3.862		0.0000	0.000	161570.7	2.445	BB	0.026	103813.40	2.799
26	3.937		0.0000	0.000	437.9	0.007	BB	0.014	528.62	0.014
27	3.969		0.0000	0.000	4093.4	0.062	BB	0.022	3146.18	0.085
28	4.030		0.0000	0.000	85966.3	1.301	BB	0.038	37259.23	1.005
29	4.125		0.0000	0.000	448734.6	6.791	BB	0.033	223466.70	6.026
30	4.180		0.0000	0.000	146805.2	2.222	BB	0.024	102847.80	2.773
31	4.265		0.0000	0.000	155131.4	2.348	BB	0.026	98327.08	2.651
32	4.315		0.0000	0.000	5514.1	0.083	BB	0.016	5832.51	0.157
33	4.359		0.0000	0.000	69076.1	1.045	BB	0.020	58655.63	1.582
34	4.395		0.0000	0.000	118960.5	1.800	BB	0.032	61335.57	1.654
35	4.527		0.0000	0.000	201414.0	3.048	BB	0.036	93356.74	2.517
36	4.598		0.0000	0.000	19091.8	0.289	BB	0.016	19516.67	0.526
37	4.648		0.0000	0.000	118320.2	1.791	BB	0.028	70539.58	1.902
38	4.698		0.0000	0.000	25250.5	0.382	BB	0.021	19917.15	0.537
39	4.763		0.0000	0.000	3821.5	0.058	BB	0.020	3120.03	0.084
40	4.815		0.0000	0.000	3356.9	0.051	BB	0.022	2533.26	0.068
41	4.856		0.0000	0.000	119273.0	1.805	BB	0.028	70363.30	1.897
42	4.913		0.0000	0.000	12889.2	0.195	BB	0.017	12788.77	0.345
43	5.002		0.0000	0.000	10384.5	0.157	BB	0.022	7718.79	0.208
44	5.058		0.0000	0.000	120624.9	1.826	BB	0.030	66800.38	1.801
45	5.232		0.0000	0.000	4303.8	0.065	BB	0.033	2166.01	0.058
46	5.309		0.0000	0.000	67462.2	1.021	BB	0.030	37653.16	1.015
47	5.375		0.0000	0.000	42844.5	0.648	BB	0.023	30458.48	0.821
48	5.428		0.0000	0.000	125324.0	1.897	BB	0.027	76745.81	2.069
49	5.469		0.0000	0.000	6141.9	0.093	BB	0.013	7729.71	0.208
50	5.521		0.0000	0.000	245645.6	3.718	BB	0.027	154469.80	4.165
51	5.636		0.0000	0.000	34455.0	0.521	BB	0.037	15435.46	0.416
52	5.703		0.0000	0.000	41414.1	0.627	BB	0.025	27480.47	0.741
53	5.766		0.0000	0.000	288019.8	4.359	BB	0.028	168612.60	4.546
54	5.811		0.0000	0.000	21568.9	0.326	BB	0.018	20006.59	0.539

ASX22 0330

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.876		0.0000	0.000	151302.6	2.290	BB	0.028	88676.07	2.391
56	5.968		0.0000	0.000	11517.6	0.174	BB	0.021	9318.15	0.251
57	6.012		0.0000	0.000	280018.8	4.238	BB	0.029	160307.40	4.323
58	6.083		0.0000	0.000	160602.5	2.431	BB	0.023	114429.00	3.085
59	6.135		0.0000	0.000	90664.9	1.372	BB	0.024	62818.77	1.694
60	6.280		0.0000	0.000	277720.2	4.203	BB	0.035	131126.60	3.536
61	6.353		0.0000	0.000	89570.0	1.356	BB	0.026	58256.53	1.571
62	6.401		0.0000	0.000	21803.5	0.330	BB	0.021	17571.06	0.474
63	6.447		0.0000	0.000	4373.2	0.066	BB	0.019	3806.49	0.103
64	6.511		0.0000	0.000	54967.6	0.832	BB	0.026	34749.17	0.937
65	6.579		0.0000	0.000	497466.2	7.529	BB	0.030	273172.00	7.366
66	6.833		0.0000	0.000	220042.6	3.330	BB	0.029	124921.80	3.368
67	6.894		0.0000	0.000	123070.2	1.863	BB	0.036	56759.64	1.530
68	7.057		0.0000	0.000	7171.1	0.109	BB	0.030	3924.24	0.106
69	7.183		0.0000	0.000	70535.4	1.067	BB	0.035	33867.52	0.913
70	7.417		0.0000	0.000	133117.5	2.015	BB	0.031	72069.73	1.943
71	7.488		0.0000	0.000	7522.9	0.114	BB	0.034	3727.34	0.101
72	7.648		0.0000	0.000	3761.9	0.057	BB	0.102	613.12	0.017
73	7.732		0.0000	0.000	48118.7	0.728	BB	0.035	23136.27	0.624
74	7.868		0.0000	0.000	576.3	0.009	BB	0.044	220.10	0.006
75	7.978	DCB	68.4287	46.452	370189.2	5.602	BB	0.036	171104.40	4.614
76	8.090		0.0000	0.000	132.8	0.002	BB	0.025	89.80	0.002
77	8.146		0.0000	0.000	335.7	0.005	BB	0.026	211.18	0.006
78	8.396		0.0000	0.000	395.8	0.006	BB	0.077	85.84	0.002
79	8.609		0.0000	0.000	278.3	0.004	BB	0.031	149.38	0.004
80	8.695		0.0000	0.000	3468.9	0.052	BB	0.069	834.09	0.022
81	8.964		0.0000	0.000	877.6	0.013	BB	0.071	205.36	0.006

Total Area = 6607701.0, Total Amount = 147.312, Total Height = 3708669.0, Sample Units = ug/l

ASX22 0331

Sample Name: **AR1650924A AAAR165AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 05:00:55pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.11R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.585		0.0000	0.000	10462.9	0.138	BV	0.053	3291.35	0.100
2	1.636		0.0000	0.000	7614.0	0.101	VV	0.045	2814.27	0.085
3	1.701		0.0000	0.000	3994.7	0.053	VB	0.051	1313.73	0.040
4	1.964		0.0000	0.000	3590.0	0.047	BB	0.055	1095.24	0.033
5	2.340	TCX	73.4582	48.110	773372.1	10.210	BB	0.040	319994.70	9.691
6	2.563		0.0000	0.000	25473.9	0.336	BB	0.033	12892.60	0.390
7	2.643		0.0000	0.000	29928.0	0.395	BB	0.035	14297.92	0.433
8	2.714		0.0000	0.000	171821.6	2.268	BB	0.042	68793.28	2.083
9	2.860		0.0000	0.000	27387.3	0.362	BB	0.036	12626.36	0.382
10	3.006		0.0000	0.000	413961.0	5.465	BB	0.048	144314.20	4.371
11	3.170		0.0000	0.000	241397.5	3.187	BB	0.039	102030.00	3.090
12	3.312		0.0000	0.000	78896.7	1.042	BB	0.045	29329.14	0.888
13	3.403		0.0000	0.000	318772.1	4.209	BB	0.033	161059.80	4.878
14	3.448		0.0000	0.000	56440.9	0.745	BB	0.030	31855.62	0.965
15	3.557		0.0000	0.000	121151.3	1.599	BB	0.030	67683.09	2.050
16	3.617		0.0000	0.000	22805.6	0.301	BB	0.025	15447.67	0.468
17	3.643		0.0000	0.000	99361.1	1.312	BB	0.038	43315.34	1.312
18	3.797		0.0000	0.000	265441.7	3.504	BB	0.051	86417.91	2.617
19	3.886		0.0000	0.000	1769.8	0.023	BB	0.012	2462.07	0.075
20	3.923		0.0000	0.000	68105.9	0.899	BB	0.028	40283.93	1.220
21	3.971		0.0000	0.000	16265.1	0.215	BB	0.025	10778.15	0.326
22	4.108		0.0000	0.000	278981.6	3.683	BB	0.054	85440.84	2.588
23	4.252		0.0000	0.000	151770.4	2.004	BB	0.044	56977.18	1.726
24	4.432		0.0000	0.000	2995.2	0.040	BB	0.030	1640.35	0.050
25	4.505		0.0000	0.000	6014.0	0.079	BB	0.022	4601.96	0.139
26	4.542		0.0000	0.000	24906.6	0.329	BB	0.023	18069.51	0.547
27	4.597		0.0000	0.000	263045.3	3.473	BB	0.050	88476.58	2.680
28	4.693		0.0000	0.000	321162.8	4.240	BB	0.034	159759.40	4.839
29	4.820		0.0000	0.000	45029.4	0.594	BB	0.034	22121.49	0.670
30	4.873		0.0000	0.000	364674.8	4.815	BB	0.035	174054.80	5.271
31	4.986		0.0000	0.000	433263.1	5.720	BB	0.054	133147.20	4.033
32	5.100		0.0000	0.000	29758.5	0.393	BB	0.022	22408.47	0.679
33	5.142		0.0000	0.000	126570.2	1.671	BB	0.025	85424.25	2.587
34	5.176		0.0000	0.000	81499.6	1.076	BB	0.020	69187.39	2.095
35	5.218		0.0000	0.000	34921.1	0.461	BB	0.022	26332.21	0.798
36	5.337		0.0000	0.000	29564.2	0.390	BB	0.025	20046.38	0.607
37	5.391		0.0000	0.000	290880.3	3.840	BB	0.034	140618.80	4.259
38	5.474		0.0000	0.000	70160.4	0.926	BB	0.024	48874.36	1.480
39	5.510		0.0000	0.000	15039.2	0.199	BB	0.020	12574.18	0.381
40	5.561		0.0000	0.000	16743.0	0.221	BB	0.022	12932.49	0.392
41	5.628		0.0000	0.000	708979.3	9.360	BB	0.041	287997.40	8.722
42	5.868		0.0000	0.000	84628.9	1.117	BB	0.024	58164.96	1.762
43	5.923		0.0000	0.000	430701.3	5.686	BB	0.038	190833.30	5.780
44	6.089		0.0000	0.000	8120.9	0.107	BB	0.033	4128.87	0.125
45	6.175		0.0000	0.000	11676.8	0.154	BB	0.031	6354.75	0.192
46	6.240		0.0000	0.000	91671.6	1.210	BB	0.038	40167.85	1.217
47	6.384		0.0000	0.000	249158.5	3.289	BB	0.045	91364.47	2.767
48	6.622		0.0000	0.000	71585.2	0.945	BB	0.040	29717.45	0.900
49	6.801	DCB	79.2286	51.890	558803.4	7.378	BB	0.039	236992.30	7.178
50	7.233		0.0000	0.000	194.9	0.003	BB	0.031	104.31	0.003
51	7.483		0.0000	0.000	1648.0	0.022	BB	0.094	291.19	0.009
52	7.638		0.0000	0.000	10774.0	0.142	BB	0.332	540.98	0.016
53	8.310		0.0000	0.000	663.5	0.009	BB	0.068	161.71	0.005
54	8.550		0.0000	0.000	334.4	0.004	BB	0.041	135.06	0.004

ASX22 0332

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	9.020		0.0000	0.000	486.1	0.006	BB	0.087	93.61	0.003

Total Area = 7574420.0, Total Amount = 152.687, Total Height = 3301832.0, Sample Units = ug/l

ASX22 8333

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.11R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 4:51:46 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.11R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 4:51:46 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.998	346348	78.88287	TCX	2.34	319995	73.45821	TCX
7.978	171104	68.42866	DCB	6.801	236992	79.2286	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.11A
Area File: C:\CPWIN\DATA1\1P7307B.11A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 5:01:04 PM
File Reported On: 11/3/2009 at 5:01:18 PM

10X REPLOTS

ASX22 8335

C:\CPWIN\DATA1\1P7307.11R



C:\CPWIN\DATA1\1P7307B.11R



ASX22 0336

Sample Name: **AR4810924B AAAR481AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 05:13:55pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.12R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.581		0.0000	0.000	2971.3	0.695	BV	0.048	1032.01	0.587
2	1.685		0.0000	0.000	8233.6	1.925	VV	0.097	1416.35	0.805
3	1.788		0.0000	0.000	10525.4	2.461	VV	0.085	2066.26	1.174
4	1.864		0.0000	0.000	9899.5	2.315	VV	0.065	2547.58	1.448
5	1.930		0.0000	0.000	10112.3	2.365	VB	0.055	3050.86	1.734
6	2.031		0.0000	0.000	14401.5	3.368	BV	0.080	3013.46	1.713
7	2.175		0.0000	0.000	32890.4	7.691	VV	0.180	3042.38	1.729
8	2.374		0.0000	0.000	6346.4	1.484	VV	0.120	879.76	0.500
9	2.441		0.0000	0.000	1906.9	0.446	VV	0.052	614.42	0.349
10	2.528		0.0000	0.000	1684.8	0.394	VB	0.038	735.08	0.418
11	2.773		0.0000	0.000	2185.2	0.511	BB	0.050	728.17	0.414
12	2.869		0.0000	0.000	4015.7	0.939	BB	0.064	1044.94	0.594
13	2.999	TCX	1.9851	26.095	51110.4	11.952	BB	0.031	27567.22	15.669
14	3.151		0.0000	0.000	99.1	0.023	BB	0.032	51.47	0.029
15	3.212		0.0000	0.000	323.6	0.076	BB	0.033	163.82	0.093
16	3.288		0.0000	0.000	1064.6	0.249	BB	0.039	452.52	0.257
17	3.381		0.0000	0.000	752.5	0.176	BB	0.023	533.71	0.303
18	3.448		0.0000	0.000	574.0	0.134	BB	0.046	209.82	0.119
19	3.515		0.0000	0.000	149.1	0.035	BB	0.021	117.06	0.067
20	3.600		0.0000	0.000	233.4	0.055	BB	0.050	77.21	0.044
21	3.694		0.0000	0.000	7639.2	1.786	BB	0.034	3717.91	2.113
22	3.771		0.0000	0.000	408.8	0.096	BB	0.027	251.76	0.143
23	3.863		0.0000	0.000	4732.6	1.107	BB	0.028	2787.85	1.585
24	3.973		0.0000	0.000	161.8	0.038	BB	0.030	89.52	0.051
25	4.028		0.0000	0.000	1649.0	0.386	BB	0.035	784.58	0.446
26	4.126		0.0000	0.000	18213.8	4.259	BB	0.036	8520.01	4.843
27	4.181		0.0000	0.000	3842.9	0.899	BB	0.024	2658.92	1.511
28	4.263		0.0000	0.000	6116.4	1.430	BB	0.028	3685.47	2.095
29	4.316		0.0000	0.000	442.5	0.103	BB	0.018	409.61	0.233
30	4.360		0.0000	0.000	7401.7	1.731	BB	0.020	6196.58	3.522
31	4.396		0.0000	0.000	9828.0	2.298	BB	0.032	5165.09	2.936
32	4.527		0.0000	0.000	18284.7	4.276	BB	0.033	9118.68	5.183
33	4.599		0.0000	0.000	1804.6	0.422	BB	0.018	1716.92	0.976
34	4.647		0.0000	0.000	9199.9	2.151	BB	0.026	5956.03	3.385
35	4.698		0.0000	0.000	2300.0	0.538	BB	0.022	1704.58	0.969
36	4.760		0.0000	0.000	66.8	0.016	BB	0.015	74.69	0.042
37	4.815		0.0000	0.000	1016.6	0.238	BB	0.028	608.18	0.346
38	4.858		0.0000	0.000	4330.3	1.013	BB	0.017	4137.54	2.352
39	4.887		0.0000	0.000	3260.4	0.762	BB	0.015	3565.32	2.026
40	4.914		0.0000	0.000	5769.1	1.349	BB	0.019	5078.09	2.886
41	5.000		0.0000	0.000	1213.0	0.284	BB	0.025	796.71	0.453
42	5.053		0.0000	0.000	9676.4	2.263	BB	0.028	5675.51	3.226
43	5.101		0.0000	0.000	3189.4	0.746	BB	0.020	2618.62	1.488
44	5.233		0.0000	0.000	4399.6	1.029	BB	0.036	2009.72	1.142
45	5.299		0.0000	0.000	3104.7	0.726	BB	0.023	2237.00	1.271
46	5.336		0.0000	0.000	1018.3	0.238	BB	0.017	1013.28	0.576
47	5.379		0.0000	0.000	11994.4	2.805	BB	0.028	7248.46	4.120
48	5.478		0.0000	0.000	2378.3	0.556	BB	0.025	1603.77	0.912
49	5.587		0.0000	0.000	345.4	0.081	BB	0.019	304.89	0.173
50	5.639		0.0000	0.000	5484.2	1.282	BB	0.027	3336.48	1.896
51	5.735		0.0000	0.000	286.4	0.067	BB	0.029	166.76	0.095
52	5.774		0.0000	0.000	900.2	0.211	BB	0.043	348.59	0.198
53	5.878		0.0000	0.000	5441.5	1.273	BB	0.033	2718.92	1.545
54	5.967		0.0000	0.000	898.7	0.210	BB	0.031	481.11	0.273

ASX22 8337

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.038		0.0000	0.000	470.9	0.110	BB	0.043	184.55	0.105
56	6.154		0.0000	0.000	619.4	0.145	BB	0.055	187.31	0.106
57	6.347		0.0000	0.000	1147.7	0.268	BB	0.070	273.24	0.155
58	6.394		0.0000	0.000	24901.8	5.823	BB	0.088	4714.38	2.680
59	6.728		0.0000	0.000	6365.8	1.489	BB	0.069	1528.33	0.869
60	6.907		0.0000	0.000	7713.0	1.804	BB	0.236	545.44	0.310
61	6.971		0.0000	0.000	210.3	0.049	BB	0.040	86.85	0.049
62	7.108		0.0000	0.000	6890.5	1.611	BB	0.070	1640.58	0.932
63	7.213		0.0000	0.000	4910.8	1.148	BB	0.076	1072.77	0.610
64	7.290		0.0000	0.000	506.3	0.118	BB	0.037	230.43	0.131
65	7.351		0.0000	0.000	1475.5	0.345	BB	0.086	287.12	0.163
66	7.524		0.0000	0.000	9306.7	2.176	BB	0.088	1770.71	1.006
67	7.655		0.0000	0.000	4868.3	1.138	BB	0.121	670.48	0.381
68	7.732		0.0000	0.000	504.4	0.118	BB	0.045	186.99	0.106
69	7.878		0.0000	0.000	1048.9	0.245	BB	0.055	315.44	0.179
70	7.979	DCB	5.6219	73.905	33358.6	7.801	BB	0.039	14211.97	8.078
71	8.127		0.0000	0.000	1628.3	0.381	BB	0.072	374.57	0.213
72	8.209		0.0000	0.000	351.6	0.082	BB	0.066	88.44	0.050
73	8.383		0.0000	0.000	812.0	0.190	BB	0.053	254.75	0.145
74	8.606		0.0000	0.000	415.6	0.097	BB	0.040	174.27	0.099
75	8.691		0.0000	0.000	3241.3	0.758	BB	0.061	884.02	0.502
76	8.962		0.0000	0.000	595.0	0.139	BB	0.079	126.22	0.072

Total Area = 427622.1, Total Amount = 7.607, Total Height = 175940.1, Sample Units = ug/l

65X22. 8338

Sample Name: **AR4810924B AAAR481AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 05:13:55pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.12R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.637		0.0000	0.000	19476.1	5.812	BB	0.105	3104.64	2.696
2	2.014		0.0000	0.000	778.5	0.232	BB	0.044	293.57	0.255
3	2.107		0.0000	0.000	183.2	0.055	BB	0.023	131.38	0.114
4	2.183		0.0000	0.000	856.5	0.256	BB	0.061	235.29	0.204
5	2.340	TCX	4.8795	68.374	50624.0	15.107	BB	0.040	21161.69	18.376
6	2.623		0.0000	0.000	524.8	0.157	BB	0.037	234.07	0.203
7	2.714		0.0000	0.000	2312.1	0.690	BB	0.099	388.69	0.338
8	3.006		0.0000	0.000	11334.9	3.382	BB	0.047	4045.79	3.513
9	3.168		0.0000	0.000	5604.3	1.672	BB	0.040	2307.71	2.004
10	3.307		0.0000	0.000	537.4	0.160	BB	0.028	315.07	0.274
11	3.350		0.0000	0.000	149.3	0.045	BB	0.023	110.47	0.096
12	3.404		0.0000	0.000	27162.0	8.105	BB	0.068	6624.27	5.752
13	3.556		0.0000	0.000	4482.2	1.338	BB	0.032	2367.44	2.056
14	3.621		0.0000	0.000	2307.8	0.689	BB	0.045	860.58	0.747
15	3.643		0.0000	0.000	7832.3	2.337	BB	0.039	3319.80	2.883
16	3.797		0.0000	0.000	25355.1	7.566	BB	0.055	7703.42	6.689
17	3.921		0.0000	0.000	5725.3	1.708	BB	0.029	3346.69	2.906
18	3.971		0.0000	0.000	1142.4	0.341	BB	0.026	742.60	0.645
19	4.116		0.0000	0.000	6303.4	1.881	BB	0.034	3066.30	2.663
20	4.143		0.0000	0.000	2590.1	0.773	BB	0.030	1463.11	1.271
21	4.252		0.0000	0.000	1431.6	0.427	BB	0.026	918.48	0.798
22	4.295		0.0000	0.000	19418.4	5.795	BB	0.050	6416.21	5.572
23	4.431		0.0000	0.000	2432.4	0.726	BB	0.032	1285.36	1.116
24	4.506		0.0000	0.000	2358.2	0.704	BB	0.026	1503.45	1.306
25	4.537		0.0000	0.000	394.9	0.118	BB	0.016	405.83	0.352
26	4.582		0.0000	0.000	6875.4	2.052	BB	0.033	3430.93	2.979
27	4.691		0.0000	0.000	3701.5	1.105	BB	0.038	1640.08	1.424
28	4.794		0.0000	0.000	1594.1	0.476	BB	0.026	1033.97	0.898
29	4.830		0.0000	0.000	2820.8	0.842	BB	0.026	1787.96	1.553
30	4.968		0.0000	0.000	603.1	0.180	BB	0.041	247.86	0.215
31	5.049		0.0000	0.000	4372.7	1.305	BB	0.041	1786.33	1.551
32	5.143		0.0000	0.000	3230.5	0.964	BB	0.071	763.53	0.663
33	5.303		0.0000	0.000	763.5	0.228	BB	0.064	198.82	0.173
34	5.381		0.0000	0.000	1153.7	0.344	BB	0.063	307.60	0.267
35	5.743		0.0000	0.000	1972.3	0.589	BB	0.239	137.46	0.119
36	5.804		0.0000	0.000	37212.7	11.105	BB	0.061	10099.32	8.770
37	6.017		0.0000	0.000	8394.7	2.505	BB	0.359	390.11	0.339
38	6.395		0.0000	0.000	146.8	0.044	BB	0.079	31.11	0.027
39	6.457		0.0000	0.000	161.4	0.048	BB	0.062	43.16	0.037
40	6.520		0.0000	0.000	1325.1	0.395	BB	0.048	457.02	0.397
41	6.653		0.0000	0.000	905.9	0.270	BB	0.050	303.38	0.263
42	6.803	DCB	2.2570	31.626	47110.1	14.058	BB	0.042	18907.61	16.419
43	7.157		0.0000	0.000	633.9	0.189	BB	0.093	113.35	0.098
44	7.487		0.0000	0.000	1562.0	0.466	BB	0.087	299.29	0.260
45	7.650		0.0000	0.000	6463.0	1.929	BB	0.213	505.35	0.439
46	8.063		0.0000	0.000	708.7	0.211	BB	0.083	143.11	0.124
47	8.274		0.0000	0.000	38.2	0.011	BB	0.019	33.58	0.029
48	8.717		0.0000	0.000	2043.9	0.610	BB	0.231	147.27	0.128

Total Area = 335111.6, Total Amount = 7.136, Total Height = 115160.1, Sample Units = ug/l

ASX22 #339

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.12R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 5:04:46 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.12R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 5:04:46 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	27567	1.98509	TCX	2.34	21162	4.87952	TCX
7.979	14212	5.62195	DCB	6.803	18908	2.25697	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.12A
Area File: C:\CPWIN\DATA1\1P7307B.12A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 5:14:02 PM
File Reported On: 11/3/2009 at 5:14:15 PM

10X REPLOTS

ASX22 0341

C:\CPWIN\DATA1\1P7307.12R



C:\CPWIN\DATA1\1P7307B.12R



ASX22 8342

Sample Name: **AR4820924B AAAR482AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 05:26:51pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.13R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.582		0.0000	0.000	3497.4	0.428	BV	0.051	1134.71	0.290
2	1.796		0.0000	0.000	22993.0	2.817	VV	0.147	2599.38	0.664
3	1.862		0.0000	0.000	11278.9	1.382	VV	0.060	3145.67	0.803
4	1.929		0.0000	0.000	13110.1	1.606	VB	0.056	3872.89	0.989
5	2.036		0.0000	0.000	17519.5	2.146	BV	0.080	3632.98	0.928
6	2.074		0.0000	0.000	17668.3	2.164	VV	0.074	3971.19	1.014
7	2.174		0.0000	0.000	31511.9	3.860	VV	0.140	3761.52	0.961
8	2.371		0.0000	0.000	6268.7	0.768	VV	0.068	1542.76	0.394
9	2.446		0.0000	0.000	5658.4	0.693	VV	0.079	1193.45	0.305
10	2.525		0.0000	0.000	5039.1	0.617	VV	0.073	1146.38	0.293
11	2.623		0.0000	0.000	502.5	0.062	VB	0.038	217.88	0.056
12	2.694		0.0000	0.000	262.1	0.032	BB	0.042	103.97	0.027
13	2.784		0.0000	0.000	1251.4	0.153	BB	0.053	391.28	0.100
14	2.883		0.0000	0.000	9124.4	1.118	BB	0.054	2793.56	0.713
15	2.998	TCX	23.0330	50.994	193889.2	23.752	BB	0.028	113507.90	28.988
16	3.156		0.0000	0.000	130.8	0.016	BB	0.047	46.27	0.012
17	3.217		0.0000	0.000	205.8	0.025	BB	0.030	115.76	0.030
18	3.287		0.0000	0.000	2865.6	0.351	BB	0.028	1691.05	0.432
19	3.380		0.0000	0.000	1587.9	0.195	BB	0.025	1058.49	0.270
20	3.471		0.0000	0.000	452.4	0.055	BB	0.028	267.04	0.068
21	3.514		0.0000	0.000	407.6	0.050	BB	0.023	299.22	0.076
22	3.578		0.0000	0.000	301.2	0.037	BB	0.055	91.57	0.023
23	3.693		0.0000	0.000	15095.1	1.849	BB	0.034	7292.73	1.862
24	3.770		0.0000	0.000	607.6	0.074	BB	0.024	425.05	0.109
25	3.862		0.0000	0.000	9084.9	1.113	BB	0.028	5479.82	1.399
26	3.939		0.0000	0.000	162.0	0.020	BB	0.017	154.34	0.039
27	3.970		0.0000	0.000	195.1	0.024	BB	0.022	151.04	0.039
28	4.028		0.0000	0.000	2731.7	0.335	BB	0.034	1350.38	0.345
29	4.125		0.0000	0.000	34628.8	4.242	BB	0.036	16240.37	4.147
30	4.180		0.0000	0.000	7396.3	0.906	BB	0.024	5200.81	1.328
31	4.263		0.0000	0.000	11523.0	1.412	BB	0.028	6894.41	1.761
32	4.316		0.0000	0.000	848.3	0.104	BB	0.018	794.02	0.203
33	4.359		0.0000	0.000	14003.3	1.715	BB	0.020	11642.45	2.973
34	4.395		0.0000	0.000	18099.6	2.217	BB	0.031	9688.85	2.474
35	4.527		0.0000	0.000	35099.1	4.300	BB	0.034	17205.63	4.394
36	4.599		0.0000	0.000	3531.8	0.433	BB	0.018	3272.66	0.836
37	4.647		0.0000	0.000	17366.1	2.127	BB	0.025	11408.50	2.913
38	4.698		0.0000	0.000	4414.5	0.541	BB	0.022	3302.02	0.843
39	4.762		0.0000	0.000	226.9	0.028	BB	0.018	211.22	0.054
40	4.814		0.0000	0.000	1291.9	0.158	BB	0.025	877.14	0.224
41	4.858		0.0000	0.000	8425.1	1.032	BB	0.018	7663.56	1.957
42	4.887		0.0000	0.000	6266.2	0.768	BB	0.016	6693.46	1.709
43	4.914		0.0000	0.000	10075.9	1.234	BB	0.018	9121.99	2.330
44	5.001		0.0000	0.000	1758.9	0.215	BB	0.023	1286.21	0.328
45	5.052		0.0000	0.000	20027.9	2.453	BB	0.030	11186.07	2.857
46	5.098		0.0000	0.000	2910.7	0.357	BB	0.020	2474.05	0.632
47	5.232		0.0000	0.000	7381.5	0.904	BB	0.034	3583.01	0.915
48	5.299		0.0000	0.000	6150.1	0.753	BB	0.023	4406.35	1.125
49	5.335		0.0000	0.000	1903.3	0.233	BB	0.016	1924.85	0.492
50	5.375		0.0000	0.000	12603.7	1.544	BB	0.024	8655.25	2.210
51	5.477		0.0000	0.000	4669.0	0.572	BB	0.025	3115.63	0.796
52	5.519		0.0000	0.000	522.9	0.064	BB	0.029	302.35	0.077
53	5.588		0.0000	0.000	701.2	0.086	BB	0.019	603.28	0.154
54	5.639		0.0000	0.000	10327.7	1.265	BB	0.027	6341.37	1.619

ASX22 0343

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.734		0.0000	0.000	515.9	0.063	BB	0.028	304.45	0.078
56	5.775		0.0000	0.000	1174.0	0.144	BB	0.030	654.27	0.167
57	5.878		0.0000	0.000	10136.9	1.242	BB	0.032	5336.93	1.363
58	5.967		0.0000	0.000	467.4	0.057	BB	0.025	305.92	0.078
59	6.014		0.0000	0.000	1296.0	0.159	BB	0.033	661.83	0.169
60	6.153		0.0000	0.000	303.6	0.037	BB	0.039	131.04	0.033
61	6.353		0.0000	0.000	6945.3	0.851	BB	0.092	1253.66	0.320
62	6.452		0.0000	0.000	6331.2	0.776	BB	0.094	1122.34	0.287
63	6.730		0.0000	0.000	6111.7	0.749	BB	0.067	1513.38	0.386
64	6.903		0.0000	0.000	7722.4	0.946	BB	0.193	666.50	0.170
65	6.966		0.0000	0.000	141.0	0.017	BB	0.031	75.19	0.019
66	7.110		0.0000	0.000	7445.4	0.912	BB	0.072	1714.75	0.438
67	7.215		0.0000	0.000	5077.5	0.622	BB	0.079	1074.85	0.274
68	7.291		0.0000	0.000	627.4	0.077	BB	0.039	264.79	0.068
69	7.524		0.0000	0.000	8320.6	1.019	BB	0.086	1616.25	0.413
70	7.654		0.0000	0.000	4467.0	0.547	BB	0.112	666.50	0.170
71	7.734		0.0000	0.000	1419.8	0.174	BB	0.037	639.18	0.163
72	7.874		0.0000	0.000	1022.5	0.125	BB	0.055	307.18	0.078
73	7.979	DCB	22.1351	49.006	122601.3	15.019	BB	0.037	55956.20	14.290
74	8.120		0.0000	0.000	1499.0	0.184	BB	0.070	357.73	0.091
75	8.385		0.0000	0.000	3574.9	0.438	BB	0.145	410.94	0.105
76	8.605		0.0000	0.000	342.2	0.042	BB	0.036	158.30	0.040
77	8.690		0.0000	0.000	2574.8	0.315	BB	0.060	711.67	0.182
78	8.963		0.0000	0.000	628.4	0.077	BB	0.076	137.45	0.035

Total Area = 816302.4, Total Amount = 45.168, Total Height = 391575.1, Sample Units = ug/l

ASX22 0344

Sample Name: **AR4820924B AAAR482AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 05:26:51pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.13R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.590		0.0000	0.000	3818.3	0.496	BV	0.036	1787.43	0.604
2	1.635		0.0000	0.000	13282.5	1.726	VB	0.082	2711.90	0.917
3	2.017		0.0000	0.000	980.1	0.127	BB	0.048	340.00	0.115
4	2.109		0.0000	0.000	3430.4	0.446	BB	0.360	159.02	0.054
5	2.340	TCX	22.2285	49.863	230319.8	29.930	BB	0.040	96401.52	32.601
6	2.621		0.0000	0.000	2473.5	0.321	BB	0.037	1126.47	0.381
7	2.716		0.0000	0.000	2105.5	0.274	BB	0.049	712.56	0.241
8	2.857		0.0000	0.000	744.3	0.097	BB	0.038	323.36	0.109
9	3.005		0.0000	0.000	22008.7	2.860	BB	0.047	7803.00	2.639
10	3.170		0.0000	0.000	10857.6	1.411	BB	0.039	4584.63	1.550
11	3.304		0.0000	0.000	2065.1	0.268	BB	0.043	802.82	0.272
12	3.404		0.0000	0.000	55623.3	7.228	BB	0.070	13318.64	4.504
13	3.555		0.0000	0.000	8463.1	1.100	BB	0.031	4531.71	1.533
14	3.620		0.0000	0.000	4468.6	0.581	BB	0.047	1582.20	0.535
15	3.643		0.0000	0.000	14689.6	1.909	BB	0.039	6285.11	2.126
16	3.797		0.0000	0.000	48501.9	6.303	BB	0.054	14921.05	5.046
17	3.922		0.0000	0.000	11464.8	1.490	BB	0.029	6662.06	2.253
18	3.972		0.0000	0.000	2424.3	0.315	BB	0.023	1731.70	0.586
19	4.115		0.0000	0.000	10892.2	1.415	BB	0.035	5245.34	1.774
20	4.142		0.0000	0.000	6080.2	0.790	BB	0.029	3516.14	1.189
21	4.252		0.0000	0.000	2858.4	0.371	BB	0.025	1912.64	0.647
22	4.293		0.0000	0.000	33283.1	4.325	BB	0.049	11260.76	3.808
23	4.432		0.0000	0.000	5051.0	0.656	BB	0.032	2657.00	0.899
24	4.505		0.0000	0.000	4521.6	0.588	BB	0.024	3103.81	1.050
25	4.537		0.0000	0.000	737.3	0.096	BB	0.016	792.55	0.268
26	4.582		0.0000	0.000	14387.6	1.870	BB	0.036	6585.21	2.227
27	4.692		0.0000	0.000	7379.9	0.959	BB	0.038	3256.43	1.101
28	4.795		0.0000	0.000	10571.7	1.374	BB	0.042	4182.05	1.414
29	4.886		0.0000	0.000	1152.4	0.150	BB	0.041	466.07	0.158
30	4.970		0.0000	0.000	1336.1	0.174	BB	0.040	558.74	0.189
31	5.049		0.0000	0.000	8677.7	1.128	BB	0.041	3520.19	1.190
32	5.143		0.0000	0.000	4737.0	0.616	BB	0.060	1325.91	0.448
33	5.384		0.0000	0.000	2983.8	0.388	BB	0.067	741.02	0.251
34	5.563		0.0000	0.000	924.3	0.120	BB	0.038	409.56	0.139
35	5.716		0.0000	0.000	1978.4	0.257	BB	0.164	201.30	0.068
36	5.805		0.0000	0.000	2740.5	0.356	BB	0.042	1080.68	0.365
37	5.926		0.0000	0.000	1072.8	0.139	BB	0.035	512.44	0.173
38	6.023		0.0000	0.000	3407.2	0.443	BB	0.113	504.35	0.171
39	6.247		0.0000	0.000	1942.1	0.252	BB	0.094	342.82	0.116
40	6.491		0.0000	0.000	799.2	0.104	BB	0.399	33.38	0.011
41	6.526		0.0000	0.000	527.7	0.069	BB	0.033	266.38	0.090
42	6.587		0.0000	0.000	58.9	0.008	BB	0.020	48.98	0.017
43	6.624		0.0000	0.000	1935.6	0.252	BB	0.047	682.84	0.231
44	6.802	DCB	22.3503	50.137	187243.8	24.332	BB	0.041	75482.79	25.527
45	7.160		0.0000	0.000	905.3	0.118	BB	0.103	146.04	0.049
46	7.377		0.0000	0.000	218.8	0.028	BB	0.051	71.96	0.024
47	7.634		0.0000	0.000	10835.7	1.408	BB	0.337	535.66	0.181
48	8.220		0.0000	0.000	214.4	0.028	BB	0.059	60.52	0.020
49	8.323		0.0000	0.000	1128.3	0.147	BB	0.104	180.34	0.061
50	8.693		0.0000	0.000	682.7	0.089	BB	0.087	130.08	0.044
51	8.884		0.0000	0.000	535.2	0.070	BB	0.089	100.36	0.034

ASX22 8345

Total Area = 769522.1, Total Amount = 44.579, Total Height = 295699.5, Sample Units = ug/l

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.13R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 5:17:42 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.13R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 5:17:42 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.998	113508	23.033	TCX	2.34	96402	22.22853	TCX
7.979	55956	22.13505	DCB	6.802	75483	22.35027	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.13A
Area File: C:\CPWIN\DATA1\1P7307B.13A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 5:26:58 PM
File Reported On: 11/3/2009 at 5:27:11 PM

10X REPLOTS

ASX22 0347

C:\CPWIN\DATA1\1P7307.13R



C:\CPWIN\DATA1\1P7307B.13R



ASX22 0348

Sample Name: **AR4830924B AAAR483AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 05:39:51pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.14R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.580		0.0000	0.000	2173.8	0.170	BV	0.039	928.53	0.139
2	1.664		0.0000	0.000	6669.7	0.521	VV	0.095	1174.68	0.176
3	1.788		0.0000	0.000	10500.9	0.820	VV	0.090	1942.78	0.291
4	1.864		0.0000	0.000	9142.9	0.714	VV	0.065	2353.64	0.352
5	1.930		0.0000	0.000	9238.7	0.722	VB	0.054	2846.39	0.426
6	2.031		0.0000	0.000	12921.1	1.009	BV	0.078	2768.34	0.415
7	2.073		0.0000	0.000	12663.8	0.989	VV	0.058	3630.45	0.544
8	2.174		0.0000	0.000	10407.5	0.813	VV	0.063	2746.14	0.411
9	2.207		0.0000	0.000	6261.9	0.489	VV	0.040	2605.74	0.390
10	2.272		0.0000	0.000	4830.5	0.377	VV	0.060	1342.97	0.201
11	2.369		0.0000	0.000	4035.5	0.315	VV	0.078	862.12	0.129
12	2.442		0.0000	0.000	2009.0	0.157	VV	0.059	570.06	0.085
13	2.527		0.0000	0.000	1750.6	0.137	VB	0.046	633.34	0.095
14	2.771		0.0000	0.000	1801.0	0.141	BB	0.051	587.45	0.088
15	2.875		0.0000	0.000	10740.3	0.839	BB	0.059	3049.47	0.457
16	2.998	TCX	45.4210	53.603	344902.8	26.935	BB	0.028	204920.50	30.686
17	3.222		0.0000	0.000	363.1	0.028	BB	0.027	228.11	0.034
18	3.287		0.0000	0.000	5655.2	0.442	BB	0.028	3388.99	0.507
19	3.380		0.0000	0.000	3262.9	0.255	BB	0.027	2031.37	0.304
20	3.472		0.0000	0.000	609.1	0.048	BB	0.021	483.36	0.072
21	3.513		0.0000	0.000	781.2	0.061	BB	0.021	611.94	0.092
22	3.612		0.0000	0.000	428.7	0.033	BB	0.066	109.08	0.016
23	3.693		0.0000	0.000	27158.6	2.121	BB	0.035	12943.93	1.938
24	3.770		0.0000	0.000	1028.3	0.080	BB	0.025	698.18	0.105
25	3.824		0.0000	0.000	71.3	0.006	BB	0.013	88.65	0.013
26	3.863		0.0000	0.000	15221.2	1.189	BB	0.027	9435.72	1.413
27	3.939		0.0000	0.000	831.8	0.065	BB	0.029	472.70	0.071
28	4.028		0.0000	0.000	4728.0	0.369	BB	0.033	2369.68	0.355
29	4.125		0.0000	0.000	61112.6	4.773	BB	0.035	28940.02	4.334
30	4.180		0.0000	0.000	13081.2	1.022	BB	0.024	9212.22	1.380
31	4.263		0.0000	0.000	20315.3	1.587	BB	0.028	12130.04	1.816
32	4.316		0.0000	0.000	1571.1	0.123	BB	0.018	1468.15	0.220
33	4.359		0.0000	0.000	24092.9	1.882	BB	0.020	19983.44	2.992
34	4.395		0.0000	0.000	31352.7	2.449	BB	0.032	16396.18	2.455
35	4.527		0.0000	0.000	61330.1	4.790	BB	0.034	30092.73	4.506
36	4.598		0.0000	0.000	6460.9	0.505	BB	0.018	5938.63	0.889
37	4.647		0.0000	0.000	30131.9	2.353	BB	0.026	19650.12	2.943
38	4.698		0.0000	0.000	7852.2	0.613	BB	0.022	5930.27	0.888
39	4.761		0.0000	0.000	429.3	0.034	BB	0.017	422.83	0.063
40	4.814		0.0000	0.000	2141.9	0.167	BB	0.024	1513.01	0.227
41	4.857		0.0000	0.000	15223.0	1.189	BB	0.019	13668.52	2.047
42	4.886		0.0000	0.000	11346.3	0.886	BB	0.016	12092.35	1.811
43	4.913		0.0000	0.000	17382.9	1.358	BB	0.018	16251.29	2.434
44	5.001		0.0000	0.000	2629.5	0.205	BB	0.022	1970.89	0.295
45	5.052		0.0000	0.000	86225.6	6.734	BB	0.051	28146.24	4.215
46	5.179		0.0000	0.000	1099.6	0.086	BB	0.023	787.10	0.118
47	5.231		0.0000	0.000	9497.7	0.742	BB	0.026	6151.25	0.921
48	5.299		0.0000	0.000	11009.4	0.860	BB	0.023	8018.80	1.201
49	5.334		0.0000	0.000	3247.2	0.254	BB	0.016	3290.09	0.493
50	5.375		0.0000	0.000	22294.1	1.741	BB	0.025	15109.87	2.263
51	5.477		0.0000	0.000	8469.4	0.661	BB	0.025	5635.10	0.844
52	5.519		0.0000	0.000	746.4	0.058	BB	0.026	486.87	0.073
53	5.589		0.0000	0.000	1386.4	0.108	BB	0.020	1155.12	0.173
54	5.640		0.0000	0.000	18637.1	1.455	BB	0.027	11316.81	1.695

ASX22 0349

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.733		0.0000	0.000	748.5	0.058	BB	0.023	543.84	0.081
56	5.776		0.0000	0.000	2433.9	0.190	BB	0.031	1320.85	0.198
57	5.877		0.0000	0.000	17608.4	1.375	BB	0.030	9895.79	1.482
58	5.966		0.0000	0.000	507.0	0.040	BB	0.024	357.99	0.054
59	6.014		0.0000	0.000	3604.0	0.281	BB	0.034	1791.49	0.268
60	6.156		0.0000	0.000	384.8	0.030	BB	0.045	141.45	0.021
61	6.269		0.0000	0.000	729.3	0.057	BB	0.027	443.37	0.066
62	6.350		0.0000	0.000	342.7	0.027	BB	0.024	239.16	0.036
63	6.388		0.0000	0.000	633.8	0.049	BB	0.018	596.26	0.089
64	6.446		0.0000	0.000	1981.0	0.155	BB	0.055	598.07	0.090
65	6.577		0.0000	0.000	434.1	0.034	BB	0.025	291.11	0.044
66	6.729		0.0000	0.000	5653.3	0.441	BB	0.066	1424.15	0.213
67	6.900		0.0000	0.000	6815.2	0.532	BB	0.193	589.90	0.088
68	6.959		0.0000	0.000	119.0	0.009	BB	0.035	57.34	0.009
69	7.109		0.0000	0.000	6743.5	0.527	BB	0.069	1621.07	0.243
70	7.220		0.0000	0.000	4472.6	0.349	BB	0.085	880.12	0.132
71	7.287		0.0000	0.000	390.4	0.030	BB	0.038	169.92	0.025
72	7.525		0.0000	0.000	12286.8	0.960	BB	0.103	1989.52	0.298
73	7.661		0.0000	0.000	4306.7	0.336	BB	0.131	549.24	0.082
74	7.734		0.0000	0.000	2529.7	0.198	BB	0.035	1206.03	0.181
75	7.874		0.0000	0.000	953.1	0.074	BB	0.058	276.04	0.041
76	7.979	DCB	39.3149	46.397	218606.2	17.072	BB	0.037	99385.99	14.883
77	8.106		0.0000	0.000	2226.6	0.174	BB	0.105	354.98	0.053
78	8.384		0.0000	0.000	3221.1	0.252	BB	0.113	475.69	0.071
79	8.606		0.0000	0.000	352.4	0.028	BB	0.036	164.26	0.025
80	8.689		0.0000	0.000	2551.2	0.199	BB	0.059	722.80	0.108
81	8.970		0.0000	0.000	590.1	0.046	BB	0.081	121.14	0.018

Total Area = 1280481.0, Total Amount = 84.736, Total Height = 667789.9, Sample Units = ug/l

ASX22 0358

Sample Name: **AR4830924B AAAR483AA ICAL 0930699999**
Acquired from CP07–H5088B via port 4 on 11/3/09 05:39:51pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.14R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.538		0.0000	0.000	225.2	0.017	BV	0.023	165.78	0.031
2	1.593		0.0000	0.000	5754.1	0.429	VV	0.042	2306.15	0.435
3	1.636		0.0000	0.000	10436.0	0.778	VV	0.061	2851.52	0.537
4	1.697		0.0000	0.000	3541.1	0.264	VB	0.056	1051.22	0.198
5	1.930		0.0000	0.000	109.5	0.008	BB	0.028	65.10	0.012
6	2.020		0.0000	0.000	573.3	0.043	BB	0.048	201.06	0.038
7	2.339	TCX	41.7912	48.733	431401.0	32.179	BB	0.040	181241.60	34.161
8	2.621		0.0000	0.000	5083.9	0.379	BB	0.037	2293.84	0.432
9	2.714		0.0000	0.000	2336.2	0.174	BB	0.034	1145.47	0.216
10	2.770		0.0000	0.000	289.4	0.022	BB	0.023	210.01	0.040
11	2.859		0.0000	0.000	1282.2	0.096	BB	0.036	588.11	0.111
12	3.005		0.0000	0.000	38795.1	2.894	BB	0.047	13801.38	2.601
13	3.170		0.0000	0.000	18530.1	1.382	BB	0.039	7860.37	1.482
14	3.264		0.0000	0.000	110.0	0.008	BB	0.016	112.48	0.021
15	3.306		0.0000	0.000	3069.3	0.229	BB	0.042	1229.27	0.232
16	3.403		0.0000	0.000	99110.8	7.393	BB	0.069	24014.46	4.526
17	3.555		0.0000	0.000	14926.9	1.113	BB	0.031	8023.58	1.512
18	3.620		0.0000	0.000	7901.8	0.589	BB	0.047	2792.84	0.526
19	3.642		0.0000	0.000	25377.4	1.893	BB	0.038	11109.59	2.094
20	3.796		0.0000	0.000	86264.4	6.435	BB	0.055	26223.83	4.943
21	3.921		0.0000	0.000	20092.4	1.499	BB	0.028	11818.38	2.228
22	3.971		0.0000	0.000	4449.4	0.332	BB	0.026	2887.35	0.544
23	4.115		0.0000	0.000	19100.8	1.425	BB	0.035	9196.44	1.733
24	4.142		0.0000	0.000	10500.7	0.783	BB	0.027	6370.92	1.201
25	4.251		0.0000	0.000	5175.3	0.386	BB	0.022	3978.04	0.750
26	4.292		0.0000	0.000	53230.5	3.971	BB	0.049	18111.63	3.414
27	4.431		0.0000	0.000	9143.6	0.682	BB	0.032	4748.12	0.895
28	4.505		0.0000	0.000	8054.1	0.601	BB	0.026	5066.65	0.955
29	4.537		0.0000	0.000	1390.1	0.104	BB	0.016	1451.15	0.274
30	4.582		0.0000	0.000	25560.9	1.907	BB	0.036	11682.65	2.202
31	4.692		0.0000	0.000	13135.2	0.980	BB	0.038	5833.42	1.100
32	4.794		0.0000	0.000	18674.7	1.393	BB	0.041	7618.13	1.436
33	4.884		0.0000	0.000	2166.8	0.162	BB	0.041	889.20	0.168
34	4.972		0.0000	0.000	2289.8	0.171	BB	0.039	976.26	0.184
35	5.050		0.0000	0.000	16237.9	1.211	BB	0.041	6640.65	1.252
36	5.142		0.0000	0.000	6719.0	0.501	BB	0.053	2114.93	0.399
37	5.287		0.0000	0.000	453.8	0.034	BB	0.052	146.54	0.028
38	5.384		0.0000	0.000	205.4	0.015	BB	0.022	153.26	0.029
39	5.414		0.0000	0.000	926.0	0.069	BB	0.030	517.79	0.098
40	5.634		0.0000	0.000	2188.1	0.163	BB	0.080	456.35	0.086
41	5.804		0.0000	0.000	5474.4	0.408	BB	0.043	2102.77	0.396
42	5.920		0.0000	0.000	1072.1	0.080	BB	0.039	458.09	0.086
43	6.030		0.0000	0.000	12856.4	0.959	BB	0.442	484.99	0.091
44	6.408		0.0000	0.000	423.4	0.032	BB	0.162	43.49	0.008
45	6.470		0.0000	0.000	158.2	0.012	BB	0.074	35.77	0.007
46	6.527		0.0000	0.000	746.3	0.056	BB	0.040	313.79	0.059
47	6.624		0.0000	0.000	4014.4	0.299	BB	0.046	1467.70	0.277
48	6.757		0.0000	0.000	45.9	0.003	BB	0.021	36.29	0.007
49	6.801	DCB	43.9648	51.267	333876.3	24.904	BB	0.041	136341.30	25.698
50	7.370		0.0000	0.000	347.3	0.026	BB	0.046	125.25	0.024
51	7.509		0.0000	0.000	1691.8	0.126	BB	0.078	361.95	0.068
52	7.640		0.0000	0.000	3204.9	0.239	BB	0.126	422.41	0.080
53	7.990		0.0000	0.000	200.0	0.015	BB	0.037	90.95	0.017
54	8.066		0.0000	0.000	880.7	0.066	BB	0.078	188.22	0.035

ASXZ2 0351

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	8.917		0.0000	0.000	842.9	0.063	BB	0.110	127.85	0.024

Total Area = 1340647.0, Total Amount = 85.756, Total Height = 530546.4, Sample Units = ug/l

ASX22 0352

AR4830924B AAAR483AA ICAL 0930699999

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.14R



Instrument ID: CP07–H5088A Injected On: 11/3/2009 5:30:42 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.14R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 5:30:42 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.998	204921	45.42105	TCX	2.339	181242	41.7912	TCX
7.979	99386	39.31493	DCB	6.801	136341	43.96484	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.14A
Area File: C:\CPWIN\DATA1\1P7307B.14A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 5:39:58 PM
File Reported On: 11/3/2009 at 5:40:10 PM

10X REPLOTS

ASX22 0354

C:\CPWIN\DATA1\1P7307.14R



C:\CPWIN\DATA1\1P7307B.14R



ASX22 0355

Sample Name: **AR4840924B AAAR484AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 05:52:49pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.15R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.583		0.0000	0.000	6234.2	0.258	BV	0.058	1801.46	0.139
2	1.686		0.0000	0.000	15815.7	0.655	VV	0.098	2701.32	0.209
3	1.799		0.0000	0.000	20770.1	0.860	VV	0.082	4209.58	0.325
4	1.864		0.0000	0.000	19152.4	0.793	VV	0.063	5060.83	0.391
5	1.928		0.0000	0.000	20283.7	0.839	VB	0.056	6043.14	0.467
6	2.032		0.0000	0.000	26271.2	1.087	BV	0.076	5776.49	0.446
7	2.075		0.0000	0.000	29976.9	1.241	VV	0.070	7136.45	0.551
8	2.175		0.0000	0.000	16947.5	0.701	VV	0.054	5234.58	0.404
9	2.210		0.0000	0.000	25712.9	1.064	VV	0.085	5023.90	0.388
10	2.368		0.0000	0.000	6423.0	0.266	VV	0.047	2292.87	0.177
11	2.444		0.0000	0.000	6087.4	0.252	VV	0.063	1605.25	0.124
12	2.529		0.0000	0.000	3421.2	0.142	VV	0.044	1302.14	0.101
13	2.590		0.0000	0.000	448.8	0.019	VB	0.040	186.34	0.014
14	2.774		0.0000	0.000	4939.2	0.204	BB	0.083	997.63	0.077
15	2.889		0.0000	0.000	6566.5	0.272	BB	0.059	1857.12	0.143
16	2.999	TCX	65.8599	54.130	466442.8	19.304	BB	0.027	288374.50	22.265
17	3.154		0.0000	0.000	210.6	0.009	BB	0.031	112.86	0.009
18	3.226		0.0000	0.000	859.4	0.036	BB	0.025	573.32	0.044
19	3.288		0.0000	0.000	9027.4	0.374	BB	0.030	5069.90	0.391
20	3.382		0.0000	0.000	8154.5	0.337	BB	0.026	5162.14	0.399
21	3.473		0.0000	0.000	860.5	0.036	BB	0.020	715.03	0.055
22	3.515		0.0000	0.000	2394.7	0.099	BB	0.023	1724.74	0.133
23	3.609		0.0000	0.000	407.8	0.017	BB	0.059	115.74	0.009
24	3.694		0.0000	0.000	68790.1	2.847	BB	0.035	32705.21	2.525
25	3.772		0.0000	0.000	2385.9	0.099	BB	0.023	1738.23	0.134
26	3.825		0.0000	0.000	152.2	0.006	BB	0.016	161.17	0.012
27	3.864		0.0000	0.000	38525.7	1.594	BB	0.027	23842.83	1.841
28	3.941		0.0000	0.000	1435.6	0.059	BB	0.034	699.10	0.054
29	4.030		0.0000	0.000	12124.1	0.502	BB	0.033	6126.54	0.473
30	4.126		0.0000	0.000	150583.5	6.232	BB	0.035	71780.47	5.542
31	4.181		0.0000	0.000	32136.0	1.330	BB	0.023	22990.12	1.775
32	4.264		0.0000	0.000	51394.3	2.127	BB	0.028	31043.31	2.397
33	4.317		0.0000	0.000	4303.8	0.178	BB	0.018	4008.37	0.309
34	4.360		0.0000	0.000	55994.8	2.317	BB	0.019	48508.95	3.745
35	4.396		0.0000	0.000	75839.4	3.139	BB	0.032	39665.30	3.062
36	4.528		0.0000	0.000	148200.1	6.133	BB	0.034	73402.13	5.667
37	4.600		0.0000	0.000	15766.7	0.653	BB	0.018	14936.68	1.153
38	4.648		0.0000	0.000	74338.1	3.077	BB	0.025	48604.10	3.753
39	4.699		0.0000	0.000	20399.0	0.844	BB	0.022	15720.54	1.214
40	4.764		0.0000	0.000	1561.6	0.065	BB	0.019	1402.51	0.108
41	4.816		0.0000	0.000	5272.4	0.218	BB	0.024	3667.39	0.283
42	4.859		0.0000	0.000	34744.9	1.438	BB	0.020	29686.94	2.292
43	4.887		0.0000	0.000	26489.7	1.096	BB	0.015	29852.17	2.305
44	4.914		0.0000	0.000	44179.7	1.828	BB	0.019	39576.71	3.056
45	5.003		0.0000	0.000	6019.3	0.249	BB	0.021	4762.14	0.368
46	5.053		0.0000	0.000	217313.6	8.994	BB	0.053	67872.54	5.240
47	5.179		0.0000	0.000	2870.6	0.119	BB	0.021	2237.88	0.173
48	5.233		0.0000	0.000	25339.9	1.049	BB	0.026	16531.01	1.276
49	5.300		0.0000	0.000	28060.5	1.161	BB	0.023	20781.77	1.605
50	5.336		0.0000	0.000	8614.4	0.357	BB	0.016	8829.14	0.682
51	5.376		0.0000	0.000	53967.9	2.233	BB	0.024	37587.39	2.902
52	5.478		0.0000	0.000	22419.5	0.928	BB	0.025	14727.97	1.137
53	5.522		0.0000	0.000	2268.8	0.094	BB	0.028	1374.75	0.106
54	5.590		0.0000	0.000	3893.1	0.161	BB	0.021	3124.38	0.241

ASX22. 8356

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.640		0.0000	0.000	46913.3	1.942	BB	0.027	28873.70	2.229
56	5.734		0.0000	0.000	1968.5	0.081	BB	0.022	1491.54	0.115
57	5.777		0.0000	0.000	6426.5	0.266	BB	0.031	3439.24	0.266
58	5.878		0.0000	0.000	45704.8	1.892	BB	0.029	26339.78	2.034
59	5.966		0.0000	0.000	736.7	0.030	BB	0.023	542.97	0.042
60	6.015		0.0000	0.000	11995.3	0.496	BB	0.035	5645.87	0.436
61	6.140		0.0000	0.000	439.2	0.018	BB	0.036	203.99	0.016
62	6.269		0.0000	0.000	4130.3	0.171	BB	0.035	1971.89	0.152
63	6.354		0.0000	0.000	631.4	0.026	BB	0.024	446.45	0.034
64	6.444		0.0000	0.000	1682.7	0.070	BB	0.049	566.60	0.044
65	6.514		0.0000	0.000	1448.0	0.060	BB	0.027	905.07	0.070
66	6.580		0.0000	0.000	1432.2	0.059	BB	0.025	942.52	0.073
67	6.732		0.0000	0.000	6077.2	0.252	BB	0.067	1504.54	0.116
68	6.840		0.0000	0.000	2661.4	0.110	BB	0.051	877.01	0.068
69	6.900		0.0000	0.000	761.5	0.032	BB	0.042	305.44	0.024
70	7.112		0.0000	0.000	7264.8	0.301	BB	0.074	1642.01	0.127
71	7.222		0.0000	0.000	4802.7	0.199	BB	0.088	905.74	0.070
72	7.289		0.0000	0.000	557.3	0.023	BB	0.039	235.53	0.018
73	7.526		0.0000	0.000	15421.5	0.638	BB	0.115	2233.94	0.172
74	7.663		0.0000	0.000	4155.9	0.172	BB	0.136	508.33	0.039
75	7.734		0.0000	0.000	3830.7	0.159	BB	0.035	1822.76	0.141
76	7.878		0.0000	0.000	996.7	0.041	BB	0.056	298.38	0.023
77	7.980	DCB	55.8092	45.870	307461.4	12.724	BB	0.036	141082.70	10.893
78	8.102		0.0000	0.000	681.0	0.028	BB	0.048	236.35	0.018
79	8.381		0.0000	0.000	2217.3	0.092	BB	0.137	270.30	0.021
80	8.603		0.0000	0.000	341.0	0.014	BB	0.041	140.12	0.011
81	8.692		0.0000	0.000	2142.5	0.089	BB	0.059	600.84	0.046
82	8.968		0.0000	0.000	633.1	0.026	BB	0.077	136.33	0.011

Total Area = 2416309.0, Total Amount = 121.669, Total Height = 1295195.0, Sample Units = ug/l

ASX22 0357

Sample Name: **AR4840924B AAAR484AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 05:52:49pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.15R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.587		0.0000	0.000	10005.7	0.416	BV	0.054	3059.97	0.307
2	1.638		0.0000	0.000	9550.0	0.398	VV	0.052	3085.27	0.309
3	1.704		0.0000	0.000	7907.0	0.329	VV	0.095	1389.42	0.139
4	2.014		0.0000	0.000	2159.6	0.090	VB	0.053	678.76	0.068
5	2.119		0.0000	0.000	762.5	0.032	BB	0.049	256.93	0.026
6	2.273		0.0000	0.000	192.5	0.008	BB	0.023	140.01	0.014
7	2.341	TCX	61.0514	48.688	615254.4	25.610	BB	0.039	264770.00	26.555
8	2.560		0.0000	0.000	568.2	0.024	BB	0.026	367.12	0.037
9	2.624		0.0000	0.000	7890.0	0.328	BB	0.038	3443.09	0.345
10	2.716		0.0000	0.000	9626.6	0.401	BB	0.047	3438.73	0.345
11	2.860		0.0000	0.000	3612.7	0.150	BB	0.037	1626.97	0.163
12	3.007		0.0000	0.000	98174.6	4.087	BB	0.047	34944.00	3.505
13	3.172		0.0000	0.000	47820.3	1.991	BB	0.039	20573.63	2.063
14	3.266		0.0000	0.000	87.1	0.004	BB	0.018	78.78	0.008
15	3.308		0.0000	0.000	9484.0	0.395	BB	0.042	3792.26	0.380
16	3.404		0.0000	0.000	262919.4	10.944	BB	0.068	64310.52	6.450
17	3.557		0.0000	0.000	38938.1	1.621	BB	0.030	21554.26	2.162
18	3.620		0.0000	0.000	23227.9	0.967	BB	0.030	12798.84	1.284
19	3.644		0.0000	0.000	58054.4	2.417	BB	0.038	25362.49	2.544
20	3.797		0.0000	0.000	123783.0	5.152	BB	0.036	56549.02	5.672
21	3.849		0.0000	0.000	4147.8	0.173	BB	0.015	4470.50	0.448
22	3.923		0.0000	0.000	52546.3	2.187	BB	0.028	31211.03	3.130
23	3.973		0.0000	0.000	12682.4	0.528	BB	0.025	8499.47	0.852
24	4.116		0.0000	0.000	49293.0	2.052	BB	0.033	24811.24	2.488
25	4.143		0.0000	0.000	29778.1	1.240	BB	0.029	17380.38	1.743
26	4.252		0.0000	0.000	14574.6	0.607	BB	0.023	10520.67	1.055
27	4.294		0.0000	0.000	141853.1	5.905	BB	0.050	47648.29	4.779
28	4.433		0.0000	0.000	23970.2	0.998	BB	0.032	12557.46	1.259
29	4.507		0.0000	0.000	19675.0	0.819	BB	0.024	13736.58	1.378
30	4.538		0.0000	0.000	3961.2	0.165	BB	0.017	3905.17	0.392
31	4.583		0.0000	0.000	62069.8	2.584	BB	0.034	30597.33	3.069
32	4.694		0.0000	0.000	35475.5	1.477	BB	0.037	15854.27	1.590
33	4.795		0.0000	0.000	48823.3	2.032	BB	0.040	20577.71	2.064
34	4.884		0.0000	0.000	6666.7	0.278	BB	0.042	2644.93	0.265
35	4.974		0.0000	0.000	6398.8	0.266	BB	0.041	2598.50	0.261
36	5.051		0.0000	0.000	45099.9	1.877	BB	0.040	18636.13	1.869
37	5.144		0.0000	0.000	16142.0	0.672	BB	0.048	5606.65	0.562
38	5.387		0.0000	0.000	235.7	0.010	BB	0.033	120.58	0.012
39	5.417		0.0000	0.000	2888.9	0.120	BB	0.031	1553.10	0.156
40	5.631		0.0000	0.000	3465.1	0.144	BB	0.053	1097.01	0.110
41	5.806		0.0000	0.000	2552.5	0.106	BB	0.033	1271.29	0.128
42	5.870		0.0000	0.000	162.1	0.007	BB	0.021	131.10	0.013
43	5.925		0.0000	0.000	2624.8	0.109	BB	0.040	1095.29	0.110
44	6.030		0.0000	0.000	890.9	0.037	BB	0.049	303.06	0.030
45	6.100		0.0000	0.000	824.7	0.034	BB	0.053	261.59	0.026
46	6.240		0.0000	0.000	3603.1	0.150	BB	0.097	620.68	0.062
47	6.518		0.0000	0.000	3976.4	0.166	BB	1.786	37.11	0.004
48	6.532		0.0000	0.000	188.3	0.008	BB	0.021	146.65	0.015
49	6.625		0.0000	0.000	5673.1	0.236	BB	0.043	2215.03	0.222
50	6.803	DCB	64.3424	51.312	467579.2	19.463	BB	0.040	193716.90	19.429
51	7.516		0.0000	0.000	1818.0	0.076	BB	0.078	389.18	0.039
52	7.640		0.0000	0.000	956.3	0.040	BB	0.087	183.21	0.018
53	7.990		0.0000	0.000	189.3	0.008	BB	0.037	86.20	0.009
54	8.330		0.0000	0.000	746.6	0.031	BB	0.084	148.33	0.015

ASX22 0358

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	8.560		0.0000	0.000	375.2	0.016	BB	0.047	133.60	0.013
56	9.096		0.0000	0.000	476.6	0.020	BB	0.092	86.47	0.009

Total Area = 2402402.0, Total Amount = 125.394, Total Height = 997072.8, Sample Units = ug/l

ASX22 0359

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.15R

C:\CPWIN\DATA1\1P7307.15R



Instrument ID: CP07–H5088A Injected On: 11/3/2009 5:43:40 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.15R

C:\CPWIN\DATA1\1P7307B.15R



Instrument ID: CP07–H5088B Injected On: 11/3/2009 5:43:40 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A
2.999	288375	65.85993	TCX
7.98	141083	55.80925	DCB

RT B	Height B	Amount B	Compound B
2.341	264770	61.05141	TCX
6.803	193717	64.34241	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.15A
Area File: C:\CPWIN\DATA1\1P7307B.15A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 5:52:58 PM
File Reported On: 11/3/2009 at 5:53:09 PM

10X REPLOTS

ASX22 0361

C:\CPWIN\DATA1\1P7307.15R



C:\CPWIN\DATA1\1P7307B.15R



ASX22 8362

Sample Name: **AR4850924B AAAR485AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 06:05:49pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.16R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.581		0.0000	0.000	3170.8	0.087	BV	0.047	1119.58	0.055
2	1.698		0.0000	0.000	10242.8	0.281	VV	0.099	1732.92	0.085
3	1.793		0.0000	0.000	13072.7	0.358	VV	0.087	2509.67	0.123
4	1.865		0.0000	0.000	11292.2	0.310	VV	0.063	3009.89	0.147
5	1.929		0.0000	0.000	11785.8	0.323	VB	0.054	3611.18	0.176
6	2.031		0.0000	0.000	15839.5	0.434	BV	0.075	3543.43	0.173
7	2.074		0.0000	0.000	19412.8	0.532	VV	0.057	5644.59	0.276
8	2.173		0.0000	0.000	12457.9	0.342	VV	0.059	3548.11	0.173
9	2.209		0.0000	0.000	17230.6	0.473	VV	0.078	3661.58	0.179
10	2.390		0.0000	0.000	9206.8	0.252	VV	0.118	1298.81	0.063
11	2.450		0.0000	0.000	3413.6	0.094	VV	0.054	1052.44	0.051
12	2.524		0.0000	0.000	3864.4	0.106	VV	0.059	1082.73	0.053
13	2.623		0.0000	0.000	478.2	0.013	VB	0.038	210.35	0.010
14	2.772		0.0000	0.000	1789.4	0.049	BB	0.070	425.12	0.021
15	2.885		0.0000	0.000	2235.5	0.061	BB	0.063	592.46	0.029
16	2.998	TCX	84.0539	54.371	580343.9	15.915	BB	0.027	362662.20	17.703
17	3.226		0.0000	0.000	1706.7	0.047	BB	0.025	1129.51	0.055
18	3.287		0.0000	0.000	12171.8	0.334	BB	0.031	6538.40	0.319
19	3.380		0.0000	0.000	15333.7	0.421	BB	0.026	9790.90	0.478
20	3.472		0.0000	0.000	1014.1	0.028	BB	0.020	860.51	0.042
21	3.514		0.0000	0.000	4754.4	0.130	BB	0.024	3287.84	0.160
22	3.693		0.0000	0.000	124314.8	3.409	BB	0.035	58407.74	2.851
23	3.771		0.0000	0.000	4378.4	0.120	BB	0.023	3169.90	0.155
24	3.824		0.0000	0.000	279.0	0.008	BB	0.015	315.50	0.015
25	3.863		0.0000	0.000	70848.5	1.943	BB	0.027	43836.95	2.140
26	3.940		0.0000	0.000	1798.1	0.049	BB	0.033	908.52	0.044
27	4.028		0.0000	0.000	23058.2	0.632	BB	0.033	11605.13	0.566
28	4.124		0.0000	0.000	269100.3	7.380	BB	0.035	128609.80	6.278
29	4.179		0.0000	0.000	59541.3	1.633	BB	0.023	42420.66	2.071
30	4.263		0.0000	0.000	93141.5	2.554	BB	0.028	55748.70	2.721
31	4.315		0.0000	0.000	8263.4	0.227	BB	0.018	7708.08	0.376
32	4.359		0.0000	0.000	101640.1	2.787	BB	0.020	84452.19	4.122
33	4.394		0.0000	0.000	134216.4	3.681	BB	0.032	68860.27	3.361
34	4.527		0.0000	0.000	267768.1	7.343	BB	0.034	129744.90	6.333
35	4.598		0.0000	0.000	29576.5	0.811	BB	0.018	28143.78	1.374
36	4.647		0.0000	0.000	133112.8	3.650	BB	0.025	87536.82	4.273
37	4.698		0.0000	0.000	37388.0	1.025	BB	0.022	28465.15	1.389
38	4.762		0.0000	0.000	3048.2	0.084	BB	0.018	2746.28	0.134
39	4.815		0.0000	0.000	9652.9	0.265	BB	0.023	7018.29	0.343
40	4.857		0.0000	0.000	67106.7	1.840	BB	0.019	60427.27	2.950
41	4.885		0.0000	0.000	50610.5	1.388	BB	0.016	53331.59	2.603
42	4.913		0.0000	0.000	76299.5	2.092	BB	0.017	72704.14	3.549
43	5.001		0.0000	0.000	12078.1	0.331	BB	0.021	9587.65	0.468
44	5.052		0.0000	0.000	381068.8	10.450	BB	0.052	123318.30	6.020
45	5.178		0.0000	0.000	6101.6	0.167	BB	0.021	4939.51	0.241
46	5.232		0.0000	0.000	46985.4	1.289	BB	0.026	30457.87	1.487
47	5.298		0.0000	0.000	52123.3	1.429	BB	0.022	39547.84	1.930
48	5.334		0.0000	0.000	15930.5	0.437	BB	0.016	16260.91	0.794
49	5.374		0.0000	0.000	97525.9	2.674	BB	0.024	68420.48	3.340
50	5.477		0.0000	0.000	41239.0	1.131	BB	0.025	27679.72	1.351
51	5.520		0.0000	0.000	4503.9	0.124	BB	0.027	2827.44	0.138
52	5.589		0.0000	0.000	7956.5	0.218	BB	0.021	6367.32	0.311
53	5.639		0.0000	0.000	85466.3	2.344	BB	0.027	52812.15	2.578
54	5.733		0.0000	0.000	3607.5	0.099	BB	0.021	2844.15	0.139

ASX22 0363

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.775		0.0000	0.000	12734.2	0.349	BB	0.031	6817.47	0.333
56	5.877		0.0000	0.000	84999.3	2.331	BB	0.028	50076.24	2.444
57	5.961		0.0000	0.000	1340.2	0.037	BB	0.025	887.34	0.043
58	6.013		0.0000	0.000	25717.6	0.705	BB	0.036	11877.22	0.580
59	6.137		0.0000	0.000	901.4	0.025	BB	0.041	365.04	0.018
60	6.267		0.0000	0.000	8838.6	0.242	BB	0.036	4143.71	0.202
61	6.352		0.0000	0.000	595.5	0.016	BB	0.023	437.00	0.021
62	6.388		0.0000	0.000	364.2	0.010	BB	0.016	368.57	0.018
63	6.444		0.0000	0.000	1212.5	0.033	BB	0.046	436.86	0.021
64	6.514		0.0000	0.000	3428.0	0.094	BB	0.028	2024.20	0.099
65	6.579		0.0000	0.000	2864.5	0.079	BB	0.025	1937.55	0.095
66	6.728		0.0000	0.000	5205.8	0.143	BB	0.070	1234.41	0.060
67	6.837		0.0000	0.000	3778.0	0.104	BB	0.041	1534.07	0.075
68	6.896		0.0000	0.000	1110.5	0.030	BB	0.037	499.59	0.024
69	6.973		0.0000	0.000	86.4	0.002	BB	0.028	51.14	0.002
70	7.109		0.0000	0.000	6872.8	0.188	BB	0.075	1525.16	0.074
71	7.201		0.0000	0.000	5401.1	0.148	BB	0.073	1239.49	0.061
72	7.292		0.0000	0.000	828.2	0.023	BB	0.039	353.24	0.017
73	7.522		0.0000	0.000	8487.6	0.233	BB	0.088	1613.13	0.079
74	7.651		0.0000	0.000	4349.4	0.119	BB	0.108	673.26	0.033
75	7.733		0.0000	0.000	5009.9	0.137	BB	0.034	2445.88	0.119
76	7.869		0.0000	0.000	719.5	0.020	BB	0.052	230.89	0.011
77	7.977	DCB	70.5405	45.629	387855.6	10.636	BB	0.036	178322.50	8.705
78	8.384		0.0000	0.000	1075.1	0.029	BB	0.060	299.16	0.015
79	8.602		0.0000	0.000	267.8	0.007	BB	0.042	105.11	0.005
80	8.683		0.0000	0.000	1925.3	0.053	BB	0.057	562.54	0.027

Total Area = 3646517.0, Total Amount = 154.594, Total Height = 2048598.0, Sample Units = ug/l

ASX22 0364

Sample Name: **AR4850924B AAAR485AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 06:05:49pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.16R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.587		0.0000	0.000	10611.0	0.269	BV	0.053	3320.74	0.208
2	1.636		0.0000	0.000	7916.9	0.201	VV	0.045	2923.22	0.183
3	1.701		0.0000	0.000	4482.0	0.114	VB	0.053	1406.86	0.088
4	2.013		0.0000	0.000	733.2	0.019	BB	0.044	279.13	0.017
5	2.276		0.0000	0.000	336.8	0.009	BB	0.031	180.21	0.011
6	2.339	TCX	78.1885	48.485	788609.8	20.028	BB	0.039	339090.90	21.258
7	2.558		0.0000	0.000	1220.0	0.031	BB	0.029	700.07	0.044
8	2.623		0.0000	0.000	10761.4	0.273	BB	0.040	4467.41	0.280
9	2.715		0.0000	0.000	17913.9	0.455	BB	0.045	6580.11	0.413
10	2.859		0.0000	0.000	7063.8	0.179	BB	0.037	3154.31	0.198
11	3.004		0.0000	0.000	178902.1	4.544	BB	0.047	63807.45	4.000
12	3.170		0.0000	0.000	88648.9	2.251	BB	0.039	38285.57	2.400
13	3.262		0.0000	0.000	34.0	0.001	BB	0.008	66.64	0.004
14	3.306		0.0000	0.000	18877.1	0.479	BB	0.042	7479.48	0.469
15	3.402		0.0000	0.000	489632.2	12.435	BB	0.067	121498.70	7.617
16	3.555		0.0000	0.000	71612.3	1.819	BB	0.030	39843.25	2.498
17	3.617		0.0000	0.000	38321.9	0.973	BB	0.025	25496.98	1.598
18	3.641		0.0000	0.000	110242.0	2.800	BB	0.039	46908.13	2.941
19	3.795		0.0000	0.000	391994.6	9.955	BB	0.054	121434.70	7.613
20	3.921		0.0000	0.000	95723.6	2.431	BB	0.027	58041.95	3.639
21	3.971		0.0000	0.000	25089.2	0.637	BB	0.026	16125.56	1.011
22	4.114		0.0000	0.000	90082.5	2.288	BB	0.031	47960.94	3.007
23	4.141		0.0000	0.000	59415.3	1.509	BB	0.030	32904.11	2.063
24	4.250		0.0000	0.000	25249.8	0.641	BB	0.021	20052.59	1.257
25	4.291		0.0000	0.000	259750.2	6.597	BB	0.050	86702.00	5.435
26	4.431		0.0000	0.000	44524.7	1.131	BB	0.032	23398.27	1.467
27	4.505		0.0000	0.000	39183.2	0.995	BB	0.023	28252.77	1.771
28	4.536		0.0000	0.000	7582.6	0.193	BB	0.016	8056.19	0.505
29	4.581		0.0000	0.000	113239.2	2.876	BB	0.033	57032.38	3.575
30	4.692		0.0000	0.000	57998.3	1.473	BB	0.033	29090.55	1.824
31	4.745		0.0000	0.000	678.1	0.017	BB	0.011	1030.94	0.065
32	4.793		0.0000	0.000	91148.0	2.315	BB	0.039	39170.87	2.456
33	4.882		0.0000	0.000	13996.1	0.355	BB	0.044	5272.70	0.331
34	4.972		0.0000	0.000	12615.8	0.320	BB	0.041	5073.96	0.318
35	5.050		0.0000	0.000	89784.8	2.280	BB	0.040	37338.28	2.341
36	5.143		0.0000	0.000	29353.0	0.745	BB	0.046	10656.64	0.668
37	5.417		0.0000	0.000	12239.7	0.311	BB	0.049	4137.24	0.259
38	5.629		0.0000	0.000	8214.6	0.209	BB	0.055	2484.72	0.156
39	5.803		0.0000	0.000	2464.7	0.063	BB	0.033	1227.88	0.077
40	5.870		0.0000	0.000	352.0	0.009	BB	0.021	272.94	0.017
41	5.925		0.0000	0.000	4502.3	0.114	BB	0.039	1924.14	0.121
42	6.030		0.0000	0.000	2799.8	0.071	BB	0.114	410.40	0.026
43	6.240		0.0000	0.000	1850.0	0.047	BB	0.092	335.17	0.021
44	6.388		0.0000	0.000	5480.8	0.139	BB	0.120	758.32	0.048
45	6.624		0.0000	0.000	7457.5	0.189	BB	0.042	2956.98	0.185
46	6.801	DCB	83.0736	51.515	590407.4	14.994	BB	0.040	246456.80	15.451
47	7.640		0.0000	0.000	5504.1	0.140	BB	0.183	500.50	0.031
48	7.977		0.0000	0.000	2090.0	0.053	BB	0.154	226.74	0.014
49	8.193		0.0000	0.000	341.7	0.009	BB	0.047	120.26	0.008
50	8.396		0.0000	0.000	185.2	0.005	BB	0.030	103.54	0.006
51	8.563		0.0000	0.000	317.3	0.008	BB	0.042	126.06	0.008

ASX22 0365

Total Area = 3937536.0, Total Amount = 161.262, Total Height = 1595126.0, Sample Units = ug/l

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.16R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 5:56:40 PM Column ID: RTX-CLP1 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.16R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 5:56:40 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.998	362662	84.05388	TCX	2.339	339091	78.18854	TCX
7.977	178322	70.54048	DCB	6.801	246457	83.07357	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.16A
Area File: C:\CPWIN\DATA1\1P7307B.16A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 6:05:58 PM
File Reported On: 11/3/2009 at 6:06:10 PM

10X REPLOTS

6SX22 8367

C:\CPWIN\DATA1\1P7307.16R



C:\CPWIN\DATA1\1P7307B.16R



ASX22 0368

Sample Name: **AR5410924B AAAR541AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 06:18:47pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.17R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.583		0.0000	0.000	1717.8	0.306	BV	0.043	670.81	0.300
2	1.663		0.0000	0.000	6682.8	1.189	VV	0.078	1419.50	0.636
3	1.794		0.0000	0.000	12247.8	2.180	VV	0.088	2331.29	1.044
4	1.865		0.0000	0.000	10492.1	1.867	VV	0.063	2789.33	1.249
5	1.930		0.0000	0.000	11417.5	2.032	VB	0.054	3509.26	1.571
6	2.038		0.0000	0.000	17614.3	3.135	BV	0.086	3416.67	1.530
7	2.175		0.0000	0.000	30537.2	5.434	VV	0.145	3521.21	1.576
8	2.283		0.0000	0.000	5742.1	1.022	VV	0.059	1628.63	0.729
9	2.371		0.0000	0.000	5042.3	0.897	VV	0.069	1213.27	0.543
10	2.447		0.0000	0.000	3846.8	0.685	VV	0.077	835.18	0.374
11	2.532		0.0000	0.000	3472.1	0.618	VB	0.064	898.32	0.402
12	2.770		0.0000	0.000	3889.9	0.692	BB	0.057	1142.84	0.512
13	2.864		0.0000	0.000	23278.5	4.143	BB	0.079	4935.67	2.210
14	2.999	TCX	3.0934	33.067	79063.3	14.070	BB	0.041	32092.65	14.368
15	3.132		0.0000	0.000	255.0	0.045	BB	0.027	154.80	0.069
16	3.214		0.0000	0.000	199.5	0.036	BB	0.029	115.05	0.052
17	3.288		0.0000	0.000	2851.2	0.507	BB	0.049	977.21	0.438
18	3.386		0.0000	0.000	2147.4	0.382	BB	0.064	561.41	0.251
19	3.469		0.0000	0.000	190.1	0.034	BB	0.040	78.77	0.035
20	3.607		0.0000	0.000	296.2	0.053	BB	0.048	102.23	0.046
21	3.724		0.0000	0.000	274.7	0.049	BB	0.026	175.18	0.078
22	3.770		0.0000	0.000	2751.0	0.490	BB	0.101	455.65	0.204
23	3.880		0.0000	0.000	1565.8	0.279	BB	0.047	561.09	0.251
24	3.937		0.0000	0.000	89.7	0.016	BB	0.023	64.98	0.029
25	3.970		0.0000	0.000	432.9	0.077	BB	0.042	171.13	0.077
26	4.034		0.0000	0.000	646.2	0.115	BB	0.040	271.63	0.122
27	4.105		0.0000	0.000	1284.4	0.229	BB	0.030	713.29	0.319
28	4.228		0.0000	0.000	515.4	0.092	BB	0.087	98.27	0.044
29	4.266		0.0000	0.000	2559.0	0.455	BB	0.046	929.59	0.416
30	4.361		0.0000	0.000	7759.4	1.381	BB	0.022	5883.40	2.634
31	4.528		0.0000	0.000	6085.2	1.083	BB	0.027	3764.83	1.686
32	4.597		0.0000	0.000	153.5	0.027	BB	0.015	172.46	0.077
33	4.649		0.0000	0.000	2192.1	0.390	BB	0.026	1406.39	0.630
34	4.699		0.0000	0.000	2671.9	0.475	BB	0.060	747.45	0.335
35	4.811		0.0000	0.000	3063.3	0.545	BB	0.066	776.00	0.347
36	4.857		0.0000	0.000	9097.5	1.619	BB	0.020	7583.88	3.395
37	4.886		0.0000	0.000	888.6	0.158	BB	0.011	1330.19	0.596
38	4.914		0.0000	0.000	1494.5	0.266	BB	0.018	1388.00	0.621
39	5.003		0.0000	0.000	2910.9	0.518	BB	0.022	2190.67	0.981
40	5.058		0.0000	0.000	17695.4	3.149	BB	0.026	11310.23	5.064
41	5.104		0.0000	0.000	14173.0	2.522	BB	0.027	8686.23	3.889
42	5.234		0.0000	0.000	10741.9	1.912	BB	0.036	4984.18	2.231
43	5.300		0.0000	0.000	16558.0	2.947	BB	0.032	8505.95	3.808
44	5.376		0.0000	0.000	22723.0	4.044	BB	0.025	15036.85	6.732
45	5.435		0.0000	0.000	1061.9	0.189	BB	0.027	648.05	0.290
46	5.476		0.0000	0.000	2100.3	0.374	BB	0.021	1694.43	0.759
47	5.523		0.0000	0.000	13990.0	2.490	BB	0.041	5755.23	2.577
48	5.641		0.0000	0.000	22680.2	4.036	BB	0.032	11943.14	5.347
49	5.707		0.0000	0.000	886.1	0.158	BB	0.024	615.24	0.275
50	5.774		0.0000	0.000	17444.9	3.104	BB	0.038	7700.39	3.448
51	5.878		0.0000	0.000	13379.1	2.381	BB	0.030	7315.64	3.275
52	5.966		0.0000	0.000	2754.0	0.490	BB	0.029	1592.63	0.713
53	6.017		0.0000	0.000	19565.7	3.482	BB	0.039	8294.17	3.713
54	6.139		0.0000	0.000	768.8	0.137	BB	0.039	327.49	0.147

ASX22 0369

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.268		0.0000	0.000	6663.7	1.186	BB	0.036	3073.01	1.376
56	6.353		0.0000	0.000	2561.7	0.456	BB	0.039	1106.81	0.496
57	6.449		0.0000	0.000	3526.8	0.628	BB	0.058	1009.04	0.452
58	6.514		0.0000	0.000	2676.7	0.476	BB	0.028	1604.14	0.718
59	6.580		0.0000	0.000	1779.8	0.317	BB	0.024	1213.47	0.543
60	6.732		0.0000	0.000	7951.3	1.415	BB	0.062	2151.45	0.963
61	6.839		0.0000	0.000	4695.4	0.836	BB	0.051	1532.42	0.686
62	6.893		0.0000	0.000	368.0	0.065	BB	0.038	162.89	0.073
63	6.973		0.0000	0.000	143.7	0.026	BB	0.032	75.31	0.034
64	7.106		0.0000	0.000	9114.0	1.622	BB	0.070	2166.30	0.970
65	7.221		0.0000	0.000	7883.6	1.403	BB	0.083	1587.31	0.711
66	7.297		0.0000	0.000	1212.5	0.216	BB	0.043	472.15	0.211
67	7.518		0.0000	0.000	14541.4	2.588	BB	0.095	2551.03	1.142
68	7.662		0.0000	0.000	6878.4	1.224	BB	0.135	846.78	0.379
69	7.735		0.0000	0.000	501.5	0.089	BB	0.043	194.20	0.087
70	7.871		0.0000	0.000	1076.2	0.192	BB	0.065	276.20	0.124
71	7.980	DCB	6.2615	66.933	37184.9	6.617	BB	0.039	15828.62	7.087
72	8.134		0.0000	0.000	2183.5	0.389	BB	0.108	336.64	0.151
73	8.211		0.0000	0.000	365.9	0.065	BB	0.067	90.35	0.040
74	8.384		0.0000	0.000	3749.9	0.667	BB	0.102	613.11	0.274
75	8.602		0.0000	0.000	828.3	0.147	BB	0.037	376.95	0.169
76	8.692		0.0000	0.000	2104.6	0.375	BB	0.058	603.89	0.270

Total Area = 561933.6, Total Amount = 9.355, Total Height = 223360.1, Sample Units = ug/l

ASX22 0376

Sample Name: **AR5410924B AAAR541AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 06:18:47pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.17R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.578		0.0000	0.000	3770.5	0.855	BV	0.038	1648.64	1.041
2	1.637		0.0000	0.000	17665.7	4.008	VB	0.105	2813.35	1.777
3	1.981		0.0000	0.000	1138.7	0.258	BV	0.131	144.73	0.091
4	2.014		0.0000	0.000	973.7	0.221	VB	0.047	346.76	0.219
5	2.113		0.0000	0.000	228.0	0.052	BB	0.027	140.86	0.089
6	2.183		0.0000	0.000	6046.8	1.372	BB	0.136	741.49	0.468
7	2.341	TCX	5.2924	65.638	69686.9	15.810	BB	0.051	22952.47	14.497
8	2.623		0.0000	0.000	2556.6	0.580	BB	0.137	309.90	0.196
9	3.007		0.0000	0.000	2380.1	0.540	BB	0.162	244.19	0.154
10	3.160		0.0000	0.000	1874.9	0.425	BB	0.067	464.64	0.293
11	3.281		0.0000	0.000	94.0	0.021	BB	0.039	40.14	0.025
12	3.325		0.0000	0.000	208.1	0.047	BB	0.097	35.89	0.023
13	3.351		0.0000	0.000	3466.9	0.787	BB	0.074	777.08	0.491
14	3.560		0.0000	0.000	591.2	0.134	BB	0.036	270.38	0.171
15	3.621		0.0000	0.000	18828.2	4.271	BB	0.052	6079.00	3.840
16	3.797		0.0000	0.000	8160.5	1.851	BB	0.044	3119.52	1.970
17	3.921		0.0000	0.000	1280.5	0.291	BB	0.031	688.22	0.435
18	4.112		0.0000	0.000	29940.6	6.793	BB	0.045	10998.25	6.947
19	4.202		0.0000	0.000	1206.4	0.274	BB	0.021	957.51	0.605
20	4.254		0.0000	0.000	9185.7	2.084	BB	0.025	6035.71	3.812
21	4.295		0.0000	0.000	11924.2	2.705	BB	0.031	6317.99	3.991
22	4.382		0.0000	0.000	248.7	0.056	BB	0.017	240.71	0.152
23	4.434		0.0000	0.000	7217.9	1.638	BB	0.034	3547.31	2.241
24	4.507		0.0000	0.000	7260.5	1.647	BB	0.026	4659.39	2.943
25	4.536		0.0000	0.000	478.5	0.109	BB	0.013	623.86	0.394
26	4.584		0.0000	0.000	29284.0	6.644	BB	0.040	12194.53	7.702
27	4.695		0.0000	0.000	15220.1	3.453	BB	0.036	7055.26	4.456
28	4.797		0.0000	0.000	20646.9	4.684	BB	0.044	7759.59	4.901
29	4.877		0.0000	0.000	11605.8	2.633	BB	0.038	5083.27	3.211
30	4.975		0.0000	0.000	10867.7	2.466	BB	0.045	4062.30	2.566
31	5.050		0.0000	0.000	5554.4	1.260	BB	0.029	3198.07	2.020
32	5.091		0.0000	0.000	366.6	0.083	BB	0.018	332.61	0.210
33	5.146		0.0000	0.000	30837.0	6.996	BB	0.048	10699.84	6.758
34	5.418		0.0000	0.000	11041.6	2.505	BB	0.054	3410.11	2.154
35	5.557		0.0000	0.000	402.5	0.091	BB	0.022	299.15	0.189
36	5.632		0.0000	0.000	7349.5	1.667	BB	0.065	1895.86	1.197
37	5.805		0.0000	0.000	7141.7	1.620	BB	0.042	2817.70	1.780
38	5.920		0.0000	0.000	3645.1	0.827	BB	0.038	1611.96	1.018
39	6.033		0.0000	0.000	1152.5	0.261	BB	0.053	365.82	0.231
40	6.097		0.0000	0.000	9824.1	2.229	BB	0.395	414.23	0.262
41	6.498		0.0000	0.000	2527.0	0.573	BB	1.154	36.49	0.023
42	6.527		0.0000	0.000	571.4	0.130	BB	0.036	267.64	0.169
43	6.590		0.0000	0.000	103.8	0.024	BB	0.021	81.94	0.052
44	6.645		0.0000	0.000	634.7	0.144	BB	0.039	269.71	0.170
45	6.804	DCB	2.7706	34.362	50915.9	11.551	BB	0.042	20353.75	12.856
46	7.510		0.0000	0.000	4030.2	0.914	BB	0.136	492.39	0.311
47	7.637		0.0000	0.000	3298.8	0.748	BB	0.122	449.28	0.284
48	7.833		0.0000	0.000	421.2	0.096	BB	0.060	117.01	0.074
49	8.050		0.0000	0.000	2197.0	0.498	BB	0.129	283.28	0.179
50	8.340		0.0000	0.000	1599.2	0.363	BB	0.116	230.74	0.146
51	8.564		0.0000	0.000	2291.3	0.520	BB	0.173	220.19	0.139
52	8.907		0.0000	0.000	844.1	0.192	BB	0.117	120.14	0.076

ASX22[illegible]0371

Total Area = 440788.3, Total Amount = 8.063, Total Height = 158320.9, Sample Units = ug/l

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.17R

C:\CPWIN\DATA1\1P7307.17R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 6:09:38 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.17R

C:\CPWIN\DATA1\1P7307B.17R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 6:09:38 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	32093	3.09342	TCX	2.341	22952	5.29245	TCX
7.98	15829	6.26146	DCB	6.804	20354	2.77058	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.17A
Area File: C:\CPWIN\DATA1\1P7307B.17A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 6:18:54 PM
File Reported On: 11/3/2009 at 6:19:07 PM

10X REPLOTS

ASX22 0373

C:\CPWIN\DATA1\1P7307.17R



C:\CPWIN\DATA1\1P7307B.17R



ASX22 0374

Sample Name: **AR5420924B AAAR542AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 06:31:46pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.18R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.581		0.0000	0.000	1344.9	0.159	BV	0.034	664.13	0.157
2	1.672		0.0000	0.000	8749.1	1.036	VB	0.076	1915.86	0.452
3	1.785		0.0000	0.000	12795.5	1.515	BV	0.092	2315.32	0.547
4	1.866		0.0000	0.000	9085.2	1.076	VV	0.068	2217.07	0.523
5	1.930		0.0000	0.000	7577.2	0.897	VB	0.052	2416.01	0.570
6	2.000		0.0000	0.000	4827.4	0.572	BV	0.041	1953.82	0.461
7	2.073		0.0000	0.000	8859.5	1.049	VV	0.086	1710.29	0.404
8	2.174		0.0000	0.000	1772.4	0.210	VB	0.055	540.90	0.128
9	2.358		0.0000	0.000	20984.2	2.485	BV	0.173	2027.44	0.479
10	2.559		0.0000	0.000	13497.7	1.598	VB	0.188	1196.70	0.283
11	2.727		0.0000	0.000	306.1	0.036	BB	0.081	62.67	0.015
12	2.765		0.0000	0.000	2582.4	0.306	BB	0.028	1517.81	0.358
13	2.870		0.0000	0.000	10676.0	1.264	BB	0.103	1721.64	0.406
14	2.998	TCX	22.0980	49.728	174601.6	20.677	BB	0.027	109690.20	25.898
15	3.214		0.0000	0.000	207.2	0.025	BB	0.035	99.96	0.024
16	3.286		0.0000	0.000	7257.7	0.859	BB	0.041	2933.12	0.693
17	3.386		0.0000	0.000	4236.3	0.502	BB	0.069	1019.87	0.241
18	3.471		0.0000	0.000	484.4	0.057	BB	0.030	273.37	0.065
19	3.521		0.0000	0.000	80.8	0.010	BB	0.024	55.50	0.013
20	3.600		0.0000	0.000	305.3	0.036	BB	0.042	120.91	0.029
21	3.691		0.0000	0.000	110.4	0.013	BB	0.016	118.33	0.028
22	3.726		0.0000	0.000	221.0	0.026	BB	0.029	126.33	0.030
23	3.774		0.0000	0.000	4691.3	0.556	BB	0.107	733.85	0.173
24	3.887		0.0000	0.000	2350.8	0.278	BB	0.071	554.33	0.131
25	3.939		0.0000	0.000	448.3	0.053	BB	0.028	269.38	0.064
26	4.027		0.0000	0.000	693.8	0.082	BB	0.043	267.05	0.063
27	4.107		0.0000	0.000	1957.6	0.232	BB	0.045	728.75	0.172
28	4.186		0.0000	0.000	305.8	0.036	BB	0.030	171.73	0.041
29	4.264		0.0000	0.000	3244.2	0.384	BB	0.049	1092.58	0.258
30	4.359		0.0000	0.000	35460.2	4.199	BB	0.043	13616.53	3.215
31	4.527		0.0000	0.000	11112.1	1.316	BB	0.027	6825.68	1.612
32	4.597		0.0000	0.000	331.4	0.039	BB	0.017	325.27	0.077
33	4.648		0.0000	0.000	3858.6	0.457	BB	0.025	2532.11	0.598
34	4.698		0.0000	0.000	2374.0	0.281	BB	0.056	704.70	0.166
35	4.807		0.0000	0.000	2781.9	0.329	BB	0.055	836.81	0.198
36	4.856		0.0000	0.000	15905.4	1.884	BB	0.020	13055.77	3.082
37	4.885		0.0000	0.000	1945.1	0.230	BB	0.012	2763.98	0.653
38	4.912		0.0000	0.000	1768.6	0.209	BB	0.014	2071.36	0.489
39	5.002		0.0000	0.000	4673.9	0.554	BB	0.022	3611.93	0.853
40	5.056		0.0000	0.000	32249.0	3.819	BB	0.027	19951.84	4.711
41	5.100		0.0000	0.000	10505.1	1.244	BB	0.022	7924.79	1.871
42	5.232		0.0000	0.000	12918.2	1.530	BB	0.028	7681.40	1.814
43	5.299		0.0000	0.000	19070.3	2.258	BB	0.024	13068.37	3.085
44	5.332		0.0000	0.000	984.3	0.117	BB	0.011	1462.19	0.345
45	5.375		0.0000	0.000	38699.4	4.583	BB	0.025	26042.99	6.149
46	5.436		0.0000	0.000	1955.1	0.232	BB	0.027	1194.99	0.282
47	5.475		0.0000	0.000	3735.1	0.442	BB	0.020	3055.60	0.721
48	5.521		0.0000	0.000	14847.5	1.758	BB	0.028	8971.24	2.118
49	5.583		0.0000	0.000	691.9	0.082	BB	0.014	842.48	0.199
50	5.639		0.0000	0.000	33649.3	3.985	BB	0.029	19443.14	4.591
51	5.704		0.0000	0.000	2123.8	0.252	BB	0.024	1490.27	0.352
52	5.772		0.0000	0.000	30491.2	3.611	BB	0.038	13540.03	3.197
53	5.876		0.0000	0.000	19470.8	2.306	BB	0.025	12834.62	3.030
54	5.920		0.0000	0.000	411.6	0.049	BB	0.012	563.46	0.133

ASX22 0375

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.963		0.0000	0.000	2358.4	0.279	BB	0.026	1515.37	0.358
56	6.013		0.0000	0.000	37424.3	4.432	BB	0.036	17383.72	4.104
57	6.136		0.0000	0.000	1268.0	0.150	BB	0.035	609.57	0.144
58	6.266		0.0000	0.000	12303.6	1.457	BB	0.036	5752.54	1.358
59	6.351		0.0000	0.000	1848.9	0.219	BB	0.035	883.35	0.209
60	6.445		0.0000	0.000	1646.4	0.195	BB	0.047	582.50	0.138
61	6.514		0.0000	0.000	5225.0	0.619	BB	0.029	3033.81	0.716
62	6.579		0.0000	0.000	3257.8	0.386	BB	0.025	2210.88	0.522
63	6.728		0.0000	0.000	5335.2	0.632	BB	0.067	1331.76	0.314
64	6.838		0.0000	0.000	4364.4	0.517	BB	0.034	2119.55	0.500
65	6.881		0.0000	0.000	455.6	0.054	BB	0.034	225.64	0.053
66	6.963		0.0000	0.000	133.0	0.016	BB	0.033	66.25	0.016
67	7.106		0.0000	0.000	7445.8	0.882	BB	0.073	1710.25	0.404
68	7.220		0.0000	0.000	6353.2	0.752	BB	0.082	1296.88	0.306
69	7.293		0.0000	0.000	1040.5	0.123	BB	0.041	424.28	0.100
70	7.519		0.0000	0.000	10951.1	1.297	BB	0.094	1936.81	0.457
71	7.652		0.0000	0.000	5346.0	0.633	BB	0.118	757.53	0.179
72	7.734		0.0000	0.000	1330.3	0.158	BB	0.036	611.67	0.144
73	7.868		0.0000	0.000	785.7	0.093	BB	0.054	242.47	0.057
74	7.979	DCB	22.3397	50.272	124123.4	14.699	BB	0.037	56473.49	13.333
75	8.121		0.0000	0.000	1636.1	0.194	BB	0.070	388.34	0.092
76	8.201		0.0000	0.000	342.0	0.041	BB	0.069	82.81	0.020
77	8.386		0.0000	0.000	822.8	0.097	BB	0.059	230.53	0.054
78	8.603		0.0000	0.000	588.3	0.070	BB	0.039	251.12	0.059
79	8.685		0.0000	0.000	1762.6	0.209	BB	0.058	503.81	0.119

Total Area = 844416.1, Total Amount = 44.438, Total Height = 423551.3, Sample Units = ug/l

ASX22 8376

Sample Name: **AR5420924B AAAR542AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 06:31:46pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.18R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.591		0.0000	0.000	5346.0	0.582	BV	0.042	2136.51	0.594
2	1.637		0.0000	0.000	23386.0	2.544	VB	0.132	2956.33	0.822
3	1.943		0.0000	0.000	2069.5	0.225	BV	0.087	397.35	0.110
4	2.014		0.0000	0.000	1134.7	0.123	VB	0.065	291.07	0.081
5	2.340	TCX	21.3036	48.638	218735.0	23.798	BB	0.039	92390.32	25.675
6	2.621		0.0000	0.000	2420.3	0.263	BB	0.036	1107.71	0.308
7	2.710		0.0000	0.000	1132.5	0.123	BB	0.132	143.28	0.040
8	2.967		0.0000	0.000	1886.7	0.205	BB	0.878	35.80	0.010
9	3.161		0.0000	0.000	7227.5	0.786	BB	0.132	909.50	0.253
10	3.328		0.0000	0.000	794.3	0.086	BB	0.355	37.33	0.010
11	3.356		0.0000	0.000	142.5	0.016	BB	0.033	71.33	0.020
12	3.403		0.0000	0.000	2657.3	0.289	BB	0.069	641.70	0.178
13	3.557		0.0000	0.000	493.0	0.054	BB	0.033	252.35	0.070
14	3.619		0.0000	0.000	34096.9	3.710	BB	0.052	10872.97	3.022
15	3.795		0.0000	0.000	15126.8	1.646	BB	0.044	5701.62	1.584
16	3.921		0.0000	0.000	3654.3	0.398	BB	0.038	1599.53	0.445
17	4.036		0.0000	0.000	309.2	0.034	BB	0.026	197.89	0.055
18	4.110		0.0000	0.000	53574.2	5.829	BB	0.045	19777.69	5.496
19	4.202		0.0000	0.000	2182.2	0.237	BB	0.020	1823.59	0.507
20	4.252		0.0000	0.000	15553.1	1.692	BB	0.026	10095.59	2.806
21	4.292		0.0000	0.000	15039.2	1.636	BB	0.030	8317.42	2.311
22	4.384		0.0000	0.000	630.4	0.069	BB	0.020	531.45	0.148
23	4.433		0.0000	0.000	12289.5	1.337	BB	0.034	6099.15	1.695
24	4.505		0.0000	0.000	25430.1	2.767	BB	0.038	11189.35	3.109
25	4.583		0.0000	0.000	53972.5	5.872	BB	0.040	22226.34	6.177
26	4.694		0.0000	0.000	26970.9	2.934	BB	0.036	12552.79	3.488
27	4.795		0.0000	0.000	34490.1	3.752	BB	0.042	13713.38	3.811
28	4.874		0.0000	0.000	19707.7	2.144	BB	0.036	9076.76	2.522
29	4.976		0.0000	0.000	19335.8	2.104	BB	0.044	7256.21	2.016
30	5.049		0.0000	0.000	10467.3	1.139	BB	0.030	5894.41	1.638
31	5.090		0.0000	0.000	623.3	0.068	BB	0.017	619.02	0.172
32	5.144		0.0000	0.000	55148.6	6.000	BB	0.048	19188.07	5.332
33	5.417		0.0000	0.000	17305.0	1.883	BB	0.048	5974.60	1.660
34	5.513		0.0000	0.000	362.5	0.039	BB	0.021	291.17	0.081
35	5.627		0.0000	0.000	11292.0	1.229	BB	0.057	3289.40	0.914
36	5.801		0.0000	0.000	2227.0	0.242	BB	0.040	932.23	0.259
37	5.921		0.0000	0.000	6290.7	0.684	BB	0.041	2531.23	0.703
38	6.024		0.0000	0.000	14299.8	1.556	BB	0.383	622.82	0.173
39	6.527		0.0000	0.000	5776.4	0.628	BB	0.184	522.48	0.145
40	6.627		0.0000	0.000	1968.5	0.214	BB	0.045	721.68	0.201
41	6.802	DCB	22.4966	51.362	186326.8	20.272	BB	0.041	75894.91	21.091
42	7.509		0.0000	0.000	499.4	0.054	BB	0.045	183.49	0.051
43	7.620		0.0000	0.000	5319.8	0.579	BB	0.179	494.87	0.138
44	8.316		0.0000	0.000	1248.6	0.136	BB	0.130	160.39	0.045
45	8.393		0.0000	0.000	186.2	0.020	BB	0.026	121.54	0.034

Total Area = 919129.9, Total Amount = 43.8, Total Height = 359844.6, Sample Units = ug/l

ASX22 0377

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.18R

C:\CPWIN\DATA1\1P7307.18R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 6:22:37 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.18R

C:\CPWIN\DATA1\1P7307B.18R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 6:22:37 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.998	109690	22.09799	TCX	2.34	92390	21.30362	TCX
7.979	56473	22.33968	DCB	6.802	75895	22.49664	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.18A
Area File: C:\CPWIN\DATA1\1P7307B.18A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 6:31:54 PM
File Reported On: 11/3/2009 at 6:32:06 PM

10X REPLOTS

ASX22 0379

C:\CPWIN\DATA1\1P7307.18R



C:\CPWIN\DATA1\1P7307B.18R



ASX22 0388

Sample Name: **AR5430924B AAAR543AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 06:44:42pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.19R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.670		0.0000	0.000	12866.8	0.800	BV	0.105	2049.42	0.248
2	1.796		0.0000	0.000	20825.0	1.295	VV	0.099	3515.30	0.426
3	1.928		0.0000	0.000	33112.5	2.058	VB	0.106	5212.21	0.631
4	2.074		0.0000	0.000	45302.5	2.816	BV	0.130	5827.99	0.706
5	2.175		0.0000	0.000	14240.6	0.885	VV	0.050	4730.97	0.573
6	2.208		0.0000	0.000	9708.6	0.604	VV	0.037	4334.88	0.525
7	2.298		0.0000	0.000	13570.4	0.844	VV	0.088	2562.28	0.310
8	2.359		0.0000	0.000	6923.3	0.430	VV	0.055	2083.53	0.252
9	2.578		0.0000	0.000	11941.8	0.742	VV	0.309	643.59	0.078
10	2.625		0.0000	0.000	631.8	0.039	VB	0.040	263.76	0.032
11	2.769		0.0000	0.000	5830.1	0.362	BB	0.076	1282.81	0.155
12	2.883		0.0000	0.000	13222.8	0.822	BB	0.073	3020.13	0.366
13	2.999	TCX	42.0045	51.803	318114.7	19.775	BB	0.028	190970.50	23.128
14	3.135		0.0000	0.000	283.4	0.018	BB	0.048	98.54	0.012
15	3.288		0.0000	0.000	6373.3	0.396	BB	0.031	3475.09	0.421
16	3.384		0.0000	0.000	1859.1	0.116	BB	0.042	741.13	0.090
17	3.473		0.0000	0.000	831.8	0.052	BB	0.029	475.63	0.058
18	3.608		0.0000	0.000	244.9	0.015	BB	0.042	97.51	0.012
19	3.694		0.0000	0.000	1214.4	0.075	BB	0.036	560.87	0.068
20	3.771		0.0000	0.000	298.8	0.019	BB	0.028	176.52	0.021
21	3.867		0.0000	0.000	1863.8	0.116	BB	0.034	911.97	0.110
22	3.940		0.0000	0.000	792.0	0.049	BB	0.024	551.27	0.067
23	4.030		0.0000	0.000	834.6	0.052	BB	0.040	347.83	0.042
24	4.130		0.0000	0.000	2826.4	0.176	BB	0.037	1287.24	0.156
25	4.183		0.0000	0.000	1007.3	0.063	BB	0.032	524.09	0.063
26	4.264		0.0000	0.000	2301.2	0.143	BB	0.039	989.75	0.120
27	4.360		0.0000	0.000	62217.1	3.868	BB	0.038	27325.98	3.309
28	4.528		0.0000	0.000	25174.6	1.565	BB	0.027	15322.45	1.856
29	4.598		0.0000	0.000	727.1	0.045	BB	0.018	682.19	0.083
30	4.648		0.0000	0.000	8343.0	0.519	BB	0.025	5604.18	0.679
31	4.699		0.0000	0.000	1835.7	0.114	BB	0.031	975.46	0.118
32	4.807		0.0000	0.000	1949.1	0.121	BB	0.034	955.37	0.116
33	4.857		0.0000	0.000	32201.2	2.002	BB	0.020	27272.43	3.303
34	4.885		0.0000	0.000	4054.6	0.252	BB	0.013	5100.53	0.618
35	4.912		0.0000	0.000	3812.3	0.237	BB	0.014	4613.23	0.559
36	5.003		0.0000	0.000	10224.2	0.636	BB	0.021	8015.24	0.971
37	5.057		0.0000	0.000	68832.7	4.279	BB	0.026	43435.94	5.260
38	5.098		0.0000	0.000	14789.3	0.919	BB	0.020	12305.22	1.490
39	5.180		0.0000	0.000	2309.2	0.144	BB	0.023	1689.88	0.205
40	5.233		0.0000	0.000	24846.0	1.545	BB	0.025	16302.62	1.974
41	5.299		0.0000	0.000	41149.1	2.558	BB	0.024	28418.17	3.442
42	5.332		0.0000	0.000	2221.7	0.138	BB	0.011	3380.47	0.409
43	5.374		0.0000	0.000	82261.4	5.114	BB	0.024	56379.27	6.828
44	5.436		0.0000	0.000	4652.6	0.289	BB	0.028	2808.48	0.340
45	5.475		0.0000	0.000	8457.5	0.526	BB	0.021	6796.33	0.823
46	5.522		0.0000	0.000	33220.4	2.065	BB	0.028	20106.92	2.435
47	5.587		0.0000	0.000	1949.7	0.121	BB	0.016	1971.65	0.239
48	5.638		0.0000	0.000	69596.8	4.326	BB	0.028	41886.26	5.073
49	5.703		0.0000	0.000	5075.4	0.316	BB	0.024	3505.94	0.425
50	5.772		0.0000	0.000	67357.8	4.187	BB	0.037	30460.97	3.689
51	5.876		0.0000	0.000	44699.4	2.779	BB	0.026	29106.72	3.525
52	5.922		0.0000	0.000	1178.7	0.073	BB	0.013	1540.18	0.187
53	5.963		0.0000	0.000	4950.5	0.308	BB	0.027	3061.46	0.371
54	6.013		0.0000	0.000	88532.7	5.504	BB	0.036	41458.47	5.021

ASX22 0381

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.136		0.0000	0.000	2860.9	0.178	BB	0.031	1515.23	0.184
56	6.266		0.0000	0.000	29906.9	1.859	BB	0.036	13706.39	1.660
57	6.352		0.0000	0.000	3159.5	0.196	BB	0.031	1690.14	0.205
58	6.447		0.0000	0.000	2812.2	0.175	BB	0.058	802.16	0.097
59	6.514		0.0000	0.000	12867.8	0.800	BB	0.029	7466.32	0.904
60	6.579		0.0000	0.000	7991.9	0.497	BB	0.024	5446.25	0.660
61	6.721		0.0000	0.000	4847.4	0.301	BB	0.066	1228.55	0.149
62	6.837		0.0000	0.000	9314.5	0.579	BB	0.033	4720.77	0.572
63	6.878		0.0000	0.000	673.0	0.042	BB	0.026	433.72	0.053
64	6.963		0.0000	0.000	181.9	0.011	BB	0.026	116.68	0.014
65	7.098		0.0000	0.000	10862.2	0.675	BB	0.069	2614.62	0.317
66	7.226		0.0000	0.000	28558.1	1.775	BB	0.176	2706.33	0.328
67	7.506		0.0000	0.000	13738.1	0.854	BB	0.080	2877.78	0.349
68	7.659		0.0000	0.000	9715.7	0.604	BB	0.152	1063.12	0.129
69	7.731		0.0000	0.000	2632.6	0.164	BB	0.035	1244.40	0.151
70	7.873		0.0000	0.000	736.6	0.046	BB	0.047	261.61	0.032
71	7.977	DCB	39.0804	48.197	215913.2	13.422	BB	0.036	98793.05	11.964
72	8.143		0.0000	0.000	3399.1	0.211	BB	0.181	312.38	0.038
73	8.189		0.0000	0.000	368.2	0.023	BB	0.055	112.35	0.014
74	8.334		0.0000	0.000	115.8	0.007	BB	0.039	49.63	0.006
75	8.583		0.0000	0.000	2435.0	0.151	BB	0.048	837.07	0.101
76	8.703		0.0000	0.000	1901.6	0.118	BB	0.062	509.56	0.062

Total Area = 1608634.0, Total Amount = 81.085, Total Height = 825724.9, Sample Units = ug/l

ASX22 0382

Sample Name: **AR5430924B AAAR543AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 06:44:42pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.19R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.585		0.0000	0.000	6881.6	0.385	BV	0.049	2328.07	0.315
2	1.639		0.0000	0.000	9343.7	0.523	VV	0.059	2623.48	0.355
3	1.703		0.0000	0.000	7282.9	0.408	VB	0.090	1345.12	0.182
4	1.982		0.0000	0.000	1594.4	0.089	BV	0.107	249.27	0.034
5	2.017		0.0000	0.000	1347.8	0.075	VB	0.049	454.98	0.062
6	2.107		0.0000	0.000	1609.4	0.090	BB	0.235	113.94	0.015
7	2.341	TCX	39.4480	47.724	403009.0	22.559	BB	0.039	171079.40	23.159
8	2.623		0.0000	0.000	4625.4	0.259	BB	0.036	2135.68	0.289
9	2.713		0.0000	0.000	288.4	0.016	BB	0.025	190.84	0.026
10	2.769		0.0000	0.000	275.8	0.015	BB	0.018	250.54	0.034
11	3.007		0.0000	0.000	1685.5	0.094	BB	0.040	710.27	0.096
12	3.097		0.0000	0.000	103.4	0.006	BB	0.028	61.87	0.008
13	3.167		0.0000	0.000	2273.5	0.127	BB	0.052	722.31	0.098
14	3.270		0.0000	0.000	437.0	0.024	BB	0.028	264.37	0.036
15	3.317		0.0000	0.000	18.8	0.001	N	-0.010	- 30.46	0.004
16	3.416		0.0000	0.000	5410.3	0.303	BB	0.073	1241.94	0.168
17	3.557		0.0000	0.000	731.3	0.041	BB	0.030	400.98	0.054
18	3.620		0.0000	0.000	78167.9	4.376	BB	0.053	24357.40	3.297
19	3.796		0.0000	0.000	35313.9	1.977	BB	0.044	13365.61	1.809
20	3.923		0.0000	0.000	9722.1	0.544	BB	0.040	4062.75	0.550
21	4.036		0.0000	0.000	724.2	0.041	BB	0.029	411.86	0.056
22	4.111		0.0000	0.000	122179.4	6.839	BB	0.045	45423.32	6.149
23	4.202		0.0000	0.000	5208.1	0.292	BB	0.020	4250.52	0.575
24	4.252		0.0000	0.000	34301.8	1.920	BB	0.025	23106.87	3.128
25	4.291		0.0000	0.000	29341.7	1.642	BB	0.028	17498.57	2.369
26	4.385		0.0000	0.000	1604.0	0.090	BB	0.020	1333.74	0.181
27	4.433		0.0000	0.000	27461.8	1.537	BB	0.033	13906.77	1.883
28	4.506		0.0000	0.000	57888.2	3.240	BB	0.036	26449.98	3.581
29	4.583		0.0000	0.000	117913.0	6.600	BB	0.039	50646.68	6.856
30	4.694		0.0000	0.000	60699.5	3.398	BB	0.035	28670.05	3.881
31	4.794		0.0000	0.000	80070.2	4.482	BB	0.041	32769.92	4.436
32	4.875		0.0000	0.000	47015.8	2.632	BB	0.036	21867.43	2.960
33	4.979		0.0000	0.000	44914.6	2.514	BB	0.045	16782.24	2.272
34	5.049		0.0000	0.000	26265.4	1.470	BB	0.028	15359.88	2.079
35	5.091		0.0000	0.000	1470.3	0.082	BB	0.017	1410.59	0.191
36	5.144		0.0000	0.000	123898.8	6.936	BB	0.047	44051.09	5.963
37	5.417		0.0000	0.000	39229.4	2.196	BB	0.047	13993.58	1.894
38	5.512		0.0000	0.000	770.8	0.043	BB	0.021	608.19	0.082
39	5.557		0.0000	0.000	376.5	0.021	BB	0.015	409.79	0.055
40	5.627		0.0000	0.000	32776.4	1.835	BB	0.060	9059.73	1.226
41	5.864		0.0000	0.000	226.7	0.013	BB	0.020	191.99	0.026
42	5.920		0.0000	0.000	12013.3	0.672	BB	0.036	5514.90	0.747
43	6.023		0.0000	0.000	669.4	0.037	BB	0.044	254.30	0.034
44	6.100		0.0000	0.000	644.5	0.036	BB	0.040	270.78	0.037
45	6.176		0.0000	0.000	6335.5	0.355	BB	0.225	469.08	0.063
46	6.390		0.0000	0.000	3283.5	0.184	BB	0.077	710.37	0.096
47	6.530		0.0000	0.000	689.0	0.039	BB	0.036	322.45	0.044
48	6.625		0.0000	0.000	4221.8	0.236	BB	0.040	1746.78	0.236
49	6.801	DCB	43.2104	52.276	321658.6	18.006	BB	0.040	134217.00	18.169
50	7.510		0.0000	0.000	873.1	0.049	BB	0.047	308.03	0.042
51	7.627		0.0000	0.000	10716.6	0.600	BB	0.332	538.44	0.073
52	8.197		0.0000	0.000	407.7	0.023	BB	0.051	133.27	0.018
53	9.067		0.0000	0.000	465.5	0.026	BB	0.075	103.50	0.014

ASX22 0383

Total Area = 1786437.0, Total Amount = 82.658, Total Height = 738720.1, Sample Units = ug/l

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.19R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 6:35:33 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.19R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 6:35:33 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	190970	42.00452	TCX	2.341	171079	39.44796	TCX
7.977	98793	39.08038	DCB	6.801	134217	43.21037	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.19A
Area File: C:\CPWIN\DATA1\1P7307B.19A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 6:44:50 PM
File Reported On: 11/3/2009 at 6:45:02 PM

10X REPLOTS

ASX22 0386

C:\CPWIN\DATA1\1P7307.19R



C:\CPWIN\DATA1\1P7307B.19R



ASX22 0387

Sample Name: **AR5440924B AAAR544AA ICAL 0930699999**
Acquired from CP07–H5088A via port 3 on 11/3/09 06:57:42pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.20R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.580		0.0000	0.000	3920.8	0.155	BV	0.058	1119.98	0.081
2	1.792		0.0000	0.000	34982.6	1.384	VV	0.148	3936.78	0.285
3	1.863		0.0000	0.000	18914.5	0.748	VV	0.066	4784.90	0.347
4	1.926		0.0000	0.000	18593.5	0.735	VB	0.055	5663.87	0.410
5	2.032		0.0000	0.000	24184.4	0.957	BV	0.075	5406.04	0.392
6	2.073		0.0000	0.000	27199.6	1.076	VV	0.067	6724.46	0.487
7	2.175		0.0000	0.000	16384.8	0.648	VV	0.058	4717.64	0.342
8	2.208		0.0000	0.000	25250.8	0.999	VV	0.092	4581.94	0.332
9	2.373		0.0000	0.000	7739.4	0.306	VV	0.071	1805.59	0.131
10	2.448		0.0000	0.000	4885.6	0.193	VV	0.071	1145.39	0.083
11	2.527		0.0000	0.000	2556.2	0.101	VB	0.062	683.11	0.049
12	2.771		0.0000	0.000	3914.6	0.155	BB	0.075	868.01	0.063
13	2.890		0.0000	0.000	4472.0	0.177	BB	0.065	1153.45	0.084
14	2.997	TCX	64.0252	54.222	452615.6	17.904	BB	0.027	280883.10	20.342
15	3.157		0.0000	0.000	563.9	0.022	BB	0.049	190.47	0.014
16	3.287		0.0000	0.000	8004.1	0.317	BB	0.027	4922.05	0.356
17	3.381		0.0000	0.000	1180.7	0.047	BB	0.025	793.56	0.057
18	3.472		0.0000	0.000	2019.5	0.080	BB	0.032	1057.48	0.077
19	3.606		0.0000	0.000	1269.1	0.050	BB	0.148	143.20	0.010
20	3.693		0.0000	0.000	2734.0	0.108	BB	0.039	1180.28	0.085
21	3.768		0.0000	0.000	480.9	0.019	BB	0.029	281.19	0.020
22	3.823		0.0000	0.000	53.7	0.002	BB	0.016	57.56	0.004
23	3.862		0.0000	0.000	1786.9	0.071	BB	0.028	1081.18	0.078
24	3.939		0.0000	0.000	1064.8	0.042	BB	0.023	774.75	0.056
25	4.027		0.0000	0.000	937.5	0.037	BB	0.035	445.87	0.032
26	4.127		0.0000	0.000	8198.3	0.324	BB	0.044	3115.67	0.226
27	4.180		0.0000	0.000	1446.9	0.057	BB	0.025	982.34	0.071
28	4.263		0.0000	0.000	2213.0	0.088	BB	0.028	1304.86	0.094
29	4.358		0.0000	0.000	111642.7	4.416	BB	0.036	51270.24	3.713
30	4.526		0.0000	0.000	49592.2	1.962	BB	0.028	29882.79	2.164
31	4.598		0.0000	0.000	1800.8	0.071	BB	0.019	1585.46	0.115
32	4.646		0.0000	0.000	16630.3	0.658	BB	0.025	11225.24	0.813
33	4.697		0.0000	0.000	2509.0	0.099	BB	0.024	1732.06	0.125
34	4.805		0.0000	0.000	2781.3	0.110	BB	0.026	1775.66	0.129
35	4.854		0.0000	0.000	64921.1	2.568	BB	0.020	54647.31	3.958
36	4.884		0.0000	0.000	22873.2	0.905	BB	0.030	12536.35	0.908
37	5.001		0.0000	0.000	19131.5	0.757	BB	0.020	15648.25	1.133
38	5.055		0.0000	0.000	128723.7	5.092	BB	0.026	81029.57	5.868
39	5.097		0.0000	0.000	26902.1	1.064	BB	0.019	23768.62	1.721
40	5.177		0.0000	0.000	5839.6	0.231	BB	0.022	4456.03	0.323
41	5.231		0.0000	0.000	48059.1	1.901	BB	0.026	31247.85	2.263
42	5.297		0.0000	0.000	123086.0	4.869	BB	0.032	63863.59	4.625
43	5.373		0.0000	0.000	150428.5	5.950	BB	0.024	103120.40	7.468
44	5.433		0.0000	0.000	9169.9	0.363	BB	0.027	5723.14	0.414
45	5.474		0.0000	0.000	15964.7	0.631	BB	0.020	12991.79	0.941
46	5.520		0.0000	0.000	60644.1	2.399	BB	0.028	36544.14	2.647
47	5.586		0.0000	0.000	4796.1	0.190	BB	0.017	4602.53	0.333
48	5.637		0.0000	0.000	126941.7	5.021	BB	0.027	77110.50	5.584
49	5.702		0.0000	0.000	9113.0	0.360	BB	0.023	6585.84	0.477
50	5.771		0.0000	0.000	123942.1	4.903	BB	0.037	55330.88	4.007
51	5.875		0.0000	0.000	84676.7	3.349	BB	0.025	55838.50	4.044
52	5.920		0.0000	0.000	2100.8	0.083	BB	0.013	2739.49	0.198
53	5.961		0.0000	0.000	8783.1	0.347	BB	0.026	5689.97	0.412
54	6.012		0.0000	0.000	167241.8	6.615	BB	0.036	77284.59	5.597

ASX22 0388

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.134		0.0000	0.000	5214.0	0.206	BB	0.031	2811.68	0.204
56	6.265		0.0000	0.000	56754.8	2.245	BB	0.036	26398.55	1.912
57	6.351		0.0000	0.000	3448.2	0.136	BB	0.022	2612.85	0.189
58	6.399		0.0000	0.000	2048.5	0.081	BB	0.023	1465.66	0.106
59	6.444		0.0000	0.000	354.6	0.014	BB	0.021	276.37	0.020
60	6.513		0.0000	0.000	24878.0	0.984	BB	0.029	14486.39	1.049
61	6.578		0.0000	0.000	14305.6	0.566	BB	0.025	9700.32	0.703
62	6.724		0.0000	0.000	3279.1	0.130	BB	0.066	825.35	0.060
63	6.836		0.0000	0.000	12554.2	0.497	BB	0.027	7836.92	0.568
64	6.877		0.0000	0.000	913.5	0.036	BB	0.020	771.96	0.056
65	6.959		0.0000	0.000	178.0	0.007	BB	0.027	108.21	0.008
66	7.107		0.0000	0.000	6598.4	0.261	BB	0.072	1519.95	0.110
67	7.186		0.0000	0.000	4778.5	0.189	BB	0.069	1159.72	0.084
68	7.520		0.0000	0.000	13344.7	0.528	BB	0.104	2131.46	0.154
69	7.644		0.0000	0.000	4748.1	0.188	BB	0.107	739.89	0.054
70	7.731		0.0000	0.000	3607.7	0.143	BB	0.035	1731.53	0.125
71	7.866		0.0000	0.000	594.4	0.024	BB	0.045	218.51	0.016
72	7.976	DCB	54.0545	45.778	299749.3	11.857	BB	0.037	136646.80	9.896
73	8.091		0.0000	0.000	1720.9	0.068	BB	0.094	303.67	0.022
74	8.381		0.0000	0.000	599.4	0.024	BB	0.065	154.70	0.011
75	8.603		0.0000	0.000	463.3	0.018	BB	0.036	216.05	0.016
76	8.674		0.0000	0.000	2067.3	0.082	BB	0.051	669.79	0.049

Total Area = 2528063.0, Total Amount = 118.08, Total Height = 1380822.0, Sample Units = ug/l

ASX22 0389

Sample Name: **AR5440924B AAAR544AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 06:57:42pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.20R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.534		0.0000	0.000	298.4	0.010	BV	0.022	222.31	0.018
2	1.589		0.0000	0.000	9221.0	0.307	VV	0.056	2732.35	0.217
3	1.636		0.0000	0.000	7285.0	0.243	VV	0.046	2648.15	0.210
4	1.701		0.0000	0.000	7803.7	0.260	VV	0.099	1312.80	0.104
5	2.014		0.0000	0.000	1555.7	0.052	VB	0.044	591.61	0.047
6	2.113		0.0000	0.000	262.1	0.009	BB	0.028	153.66	0.012
7	2.279		0.0000	0.000	814.8	0.027	BB	0.034	402.44	0.032
8	2.339	TCX	59.7397	48.718	605076.6	20.152	BB	0.039	259081.50	20.543
9	2.621		0.0000	0.000	7273.7	0.242	BB	0.036	3372.86	0.267
10	2.714		0.0000	0.000	853.0	0.028	BB	0.033	432.02	0.034
11	2.773		0.0000	0.000	766.9	0.026	BB	0.028	454.90	0.036
12	3.005		0.0000	0.000	4241.9	0.141	BB	0.048	1465.30	0.116
13	3.169		0.0000	0.000	2530.6	0.084	BB	0.041	1027.16	0.081
14	3.267		0.0000	0.000	1132.7	0.038	BB	0.039	487.64	0.039
15	3.352		0.0000	0.000	236.6	0.008	BB	0.034	115.43	0.009
16	3.406		0.0000	0.000	11084.6	0.369	BB	0.076	2418.39	0.192
17	3.554		0.0000	0.000	1342.8	0.045	BB	0.029	765.45	0.061
18	3.618		0.0000	0.000	150357.7	5.008	BB	0.054	46531.94	3.690
19	3.795		0.0000	0.000	68542.2	2.283	BB	0.044	26036.33	2.064
20	3.923		0.0000	0.000	20367.8	0.678	BB	0.041	8205.48	0.651
21	4.034		0.0000	0.000	1459.7	0.049	BB	0.026	918.21	0.073
22	4.109		0.0000	0.000	229018.1	7.628	BB	0.045	85112.61	6.749
23	4.200		0.0000	0.000	9821.4	0.327	BB	0.020	8081.13	0.641
24	4.250		0.0000	0.000	65553.0	2.183	BB	0.023	46718.90	3.704
25	4.289		0.0000	0.000	51655.8	1.720	BB	0.026	33156.39	2.629
26	4.384		0.0000	0.000	3194.2	0.106	BB	0.020	2609.67	0.207
27	4.431		0.0000	0.000	51528.7	1.716	BB	0.032	26425.72	2.095
28	4.504		0.0000	0.000	111147.9	3.702	BB	0.036	51051.69	4.048
29	4.581		0.0000	0.000	217543.8	7.245	BB	0.039	94029.50	7.456
30	4.693		0.0000	0.000	111079.8	3.700	BB	0.035	53222.17	4.220
31	4.793		0.0000	0.000	146945.7	4.894	BB	0.040	60998.71	4.837
32	4.873		0.0000	0.000	86501.6	2.881	BB	0.035	40655.18	3.224
33	4.977		0.0000	0.000	85980.5	2.864	BB	0.044	32219.37	2.555
34	5.048		0.0000	0.000	50375.3	1.678	BB	0.028	30113.70	2.388
35	5.089		0.0000	0.000	2725.3	0.091	BB	0.016	2863.82	0.227
36	5.143		0.0000	0.000	231724.6	7.718	BB	0.047	82180.04	6.516
37	5.417		0.0000	0.000	75936.1	2.529	BB	0.047	26864.99	2.130
38	5.511		0.0000	0.000	1365.4	0.045	BB	0.021	1078.72	0.086
39	5.559		0.0000	0.000	337.8	0.011	BB	0.015	387.86	0.031
40	5.624		0.0000	0.000	61521.5	2.049	BB	0.059	17268.29	1.369
41	5.803		0.0000	0.000	1747.7	0.058	BB	0.031	938.63	0.074
42	5.866		0.0000	0.000	253.8	0.008	BB	0.018	229.00	0.018
43	5.921		0.0000	0.000	27338.1	0.911	BB	0.043	10477.79	0.831
44	6.174		0.0000	0.000	749.0	0.025	BB	0.032	394.94	0.031
45	6.243		0.0000	0.000	3052.2	0.102	BB	0.072	709.01	0.056
46	6.390		0.0000	0.000	1581.3	0.053	BB	0.046	570.67	0.045
47	6.453		0.0000	0.000	61.2	0.002	BB	0.021	47.45	0.004
48	6.523		0.0000	0.000	629.7	0.021	BB	0.038	279.53	0.022
49	6.623		0.0000	0.000	5789.7	0.193	BB	0.044	2206.43	0.175
50	6.801	DCB	62.8845	51.282	453529.3	15.105	BB	0.040	189612.10	15.034
51	7.500		0.0000	0.000	1327.4	0.044	BB	0.074	298.65	0.024
52	7.633		0.0000	0.000	6090.4	0.203	BB	0.211	480.96	0.038
53	7.983		0.0000	0.000	3031.6	0.101	BB	0.149	340.08	0.027
54	8.333		0.0000	0.000	866.2	0.029	BB	0.077	188.09	0.015

ASX22 8398

Total Area = 3002512.0, Total Amount = 122.624, Total Height = 1261188.0, Sample Units = ug/l

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.20R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 6:48:33 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.20R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 6:48:33 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A
2.997	280883	64.02521	TCX
7.976	136647	54.0545	DCB

RT B	Height B	Amount B	Compound B
2.339	259082	59.73975	TCX
6.801	189612	62.88453	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.20A
Area File: C:\CPWIN\DATA1\1P7307B.20A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 6:57:50 PM
File Reported On: 11/3/2009 at 6:58:02 PM

10X REPLOTS

ASX22 2393

C:\CPWIN\DATA1\1P7307.20R



C:\CPWIN\DATA1\1P7307B.20R



ASX22 0394

Sample Name: **AR5450924B AAAR545AA ICAL 0930699999**
Acquired from CP07–H5088A via port 3 on 11/3/09 07:10:43pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.21R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.621		0.0000	0.000	2928.9	0.077	BV	0.076	642.26	0.028
2	1.799		0.0000	0.000	7153.8	0.187	VV	0.111	1078.67	0.047
3	1.867		0.0000	0.000	5179.0	0.135	VV	0.064	1357.21	0.060
4	1.931		0.0000	0.000	5777.8	0.151	VB	0.051	1875.86	0.082
5	2.073		0.0000	0.000	21984.5	0.574	BV	0.084	4353.47	0.191
6	2.173		0.0000	0.000	7640.8	0.200	VV	0.051	2500.57	0.110
7	2.209		0.0000	0.000	9739.6	0.254	VV	0.062	2600.35	0.114
8	2.321		0.0000	0.000	2131.4	0.056	VV	0.044	810.90	0.036
9	2.370		0.0000	0.000	2266.4	0.059	VV	0.060	632.43	0.028
10	2.469		0.0000	0.000	1884.1	0.049	VV	0.068	461.64	0.020
11	2.526		0.0000	0.000	1676.2	0.044	VV	0.045	619.27	0.027
12	2.620		0.0000	0.000	254.3	0.007	VB	0.030	142.97	0.006
13	2.708		0.0000	0.000	308.3	0.008	BB	0.044	116.27	0.005
14	2.787		0.0000	0.000	481.9	0.013	BB	0.039	204.51	0.009
15	2.863		0.0000	0.000	2017.8	0.053	BB	0.070	481.06	0.021
16	2.997	TCX	81.5228	53.821	569176.9	14.871	BB	0.027	352327.60	15.491
17	3.224		0.0000	0.000	452.5	0.012	BB	0.029	259.98	0.011
18	3.287		0.0000	0.000	10249.2	0.268	BB	0.027	6351.19	0.279
19	3.380		0.0000	0.000	2562.1	0.067	BB	0.026	1614.44	0.071
20	3.472		0.0000	0.000	1198.1	0.031	BB	0.022	920.90	0.040
21	3.513		0.0000	0.000	237.3	0.006	BB	0.018	216.27	0.010
22	3.600		0.0000	0.000	564.9	0.015	BB	0.050	187.34	0.008
23	3.693		0.0000	0.000	5895.9	0.154	BB	0.039	2491.06	0.110
24	3.770		0.0000	0.000	866.7	0.023	BB	0.028	519.54	0.023
25	3.862		0.0000	0.000	4013.3	0.105	BB	0.028	2369.71	0.104
26	3.939		0.0000	0.000	1411.3	0.037	BB	0.024	996.36	0.044
27	4.027		0.0000	0.000	1645.2	0.043	BB	0.037	746.15	0.033
28	4.127		0.0000	0.000	13912.9	0.364	BB	0.037	6267.38	0.276
29	4.180		0.0000	0.000	2940.5	0.077	BB	0.024	2039.59	0.090
30	4.263		0.0000	0.000	4084.3	0.107	BB	0.028	2439.39	0.107
31	4.316		0.0000	0.000	211.0	0.006	BB	0.016	213.80	0.009
32	4.358		0.0000	0.000	94642.7	2.473	BB	0.021	76193.77	3.350
33	4.389		0.0000	0.000	6557.5	0.171	BB	0.013	8517.98	0.375
34	4.526		0.0000	0.000	92621.5	2.420	BB	0.028	55694.98	2.449
35	4.598		0.0000	0.000	3509.7	0.092	BB	0.019	3099.02	0.136
36	4.646		0.0000	0.000	31410.9	0.821	BB	0.024	21383.91	0.940
37	4.697		0.0000	0.000	4074.4	0.106	BB	0.021	3203.58	0.141
38	4.805		0.0000	0.000	4970.5	0.130	BB	0.025	3290.63	0.145
39	4.854		0.0000	0.000	116222.4	3.037	BB	0.019	99638.84	4.381
40	4.883		0.0000	0.000	15849.3	0.414	BB	0.012	21911.75	0.963
41	4.910		0.0000	0.000	13814.3	0.361	BB	0.015	15658.13	0.688
42	5.001		0.0000	0.000	34753.3	0.908	BB	0.020	28495.13	1.253
43	5.055		0.0000	0.000	228481.5	5.970	BB	0.026	145932.80	6.416
44	5.096		0.0000	0.000	52274.2	1.366	BB	0.019	45697.12	2.009
45	5.177		0.0000	0.000	12778.4	0.334	BB	0.022	9755.83	0.429
46	5.230		0.0000	0.000	89644.7	2.342	BB	0.025	59209.21	2.603
47	5.297		0.0000	0.000	227920.9	5.955	BB	0.032	118353.70	5.204
48	5.372		0.0000	0.000	276780.7	7.232	BB	0.024	191746.20	8.431
49	5.432		0.0000	0.000	18499.7	0.483	BB	0.027	11357.35	0.499
50	5.473		0.0000	0.000	30865.6	0.806	BB	0.021	25006.56	1.099
51	5.520		0.0000	0.000	113222.5	2.958	BB	0.027	68777.13	3.024
52	5.587		0.0000	0.000	9951.9	0.260	BB	0.018	9384.68	0.413
53	5.637		0.0000	0.000	235864.5	6.163	BB	0.027	143641.40	6.316
54	5.701		0.0000	0.000	18228.4	0.476	BB	0.023	13154.66	0.578

ASX22 0395

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.770		0.0000	0.000	228900.9	5.981	BB	0.037	103123.50	4.534
56	5.874		0.0000	0.000	162324.2	4.241	BB	0.025	108676.30	4.778
57	5.920		0.0000	0.000	4613.3	0.121	BB	0.013	5740.05	0.252
58	5.960		0.0000	0.000	17163.2	0.448	BB	0.026	11026.37	0.485
59	6.011		0.0000	0.000	321606.6	8.403	BB	0.036	149743.90	6.584
60	6.134		0.0000	0.000	10491.0	0.274	BB	0.030	5750.69	0.253
61	6.264		0.0000	0.000	111040.5	2.901	BB	0.036	50983.61	2.242
62	6.350		0.0000	0.000	6871.4	0.180	BB	0.022	5113.08	0.225
63	6.399		0.0000	0.000	5055.1	0.132	BB	0.023	3681.60	0.162
64	6.443		0.0000	0.000	450.2	0.012	BB	0.017	430.25	0.019
65	6.511		0.0000	0.000	50050.1	1.308	BB	0.029	29229.00	1.285
66	6.577		0.0000	0.000	28466.6	0.744	BB	0.024	19628.04	0.863
67	6.835		0.0000	0.000	24020.2	0.628	BB	0.025	16107.56	0.708
68	6.876		0.0000	0.000	1910.6	0.050	BB	0.019	1678.11	0.074
69	7.091		0.0000	0.000	7339.5	0.192	BB	0.079	1547.46	0.068
70	7.179		0.0000	0.000	2648.1	0.069	BB	0.046	968.38	0.043
71	7.292		0.0000	0.000	10853.4	0.284	BB	0.164	1103.52	0.049
72	7.518		0.0000	0.000	13458.2	0.352	BB	0.106	2119.29	0.093
73	7.650		0.0000	0.000	4514.7	0.118	BB	0.119	633.04	0.028
74	7.730		0.0000	0.000	4806.4	0.126	BB	0.035	2276.49	0.100
75	7.872		0.0000	0.000	386.7	0.010	BB	0.035	181.96	0.008
76	7.975	DCB	69.9479	46.179	382928.0	10.005	BB	0.036	176824.40	7.775
77	8.086		0.0000	0.000	642.4	0.017	BB	0.047	230.18	0.010
78	8.376		0.0000	0.000	2229.8	0.058	BB	0.115	321.77	0.014
79	8.600		0.0000	0.000	792.6	0.021	BB	0.039	338.23	0.015

Total Area = 3827351.0, Total Amount = 151.471, Total Height = 2274399.0, Sample Units = ug/l

ASX22 0396

Sample Name: **AR5450924B AAAR545AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 07:10:43pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.21R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.585		0.0000	0.000	11265.7	0.222	BV	0.057	3288.76	0.154
2	1.634		0.0000	0.000	7086.6	0.140	VV	0.043	2740.75	0.128
3	1.703		0.0000	0.000	7334.0	0.145	VB	0.083	1466.68	0.069
4	2.109		0.0000	0.000	125.3	0.002	BB	0.021	98.01	0.005
5	2.339	TCX	76.1464	47.998	777905.7	15.336	BB	0.039	330234.60	15.471
6	2.621		0.0000	0.000	9469.5	0.187	BB	0.036	4371.35	0.205
7	2.716		0.0000	0.000	1575.3	0.031	BB	0.032	811.49	0.038
8	2.773		0.0000	0.000	767.1	0.015	BB	0.025	513.87	0.024
9	3.006		0.0000	0.000	8142.9	0.161	BB	0.048	2837.98	0.133
10	3.169		0.0000	0.000	5275.3	0.104	BB	0.042	2109.69	0.099
11	3.268		0.0000	0.000	2826.7	0.056	BB	0.065	722.12	0.034
12	3.416		0.0000	0.000	22972.0	0.453	BB	0.077	4975.79	0.233
13	3.555		0.0000	0.000	2557.4	0.050	BB	0.029	1467.65	0.069
14	3.617		0.0000	0.000	279254.7	5.505	BB	0.054	86736.97	4.063
15	3.794		0.0000	0.000	130684.4	2.576	BB	0.044	50018.70	2.343
16	3.922		0.0000	0.000	41525.0	0.819	BB	0.042	16469.06	0.772
17	4.035		0.0000	0.000	2842.2	0.056	BB	0.028	1676.07	0.079
18	4.108		0.0000	0.000	430124.0	8.480	BB	0.044	163561.90	7.663
19	4.200		0.0000	0.000	20007.9	0.394	BB	0.020	16394.72	0.768
20	4.250		0.0000	0.000	124510.2	2.455	BB	0.023	89380.04	4.187
21	4.288		0.0000	0.000	95519.6	1.883	BB	0.027	58958.84	2.762
22	4.384		0.0000	0.000	6505.0	0.128	BB	0.020	5302.61	0.248
23	4.431		0.0000	0.000	97334.5	1.919	BB	0.032	50820.07	2.381
24	4.503		0.0000	0.000	218196.8	4.302	BB	0.036	100055.60	4.687
25	4.580		0.0000	0.000	413258.1	8.147	BB	0.038	183308.40	8.588
26	4.692		0.0000	0.000	207069.7	4.082	BB	0.034	101874.10	4.773
27	4.792		0.0000	0.000	281121.2	5.542	BB	0.039	120600.80	5.650
28	4.872		0.0000	0.000	164034.7	3.234	BB	0.035	78714.75	3.688
29	4.979		0.0000	0.000	164403.6	3.241	BB	0.045	60971.09	2.856
30	5.047		0.0000	0.000	155657.7	3.069	BB	0.037	69464.57	3.254
31	5.142		0.0000	0.000	442127.6	8.716	BB	0.046	158513.70	7.426
32	5.416		0.0000	0.000	149909.4	2.955	BB	0.047	53503.88	2.507
33	5.511		0.0000	0.000	2855.2	0.056	BB	0.021	2234.65	0.105
34	5.559		0.0000	0.000	632.2	0.012	BB	0.015	715.71	0.034
35	5.624		0.0000	0.000	126769.8	2.499	BB	0.060	35252.56	1.652
36	5.799		0.0000	0.000	852.1	0.017	BB	0.027	520.89	0.024
37	5.867		0.0000	0.000	583.5	0.012	BB	0.022	451.85	0.021
38	5.921		0.0000	0.000	55575.7	1.096	BB	0.044	21226.62	0.994
39	6.179		0.0000	0.000	1684.7	0.033	BB	0.034	829.70	0.039
40	6.238		0.0000	0.000	4176.4	0.082	BB	0.070	993.12	0.047
41	6.389		0.0000	0.000	5741.4	0.113	BB	0.066	1455.26	0.068
42	6.522		0.0000	0.000	461.4	0.009	BB	0.036	215.88	0.010
43	6.622		0.0000	0.000	7112.8	0.140	BB	0.041	2910.35	0.136
44	6.801	DCB	82.4976	52.002	578803.3	11.411	BB	0.039	244835.00	11.470
45	7.643		0.0000	0.000	2949.9	0.058	BB	0.121	407.64	0.019
46	7.817		0.0000	0.000	411.2	0.008	BB	0.062	110.40	0.005
47	7.977		0.0000	0.000	321.0	0.006	BB	0.036	149.49	0.007
48	8.047		0.0000	0.000	874.8	0.017	BB	0.117	125.03	0.006
49	8.290		0.0000	0.000	1136.8	0.022	BB	0.128	148.12	0.007

Total Area = 5072332.0, Total Amount = 158.644, Total Height = 2134547.0, Sample Units = ug/l

ASX22 0397

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.21R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 7:01:34 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.21R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 7:01:34 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A
2.997	352328	81.52281	TCX
7.975	176824	69.94788	DCB

RT B	Height B	Amount B	Compound B
2.339	330235	76.14642	TCX
6.801	244835	82.49756	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.21A
Area File: C:\CPWIN\DATA1\1P7307B.21A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 7:10:50 PM
File Reported On: 11/3/2009 at 7:11:02 PM

10X REPLOTS

ASX22 0399

C:\CPWIN\DATA1\1P7307.21R



C:\CPWIN\DATA1\1P7307B.21R



ASX22 0489

Sample Name: **AR6210924B AAAR621AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 07:23:43pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.22R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.582		0.0000	0.000	4452.0	0.879	BV	0.059	1256.49	0.632
2	1.663		0.0000	0.000	7943.1	1.568	VV	0.057	2336.26	1.176
3	1.715		0.0000	0.000	9161.5	1.809	VV	0.050	3065.98	1.543
4	1.798		0.0000	0.000	19162.3	3.783	VV	0.076	4218.91	2.123
5	1.866		0.0000	0.000	19393.6	3.829	VV	0.063	5161.81	2.598
6	1.926		0.0000	0.000	21552.7	4.255	VB	0.058	6205.08	3.123
7	2.036		0.0000	0.000	35785.9	7.065	BV	0.099	6041.07	3.040
8	2.173		0.0000	0.000	56300.1	11.114	VV	0.128	7302.65	3.675
9	2.372		0.0000	0.000	9348.7	1.846	VV	0.069	2266.23	1.140
10	2.445		0.0000	0.000	8494.6	1.677	VV	0.085	1661.85	0.836
11	2.585		0.0000	0.000	3045.5	0.601	VB	0.166	305.45	0.154
12	2.777		0.0000	0.000	7488.0	1.478	BB	0.087	1437.16	0.723
13	2.885		0.0000	0.000	8022.1	1.584	BB	0.061	2207.22	1.111
14	3.020	TCX	-4.6608	111.438	6373.9	1.258	BB	0.246	431.19	0.217
15	3.062		0.0000	0.000	263.8	0.052	BB	0.041	107.18	0.054
16	3.209		0.0000	0.000	541.0	0.107	BB	0.071	126.53	0.064
17	3.334		0.0000	0.000	730.3	0.144	BB	0.023	520.38	0.262
18	3.388		0.0000	0.000	627.8	0.124	BB	0.028	376.86	0.190
19	3.529		0.0000	0.000	1200.0	0.237	BB	0.219	91.16	0.046
20	3.609		0.0000	0.000	484.2	0.096	BB	0.074	108.46	0.055
21	3.720		0.0000	0.000	405.6	0.080	BB	0.042	160.88	0.081
22	3.773		0.0000	0.000	187.2	0.037	BB	0.036	86.71	0.044
23	3.830		0.0000	0.000	68.4	0.014	BB	0.018	62.92	0.032
24	3.862		0.0000	0.000	362.8	0.072	BB	0.027	226.40	0.114
25	4.027		0.0000	0.000	1606.9	0.317	BB	0.065	409.08	0.206
26	4.125		0.0000	0.000	738.5	0.146	BB	0.044	282.35	0.142
27	4.182		0.0000	0.000	604.2	0.119	BB	0.042	241.27	0.121
28	4.266		0.0000	0.000	1236.5	0.244	BB	0.029	701.91	0.353
29	4.359		0.0000	0.000	207.6	0.041	BB	0.019	178.52	0.090
30	4.400		0.0000	0.000	1028.9	0.203	BB	0.067	257.40	0.130
31	4.526		0.0000	0.000	177.3	0.035	BB	0.019	156.90	0.079
32	4.655		0.0000	0.000	434.0	0.086	BB	0.038	192.47	0.097
33	4.810		0.0000	0.000	563.9	0.111	BB	0.041	231.67	0.117
34	4.854		0.0000	0.000	1535.3	0.303	BB	0.020	1248.69	0.628
35	4.919		0.0000	0.000	717.6	0.142	BB	0.025	476.14	0.240
36	4.997		0.0000	0.000	1002.1	0.198	BB	0.033	500.56	0.252
37	5.058		0.0000	0.000	2001.7	0.395	BB	0.022	1497.52	0.754
38	5.105		0.0000	0.000	2737.2	0.540	BB	0.032	1426.39	0.718
39	5.203		0.0000	0.000	1970.2	0.389	BB	0.043	763.96	0.384
40	5.310		0.0000	0.000	3180.7	0.628	BB	0.029	1797.63	0.905
41	5.376		0.0000	0.000	1212.0	0.239	BB	0.024	846.51	0.426
42	5.427		0.0000	0.000	13079.2	2.582	BB	0.033	6528.04	3.285
43	5.521		0.0000	0.000	16240.6	3.206	BB	0.027	10027.12	5.046
44	5.635		0.0000	0.000	1179.7	0.233	BB	0.033	588.55	0.296
45	5.703		0.0000	0.000	1534.5	0.303	BB	0.026	994.49	0.500
46	5.766		0.0000	0.000	16594.3	3.276	BB	0.027	10342.07	5.205
47	5.813		0.0000	0.000	4998.8	0.987	BB	0.021	3887.34	1.956
48	5.877		0.0000	0.000	8171.5	1.613	BB	0.031	4387.32	2.208
49	5.969		0.0000	0.000	334.9	0.066	BB	0.026	215.01	0.108
50	6.010		0.0000	0.000	13799.2	2.724	BB	0.030	7627.69	3.839
51	6.083		0.0000	0.000	20710.0	4.088	BB	0.025	13849.49	6.970
52	6.134		0.0000	0.000	6886.0	1.359	BB	0.024	4853.77	2.443
53	6.250		0.0000	0.000	796.0	0.157	BB	0.032	417.85	0.210
54	6.281		0.0000	0.000	17517.5	3.458	BB	0.027	10666.57	5.368

ASX22 8481

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.352		0.0000	0.000	7168.5	1.415	BB	0.028	4342.07	2.185
56	6.397		0.0000	0.000	618.9	0.122	BB	0.018	584.45	0.294
57	6.445		0.0000	0.000	513.5	0.101	BB	0.024	353.78	0.178
58	6.508		0.0000	0.000	2249.3	0.444	BB	0.025	1505.89	0.758
59	6.579		0.0000	0.000	43640.6	8.615	BB	0.030	24216.40	12.187
60	6.728		0.0000	0.000	2694.3	0.532	BB	0.060	747.62	0.376
61	6.828		0.0000	0.000	23468.2	4.633	BB	0.031	12636.00	6.359
62	6.895		0.0000	0.000	16956.3	3.347	BB	0.036	7836.37	3.944
63	7.056		0.0000	0.000	1036.4	0.205	BB	0.028	616.10	0.310
64	7.183		0.0000	0.000	8256.7	1.630	BB	0.037	3737.86	1.881
65	7.420		0.0000	0.000	4378.3	0.864	BB	0.025	2970.23	1.495
66	7.497		0.0000	0.000	5952.7	1.175	BB	0.072	1374.09	0.692
67	7.649		0.0000	0.000	4597.8	0.908	BB	0.118	651.01	0.328
68	7.731		0.0000	0.000	6542.2	1.292	BB	0.035	3129.24	1.575
69	7.863		0.0000	0.000	396.1	0.078	BB	0.044	150.50	0.076
70	7.972	DCB	0.4784	-11.438	5779.9	1.141	BB	0.080	1209.35	0.609
71	8.198		0.0000	0.000	4979.5	0.983	BB	0.262	316.61	0.159
72	8.381		0.0000	0.000	1218.3	0.241	BB	0.076	267.43	0.135
73	8.500		0.0000	0.000	326.9	0.065	BB	0.058	93.49	0.047
74	8.598		0.0000	0.000	926.5	0.183	BB	0.041	376.65	0.190
75	8.668		0.0000	0.000	655.6	0.129	BB	0.047	232.78	0.117

Total Area = 506548.3, Total Amount = -4.182, Total Height = 198707.0, Sample Units = ug/l

ASX22 0482

Sample Name: **AR6210924B AAAR621AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 07:23:43pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.22R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.568		0.0000	0.000	7794.4	1.911	BV	0.045	2874.20	1.831
2	1.637		0.0000	0.000	16737.0	4.104	VV	0.109	2565.14	1.635
3	1.950		0.0000	0.000	2194.6	0.538	VV	0.065	566.50	0.361
4	2.014		0.0000	0.000	3193.6	0.783	VB	0.055	961.77	0.613
5	2.187		0.0000	0.000	1214.4	0.298	BB	0.103	196.71	0.125
6	2.277		0.0000	0.000	1920.5	0.471	BB	0.066	484.56	0.309
7	2.494		0.0000	0.000	572.1	0.140	BB	0.070	135.65	0.086
8	3.004		0.0000	0.000	486.4	0.119	BB	0.036	227.54	0.145
9	3.163		0.0000	0.000	1948.5	0.478	BB	0.054	601.10	0.383
10	3.353		0.0000	0.000	245.6	0.060	BB	0.029	142.10	0.091
11	3.400		0.0000	0.000	920.4	0.226	BB	0.049	316.03	0.201
12	3.559		0.0000	0.000	4166.6	1.022	BB	0.103	673.37	0.429
13	3.797		0.0000	0.000	1408.1	0.345	BB	0.146	160.62	0.102
14	4.109		0.0000	0.000	4623.8	1.134	BB	0.045	1709.95	1.090
15	4.256		0.0000	0.000	1828.9	0.448	BB	0.032	955.01	0.609
16	4.299		0.0000	0.000	826.8	0.203	BB	0.034	400.54	0.255
17	4.507		0.0000	0.000	142.8	0.035	BB	0.036	66.60	0.042
18	4.543		0.0000	0.000	1942.9	0.476	BB	0.027	1212.86	0.773
19	4.599		0.0000	0.000	16143.9	3.959	BB	0.045	5989.06	3.816
20	4.694		0.0000	0.000	21799.6	5.346	BB	0.037	9850.76	6.277
21	4.825		0.0000	0.000	1389.7	0.341	BB	0.032	721.57	0.460
22	4.875		0.0000	0.000	24244.5	5.945	BB	0.038	10603.59	6.757
23	4.979		0.0000	0.000	25596.7	6.277	BB	0.053	8013.76	5.106
24	5.101		0.0000	0.000	2304.6	0.565	BB	0.023	1670.28	1.064
25	5.144		0.0000	0.000	2921.5	0.716	BB	0.024	2004.33	1.277
26	5.180		0.0000	0.000	33855.6	8.302	BB	0.044	12832.34	8.177
27	5.337		0.0000	0.000	2882.0	0.707	BB	0.026	1856.72	1.183
28	5.391		0.0000	0.000	27485.5	6.740	BB	0.035	13260.37	8.450
29	5.474		0.0000	0.000	12647.1	3.101	BB	0.041	5166.07	3.292
30	5.563		0.0000	0.000	1552.9	0.381	BB	0.025	1041.43	0.664
31	5.629		0.0000	0.000	63912.2	15.672	BB	0.045	23646.07	15.067
32	5.801		0.0000	0.000	841.5	0.206	BB	0.024	587.73	0.374
33	5.869		0.0000	0.000	14715.2	3.608	BB	0.028	8674.83	5.528
34	5.926		0.0000	0.000	32901.9	8.068	BB	0.035	15492.40	9.872
35	6.091		0.0000	0.000	1752.0	0.430	BB	0.039	753.15	0.480
36	6.172		0.0000	0.000	651.3	0.160	BB	0.027	409.17	0.261
37	6.240		0.0000	0.000	13789.5	3.381	BB	0.048	4739.29	3.020
38	6.385		0.0000	0.000	37378.0	9.166	BB	0.061	10150.80	6.468
39	6.623		0.0000	0.000	10110.7	2.479	BB	0.042	4052.28	2.582
40	6.806	DCB	-4.3014	100.000	1801.7	0.442	BB	0.068	441.72	0.281
41	7.487		0.0000	0.000	1246.1	0.306	BB	0.082	253.94	0.162
42	7.610		0.0000	0.000	315.9	0.077	BB	0.062	85.32	0.054
43	8.303		0.0000	0.000	3268.3	0.801	BB	0.166	327.88	0.209
44	8.933		0.0000	0.000	133.2	0.033	BB	0.036	61.23	0.039

Total Area = 407808.7, Total Amount = -4.301, Total Height = 156936.3, Sample Units = ug/l

ASX22 0403

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.22R

C:\CPWIN\DATA1\1P7307.22R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 7:14:34 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.22R

C:\CPWIN\DATA1\1P7307B.22R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 7:14:34 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
				.	0	.	TCX
3.02	431	-4.66085	TCX				
7.972	1209	.47839	DCB	6.806	442	-4.3014	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.22A
Area File: C:\CPWIN\DATA1\1P7307B.22A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 7:23:50 PM
File Reported On: 11/3/2009 at 7:24:03 PM

10X REPLOTS

ASX22 0485

C:\CPWIN\DATA1\1P7307.22R



C:\CPWIN\DATA1\1P7307B.22R



ASX22 0486

Sample Name: **AR6220924B AAAR622AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 07:36:40pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.23R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.554		0.0000	0.000	4172.9	0.565	BV	0.053	1300.61	0.380
2	1.627		0.0000	0.000	2804.3	0.380	VV	0.038	1239.63	0.362
3	1.756		0.0000	0.000	16652.1	2.254	VV	0.094	2939.77	0.859
4	1.868		0.0000	0.000	23991.6	3.247	VV	0.098	4092.98	1.196
5	1.929		0.0000	0.000	16566.7	2.242	VB	0.056	4936.42	1.443
6	2.089		0.0000	0.000	37013.7	5.009	BV	0.140	4410.17	1.289
7	2.175		0.0000	0.000	30040.3	4.066	VV	0.102	4905.92	1.434
8	2.269		0.0000	0.000	8104.2	1.097	VV	0.049	2752.22	0.804
9	2.319		0.0000	0.000	5662.1	0.766	VV	0.040	2364.01	0.691
10	2.380		0.0000	0.000	9803.2	1.327	VV	0.086	1910.56	0.558
11	2.460		0.0000	0.000	6155.1	0.833	VV	0.072	1416.28	0.414
12	2.524		0.0000	0.000	2119.3	0.287	VV	0.036	986.12	0.288
13	2.579		0.0000	0.000	3677.0	0.498	VB	0.082	743.13	0.217
14	2.732		0.0000	0.000	219.3	0.030	BB	0.060	60.99	0.018
15	2.776		0.0000	0.000	2110.5	0.286	BB	0.054	647.24	0.189
16	2.879		0.0000	0.000	15925.3	2.155	BB	0.051	5154.60	1.507
17	2.990	TCX	-3.9630	112.982	18274.0	2.473	BB	0.093	3280.69	0.959
18	3.063		0.0000	0.000	440.0	0.060	BB	0.051	144.63	0.042
19	3.218		0.0000	0.000	944.7	0.128	BB	0.123	127.72	0.037
20	3.301		0.0000	0.000	173.3	0.023	BB	0.035	83.58	0.024
21	3.383		0.0000	0.000	549.0	0.074	BB	0.026	348.63	0.102
22	3.445		0.0000	0.000	109.8	0.015	BB	0.034	53.97	0.016
23	3.534		0.0000	0.000	320.3	0.043	BB	0.040	132.74	0.039
24	3.604		0.0000	0.000	319.8	0.043	BB	0.040	133.08	0.039
25	3.717		0.0000	0.000	1147.4	0.155	BB	0.048	395.92	0.116
26	3.775		0.0000	0.000	343.1	0.046	BB	0.032	180.11	0.053
27	3.865		0.0000	0.000	817.9	0.111	BB	0.028	482.26	0.141
28	3.976		0.0000	0.000	305.9	0.041	BB	0.029	174.65	0.051
29	4.030		0.0000	0.000	879.1	0.119	BB	0.042	349.95	0.102
30	4.112		0.0000	0.000	2823.7	0.382	BB	0.041	1149.78	0.336
31	4.182		0.0000	0.000	861.2	0.117	BB	0.033	441.31	0.129
32	4.266		0.0000	0.000	1434.5	0.194	BB	0.028	840.46	0.246
33	4.361		0.0000	0.000	473.7	0.064	BB	0.020	389.88	0.114
34	4.399		0.0000	0.000	207.6	0.028	BB	0.020	170.84	0.050
35	4.527		0.0000	0.000	720.7	0.098	BB	0.029	407.35	0.119
36	4.650		0.0000	0.000	577.3	0.078	BB	0.031	307.81	0.090
37	4.813		0.0000	0.000	483.1	0.065	BB	0.037	220.44	0.064
38	4.856		0.0000	0.000	4053.5	0.549	BB	0.023	2914.80	0.852
39	4.918		0.0000	0.000	652.2	0.088	BB	0.022	494.91	0.145
40	4.999		0.0000	0.000	844.6	0.114	BB	0.027	513.74	0.150
41	5.059		0.0000	0.000	4849.9	0.656	BB	0.024	3301.59	0.965
42	5.105		0.0000	0.000	1379.3	0.187	BB	0.031	737.63	0.216
43	5.204		0.0000	0.000	641.7	0.087	BB	0.026	408.57	0.119
44	5.311		0.0000	0.000	6368.7	0.862	BB	0.029	3652.06	1.067
45	5.376		0.0000	0.000	2130.4	0.288	BB	0.023	1512.95	0.442
46	5.428		0.0000	0.000	26344.2	3.565	BB	0.034	12959.45	3.788
47	5.522		0.0000	0.000	31377.0	4.246	BB	0.027	19393.95	5.669
48	5.637		0.0000	0.000	2152.0	0.291	BB	0.034	1046.84	0.306
49	5.704		0.0000	0.000	3257.3	0.441	BB	0.025	2131.16	0.623
50	5.767		0.0000	0.000	32208.4	4.359	BB	0.027	20201.48	5.905
51	5.814		0.0000	0.000	9806.5	1.327	BB	0.021	7651.67	2.236
52	5.878		0.0000	0.000	17484.4	2.366	BB	0.031	9474.82	2.769
53	5.970		0.0000	0.000	437.5	0.059	BB	0.022	327.58	0.096
54	6.010		0.0000	0.000	27222.2	3.684	BB	0.030	15094.34	4.412

ASX22 0487

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.084		0.0000	0.000	39949.8	5.407	BB	0.025	26901.39	7.863
56	6.136		0.0000	0.000	13712.8	1.856	BB	0.023	9774.65	2.857
57	6.251		0.0000	0.000	1889.3	0.256	BB	0.022	1436.93	0.420
58	6.282		0.0000	0.000	33675.1	4.557	BB	0.027	20750.11	6.065
59	6.353		0.0000	0.000	14848.8	2.010	BB	0.028	8813.65	2.576
60	6.398		0.0000	0.000	1473.5	0.199	BB	0.016	1508.46	0.441
61	6.446		0.0000	0.000	878.7	0.119	BB	0.021	696.95	0.204
62	6.510		0.0000	0.000	5207.0	0.705	BB	0.026	3310.54	0.968
63	6.580		0.0000	0.000	85408.1	11.559	BB	0.030	47162.34	13.785
64	6.829		0.0000	0.000	44719.1	6.052	BB	0.031	24241.76	7.086
65	6.896		0.0000	0.000	33069.0	4.475	BB	0.035	15638.93	4.571
66	7.057		0.0000	0.000	2504.3	0.339	BB	0.028	1494.74	0.437
67	7.184		0.0000	0.000	19118.3	2.587	BB	0.045	7123.26	2.082
68	7.418		0.0000	0.000	18911.4	2.559	BB	0.029	10989.64	3.212
69	7.491		0.0000	0.000	4613.1	0.624	BB	0.063	1217.07	0.356
70	7.651		0.0000	0.000	4159.6	0.563	BB	0.118	587.08	0.172
71	7.733		0.0000	0.000	14139.8	1.914	BB	0.034	6899.63	2.017
72	7.973	DCB	0.4554	-12.982	5495.2	0.744	BB	0.080	1151.11	0.336
73	8.206		0.0000	0.000	4740.6	0.642	BB	0.291	271.70	0.079
74	8.384		0.0000	0.000	595.8	0.081	BB	0.054	183.20	0.054
75	8.610		0.0000	0.000	340.4	0.046	BB	0.035	161.44	0.047
76	8.664		0.0000	0.000	1017.7	0.138	BB	0.053	318.84	0.093

Total Area = 738900.8, Total Amount = -3.508, Total Height = 342127.4, Sample Units = ug/l

ASX22 0488

Sample Name: **AR6220924B AAAR622AA ICAL 0930699999**
Acquired from CP07–H5088B via port 4 on 11/3/09 07:36:40pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.23R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.576		0.0000	0.000	3617.1	0.494	BV	0.038	1566.61	0.525
2	1.639		0.0000	0.000	16388.2	2.236	VV	0.112	2445.90	0.819
3	2.010		0.0000	0.000	2140.5	0.292	VB	0.053	672.23	0.225
4	2.113		0.0000	0.000	9758.0	1.331	BB	0.422	385.51	0.129
5	2.463		0.0000	0.000	8897.5	1.214	BB	0.932	159.19	0.053
6	2.533		0.0000	0.000	194.2	0.026	BB	0.061	53.43	0.018
7	2.556		0.0000	0.000	82.2	0.011	BB	0.027	51.59	0.017
8	2.717		0.0000	0.000	1028.7	0.140	BB	0.074	232.42	0.078
9	3.010		0.0000	0.000	1987.2	0.271	BB	0.057	583.78	0.196
10	3.169		0.0000	0.000	1231.1	0.168	BB	0.041	504.46	0.169
11	3.353		0.0000	0.000	621.9	0.085	BB	0.027	391.05	0.131
12	3.403		0.0000	0.000	1893.7	0.258	BB	0.054	589.11	0.197
13	3.557		0.0000	0.000	1153.7	0.157	BB	0.036	535.57	0.179
14	3.620		0.0000	0.000	1925.6	0.263	BB	0.067	477.83	0.160
15	3.796		0.0000	0.000	2154.3	0.294	BB	0.094	380.80	0.128
16	4.109		0.0000	0.000	9346.4	1.275	BB	0.045	3449.14	1.155
17	4.255		0.0000	0.000	13487.2	1.840	BB	0.064	3498.56	1.172
18	4.433		0.0000	0.000	211.5	0.029	BB	0.025	139.28	0.047
19	4.509		0.0000	0.000	270.6	0.037	BB	0.041	110.87	0.037
20	4.544		0.0000	0.000	3706.6	0.506	BB	0.025	2431.00	0.814
21	4.600		0.0000	0.000	32504.9	4.435	BB	0.045	12044.60	4.034
22	4.695		0.0000	0.000	42324.6	5.775	BB	0.036	19326.65	6.473
23	4.823		0.0000	0.000	2682.6	0.366	BB	0.029	1517.25	0.508
24	4.875		0.0000	0.000	47536.9	6.486	BB	0.038	20736.84	6.945
25	4.981		0.0000	0.000	53682.2	7.324	BB	0.054	16531.15	5.537
26	5.102		0.0000	0.000	4517.9	0.616	BB	0.024	3092.17	1.036
27	5.145		0.0000	0.000	6408.2	0.874	BB	0.023	4695.82	1.573
28	5.181		0.0000	0.000	30565.0	4.170	BB	0.025	20414.09	6.837
29	5.219		0.0000	0.000	4266.1	0.582	BB	0.019	3728.99	1.249
30	5.338		0.0000	0.000	5823.2	0.795	BB	0.025	3893.27	1.304
31	5.392		0.0000	0.000	53244.1	7.264	BB	0.034	25983.53	8.702
32	5.475		0.0000	0.000	25903.4	3.534	BB	0.041	10458.41	3.503
33	5.562		0.0000	0.000	2867.2	0.391	BB	0.025	1874.95	0.628
34	5.630		0.0000	0.000	126565.1	17.268	BB	0.045	46928.50	15.717
35	5.803		0.0000	0.000	666.1	0.091	BB	0.023	486.33	0.163
36	5.870		0.0000	0.000	28405.6	3.876	BB	0.027	17356.73	5.813
37	5.927		0.0000	0.000	65490.3	8.935	BB	0.036	30362.42	10.169
38	6.091		0.0000	0.000	3833.4	0.523	BB	0.036	1795.17	0.601
39	6.174		0.0000	0.000	1948.2	0.266	BB	0.030	1072.03	0.359
40	6.241		0.0000	0.000	20965.0	2.860	BB	0.042	8248.69	2.763
41	6.386		0.0000	0.000	58883.5	8.034	BB	0.051	19326.27	6.473
42	6.624		0.0000	0.000	21085.3	2.877	BB	0.040	8783.57	2.942
43	6.803	DCB	-4.2908	100.000	1806.2	0.246	BB	0.064	471.51	0.158
44	7.377		0.0000	0.000	159.0	0.022	BB	0.035	75.78	0.025
45	7.620		0.0000	0.000	8495.7	1.159	BB	0.324	437.22	0.146
46	8.303		0.0000	0.000	1649.2	0.225	BB	0.141	195.18	0.065
47	8.873		0.0000	0.000	565.8	0.077	BB	0.110	85.35	0.029

Total Area = 732940.9, Total Amount = -4.291, Total Height = 298580.8, Sample Units = ug/l

ASX22 0489

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.23R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 7:27:31 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.23R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 7:27:31 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.99	3281	-3.96297	TCX	.	0	.	TCX
7.973	1151	.45535	DCB	6.803	472	-4.29082	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.23A
Area File: C:\CPWIN\DATA1\1P7307B.23A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 7:36:48 PM
File Reported On: 11/3/2009 at 7:37:00 PM

10X REPLOTS

ASX22 0411

C:\CPWIN\DATA1\1P7307.23R



C:\CPWIN\DATA1\1P7307B.23R



ASX22 8412:

Sample Name: **AR6230924B AAAR623AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 07:49:39pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.24R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.512		0.0000	0.000	169.7	0.014	BV	0.022	126.76	0.022
2	1.551		0.0000	0.000	7303.2	0.607	VV	0.048	2518.99	0.442
3	1.668		0.0000	0.000	8571.6	0.712	VV	0.081	1758.18	0.309
4	1.795		0.0000	0.000	15162.3	1.260	VV	0.092	2758.31	0.484
5	1.926		0.0000	0.000	26959.5	2.241	VB	0.110	4093.04	0.719
6	2.049		0.0000	0.000	24566.3	2.042	BV	0.106	3862.81	0.678
7	2.172		0.0000	0.000	28958.6	2.407	VV	0.113	4270.11	0.750
8	2.277		0.0000	0.000	6046.4	0.503	VV	0.053	1892.60	0.332
9	2.365		0.0000	0.000	7959.3	0.662	VV	0.098	1354.73	0.238
10	2.453		0.0000	0.000	4322.6	0.359	VV	0.086	835.04	0.147
11	2.587		0.0000	0.000	2443.8	0.203	VB	0.140	290.02	0.051
12	2.766		0.0000	0.000	4200.2	0.349	BB	0.070	1004.75	0.176
13	2.873		0.0000	0.000	12245.7	1.018	BB	0.059	3450.43	0.606
14	3.016	TCX	-4.5691	130.093	12526.6	1.041	BB	0.259	805.64	0.141
15	3.060		0.0000	0.000	704.4	0.059	BB	0.058	203.78	0.036
16	3.217		0.0000	0.000	1263.7	0.105	BB	0.127	166.30	0.029
17	3.268		0.0000	0.000	1904.0	0.158	BB	0.095	333.89	0.059
18	3.380		0.0000	0.000	1431.9	0.119	BB	0.030	799.34	0.140
19	3.516		0.0000	0.000	602.6	0.050	BB	0.072	139.06	0.024
20	3.610		0.0000	0.000	382.1	0.032	BB	0.048	133.76	0.023
21	3.692		0.0000	0.000	2122.5	0.176	BB	0.037	946.04	0.166
22	3.766		0.0000	0.000	3047.8	0.253	BB	0.084	602.28	0.106
23	3.866		0.0000	0.000	2849.8	0.237	BB	0.033	1426.81	0.251
24	3.935		0.0000	0.000	75.9	0.006	BB	0.027	46.75	0.008
25	3.969		0.0000	0.000	620.1	0.052	BB	0.025	408.04	0.072
26	4.028		0.0000	0.000	1272.9	0.106	BB	0.042	509.85	0.090
27	4.121		0.0000	0.000	3847.4	0.320	BB	0.038	1672.88	0.294
28	4.179		0.0000	0.000	1584.7	0.132	BB	0.031	863.79	0.152
29	4.262		0.0000	0.000	3781.0	0.314	BB	0.037	1711.76	0.301
30	4.357		0.0000	0.000	2197.3	0.183	BB	0.034	1075.69	0.189
31	4.394		0.0000	0.000	1131.0	0.094	BB	0.044	426.99	0.075
32	4.524		0.0000	0.000	1360.8	0.113	BB	0.031	739.37	0.130
33	4.648		0.0000	0.000	1025.6	0.085	BB	0.029	581.77	0.102
34	4.693		0.0000	0.000	2448.8	0.204	BB	0.063	643.17	0.113
35	4.810		0.0000	0.000	3089.0	0.257	BB	0.099	517.85	0.091
36	4.852		0.0000	0.000	7642.0	0.635	BB	0.024	5371.78	0.943
37	4.914		0.0000	0.000	737.3	0.061	BB	0.024	521.06	0.091
38	4.998		0.0000	0.000	1223.5	0.102	BB	0.027	753.33	0.132
39	5.055		0.0000	0.000	9514.0	0.791	BB	0.025	6464.16	1.135
40	5.100		0.0000	0.000	3886.0	0.323	BB	0.062	1040.78	0.183
41	5.230		0.0000	0.000	6287.4	0.523	BB	0.071	1472.15	0.258
42	5.307		0.0000	0.000	11572.1	0.962	BB	0.029	6612.29	1.161
43	5.372		0.0000	0.000	3737.6	0.311	BB	0.023	2711.63	0.476
44	5.424		0.0000	0.000	47541.9	3.952	BB	0.034	23161.30	4.066
45	5.518		0.0000	0.000	60697.1	5.045	BB	0.029	34766.57	6.104
46	5.634		0.0000	0.000	7590.9	0.631	BB	0.042	2981.10	0.523
47	5.700		0.0000	0.000	5894.4	0.490	BB	0.025	3899.38	0.685
48	5.762		0.0000	0.000	56540.9	4.700	BB	0.026	36053.05	6.330
49	5.810		0.0000	0.000	16979.3	1.411	BB	0.021	13386.97	2.350
50	5.874		0.0000	0.000	30883.6	2.567	BB	0.030	17447.69	3.063
51	5.966		0.0000	0.000	1002.6	0.083	BB	0.020	830.40	0.146
52	6.008		0.0000	0.000	40880.1	3.398	BB	0.029	23322.86	4.095
53	6.080		0.0000	0.000	69222.1	5.754	BB	0.024	47282.01	8.301
54	6.131		0.0000	0.000	24909.2	2.070	BB	0.024	17497.49	3.072

ASX22 0413

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.247		0.0000	0.000	3650.8	0.303	BB	0.023	2605.43	0.457
56	6.278		0.0000	0.000	59539.3	4.949	BB	0.027	37085.41	6.511
57	6.350		0.0000	0.000	30600.0	2.543	BB	0.029	17485.91	3.070
58	6.399		0.0000	0.000	2089.0	0.174	BB	0.017	2079.66	0.365
59	6.444		0.0000	0.000	1851.7	0.154	BB	0.022	1373.95	0.241
60	6.507		0.0000	0.000	9840.3	0.818	BB	0.026	6300.83	1.106
61	6.576		0.0000	0.000	153193.7	12.733	BB	0.030	85373.50	14.989
62	6.716		0.0000	0.000	5390.5	0.448	BB	0.060	1488.25	0.261
63	6.824		0.0000	0.000	80794.9	6.716	BB	0.031	43963.83	7.719
64	6.892		0.0000	0.000	60659.7	5.042	BB	0.036	27879.68	4.895
65	7.053		0.0000	0.000	13181.1	1.096	BB	0.058	3789.49	0.665
66	7.180		0.0000	0.000	60533.8	5.032	BB	0.069	14697.02	2.580
67	7.415		0.0000	0.000	26717.3	2.221	BB	0.028	15777.03	2.770
68	7.487		0.0000	0.000	8365.6	0.695	BB	0.050	2809.22	0.493
69	7.656		0.0000	0.000	10638.0	0.884	BB	0.174	1021.54	0.179
70	7.728		0.0000	0.000	24227.8	2.014	BB	0.034	11814.65	2.074
71	7.968	DCB	1.0569	-30.093	11101.1	0.923	BB	0.069	2671.87	0.469
72	8.219		0.0000	0.000	21653.5	1.800	BB	0.350	1030.98	0.181
73	8.319		0.0000	0.000	108.1	0.009	BB	0.030	59.10	0.010
74	8.373		0.0000	0.000	890.3	0.074	BB	0.077	193.53	0.034
75	8.512		0.0000	0.000	13147.2	1.093	BB	0.144	1523.35	0.267

Total Area = 1203095.0, Total Amount = -3.512, Total Height = 569571.0, Sample Units = ug/l

ASX22 0414

Sample Name: **AR6230924B AAAR623AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 07:49:39pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.24R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.570		0.0000	0.000	4031.7	0.310	BV	0.042	1608.64	0.301
2	1.636		0.0000	0.000	12955.5	0.995	VV	0.092	2336.55	0.437
3	1.971		0.0000	0.000	430.9	0.033	VV	0.062	115.38	0.022
4	2.013		0.0000	0.000	1246.3	0.096	VB	0.048	431.88	0.081
5	2.110		0.0000	0.000	118.3	0.009	BB	0.018	110.33	0.021
6	2.186		0.0000	0.000	3932.2	0.302	BB	0.103	635.93	0.119
7	2.456		0.0000	0.000	4822.9	0.370	BB	0.444	180.99	0.034
8	2.506		0.0000	0.000	231.2	0.018	BB	0.030	127.98	0.024
9	2.680		0.0000	0.000	220.7	0.017	BB	0.072	51.13	0.010
10	2.714		0.0000	0.000	1492.4	0.115	BB	0.055	455.42	0.085
11	2.950		0.0000	0.000	1003.9	0.077	BB	0.518	32.32	0.006
12	3.007		0.0000	0.000	4306.2	0.331	BB	0.059	1216.03	0.227
13	3.166		0.0000	0.000	3517.8	0.270	BB	0.051	1139.39	0.213
14	3.314		0.0000	0.000	297.9	0.023	BB	0.074	67.48	0.013
15	3.347		0.0000	0.000	755.8	0.058	BB	0.025	497.53	0.093
16	3.403		0.0000	0.000	4078.1	0.313	BB	0.060	1138.89	0.213
17	3.556		0.0000	0.000	1551.9	0.119	BB	0.036	721.73	0.135
18	3.619		0.0000	0.000	2715.5	0.209	BB	0.051	887.18	0.166
19	3.794		0.0000	0.000	2164.5	0.166	BB	0.050	717.47	0.134
20	3.920		0.0000	0.000	509.4	0.039	BB	0.027	318.02	0.059
21	4.106		0.0000	0.000	15781.8	1.212	BB	0.043	6185.13	1.156
22	4.199		0.0000	0.000	248.2	0.019	BB	0.019	213.42	0.040
23	4.251		0.0000	0.000	23235.7	1.785	BB	0.060	6484.82	1.212
24	4.430		0.0000	0.000	1203.8	0.092	BB	0.031	650.39	0.122
25	4.506		0.0000	0.000	614.8	0.047	BB	0.051	200.58	0.037
26	4.540		0.0000	0.000	6791.5	0.522	BB	0.026	4422.30	0.826
27	4.597		0.0000	0.000	59205.0	4.548	BB	0.045	21848.87	4.083
28	4.691		0.0000	0.000	74846.1	5.749	BB	0.036	34701.82	6.484
29	4.818		0.0000	0.000	6289.9	0.483	BB	0.030	3456.76	0.646
30	4.872		0.0000	0.000	84102.2	6.460	BB	0.038	37163.71	6.944
31	4.979		0.0000	0.000	97234.4	7.469	BB	0.055	29612.49	5.533
32	5.099		0.0000	0.000	8029.7	0.617	BB	0.024	5561.64	1.039
33	5.140		0.0000	0.000	12432.3	0.955	BB	0.023	8823.47	1.649
34	5.177		0.0000	0.000	105660.0	8.116	BB	0.041	43305.41	8.092
35	5.334		0.0000	0.000	11063.5	0.850	BB	0.025	7329.74	1.370
36	5.388		0.0000	0.000	94136.2	7.231	BB	0.034	46703.19	8.727
37	5.471		0.0000	0.000	46911.4	3.603	BB	0.041	19073.17	3.564
38	5.557		0.0000	0.000	4299.4	0.330	BB	0.022	3281.29	0.613
39	5.626		0.0000	0.000	229239.4	17.609	BB	0.044	86270.26	16.120
40	5.795		0.0000	0.000	374.9	0.029	BB	0.019	332.34	0.062
41	5.866		0.0000	0.000	51439.5	3.951	BB	0.027	31307.40	5.850
42	5.922		0.0000	0.000	108168.6	8.309	BB	0.035	52026.93	9.721
43	6.087		0.0000	0.000	6312.2	0.485	BB	0.035	3021.45	0.565
44	6.170		0.0000	0.000	3623.6	0.278	BB	0.031	1970.53	0.368
45	6.237		0.0000	0.000	40020.9	3.074	BB	0.041	16211.27	3.029
46	6.382		0.0000	0.000	111777.8	8.586	BB	0.052	35621.26	6.656
47	6.620		0.0000	0.000	33317.0	2.559	BB	0.037	14848.40	2.774
48	6.802	DCB	-4.3102	100.000	4442.3	0.341	BB	0.178	417.06	0.078
49	7.090		0.0000	0.000	302.2	0.023	BB	0.173	29.09	0.005
50	7.510		0.0000	0.000	1718.7	0.132	BB	0.082	349.14	0.065
51	7.620		0.0000	0.000	4813.7	0.370	BB	0.184	436.25	0.082
52	8.033		0.0000	0.000	2787.6	0.214	BB	0.144	323.26	0.060
53	8.670		0.0000	0.000	741.7	0.057	BB	0.092	134.34	0.025
54	9.020		0.0000	0.000	293.3	0.023	BB	0.073	66.69	0.012

ASX22 8415

Total Area = 1301843.0, Total Amount = -4.31, Total Height = 535174.1, Sample Units = ug/l

ASX22 0416

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.24R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 7:40:30 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.24R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 7:40:30 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.016	806	-4.56914	TCX	.	0	.	TCX
7.968	2672	1.05693	DCB	6.802	417	-4.31016	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.24A
Area File: C:\CPWIN\DATA1\1P7307B.24A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 7:49:46 PM
File Reported On: 11/3/2009 at 7:49:58 PM

10X REPLOTS

ASXZ2 8418

C:\CPWIN\DATA1\1P7307.24R



C:\CPWIN\DATA1\1P7307B.24R



ASX22 0419

Sample Name: **AR6240924B AAAR624AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 08:02:37pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.25R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.554		0.0000	0.000	3750.3	0.188	BV	0.041	1534.38	0.138
2	1.873		0.0000	0.000	15923.1	0.797	VV	0.213	1247.20	0.112
3	1.932		0.0000	0.000	3977.4	0.199	VB	0.046	1439.84	0.129
4	2.034		0.0000	0.000	6745.7	0.338	BV	0.077	1466.45	0.132
5	2.173		0.0000	0.000	19602.4	0.982	VB	0.061	5397.92	0.484
6	2.322		0.0000	0.000	3406.3	0.171	BV	0.072	793.54	0.071
7	2.370		0.0000	0.000	2791.1	0.140	VV	0.067	697.92	0.063
8	2.457		0.0000	0.000	2477.1	0.124	VV	0.068	607.09	0.054
9	2.523		0.0000	0.000	2192.1	0.110	VV	0.053	685.30	0.061
10	2.627		0.0000	0.000	283.2	0.014	VB	0.034	137.32	0.012
11	2.726		0.0000	0.000	209.8	0.011	BB	0.062	56.33	0.005
12	2.771		0.0000	0.000	1608.1	0.081	BB	0.036	735.55	0.066
13	2.824		0.0000	0.000	14306.4	0.716	BB	0.072	3325.82	0.298
14	3.016	TCX	-4.6269	114.765	6743.6	0.338	BB	0.197	569.69	0.051
15	3.057		0.0000	0.000	81.1	0.004	BB	0.023	59.76	0.005
16	3.224		0.0000	0.000	695.3	0.035	BB	0.051	227.89	0.020
17	3.308		0.0000	0.000	888.2	0.044	BB	0.031	471.11	0.042
18	3.380		0.0000	0.000	2721.2	0.136	BB	0.029	1539.22	0.138
19	3.456		0.0000	0.000	89.0	0.004	BB	0.030	49.69	0.004
20	3.515		0.0000	0.000	316.9	0.016	BB	0.028	187.58	0.017
21	3.583		0.0000	0.000	127.8	0.006	BB	0.043	50.03	0.004
22	3.693		0.0000	0.000	4975.6	0.249	BB	0.038	2163.22	0.194
23	3.769		0.0000	0.000	404.8	0.020	BB	0.028	242.06	0.022
24	3.820		0.0000	0.000	178.7	0.009	BB	0.016	191.06	0.017
25	3.863		0.0000	0.000	4441.9	0.222	BB	0.032	2337.51	0.210
26	3.973		0.0000	0.000	387.3	0.019	BB	0.024	264.90	0.024
27	4.029		0.0000	0.000	1569.9	0.079	BB	0.038	694.79	0.062
28	4.125		0.0000	0.000	8481.8	0.425	BB	0.035	4059.55	0.364
29	4.180		0.0000	0.000	2891.3	0.145	BB	0.026	1876.09	0.168
30	4.265		0.0000	0.000	3431.7	0.172	BB	0.029	1958.34	0.176
31	4.359		0.0000	0.000	2414.3	0.121	BB	0.021	1901.10	0.170
32	4.395		0.0000	0.000	1488.4	0.075	BB	0.030	832.67	0.075
33	4.526		0.0000	0.000	3555.4	0.178	BB	0.032	1852.54	0.166
34	4.595		0.0000	0.000	235.7	0.012	BB	0.016	253.30	0.023
35	4.649		0.0000	0.000	2077.6	0.104	BB	0.029	1208.93	0.108
36	4.696		0.0000	0.000	788.0	0.039	BB	0.036	360.35	0.032
37	4.812		0.0000	0.000	620.3	0.031	BB	0.054	190.49	0.017
38	4.855		0.0000	0.000	19471.2	0.975	BB	0.026	12322.16	1.105
39	4.914		0.0000	0.000	1235.7	0.062	BB	0.019	1059.86	0.095
40	5.001		0.0000	0.000	1753.9	0.088	BB	0.023	1256.47	0.113
41	5.057		0.0000	0.000	32716.4	1.638	BB	0.034	15805.19	1.417
42	5.232		0.0000	0.000	2908.1	0.146	BB	0.045	1083.50	0.097
43	5.309		0.0000	0.000	24790.8	1.241	BB	0.029	14135.10	1.268
44	5.374		0.0000	0.000	7962.0	0.399	BB	0.023	5863.59	0.526
45	5.427		0.0000	0.000	96190.1	4.817	BB	0.034	46516.31	4.172
46	5.520		0.0000	0.000	109514.5	5.484	BB	0.027	68018.80	6.100
47	5.636		0.0000	0.000	8277.7	0.414	BB	0.035	3986.22	0.357
48	5.702		0.0000	0.000	12933.7	0.648	BB	0.025	8514.60	0.764
49	5.765		0.0000	0.000	111059.5	5.561	BB	0.026	72584.15	6.509
50	5.812		0.0000	0.000	34569.8	1.731	BB	0.021	27791.60	2.492
51	5.876		0.0000	0.000	64305.3	3.220	BB	0.029	36750.79	3.296
52	5.968		0.0000	0.000	1625.1	0.081	BB	0.018	1492.99	0.134
53	6.009		0.0000	0.000	100408.2	5.028	BB	0.030	55695.25	4.995
54	6.082		0.0000	0.000	137481.2	6.884	BB	0.024	93938.00	8.424

ASX22 0420

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.134		0.0000	0.000	50893.4	2.548	BB	0.024	36065.36	3.234
56	6.249		0.0000	0.000	8195.8	0.410	BB	0.025	5464.20	0.490
57	6.280		0.0000	0.000	116796.4	5.849	BB	0.026	74482.07	6.680
58	6.352		0.0000	0.000	55513.7	2.780	BB	0.028	33101.70	2.969
59	6.399		0.0000	0.000	6950.3	0.348	BB	0.017	6914.55	0.620
60	6.445		0.0000	0.000	3098.5	0.155	BB	0.020	2628.31	0.236
61	6.510		0.0000	0.000	21398.8	1.072	BB	0.026	13465.79	1.208
62	6.578		0.0000	0.000	317828.4	15.915	BB	0.031	171805.60	15.408
63	6.827		0.0000	0.000	158908.4	7.957	BB	0.030	88887.17	7.971
64	6.895		0.0000	0.000	120895.4	6.054	BB	0.036	55968.61	5.019
65	7.056		0.0000	0.000	10305.6	0.516	BB	0.030	5807.39	0.521
66	7.182		0.0000	0.000	53201.9	2.664	BB	0.035	25637.31	2.299
67	7.416		0.0000	0.000	101649.8	5.090	BB	0.031	54741.36	4.909
68	7.487		0.0000	0.000	6224.3	0.312	BB	0.036	2898.15	0.260
69	7.633		0.0000	0.000	4532.2	0.227	BB	0.097	780.71	0.070
70	7.731		0.0000	0.000	50644.7	2.536	BB	0.034	24636.59	2.209
71	7.870		0.0000	0.000	477.9	0.024	BB	0.041	196.52	0.018
72	7.974	DCB	0.5953	-14.765	6437.9	0.322	BB	0.071	1504.81	0.135
73	8.173		0.0000	0.000	5883.3	0.295	BB	0.179	549.11	0.049
74	8.305		0.0000	0.000	44.8	0.002	BB	0.023	32.08	0.003
75	8.385		0.0000	0.000	709.3	0.036	BB	0.052	225.36	0.020
76	8.486		0.0000	0.000	231.0	0.012	BB	0.056	68.78	0.006
77	8.595		0.0000	0.000	837.2	0.042	BB	0.041	339.98	0.030
78	8.673		0.0000	0.000	1290.5	0.065	BB	0.055	389.58	0.035

Total Area = 1997032.0, Total Amount = -4.032, Total Height = 1115072.0, Sample Units = ug/l

ASX22 0421

Sample Name: **AR6240924B AAAR624AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 08:02:37pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.25R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.568		0.0000	0.000	10420.7	0.417	BV	0.043	4065.09	0.366
2	1.634		0.0000	0.000	10379.0	0.416	VB	0.072	2408.62	0.217
3	1.967		0.0000	0.000	239.5	0.010	BB	0.135	29.47	0.003
4	2.064		0.0000	0.000	601.2	0.024	BB	0.200	50.00	0.005
5	2.280		0.0000	0.000	13071.0	0.523	BB	0.066	3286.14	0.296
6	2.563		0.0000	0.000	296.5	0.012	BB	0.035	143.14	0.013
7	2.653		0.0000	0.000	300.6	0.012	BB	0.036	140.10	0.013
8	2.716		0.0000	0.000	2743.9	0.110	BB	0.045	1020.12	0.092
9	2.834		0.0000	0.000	28.6	0.001	BB	0.018	26.37	0.002
10	2.860		0.0000	0.000	316.9	0.013	BB	0.031	169.60	0.015
11	3.007		0.0000	0.000	7985.2	0.320	BB	0.051	2592.57	0.234
12	3.168		0.0000	0.000	4806.3	0.192	BB	0.043	1853.15	0.167
13	3.347		0.0000	0.000	1294.3	0.052	BB	0.047	455.99	0.041
14	3.404		0.0000	0.000	4147.4	0.166	BB	0.034	2061.78	0.186
15	3.450		0.0000	0.000	1040.5	0.042	BB	0.031	567.48	0.051
16	3.559		0.0000	0.000	1986.8	0.080	BB	0.033	998.77	0.090
17	3.618		0.0000	0.000	8131.2	0.326	BB	0.067	2011.20	0.181
18	3.797		0.0000	0.000	5079.7	0.203	BB	0.052	1624.72	0.146
19	3.921		0.0000	0.000	1183.3	0.047	BB	0.035	562.71	0.051
20	4.108		0.0000	0.000	34124.0	1.367	BB	0.042	13660.31	1.231
21	4.201		0.0000	0.000	518.5	0.021	BB	0.019	445.55	0.040
22	4.253		0.0000	0.000	42070.1	1.685	BB	0.050	13897.62	1.252
23	4.433		0.0000	0.000	1443.2	0.058	BB	0.033	736.62	0.066
24	4.505		0.0000	0.000	1086.7	0.044	BB	0.022	812.29	0.073
25	4.542		0.0000	0.000	14831.3	0.594	BB	0.026	9519.37	0.858
26	4.599		0.0000	0.000	123657.6	4.952	BB	0.045	46213.13	4.165
27	4.693		0.0000	0.000	149972.8	6.006	BB	0.035	71501.55	6.444
28	4.821		0.0000	0.000	11439.4	0.458	BB	0.030	6381.35	0.575
29	4.874		0.0000	0.000	170334.2	6.822	BB	0.036	78319.04	7.058
30	4.981		0.0000	0.000	205577.0	8.233	BB	0.055	62575.82	5.639
31	5.101		0.0000	0.000	17684.9	0.708	BB	0.022	13372.91	1.205
32	5.143		0.0000	0.000	24106.5	0.965	BB	0.024	16883.83	1.522
33	5.179		0.0000	0.000	103941.3	4.163	BB	0.024	73140.27	6.592
34	5.218		0.0000	0.000	21257.1	0.851	BB	0.023	15737.42	1.418
35	5.337		0.0000	0.000	24088.6	0.965	BB	0.025	16135.56	1.454
36	5.390		0.0000	0.000	187587.5	7.513	BB	0.033	95742.46	8.628
37	5.473		0.0000	0.000	98634.7	3.950	BB	0.042	39570.76	3.566
38	5.562		0.0000	0.000	10553.6	0.423	BB	0.023	7557.57	0.681
39	5.628		0.0000	0.000	465098.5	18.627	BB	0.043	180114.00	16.232
40	5.799		0.0000	0.000	515.2	0.021	BB	0.018	481.87	0.043
41	5.868		0.0000	0.000	98012.0	3.925	BB	0.026	62206.05	5.606
42	5.925		0.0000	0.000	239327.7	9.585	BB	0.035	113118.70	10.194
43	6.090		0.0000	0.000	13283.4	0.532	BB	0.034	6469.87	0.583
44	6.173		0.0000	0.000	8300.7	0.332	BB	0.030	4562.88	0.411
45	6.240		0.0000	0.000	70630.0	2.829	BB	0.039	30349.67	2.735
46	6.384		0.0000	0.000	198417.4	7.946	BB	0.045	72691.09	6.551
47	6.623		0.0000	0.000	75981.1	3.043	BB	0.040	31578.68	2.846
48	6.803	DCB	-4.1820	100.000	1673.8	0.067	BB	0.036	777.77	0.070
49	7.473		0.0000	0.000	1415.5	0.057	BB	0.090	260.85	0.024
50	7.620		0.0000	0.000	4671.2	0.187	BB	0.194	400.80	0.036
51	8.033		0.0000	0.000	2648.3	0.106	BB	0.136	325.46	0.029

Total Area = 2496936.0, Total Amount = -4.182, Total Height = 1109608.0, Sample Units = ug/l

ASX22 0422

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.25R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 7:53:28 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.25R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 7:53:28 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.016	570	-4.62693	TCX	.	0	.	TCX
7.974	1505	.59527	DCB	6.803	778	-4.18204	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.25A
Area File: C:\CPWIN\DATA1\1P7307B.25A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 8:02:46 PM
File Reported On: 11/3/2009 at 8:02:58 PM

10X REPLOTS

ASX22 0424

C:\CPWIN\DATA1\1P7307.25R



C:\CPWIN\DATA1\1P7307B.25R



ASX22 #425

Sample Name: **AR6250924B AAAR625AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 08:15:39pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.26R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.583		0.0000	0.000	4909.8	0.133	BV	0.064	1277.82	0.062
2	1.928		0.0000	0.000	64832.8	1.760	VB	0.212	5095.10	0.247
3	2.037		0.0000	0.000	30141.9	0.818	BV	0.101	4993.53	0.242
4	2.174		0.0000	0.000	42397.5	1.151	VV	0.138	5112.02	0.248
5	2.373		0.0000	0.000	22809.8	0.619	VV	0.192	1976.56	0.096
6	2.452		0.0000	0.000	6574.5	0.178	VV	0.079	1395.71	0.068
7	2.527		0.0000	0.000	2267.5	0.062	VV	0.040	944.59	0.046
8	2.589		0.0000	0.000	1856.0	0.050	VV	0.066	467.99	0.023
9	2.628		0.0000	0.000	439.7	0.012	VB	0.032	228.62	0.011
10	2.772		0.0000	0.000	2882.1	0.078	BB	0.059	811.55	0.039
11	2.863		0.0000	0.000	5350.9	0.145	BB	0.085	1050.18	0.051
12	3.012	TCX	-4.6256	121.833	3830.0	0.104	BB	0.111	575.17	0.028
13	3.058		0.0000	0.000	242.7	0.007	BB	0.040	100.36	0.005
14	3.226		0.0000	0.000	835.8	0.023	BB	0.035	393.95	0.019
15	3.308		0.0000	0.000	1184.3	0.032	BB	0.033	593.68	0.029
16	3.381		0.0000	0.000	4436.4	0.120	BB	0.026	2866.95	0.139
17	3.462		0.0000	0.000	105.4	0.003	BB	0.036	48.71	0.002
18	3.514		0.0000	0.000	660.7	0.018	BB	0.025	436.45	0.021
19	3.606		0.0000	0.000	385.7	0.010	BB	0.051	126.53	0.006
20	3.694		0.0000	0.000	9924.3	0.269	BB	0.040	4160.36	0.202
21	3.772		0.0000	0.000	663.8	0.018	BB	0.028	391.37	0.019
22	3.821		0.0000	0.000	304.3	0.008	BB	0.015	329.87	0.016
23	3.863		0.0000	0.000	7412.4	0.201	BB	0.030	4116.56	0.200
24	3.977		0.0000	0.000	142.7	0.004	BB	0.025	93.81	0.005
25	4.030		0.0000	0.000	3026.8	0.082	BB	0.038	1311.74	0.064
26	4.126		0.0000	0.000	16423.7	0.446	BB	0.034	8134.23	0.395
27	4.181		0.0000	0.000	5361.0	0.146	BB	0.025	3615.20	0.175
28	4.265		0.0000	0.000	6264.8	0.170	BB	0.028	3677.29	0.178
29	4.360		0.0000	0.000	4680.3	0.127	BB	0.021	3738.23	0.181
30	4.395		0.0000	0.000	2636.9	0.072	BB	0.026	1661.39	0.081
31	4.528		0.0000	0.000	7351.1	0.200	BB	0.032	3784.66	0.184
32	4.598		0.0000	0.000	515.4	0.014	BB	0.016	528.32	0.026
33	4.650		0.0000	0.000	3937.9	0.107	BB	0.028	2327.79	0.113
34	4.698		0.0000	0.000	1468.1	0.040	BB	0.035	706.56	0.034
35	4.813		0.0000	0.000	961.1	0.026	BB	0.059	270.14	0.013
36	4.856		0.0000	0.000	38295.1	1.040	BB	0.027	23997.20	1.164
37	4.913		0.0000	0.000	1977.6	0.054	BB	0.018	1816.72	0.088
38	5.002		0.0000	0.000	3151.3	0.086	BB	0.023	2304.36	0.112
39	5.057		0.0000	0.000	60060.6	1.631	BB	0.033	30330.12	1.471
40	5.232		0.0000	0.000	5451.0	0.148	BB	0.038	2365.88	0.115
41	5.310		0.0000	0.000	48060.1	1.305	BB	0.029	27438.44	1.331
42	5.374		0.0000	0.000	15392.6	0.418	BB	0.023	11360.51	0.551
43	5.428		0.0000	0.000	183841.2	4.991	BB	0.034	88924.80	4.313
44	5.520		0.0000	0.000	203185.7	5.516	BB	0.027	127541.10	6.186
45	5.636		0.0000	0.000	16850.5	0.457	BB	0.035	8115.17	0.394
46	5.702		0.0000	0.000	25267.4	0.686	BB	0.025	16782.99	0.814
47	5.765		0.0000	0.000	207085.7	5.622	BB	0.026	132311.40	6.417
48	5.812		0.0000	0.000	61098.1	1.659	BB	0.020	50545.21	2.452
49	5.876		0.0000	0.000	119942.7	3.256	BB	0.029	69618.34	3.377
50	5.968		0.0000	0.000	3241.6	0.088	BB	0.018	2991.62	0.145
51	6.010		0.0000	0.000	187205.0	5.083	BB	0.030	104115.40	5.050
52	6.082		0.0000	0.000	252491.3	6.855	BB	0.024	176538.40	8.562
53	6.134		0.0000	0.000	98401.8	2.672	BB	0.024	69713.31	3.381
54	6.249		0.0000	0.000	16153.2	0.439	BB	0.024	11247.07	0.545

ASX22 0426

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.280		0.0000	0.000	215519.7	5.851	BB	0.026	137006.50	6.645
56	6.352		0.0000	0.000	101383.2	2.753	BB	0.028	61359.50	2.976
57	6.398		0.0000	0.000	14880.3	0.404	BB	0.017	14222.16	0.690
58	6.446		0.0000	0.000	6383.4	0.173	BB	0.020	5295.69	0.257
59	6.510		0.0000	0.000	41590.7	1.129	BB	0.026	26502.28	1.285
60	6.578		0.0000	0.000	582394.3	15.812	BB	0.030	322251.20	15.630
61	6.828		0.0000	0.000	293026.1	7.956	BB	0.030	161888.70	7.852
62	6.895		0.0000	0.000	222574.5	6.043	BB	0.036	103491.90	5.020
63	7.056		0.0000	0.000	20245.8	0.550	BB	0.030	11265.42	0.546
64	7.146		0.0000	0.000	2123.1	0.058	BB	0.027	1305.67	0.063
65	7.182		0.0000	0.000	72878.6	1.979	BB	0.028	42766.07	2.074
66	7.416		0.0000	0.000	168897.9	4.586	BB	0.030	93023.08	4.512
67	7.487		0.0000	0.000	9717.9	0.264	BB	0.030	5416.39	0.263
68	7.628		0.0000	0.000	4597.8	0.125	BB	0.088	872.47	0.042
69	7.730		0.0000	0.000	96349.7	2.616	BB	0.035	46206.84	2.241
70	7.977	DCB	0.8289	-21.833	7628.7	0.207	BB	0.061	2095.44	0.102
71	8.149		0.0000	0.000	7157.8	0.194	BB	0.131	913.82	0.044
72	8.320		0.0000	0.000	76.0	0.002	BB	0.032	40.15	0.002
73	8.377		0.0000	0.000	773.2	0.021	BB	0.054	237.61	0.012
74	8.683		0.0000	0.000	689.1	0.019	BB	0.049	234.35	0.011

Total Area = 3683258.0, Total Amount = -3.797, Total Height = 2061796.0, Sample Units = ug/l

ASX22 8427

Sample Name: **AR6250924B AAAR625AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 08:15:39pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.26R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.576		0.0000	0.000	4043.6	0.088	BV	0.042	1614.02	0.077
2	1.639		0.0000	0.000	13813.9	0.302	VV	0.091	2534.67	0.121
3	2.017		0.0000	0.000	1171.0	0.026	VB	0.049	395.54	0.019
4	2.113		0.0000	0.000	186.6	0.004	BB	0.023	135.50	0.006
5	2.280		0.0000	0.000	3629.4	0.079	BB	0.046	1313.86	0.063
6	2.393		0.0000	0.000	535.7	0.012	BB	0.040	225.21	0.011
7	2.560		0.0000	0.000	379.1	0.008	BB	0.029	215.07	0.010
8	2.650		0.0000	0.000	613.5	0.013	BB	0.033	314.27	0.015
9	2.715		0.0000	0.000	5049.6	0.110	BB	0.044	1922.38	0.092
10	2.860		0.0000	0.000	800.9	0.017	BB	0.038	352.03	0.017
11	3.008		0.0000	0.000	14313.4	0.313	BB	0.049	4900.43	0.234
12	3.172		0.0000	0.000	8696.8	0.190	BB	0.042	3459.48	0.165
13	3.314		0.0000	0.000	963.5	0.021	BB	0.039	414.63	0.020
14	3.347		0.0000	0.000	158.4	0.003	BB	0.014	184.15	0.009
15	3.405		0.0000	0.000	8734.4	0.191	BB	0.034	4342.76	0.207
16	3.452		0.0000	0.000	1801.8	0.039	BB	0.028	1086.10	0.052
17	3.557		0.0000	0.000	3467.6	0.076	BB	0.031	1866.44	0.089
18	3.620		0.0000	0.000	15954.0	0.348	BB	0.067	3974.66	0.190
19	3.799		0.0000	0.000	10037.3	0.219	BB	0.052	3221.68	0.154
20	3.924		0.0000	0.000	1952.1	0.043	BB	0.030	1080.62	0.052
21	3.973		0.0000	0.000	298.1	0.007	BB	0.021	241.07	0.012
22	4.108		0.0000	0.000	64611.2	1.411	BB	0.042	25829.68	1.233
23	4.203		0.0000	0.000	1103.7	0.024	BB	0.019	966.20	0.046
24	4.253		0.0000	0.000	77842.6	1.700	BB	0.049	26483.93	1.264
25	4.433		0.0000	0.000	3122.7	0.068	BB	0.033	1581.08	0.075
26	4.506		0.0000	0.000	1941.7	0.042	BB	0.028	1153.43	0.055
27	4.543		0.0000	0.000	26709.5	0.583	BB	0.025	18021.02	0.860
28	4.599		0.0000	0.000	229973.8	5.022	BB	0.044	86695.33	4.139
29	4.694		0.0000	0.000	270861.8	5.915	BB	0.034	131799.10	6.292
30	4.822		0.0000	0.000	22078.3	0.482	BB	0.030	12385.85	0.591
31	4.874		0.0000	0.000	308409.8	6.735	BB	0.036	144424.70	6.895
32	4.982		0.0000	0.000	382545.8	8.353	BB	0.054	117348.20	5.602
33	5.101		0.0000	0.000	33340.9	0.728	BB	0.021	25909.62	1.237
34	5.143		0.0000	0.000	43084.2	0.941	BB	0.022	32815.67	1.567
35	5.179		0.0000	0.000	192474.0	4.203	BB	0.023	137157.80	6.548
36	5.218		0.0000	0.000	42073.3	0.919	BB	0.019	36661.49	1.750
37	5.338		0.0000	0.000	46098.1	1.007	BB	0.024	32002.66	1.528
38	5.390		0.0000	0.000	344014.1	7.512	BB	0.032	179029.10	8.547
39	5.473		0.0000	0.000	183277.5	4.002	BB	0.041	75088.75	3.585
40	5.562		0.0000	0.000	20463.9	0.447	BB	0.023	14723.40	0.703
41	5.628		0.0000	0.000	862440.9	18.833	BB	0.042	344198.30	16.431
42	5.795		0.0000	0.000	349.6	0.008	BB	0.015	377.74	0.018
43	5.868		0.0000	0.000	182405.0	3.983	BB	0.025	122366.10	5.842
44	5.925		0.0000	0.000	451888.2	9.868	BB	0.035	216480.50	10.334
45	6.091		0.0000	0.000	25074.9	0.548	BB	0.034	12309.59	0.588
46	6.173		0.0000	0.000	15975.9	0.349	BB	0.030	8862.55	0.423
47	6.240		0.0000	0.000	132251.2	2.888	BB	0.037	59013.20	2.817
48	6.384		0.0000	0.000	364756.6	7.965	BB	0.045	136268.20	6.505
49	6.622		0.0000	0.000	140946.2	3.078	BB	0.040	58595.79	2.797
50	6.803	DCB	-3.9253	100.000	3073.3	0.067	BB	0.034	1500.73	0.072
51	7.060		0.0000	0.000	434.1	0.009	BB	0.034	210.83	0.010
52	7.510		0.0000	0.000	1058.9	0.023	BB	0.067	263.10	0.013
53	7.623		0.0000	0.000	587.4	0.013	BB	0.068	143.36	0.007
54	8.283		0.0000	0.000	7614.5	0.166	BB	0.436	290.92	0.014

ASX22 0428

Total Area = 4579485.0, Total Amount = -3.925, Total Height = 2094753.0, Sample Units = ug/l

ASX22 0429

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.26R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 8:06:30 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.26R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 8:06:30 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
				.	0	.	TCX
3.012	575	-4.62558	TCX				
7.977	2095	.82891	DCB	6.803	1501	-3.92528	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.26A
Area File: C:\CPWIN\DATA1\1P7307B.26A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 8:15:46 PM
File Reported On: 11/3/2009 at 8:15:59 PM

10X REPLOTS

ASX22 0431

C:\CPWIN\DATA1\1P7307.26R



C:\CPWIN\DATA1\1P7307B.26R



ASX22: 0432:

Sample Name: **AR6810924C AAAR681AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 08:28:38pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.27R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.581		0.0000	0.000	2776.3	0.501	BV	0.056	822.63	0.392
2	1.665		0.0000	0.000	7182.7	1.295	VV	0.069	1745.89	0.832
3	1.787		0.0000	0.000	19394.0	3.497	VV	0.113	2861.14	1.364
4	1.861		0.0000	0.000	13013.6	2.347	VV	0.063	3451.96	1.645
5	1.929		0.0000	0.000	14109.2	2.544	VB	0.054	4370.40	2.083
6	2.056		0.0000	0.000	27945.2	5.039	BV	0.110	4243.99	2.023
7	2.174		0.0000	0.000	40387.9	7.283	VV	0.097	6973.86	3.324
8	2.314		0.0000	0.000	11865.2	2.140	VV	0.099	2001.65	0.954
9	2.582		0.0000	0.000	8488.9	1.531	VB	0.287	493.28	0.235
10	2.727		0.0000	0.000	255.2	0.046	BB	0.047	90.86	0.043
11	2.777		0.0000	0.000	1645.3	0.297	BB	0.035	777.24	0.370
12	2.825		0.0000	0.000	14102.8	2.543	BB	0.088	2671.85	1.274
13	2.983	TCX	-4.5409	-462.985	5067.0	0.914	BB	0.092	920.94	0.439
14	3.023		0.0000	0.000	115.4	0.021	BB	0.017	116.31	0.055
15	3.065		0.0000	0.000	154.8	0.028	BB	0.026	99.82	0.048
16	3.212		0.0000	0.000	311.7	0.056	BB	0.054	96.47	0.046
17	3.389		0.0000	0.000	306.4	0.055	BB	0.029	174.58	0.083
18	3.606		0.0000	0.000	2132.1	0.384	BB	0.170	209.18	0.100
19	3.773		0.0000	0.000	630.9	0.114	BB	0.068	153.82	0.073
20	3.861		0.0000	0.000	1045.0	0.188	BB	0.036	477.77	0.228
21	4.032		0.0000	0.000	2515.8	0.454	BB	0.083	504.27	0.240
22	4.265		0.0000	0.000	3391.0	0.612	BB	0.065	875.18	0.417
23	4.431		0.0000	0.000	313.3	0.056	BB	0.062	84.86	0.040
24	4.460		0.0000	0.000	356.2	0.064	BB	0.030	197.57	0.094
25	4.567		0.0000	0.000	651.2	0.117	BB	0.058	187.23	0.089
26	4.655		0.0000	0.000	1667.2	0.301	BB	0.090	310.22	0.148
27	4.811		0.0000	0.000	4015.1	0.724	BB	0.071	940.64	0.448
28	4.926		0.0000	0.000	1099.5	0.198	BB	0.043	427.18	0.204
29	4.994		0.0000	0.000	723.3	0.130	BB	0.034	353.71	0.169
30	5.107		0.0000	0.000	4935.0	0.890	BB	0.045	1810.02	0.863
31	5.205		0.0000	0.000	1846.5	0.333	BB	0.039	781.10	0.372
32	5.341		0.0000	0.000	263.2	0.047	BB	0.035	126.35	0.060
33	5.383		0.0000	0.000	690.0	0.124	BB	0.028	411.79	0.196
34	5.482		0.0000	0.000	2216.2	0.400	BB	0.090	412.55	0.197
35	5.633		0.0000	0.000	977.1	0.176	BB	0.045	360.02	0.172
36	5.814		0.0000	0.000	7158.7	1.291	BB	0.092	1302.53	0.621
37	5.900		0.0000	0.000	2236.3	0.403	BB	0.034	1081.37	0.515
38	5.969		0.0000	0.000	1059.6	0.191	BB	0.034	521.60	0.249
39	6.084		0.0000	0.000	8395.3	1.514	BB	0.029	4854.92	2.314
40	6.292		0.0000	0.000	11382.7	2.053	BB	0.030	6376.74	3.039
41	6.372		0.0000	0.000	2787.0	0.503	BB	0.028	1680.12	0.801
42	6.442		0.0000	0.000	4359.4	0.786	BB	0.083	875.84	0.417
43	6.583		0.0000	0.000	6176.3	1.114	BB	0.032	3169.14	1.511
44	6.728		0.0000	0.000	6080.1	1.096	BB	0.073	1387.43	0.661
45	6.824		0.0000	0.000	41678.8	7.516	BB	0.028	24549.96	11.702
46	6.900		0.0000	0.000	36820.7	6.640	BB	0.029	20923.16	9.973
47	7.056		0.0000	0.000	36445.8	6.572	BB	0.030	20324.99	9.688
48	7.145		0.0000	0.000	4733.7	0.854	BB	0.032	2452.48	1.169
49	7.417		0.0000	0.000	12586.7	2.270	BB	0.031	6844.10	3.262
50	7.509		0.0000	0.000	3226.1	0.582	BB	0.044	1210.51	0.577
51	7.661		0.0000	0.000	6836.3	1.233	BB	0.176	646.94	0.308
52	7.731		0.0000	0.000	113918.8	20.544	BB	0.035	54468.18	25.962
53	7.978	DCB	5.5217	562.985	33400.5	6.023	BB	0.040	13958.53	6.653
54	8.173		0.0000	0.000	9932.0	1.791	BB	0.181	912.59	0.435

ASX22 8433

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	8.303		0.0000	0.000	617.4	0.111	BB	0.100	103.34	0.049
56	8.382		0.0000	0.000	1855.5	0.335	BB	0.128	241.40	0.115
57	8.582		0.0000	0.000	1839.0	0.332	BB	0.042	726.44	0.346
58	8.693		0.0000	0.000	2101.5	0.379	BB	0.067	522.80	0.249
59	9.111		0.0000	0.000	2305.5	0.416	BB	0.296	129.62	0.062

Total Area = 554523.6, Total Amount = 0.981, Total Height = 209801.0, Sample Units = ug/l

ASX22 0434

Sample Name: **AR6810924C AAAR681AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 08:28:38pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.27R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.569		0.0000	0.000	9687.2	1.930	BV	0.043	3766.40	1.816
2	1.639		0.0000	0.000	13901.5	2.769	VV	0.098	2363.59	1.140
3	2.017		0.0000	0.000	758.2	0.151	VB	0.044	285.06	0.137
4	2.108		0.0000	0.000	94.0	0.019	BB	0.024	64.41	0.031
5	2.186		0.0000	0.000	577.1	0.115	BB	0.058	166.98	0.081
6	2.282		0.0000	0.000	10365.3	2.065	BB	0.061	2840.64	1.370
7	2.527		0.0000	0.000	669.4	0.133	BB	0.065	171.02	0.082
8	3.037		0.0000	0.000	159.8	0.032	BB	0.067	39.91	0.019
9	3.107		0.0000	0.000	94.8	0.019	BB	0.041	38.11	0.018
10	3.157		0.0000	0.000	2217.4	0.442	BB	0.057	646.18	0.312
11	3.356		0.0000	0.000	2048.1	0.408	BB	0.075	452.62	0.218
12	3.560		0.0000	0.000	9299.5	1.852	BB	0.174	890.84	0.430
13	4.146		0.0000	0.000	1698.1	0.338	BB	0.086	328.35	0.158
14	4.307		0.0000	0.000	7675.9	1.529	BB	0.094	1365.94	0.659
15	4.697		0.0000	0.000	3540.1	0.705	BB	0.214	275.13	0.133
16	4.830		0.0000	0.000	876.6	0.175	BB	0.035	417.58	0.201
17	4.914		0.0000	0.000	407.4	0.081	BB	0.040	170.06	0.082
18	4.969		0.0000	0.000	346.8	0.069	BB	0.026	218.66	0.105
19	5.050		0.0000	0.000	8.1	0.002	BB	0.022	6.26	0.003
20	5.128		0.0000	0.000	559.9	0.112	BB	0.034	273.16	0.132
21	5.183		0.0000	0.000	12035.4	2.397	BB	0.038	5313.52	2.562
22	5.384		0.0000	0.000	16259.8	3.239	BB	0.036	7626.38	3.677
23	5.458		0.0000	0.000	4111.6	0.819	BB	0.048	1417.99	0.684
24	5.630		0.0000	0.000	4259.5	0.848	BB	0.032	2242.05	1.081
25	5.685		0.0000	0.000	1866.9	0.372	BB	0.029	1073.32	0.518
26	5.804		0.0000	0.000	1278.8	0.255	BB	0.026	809.64	0.390
27	5.870		0.0000	0.000	43997.4	8.764	BB	0.030	24426.65	11.778
28	5.928		0.0000	0.000	39949.5	7.958	BB	0.031	21469.35	10.352
29	6.091		0.0000	0.000	49800.7	9.920	BB	0.035	23465.08	11.314
30	6.173		0.0000	0.000	9817.5	1.956	BB	0.032	5185.16	2.500
31	6.237		0.0000	0.000	1620.9	0.323	BB	0.103	263.40	0.127
32	6.386		0.0000	0.000	29579.2	5.892	BB	0.051	9704.78	4.679
33	6.622		0.0000	0.000	169227.2	33.709	BB	0.040	70220.88	33.859
34	6.803	DCB	1.8125	100.000	43496.1	8.664	BB	0.041	17656.19	8.513
35	7.356		0.0000	0.000	443.4	0.088	BB	0.041	180.86	0.087
36	7.480		0.0000	0.000	1507.8	0.300	BB	0.081	311.62	0.150
37	7.606		0.0000	0.000	3977.3	0.792	BB	0.184	359.36	0.173
38	8.017		0.0000	0.000	567.2	0.113	BB	0.100	94.64	0.046
39	8.080		0.0000	0.000	529.3	0.105	BB	0.048	184.78	0.089
40	8.392		0.0000	0.000	254.7	0.051	BB	0.029	147.70	0.071
41	8.546		0.0000	0.000	1080.7	0.215	BB	0.082	218.73	0.105
42	8.670		0.0000	0.000	1011.8	0.202	BB	0.106	159.16	0.077
43	9.040		0.0000	0.000	358.9	0.071	BB	0.072	82.55	0.040

Total Area = 502016.6, Total Amount = 1.813, Total Height = 207394.7, Sample Units = ug/l

ASX22 0435

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.27R

C:\CPWIN\DATA1\1P7307.27R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 8:19:29 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.27R

C:\CPWIN\DATA1\1P7307B.27R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 8:19:29 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.983	921	-4.5409	TCX	.	0	.	TCX
7.978	13959	5.52169	DCB	6.803	17656	1.81251	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.27A
Area File: C:\CPWIN\DATA1\1P7307B.27A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 8:28:46 PM
File Reported On: 11/3/2009 at 8:28:58 PM

10X REPLOTS

ASX22 0437

C:\CPWIN\DATA1\1P7307.27R



C:\CPWIN\DATA1\1P7307B.27R



ASX22 9438

Sample Name: **AR6820924C AAAR682AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 08:41:39pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.28R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.580		0.0000	0.000	2930.4	0.385	BV	0.062	784.84	0.235
2	1.797		0.0000	0.000	20954.6	2.752	VV	0.150	2323.14	0.696
3	1.927		0.0000	0.000	22224.2	2.919	VB	0.108	3439.95	1.030
4	2.034		0.0000	0.000	18783.4	2.467	BV	0.095	3303.51	0.989
5	2.174		0.0000	0.000	26805.5	3.521	VV	0.131	3408.39	1.021
6	2.318		0.0000	0.000	6984.1	0.917	VV	0.104	1121.71	0.336
7	2.364		0.0000	0.000	2636.5	0.346	VV	0.052	844.53	0.253
8	2.451		0.0000	0.000	1856.1	0.244	VV	0.084	368.29	0.110
9	2.526		0.0000	0.000	531.4	0.070	VB	0.041	218.09	0.065
10	2.714		0.0000	0.000	143.4	0.019	BB	0.033	72.54	0.022
11	2.774		0.0000	0.000	1008.8	0.133	BB	0.038	445.18	0.133
12	2.841		0.0000	0.000	14742.0	1.936	BB	0.110	2230.68	0.668
13	3.058		0.0000	0.000	24055.7	3.160	BB	0.274	1461.29	0.438
14	3.209		0.0000	0.000	389.3	0.051	BB	0.059	109.60	0.033
15	3.287		0.0000	0.000	236.0	0.031	BB	0.035	110.99	0.033
16	3.388		0.0000	0.000	144.8	0.019	BB	0.025	97.71	0.029
17	3.463		0.0000	0.000	176.4	0.023	BB	0.050	59.36	0.018
18	3.596		0.0000	0.000	725.7	0.095	BB	0.098	123.69	0.037
19	3.722		0.0000	0.000	273.1	0.036	BB	0.033	138.05	0.041
20	3.772		0.0000	0.000	233.3	0.031	BB	0.042	92.83	0.028
21	3.859		0.0000	0.000	1264.3	0.166	BB	0.053	394.31	0.118
22	3.984		0.0000	0.000	204.3	0.027	BB	0.027	124.40	0.037
23	4.031		0.0000	0.000	591.0	0.078	BB	0.041	239.03	0.072
24	4.104		0.0000	0.000	106.1	0.014	BB	0.020	89.87	0.027
25	4.201		0.0000	0.000	576.5	0.076	BB	0.055	173.71	0.052
26	4.264		0.0000	0.000	898.5	0.118	BB	0.028	536.15	0.161
27	4.360		0.0000	0.000	87.2	0.011	BB	0.020	74.23	0.022
28	4.426		0.0000	0.000	1216.9	0.160	BB	0.066	305.30	0.091
29	4.527		0.0000	0.000	504.5	0.066	BB	0.055	152.53	0.046
30	4.651		0.0000	0.000	1117.0	0.147	BB	0.074	250.73	0.075
31	4.810		0.0000	0.000	1479.5	0.194	BB	0.056	442.06	0.132
32	4.919		0.0000	0.000	1016.8	0.134	BB	0.032	525.88	0.157
33	4.992		0.0000	0.000	449.9	0.059	BB	0.029	254.26	0.076
34	5.103		0.0000	0.000	4534.6	0.596	BB	0.050	1500.53	0.449
35	5.203		0.0000	0.000	889.1	0.117	BB	0.029	509.96	0.153
36	5.380		0.0000	0.000	2808.8	0.369	BB	0.071	657.85	0.197
37	5.481		0.0000	0.000	298.4	0.039	BB	0.026	191.20	0.057
38	5.635		0.0000	0.000	317.2	0.042	BB	0.027	193.04	0.058
39	5.768		0.0000	0.000	805.8	0.106	BB	0.084	160.50	0.048
40	5.812		0.0000	0.000	1298.8	0.171	BB	0.029	753.97	0.226
41	5.898		0.0000	0.000	1611.4	0.212	BB	0.045	597.45	0.179
42	5.968		0.0000	0.000	181.2	0.024	BB	0.019	155.51	0.047
43	6.083		0.0000	0.000	15318.6	2.012	BB	0.027	9556.21	2.862
44	6.291		0.0000	0.000	22593.7	2.968	BB	0.030	12657.97	3.791
45	6.371		0.0000	0.000	5566.6	0.731	BB	0.029	3168.16	0.949
46	6.441		0.0000	0.000	1572.2	0.206	BB	0.066	394.88	0.118
47	6.583		0.0000	0.000	12054.2	1.583	BB	0.033	6135.71	1.838
48	6.725		0.0000	0.000	1879.1	0.247	BB	0.066	476.92	0.143
49	6.823		0.0000	0.000	76306.0	10.022	BB	0.028	45966.43	13.766
50	6.898		0.0000	0.000	70061.7	9.202	BB	0.029	39822.66	11.926
51	7.054		0.0000	0.000	67438.6	8.858	BB	0.029	38546.81	11.544
52	7.143		0.0000	0.000	14966.0	1.966	BB	0.032	7780.77	2.330
53	7.416		0.0000	0.000	27320.2	3.588	BB	0.032	14412.14	4.316
54	7.637		0.0000	0.000	4434.4	0.582	BB	0.103	720.79	0.216

ASX22 0439

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	7.729		0.0000	0.000	211696.1	27.805	BB	0.036	99108.55	29.681
56	7.977	DCB	9.9285	100.000	55782.3	7.327	BB	0.037	25098.77	7.516
57	8.150		0.0000	0.000	4603.8	0.605	BB	0.139	553.32	0.166
58	8.381		0.0000	0.000	1043.3	0.137	BB	0.086	201.27	0.060
59	8.512		0.0000	0.000	84.4	0.011	BB	0.038	36.91	0.011
60	8.604		0.0000	0.000	545.6	0.072	BB	0.038	240.99	0.072

Total Area = 761359.4, Total Amount = 9.929, Total Height = 333916.1, Sample Units = ug/l

ASX22 0440

Sample Name: **AR6820924C AAAR682AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 08:41:39pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.28R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.571		0.0000	0.000	3467.0	0.408	BV	0.040	1436.50	0.385
2	1.638		0.0000	0.000	11280.5	1.329	VB	0.085	2202.30	0.590
3	2.014		0.0000	0.000	412.0	0.049	BB	0.038	181.94	0.049
4	2.113		0.0000	0.000	137.8	0.016	BB	0.021	107.90	0.029
5	2.190		0.0000	0.000	1880.4	0.222	BB	0.095	329.70	0.088
6	2.391		0.0000	0.000	2042.2	0.241	BB	0.892	38.15	0.010
7	2.468		0.0000	0.000	112.4	0.013	BB	0.112	16.71	0.004
8	2.517		0.0000	0.000	506.1	0.060	BB	0.062	136.74	0.037
9	2.747		0.0000	0.000	149.7	0.018	BB	0.042	59.26	0.016
10	2.963		0.0000	0.000	274.8	0.032	BB	0.084	54.33	0.015
11	3.040		0.0000	0.000	112.1	0.013	BB	0.049	38.33	0.010
12	3.159		0.0000	0.000	988.5	0.116	BB	0.048	343.07	0.092
13	3.353		0.0000	0.000	471.7	0.056	BB	0.033	237.09	0.064
14	3.396		0.0000	0.000	127.4	0.015	BB	0.028	76.19	0.020
15	3.556		0.0000	0.000	3416.5	0.402	BB	0.097	587.49	0.158
16	3.890		0.0000	0.000	1434.3	0.169	BB	0.108	220.47	0.059
17	4.023		0.0000	0.000	92.6	0.011	BB	0.024	65.24	0.017
18	4.141		0.0000	0.000	2002.7	0.236	BB	0.084	396.15	0.106
19	4.302		0.0000	0.000	5587.4	0.658	BB	0.088	1060.99	0.284
20	4.595		0.0000	0.000	1341.0	0.158	BB	0.075	297.02	0.080
21	4.693		0.0000	0.000	412.9	0.049	BB	0.038	180.71	0.048
22	4.880		0.0000	0.000	492.1	0.058	BB	0.103	79.31	0.021
23	4.944		0.0000	0.000	70.7	0.008	BB	0.047	24.86	0.007
24	4.971		0.0000	0.000	1518.5	0.179	BB	0.034	753.63	0.202
25	5.103		0.0000	0.000	509.0	0.060	BB	0.033	255.71	0.069
26	5.182		0.0000	0.000	27995.6	3.298	BB	0.042	11127.22	2.983
27	5.334		0.0000	0.000	240.9	0.028	BB	0.018	221.76	0.059
28	5.383		0.0000	0.000	31187.4	3.674	BB	0.035	14909.37	3.998
29	5.459		0.0000	0.000	5973.1	0.704	BB	0.031	3195.25	0.857
30	5.628		0.0000	0.000	7782.2	0.917	BB	0.031	4176.32	1.120
31	5.684		0.0000	0.000	3863.6	0.455	BB	0.029	2225.65	0.597
32	5.801		0.0000	0.000	761.9	0.090	BB	0.023	546.79	0.147
33	5.869		0.0000	0.000	83433.4	9.828	BB	0.030	46986.57	12.598
34	5.928		0.0000	0.000	78841.9	9.287	BB	0.032	41071.02	11.012
35	6.090		0.0000	0.000	95007.7	11.192	BB	0.035	45326.94	12.153
36	6.172		0.0000	0.000	19767.3	2.329	BB	0.032	10414.36	2.792
37	6.385		0.0000	0.000	50247.8	5.919	BB	0.046	18393.44	4.932
38	6.621		0.0000	0.000	314334.1	37.028	BB	0.040	130720.40	35.049
39	6.802	DCB	7.3720	100.000	77038.0	9.075	BB	0.039	33309.75	8.931
40	7.487		0.0000	0.000	1514.3	0.178	BB	0.088	285.28	0.076
41	7.620		0.0000	0.000	1876.6	0.221	BB	0.105	298.74	0.080
42	7.946		0.0000	0.000	90.6	0.011	BB	0.026	57.09	0.015
43	8.017		0.0000	0.000	1731.7	0.204	BB	0.130	221.32	0.059
44	8.260		0.0000	0.000	8376.0	0.987	BB	0.470	297.13	0.080

Total Area = 848904.2, Total Amount = 7.372, Total Height = 372964.2, Sample Units = ug/l

ASX22 8441

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.28R

C:\CPWIN\DATA1\1P7307.28R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 8:32:30 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.28R

C:\CPWIN\DATA1\1P7307B.28R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 8:32:30 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
7.977	25099	9.92853	DCB	6.802	33310	7.37205	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.28A
Area File: C:\CPWIN\DATA1\1P7307B.28A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 8:41:46 PM
File Reported On: 11/3/2009 at 8:41:59 PM

10X REPLOTS

ASX22 0443

C:\CPWIN\DATA1\1P7307.28R



C:\CPWIN\DATA1\1P7307B.28R



ASX22 8444

Sample Name: **AR6830924C AAAR683AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 08:54:36pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.29R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.584		0.0000	0.000	2774.1	0.222	BV	0.055	844.54	0.142
2	1.778		0.0000	0.000	25828.8	2.065	VV	0.137	3134.39	0.528
3	1.870		0.0000	0.000	14621.6	1.169	VV	0.059	4120.97	0.695
4	1.928		0.0000	0.000	17489.6	1.398	VB	0.058	5038.54	0.849
5	2.030		0.0000	0.000	27855.3	2.227	BV	0.094	4956.64	0.835
6	2.126		0.0000	0.000	18374.1	1.469	VV	0.075	4096.49	0.690
7	2.175		0.0000	0.000	24145.2	1.931	VV	0.088	4587.97	0.773
8	2.273		0.0000	0.000	6787.2	0.543	VV	0.043	2619.63	0.442
9	2.369		0.0000	0.000	7710.2	0.616	VV	0.072	1777.51	0.300
10	2.450		0.0000	0.000	4175.0	0.334	VV	0.059	1181.46	0.199
11	2.581		0.0000	0.000	2699.2	0.216	VB	0.155	290.26	0.049
12	2.692		0.0000	0.000	1310.6	0.105	BB	0.044	492.23	0.083
13	2.773		0.0000	0.000	1188.6	0.095	BB	0.038	518.65	0.087
14	2.869		0.0000	0.000	6776.8	0.542	BB	0.064	1758.31	0.296
15	2.977	TCX	-4.5412	-35.242	4221.9	0.338	BB	0.077	919.64	0.155
16	3.063		0.0000	0.000	165.0	0.013	BB	0.030	91.15	0.015
17	3.210		0.0000	0.000	582.1	0.047	BB	0.073	133.01	0.022
18	3.286		0.0000	0.000	479.3	0.038	BB	0.033	242.00	0.041
19	3.333		0.0000	0.000	387.1	0.031	BB	0.023	284.37	0.048
20	3.392		0.0000	0.000	362.0	0.029	BB	0.042	145.31	0.024
21	3.522		0.0000	0.000	861.4	0.069	BB	0.126	114.12	0.019
22	3.601		0.0000	0.000	177.0	0.014	BB	0.042	71.02	0.012
23	3.722		0.0000	0.000	392.9	0.031	BB	0.044	149.09	0.025
24	3.767		0.0000	0.000	249.5	0.020	BB	0.031	132.89	0.022
25	3.863		0.0000	0.000	622.1	0.050	BB	0.034	301.90	0.051
26	3.970		0.0000	0.000	368.2	0.029	BB	0.027	224.52	0.038
27	4.030		0.0000	0.000	483.4	0.039	BB	0.042	190.70	0.032
28	4.106		0.0000	0.000	809.1	0.065	BB	0.037	362.05	0.061
29	4.185		0.0000	0.000	490.5	0.039	BB	0.031	259.68	0.044
30	4.265		0.0000	0.000	1160.1	0.093	BB	0.028	683.93	0.115
31	4.361		0.0000	0.000	152.3	0.012	BB	0.018	142.65	0.024
32	4.399		0.0000	0.000	175.9	0.014	BB	0.024	120.80	0.020
33	4.432		0.0000	0.000	953.4	0.076	BB	0.031	508.13	0.086
34	4.527		0.0000	0.000	498.6	0.040	BB	0.031	269.66	0.045
35	4.650		0.0000	0.000	475.8	0.038	BB	0.030	260.73	0.044
36	4.814		0.0000	0.000	2288.3	0.183	BB	0.060	634.67	0.107
37	4.919		0.0000	0.000	1317.0	0.105	BB	0.033	656.42	0.111
38	4.994		0.0000	0.000	663.3	0.053	BB	0.035	318.62	0.054
39	5.104		0.0000	0.000	5286.1	0.423	BB	0.052	1703.42	0.287
40	5.204		0.0000	0.000	1780.9	0.142	BB	0.040	744.12	0.125
41	5.379		0.0000	0.000	1947.3	0.156	BB	0.056	581.64	0.098
42	5.494		0.0000	0.000	2441.9	0.195	BB	0.064	634.75	0.107
43	5.637		0.0000	0.000	993.4	0.079	BB	0.039	420.60	0.071
44	5.765		0.0000	0.000	897.0	0.072	BB	0.049	302.85	0.051
45	5.813		0.0000	0.000	3281.4	0.262	BB	0.033	1644.36	0.277
46	5.899		0.0000	0.000	2132.4	0.170	BB	0.042	838.55	0.141
47	5.968		0.0000	0.000	183.3	0.015	BB	0.018	166.83	0.028
48	5.997		0.0000	0.000	181.3	0.014	BB	0.019	161.27	0.027
49	6.083		0.0000	0.000	29120.0	2.328	BB	0.026	18323.58	3.089
50	6.291		0.0000	0.000	41303.4	3.302	BB	0.029	23565.46	3.972
51	6.371		0.0000	0.000	10731.6	0.858	BB	0.028	6326.52	1.066
52	6.444		0.0000	0.000	1788.3	0.143	BB	0.056	535.62	0.090
53	6.583		0.0000	0.000	24726.6	1.977	BB	0.034	12117.49	2.043
54	6.730		0.0000	0.000	4652.9	0.372	BB	0.081	959.18	0.162

ASX22 0445

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.822		0.0000	0.000	136628.8	10.924	BB	0.027	83034.80	13.996
56	6.898		0.0000	0.000	127943.9	10.230	BB	0.029	72532.77	12.226
57	7.055		0.0000	0.000	122451.5	9.790	BB	0.029	69595.78	11.731
58	7.143		0.0000	0.000	37979.4	3.037	BB	0.040	15713.39	2.649
59	7.417		0.0000	0.000	27382.5	2.189	BB	0.028	16038.30	2.703
60	7.518		0.0000	0.000	3730.6	0.298	BB	0.046	1363.90	0.230
61	7.641		0.0000	0.000	4596.5	0.368	BB	0.107	715.26	0.121
62	7.728		0.0000	0.000	373115.7	29.832	BB	0.035	177005.80	29.836
63	7.976	DCB	17.4270	135.242	97318.4	7.781	BB	0.037	44054.38	7.426
64	8.145		0.0000	0.000	5512.5	0.441	BB	0.131	703.99	0.119
65	8.309		0.0000	0.000	58.0	0.005	BB	0.023	41.80	0.007
66	8.374		0.0000	0.000	826.8	0.066	BB	0.063	217.18	0.037
67	8.521		0.0000	0.000	342.3	0.027	BB	0.063	90.62	0.015
68	8.591		0.0000	0.000	877.4	0.070	BB	0.042	347.65	0.059
69	8.671		0.0000	0.000	434.7	0.035	BB	0.048	151.16	0.025

Total Area = 1250719.0, Total Amount = 12.886, Total Height = 593267.8, Sample Units = ug/l

ASX22 0246

Sample Name: **AR6830924C AAAR683AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 08:54:36pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.29R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.576		0.0000	0.000	3427.0	0.228	BV	0.039	1452.07	0.215
2	1.639		0.0000	0.000	14355.4	0.957	VV	0.102	2347.77	0.348
3	1.978		0.0000	0.000	770.4	0.051	VV	0.051	250.82	0.037
4	2.020		0.0000	0.000	1595.6	0.106	VB	0.049	541.50	0.080
5	2.200		0.0000	0.000	1312.8	0.088	BB	0.083	264.78	0.039
6	2.291		0.0000	0.000	1064.7	0.071	BB	0.053	334.25	0.050
7	2.397		0.0000	0.000	74.5	0.005	BB	0.037	33.45	0.005
8	2.500		0.0000	0.000	905.1	0.060	BB	0.091	166.47	0.025
9	2.740		0.0000	0.000	105.7	0.007	BB	0.023	77.04	0.011
10	3.007		0.0000	0.000	278.9	0.019	BB	0.035	133.94	0.020
11	3.160		0.0000	0.000	1213.0	0.081	BB	0.049	411.81	0.061
12	3.320		0.0000	0.000	201.0	0.013	BB	0.066	50.63	0.008
13	3.353		0.0000	0.000	2550.7	0.170	BB	0.082	515.70	0.076
14	3.557		0.0000	0.000	1315.8	0.088	BB	0.040	546.17	0.081
15	3.799		0.0000	0.000	645.4	0.043	BB	0.040	266.92	0.040
16	3.907		0.0000	0.000	474.0	0.032	BB	0.045	175.43	0.026
17	4.144		0.0000	0.000	2891.7	0.193	BB	0.083	583.22	0.086
18	4.303		0.0000	0.000	5484.4	0.366	BB	0.074	1229.41	0.182
19	4.513		0.0000	0.000	558.0	0.037	BB	0.046	204.04	0.030
20	4.596		0.0000	0.000	508.4	0.034	BB	0.034	248.88	0.037
21	4.693		0.0000	0.000	1283.2	0.086	BB	0.047	451.38	0.067
22	4.833		0.0000	0.000	625.6	0.042	BB	0.043	241.00	0.036
23	4.903		0.0000	0.000	459.5	0.031	BB	0.057	134.46	0.020
24	4.973		0.0000	0.000	2670.0	0.178	BB	0.033	1349.54	0.200
25	5.057		0.0000	0.000	215.2	0.014	BB	0.031	116.22	0.017
26	5.106		0.0000	0.000	911.2	0.061	BB	0.038	397.18	0.059
27	5.182		0.0000	0.000	51220.6	3.415	BB	0.041	20753.41	3.076
28	5.337		0.0000	0.000	384.6	0.026	BB	0.021	310.45	0.046
29	5.383		0.0000	0.000	56834.6	3.789	BB	0.034	27485.46	4.073
30	5.459		0.0000	0.000	11633.4	0.776	BB	0.031	6173.29	0.915
31	5.629		0.0000	0.000	14610.3	0.974	BB	0.031	7829.93	1.160
32	5.684		0.0000	0.000	7673.6	0.512	BB	0.029	4337.99	0.643
33	5.803		0.0000	0.000	890.1	0.059	BB	0.025	593.02	0.088
34	5.869		0.0000	0.000	149953.7	9.997	BB	0.029	86036.64	12.750
35	5.928		0.0000	0.000	142669.6	9.511	BB	0.031	75635.23	11.209
36	6.090		0.0000	0.000	170192.2	11.346	BB	0.034	83181.65	12.327
37	6.172		0.0000	0.000	50394.3	3.360	BB	0.041	20250.27	3.001
38	6.385		0.0000	0.000	101858.3	6.790	BB	0.049	34823.77	5.161
39	6.621		0.0000	0.000	553185.9	36.879	BB	0.039	234729.20	34.786
40	6.802	DCB	16.5980	100.000	137293.2	9.153	BB	0.039	59286.45	8.786
41	7.053		0.0000	0.000	279.2	0.019	BB	0.033	141.77	0.021
42	7.473		0.0000	0.000	1028.3	0.069	BB	0.099	173.69	0.026
43	7.633		0.0000	0.000	415.0	0.028	BB	0.061	113.16	0.017
44	8.217		0.0000	0.000	3199.5	0.213	BB	0.219	242.95	0.036
45	8.913		0.0000	0.000	406.0	0.027	BB	0.043	157.30	0.023

Total Area = 1500020.0, Total Amount = 16.598, Total Height = 674779.7, Sample Units = ug/l

ASX22 0447

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.29R

C:\CPWIN\DATA1\1P7307.29R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 8:45:27 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.29R

C:\CPWIN\DATA1\1P7307B.29R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 8:45:27 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.977	920	-4.54122	TCX	.	0	.	TCX
7.976	44054	17.42695	DCB	6.802	59286	16.59796	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.29A
Area File: C:\CPWIN\DATA1\1P7307B.29A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 8:54:44 PM
File Reported On: 11/3/2009 at 8:54:56 PM

10X REPLOTS

ASX22 0449

C:\CPWIN\DATA1\1P7307.29R



C:\CPWIN\DATA1\1P7307B.29R



ASX22 0450

Sample Name: **AR6840924C AAAR684AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 09:07:36pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.30R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.581		0.0000	0.000	3023.8	0.118	BV	0.060	842.15	0.065
2	1.867		0.0000	0.000	36254.4	1.420	VV	0.198	3056.50	0.237
3	1.927		0.0000	0.000	10590.8	0.415	VB	0.049	3602.34	0.279
4	2.042		0.0000	0.000	20603.3	0.807	BV	0.100	3423.19	0.265
5	2.173		0.0000	0.000	33359.8	1.307	VV	0.146	3798.31	0.294
6	2.455		0.0000	0.000	9480.4	0.371	VV	0.201	786.59	0.061
7	2.522		0.0000	0.000	1782.3	0.070	VV	0.062	481.16	0.037
8	2.583		0.0000	0.000	867.8	0.034	VB	0.080	181.47	0.014
9	2.780		0.0000	0.000	3230.3	0.127	BB	0.085	632.63	0.049
10	2.876		0.0000	0.000	8820.1	0.345	BB	0.056	2621.02	0.203
11	3.064		0.0000	0.000	18733.8	0.734	BB	0.302	1033.66	0.080
12	3.209		0.0000	0.000	441.3	0.017	BB	0.077	95.15	0.007
13	3.292		0.0000	0.000	266.3	0.010	BB	0.038	115.68	0.009
14	3.333		0.0000	0.000	319.6	0.013	BB	0.023	232.38	0.018
15	3.384		0.0000	0.000	391.1	0.015	BB	0.026	247.17	0.019
16	3.456		0.0000	0.000	372.8	0.015	BB	0.047	132.32	0.010
17	3.534		0.0000	0.000	206.2	0.008	BB	0.031	111.55	0.009
18	3.610		0.0000	0.000	134.1	0.005	BB	0.050	44.99	0.003
19	3.693		0.0000	0.000	184.7	0.007	BB	0.020	155.21	0.012
20	3.770		0.0000	0.000	155.7	0.006	BB	0.027	95.21	0.007
21	3.862		0.0000	0.000	1009.4	0.040	BB	0.034	490.55	0.038
22	3.967		0.0000	0.000	383.2	0.015	BB	0.038	166.60	0.013
23	4.028		0.0000	0.000	855.4	0.033	BB	0.044	326.93	0.025
24	4.124		0.0000	0.000	2205.1	0.086	BB	0.037	994.48	0.077
25	4.179		0.0000	0.000	812.6	0.032	BB	0.028	479.96	0.037
26	4.263		0.0000	0.000	1535.9	0.060	BB	0.028	902.29	0.070
27	4.358		0.0000	0.000	559.5	0.022	BB	0.020	477.19	0.037
28	4.395		0.0000	0.000	1501.7	0.059	BB	0.044	563.74	0.044
29	4.525		0.0000	0.000	1539.9	0.060	BB	0.034	760.97	0.059
30	4.594		0.0000	0.000	142.7	0.006	BB	0.015	154.07	0.012
31	4.648		0.0000	0.000	1041.0	0.041	BB	0.029	604.37	0.047
32	4.695		0.0000	0.000	294.7	0.012	BB	0.027	182.11	0.014
33	4.811		0.0000	0.000	573.4	0.022	BB	0.035	271.56	0.021
34	4.854		0.0000	0.000	291.4	0.011	BB	0.014	347.59	0.027
35	4.887		0.0000	0.000	329.4	0.013	BB	0.017	320.86	0.025
36	4.914		0.0000	0.000	1229.5	0.048	BB	0.022	927.85	0.072
37	4.992		0.0000	0.000	518.7	0.020	BB	0.032	273.52	0.021
38	5.051		0.0000	0.000	611.3	0.024	BB	0.021	489.72	0.038
39	5.101		0.0000	0.000	2872.6	0.113	BB	0.042	1138.77	0.088
40	5.202		0.0000	0.000	644.6	0.025	BB	0.029	371.47	0.029
41	5.300		0.0000	0.000	239.9	0.009	BB	0.024	164.98	0.013
42	5.337		0.0000	0.000	202.0	0.008	BB	0.021	161.14	0.012
43	5.375		0.0000	0.000	976.5	0.038	BB	0.025	641.63	0.050
44	5.425		0.0000	0.000	269.0	0.011	BB	0.021	208.86	0.016
45	5.478		0.0000	0.000	469.8	0.018	BB	0.023	335.88	0.026
46	5.637		0.0000	0.000	1045.7	0.041	BB	0.028	633.14	0.049
47	5.764		0.0000	0.000	1130.5	0.044	BB	0.028	682.59	0.053
48	5.812		0.0000	0.000	5193.8	0.203	BB	0.027	3229.04	0.250
49	5.883		0.0000	0.000	2280.5	0.089	BB	0.048	796.20	0.062
50	5.999		0.0000	0.000	2510.3	0.098	BB	0.035	1204.98	0.093
51	6.081		0.0000	0.000	80434.1	3.150	BB	0.030	44207.56	3.427
52	6.248		0.0000	0.000	919.1	0.036	BB	0.019	788.41	0.061
53	6.290		0.0000	0.000	90915.4	3.561	BB	0.028	54689.73	4.240
54	6.369		0.0000	0.000	25448.2	0.997	BB	0.028	15342.50	1.189

ASX22 9451

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.445		0.0000	0.000	1594.9	0.062	BB	0.033	799.39	0.062
56	6.581		0.0000	0.000	59208.8	2.319	BB	0.034	28956.67	2.245
57	6.821		0.0000	0.000	302758.0	11.857	BB	0.027	188223.90	14.593
58	6.896		0.0000	0.000	287228.6	11.249	BB	0.029	163420.90	12.670
59	7.053		0.0000	0.000	274929.8	10.767	BB	0.029	158498.90	12.288
60	7.141		0.0000	0.000	72775.0	2.850	BB	0.031	38873.05	3.014
61	7.414		0.0000	0.000	127194.5	4.981	BB	0.035	61265.70	4.750
62	7.625		0.0000	0.000	4185.2	0.164	BB	0.086	814.25	0.063
63	7.727		0.0000	0.000	829707.3	32.495	BB	0.035	396423.00	30.734
64	7.974	DCB	38.2166	100.000	209868.7	8.219	BB	0.036	96609.35	7.490
65	8.094		0.0000	0.000	450.3	0.018	BB	0.042	179.54	0.014
66	8.143		0.0000	0.000	872.4	0.034	BB	0.043	337.10	0.026
67	8.598		0.0000	0.000	2374.9	0.093	BB	0.094	421.83	0.033

Total Area = 2553354.0, Total Amount = 38.217, Total Height = 1289844.0, Sample Units = ug/l

ASX22 8452

Sample Name: **AR6840924C AAAR684AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 09:07:36pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.30R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.571		0.0000	0.000	3565.9	0.109	BV	0.041	1439.19	0.093
2	1.639		0.0000	0.000	12034.7	0.368	VV	0.090	2220.16	0.143
3	2.016		0.0000	0.000	1377.6	0.042	VB	0.050	458.73	0.030
4	2.107		0.0000	0.000	117.8	0.004	BB	0.018	110.30	0.007
5	2.183		0.0000	0.000	1682.6	0.051	BB	0.111	253.37	0.016
6	2.385		0.0000	0.000	1217.2	0.037	BB	0.577	35.15	0.002
7	2.643		0.0000	0.000	497.2	0.015	BB	0.143	57.77	0.004
8	2.830		0.0000	0.000	895.1	0.027	BB	0.161	92.72	0.006
9	2.964		0.0000	0.000	48.1	0.001	BB	0.019	41.28	0.003
10	3.006		0.0000	0.000	1037.5	0.032	BB	0.039	440.30	0.028
11	3.167		0.0000	0.000	1846.6	0.056	BB	0.053	581.33	0.038
12	3.327		0.0000	0.000	221.1	0.007	BB	0.088	41.66	0.003
13	3.353		0.0000	0.000	174.9	0.005	BB	0.024	123.58	0.008
14	3.403		0.0000	0.000	1197.2	0.037	BB	0.034	587.61	0.038
15	3.556		0.0000	0.000	1346.2	0.041	BB	0.036	624.12	0.040
16	3.628		0.0000	0.000	256.3	0.008	BB	0.077	55.65	0.004
17	3.641		0.0000	0.000	616.1	0.019	BB	0.045	230.58	0.015
18	3.794		0.0000	0.000	1921.7	0.059	BB	0.047	679.70	0.044
19	3.920		0.0000	0.000	590.1	0.018	BB	0.032	307.95	0.020
20	4.144		0.0000	0.000	5860.4	0.179	BB	0.082	1189.63	0.077
21	4.297		0.0000	0.000	7639.4	0.233	BB	0.083	1534.20	0.099
22	4.507		0.0000	0.000	740.1	0.023	BB	0.040	306.05	0.020
23	4.590		0.0000	0.000	1535.7	0.047	BB	0.043	596.29	0.038
24	4.695		0.0000	0.000	1126.6	0.034	BB	0.040	464.70	0.030
25	4.800		0.0000	0.000	852.7	0.026	BB	0.048	294.43	0.019
26	4.876		0.0000	0.000	1666.9	0.051	BB	0.039	705.15	0.046
27	4.972		0.0000	0.000	7464.7	0.228	BB	0.035	3583.83	0.231
28	5.046		0.0000	0.000	347.3	0.011	BB	0.027	217.84	0.014
29	5.101		0.0000	0.000	2238.0	0.068	BB	0.033	1128.50	0.073
30	5.181		0.0000	0.000	121175.4	3.701	BB	0.041	49304.07	3.182
31	5.336		0.0000	0.000	1136.7	0.035	BB	0.024	795.86	0.051
32	5.382		0.0000	0.000	126565.5	3.866	BB	0.033	63025.07	4.068
33	5.457		0.0000	0.000	28497.2	0.870	BB	0.031	15097.66	0.974
34	5.628		0.0000	0.000	35854.9	1.095	BB	0.030	19810.25	1.279
35	5.683		0.0000	0.000	18778.2	0.574	BB	0.029	10823.97	0.699
36	5.800		0.0000	0.000	194.1	0.006	BB	0.018	182.54	0.012
37	5.867		0.0000	0.000	332384.9	10.153	BB	0.028	201376.20	12.998
38	5.926		0.0000	0.000	327049.7	9.990	BB	0.031	175987.20	11.359
39	6.088		0.0000	0.000	377924.8	11.544	BB	0.033	192181.70	12.404
40	6.170		0.0000	0.000	86305.2	2.636	BB	0.031	46288.47	2.988
41	6.383		0.0000	0.000	207746.0	6.346	BB	0.043	80888.99	5.221
42	6.619		0.0000	0.000	1235192.0	37.730	BB	0.038	542144.80	34.992
43	6.800	DCB	42.3091	100.000	296335.6	9.052	BB	0.038	131679.20	8.499
44	7.053		0.0000	0.000	526.2	0.016	BB	0.032	277.08	0.018
45	7.240		0.0000	0.000	733.1	0.022	BB	0.035	347.79	0.022
46	7.617		0.0000	0.000	10025.8	0.306	BB	0.394	424.43	0.027
47	8.280		0.0000	0.000	7187.6	0.220	BB	0.408	293.55	0.019

Total Area = 3273728.0, Total Amount = 42.309, Total Height = 1549331.0, Sample Units = ug/l

ASX22 8453

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.30R

C:\CPWIN\DATA1\1P7307.30R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 8:58:27 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.30R

C:\CPWIN\DATA1\1P7307B.30R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 8:58:27 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
7.974	96609	38.21656	DCB	6.8	131679	42.30906	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.30A
Area File: C:\CPWIN\DATA1\1P7307B.30A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 9:07:44 PM
File Reported On: 11/3/2009 at 9:07:56 PM

10X REPLOTS

ASX22 0455

C:\CPWIN\DATA1\1P7307.30R



C:\CPWIN\DATA1\1P7307B.30R



ASX22 0456

Sample Name: **AR6850924C AAAR685AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 09:20:36pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.31R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.581		0.0000	0.000	1604.6	0.036	BV	0.050	535.62	0.023
2	1.790		0.0000	0.000	17958.6	0.400	VV	0.154	1943.15	0.082
3	1.862		0.0000	0.000	7718.6	0.172	VV	0.057	2273.99	0.096
4	1.929		0.0000	0.000	10291.7	0.229	VB	0.059	2925.45	0.123
5	2.056		0.0000	0.000	17780.0	0.396	BV	0.113	2614.61	0.110
6	2.173		0.0000	0.000	17449.9	0.389	VV	0.090	3236.57	0.136
7	2.271		0.0000	0.000	1451.7	0.032	VV	0.041	594.24	0.025
8	2.321		0.0000	0.000	842.2	0.019	VB	0.058	241.94	0.010
9	2.564		0.0000	0.000	1777.0	0.040	BB	0.053	561.26	0.024
10	2.782		0.0000	0.000	2178.3	0.049	BB	0.071	512.35	0.022
11	2.852		0.0000	0.000	6147.2	0.137	BB	0.083	1233.16	0.052
12	2.986	TCX	-4.5045	-7.054	5034.5	0.112	BB	0.078	1069.41	0.045
13	3.057		0.0000	0.000	209.5	0.005	BB	0.027	128.37	0.005
14	3.138		0.0000	0.000	115.3	0.003	BB	0.038	50.36	0.002
15	3.220		0.0000	0.000	107.2	0.002	BB	0.036	49.60	0.002
16	3.289		0.0000	0.000	698.9	0.016	BB	0.057	205.75	0.009
17	3.381		0.0000	0.000	495.6	0.011	BB	0.025	330.63	0.014
18	3.450		0.0000	0.000	164.7	0.004	BB	0.027	101.60	0.004
19	3.526		0.0000	0.000	183.7	0.004	BB	0.040	76.80	0.003
20	3.608		0.0000	0.000	422.4	0.009	BB	0.043	165.56	0.007
21	3.694		0.0000	0.000	355.0	0.008	BB	0.020	293.14	0.012
22	3.769		0.0000	0.000	338.0	0.008	BB	0.031	181.05	0.008
23	3.861		0.0000	0.000	2309.0	0.051	BB	0.035	1112.48	0.047
24	4.028		0.0000	0.000	1490.8	0.033	BB	0.054	463.41	0.020
25	4.124		0.0000	0.000	4446.0	0.099	BB	0.036	2045.16	0.086
26	4.180		0.0000	0.000	1538.5	0.034	BB	0.028	900.84	0.038
27	4.264		0.0000	0.000	2263.6	0.050	BB	0.027	1381.34	0.058
28	4.359		0.0000	0.000	1092.8	0.024	BB	0.021	878.15	0.037
29	4.395		0.0000	0.000	2137.3	0.048	BB	0.040	892.35	0.038
30	4.526		0.0000	0.000	2993.0	0.067	BB	0.034	1488.80	0.063
31	4.596		0.0000	0.000	280.9	0.006	BB	0.016	286.41	0.012
32	4.648		0.0000	0.000	1866.9	0.042	BB	0.027	1132.06	0.048
33	4.696		0.0000	0.000	531.1	0.012	BB	0.024	363.60	0.015
34	4.814		0.0000	0.000	677.4	0.015	BB	0.035	319.75	0.013
35	4.855		0.0000	0.000	715.6	0.016	BB	0.016	723.54	0.031
36	4.886		0.0000	0.000	653.0	0.015	BB	0.016	673.08	0.028
37	4.913		0.0000	0.000	1897.2	0.042	BB	0.021	1537.03	0.065
38	4.994		0.0000	0.000	704.1	0.016	BB	0.033	358.50	0.015
39	5.051		0.0000	0.000	1248.4	0.028	BB	0.020	1015.19	0.043
40	5.103		0.0000	0.000	3321.2	0.074	BB	0.039	1425.68	0.060
41	5.204		0.0000	0.000	483.1	0.011	BB	0.029	276.98	0.012
42	5.229		0.0000	0.000	244.0	0.005	BB	0.018	229.60	0.010
43	5.298		0.0000	0.000	492.5	0.011	BB	0.021	395.29	0.017
44	5.335		0.0000	0.000	283.3	0.006	BB	0.018	259.48	0.011
45	5.375		0.0000	0.000	1830.2	0.041	BB	0.026	1188.46	0.050
46	5.426		0.0000	0.000	506.9	0.011	BB	0.021	396.03	0.017
47	5.479		0.0000	0.000	2443.6	0.054	BB	0.043	955.90	0.040
48	5.637		0.0000	0.000	2384.9	0.053	BB	0.030	1311.44	0.055
49	5.764		0.0000	0.000	2107.3	0.047	BB	0.027	1278.80	0.054
50	5.812		0.0000	0.000	9561.1	0.213	BB	0.026	6060.02	0.256
51	5.877		0.0000	0.000	3203.4	0.071	BB	0.038	1409.21	0.059
52	5.998		0.0000	0.000	4183.2	0.093	BB	0.032	2201.55	0.093
53	6.081		0.0000	0.000	147336.1	3.285	BB	0.030	81074.82	3.418
54	6.248		0.0000	0.000	1697.0	0.038	BB	0.024	1172.09	0.049

ASX22 8457

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.289		0.0000	0.000	163743.7	3.651	BB	0.028	97662.18	4.118
56	6.369		0.0000	0.000	47286.5	1.054	BB	0.028	28448.96	1.200
57	6.445		0.0000	0.000	2372.2	0.053	BB	0.028	1401.89	0.059
58	6.582		0.0000	0.000	111789.8	2.492	BB	0.034	54724.26	2.307
59	6.821		0.0000	0.000	554257.7	12.357	BB	0.027	347969.80	14.672
60	6.897		0.0000	0.000	527349.4	11.757	BB	0.029	299262.00	12.618
61	7.053		0.0000	0.000	500680.8	11.163	BB	0.028	294360.50	12.411
62	7.141		0.0000	0.000	135996.5	3.032	BB	0.031	73471.64	3.098
63	7.414		0.0000	0.000	247116.5	5.509	BB	0.034	121392.10	5.118
64	7.620		0.0000	0.000	4118.4	0.092	BB	0.075	915.09	0.039
65	7.728		0.0000	0.000	1513881.0	33.752	BB	0.034	743437.50	31.346
66	7.975	DCB	68.3583	107.054	373289.8	8.322	BB	0.036	172806.00	7.286
67	8.085		0.0000	0.000	229.7	0.005	BB	0.025	151.46	0.006
68	8.145		0.0000	0.000	1311.4	0.029	BB	0.037	590.01	0.025
69	8.372		0.0000	0.000	378.5	0.008	BB	0.034	184.71	0.008
70	8.601		0.0000	0.000	1309.4	0.029	BB	0.059	370.91	0.016

Total Area = 4485360.0, Total Amount = 63.854, Total Height = 2371681.0, Sample Units = ug/l

ASX22 0458

Sample Name: **AR6850924C AAAR685AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 09:20:36pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.31R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.572		0.0000	0.000	4529.3	0.076	BV	0.042	1785.26	0.064
2	1.639		0.0000	0.000	12919.0	0.218	VV	0.096	2243.68	0.080
3	2.017		0.0000	0.000	777.8	0.013	VB	0.047	276.92	0.010
4	2.282		0.0000	0.000	2351.5	0.040	BB	0.050	783.35	0.028
5	2.394		0.0000	0.000	42.7	0.001	BB	0.016	43.29	0.002
6	2.471		0.0000	0.000	130.0	0.002	BB	0.064	33.75	0.001
7	2.520		0.0000	0.000	275.6	0.005	BB	0.059	78.32	0.003
8	3.007		0.0000	0.000	2490.5	0.042	BB	0.048	872.49	0.031
9	3.167		0.0000	0.000	3314.7	0.056	BB	0.048	1152.45	0.041
10	3.313		0.0000	0.000	370.4	0.006	BB	0.037	167.60	0.006
11	3.350		0.0000	0.000	162.0	0.003	BB	0.023	119.34	0.004
12	3.404		0.0000	0.000	6448.5	0.109	BB	0.066	1629.33	0.058
13	3.557		0.0000	0.000	1904.4	0.032	BB	0.035	895.88	0.032
14	3.620		0.0000	0.000	332.5	0.006	BB	0.037	148.59	0.005
15	3.643		0.0000	0.000	1323.4	0.022	BB	0.041	536.88	0.019
16	3.797		0.0000	0.000	4281.1	0.072	BB	0.055	1307.62	0.047
17	3.920		0.0000	0.000	895.2	0.015	BB	0.029	506.48	0.018
18	4.117		0.0000	0.000	658.9	0.011	BB	0.025	444.85	0.016
19	4.144		0.0000	0.000	1210.0	0.020	BB	0.032	634.54	0.023
20	4.299		0.0000	0.000	11200.7	0.189	BB	0.077	2439.97	0.087
21	4.431		0.0000	0.000	384.8	0.006	BB	0.030	213.32	0.008
22	4.507		0.0000	0.000	1311.4	0.022	BB	0.042	525.53	0.019
23	4.593		0.0000	0.000	2805.1	0.047	BB	0.041	1130.39	0.040
24	4.691		0.0000	0.000	1961.1	0.033	BB	0.039	836.79	0.030
25	4.797		0.0000	0.000	1947.1	0.033	BB	0.047	690.54	0.025
26	4.874		0.0000	0.000	2921.6	0.049	BB	0.037	1303.30	0.047
27	4.972		0.0000	0.000	14152.5	0.239	BB	0.035	6743.74	0.241
28	5.046		0.0000	0.000	920.1	0.016	BB	0.026	585.26	0.021
29	5.102		0.0000	0.000	4004.8	0.068	BB	0.032	2076.00	0.074
30	5.180		0.0000	0.000	220430.8	3.719	BB	0.040	90717.07	3.238
31	5.336		0.0000	0.000	2187.1	0.037	BB	0.023	1587.74	0.057
32	5.381		0.0000	0.000	223616.5	3.772	BB	0.033	114246.40	4.078
33	5.458		0.0000	0.000	52783.5	0.890	BB	0.031	28276.30	1.009
34	5.628		0.0000	0.000	68933.3	1.163	BB	0.030	38492.63	1.374
35	5.683		0.0000	0.000	36035.9	0.608	BB	0.029	21018.63	0.750
36	5.798		0.0000	0.000	592.6	0.010	BB	0.019	531.66	0.019
37	5.867		0.0000	0.000	630434.8	10.635	BB	0.027	384950.40	13.739
38	5.926		0.0000	0.000	617238.1	10.413	BB	0.030	338396.90	12.078
39	6.088		0.0000	0.000	694235.3	11.712	BB	0.032	360481.30	12.866
40	6.170		0.0000	0.000	158398.3	2.672	BB	0.031	86385.53	3.083
41	6.232		0.0000	0.000	2797.5	0.047	BB	0.020	2289.37	0.082
42	6.383		0.0000	0.000	377556.4	6.369	BB	0.041	153214.70	5.468
43	6.618		0.0000	0.000	2221474.0	37.476	BB	0.041	910229.90	32.487
44	6.800	DCB	80.5199	100.000	527156.1	8.893	BB	0.037	239266.70	8.540
45	7.053		0.0000	0.000	818.9	0.014	BB	0.027	508.76	0.018
46	7.238		0.0000	0.000	1388.5	0.023	BB	0.035	656.24	0.023
47	7.616		0.0000	0.000	674.4	0.011	BB	0.084	133.89	0.005
48	8.227		0.0000	0.000	2386.0	0.040	BB	0.272	146.16	0.005
49	8.487		0.0000	0.000	2506.9	0.042	BB	0.318	131.43	0.005

Total Area = 5927671.0, Total Amount = 80.52, Total Height = 2801868.0, Sample Units = ug/l

ASX22 0459

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.31R

C:\CPWIN\DATA1\1P7307.31R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 9:11:27 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.31R

C:\CPWIN\DATA1\1P7307B.31R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 9:11:27 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.986	1069	-4.50454	TCX	.	0	.	TCX
7.975	172806	68.35829	DCB	6.8	239267	80.51991	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.31A
Area File: C:\CPWIN\DATA1\1P7307B.31A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 9:20:42 PM
File Reported On: 11/3/2009 at 9:20:55 PM

10X REPLOTS

ASX22 0461

C:\CPWIN\DATA1\1P7307.31R



C:\CPWIN\DATA1\1P7307B.31R



ASX22 8462

Sample Name: **AR2130924B AAAR213AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 09:33:35pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.32R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.586		0.0000	0.000	3172.4	0.368	BV	0.052	1008.69	0.240
2	1.675		0.0000	0.000	8915.2	1.033	VV	0.080	1849.19	0.440
3	1.800		0.0000	0.000	14578.3	1.689	VV	0.086	2825.71	0.673
4	1.929		0.0000	0.000	27070.5	3.137	VB	0.111	4077.63	0.971
5	2.035		0.0000	0.000	19345.5	2.242	BV	0.083	3906.47	0.930
6	2.078		0.0000	0.000	16159.0	1.872	VV	0.065	4149.70	0.988
7	2.177		0.0000	0.000	13011.7	1.508	VV	0.059	3677.25	0.875
8	2.209		0.0000	0.000	16488.1	1.910	VV	0.080	3453.33	0.822
9	2.450		0.0000	0.000	9721.4	1.126	VV	0.158	1027.65	0.245
10	2.519		0.0000	0.000	19729.3	2.286	VB	0.034	9788.32	2.330
11	2.775		0.0000	0.000	3284.7	0.381	BB	0.093	586.50	0.140
12	2.896		0.0000	0.000	8463.8	0.981	BB	0.036	3951.56	0.941
13	2.999	TCX	42.5514	52.274	316236.6	36.642	BB	0.027	193203.40	45.986
14	3.228		0.0000	0.000	19806.6	2.295	BB	0.027	12424.89	2.957
15	3.310		0.0000	0.000	16641.9	1.928	BB	0.032	8765.26	2.086
16	3.380		0.0000	0.000	46716.7	5.413	BB	0.027	29071.38	6.920
17	3.472		0.0000	0.000	553.9	0.064	BB	0.021	435.17	0.104
18	3.513		0.0000	0.000	198.4	0.023	BB	0.019	176.54	0.042
19	3.604		0.0000	0.000	330.5	0.038	BB	0.059	93.60	0.022
20	3.694		0.0000	0.000	871.5	0.101	BB	0.019	778.03	0.185
21	3.730		0.0000	0.000	2424.6	0.281	BB	0.051	797.63	0.190
22	3.825		0.0000	0.000	4266.5	0.494	BB	0.021	3334.01	0.794
23	3.861		0.0000	0.000	1432.2	0.166	BB	0.020	1186.96	0.283
24	3.940		0.0000	0.000	476.1	0.055	BB	0.019	409.14	0.097
25	4.042		0.0000	0.000	2095.4	0.243	BB	0.039	903.51	0.215
26	4.126		0.0000	0.000	6763.8	0.784	BB	0.032	3568.55	0.849
27	4.181		0.0000	0.000	3313.1	0.384	BB	0.026	2104.42	0.501
28	4.266		0.0000	0.000	2703.9	0.313	BB	0.026	1728.15	0.411
29	4.360		0.0000	0.000	690.4	0.080	BB	0.020	563.03	0.134
30	4.397		0.0000	0.000	939.1	0.109	BB	0.045	346.42	0.082
31	4.528		0.0000	0.000	573.8	0.066	BB	0.021	446.79	0.106
32	4.656		0.0000	0.000	2225.9	0.258	BB	0.045	819.49	0.195
33	4.807		0.0000	0.000	696.0	0.081	BB	0.040	288.39	0.069
34	4.854		0.0000	0.000	457.7	0.053	BB	0.017	451.55	0.107
35	4.888		0.0000	0.000	83.2	0.010	BB	0.013	109.91	0.026
36	4.918		0.0000	0.000	767.0	0.089	BB	0.024	540.54	0.129
37	4.997		0.0000	0.000	624.4	0.072	BB	0.029	357.24	0.085
38	5.058		0.0000	0.000	861.7	0.100	BB	0.025	585.96	0.139
39	5.101		0.0000	0.000	1856.5	0.215	BB	0.035	893.50	0.213
40	5.203		0.0000	0.000	599.0	0.069	BB	0.026	389.68	0.093
41	5.234		0.0000	0.000	320.6	0.037	BB	0.020	262.74	0.063
42	5.300		0.0000	0.000	546.4	0.063	BB	0.025	359.25	0.086
43	5.339		0.0000	0.000	87.1	0.010	BB	0.018	78.97	0.019
44	5.376		0.0000	0.000	1727.2	0.200	BB	0.027	1086.10	0.259
45	5.479		0.0000	0.000	273.7	0.032	BB	0.024	193.73	0.046
46	5.520		0.0000	0.000	542.7	0.063	BB	0.030	305.95	0.073
47	5.637		0.0000	0.000	1070.8	0.124	BB	0.030	587.72	0.140
48	5.776		0.0000	0.000	3102.9	0.360	BB	0.083	621.79	0.148
49	5.900		0.0000	0.000	1557.7	0.180	BB	0.046	563.85	0.134
50	5.969		0.0000	0.000	375.0	0.043	BB	0.025	255.08	0.061
51	6.015		0.0000	0.000	583.1	0.068	BB	0.031	310.79	0.074
52	6.156		0.0000	0.000	122.4	0.014	BB	0.036	56.80	0.014
53	6.393		0.0000	0.000	1529.9	0.177	BB	0.094	272.62	0.065
54	6.431		0.0000	0.000	2457.2	0.285	BB	0.124	330.61	0.079

ASX22 0463

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.733		0.0000	0.000	2043.2	0.237	BB	0.056	612.53	0.146
56	6.827		0.0000	0.000	2881.9	0.334	BB	0.045	1066.78	0.254
57	6.899		0.0000	0.000	1207.7	0.140	BB	0.033	611.27	0.145
58	6.966		0.0000	0.000	146.5	0.017	BB	0.024	103.84	0.025
59	7.057		0.0000	0.000	1019.0	0.118	BB	0.030	559.28	0.133
60	7.207		0.0000	0.000	3787.8	0.439	BB	0.078	813.25	0.194
61	7.279		0.0000	0.000	2038.1	0.236	BB	0.138	245.58	0.058
62	7.517		0.0000	0.000	7779.5	0.901	BB	0.085	1516.89	0.361
63	7.638		0.0000	0.000	4453.1	0.516	BB	0.100	740.91	0.176
64	7.731		0.0000	0.000	9495.5	1.100	BB	0.038	4164.70	0.991
65	7.976	DCB	38.8497	47.726	213741.7	24.766	BB	0.036	98209.79	23.376
66	8.102		0.0000	0.000	3097.6	0.359	BB	0.112	460.71	0.110
67	8.372		0.0000	0.000	287.0	0.033	BB	0.047	101.48	0.024
68	8.531		0.0000	0.000	88.2	0.010	BB	0.046	31.77	0.008
69	8.594		0.0000	0.000	1062.0	0.123	BB	0.046	382.25	0.091
70	8.675		0.0000	0.000	469.5	0.054	BB	0.051	154.41	0.037

Total Area = 863041.9, Total Amount = 81.401, Total Height = 420136.4, Sample Units = ug/l

ASX22 0464

Sample Name: **AR2130924B AAAR213AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 09:33:35pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.32R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.589		0.0000	0.000	6113.3	0.646	BV	0.048	2108.11	0.546
2	1.639		0.0000	0.000	7871.6	0.832	VV	0.056	2341.49	0.606
3	1.704		0.0000	0.000	3720.4	0.393	VB	0.064	972.92	0.252
4	1.965		0.0000	0.000	25649.1	2.710	BB	0.044	9674.68	2.504
5	2.107		0.0000	0.000	1256.4	0.133	BB	0.173	120.71	0.031
6	2.294		0.0000	0.000	1388.3	0.147	BB	0.028	840.48	0.218
7	2.341	TCX	40.0584	48.039	404285.1	42.708	BB	0.039	173726.60	44.967
8	2.564	—	0.0000	0.000	17455.7	1.844	BB	0.033	8778.73	2.272
9	2.644	—	0.0000	0.000	11943.5	1.262	BB	0.035	5674.89	1.469
10	2.716	—	0.0000	0.000	53696.2	5.672	BB	0.042	21361.05	5.529
11	3.011		0.0000	0.000	4864.3	0.514	BB	0.042	1933.11	0.500
12	3.067		0.0000	0.000	463.5	0.049	BB	0.033	234.39	0.061
13	3.167		0.0000	0.000	9176.2	0.969	BB	0.051	3015.03	0.780
14	3.267		0.0000	0.000	212.3	0.022	BB	0.018	192.73	0.050
15	3.317		0.0000	0.000	1079.9	0.114	BB	0.043	421.32	0.109
16	3.406		0.0000	0.000	3998.8	0.422	BB	0.033	1990.94	0.515
17	3.454		0.0000	0.000	1347.2	0.142	BB	0.032	710.66	0.184
18	3.560		0.0000	0.000	1842.0	0.195	BB	0.034	909.18	0.235
19	3.619		0.0000	0.000	2002.2	0.212	BB	0.064	521.40	0.135
20	3.797		0.0000	0.000	3008.2	0.318	BB	0.101	497.69	0.129
21	4.114		0.0000	0.000	4193.1	0.443	BB	0.064	1092.59	0.283
22	4.260		0.0000	0.000	474.1	0.050	BB	0.037	210.79	0.055
23	4.296		0.0000	0.000	1966.7	0.208	BB	0.042	779.22	0.202
24	4.434		0.0000	0.000	282.7	0.030	BB	0.028	167.50	0.043
25	4.509		0.0000	0.000	1137.2	0.120	BB	0.040	475.37	0.123
26	4.586		0.0000	0.000	1818.7	0.192	BB	0.041	743.61	0.192
27	4.696		0.0000	0.000	1330.8	0.141	BB	0.040	550.09	0.142
28	4.830		0.0000	0.000	1115.5	0.118	BB	0.060	311.77	0.081
29	4.879		0.0000	0.000	337.1	0.036	BB	0.039	144.63	0.037
30	4.969		0.0000	0.000	495.5	0.052	BB	0.040	208.20	0.054
31	5.053		0.0000	0.000	243.1	0.026	BB	0.029	141.77	0.037
32	5.144		0.0000	0.000	2326.0	0.246	BB	0.077	500.70	0.130
33	5.386		0.0000	0.000	1340.4	0.142	BB	0.050	450.52	0.117
34	5.671		0.0000	0.000	324.9	0.034	BB	0.229	23.61	0.006
35	5.804		0.0000	0.000	1398.9	0.148	BB	0.030	766.56	0.198
36	5.867		0.0000	0.000	751.1	0.079	BB	0.029	435.26	0.113
37	5.928		0.0000	0.000	1737.1	0.184	BB	0.035	834.13	0.216
38	6.027		0.0000	0.000	625.5	0.066	BB	0.049	211.91	0.055
39	6.090		0.0000	0.000	2052.5	0.217	BB	0.041	833.83	0.216
40	6.166		0.0000	0.000	3176.4	0.336	BB	0.246	215.40	0.056
41	6.527		0.0000	0.000	4884.6	0.516	BB	0.251	323.70	0.084
42	6.624		0.0000	0.000	15084.8	1.594	BB	0.047	5326.86	1.379
43	6.802	DCB	43.3291	51.961	324812.8	34.313	BB	0.040	134551.30	34.827
44	7.497		0.0000	0.000	1466.6	0.155	BB	0.083	293.57	0.076
45	7.637		0.0000	0.000	8030.0	0.848	BB	0.348	384.68	0.100
46	8.217		0.0000	0.000	740.4	0.078	BB	0.078	158.24	0.041
47	8.547		0.0000	0.000	3110.1	0.329	BB	0.286	181.45	0.047

Total Area = 946630.8, Total Amount = 83.387, Total Height = 386343.4, Sample Units = ug/l

ASX22 0465

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.32R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 9:24:26 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.32R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 9:24:26 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	193203	42.55138	TCX	2.341	173727	40.05836	TCX
7.976	98210	38.84965	DCB	6.802	134551	43.32912	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.32A
Area File: C:\CPWIN\DATA1\1P7307B.32A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 9:33:44 PM
File Reported On: 11/3/2009 at 9:33:56 PM

10X REPLOTS

ASX22 0467

C:\CPWIN\DATA1\1P7307.32R



C:\CPWIN\DATA1\1P7307B.32R



ASX22 0468

Sample Name: **AR32X0924B AAAR32XAA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 09:46:33pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.33R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.581		0.0000	0.000	2884.2	0.283	BV	0.053	906.33	0.173
2	1.690		0.0000	0.000	6711.4	0.658	VV	0.081	1377.50	0.263
3	1.868		0.0000	0.000	20073.4	1.969	VV	0.133	2515.48	0.480
4	1.929		0.0000	0.000	10662.5	1.046	VB	0.058	3080.11	0.588
5	2.030		0.0000	0.000	13788.6	1.353	BV	0.077	2976.61	0.568
6	2.075		0.0000	0.000	11804.8	1.158	VV	0.053	3678.27	0.702
7	2.175		0.0000	0.000	9227.9	0.905	VV	0.047	3292.22	0.629
8	2.207		0.0000	0.000	13800.6	1.354	VV	0.077	2968.09	0.567
9	2.364		0.0000	0.000	3735.2	0.366	VV	0.052	1198.08	0.229
10	2.452		0.0000	0.000	4616.0	0.453	VV	0.092	834.15	0.159
11	2.518		0.0000	0.000	13041.8	1.279	VV	0.034	6464.78	1.234
12	2.618		0.0000	0.000	510.2	0.050	VB	0.036	236.14	0.045
13	2.768		0.0000	0.000	2325.1	0.228	BB	0.088	438.67	0.084
14	2.891		0.0000	0.000	7807.5	0.766	BB	0.041	3139.76	0.599
15	2.998	TCX	43.7273	52.508	327894.9	32.163	BB	0.028	198004.60	37.802
16	3.227		0.0000	0.000	13466.2	1.321	BB	0.027	8457.26	1.615
17	3.309		0.0000	0.000	13908.2	1.364	BB	0.034	6821.62	1.302
18	3.380		0.0000	0.000	38999.4	3.825	BB	0.026	24551.39	4.687
19	3.471		0.0000	0.000	425.6	0.042	BB	0.018	386.23	0.074
20	3.513		0.0000	0.000	1693.7	0.166	BB	0.024	1189.76	0.227
21	3.693		0.0000	0.000	30799.2	3.021	BB	0.039	13017.18	2.485
22	3.768		0.0000	0.000	1048.6	0.103	BB	0.022	787.52	0.150
23	3.824		0.0000	0.000	1896.4	0.186	BB	0.018	1761.21	0.336
24	3.862		0.0000	0.000	17097.9	1.677	BB	0.025	11284.68	2.154
25	3.939		0.0000	0.000	367.0	0.036	BB	0.018	339.27	0.065
26	3.968		0.0000	0.000	281.3	0.028	BB	0.018	255.81	0.049
27	4.030		0.0000	0.000	9095.8	0.892	BB	0.039	3867.88	0.738
28	4.125		0.0000	0.000	51741.7	5.075	BB	0.033	26026.24	4.969
29	4.180		0.0000	0.000	16735.3	1.642	BB	0.024	11568.92	2.209
30	4.264		0.0000	0.000	16909.7	1.659	BB	0.026	10736.05	2.050
31	4.315		0.0000	0.000	410.4	0.040	BB	0.016	421.77	0.081
32	4.359		0.0000	0.000	6928.4	0.680	BB	0.020	5681.98	1.085
33	4.395		0.0000	0.000	12142.0	1.191	BB	0.032	6242.62	1.192
34	4.527		0.0000	0.000	20083.3	1.970	BB	0.036	9269.13	1.770
35	4.597		0.0000	0.000	1819.6	0.178	BB	0.017	1811.72	0.346
36	4.649		0.0000	0.000	12106.5	1.188	BB	0.028	7084.58	1.353
37	4.697		0.0000	0.000	2194.3	0.215	BB	0.020	1786.97	0.341
38	4.761		0.0000	0.000	364.7	0.036	BB	0.020	304.87	0.058
39	4.815		0.0000	0.000	689.8	0.068	BB	0.022	511.05	0.098
40	4.857		0.0000	0.000	3887.4	0.381	BB	0.019	3474.47	0.663
41	4.886		0.0000	0.000	3281.5	0.322	BB	0.016	3343.50	0.638
42	4.913		0.0000	0.000	5783.6	0.567	BB	0.018	5305.84	1.013
43	4.995		0.0000	0.000	1025.3	0.101	BB	0.029	586.01	0.112
44	5.051		0.0000	0.000	4946.7	0.485	BB	0.021	3944.43	0.753
45	5.075		0.0000	0.000	1573.0	0.154	BB	0.020	1280.93	0.245
46	5.231		0.0000	0.000	3594.3	0.353	BB	0.046	1306.22	0.249
47	5.298		0.0000	0.000	1690.6	0.166	BB	0.022	1269.63	0.242
48	5.334		0.0000	0.000	619.1	0.061	BB	0.017	599.19	0.114
49	5.375		0.0000	0.000	3982.7	0.391	BB	0.025	2616.65	0.500
50	5.478		0.0000	0.000	1856.1	0.182	BB	0.025	1217.45	0.232
51	5.586		0.0000	0.000	155.9	0.015	BB	0.017	156.77	0.030
52	5.638		0.0000	0.000	2915.5	0.286	BB	0.027	1775.65	0.339
53	5.770		0.0000	0.000	1350.4	0.132	BB	0.061	367.39	0.070
54	5.877		0.0000	0.000	3550.1	0.348	BB	0.039	1536.20	0.293

ASX22 8469

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.967		0.0000	0.000	329.0	0.032	BB	0.023	237.20	0.045
56	6.013		0.0000	0.000	611.9	0.060	BB	0.034	304.40	0.058
57	6.087		0.0000	0.000	154.3	0.015	BB	0.027	93.54	0.018
58	6.152		0.0000	0.000	190.1	0.019	BB	0.034	93.22	0.018
59	6.393		0.0000	0.000	826.3	0.081	BB	0.025	540.72	0.103
60	6.727		0.0000	0.000	1589.8	0.156	BB	0.062	428.68	0.082
61	6.835		0.0000	0.000	1543.1	0.151	BB	0.065	398.02	0.076
62	6.895		0.0000	0.000	282.1	0.028	BB	0.037	126.56	0.024
63	6.970		0.0000	0.000	117.8	0.012	BB	0.036	54.67	0.010
64	7.107		0.0000	0.000	5581.7	0.548	BB	0.077	1203.67	0.230
65	7.213		0.0000	0.000	11440.4	1.122	BB	0.149	1276.93	0.244
66	7.517		0.0000	0.000	6205.1	0.609	BB	0.080	1292.49	0.247
67	7.639		0.0000	0.000	4543.7	0.446	BB	0.108	704.05	0.134
68	7.730		0.0000	0.000	3971.6	0.390	BB	0.035	1882.49	0.359
69	7.975	DCB	39.5498	47.492	217250.9	21.310	BB	0.036	99979.83	19.088
70	8.100		0.0000	0.000	2298.9	0.225	BB	0.098	392.62	0.075
71	8.372		0.0000	0.000	326.2	0.032	BB	0.042	130.46	0.025
72	8.588		0.0000	0.000	1311.1	0.129	BB	0.054	403.87	0.077
73	8.683		0.0000	0.000	591.1	0.058	BB	0.052	188.55	0.036

Total Area = 1019466.0, Total Amount = 83.277, Total Height = 523789.0, Sample Units = ug/l

ASX22 0478

Sample Name: **AR32X0924B AAAR32XAA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 09:46:33pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.33R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.586		0.0000	0.000	6161.4	0.567	BV	0.044	2342.66	0.523
2	1.637		0.0000	0.000	9132.6	0.841	VV	0.060	2544.37	0.568
3	1.700		0.0000	0.000	3503.5	0.322	VB	0.061	955.07	0.213
4	1.923		0.0000	0.000	23.0	0.002	BB	0.012	32.97	0.007
5	1.964		0.0000	0.000	14635.3	1.347	BB	0.044	5546.92	1.239
6	2.301		0.0000	0.000	631.6	0.058	BB	0.068	154.19	0.034
7	2.340	TCX	40.7375	47.990	413880.7	38.094	BB	0.039	176672.10	39.465
8	2.563		0.0000	0.000	11460.6	1.055	BB	0.033	5793.60	1.294
9	2.643		0.0000	0.000	10057.6	0.926	BB	0.037	4481.64	1.001
10	2.715		0.0000	0.000	45265.9	4.166	BB	0.043	17652.90	3.943
11	2.858		0.0000	0.000	2268.3	0.209	BB	0.035	1073.52	0.240
12	3.008		0.0000	0.000	46371.7	4.268	BB	0.050	15523.43	3.468
13	3.171		0.0000	0.000	27135.5	2.498	BB	0.041	11034.04	2.465
14	3.311		0.0000	0.000	6481.7	0.597	BB	0.045	2423.21	0.541
15	3.405		0.0000	0.000	30518.6	2.809	BB	0.035	14465.40	3.231
16	3.451		0.0000	0.000	5611.6	0.516	BB	0.030	3084.37	0.689
17	3.558		0.0000	0.000	11823.4	1.088	BB	0.032	6224.37	1.390
18	3.620		0.0000	0.000	1787.2	0.164	BB	0.045	668.60	0.149
19	3.645		0.0000	0.000	10191.8	0.938	BB	0.042	4084.97	0.912
20	3.798		0.0000	0.000	24906.1	2.292	BB	0.052	7951.65	1.776
21	3.887		0.0000	0.000	113.7	0.010	BB	0.011	178.55	0.040
22	3.924		0.0000	0.000	5961.5	0.549	BB	0.029	3391.02	0.757
23	3.972		0.0000	0.000	1578.5	0.145	BB	0.025	1040.44	0.232
24	4.118		0.0000	0.000	4185.4	0.385	BB	0.033	2105.63	0.470
25	4.146		0.0000	0.000	4254.3	0.392	BB	0.027	2600.76	0.581
26	4.251		0.0000	0.000	620.3	0.057	BB	0.019	542.55	0.121
27	4.294		0.0000	0.000	16689.8	1.536	BB	0.058	4783.74	1.069
28	4.431		0.0000	0.000	1353.4	0.125	BB	0.032	713.47	0.159
29	4.506		0.0000	0.000	1439.1	0.132	BB	0.026	906.22	0.202
30	4.537		0.0000	0.000	218.9	0.020	BB	0.015	241.83	0.054
31	4.583		0.0000	0.000	3667.7	0.338	BB	0.034	1779.81	0.398
32	4.695		0.0000	0.000	2995.5	0.276	BB	0.039	1268.88	0.283
33	4.795		0.0000	0.000	2950.5	0.272	BB	0.046	1068.55	0.239
34	4.877		0.0000	0.000	1393.9	0.128	BB	0.125	186.45	0.042
35	5.051		0.0000	0.000	2170.6	0.200	BB	0.040	898.68	0.201
36	5.143		0.0000	0.000	1747.0	0.161	BB	0.071	410.75	0.092
37	5.386		0.0000	0.000	1028.9	0.095	BB	0.047	365.85	0.082
38	5.636		0.0000	0.000	1536.7	0.141	BB	0.072	355.74	0.079
39	5.803		0.0000	0.000	3204.7	0.295	BB	0.044	1219.51	0.272
40	5.926		0.0000	0.000	780.7	0.072	BB	0.034	382.90	0.086
41	6.020		0.0000	0.000	840.4	0.077	BB	0.056	248.38	0.055
42	6.096		0.0000	0.000	3349.2	0.308	BB	0.309	180.40	0.040
43	6.393		0.0000	0.000	272.6	0.025	BB	0.084	54.16	0.012
44	6.444		0.0000	0.000	25.5	0.002	BB	0.012	35.95	0.008
45	6.533		0.0000	0.000	514.8	0.047	BB	0.040	212.01	0.047
46	6.623		0.0000	0.000	6524.3	0.601	BB	0.044	2461.72	0.550
47	6.801	DCB	44.1497	52.010	331334.7	30.496	BB	0.040	136861.70	30.572
48	7.640		0.0000	0.000	2312.7	0.213	BB	0.129	298.34	0.067
49	8.017		0.0000	0.000	1564.2	0.144	BB	0.156	167.61	0.037

Total Area = 1086478.0, Total Amount = 84.887, Total Height = 447671.5, Sample Units = ug/l

ASX22 8471

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.33R

C:\CPWIN\DATA1\1P7307.33R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 9:37:24 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.33R

C:\CPWIN\DATA1\1P7307B.33R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 9:37:24 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.998	198005	43.72726	TCX	2.34	176672	40.73755	TCX
7.975	99980	39.54984	DCB	6.801	136862	44.14965	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.33A
Area File: C:\CPWIN\DATA1\1P7307B.33A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 9:46:40 PM
File Reported On: 11/3/2009 at 9:46:53 PM

10X REPLOTS

ASX22 0473

C:\CPWIN\DATA1\1P7307.33R



C:\CPWIN\DATA1\1P7307B.33R



ASX22. 0474

Sample Name: **AR4230924B AAAR423AA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 09:59:32pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.34R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.581		0.0000	0.000	5988.3	0.435	BV	0.063	1590.65	0.240
2	1.867		0.0000	0.000	57784.6	4.196	VV	0.182	5277.34	0.797
3	1.925		0.0000	0.000	21384.7	1.553	VB	0.057	6236.19	0.942
4	2.030		0.0000	0.000	26538.2	1.927	BV	0.074	5993.27	0.905
5	2.073		0.0000	0.000	25600.8	1.859	VV	0.065	6588.06	0.995
6	2.174		0.0000	0.000	20184.0	1.466	VV	0.064	5264.59	0.795
7	2.207		0.0000	0.000	25341.5	1.840	VV	0.084	5008.60	0.757
8	2.394		0.0000	0.000	8493.4	0.617	VV	0.071	1990.54	0.301
9	2.447		0.0000	0.000	5671.1	0.412	VV	0.061	1558.74	0.235
10	2.522		0.0000	0.000	4767.4	0.346	VV	0.065	1225.35	0.185
11	2.587		0.0000	0.000	446.7	0.032	VB	0.035	212.30	0.032
12	2.765		0.0000	0.000	10225.7	0.742	BB	0.089	1912.13	0.289
13	2.887		0.0000	0.000	5095.3	0.370	BB	0.086	984.53	0.149
14	2.997	TCX	43.8555	52.180	316528.5	22.983	BB	0.027	198528.40	29.987
15	3.225		0.0000	0.000	5096.2	0.370	BB	0.026	3252.82	0.491
16	3.287		0.0000	0.000	1109.8	0.081	BB	0.021	866.77	0.131
17	3.307		0.0000	0.000	2200.6	0.160	BB	0.018	2046.62	0.309
18	3.379		0.0000	0.000	28297.0	2.055	BB	0.026	17884.59	2.701
19	3.470		0.0000	0.000	362.5	0.026	BB	0.017	355.68	0.054
20	3.512		0.0000	0.000	3453.2	0.251	BB	0.024	2394.05	0.362
21	3.601		0.0000	0.000	168.8	0.012	BB	0.046	61.61	0.009
22	3.692		0.0000	0.000	57257.7	4.158	BB	0.039	24472.86	3.697
23	3.769		0.0000	0.000	2956.1	0.215	BB	0.027	1853.39	0.280
24	3.822		0.0000	0.000	1017.0	0.074	BB	0.017	993.32	0.150
25	3.861		0.0000	0.000	32487.3	2.359	BB	0.026	20724.35	3.130
26	3.937		0.0000	0.000	326.5	0.024	BB	0.016	334.63	0.051
27	3.967		0.0000	0.000	524.3	0.038	BB	0.020	442.05	0.067
28	4.028		0.0000	0.000	15916.2	1.156	BB	0.039	6877.70	1.039
29	4.123		0.0000	0.000	93392.0	6.781	BB	0.033	47031.33	7.104
30	4.178		0.0000	0.000	30063.5	2.183	BB	0.024	20675.10	3.123
31	4.263		0.0000	0.000	31143.0	2.261	BB	0.026	19635.00	2.966
32	4.314		0.0000	0.000	948.0	0.069	BB	0.016	983.97	0.149
33	4.357		0.0000	0.000	14473.0	1.051	BB	0.020	12228.20	1.847
34	4.394		0.0000	0.000	23907.7	1.736	BB	0.032	12416.20	1.875
35	4.526		0.0000	0.000	41112.4	2.985	BB	0.036	18991.18	2.869
36	4.596		0.0000	0.000	3780.9	0.275	BB	0.017	3674.45	0.555
37	4.647		0.0000	0.000	24232.7	1.760	BB	0.028	14297.02	2.159
38	4.696		0.0000	0.000	4507.1	0.327	BB	0.020	3690.12	0.557
39	4.760		0.0000	0.000	872.1	0.063	BB	0.020	719.71	0.109
40	4.814		0.0000	0.000	1203.6	0.087	BB	0.021	939.15	0.142
41	4.856		0.0000	0.000	7649.4	0.555	BB	0.019	6580.20	0.994
42	4.884		0.0000	0.000	6179.4	0.449	BB	0.015	6921.53	1.045
43	4.911		0.0000	0.000	12175.5	0.884	BB	0.019	10560.06	1.595
44	4.996		0.0000	0.000	1730.0	0.126	BB	0.027	1068.50	0.161
45	5.050		0.0000	0.000	49369.6	3.585	BB	0.054	15105.65	2.282
46	5.229		0.0000	0.000	6374.9	0.463	BB	0.038	2825.06	0.427
47	5.296		0.0000	0.000	4486.3	0.326	BB	0.023	3289.82	0.497
48	5.333		0.0000	0.000	1503.3	0.109	BB	0.017	1508.58	0.228
49	5.373		0.0000	0.000	9199.1	0.668	BB	0.025	6191.31	0.935
50	5.477		0.0000	0.000	6321.4	0.459	BB	0.035	3011.60	0.455
51	5.586		0.0000	0.000	582.2	0.042	BB	0.020	480.90	0.073
52	5.637		0.0000	0.000	8071.1	0.586	BB	0.028	4876.60	0.737
53	5.732		0.0000	0.000	401.0	0.029	BB	0.031	218.79	0.033
54	5.771		0.0000	0.000	1128.6	0.082	BB	0.035	539.76	0.082

ASX22 8475

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.875		0.0000	0.000	7737.8	0.562	BB	0.032	4036.52	0.610
56	5.965		0.0000	0.000	371.7	0.027	BB	0.024	257.31	0.039
57	6.011		0.0000	0.000	2047.4	0.149	BB	0.035	971.44	0.147
58	6.155		0.0000	0.000	120.4	0.009	BB	0.028	70.51	0.011
59	6.393		0.0000	0.000	35383.5	2.569	BB	0.189	3112.69	0.470
60	6.730		0.0000	0.000	1803.7	0.131	BB	0.059	509.95	0.077
61	6.839		0.0000	0.000	1530.5	0.111	BB	0.082	309.33	0.047
62	6.899		0.0000	0.000	192.5	0.014	BB	0.036	89.35	0.013
63	6.973		0.0000	0.000	110.2	0.008	BB	0.041	45.23	0.007
64	7.107		0.0000	0.000	5384.9	0.391	BB	0.073	1225.73	0.185
65	7.203		0.0000	0.000	3736.6	0.271	BB	0.076	819.04	0.124
66	7.276		0.0000	0.000	4646.7	0.337	BB	0.241	321.24	0.049
67	7.517		0.0000	0.000	9133.6	0.663	BB	0.096	1589.89	0.240
68	7.629		0.0000	0.000	5000.3	0.363	BB	0.095	874.12	0.132
69	7.729		0.0000	0.000	2623.4	0.190	BB	0.034	1301.73	0.197
70	7.866		0.0000	0.000	518.8	0.038	BB	0.047	182.54	0.028
71	7.974	DCB	40.1916	47.820	220680.3	16.024	BB	0.036	101602.10	15.347
72	8.111		0.0000	0.000	3521.5	0.256	BB	0.137	428.25	0.065
73	8.376		0.0000	0.000	528.3	0.038	BB	0.045	197.70	0.030
74	8.520		0.0000	0.000	87.5	0.006	BB	0.036	41.03	0.006
75	8.589		0.0000	0.000	1196.7	0.087	BB	0.045	444.01	0.067
76	8.686		0.0000	0.000	820.9	0.060	BB	0.059	231.84	0.035

Total Area = 1377208.0, Total Amount = 84.047, Total Height = 662052.5, Sample Units = ug/l

ASX22 0476

Sample Name: **AR4230924B AAAR423AA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 09:59:32pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.34R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.534		0.0000	0.000	288.1	0.022	BV	0.021	225.70	0.043
2	1.589		0.0000	0.000	6811.3	0.529	VV	0.047	2406.59	0.457
3	1.637		0.0000	0.000	17631.6	1.368	VV	0.100	2926.31	0.556
4	1.932		0.0000	0.000	3214.9	0.249	VV	0.080	673.14	0.128
5	1.981		0.0000	0.000	2399.3	0.186	VV	0.041	970.68	0.184
6	2.016		0.0000	0.000	3263.5	0.253	VB	0.047	1165.51	0.221
7	2.113		0.0000	0.000	620.4	0.048	BB	0.053	193.74	0.037
8	2.339	TCX	41.5858	48.211	418799.0	32.501	BB	0.039	180350.60	34.273
9	2.561		0.0000	0.000	4041.6	0.314	BB	0.031	2151.83	0.409
10	2.639		0.0000	0.000	8133.9	0.631	BB	0.043	3143.15	0.597
11	2.714		0.0000	0.000	33569.8	2.605	BB	0.044	12805.05	2.433
12	2.858		0.0000	0.000	4632.9	0.360	BB	0.035	2175.15	0.413
13	3.006		0.0000	0.000	81523.1	6.327	BB	0.049	27895.51	5.301
14	3.170		0.0000	0.000	45766.8	3.552	BB	0.040	19148.04	3.639
15	3.310		0.0000	0.000	11965.0	0.929	BB	0.044	4494.41	0.854
16	3.403		0.0000	0.000	55146.4	4.280	BB	0.034	27136.71	5.157
17	3.449		0.0000	0.000	10572.4	0.820	BB	0.028	6234.74	1.185
18	3.556		0.0000	0.000	21851.1	1.696	BB	0.031	11621.85	2.209
19	3.618		0.0000	0.000	4237.2	0.329	BB	0.025	2772.47	0.527
20	3.643		0.0000	0.000	20473.3	1.589	BB	0.040	8538.08	1.623
21	3.797		0.0000	0.000	50502.2	3.919	BB	0.052	16223.10	3.083
22	3.885		0.0000	0.000	201.6	0.016	BB	0.010	326.89	0.062
23	3.923		0.0000	0.000	12601.3	0.978	BB	0.029	7150.00	1.359
24	3.971		0.0000	0.000	3191.8	0.248	BB	0.027	2006.94	0.381
25	4.115		0.0000	0.000	8509.2	0.660	BB	0.037	3799.36	0.722
26	4.143		0.0000	0.000	9113.5	0.707	BB	0.030	5137.07	0.976
27	4.251		0.0000	0.000	1878.8	0.146	BB	0.022	1442.14	0.274
28	4.292		0.0000	0.000	33552.5	2.604	BB	0.056	9989.44	1.898
29	4.431		0.0000	0.000	3586.3	0.278	BB	0.032	1857.24	0.353
30	4.504		0.0000	0.000	3354.6	0.260	BB	0.024	2286.85	0.435
31	4.536		0.0000	0.000	530.7	0.041	BB	0.015	574.99	0.109
32	4.581		0.0000	0.000	8935.3	0.693	BB	0.034	4403.25	0.837
33	4.692		0.0000	0.000	6134.4	0.476	BB	0.039	2605.23	0.495
34	4.795		0.0000	0.000	7974.5	0.619	BB	0.043	3120.69	0.593
35	4.964		0.0000	0.000	951.3	0.074	BB	0.043	369.77	0.070
36	5.049		0.0000	0.000	6449.6	0.501	BB	0.041	2611.69	0.496
37	5.142		0.0000	0.000	3031.1	0.235	BB	0.049	1022.66	0.194
38	5.410		0.0000	0.000	1717.6	0.133	BB	0.060	480.27	0.091
39	5.570		0.0000	0.000	525.7	0.041	BB	0.052	167.23	0.032
40	5.801		0.0000	0.000	9463.0	0.734	BB	0.072	2182.00	0.415
41	6.024		0.0000	0.000	11411.6	0.886	BB	0.385	493.50	0.094
42	6.415		0.0000	0.000	436.1	0.034	BB	0.201	36.18	0.007
43	6.457		0.0000	0.000	70.9	0.005	BB	0.028	42.26	0.008
44	6.527		0.0000	0.000	446.7	0.035	BB	0.038	197.86	0.038
45	6.623		0.0000	0.000	4594.7	0.357	BB	0.043	1800.98	0.342
46	6.801	DCB	44.6723	51.789	335612.3	26.046	BB	0.040	138333.30	26.289
47	7.633		0.0000	0.000	5735.4	0.445	BB	0.392	243.77	0.046
48	8.321		0.0000	0.000	3099.8	0.241	BB	0.186	278.05	0.053

Total Area = 1288554.0, Total Amount = 86.258, Total Height = 526211.9, Sample Units = ug/l

GSX22 0477

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.34R

C:\CPWIN\DATA1\1P7307.34R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 9:50:23 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.34R

C:\CPWIN\DATA1\1P7307B.34R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 9:50:23 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.997	198528	43.85554	TCX	2.339	180351	41.58575	TCX
7.974	101602	40.19157	DCB	6.801	138333	44.67231	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.34A
Area File: C:\CPWIN\DATA1\1P7307B.34A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 9:59:40 PM
File Reported On: 11/3/2009 at 9:59:53 PM

10X REPLOTS

ASX22 0479

C:\CPWIN\DATA1\1P7307.34R



C:\CPWIN\DATA1\1P7307B.34R



ASX22 0488

Sample Name: **AR16X0924A AAAR16XAA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 10:12:33pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.35R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.581		0.0000	0.000	1148.5	0.145	BV	0.039	493.66	0.128
2	1.667		0.0000	0.000	4085.9	0.517	VV	0.080	850.21	0.221
3	1.931		0.0000	0.000	19845.6	2.511	VB	0.175	1889.12	0.491
4	2.033		0.0000	0.000	9280.0	1.174	BV	0.079	1964.09	0.510
5	2.074		0.0000	0.000	5341.7	0.676	VV	0.048	1872.24	0.486
6	2.174		0.0000	0.000	21158.1	2.677	VV	0.057	6173.62	1.603
7	2.366		0.0000	0.000	1469.0	0.186	VV	0.049	497.58	0.129
8	2.459		0.0000	0.000	898.2	0.114	VV	0.058	258.74	0.067
9	2.526		0.0000	0.000	2143.9	0.271	VB	0.055	650.36	0.169
10	2.770		0.0000	0.000	2210.8	0.280	BB	0.059	620.79	0.161
11	2.824		0.0000	0.000	8663.4	1.096	BB	0.104	1391.65	0.361
12	2.999	TCX	7.2963	41.838	86536.2	10.949	BB	0.029	49253.56	12.789
13	3.226		0.0000	0.000	5785.1	0.732	BB	0.026	3641.16	0.945
14	3.310		0.0000	0.000	7356.5	0.931	BB	0.029	4262.78	1.107
15	3.380		0.0000	0.000	32534.6	4.117	BB	0.027	20318.86	5.276
16	3.513		0.0000	0.000	4379.6	0.554	BB	0.026	2837.59	0.737
17	3.604		0.0000	0.000	94.1	0.012	BB	0.021	75.38	0.020
18	3.693		0.0000	0.000	65290.9	8.261	BB	0.040	27479.20	7.135
19	3.771		0.0000	0.000	3342.2	0.423	BB	0.027	2098.67	0.545
20	3.823		0.0000	0.000	1115.8	0.141	BB	0.017	1116.58	0.290
21	3.862		0.0000	0.000	38526.9	4.875	BB	0.026	24581.88	6.383
22	3.969		0.0000	0.000	817.3	0.103	BB	0.023	589.34	0.153
23	4.030		0.0000	0.000	18660.9	2.361	BB	0.039	8037.10	2.087
24	4.125		0.0000	0.000	107913.9	13.654	BB	0.033	54742.98	14.215
25	4.179		0.0000	0.000	33994.6	4.301	BB	0.024	23724.26	6.160
26	4.264		0.0000	0.000	35299.4	4.466	BB	0.026	22208.75	5.767
27	4.314		0.0000	0.000	1107.2	0.140	BB	0.017	1116.24	0.290
28	4.358		0.0000	0.000	16007.6	2.025	BB	0.021	12994.94	3.374
29	4.395		0.0000	0.000	27303.0	3.455	BB	0.033	13905.20	3.611
30	4.527		0.0000	0.000	46385.8	5.869	BB	0.036	21499.21	5.583
31	4.598		0.0000	0.000	4154.6	0.526	BB	0.017	4107.40	1.067
32	4.648		0.0000	0.000	26164.7	3.311	BB	0.028	15508.56	4.027
33	4.697		0.0000	0.000	5149.8	0.652	BB	0.021	4149.62	1.078
34	4.761		0.0000	0.000	635.7	0.080	BB	0.019	548.03	0.142
35	4.814		0.0000	0.000	861.5	0.109	BB	0.023	631.95	0.164
36	4.856		0.0000	0.000	4938.4	0.625	BB	0.019	4335.10	1.126
37	4.885		0.0000	0.000	2224.5	0.281	BB	0.015	2546.28	0.661
38	4.913		0.0000	0.000	2830.5	0.358	BB	0.017	2808.04	0.729
39	4.996		0.0000	0.000	573.0	0.072	BB	0.029	333.68	0.087
40	5.049		0.0000	0.000	1270.0	0.161	BB	0.024	892.72	0.232
41	5.101		0.0000	0.000	1751.8	0.222	BB	0.039	744.58	0.193
42	5.202		0.0000	0.000	1769.0	0.224	BB	0.043	688.30	0.179
43	5.381		0.0000	0.000	2021.6	0.256	BB	0.057	596.01	0.155
44	5.483		0.0000	0.000	1446.1	0.183	BB	0.065	368.07	0.096
45	5.634		0.0000	0.000	414.6	0.052	BB	0.039	177.36	0.046
46	5.812		0.0000	0.000	3316.3	0.420	BB	0.114	485.90	0.126
47	5.903		0.0000	0.000	1054.4	0.133	BB	0.041	431.82	0.112
48	5.968		0.0000	0.000	510.3	0.065	BB	0.028	301.50	0.078
49	6.031		0.0000	0.000	231.4	0.029	BB	0.033	117.73	0.031
50	6.172		0.0000	0.000	126.0	0.016	BB	0.048	44.18	0.011
51	6.395		0.0000	0.000	626.8	0.079	BB	0.025	426.31	0.111
52	6.722		0.0000	0.000	2436.6	0.308	BB	0.081	500.75	0.130
53	6.948		0.0000	0.000	6190.6	0.783	BB	0.340	303.21	0.079
54	7.111		0.0000	0.000	5299.1	0.670	BB	0.074	1193.85	0.310

ASX22 0481

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	7.202		0.0000	0.000	21242.9	2.688	BB	0.238	1487.74	0.386
56	7.516		0.0000	0.000	10209.5	1.292	BB	0.100	1697.98	0.441
57	7.626		0.0000	0.000	5171.1	0.654	BB	0.094	912.83	0.237
58	7.731		0.0000	0.000	682.2	0.086	BB	0.039	288.32	0.075
59	7.872		0.0000	0.000	408.7	0.052	BB	0.041	164.75	0.043
60	7.976	DCB	10.1433	58.162	57767.6	7.309	BB	0.038	25641.62	6.658
61	8.183		0.0000	0.000	4977.9	0.630	BB	0.212	392.25	0.102
62	8.371		0.0000	0.000	331.3	0.042	BB	0.043	128.81	0.033
63	8.587		0.0000	0.000	3754.8	0.475	BB	0.087	721.98	0.187
64	8.690		0.0000	0.000	1092.4	0.138	BB	0.063	289.66	0.075

Total Area = 790332.6, Total Amount = 17.44, Total Height = 385112.6, Sample Units = ug/l

ASX22 0482

Sample Name: **AR16X0924A AAAR16XAA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 10:12:33pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.35R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.569		0.0000	0.000	13684.5	1.920	BV	0.047	4809.37	1.763
2	1.639		0.0000	0.000	10583.8	1.485	VB	0.075	2362.34	0.866
3	2.047		0.0000	0.000	607.6	0.085	BB	0.248	40.87	0.015
4	2.110		0.0000	0.000	132.1	0.019	BB	0.022	102.25	0.037
5	2.186		0.0000	0.000	800.0	0.112	BB	0.069	193.84	0.071
6	2.283		0.0000	0.000	1187.2	0.167	BB	0.035	569.78	0.209
7	2.341	TCX	9.0243	53.751	93468.6	13.111	BB	0.040	39137.03	14.346
8	2.563		0.0000	0.000	5180.4	0.727	BB	0.034	2544.25	0.933
9	2.645		0.0000	0.000	5373.8	0.754	BB	0.033	2674.86	0.981
10	2.716		0.0000	0.000	37486.7	5.258	BB	0.043	14569.90	5.341
11	2.860		0.0000	0.000	5443.6	0.764	BB	0.037	2474.75	0.907
12	3.008		0.0000	0.000	95225.9	13.358	BB	0.049	32603.19	11.951
13	3.171		0.0000	0.000	54559.6	7.653	BB	0.041	22393.70	8.209
14	3.312		0.0000	0.000	14998.4	2.104	BB	0.046	5403.80	1.981
15	3.405		0.0000	0.000	66519.5	9.331	BB	0.034	33056.54	12.117
16	3.450		0.0000	0.000	11240.0	1.577	BB	0.029	6421.38	2.354
17	3.558		0.0000	0.000	24995.0	3.506	BB	0.031	13483.72	4.943
18	3.620		0.0000	0.000	4282.8	0.601	BB	0.042	1701.96	0.624
19	3.645		0.0000	0.000	24101.1	3.381	BB	0.041	9705.08	3.558
20	3.798		0.0000	0.000	58339.7	8.183	BB	0.052	18688.82	6.851
21	3.887		0.0000	0.000	283.9	0.040	BB	0.011	436.40	0.160
22	3.924		0.0000	0.000	14256.8	2.000	BB	0.029	8170.82	2.995
23	3.971		0.0000	0.000	3410.5	0.478	BB	0.027	2107.36	0.772
24	4.115		0.0000	0.000	37776.7	5.299	BB	0.073	8604.35	3.154
25	4.264		0.0000	0.000	74.8	0.010	BB	0.033	37.28	0.014
26	4.297		0.0000	0.000	4631.8	0.650	BB	0.050	1529.61	0.561
27	4.599		0.0000	0.000	122.9	0.017	BB	0.017	120.81	0.044
28	4.836		0.0000	0.000	825.1	0.116	BB	0.040	347.10	0.127
29	4.946		0.0000	0.000	1087.5	0.153	BB	0.066	275.21	0.101
30	5.570		0.0000	0.000	1484.5	0.208	BB	0.072	345.40	0.127
31	5.680		0.0000	0.000	895.6	0.126	BB	0.084	178.38	0.065
32	5.801		0.0000	0.000	6645.8	0.932	BB	0.091	1216.20	0.446
33	6.016		0.0000	0.000	13762.8	1.931	BB	0.384	597.93	0.219
34	6.414		0.0000	0.000	620.0	0.087	BB	0.191	54.17	0.020
35	6.527		0.0000	0.000	1417.1	0.199	BB	0.079	299.07	0.110
36	6.598		0.0000	0.000	65.5	0.009	BB	0.036	30.72	0.011
37	6.630		0.0000	0.000	1017.8	0.143	BB	0.042	404.58	0.148
38	6.803	DCB	7.7647	46.249	84955.2	11.917	BB	0.041	34415.30	12.615
39	7.620		0.0000	0.000	7977.7	1.119	BB	0.365	364.04	0.133
40	8.283		0.0000	0.000	2973.1	0.417	BB	0.174	285.35	0.105
41	8.639		0.0000	0.000	402.5	0.056	BB	0.149	45.12	0.017

Total Area = 712898.0, Total Amount = 16.789, Total Height = 272802.6, Sample Units = ug/l

ASX22 0483

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.35R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 10:03:24 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.35R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 10:03:24 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	49254	7.29634	TCX	2.341	39137	9.02433	TCX
7.976	25642	10.14326	DCB	6.803	34415	7.76469	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.35A
Area File: C:\CPWIN\DATA1\1P7307B.35A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 10:12:40 PM
File Reported On: 11/3/2009 at 10:12:54 PM

10X REPLOTS

ASX22 0485

C:\CPWIN\DATA1\1P7307.35R



C:\CPWIN\DATA1\1P7307B.35R



ASX22 0486

Sample Name: **MD16X0924A AAMD16XAA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 10:25:34pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.36R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.576		0.0000	0.000	2823.0	0.815	BV	0.053	881.38	0.875
2	1.659		0.0000	0.000	6998.7	2.020	VV	0.062	1878.48	1.865
3	1.710		0.0000	0.000	6625.4	1.912	VV	0.048	2306.44	2.289
4	1.789		0.0000	0.000	12616.7	3.641	VV	0.072	2940.15	2.918
5	1.929		0.0000	0.000	29298.2	8.456	VB	0.104	4709.26	4.675
6	2.038		0.0000	0.000	28229.0	8.147	BV	0.098	4819.73	4.784
7	2.128		0.0000	0.000	14553.6	4.200	VV	0.064	3779.71	3.752
8	2.174		0.0000	0.000	26448.5	7.633	VV	0.114	3863.15	3.835
9	2.313		0.0000	0.000	5675.0	1.638	VV	0.049	1911.41	1.897
10	2.370		0.0000	0.000	5212.6	1.504	VV	0.061	1435.85	1.425
11	2.453		0.0000	0.000	5007.4	1.445	VV	0.086	975.89	0.969
12	2.541		0.0000	0.000	3582.5	1.034	VB	0.067	888.64	0.882
13	2.714		0.0000	0.000	356.4	0.103	BB	0.033	182.32	0.181
14	2.777		0.0000	0.000	1916.0	0.553	BB	0.048	671.53	0.667
15	2.860		0.0000	0.000	11153.2	3.219	BB	0.071	2624.77	2.605
16	2.999	TCX	-3.1226	200.363	21069.6	6.081	BB	0.052	6712.00	6.662
17	3.059		0.0000	0.000	241.4	0.070	BB	0.031	131.28	0.130
18	3.227		0.0000	0.000	515.6	0.149	BB	0.030	289.41	0.287
19	3.307		0.0000	0.000	1039.0	0.300	BB	0.052	331.60	0.329
20	3.381		0.0000	0.000	1932.8	0.558	BB	0.026	1262.02	1.253
21	3.511		0.0000	0.000	850.2	0.245	BB	0.037	381.83	0.379
22	3.602		0.0000	0.000	728.0	0.210	BB	0.053	229.44	0.228
23	3.695		0.0000	0.000	737.4	0.213	BB	0.026	475.28	0.472
24	3.768		0.0000	0.000	403.1	0.116	BB	0.029	228.80	0.227
25	3.863		0.0000	0.000	2730.0	0.788	BB	0.029	1555.06	1.544
26	3.979		0.0000	0.000	96.3	0.028	BB	0.021	75.58	0.075
27	4.032		0.0000	0.000	1550.9	0.448	BB	0.038	678.28	0.673
28	4.126		0.0000	0.000	6797.7	1.962	BB	0.034	3323.90	3.299
29	4.182		0.0000	0.000	2738.6	0.790	BB	0.027	1693.80	1.681
30	4.265		0.0000	0.000	2165.7	0.625	BB	0.026	1379.34	1.369
31	4.320		0.0000	0.000	129.6	0.037	BB	0.019	114.24	0.113
32	4.360		0.0000	0.000	921.4	0.266	BB	0.020	774.04	0.768
33	4.397		0.0000	0.000	1759.7	0.508	BB	0.033	897.08	0.890
34	4.527		0.0000	0.000	2595.4	0.749	BB	0.035	1248.11	1.239
35	4.597		0.0000	0.000	220.7	0.064	BB	0.016	231.10	0.229
36	4.649		0.0000	0.000	1558.9	0.450	BB	0.028	934.19	0.927
37	4.698		0.0000	0.000	376.3	0.109	BB	0.025	255.69	0.254
38	4.811		0.0000	0.000	590.5	0.170	BB	0.036	269.98	0.268
39	4.854		0.0000	0.000	786.9	0.227	BB	0.019	697.03	0.692
40	4.920		0.0000	0.000	623.6	0.180	BB	0.028	367.62	0.365
41	4.986		0.0000	0.000	361.7	0.104	BB	0.033	184.73	0.183
42	5.060		0.0000	0.000	797.4	0.230	BB	0.025	541.01	0.537
43	5.106		0.0000	0.000	3236.3	0.934	BB	0.032	1701.37	1.689
44	5.204		0.0000	0.000	1990.0	0.574	BB	0.043	768.91	0.763
45	5.313		0.0000	0.000	789.1	0.228	BB	0.033	400.04	0.397
46	5.379		0.0000	0.000	1006.8	0.291	BB	0.026	641.76	0.637
47	5.429		0.0000	0.000	1517.7	0.438	BB	0.027	933.26	0.926
48	5.477		0.0000	0.000	288.6	0.083	BB	0.018	269.07	0.267
49	5.521		0.0000	0.000	3696.8	1.067	BB	0.027	2259.20	2.243
50	5.634		0.0000	0.000	727.6	0.210	BB	0.041	297.48	0.295
51	5.709		0.0000	0.000	655.6	0.189	BB	0.030	364.13	0.361
52	5.767		0.0000	0.000	4336.3	1.251	BB	0.029	2505.32	2.487
53	5.812		0.0000	0.000	433.5	0.125	BB	0.018	412.35	0.409
54	5.879		0.0000	0.000	2678.4	0.773	BB	0.040	1108.49	1.100

ASX22 0487

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.969		0.0000	0.000	402.4	0.116	BB	0.023	291.42	0.289
56	6.013		0.0000	0.000	3666.2	1.058	BB	0.029	2111.24	2.096
57	6.084		0.0000	0.000	2485.9	0.717	BB	0.024	1693.64	1.681
58	6.136		0.0000	0.000	1173.5	0.339	BB	0.025	783.91	0.778
59	6.282		0.0000	0.000	3753.0	1.083	BB	0.034	1846.27	1.833
60	6.354		0.0000	0.000	1010.4	0.292	BB	0.025	676.07	0.671
61	6.396		0.0000	0.000	2683.1	0.774	BB	0.063	711.82	0.707
62	6.506		0.0000	0.000	433.7	0.125	BB	0.023	311.55	0.309
63	6.579		0.0000	0.000	6847.6	1.976	BB	0.028	4045.87	4.016
64	6.677		0.0000	0.000	90.7	0.026	BB	0.022	69.18	0.069
65	6.726		0.0000	0.000	1842.3	0.532	BB	0.060	510.82	0.507
66	6.833		0.0000	0.000	4765.1	1.375	BB	0.036	2190.84	2.175
67	6.893		0.0000	0.000	1821.2	0.526	BB	0.036	850.27	0.844
68	7.106		0.0000	0.000	6181.1	1.784	BB	0.072	1421.45	1.411
69	7.186		0.0000	0.000	5750.7	1.660	BB	0.065	1474.02	1.463
70	7.283		0.0000	0.000	17774.8	5.130	BB	0.276	1071.53	1.064
71	7.517		0.0000	0.000	19090.5	5.510	BB	0.119	2681.99	2.662
72	7.641		0.0000	0.000	6375.7	1.840	BB	0.108	983.00	0.976
73	7.730		0.0000	0.000	817.5	0.236	BB	0.041	332.41	0.330
74	7.844		0.0000	0.000	553.5	0.160	BB	0.080	115.34	0.114
75	7.976	DCB	1.5641	-100.363	11751.6	3.392	BB	0.050	3954.02	3.925
76	8.177		0.0000	0.000	7259.2	2.095	BB	0.192	630.03	0.625
77	8.376		0.0000	0.000	682.3	0.197	BB	0.047	239.81	0.238
78	8.522		0.0000	0.000	112.9	0.033	BB	0.033	57.61	0.057
79	8.585		0.0000	0.000	1379.1	0.398	BB	0.047	489.80	0.486
80	8.694		0.0000	0.000	1618.1	0.467	BB	0.066	406.70	0.404

Total Area = 346493.8, Total Amount = -1.558, Total Height = 100743.1, Sample Units = ug/l

ASX22 0488

Sample Name: **MD16X0924A AAMD16XAA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 10:25:34pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.36R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.584		0.0000	0.000	2517.4	1.435	BV	0.036	1175.41	2.207
2	1.639		0.0000	0.000	17374.5	9.904	VV	0.116	2494.00	4.682
3	1.933		0.0000	0.000	589.3	0.336	VV	0.051	192.10	0.361
4	2.016		0.0000	0.000	1171.1	0.668	VB	0.046	427.60	0.803
5	2.177		0.0000	0.000	6794.8	3.873	BB	0.131	867.47	1.629
6	2.341	TCX	1.0496	-53.449	14609.1	8.328	BB	0.053	4551.89	8.546
7	2.507		0.0000	0.000	377.0	0.215	BB	0.038	164.59	0.309
8	2.647		0.0000	0.000	361.8	0.206	BB	0.035	170.87	0.321
9	2.717		0.0000	0.000	2344.8	1.337	BB	0.048	805.96	1.513
10	3.010		0.0000	0.000	5615.0	3.201	BB	0.048	1961.12	3.682
11	3.171		0.0000	0.000	2967.1	1.691	BB	0.037	1337.50	2.511
12	3.353		0.0000	0.000	945.9	0.539	BB	0.138	114.61	0.215
13	3.405		0.0000	0.000	4061.4	2.315	BB	0.035	1952.70	3.666
14	3.451		0.0000	0.000	498.8	0.284	BB	0.030	281.43	0.528
15	3.562		0.0000	0.000	1564.7	0.892	BB	0.034	777.27	1.459
16	3.620		0.0000	0.000	399.4	0.228	BB	0.025	261.18	0.490
17	3.646		0.0000	0.000	1216.1	0.693	BB	0.037	549.94	1.032
18	3.800		0.0000	0.000	2939.8	1.676	BB	0.048	1017.11	1.910
19	3.887		0.0000	0.000	48.7	0.028	BB	0.013	63.88	0.120
20	3.921		0.0000	0.000	673.0	0.384	BB	0.029	391.38	0.735
21	4.111		0.0000	0.000	4080.8	2.326	BB	0.059	1151.25	2.161
22	4.263		0.0000	0.000	659.6	0.376	BB	0.042	262.11	0.492
23	4.306		0.0000	0.000	2303.9	1.313	BB	0.048	796.10	1.495
24	4.508		0.0000	0.000	149.5	0.085	BB	0.028	88.12	0.165
25	4.544		0.0000	0.000	348.6	0.199	BB	0.024	242.23	0.455
26	4.600		0.0000	0.000	3317.8	1.891	BB	0.050	1099.60	2.064
27	4.696		0.0000	0.000	5345.9	3.047	BB	0.037	2393.48	4.494
28	4.830		0.0000	0.000	854.5	0.487	BB	0.033	428.58	0.805
29	4.876		0.0000	0.000	5074.5	2.893	BB	0.037	2277.97	4.277
30	4.984		0.0000	0.000	4993.9	2.847	BB	0.051	1616.47	3.035
31	5.103		0.0000	0.000	377.2	0.215	BB	0.025	254.44	0.478
32	5.144		0.0000	0.000	1617.2	0.922	BB	0.024	1124.80	2.112
33	5.180		0.0000	0.000	2217.8	1.264	BB	0.034	1091.16	2.049
34	5.339		0.0000	0.000	337.7	0.193	BB	0.024	235.32	0.442
35	5.392		0.0000	0.000	4331.7	2.469	BB	0.037	1967.62	3.694
36	5.475		0.0000	0.000	969.0	0.552	BB	0.025	640.08	1.202
37	5.514		0.0000	0.000	910.6	0.519	BB	0.088	171.65	0.322
38	5.631		0.0000	0.000	10640.2	6.065	BB	0.045	3916.13	7.352
39	5.805		0.0000	0.000	2234.5	1.274	BB	0.033	1129.88	2.121
40	5.870		0.0000	0.000	1048.0	0.597	BB	0.026	684.26	1.285
41	5.927		0.0000	0.000	6140.8	3.501	BB	0.037	2745.83	5.155
42	6.032		0.0000	0.000	124.7	0.071	BB	0.022	95.10	0.179
43	6.241		0.0000	0.000	6144.3	3.502	BB	0.072	1428.40	2.682
44	6.390		0.0000	0.000	17076.7	9.734	BB	0.122	2324.08	4.363
45	6.633		0.0000	0.000	1035.0	0.590	BB	0.038	457.47	0.859
46	6.737		0.0000	0.000	189.4	0.108	BB	0.060	52.85	0.099
47	6.804	DCB	-3.0133	153.449	11508.4	6.560	BB	0.047	4068.52	7.638
48	7.080		0.0000	0.000	116.9	0.067	BB	0.040	49.23	0.092
49	7.490		0.0000	0.000	1528.2	0.871	BB	0.084	301.49	0.566
50	7.616		0.0000	0.000	8173.4	4.659	BB	0.341	399.07	0.749
51	8.487		0.0000	0.000	4506.4	2.569	BB	0.352	213.34	0.401

ASX22 B489

Total Area = 175426.9, Total Amount = -1.964, Total Height = 53264.66, Sample Units = ug/l

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\1P7307.36R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 10:16:25 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.36R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 10:16:25 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	6712	-3.1226	TCX	2.341	4552	1.04959	TCX
7.976	3954	1.56413	DCB	6.804	4069	-3.0133	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.36A
Area File: C:\CPWIN\DATA1\1P7307B.36A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 10:25:42 PM
File Reported On: 11/3/2009 at 10:25:54 PM

10X REPLOTS

ASX22 6491

C:\CPWIN\DATA1\1P7307.36R



C:\CPWIN\DATA1\1P7307B.36R



ASX22 0492

Sample Name: **MD62X0924B AAMD62XAA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 10:38:33pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.37R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.510		0.0000	0.000	83.2	0.036	BV	0.025	56.24	0.087
2	1.586		0.0000	0.000	1714.6	0.747	VV	0.044	643.56	0.994
3	1.665		0.0000	0.000	4287.9	1.868	VV	0.070	1015.47	1.569
4	1.865		0.0000	0.000	17259.5	7.520	VV	0.151	1910.84	2.953
5	1.932		0.0000	0.000	8525.9	3.715	VB	0.061	2321.20	3.587
6	2.027		0.0000	0.000	12681.2	5.525	BV	0.096	2194.19	3.391
7	2.175		0.0000	0.000	21420.2	9.333	VV	0.104	3448.72	5.329
8	2.297		0.0000	0.000	4066.0	1.772	VV	0.065	1038.25	1.604
9	2.366		0.0000	0.000	1919.3	0.836	VV	0.039	813.20	1.257
10	2.446		0.0000	0.000	3076.2	1.340	VV	0.082	627.20	0.969
11	2.526		0.0000	0.000	2153.6	0.938	VV	0.054	664.05	1.026
12	2.589		0.0000	0.000	706.4	0.308	VB	0.049	242.67	0.375
13	2.661		0.0000	0.000	64.7	0.028	BB	0.017	64.43	0.100
14	2.780		0.0000	0.000	1170.4	0.510	BB	0.074	263.31	0.407
15	2.826		0.0000	0.000	21516.3	9.375	BB	0.147	2437.79	3.767
16	3.000	TCX	-4.1321	119.113	19961.0	8.697	BB	0.128	2590.29	4.003
17	3.063		0.0000	0.000	153.6	0.067	BB	0.030	86.21	0.133
18	3.207		0.0000	0.000	902.0	0.393	BB	0.108	138.91	0.215
19	3.290		0.0000	0.000	179.8	0.078	BB	0.035	85.92	0.133
20	3.454		0.0000	0.000	663.4	0.289	BB	0.074	149.06	0.230
21	3.536		0.0000	0.000	1078.7	0.470	BB	0.031	588.02	0.909
22	3.861		0.0000	0.000	1193.7	0.520	BB	0.055	359.48	0.556
23	3.985		0.0000	0.000	252.1	0.110	BB	0.025	167.30	0.259
24	4.036		0.0000	0.000	506.0	0.220	BB	0.039	217.90	0.337
25	4.102		0.0000	0.000	637.5	0.278	BB	0.023	462.21	0.714
26	4.205		0.0000	0.000	558.2	0.243	BB	0.060	155.22	0.240
27	4.266		0.0000	0.000	449.6	0.196	BB	0.028	269.09	0.416
28	4.430		0.0000	0.000	1578.0	0.688	BB	0.072	363.39	0.562
29	4.567		0.0000	0.000	210.7	0.092	BB	0.038	91.22	0.141
30	4.670		0.0000	0.000	898.6	0.392	BB	0.098	152.23	0.235
31	4.814		0.0000	0.000	649.3	0.283	BB	0.039	274.36	0.424
32	4.921		0.0000	0.000	882.6	0.385	BB	0.030	485.00	0.749
33	4.996		0.0000	0.000	830.8	0.362	BB	0.044	315.35	0.487
34	5.103		0.0000	0.000	3850.6	1.678	BB	0.061	1044.98	1.615
35	5.203		0.0000	0.000	1273.9	0.555	BB	0.035	608.80	0.941
36	5.313		0.0000	0.000	514.2	0.224	BB	0.028	307.81	0.476
37	5.381		0.0000	0.000	553.3	0.241	BB	0.024	387.51	0.599
38	5.428		0.0000	0.000	1910.9	0.833	BB	0.026	1231.80	1.904
39	5.521		0.0000	0.000	3661.0	1.595	BB	0.028	2195.43	3.393
40	5.636		0.0000	0.000	458.2	0.200	BB	0.039	196.24	0.303
41	5.766		0.0000	0.000	3757.2	1.637	BB	0.027	2314.91	3.577
42	5.813		0.0000	0.000	1113.8	0.485	BB	0.021	866.90	1.340
43	5.879		0.0000	0.000	2674.8	1.165	BB	0.043	1039.90	1.607
44	5.970		0.0000	0.000	185.7	0.081	BB	0.023	135.27	0.209
45	6.010		0.0000	0.000	2694.4	1.174	BB	0.029	1551.95	2.398
46	6.084		0.0000	0.000	4783.9	2.084	BB	0.025	3183.74	4.920
47	6.136		0.0000	0.000	1533.5	0.668	BB	0.024	1046.31	1.617
48	6.281		0.0000	0.000	5494.6	2.394	BB	0.034	2673.44	4.131
49	6.354		0.0000	0.000	1437.3	0.626	BB	0.025	940.69	1.454
50	6.397		0.0000	0.000	4803.1	2.093	BB	0.088	908.12	1.403
51	6.579		0.0000	0.000	9217.6	4.016	BB	0.029	5319.01	8.220
52	6.727		0.0000	0.000	1188.7	0.518	BB	0.056	354.37	0.548
53	6.828		0.0000	0.000	6169.6	2.688	BB	0.034	3040.24	4.698
54	6.896		0.0000	0.000	3783.0	1.648	BB	0.035	1797.11	2.777

ASX22 9493

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	7.109		0.0000	0.000	4304.5	1.876	BB	0.072	992.98	1.534
56	7.184		0.0000	0.000	4525.5	1.972	BB	0.055	1370.11	2.117
57	7.282		0.0000	0.000	2337.8	1.019	BB	0.117	333.83	0.516
58	7.422		0.0000	0.000	637.3	0.278	BB	0.023	469.54	0.726
59	7.520		0.0000	0.000	4397.1	1.916	BB	0.063	1161.31	1.795
60	7.629		0.0000	0.000	4767.8	2.077	BB	0.093	858.74	1.327
61	7.731		0.0000	0.000	1731.4	0.754	BB	0.035	815.17	1.260
62	7.975	DCB	0.6630	-19.113	6419.3	2.797	BB	0.064	1676.09	2.590
63	8.186		0.0000	0.000	4416.3	1.924	BB	0.229	320.99	0.496
64	8.375		0.0000	0.000	402.5	0.175	BB	0.043	154.90	0.239
65	8.539		0.0000	0.000	95.5	0.042	BB	0.050	31.81	0.049
66	8.592		0.0000	0.000	1295.5	0.564	BB	0.047	457.69	0.707
67	8.687		0.0000	0.000	883.0	0.385	BB	0.063	231.86	0.358

Total Area = 229503.8, Total Amount = -3.469, Total Height = 64711.78, Sample Units = ug/l

ASX22 0494

Sample Name: **MD62X0924B AAMD62XAA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 10:38:33pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.37R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.571		0.0000	0.000	5051.6	3.589	BV	0.043	1966.20	4.611
2	1.639		0.0000	0.000	9692.9	6.887	VB	0.076	2127.00	4.989
3	1.883		0.0000	0.000	189.0	0.134	BB	0.034	91.63	0.215
4	1.967		0.0000	0.000	71.6	0.051	BB	0.032	37.29	0.087
5	2.023		0.0000	0.000	578.5	0.411	BB	0.042	229.95	0.539
6	2.107		0.0000	0.000	10873.0	7.725	BB	0.572	316.86	0.743
7	2.384		0.0000	0.000	5325.4	3.784	BB	0.928	95.65	0.224
8	2.580		0.0000	0.000	997.9	0.709	BB	0.413	40.31	0.095
9	2.880		0.0000	0.000	600.4	0.427	BB	0.193	51.92	0.122
10	3.053		0.0000	0.000	180.6	0.128	BB	0.078	38.53	0.090
11	3.153		0.0000	0.000	2130.8	1.514	BB	0.069	515.87	1.210
12	3.321		0.0000	0.000	21.0	0.015	BB	0.009	37.11	0.087
13	3.353		0.0000	0.000	4992.7	3.547	BB	0.135	615.33	1.443
14	3.887		0.0000	0.000	437.4	0.311	BB	0.044	165.37	0.388
15	4.110		0.0000	0.000	1501.5	1.067	BB	0.055	454.54	1.066
16	4.299		0.0000	0.000	4131.3	2.935	BB	0.095	723.35	1.697
17	4.517		0.0000	0.000	72.0	0.051	BB	0.041	29.38	0.069
18	4.544		0.0000	0.000	401.8	0.285	BB	0.025	265.52	0.623
19	4.600		0.0000	0.000	3286.4	2.335	BB	0.045	1230.34	2.886
20	4.696		0.0000	0.000	5146.0	3.656	BB	0.038	2283.09	5.355
21	4.830		0.0000	0.000	428.9	0.305	BB	0.028	253.98	0.596
22	4.877		0.0000	0.000	5094.5	3.620	BB	0.038	2207.00	5.176
23	4.979		0.0000	0.000	5380.6	3.823	BB	0.050	1777.28	4.168
24	5.103		0.0000	0.000	479.3	0.341	BB	0.024	329.65	0.773
25	5.146		0.0000	0.000	744.7	0.529	BB	0.025	487.29	1.143
26	5.182		0.0000	0.000	7855.7	5.581	BB	0.044	2948.08	6.914
27	5.337		0.0000	0.000	579.9	0.412	BB	0.025	388.23	0.911
28	5.392		0.0000	0.000	6661.4	4.733	BB	0.036	3105.22	7.283
29	5.476		0.0000	0.000	3781.0	2.686	BB	0.056	1121.55	2.630
30	5.631		0.0000	0.000	14370.8	10.210	BB	0.046	5194.58	12.183
31	5.804		0.0000	0.000	1833.9	1.303	BB	0.030	1011.60	2.373
32	5.870		0.0000	0.000	3222.3	2.289	BB	0.028	1911.66	4.484
33	5.928		0.0000	0.000	7385.9	5.248	BB	0.035	3488.04	8.181
34	6.023		0.0000	0.000	352.5	0.250	BB	0.035	166.05	0.389
35	6.093		0.0000	0.000	1102.7	0.783	BB	0.061	298.86	0.701
36	6.241		0.0000	0.000	3823.7	2.717	BB	0.059	1087.16	2.550
37	6.387		0.0000	0.000	7166.4	5.092	BB	0.052	2281.99	5.352
38	6.536		0.0000	0.000	178.4	0.127	BB	0.023	127.04	0.298
39	6.626		0.0000	0.000	2679.8	1.904	BB	0.041	1087.57	2.551
40	6.804	DCB	-4.0335	100.000	2732.5	1.941	BB	0.038	1196.12	2.805
41	7.190		0.0000	0.000	218.7	0.155	BB	0.055	66.70	0.156
42	7.520		0.0000	0.000	1218.8	0.866	BB	0.073	277.36	0.651
43	7.617		0.0000	0.000	7361.0	5.230	BB	0.331	370.11	0.868
44	8.201		0.0000	0.000	100.6	0.071	BB	0.061	27.39	0.064
45	8.490		0.0000	0.000	310.7	0.221	BB	0.047	111.25	0.261

Total Area = 140746.4, Total Amount = -4.033, Total Height = 42637.02, Sample Units = ug/l

ASX22 0495

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.37R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 10:29:24 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.37R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 10:29:24 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.	2590	-4.13206	TCX	.	0	.	TCX
7.975	1676	.66302	DCB	6.804	1196	-4.03347	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.37A
Area File: C:\CPWIN\DATA1\1P7307B.37A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 10:38:40 PM
File Reported On: 11/3/2009 at 10:38:53 PM

10X REPLOTS

03X22 0497

C:\CPWIN\DATA1\1P7307.37R



C:\CPWIN\DATA1\1P7307B.37R



ASX22 0498

Sample Name: **MD68X0924C AAMD68XAA ICAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 10:51:33pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.38R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.554		0.0000	0.000	3466.0	1.156	BV	0.060	957.74	1.163
2	1.669		0.0000	0.000	7395.3	2.467	VV	0.065	1896.66	2.303
3	1.800		0.0000	0.000	21857.2	7.292	VV	0.110	3317.49	4.029
4	1.862		0.0000	0.000	14419.4	4.811	VV	0.060	3990.90	4.846
5	1.928		0.0000	0.000	16687.0	5.567	VB	0.059	4743.36	5.760
6	2.037		0.0000	0.000	23378.0	7.799	BV	0.088	4430.05	5.380
7	2.174		0.0000	0.000	47562.4	15.868	VV	0.113	7023.35	8.529
8	2.450		0.0000	0.000	11643.1	3.884	VV	0.194	998.51	1.213
9	2.527		0.0000	0.000	2482.5	0.828	VB	0.061	676.07	0.821
10	2.770		0.0000	0.000	4908.8	1.638	BB	0.084	975.37	1.184
11	2.884		0.0000	0.000	4291.7	1.432	BB	0.058	1225.08	1.488
12	3.055		0.0000	0.000	7299.5	2.435	BB	0.219	556.15	0.675
13	3.205		0.0000	0.000	492.7	0.164	BB	0.075	109.21	0.133
14	3.291		0.0000	0.000	326.8	0.109	BB	0.039	138.30	0.168
15	3.335		0.0000	0.000	305.2	0.102	BB	0.023	218.20	0.265
16	3.456		0.0000	0.000	956.6	0.319	BB	0.115	139.09	0.169
17	3.536		0.0000	0.000	196.6	0.066	BB	0.029	111.44	0.135
18	3.615		0.0000	0.000	60.3	0.020	BB	0.023	44.18	0.054
19	3.775		0.0000	0.000	384.7	0.128	BB	0.104	61.55	0.075
20	3.861		0.0000	0.000	1782.2	0.595	BB	0.048	620.42	0.753
21	3.979		0.0000	0.000	243.2	0.081	BB	0.035	114.84	0.139
22	4.103		0.0000	0.000	370.1	0.123	BB	0.021	297.65	0.361
23	4.205		0.0000	0.000	363.0	0.121	BB	0.055	109.52	0.133
24	4.266		0.0000	0.000	2244.5	0.749	BB	0.030	1239.98	1.506
25	4.432		0.0000	0.000	1393.9	0.465	BB	0.064	363.40	0.441
26	4.571		0.0000	0.000	309.2	0.103	BB	0.036	141.28	0.172
27	4.664		0.0000	0.000	159.9	0.053	BB	0.039	67.97	0.083
28	4.816		0.0000	0.000	1977.3	0.660	BB	0.060	547.78	0.665
29	4.921		0.0000	0.000	957.5	0.319	BB	0.032	503.73	0.612
30	4.997		0.0000	0.000	860.6	0.287	BB	0.045	318.74	0.387
31	5.102		0.0000	0.000	2871.5	0.958	BB	0.051	929.34	1.129
32	5.203		0.0000	0.000	1325.6	0.442	BB	0.036	612.40	0.744
33	5.383		0.0000	0.000	2154.3	0.719	BB	0.059	605.58	0.735
34	5.514		0.0000	0.000	795.1	0.265	BB	0.092	144.76	0.176
35	5.635		0.0000	0.000	409.8	0.137	BB	0.051	133.06	0.162
36	5.817		0.0000	0.000	2244.8	0.749	BB	0.116	322.23	0.391
37	5.900		0.0000	0.000	1177.3	0.393	BB	0.034	578.86	0.703
38	5.970		0.0000	0.000	463.2	0.155	BB	0.031	247.06	0.300
39	6.034		0.0000	0.000	92.3	0.031	BB	0.022	69.62	0.085
40	6.084		0.0000	0.000	1842.1	0.615	BB	0.027	1129.73	1.372
41	6.162		0.0000	0.000	130.1	0.043	BB	0.032	67.65	0.082
42	6.292		0.0000	0.000	2649.2	0.884	BB	0.029	1502.68	1.825
43	6.394		0.0000	0.000	11896.1	3.969	BB	0.108	1832.57	2.225
44	6.580		0.0000	0.000	1140.3	0.380	BB	0.030	625.54	0.760
45	6.728		0.0000	0.000	1635.4	0.546	BB	0.069	393.09	0.477
46	6.824		0.0000	0.000	10786.1	3.598	BB	0.029	6160.86	7.481
47	6.899		0.0000	0.000	8909.4	2.972	BB	0.030	4975.88	6.042
48	7.056		0.0000	0.000	10304.7	3.438	BB	0.035	4852.95	5.893
49	7.194		0.0000	0.000	5554.9	1.853	BB	0.243	380.84	0.462
50	7.422		0.0000	0.000	550.6	0.184	BB	0.023	406.36	0.493
51	7.521		0.0000	0.000	4012.2	1.339	BB	0.060	1105.49	1.342
52	7.631		0.0000	0.000	4038.0	1.347	BB	0.092	732.00	0.889
53	7.730		0.0000	0.000	27700.9	9.242	BB	0.034	13549.46	16.454
54	7.977 DCB		1.5054	100.000	10699.4	3.570	BB	0.047	3805.60	4.621

ASX22: 8499

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	8.190		0.0000	0.000	4656.7	1.554	BB	0.221	351.33	0.427
56	8.277		0.0000	0.000	280.4	0.094	BB	0.066	70.87	0.086
57	8.378		0.0000	0.000	515.0	0.172	BB	0.040	212.06	0.258
58	8.517		0.0000	0.000	101.2	0.034	BB	0.035	48.25	0.059
59	8.584		0.0000	0.000	1250.0	0.417	BB	0.057	368.33	0.447
60	8.688		0.0000	0.000	781.3	0.261	BB	0.066	197.04	0.239

Total Area = 299743.4, Total Amount = 1.505, Total Height = 82349.52, Sample Units = ug/l

ASX22 9500

Sample Name: **MD68X0924C AAMD68XAA ICAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 10:51:33pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.38R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.568		0.0000	0.000	10087.0	6.141	BV	0.045	3772.87	6.261
2	1.637		0.0000	0.000	12612.1	7.678	VV	0.090	2334.53	3.874
3	2.014		0.0000	0.000	1690.8	1.029	VB	0.049	577.05	0.958
4	2.112		0.0000	0.000	1124.3	0.684	BB	0.097	193.55	0.321
5	2.286		0.0000	0.000	1636.6	0.996	BB	0.063	430.71	0.715
6	2.460		0.0000	0.000	128.3	0.078	BB	0.051	42.24	0.070
7	2.657		0.0000	0.000	501.2	0.305	BB	0.187	44.72	0.074
8	2.873		0.0000	0.000	511.2	0.311	BB	0.155	55.07	0.091
9	3.159		0.0000	0.000	2557.3	1.557	BB	0.057	752.60	1.249
10	3.313		0.0000	0.000	47.2	0.029	BB	0.021	37.18	0.062
11	3.353		0.0000	0.000	1863.7	1.135	BB	0.064	481.90	0.800
12	3.557		0.0000	0.000	6561.4	3.994	BB	0.076	1446.78	2.401
13	3.893		0.0000	0.000	634.7	0.386	BB	0.052	204.27	0.339
14	4.140		0.0000	0.000	2161.6	1.316	BB	0.163	221.16	0.367
15	4.300		0.0000	0.000	2823.5	1.719	BB	0.080	588.08	0.976
16	4.602		0.0000	0.000	308.2	0.188	BB	0.020	251.15	0.417
17	4.840		0.0000	0.000	767.7	0.467	BB	0.045	287.52	0.477
18	4.973		0.0000	0.000	605.5	0.369	BB	0.051	197.95	0.328
19	5.130		0.0000	0.000	228.2	0.139	BB	0.033	114.15	0.189
20	5.183		0.0000	0.000	3556.3	2.165	BB	0.048	1236.22	2.051
21	5.385		0.0000	0.000	4678.2	2.848	BB	0.037	2109.06	3.500
22	5.458		0.0000	0.000	525.6	0.320	BB	0.030	291.37	0.484
23	5.633		0.0000	0.000	824.4	0.502	BB	0.031	437.91	0.727
24	5.683		0.0000	0.000	601.6	0.366	BB	0.034	298.82	0.496
25	5.804		0.0000	0.000	2644.4	1.610	BB	0.030	1461.94	2.426
26	5.871		0.0000	0.000	10368.0	6.312	BB	0.031	5637.11	9.355
27	5.930		0.0000	0.000	9745.0	5.932	BB	0.032	5111.46	8.482
28	6.091		0.0000	0.000	11871.0	7.227	BB	0.036	5536.69	9.188
29	6.172		0.0000	0.000	4915.3	2.992	BB	0.067	1224.71	2.032
30	6.387		0.0000	0.000	6486.2	3.949	BB	0.049	2218.63	3.682
31	6.533		0.0000	0.000	417.8	0.254	BB	0.039	178.73	0.297
32	6.624		0.0000	0.000	42184.3	25.680	BB	0.041	17205.37	28.552
33	6.804	DCB	-2.9611	100.000	9757.1	5.940	BB	0.039	4215.45	6.995
34	7.490		0.0000	0.000	2627.4	1.599	BB	0.138	316.92	0.526
35	7.637		0.0000	0.000	1807.3	1.100	BB	0.134	224.70	0.373
36	8.073		0.0000	0.000	1696.0	1.032	BB	0.096	293.83	0.488
37	8.220		0.0000	0.000	2712.1	1.651	BB	0.199	227.52	0.378

Total Area = 164268.5, Total Amount = -2.961, Total Height = 60259.91, Sample Units = ug/l

ASX22 0581

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.38R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 10:42:24 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.38R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 10:42:24 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
7.977	3806	1.50541	DCB	6.804	4215	-2.96111	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.38A
Area File: C:\CPWIN\DATA1\1P7307B.38A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/3/2009 10:51:40 PM
File Reported On: 11/3/2009 at 10:51:53 PM

10X REPLOTS

&SX22 0503

C:\CPWIN\DATA1\1P7307.38R



C:\CPWIN\DATA1\1P7307B.38R



ASX22: 0504

Sample Name: **IC16X0924A AAIC16XAA CCAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 11:04:29pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.39R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.584		0.0000	0.000	4935.4	0.224	BV	0.054	1514.50	0.133
2	1.666		0.0000	0.000	12583.8	0.570	VV	0.067	3149.04	0.276
3	1.865		0.0000	0.000	60425.8	2.737	VV	0.149	6752.78	0.591
4	1.925		0.0000	0.000	28869.6	1.308	VB	0.061	7896.21	0.691
5	2.073		0.0000	0.000	45315.4	2.052	BV	0.147	5142.06	0.450
6	2.173		0.0000	0.000	4815.1	0.218	VB	0.041	1959.15	0.171
7	2.447		0.0000	0.000	363.7	0.016	BB	0.066	92.36	0.008
8	2.520		0.0000	0.000	610.9	0.028	BV	0.026	395.66	0.035
9	2.593		0.0000	0.000	354.2	0.016	VB	0.056	104.75	0.009
10	2.770		0.0000	0.000	16929.3	0.767	BB	0.096	2935.57	0.257
11	2.883		0.0000	0.000	11368.5	0.515	BB	0.062	3032.82	0.265
12	2.997	TCX	44.3564	49.046	337113.8	15.269	BB	0.028	200573.40	17.550
13	3.225		0.0000	0.000	6263.0	0.284	BB	0.026	3943.61	0.345
14	3.308		0.0000	0.000	12344.7	0.559	BB	0.040	5181.30	0.453
15	3.380		0.0000	0.000	36282.4	1.643	BB	0.027	22516.93	1.970
16	3.471		0.0000	0.000	472.5	0.021	BB	0.019	420.16	0.037
17	3.513		0.0000	0.000	5054.0	0.229	BB	0.026	3288.89	0.288
18	3.692		0.0000	0.000	76442.1	3.462	BB	0.040	32047.41	2.804
19	3.770		0.0000	0.000	3693.7	0.167	BB	0.027	2315.76	0.203
20	3.822		0.0000	0.000	1438.5	0.065	BB	0.017	1425.39	0.125
21	3.861		0.0000	0.000	42475.4	1.924	BB	0.026	27105.21	2.372
22	3.937		0.0000	0.000	305.7	0.014	BB	0.016	328.49	0.029
23	3.968		0.0000	0.000	735.4	0.033	BB	0.021	595.77	0.052
24	4.029		0.0000	0.000	21956.9	0.994	BB	0.039	9391.72	0.822
25	4.124		0.0000	0.000	125279.8	5.674	BB	0.033	63262.36	5.535
26	4.179		0.0000	0.000	38950.9	1.764	BB	0.024	27258.50	2.385
27	4.264		0.0000	0.000	41158.4	1.864	BB	0.027	25834.61	2.261
28	4.314		0.0000	0.000	1306.4	0.059	BB	0.016	1336.07	0.117
29	4.358		0.0000	0.000	19643.5	0.890	BB	0.020	16619.03	1.454
30	4.394		0.0000	0.000	31505.0	1.427	BB	0.033	16073.80	1.406
31	4.526		0.0000	0.000	53817.1	2.437	BB	0.036	24661.09	2.158
32	4.597		0.0000	0.000	4494.9	0.204	BB	0.017	4427.78	0.387
33	4.646		0.0000	0.000	24596.4	1.114	BB	0.028	14847.54	1.299
34	4.696		0.0000	0.000	4483.8	0.203	BB	0.021	3615.37	0.316
35	4.760		0.0000	0.000	239.4	0.011	BB	0.017	228.48	0.020
36	4.813		0.0000	0.000	490.4	0.022	BB	0.024	340.82	0.030
37	4.855		0.0000	0.000	28511.2	1.291	BB	0.027	17543.62	1.535
38	4.912		0.0000	0.000	1872.8	0.085	BB	0.017	1801.86	0.158
39	5.001		0.0000	0.000	3303.0	0.150	BB	0.023	2415.53	0.211
40	5.057		0.0000	0.000	36727.3	1.663	BB	0.031	19966.05	1.747
41	5.202		0.0000	0.000	335.2	0.015	BB	0.030	187.17	0.016
42	5.231		0.0000	0.000	828.4	0.038	BB	0.022	617.52	0.054
43	5.307		0.0000	0.000	19311.4	0.875	BB	0.031	10543.83	0.923
44	5.374		0.0000	0.000	14258.6	0.646	BB	0.024	10035.16	0.878
45	5.427		0.0000	0.000	48645.8	2.203	BB	0.038	21572.52	1.888
46	5.520		0.0000	0.000	68690.3	3.111	BB	0.027	42378.84	3.708
47	5.635		0.0000	0.000	10892.0	0.493	BB	0.037	4931.13	0.431
48	5.701		0.0000	0.000	10607.5	0.480	BB	0.025	7019.75	0.614
49	5.765		0.0000	0.000	83356.0	3.775	BB	0.030	46970.68	4.110
50	5.809		0.0000	0.000	3468.9	0.157	BB	0.015	3951.66	0.346
51	5.874		0.0000	0.000	40457.2	1.832	BB	0.029	23338.60	2.042
52	5.966		0.0000	0.000	3630.3	0.164	BB	0.022	2755.18	0.241
53	6.010		0.0000	0.000	83914.1	3.801	BB	0.029	47742.73	4.177
54	6.081		0.0000	0.000	36534.1	1.655	BB	0.024	25844.70	2.261

ASX22 0505

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.133		0.0000	0.000	21030.7	0.953	BB	0.024	14493.65	1.268
56	6.278		0.0000	0.000	67944.9	3.077	BB	0.035	32323.42	2.828
57	6.351		0.0000	0.000	21373.2	0.968	BB	0.025	14068.58	1.231
58	6.399		0.0000	0.000	8090.9	0.366	BB	0.022	6266.00	0.548
59	6.443		0.0000	0.000	670.1	0.030	BB	0.016	681.22	0.060
60	6.509		0.0000	0.000	14293.6	0.647	BB	0.027	8976.02	0.785
61	6.577		0.0000	0.000	115786.7	5.244	BB	0.030	63986.50	5.599
62	6.832		0.0000	0.000	53519.4	2.424	BB	0.029	30987.95	2.711
63	6.888		0.0000	0.000	25274.7	1.145	BB	0.037	11422.68	0.999
64	7.056		0.0000	0.000	2261.3	0.102	BB	0.029	1278.02	0.112
65	7.180		0.0000	0.000	38189.2	1.730	BB	0.079	8041.68	0.704
66	7.417		0.0000	0.000	10467.5	0.474	BB	0.026	6833.14	0.598
67	7.491		0.0000	0.000	5566.2	0.252	BB	0.061	1511.65	0.132
68	7.634		0.0000	0.000	4249.4	0.192	BB	0.097	728.31	0.064
69	7.729		0.0000	0.000	13456.9	0.609	BB	0.034	6629.18	0.580
70	7.860		0.0000	0.000	413.8	0.019	BB	0.052	133.67	0.012
71	7.975	DCB	46.0823	50.954	224960.5	10.189	BB	0.036	102859.40	9.000
72	8.097		0.0000	0.000	2076.9	0.094	BB	0.108	320.94	0.028
73	8.376		0.0000	0.000	697.5	0.032	BB	0.058	199.27	0.017
74	8.586		0.0000	0.000	4440.4	0.201	BB	0.104	714.65	0.063
75	8.686		0.0000	0.000	672.3	0.030	BB	0.063	176.79	0.015

Total Area = 2207896.0, Total Amount = 90.439, Total Height = 1142864.0, Sample Units = ug/l

ASXZ2 0586

Sample Name: **IC16X0924A AAIC16XAA CCAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 11:04:29pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.39R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.579		0.0000	0.000	11901.3	0.506	BV	0.061	3241.05	0.330
2	1.635		0.0000	0.000	9539.0	0.405	VV	0.055	2908.92	0.296
3	1.703		0.0000	0.000	18035.2	0.767	VV	0.182	1652.65	0.168
4	2.010		0.0000	0.000	9727.8	0.413	VB	0.087	1853.49	0.189
5	2.117		0.0000	0.000	414.4	0.018	BB	0.033	211.26	0.022
6	2.340	TCX	41.5473	50.599	429481.9	18.255	BB	0.040	180183.80	18.352
7	2.563		0.0000	0.000	5143.7	0.219	BB	0.032	2638.82	0.269
8	2.641		0.0000	0.000	9084.1	0.386	BB	0.041	3670.50	0.374
9	2.715		0.0000	0.000	42019.9	1.786	BB	0.044	16086.06	1.638
10	2.859		0.0000	0.000	6442.2	0.274	BB	0.036	2951.72	0.301
11	3.007		0.0000	0.000	107987.1	4.590	BB	0.049	36600.43	3.728
12	3.171		0.0000	0.000	61706.1	2.623	BB	0.040	25491.85	2.596
13	3.311		0.0000	0.000	16858.7	0.717	BB	0.045	6203.58	0.632
14	3.404		0.0000	0.000	76233.8	3.240	BB	0.034	37554.93	3.825
15	3.449		0.0000	0.000	11944.8	0.508	BB	0.028	7150.99	0.728
16	3.557		0.0000	0.000	28193.1	1.198	BB	0.031	15340.16	1.562
17	3.618		0.0000	0.000	5516.8	0.234	BB	0.047	1965.66	0.200
18	3.644		0.0000	0.000	26992.1	1.147	BB	0.040	11157.25	1.136
19	3.798		0.0000	0.000	66698.1	2.835	BB	0.052	21398.51	2.179
20	3.886		0.0000	0.000	252.7	0.011	BB	0.010	438.57	0.045
21	3.922		0.0000	0.000	13470.3	0.573	BB	0.028	7927.17	0.807
22	3.970		0.0000	0.000	2457.3	0.104	BB	0.025	1630.22	0.166
23	4.108		0.0000	0.000	47557.8	2.021	BB	0.043	18368.43	1.871
24	4.202		0.0000	0.000	1223.8	0.052	BB	0.020	1029.10	0.105
25	4.252		0.0000	0.000	44654.9	1.898	BB	0.045	16563.34	1.687
26	4.433		0.0000	0.000	930.3	0.040	BB	0.030	521.46	0.053
27	4.505		0.0000	0.000	1999.3	0.085	BB	0.023	1480.82	0.151
28	4.542		0.0000	0.000	6128.1	0.260	BB	0.023	4355.39	0.444
29	4.598		0.0000	0.000	68101.7	2.895	BB	0.051	22123.01	2.253
30	4.694		0.0000	0.000	88799.3	3.774	BB	0.035	42036.25	4.281
31	4.820		0.0000	0.000	12949.2	0.550	BB	0.036	5946.96	0.606
32	4.874		0.0000	0.000	100431.4	4.269	BB	0.037	45588.40	4.643
33	4.986		0.0000	0.000	109329.0	4.647	BB	0.054	33777.84	3.440
34	5.100		0.0000	0.000	7225.1	0.307	BB	0.023	5151.92	0.525
35	5.143		0.0000	0.000	34781.0	1.478	BB	0.025	23231.47	2.366
36	5.175		0.0000	0.000	15479.8	0.658	BB	0.021	12022.78	1.225
37	5.219		0.0000	0.000	9228.9	0.392	BB	0.022	7110.01	0.724
38	5.337		0.0000	0.000	5767.4	0.245	BB	0.025	3802.81	0.387
39	5.391		0.0000	0.000	71610.7	3.044	BB	0.037	32459.89	3.306
40	5.474		0.0000	0.000	17322.0	0.736	BB	0.025	11567.36	1.178
41	5.510		0.0000	0.000	2888.9	0.123	BB	0.019	2478.32	0.252
42	5.561		0.0000	0.000	3347.4	0.142	BB	0.021	2649.27	0.270
43	5.628		0.0000	0.000	166485.6	7.076	BB	0.042	65989.70	6.721
44	5.802		0.0000	0.000	974.3	0.041	BB	0.023	702.04	0.072
45	5.869		0.0000	0.000	17281.9	0.735	BB	0.026	10927.03	1.113
46	5.924		0.0000	0.000	103989.9	4.420	BB	0.040	43617.61	4.442
47	6.090		0.0000	0.000	3000.2	0.128	BB	0.034	1474.73	0.150
48	6.174		0.0000	0.000	2449.5	0.104	BB	0.030	1344.67	0.137
49	6.239		0.0000	0.000	24691.3	1.049	BB	0.044	9398.63	0.957
50	6.385		0.0000	0.000	62417.8	2.653	BB	0.055	18983.79	1.934
51	6.622		0.0000	0.000	20311.3	0.863	BB	0.040	8391.61	0.855
52	6.801	DCB	40.5632	49.401	338213.7	14.375	BB	0.040	140032.80	14.262
53	7.503		0.0000	0.000	766.6	0.033	BB	0.060	211.88	0.022
54	7.987		0.0000	0.000	2304.8	0.098	BB	0.167	230.64	0.023

ASX22 0587

Total Area = 2352743.0, Total Amount = 82.11, Total Height = 981827.6, Sample Units = ug/l

ASX22 0588

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.39R

C:\CPWIN\DATA1\1P7307.39R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 10:55:20 PM Column ID: RTX-CLP1 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.39R

C:\CPWIN\DATA1\1P7307B.39R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 10:55:20 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.997	200573	44.35638	TCX	2.34	180184	41.54728	TCX
7.975	102859	46.0823	DCB	6.801	140033	40.56315	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.39A
Area File: C:\CPWIN\DATA1\1P7307B.39A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/4/2009 9:15:32 AM
File Reported On: 11/4/2009 at 9:15:44 AM

10X REPLOTS

ASX22 8510

C:\CPWIN\DATA1\1P7307.39R



C:\CPWIN\DATA1\1P7307B.39R



ASX22 0511

Sample Name: **IC48X0924B AAIC48XAA CCAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 11:17:30pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.40R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.580		0.0000	0.000	4551.0	0.323	BV	0.051	1475.57	0.211
2	1.675		0.0000	0.000	6384.8	0.453	VV	0.064	1655.04	0.237
3	1.927		0.0000	0.000	37711.7	2.676	VB	0.172	3647.90	0.522
4	2.073		0.0000	0.000	33220.1	2.358	BV	0.123	4513.19	0.645
5	2.175		0.0000	0.000	12193.3	0.865	VV	0.058	3477.22	0.497
6	2.207		0.0000	0.000	7451.0	0.529	VV	0.037	3355.62	0.480
7	2.447		0.0000	0.000	20878.2	1.482	VV	0.317	1098.72	0.157
8	2.523		0.0000	0.000	3172.4	0.225	VV	0.059	895.13	0.128
9	2.623		0.0000	0.000	527.5	0.037	VB	0.039	224.23	0.032
10	2.766		0.0000	0.000	5888.7	0.418	BB	0.065	1499.90	0.214
11	2.826		0.0000	0.000	16582.8	1.177	BB	0.102	2716.82	0.389
12	2.997	TCX	45.4694	49.409	347750.9	24.680	BB	0.028	205117.80	29.332
13	3.225		0.0000	0.000	283.7	0.020	BB	0.023	203.35	0.029
14	3.286		0.0000	0.000	7556.1	0.536	BB	0.031	4041.23	0.578
15	3.380		0.0000	0.000	5611.8	0.398	BB	0.035	2652.89	0.379
16	3.471		0.0000	0.000	610.0	0.043	BB	0.020	502.00	0.072
17	3.513		0.0000	0.000	965.0	0.068	BB	0.024	657.47	0.094
18	3.607		0.0000	0.000	176.0	0.012	BB	0.042	70.43	0.010
19	3.692		0.0000	0.000	30083.9	2.135	BB	0.036	14041.46	2.008
20	3.770		0.0000	0.000	3220.5	0.229	BB	0.059	914.44	0.131
21	3.862		0.0000	0.000	18099.0	1.285	BB	0.028	10617.14	1.518
22	3.939		0.0000	0.000	975.4	0.069	BB	0.032	513.87	0.073
23	4.028		0.0000	0.000	5128.5	0.364	BB	0.033	2564.16	0.367
24	4.124		0.0000	0.000	67987.7	4.825	BB	0.035	32136.78	4.596
25	4.179		0.0000	0.000	14213.6	1.009	BB	0.024	10069.90	1.440
26	4.262		0.0000	0.000	23930.1	1.698	BB	0.029	13726.71	1.963
27	4.317		0.0000	0.000	1003.2	0.071	BB	0.016	1014.28	0.145
28	4.358		0.0000	0.000	26568.9	1.886	BB	0.019	22751.29	3.253
29	4.394		0.0000	0.000	32507.4	2.307	BB	0.032	17069.80	2.441
30	4.525		0.0000	0.000	67026.4	4.757	BB	0.034	32656.74	4.670
31	4.597		0.0000	0.000	6799.3	0.483	BB	0.017	6499.40	0.929
32	4.646		0.0000	0.000	32652.3	2.317	BB	0.026	21204.66	3.032
33	4.696		0.0000	0.000	9931.3	0.705	BB	0.024	6773.03	0.969
34	4.769		0.0000	0.000	776.4	0.055	BB	0.035	368.97	0.053
35	4.813		0.0000	0.000	2357.4	0.167	BB	0.025	1572.18	0.225
36	4.856		0.0000	0.000	14843.1	1.053	BB	0.019	12981.69	1.856
37	4.884		0.0000	0.000	11649.1	0.827	BB	0.015	13036.65	1.864
38	4.912		0.0000	0.000	20064.2	1.424	BB	0.018	18879.29	2.700
39	5.000		0.0000	0.000	2692.8	0.191	BB	0.021	2090.33	0.299
40	5.051		0.0000	0.000	97798.9	6.941	BB	0.053	30477.67	4.358
41	5.230		0.0000	0.000	14718.5	1.045	BB	0.031	7842.10	1.121
42	5.297		0.0000	0.000	12079.2	0.857	BB	0.022	8982.99	1.285
43	5.333		0.0000	0.000	3647.7	0.259	BB	0.016	3757.22	0.537
44	5.373		0.0000	0.000	24047.6	1.707	BB	0.024	16455.09	2.353
45	5.476		0.0000	0.000	9552.6	0.678	BB	0.025	6253.91	0.894
46	5.518		0.0000	0.000	800.9	0.057	BB	0.027	491.33	0.070
47	5.579		0.0000	0.000	597.5	0.042	BB	0.014	734.93	0.105
48	5.637		0.0000	0.000	22291.8	1.582	BB	0.029	12849.53	1.837
49	5.732		0.0000	0.000	804.3	0.057	BB	0.024	555.52	0.079
50	5.773		0.0000	0.000	2774.1	0.197	BB	0.031	1487.33	0.213
51	5.875		0.0000	0.000	19965.7	1.417	BB	0.030	11171.10	1.597
52	5.964		0.0000	0.000	1007.6	0.072	BB	0.029	582.59	0.083
53	6.015		0.0000	0.000	5285.1	0.375	BB	0.047	1884.85	0.270
54	6.154		0.0000	0.000	161.0	0.011	BB	0.038	69.79	0.010

ASX22 8512

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.267		0.0000	0.000	1738.5	0.123	BB	0.035	827.59	0.118
56	6.352		0.0000	0.000	1650.0	0.117	BB	0.047	583.20	0.083
57	6.439		0.0000	0.000	2451.7	0.174	BB	0.058	704.50	0.101
58	6.507		0.0000	0.000	414.0	0.029	BB	0.023	295.23	0.042
59	6.575		0.0000	0.000	618.3	0.044	BB	0.024	435.24	0.062
60	6.739		0.0000	0.000	2048.6	0.145	BB	0.045	765.76	0.110
61	6.834		0.0000	0.000	2827.3	0.201	BB	0.059	804.84	0.115
62	6.891		0.0000	0.000	275.5	0.020	BB	0.037	125.03	0.018
63	7.107		0.0000	0.000	5364.7	0.381	BB	0.075	1194.93	0.171
64	7.200		0.0000	0.000	20417.0	1.449	BB	0.217	1564.83	0.224
65	7.515		0.0000	0.000	9599.3	0.681	BB	0.093	1717.05	0.246
66	7.653		0.0000	0.000	5124.2	0.364	BB	0.142	600.74	0.086
67	7.729		0.0000	0.000	2512.0	0.178	BB	0.034	1228.34	0.176
68	7.870		0.0000	0.000	425.5	0.030	BB	0.045	158.05	0.023
69	7.974	DCB	46.5579	50.591	225931.3	16.035	BB	0.036	103822.70	14.847
70	8.092		0.0000	0.000	2827.1	0.201	BB	0.096	490.91	0.070
71	8.580		0.0000	0.000	4040.2	0.287	BB	0.087	776.02	0.111
72	8.689		0.0000	0.000	1190.6	0.085	BB	0.063	316.13	0.045

Total Area = 1409014.0, Total Amount = 92.027, Total Height = 699294.3, Sample Units = ug/l

ASX22 8513

Sample Name: **IC48X0924B AAIC48XAA CCAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 11:17:30pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.40R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.585		0.0000	0.000	6491.4	0.474	BV	0.049	2230.26	0.407
2	1.639		0.0000	0.000	14154.9	1.033	VB	0.094	2511.82	0.459
3	1.978		0.0000	0.000	1658.0	0.121	BV	0.108	255.02	0.047
4	2.019		0.0000	0.000	1214.5	0.089	VB	0.049	409.54	0.075
5	2.286		0.0000	0.000	1685.8	0.123	BB	0.036	770.93	0.141
6	2.340	TCX	41.3650	50.345	422135.7	30.818	BB	0.039	179393.40	32.762
7	2.554		0.0000	0.000	214.1	0.016	BB	0.030	120.48	0.022
8	2.622		0.0000	0.000	5093.8	0.372	BB	0.037	2294.51	0.419
9	2.714		0.0000	0.000	4055.6	0.296	BB	0.047	1431.44	0.261
10	2.860		0.0000	0.000	1932.0	0.141	BB	0.043	749.10	0.137
11	3.006		0.0000	0.000	42665.1	3.115	BB	0.047	14975.73	2.735
12	3.170		0.0000	0.000	20748.2	1.515	BB	0.040	8650.52	1.580
13	3.274		0.0000	0.000	163.3	0.012	BB	0.046	58.97	0.011
14	3.306		0.0000	0.000	3497.0	0.255	BB	0.042	1379.25	0.252
15	3.404		0.0000	0.000	109239.3	7.975	BB	0.070	26096.05	4.766
16	3.556		0.0000	0.000	16031.7	1.170	BB	0.031	8649.72	1.580
17	3.621		0.0000	0.000	8119.4	0.593	BB	0.045	3026.03	0.553
18	3.643		0.0000	0.000	25854.5	1.888	BB	0.036	11912.91	2.176
19	3.797		0.0000	0.000	93218.6	6.805	BB	0.055	28381.74	5.183
20	3.922		0.0000	0.000	22029.4	1.608	BB	0.029	12845.42	2.346
21	3.972		0.0000	0.000	5041.3	0.368	BB	0.024	3477.44	0.635
22	4.116		0.0000	0.000	18578.1	1.356	BB	0.032	9694.31	1.770
23	4.143		0.0000	0.000	12126.8	0.885	BB	0.030	6767.81	1.236
24	4.252		0.0000	0.000	6062.6	0.443	BB	0.024	4185.96	0.764
25	4.293		0.0000	0.000	56506.6	4.125	BB	0.049	19242.01	3.514
26	4.432		0.0000	0.000	10733.0	0.784	BB	0.033	5482.53	1.001
27	4.506		0.0000	0.000	8524.7	0.622	BB	0.026	5480.41	1.001
28	4.538		0.0000	0.000	1619.1	0.118	BB	0.017	1613.80	0.295
29	4.582		0.0000	0.000	26241.5	1.916	BB	0.034	12870.45	2.350
30	4.693		0.0000	0.000	14828.2	1.083	BB	0.038	6538.85	1.194
31	4.794		0.0000	0.000	20998.6	1.533	BB	0.041	8458.89	1.545
32	4.884		0.0000	0.000	3180.2	0.232	BB	0.043	1226.85	0.224
33	4.972		0.0000	0.000	2674.6	0.195	BB	0.040	1104.11	0.202
34	5.050		0.0000	0.000	17856.2	1.304	BB	0.041	7190.56	1.313
35	5.144		0.0000	0.000	6986.2	0.510	BB	0.049	2357.96	0.431
36	5.387		0.0000	0.000	158.8	0.012	BB	0.026	100.37	0.018
37	5.416		0.0000	0.000	973.7	0.071	BB	0.028	574.15	0.105
38	5.633		0.0000	0.000	2295.5	0.168	BB	0.061	626.26	0.114
39	5.801		0.0000	0.000	1694.9	0.124	BB	0.044	635.62	0.116
40	5.921		0.0000	0.000	1368.2	0.100	BB	0.039	587.51	0.107
41	6.424		0.0000	0.000	882.3	0.064	BB	0.341	43.11	0.008
42	6.467		0.0000	0.000	82.1	0.006	BB	0.029	47.17	0.009
43	6.530		0.0000	0.000	380.8	0.028	BB	0.034	185.20	0.034
44	6.623		0.0000	0.000	3988.0	0.291	BB	0.041	1610.31	0.294
45	6.801	DCB	40.7986	49.655	342473.3	25.002	BB	0.041	140845.80	25.722
46	7.363		0.0000	0.000	228.7	0.017	BB	0.036	105.60	0.019
47	7.510		0.0000	0.000	590.3	0.043	BB	0.049	201.82	0.037
48	8.230		0.0000	0.000	213.2	0.016	BB	0.067	53.04	0.010
49	8.497		0.0000	0.000	2271.6	0.166	BB	0.315	120.09	0.022

Total Area = 1369762.0, Total Amount = 82.164, Total Height = 547570.8, Sample Units = ug/l

ASX22 0514

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.40R

C:\CPWIN\DATA1\1P7307.40R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 11:08:21 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.40R

C:\CPWIN\DATA1\1P7307B.40R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 11:08:21 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.997	205118	45.46937	TCX	2.34	179393	41.36504	TCX
7.974	103823	46.55792	DCB	6.801	140846	40.79864	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.40A
Area File: C:\CPWIN\DATA1\1P7307B.40A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/4/2009 9:15:58 AM
File Reported On: 11/4/2009 at 9:16:13 AM

10X REPLOTS

ASX22 0516

C:\CPWIN\DATA1\1P7307.40R



C:\CPWIN\DATA1\1P7307B.40R



RSX22 0517

Sample Name: **IC54X0924B AAIC54XAA CCAL 0930699999**
Acquired from CP07--H5088A via port 3 on 11/3/09 11:30:31pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307.41R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.582		0.0000	0.000	2382.9	0.151	BV	0.048	828.23	0.095
2	1.671		0.0000	0.000	10018.3	0.635	VV	0.094	1783.48	0.205
3	1.787		0.0000	0.000	16594.0	1.051	VV	0.097	2845.14	0.327
4	1.868		0.0000	0.000	13828.4	0.876	VV	0.066	3485.35	0.401
5	1.929		0.0000	0.000	14135.7	0.896	VB	0.056	4243.39	0.488
6	2.038		0.0000	0.000	16330.8	1.035	BV	0.079	3452.94	0.397
7	2.075		0.0000	0.000	14450.7	0.916	VV	0.060	4006.44	0.461
8	2.176		0.0000	0.000	7294.2	0.462	VV	0.059	2060.62	0.237
9	2.210		0.0000	0.000	3222.4	0.204	VB	0.034	1581.34	0.182
10	2.364		0.0000	0.000	104.4	0.007	BB	0.029	60.77	0.007
11	2.456		0.0000	0.000	169.2	0.011	BB	0.042	67.49	0.008
12	2.531		0.0000	0.000	1112.4	0.070	BB	0.051	363.22	0.042
13	2.770		0.0000	0.000	3015.8	0.191	BB	0.085	593.27	0.068
14	2.868		0.0000	0.000	3772.2	0.239	BB	0.068	930.99	0.107
15	2.999	TCX	44.1918	48.594	327890.5	20.777	BB	0.027	199901.20	22.984
16	3.288		0.0000	0.000	5590.8	0.354	BB	0.027	3452.98	0.397
17	3.383		0.0000	0.000	685.2	0.043	BB	0.025	454.84	0.052
18	3.473		0.0000	0.000	1138.7	0.072	BB	0.033	582.80	0.067
19	3.607		0.0000	0.000	364.6	0.023	BB	0.058	104.25	0.012
20	3.694		0.0000	0.000	1555.7	0.099	BB	0.040	642.87	0.074
21	3.770		0.0000	0.000	311.5	0.020	BB	0.028	184.45	0.021
22	3.864		0.0000	0.000	1477.1	0.094	BB	0.030	807.91	0.093
23	3.941		0.0000	0.000	743.4	0.047	BB	0.024	524.61	0.060
24	4.031		0.0000	0.000	750.5	0.048	BB	0.041	303.57	0.035
25	4.129		0.0000	0.000	3841.0	0.243	BB	0.039	1620.74	0.186
26	4.182		0.0000	0.000	874.5	0.055	BB	0.026	554.81	0.064
27	4.264		0.0000	0.000	1669.4	0.106	BB	0.031	885.29	0.102
28	4.360		0.0000	0.000	65959.0	4.180	BB	0.037	29682.97	3.413
29	4.527		0.0000	0.000	28140.5	1.783	BB	0.028	16892.97	1.942
30	4.598		0.0000	0.000	902.2	0.057	BB	0.018	833.42	0.096
31	4.648		0.0000	0.000	9097.3	0.576	BB	0.025	6151.75	0.707
32	4.699		0.0000	0.000	1628.3	0.103	BB	0.027	1018.81	0.117
33	4.807		0.0000	0.000	1580.4	0.100	BB	0.027	963.89	0.111
34	4.856		0.0000	0.000	36116.6	2.289	BB	0.020	29940.44	3.443
35	4.885		0.0000	0.000	4598.6	0.291	BB	0.012	6525.69	0.750
36	4.912		0.0000	0.000	3905.0	0.247	BB	0.013	4928.76	0.567
37	5.003		0.0000	0.000	11138.2	0.706	BB	0.021	8710.96	1.002
38	5.056		0.0000	0.000	75805.4	4.803	BB	0.027	47432.11	5.454
39	5.099		0.0000	0.000	18476.3	1.171	BB	0.020	15517.33	1.784
40	5.179		0.0000	0.000	2721.0	0.172	BB	0.022	2093.59	0.241
41	5.232		0.0000	0.000	28081.3	1.779	BB	0.026	18065.93	2.077
42	5.299		0.0000	0.000	45451.7	2.880	BB	0.024	31177.90	3.585
43	5.332		0.0000	0.000	2437.4	0.154	BB	0.011	3775.38	0.434
44	5.375		0.0000	0.000	89138.0	5.648	BB	0.024	60694.52	6.979
45	5.433		0.0000	0.000	4816.4	0.305	BB	0.028	2896.92	0.333
46	5.475		0.0000	0.000	9297.3	0.589	BB	0.020	7576.96	0.871
47	5.521		0.0000	0.000	35646.9	2.259	BB	0.028	21521.43	2.475
48	5.587		0.0000	0.000	2261.0	0.143	BB	0.017	2270.26	0.261
49	5.638		0.0000	0.000	76045.8	4.819	BB	0.028	45261.85	5.204
50	5.703		0.0000	0.000	5306.3	0.336	BB	0.023	3770.28	0.434
51	5.772		0.0000	0.000	73461.8	4.655	BB	0.037	33019.95	3.797
52	5.876		0.0000	0.000	49261.0	3.121	BB	0.025	32383.37	3.723
53	5.921		0.0000	0.000	1212.7	0.077	BB	0.013	1508.93	0.173
54	5.963		0.0000	0.000	5142.3	0.326	BB	0.025	3385.97	0.389

ASX22 0518

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.013		0.0000	0.000	95892.7	6.076	BB	0.036	44223.94	5.085
56	6.136		0.0000	0.000	2720.3	0.172	BB	0.030	1516.95	0.174
57	6.266		0.0000	0.000	33082.1	2.096	BB	0.036	15214.64	1.749
58	6.353		0.0000	0.000	2188.5	0.139	BB	0.023	1603.95	0.184
59	6.400		0.0000	0.000	1311.7	0.083	BB	0.025	873.51	0.100
60	6.513		0.0000	0.000	13884.0	0.880	BB	0.029	8048.73	0.925
61	6.579		0.0000	0.000	8240.4	0.522	BB	0.025	5576.77	0.641
62	6.736		0.0000	0.000	1578.0	0.100	BB	0.047	561.83	0.065
63	6.836		0.0000	0.000	7957.4	0.504	BB	0.028	4746.64	0.546
64	6.970		0.0000	0.000	91.9	0.006	BB	0.028	54.70	0.006
65	7.107		0.0000	0.000	5468.2	0.346	BB	0.074	1232.05	0.142
66	7.189		0.0000	0.000	6207.6	0.393	BB	0.079	1309.52	0.151
67	7.290		0.0000	0.000	5064.0	0.321	BB	0.108	780.24	0.090
68	7.518		0.0000	0.000	7944.2	0.503	BB	0.084	1570.63	0.181
69	7.640		0.0000	0.000	5444.8	0.345	BB	0.108	840.26	0.097
70	7.731		0.0000	0.000	2503.2	0.159	BB	0.034	1228.36	0.141
71	7.976	DCB	46.7484	51.406	228424.6	14.474	BB	0.037	104208.30	11.982
72	8.100		0.0000	0.000	3836.5	0.243	BB	0.118	543.58	0.063
73	8.375		0.0000	0.000	465.6	0.030	BB	0.041	190.53	0.022
74	8.585		0.0000	0.000	3900.8	0.247	BB	0.083	782.47	0.090
75	8.694		0.0000	0.000	992.9	0.063	BB	0.057	290.54	0.033

Total Area = 1578153.0, Total Amount = 90.94, Total Height = 869723.3, Sample Units = ug/l

ASX22 0519

Sample Name: **IC54X0924B AAIC54XAA CCAL 0930699999**
Acquired from CP07--H5088B via port 4 on 11/3/09 11:30:31pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\1P7307B.41R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.589		0.0000	0.000	5423.2	0.286	BV	0.043	2083.60	0.267
2	1.639		0.0000	0.000	8118.8	0.428	VV	0.060	2251.82	0.288
3	1.703		0.0000	0.000	4204.3	0.222	VV	0.070	1005.67	0.129
4	2.017		0.0000	0.000	1052.0	0.056	VB	0.046	383.71	0.049
5	2.116		0.0000	0.000	252.2	0.013	BB	0.028	149.53	0.019
6	2.341	TCX	41.8602	50.478	431234.9	22.758	BB	0.040	181541.00	23.252
7	2.624		0.0000	0.000	4898.3	0.259	BB	0.036	2249.75	0.288
8	2.773		0.0000	0.000	441.1	0.023	BB	0.026	286.37	0.037
9	3.010		0.0000	0.000	2099.0	0.111	BB	0.047	737.78	0.094
10	3.168		0.0000	0.000	2459.0	0.130	BB	0.050	822.76	0.105
11	3.270		0.0000	0.000	1055.1	0.056	BB	0.047	373.04	0.048
12	3.417		0.0000	0.000	5948.3	0.314	BB	0.075	1317.04	0.169
13	3.557		0.0000	0.000	803.4	0.042	BB	0.029	459.07	0.059
14	3.621		0.0000	0.000	85935.0	4.535	BB	0.055	26220.85	3.358
15	3.798		0.0000	0.000	38128.3	2.012	BB	0.044	14338.77	1.837
16	3.925		0.0000	0.000	10976.4	0.579	BB	0.041	4409.37	0.565
17	4.038		0.0000	0.000	764.6	0.040	BB	0.028	448.26	0.057
18	4.112		0.0000	0.000	130480.8	6.886	BB	0.045	48133.38	6.165
19	4.204		0.0000	0.000	5282.4	0.279	BB	0.020	4357.17	0.558
20	4.253		0.0000	0.000	37670.4	1.988	BB	0.024	26021.14	3.333
21	4.292		0.0000	0.000	31246.8	1.649	BB	0.027	19162.81	2.454
22	4.386		0.0000	0.000	1618.1	0.085	BB	0.019	1387.38	0.178
23	4.434		0.0000	0.000	29437.2	1.554	BB	0.033	14776.29	1.893
24	4.507		0.0000	0.000	61615.1	3.252	BB	0.037	27952.47	3.580
25	4.584		0.0000	0.000	125987.5	6.649	BB	0.040	52770.14	6.759
26	4.695		0.0000	0.000	64667.2	3.413	BB	0.035	30392.05	3.893
27	4.796		0.0000	0.000	83291.8	4.396	BB	0.041	33645.09	4.309
28	4.876		0.0000	0.000	48942.0	2.583	BB	0.036	22830.58	2.924
29	4.980		0.0000	0.000	48872.4	2.579	BB	0.044	18356.78	2.351
30	5.051		0.0000	0.000	26935.4	1.421	BB	0.028	15878.60	2.034
31	5.092		0.0000	0.000	1434.1	0.076	BB	0.015	1637.85	0.210
32	5.146		0.0000	0.000	132449.2	6.990	BB	0.047	46494.80	5.955
33	5.419		0.0000	0.000	42600.0	2.248	BB	0.048	14870.40	1.905
34	5.513		0.0000	0.000	676.2	0.036	BB	0.019	582.70	0.075
35	5.561		0.0000	0.000	125.2	0.007	BB	0.013	155.00	0.020
36	5.628		0.0000	0.000	34584.6	1.825	BB	0.060	9587.23	1.228
37	5.805		0.0000	0.000	2210.1	0.117	BB	0.033	1120.64	0.144
38	5.864		0.0000	0.000	124.2	0.007	BB	0.021	99.12	0.013
39	5.923		0.0000	0.000	14590.7	0.770	BB	0.042	5849.18	0.749
40	6.023		0.0000	0.000	430.3	0.023	BB	0.038	190.91	0.024
41	6.093		0.0000	0.000	363.7	0.019	BB	0.044	139.07	0.018
42	6.177		0.0000	0.000	377.6	0.020	BB	0.028	223.01	0.029
43	6.246		0.0000	0.000	1848.4	0.098	BB	0.078	395.77	0.051
44	6.533		0.0000	0.000	3937.6	0.208	BB	0.244	268.82	0.034
45	6.624		0.0000	0.000	3645.0	0.192	BB	0.038	1590.71	0.204
46	6.803	DCB	41.0676	49.522	344505.7	18.181	BB	0.040	141774.30	18.159
47	7.637		0.0000	0.000	2565.9	0.135	BB	0.121	352.91	0.045
48	8.006		0.0000	0.000	2165.8	0.114	BB	0.152	236.75	0.030
49	8.210		0.0000	0.000	3256.2	0.172	BB	0.214	253.87	0.033
50	8.490		0.0000	0.000	3130.0	0.165	BB	0.279	186.92	0.024

Total Area = 1894861.0, Total Amount = 82.928, Total Height = 780752.3, Sample Units = ug/l

ASX22 8528

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1P7307.41R



Instrument ID: CP07--H5088A Injected On: 11/3/2009 11:21:22 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\1P7307B.41R



Instrument ID: CP07--H5088B Injected On: 11/3/2009 11:21:22 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	199901	44.19175	TCX	2.341	181541	41.86023	TCX
7.976	104208	46.74836	DCB	6.803	141774	41.0676	DCB

Files:
Area File: C:\CPWIN\DATA1\1P7307.41A
Area File: C:\CPWIN\DATA1\1P7307B.41A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/4/2009 9:16:30 AM
File Reported On: 11/4/2009 at 9:17:08 AM

10X REPLOTS

ASX22 8522

C:\CPWIN\DATA1\1P7307.41R



C:\CPWIN\DATA1\1P7307B.41R



ASX22 8523

Sample Name: **AR1640924A FNAR164FN CCAL 0930899999**
Acquired from CP07--H5088A via port 3 on 11/5/09 11:50:27pm by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.40R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.928		0.0000	0.000	45483.9	0.828	BB	0.256	2963.62	0.133
2	2.076		0.0000	0.000	28937.8	0.527	BV	0.104	4618.73	0.208
3	2.174		0.0000	0.000	19205.6	0.350	VV	0.066	4826.56	0.217
4	2.523		0.0000	0.000	12825.2	0.234	VB	0.262	816.90	0.037
5	2.772		0.0000	0.000	1522.8	0.028	BB	0.059	428.78	0.019
6	2.840		0.0000	0.000	5935.5	0.108	BB	0.082	1205.24	0.054
7	2.999	TCX	64.3963	57.595	469406.2	8.548	BB	0.028	282398.60	12.707
8	3.227		0.0000	0.000	16627.4	0.303	BB	0.027	10267.59	0.462
9	3.310		0.0000	0.000	25059.8	0.456	BB	0.034	12320.98	0.554
10	3.380		0.0000	0.000	82754.9	1.507	BB	0.027	51259.74	2.307
11	3.472		0.0000	0.000	596.9	0.011	BB	0.018	566.36	0.025
12	3.513		0.0000	0.000	11833.0	0.215	BB	0.025	7777.51	0.350
13	3.693		0.0000	0.000	162637.2	2.962	BB	0.040	68344.07	3.075
14	3.771		0.0000	0.000	9495.6	0.173	BB	0.027	5923.40	0.267
15	3.824		0.0000	0.000	3064.7	0.056	BB	0.016	3110.35	0.140
16	3.862		0.0000	0.000	95422.9	1.738	BB	0.026	60775.39	2.735
17	3.937		0.0000	0.000	330.3	0.006	BB	0.014	382.44	0.017
18	3.969		0.0000	0.000	2096.5	0.038	BB	0.021	1660.79	0.075
19	4.030		0.0000	0.000	49780.5	0.907	BB	0.039	21165.01	0.952
20	4.125		0.0000	0.000	272394.2	4.960	BB	0.033	136199.40	6.129
21	4.179		0.0000	0.000	85862.6	1.564	BB	0.024	60494.62	2.722
22	4.264		0.0000	0.000	89814.0	1.636	BB	0.027	56029.85	2.521
23	4.314		0.0000	0.000	3072.8	0.056	BB	0.016	3104.86	0.140
24	4.358		0.0000	0.000	41431.6	0.754	BB	0.019	36047.91	1.622
25	4.395		0.0000	0.000	67426.7	1.228	BB	0.033	34160.90	1.537
26	4.526		0.0000	0.000	117125.9	2.133	BB	0.036	53670.46	2.415
27	4.598		0.0000	0.000	10986.3	0.200	BB	0.017	10794.18	0.486
28	4.648		0.0000	0.000	64758.8	1.179	BB	0.028	38254.73	1.721
29	4.697		0.0000	0.000	11500.9	0.209	BB	0.020	9531.09	0.429
30	4.814		0.0000	0.000	11362.6	0.207	BB	0.096	1967.80	0.089
31	4.856		0.0000	0.000	34711.9	0.632	BB	0.021	27578.76	1.241
32	4.885		0.0000	0.000	2202.3	0.040	BB	0.012	3105.58	0.140
33	4.915		0.0000	0.000	7872.3	0.143	BB	0.020	6422.11	0.289
34	5.001		0.0000	0.000	6872.0	0.125	BB	0.027	4172.56	0.188
35	5.057		0.0000	0.000	61406.9	1.118	BB	0.036	28154.67	1.267
36	5.219		0.0000	0.000	5702.1	0.104	BB	0.043	2191.37	0.099
37	5.307		0.0000	0.000	25121.2	0.457	BB	0.028	14827.72	0.667
38	5.374		0.0000	0.000	16522.8	0.301	BB	0.023	11961.26	0.538
39	5.427		0.0000	0.000	78854.1	1.436	BB	0.036	36284.77	1.633
40	5.520		0.0000	0.000	107448.0	1.957	BB	0.026	67906.37	3.056
41	5.635		0.0000	0.000	14425.5	0.263	BB	0.038	6304.45	0.284
42	5.702		0.0000	0.000	17732.4	0.323	BB	0.025	11783.19	0.530
43	5.765		0.0000	0.000	128678.2	2.343	BB	0.028	75841.77	3.413
44	5.810		0.0000	0.000	8428.1	0.153	BB	0.016	8662.89	0.390
45	5.874		0.0000	0.000	65511.1	1.193	BB	0.028	38643.48	1.739
46	5.967		0.0000	0.000	5030.7	0.092	BB	0.022	3815.70	0.172
47	6.010		0.0000	0.000	129025.5	2.350	BB	0.029	74050.80	3.332
48	6.082		0.0000	0.000	70829.6	1.290	BB	0.023	50250.65	2.261
49	6.134		0.0000	0.000	36996.2	0.674	BB	0.024	25940.07	1.167
50	6.278		0.0000	0.000	115610.2	2.105	BB	0.034	56405.03	2.538
51	6.351		0.0000	0.000	37287.2	0.679	BB	0.025	24789.45	1.115
52	6.399		0.0000	0.000	11661.4	0.212	BB	0.021	9375.27	0.422
53	6.444		0.0000	0.000	1293.4	0.024	BB	0.017	1233.55	0.056
54	6.510		0.0000	0.000	21693.1	0.395	BB	0.027	13594.80	0.612

ASX22 0524

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.577		0.0000	0.000	198725.4	3.619	BB	0.029	114246.10	5.141
56	6.831		0.0000	0.000	92948.0	1.693	BB	0.029	54121.50	2.435
57	6.890		0.0000	0.000	51592.2	0.940	BB	0.036	23815.82	1.072
58	7.055		0.0000	0.000	4544.4	0.083	BB	0.039	1947.50	0.088
59	7.181		0.0000	0.000	33812.0	0.616	BB	0.039	14470.29	0.651
60	7.414		0.0000	0.000	56061.7	1.021	BB	0.031	30406.43	1.368
61	7.486		0.0000	0.000	2204.4	0.040	BB	0.026	1435.43	0.065
62	7.747		0.0000	0.000	1988872.0	36.219	BB	0.114	290440.10	13.069
63	7.973	DCB	47.4118	42.405	223885.9	4.077	BB	0.035	105551.90	4.749
64	8.172		0.0000	0.000	4379.2	0.080	BB	0.092	791.65	0.036
65	8.380		0.0000	0.000	893.0	0.016	BB	0.069	216.22	0.010
66	8.611		0.0000	0.000	2075.8	0.038	BB	0.121	285.77	0.013
67	8.783		0.0000	0.000	256.7	0.005	BB	0.047	91.18	0.004
68	9.013		0.0000	0.000	1354.8	0.025	BB	0.108	209.08	0.009

Total Area = 5491275.0, Total Amount = 111.808, Total Height = 2222393.0, Sample Units = ug/l

ASX22 08525

Sample Name: **AR1640924A FNAR164FN CCAL 0930899999**
Acquired from CP07--H5088B via port 4 on 11/5/09 11:50:27pm by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.40R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.582		0.0000	0.000	12479.9	0.277	BV	0.067	3098.60	0.154
2	1.707		0.0000	0.000	2190.5	0.049	VV	0.044	823.79	0.041
3	1.828		0.0000	0.000	515.4	0.011	VV	0.111	77.65	0.004
4	1.934		0.0000	0.000	508.7	0.011	VV	0.128	66.10	0.003
5	1.968		0.0000	0.000	2278.8	0.051	VB	0.059	643.64	0.032
6	2.200		0.0000	0.000	2937.3	0.065	BB	0.077	632.18	0.031
7	2.284		0.0000	0.000	1569.9	0.035	BB	0.037	707.48	0.035
8	2.341	TCX	60.3855	54.328	606495.6	13.475	BB	0.039	261881.90	12.996
9	2.563		0.0000	0.000	14392.9	0.320	BB	0.033	7380.88	0.366
10	2.642		0.0000	0.000	19429.7	0.432	BB	0.037	8785.55	0.436
11	2.714		0.0000	0.000	98632.6	2.191	BB	0.041	39783.74	1.974
12	2.859		0.0000	0.000	15511.7	0.345	BB	0.036	7209.20	0.358
13	3.006		0.0000	0.000	238572.6	5.300	BB	0.047	84225.78	4.180
14	3.169		0.0000	0.000	140078.0	3.112	BB	0.039	59669.85	2.961
15	3.310		0.0000	0.000	45535.9	1.012	BB	0.044	17072.86	0.847
16	3.402		0.0000	0.000	186896.1	4.152	BB	0.032	97414.91	4.834
17	3.447		0.0000	0.000	32114.9	0.714	BB	0.028	18850.35	0.935
18	3.555		0.0000	0.000	72781.8	1.617	BB	0.030	40067.36	1.988
19	3.616		0.0000	0.000	14823.1	0.329	BB	0.027	9198.95	0.456
20	3.642		0.0000	0.000	57977.8	1.288	BB	0.037	25982.93	1.289
21	3.795		0.0000	0.000	157111.8	3.491	BB	0.051	51255.54	2.544
22	3.886		0.0000	0.000	1066.0	0.024	BB	0.012	1422.98	0.071
23	3.921		0.0000	0.000	40746.3	0.905	BB	0.028	24165.96	1.199
24	3.970		0.0000	0.000	11164.8	0.248	BB	0.025	7487.37	0.372
25	4.107		0.0000	0.000	147277.7	3.272	BB	0.049	50066.81	2.485
26	4.198		0.0000	0.000	1548.1	0.034	BB	0.019	1371.21	0.068
27	4.250		0.0000	0.000	90287.3	2.006	BB	0.043	34613.61	1.718
28	4.428		0.0000	0.000	1825.9	0.041	BB	0.031	969.33	0.048
29	4.503		0.0000	0.000	3563.4	0.079	BB	0.025	2415.33	0.120
30	4.540		0.0000	0.000	14376.3	0.319	BB	0.022	10855.64	0.539
31	4.596		0.0000	0.000	153486.7	3.410	BB	0.050	51625.96	2.562
32	4.691		0.0000	0.000	189432.6	4.209	BB	0.033	95209.17	4.725
33	4.819		0.0000	0.000	28182.4	0.626	BB	0.035	13397.55	0.665
34	4.871		0.0000	0.000	212556.5	4.722	BB	0.034	103833.10	5.153
35	4.984		0.0000	0.000	241983.6	5.376	BB	0.053	75884.02	3.766
36	5.098		0.0000	0.000	17660.2	0.392	BB	0.021	13879.45	0.689
37	5.141		0.0000	0.000	69435.0	1.543	BB	0.023	51326.04	2.547
38	5.174		0.0000	0.000	46756.7	1.039	BB	0.021	37375.44	1.855
39	5.218		0.0000	0.000	26102.8	0.580	BB	0.023	18644.62	0.925
40	5.336		0.0000	0.000	15283.0	0.340	BB	0.025	10155.38	0.504
41	5.389		0.0000	0.000	154488.8	3.432	BB	0.034	75410.80	3.742
42	5.472		0.0000	0.000	40286.0	0.895	BB	0.025	27018.15	1.341
43	5.508		0.0000	0.000	7437.8	0.165	BB	0.018	6905.75	0.343
44	5.560		0.0000	0.000	8626.8	0.192	BB	0.021	6828.30	0.339
45	5.625		0.0000	0.000	365841.2	8.128	BB	0.038	159721.10	7.926
46	5.867		0.0000	0.000	46346.2	1.030	BB	0.025	30896.56	1.533
47	5.921		0.0000	0.000	223226.2	4.959	BB	0.037	101180.70	5.021
48	6.087		0.0000	0.000	4390.0	0.098	BB	0.033	2219.66	0.110
49	6.172		0.0000	0.000	6486.7	0.144	BB	0.030	3557.88	0.177
50	6.237		0.0000	0.000	43261.8	0.961	BB	0.035	20388.90	1.012
51	6.382		0.0000	0.000	125117.1	2.780	BB	0.043	48309.03	2.397
52	6.620		0.0000	0.000	36478.7	0.810	BB	0.038	15863.44	0.787
53	6.798	DCB	50.7637	45.672	388158.7	8.624	BB	0.037	175247.20	8.697
54	7.467		0.0000	0.000	166.9	0.004	BB	0.073	37.85	0.002

ASX22 8526

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	7.748		0.0000	0.000	18.9	0.000	BB	0.009	36.39	0.002
56	7.850		0.0000	0.000	7741.9	0.172	BB	0.143	903.00	0.045
57	9.003		0.0000	0.000	7387.8	0.164	BB	0.115	1069.66	0.053

Total Area = 4501033.0, Total Amount = 111.149, Total Height = 2015123.0, Sample Units = ug/l

ASX22 6527

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3P7307.40R



Instrument ID: CP07--H5088A Injected On: 11/5/2009 11:41:18 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.40R



Instrument ID: CP07--H5088B Injected On: 11/5/2009 11:41:18 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	282399	64.39635	TCX	2.341	261882	60.38545	TCX
7.973	105552	47.41177	DCB	6.798	175247	50.76368	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.40A
Area File: C:\CPWIN\DATA1\3P7307B.40A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\IP7307.CAL
Calibration File B: C:\CPWIN\DATA1\IP7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/5/2009 11:50:36 PM
File Reported On: 11/5/2009 at 11:50:48 PM

10X REPLOTS

ASX22 0529

C:\CPWIN\DATA1\3P7307.40R



C:\CPWIN\DATA1\3P7307B.40R



ASX22 0538

Sample Name: **IBLKX0924B FHPIBLKFH PIBLK0930899999**
Acquired from CP07--H5088A via port 3 on 11/6/09 12:03:29am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.41R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.930		0.0000	0.000	22876.6	5.716	BB	0.274	1393.99	0.803
2	2.041		0.0000	0.000	7344.8	1.835	BV	0.076	1606.46	0.926
3	2.077		0.0000	0.000	5897.1	1.474	VV	0.049	1986.42	1.145
4	2.175		0.0000	0.000	9883.7	2.470	VV	0.071	2318.21	1.336
5	2.554		0.0000	0.000	3232.3	0.808	VB	0.368	146.50	0.084
6	2.778		0.0000	0.000	1197.2	0.299	BB	0.070	284.56	0.164
7	2.825		0.0000	0.000	7989.1	1.996	BB	0.098	1353.90	0.780
8	3.000	TCX	227.0609	59.025	190439.4	47.586	BB	0.028	112173.10	64.653
9	3.288		0.0000	0.000	3547.8	0.887	BB	0.031	1898.74	1.094
10	3.384		0.0000	0.000	186.4	0.047	BB	0.029	108.72	0.063
11	3.473		0.0000	0.000	450.9	0.113	BB	0.028	266.33	0.154
12	3.697		0.0000	0.000	458.2	0.115	BB	0.054	141.81	0.082
13	3.773		0.0000	0.000	264.1	0.066	BB	0.037	118.37	0.068
14	3.859		0.0000	0.000	846.5	0.212	BB	0.032	446.39	0.257
15	3.940		0.0000	0.000	516.1	0.129	BB	0.030	284.92	0.164
16	4.101		0.0000	0.000	3463.9	0.866	BB	0.060	956.78	0.551
17	4.204		0.0000	0.000	566.0	0.141	BB	0.052	180.08	0.104
18	4.264		0.0000	0.000	416.1	0.104	BB	0.031	222.15	0.128
19	4.360		0.0000	0.000	54.6	0.014	BB	0.018	51.57	0.030
20	4.459		0.0000	0.000	907.5	0.227	BB	0.067	227.27	0.131
21	4.649		0.0000	0.000	927.6	0.232	BB	0.085	182.30	0.105
22	5.243		0.0000	0.000	282.4	0.071	BB	0.059	79.69	0.046
23	5.380		0.0000	0.000	1856.5	0.464	BB	0.057	539.80	0.311
24	5.477		0.0000	0.000	154.9	0.039	BB	0.024	105.38	0.061
25	5.520		0.0000	0.000	141.2	0.035	BB	0.023	104.53	0.060
26	5.657		0.0000	0.000	216.8	0.054	BB	0.045	80.03	0.046
27	5.773		0.0000	0.000	607.7	0.152	BB	0.052	195.87	0.113
28	5.897		0.0000	0.000	322.3	0.081	BB	0.053	101.92	0.059
29	5.961		0.0000	0.000	80.0	0.020	BB	0.027	48.50	0.028
30	6.009		0.0000	0.000	227.6	0.057	BB	0.026	147.53	0.085
31	6.426		0.0000	0.000	17305.1	4.324	BB	0.180	1604.17	0.925
32	6.833		0.0000	0.000	720.5	0.180	BB	0.063	191.17	0.110
33	7.174		0.0000	0.000	22061.9	5.513	BB	0.241	1528.00	0.881
34	7.588		0.0000	0.000	2384.0	0.596	BB	0.169	235.31	0.136
35	7.733		0.0000	0.000	2172.7	0.543	BB	0.054	675.32	0.389
36	7.976	DCB	157.6228	40.975	90009.4	22.491	BB	0.036	41453.66	23.892
37	8.539		0.0000	0.000	194.1	0.049	BB	0.052	62.08	0.036

Total Area = 400202.9, Total Amount = 384.684, Total Height = 173501.5, Sample Units = ug/l

ASX22 6531

Sample Name: **IBLKX0924B FHPIBLKFH PIBLK0930899999**
Acquired from CP07--H5088B via port 4 on 11/6/09 12:03:29am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.41R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.578		0.0000	0.000	7188.4	1.719	BV	0.084	1433.45	0.822
2	1.934		0.0000	0.000	2347.2	0.561	VB	0.978	40.01	0.023
3	2.123		0.0000	0.000	71.5	0.017	BB	0.022	53.01	0.030
4	2.193		0.0000	0.000	1449.9	0.347	BB	0.052	465.11	0.267
5	2.289		0.0000	0.000	1441.7	0.345	BB	0.050	479.47	0.275
6	2.342	TCX	223.0951	52.571	217550.3	52.020	BB	0.037	96752.73	55.466
7	2.621		0.0000	0.000	2796.9	0.669	BB	0.037	1260.05	0.722
8	2.774		0.0000	0.000	371.2	0.089	BB	0.033	189.85	0.109
9	2.900		0.0000	0.000	74.1	0.018	BB	0.025	49.67	0.028
10	2.963		0.0000	0.000	45.7	0.011	BB	0.018	42.37	0.024
11	3.013		0.0000	0.000	223.0	0.053	BB	0.036	104.37	0.060
12	3.103		0.0000	0.000	60.2	0.014	BB	0.022	46.50	0.027
13	3.160		0.0000	0.000	1737.6	0.415	BB	0.055	529.40	0.303
14	3.269		0.0000	0.000	376.7	0.090	BB	0.034	184.82	0.106
15	3.349		0.0000	0.000	2183.1	0.522	BB	0.084	433.07	0.248
16	3.553		0.0000	0.000	968.4	0.232	BB	0.079	204.49	0.117
17	4.310		0.0000	0.000	1205.6	0.288	BB	0.066	304.14	0.174
18	4.824		0.0000	0.000	2063.5	0.493	BB	0.076	450.74	0.258
19	5.627		0.0000	0.000	1405.2	0.336	BB	0.057	410.47	0.235
20	5.923		0.0000	0.000	4655.1	1.113	BB	0.373	208.17	0.119
21	6.401		0.0000	0.000	139.6	0.033	BB	0.064	36.47	0.021
22	6.451		0.0000	0.000	31.7	0.008	BB	0.016	32.85	0.019
23	6.557		0.0000	0.000	240.2	0.057	BB	0.077	52.08	0.030
24	6.621		0.0000	0.000	2048.2	0.490	BB	0.043	800.78	0.459
25	6.800	DCB	201.2772	47.429	162075.7	38.755	BB	0.039	69485.26	39.834
26	7.240		0.0000	0.000	119.8	0.029	BB	0.033	60.58	0.035
27	7.457		0.0000	0.000	107.2	0.026	BB	0.068	26.17	0.015
28	7.640		0.0000	0.000	3125.6	0.747	BB	0.237	219.95	0.126
29	7.980		0.0000	0.000	2101.8	0.503	BB	0.433	80.97	0.046

Total Area = 418205.1, Total Amount = 424.372, Total Height = 174437.0, Sample Units = ug/l

ASX22 0532

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3P7307.41R



Instrument ID: CP07--H5088A Injected On: 11/5/2009 11:54:20 PM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.41R



Instrument ID: CP07--H5088B Injected On: 11/5/2009 11:54:20 PM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1000
Analyst: 2280

Dilution Factor: 10

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.	112173	.22706	TCX	2.342	96753	.2231	TCX
7.976	41454	.15762	DCB	6.8	69485	.20128	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.41A
Area File: C:\CPWIN\DATA1\3P7307B.41A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 12:03:36 AM
File Reported On: 11/6/2009 at 12:03:49 AM

10X REPLOTS

ASX22 0534

C:\CPWIN\DATA1\3P7307.41R



C:\CPWIN\DATA1\3P7307B.41R



ASX22 0535

Sample Name: **AR1630924A OUAR163OU CCAL 0930899999**
Acquired from CP07--H5088A via port 3 on 11/6/09 02:26:47am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.52R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.927		0.0000	0.000	31947.3	1.883	BB	0.259	2056.27	0.224
2	2.074		0.0000	0.000	19949.2	1.176	BV	0.107	3121.21	0.340
3	2.172		0.0000	0.000	10503.1	0.619	VV	0.054	3216.94	0.351
4	2.531		0.0000	0.000	8214.3	0.484	VB	0.229	596.89	0.065
5	2.764		0.0000	0.000	1931.0	0.114	BB	0.048	663.70	0.072
6	2.827		0.0000	0.000	4842.9	0.285	BB	0.091	882.93	0.096
7	2.998	TCX	40.9124	55.018	304970.8	17.972	BB	0.027	186511.40	20.334
8	3.127		0.0000	0.000	248.6	0.015	BB	0.019	219.72	0.024
9	3.226		0.0000	0.000	5869.4	0.346	BB	0.027	3664.54	0.400
10	3.308		0.0000	0.000	10371.6	0.611	BB	0.038	4589.74	0.500
11	3.380		0.0000	0.000	31401.8	1.851	BB	0.027	19629.80	2.140
12	3.471		0.0000	0.000	334.8	0.020	BB	0.019	300.93	0.033
13	3.513		0.0000	0.000	3807.9	0.224	BB	0.025	2543.26	0.277
14	3.692		0.0000	0.000	62516.9	3.684	BB	0.039	26504.64	2.890
15	3.770		0.0000	0.000	3344.4	0.197	BB	0.027	2096.32	0.229
16	3.823		0.0000	0.000	992.4	0.058	BB	0.017	1001.67	0.109
17	3.861		0.0000	0.000	36865.4	2.172	BB	0.026	23499.34	2.562
18	3.937		0.0000	0.000	243.1	0.014	BB	0.016	254.18	0.028
19	3.967		0.0000	0.000	563.9	0.033	BB	0.020	474.66	0.052
20	4.028		0.0000	0.000	17066.3	1.006	BB	0.038	7495.51	0.817
21	4.124		0.0000	0.000	106625.8	6.283	BB	0.033	53354.37	5.817
22	4.179		0.0000	0.000	32576.0	1.920	BB	0.024	23029.22	2.511
23	4.264		0.0000	0.000	32598.4	1.921	BB	0.026	20692.70	2.256
24	4.314		0.0000	0.000	978.6	0.058	BB	0.016	996.14	0.109
25	4.358		0.0000	0.000	15594.5	0.919	BB	0.019	13367.06	1.457
26	4.394		0.0000	0.000	24455.2	1.441	BB	0.032	12607.81	1.375
27	4.526		0.0000	0.000	43400.8	2.558	BB	0.037	19815.99	2.160
28	4.598		0.0000	0.000	3982.6	0.235	BB	0.017	3909.38	0.426
29	4.647		0.0000	0.000	24496.7	1.444	BB	0.028	14656.44	1.598
30	4.696		0.0000	0.000	4786.8	0.282	BB	0.020	3927.51	0.428
31	4.761		0.0000	0.000	626.9	0.037	BB	0.020	518.63	0.057
32	4.814		0.0000	0.000	561.0	0.033	BB	0.021	451.45	0.049
33	4.855		0.0000	0.000	27154.9	1.600	BB	0.028	15956.86	1.740
34	4.912		0.0000	0.000	2544.0	0.150	BB	0.016	2725.87	0.297
35	5.001		0.0000	0.000	1966.6	0.116	BB	0.022	1484.90	0.162
36	5.057		0.0000	0.000	26746.4	1.576	BB	0.031	14521.70	1.583
37	5.232		0.0000	0.000	924.1	0.054	BB	0.030	510.18	0.056
38	5.308		0.0000	0.000	13114.0	0.773	BB	0.029	7412.44	0.808
39	5.373		0.0000	0.000	8500.3	0.501	BB	0.024	5933.33	0.647
40	5.427		0.0000	0.000	36917.1	2.176	BB	0.036	16909.63	1.844
41	5.520		0.0000	0.000	49984.5	2.946	BB	0.027	30865.37	3.365
42	5.635		0.0000	0.000	5541.2	0.327	BB	0.037	2493.43	0.272
43	5.701		0.0000	0.000	7130.1	0.420	BB	0.025	4698.64	0.512
44	5.766		0.0000	0.000	56640.9	3.338	BB	0.028	33331.61	3.634
45	5.810		0.0000	0.000	3391.9	0.200	BB	0.016	3574.81	0.390
46	5.874		0.0000	0.000	27343.2	1.611	BB	0.029	15937.12	1.738
47	5.966		0.0000	0.000	2133.5	0.126	BB	0.023	1563.76	0.170
48	6.010		0.0000	0.000	55912.0	3.295	BB	0.029	32387.08	3.531
49	6.082		0.0000	0.000	32153.1	1.895	BB	0.024	22310.85	2.432
50	6.134		0.0000	0.000	16355.4	0.964	BB	0.024	11379.57	1.241
51	6.279		0.0000	0.000	51972.0	3.063	BB	0.035	24847.27	2.709
52	6.351		0.0000	0.000	16519.8	0.974	BB	0.025	10922.07	1.191
53	6.399		0.0000	0.000	5356.2	0.316	BB	0.021	4248.21	0.463
54	6.509		0.0000	0.000	9218.0	0.543	BB	0.026	5903.72	0.644

ASX22 8536

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.577		0.0000	0.000	90815.5	5.352	BB	0.029	51836.91	5.651
56	6.831		0.0000	0.000	42731.1	2.518	BB	0.029	24306.55	2.650
57	6.890		0.0000	0.000	23468.8	1.383	BB	0.037	10596.14	1.155
58	7.056		0.0000	0.000	1288.5	0.076	BB	0.031	684.53	0.075
59	7.180		0.0000	0.000	19150.5	1.129	BB	0.049	6450.46	0.703
60	7.415		0.0000	0.000	23480.0	1.384	BB	0.031	12781.76	1.394
61	7.486		0.0000	0.000	930.4	0.055	BB	0.027	573.09	0.062
62	7.731		0.0000	0.000	15979.2	0.942	BB	0.049	5420.06	0.591
63	7.974	DCB	33.4490	44.982	170359.7	10.039	BB	0.037	77273.85	8.425
64	8.144		0.0000	0.000	1293.5	0.076	BB	0.152	141.89	0.015
65	8.370		0.0000	0.000	253.8	0.015	BB	0.045	93.84	0.010
66	8.525		0.0000	0.000	894.1	0.053	BB	0.080	187.20	0.020
67	8.630		0.0000	0.000	2102.7	0.124	BB	0.111	315.19	0.034

Total Area = 1696935.0, Total Amount = 74.361, Total Height = 917230.3, Sample Units = ug/l

ASX22 8537

Sample Name: **AR1630924A OUAR163OU CCAL 0930899999**
Acquired from CP07--H5088B via port 4 on 11/6/09 02:26:47am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.52R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.578		0.0000	0.000	8877.7	0.409	BV	0.072	2050.61	0.213
2	1.706		0.0000	0.000	2202.9	0.101	VV	0.053	693.37	0.072
3	1.841		0.0000	0.000	403.3	0.019	VV	0.065	102.98	0.011
4	1.918		0.0000	0.000	377.8	0.017	VV	0.101	62.47	0.006
5	1.988		0.0000	0.000	600.5	0.028	VV	0.057	176.47	0.018
6	2.017		0.0000	0.000	388.8	0.018	VB	0.038	168.80	0.018
7	2.170		0.0000	0.000	3535.7	0.163	BB	0.084	697.98	0.072
8	2.283		0.0000	0.000	1506.9	0.069	BB	0.044	572.68	0.059
9	2.339	TCX	39.6785	52.179	389622.6	17.950	BB	0.038	172079.00	17.840
10	2.560		0.0000	0.000	5294.1	0.244	BB	0.032	2743.11	0.284
11	2.639		0.0000	0.000	9268.7	0.427	BB	0.040	3820.64	0.396
12	2.713		0.0000	0.000	40706.5	1.875	BB	0.042	16305.56	1.690
13	2.857		0.0000	0.000	5943.8	0.274	BB	0.036	2719.76	0.282
14	3.005		0.0000	0.000	100355.3	4.623	BB	0.047	35402.91	3.670
15	3.168		0.0000	0.000	57590.9	2.653	BB	0.039	24535.81	2.544
16	3.308		0.0000	0.000	16999.9	0.783	BB	0.046	6189.92	0.642
17	3.401		0.0000	0.000	77603.2	3.575	BB	0.033	39409.39	4.086
18	3.447		0.0000	0.000	12859.9	0.592	BB	0.028	7638.05	0.792
19	3.554		0.0000	0.000	29555.4	1.362	BB	0.030	16179.68	1.677
20	3.616		0.0000	0.000	6452.6	0.297	BB	0.030	3578.89	0.371
21	3.641		0.0000	0.000	23765.9	1.095	BB	0.039	10081.88	1.045
22	3.795		0.0000	0.000	64027.9	2.950	BB	0.051	21057.50	2.183
23	3.884		0.0000	0.000	374.7	0.017	BB	0.011	549.15	0.057
24	3.920		0.0000	0.000	16661.6	0.768	BB	0.029	9670.84	1.003
25	3.969		0.0000	0.000	4203.5	0.194	BB	0.024	2891.17	0.300
26	4.107		0.0000	0.000	59552.4	2.744	BB	0.049	20245.92	2.099
27	4.195		0.0000	0.000	629.1	0.029	BB	0.019	558.75	0.058
28	4.250		0.0000	0.000	36541.4	1.683	BB	0.043	14025.32	1.454
29	4.433		0.0000	0.000	580.1	0.027	BB	0.029	328.84	0.034
30	4.502		0.0000	0.000	1431.3	0.066	BB	0.026	927.86	0.096
31	4.540		0.0000	0.000	5845.6	0.269	BB	0.023	4243.76	0.440
32	4.595		0.0000	0.000	61459.7	2.831	BB	0.049	20918.53	2.169
33	4.691		0.0000	0.000	81832.6	3.770	BB	0.034	39912.44	4.138
34	4.819		0.0000	0.000	11375.6	0.524	BB	0.035	5407.00	0.561
35	4.871		0.0000	0.000	93633.1	4.314	BB	0.035	44688.48	4.633
36	4.983		0.0000	0.000	102085.8	4.703	BB	0.054	31790.50	3.296
37	5.098		0.0000	0.000	6812.4	0.314	BB	0.021	5286.18	0.548
38	5.140		0.0000	0.000	32390.1	1.492	BB	0.024	22488.46	2.331
39	5.174		0.0000	0.000	21301.4	0.981	BB	0.021	16660.14	1.727
40	5.217		0.0000	0.000	10796.8	0.497	BB	0.023	7736.03	0.802
41	5.334		0.0000	0.000	6568.0	0.303	BB	0.025	4419.85	0.458
42	5.388		0.0000	0.000	70564.3	3.251	BB	0.035	33669.83	3.491
43	5.472		0.0000	0.000	18539.9	0.854	BB	0.025	12504.36	1.296
44	5.508		0.0000	0.000	2915.5	0.134	BB	0.021	2306.03	0.239
45	5.559		0.0000	0.000	3537.2	0.163	BB	0.021	2815.45	0.292
46	5.625		0.0000	0.000	164654.0	7.586	BB	0.039	70816.16	7.342
47	5.793		0.0000	0.000	316.6	0.015	BB	0.019	271.24	0.028
48	5.866		0.0000	0.000	20684.4	0.953	BB	0.026	13306.43	1.380
49	5.921		0.0000	0.000	98477.1	4.537	BB	0.037	43945.10	4.556
50	6.084		0.0000	0.000	1843.6	0.085	BB	0.035	890.25	0.092
51	6.172		0.0000	0.000	2767.8	0.128	BB	0.031	1481.02	0.154
52	6.236		0.0000	0.000	18624.1	0.858	BB	0.036	8627.86	0.894
53	6.381		0.0000	0.000	57334.7	2.641	BB	0.046	20860.22	2.163
54	6.620		0.0000	0.000	16982.5	0.782	BB	0.039	7307.48	0.758

ASX22 0538

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.798	DCB	36.3638	47.821	276687.1	12.747	BB	0.037	125535.70	13.015
56	7.043		0.0000	0.000	217.1	0.010	BB	0.027	135.93	0.014
57	8.170		0.0000	0.000	236.0	0.011	BB	0.044	88.78	0.009
58	8.283		0.0000	0.000	2337.3	0.108	BB	0.075	516.99	0.054
59	8.958		0.0000	0.000	1869.3	0.086	BB	0.067	467.09	0.048

Total Area = 2170603.0, Total Amount = 76.042, Total Height = 964562.8, Sample Units = ug/l

ASX22 8539

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3P7307.52R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 2:17:38 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.52R



Instrument ID: CP07--H5088B Injected On: 11/6/2009 2:17:38 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.998	186511	40.91243	TCX	2.339	172079	39.67846	TCX
7.974	77274	33.44904	DCB	6.798	125536	36.3638	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.52A
Area File: C:\CPWIN\DATA1\3P7307B.52A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 2:26:54 AM
File Reported On: 11/6/2009 at 2:27:07 AM

10X REPLOTS

ASX22 8541

C:\CPWIN\DATA1\3P7307.52R



C:\CPWIN\DATA1\3P7307B.52R



ASX22 0542

Sample Name: **IBLKX0924B FHPIBLKFH PIBLK0930899999**
Acquired from CP07–H5088A via port 3 on 11/6/09 02:39:47am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.53R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.926		0.0000	0.000	54127.6	10.850	BB	0.245	3677.44	2.139
2	2.075		0.0000	0.000	28589.4	5.731	BV	0.123	3881.59	2.258
3	2.175		0.0000	0.000	15164.8	3.040	VV	0.071	3566.57	2.075
4	2.596		0.0000	0.000	23219.0	4.654	VB	1.735	223.05	0.130
5	2.781		0.0000	0.000	4502.2	0.902	BB	0.093	808.19	0.470
6	2.864		0.0000	0.000	9114.3	1.827	BB	0.048	3190.22	1.856
7	2.999	TCX	204.9429	56.225	187121.0	37.508	BB	0.030	103142.10	60.000
8	3.213		0.0000	0.000	262.6	0.053	BB	0.058	75.14	0.044
9	3.288		0.0000	0.000	3253.7	0.652	BB	0.032	1719.30	1.000
10	3.381		0.0000	0.000	137.1	0.027	BB	0.025	90.42	0.053
11	3.472		0.0000	0.000	430.5	0.086	BB	0.029	251.08	0.146
12	3.689		0.0000	0.000	346.1	0.069	BB	0.052	111.75	0.065
13	3.770		0.0000	0.000	523.2	0.105	BB	0.052	166.14	0.097
14	3.864		0.0000	0.000	250.2	0.050	BB	0.038	109.48	0.064
15	3.940		0.0000	0.000	479.7	0.096	BB	0.029	272.82	0.159
16	4.027		0.0000	0.000	129.4	0.026	BB	0.033	64.62	0.038
17	4.101		0.0000	0.000	1129.9	0.226	BB	0.031	604.21	0.351
18	4.206		0.0000	0.000	676.3	0.136	BB	0.053	213.06	0.124
19	4.264		0.0000	0.000	129.6	0.026	BB	0.024	89.62	0.052
20	4.337		0.0000	0.000	71.2	0.014	BB	0.026	46.45	0.027
21	4.432		0.0000	0.000	3089.1	0.619	BB	0.154	334.43	0.195
22	4.810		0.0000	0.000	419.7	0.084	BB	0.060	117.13	0.068
23	4.919		0.0000	0.000	1795.3	0.360	BB	0.087	344.07	0.200
24	5.102		0.0000	0.000	993.6	0.199	BB	0.037	444.08	0.258
25	5.342		0.0000	0.000	847.7	0.170	BB	0.212	66.66	0.039
26	5.380		0.0000	0.000	2114.4	0.424	BB	0.074	478.58	0.278
27	5.770		0.0000	0.000	353.4	0.071	BB	0.066	88.79	0.052
28	5.967		0.0000	0.000	105.4	0.021	BB	0.033	53.70	0.031
29	6.024		0.0000	0.000	251.3	0.050	BB	0.044	96.20	0.056
30	6.427		0.0000	0.000	33857.4	6.787	BB	0.226	2500.81	1.455
31	6.734		0.0000	0.000	306.2	0.061	BB	0.047	109.38	0.064
32	6.897		0.0000	0.000	973.1	0.195	BB	0.089	182.89	0.106
33	6.987		0.0000	0.000	166.4	0.033	BB	0.044	63.12	0.037
34	7.153		0.0000	0.000	17824.6	3.573	BB	0.233	1272.52	0.740
35	7.584		0.0000	0.000	4458.6	0.894	BB	0.200	371.23	0.216
36	7.732		0.0000	0.000	2070.2	0.415	BB	0.052	664.42	0.387
37	7.975	DCB	159.5650	43.775	93942.1	18.830	BB	0.037	41847.02	24.343
38	8.610		0.0000	0.000	5659.2	1.134	BB	0.167	565.10	0.329

Total Area = 498885.5, Total Amount = 364.508, Total Height = 171903.4, Sample Units = ug/l

ASX22 8543

Sample Name: **IBLKX0924B FHPIBLKFH PIBLK0930899999**
Acquired from CP07--H5088B via port 4 on 11/6/09 02:39:47am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.53R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.580		0.0000	0.000	6723.8	1.589	BV	0.097	1150.82	0.682
2	1.938		0.0000	0.000	1346.9	0.318	VV	0.068	332.43	0.197
3	2.014		0.0000	0.000	2261.3	0.535	VB	0.073	519.57	0.308
4	2.197		0.0000	0.000	5483.7	1.296	BB	0.073	1259.78	0.747
5	2.340	TCX	210.2704	51.538	219091.3	51.788	BB	0.040	91190.88	54.062
6	2.622		0.0000	0.000	2580.8	0.610	BB	0.037	1152.87	0.683
7	2.777		0.0000	0.000	339.4	0.080	BB	0.033	172.48	0.102
8	3.107		0.0000	0.000	334.4	0.079	BB	0.157	35.47	0.021
9	3.267		0.0000	0.000	348.5	0.082	BB	0.032	182.52	0.108
10	3.349		0.0000	0.000	3500.4	0.827	BB	0.112	522.32	0.310
11	3.890		0.0000	0.000	269.3	0.064	BB	0.040	113.54	0.067
12	4.114		0.0000	0.000	746.0	0.176	BB	0.080	155.57	0.092
13	4.310		0.0000	0.000	545.7	0.129	BB	0.047	194.46	0.115
14	4.687		0.0000	0.000	739.0	0.175	BB	0.061	203.38	0.121
15	4.827		0.0000	0.000	4266.0	1.008	BB	0.116	614.12	0.364
16	5.143		0.0000	0.000	149.7	0.035	BB	0.030	83.67	0.050
17	5.226		0.0000	0.000	2743.0	0.648	BB	0.220	208.00	0.123
18	5.620		0.0000	0.000	345.7	0.082	BB	0.033	172.77	0.102
19	5.794		0.0000	0.000	10205.6	2.412	BB	0.542	314.07	0.186
20	6.438		0.0000	0.000	727.9	0.172	BB	0.437	27.78	0.016
21	6.474		0.0000	0.000	47.4	0.011	BB	0.016	49.43	0.029
22	6.543		0.0000	0.000	130.8	0.031	BB	0.032	68.26	0.040
23	6.621		0.0000	0.000	2397.1	0.567	BB	0.050	800.42	0.475
24	6.799	DCB	197.7223	48.462	153365.9	36.252	BB	0.037	68258.02	40.466
25	6.992		0.0000	0.000	1153.5	0.273	BB	0.076	252.37	0.150
26	7.350		0.0000	0.000	441.1	0.104	BB	0.054	135.65	0.080
27	7.610		0.0000	0.000	649.7	0.154	BB	0.105	103.11	0.061
28	8.980		0.0000	0.000	2118.9	0.501	BB	0.087	404.22	0.240

Total Area = 423052.9, Total Amount = 407.993, Total Height = 168678.0, Sample Units = ug/l

ASX22 0544

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3P7307.53R

C:\CPWIN\DATA1\3P7307.53R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 2:30:38 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.53R

C:\CPWIN\DATA1\3P7307B.53R



Instrument ID: CP07--H5088B Injected On: 11/6/2009 2:30:38 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1000
Analyst: 2280

Dilution Factor: 10

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	103142	.20494	TCX	2.34	91191	.21027	TCX
7.975	41847	.15957	DCB	6.799	68258	.19772	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.53A
Area File: C:\CPWIN\DATA1\3P7307B.53A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 2:39:54 AM
File Reported On: 11/6/2009 at 2:40:06 AM

10X REPLOTS

ASX22 546

C:\CPWIN\DATA1\3P7307.53R



C:\CPWIN\DATA1\3P7307B.53R



ASX22 0547

Sample Name: **AR1640924A FOAR164FO CCAL 0930899999**
Acquired from CP07--H5088A via port 3 on 11/6/09 04:50:15am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.63R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.611		0.0000	0.000	3753.9	0.085	BV	0.056	1109.71	0.045
2	1.669		0.0000	0.000	4948.9	0.112	VV	0.066	1246.99	0.050
3	1.795		0.0000	0.000	8396.8	0.191	VV	0.083	1693.47	0.068
4	1.866		0.0000	0.000	7570.8	0.172	VV	0.067	1883.68	0.076
5	1.927		0.0000	0.000	7787.2	0.177	VB	0.059	2185.21	0.088
6	2.074		0.0000	0.000	25766.4	0.585	BV	0.099	4352.89	0.175
7	2.172		0.0000	0.000	11768.8	0.267	VV	0.049	4009.22	0.162
8	2.207		0.0000	0.000	4244.8	0.096	VV	0.030	2395.47	0.097
9	2.579		0.0000	0.000	10061.3	0.228	VB	0.415	404.07	0.016
10	2.768		0.0000	0.000	1505.0	0.034	BB	0.070	356.01	0.014
11	2.835		0.0000	0.000	9730.8	0.221	BB	0.087	1865.37	0.075
12	2.997	TCX	73.3933	52.935	531761.9	12.067	BB	0.028	319134.00	12.860
13	3.159		0.0000	0.000	305.7	0.007	BB	0.058	87.57	0.004
14	3.225		0.0000	0.000	17448.6	0.396	BB	0.027	10737.69	0.433
15	3.308		0.0000	0.000	26010.1	0.590	BB	0.033	12960.56	0.522
16	3.378		0.0000	0.000	89171.5	2.023	BB	0.027	55628.59	2.242
17	3.471		0.0000	0.000	625.3	0.014	BB	0.018	576.49	0.023
18	3.512		0.0000	0.000	12260.1	0.278	BB	0.025	8084.86	0.326
19	3.691		0.0000	0.000	177280.3	4.023	BB	0.039	75646.99	3.048
20	3.769		0.0000	0.000	9941.8	0.226	BB	0.027	6156.77	0.248
21	3.822		0.0000	0.000	2937.5	0.067	BB	0.015	3177.62	0.128
22	3.860		0.0000	0.000	105421.0	2.392	BB	0.026	66931.29	2.697
23	3.935		0.0000	0.000	372.9	0.008	BB	0.015	419.50	0.017
24	3.967		0.0000	0.000	2021.3	0.046	BB	0.021	1642.33	0.066
25	4.028		0.0000	0.000	53027.8	1.203	BB	0.039	22836.97	0.920
26	4.123		0.0000	0.000	306455.1	6.954	BB	0.033	155164.60	6.252
27	4.178		0.0000	0.000	93108.3	2.113	BB	0.024	65124.75	2.624
28	4.262		0.0000	0.000	97184.0	2.205	BB	0.026	61314.00	2.471
29	4.312		0.0000	0.000	3018.5	0.068	BB	0.016	3234.71	0.130
30	4.357		0.0000	0.000	43786.3	0.994	BB	0.019	37834.64	1.525
31	4.393		0.0000	0.000	75176.9	1.706	BB	0.032	38974.60	1.570
32	4.525		0.0000	0.000	130002.4	2.950	BB	0.036	60239.19	2.427
33	4.596		0.0000	0.000	11900.6	0.270	BB	0.017	11532.04	0.465
34	4.646		0.0000	0.000	74156.3	1.683	BB	0.028	44450.97	1.791
35	4.695		0.0000	0.000	14702.0	0.334	BB	0.020	12314.84	0.496
36	4.761		0.0000	0.000	2424.2	0.055	BB	0.021	1910.85	0.077
37	4.812		0.0000	0.000	1867.5	0.042	BB	0.022	1437.74	0.058
38	4.854		0.0000	0.000	81910.1	1.859	BB	0.029	47496.72	1.914
39	4.910		0.0000	0.000	6227.9	0.141	BB	0.015	6867.96	0.277
40	5.000		0.0000	0.000	6080.1	0.138	BB	0.022	4553.16	0.183
41	5.055		0.0000	0.000	81453.2	1.848	BB	0.032	42799.16	1.725
42	5.230		0.0000	0.000	3742.0	0.085	BB	0.040	1571.98	0.063
43	5.306		0.0000	0.000	42004.8	0.953	BB	0.030	23674.63	0.954
44	5.372		0.0000	0.000	27068.0	0.614	BB	0.024	18952.05	0.764
45	5.424		0.0000	0.000	81344.3	1.846	BB	0.027	49349.84	1.989
46	5.466		0.0000	0.000	3963.7	0.090	BB	0.014	4874.82	0.196
47	5.518		0.0000	0.000	161170.3	3.657	BB	0.026	102614.20	4.135
48	5.633		0.0000	0.000	22009.3	0.499	BB	0.040	9226.07	0.372
49	5.700		0.0000	0.000	24711.3	0.561	BB	0.025	16484.66	0.664
50	5.763		0.0000	0.000	182208.9	4.135	BB	0.028	106633.20	4.297
51	5.808		0.0000	0.000	12751.1	0.289	BB	0.017	12414.40	0.500
52	5.873		0.0000	0.000	91419.6	2.074	BB	0.028	54160.77	2.182
53	5.964		0.0000	0.000	7387.7	0.168	BB	0.022	5654.12	0.228
54	6.008		0.0000	0.000	182334.3	4.137	BB	0.029	105101.10	4.235

ASX22 8548

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.080		0.0000	0.000	100077.6	2.271	BB	0.024	70732.68	2.850
56	6.131		0.0000	0.000	53077.9	1.204	BB	0.024	37455.46	1.509
57	6.276		0.0000	0.000	168531.3	3.824	BB	0.035	81409.95	3.280
58	6.349		0.0000	0.000	54055.7	1.227	BB	0.025	35520.34	1.431
59	6.397		0.0000	0.000	18007.2	0.409	BB	0.021	14108.75	0.569
60	6.442		0.0000	0.000	2179.7	0.049	BB	0.017	2159.64	0.087
61	6.508		0.0000	0.000	32332.4	0.734	BB	0.026	20585.41	0.829
62	6.575		0.0000	0.000	301087.2	6.832	BB	0.029	171240.90	6.900
63	6.829		0.0000	0.000	135013.1	3.064	BB	0.029	78852.95	3.177
64	6.889		0.0000	0.000	73188.0	1.661	BB	0.036	33670.94	1.357
65	7.053		0.0000	0.000	4561.0	0.103	BB	0.034	2222.02	0.090
66	7.178		0.0000	0.000	50288.1	1.141	BB	0.041	20413.73	0.823
67	7.412		0.0000	0.000	77931.5	1.768	BB	0.030	42779.95	1.724
68	7.484		0.0000	0.000	3259.4	0.074	BB	0.025	2180.62	0.088
69	7.577		0.0000	0.000	3159.8	0.072	BB	0.099	531.98	0.021
70	7.727		0.0000	0.000	26420.8	0.600	BB	0.033	13344.26	0.538
71	7.893		0.0000	0.000	80.8	0.002	BB	0.029	47.04	0.002
72	7.972	DCB	65.2538	47.065	299125.3	6.788	BB	0.035	141686.60	5.709
73	8.143		0.0000	0.000	281.1	0.006	BB	0.032	147.46	0.006
74	8.300		0.0000	0.000	629.9	0.014	BB	0.074	142.29	0.006
75	8.627		0.0000	0.000	5379.0	0.122	BB	0.106	841.98	0.034
76	9.073		0.0000	0.000	2548.9	0.058	BB	0.348	122.09	0.005

Total Area = 4406905.0, Total Amount = 138.647, Total Height = 2481682.0, Sample Units = ug/l

ASX22 8549

Sample Name: **AR1640924A FOAR164FO CCAL 0930899999**
Acquired from CP07--H5088B via port 4 on 11/6/09 04:50:15am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.63R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.580		0.0000	0.000	11959.8	0.217	BV	0.063	3153.72	0.127
2	1.708		0.0000	0.000	4420.9	0.080	VV	0.062	1189.85	0.048
3	1.834		0.0000	0.000	1358.0	0.025	VV	0.092	245.02	0.010
4	1.897		0.0000	0.000	1044.5	0.019	VV	0.089	195.03	0.008
5	1.966		0.0000	0.000	2465.1	0.045	VB	0.058	711.61	0.029
6	2.200		0.0000	0.000	4668.1	0.085	BB	0.072	1077.22	0.043
7	2.284		0.0000	0.000	1783.2	0.032	BB	0.040	743.50	0.030
8	2.339	TCX	71.3195	51.304	722265.2	13.113	BB	0.039	309301.00	12.423
9	2.561		0.0000	0.000	15829.7	0.287	BB	0.033	8080.18	0.325
10	2.640		0.0000	0.000	21767.9	0.395	BB	0.037	9783.04	0.393
11	2.712		0.0000	0.000	112029.4	2.034	BB	0.041	45164.27	1.814
12	2.857		0.0000	0.000	16896.9	0.307	BB	0.036	7828.07	0.314
13	3.004		0.0000	0.000	271323.4	4.926	BB	0.047	96504.67	3.876
14	3.167		0.0000	0.000	161323.1	2.929	BB	0.039	68818.72	2.764
15	3.308		0.0000	0.000	50770.2	0.922	BB	0.044	19073.25	0.766
16	3.400		0.0000	0.000	217391.3	3.947	BB	0.032	111527.30	4.479
17	3.445		0.0000	0.000	39037.9	0.709	BB	0.026	24579.97	0.987
18	3.553		0.0000	0.000	83812.2	1.522	BB	0.030	46628.46	1.873
19	3.614		0.0000	0.000	15322.5	0.278	BB	0.024	10792.06	0.433
20	3.640		0.0000	0.000	68400.2	1.242	BB	0.038	29809.21	1.197
21	3.793		0.0000	0.000	179179.0	3.253	BB	0.051	58729.18	2.359
22	3.883		0.0000	0.000	1271.2	0.023	BB	0.012	1718.41	0.069
23	3.919		0.0000	0.000	48355.8	0.878	BB	0.028	28492.52	1.144
24	3.967		0.0000	0.000	10954.4	0.199	BB	0.025	7330.78	0.294
25	4.105		0.0000	0.000	168883.8	3.066	BB	0.049	57363.85	2.304
26	4.195		0.0000	0.000	1767.1	0.032	BB	0.020	1459.51	0.059
27	4.248		0.0000	0.000	103427.4	1.878	BB	0.044	39142.73	1.572
28	4.430		0.0000	0.000	1215.0	0.022	BB	0.029	694.94	0.028
29	4.501		0.0000	0.000	4435.9	0.081	BB	0.023	3260.21	0.131
30	4.538		0.0000	0.000	16810.5	0.305	BB	0.022	12937.33	0.520
31	4.594		0.0000	0.000	180055.9	3.269	BB	0.049	61111.07	2.454
32	4.689		0.0000	0.000	223657.9	4.060	BB	0.033	112741.30	4.528
33	4.817		0.0000	0.000	30896.5	0.561	BB	0.034	15043.52	0.604
34	4.869		0.0000	0.000	258765.6	4.698	BB	0.034	126748.00	5.091
35	4.983		0.0000	0.000	296662.8	5.386	BB	0.054	92007.85	3.695
36	5.096		0.0000	0.000	21342.8	0.387	BB	0.023	15499.42	0.623
37	5.138		0.0000	0.000	98808.7	1.794	BB	0.023	71872.82	2.887
38	5.172		0.0000	0.000	51875.1	0.942	BB	0.019	45780.55	1.839
39	5.216		0.0000	0.000	35825.6	0.650	BB	0.022	27218.42	1.093
40	5.333		0.0000	0.000	21146.9	0.384	BB	0.025	14010.87	0.563
41	5.386		0.0000	0.000	204813.9	3.718	BB	0.034	100046.80	4.018
42	5.469		0.0000	0.000	53871.0	0.978	BB	0.025	36345.79	1.460
43	5.506		0.0000	0.000	11733.4	0.213	BB	0.020	9876.41	0.397
44	5.557		0.0000	0.000	10530.2	0.191	BB	0.021	8481.35	0.341
45	5.623		0.0000	0.000	494811.6	8.983	BB	0.038	217765.00	8.746
46	5.788		0.0000	0.000	378.7	0.007	BB	0.020	317.22	0.013
47	5.864		0.0000	0.000	62870.6	1.141	BB	0.025	41848.52	1.681
48	5.919		0.0000	0.000	294602.2	5.348	BB	0.037	131655.00	5.288
49	6.085		0.0000	0.000	6006.2	0.109	BB	0.033	3054.00	0.123
50	6.169		0.0000	0.000	8823.4	0.160	BB	0.030	4823.64	0.194
51	6.234		0.0000	0.000	56742.5	1.030	BB	0.035	27263.20	1.095
52	6.379		0.0000	0.000	162059.7	2.942	BB	0.041	65207.59	2.619
53	6.617		0.0000	0.000	46719.7	0.848	BB	0.038	20368.74	0.818
54	6.796	DCB	67.6935	48.696	512977.2	9.313	BB	0.037	233692.60	9.386

ASX22 0550

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	7.323		0.0000	0.000	1424.1	0.026	BB	0.052	457.23	0.018
56	7.750		0.0000	0.000	599.9	0.011	BB	0.045	223.53	0.009

Total Area = 5508171.0, Total Amount = 139.013, Total Height = 2489796.0, Sample Units = ug/l

ASX22 8551

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3P7307.63R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 4:41:06 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.63R



Instrument ID: CP07--H5088B Injected On: 11/6/2009 4:41:06 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1
Analyst: 2280

Dilution Factor: 1

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.997	319134	73.39328	TCX	2.339	309301	71.31948	TCX
7.972	141687	65.25384	DCB	6.796	233693	67.69348	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.63A
Area File: C:\CPWIN\DATA1\3P7307B.63A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 4:50:22 AM
File Reported On: 11/6/2009 at 4:50:35 AM

10X REPLOTS

ASX22 0553

C:\CPWIN\DATA1\3P7307.63R



C:\CPWIN\DATA1\3P7307B.63R



ASX22. 0554.

Sample Name: **IBLKX0924B FHPIBLKFH PIBLK0930899999**
Acquired from CP07--H5088A via port 3 on 11/6/09 05:03:17am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.64R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.929		0.0000	0.000	30324.8	6.750	BB	0.262	1928.72	1.170
2	2.076		0.0000	0.000	16516.9	3.677	BV	0.118	2324.08	1.410
3	2.175		0.0000	0.000	13877.2	3.089	VV	0.091	2547.29	1.545
4	2.584		0.0000	0.000	12561.1	2.796	VB	0.574	364.62	0.221
5	2.782		0.0000	0.000	1326.2	0.295	BB	0.087	253.70	0.154
6	2.839		0.0000	0.000	5150.3	1.146	BB	0.078	1095.32	0.664
7	2.999	TCX	195.1024	54.086	167836.4	37.361	BB	0.028	99124.13	60.117
8	3.212		0.0000	0.000	90.1	0.020	BB	0.026	58.62	0.036
9	3.288		0.0000	0.000	2792.1	0.622	BB	0.030	1566.92	0.950
10	3.381		0.0000	0.000	114.8	0.026	BB	0.027	69.87	0.042
11	3.473		0.0000	0.000	338.8	0.075	BB	0.027	207.80	0.126
12	3.780		0.0000	0.000	1283.7	0.286	BB	0.140	153.07	0.093
13	3.859		0.0000	0.000	415.0	0.092	BB	0.032	214.54	0.130
14	3.940		0.0000	0.000	444.9	0.099	BB	0.029	255.75	0.155
15	4.101		0.0000	0.000	1395.0	0.311	BB	0.033	701.33	0.425
16	4.208		0.0000	0.000	608.7	0.135	BB	0.055	184.40	0.112
17	4.267		0.0000	0.000	224.4	0.050	BB	0.029	129.97	0.079
18	4.459		0.0000	0.000	969.0	0.216	BB	0.083	193.60	0.117
19	4.597		0.0000	0.000	253.1	0.056	BB	0.061	69.19	0.042
20	4.689		0.0000	0.000	1123.1	0.250	BB	0.057	327.05	0.198
21	4.811		0.0000	0.000	980.1	0.218	BB	0.071	229.98	0.139
22	4.922		0.0000	0.000	111.2	0.025	BB	0.022	84.73	0.051
23	4.982		0.0000	0.000	568.1	0.126	BB	0.044	213.40	0.129
24	5.104		0.0000	0.000	1402.5	0.312	BB	0.044	536.69	0.325
25	5.262		0.0000	0.000	1449.5	0.323	BB	0.091	265.94	0.161
26	5.380		0.0000	0.000	2792.0	0.622	BB	0.062	750.33	0.455
27	5.474		0.0000	0.000	3625.1	0.807	BB	0.164	368.00	0.223
28	5.715		0.0000	0.000	225.1	0.050	BB	0.027	137.63	0.083
29	5.765		0.0000	0.000	911.1	0.203	BB	0.065	231.85	0.141
30	6.010		0.0000	0.000	310.7	0.069	BB	0.029	177.46	0.108
31	6.086		0.0000	0.000	179.5	0.040	BB	0.026	113.24	0.069
32	6.397		0.0000	0.000	20847.0	4.641	BB	0.176	1973.12	1.197
33	6.729		0.0000	0.000	117.2	0.026	BB	0.026	74.67	0.045
34	6.832		0.0000	0.000	6370.8	1.418	BB	0.163	650.51	0.395
35	7.152		0.0000	0.000	35306.9	7.860	BB	0.320	1838.86	1.115
36	7.584		0.0000	0.000	17180.7	3.825	BB	0.237	1207.62	0.732
37	7.730		0.0000	0.000	1480.6	0.330	BB	0.046	535.68	0.325
38	7.891		0.0000	0.000	66.4	0.015	BB	0.027	40.45	0.025
39	7.975	DCB	165.6241	45.914	94039.4	20.934	BB	0.036	43074.13	26.124
40	8.179		0.0000	0.000	173.2	0.039	BB	0.042	67.93	0.041
41	8.534		0.0000	0.000	174.1	0.039	BB	0.045	64.46	0.039
42	8.626		0.0000	0.000	3268.9	0.728	BB	0.114	477.84	0.290

Total Area = 449225.1, Total Amount = 360.727, Total Height = 164884.5, Sample Units = ug/l

A3X22 0555

Sample Name: **IBLKX0924B FHPIBLKFH PIBLK0930899999**
Acquired from CP07–H5088B via port 4 on 11/6/09 05:03:17am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.64R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.580		0.0000	0.000	6317.3	1.512	BV	0.085	1245.05	0.733
2	1.709		0.0000	0.000	1588.6	0.380	VV	0.057	465.73	0.274
3	1.820		0.0000	0.000	343.7	0.082	VV	0.105	54.80	0.032
4	1.891		0.0000	0.000	159.4	0.038	VV	0.058	45.52	0.027
5	1.955		0.0000	0.000	135.8	0.032	VB	0.102	22.18	0.013
6	2.025		0.0000	0.000	411.2	0.098	BB	0.055	125.11	0.074
7	2.200		0.0000	0.000	1561.3	0.374	BB	0.058	451.91	0.266
8	2.286		0.0000	0.000	1427.3	0.342	BB	0.047	502.32	0.296
9	2.341	TCX	205.7129	50.231	200448.7	47.985	BB	0.037	89214.34	52.539
10	2.623		0.0000	0.000	2575.6	0.617	BB	0.038	1125.03	0.663
11	2.777		0.0000	0.000	390.3	0.093	BB	0.038	170.49	0.100
12	2.960		0.0000	0.000	85.1	0.020	BB	0.038	36.87	0.022
13	3.047		0.0000	0.000	388.4	0.093	BB	0.208	31.13	0.018
14	3.125		0.0000	0.000	9.9	0.002	BB	0.006	27.65	0.016
15	3.166		0.0000	0.000	606.6	0.145	BB	0.054	188.88	0.111
16	3.244		0.0000	0.000	28.2	0.007	BB	0.019	24.55	0.014
17	3.270		0.0000	0.000	377.2	0.090	BB	0.034	184.00	0.108
18	3.348		0.0000	0.000	2625.8	0.629	BB	0.088	496.63	0.292
19	3.560		0.0000	0.000	130.1	0.031	BB	0.025	86.66	0.051
20	3.884		0.0000	0.000	846.2	0.203	BB	0.072	196.97	0.116
21	4.123		0.0000	0.000	1590.7	0.381	BB	0.094	280.97	0.165
22	4.304		0.0000	0.000	1650.3	0.395	BB	0.085	323.03	0.190
23	4.557		0.0000	0.000	505.1	0.121	BB	0.317	26.56	0.016
24	4.620		0.0000	0.000	87.1	0.021	BB	0.052	28.05	0.017
25	4.664		0.0000	0.000	68.0	0.016	BB	0.034	33.79	0.020
26	4.691		0.0000	0.000	473.5	0.113	BB	0.038	206.37	0.122
27	4.827		0.0000	0.000	2625.8	0.629	BB	0.073	595.86	0.351
28	4.987		0.0000	0.000	379.7	0.091	BB	0.040	158.67	0.093
29	5.070		0.0000	0.000	57.3	0.014	BB	0.028	34.22	0.020
30	5.114		0.0000	0.000	26.2	0.006	BB	0.012	35.96	0.021
31	5.177		0.0000	0.000	380.5	0.091	BB	0.085	74.88	0.044
32	5.380		0.0000	0.000	2486.8	0.595	BB	0.106	390.33	0.230
33	5.626		0.0000	0.000	859.0	0.206	BB	0.041	349.94	0.206
34	5.796		0.0000	0.000	4876.1	1.167	BB	0.076	1065.18	0.627
35	6.588		0.0000	0.000	11054.4	2.646	BB	3.871	47.60	0.028
36	6.622		0.0000	0.000	2388.8	0.572	BB	0.052	773.04	0.455
37	6.799	DCB	203.8212	49.769	164926.6	39.482	BB	0.039	70363.51	41.437
38	7.336		0.0000	0.000	432.0	0.103	BB	0.053	134.94	0.079
39	7.490		0.0000	0.000	2230.8	0.534	BB	0.340	109.48	0.064
40	7.763		0.0000	0.000	174.1	0.042	BB	0.037	79.30	0.047

Total Area = 417729.7, Total Amount = 409.534, Total Height = 169807.5, Sample Units = ug/l

ASX22 8556

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3P7307.64R

C:\CPWIN\DATA1\3P7307.64R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 4:54:08 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.64R

C:\CPWIN\DATA1\3P7307B.64R



Instrument ID: CP07--H5088B Injected On: 11/6/2009 4:54:08 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1000
Analyst: 2280

Dilution Factor: 10

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
2.999	99124	.1951	TCX	2.341	89214	.20571	TCX
7.975	43074	.16562	DCB	6.799	70364	.20382	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.64A
Area File: C:\CPWIN\DATA1\3P7307B.64A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 5:03:24 AM
File Reported On: 11/6/2009 at 5:03:37 AM

10X REPLOTS

ASX22 0558

C:\CPWIN\DATA1\3P7307.64R



C:\CPWIN\DATA1\3P7307B.64R



ASXZ2 0559

Lancaster Laboratories

CHROM PERFECT SEQUENCE FILE

Sequence File: \\d35pc_ch\C-Drive\CPWIN\DATA1\1c6310.seq

Chromatography Directory: \\d35pc_ch\C-Drive\cpwin\DATA1

Method Directory: \\d35pc_ch\C-Drive\cpwin\DATA1

Number of Entries: 94

	Samplename	Code	ID	FileName	Method	Samp Wt	DF	Int Std	Batch Number	Analysis
1	CONDITIONER	MISC	AA	1c6310.01R	PESTD.MET	1	1	1	0930999999	
2	CONDITIONER	MISC	AA	1c6310.02R	PESTD.MET	1	1	1	0930999999	
3	CONDITIONER	MISC	AA	1c6310.03R	PESTD.MET	1	1	1	0930999999	
4	PBLKX0924B	PIBLK	AA	1c6310.04R	PESTD.MET	1000	10	1	0930999999	
5	AR1610924A	ICAL	AA	1c6310.05R	PESTD.MET	1	1	1	0930999999	
6	AR1620924A	ICAL	AA	1c6310.06R	PESTD.MET	1	1	1	0930999999	
7	AR1630924A	ICAL	AA	1c6310.07R	PESTD.MET	1	1	1	0930999999	
8	AR1640924A	ICAL	AA	1c6310.08R	PESTD.MET	1	1	1	0930999999	
9	AR1650924A	ICAL	AA	1c6310.09R	PESTD.MET	1	1	1	0930999999	
10	AR4810924B	ICAL	AA	1c6310.10R	PESTD.MET	1	1	1	0930999999	
11	AR4820924B	ICAL	AA	1c6310.11R	PESTD.MET	1	1	1	0930999999	
12	AR4830924B	ICAL	AA	1c6310.12R	PESTD.MET	1	1	1	0930999999	
13	AR4840924B	ICAL	AA	1c6310.13R	PESTD.MET	1	1	1	0930999999	
14	AR4850924B	ICAL	AA	1c6310.14R	PESTD.MET	1	1	1	0930999999	
15	AR5410924B	ICAL	AA	1c6310.15R	PESTD.MET	1	1	1	0930999999	
16	AR5420924B	ICAL	AA	1c6310.16R	PESTD.MET	1	1	1	0930999999	
17	AR5430924B	ICAL	AA	1c6310.17R	PESTD.MET	1	1	1	0930999999	
18	AR5440924B	ICAL	AA	1c6310.18R	PESTD.MET	1	1	1	0930999999	
19	AR5450924B	ICAL	AA	1c6310.19R	PESTD.MET	1	1	1	0930999999	
20	AR6210924B	ICAL	AA	1c6310.20R	PESTD.MET	1	1	1	0930999999	
21	AR6220924B	ICAL	AA	1c6310.21R	PESTD.MET	1	1	1	0930999999	
22	AR6230924B	ICAL	AA	1c6310.22R	PESTD.MET	1	1	1	0930999999	
23	AR6240924B	ICAL	AA	1c6310.23R	PESTD.MET	1	1	1	0930999999	
24	AR6250924B	ICAL	AA	1c6310.24R	PESTD.MET	1	1	1	0930999999	
25	AR6810924C	ICAL	AA	1c6310.25R	PESTD.MET	1	1	1	0930999999	
26	AR6820924C	ICAL	AA	1c6310.26R	PESTD.MET	1	1	1	0930999999	
27	AR6830924C	ICAL	AA	1c6310.27R	PESTD.MET	1	1	1	0930999999	
28	AR6840924C	ICAL	AA	1c6310.28R	PESTD.MET	1	1	1	0930999999	
29	AR6850924C	ICAL	AA	1c6310.29R	PESTD.MET	1	1	1	0930999999	
30	AR2130924B	ICAL	AA	1c6310.30R	PESTD.MET	1	1	1	0930999999	
31	AR32X0924B	ICAL	AA	1c6310.31R	PESTD.MET	1	1	1	0930999999	
32	AR4230924B	ICAL	AA	1c6310.32R	PESTD.MET	1	1	1	0930999999	
33	AR16X0924A	ICAL	AA	1c6310.33R	PESTD.MET	1	1	1	0930999999	
34	MD16X0924A	ICAL	AA	1c6310.34R	PESTD.MET	1	1	1	0930999999	
35	MD62X0924B	ICAL	AA	1c6310.35R	PESTD.MET	1	1	1	0930999999	
36	MD68X0924C	ICAL	AA	1c6310.36R	PESTD.MET	1	1	1	0930999999	
37	IC16X0924A	CCAL	AA	1c6310.37R	PESTD.MET	1	1	1	0930999999	
38	IC48X0924B	CCAL	AA	1c6310.38R	PESTD.MET	1	1	1	0930999999	
39	IC54X0924B	CCAL	AA	1c6310.39R	PESTD.MET	1	1	1	0930999999	
40	BLANKA 11/5/09 RI A	BLK	AB	1c6310.40R	PESTD.MET	1000	10	1	093080009A	00173
41	LCSA 11/5/09 RI A	LCS	AB	1c6310.41R	PESTD.MET	1000	10	1	093080009A	00173
42	LCSDA 11/5/09 RI A	LCSD	AB	1c6310.42R	PESTD.MET	1000	10	1	093080009A	00173
43	5826632AF DF50	T	AC	1c6310.43R	PESTD.MET	1013	500	1	093080009A	00173
44	5826632AF DF250	T	AD	1c6310.44R	PESTD.MET	1013	2500	1	093080009A	00173
45	5826633AF DF50	T	AC	1c6310.45R	PESTD.MET	982	500	1	093080009A	00173
46	5826633AF DF250	T	AD	1c6310.46R	PESTD.MET	982	2500	1	093080009A	00173
47	5826635AF DF50	T	AC	1c6310.47R	PESTD.MET	992	500	1	093080009A	00173
48	5826635AF DF250	T	AD	1c6310.48R	PESTD.MET	992	2500	1	093080009A	00173
49	5826636AF DF10	T	AC	1c6310.49R	PESTD.MET	985	100	1	093080009A	00173
50	AR1630924A	CCAL	PB	1c6310.50R	PESTD.MET	1	1	1	0930999999	ASX22 8568

Lancaster Laboratories
CHROM PERFECT SEQUENCE FILE

Sequence File: \\d35pc_ch\C-Drive\CPWIN\DATA1\1c6310.seq
Chromatography Directory: \\d35pc_ch\C-Drive\cpwin\DATA1
Method Directory: \\d35pc_ch\C-Drive\cpwin\DATA1
Number of Entries: 94

Samplename	Code	ID	FileName	Method	Samp Wt	DF	Int Std	Batch Number	Analysis
51 IBLKX0924B	PIBLK	FJ	1c6310.51R	PESTD.MET	1000	10	1	0930999999	
52 5826636AF DF50	T	AD	1c6310.52R	PESTD.MET	985	500	1	093080009A	00173
53 5826637 RI AF	T	AB	1c6310.53R	PESTD.MET	973	10	1	093080009A	00173
54 BLANKA 11/6/09AF	BLK	AB	1c6310.54R	PESTD.MET	1000	10	1	093080011A	02034
55 LCSA 11/6/09AF	LCS	AB	1c6310.55R	PESTD.MET	1000	10	1	093080011A	02034
56 5813035RCAF	T	AC	1c6310.56R	PESTD.MET	1057	10	1	093080011A	02034
57 5813036RCAF	T	AC	1c6310.57R	PESTD.MET	1053	10	1	093080011A	02034
58 5813037RCAF	T	AC	1c6310.58R	PESTD.MET	1054	10	1	093080011A	02034
59 5813038RCAF	T	AC	1c6310.59R	PESTD.MET	1055	10	1	093080011A	02034
60 5813039RCAF	T	AC	1c6310.60R	PESTD.MET	1059	10	1	093080011A	02034
61 5813040RCAF	T	AC	1c6310.61R	PESTD.MET	1059	10	1	093080011A	02034
62 AR1620924A	CCAL	ON	1c6310.62R	PESTD.MET	1	1	1	0930999999	
63 IBLKX0924B	PIBLK	FJ	1c6310.63R	PESTD.MET	1000	10	1	0930999999	
64 5813041RCAF	T	AC	1c6310.64R	PESTD.MET	1056	10	1	093080011A	02034
65 5813042RCAF	T	AC	1c6310.65R	PESTD.MET	1054	10	1	093080011A	02034
66 5816066RCAF	T	AC	1c6310.66R	PESTD.MET	1051	10	1	093080011A	02034
67 5816067RCAF	T	AC	1c6310.67R	PESTD.MET	1055	10	1	093080011A	02034
68 5816068RCAF DF5	T	AD	1c6310.68R	PESTD.MET	100	50	1	093080011A	02034
69 5816069RCAF	T	AC	1c6310.69R	PESTD.MET	1053	10	1	093080011A	02034
70 5816070RMSCAF	MS	AC	1c6310.70R	PESTD.MET	1049	10	1	093080011A	02034
71 5816071RMSDCAF	MSD	AC	1c6310.71R	PESTD.MET	1050	10	1	093080011A	02034
72 5816072RCAF	T	AC	1c6310.72R	PESTD.MET	1057	10	1	093080011A	02034
73 5820183RCAF	T	AC	1c6310.73R	PESTD.MET	1057	10	1	093080011A	02034
74 AR1630924A	CCAL	PB	1c6310.74R	PESTD.MET	1	1	1	0930999999	
75 IBLKX0924B	PIBLK	FJ	1c6310.75R	PESTD.MET	1000	10	1	0930999999	
76 5820184RCAF	T	AC	1c6310.76R	PESTD.MET	1057	10	1	093080011A	02034
77 5820185RCAF	T	AC	1c6310.77R	PESTD.MET	1055	10	1	093080011A	02034
78 5820186RCAF	T	AC	1c6310.78R	PESTD.MET	1050	10	1	093080011A	02034
79 5820187RCAF	T	AC	1c6310.79R	PESTD.MET	1045	10	1	093080011A	02034
80 5820188RCAF	T	AC	1c6310.80R	PESTD.MET	1048	10	1	093080011A	02034
81 5820189RCAF	T	AC	1c6310.81R	PESTD.MET	1050	10	1	093080011A	02034
82 BLANKA 11/4/09ACF	BLK	AB	1c6310.82R	PESTD.MET	30	10	1	093080004A	02033
83 LCSA 11/4/09ACF	LCS	AB	1c6310.83R	PESTD.MET	30	10	1	093080004A	02033
84 5823539ACF	T	AB	1c6310.84R	PESTD.MET	30	10	1	093080004A	02033
85 5823540ACF DF100	T	AC	1c6310.85R	PESTD.MET	30	1000	1	093080004A	02033
86 AR1620924A	CCAL	OO	1c6310.86R	PESTD.MET	1	1	1	0930999999	
87 IBLKX0924B	PIBLK	FJ	1c6310.87R	PESTD.MET	1000	10	1	0930999999	
88 5825593ACF	T	AB	1c6310.88R	PESTD.MET	30	10	1	093080004A	02033
89 5825593MSACF	MS	AB	1c6310.89R	PESTD.MET	30	10	1	093080004A	02033
90 5825593MSDACF	MSD	AB	1c6310.90R	PESTD.MET	30	10	1	093080004A	02033
91 5825594ACF	T	AB	1c6310.91R	PESTD.MET	30	10	1	093080004A	02033
92 5825595ACF	T	AB	1c6310.92R	PESTD.MET	30	10	1	093080004A	02033
93 AR1640924A	CCAL	FS	1c6310.93R	PESTD.MET	1	1	1	0930999999	
94 IBLKX0924B	PIBLK	FJ	1c6310.94R	PESTD.MET	1000	10	1	0930999999	

ASX22 #561

Set-up by: Andrea Carey 1157 Date: 11-709

11/7/2009

Lancaster Laboratories

CHROM PERFECT SEQUENCE FILE

Sequence File: \\D35PC_CH\C-Drive\cpwin\DATA1\2c6310.seq

Chromatography Directory: \\D35PC_CH\C-Drive\cpwin\DATA1

Method Directory: \\D35PC_CH\C-Drive\cpwin\DATA1

Number of Entries: 59

	Samplename	Code	ID	FileName	Method	Samp Wt	DF	Int Std	Batch Number	Analysis
1	5822189 RI AF	T	AB	2c6310.01R	PESTD.MET	1057	10	1	093060018A	00173
2	5822191 RI AF	T	AB	2c6310.02R	PESTD.MET	1046	10	1	093060018A	00173
3	BLANKA 11/4/09ACF	BLK	AB	2c6310.03R	PESTD.MET	30	10	1	093070028A	01216
4	LCSA 11/4/09ACF	LCS	AB	2c6310.04R	PESTD.MET	30	10	1	093070028A	01216
5	5814908ACF	T	AB	2c6310.05R	PESTD.MET	30	10	1	093070028A	01216
6	5822069ACF DF50	T	AC	2c6310.06R	PESTD.MET	30	500	1	093070028A	01216
7	5822070ACF DF100	T	AC	2c6310.07R	PESTD.MET	30	1000	1	093070028A	01216
8	5822070ACF DF1000	T	AD	2c6310.08R	PESTD.MET	30	10000	1	093070028A	01216
9	5822071MSACF DF100	MS	AC	2c6310.09R	PESTD.MET	30	1000	1	093070028A	01216
10	5822071MSACF DF1000	MS	AD	2c6310.10R	PESTD.MET	30	10000	1	093070028A	01216
11	AR1630924A	CCAL	PM	2c6310.11R	PESTD.MET	1	1	1	0931099999	
12	IBLKX0924B	PIBLK	FU	2c6310.12R	PESTD.MET	1000	10	1	0931099999	
13	5822072MSDACF DF100	MSD	AC	2c6310.13R	PESTD.MET	30	1000	1	093070028A	01216
14	5822072MSDACF DF1000	MSD	AD	2c6310.14R	PESTD.MET	30	10000	1	093070028A	01216
15	5822073ACF DF100	T	AC	2c6310.15R	PESTD.MET	30	1000	1	093070028A	01216
16	5822073ACF DF1000	T	AD	2c6310.16R	PESTD.MET	30	10000	1	093070028A	01216
17	5822074ACF DF100	T	AC	2c6310.17R	PESTD.MET	30	1000	1	093070028A	01216
18	5822074ACF DF1000	T	AD	2c6310.18R	PESTD.MET	30	10000	1	093070028A	01216
19	5822075ACF DF20	T	AC	2c6310.19R	PESTD.MET	30	200	1	093070028A	01216
20	5822076ACF	T	AB	2c6310.20R	PESTD.MET	30	25	1	093070028A	01216
21	5822077ACF DF100	T	AC	2c6310.21R	PESTD.MET	30	1000	1	093070028A	01216
22	5822077ACF DF1000	T	AD	2c6310.22R	PESTD.MET	30	10000	1	093070028A	01216
23	AR1620924A	CCAL	OS	2c6310.23R	PESTD.MET	1	1	1	0931099999	
24	IBLKX0924B	PIBLK	FV	2c6310.24R	PESTD.MET	1000	10	1	0931099999	
25	BLANKA 11/4/09ACF	BLK	AB	2c6310.25R	PESTD.MET	30	10	1	093070027A	01216
26	LCSA 11/4/09ACF	LCS	AB	2c6310.26R	PESTD.MET	30	10	1	093070027A	01216
27	5821274ACF	T	AB	2c6310.27R	PESTD.MET	30	10	1	093070027A	01216
28	5821275MSACF	MS	AB	2c6310.28R	PESTD.MET	30	10	1	093070027A	01216
29	5821276MSDACF	MSD	AB	2c6310.29R	PESTD.MET	30	10	1	093070027A	01216
30	5821278ACF	T	AB	2c6310.30R	PESTD.MET	30	10	1	093070027A	01216
31	5821279ACF	T	AB	2c6310.31R	PESTD.MET	30	10	1	093070027A	01216
32	5821280ACF	T	AB	2c6310.32R	PESTD.MET	30	10	1	093070027A	01216
33	5821281ACF	T	AB	2c6310.33R	PESTD.MET	30	10	1	093070027A	01216
34	5821282ACF	T	AB	2c6310.34R	PESTD.MET	30	10	1	093070027A	01216
35	AR1630924A	CCAL	PO	2c6310.35R	PESTD.MET	1	1	1	0931099999	
36	IBLKX0924B	PIBLK	FW	2c6310.36R	PESTD.MET	1000	10	1	0931099999	
37	5821283ACF	T	AB	2c6310.37R	PESTD.MET	30	10	1	093070027A	01216
38	5821284ACF	T	AB	2c6310.38R	PESTD.MET	30	10	1	093070027A	01216
39	5821285ACF	T	AB	2c6310.39R	PESTD.MET	30	10	1	093070027A	01216
40	5821286ACF	T	AB	2c6310.40R	PESTD.MET	30	10	1	093070027A	01216
41	5821287ACF	T	AB	2c6310.41R	PESTD.MET	30	10	1	093070027A	01216
42	5821288ACF	T	AB	2c6310.42R	PESTD.MET	30	10	1	093070027A	01216
43	5821289ACF	T	AB	2c6310.43R	PESTD.MET	30	10	1	093070027A	01216
44	5821290ACF	T	AB	2c6310.44R	PESTD.MET	30	10	1	093070027A	01216
45	5821291ACF	T	AB	2c6310.45R	PESTD.MET	30	10	1	093070027A	01216
46	5821292ACF	T	AB	2c6310.46R	PESTD.MET	30	10	1	093070027A	01216
47	AR1640924A	CCAL	FU	2c6310.47R	PESTD.MET	1	1	1	0931099999	
48	IBLKX0924B	PIBLK	FX	2c6310.48R	PESTD.MET	1000	10	1	0931099999	
49	5821293ACF	T	AB	2c6310.49R	PESTD.MET	30	10	1	093070027A	01216
50	5821294ACF	T	AB	2c6310.50R	PESTD.MET	30	10	1	093070027A	01216

A3X22 0562

Lancaster Laboratories

CHROM PERFECT SEQUENCE FILE

Sequence File: \\D35PC_CH\C-Drive\cpwin\DATA1\2c6310.seq
Chromatography Directory: \\D35PC_CH\C-Drive\cpwin\DATA1
Method Directory: \\D35PC_CH\C-Drive\cpwin\DATA1
Number of Entries: 59

Samplename	Code	ID	FileName	Method	Samp Wt	DF	Int Std	Batch Number	Analysis
51 5821295ACF	T	AB	2c6310.51R	PESTD.MET	30	10	1	093070027A	01216
52 5821296ACF	T	AB	2c6310.52R	PESTD.MET	30	10	1	093070027A	01216
53 BLANKA 11/5/09ACF	BLK	AB	2c6310.53R	PESTD.MET	30	10	1	093090009A	01216
54 LCSA 11/5/09ACF	LCS	AB	2c6310.54R	PESTD.MET	30	10	1	093090009A	01216
55 5826082ACF	T	AB	2c6310.55R	PESTD.MET	30	10	1	093090009A	01216
56 5826082MSACF	MS	AB	2c6310.56R	PESTD.MET	30	10	1	093090009A	01216
57 5826082MSDACF	MSD	AB	2c6310.57R	PESTD.MET	30	10	1	093090009A	01216
58 AR1630924A	CCAL	PQ	2c6310.58R	PESTD.MET	1	1	1	0931099999	
59 IBLKX0924B	PIBLK	FY	2c6310.59R	PESTD.MET	1000	10	1	0931099999	

ASX22 0563

Set-up by: Andrea Carey Date: 11-9-09

11/7/09

Lancaster Laboratories

CHROM PERFECT SEQUENCE FILE

Sequence File: \\D35PC_CH\C-Drive\CPWIN\DATA1\3c6310.seq
Chromatography Directory: \\D35PC_CH\C-Drive\cpwin\DATA1
Method Directory: \\D35PC_CH\C-Drive\cpwin\DATA1
Number of Entries: 58

Samplename	Code	ID	FileName	Method	Samp Wt	DF	Int Std	Batch Number	Analysis
1 CONDITIONER	MISC	AA	3c6310.01R	PESTD.MET	1	1	1	0931299999	
2 CONDITIONER	MISC	AA	3c6310.02R	PESTD.MET	1	1	1	0931299999	
3 CONDITIONER	MISC	AA	3c6310.03R	PESTD.MET	1	1	1	0931299999	
4 AR1640924A	CCAL	FY	3c6310.04R	PESTD.MET	1	1	1	0931299999	
5 IBLKX0924B	PIBLK	GD	3c6310.05R	PESTD.MET	1000	10	1	0931299999	
6 BLANKA 11/4/09 RI A	BLK	AB	3c6310.06R	PESTD.MET	30	10	1	093070027A	01216
7 LCSA 11/4/09 RI A	LCS	AB	3c6310.07R	PESTD.MET	30	10	1	093070027A	01216
8 5821274 RI ACF	T	AB	3c6310.08R	PESTD.MET	30	10	1	093070027A	01216
9 5821275MS RI ACF	MS	AB	3c6310.09R	PESTD.MET	30	10	1	093070027A	01216
10 5821276MSD RI ACF	MSD	AB	3c6310.10R	PESTD.MET	30	10	1	093070027A	01216
11 5821278 RI ACF	T	AB	3c6310.11R	PESTD.MET	30	10	1	093070027A	01216
12 5821279 RI ACF	T	AB	3c6310.12R	PESTD.MET	30	10	1	093070027A	01216
13 5821280 RI ACF	T	AB	3c6310.13R	PESTD.MET	30	10	1	093070027A	01216
14 5821281 RI ACF	T	AB	3c6310.14R	PESTD.MET	30	10	1	093070027A	01216
15 5821282 RI ACF	T	AB	3c6310.15R	PESTD.MET	30	10	1	093070027A	01216
16 5821283ACF DF5	T	AC	3c6310.16R	PESTD.MET	30	50	1	093070027A	01216
17 AR1630924A	CCAL	PZ	3c6310.17R	PESTD.MET	1	1	1	0931299999	
18 IBLKX0924B	PIBLK	GE	3c6310.18R	PESTD.MET	1000	10	1	0931299999	
19 BLANKA 11/6/09ACF	BLK	AB	3c6310.19R	PESTD.MET	30	10	1	093090023A	01216
20 LCSA 11/6/09ACF	LCS	AB	3c6310.20R	PESTD.MET	30	10	1	093090023A	01216
21 5827102ACF	T	AB	3c6310.21R	PESTD.MET	30	10	1	093090023A	01216
22 5827102MSACF	MS	AB	3c6310.22R	PESTD.MET	30	10	1	093090023A	01216
23 5827102MSDACF	MSD	AB	3c6310.23R	PESTD.MET	30	10	1	093090023A	01216
24 5827103ACF	T	AB	3c6310.24R	PESTD.MET	30	10	1	093090023A	01216
25 5827104ACF DF20	T	AC	3c6310.25R	PESTD.MET	30	200	1	093090023A	01216
26 5827105ACF DF100	T	AC	3c6310.26R	PESTD.MET	30	1000	1	093090023A	01216
27 5827106ACF DF20	T	AC	3c6310.27R	PESTD.MET	30	200	1	093090023A	01216
28 5827107ACF DF20	T	AC	3c6310.28R	PESTD.MET	30	200	1	093090023A	01216
29 5827108ACF DF10	T	AC	3c6310.29R	PESTD.MET	30	100	1	093090023A	01216
30 5827109ACF DF10	T	AC	3c6310.30R	PESTD.MET	30	100	1	093090023A	01216
31 5827110ACF DF20	T	AC	3c6310.31R	PESTD.MET	30	200	1	093090023A	01216
32 5827111ACF DF20	T	AC	3c6310.32R	PESTD.MET	30	200	1	093090023A	01216
33 5827112ACF DF20	T	AC	3c6310.33R	PESTD.MET	30	200	1	093090023A	01216
34 5827113ACF DF10	T	AC	3c6310.34R	PESTD.MET	30	100	1	093090023A	01216
35 5827114ACF DF20	T	AC	3c6310.35R	PESTD.MET	30	200	1	093090023A	01216
36 5827115ACF DF20	T	AC	3c6310.36R	PESTD.MET	30	200	1	093090023A	01216
37 5827116ACF DF20	T	AC	3c6310.37R	PESTD.MET	30	200	1	093090023A	01216
38 AR1630924A	CCAL	QC	3c6310.38R	PESTD.MET	1	1	1	0931299999	
39 IBLKX0924B	PIBLK	GG	3c6310.39R	PESTD.MET	1000	10	1	0931299999	
40 5827117ACF DF100	T	AC	3c6310.40R	PESTD.MET	30	1000	1	093090023A	01216
41 5827117ACF DF1000	T	AD	3c6310.41R	PESTD.MET	30	10000	1	093090023A	01216
42 5827118ACF DF100	T	AC	3c6310.42R	PESTD.MET	30	1000	1	093090023A	01216
43 5827118ACF DF1000	T	AD	3c6310.43R	PESTD.MET	30	10000	1	093090023A	01216
44 5821278 RI ACF	T	AB	3c6310.44R	PESTD.MET	30	10	1	093070027A	01216
45 AR1630924A	CCAL	QD	3c6310.45R	PESTD.MET	1	1	1	0931299999	
46 IBLKX0924B	PIBLK	GH	3c6310.46R	PESTD.MET	1000	10	1	0931299999	
47 5822072MSD RI ACF DF	MSD	AD	3c6310.47R	PESTD.MET	30	10000	1	093070028A	01216
48 5822073 RI ACF DF100	T	AD	3c6310.48R	PESTD.MET	30	10000	1	093070028A	01216
49 5822074 RI ACF DF100	T	AC	3c6310.49R	PESTD.MET	30	1000	1	093070028A	01216
50 5822075 RI ACF DF20	T	AC	3c6310.50R	PESTD.MET	30	200	1	093070028A	01216

HSX22 0564

Lancaster Laboratories

CHROM PERFECT SEQUENCE FILE

Sequence File: \\D35PC_CH\C-Drive\CPWIN\DATA1\3c6310.seq
Chromatography Directory: \\D35PC_CH\C-Drive\cpwin\DATA1
Method Directory: \\D35PC_CH\C-Drive\cpwin\DATA1
Number of Entries: 58

	Samplename	Code	ID	FileName	Method	Samp Wt	DF	Int Std	Batch Number	Analysis
51	5822076 RI ACF	T	AB	3c6310.51R	PESTD.MET	30	25	1	093070028A	01216
52	5822077 RI ACF DF100	T	AC	3c6310.52R	PESTD.MET	30	1000	1	093070028A	01216
53	5822077 RI ACF DF100	T	AD	3c6310.53R	PESTD.MET	30	10000	1	093070028A	01216
54	CONDITIONER	MISC	AA	3c6310.54R	PESTD.MET	1	1	1	093070028A	
55	CONDITIONER	MISC	AA	3c6310.55R	PESTD.MET	1	1	1	093070028A	
56	CONDITIONER	MISC	AA	3c6310.56R	PESTD.MET	1	1	1	093070028A	
57	AR1620924A	CCAL	OT	3c6310.57R	PESTD.MET	1	1	1	0931299999	
58	IBLKX0924B	PIBLK	GI	3c6310.58R	PESTD.MET	1000	10	1	0931299999	

ASX22 0565

Set-up by [illegible] Date: 11-10-09

11/10/09

Sample Name: **IBLKX0924B AAPIBLKAA PIBLK0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 03:48:57pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.04R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.232		0.0000	0.000	391.7	0.659	BB	0.061	106.28	0.205
2	3.310 TCX		201.3141	56.321	39282.7	66.135	BB	0.015	43396.96	83.596
3	3.512		0.0000	0.000	609.5	1.026	BB	0.016	646.77	1.246
4	3.647		0.0000	0.000	321.5	0.541	BB	0.046	115.79	0.223
5	3.980		0.0000	0.000	84.2	0.142	BB	0.015	91.48	0.176
6	4.072		0.0000	0.000	96.1	0.162	BB	0.020	81.53	0.157
7	4.780		0.0000	0.000	288.7	0.486	BB	0.044	110.02	0.212
8	5.141		0.0000	0.000	161.1	0.271	BB	0.024	111.42	0.215
9	5.442		0.0000	0.000	206.8	0.348	BB	0.037	93.32	0.180
10	6.011		0.0000	0.000	115.8	0.195	BB	0.029	66.77	0.129
11	6.984		0.0000	0.000	160.2	0.270	BB	0.033	80.87	0.156
12	7.159 DCB		156.1253	43.679	17679.5	29.765	BB	0.042	7011.45	13.506

Total Area = 59397.6, Total Amount = 357.439, Total Height = 51912.66, Sample Units = PPB

ASX22 0566

Sample Name: **IBLKX0924B AAPIBLKAA PIBLK0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 03:48:57pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.04R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.612		0.0000	0.000	338.9	0.201	BV	0.043	132.53	0.080
2	2.699		0.0000	0.000	1097.2	0.652	VB	0.017	1048.44	0.635
3	2.807	TCX	233.4886	50.352	88436.0	52.532	BB	0.014	104439.20	63.265
4	3.029		0.0000	0.000	1292.6	0.768	BB	0.014	1559.40	0.945
5	3.147		0.0000	0.000	266.9	0.159	BB	0.014	321.36	0.195
6	3.300		0.0000	0.000	162.6	0.097	BB	0.043	62.94	0.038
7	3.494		0.0000	0.000	193.3	0.115	BV	0.012	263.97	0.160
8	3.530		0.0000	0.000	615.5	0.366	VB	0.016	641.05	0.388
9	3.761		0.0000	0.000	222.6	0.132	BB	0.026	140.10	0.085
10	3.812		0.0000	0.000	1141.5	0.678	BB	0.125	152.58	0.092
11	3.931		0.0000	0.000	1755.8	1.043	BV	0.105	279.58	0.169
12	4.118		0.0000	0.000	295.4	0.175	VB	0.037	131.45	0.080
13	4.306		0.0000	0.000	1351.3	0.803	BB	0.059	380.25	0.230
14	4.445		0.0000	0.000	656.7	0.390	BB	0.070	155.98	0.094
15	4.692		0.0000	0.000	785.4	0.467	BV	0.026	503.12	0.305
16	4.834		0.0000	0.000	914.1	0.543	VB	0.036	427.20	0.259
17	5.131		0.0000	0.000	31.3	0.019	BB	0.018	29.30	0.018
18	5.548		0.0000	0.000	248.4	0.148	BB	0.024	170.66	0.103
19	5.653		0.0000	0.000	291.4	0.173	BB	0.041	119.77	0.073
20	5.963		0.0000	0.000	211.6	0.126	BB	0.101	34.77	0.021
21	6.090		0.0000	0.000	133.2	0.079	BB	0.035	64.33	0.039
22	6.199		0.0000	0.000	639.9	0.380	BB	0.020	535.31	0.324
23	6.358	DCB	230.2278	49.648	66875.3	39.725	BB	0.021	53339.45	32.311
24	6.753		0.0000	0.000	112.4	0.067	BB	0.028	67.90	0.041
25	6.797		0.0000	0.000	52.9	0.031	BB	0.025	35.88	0.022
26	7.184		0.0000	0.000	224.7	0.133	BB	0.080	46.61	0.028

Total Area = 168346.9, Total Amount = 463.716, Total Height = 165083.1, Sample Units = PPB

ASX22 8567

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.04R

IBLKX0924B AAPIBLKAA PIBLK0930999999

C:\CPWIN\DATA1\1C6310.04R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 3:41:27 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.04R

IBLKX0924B AAPIBLKAA PIBLK0930999999

C:\CPWIN\DATA1\1C6310B.04R



ASX22 #568

Instrument ID: CP06--S9147B Injected On: 11/6/2009 3:41:27 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1000 Dilution Factor: 10
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.31	43397	.20131	TCX	2.807	104439	.23349	TCX
7.159	7011	.15613	DCB	6.358	53339	.23023	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.04A
Area File: C:\CPWIN\DATA1\1C6310B.04A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 3:49:02 PM
File Reported On: 11/6/2009 at 3:49:11 PM

REPLOTS

ASX22 0569

IBLKX0924B AAPIBLKAA PIBLK0930999999
C:\CPWIN\DATA1\1C6310.04R



IBLKX0924B AAPIBLKAA PIBLK0930999999
C:\CPWIN\DATA1\1C6310B.04R



ASX22 0578

Sample Name: **AR1610924A AAAR161AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 04:00:59pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.05R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.149		0.0000	0.000	69.4	0.105	BB	0.032	35.90	0.070
2	3.223		0.0000	0.000	156.5	0.236	BB	0.044	59.49	0.116
3	3.311	TCX	4.5854	56.708	9365.5	14.110	BB	0.016	9884.64	19.286
4	3.457		0.0000	0.000	388.1	0.585	BB	0.018	362.25	0.707
5	3.515		0.0000	0.000	477.9	0.720	BB	0.014	554.70	1.082
6	3.564		0.0000	0.000	1758.4	2.649	BB	0.015	1936.64	3.779
7	3.662		0.0000	0.000	317.4	0.478	BB	0.016	324.50	0.633
8	3.788		0.0000	0.000	3608.0	5.436	BB	0.023	2571.34	5.017
9	3.846		0.0000	0.000	198.6	0.299	BB	0.015	215.59	0.421
10	3.914		0.0000	0.000	2264.5	3.412	BB	0.017	2231.51	4.354
11	4.033		0.0000	0.000	746.4	1.124	BB	0.024	516.10	1.007
12	4.112		0.0000	0.000	4876.2	7.347	BB	0.028	2894.69	5.648
13	4.155		0.0000	0.000	1656.3	2.495	BB	0.017	1636.92	3.194
14	4.221		0.0000	0.000	1504.5	2.267	BB	0.017	1472.59	2.873
15	4.265		0.0000	0.000	64.1	0.097	BB	0.012	91.29	0.178
16	4.292		0.0000	0.000	1005.6	1.515	BB	0.015	1101.43	2.149
17	4.323		0.0000	0.000	1591.8	2.398	BB	0.029	927.84	1.810
18	4.432		0.0000	0.000	1107.1	1.668	BB	0.016	1122.76	2.191
19	4.459		0.0000	0.000	221.2	0.333	BB	0.012	312.37	0.609
20	4.493		0.0000	0.000	306.1	0.461	BB	0.015	338.92	0.661
21	4.532		0.0000	0.000	1158.0	1.745	BB	0.020	941.90	1.838
22	4.579		0.0000	0.000	351.0	0.529	BB	0.017	344.56	0.672
23	4.669		0.0000	0.000	35.2	0.053	BB	0.011	51.51	0.100
24	4.708		0.0000	0.000	1655.1	2.494	BV	0.025	1117.04	2.179
25	4.759		0.0000	0.000	609.7	0.919	VV	0.035	286.71	0.559
26	4.827		0.0000	0.000	212.9	0.321	VV	0.022	160.87	0.314
27	4.877		0.0000	0.000	1478.5	2.228	VB	0.031	803.57	1.568
28	5.107		0.0000	0.000	605.2	0.912	BV	0.024	421.99	0.823
29	5.153		0.0000	0.000	380.8	0.574	VB	0.022	285.02	0.556
30	5.203		0.0000	0.000	1786.1	2.691	BB	0.031	973.88	1.900
31	5.284		0.0000	0.000	2488.0	3.748	BB	0.023	1784.09	3.481
32	5.372		0.0000	0.000	214.3	0.323	BB	0.020	175.23	0.342
33	5.431		0.0000	0.000	351.1	0.529	BB	0.022	266.05	0.519
34	5.485		0.0000	0.000	2529.2	3.810	BB	0.027	1567.22	3.058
35	5.539		0.0000	0.000	394.5	0.594	BB	0.015	434.96	0.849
36	5.578		0.0000	0.000	1046.5	1.577	BB	0.025	690.92	1.348
37	5.693		0.0000	0.000	2218.9	3.343	BB	0.023	1600.54	3.123
38	5.755		0.0000	0.000	1312.1	1.977	BB	0.019	1160.90	2.265
39	5.798		0.0000	0.000	701.3	1.057	BB	0.020	587.61	1.146
40	5.920		0.0000	0.000	2086.4	3.143	BB	0.027	1284.06	2.505
41	5.978		0.0000	0.000	448.6	0.676	BB	0.017	438.83	0.856
42	6.016		0.0000	0.000	300.6	0.453	BB	0.018	272.14	0.531
43	6.097		0.0000	0.000	296.7	0.447	BB	0.019	263.54	0.514
44	6.152		0.0000	0.000	4735.7	7.135	BB	0.032	2429.79	4.741
45	6.357		0.0000	0.000	1508.2	2.272	BB	0.021	1205.80	2.353
46	6.405		0.0000	0.000	759.0	1.143	BB	0.027	472.78	0.922
47	6.626		0.0000	0.000	449.5	0.677	BB	0.026	284.69	0.555
48	6.775		0.0000	0.000	1059.1	1.596	BB	0.028	624.51	1.218
49	6.990		0.0000	0.000	315.8	0.476	BB	0.033	159.97	0.312
50	7.160	DCB	3.5005	43.292	3203.3	4.826	BB	0.034	1572.04	3.067

Total Area = 66374.6, Total Amount = 8.086, Total Height = 51254.18, Sample Units = PPB

ASX22 8571

Sample Name: **AR1610924A AAAR161AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 04:00:59pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.05R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.607		0.0000	0.000	461.4	0.177	BV	0.057	134.29	0.056
2	2.702		0.0000	0.000	494.4	0.190	VB	0.029	283.71	0.118
3	2.808 TCX		5.1147	51.182	18606.6	7.155	BB	0.014	22877.87	9.536
4	2.853		0.0000	0.000	211.6	0.081	BB	0.009	377.82	0.157
5	2.929		0.0000	0.000	121.5	0.047	BB	0.013	160.70	0.067
6	2.983		0.0000	0.000	776.1	0.298	BB	0.013	960.93	0.401
7	3.045		0.0000	0.000	1060.2	0.408	BB	0.016	1090.37	0.455
8	3.098		0.0000	0.000	4098.8	1.576	BB	0.013	5124.06	2.136
9	3.212		0.0000	0.000	681.8	0.262	BB	0.014	807.87	0.337
10	3.308		0.0000	0.000	9201.2	3.538	BV	0.020	7592.66	3.165
11	3.360		0.0000	0.000	1093.0	0.420	VV	0.019	948.35	0.395
12	3.405		0.0000	0.000	376.9	0.145	VV	0.013	468.10	0.195
13	3.432		0.0000	0.000	5797.9	2.229	VV	0.014	6881.89	2.869
14	3.467		0.0000	0.000	225.5	0.087	VV	0.014	260.80	0.109
15	3.513		0.0000	0.000	2644.0	1.017	VB	0.022	1988.13	0.829
16	3.588		0.0000	0.000	15343.3	5.900	BV	0.021	11916.80	4.967
17	3.624		0.0000	0.000	8268.4	3.179	VV	0.021	6654.28	2.774
18	3.701		0.0000	0.000	5261.4	2.023	VV	0.014	6253.93	2.607
19	3.736		0.0000	0.000	4854.5	1.867	VV	0.015	5345.77	2.228
20	3.761		0.0000	0.000	6987.9	2.687	VB	0.020	5893.66	2.457
21	3.870		0.0000	0.000	6844.5	2.632	BB	0.019	5872.71	2.448
22	3.918		0.0000	0.000	1220.1	0.469	BV	0.013	1618.13	0.674
23	3.949		0.0000	0.000	986.2	0.379	VV	0.012	1393.37	0.581
24	3.965		0.0000	0.000	3685.7	1.417	VV	0.015	4227.21	1.762
25	4.011		0.0000	0.000	1396.2	0.537	VB	0.015	1510.57	0.630
26	4.055		0.0000	0.000	141.7	0.054	BV	0.014	165.27	0.069
27	4.080		0.0000	0.000	831.0	0.320	VV	0.014	1004.21	0.419
28	4.104		0.0000	0.000	4815.2	1.852	VV	0.015	5401.65	2.252
29	4.129		0.0000	0.000	1552.2	0.597	VV	0.021	1211.25	0.505
30	4.169		0.0000	0.000	525.8	0.202	VV	0.016	561.89	0.234
31	4.207		0.0000	0.000	3483.0	1.339	VV	0.016	3618.28	1.508
32	4.260		0.0000	0.000	619.7	0.238	VV	0.021	498.94	0.208
33	4.306		0.0000	0.000	816.3	0.314	VV	0.039	345.85	0.144
34	4.415		0.0000	0.000	244.4	0.094	VB	0.012	334.95	0.140
35	4.463		0.0000	0.000	1484.9	0.571	BV	0.015	1621.78	0.676
36	4.493		0.0000	0.000	4534.2	1.743	VB	0.016	4612.23	1.923
37	4.522		0.0000	0.000	1493.7	0.574	BB	0.016	1580.16	0.659
38	4.575		0.0000	0.000	9883.4	3.800	BV	0.017	9428.74	3.930
39	4.645		0.0000	0.000	246.2	0.095	VV	0.013	304.88	0.127
40	4.671		0.0000	0.000	1754.5	0.675	VV	0.016	1802.19	0.751
41	4.714		0.0000	0.000	11957.4	4.598	VB	0.020	9853.57	4.107
42	4.822		0.0000	0.000	11175.0	4.297	BV	0.020	9115.07	3.799
43	4.879		0.0000	0.000	1169.9	0.450	VB	0.017	1162.22	0.484
44	4.917		0.0000	0.000	1955.7	0.752	BB	0.015	2109.46	0.879
45	4.953		0.0000	0.000	9471.3	3.642	BV	0.019	8186.40	3.412
46	4.986		0.0000	0.000	8284.7	3.186	VB	0.018	7832.93	3.265
47	5.026		0.0000	0.000	3396.7	1.306	BB	0.015	3653.34	1.523
48	5.128		0.0000	0.000	1520.0	0.584	BB	0.018	1379.57	0.575
49	5.180		0.0000	0.000	9652.5	3.712	BV	0.020	8031.29	3.348
50	5.253		0.0000	0.000	4433.4	1.705	VV	0.018	4027.76	1.679
51	5.288		0.0000	0.000	1912.8	0.736	VV	0.018	1768.14	0.737
52	5.314		0.0000	0.000	1197.6	0.461	VB	0.016	1243.27	0.518
53	5.368		0.0000	0.000	18681.0	7.183	BB	0.021	14779.34	6.160
54	5.445		0.0000	0.000	552.9	0.213	BB	0.016	592.61	0.247
55	5.558		0.0000	0.000	197.3	0.076	BB	0.023	145.70	0.061
56	5.583		0.0000	0.000	3862.7	1.485	BB	0.016	3946.99	1.645
57	5.627		0.0000	0.000	13201.6	5.076	BB	0.023	9755.66	4.066
58	5.783		0.0000	0.000	205.5	0.079	BB	0.017	197.54	0.082
59	5.821		0.0000	0.000	120.7	0.046	BB	0.015	133.97	0.056
60	5.852		0.0000	0.000	147.2	0.057	BB	0.017	143.81	0.060
61	5.898		0.0000	0.000	2392.5	0.920	BB	0.021	1913.15	0.797
62	6.001		0.0000	0.000	5840.6	2.246	BB	0.024	4119.24	1.717

ASX22 0552

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.200		0.0000	0.000	1444.2	0.555	BB	0.020	1185.54	0.494
64	6.360	DCB	4.8785	48.818	13821.7	5.315	BB	0.020	11302.58	4.711
65	6.648		0.0000	0.000	111.8	0.043	BB	0.039	47.58	0.020
66	6.756		0.0000	0.000	132.0	0.051	BB	0.023	94.52	0.039
67	7.297		0.0000	0.000	102.5	0.039	BB	0.035	48.15	0.020

Total Area = 260066.5, Total Amount = 9.993, Total Height = 239905.6, Sample Units = PPB

ASX22 0573

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.05R

AR1610924A AAAR161AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.05R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 3:53:29 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.05R

AR1610924A AAAR161AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.05R



ASX22 6574

Instrument ID: CP06--S9147B Injected On: 11/6/2009 3:53:29 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	9885	4.58538	TCX	2.808	22878	5.11467	TCX
7.16	1572	3.5005	DCB	6.36	11303	4.87851	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.05A
Area File: C:\CPWIN\DATA1\1C6310B.05A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 4:01:04 PM
File Reported On: 11/6/2009 at 4:01:15 PM

REPLOTS

ASX22 0575

AR1610924A AAAR161AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.05R



AR1610924A AAAR161AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.05R



ASX22 8576

Sample Name: **AR1620924A AAAR162AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 04:12:59pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.06R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.091		0.0000	0.000	40.1	0.026	BB	0.020	33.97	0.028
2	3.221		0.0000	0.000	286.8	0.188	BB	0.047	100.81	0.084
3	3.311	TCX	18.6456	54.848	38323.5	25.076	BB	0.016	40131.78	33.534
4	3.458		0.0000	0.000	639.3	0.418	BB	0.017	628.78	0.525
5	3.515		0.0000	0.000	1129.2	0.739	BB	0.015	1280.92	1.070
6	3.565		0.0000	0.000	3131.5	2.049	BB	0.015	3466.82	2.897
7	3.662		0.0000	0.000	629.9	0.412	BB	0.018	596.48	0.498
8	3.788		0.0000	0.000	6566.7	4.297	BB	0.024	4634.54	3.873
9	3.847		0.0000	0.000	725.7	0.475	BB	0.025	478.97	0.400
10	3.915		0.0000	0.000	3888.3	2.544	BB	0.016	4086.61	3.415
11	4.033		0.0000	0.000	1447.7	0.947	BB	0.024	991.82	0.829
12	4.112		0.0000	0.000	8989.3	5.882	BB	0.028	5306.81	4.434
13	4.155		0.0000	0.000	3022.0	1.977	BB	0.017	2967.84	2.480
14	4.221		0.0000	0.000	2742.7	1.795	BB	0.017	2690.60	2.248
15	4.266		0.0000	0.000	112.5	0.074	BB	0.012	151.44	0.127
16	4.292		0.0000	0.000	1798.8	1.177	BB	0.015	1966.98	1.644
17	4.324		0.0000	0.000	2830.4	1.852	BB	0.028	1711.81	1.430
18	4.432		0.0000	0.000	1941.5	1.270	BB	0.016	2000.35	1.672
19	4.460		0.0000	0.000	356.6	0.233	BB	0.011	546.17	0.456
20	4.494		0.0000	0.000	539.4	0.353	BB	0.014	639.10	0.534
21	4.531		0.0000	0.000	2062.0	1.349	BB	0.019	1764.87	1.475
22	4.578		0.0000	0.000	567.5	0.371	BB	0.017	556.49	0.465
23	4.708		0.0000	0.000	3057.5	2.001	BV	0.026	1977.19	1.652
24	4.757		0.0000	0.000	1054.6	0.690	VV	0.032	552.99	0.462
25	4.828		0.0000	0.000	395.8	0.259	VV	0.025	261.03	0.218
26	4.876		0.0000	0.000	2895.1	1.894	VB	0.032	1487.68	1.243
27	5.107		0.0000	0.000	1101.0	0.720	BV	0.024	764.53	0.639
28	5.153		0.0000	0.000	650.2	0.425	VB	0.021	514.37	0.430
29	5.203		0.0000	0.000	3269.1	2.139	BB	0.031	1764.30	1.474
30	5.283		0.0000	0.000	4530.9	2.965	BB	0.023	3243.08	2.710
31	5.376		0.0000	0.000	379.0	0.248	BB	0.020	314.75	0.263
32	5.433		0.0000	0.000	601.0	0.393	BB	0.022	455.70	0.381
33	5.485		0.0000	0.000	4642.4	3.038	BB	0.026	2976.88	2.487
34	5.538		0.0000	0.000	757.1	0.495	BB	0.017	745.89	0.623
35	5.578		0.0000	0.000	1956.1	1.280	BB	0.025	1311.98	1.096
36	5.657		0.0000	0.000	145.7	0.095	BB	0.016	148.66	0.124
37	5.693		0.0000	0.000	4076.4	2.667	BB	0.024	2880.72	2.407
38	5.755		0.0000	0.000	2403.3	1.573	BB	0.019	2149.74	1.796
39	5.797		0.0000	0.000	1284.2	0.840	BB	0.019	1108.05	0.926
40	5.920		0.0000	0.000	3781.6	2.474	BB	0.027	2295.14	1.918
41	5.978		0.0000	0.000	1004.8	0.657	BB	0.018	946.09	0.791
42	6.015		0.0000	0.000	460.3	0.301	BB	0.017	461.22	0.385
43	6.098		0.0000	0.000	524.3	0.343	BB	0.019	465.71	0.389
44	6.153		0.0000	0.000	8589.6	5.620	BB	0.032	4509.43	3.768
45	6.357		0.0000	0.000	2842.5	1.860	BB	0.021	2258.24	1.887
46	6.405		0.0000	0.000	1308.2	0.856	BB	0.026	850.20	0.710
47	6.626		0.0000	0.000	975.1	0.638	BB	0.030	547.30	0.457
48	6.776		0.0000	0.000	2053.2	1.343	BB	0.029	1163.63	0.972
49	6.989		0.0000	0.000	602.6	0.394	BB	0.030	336.35	0.281
50	7.161	DCB	15.3498	45.152	15717.6	10.284	BB	0.041	6448.83	5.389

Total Area = 152830.5, Total Amount = 33.995, Total Height = 119673.6, Sample Units = PPB

ASX22 8577

Sample Name: **AR1620924A AAAR162AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 04:12:59pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.06R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.613		0.0000	0.000	517.1	0.092	BV	0.047	182.58	0.034
2	2.701		0.0000	0.000	1081.7	0.192	VB	0.022	824.92	0.155
3	2.808 TCX		20.3680	49.838	84835.5	15.071	BB	0.015	94601.83	17.826
4	2.929		0.0000	0.000	100.1	0.018	BB	0.011	151.58	0.029
5	2.983		0.0000	0.000	1488.1	0.264	BB	0.014	1783.17	0.336
6	3.030		0.0000	0.000	493.0	0.088	BB	0.010	822.66	0.155
7	3.046		0.0000	0.000	815.0	0.145	BB	0.009	1497.23	0.282
8	3.098		0.0000	0.000	7463.1	1.326	BB	0.014	9193.52	1.732
9	3.149		0.0000	0.000	179.4	0.032	BB	0.011	266.42	0.050
10	3.213		0.0000	0.000	1198.1	0.213	BB	0.013	1510.34	0.285
11	3.309		0.0000	0.000	16739.8	2.974	BV	0.020	13865.74	2.613
12	3.361		0.0000	0.000	1450.7	0.258	VV	0.014	1754.01	0.331
13	3.382		0.0000	0.000	583.2	0.104	VV	0.012	778.78	0.147
14	3.406		0.0000	0.000	712.4	0.127	VV	0.013	892.88	0.168
15	3.432		0.0000	0.000	10935.0	1.943	VV	0.015	12475.27	2.351
16	3.514		0.0000	0.000	4764.5	0.846	VB	0.021	3747.92	0.706
17	3.588		0.0000	0.000	28250.1	5.019	BV	0.021	22437.21	4.228
18	3.625		0.0000	0.000	15114.8	2.685	VV	0.021	12101.50	2.280
19	3.701		0.0000	0.000	9723.5	1.727	VV	0.014	11393.51	2.147
20	3.736		0.0000	0.000	8862.7	1.574	VV	0.015	9944.87	1.874
21	3.761		0.0000	0.000	13047.7	2.318	VB	0.020	10795.81	2.034
22	3.871		0.0000	0.000	7824.7	1.390	BB	0.013	10158.91	1.914
23	3.895		0.0000	0.000	1806.4	0.321	BB	0.009	3294.07	0.621
24	3.918		0.0000	0.000	1963.5	0.349	BB	0.012	2836.93	0.535
25	3.950		0.0000	0.000	1670.3	0.297	BV	0.011	2499.47	0.471
26	3.966		0.0000	0.000	6754.2	1.200	VV	0.014	7953.77	1.499
27	4.011		0.0000	0.000	2527.0	0.449	VB	0.015	2814.58	0.530
28	4.056		0.0000	0.000	255.8	0.045	BV	0.014	312.42	0.059
29	4.080		0.0000	0.000	1568.2	0.279	VV	0.014	1901.99	0.358
30	4.104		0.0000	0.000	8829.8	1.569	VV	0.015	10062.92	1.896
31	4.129		0.0000	0.000	2820.6	0.501	VV	0.021	2281.59	0.430
32	4.170		0.0000	0.000	1008.9	0.179	VV	0.016	1072.64	0.202
33	4.207		0.0000	0.000	6347.3	1.128	VV	0.016	6675.58	1.258
34	4.244		0.0000	0.000	577.8	0.103	VV	0.015	626.60	0.118
35	4.261		0.0000	0.000	1118.8	0.199	VV	0.021	891.81	0.168
36	4.306		0.0000	0.000	630.8	0.112	VV	0.028	375.10	0.071
37	4.359		0.0000	0.000	269.9	0.048	VB	0.019	239.49	0.045
38	4.416		0.0000	0.000	633.9	0.113	BB	0.014	753.68	0.142
39	4.463		0.0000	0.000	2832.4	0.503	BV	0.015	3143.97	0.592
40	4.494		0.0000	0.000	8394.9	1.491	VB	0.016	8498.79	1.601
41	4.522		0.0000	0.000	2822.3	0.501	BB	0.016	3029.27	0.571
42	4.575		0.0000	0.000	18189.4	3.231	BV	0.017	17419.56	3.282
43	4.644		0.0000	0.000	534.1	0.095	VV	0.015	587.89	0.111
44	4.671		0.0000	0.000	3369.5	0.599	VV	0.016	3438.21	0.648
45	4.714		0.0000	0.000	21738.7	3.862	VB	0.020	18390.02	3.465
46	4.822		0.0000	0.000	20495.1	3.641	BV	0.020	16679.71	3.143
47	4.879		0.0000	0.000	2192.5	0.389	VB	0.017	2208.67	0.416
48	4.917		0.0000	0.000	3788.4	0.673	BB	0.015	4123.10	0.777
49	4.953		0.0000	0.000	18047.0	3.206	BV	0.019	15453.46	2.912
50	4.986		0.0000	0.000	15389.4	2.734	VB	0.017	14705.42	2.771
51	5.026		0.0000	0.000	6485.1	1.152	BB	0.015	7052.06	1.329
52	5.127		0.0000	0.000	3057.4	0.543	BB	0.019	2716.80	0.512
53	5.180		0.0000	0.000	18026.7	3.202	BV	0.020	14846.19	2.798
54	5.252		0.0000	0.000	8328.1	1.479	VV	0.019	7440.39	1.402
55	5.288		0.0000	0.000	3642.7	0.647	VV	0.018	3434.16	0.647
56	5.314		0.0000	0.000	2302.9	0.409	VB	0.016	2421.02	0.456
57	5.368		0.0000	0.000	35246.3	6.262	BB	0.021	28328.94	5.338
58	5.444		0.0000	0.000	1032.9	0.183	BB	0.015	1184.17	0.223
59	5.582		0.0000	0.000	7750.2	1.377	BB	0.017	7418.67	1.398
60	5.628		0.0000	0.000	25375.3	4.508	BB	0.023	18791.85	3.541
61	5.782		0.0000	0.000	347.8	0.062	BB	0.015	385.20	0.073
62	5.821		0.0000	0.000	227.7	0.040	BB	0.013	282.15	0.053

ASX22 8578

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.850		0.0000	0.000	251.6	0.045	BB	0.015	284.44	0.054
64	5.898		0.0000	0.000	4435.6	0.788	BB	0.020	3650.24	0.688
65	6.001		0.0000	0.000	11075.0	1.967	BB	0.023	8133.22	1.533
66	6.200		0.0000	0.000	2925.0	0.520	BB	0.019	2554.79	0.481
67	6.361	DCB	20.5003	50.162	59300.5	10.535	BB	0.021	48138.05	9.071
68	6.750		0.0000	0.000	86.2	0.015	BB	0.025	57.73	0.011
69	6.803		0.0000	0.000	82.7	0.015	BB	0.030	46.05	0.009
70	6.960		0.0000	0.000	33.2	0.006	BB	0.025	22.33	0.004
71	7.173		0.0000	0.000	133.0	0.024	BB	0.046	47.87	0.009

Total Area = 562903.1, Total Amount = 40.868, Total Height = 530687.8, Sample Units = PPB

ASX22 0579

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.06R

AR1620924A AAAR162AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.06R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 4:05:29 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.06R

AR1620924A AAAR162AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.06R



ASX22 0580

Instrument ID: CP06--S9147B Injected On: 11/6/2009 4:05:29 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min\Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	40132	18.64563	TCX	2.808	94602	20.36802	TCX
7.161	6449	15.34976	DCB	6.361	48138	20.50034	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.06A
Area File: C:\CPWIN\DATA1\1C6310B.06A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 4:13:04 PM
File Reported On: 11/6/2009 at 4:13:14 PM

REPLOTS

ASX22 0581

AR1620924A AAAR162AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.06R



AR1620924A AAAR162AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.06R



ASX22 0582

Sample Name: **AR1630924A AAAR163AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 04:24:55pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.07R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.219		0.0000	0.000	192.9	0.067	BB	0.036	88.21	0.039
2	3.311	TCX	36.6033	53.685	74766.0	25.799	BB	0.016	77543.09	33.867
3	3.457		0.0000	0.000	1249.9	0.431	BB	0.017	1237.96	0.541
4	3.515		0.0000	0.000	2184.2	0.754	BB	0.015	2451.71	1.071
5	3.565		0.0000	0.000	6039.1	2.084	BB	0.016	6445.21	2.815
6	3.661		0.0000	0.000	1241.6	0.428	BB	0.018	1141.38	0.498
7	3.788		0.0000	0.000	12470.4	4.303	BB	0.024	8803.10	3.845
8	3.846		0.0000	0.000	1331.6	0.459	BB	0.025	887.93	0.388
9	3.914		0.0000	0.000	7395.2	2.552	BB	0.016	7734.30	3.378
10	3.994		0.0000	0.000	230.1	0.079	BB	0.020	188.09	0.082
11	4.033		0.0000	0.000	2615.0	0.902	BB	0.024	1840.69	0.804
12	4.112		0.0000	0.000	17366.1	5.992	BB	0.028	10320.64	4.507
13	4.155		0.0000	0.000	5718.2	1.973	BB	0.017	5736.71	2.505
14	4.221		0.0000	0.000	5197.6	1.794	BB	0.017	5168.50	2.257
15	4.266		0.0000	0.000	205.5	0.071	BB	0.013	268.11	0.117
16	4.291		0.0000	0.000	3415.4	1.179	BB	0.015	3885.88	1.697
17	4.323		0.0000	0.000	5403.9	1.865	BB	0.028	3220.48	1.407
18	4.432		0.0000	0.000	3766.3	1.300	BB	0.017	3752.66	1.639
19	4.459		0.0000	0.000	708.3	0.244	BB	0.011	1084.17	0.474
20	4.493		0.0000	0.000	1011.0	0.349	BB	0.014	1181.92	0.516
21	4.532		0.0000	0.000	4005.0	1.382	BB	0.020	3299.60	1.441
22	4.578		0.0000	0.000	1083.1	0.374	BB	0.017	1069.04	0.467
23	4.670		0.0000	0.000	129.9	0.045	BB	0.015	139.72	0.061
24	4.707		0.0000	0.000	5750.2	1.984	BV	0.025	3812.79	1.665
25	4.756		0.0000	0.000	1820.5	0.628	VV	0.029	1038.28	0.453
26	4.827		0.0000	0.000	651.7	0.225	VV	0.025	436.94	0.191
27	4.876		0.0000	0.000	4875.3	1.682	VB	0.030	2698.47	1.179
28	5.106		0.0000	0.000	2101.4	0.725	BV	0.024	1446.66	0.632
29	5.153		0.0000	0.000	1165.8	0.402	VB	0.020	953.46	0.416
30	5.203		0.0000	0.000	4174.8	1.441	BB	0.023	3025.50	1.321
31	5.240		0.0000	0.000	285.5	0.099	BB	0.012	390.94	0.171
32	5.283		0.0000	0.000	8533.9	2.945	BB	0.023	6198.19	2.707
33	5.373		0.0000	0.000	623.3	0.215	BB	0.019	540.19	0.236
34	5.431		0.0000	0.000	982.8	0.339	BB	0.019	848.32	0.371
35	5.484		0.0000	0.000	8883.2	3.065	BB	0.026	5595.91	2.444
36	5.538		0.0000	0.000	1368.8	0.472	BB	0.017	1373.81	0.600
37	5.578		0.0000	0.000	3621.5	1.250	BB	0.025	2406.56	1.051
38	5.657		0.0000	0.000	253.9	0.088	BB	0.016	256.48	0.112
39	5.692		0.0000	0.000	7744.1	2.672	BB	0.023	5508.61	2.406
40	5.754		0.0000	0.000	4574.4	1.578	BB	0.019	4091.02	1.787
41	5.798		0.0000	0.000	2402.2	0.829	BB	0.019	2083.89	0.910
42	5.919		0.0000	0.000	7191.2	2.481	BB	0.026	4545.77	1.985
43	5.976		0.0000	0.000	1953.9	0.674	BB	0.018	1813.25	0.792
44	6.015		0.0000	0.000	741.6	0.256	BB	0.017	731.30	0.319
45	6.096		0.0000	0.000	1025.0	0.354	BB	0.019	906.42	0.396
46	6.152		0.0000	0.000	16269.9	5.614	BB	0.031	8858.56	3.869
47	6.356		0.0000	0.000	5313.9	1.834	BB	0.021	4256.70	1.859
48	6.405		0.0000	0.000	2458.0	0.848	BB	0.027	1514.59	0.661
49	6.625		0.0000	0.000	1838.0	0.634	BB	0.029	1046.90	0.457
50	6.775		0.0000	0.000	4187.2	1.445	BB	0.033	2143.81	0.936
51	6.989		0.0000	0.000	1098.7	0.379	BB	0.028	645.93	0.282
52	7.160	DCB	31.5789	46.315	30179.9	10.414	BB	0.041	12307.60	5.375

Total Area = 289797.1, Total Amount = 68.182, Total Height = 228966.0, Sample Units = PPB

ASX22 583

Sample Name: **AR1630924A AAAR163AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 04:24:55pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.07R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.612		0.0000	0.000	240.5	0.022	BV	0.037	109.29	0.011
2	2.700		0.0000	0.000	1076.4	0.099	VB	0.017	1030.23	0.101
3	2.808	TCX	39.6035	49.023	167678.7	15.345	BB	0.015	184135.80	17.995
4	2.983		0.0000	0.000	2848.4	0.261	BB	0.013	3543.01	0.346
5	3.045		0.0000	0.000	5213.8	0.477	BB	0.022	4021.11	0.393
6	3.098		0.0000	0.000	14269.3	1.306	BB	0.013	17700.31	1.730
7	3.148		0.0000	0.000	363.5	0.033	BB	0.012	523.10	0.051
8	3.212		0.0000	0.000	2387.4	0.218	BB	0.013	2947.69	0.288
9	3.309		0.0000	0.000	31807.0	2.911	BV	0.020	26522.50	2.592
10	3.360		0.0000	0.000	2941.1	0.269	VV	0.014	3383.83	0.331
11	3.382		0.0000	0.000	1115.4	0.102	VV	0.012	1532.56	0.150
12	3.405		0.0000	0.000	1381.3	0.126	VV	0.013	1729.44	0.169
13	3.432		0.0000	0.000	20317.6	1.859	VV	0.014	23872.24	2.333
14	3.467		0.0000	0.000	733.4	0.067	VV	0.014	871.16	0.085
15	3.495		0.0000	0.000	305.2	0.028	VV	0.012	424.80	0.042
16	3.513		0.0000	0.000	9325.0	0.853	VB	0.022	7093.66	0.693
17	3.588		0.0000	0.000	54512.0	4.989	BV	0.022	42214.04	4.125
18	3.624		0.0000	0.000	28637.7	2.621	VV	0.020	23459.72	2.293
19	3.700		0.0000	0.000	18780.1	1.719	VV	0.014	22304.11	2.180
20	3.736		0.0000	0.000	17003.2	1.556	VV	0.015	18912.12	1.848
21	3.761		0.0000	0.000	24596.1	2.251	VB	0.020	20778.91	2.031
22	3.871		0.0000	0.000	22968.6	2.102	BB	0.018	20783.74	2.031
23	3.917		0.0000	0.000	3847.1	0.352	BB	0.011	5685.17	0.556
24	3.949		0.0000	0.000	3488.1	0.319	BV	0.012	5044.88	0.493
25	3.965		0.0000	0.000	13130.8	1.202	VV	0.014	15285.48	1.494
26	4.010		0.0000	0.000	4960.6	0.454	VB	0.015	5541.05	0.541
27	4.056		0.0000	0.000	553.0	0.051	BV	0.014	659.36	0.064
28	4.080		0.0000	0.000	3140.6	0.287	VV	0.014	3770.52	0.368
29	4.104		0.0000	0.000	16906.2	1.547	VV	0.014	19453.21	1.901
30	4.129		0.0000	0.000	5316.7	0.487	VV	0.020	4401.01	0.430
31	4.169		0.0000	0.000	1826.2	0.167	VV	0.015	2043.48	0.200
32	4.207		0.0000	0.000	12223.0	1.119	VV	0.016	12784.13	1.249
33	4.243		0.0000	0.000	1084.8	0.099	VV	0.015	1202.26	0.117
34	4.260		0.0000	0.000	2125.7	0.195	VV	0.021	1677.80	0.164
35	4.306		0.0000	0.000	655.5	0.060	VV	0.028	386.39	0.038
36	4.358		0.0000	0.000	385.7	0.035	VB	0.016	397.86	0.039
37	4.417		0.0000	0.000	1629.1	0.149	BB	0.014	1899.90	0.186
38	4.463		0.0000	0.000	5392.8	0.494	BV	0.015	5873.33	0.574
39	4.493		0.0000	0.000	15910.0	1.456	VB	0.016	16502.65	1.613
40	4.521		0.0000	0.000	5697.8	0.521	BB	0.016	5809.13	0.568
41	4.575		0.0000	0.000	34766.1	3.182	BV	0.018	32835.74	3.209
42	4.644		0.0000	0.000	1011.9	0.093	VV	0.014	1216.07	0.119
43	4.670		0.0000	0.000	6565.4	0.601	VV	0.016	6653.56	0.650
44	4.714		0.0000	0.000	40892.1	3.742	VB	0.019	35336.33	3.453
45	4.821		0.0000	0.000	38836.2	3.554	BV	0.020	32109.61	3.138
46	4.879		0.0000	0.000	4071.3	0.373	VB	0.016	4165.77	0.407
47	4.916		0.0000	0.000	7150.0	0.654	BB	0.016	7560.93	0.739
48	4.952		0.0000	0.000	34838.2	3.188	BV	0.020	29660.37	2.899
49	4.985		0.0000	0.000	29251.6	2.677	VB	0.017	27962.61	2.733
50	5.025		0.0000	0.000	12576.6	1.151	BB	0.015	13544.14	1.324
51	5.127		0.0000	0.000	5877.7	0.538	BB	0.018	5447.88	0.532
52	5.179		0.0000	0.000	34337.3	3.142	BV	0.020	28757.50	2.810
53	5.252		0.0000	0.000	15919.6	1.457	VV	0.019	14322.11	1.400
54	5.287		0.0000	0.000	6573.8	0.602	VV	0.017	6632.90	0.648
55	5.313		0.0000	0.000	4855.9	0.444	VB	0.017	4696.31	0.459
56	5.368		0.0000	0.000	68066.7	6.229	BB	0.020	55924.63	5.465
57	5.443		0.0000	0.000	2181.1	0.200	BB	0.015	2423.11	0.237
58	5.558		0.0000	0.000	270.3	0.025	BB	0.012	385.43	0.038
59	5.582		0.0000	0.000	14006.5	1.282	BB	0.016	14380.95	1.405
60	5.626		0.0000	0.000	49126.9	4.496	BB	0.022	37031.73	3.619
61	5.781		0.0000	0.000	685.0	0.063	BB	0.016	729.15	0.071
62	5.820		0.0000	0.000	469.4	0.043	BB	0.014	555.63	0.054

ASX22 0504

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.849		0.0000	0.000	438.3	0.040	BB	0.014	539.28	0.053
64	5.897		0.0000	0.000	8397.6	0.769	BB	0.020	7004.27	0.684
65	6.000		0.0000	0.000	21281.7	1.948	BB	0.023	15661.08	1.530
66	6.199		0.0000	0.000	5723.2	0.524	BB	0.020	4874.66	0.476
67	6.359	DCB	41.1826	50.977	117531.6	10.756	BB	0.020	96410.52	9.422
68	6.750		0.0000	0.000	134.9	0.012	BB	0.028	79.80	0.008
69	6.996		0.0000	0.000	62.8	0.006	BB	0.022	48.34	0.005
70	7.353		0.0000	0.000	22.9	0.002	BB	0.015	25.80	0.003

Total Area = 1092698.0, Total Amount = 80.786, Total Height = 1023287.0, Sample Units = PPB

ASX22. 0585

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.07R
AR1630924A AAAR163AA ICAL 0930999999
——— C:\CPWIN\DATA1\1C6310.07R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 4:17:25 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.07R
AR1630924A AAAR163AA ICAL 0930999999
——— C:\CPWIN\DATA1\1C6310B.07R



ASX22 0586

Instrument ID: CP06--S9147B Injected On: 11/6/2009 4:17:25 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	77543	36.60335	TCX	2.808	184136	39.60349	TCX
7.16	12308	31.57893	DCB	6.359	96411	41.18259	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.07A
Area File: C:\CPWIN\DATA1\1C6310B.07A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 4:25:02 PM
File Reported On: 11/6/2009 at 4:25:11 PM

REPLOTS

ASX22 8587

AR1630924A AAAR163AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.07R



AR1630924A AAAR163AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.07R



ASX22 8588

Sample Name: **AR1640924A AAAR164AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 04:36:54pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.08R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.218		0.0000	0.000	153.8	0.026	BB	0.029	87.12	0.019
2	3.312 TCX		53.9879	52.225	108375.8	18.449	BB	0.016	113177.80	24.857
3	3.458		0.0000	0.000	2931.8	0.499	BB	0.017	2848.42	0.626
4	3.516		0.0000	0.000	4192.3	0.714	BB	0.015	4714.88	1.036
5	3.565		0.0000	0.000	13580.7	2.312	BB	0.015	14810.57	3.253
6	3.662		0.0000	0.000	2469.2	0.420	BB	0.016	2565.03	0.563
7	3.788		0.0000	0.000	28282.6	4.815	BB	0.024	19615.62	4.308
8	3.846		0.0000	0.000	3024.9	0.515	BB	0.025	2008.63	0.441
9	3.914		0.0000	0.000	17040.0	2.901	BB	0.016	17490.71	3.841
10	3.994		0.0000	0.000	516.3	0.088	BB	0.018	465.84	0.102
11	4.033		0.0000	0.000	6446.5	1.097	BB	0.025	4293.23	0.943
12	4.113		0.0000	0.000	39932.3	6.798	BB	0.028	23766.62	5.220
13	4.156		0.0000	0.000	13124.6	2.234	BB	0.017	13188.62	2.897
14	4.221		0.0000	0.000	11954.7	2.035	BB	0.016	12081.13	2.653
15	4.266		0.0000	0.000	442.6	0.075	BB	0.012	623.07	0.137
16	4.292		0.0000	0.000	7745.9	1.319	BB	0.015	8737.64	1.919
17	4.324		0.0000	0.000	12080.3	2.056	BB	0.027	7326.13	1.609
18	4.433		0.0000	0.000	8670.9	1.476	BB	0.016	8989.32	1.974
19	4.460		0.0000	0.000	1608.0	0.274	BB	0.012	2312.77	0.508
20	4.494		0.0000	0.000	2227.4	0.379	BB	0.014	2615.94	0.575
21	4.532		0.0000	0.000	9224.1	1.570	BB	0.020	7624.43	1.675
22	4.579		0.0000	0.000	2389.4	0.407	BB	0.017	2399.28	0.527
23	4.670		0.0000	0.000	278.9	0.047	BB	0.014	325.43	0.071
24	4.708		0.0000	0.000	13150.8	2.239	BV	0.025	8650.55	1.900
25	4.756		0.0000	0.000	4211.3	0.717	VV	0.029	2411.84	0.530
26	4.828		0.0000	0.000	1385.1	0.236	VV	0.024	981.00	0.215
27	4.876		0.0000	0.000	10796.2	1.838	VV	0.028	6343.66	1.393
28	5.030		0.0000	0.000	259.1	0.044	VB	0.021	209.43	0.046
29	5.107		0.0000	0.000	4798.6	0.817	BV	0.024	3328.32	0.731
30	5.154		0.0000	0.000	2589.5	0.441	VB	0.019	2232.09	0.490
31	5.203		0.0000	0.000	14740.3	2.509	BB	0.031	8023.81	1.762
32	5.283		0.0000	0.000	20146.4	3.430	BB	0.023	14675.88	3.223
33	5.373		0.0000	0.000	1344.3	0.229	BB	0.018	1261.37	0.277
34	5.432		0.0000	0.000	2225.7	0.379	BB	0.019	1935.77	0.425
35	5.485		0.0000	0.000	21459.7	3.653	BB	0.026	13688.03	3.006
36	5.539		0.0000	0.000	3338.9	0.568	BB	0.016	3499.05	0.768
37	5.578		0.0000	0.000	8851.8	1.507	BB	0.024	6118.81	1.344
38	5.657		0.0000	0.000	645.9	0.110	BB	0.017	644.99	0.142
39	5.693		0.0000	0.000	18808.1	3.202	BB	0.024	13165.88	2.892
40	5.756		0.0000	0.000	11374.5	1.936	BB	0.019	9790.16	2.150
41	5.797		0.0000	0.000	5844.7	0.995	BB	0.018	5360.82	1.177
42	5.919		0.0000	0.000	17383.6	2.959	BB	0.027	10925.13	2.399
43	5.977		0.0000	0.000	4693.7	0.799	BB	0.018	4397.25	0.966
44	6.015		0.0000	0.000	1689.9	0.288	BB	0.015	1836.29	0.403
45	6.057		0.0000	0.000	242.5	0.041	BB	0.012	331.57	0.073
46	6.097		0.0000	0.000	2451.3	0.417	BB	0.020	2094.02	0.460
47	6.152		0.0000	0.000	38551.5	6.563	BB	0.032	20381.40	4.476
48	6.357		0.0000	0.000	12891.1	2.195	BB	0.021	10047.01	2.207
49	6.405		0.0000	0.000	5978.5	1.018	BB	0.027	3640.90	0.800
50	6.541		0.0000	0.000	287.3	0.049	BB	0.020	241.96	0.053
51	6.625		0.0000	0.000	3980.7	0.678	BB	0.027	2419.72	0.531
52	6.775		0.0000	0.000	9740.9	1.658	BB	0.033	4924.45	1.082
53	6.989		0.0000	0.000	2501.7	0.426	BB	0.028	1502.01	0.330
54	7.160 DCB		49.3872	47.775	44366.4	7.553	BB	0.041	18185.75	3.994

Total Area = 587422.5, Total Amount = 103.375, Total Height = 455317.3, Sample Units = PPB

ASX22 0589

Sample Name: **AR1640924A AAAR164AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 04:36:54pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.08R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.608		0.0000	0.000	275.4	0.012	BV	0.037	125.19	0.006
2	2.703		0.0000	0.000	406.0	0.018	VV	0.025	273.68	0.013
3	2.764		0.0000	0.000	96.6	0.004	VB	0.016	103.45	0.005
4	2.809 TCX		57.8153	49.288	224464.0	9.818	BB	0.014	271897.90	12.561
5	2.853		0.0000	0.000	1764.6	0.077	BB	0.010	2805.92	0.130
6	2.983		0.0000	0.000	6785.5	0.297	BB	0.013	8407.03	0.388
7	3.031		0.0000	0.000	1417.7	0.062	BB	0.010	2380.80	0.110
8	3.046		0.0000	0.000	4371.6	0.191	BB	0.010	7371.91	0.341
9	3.098		0.0000	0.000	32151.5	1.406	BB	0.013	40357.61	1.864
10	3.149		0.0000	0.000	559.2	0.024	BB	0.011	830.08	0.038
11	3.213		0.0000	0.000	5482.7	0.240	BB	0.013	7017.38	0.324
12	3.309		0.0000	0.000	71464.9	3.126	BV	0.020	59664.94	2.756
13	3.361		0.0000	0.000	6894.9	0.302	VV	0.014	8092.60	0.374
14	3.382		0.0000	0.000	2727.4	0.119	VV	0.012	3694.82	0.171
15	3.405		0.0000	0.000	3294.0	0.144	VV	0.013	4213.10	0.195
16	3.433		0.0000	0.000	46252.8	2.023	VV	0.014	55435.98	2.561
17	3.467		0.0000	0.000	1835.2	0.080	VV	0.015	2106.34	0.097
18	3.514		0.0000	0.000	20688.0	0.905	VB	0.021	16791.99	0.776
19	3.589		0.0000	0.000	127085.6	5.559	BV	0.021	99758.00	4.609
20	3.625		0.0000	0.000	64078.8	2.803	VV	0.020	52435.38	2.422
21	3.701		0.0000	0.000	43251.7	1.892	VV	0.014	50793.97	2.347
22	3.736		0.0000	0.000	38952.8	1.704	VV	0.015	43466.30	2.008
23	3.761		0.0000	0.000	57445.1	2.513	VB	0.020	47918.35	2.214
24	3.871		0.0000	0.000	53067.8	2.321	BB	0.019	47761.12	2.206
25	3.918		0.0000	0.000	8930.0	0.391	BB	0.011	13047.29	0.603
26	3.950		0.0000	0.000	7790.9	0.341	BV	0.011	11710.21	0.541
27	3.966		0.0000	0.000	31172.0	1.363	VV	0.014	36445.68	1.684
28	4.011		0.0000	0.000	11622.8	0.508	VB	0.015	12859.53	0.594
29	4.056		0.0000	0.000	1311.3	0.057	BV	0.014	1511.94	0.070
30	4.080		0.0000	0.000	7317.4	0.320	VV	0.014	8896.93	0.411
31	4.104		0.0000	0.000	39356.4	1.721	VV	0.015	45059.17	2.082
32	4.129		0.0000	0.000	12385.5	0.542	VV	0.020	10197.45	0.471
33	4.170		0.0000	0.000	4333.6	0.190	VV	0.016	4655.44	0.215
34	4.207		0.0000	0.000	28694.5	1.255	VV	0.016	30058.09	1.389
35	4.244		0.0000	0.000	2503.0	0.109	VV	0.015	2796.74	0.129
36	4.261		0.0000	0.000	5041.9	0.221	VV	0.022	3878.85	0.179
37	4.359		0.0000	0.000	937.6	0.041	VB	0.016	972.88	0.045
38	4.417		0.0000	0.000	4358.5	0.191	BB	0.015	4864.40	0.225
39	4.464		0.0000	0.000	12759.8	0.558	BV	0.015	13805.67	0.638
40	4.494		0.0000	0.000	39029.7	1.707	VB	0.017	39204.03	1.811
41	4.522		0.0000	0.000	13052.7	0.571	BB	0.016	13638.12	0.630
42	4.575		0.0000	0.000	80893.8	3.538	BV	0.017	78276.23	3.616
43	4.645		0.0000	0.000	2615.4	0.114	VV	0.015	2913.82	0.135
44	4.671		0.0000	0.000	15356.1	0.672	VV	0.016	16345.90	0.755
45	4.714		0.0000	0.000	96118.2	4.204	VB	0.019	83768.54	3.870
46	4.822		0.0000	0.000	91312.1	3.994	BV	0.020	77207.54	3.567
47	4.879		0.0000	0.000	9657.0	0.422	VB	0.016	9849.86	0.455
48	4.917		0.0000	0.000	17488.9	0.765	BB	0.015	18832.08	0.870
49	4.953		0.0000	0.000	83526.5	3.653	BV	0.019	72388.70	3.344
50	4.986		0.0000	0.000	69528.7	3.041	VB	0.017	66448.62	3.070
51	5.025		0.0000	0.000	30376.4	1.329	BB	0.016	32081.11	1.482
52	5.127		0.0000	0.000	14248.5	0.623	BB	0.018	13369.06	0.618
53	5.180		0.0000	0.000	80765.8	3.533	BV	0.020	68510.35	3.165
54	5.253		0.0000	0.000	37646.6	1.647	VV	0.018	34550.15	1.596
55	5.288		0.0000	0.000	15614.2	0.683	VV	0.016	16231.10	0.750
56	5.314		0.0000	0.000	11868.7	0.519	VB	0.018	11156.92	0.515
57	5.368		0.0000	0.000	161312.3	7.056	BB	0.020	135537.20	6.262
58	5.444		0.0000	0.000	5216.9	0.228	BB	0.015	5898.67	0.273
59	5.560		0.0000	0.000	597.7	0.026	BB	0.019	532.41	0.025
60	5.583		0.0000	0.000	33693.3	1.474	BB	0.016	34708.23	1.603
61	5.627		0.0000	0.000	118297.0	5.174	BB	0.023	87410.43	4.038
62	5.782		0.0000	0.000	1695.4	0.074	BB	0.015	1846.56	0.085

ASX22 8598

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.821		0.0000	0.000	1141.0	0.050	BB	0.015	1304.71	0.060
64	5.850		0.0000	0.000	1091.6	0.048	BB	0.014	1284.40	0.059
65	5.898		0.0000	0.000	20001.2	0.875	BB	0.019	17125.65	0.791
66	6.001		0.0000	0.000	51009.1	2.231	BB	0.023	37112.57	1.715
67	6.200		0.0000	0.000	13036.8	0.570	BB	0.020	11086.28	0.512
68	6.360	DCB	59.4868	50.712	176282.9	7.710	BB	0.021	141265.90	6.526
69	6.745		0.0000	0.000	109.0	0.005	BB	0.032	57.47	0.003
70	6.801		0.0000	0.000	272.9	0.012	BB	0.037	122.71	0.006
71	7.171		0.0000	0.000	113.4	0.005	BB	0.027	71.08	0.003

Total Area = 2286299.0, Total Amount = 117.302, Total Height = 2164600.0, Sample Units = PPB

ASX22 0591

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.08R

AR1640924A AAAR164AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.08R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 4:29:24 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.08R

AR1640924A AAAR164AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.08R



ASX22 8592

Instrument ID: CP06--S9147B Injected On: 11/6/2009 4:29:24 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Valve Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.312	113178	53.98792	TCX	2.809	271898	57.8153	TCX
7.16	18186	49.38717	DCB	6.36	141266	59.48684	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.08A
Area File: C:\CPWIN\DATA1\1C6310B.08A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 4:37:00 PM
File Reported On: 11/6/2009 at 4:37:10 PM

REPLOTS

ASX22 0593

AR1640924A AAAR164AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.08R



AR1640924A AAAR164AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.08R



ASX22 0594

Sample Name: **AR1650924A AAAR165AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 04:48:53pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.09R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.220		0.0000	0.000	232.1	0.026	BB	0.032	119.70	0.016
2	3.312 TCX		67.4828	50.332	133406.3	14.674	BB	0.016	136617.20	18.563
3	3.457		0.0000	0.000	5195.2	0.571	BB	0.017	4951.12	0.673
4	3.515		0.0000	0.000	6791.8	0.747	BB	0.015	7531.53	1.023
5	3.564		0.0000	0.000	23797.6	2.618	BB	0.015	25796.96	3.505
6	3.662		0.0000	0.000	4234.0	0.466	BB	0.016	4367.07	0.593
7	3.788		0.0000	0.000	49487.4	5.444	BB	0.024	34108.88	4.635
8	3.846		0.0000	0.000	2717.2	0.299	BB	0.015	3020.69	0.410
9	3.879		0.0000	0.000	489.3	0.054	BB	0.008	998.89	0.136
10	3.914		0.0000	0.000	30183.8	3.320	BB	0.016	31347.66	4.259
11	3.994		0.0000	0.000	530.8	0.058	BB	0.013	666.61	0.091
12	4.033		0.0000	0.000	11691.8	1.286	BB	0.025	7878.86	1.071
13	4.101		0.0000	0.000	8451.7	0.930	BB	0.019	7588.15	1.031
14	4.112		0.0000	0.000	10669.7	1.174	BB	0.009	19724.42	2.680
15	4.155		0.0000	0.000	24059.6	2.647	BB	0.017	23864.70	3.243
16	4.221		0.0000	0.000	22195.7	2.441	BB	0.017	21847.64	2.969
17	4.266		0.0000	0.000	786.9	0.087	BB	0.012	1074.05	0.146
18	4.292		0.0000	0.000	14043.9	1.545	BB	0.015	15624.84	2.123
19	4.323		0.0000	0.000	22598.7	2.486	BB	0.028	13285.47	1.805
20	4.432		0.0000	0.000	15819.1	1.740	BB	0.016	16162.10	2.196
21	4.459		0.0000	0.000	2940.4	0.323	BB	0.011	4596.44	0.625
22	4.493		0.0000	0.000	4364.5	0.480	BB	0.014	5074.59	0.690
23	4.532		0.0000	0.000	17466.1	1.921	BB	0.020	14703.25	1.998
24	4.578		0.0000	0.000	4560.7	0.502	BB	0.017	4571.41	0.621
25	4.624		0.0000	0.000	256.2	0.028	BB	0.012	360.31	0.049
26	4.668		0.0000	0.000	572.4	0.063	BB	0.015	619.18	0.084
27	4.708		0.0000	0.000	24635.4	2.710	BV	0.026	15562.71	2.115
28	4.756		0.0000	0.000	7169.5	0.789	VV	0.026	4546.11	0.618
29	4.827		0.0000	0.000	2527.5	0.278	VV	0.024	1727.54	0.235
30	4.876		0.0000	0.000	19691.9	2.166	VV	0.027	12030.15	1.635
31	5.030		0.0000	0.000	500.7	0.055	VB	0.023	361.25	0.049
32	5.107		0.0000	0.000	8840.0	0.972	BV	0.025	5924.48	0.8[illegible]
33	5.153		0.0000	0.000	4853.6	0.534	VB	0.019	4274.48	0.581
34	5.203		0.0000	0.000	17961.3	1.976	BB	0.023	13156.17	1.788
35	5.241		0.0000	0.000	1233.9	0.136	BB	0.011	1789.66	0.243
36	5.283		0.0000	0.000	36364.5	4.000	BB	0.022	27125.09	3.686
37	5.373		0.0000	0.000	2462.0	0.271	BB	0.018	2291.48	0.311
38	5.431		0.0000	0.000	3989.8	0.439	BB	0.019	3537.90	0.481
39	5.484		0.0000	0.000	39074.7	4.298	BB	0.026	25508.17	3.466
40	5.538		0.0000	0.000	6124.3	0.674	BB	0.016	6476.91	0.880
41	5.577		0.0000	0.000	16418.3	1.806	BB	0.024	11376.98	1.546
42	5.656		0.0000	0.000	1274.6	0.140	BB	0.017	1240.28	0.169
43	5.692		0.0000	0.000	34567.7	3.802	BB	0.023	25188.33	3.422
44	5.755		0.0000	0.000	19963.6	2.196	BB	0.019	17730.34	2.409
45	5.798		0.0000	0.000	10501.6	1.155	BB	0.019	9291.35	1.262
46	5.919		0.0000	0.000	31788.5	3.497	BB	0.027	19981.32	2.715
47	5.977		0.0000	0.000	8718.1	0.959	BB	0.019	7851.26	1.067
48	6.015		0.0000	0.000	3166.9	0.348	BB	0.016	3323.59	0.452
49	6.056		0.0000	0.000	352.0	0.039	BB	0.012	481.51	0.065
50	6.096		0.0000	0.000	4402.7	0.484	BB	0.019	3884.95	0.528
51	6.151		0.0000	0.000	69823.8	7.680	BB	0.029	39459.41	5.362
52	6.356		0.0000	0.000	22683.9	2.495	BB	0.020	18715.46	2.543
53	6.406		0.0000	0.000	11497.0	1.265	BB	0.026	7416.69	1.008
54	6.542		0.0000	0.000	448.1	0.049	BB	0.017	430.45	0.058
55	6.626		0.0000	0.000	7101.9	0.781	BB	0.028	4246.10	0.577
56	6.776		0.0000	0.000	17010.9	1.871	BB	0.031	9252.87	1.257
57	6.990		0.0000	0.000	4152.0	0.457	BB	0.029	2380.42	0.323
58	7.161 DCB		66.5932	49.668	52264.6	5.749	BB	0.038	22902.50	3.112

Not Used
See Reintegration

Total Area = 909108.0, Total Amount = 134.076, Total Height = 735967.6, Sample Units = PPB

ASX22 8595

Sample Name: **AR1650924A AAAR165AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 04:48:53pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.09R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.607		0.0000	0.000	298.6	0.008	BV	0.037	133.18	0.004
2	2.701		0.0000	0.000	441.8	0.012	VB	0.022	336.59	0.009
3	2.763		0.0000	0.000	96.7	0.003	BB	0.012	135.38	0.004
4	2.808	TCX	65.7286	47.839	271033.1	7.108	BB	0.015	306404.70	8.516
5	2.852		0.0000	0.000	2312.2	0.061	BB	0.009	4458.67	0.124
6	2.982		0.0000	0.000	12099.8	0.317	BB	0.014	14344.83	0.399
7	3.030		0.0000	0.000	1128.2	0.030	BB	0.012	1558.99	0.043
8	3.045		0.0000	0.000	8548.9	0.224	BB	0.011	12847.59	0.357
9	3.098		0.0000	0.000	56509.6	1.482	BB	0.014	69005.74	1.918
10	3.148		0.0000	0.000	592.8	0.016	BB	0.011	916.20	0.025
11	3.212		0.0000	0.000	9625.7	0.252	BB	0.013	12005.09	0.334
12	3.308		0.0000	0.000	124466.5	3.264	BV	0.020	102523.70	2.850
13	3.360		0.0000	0.000	17162.9	0.450	VV	0.020	14071.03	0.391
14	3.405		0.0000	0.000	6095.3	0.160	VV	0.014	7438.06	0.207
15	3.432		0.0000	0.000	81326.7	2.133	VV	0.014	94506.58	2.627
16	3.467		0.0000	0.000	3216.7	0.084	VV	0.014	3794.91	0.105
17	3.513		0.0000	0.000	36889.7	0.967	VB	0.020	30099.10	0.837
18	3.588		0.0000	0.000	222380.0	5.832	BV	0.022	171904.90	4.778
19	3.625		0.0000	0.000	115094.8	3.019	VV	0.021	91870.64	2.553
20	3.701		0.0000	0.000	77699.9	2.038	VV	0.014	90882.59	2.526
21	3.736		0.0000	0.000	69108.0	1.812	VV	0.015	74907.88	2.082
22	3.761		0.0000	0.000	100304.8	2.631	VB	0.020	83587.27	2.323
23	3.870		0.0000	0.000	71684.2	1.880	BV	0.014	84833.91	2.358
24	3.894		0.0000	0.000	22770.9	0.597	VB	0.012	30857.05	0.858
25	3.917		0.0000	0.000	16028.6	0.420	BB	0.011	24039.13	0.668
26	3.949		0.0000	0.000	15510.0	0.407	BV	0.012	21473.88	0.597
27	3.965		0.0000	0.000	56144.2	1.472	VV	0.014	65805.83	1.829
28	4.010		0.0000	0.000	21016.6	0.551	VB	0.015	23545.26	0.654
29	4.055		0.0000	0.000	2533.2	0.066	BV	0.015	2827.67	0.079
30	4.079		0.0000	0.000	14126.5	0.370	VV	0.014	16664.09	0.463
31	4.103		0.0000	0.000	70054.4	1.837	VV	0.015	80148.41	2.228
32	4.129		0.0000	0.000	21602.5	0.567	VV	0.019	18728.35	0.521
33	4.169		0.0000	0.000	7388.1	0.194	VV	0.015	8380.08	0.233
34	4.206		0.0000	0.000	50327.5	1.320	VV	0.016	53244.87	1.480
35	4.243		0.0000	0.000	4916.7	0.129	VV	0.017	4798.15	0.133
36	4.260		0.0000	0.000	9894.4	0.259	VV	0.024	6814.48	0.189
37	4.358		0.0000	0.000	1645.3	0.043	VB	0.016	1747.53	0.049
38	4.416		0.0000	0.000	7994.7	0.210	BB	0.015	9182.81	0.255
39	4.463		0.0000	0.000	20674.0	0.542	BV	0.014	24584.81	0.683
40	4.493		0.0000	0.000	71421.6	1.873	VB	0.017	68360.70	1.900
41	4.521		0.0000	0.000	24866.2	0.652	BB	0.016	25300.77	0.703
42	4.574		0.0000	0.000	140756.7	3.692	BV	0.017	135980.90	3.780
43	4.643		0.0000	0.000	4769.6	0.125	VV	0.015	5474.86	0.152
44	4.670		0.0000	0.000	27614.5	0.724	VV	0.016	29208.29	0.812
45	4.713		0.0000	0.000	167470.5	4.392	VB	0.019	150311.40	4.178
46	4.821		0.0000	0.000	159187.9	4.175	BV	0.019	136563.20	3.796
47	4.878		0.0000	0.000	17112.3	0.449	VB	0.016	17484.80	0.486
48	4.916		0.0000	0.000	31325.2	0.822	BV	0.016	33274.52	0.925
49	4.952		0.0000	0.000	150778.1	3.954	VV	0.020	127827.50	3.553
50	4.985		0.0000	0.000	119762.9	3.141	VB	0.017	117176.10	3.257
51	5.025		0.0000	0.000	54881.3	1.439	BB	0.016	58479.17	1.625
52	5.127		0.0000	0.000	26134.3	0.685	BB	0.018	24428.75	0.679
53	5.179		0.0000	0.000	142241.6	3.730	BV	0.020	121369.30	3.373
54	5.252		0.0000	0.000	67143.9	1.761	VV	0.018	61071.29	1.697
55	5.287		0.0000	0.000	28210.3	0.740	VV	0.016	28574.21	0.794
56	5.314		0.0000	0.000	21101.7	0.553	VB	0.017	20226.80	0.562
57	5.368		0.0000	0.000	285627.1	7.491	BB	0.020	238134.10	6.619
58	5.443		0.0000	0.000	8844.8	0.232	BB	0.014	10324.06	0.287
59	5.558		0.0000	0.000	1214.0	0.032	BB	0.012	1707.84	0.047
60	5.581		0.0000	0.000	60140.8	1.577	BB	0.016	62427.02	1.735
61	5.627		0.0000	0.000	210723.8	5.526	BB	0.022	160111.60	4.450
62	5.781		0.0000	0.000	2938.0	0.077	BB	0.015	3182.63	0.088

Not Used
See Reintegration

ASX22. 0596

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.820		0.0000	0.000	2039.7	0.053	BB	0.014	2350.07	0.065
64	5.849		0.0000	0.000	1777.8	0.047	BB	0.013	2275.60	0.063
65	5.897		0.0000	0.000	35693.7	0.936	BB	0.019	31793.76	0.884
66	6.000		0.0000	0.000	91110.4	2.389	BB	0.022	68649.24	1.908
67	6.199		0.0000	0.000	22482.6	0.590	BB	0.019	19327.61	0.537
68	6.360	DCB	71.6670	52.161	204299.2	5.358	BB	0.020	170656.80	4.743
69	6.528		0.0000	0.000	111.5	0.003	BB	0.011	166.28	0.005
70	6.797		0.0000	0.000	104.1	0.003	BB	0.031	55.60	0.002
71	6.989		0.0000	0.000	257.9	0.007	BB	0.051	83.54	0.002
72	7.171		0.0000	0.000	84.2	0.002	BB	0.019	75.11	0.002

Total Area = 3812973.0, Total Amount = 137.396, Total Height = 3597842.0, Sample Units = PPB

Not Used
See Reintegration

ASX22 0597

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.09R

AR1650924A AAAR165AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.09R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 4:41:23 PM Column ID: RTX CLP

Not Used
See Reintegration

File: C:\CPWIN\DATA1\1C6310B.09R

AR1650924A AAAR165AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.09R



ASX22 0598

Instrument ID: CP06--S9147B Injected On: 11/6/2009 4:41:23 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/minMultiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.312	136617	67.48285	TCX	2.808	306405	65.72858	TCX
7.161	22902	66.59321	DCB	6.36	170657	71.66696	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.09A
Area File: C:\CPWIN\DATA1\1C6310B.09A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 4:48:58 PM
File Reported On: 11/6/2009 at 4:49:08 PM

REPLOTS

Not Used
See Reintegration

ASX22: 0599

AR1650924A AAAR165AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.09R



AR1650924A AAAR165AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.09R



ASX22 0688

Lancaster Laboratories

Sample Name: **AR1650924A AAAR165AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 04:48:53pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.09R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.220		0.0000	0.000	232.1	0.024	BB	0.032	119.70	0.015
2	3.312	TCX	69.8416	49.175	133406.3	13.699	BB	0.016	136617.20	17.169
3	3.457		0.0000	0.000	5195.2	0.533	BB	0.017	4951.12	0.622
4	3.515		0.0000	0.000	6791.8	0.697	BB	0.015	7531.53	0.947
5	3.564		0.0000	0.000	23797.6	2.444	BB	0.015	25796.96	3.242
6	3.662		0.0000	0.000	4234.0	0.435	BB	0.016	4367.07	0.549
7	3.788		0.0000	0.000	49487.4	5.082	BB	0.024	34108.88	4.287
8	3.846		0.0000	0.000	2717.2	0.279	BB	0.015	3020.69	0.380
9	3.879		0.0000	0.000	489.3	0.050	BB	0.008	998.89	0.126
10	3.914		0.0000	0.000	30183.8	3.100	BB	0.016	31347.66	3.940
11	3.994		0.0000	0.000	530.8	0.055	BB	0.013	666.61	0.084
12	4.033		0.0000	0.000	11691.8	1.201	BB	0.025	7878.86	0.990
13	4.101		0.0000	0.000	31407.8	3.225	BV	0.013	39124.26	4.917
14	4.112		0.0000	0.000	47454.6	4.873	VV	0.017	45863.87	5.764
15	4.155		0.0000	0.000	29027.8	2.981	VB	0.019	25932.40	3.259
16	4.221		0.0000	0.000	22195.7	2.279	BB	0.017	21847.64	2.746
17	4.266		0.0000	0.000	786.9	0.081	BB	0.012	1074.05	0.135
18	4.292		0.0000	0.000	14043.9	1.442	BB	0.015	15624.84	1.964
19	4.323		0.0000	0.000	22598.7	2.321	BB	0.028	13285.47	1.670
20	4.432		0.0000	0.000	15819.1	1.624	BB	0.016	16162.10	2.031
21	4.459		0.0000	0.000	2940.4	0.302	BB	0.011	4596.44	0.578
22	4.493		0.0000	0.000	4364.5	0.448	BB	0.014	5074.59	0.638
23	4.532		0.0000	0.000	17466.1	1.794	BB	0.020	14703.25	1.848
24	4.578		0.0000	0.000	4560.7	0.468	BB	0.017	4571.41	0.575
25	4.624		0.0000	0.000	256.2	0.026	BB	0.012	360.31	0.045
26	4.668		0.0000	0.000	572.4	0.059	BB	0.015	619.18	0.078
27	4.708		0.0000	0.000	24635.4	2.530	BV	0.026	15562.71	1.956
28	4.756		0.0000	0.000	7169.5	0.736	VV	0.026	4546.11	0.571
29	4.827		0.0000	0.000	2527.5	0.260	VV	0.024	1727.54	0.217
30	4.876		0.0000	0.000	19691.9	2.022	VV	0.027	12030.15	1.512
31	5.030		0.0000	0.000	500.7	0.051	VB	0.023	361.25	0.045
32	5.107		0.0000	0.000	8840.0	0.908	BV	0.025	5924.48	0.745
33	5.153		0.0000	0.000	4853.6	0.498	VB	0.019	4274.48	0.537
34	5.203		0.0000	0.000	17961.3	1.844	BB	0.023	13156.17	1.653
35	5.241		0.0000	0.000	1233.9	0.127	BB	0.011	1789.66	0.225
36	5.283		0.0000	0.000	36364.5	3.734	BB	0.022	27125.09	3.409
37	5.373		0.0000	0.000	2462.0	0.253	BB	0.018	2291.48	0.288
38	5.431		0.0000	0.000	3989.8	0.410	BB	0.019	3537.90	0.445
39	5.484		0.0000	0.000	39074.7	4.013	BB	0.026	25508.17	3.206
40	5.538		0.0000	0.000	6124.3	0.629	BB	0.016	6476.91	0.814
41	5.577		0.0000	0.000	16418.3	1.686	BB	0.024	11376.98	1.430
42	5.656		0.0000	0.000	1274.6	0.131	BB	0.017	1240.28	0.156
43	5.692		0.0000	0.000	34567.7	3.550	BB	0.023	25188.33	3.166
44	5.755		0.0000	0.000	19963.6	2.050	BB	0.019	17730.34	2.228
45	5.798		0.0000	0.000	10501.6	1.078	BB	0.019	9291.35	1.168
46	5.919		0.0000	0.000	31788.5	3.264	BB	0.027	19981.32	2.511
47	5.977		0.0000	0.000	8718.1	0.895	BB	0.019	7851.26	0.987
48	6.015		0.0000	0.000	3166.9	0.325	BB	0.016	3323.59	0.418
49	6.056		0.0000	0.000	352.0	0.036	BB	0.012	481.51	0.061
50	6.096		0.0000	0.000	4402.7	0.452	BB	0.019	3884.95	0.488
51	6.151		0.0000	0.000	69823.8	7.170	BB	0.029	39459.41	4.959
52	6.356		0.0000	0.000	22683.9	2.329	BB	0.020	18715.46	2.352
53	6.406		0.0000	0.000	11497.0	1.181	BB	0.026	7416.69	0.932
54	6.542		0.0000	0.000	448.1	0.046	BB	0.017	430.45	0.054
55	6.626		0.0000	0.000	7101.9	0.729	BB	0.028	4246.10	0.534
56	6.776		0.0000	0.000	17010.9	1.747	BB	0.031	9252.87	1.163
57	6.990		0.0000	0.000	4152.0	0.426	BB	0.029	2380.42	0.299
58	7.161	DCB	72.1864	50.825	52264.6	5.367	BB	0.038	22902.50	2.878

Total Area = 973817.2, Total Amount = 142.028, Total Height = 795710.9, Sample Units = PPB

ASX22. 0601

Printed on 11/7/2009 6:22:46 AM

Sample Name: **AR1650924A AAAR165AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 04:48:53pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.09R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.607		0.0000	0.000	298.6	0.008	BV	0.037	133.18	0.004
2	2.701		0.0000	0.000	441.8	0.012	VB	0.022	336.59	0.009
3	2.763		0.0000	0.000	96.7	0.003	BB	0.012	135.38	0.004
4	2.808 TCX		66.9677	48.362	271033.1	7.108	BB	0.015	306404.70	8.516
5	2.852		0.0000	0.000	2312.2	0.061	BB	0.009	4458.67	0.124
6	2.982		0.0000	0.000	12099.8	0.317	BB	0.014	14344.83	0.399
7	3.030		0.0000	0.000	1128.2	0.030	BB	0.012	1558.99	0.043
8	3.045		0.0000	0.000	8548.9	0.224	BB	0.011	12847.59	0.357
9	3.098		0.0000	0.000	56509.6	1.482	BB	0.014	69005.74	1.918
10	3.148		0.0000	0.000	592.8	0.016	BB	0.011	916.20	0.025
11	3.212		0.0000	0.000	9625.7	0.252	BB	0.013	12005.09	0.334
12	3.308		0.0000	0.000	124466.5	3.264	BV	0.020	102523.70	2.850
13	3.360		0.0000	0.000	17162.9	0.450	VV	0.020	14071.03	0.391
14	3.405		0.0000	0.000	6095.3	0.160	VV	0.014	7438.06	0.207
15	3.432		0.0000	0.000	81326.7	2.133	VV	0.014	94506.58	2.627
16	3.467		0.0000	0.000	3216.7	0.084	VV	0.014	3794.91	0.105
17	3.513		0.0000	0.000	36889.7	0.967	VB	0.020	30099.10	0.837
18	3.588		0.0000	0.000	222380.0	5.832	BV	0.022	171904.90	4.778
19	3.625		0.0000	0.000	115094.8	3.019	VV	0.021	91870.64	2.553
20	3.701		0.0000	0.000	77699.9	2.038	VV	0.014	90882.59	2.526
21	3.736		0.0000	0.000	69108.0	1.812	VV	0.015	74907.88	2.082
22	3.761		0.0000	0.000	100304.8	2.631	VB	0.020	83587.27	2.323
23	3.870		0.0000	0.000	71684.2	1.880	BV	0.014	84833.91	2.358
24	3.894		0.0000	0.000	22770.9	0.597	VB	0.012	30857.05	0.858
25	3.917		0.0000	0.000	16028.6	0.420	BB	0.011	24039.13	0.668
26	3.949		0.0000	0.000	15510.0	0.407	BV	0.012	21473.88	0.597
27	3.965		0.0000	0.000	56144.2	1.472	VV	0.014	65805.83	1.829
28	4.010		0.0000	0.000	21016.6	0.551	VB	0.015	23545.26	0.654
29	4.055		0.0000	0.000	2533.2	0.066	BV	0.015	2827.67	0.079
30	4.079		0.0000	0.000	14126.5	0.370	VV	0.014	16664.09	0.463
31	4.103		0.0000	0.000	70054.4	1.837	VV	0.015	80148.41	2.228
32	4.129		0.0000	0.000	21602.5	0.567	VV	0.019	18728.35	0.521
33	4.169		0.0000	0.000	7388.1	0.194	VV	0.015	8380.08	0.233
34	4.206		0.0000	0.000	50327.5	1.320	VV	0.016	53244.87	1.480
35	4.243		0.0000	0.000	4916.7	0.129	VV	0.017	4798.15	0.133
36	4.260		0.0000	0.000	9894.4	0.259	VV	0.024	6814.48	0.189
37	4.358		0.0000	0.000	1645.3	0.043	VB	0.016	1747.53	0.049
38	4.416		0.0000	0.000	7994.7	0.210	BB	0.015	9182.81	0.255
39	4.463		0.0000	0.000	20674.0	0.542	BV	0.014	24584.81	0.683
40	4.493		0.0000	0.000	71421.6	1.873	VB	0.017	68360.70	1.900
41	4.521		0.0000	0.000	24866.2	0.652	BB	0.016	25300.77	0.703
42	4.574		0.0000	0.000	140756.7	3.692	BV	0.017	135980.90	3.780
43	4.643		0.0000	0.000	4769.6	0.125	VV	0.015	5474.86	0.152
44	4.670		0.0000	0.000	27614.5	0.724	VV	0.016	29208.29	0.812
45	4.713		0.0000	0.000	167470.5	4.392	VB	0.019	150311.40	4.178
46	4.821		0.0000	0.000	159187.9	4.175	BV	0.019	136563.20	3.796
47	4.878		0.0000	0.000	17112.3	0.449	VB	0.016	17484.80	0.486
48	4.916		0.0000	0.000	31325.2	0.822	BV	0.016	33274.52	0.925
49	4.952		0.0000	0.000	150778.1	3.954	VV	0.020	127827.50	3.553
50	4.985		0.0000	0.000	119762.9	3.141	VB	0.017	117176.10	3.257
51	5.025		0.0000	0.000	54881.3	1.439	BB	0.016	58479.17	1.625
52	5.127		0.0000	0.000	26134.3	0.685	BB	0.018	24428.75	0.679
53	5.179		0.0000	0.000	142241.6	3.730	BV	0.020	121369.30	3.373
54	5.252		0.0000	0.000	67143.9	1.761	VV	0.018	61071.29	1.697
55	5.287		0.0000	0.000	28210.3	0.740	VV	0.016	28574.21	0.794
56	5.314		0.0000	0.000	21101.7	0.553	VB	0.017	20226.80	0.562
57	5.368		0.0000	0.000	285627.1	7.491	BB	0.020	238134.10	6.619
58	5.443		0.0000	0.000	8844.8	0.232	BB	0.014	10324.06	0.287
59	5.558		0.0000	0.000	1214.0	0.032	BB	0.012	1707.84	0.047
60	5.581		0.0000	0.000	60140.8	1.577	BB	0.016	62427.02	1.735
61	5.627		0.0000	0.000	210723.8	5.526	BB	0.022	160111.60	4.450
62	5.781		0.0000	0.000	2938.0	0.077	BB	0.015	3182.63	0.088

ASX22 0682

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.820		0.0000	0.000	2039.7	0.053	BB	0.014	2350.07	0.065
64	5.849		0.0000	0.000	1777.8	0.047	BB	0.013	2275.60	0.063
65	5.897		0.0000	0.000	35693.7	0.936	BB	0.019	31793.76	0.884
66	6.000		0.0000	0.000	91110.4	2.389	BB	0.022	68649.24	1.908
67	6.199		0.0000	0.000	22482.6	0.590	BB	0.019	19327.61	0.537
68	6.360	DCB	71.5029	51.638	204299.2	5.358	BB	0.020	170656.80	4.743
69	6.528		0.0000	0.000	111.5	0.003	BB	0.011	166.28	0.005
70	6.797		0.0000	0.000	104.1	0.003	BB	0.031	55.60	0.002
71	6.989		0.0000	0.000	257.9	0.007	BB	0.051	83.54	0.002
72	7.171		0.0000	0.000	84.2	0.002	BB	0.019	75.11	0.002

Total Area = 3812973.0, Total Amount = 138.471, Total Height = 3597842.0, Sample Units = PPB

ASX22 8683

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.09R

AR1650924A AAAR165AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.09R



M = Manually Integrated
Analyst
Approved by
Circle Reason 1 2 3 4
1 = Missed Peak
2 = Improper Baseline
3 = RT Update
4 = Other

Instrument ID: CP06–S9147A Injected On: 11/6/2009 4:41:23 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.09R

AR1650924A AAAR165AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.09R



ASX22 0604

Instrument ID: CP06–S9147B Injected On: 11/6/2009 4:41:23 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.312	136617	69.84163	TCX	2.808	306405	66.96774	TCX
7.161	22902	72.18638	DCB	6.36	170657	71.50294	DCB

Files:
Area File: C:\CPWIN\~Dualcha.00A
Area File: C:\CPWIN\~Dualchb.00A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/7/2009 6:24:00 AM
File Reported On: 11/7/2009 at 6:23:59 AM

REPLOTS

ASX22 0685

AR1650924A AAAR165AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.09R



AR1650924A AAAR165AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.09R



ASX22 0686

Sample Name: **AR4810924B AAAR481AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 05:00:54pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.10R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.260		0.0000	0.000	226.7	0.480	BB	0.090	41.97	0.120
2	3.311	TCX	5.0932	46.909	9142.2	19.378	BB	0.015	9962.83	28.391
3	3.513		0.0000	0.000	124.2	0.263	BB	0.013	160.79	0.458
4	3.564		0.0000	0.000	304.9	0.646	BB	0.021	247.90	0.706
5	3.659		0.0000	0.000	128.5	0.272	BB	0.027	80.62	0.230
6	3.787		0.0000	0.000	1583.6	3.357	BB	0.021	1262.74	3.598
7	3.914		0.0000	0.000	830.5	1.760	BB	0.016	855.65	2.438
8	4.032		0.0000	0.000	179.6	0.381	BB	0.018	163.69	0.466
9	4.112		0.0000	0.000	2869.6	6.082	BB	0.031	1520.52	4.333
10	4.155		0.0000	0.000	686.9	1.456	BB	0.017	662.53	1.888
11	4.220		0.0000	0.000	993.9	2.107	BB	0.018	942.19	2.685
12	4.266		0.0000	0.000	78.6	0.167	BB	0.013	99.40	0.283
13	4.292		0.0000	0.000	1522.3	3.227	BB	0.015	1646.59	4.692
14	4.323		0.0000	0.000	1958.8	4.152	BB	0.027	1199.14	3.417
15	4.432		0.0000	0.000	1613.2	3.419	BB	0.016	1654.89	4.716
16	4.459		0.0000	0.000	261.9	0.555	BB	0.012	378.80	1.079
17	4.493		0.0000	0.000	395.2	0.838	BB	0.015	450.10	1.283
18	4.531		0.0000	0.000	1377.7	2.920	BB	0.019	1234.98	3.519
19	4.577		0.0000	0.000	403.7	0.856	BB	0.017	407.68	1.162
20	4.667		0.0000	0.000	101.1	0.214	BB	0.017	98.16	0.280
21	4.708		0.0000	0.000	1448.3	3.070	BV	0.018	1349.40	3.845
22	4.729		0.0000	0.000	1756.4	3.723	VV	0.020	1494.84	4.260
23	4.756		0.0000	0.000	2362.0	5.006	VV	0.029	1377.81	3.926
24	4.826		0.0000	0.000	513.8	1.089	VV	0.033	259.07	0.738
25	4.874		0.0000	0.000	4058.4	8.602	VV	0.047	1438.82	4.100
26	5.031		0.0000	0.000	412.7	0.875	VV	0.023	299.65	0.854
27	5.088		0.0000	0.000	603.0	1.278	VV	0.022	451.99	1.288
28	5.123		0.0000	0.000	373.8	0.792	VV	0.021	291.93	0.832
29	5.154		0.0000	0.000	1339.0	2.838	VB	0.030	754.69	2.151
30	5.245		0.0000	0.000	193.7	0.411	BB	0.017	192.43	0.548
31	5.282		0.0000	0.000	169.9	0.360	BB	0.020	141.69	0.404
32	5.376		0.0000	0.000	404.7	0.858	BB	0.020	344.35	0.981
33	5.482		0.0000	0.000	307.8	0.652	BB	0.034	149.63	0.426
34	5.579		0.0000	0.000	636.8	1.350	BB	0.029	361.67	1.031
35	5.692		0.0000	0.000	269.8	0.572	BB	0.022	201.79	0.575
36	5.750		0.0000	0.000	108.1	0.229	BB	0.018	98.72	0.281
37	5.915		0.0000	0.000	296.7	0.629	BB	0.033	147.74	0.421
38	6.015		0.0000	0.000	1235.6	2.619	BB	0.073	280.26	0.799
39	6.150		0.0000	0.000	431.8	0.915	BB	0.029	248.64	0.709
40	6.354		0.0000	0.000	666.3	1.412	BB	0.054	205.15	0.585
41	6.770		0.0000	0.000	261.3	0.554	BB	0.043	101.53	0.289
42	7.160	DCB	5.7645	53.091	4546.2	9.636	BB	0.041	1828.91	5.212

Total Area = 47179.2, Total Amount = 10.858, Total Height = 35091.88, Sample Units = PPB

ASX22 8687

Sample Name: **AR4810924B AAAR481AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 05:00:54pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.10R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.609		0.0000	0.000	581.9	0.354	BV	0.037	260.55	0.172
2	2.700		0.0000	0.000	863.1	0.524	VB	0.020	709.12	0.469
3	2.808	TCX	4.9434	47.589	20417.7	12.407	BB	0.015	22618.24	14.971
4	2.928		0.0000	0.000	261.7	0.159	BB	0.014	318.97	0.211
5	3.030		0.0000	0.000	300.6	0.183	BB	0.016	317.56	0.210
6	3.098		0.0000	0.000	672.2	0.408	BB	0.018	628.21	0.416
7	3.212		0.0000	0.000	165.8	0.101	BB	0.013	220.79	0.146
8	3.308		0.0000	0.000	3907.5	2.374	BV	0.018	3677.50	2.434
9	3.360		0.0000	0.000	221.9	0.135	VV	0.014	261.93	0.173
10	3.431		0.0000	0.000	2245.1	1.364	VV	0.014	2597.62	1.719
11	3.513		0.0000	0.000	463.9	0.282	VV	0.013	595.54	0.394
12	3.530		0.0000	0.000	523.9	0.318	VB	0.013	681.53	0.451
13	3.580		0.0000	0.000	9293.2	5.647	BV	0.025	6235.33	4.127
14	3.624		0.0000	0.000	3092.3	1.879	VV	0.019	2735.00	1.810
15	3.701		0.0000	0.000	2997.4	1.821	VV	0.014	3570.82	2.364
16	3.736		0.0000	0.000	6226.1	3.783	VV	0.013	7743.55	5.126
17	3.760		0.0000	0.000	8956.2	5.442	VB	0.020	7287.83	4.824
18	3.870		0.0000	0.000	6774.4	4.116	BV	0.014	8306.78	5.498
19	3.894		0.0000	0.000	1881.0	1.143	VB	0.012	2557.55	1.693
20	3.917		0.0000	0.000	1419.4	0.862	BB	0.011	2065.42	1.367
21	3.949		0.0000	0.000	896.8	0.545	BV	0.012	1253.31	0.830
22	3.965		0.0000	0.000	4375.0	2.658	VV	0.014	5222.42	3.457
23	4.010		0.0000	0.000	1623.1	0.986	VB	0.014	1912.84	1.266
24	4.055		0.0000	0.000	349.3	0.212	BV	0.015	387.01	0.256
25	4.079		0.0000	0.000	3105.9	1.887	VV	0.014	3588.00	2.375
26	4.103		0.0000	0.000	8559.4	5.201	VV	0.015	9632.92	6.376
27	4.129		0.0000	0.000	8969.5	5.450	VV	0.025	6088.13	4.030
28	4.207		0.0000	0.000	3210.9	1.951	VV	0.017	3140.43	2.079
29	4.243		0.0000	0.000	6393.4	3.885	VV	0.023	4544.81	3.008
30	4.272		0.0000	0.000	3396.8	2.064	VV	0.018	3230.21	2.138
31	4.303		0.0000	0.000	1604.1	0.975	VV	0.038	708.74	0.469
32	4.359		0.0000	0.000	1901.4	1.155	VV	0.019	1639.15	1.085
33	4.417		0.0000	0.000	2324.4	1.412	VV	0.017	2321.87	1.537
34	4.452		0.0000	0.000	1572.9	0.956	VV	0.016	1590.82	1.053
35	4.480		0.0000	0.000	4888.6	2.970	VB	0.021	3924.80	2.598
36	4.577		0.0000	0.000	2125.8	1.292	BV	0.020	1807.87	1.197
37	4.610		0.0000	0.000	324.2	0.197	VV	0.016	332.03	0.220
38	4.645		0.0000	0.000	2475.2	1.504	VV	0.020	2012.75	1.332
39	4.694		0.0000	0.000	1079.0	0.656	VV	0.020	921.76	0.610
40	4.712		0.0000	0.000	1532.3	0.931	VB	0.025	1004.11	0.665
41	4.821		0.0000	0.000	1425.0	0.866	BV	0.026	926.96	0.614
42	4.872		0.0000	0.000	2633.0	1.600	VV	0.024	1865.50	1.235
43	4.953		0.0000	0.000	1338.0	0.813	VV	0.022	1022.26	0.677
44	4.984		0.0000	0.000	567.3	0.345	VB	0.018	538.72	0.357
45	5.024		0.0000	0.000	161.0	0.098	BB	0.015	180.27	0.119
46	5.125		0.0000	0.000	75.4	0.046	BB	0.015	83.44	0.055
47	5.179		0.0000	0.000	693.7	0.422	BV	0.020	588.10	0.389
48	5.252		0.0000	0.000	293.9	0.179	VV	0.019	256.84	0.170
49	5.285		0.0000	0.000	101.1	0.061	VV	0.016	102.55	0.068
50	5.313		0.0000	0.000	46.4	0.028	VB	0.013	58.89	0.039
51	5.367		0.0000	0.000	2305.0	1.401	BB	0.028	1354.74	0.897
52	5.546		0.0000	0.000	407.6	0.248	BB	0.021	324.42	0.215
53	5.581		0.0000	0.000	297.8	0.181	BB	0.015	325.71	0.216
54	5.626		0.0000	0.000	870.3	0.529	BB	0.019	744.03	0.492
55	5.724		0.0000	0.000	2667.8	1.621	BB	0.371	119.80	0.079
56	5.941		0.0000	0.000	939.9	0.571	BB	0.227	69.07	0.046
57	6.001		0.0000	0.000	715.8	0.435	BB	0.027	448.61	0.297
58	6.085		0.0000	0.000	218.4	0.133	BB	0.030	120.97	0.080
59	6.199		0.0000	0.000	213.0	0.129	BB	0.017	205.85	0.136
60	6.360	DCB	5.4444	52.411	16391.7	9.960	BB	0.021	12994.18	8.601
61	6.751		0.0000	0.000	237.0	0.144	BB	0.043	92.61	0.061

ASX22 0608

Total Area = 164572.3, Total Amount = 10.388, Total Height = 151077.3, Sample Units = PPB

ASX22. 0689

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.10R

AR4810924B AAAR481AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.10R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 4:53:24 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.10R

AR4810924B AAAR481AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.10R



ASX22 0618

Instrument ID: CP06--S9147B Injected On: 11/6/2009 4:53:24 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	9963	5.09321	TCX	2.808	22618	4.94344	TCX
7.16	1829	5.76453	DCB	6.36	12994	5.44439	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.10A
Area File: C:\CPWIN\DATA1\1C6310B.10A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 5:01:00 PM
File Reported On: 11/6/2009 at 5:01:09 PM

REPLOTS

ASX22 8611

AR4810924B AAAR481AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.10R



AR4810924B AAAR481AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.10R



ASX22 0612

Sample Name: **AR4820924B AAAR482AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 05:12:51pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.11R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.180		0.0000	0.000	70.8	0.062	BB	0.023	51.92	0.056
2	3.311	TCX	22.3716	49.826	40523.5	35.205	BB	0.015	43760.98	47.177
3	3.513		0.0000	0.000	575.9	0.500	BB	0.014	664.04	0.716
4	3.565		0.0000	0.000	364.0	0.316	BB	0.017	357.42	0.385
5	3.662		0.0000	0.000	302.1	0.262	BB	0.028	177.75	0.192
6	3.787		0.0000	0.000	2833.1	2.461	BB	0.020	2308.96	2.489
7	3.844		0.0000	0.000	123.3	0.107	BB	0.017	119.38	0.129
8	3.914		0.0000	0.000	1486.4	1.291	BB	0.015	1610.60	1.736
9	3.980		0.0000	0.000	99.7	0.087	BB	0.016	102.37	0.110
10	4.032		0.0000	0.000	358.3	0.311	BB	0.020	302.00	0.326
11	4.100		0.0000	0.000	5436.8	4.723	BB	0.031	2892.29	3.118
12	4.155		0.0000	0.000	1248.5	1.085	BB	0.017	1243.34	1.340
13	4.220		0.0000	0.000	1761.5	1.530	BB	0.017	1736.83	1.872
14	4.265		0.0000	0.000	159.1	0.138	BB	0.015	182.67	0.197
15	4.291		0.0000	0.000	2881.9	2.504	BB	0.015	3177.86	3.426
16	4.322		0.0000	0.000	3634.7	3.158	BB	0.028	2191.83	2.363
17	4.432		0.0000	0.000	2970.1	2.580	BB	0.016	3013.83	3.249
18	4.459		0.0000	0.000	437.1	0.380	BB	0.011	673.08	0.726
19	4.493		0.0000	0.000	768.1	0.667	BB	0.014	885.84	0.955
20	4.531		0.0000	0.000	2602.3	2.261	BB	0.019	2295.10	2.474
21	4.578		0.0000	0.000	743.6	0.646	BB	0.017	721.05	0.777
22	4.668		0.0000	0.000	164.9	0.143	BB	0.015	178.45	0.192
23	4.707		0.0000	0.000	2673.9	2.323	BV	0.018	2411.54	2.600
24	4.728		0.0000	0.000	3257.1	2.830	VV	0.019	2814.88	3.035
25	4.756		0.0000	0.000	4315.9	3.749	VV	0.028	2587.77	2.790
26	4.827		0.0000	0.000	805.6	0.700	VV	0.027	491.92	0.530
27	4.874		0.0000	0.000	7674.1	6.667	VV	0.048	2658.98	2.867
28	5.031		0.0000	0.000	894.8	0.777	VV	0.024	614.53	0.663
29	5.087		0.0000	0.000	1208.3	1.050	VV	0.023	876.61	0.945
30	5.122		0.0000	0.000	797.8	0.693	VV	0.022	604.09	0.651
31	5.153		0.0000	0.000	2147.1	1.865	VB	0.027	1349.65	1.455
32	5.244		0.0000	0.000	399.3	0.347	BB	0.018	376.54	0.406
33	5.282		0.0000	0.000	117.9	0.102	BB	0.018	110.11	0.119
34	5.376		0.0000	0.000	837.7	0.728	BB	0.022	640.21	0.690
35	5.483		0.0000	0.000	42.1	0.037	BB	0.013	56.04	0.060
36	5.578		0.0000	0.000	1069.6	0.929	BB	0.029	611.65	0.659
37	5.692		0.0000	0.000	447.2	0.388	BB	0.035	212.42	0.229
38	5.905		0.0000	0.000	183.7	0.160	BB	0.036	86.14	0.093
39	6.013		0.0000	0.000	458.7	0.398	BB	0.063	121.18	0.131
40	6.150		0.0000	0.000	135.1	0.117	BB	0.022	101.02	0.109
41	6.353		0.0000	0.000	134.2	0.117	BB	0.027	81.75	0.088
42	6.770		0.0000	0.000	257.4	0.224	BB	0.063	68.38	0.074
43	6.995		0.0000	0.000	175.5	0.153	BB	0.033	89.04	0.096
44	7.161	DCB	22.5275	50.174	17529.1	15.228	BB	0.041	7147.26	7.705

Total Area = 115107.8, Total Amount = 44.899, Total Height = 92759.36, Sample Units = PPB

ASX22 0613

Sample Name: **AR4820924B AAAR482AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 05:12:51pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.11R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.611		0.0000	0.000	1264.5	0.338	BV	0.065	322.27	0.089
2	2.701		0.0000	0.000	553.8	0.148	VB	0.029	315.54	0.087
3	2.808	TCX	22.2178	49.619	88708.7	23.711	BB	0.015	101655.30	28.105
4	2.929		0.0000	0.000	101.5	0.027	BB	0.012	144.99	0.040
5	2.982		0.0000	0.000	76.4	0.020	BB	0.011	113.39	0.031
6	3.030		0.0000	0.000	1261.0	0.337	BB	0.014	1509.71	0.417
7	3.098		0.0000	0.000	734.5	0.196	BB	0.013	924.69	0.256
8	3.148		0.0000	0.000	232.1	0.062	BB	0.013	303.61	0.084
9	3.212		0.0000	0.000	308.0	0.082	BB	0.013	396.71	0.110
10	3.308		0.0000	0.000	7048.4	1.884	BV	0.017	6808.08	1.882
11	3.360		0.0000	0.000	498.7	0.133	VV	0.017	486.06	0.134
12	3.431		0.0000	0.000	4071.6	1.088	VV	0.014	4732.54	1.308
13	3.495		0.0000	0.000	288.8	0.077	VV	0.017	287.66	0.080
14	3.513		0.0000	0.000	1248.6	0.334	VB	0.019	1121.07	0.310
15	3.580		0.0000	0.000	17440.1	4.662	BV	0.025	11646.79	3.220
16	3.624		0.0000	0.000	5751.0	1.537	VV	0.019	5131.32	1.419
17	3.701		0.0000	0.000	5574.4	1.490	VV	0.014	6737.82	1.863
18	3.736		0.0000	0.000	12265.0	3.278	VV	0.014	14409.92	3.984
19	3.760		0.0000	0.000	15862.8	4.240	VB	0.020	13556.54	3.748
20	3.870		0.0000	0.000	12699.8	3.395	BV	0.014	15344.74	4.242
21	3.894		0.0000	0.000	3582.9	0.958	VB	0.012	4917.84	1.360
22	3.917		0.0000	0.000	2617.6	0.700	BB	0.012	3760.49	1.040
23	3.950		0.0000	0.000	1752.8	0.469	BV	0.012	2487.71	0.688
24	3.965		0.0000	0.000	8420.5	2.251	VV	0.014	10082.39	2.788
25	4.010		0.0000	0.000	3195.9	0.854	VB	0.014	3681.29	1.018
26	4.055		0.0000	0.000	761.8	0.204	BV	0.015	850.56	0.235
27	4.079		0.0000	0.000	6140.6	1.641	VV	0.015	7051.57	1.950
28	4.103		0.0000	0.000	16065.4	4.294	VV	0.015	18188.91	5.029
29	4.129		0.0000	0.000	16857.7	4.506	VV	0.024	11918.60	3.295
30	4.206		0.0000	0.000	5974.6	1.597	VV	0.017	5870.38	1.623
31	4.243		0.0000	0.000	9329.2	2.494	VV	0.017	8888.40	2.457
32	4.271		0.0000	0.000	10735.7	2.870	VV	0.029	6201.48	1.715
33	4.321		0.0000	0.000	1679.0	0.449	VV	0.022	1274.70	0.352
34	4.358		0.0000	0.000	3699.2	0.989	VV	0.019	3195.03	0.883
35	4.416		0.0000	0.000	4466.2	1.194	VV	0.017	4404.15	1.218
36	4.451		0.0000	0.000	2988.4	0.799	VV	0.017	2977.46	0.823
37	4.480		0.0000	0.000	8582.6	2.294	VB	0.019	7669.01	2.120
38	4.578		0.0000	0.000	3316.2	0.886	BV	0.019	2870.80	0.794
39	4.610		0.0000	0.000	646.2	0.173	VV	0.016	652.75	0.180
40	4.644		0.0000	0.000	4826.0	1.290	VV	0.021	3918.36	1.083
41	4.692		0.0000	0.000	688.5	0.184	VV	0.017	671.98	0.186
42	4.712		0.0000	0.000	1077.4	0.288	VV	0.021	860.89	0.238
43	4.740		0.0000	0.000	663.3	0.177	VB	0.020	540.36	0.149
44	4.820		0.0000	0.000	1430.7	0.382	BV	0.026	933.55	0.258
45	4.872		0.0000	0.000	5516.8	1.475	VV	0.025	3679.16	1.017
46	4.953		0.0000	0.000	2999.1	0.802	VB	0.036	1381.70	0.382
47	5.179		0.0000	0.000	270.6	0.072	BV	0.018	246.42	0.068
48	5.202		0.0000	0.000	437.9	0.117	VV	0.019	384.65	0.106
49	5.252		0.0000	0.000	130.1	0.035	VB	0.018	120.82	0.033
50	5.367		0.0000	0.000	1089.5	0.291	BB	0.030	609.80	0.169
51	5.545		0.0000	0.000	111.7	0.030	BB	0.016	114.85	0.032
52	5.581		0.0000	0.000	141.4	0.038	BB	0.017	135.80	0.038
53	5.625		0.0000	0.000	389.7	0.104	BB	0.021	307.43	0.085
54	5.913		0.0000	0.000	424.2	0.113	BB	0.048	148.74	0.041
55	6.000		0.0000	0.000	358.8	0.096	BB	0.028	211.81	0.059
56	6.200		0.0000	0.000	709.2	0.190	BB	0.020	580.31	0.160
57	6.360	DCB	22.5592	50.381	65828.0	17.595	BB	0.020	53842.26	14.886
58	6.752		0.0000	0.000	160.7	0.043	BB	0.029	93.16	0.026
59	7.305		0.0000	0.000	64.9	0.017	BB	0.061	17.66	0.005

ASX22. 06.14

Total Area = 374120.5, Total Amount = 44.777, Total Height = 361691.9, Sample Units = PPB

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.11R
AR4820924B AAAR482AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.11R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 5:05:21 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.11R

AR4820924B AAAR482AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.11R



ASX22 0615

Instrument ID: CP06--S9147B Injected On: 11/6/2009 5:05:21 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	43761	22.37156	TCX	2.808	101655	22.21777	TCX
7.161	7147	22.52746	DCB	6.36	53842	22.5592	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.11A
Area File: C:\CPWIN\DATA1\1C6310B.11A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 5:12:56 PM
File Reported On: 11/6/2009 at 5:13:06 PM

REPLOTS

ASX22 0616

AR4820924B AAAR482AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.11R



AR4820924B AAAR482AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.11R



ASX22 0617

Sample Name: **AR4830924B AAAR483AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 05:24:44pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.12R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.171		0.0000	0.000	82.6	0.040	BB	0.019	71.14	0.043
2	3.227		0.0000	0.000	262.7	0.127	BB	0.048	91.87	0.055
3	3.311	TCX	41.8723	50.410	78254.5	37.787	BB	0.016	81906.31	49.119
4	3.513		0.0000	0.000	1053.9	0.509	BB	0.014	1221.06	0.732
5	3.564		0.0000	0.000	539.2	0.260	BB	0.016	579.29	0.347
6	3.660		0.0000	0.000	425.1	0.205	BB	0.025	287.14	0.172
7	3.787		0.0000	0.000	4857.1	2.345	BB	0.021	3841.38	2.304
8	3.844		0.0000	0.000	239.6	0.116	BB	0.018	222.21	0.133
9	3.914		0.0000	0.000	2585.0	1.248	BB	0.016	2709.20	1.625
10	3.980		0.0000	0.000	155.7	0.075	BB	0.017	151.48	0.091
11	4.032		0.0000	0.000	618.8	0.299	BB	0.020	512.38	0.307
12	4.111		0.0000	0.000	9313.7	4.497	BB	0.032	4896.98	2.937
13	4.154		0.0000	0.000	2139.3	1.033	BB	0.017	2094.22	1.256
14	4.219		0.0000	0.000	3005.3	1.451	BB	0.017	2887.15	1.731
15	4.265		0.0000	0.000	252.8	0.122	BB	0.014	295.68	0.177
16	4.291		0.0000	0.000	4879.5	2.356	BB	0.015	5351.84	3.209
17	4.322		0.0000	0.000	6114.3	2.952	BB	0.027	3744.55	2.246
18	4.431		0.0000	0.000	5022.7	2.425	BB	0.015	5403.12	3.240
19	4.459		0.0000	0.000	886.8	0.428	BB	0.012	1284.88	0.771
20	4.492		0.0000	0.000	1302.0	0.629	BB	0.015	1484.66	0.890
21	4.531		0.0000	0.000	4462.7	2.155	BB	0.019	3866.58	2.319
22	4.577		0.0000	0.000	1242.0	0.600	BB	0.017	1253.90	0.752
23	4.668		0.0000	0.000	332.9	0.161	BB	0.017	325.72	0.195
24	4.707		0.0000	0.000	4743.3	2.290	BV	0.019	4184.83	2.510
25	4.727		0.0000	0.000	5101.0	2.463	VV	0.018	4708.98	2.824
26	4.755		0.0000	0.000	7528.9	3.636	VV	0.028	4486.93	2.691
27	4.825		0.0000	0.000	1557.3	0.752	VV	0.032	822.65	0.493
28	4.873		0.0000	0.000	12089.5	5.838	VV	0.044	4561.34	2.735
29	4.983		0.0000	0.000	866.2	0.418	VV	0.030	487.68	0.292
30	5.031		0.0000	0.000	1500.9	0.725	VV	0.024	1060.94	0.636
31	5.087		0.0000	0.000	2111.5	1.020	VV	0.023	1511.36	0.906
32	5.121		0.0000	0.000	1231.6	0.595	VV	0.020	1020.39	0.612
33	5.153		0.0000	0.000	3626.2	1.751	VB	0.026	2335.70	1.401
34	5.244		0.0000	0.000	715.2	0.345	BB	0.019	628.63	0.377
35	5.281		0.0000	0.000	177.5	0.086	BB	0.021	140.80	0.084
36	5.331		0.0000	0.000	119.5	0.058	BB	0.016	126.59	0.076
37	5.375		0.0000	0.000	1351.4	0.653	BB	0.021	1053.02	0.631
38	5.480		0.0000	0.000	45.2	0.022	BB	0.011	71.38	0.043
39	5.578		0.0000	0.000	1806.4	0.872	BB	0.029	1043.38	0.626
40	5.692		0.0000	0.000	723.8	0.350	BB	0.038	319.53	0.192
41	5.901		0.0000	0.000	227.3	0.110	BB	0.035	109.57	0.066
42	6.012		0.0000	0.000	459.3	0.222	BB	0.062	123.10	0.074
43	6.150		0.0000	0.000	89.8	0.043	BB	0.017	85.89	0.052
44	6.350		0.0000	0.000	353.9	0.171	BB	0.060	99.07	0.059
45	6.538		0.0000	0.000	46.4	0.022	BB	0.029	26.78	0.016
46	6.776		0.0000	0.000	71.8	0.035	BB	0.026	46.01	0.028
47	6.989		0.0000	0.000	249.8	0.121	BB	0.029	144.85	0.087
48	7.160	DCB	41.1904	49.590	32272.6	15.584	BB	0.041	13068.43	7.837

Total Area = 207094.3, Total Amount = 83.063, Total Height = 166750.6, Sample Units = PPB

ASX22 6618.

Sample Name: **AR4830924B AAAR483AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 05:24:44pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.12R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.610		0.0000	0.000	936.1	0.141	BV	0.068	231.07	0.036
2	2.700		0.0000	0.000	1072.4	0.161	VB	0.022	823.32	0.128
3	2.807 TCX		41.7656	50.304	169813.5	25.500	BB	0.015	191094.60	29.763
4	2.927		0.0000	0.000	91.8	0.014	BB	0.011	138.02	0.021
5	2.980		0.0000	0.000	151.4	0.023	BB	0.012	210.64	0.033
6	3.029		0.0000	0.000	2463.5	0.370	BB	0.014	2909.89	0.453
7	3.097		0.0000	0.000	1233.5	0.185	BB	0.013	1574.93	0.245
8	3.147		0.0000	0.000	450.7	0.068	BB	0.013	591.92	0.092
9	3.211		0.0000	0.000	574.6	0.086	BB	0.014	703.40	0.110
10	3.308		0.0000	0.000	12162.1	1.826	BV	0.018	11574.68	1.803
11	3.360		0.0000	0.000	812.1	0.122	VV	0.016	827.49	0.129
12	3.403		0.0000	0.000	159.1	0.024	VV	0.015	173.80	0.027
13	3.430		0.0000	0.000	7025.3	1.055	VV	0.014	8246.91	1.284
14	3.494		0.0000	0.000	402.1	0.060	VV	0.012	547.62	0.085
15	3.513		0.0000	0.000	2216.2	0.333	VB	0.018	2006.16	0.312
16	3.586		0.0000	0.000	30395.3	4.564	BV	0.025	19876.14	3.096
17	3.623		0.0000	0.000	9913.0	1.489	VV	0.018	9099.30	1.417
18	3.700		0.0000	0.000	9765.4	1.466	VV	0.014	11606.01	1.808
19	3.735		0.0000	0.000	20109.8	3.020	VV	0.013	24920.80	3.881
20	3.760		0.0000	0.000	28420.4	4.268	VB	0.020	23428.36	3.649
21	3.869		0.0000	0.000	21827.7	3.278	BV	0.014	26273.10	4.092
22	3.894		0.0000	0.000	5910.1	0.887	VB	0.012	8293.02	1.292
23	3.917		0.0000	0.000	4504.3	0.676	BB	0.011	6816.01	1.062
24	3.948		0.0000	0.000	3269.3	0.491	BV	0.012	4454.01	0.694
25	3.964		0.0000	0.000	14551.5	2.185	VV	0.014	17628.88	2.746
26	4.009		0.0000	0.000	5615.4	0.843	VB	0.015	6317.30	0.984
27	4.054		0.0000	0.000	1373.2	0.206	BV	0.015	1501.17	0.234
28	4.078		0.0000	0.000	10886.3	1.635	VV	0.015	12340.64	1.922
29	4.102		0.0000	0.000	27690.2	4.158	VV	0.015	31096.71	4.843
30	4.128		0.0000	0.000	28351.2	4.257	VV	0.023	20445.67	3.184
31	4.205		0.0000	0.000	10099.3	1.517	VV	0.017	9781.23	1.523
32	4.242		0.0000	0.000	16237.4	2.438	VV	0.018	15167.96	2.362
33	4.271		0.0000	0.000	17851.4	2.681	VV	0.027	10856.03	1.691
34	4.320		0.0000	0.000	2462.1	0.370	VV	0.020	2060.01	0.321
35	4.357		0.0000	0.000	6445.1	0.968	VV	0.019	5543.04	0.863
36	4.416		0.0000	0.000	7853.3	1.179	VV	0.017	7719.04	1.202
37	4.450		0.0000	0.000	5190.7	0.779	VV	0.016	5360.66	0.835
38	4.479		0.0000	0.000	14504.6	2.178	VB	0.019	12739.66	1.984
39	4.577		0.0000	0.000	5518.6	0.829	BV	0.019	4825.60	0.752
40	4.610		0.0000	0.000	1210.2	0.182	VV	0.018	1117.61	0.174
41	4.644		0.0000	0.000	9089.6	1.365	VV	0.023	6653.54	1.036
42	4.712		0.0000	0.000	1472.0	0.221	VV	0.019	1299.81	0.202
43	4.739		0.0000	0.000	1178.3	0.177	VB	0.022	894.86	0.139
44	4.819		0.0000	0.000	1924.0	0.289	BV	0.022	1430.17	0.223
45	4.871		0.0000	0.000	9041.9	1.358	VV	0.024	6267.38	0.976
46	4.952		0.0000	0.000	4470.3	0.671	VB	0.034	2191.80	0.341
47	5.179		0.0000	0.000	202.2	0.030	BV	0.015	229.26	0.036
48	5.202		0.0000	0.000	715.6	0.107	VV	0.019	638.93	0.100
49	5.251		0.0000	0.000	125.7	0.019	VB	0.017	119.84	0.019
50	5.346		0.0000	0.000	313.8	0.047	BV	0.017	314.21	0.049
51	5.367		0.0000	0.000	917.1	0.138	VB	0.025	613.45	0.096
52	5.550		0.0000	0.000	121.4	0.018	BB	0.025	81.63	0.013
53	5.578		0.0000	0.000	139.1	0.021	BB	0.016	145.29	0.023
54	5.626		0.0000	0.000	390.3	0.059	BB	0.019	336.95	0.052
55	5.911		0.0000	0.000	194.8	0.029	BB	0.038	86.25	0.013
56	6.001		0.0000	0.000	249.8	0.038	BB	0.025	169.37	0.026
57	6.088		0.0000	0.000	96.7	0.015	BB	0.026	62.33	0.010
58	6.200		0.0000	0.000	1173.9	0.176	BB	0.019	1019.83	0.159
59	6.359 DCB		41.2602	49.696	124450.5	18.688	BB	0.021	98476.05	15.338
60	6.749		0.0000	0.000	152.6	0.023	BB	0.026	96.98	0.015

ASX22 0619

Total Area = 665939.8, Total Amount = 83.026, Total Height = 642050.3, Sample Units = PPB

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.12R
AR4830924B AAAR483AA ICAL 0930999999
——— C:\CPWIN\DATA1\1C6310.12R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 5:17:14 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.12R
AR4830924B AAAR483AA ICAL 0930999999
——— C:\CPWIN\DATA1\1C6310B.12R



Instrument ID: CP06--S9147B Injected On: 11/6/2009 5:17:14 PM Column ID: RTX CLPII

ASX22 0628

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min\nmldrle Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	81906	41.87227	TCX	2.807	191095	41.76559	TCX
7.16	13068	41.1904	DCB	6.359	98476	41.26018	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.12A
Area File: C:\CPWIN\DATA1\1C6310B.12A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 5:24:50 PM
File Reported On: 11/6/2009 at 5:24:58 PM

REPLOTS

ASX22 0621

AR4830924B AAAR483AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.12R



AR4830924B AAAR483AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.12R



ASX22 0622

Sample Name: **AR4840924B AAAR484AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 05:36:44pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.13R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.220		0.0000	0.000	230.3	0.059	BB	0.047	81.05	0.026
2	3.311 TCX		59.3881	50.319	112039.0	28.486	BB	0.016	116169.00	37.205
3	3.457		0.0000	0.000	146.8	0.037	BB	0.014	170.43	0.055
4	3.513		0.0000	0.000	1604.1	0.408	BB	0.014	1877.80	0.601
5	3.564		0.0000	0.000	1331.0	0.338	BB	0.015	1468.97	0.470
6	3.661		0.0000	0.000	866.8	0.220	BB	0.022	661.24	0.212
7	3.787		0.0000	0.000	11830.1	3.008	BB	0.021	9272.20	2.970
8	3.845		0.0000	0.000	558.4	0.142	BB	0.016	566.69	0.181
9	3.914		0.0000	0.000	6372.4	1.620	BB	0.016	6641.12	2.127
10	3.980		0.0000	0.000	214.6	0.055	BB	0.018	200.15	0.064
11	4.032		0.0000	0.000	1627.4	0.414	BB	0.021	1295.35	0.415
12	4.112		0.0000	0.000	23357.3	5.939	BB	0.032	12344.61	3.954
13	4.155		0.0000	0.000	5245.5	1.334	BB	0.017	5189.21	1.662
14	4.220		0.0000	0.000	7335.4	1.865	BB	0.017	7233.54	2.317
15	4.266		0.0000	0.000	528.1	0.134	BB	0.013	670.70	0.215
16	4.291		0.0000	0.000	11868.5	3.018	BB	0.015	13096.45	4.194
17	4.323		0.0000	0.000	15141.7	3.850	BB	0.027	9504.84	3.044
18	4.432		0.0000	0.000	12751.6	3.242	BB	0.017	12770.16	4.090
19	4.459		0.0000	0.000	1893.8	0.482	BB	0.011	2911.83	0.933
20	4.493		0.0000	0.000	3181.3	0.809	BB	0.015	3621.81	1.160
21	4.531		0.0000	0.000	11428.0	2.906	BB	0.019	9865.12	3.159
22	4.578		0.0000	0.000	3154.1	0.802	BB	0.017	3145.66	1.007
23	4.622		0.0000	0.000	152.0	0.039	BB	0.012	210.96	0.068
24	4.670		0.0000	0.000	801.8	0.204	BB	0.016	855.36	0.274
25	4.707		0.0000	0.000	11627.7	2.956	BV	0.018	10692.85	3.425
26	4.728		0.0000	0.000	13365.7	3.398	VV	0.018	12535.80	4.015
27	4.755		0.0000	0.000	19143.1	4.867	VV	0.028	11481.87	3.677
28	4.826		0.0000	0.000	3165.3	0.805	VV	0.028	1901.56	0.609
29	4.873		0.0000	0.000	30293.8	7.702	VV	0.043	11623.68	3.723
30	4.984		0.0000	0.000	1703.4	0.433	VV	0.025	1148.66	0.368
31	5.030		0.0000	0.000	3773.8	0.960	VV	0.024	2625.55	0.841
32	5.087		0.0000	0.000	5114.5	1.300	VV	0.023	3747.18	1.200
33	5.122		0.0000	0.000	3277.5	0.833	VV	0.022	2497.67	0.800
34	5.154		0.0000	0.000	8466.3	2.153	VB	0.024	5800.30	1.858
35	5.244		0.0000	0.000	1743.5	0.443	BB	0.018	1609.72	0.516
36	5.282		0.0000	0.000	347.9	0.088	BB	0.018	329.58	0.106
37	5.332		0.0000	0.000	283.9	0.072	BB	0.016	303.16	0.097
38	5.375		0.0000	0.000	3659.3	0.930	BB	0.023	2662.22	0.853
39	5.457		0.0000	0.000	95.7	0.024	BB	0.014	111.24	0.036
40	5.483		0.0000	0.000	109.8	0.028	BB	0.012	155.97	0.050
41	5.504		0.0000	0.000	159.9	0.041	BB	0.013	203.02	0.065
42	5.579		0.0000	0.000	4514.0	1.148	BB	0.028	2673.62	0.856
43	5.693		0.0000	0.000	1403.8	0.357	BB	0.031	759.94	0.243
44	5.903		0.0000	0.000	464.6	0.118	BB	0.028	274.76	0.088
45	6.096		0.0000	0.000	142.3	0.036	BB	0.021	113.96	0.036
46	6.151		0.0000	0.000	189.8	0.048	BB	0.020	157.66	0.050
47	6.354		0.0000	0.000	453.1	0.115	BB	0.050	152.25	0.049
48	6.909		0.0000	0.000	75.7	0.019	BB	0.034	37.46	0.012
49	6.988		0.0000	0.000	426.0	0.108	BB	0.033	214.95	0.069
50	7.160 DCB		58.6357	49.681	45650.5	11.607	BB	0.041	18603.27	5.958

Total Area = 393310.8, Total Amount = 118.024, Total Height = 312242.1, Sample Units = PPB

ASX22. 0623

Sample Name: **AR4840924B AAAR484AA ICAL 0930999999**
Acquired from CP06–S9147B via port 2 on 11/6/09 05:36:44pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.13R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.609		0.0000	0.000	478.9	0.036	BV	0.042	189.62	0.015
2	2.700		0.0000	0.000	881.0	0.066	VB	0.021	693.36	0.054
3	2.808 TCX		57.6943	49.388	248119.4	18.587	BB	0.016	263975.10	20.704
4	2.982		0.0000	0.000	387.4	0.029	BB	0.013	508.35	0.040
5	3.029		0.0000	0.000	3806.6	0.285	BB	0.015	4161.23	0.326
6	3.097		0.0000	0.000	3192.3	0.239	BB	0.013	3989.00	0.313
7	3.148		0.0000	0.000	699.0	0.052	BB	0.014	849.43	0.067
8	3.212		0.0000	0.000	1329.6	0.100	BB	0.013	1681.31	0.132
9	3.308		0.0000	0.000	30001.6	2.247	BV	0.017	28575.19	2.241
10	3.360		0.0000	0.000	2126.8	0.159	VV	0.017	2141.25	0.168
11	3.403		0.0000	0.000	389.8	0.029	VV	0.016	416.45	0.033
12	3.431		0.0000	0.000	17588.0	1.318	VV	0.015	19978.20	1.567
13	3.494		0.0000	0.000	559.3	0.042	VV	0.012	781.53	0.061
14	3.513		0.0000	0.000	5321.4	0.399	VB	0.017	5086.16	0.399
15	3.586		0.0000	0.000	75293.4	5.640	BV	0.026	49099.13	3.851
16	3.624		0.0000	0.000	24923.8	1.867	VV	0.018	22569.85	1.770
17	3.701		0.0000	0.000	24311.9	1.821	VV	0.014	28793.86	2.258
18	3.736		0.0000	0.000	48969.5	3.668	VV	0.014	59001.78	4.628
19	3.760		0.0000	0.000	70063.3	5.249	VB	0.020	57701.82	4.526
20	3.870		0.0000	0.000	53178.7	3.984	BV	0.014	64966.26	5.095
21	3.894		0.0000	0.000	14499.8	1.086	VB	0.012	20286.19	1.591
22	3.917		0.0000	0.000	11482.0	0.860	BB	0.011	16963.14	1.330
23	3.949		0.0000	0.000	8147.0	0.610	BV	0.012	10925.36	0.857
24	3.964		0.0000	0.000	37098.3	2.779	VV	0.014	43729.29	3.430
25	4.010		0.0000	0.000	14103.7	1.057	VB	0.015	16142.77	1.266
26	4.055		0.0000	0.000	3497.1	0.262	BV	0.015	3867.27	0.303
27	4.079		0.0000	0.000	27301.1	2.045	VV	0.015	31306.41	2.455
28	4.103		0.0000	0.000	68664.6	5.144	VV	0.015	77165.21	6.052
29	4.129		0.0000	0.000	70085.4	5.250	VV	0.022	52516.38	4.119
30	4.206		0.0000	0.000	24609.6	1.844	VV	0.017	24497.80	1.921
31	4.243		0.0000	0.000	40077.5	3.002	VV	0.017	38939.12	3.054
32	4.272		0.0000	0.000	43951.1	3.292	VV	0.027	27591.49	2.164
33	4.321		0.0000	0.000	5633.9	0.422	VV	0.020	4808.44	0.377
34	4.358		0.0000	0.000	15554.8	1.165	VV	0.019	13465.48	1.056
35	4.416		0.0000	0.000	20659.2	1.548	VV	0.017	20542.27	1.611
36	4.451		0.0000	0.000	13144.5	0.985	VV	0.016	13691.95	1.074
37	4.479		0.0000	0.000	36213.3	2.713	VB	0.019	32016.95	2.511
38	4.577		0.0000	0.000	13879.2	1.040	BV	0.019	12138.74	0.952
39	4.611		0.0000	0.000	3106.0	0.233	VV	0.018	2889.10	0.227
40	4.644		0.0000	0.000	22237.7	1.666	VV	0.021	17522.28	1.374
41	4.713		0.0000	0.000	3353.5	0.251	VV	0.019	3010.98	0.236
42	4.739		0.0000	0.000	2979.1	0.223	VB	0.022	2247.99	0.176
43	4.819		0.0000	0.000	3959.8	0.297	BV	0.020	3304.36	0.259
44	4.871		0.0000	0.000	22343.9	1.674	VV	0.023	16174.33	1.269
45	4.953		0.0000	0.000	10268.0	0.769	VB	0.033	5260.93	0.413
46	5.140		0.0000	0.000	166.6	0.012	BB	0.018	155.48	0.012
47	5.179		0.0000	0.000	360.0	0.027	BV	0.015	403.08	0.032
48	5.203		0.0000	0.000	1851.8	0.139	VV	0.018	1694.43	0.133
49	5.251		0.0000	0.000	237.6	0.018	VB	0.017	239.48	0.019
50	5.285		0.0000	0.000	94.4	0.007	BB	0.014	115.23	0.009
51	5.346		0.0000	0.000	688.4	0.052	BV	0.015	754.59	0.059
52	5.367		0.0000	0.000	1722.4	0.129	VB	0.025	1163.26	0.091
53	5.549		0.0000	0.000	110.2	0.008	BB	0.028	65.52	0.005
54	5.581		0.0000	0.000	259.0	0.019	BB	0.021	201.26	0.016
55	5.626		0.0000	0.000	687.7	0.052	BB	0.019	617.60	0.048
56	5.786		0.0000	0.000	63.5	0.005	BB	0.016	64.66	0.005
57	5.896		0.0000	0.000	267.1	0.020	BB	0.032	137.68	0.011
58	6.000		0.0000	0.000	245.3	0.018	BB	0.017	240.92	0.019
59	6.085		0.0000	0.000	102.8	0.008	BB	0.038	45.01	0.004
60	6.200		0.0000	0.000	1661.9	0.124	BB	0.019	1495.55	0.117
61	6.360 DCB		59.1232	50.612	177217.5	13.276	BB	0.021	141110.00	11.067
62	6.692		0.0000	0.000	78.6	0.006	BB	0.016	82.34	0.006

ASX22. 0624.

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.746		0.0000	0.000	58.4	0.004	BB	0.018	55.56	0.004
64	6.986		0.0000	0.000	232.6	0.017	BB	0.047	82.79	0.006
65	7.173		0.0000	0.000	230.7	0.017	BB	0.041	93.40	0.007
66	7.334		0.0000	0.000	105.2	0.008	BB	0.046	38.03	0.003

Total Area = 1334914.0, Total Amount = 116.818, Total Height = 1275019.0, Sample Units = PPB

ASX22. 0525.

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.13R

AR4840924B AAAR484AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.13R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 5:29:14 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.13R

AR4840924B AAAR484AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.13R



Instrument ID: CP06--S9147B Injected On: 11/6/2009 5:29:14 PM Column ID: RTX CLPII

ASX22 8626.

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	116169	59.38811	TCX	2.808	263975	57.69434	TCX
7.16	18603	58.63566	DCB	6.36	141110	59.12323	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.13A
Area File: C:\CPWIN\DATA1\1C6310B.13A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 5:36:50 PM
File Reported On: 11/6/2009 at 5:36:58 PM

REPLOTS

ASX22 0627

AR4840924B AAAR484AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.13R



AR4840924B AAAR484AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.13R



ASX22 0620

Sample Name: **AR4850924B AAAR485AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 05:48:45pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.14R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.046		0.0000	0.000	65.0	0.010	BB	0.027	40.30	0.008
2	3.207		0.0000	0.000	315.6	0.048	BB	0.070	75.53	0.014
3	3.310	TCX	77.8617	50.436	147043.4	22.447	BB	0.016	152305.30	28.926
4	3.455		0.0000	0.000	378.0	0.058	BB	0.019	339.43	0.064
5	3.513		0.0000	0.000	2286.4	0.349	BB	0.015	2604.81	0.495
6	3.563		0.0000	0.000	2572.9	0.393	BB	0.015	2825.99	0.537
7	3.660		0.0000	0.000	1392.7	0.213	BB	0.019	1253.93	0.238
8	3.740		0.0000	0.000	25.2	0.004	BB	0.012	36.21	0.007
9	3.787		0.0000	0.000	22072.6	3.370	BB	0.021	17407.58	3.306
10	3.844		0.0000	0.000	1027.5	0.157	BB	0.017	1037.41	0.197
11	3.913		0.0000	0.000	12287.7	1.876	BB	0.017	12397.29	2.355
12	3.978		0.0000	0.000	325.2	0.050	BB	0.020	271.02	0.051
13	4.031		0.0000	0.000	3192.1	0.487	BB	0.021	2492.34	0.473
14	4.111		0.0000	0.000	44423.1	6.781	BB	0.031	23511.83	4.465
15	4.153		0.0000	0.000	10165.4	1.552	BB	0.017	9979.35	1.895
16	4.219		0.0000	0.000	14105.3	2.153	BB	0.017	13731.63	2.608
17	4.265		0.0000	0.000	1051.7	0.161	BB	0.012	1428.01	0.271
18	4.290		0.0000	0.000	22360.2	3.413	BB	0.015	24113.33	4.580
19	4.321		0.0000	0.000	28674.6	4.377	BB	0.027	17836.41	3.388
20	4.431		0.0000	0.000	23169.1	3.537	BB	0.016	24346.96	4.624
21	4.458		0.0000	0.000	4071.8	0.622	BB	0.012	5503.60	1.045
22	4.492		0.0000	0.000	6359.6	0.971	BB	0.015	7201.56	1.368
23	4.530		0.0000	0.000	22032.4	3.363	BB	0.019	19157.57	3.638
24	4.577		0.0000	0.000	6142.1	0.938	BB	0.016	6297.43	1.196
25	4.621		0.0000	0.000	329.1	0.050	BB	0.016	345.49	0.066
26	4.668		0.0000	0.000	1512.5	0.231	BB	0.017	1519.87	0.289
27	4.706		0.0000	0.000	23243.1	3.548	BV	0.019	20150.21	3.827
28	4.727		0.0000	0.000	25226.5	3.851	VV	0.018	23784.92	4.517
29	4.755		0.0000	0.000	35291.4	5.387	VV	0.026	22442.03	4.262
30	4.825		0.0000	0.000	6256.2	0.955	VV	0.029	3572.16	0.678
31	4.873		0.0000	0.000	29179.5	4.454	VV	0.022	22257.04	4.227
32	4.896		0.0000	0.000	28460.9	4.345	VV	0.041	11614.86	2.206
33	4.982		0.0000	0.000	3410.4	0.521	VV	0.027	2088.38	0.397
34	5.030		0.0000	0.000	7210.0	1.101	VV	0.024	4989.65	0.948
35	5.087		0.0000	0.000	10240.9	1.563	VV	0.022	7634.65	1.450
36	5.121		0.0000	0.000	5983.5	0.913	VV	0.020	5039.30	0.957
37	5.152		0.0000	0.000	16764.3	2.559	VB	0.024	11644.58	2.212
38	5.243		0.0000	0.000	3449.4	0.527	BB	0.017	3296.39	0.626
39	5.282		0.0000	0.000	734.3	0.112	BB	0.021	576.74	0.110
40	5.331		0.0000	0.000	620.7	0.095	BB	0.017	614.22	0.117
41	5.375		0.0000	0.000	7312.2	1.116	BB	0.022	5508.10	1.046
42	5.458		0.0000	0.000	151.1	0.023	BB	0.014	186.02	0.035
43	5.482		0.0000	0.000	205.0	0.031	BB	0.011	298.51	0.057
44	5.503		0.0000	0.000	298.4	0.046	BB	0.013	373.56	0.071
45	5.578		0.0000	0.000	8670.4	1.324	BB	0.028	5140.36	0.976
46	5.692		0.0000	0.000	2701.6	0.412	BB	0.031	1452.10	0.276
47	5.902		0.0000	0.000	869.1	0.133	BB	0.029	507.89	0.096
48	6.095		0.0000	0.000	234.2	0.036	BB	0.020	198.62	0.038
49	6.150		0.0000	0.000	321.9	0.049	BB	0.022	245.79	0.047
50	6.356		0.0000	0.000	606.3	0.093	BB	0.042	239.00	0.045
51	6.771		0.0000	0.000	146.1	0.022	BB	0.037	65.33	0.012
52	6.989		0.0000	0.000	467.5	0.071	BB	0.029	268.26	0.051
53	7.159	DCB	76.5168	49.564	59632.9	9.103	BB	0.041	24276.39	4.611

Total Area = 655068.6, Total Amount = 154.379, Total Height = 526525.3, Sample Units = PPB

ASX22 8629

Sample Name: **AR4850924B AAAR485AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 05:48:45pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.14R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.607		0.0000	0.000	1217.7	0.054	BV	0.072	283.73	0.013
2	2.700		0.0000	0.000	296.8	0.013	VB	0.029	168.10	0.008
3	2.807 TCX		77.1387	50.129	316779.8	14.115	BB	0.015	352940.90	16.232
4	2.981		0.0000	0.000	783.7	0.035	BB	0.013	1009.89	0.046
5	3.028		0.0000	0.000	5349.6	0.238	BB	0.016	5578.59	0.257
6	3.096		0.0000	0.000	6116.8	0.273	BB	0.013	7705.82	0.354
7	3.147		0.0000	0.000	852.6	0.038	BB	0.012	1153.92	0.053
8	3.211		0.0000	0.000	2545.6	0.113	BB	0.013	3291.47	0.151
9	3.307		0.0000	0.000	55655.7	2.480	BV	0.018	51839.71	2.384
10	3.359		0.0000	0.000	3947.1	0.176	VV	0.016	4082.61	0.188
11	3.403		0.0000	0.000	711.3	0.032	VV	0.015	791.38	0.036
12	3.430		0.0000	0.000	33028.0	1.472	VV	0.014	38649.07	1.778
13	3.493		0.0000	0.000	817.8	0.036	VV	0.013	1054.06	0.048
14	3.512		0.0000	0.000	9944.1	0.443	VB	0.017	9483.42	0.436
15	3.585		0.0000	0.000	126541.0	5.638	BV	0.023	91908.99	4.227
16	3.623		0.0000	0.000	62018.0	2.763	VV	0.025	42154.10	1.939
17	3.700		0.0000	0.000	45380.2	2.022	VV	0.014	54814.86	2.521
18	3.734		0.0000	0.000	92917.8	4.140	VV	0.014	109016.10	5.014
19	3.759		0.0000	0.000	128690.7	5.734	VB	0.020	105674.90	4.860
20	3.869		0.0000	0.000	99129.1	4.417	BV	0.014	117752.50	5.416
21	3.893		0.0000	0.000	25244.7	1.125	VB	0.011	37117.30	1.707
22	3.916		0.0000	0.000	21514.6	0.959	BB	0.011	32614.06	1.500
23	3.948		0.0000	0.000	13264.5	0.591	BV	0.010	21841.10	1.004
24	3.964		0.0000	0.000	73653.1	3.282	VV	0.014	85983.22	3.954
25	4.009		0.0000	0.000	26712.4	1.190	VB	0.015	30106.96	1.385
26	4.054		0.0000	0.000	7023.4	0.313	BV	0.015	7592.75	0.349
27	4.078		0.0000	0.000	54404.3	2.424	VV	0.015	60647.85	2.789
28	4.102		0.0000	0.000	129297.1	5.761	VV	0.015	145653.80	6.699
29	4.128		0.0000	0.000	129008.5	5.748	VV	0.022	99331.52	4.568
30	4.205		0.0000	0.000	46261.8	2.061	VV	0.017	45839.66	2.108
31	4.242		0.0000	0.000	79322.9	3.535	VV	0.018	74905.66	3.445
32	4.270		0.0000	0.000	79969.6	3.563	VV	0.026	52055.39	2.394
33	4.320		0.0000	0.000	9934.2	0.443	VV	0.019	8740.20	0.402
34	4.357		0.0000	0.000	28980.5	1.291	VV	0.018	26333.26	1.211
35	4.415		0.0000	0.000	40308.1	1.796	VV	0.016	41004.91	1.886
36	4.450		0.0000	0.000	25514.7	1.137	VV	0.016	26863.08	1.235
37	4.479		0.0000	0.000	68356.3	3.046	VB	0.018	62373.52	2.869
38	4.577		0.0000	0.000	26954.1	1.201	BV	0.019	23885.69	1.099
39	4.610		0.0000	0.000	5908.6	0.263	VV	0.018	5561.78	0.256
40	4.643		0.0000	0.000	42104.2	1.876	VV	0.021	34158.10	1.571
41	4.712		0.0000	0.000	6612.9	0.295	VV	0.020	5562.49	0.256
42	4.739		0.0000	0.000	5374.4	0.239	VB	0.021	4257.46	0.196
43	4.818		0.0000	0.000	7005.4	0.312	BB	0.019	6261.03	0.288
44	4.870		0.0000	0.000	39975.4	1.781	BV	0.021	31450.98	1.446
45	4.951		0.0000	0.000	11051.9	0.492	VB	0.021	8946.46	0.411
46	5.000		0.0000	0.000	322.0	0.014	BB	0.009	589.42	0.027
47	5.139		0.0000	0.000	383.7	0.017	BB	0.018	348.91	0.016
48	5.178		0.0000	0.000	749.6	0.033	BV	0.016	802.08	0.037
49	5.202		0.0000	0.000	3642.3	0.162	VV	0.018	3281.54	0.151
50	5.250		0.0000	0.000	483.0	0.022	VV	0.017	465.69	0.021
51	5.285		0.0000	0.000	182.6	0.008	VB	0.013	234.44	0.011
52	5.344		0.0000	0.000	696.8	0.031	BB	0.013	917.45	0.042
53	5.367		0.0000	0.000	1243.9	0.055	BB	0.015	1370.58	0.063
54	5.545		0.0000	0.000	117.6	0.005	BB	0.022	88.62	0.004
55	5.581		0.0000	0.000	360.1	0.016	BB	0.015	391.63	0.018
56	5.625		0.0000	0.000	1613.7	0.072	BB	0.021	1291.55	0.059
57	5.897		0.0000	0.000	321.6	0.014	BB	0.026	206.16	0.009
58	5.999		0.0000	0.000	566.0	0.025	BB	0.021	449.72	0.021
59	6.198		0.0000	0.000	2394.2	0.107	BB	0.020	2004.23	0.092
60	6.359 DCB		76.7428	49.871	234162.0	10.434	BB	0.021	183162.70	8.424
61	6.797		0.0000	0.000	103.6	0.005	BB	0.025	67.97	0.003
62	6.954		0.0000	0.000	51.5	0.002	BB	0.023	36.87	0.002

ASX22 0639

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.984		0.0000	0.000	108.4	0.005	BB	0.023	77.40	0.004
64	7.170		0.0000	0.000	255.6	0.011	BB	0.038	111.87	0.005

Total Area = 2244239.0, Total Amount = 153.881, Total Height = 2174341.0, Sample Units = PPB

ASX22. 0631

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.14R
AR4850924B AAAR485AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.14R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 5:41:15 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.14R
AR4850924B AAAR485AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.14R



ASX22 0632

Instrument ID: CP06--S9147B Injected On: 11/6/2009 5:41:15 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.31	152305	77.86175	TCX	2.807	352941	77.1387	TCX
7.159	24276	76.51676	DCB	6.359	183163	76.74275	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.14A
Area File: C:\CPWIN\DATA1\1C6310B.14A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 5:48:50 PM
File Reported On: 11/6/2009 at 5:48:59 PM

REPLOTS

ASX22 0633

AR4850924B AAAR485AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.14R



AR4850924B AAAR485AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.14R



ASX22 0634

Sample Name: **AR5410924B AAAR541AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 06:00:40pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.15R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.080		0.0000	0.000	191.3	0.424	BB	0.071	45.03	0.130
2	3.311 TCX		5.5202	46.943	9475.6	20.981	BB	0.015	10798.04	31.236
3	3.514		0.0000	0.000	153.1	0.339	BB	0.016	155.98	0.451
4	3.566		0.0000	0.000	113.0	0.250	BB	0.035	54.08	0.156
5	3.650		0.0000	0.000	65.1	0.144	BB	0.028	38.82	0.112
6	3.737		0.0000	0.000	36.9	0.082	BB	0.014	42.86	0.124
7	3.789		0.0000	0.000	105.5	0.234	BB	0.021	85.52	0.247
8	3.913		0.0000	0.000	60.4	0.134	BB	0.017	57.77	0.167
9	4.076		0.0000	0.000	41.9	0.093	BB	0.012	58.53	0.169
10	4.102		0.0000	0.000	166.1	0.368	BB	0.028	98.68	0.285
11	4.157		0.0000	0.000	50.8	0.113	BB	0.018	48.22	0.139
12	4.222		0.0000	0.000	91.6	0.203	BB	0.025	61.74	0.179
13	4.292		0.0000	0.000	1434.3	3.176	BB	0.016	1481.77	4.286
14	4.323		0.0000	0.000	325.8	0.721	BB	0.015	364.48	1.054
15	4.433		0.0000	0.000	1032.0	2.285	BB	0.020	879.23	2.543
16	4.533		0.0000	0.000	324.8	0.719	BB	0.017	312.14	0.903
17	4.579		0.0000	0.000	94.9	0.210	BB	0.022	70.93	0.205
18	4.709		0.0000	0.000	3603.9	7.980	BV	0.028	2170.26	6.278
19	4.759		0.0000	0.000	1205.3	2.669	VV	0.029	688.40	1.991
20	4.829		0.0000	0.000	526.2	1.165	VV	0.019	453.14	1.311
21	4.875		0.0000	0.000	3358.5	7.436	VV	0.023	2483.64	7.185
22	4.915		0.0000	0.000	1418.3	3.140	VB	0.025	951.93	2.754
23	5.033		0.0000	0.000	784.2	1.736	BB	0.021	626.11	1.811
24	5.088		0.0000	0.000	2354.6	5.213	BV	0.031	1278.62	3.699
25	5.155		0.0000	0.000	3034.1	6.718	VV	0.023	2197.99	6.358
26	5.205		0.0000	0.000	415.2	0.919	VB	0.026	267.96	0.775
27	5.244		0.0000	0.000	233.8	0.518	BB	0.016	250.63	0.725
28	5.283		0.0000	0.000	1065.8	2.360	BB	0.022	816.13	2.361
29	5.332		0.0000	0.000	96.5	0.214	BB	0.015	104.39	0.302
30	5.377		0.0000	0.000	1282.1	2.839	BB	0.022	988.01	2.858
31	5.432		0.0000	0.000	122.1	0.270	BB	0.017	123.27	0.357
32	5.486		0.0000	0.000	1956.6	4.332	BB	0.040	824.57	2.385
33	5.579		0.0000	0.000	1005.4	2.226	BB	0.020	845.30	2.445
34	5.621		0.0000	0.000	51.6	0.114	BB	0.011	79.14	0.229
35	5.694		0.0000	0.000	2672.2	5.917	BB	0.030	1497.64	4.332
36	5.904		0.0000	0.000	808.9	1.791	BB	0.031	441.53	1.277
37	6.015		0.0000	0.000	127.7	0.283	BB	0.020	105.53	0.305
38	6.063		0.0000	0.000	34.8	0.077	BB	0.013	46.09	0.133
39	6.097		0.0000	0.000	276.2	0.612	BB	0.023	199.70	0.578
40	6.153		0.0000	0.000	264.9	0.587	BB	0.021	208.05	0.602
41	6.359		0.0000	0.000	641.4	1.420	BB	0.044	244.03	0.706
42	6.780		0.0000	0.000	301.1	0.667	BB	0.116	43.30	0.125
43	7.161 DCB		6.2392	53.057	3759.0	8.323	BB	0.032	1979.52	5.726

Total Area = 45163.8, Total Amount = 11.759, Total Height = 34568.68, Sample Units = PPB

ASX22 0635

Sample Name: **AR5410924B AAAR541AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 06:00:40pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.15R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.613		0.0000	0.000	1289.1	0.648	BV	0.084	255.55	0.150
2	2.701		0.0000	0.000	615.2	0.309	VB	0.036	282.67	0.166
3	2.808	TCX	5.4986	48.543	20912.0	10.518	BB	0.014	25158.15	14.759
4	2.899		0.0000	0.000	127.1	0.064	BB	0.012	178.84	0.105
5	3.031		0.0000	0.000	278.5	0.140	BB	0.013	358.32	0.210
6	3.100		0.0000	0.000	40.4	0.020	BB	0.016	43.12	0.025
7	3.149		0.0000	0.000	64.0	0.032	BB	0.014	77.27	0.045
8	3.309		0.0000	0.000	685.7	0.345	BB	0.036	315.15	0.185
9	3.433		0.0000	0.000	751.6	0.378	BV	0.038	330.72	0.194
10	3.532		0.0000	0.000	1361.8	0.685	VV	0.031	727.93	0.427
11	3.580		0.0000	0.000	937.1	0.471	VV	0.025	615.28	0.361
12	3.612		0.0000	0.000	888.0	0.447	VB	0.040	369.71	0.217
13	3.702		0.0000	0.000	579.0	0.291	BV	0.026	373.56	0.219
14	3.736		0.0000	0.000	5334.2	2.683	VV	0.014	6524.06	3.827
15	3.761		0.0000	0.000	2854.8	1.436	VV	0.021	2242.15	1.315
16	3.870		0.0000	0.000	3670.2	1.846	VB	0.016	3853.14	2.260
17	3.920		0.0000	0.000	398.8	0.201	BB	0.016	425.61	0.250
18	3.966		0.0000	0.000	1509.7	0.759	BV	0.018	1402.12	0.823
19	4.010		0.0000	0.000	419.9	0.211	VV	0.018	393.48	0.231
20	4.052		0.0000	0.000	265.8	0.134	VV	0.016	274.37	0.161
21	4.080		0.0000	0.000	899.7	0.453	VV	0.013	1120.80	0.658
22	4.104		0.0000	0.000	13108.2	6.593	VV	0.018	12470.46	7.316
23	4.152		0.0000	0.000	1910.6	0.961	VV	0.016	2048.11	1.202
24	4.170		0.0000	0.000	2504.5	1.260	VV	0.016	2650.27	1.555
25	4.206		0.0000	0.000	8789.7	4.421	VV	0.015	9704.54	5.693
26	4.240		0.0000	0.000	4318.4	2.172	VV	0.015	4659.27	2.733
27	4.260		0.0000	0.000	5071.0	2.550	VV	0.020	4319.18	2.534
28	4.323		0.0000	0.000	1334.4	0.671	VV	0.024	932.27	0.547
29	4.359		0.0000	0.000	3506.8	1.764	VB	0.017	3520.66	2.065
30	4.417		0.0000	0.000	6272.9	3.155	BV	0.016	6494.84	3.810
31	4.453		0.0000	0.000	2505.7	1.260	VV	0.018	2286.20	1.341
32	4.480		0.0000	0.000	12702.7	6.389	VB	0.018	11874.21	6.966
33	4.523		0.0000	0.000	758.6	0.382	BB	0.016	797.92	0.468
34	4.576		0.0000	0.000	6679.0	3.359	BV	0.019	5960.72	3.497
35	4.612		0.0000	0.000	981.0	0.493	VV	0.019	868.53	0.510
36	4.646		0.0000	0.000	9754.7	4.906	VV	0.025	6446.25	3.782
37	4.714		0.0000	0.000	8707.6	4.380	VB	0.026	5562.30	3.263
38	4.820		0.0000	0.000	5559.1	2.796	BV	0.022	4168.88	2.446
39	4.873		0.0000	0.000	5864.1	2.949	VV	0.022	4497.56	2.639
40	4.913		0.0000	0.000	2310.6	1.162	VV	0.021	1800.43	1.056
41	4.954		0.0000	0.000	16124.6	8.110	VB	0.030	8845.86	5.189
42	5.143		0.0000	0.000	754.8	0.380	BV	0.025	498.17	0.292
43	5.180		0.0000	0.000	902.1	0.454	VV	0.017	863.02	0.506
44	5.203		0.0000	0.000	3938.0	1.981	VV	0.022	2947.61	1.729
45	5.253		0.0000	0.000	1080.9	0.544	VV	0.026	690.57	0.405
46	5.288		0.0000	0.000	760.3	0.382	VV	0.022	564.82	0.331
47	5.314		0.0000	0.000	487.9	0.245	VV	0.020	399.98	0.235
48	5.346		0.0000	0.000	1831.6	0.921	VV	0.020	1534.80	0.900
49	5.368		0.0000	0.000	3238.5	1.629	VB	0.030	1818.54	1.067
50	5.461		0.0000	0.000	493.9	0.248	BB	0.054	151.84	0.089
51	5.550		0.0000	0.000	92.9	0.047	BB	0.022	69.56	0.041
52	5.625		0.0000	0.000	2001.1	1.006	BB	0.029	1151.32	0.675
53	5.787		0.0000	0.000	317.2	0.160	BB	0.081	65.51	0.038
54	5.910		0.0000	0.000	153.0	0.077	BB	0.034	74.44	0.044
55	6.035		0.0000	0.000	656.8	0.330	BB	0.058	187.71	0.110
56	6.202		0.0000	0.000	186.9	0.094	BB	0.019	168.19	0.099
57	6.360	DCB	5.8286	51.457	18684.2	9.397	BB	0.022	13911.19	8.161
58	6.574		0.0000	0.000	54.3	0.027	BB	0.024	37.67	0.022
59	6.658		0.0000	0.000	146.2	0.074	BB	0.060	40.31	0.024
60	6.824		0.0000	0.000	71.5	0.036	BB	0.043	27.69	0.016
61	7.242		0.0000	0.000	325.5	0.164	BB	0.214	25.35	0.015

ASX22 5636.

Total Area = 198824.6, Total Amount = 11.327, Total Height = 170458.7, Sample Units = PPB

ASX22 0637

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.15R

AR5410924B AAAR541AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.15R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 5:53:10 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.15R

AR5410924B AAAR541AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.15R



Instrument ID: CP06--S9147B Injected On: 11/6/2009 5:53:10 PM Column ID: RTX CLPII

ASX22 8638

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min\Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	10798	5.52019	TCX	2.808	25158	5.49856	TCX
7.161	1980	6.23924	DCB	6.36	13911	5.82861	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.15A
Area File: C:\CPWIN\DATA1\1C6310B.15A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 6:00:46 PM
File Reported On: 11/6/2009 at 6:00:54 PM

REPLOTS

ASX22 0639

AR5410924B AAAR541AA ICAL 0930999999
——— C:\CPWIN\DATA1\1C6310.15R



AR5410924B AAAR541AA ICAL 0930999999
——— C:\CPWIN\DATA1\1C6310B.15R



ASX22 0648

Sample Name: **AR5420924B AAAR542AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 06:12:35pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.16R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.231		0.0000	0.000	135.1	0.120	BB	0.042	54.22	0.062
2	3.311	TCX	21.6633	50.942	40707.8	36.224	BB	0.016	42375.54	48.672
3	3.513		0.0000	0.000	597.9	0.532	BB	0.016	615.28	0.707
4	3.649		0.0000	0.000	135.7	0.121	BB	0.018	128.64	0.148
5	3.789		0.0000	0.000	148.4	0.132	BB	0.021	115.65	0.133
6	3.913		0.0000	0.000	77.4	0.069	BB	0.014	91.42	0.105
7	3.979		0.0000	0.000	95.3	0.085	BB	0.018	87.36	0.100
8	4.079		0.0000	0.000	38.3	0.034	BB	0.021	30.94	0.036
9	4.100		0.0000	0.000	265.2	0.236	BB	0.032	139.57	0.160
10	4.155		0.0000	0.000	77.2	0.069	BB	0.022	58.72	0.067
11	4.216		0.0000	0.000	116.9	0.104	BB	0.023	83.10	0.095
12	4.291		0.0000	0.000	2496.6	2.222	BB	0.016	2587.04	2.971
13	4.322		0.0000	0.000	481.4	0.428	BB	0.014	560.40	0.644
14	4.431		0.0000	0.000	1745.9	1.554	BB	0.020	1473.49	1.692
15	4.493		0.0000	0.000	74.2	0.066	BB	0.013	97.25	0.112
16	4.533		0.0000	0.000	547.2	0.487	BB	0.017	523.64	0.601
17	4.577		0.0000	0.000	132.6	0.118	BB	0.019	115.69	0.133
18	4.668		0.0000	0.000	119.1	0.106	BB	0.017	116.79	0.134
19	4.708		0.0000	0.000	6309.6	5.615	BV	0.027	3832.31	4.402
20	4.758		0.0000	0.000	2254.4	2.006	VV	0.031	1219.11	1.400
21	4.828		0.0000	0.000	1246.9	1.110	VV	0.023	904.99	1.039
22	4.875		0.0000	0.000	6136.7	5.461	VV	0.023	4484.64	5.151
23	4.914		0.0000	0.000	3409.8	3.034	VV	0.032	1789.18	2.055
24	4.983		0.0000	0.000	545.3	0.485	VV	0.025	359.14	0.413
25	5.031		0.0000	0.000	1334.3	1.187	VB	0.020	1087.83	1.249
26	5.088		0.0000	0.000	4017.0	3.575	BV	0.032	2116.40	2.431
27	5.154		0.0000	0.000	5023.3	4.470	VV	0.022	3805.36	4.371
28	5.201		0.0000	0.000	720.3	0.641	VB	0.027	447.38	0.514
29	5.244		0.0000	0.000	354.9	0.316	BB	0.016	379.47	0.436
30	5.283		0.0000	0.000	1740.8	1.549	BB	0.021	1361.09	1.563
31	5.331		0.0000	0.000	136.8	0.122	BB	0.016	142.94	0.164
32	5.376		0.0000	0.000	2406.1	2.141	BB	0.023	1782.10	2.047
33	5.431		0.0000	0.000	213.7	0.190	BB	0.017	214.50	0.246
34	5.484		0.0000	0.000	696.4	0.620	BB	0.015	787.50	0.905
35	5.505		0.0000	0.000	418.2	0.372	BB	0.014	509.76	0.586
36	5.578		0.0000	0.000	1724.5	1.535	BB	0.021	1367.85	1.571
37	5.621		0.0000	0.000	72.2	0.064	BB	0.012	98.97	0.114
38	5.646		0.0000	0.000	129.2	0.115	BB	0.022	98.12	0.113
39	5.693		0.0000	0.000	5472.5	4.870	BB	0.037	2464.42	2.831
40	5.903		0.0000	0.000	1312.7	1.168	BB	0.029	748.61	0.860
41	6.016		0.0000	0.000	184.1	0.164	BB	0.025	124.09	0.143
42	6.097		0.0000	0.000	462.7	0.412	BB	0.022	355.20	0.408
43	6.151		0.0000	0.000	407.9	0.363	BB	0.021	316.48	0.364
44	6.357		0.0000	0.000	772.7	0.688	BB	0.040	318.76	0.366
45	6.990		0.0000	0.000	151.8	0.135	BB	0.035	72.99	0.084
46	7.160	DCB	20.8623	49.058	16730.9	14.888	BB	0.042	6618.95	7.602

Total Area = 112377.7, Total Amount = 42.526, Total Height = 87062.92, Sample Units = PPB

ASX22 0641

Sample Name: **AR5420924B AAAR542AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 06:12:35pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.16R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.621		0.0000	0.000	566.8	0.141	BV	0.066	143.42	0.038
2	2.703		0.0000	0.000	852.7	0.212	VB	0.048	295.85	0.079
3	2.808	TCX	21.2386	49.817	90281.5	22.419	BB	0.015	97175.13	25.942
4	2.929		0.0000	0.000	114.2	0.028	BB	0.019	99.85	0.027
5	3.030		0.0000	0.000	1304.9	0.324	BB	0.015	1429.47	0.382
6	3.148		0.0000	0.000	229.6	0.057	BB	0.013	297.23	0.079
7	3.308		0.0000	0.000	727.3	0.181	BB	0.035	349.59	0.093
8	3.403		0.0000	0.000	154.5	0.038	BV	0.020	130.41	0.035
9	3.432		0.0000	0.000	458.8	0.114	VV	0.020	387.98	0.104
10	3.495		0.0000	0.000	419.5	0.104	VV	0.017	402.28	0.107
11	3.531		0.0000	0.000	796.4	0.198	VV	0.025	522.78	0.140
12	3.580		0.0000	0.000	1005.9	0.250	VV	0.024	696.62	0.186
13	3.611		0.0000	0.000	895.6	0.222	VB	0.033	455.22	0.122
14	3.701		0.0000	0.000	472.8	0.117	BV	0.021	376.74	0.101
15	3.736		0.0000	0.000	8950.9	2.223	VV	0.014	10931.34	2.918
16	3.760		0.0000	0.000	4085.6	1.015	VV	0.019	3547.45	0.947
17	3.870		0.0000	0.000	5776.4	1.434	VB	0.015	6334.56	1.691
18	3.919		0.0000	0.000	488.5	0.121	BB	0.016	514.23	0.137
19	3.965		0.0000	0.000	2092.5	0.520	BV	0.016	2178.75	0.582
20	4.010		0.0000	0.000	433.9	0.108	VB	0.015	495.08	0.132
21	4.051		0.0000	0.000	347.1	0.086	BV	0.014	418.12	0.112
22	4.080		0.0000	0.000	1363.6	0.339	VV	0.013	1782.84	0.476
23	4.104		0.0000	0.000	19123.6	4.749	VV	0.015	21128.23	5.640
24	4.129		0.0000	0.000	2879.5	0.715	VV	0.016	3060.58	0.817
25	4.152		0.0000	0.000	3232.0	0.803	VV	0.016	3456.48	0.923
26	4.169		0.0000	0.000	4065.3	1.010	VV	0.015	4542.81	1.213
27	4.206		0.0000	0.000	16364.9	4.064	VV	0.016	17574.80	4.692
28	4.240		0.0000	0.000	7839.2	1.947	VV	0.016	8306.46	2.217
29	4.260		0.0000	0.000	8995.4	2.234	VV	0.020	7576.04	2.022
30	4.322		0.0000	0.000	1775.2	0.441	VV	0.020	1517.18	0.405
31	4.358		0.0000	0.000	6012.3	1.493	VB	0.016	6139.77	1.639
32	4.417		0.0000	0.000	10603.9	2.633	BV	0.016	11219.55	2.995
33	4.453		0.0000	0.000	3765.5	0.935	VV	0.018	3469.91	0.926
34	4.480		0.0000	0.000	20774.8	5.159	VB	0.017	19850.42	5.299
35	4.520		0.0000	0.000	1218.9	0.303	BB	0.015	1394.59	0.372
36	4.576		0.0000	0.000	11378.8	2.826	BV	0.018	10293.70	2.748
37	4.611		0.0000	0.000	1486.5	0.369	VV	0.018	1367.05	0.365
38	4.644		0.0000	0.000	11504.6	2.857	VV	0.017	11138.25	2.973
39	4.669		0.0000	0.000	4702.1	1.168	VV	0.019	4099.89	1.095
40	4.714		0.0000	0.000	12713.5	3.157	VB	0.023	9189.13	2.453
41	4.820		0.0000	0.000	10106.2	2.510	BV	0.020	8353.33	2.230
42	4.872		0.0000	0.000	9307.2	2.311	VV	0.021	7282.55	1.944
43	4.912		0.0000	0.000	3647.8	0.906	VV	0.021	2894.29	0.773
44	4.953		0.0000	0.000	26751.9	6.643	VB	0.030	14971.46	3.997
45	5.086		0.0000	0.000	174.7	0.043	BB	0.018	166.15	0.044
46	5.142		0.0000	0.000	647.3	0.161	BV	0.019	579.63	0.155
47	5.179		0.0000	0.000	1006.0	0.250	VV	0.016	1053.89	0.281
48	5.203		0.0000	0.000	5496.2	1.365	VV	0.020	4582.40	1.223
49	5.252		0.0000	0.000	922.4	0.229	VV	0.020	773.02	0.206
50	5.287		0.0000	0.000	563.5	0.140	VV	0.017	545.97	0.146
51	5.314		0.0000	0.000	256.4	0.064	VV	0.017	255.94	0.068
52	5.346		0.0000	0.000	2385.1	0.592	VV	0.018	2235.31	0.597
53	5.367		0.0000	0.000	3917.4	0.973	VB	0.025	2577.77	0.688
54	5.544		0.0000	0.000	111.6	0.028	BB	0.019	97.93	0.026
55	5.580		0.0000	0.000	88.0	0.022	BB	0.017	86.95	0.023
56	5.624		0.0000	0.000	2560.5	0.636	BB	0.023	1895.29	0.506
57	5.822		0.0000	0.000	81.5	0.020	BB	0.020	67.54	0.018
58	5.894		0.0000	0.000	169.3	0.042	BB	0.040	70.63	0.019
59	5.998		0.0000	0.000	88.1	0.022	BB	0.016	94.23	0.025
60	6.200		0.0000	0.000	690.7	0.172	BB	0.022	526.55	0.141
61	6.360	DCB	21.3947	50.183	63276.4	15.713	BB	0.021	51062.88	13.632
62	6.756		0.0000	0.000	114.9	0.029	BB	0.026	73.21	0.020

ASX22 8642

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	7.073		0.0000	0.000	56.8	0.014	BB	0.042	22.38	0.006
64	7.119		0.0000	0.000	21.2	0.005	BB	0.012	30.00	0.008

Total Area = 402695.8, Total Amount = 42.633, Total Height = 374589.1, Sample Units = PPB

ASX22 8643

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.16R

AR5420924B AAAR542AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.16R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 6:05:05 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.16R

AR5420924B AAAR542AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.16R



ASX22 8644

Instrument ID: CP06--S9147B Injected On: 11/6/2009 6:05:05 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	42376	21.66329	TCX	2.808	97175	21.23858	TCX
7.16	6619	20.86226	DCB	6.36	51063	21.39468	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.16A
Area File: C:\CPWIN\DATA1\1C6310B.16A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 6:12:40 PM
File Reported On: 11/6/2009 at 6:12:49 PM

REPLOTS

ASX22: 0645

AR5420924B AAAR542AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.16R



AR5420924B AAAR542AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.16R



ASXZZ 0646

Sample Name: **AR5430924B AAAR543AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 06:24:37pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.17R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.250		0.0000	0.000	158.1	0.070	BB	0.084	31.37	0.018
2	3.311	TCX	41.0426	50.869	74232.0	32.858	BB	0.015	80283.48	45.027
3	3.514		0.0000	0.000	1014.4	0.449	BB	0.015	1163.32	0.652
4	3.564		0.0000	0.000	146.5	0.065	BB	0.016	155.35	0.087
5	3.649		0.0000	0.000	200.6	0.089	BB	0.015	220.91	0.124
6	3.788		0.0000	0.000	292.0	0.129	BB	0.023	212.78	0.119
7	3.914		0.0000	0.000	209.2	0.093	BB	0.018	191.91	0.108
8	3.981		0.0000	0.000	145.4	0.064	BB	0.017	145.62	0.082
9	4.100		0.0000	0.000	103.3	0.046	BB	0.014	122.75	0.069
10	4.155		0.0000	0.000	130.5	0.058	BB	0.016	136.33	0.076
11	4.220		0.0000	0.000	184.8	0.082	BB	0.017	182.83	0.103
12	4.292		0.0000	0.000	5483.5	2.427	BB	0.016	5684.23	3.188
13	4.322		0.0000	0.000	1028.0	0.455	BB	0.014	1225.55	0.687
14	4.433		0.0000	0.000	3758.7	1.664	BB	0.020	3206.82	1.799
15	4.494		0.0000	0.000	162.7	0.072	BB	0.013	201.77	0.113
16	4.533		0.0000	0.000	1164.8	0.516	BB	0.017	1140.34	0.640
17	4.578		0.0000	0.000	233.9	0.104	BB	0.016	239.34	0.134
18	4.666		0.0000	0.000	207.0	0.092	BB	0.016	212.74	0.119
19	4.708		0.0000	0.000	13975.5	6.186	BV	0.028	8228.91	4.615
20	4.759		0.0000	0.000	4627.8	2.048	VV	0.029	2651.82	1.487
21	4.829		0.0000	0.000	2590.6	1.147	VV	0.022	1965.97	1.103
22	4.876		0.0000	0.000	13878.0	6.143	VV	0.023	10065.03	5.645
23	4.915		0.0000	0.000	7233.1	3.202	VV	0.030	4040.59	2.266
24	4.984		0.0000	0.000	1149.8	0.509	VV	0.025	770.33	0.432
25	5.032		0.0000	0.000	3051.6	1.351	VB	0.021	2412.33	1.353
26	5.089		0.0000	0.000	9042.4	4.003	BV	0.030	4955.72	2.779
27	5.155		0.0000	0.000	11390.9	5.042	VV	0.022	8674.51	4.865
28	5.202		0.0000	0.000	1595.5	0.706	VB	0.028	966.76	0.542
29	5.245		0.0000	0.000	833.2	0.369	BB	0.015	910.59	0.511
30	5.284		0.0000	0.000	3943.0	1.745	BB	0.021	3087.05	1.731
31	5.333		0.0000	0.000	270.8	0.120	BB	0.013	340.45	0.191
32	5.377		0.0000	0.000	5672.1	2.511	BB	0.022	4207.47	2.360
33	5.431		0.0000	0.000	401.4	0.178	BB	0.016	430.94	0.242
34	5.485		0.0000	0.000	1750.1	0.775	BB	0.015	1958.38	1.098
35	5.506		0.0000	0.000	851.2	0.377	BB	0.012	1228.61	0.689
36	5.579		0.0000	0.000	3898.8	1.726	BB	0.021	3112.21	1.745
37	5.621		0.0000	0.000	215.1	0.095	BB	0.012	288.32	0.162
38	5.649		0.0000	0.000	356.1	0.158	BB	0.019	308.83	0.173
39	5.694		0.0000	0.000	10959.4	4.851	BB	0.032	5737.08	3.218
40	5.796		0.0000	0.000	201.2	0.089	BB	0.019	179.03	0.100
41	5.904		0.0000	0.000	3035.2	1.343	BB	0.030	1684.35	0.945
42	5.976		0.0000	0.000	164.3	0.073	BB	0.015	181.56	0.102
43	6.016		0.0000	0.000	250.4	0.111	BB	0.022	190.44	0.107
44	6.098		0.0000	0.000	1068.7	0.473	BB	0.022	809.03	0.454
45	6.152		0.0000	0.000	819.0	0.363	BB	0.019	700.95	0.393
46	6.360		0.0000	0.000	1446.1	0.640	BB	0.037	653.75	0.367
47	6.776		0.0000	0.000	94.7	0.042	BB	0.024	64.97	0.036
48	6.990		0.0000	0.000	328.2	0.145	BB	0.034	161.88	0.091
49	7.161	DCB	39.6403	49.131	31966.3	14.150	BB	0.042	12576.64	7.054

Total Area = 225915.9, Total Amount = 80.683, Total Height = 178301.9, Sample Units = PPB

ASX22: 3647

Sample Name: **AR5430924B AAAR543AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 06:24:37pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.17R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.537		0.0000	0.000	84.6	0.010	BB	0.021	68.11	0.009
2	2.611		0.0000	0.000	474.3	0.058	BV	0.064	124.47	0.016
3	2.703		0.0000	0.000	481.2	0.059	VB	0.018	439.79	0.055
4	2.808	TCX	41.5105	50.496	165046.8	20.309	BB	0.014	189927.60	23.795
5	3.030		0.0000	0.000	2299.4	0.283	BB	0.013	2916.64	0.365
6	3.098		0.0000	0.000	274.9	0.034	BB	0.013	363.67	0.046
7	3.149		0.0000	0.000	520.3	0.064	BB	0.014	604.04	0.076
8	3.309		0.0000	0.000	831.0	0.102	BV	0.019	718.49	0.090
9	3.433		0.0000	0.000	469.0	0.058	VB	0.014	578.75	0.073
10	3.496		0.0000	0.000	501.2	0.062	BV	0.016	528.98	0.066
11	3.531		0.0000	0.000	501.8	0.062	VV	0.013	622.21	0.078
12	3.581		0.0000	0.000	1644.8	0.202	VV	0.021	1331.79	0.167
13	3.612		0.0000	0.000	1144.2	0.141	VB	0.026	743.71	0.093
14	3.701		0.0000	0.000	492.4	0.061	BB	0.013	644.05	0.081
15	3.736		0.0000	0.000	19741.9	2.429	BV	0.013	24855.63	3.114
16	3.760		0.0000	0.000	8478.8	1.043	VB	0.018	7735.73	0.969
17	3.870		0.0000	0.000	12412.7	1.527	BB	0.015	14108.69	1.768
18	3.919		0.0000	0.000	815.6	0.100	BB	0.014	982.71	0.123
19	3.966		0.0000	0.000	4356.9	0.536	BV	0.015	4968.08	0.622
20	4.010		0.0000	0.000	925.6	0.114	VB	0.014	1084.08	0.136
21	4.051		0.0000	0.000	794.2	0.098	BV	0.014	966.96	0.121
22	4.080		0.0000	0.000	3065.5	0.377	VV	0.013	4037.99	0.506
23	4.104		0.0000	0.000	42986.8	5.289	VV	0.015	49194.87	6.163
24	4.129		0.0000	0.000	6545.0	0.805	VV	0.016	6864.15	0.860
25	4.151		0.0000	0.000	7362.3	0.906	VV	0.016	7871.84	0.986
26	4.169		0.0000	0.000	9324.4	1.147	VV	0.015	10111.50	1.267
27	4.207		0.0000	0.000	38645.8	4.755	VV	0.016	40222.61	5.039
28	4.240		0.0000	0.000	18420.5	2.267	VV	0.016	19528.10	2.447
29	4.260		0.0000	0.000	20630.0	2.538	VV	0.020	17140.76	2.147
30	4.323		0.0000	0.000	3799.4	0.468	VV	0.018	3427.41	0.429
31	4.359		0.0000	0.000	13785.9	1.696	VB	0.016	14091.28	1.765
32	4.417		0.0000	0.000	25110.1	3.090	BV	0.015	27003.67	3.383
33	4.453		0.0000	0.000	8147.8	1.003	VV	0.017	7879.59	0.987
34	4.481		0.0000	0.000	47143.1	5.801	VB	0.017	45532.80	5.704
35	4.520		0.0000	0.000	2808.9	0.346	BB	0.014	3248.79	0.407
36	4.576		0.0000	0.000	26365.4	3.244	BV	0.018	24026.00	3.010
37	4.611		0.0000	0.000	3453.6	0.425	VV	0.018	3241.63	0.406
38	4.645		0.0000	0.000	26692.9	3.284	VV	0.017	25753.42	3.226
39	4.670		0.0000	0.000	11039.3	1.358	VV	0.020	9345.48	1.171
40	4.714		0.0000	0.000	28131.5	3.462	VB	0.022	21273.54	2.665
41	4.820		0.0000	0.000	23620.5	2.906	BV	0.019	20715.10	2.595
42	4.873		0.0000	0.000	19167.5	2.359	VV	0.019	16449.25	2.061
43	4.913		0.0000	0.000	5351.3	0.658	VB	0.016	5504.57	0.690
44	4.954		0.0000	0.000	41109.3	5.058	BB	0.021	32338.76	4.051
45	5.002		0.0000	0.000	951.2	0.117	BB	0.010	1652.46	0.207
46	5.087		0.0000	0.000	298.0	0.037	BB	0.015	322.35	0.040
47	5.142		0.0000	0.000	1427.4	0.176	BB	0.019	1279.58	0.160
48	5.180		0.0000	0.000	2010.0	0.247	BV	0.015	2294.73	0.287
49	5.204		0.0000	0.000	12471.5	1.535	VV	0.019	10690.05	1.339
50	5.253		0.0000	0.000	1718.1	0.211	VV	0.018	1599.73	0.200
51	5.288		0.0000	0.000	928.3	0.114	VV	0.014	1082.19	0.136
52	5.316		0.0000	0.000	274.5	0.034	VB	0.013	358.58	0.045
53	5.347		0.0000	0.000	2661.3	0.327	BB	0.014	3128.15	0.392
54	5.368		0.0000	0.000	2933.1	0.361	BB	0.016	3150.97	0.395
55	5.583		0.0000	0.000	153.3	0.019	BB	0.014	178.56	0.022
56	5.625		0.0000	0.000	6240.2	0.768	BB	0.023	4458.69	0.559
57	5.822		0.0000	0.000	256.6	0.032	BB	0.023	185.13	0.023
58	5.899		0.0000	0.000	245.4	0.030	BB	0.026	158.12	0.020
59	6.000		0.0000	0.000	322.2	0.040	BB	0.021	258.25	0.032
60	6.200		0.0000	0.000	1202.0	0.148	BB	0.020	1023.95	0.128
61	6.360	DCB	40.6948	49.504	123347.1	15.178	BB	0.021	97126.73	12.168
62	6.803		0.0000	0.000	141.9	0.017	BB	0.038	62.24	0.008

ASX22. 8648

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.989		0.0000	0.000	114.0	0.014	BB	0.029	64.73	0.008

Total Area = 812694.8, Total Amount = 82.205, Total Height = 798192.4, Sample Units = PPB

ASX22 8649

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.17R

AR5430924B AAAR543AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.17R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 6:17:07 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.17R

AR5430924B AAAR543AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.17R



Instrument ID: CP06--S9147B Injected On: 11/6/2009 6:17:07 PM Column ID: RTX CLPII

ASX22 0658

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	80283	41.04264	TCX	2.808	189928	41.51053	TCX
7.161	12577	39.64031	DCB	6.36	97127	40.69483	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.17A
Area File: C:\CPWIN\DATA1\1C6310B.17A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 6:24:42 PM
File Reported On: 11/6/2009 at 6:24:51 PM

REPLOTS

ASX22 0651

AR5430924B AAAR543AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.17R



AR5430924B AAAR543AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.17R



ASX22 0652

Sample Name: **AR5440924B AAAR544AA ICAL 0930999999**
Acquired from CP06–S9147A via port 1 on 11/6/09 06:36:36pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.18R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.081		0.0000	0.000	440.3	0.105	BB	0.035	208.19	0.064
2	3.215		0.0000	0.000	93.2	0.022	BB	0.021	73.07	0.023
3	3.311	TCX	65.2744	50.600	118799.7	28.244	BB	0.016	127683.30	39.362
4	3.513		0.0000	0.000	1580.5	0.376	BB	0.014	1846.07	0.569
5	3.564		0.0000	0.000	259.9	0.062	BB	0.018	241.71	0.075
6	3.648		0.0000	0.000	356.9	0.085	BB	0.019	320.57	0.099
7	3.733		0.0000	0.000	55.8	0.013	BB	0.021	45.28	0.014
8	3.788		0.0000	0.000	229.4	0.055	BB	0.012	310.71	0.096
9	3.839		0.0000	0.000	119.4	0.028	BB	0.025	78.68	0.024
10	3.913		0.0000	0.000	333.3	0.079	BB	0.016	358.12	0.110
11	3.979		0.0000	0.000	222.4	0.053	BB	0.016	226.38	0.070
12	4.033		0.0000	0.000	93.6	0.022	BB	0.021	75.95	0.023
13	4.099		0.0000	0.000	219.1	0.052	BB	0.016	230.05	0.071
14	4.154		0.0000	0.000	250.0	0.059	BB	0.016	255.08	0.079
15	4.220		0.0000	0.000	338.6	0.081	BB	0.017	336.41	0.104
16	4.291		0.0000	0.000	10926.7	2.598	BB	0.016	11593.05	3.574
17	4.321		0.0000	0.000	1891.2	0.450	BB	0.014	2236.79	0.690
18	4.432		0.0000	0.000	7550.5	1.795	BB	0.020	6407.36	1.975
19	4.493		0.0000	0.000	333.7	0.079	BB	0.014	391.77	0.121
20	4.532		0.0000	0.000	2395.0	0.569	BB	0.018	2262.56	0.698
21	4.578		0.0000	0.000	411.2	0.098	BB	0.016	429.83	0.133
22	4.667		0.0000	0.000	399.1	0.095	BB	0.017	397.37	0.123
23	4.708		0.0000	0.000	27892.3	6.631	BV	0.028	16815.39	5.184
24	4.758		0.0000	0.000	9181.1	2.183	VV	0.030	5136.20	1.583
25	4.828		0.0000	0.000	5210.1	1.239	VV	0.022	3873.17	1.194
26	4.875		0.0000	0.000	27470.9	6.531	VV	0.022	20556.20	6.337
27	4.914		0.0000	0.000	14946.9	3.554	VV	0.030	8370.23	2.580
28	4.985		0.0000	0.000	2025.6	0.482	VV	0.023	1490.51	0.459
29	5.032		0.0000	0.000	6119.7	1.455	VB	0.020	5035.90	1.552
30	5.088		0.0000	0.000	18934.3	4.502	BV	0.030	10659.71	3.286
31	5.154		0.0000	0.000	23610.8	5.613	VV	0.022	18019.88	5.555
32	5.202		0.0000	0.000	2843.5	0.676	VB	0.027	1783.92	0.550
33	5.244		0.0000	0.000	1691.4	0.402	BB	0.015	1908.78	0.588
34	5.283		0.0000	0.000	8008.5	1.904	BB	0.020	6551.89	2.020
35	5.332		0.0000	0.000	480.7	0.114	BB	0.013	629.44	0.194
36	5.376		0.0000	0.000	11905.0	2.830	BB	0.021	9377.59	2.891
37	5.431		0.0000	0.000	862.2	0.205	BB	0.015	971.71	0.300
38	5.484		0.0000	0.000	15844.2	3.767	BB	0.039	6819.51	2.102
39	5.578		0.0000	0.000	8468.2	2.013	BB	0.020	7108.13	2.191
40	5.620		0.0000	0.000	426.0	0.101	BB	0.012	573.34	0.177
41	5.648		0.0000	0.000	793.8	0.189	BB	0.021	630.94	0.195
42	5.694		0.0000	0.000	22384.6	5.322	BB	0.030	12386.36	3.818
43	5.796		0.0000	0.000	430.5	0.102	BB	0.018	404.31	0.125
44	5.903		0.0000	0.000	6297.1	1.497	BB	0.029	3559.30	1.097
45	5.975		0.0000	0.000	324.1	0.077	BB	0.016	348.23	0.107
46	6.014		0.0000	0.000	385.9	0.092	BB	0.018	353.17	0.109
47	6.097		0.0000	0.000	2220.3	0.528	BB	0.023	1643.48	0.507
48	6.151		0.0000	0.000	1781.4	0.424	BB	0.020	1492.94	0.460
49	6.358		0.0000	0.000	2925.2	0.695	BB	0.036	1371.46	0.423
50	6.770		0.0000	0.000	148.0	0.035	BB	0.035	70.59	0.022
51	6.988		0.0000	0.000	352.9	0.084	BB	0.028	211.89	0.065
52	7.160	DCB	63.7276	49.400	49350.7	11.733	BB	0.041	20218.79	6.233

Total Area = 420615.1, Total Amount = 129.002, Total Height = 324381.2, Sample Units = PPB

ASX22 0653

Sample Name: **AR5440924B AAAR544AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 06:36:36pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.18R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.613		0.0000	0.000	994.5	0.064	BV	0.062	268.78	0.018
2	2.701		0.0000	0.000	347.8	0.022	VB	0.035	166.26	0.011
3	2.808 TCX		64.4874	50.342	252034.2	16.284	BB	0.014	295055.90	19.977
4	2.927		0.0000	0.000	128.3	0.008	BB	0.011	200.37	0.014
5	2.977		0.0000	0.000	100.4	0.006	BB	0.015	112.21	0.008
6	3.029		0.0000	0.000	3768.6	0.243	BB	0.014	4644.45	0.314
7	3.097		0.0000	0.000	598.5	0.039	BB	0.014	702.47	0.048
8	3.148		0.0000	0.000	687.9	0.044	BB	0.013	906.18	0.061
9	3.210		0.0000	0.000	101.9	0.007	BB	0.012	136.79	0.009
10	3.308		0.0000	0.000	1523.4	0.098	BV	0.019	1358.56	0.092
11	3.360		0.0000	0.000	128.4	0.008	VB	0.016	135.31	0.009
12	3.401		0.0000	0.000	120.9	0.008	BV	0.014	147.17	0.010
13	3.431		0.0000	0.000	1184.8	0.077	VV	0.016	1260.36	0.085
14	3.494		0.0000	0.000	928.5	0.060	VV	0.018	874.48	0.059
15	3.529		0.0000	0.000	404.9	0.026	VV	0.014	489.76	0.033
16	3.580		0.0000	0.000	3275.2	0.212	VV	0.021	2586.92	0.175
17	3.610		0.0000	0.000	2138.3	0.138	VB	0.024	1455.52	0.099
18	3.700		0.0000	0.000	978.2	0.063	BB	0.013	1255.67	0.085
19	3.735		0.0000	0.000	39009.8	2.520	BV	0.013	48285.46	3.269
20	3.760		0.0000	0.000	16068.8	1.038	VB	0.017	15371.50	1.041
21	3.870		0.0000	0.000	24776.9	1.601	BB	0.015	28288.42	1.915
22	3.918		0.0000	0.000	1672.0	0.108	BB	0.014	2015.12	0.136
23	3.965		0.0000	0.000	8832.3	0.571	BV	0.015	9978.87	0.676
24	4.009		0.0000	0.000	1940.3	0.125	VB	0.015	2220.61	0.150
25	4.051		0.0000	0.000	1649.9	0.107	BV	0.014	1966.59	0.133
26	4.079		0.0000	0.000	6717.7	0.434	VV	0.013	8459.87	0.573
27	4.103		0.0000	0.000	97842.5	6.322	VV	0.017	96676.34	6.545
28	4.150		0.0000	0.000	14323.9	0.925	VV	0.015	15526.80	1.051
29	4.169		0.0000	0.000	18003.7	1.163	VV	0.015	20530.73	1.390
30	4.206		0.0000	0.000	73783.3	4.767	VV	0.015	80044.55	5.419
31	4.239		0.0000	0.000	37929.0	2.451	VV	0.016	38437.35	2.602
32	4.260		0.0000	0.000	38058.8	2.459	VV	0.019	33410.79	2.262
33	4.322		0.0000	0.000	7092.7	0.458	VV	0.018	6737.03	0.456
34	4.358		0.0000	0.000	27783.1	1.795	VB	0.016	28578.23	1.935
35	4.416		0.0000	0.000	50771.2	3.280	BV	0.016	54016.47	3.657
36	4.451		0.0000	0.000	18313.0	1.183	VV	0.018	17371.31	1.176
37	4.480		0.0000	0.000	97706.6	6.313	VB	0.017	95349.38	6.456
38	4.520		0.0000	0.000	6279.6	0.406	BB	0.015	7044.86	0.477
39	4.575		0.0000	0.000	53428.7	3.452	BV	0.018	48303.57	3.270
40	4.610		0.0000	0.000	7234.9	0.467	VV	0.017	7049.14	0.477
41	4.644		0.0000	0.000	56917.2	3.677	VV	0.017	55575.37	3.763
42	4.669		0.0000	0.000	20091.0	1.298	VV	0.018	18707.95	1.267
43	4.713		0.0000	0.000	57061.9	3.687	VB	0.021	44731.79	3.029
44	4.819		0.0000	0.000	46119.8	2.980	BV	0.019	40302.61	2.729
45	4.871		0.0000	0.000	44760.4	2.892	VV	0.020	37162.57	2.516
46	4.912		0.0000	0.000	16456.9	1.063	VV	0.020	14038.29	0.950
47	4.952		0.0000	0.000	125102.9	8.083	VB	0.028	73507.62	4.977
48	5.085		0.0000	0.000	778.1	0.050	BB	0.016	801.80	0.054
49	5.141		0.0000	0.000	3069.4	0.198	BV	0.017	2944.75	0.199
50	5.179		0.0000	0.000	4549.6	0.294	VV	0.015	4918.02	0.333
51	5.203		0.0000	0.000	25555.5	1.651	VV	0.019	22627.32	1.532
52	5.251		0.0000	0.000	3384.8	0.219	VV	0.017	3282.11	0.222
53	5.287		0.0000	0.000	1967.4	0.127	VV	0.015	2217.33	0.150
54	5.314		0.0000	0.000	609.8	0.039	VB	0.013	775.44	0.053
55	5.345		0.0000	0.000	6140.4	0.397	BB	0.015	6792.80	0.460
56	5.367		0.0000	0.000	5684.5	0.367	BB	0.014	6745.40	0.457
57	5.546		0.0000	0.000	98.2	0.006	BB	0.017	95.88	0.006
58	5.581		0.0000	0.000	320.0	0.021	BB	0.014	380.91	0.026
59	5.624		0.0000	0.000	12871.0	0.832	BB	0.023	9280.56	0.628
60	5.821		0.0000	0.000	524.1	0.034	BB	0.022	395.96	0.027
61	5.896		0.0000	0.000	428.0	0.028	BB	0.023	315.93	0.021
62	5.999		0.0000	0.000	465.6	0.030	BB	0.018	424.98	0.029

ASX22 B654

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.199		0.0000	0.000	1982.1	0.128	BB	0.020	1617.52	0.110
64	6.359	DCB	63.6108	49.658	193899.3	12.528	BB	0.021	151820.50	10.279
65	6.803		0.0000	0.000	134.5	0.009	BB	0.037	61.16	0.004
66	6.982		0.0000	0.000	87.4	0.006	BB	0.025	57.23	0.004

Total Area = 1547723.0, Total Amount = 128.098, Total Height = 1477002.0, Sample Units = PPB

ASX22 8655

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.18R

AR5440924B AAAR544AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.18R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 6:29:06 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.18R

AR5440924B AAAR544AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.18R



ASX22 0656

Instrument ID: CP06--S9147B Injected On: 11/6/2009 6:29:06 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	127683	65.27444	TCX	2.808	295056	64.48737	TCX
7.16	20219	63.72761	DCB	6.359	151820	63.61079	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.18A
Area File: C:\CPWIN\DATA1\1C6310B.18A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 6:36:42 PM
File Reported On: 11/6/2009 at 6:36:50 PM

REPLOTS

ASX22. P657

AR5440924B AAAR544AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.18R



AR5440924B AAAR544AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.18R



ASX22 0658

Sample Name: **AR5450924B AAAR545AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 06:48:37pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.19R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.310	TCX	81.3885	49.973	146576.2	22.090	BB	0.015	159204.00	30.561
2	3.453		0.0000	0.000	89.0	0.013	BB	0.015	99.37	0.019
3	3.512		0.0000	0.000	2013.6	0.303	BB	0.014	2383.53	0.458
4	3.563		0.0000	0.000	456.9	0.069	BB	0.015	516.23	0.099
5	3.648		0.0000	0.000	464.9	0.070	BB	0.018	441.65	0.085
6	3.787		0.0000	0.000	482.4	0.073	BB	0.013	634.72	0.122
7	3.805		0.0000	0.000	105.8	0.016	BB	0.008	235.20	0.045
8	3.843		0.0000	0.000	148.3	0.022	BB	0.026	95.37	0.018
9	3.913		0.0000	0.000	661.7	0.100	BB	0.016	701.66	0.135
10	3.979		0.0000	0.000	310.6	0.047	BB	0.017	306.64	0.059
11	4.030		0.0000	0.000	167.8	0.025	BB	0.020	136.50	0.026
12	4.114		0.0000	0.000	2453.9	0.370	BB	0.035	1164.79	0.224
13	4.154		0.0000	0.000	463.9	0.070	BB	0.016	490.34	0.094
14	4.219		0.0000	0.000	639.3	0.096	BB	0.017	630.56	0.121
15	4.291		0.0000	0.000	19865.5	2.994	BB	0.016	20938.28	4.019
16	4.320		0.0000	0.000	3523.1	0.531	BB	0.014	4231.78	0.812
17	4.431		0.0000	0.000	14106.2	2.126	BB	0.019	12097.66	2.322
18	4.491		0.0000	0.000	618.8	0.093	BB	0.015	702.30	0.135
19	4.531		0.0000	0.000	4483.8	0.676	BB	0.018	4175.06	0.801
20	4.577		0.0000	0.000	732.4	0.110	BB	0.016	768.49	0.148
21	4.666		0.0000	0.000	727.7	0.110	BB	0.016	753.50	0.145
22	4.708		0.0000	0.000	51959.2	7.831	BV	0.028	31252.46	5.999
23	4.758		0.0000	0.000	16683.5	2.514	VV	0.028	9770.41	1.876
24	4.827		0.0000	0.000	9600.1	1.447	VV	0.021	7537.18	1.447
25	4.874		0.0000	0.000	51548.0	7.769	VV	0.022	38433.30	7.378
26	4.913		0.0000	0.000	26916.6	4.056	VV	0.030	14938.87	2.868
27	4.984		0.0000	0.000	3609.0	0.544	VV	0.022	2742.03	0.526
28	5.031		0.0000	0.000	11704.6	1.764	VB	0.020	9631.10	1.849
29	5.087		0.0000	0.000	28160.1	4.244	BV	0.024	19808.03	3.802
30	5.120		0.0000	0.000	8170.8	1.231	VV	0.018	7399.86	1.420
31	5.153		0.0000	0.000	44695.0	6.736	VV	0.022	33333.74	6.399
32	5.203		0.0000	0.000	5590.3	0.842	VB	0.027	3425.59	0.658
33	5.243		0.0000	0.000	3561.7	0.537	BB	0.015	3879.95	0.745
34	5.283		0.0000	0.000	15304.9	2.307	BB	0.020	12513.18	2.402
35	5.331		0.0000	0.000	1197.5	0.180	BB	0.015	1357.29	0.261
36	5.375		0.0000	0.000	23353.1	3.519	BB	0.022	18017.22	3.459
37	5.430		0.0000	0.000	1610.6	0.243	BB	0.015	1802.07	0.346
38	5.484		0.0000	0.000	6589.1	0.993	BB	0.016	6897.90	1.324
39	5.504		0.0000	0.000	3500.4	0.528	BB	0.013	4624.43	0.888
40	5.577		0.0000	0.000	16418.9	2.474	BB	0.020	13525.62	2.596
41	5.620		0.0000	0.000	1043.9	0.157	BB	0.013	1320.34	0.253
42	5.647		0.0000	0.000	1634.6	0.246	BB	0.020	1357.02	0.260
43	5.693		0.0000	0.000	42852.1	6.458	BB	0.031	23390.52	4.490
44	5.795		0.0000	0.000	527.8	0.080	BB	0.014	644.78	0.124
45	5.903		0.0000	0.000	11961.5	1.803	BB	0.029	6983.30	1.341
46	5.976		0.0000	0.000	664.7	0.100	BB	0.015	723.76	0.139
47	6.014		0.0000	0.000	630.5	0.095	BB	0.018	581.64	0.112
48	6.097		0.0000	0.000	4087.6	0.616	BB	0.023	2981.93	0.572
49	6.151		0.0000	0.000	3156.0	0.476	BB	0.021	2545.52	0.489
50	6.359		0.0000	0.000	5134.7	0.774	BB	0.033	2584.90	0.496
51	6.773		0.0000	0.000	191.8	0.029	BB	0.026	125.36	0.024
52	6.989		0.0000	0.000	473.9	0.071	BB	0.031	257.21	0.049
53	7.160	DCB	81.4757	50.027	61922.5	9.332	BB	0.040	25849.72	4.962

Total Area = 663546.6, Total Amount = 162.864, Total Height = 520944.0, Sample Units = PPB

ASX22. 0659

Sample Name: **AR5450924B AAAR545AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 06:48:37pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.19R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.611		0.0000	0.000	302.5	0.012	BV	0.046	109.60	0.004
2	2.702		0.0000	0.000	391.7	0.015	VB	0.031	211.95	0.008
3	2.807	TCX	82.9938	50.290	317082.2	12.521	BB	0.014	379730.20	15.156
4	2.979		0.0000	0.000	171.1	0.007	BB	0.015	195.55	0.008
5	3.029		0.0000	0.000	4777.9	0.189	BB	0.013	6103.63	0.244
6	3.097		0.0000	0.000	1076.6	0.043	BB	0.012	1471.17	0.059
7	3.147		0.0000	0.000	931.2	0.037	BB	0.012	1263.58	0.050
8	3.211		0.0000	0.000	190.7	0.008	BB	0.013	242.85	0.010
9	3.307		0.0000	0.000	3089.6	0.122	BV	0.019	2752.11	0.110
10	3.360		0.0000	0.000	312.2	0.012	VV	0.017	299.19	0.012
11	3.401		0.0000	0.000	204.5	0.008	VV	0.015	221.63	0.009
12	3.431		0.0000	0.000	1884.8	0.074	VV	0.014	2271.23	0.091
13	3.494		0.0000	0.000	783.0	0.031	VV	0.012	1083.20	0.043
14	3.513		0.0000	0.000	479.5	0.019	VV	0.012	653.12	0.026
15	3.529		0.0000	0.000	540.3	0.021	VB	0.013	710.86	0.028
16	3.579		0.0000	0.000	6370.6	0.252	BV	0.022	4937.21	0.197
17	3.610		0.0000	0.000	4180.8	0.165	VB	0.025	2808.13	0.112
18	3.700		0.0000	0.000	1886.6	0.074	BB	0.013	2465.61	0.098
19	3.735		0.0000	0.000	72445.1	2.861	BV	0.014	88941.33	3.550
20	3.759		0.0000	0.000	29445.9	1.163	VB	0.017	28927.22	1.155
21	3.869		0.0000	0.000	45762.5	1.807	BB	0.015	51781.76	2.067
22	3.917		0.0000	0.000	2871.3	0.113	BB	0.012	3862.55	0.154
23	3.964		0.0000	0.000	16834.9	0.665	BV	0.015	19330.18	0.771
24	4.009		0.0000	0.000	3702.2	0.146	VB	0.014	4306.85	0.172
25	4.050		0.0000	0.000	3297.2	0.130	BV	0.014	3791.42	0.151
26	4.079		0.0000	0.000	13215.2	0.522	VV	0.014	16182.03	0.646
27	4.103		0.0000	0.000	179879.0	7.103	VV	0.017	174728.20	6.974
28	4.150		0.0000	0.000	23212.1	0.917	VV	0.013	29305.13	1.170
29	4.168		0.0000	0.000	36848.9	1.455	VV	0.016	38722.34	1.545
30	4.205		0.0000	0.000	138695.4	5.477	VV	0.015	150858.10	6.021
31	4.238		0.0000	0.000	73478.7	2.901	VV	0.017	72631.84	2.899
32	4.259		0.0000	0.000	68504.6	2.705	VV	0.018	61998.45	2.474
33	4.321		0.0000	0.000	12973.5	0.512	VV	0.017	12646.95	0.505
34	4.357		0.0000	0.000	52244.6	2.063	VB	0.016	53618.10	2.140
35	4.416		0.0000	0.000	101477.0	4.007	BV	0.015	109270.10	4.361
36	4.451		0.0000	0.000	35311.5	1.394	VV	0.018	33629.39	1.342
37	4.479		0.0000	0.000	179905.3	7.104	VB	0.017	175707.60	7.013
38	4.518		0.0000	0.000	12851.7	0.507	BB	0.015	14390.73	0.574
39	4.574		0.0000	0.000	99433.7	3.926	BV	0.018	92756.13	3.702
40	4.609		0.0000	0.000	13800.3	0.545	VV	0.018	13036.10	0.520
41	4.643		0.0000	0.000	109707.4	4.332	VV	0.017	105603.00	4.215
42	4.669		0.0000	0.000	35578.2	1.405	VV	0.017	34515.45	1.378
43	4.713		0.0000	0.000	105107.4	4.150	VB	0.021	82999.39	3.313
44	4.818		0.0000	0.000	88566.4	3.497	BV	0.018	80269.48	3.204
45	4.871		0.0000	0.000	77648.5	3.066	VV	0.019	69843.09	2.788
46	4.912		0.0000	0.000	21820.6	0.862	VB	0.017	21813.16	0.871
47	4.952		0.0000	0.000	160256.2	6.328	BB	0.021	128665.60	5.135
48	5.000		0.0000	0.000	4767.6	0.188	BB	0.010	7819.40	0.312
49	5.084		0.0000	0.000	1056.8	0.042	BB	0.015	1171.35	0.047
50	5.139		0.0000	0.000	5755.4	0.227	BB	0.018	5396.52	0.215
51	5.179		0.0000	0.000	8500.7	0.336	BV	0.016	8956.51	0.357
52	5.202		0.0000	0.000	47852.3	1.890	VV	0.018	43328.31	1.729
53	5.251		0.0000	0.000	5929.2	0.234	VB	0.016	6044.62	0.241
54	5.286		0.0000	0.000	3478.4	0.137	BV	0.014	4050.30	0.162
55	5.313		0.0000	0.000	1087.1	0.043	VB	0.013	1423.08	0.057
56	5.345		0.0000	0.000	11705.3	0.462	BB	0.014	14137.70	0.564
57	5.366		0.0000	0.000	10054.3	0.397	BB	0.013	12674.58	0.506
58	5.580		0.0000	0.000	597.8	0.024	BB	0.015	679.23	0.027
59	5.623		0.0000	0.000	24839.1	0.981	BB	0.023	17887.01	0.714
60	5.820		0.0000	0.000	1070.4	0.042	BB	0.023	791.90	0.032
61	5.895		0.0000	0.000	719.0	0.028	BB	0.021	573.81	0.023
62	6.000		0.0000	0.000	1034.9	0.041	BB	0.020	876.97	0.035

A3X22 8668

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.199		0.0000	0.000	2418.3	0.095	BB	0.020	2062.82	0.082
64	6.359	DCB	82.0355	49.710	241732.5	9.545	BB	0.021	195794.90	7.814
65	6.803		0.0000	0.000	103.0	0.004	BB	0.028	60.62	0.002
66	6.956		0.0000	0.000	35.4	0.001	BB	0.013	45.86	0.002
67	7.172		0.0000	0.000	202.5	0.008	BB	0.030	110.68	0.004

Total Area = 2532471.0, Total Amount = 165.029, Total Height = 2505549.0, Sample Units = PPB

ASX22 0661

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.19R
AR5450924B AAAR545AA ICAL 0930999999
——— C:\CPWIN\DATA1\1C6310.19R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 6:41:07 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.19R
AR5450924B AAAR545AA ICAL 0930999999
——— C:\CPWIN\DATA1\1C6310B.19R



ASX22 0662

Instrument ID: CP06–S9147B Injected On: 11/6/2009 6:41:07 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.31	159204	81.38853	TCX	2.807	379730	82.99376	TCX
7.16	25850	81.47575	DCB	6.359	195795	82.0355	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.19A
Area File: C:\CPWIN\DATA1\1C6310B.19A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 6:48:42 PM
File Reported On: 11/6/2009 at 6:48:51 PM

REPLOTS

ASX22 0663

AR5450924B AAAR545AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.19R



AR5450924B AAAR545AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.19R



ASX22 8664

Sample Name: **AR6210924B AAAR621AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 07:00:36pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.20R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.141		0.0000	0.000	163.2	0.455	BB	0.039	69.67	0.293
2	3.217		0.0000	0.000	232.9	0.649	BB	0.046	84.24	0.354
3	3.309	TCX	0.0753	13.081	230.8	0.643	BB	0.026	147.28	0.619
4	3.563		0.0000	0.000	72.5	0.202	BB	0.017	70.42	0.296
5	3.786		0.0000	0.000	119.5	0.333	BB	0.019	106.12	0.446
6	3.915		0.0000	0.000	99.0	0.276	BB	0.018	92.42	0.388
7	4.072		0.0000	0.000	40.3	0.112	BB	0.012	56.07	0.236
8	4.154		0.0000	0.000	88.8	0.247	BB	0.017	84.58	0.355
9	4.290		0.0000	0.000	111.5	0.311	BB	0.016	118.48	0.498
10	4.432		0.0000	0.000	92.2	0.257	BB	0.019	79.20	0.333
11	4.534		0.0000	0.000	148.8	0.415	BB	0.045	55.08	0.231
12	4.708		0.0000	0.000	696.6	1.940	BV	0.023	508.48	2.137
13	4.827		0.0000	0.000	84.5	0.236	VB	0.022	65.33	0.275
14	4.875		0.0000	0.000	465.4	1.296	BB	0.019	408.35	1.716
15	5.027		0.0000	0.000	85.9	0.239	BB	0.026	54.63	0.230
16	5.107		0.0000	0.000	584.0	1.627	BV	0.024	402.48	1.692
17	5.151		0.0000	0.000	308.3	0.859	VB	0.023	224.67	0.944
18	5.202		0.0000	0.000	1439.0	4.009	BB	0.022	1073.90	4.513
19	5.282		0.0000	0.000	2373.3	6.611	BB	0.023	1696.34	7.129
20	5.372		0.0000	0.000	138.0	0.385	BB	0.017	132.29	0.556
21	5.432		0.0000	0.000	243.9	0.679	BB	0.024	172.45	0.725
22	5.484		0.0000	0.000	2170.3	6.046	BB	0.024	1520.05	6.388
23	5.538		0.0000	0.000	888.6	2.476	BB	0.018	837.11	3.518
24	5.577		0.0000	0.000	931.6	2.595	BB	0.028	549.31	2.309
25	5.691		0.0000	0.000	1698.3	4.731	BB	0.023	1225.65	5.151
26	5.754		0.0000	0.000	2512.2	6.998	BB	0.020	2140.31	8.995
27	5.796		0.0000	0.000	939.4	2.617	BB	0.019	828.70	3.483
28	5.892		0.0000	0.000	175.0	0.488	BB	0.019	157.37	0.661
29	5.918		0.0000	0.000	2141.0	5.964	BB	0.023	1581.53	6.647
30	5.976		0.0000	0.000	1548.0	4.312	BB	0.037	705.39	2.965
31	6.096		0.0000	0.000	253.6	0.707	BB	0.018	233.40	0.981
32	6.150		0.0000	0.000	6638.3	18.493	BB	0.034	3243.51	13.632
33	6.354		0.0000	0.000	2338.2	6.514	BB	0.021	1841.70	7.740
34	6.406		0.0000	0.000	1653.1	4.605	BB	0.027	1020.85	4.290
35	6.543		0.0000	0.000	149.5	0.416	BB	0.022	112.42	0.472
36	6.624		0.0000	0.000	897.7	2.501	BB	0.030	497.36	2.090
37	6.774		0.0000	0.000	2132.1	5.940	BB	0.034	1058.78	4.450
38	6.988		0.0000	0.000	720.1	2.006	BB	0.032	379.26	1.594
39	7.159	DCB	0.5003	86.919	290.9	0.811	BB	0.031	158.73	0.667

Total Area = 35896.6, Total Amount = 0.576, Total Height = 23793.92, Sample Units = PPB

ASX22 8665

Sample Name: **AR6210924B AAAR621AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 07:00:36pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.20R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.614		0.0000	0.000	487.5	0.268	BV	0.044	183.11	0.115
2	2.700		0.0000	0.000	693.7	0.381	VB	0.024	480.10	0.303
3	2.746		0.0000	0.000	163.0	0.090	BB	0.026	105.47	0.066
4	2.807	TCX	0.0714	13.742	253.3	0.139	BB	0.013	326.73	0.206
5	2.928		0.0000	0.000	231.3	0.127	BB	0.030	128.79	0.081
6	3.095		0.0000	0.000	209.9	0.115	BB	0.018	197.75	0.125
7	3.307		0.0000	0.000	313.6	0.172	BB	0.018	292.02	0.184
8	3.430		0.0000	0.000	435.8	0.239	BV	0.025	286.03	0.180
9	3.530		0.0000	0.000	698.4	0.384	VV	0.023	513.67	0.324
10	3.586		0.0000	0.000	748.9	0.411	VV	0.024	515.77	0.325
11	3.623		0.0000	0.000	505.3	0.278	VV	0.026	329.10	0.207
12	3.699		0.0000	0.000	554.3	0.305	VV	0.025	364.81	0.230
13	3.735		0.0000	0.000	594.1	0.326	VV	0.017	589.27	0.371
14	3.759		0.0000	0.000	674.0	0.370	VV	0.026	432.50	0.273
15	3.810		0.0000	0.000	793.1	0.436	VV	0.064	205.46	0.129
16	3.869		0.0000	0.000	855.9	0.470	VB	0.026	542.83	0.342
17	3.931		0.0000	0.000	475.7	0.261	BV	0.034	231.56	0.146
18	3.964		0.0000	0.000	305.3	0.168	VB	0.022	235.74	0.149
19	4.079		0.0000	0.000	89.3	0.049	BV	0.015	98.40	0.062
20	4.103		0.0000	0.000	2191.4	1.204	VV	0.017	2135.06	1.345
21	4.168		0.0000	0.000	162.5	0.089	VB	0.014	193.56	0.122
22	4.205		0.0000	0.000	1659.1	0.912	BV	0.015	1896.48	1.195
23	4.240		0.0000	0.000	137.9	0.076	VB	0.013	176.72	0.111
24	4.305		0.0000	0.000	478.9	0.263	BV	0.028	283.96	0.179
25	4.357		0.0000	0.000	220.7	0.121	VB	0.017	217.51	0.137
26	4.414		0.0000	0.000	188.5	0.104	BB	0.013	245.85	0.155
27	4.462		0.0000	0.000	1241.3	0.682	BV	0.015	1362.51	0.859
28	4.492		0.0000	0.000	4677.9	2.570	VB	0.015	5263.69	3.317
29	4.521		0.0000	0.000	1168.0	0.642	BB	0.015	1270.99	0.801
30	4.573		0.0000	0.000	9367.9	5.147	BV	0.018	8900.25	5.608
31	4.642		0.0000	0.000	325.3	0.179	VV	0.016	334.06	0.210
32	4.670		0.0000	0.000	1088.2	0.598	VV	0.016	1137.89	0.717
33	4.713		0.0000	0.000	11043.6	6.068	VB	0.020	9053.75	5.705
34	4.822		0.0000	0.000	11149.0	6.125	BV	0.019	9590.48	6.043
35	4.877		0.0000	0.000	1679.0	0.922	VB	0.017	1622.10	1.022
36	4.915		0.0000	0.000	2326.9	1.278	BB	0.015	2550.18	1.607
37	4.951		0.0000	0.000	5462.6	3.001	BV	0.018	4931.99	3.108
38	4.985		0.0000	0.000	12867.8	7.070	VB	0.017	12544.00	7.904
39	5.024		0.0000	0.000	4696.8	2.581	BB	0.016	5007.89	3.156
40	5.126		0.0000	0.000	2644.5	1.453	BB	0.017	2593.90	1.634
41	5.179		0.0000	0.000	14175.3	7.788	BV	0.019	12133.39	7.645
42	5.251		0.0000	0.000	6539.5	3.593	VV	0.020	5421.19	3.416
43	5.287		0.0000	0.000	1994.8	1.096	VV	0.018	1828.38	1.152
44	5.312		0.0000	0.000	1703.2	0.936	VB	0.018	1551.82	0.978
45	5.367		0.0000	0.000	24854.4	13.655	BB	0.020	20265.34	12.769
46	5.443		0.0000	0.000	1579.5	0.868	BB	0.016	1658.39	1.045
47	5.581		0.0000	0.000	9606.0	5.278	BB	0.018	9010.81	5.678
48	5.627		0.0000	0.000	17362.8	9.539	BB	0.021	13755.52	8.668
49	5.781		0.0000	0.000	674.2	0.370	BB	0.017	670.76	0.423
50	5.820		0.0000	0.000	43.0	0.024	BB	0.011	63.01	0.040
51	5.849		0.0000	0.000	437.2	0.240	BB	0.016	453.22	0.286
52	5.897		0.0000	0.000	3854.0	2.117	BB	0.020	3173.16	1.999
53	5.999		0.0000	0.000	10113.0	5.556	BB	0.023	7361.98	4.639
54	6.198		0.0000	0.000	3368.5	1.851	BB	0.020	2791.53	1.759
55	6.358	DCB	0.4483	86.258	1562.6	0.859	BB	0.024	1069.84	0.674
56	6.752		0.0000	0.000	92.3	0.051	BB	0.032	48.32	0.030
57	6.867		0.0000	-0.000	141.8	0.078	BB	0.068	34.51	0.022
58	6.967		0.0000	0.000	48.2	0.026	BB	0.021	39.10	0.025

Total Area = 182010.5, Total Amount = 0.52, Total Height = 158702.2, Sample Units = PPB

ASX22. 0666

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.20R

AR6210924B AAAR621AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.20R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 6:53:06 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.20R

AR6210924B AAAR621AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.20R



ASX22 0667

Instrument ID: CP06--S9147B Injected On: 11/6/2009 6:53:06 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.309	147	.07529	TCX	2.807	327	.07141	TCX
7.159	159	.50031	DCB	6.358	1070	.44825	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.20A
Area File: C:\CPWIN\DATA1\1C6310B.20A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 7:00:42 PM
File Reported On: 11/6/2009 at 7:00:51 PM

REPLOTS

ASX22 0068

AR6210924B AAAR621AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.20R



AR6210924B AAAR621AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.20R



ASX22 8669

Sample Name: **AR6220924B AAAR622AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 07:12:31pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.21R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.141		0.0000	0.000	41.4	0.063	BB	0.013	51.15	0.116
2	3.173		0.0000	0.000	59.5	0.090	BB	0.015	66.01	0.149
3	3.222		0.0000	0.000	175.3	0.266	BB	0.042	69.64	0.157
4	3.331	TCX	0.0294	8.336	74.6	0.113	BB	0.022	57.52	0.130
5	3.564		0.0000	0.000	118.6	0.180	BB	0.018	112.44	0.254
6	3.788		0.0000	0.000	111.5	0.169	BB	0.014	135.57	0.306
7	3.914		0.0000	0.000	179.1	0.272	BB	0.017	177.71	0.401
8	4.033		0.0000	0.000	42.4	0.064	BB	0.019	36.90	0.083
9	4.073		0.0000	0.000	43.2	0.065	BB	0.016	45.47	0.103
10	4.111		0.0000	0.000	281.4	0.427	BB	0.026	183.82	0.415
11	4.154		0.0000	0.000	120.8	0.183	BB	0.016	123.29	0.279
12	4.221		0.0000	0.000	129.0	0.196	BB	0.020	105.67	0.239
13	4.291		0.0000	0.000	120.0	0.182	BB	0.014	141.06	0.319
14	4.433		0.0000	0.000	162.7	0.247	BB	0.023	119.11	0.269
15	4.532		0.0000	0.000	78.2	0.119	BB	0.022	59.27	0.134
16	4.708		0.0000	0.000	1097.7	1.665	BV	0.022	819.62	1.852
17	4.757		0.0000	0.000	217.3	0.330	VB	0.033	109.78	0.248
18	4.875		0.0000	0.000	825.0	1.251	BB	0.020	694.94	1.570
19	5.034		0.0000	0.000	71.3	0.108	BB	0.023	51.84	0.117
20	5.106		0.0000	0.000	1093.5	1.659	BV	0.025	740.55	1.673
21	5.153		0.0000	0.000	414.4	0.629	VB	0.024	287.75	0.650
22	5.202		0.0000	0.000	2719.5	4.125	BB	0.022	2071.55	4.680
23	5.240		0.0000	0.000	147.1	0.223	BB	0.012	201.79	0.456
24	5.283		0.0000	0.000	4472.7	6.785	BB	0.024	3165.80	7.152
25	5.373		0.0000	0.000	178.2	0.270	BB	0.019	156.91	0.355
26	5.432		0.0000	0.000	450.9	0.684	BB	0.022	339.69	0.767
27	5.484		0.0000	0.000	4053.9	6.149	BB	0.024	2864.12	6.471
28	5.539		0.0000	0.000	1747.8	2.651	BB	0.018	1642.20	3.710
29	5.577		0.0000	0.000	1804.2	2.737	BB	0.029	1048.53	2.369
30	5.655		0.0000	0.000	38.4	0.058	BB	0.011	57.38	0.130
31	5.692		0.0000	0.000	3050.6	4.627	BB	0.023	2217.68	5.010
32	5.755		0.0000	0.000	4644.6	7.045	BB	0.020	3911.04	8.836
33	5.797		0.0000	0.000	1745.2	2.647	BB	0.019	1514.73	3.422
34	5.893		0.0000	0.000	312.6	0.474	BB	0.016	328.62	0.742
35	5.920		0.0000	0.000	4075.5	6.182	BB	0.022	3066.80	6.929
36	5.977		0.0000	0.000	2736.8	4.151	BB	0.035	1299.49	2.936
37	6.097		0.0000	0.000	454.5	0.689	BB	0.018	426.97	0.965
38	6.152		0.0000	0.000	12518.9	18.990	BB	0.034	6101.95	13.786
39	6.355		0.0000	0.000	4549.8	6.902	BB	0.021	3620.36	8.179
40	6.406		0.0000	0.000	3173.5	4.814	BB	0.028	1904.69	4.303
41	6.542		0.0000	0.000	282.8	0.429	BB	0.020	240.47	0.543
42	6.626		0.0000	0.000	1645.8	2.496	BB	0.029	945.35	2.136
43	6.775		0.0000	0.000	4063.2	6.163	BB	0.033	2057.21	4.648
44	6.990		0.0000	0.000	1445.2	2.192	BB	0.031	786.62	1.777
45	7.162	DCB	0.3233	91.664	156.3	0.237	BB	0.025	102.58	0.232

Total Area = 65924.6, Total Amount = 0.353, Total Height = 44261.65, Sample Units = PPB

ASX22 0670

Sample Name: **AR6220924B AAAR622AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 07:12:31pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.21R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.619		0.0000	0.000	565.3	0.165	BV	0.051	183.03	0.061
2	2.701		0.0000	0.000	775.7	0.227	VB	0.019	679.56	0.226
3	2.928		0.0000	0.000	135.8	0.040	BB	0.016	142.32	0.047
4	2.979		0.0000	0.000	106.7	0.031	BB	0.023	77.13	0.026
5	3.044		0.0000	0.000	131.8	0.039	BB	0.029	75.09	0.025
6	3.098		0.0000	0.000	297.2	0.087	BB	0.014	354.30	0.118
7	3.212		0.0000	0.000	70.7	0.021	BB	0.018	65.94	0.022
8	3.308		0.0000	0.000	774.5	0.227	BV	0.022	575.61	0.192
9	3.432		0.0000	0.000	637.4	0.186	VV	0.019	545.70	0.182
10	3.513		0.0000	0.000	147.0	0.043	VV	0.013	184.69	0.061
11	3.530		0.0000	0.000	640.4	0.187	VV	0.015	707.86	0.236
12	3.588		0.0000	0.000	1215.2	0.355	VV	0.023	861.84	0.287
13	3.624		0.0000	0.000	810.3	0.237	VV	0.027	498.02	0.166
14	3.700		0.0000	0.000	482.4	0.141	VV	0.017	486.88	0.162
15	3.736		0.0000	0.000	637.8	0.187	VV	0.015	697.40	0.232
16	3.761		0.0000	0.000	935.5	0.274	VV	0.030	515.94	0.172
17	3.870		0.0000	0.000	746.9	0.218	VB	0.021	590.59	0.197
18	3.930		0.0000	0.000	226.0	0.066	BV	0.025	152.18	0.051
19	3.965		0.0000	0.000	347.3	0.102	VV	0.017	339.68	0.113
20	4.010		0.0000	0.000	274.5	0.080	VV	0.035	129.34	0.043
21	4.080		0.0000	0.000	161.1	0.047	VV	0.014	188.01	0.063
22	4.103		0.0000	0.000	4019.5	1.176	VV	0.018	3724.91	1.240
23	4.169		0.0000	0.000	431.0	0.126	VV	0.018	406.31	0.135
24	4.206		0.0000	0.000	3455.1	1.011	VV	0.017	3487.28	1.160
25	4.240		0.0000	0.000	1259.4	0.368	VV	0.037	563.63	0.188
26	4.306		0.0000	0.000	680.7	0.199	VV	0.033	344.56	0.115
27	4.358		0.0000	0.000	331.6	0.097	VB	0.019	288.35	0.096
28	4.416		0.0000	0.000	382.2	0.112	BB	0.014	451.43	0.150
29	4.463		0.0000	0.000	2348.6	0.687	BV	0.014	2755.83	0.917
30	4.493		0.0000	0.000	8943.9	2.616	VB	0.014	10372.52	3.452
31	4.521		0.0000	0.000	2200.1	0.644	BB	0.016	2335.75	0.777
32	4.574		0.0000	0.000	17948.0	5.250	BV	0.017	17252.33	5.741
33	4.643		0.0000	0.000	318.7	0.093	VV	0.013	395.08	0.131
34	4.670		0.0000	0.000	2110.7	0.617	VV	0.016	2190.70	0.729
35	4.714		0.0000	0.000	20777.5	6.078	VB	0.020	17487.74	5.820
36	4.822		0.0000	0.000	21015.8	6.148	BV	0.019	18158.96	6.043
37	4.878		0.0000	0.000	3261.3	0.954	VB	0.017	3216.09	1.070
38	4.916		0.0000	0.000	4630.6	1.355	BB	0.015	5003.63	1.665
39	4.951		0.0000	0.000	10171.9	2.976	BV	0.019	9128.44	3.038
40	4.985		0.0000	0.000	24933.4	7.294	VB	0.017	24350.84	8.103
41	5.024		0.0000	0.000	9098.8	2.662	BB	0.016	9425.86	3.137
42	5.126		0.0000	0.000	5155.7	1.508	BB	0.017	5120.78	1.704
43	5.179		0.0000	0.000	27202.6	7.958	BV	0.020	23139.36	7.700
44	5.251		0.0000	0.000	12738.2	3.726	VV	0.020	10393.86	3.459
45	5.287		0.0000	0.000	3941.5	1.153	VV	0.018	3590.97	1.195
46	5.313		0.0000	0.000	3372.1	0.986	VB	0.019	2956.10	0.984
47	5.367		0.0000	0.000	47852.3	13.999	BB	0.020	39374.06	13.103
48	5.443		0.0000	0.000	2953.3	0.864	BB	0.016	3072.65	1.023
49	5.581		0.0000	0.000	18742.6	5.483	BB	0.018	17688.36	5.886
50	5.628		0.0000	0.000	33053.9	9.670	BB	0.021	26127.29	8.695
51	5.781		0.0000	0.000	1481.9	0.434	BB	0.016	1517.08	0.505
52	5.849		0.0000	0.000	855.3	0.250	BB	0.016	875.21	0.291
53	5.897		0.0000	0.000	7487.9	2.190	BB	0.020	6239.13	2.076
54	6.000		0.0000	0.000	20142.2	5.892	BB	0.023	14348.50	4.775
55	6.200		0.0000	0.000	7224.6	2.113	BB	0.020	5877.02	1.956
56	6.358	DCB	0.2962	100.000	971.1	0.284	BB	0.023	706.87	0.235
57	6.757		0.0000	0.000	79.8	0.023	BB	0.028	48.17	0.016
58	7.420		0.0000	0.000	143.6	0.042	BB	0.067	35.55	0.012

Total Area = 341836.7, Total Amount = 0.296, Total Height = 300502.3, Sample Units = PPB

ASX22 0671

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.21R

AR6220924B AAAR622AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.21R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 7:05:01 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.21R

AR6220924B AAAR622AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.21R



Instrument ID: CP06--S9147B Injected On: 11/6/2009 7:05:01 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min\Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.331	58	.0294	TCX	.	0	.	TCX
7.162	103	.32331	DCB	6.358	707	.29617	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.21A
Area File: C:\CPWIN\DATA1\1C6310B.21A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 7:12:36 PM
File Reported On: 11/6/2009 at 7:12:46 PM

REPLOTS

ASX22 0673

AR6220924B AAAR622AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.21R



AR6220924B AAAR622AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.21R



ASX22 6674

Sample Name: **AR6230924B AAAR623AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 07:24:33pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.22R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.139		0.0000	0.000	228.4	0.194	BB	0.042	89.98	0.114
2	3.253		0.0000	0.000	249.6	0.212	BB	0.074	56.26	0.071
3	3.305	TCX	0.0131	4.987	32.1	0.027	BB	0.021	25.56	0.032
4	3.514		0.0000	0.000	39.8	0.034	BB	0.015	44.71	0.057
5	3.563		0.0000	0.000	231.9	0.197	BB	0.017	226.52	0.287
6	3.787		0.0000	0.000	208.6	0.177	BB	0.013	266.26	0.337
7	3.913		0.0000	0.000	285.5	0.242	BB	0.016	296.86	0.376
8	4.030		0.0000	0.000	94.0	0.080	BB	0.022	72.70	0.092
9	4.110		0.0000	0.000	669.4	0.568	BB	0.029	378.72	0.480
10	4.154		0.0000	0.000	217.6	0.185	BB	0.017	211.74	0.268
11	4.220		0.0000	0.000	196.3	0.167	BB	0.017	193.30	0.245
12	4.290		0.0000	0.000	254.5	0.216	BB	0.015	279.06	0.353
13	4.321		0.0000	0.000	205.9	0.175	BB	0.026	131.74	0.167
14	4.431		0.0000	0.000	314.2	0.267	BB	0.026	205.03	0.260
15	4.531		0.0000	0.000	143.0	0.121	BB	0.021	113.35	0.144
16	4.707		0.0000	0.000	2067.9	1.755	BV	0.023	1474.47	1.867
17	4.754		0.0000	0.000	465.0	0.395	VV	0.036	218.04	0.276
18	4.826		0.0000	0.000	186.1	0.158	VB	0.023	134.55	0.170
19	4.876		0.0000	0.000	1471.2	1.249	BB	0.019	1275.67	1.616
20	5.031		0.0000	0.000	134.8	0.114	BB	0.021	105.75	0.134
21	5.106		0.0000	0.000	1953.1	1.658	BV	0.024	1363.59	1.727
22	5.152		0.0000	0.000	605.1	0.514	VB	0.021	489.74	0.620
23	5.202		0.0000	0.000	4878.7	4.141	BB	0.023	3573.36	4.526
24	5.240		0.0000	0.000	233.2	0.198	BB	0.011	365.60	0.463
25	5.282		0.0000	0.000	7952.0	6.750	BB	0.024	5596.12	7.087
26	5.373		0.0000	0.000	193.9	0.165	BB	0.016	202.36	0.256
27	5.431		0.0000	0.000	716.0	0.608	BB	0.020	585.99	0.742
28	5.483		0.0000	0.000	7250.2	6.154	BB	0.024	5132.79	6.501
29	5.538		0.0000	0.000	3070.2	2.606	BB	0.018	2844.29	3.602
30	5.577		0.0000	0.000	3172.6	2.693	BB	0.028	1873.97	2.373
31	5.655		0.0000	0.000	71.4	0.061	BB	0.013	93.08	0.118
32	5.690		0.0000	0.000	5362.2	4.552	BB	0.022	4073.86	5.160
33	5.755		0.0000	0.000	8360.0	7.096	BB	0.020	6914.81	8.758
34	5.797		0.0000	0.000	2993.0	2.541	BB	0.018	2759.41	3.495
35	5.893		0.0000	0.000	560.1	0.475	BB	0.018	520.05	0.659
36	5.919		0.0000	0.000	7276.0	6.176	BB	0.023	5264.62	6.668
37	5.976		0.0000	0.000	4936.3	4.190	BB	0.033	2477.76	3.138
38	6.056		0.0000	0.000	198.7	0.169	BB	0.014	237.75	0.301
39	6.095		0.0000	0.000	846.1	0.718	BB	0.019	737.34	0.934
40	6.151		0.0000	0.000	22902.2	19.441	BB	0.034	11229.98	14.223
41	6.354		0.0000	0.000	7910.4	6.715	BB	0.021	6244.02	7.908
42	6.406		0.0000	0.000	5714.2	4.851	BB	0.028	3376.66	4.277
43	6.542		0.0000	0.000	415.3	0.353	BB	0.019	366.26	0.464
44	6.625		0.0000	0.000	2966.9	2.519	BB	0.029	1704.51	2.159
45	6.774		0.0000	0.000	7093.4	6.021	BB	0.032	3677.15	4.657
46	6.989		0.0000	0.000	2340.9	1.987	BB	0.028	1374.05	1.740
47	7.157	DCB	0.2490	95.013	137.1	0.116	BB	0.029	79.00	0.100

Total Area = 117804.7, Total Amount = 0.262, Total Height = 78958.41, Sample Units = PPB

ASX22. 0675

Sample Name: **AR6230924B AAAR623AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 07:24:33pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.22R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.607		0.0000	0.000	197.3	0.033	BV	0.031	106.19	0.020
2	2.700		0.0000	0.000	768.4	0.128	VV	0.019	664.45	0.125
3	2.747		0.0000	0.000	130.8	0.022	VB	0.021	103.15	0.019
4	2.789 TCX		0.0157	5.907	145.8	0.024	BB	0.034	71.62	0.013
5	2.898		0.0000	0.000	53.9	0.009	BB	0.016	57.66	0.011
6	2.927		0.0000	0.000	269.0	0.045	BB	0.048	94.27	0.018
7	3.043		0.0000	0.000	166.9	0.028	BB	0.019	149.87	0.028
8	3.097		0.0000	0.000	486.0	0.081	BV	0.015	534.62	0.101
9	3.210		0.0000	0.000	216.8	0.036	VV	0.029	126.30	0.024
10	3.307		0.0000	0.000	1365.3	0.228	VV	0.020	1113.34	0.210
11	3.360		0.0000	0.000	242.3	0.040	VB	0.021	192.40	0.036
12	3.404		0.0000	0.000	103.5	0.017	BV	0.017	98.86	0.019
13	3.431		0.0000	0.000	977.9	0.163	VV	0.016	1017.35	0.192
14	3.513		0.0000	0.000	293.9	0.049	VV	0.014	360.16	0.068
15	3.529		0.0000	0.000	811.1	0.135	VV	0.018	741.33	0.140
16	3.586		0.0000	0.000	2252.7	0.376	VV	0.023	1648.13	0.310
17	3.623		0.0000	0.000	1300.7	0.217	VB	0.023	959.74	0.181
18	3.699		0.0000	0.000	988.3	0.165	BV	0.019	876.34	0.165
19	3.735		0.0000	0.000	1156.4	0.193	VV	0.015	1309.51	0.247
20	3.760		0.0000	0.000	1667.5	0.279	VV	0.030	931.08	0.175
21	3.869		0.0000	0.000	1102.5	0.184	VV	0.017	1076.58	0.203
22	3.893		0.0000	0.000	337.5	0.056	VB	0.013	417.99	0.079
23	3.918		0.0000	0.000	282.9	0.047	BB	0.016	290.42	0.055
24	3.964		0.0000	0.000	883.2	0.148	BV	0.022	666.65	0.126
25	4.009		0.0000	0.000	431.0	0.072	VV	0.025	286.57	0.054
26	4.079		0.0000	0.000	383.0	0.064	VV	0.017	386.53	0.073
27	4.103		0.0000	0.000	7375.0	1.232	VV	0.018	6677.79	1.258
28	4.168		0.0000	0.000	844.0	0.141	VV	0.018	788.75	0.149
29	4.205		0.0000	0.000	5676.4	0.948	VV	0.015	6175.99	1.164
30	4.240		0.0000	0.000	1071.6	0.179	VV	0.018	973.94	0.183
31	4.259		0.0000	0.000	1210.4	0.202	VV	0.026	775.06	0.146
32	4.305		0.0000	0.000	903.7	0.151	VV	0.035	425.54	0.080
33	4.357		0.0000	0.000	567.6	0.095	VB	0.019	498.23	0.094
34	4.415		0.0000	0.000	785.6	0.131	BB	0.015	890.41	0.168
35	4.462		0.0000	0.000	4584.0	0.766	BV	0.015	5008.97	0.944
36	4.492		0.0000	0.000	16253.0	2.715	VB	0.015	18061.05	3.403
37	4.522		0.0000	0.000	3937.5	0.658	BB	0.015	4368.48	0.823
38	4.574		0.0000	0.000	31277.8	5.225	BV	0.018	29426.89	5.544
39	4.643		0.0000	0.000	664.3	0.111	VV	0.016	713.33	0.134
40	4.669		0.0000	0.000	3893.4	0.650	VV	0.016	4051.57	0.763
41	4.713		0.0000	0.000	36843.5	6.155	VB	0.020	31386.74	5.913
42	4.822		0.0000	0.000	32421.2	5.416	BV	0.020	27691.01	5.217
43	4.877		0.0000	0.000	5586.2	0.933	VB	0.017	5534.78	1.043
44	4.915		0.0000	0.000	8150.9	1.362	BB	0.015	8934.65	1.683
45	4.951		0.0000	0.000	18440.6	3.081	BV	0.019	16427.34	3.095
46	4.984		0.0000	0.000	44300.0	7.401	VB	0.017	43875.11	8.266
47	5.024		0.0000	0.000	16415.7	2.742	BB	0.016	17066.78	3.215
48	5.126		0.0000	0.000	7796.8	1.303	BB	0.017	7471.95	1.408
49	5.178		0.0000	0.000	47170.7	7.881	BV	0.020	39962.08	7.529
50	5.250		0.0000	0.000	22424.0	3.746	VV	0.020	18348.58	3.457
51	5.286		0.0000	0.000	6977.0	1.166	VV	0.018	6400.99	1.206
52	5.312		0.0000	0.000	5979.4	0.999	VB	0.018	5436.00	1.024
53	5.367		0.0000	0.000	87058.6	14.544	BB	0.020	71363.63	13.445
54	5.443		0.0000	0.000	5406.5	0.903	BB	0.016	5571.36	1.050
55	5.558		0.0000	0.000	390.5	0.065	BB	0.014	466.18	0.088
56	5.581		0.0000	0.000	30977.0	5.175	BB	0.017	30904.80	5.822
57	5.627		0.0000	0.000	61809.1	10.326	BB	0.021	49634.55	9.351
58	5.781		0.0000	0.000	2390.4	0.399	BB	0.017	2391.28	0.451
59	5.818		0.0000	0.000	130.0	0.022	BB	0.011	193.72	0.036
60	5.848		0.0000	0.000	1433.1	0.239	BB	0.015	1548.01	0.292
61	5.897		0.0000	0.000	13507.0	2.257	BB	0.019	11730.42	2.210
62	5.999		0.0000	0.000	33858.4	5.657	BB	0.023	24505.15	4.617

ASX22 8675

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.198		0.0000	0.000	12244.4	2.046	BB	0.020	10147.67	1.912
64	6.301		0.0000	0.000	76.3	0.013	BB	0.022	57.44	0.011
65	6.358	DCB	0.2493	94.093	656.8	0.110	BB	0.018	595.12	0.112
66	6.570		0.0000	0.000	65.5	0.011	BB	0.030	36.63	0.007

Total Area = 598568.6, Total Amount = 0.265, Total Height = 530799.1, Sample Units = PPB

ASX22 0677

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.22R

AR6230924B AAAR623AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.22R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 7:17:03 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.22R

AR6230924B AAAR623AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.22R



ASX22 8678

Instrument ID: CP06--S9147B Injected On: 11/6/2009 7:17:03 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.305	26	.01307	TCX	2.789	72	.01565	TCX
7.157	79	.249	DCB	6.358	595	.24935	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.22A
Area File: C:\CPWIN\DATA1\1C6310B.22A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 7:24:38 PM
File Reported On: 11/6/2009 at 7:24:48 PM

REPLOTS

ASX22 8679

AR6230924B AAAR623AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.22R



AR6230924B AAAR623AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.22R



ASX22 0680

Sample Name: **AR6240924B AAAR624AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 07:36:34pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.23R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.141		0.0000	0.000	262.5	0.103	BB	0.055	79.21	0.046
2	3.228		0.0000	0.000	267.1	0.105	BB	0.043	103.46	0.060
3	3.305	TCX	0.0292	7.365	61.6	0.024	BB	0.018	57.17	0.033
4	3.457		0.0000	0.000	104.7	0.041	BB	0.022	79.40	0.046
5	3.515		0.0000	0.000	111.7	0.044	BB	0.018	102.60	0.059
6	3.564		0.0000	0.000	451.5	0.177	BB	0.015	492.45	0.285
7	3.663		0.0000	0.000	132.6	0.052	BB	0.020	108.73	0.063
8	3.787		0.0000	0.000	422.6	0.166	BB	0.012	572.95	0.331
9	3.806		0.0000	0.000	110.5	0.043	BB	0.008	226.05	0.131
10	3.846		0.0000	0.000	82.4	0.032	BB	0.016	85.41	0.049
11	3.914		0.0000	0.000	613.3	0.240	BB	0.016	653.23	0.378
12	4.035		0.0000	0.000	163.6	0.064	BB	0.021	127.87	0.074
13	4.111		0.0000	0.000	1336.3	0.524	BB	0.028	795.42	0.460
14	4.155		0.0000	0.000	493.4	0.193	BB	0.017	486.94	0.281
15	4.221		0.0000	0.000	408.5	0.160	BB	0.017	404.06	0.234
16	4.291		0.0000	0.000	547.8	0.215	BB	0.016	585.97	0.339
17	4.323		0.0000	0.000	428.7	0.168	BB	0.027	266.49	0.154
18	4.432		0.0000	0.000	399.4	0.157	BB	0.016	428.33	0.248
19	4.496		0.0000	0.000	69.0	0.027	BB	0.014	83.45	0.048
20	4.532		0.0000	0.000	335.5	0.132	BB	0.020	283.64	0.164
21	4.574		0.0000	0.000	159.2	0.062	BB	0.026	103.22	0.060
22	4.708		0.0000	0.000	4436.6	1.739	BV	0.025	2991.79	1.729
23	4.757		0.0000	0.000	1142.5	0.448	VV	0.036	532.48	0.308
24	4.829		0.0000	0.000	559.1	0.219	VV	0.025	368.29	0.213
25	4.876		0.0000	0.000	5878.9	2.304	VV	0.033	2951.46	1.706
26	5.032		0.0000	0.000	306.9	0.120	VB	0.024	214.60	0.124
27	5.108		0.0000	0.000	4109.3	1.611	BV	0.025	2786.56	1.611
28	5.154		0.0000	0.000	1163.8	0.456	VB	0.020	946.87	0.547
29	5.203		0.0000	0.000	15065.4	5.905	BB	0.029	8808.42	5.092
30	5.283		0.0000	0.000	17140.6	6.718	BB	0.023	12181.34	7.042
31	5.373		0.0000	0.000	536.3	0.210	BB	0.017	527.46	0.305
32	5.433		0.0000	0.000	1458.9	0.572	BB	0.020	1198.29	0.693
33	5.485		0.0000	0.000	15635.9	6.129	BB	0.023	11251.22	6.504
34	5.540		0.0000	0.000	6603.8	2.588	BB	0.018	6258.28	3.618
35	5.578		0.0000	0.000	6816.1	2.672	BB	0.028	4082.00	2.360
36	5.656		0.0000	0.000	187.5	0.073	BB	0.013	239.63	0.139
37	5.692		0.0000	0.000	12182.8	4.775	BB	0.022	9036.62	5.224
38	5.755		0.0000	0.000	18417.7	7.219	BB	0.020	15222.33	8.799
39	5.798		0.0000	0.000	6769.8	2.653	BB	0.019	6021.94	3.481
40	5.894		0.0000	0.000	1143.0	0.448	BB	0.016	1165.16	0.674
41	5.920		0.0000	0.000	15936.4	6.246	BB	0.023	11774.97	6.807
42	5.977		0.0000	0.000	4928.1	1.932	BB	0.018	4508.70	2.606
43	6.013		0.0000	0.000	737.7	0.289	BB	0.014	904.18	0.523
44	6.058		0.0000	0.000	349.6	0.137	BB	0.015	379.65	0.219
45	6.097		0.0000	0.000	1805.8	0.708	BB	0.018	1636.11	0.946
46	6.153		0.0000	0.000	47881.2	18.767	BB	0.032	24604.26	14.223
47	6.356		0.0000	0.000	17511.4	6.864	BB	0.021	13574.99	7.847
48	6.408		0.0000	0.000	12256.5	4.804	BB	0.027	7605.97	4.397
49	6.543		0.0000	0.000	786.5	0.308	BB	0.018	727.82	0.421
50	6.626		0.0000	0.000	5961.2	2.337	BB	0.028	3574.83	2.066
51	6.776		0.0000	0.000	14995.3	5.877	BB	0.032	7835.05	4.529
52	6.990		0.0000	0.000	5274.1	2.067	BB	0.031	2838.67	1.641
53	7.159	DCB	0.3676	92.635	191.8	0.075	BB	0.027	116.63	0.067

Total Area = 255132.3, Total Amount = 0.397, Total Height = 172992.6, Sample Units = PPB

ASX22 0681

Sample Name: **AR6240924B AAAR624AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 07:36:34pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.23R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.551		0.0000	0.000	72.7	0.006	BB	0.018	68.94	0.006
2	2.611		0.0000	0.000	473.0	0.037	BV	0.040	197.37	0.017
3	2.700		0.0000	0.000	1126.7	0.088	VV	0.018	1034.72	0.090
4	2.745		0.0000	0.000	106.2	0.008	VB	0.023	78.21	0.007
5	2.799 TCX		0.0293	7.645	202.3	0.016	BB	0.025	134.15	0.012
6	2.854		0.0000	0.000	177.8	0.014	BB	0.019	153.23	0.013
7	2.928		0.0000	0.000	97.2	0.008	BB	0.015	106.36	0.009
8	2.982		0.0000	0.000	173.9	0.014	BB	0.015	193.03	0.017
9	3.044		0.0000	0.000	200.7	0.016	BB	0.013	257.56	0.022
10	3.098		0.0000	0.000	1017.1	0.080	BB	0.013	1258.61	0.109
11	3.212		0.0000	0.000	285.8	0.022	BB	0.018	261.97	0.023
12	3.309		0.0000	0.000	2556.2	0.200	BV	0.020	2175.15	0.188
13	3.360		0.0000	0.000	295.2	0.023	VB	0.019	264.82	0.023
14	3.404		0.0000	0.000	42.1	0.003	BB	0.010	69.65	0.006
15	3.432		0.0000	0.000	1565.5	0.122	BB	0.014	1921.39	0.166
16	3.513		0.0000	0.000	373.9	0.029	BV	0.012	538.61	0.047
17	3.530		0.0000	0.000	596.2	0.047	VV	0.013	763.81	0.066
18	3.588		0.0000	0.000	4072.6	0.318	VV	0.022	3153.15	0.273
19	3.624		0.0000	0.000	2071.3	0.162	VV	0.020	1768.51	0.153
20	3.700		0.0000	0.000	1390.5	0.109	VV	0.014	1651.50	0.143
21	3.736		0.0000	0.000	2098.6	0.164	VV	0.014	2479.48	0.215
22	3.761		0.0000	0.000	2013.9	0.157	VB	0.020	1698.80	0.147
23	3.870		0.0000	0.000	2105.8	0.165	BB	0.018	1992.41	0.172
24	3.918		0.0000	0.000	354.5	0.028	BB	0.013	443.79	0.038
25	3.965		0.0000	0.000	1221.8	0.096	BV	0.018	1132.53	0.098
26	4.010		0.0000	0.000	350.5	0.027	VB	0.014	407.47	0.035
27	4.055		0.0000	0.000	46.4	0.004	BB	0.013	60.98	0.005
28	4.080		0.0000	0.000	531.7	0.042	BV	0.013	679.34	0.059
29	4.103		0.0000	0.000	14920.5	1.166	VV	0.017	14280.99	1.236
30	4.170		0.0000	0.000	1587.9	0.124	VV	0.017	1548.47	0.134
31	4.206		0.0000	0.000	12954.7	1.013	VV	0.016	13318.67	1.152
32	4.241		0.0000	0.000	4676.1	0.366	VV	0.037	2086.14	0.181
33	4.306		0.0000	0.000	1224.3	0.096	VV	0.036	572.44	0.050
34	4.358		0.0000	0.000	1097.4	0.086	VB	0.018	1010.62	0.087
35	4.417		0.0000	0.000	1857.6	0.145	BB	0.015	2068.99	0.179
36	4.463		0.0000	0.000	9106.4	0.712	BV	0.015	10451.95	0.904
37	4.493		0.0000	0.000	35440.5	2.770	VB	0.015	40186.78	3.477
38	4.522		0.0000	0.000	8457.3	0.661	BB	0.015	9331.83	0.807
39	4.575		0.0000	0.000	67058.8	5.242	BV	0.017	63994.61	5.537
40	4.644		0.0000	0.000	1198.9	0.094	VV	0.013	1481.41	0.128
41	4.670		0.0000	0.000	8152.3	0.637	VV	0.015	8793.64	0.761
42	4.714		0.0000	0.000	77206.4	6.035	VB	0.019	66901.50	5.789
43	4.823		0.0000	0.000	75449.4	5.898	BV	0.019	66664.64	5.768
44	4.878		0.0000	0.000	11868.7	0.928	VB	0.017	11719.92	1.014
45	4.917		0.0000	0.000	18157.1	1.419	BB	0.015	19760.87	1.710
46	4.952		0.0000	0.000	40969.1	3.202	BV	0.018	37530.20	3.247
47	4.986		0.0000	0.000	97285.2	7.604	VB	0.017	96627.05	8.361
48	5.025		0.0000	0.000	35755.8	2.795	BB	0.016	37788.59	3.270
49	5.127		0.0000	0.000	19365.4	1.514	BB	0.017	18972.96	1.642
50	5.179		0.0000	0.000	101604.6	7.942	BV	0.019	89938.39	7.782
51	5.252		0.0000	0.000	47897.7	3.744	VV	0.020	40742.18	3.525
52	5.288		0.0000	0.000	14707.9	1.150	VV	0.018	13573.13	1.174
53	5.313		0.0000	0.000	13144.0	1.027	VB	0.018	11869.52	1.027
54	5.368		0.0000	0.000	185543.4	14.503	BB	0.020	154697.70	13.386
55	5.444		0.0000	0.000	11749.5	0.918	BB	0.016	11989.46	1.037
56	5.558		0.0000	0.000	1228.2	0.096	BB	0.012	1670.83	0.145
57	5.582		0.0000	0.000	67481.2	5.275	BB	0.016	69184.67	5.987
58	5.628		0.0000	0.000	127146.6	9.939	BB	0.021	100196.30	8.670
59	5.782		0.0000	0.000	5139.5	0.402	BB	0.016	5300.92	0.459
60	5.820		0.0000	0.000	369.7	0.029	BB	0.012	494.55	0.043
61	5.850		0.0000	0.000	3251.1	0.254	BB	0.016	3444.51	0.298
62	5.897		0.0000	0.000	28943.6	2.262	BB	0.020	24466.81	2.117

ASX22 068Z

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.000		0.0000	0.000	74319.3	5.809	BB	0.022	55103.03	4.768
64	6.200		0.0000	0.000	26277.4	2.054	BB	0.020	22075.75	1.910
65	6.359	DCB	0.3542	92.355	1059.1	0.083	BB	0.021	845.43	0.073
66	7.377		0.0000	0.000	41.7	0.003	BB	0.019	36.65	0.003

Total Area = 1279314.0, Total Amount = 0.384, Total Height = 1155667.0, Sample Units = PPB

ASX22 0683

LANCASTER LABORATORIES
File: C:\CPWIN\DATA1\1C6310.23R
AR6240924B AAAR624AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.23R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 7:29:04 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.23R
AR6240924B AAAR624AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.23R



ASX22 6684

Instrument ID: CP06--S9147B Injected On: 11/6/2009 7:29:04 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.305	57	.02923	TCX	2.799	134	.02932	TCX
7.159	117	.36761	DCB	6.359	845	.35422	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.23A
Area File: C:\CPWIN\DATA1\1C6310B.23A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 7:36:40 PM
File Reported On: 11/6/2009 at 7:36:49 PM

REPLOTS

ASX22 0685

AR6240924B AAAR624AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.23R



AR6240924B AAAR624AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.23R



ASX22 0606

Sample Name: **AR6250924B AAAR625AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 07:48:33pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.24R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.024		0.0000	0.000	62.6	0.012	BB	0.021	50.44	0.015
2	3.142		0.0000	0.000	149.9	0.029	BB	0.042	59.53	0.017
3	3.239		0.0000	0.000	169.0	0.033	BB	0.057	49.45	0.014
4	3.304	TCX	0.0682	100.000	129.5	0.025	BB	0.016	133.37	0.039
5	3.457		0.0000	0.000	167.2	0.033	BB	0.020	142.51	0.042
6	3.516		0.0000	0.000	177.1	0.035	BB	0.014	204.14	0.060
7	3.565		0.0000	0.000	870.0	0.170	BB	0.015	959.37	0.280
8	3.663		0.0000	0.000	172.6	0.034	BB	0.015	186.20	0.054
9	3.788		0.0000	0.000	2018.7	0.396	BB	0.023	1435.65	0.419
10	3.846		0.0000	0.000	82.8	0.016	BB	0.012	117.21	0.034
11	3.914		0.0000	0.000	1181.0	0.231	BB	0.015	1270.38	0.371
12	4.033		0.0000	0.000	378.4	0.074	BB	0.023	279.09	0.082
13	4.112		0.0000	0.000	2573.7	0.504	BB	0.027	1561.25	0.456
14	4.155		0.0000	0.000	931.1	0.182	BB	0.017	901.92	0.263
15	4.222		0.0000	0.000	802.2	0.157	BB	0.017	766.59	0.224
16	4.292		0.0000	0.000	1005.3	0.197	BB	0.016	1077.84	0.315
17	4.323		0.0000	0.000	741.4	0.145	BB	0.025	501.74	0.147
18	4.433		0.0000	0.000	732.6	0.144	BB	0.016	743.39	0.217
19	4.494		0.0000	0.000	143.5	0.028	BB	0.014	170.11	0.050
20	4.532		0.0000	0.000	626.7	0.123	BB	0.019	535.76	0.157
21	4.578		0.0000	0.000	170.8	0.033	BB	0.019	151.64	0.044
22	4.668		0.0000	0.000	44.3	0.009	BB	0.015	48.78	0.014
23	4.708		0.0000	0.000	8284.3	1.623	BV	0.024	5745.85	1.679
24	4.756		0.0000	0.000	2182.3	0.428	VV	0.038	963.96	0.282
25	4.829		0.0000	0.000	1022.1	0.200	VV	0.024	700.78	0.205
26	4.877		0.0000	0.000	11030.2	2.162	VV	0.032	5661.77	1.654
27	5.032		0.0000	0.000	526.4	0.103	VB	0.023	389.36	0.114
28	5.108		0.0000	0.000	7820.8	1.533	BV	0.025	5295.28	1.547
29	5.154		0.0000	0.000	2255.7	0.442	VB	0.020	1892.70	0.553
30	5.203		0.0000	0.000	29023.3	5.688	BB	0.028	17006.79	4.968
31	5.284		0.0000	0.000	32714.7	6.411	BB	0.023	23656.29	6.911
32	5.374		0.0000	0.000	1119.9	0.219	BB	0.018	1047.85	0.306
33	5.432		0.0000	0.000	2745.7	0.538	BB	0.019	2356.87	0.689
34	5.485		0.0000	0.000	30691.5	6.015	BB	0.023	22064.84	6.446
35	5.540		0.0000	0.000	12792.8	2.507	BB	0.018	11751.04	3.433
36	5.578		0.0000	0.000	13601.4	2.665	BB	0.028	8178.07	2.389
37	5.657		0.0000	0.000	347.6	0.068	BB	0.013	434.35	0.127
38	5.692		0.0000	0.000	23643.6	4.633	BB	0.022	17841.78	5.212
39	5.756		0.0000	0.000	36077.7	7.070	BB	0.020	30625.96	8.947
40	5.798		0.0000	0.000	13340.3	2.614	BB	0.018	12161.25	3.553
41	5.894		0.0000	0.000	2349.7	0.460	BB	0.016	2428.14	0.709
42	5.921		0.0000	0.000	31655.5	6.204	BB	0.021	24796.03	7.244
43	5.977		0.0000	0.000	20778.1	4.072	BB	0.033	10506.47	3.069
44	6.058		0.0000	0.000	672.1	0.132	BB	0.015	747.19	0.218
45	6.098		0.0000	0.000	3667.5	0.719	BB	0.019	3268.85	0.955
46	6.153		0.0000	0.000	95694.5	18.753	BB	0.032	49516.20	14.466
47	6.356		0.0000	0.000	35197.4	6.898	BB	0.022	27163.28	7.935
48	6.409		0.0000	0.000	24282.4	4.759	BB	0.027	15133.52	4.421
49	6.543		0.0000	0.000	1510.8	0.296	BB	0.018	1380.46	0.403
50	6.627		0.0000	0.000	11957.7	2.343	BB	0.028	7141.82	2.086
51	6.776		0.0000	0.000	29818.6	5.844	BB	0.032	15552.23	4.543
52	6.990		0.0000	0.000	10149.7	1.989	BB	0.030	5549.75	1.621

Total Area = 510284.6, Total Amount = 0.068, Total Height = 342305.1, Sample Units = PPB

ASX22 8687

Sample Name: **AR6250924B AAAR625AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 07:48:33pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.24R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.552		0.0000	0.000	81.5	0.003	BB	0.014	95.22	0.004
2	2.611		0.0000	0.000	384.9	0.015	BV	0.051	126.37	0.006
3	2.701		0.0000	0.000	539.6	0.022	VB	0.020	454.85	0.020
4	2.746		0.0000	0.000	77.4	0.003	BB	0.025	51.71	0.002
5	2.856		0.0000	0.000	306.3	0.012	BB	0.019	273.71	0.012
6	2.983		0.0000	0.000	326.0	0.013	BB	0.016	349.58	0.015
7	3.046		0.0000	0.000	409.6	0.016	BB	0.013	515.70	0.023
8	3.099		0.0000	0.000	1946.9	0.078	BB	0.013	2437.16	0.108
9	3.213		0.0000	0.000	405.2	0.016	BB	0.015	452.92	0.020
10	3.309		0.0000	0.000	5009.5	0.200	BV	0.020	4254.55	0.188
11	3.361		0.0000	0.000	643.0	0.026	VV	0.020	524.89	0.023
12	3.405		0.0000	0.000	163.8	0.007	VV	0.013	202.94	0.009
13	3.432		0.0000	0.000	3318.6	0.133	VB	0.014	3842.10	0.170
14	3.530		0.0000	0.000	1716.1	0.069	BV	0.025	1145.91	0.051
15	3.588		0.0000	0.000	8041.2	0.322	VV	0.021	6392.99	0.282
16	3.625		0.0000	0.000	4214.1	0.169	VV	0.020	3504.88	0.155
17	3.701		0.0000	0.000	2784.7	0.111	VV	0.014	3240.98	0.143
18	3.736		0.0000	0.000	4125.6	0.165	VV	0.014	4846.07	0.214
19	3.761		0.0000	0.000	3855.7	0.154	VB	0.020	3244.19	0.143
20	3.871		0.0000	0.000	2823.5	0.113	BB	0.013	3630.45	0.160
21	3.895		0.0000	0.000	471.3	0.019	BB	0.009	897.26	0.040
22	3.918		0.0000	0.000	590.7	0.024	BB	0.012	796.76	0.035
23	3.950		0.0000	0.000	394.0	0.016	BV	0.011	616.69	0.027
24	3.966		0.0000	0.000	2057.2	0.082	VV	0.015	2322.59	0.103
25	4.011		0.0000	0.000	780.2	0.031	VB	0.016	811.85	0.036
26	4.056		0.0000	0.000	136.8	0.005	BV	0.014	161.89	0.007
27	4.080		0.0000	0.000	1095.9	0.044	VV	0.013	1419.57	0.063
28	4.104		0.0000	0.000	28705.7	1.149	VV	0.017	27785.26	1.227
29	4.169		0.0000	0.000	3141.5	0.126	VV	0.017	3080.38	0.136
30	4.207		0.0000	0.000	24954.5	0.999	VV	0.016	26251.89	1.159
31	4.241		0.0000	0.000	8962.7	0.359	VV	0.037	4030.72	0.178
32	4.321		0.0000	0.000	1657.2	0.066	VV	0.035	780.93	0.034
33	4.359		0.0000	0.000	2063.1	0.083	VB	0.017	1999.82	0.088
34	4.417		0.0000	0.000	3878.9	0.155	BB	0.015	4418.28	0.195
35	4.463		0.0000	0.000	17920.1	0.717	BV	0.015	20511.75	0.905
36	4.494		0.0000	0.000	69985.7	2.801	VB	0.015	79721.16	3.519
37	4.523		0.0000	0.000	17749.5	0.710	BB	0.015	19508.64	0.861
38	4.575		0.0000	0.000	127231.7	5.092	BV	0.018	120888.40	5.337
39	4.644		0.0000	0.000	2423.1	0.097	VV	0.013	3070.42	0.136
40	4.670		0.0000	0.000	16083.3	0.644	VV	0.015	17592.20	0.777
41	4.714		0.0000	0.000	148026.4	5.924	VB	0.019	132149.70	5.834
42	4.823		0.0000	0.000	144462.9	5.781	BV	0.018	130216.70	5.748
43	4.879		0.0000	0.000	22800.8	0.912	VB	0.017	22046.99	0.973
44	4.917		0.0000	0.000	35910.2	1.437	BB	0.015	39889.39	1.761
45	4.953		0.0000	0.000	85241.4	3.411	BV	0.019	74750.08	3.300
46	4.986		0.0000	0.000	183219.4	7.332	VB	0.017	181488.20	8.012
47	5.026		0.0000	0.000	71100.1	2.845	BB	0.015	77146.27	3.406
48	5.127		0.0000	0.000	39927.9	1.598	BB	0.017	38844.30	1.715
49	5.180		0.0000	0.000	195708.9	7.832	BV	0.019	171168.20	7.556
50	5.252		0.0000	0.000	93002.8	3.722	VV	0.020	78868.32	3.482
51	5.288		0.0000	0.000	28788.1	1.152	VV	0.018	27166.67	1.199
52	5.313		0.0000	0.000	26293.4	1.052	VB	0.018	24114.32	1.065
53	5.368		0.0000	0.000	356312.9	14.259	BB	0.020	296580.60	13.093
54	5.444		0.0000	0.000	23217.0	0.929	BB	0.016	24339.13	1.074
55	5.559		0.0000	0.000	2421.1	0.097	BB	0.023	1736.43	0.077
56	5.582		0.0000	0.000	132407.5	5.299	BB	0.017	132132.30	5.833
57	5.629		0.0000	0.000	257920.8	10.321	BB	0.021	208307.40	9.196
58	5.782		0.0000	0.000	10350.8	0.414	BB	0.017	10221.42	0.451
59	5.820		0.0000	0.000	677.7	0.027	BB	0.012	960.50	0.042
60	5.850		0.0000	0.000	6382.8	0.255	BB	0.016	6730.00	0.297
61	5.898		0.0000	0.000	57445.3	2.299	BB	0.019	50286.33	2.220
62	6.000		0.0000	0.000	151683.0	6.070	BB	0.022	114464.10	5.053

ASX22 8688

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.200		0.0000	0.000	52692.5	2.109	BB	0.020	44150.52	1.949
64	6.360	DCB	0.5226	100.000	1433.2	0.057	BB	0.019	1247.39	0.055

Total Area = 2498883.0, Total Amount = 0.523, Total Height = 2265259.0, Sample Units = PPB

ASX22 0689

LANCASTER LABORATORIES
File: C:\CPWIN\DATA1\1C6310.24R
AR6250924B AAAR625AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.24R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 7:41:03 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.24R
AR6250924B AAAR625AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.24R



ASX22. 0698

Instrument ID: CP06--S9147B Injected On: 11/6/2009 7:41:03 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.304	133	.06818	TCX	.	0	.	TCX
.	0	.	DCB	6.36	1247	.52264	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.24A
Area File: C:\CPWIN\DATA1\1C6310B.24A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 7:48:40 PM
File Reported On: 11/6/2009 at 7:48:48 PM

REPLOTS

ASX22 0691

AR6250924B AAAR625AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.24R



AR6250924B AAAR625AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.24R



ASX22: 8692

Sample Name: **AR6810924C AAAR681AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 08:00:33pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.25R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.141		0.0000	0.000	36.8	0.095	BB	0.013	46.75	0.198
2	3.291		0.0000	0.000	369.9	0.954	BB	0.192	32.03	0.135
3	3.327	TCX	0.0323	0.633	101.3	0.261	BB	0.027	63.23	0.267
4	3.653		0.0000	0.000	33.9	0.087	BB	0.028	20.20	0.085
5	4.072		0.0000	0.000	159.3	0.411	BB	0.027	97.21	0.411
6	4.290		0.0000	0.000	298.6	0.770	BB	0.091	54.52	0.231
7	4.780		0.0000	0.000	229.1	0.591	BB	0.043	89.09	0.377
8	4.908		0.0000	0.000	160.4	0.414	BB	0.052	51.52	0.218
9	5.013		0.0000	0.000	85.9	0.221	BB	0.049	29.07	0.123
10	5.143		0.0000	0.000	664.7	1.715	BB	0.138	80.39	0.340
11	5.446		0.0000	0.000	122.7	0.317	BB	0.038	53.89	0.228
12	5.540		0.0000	0.000	129.7	0.335	BB	0.021	101.50	0.429
13	5.688		0.0000	0.000	160.4	0.414	BB	0.034	77.54	0.328
14	5.755		0.0000	0.000	1147.9	2.961	BB	0.021	907.57	3.838
15	5.936		0.0000	0.000	1727.1	4.455	BB	0.024	1211.22	5.122
16	5.999		0.0000	0.000	611.6	1.578	BB	0.027	383.29	1.621
17	6.151		0.0000	0.000	367.0	0.947	BB	0.018	344.82	1.458
18	6.177		0.0000	0.000	158.0	0.408	BB	0.015	174.98	0.740
19	6.353		0.0000	0.000	4774.5	12.316	BB	0.022	3628.75	15.344
20	6.410		0.0000	0.000	4285.0	11.053	BB	0.025	2864.31	12.112
21	6.542		0.0000	0.000	4093.9	10.560	BB	0.024	2902.54	12.274
22	6.605		0.0000	0.000	1000.4	2.581	BB	0.022	749.19	3.168
23	6.775		0.0000	0.000	1818.3	4.690	BB	0.028	1079.01	4.563
24	6.989		0.0000	0.000	13353.4	34.445	BB	0.032	6997.07	29.588
25	7.160	DCB	5.0713	99.367	2877.8	7.423	BB	0.030	1608.97	6.804

Total Area = 38767.8, Total Amount = 5.104, Total Height = 23648.64, Sample Units = PPB

ASXZZ 8693

Sample Name: **AR6810924C AAAR681AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 08:00:33pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.25R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.616		0.0000	0.000	294.6	0.152	BV	0.048	101.66	0.061
2	2.701		0.0000	0.000	880.1	0.454	VV	0.019	783.31	0.472
3	2.747		0.0000	0.000	77.2	0.040	VB	0.027	46.95	0.028
4	2.807	TCX	0.0127	0.257	100.8	0.052	BB	0.029	57.95	0.035
5	2.900		0.0000	0.000	33.6	0.017	BB	0.013	43.79	0.026
6	2.928		0.0000	0.000	169.6	0.088	BB	0.025	115.03	0.069
7	3.430		0.0000	0.000	126.9	0.066	BB	0.021	100.26	0.060
8	3.531		0.0000	0.000	716.3	0.370	BV	0.014	855.66	0.516
9	3.579		0.0000	0.000	171.7	0.089	VB	0.022	128.18	0.077
10	3.699		0.0000	0.000	147.4	0.076	BV	0.024	101.71	0.061
11	3.737		0.0000	0.000	130.0	0.067	VV	0.014	149.99	0.090
12	3.761		0.0000	0.000	372.6	0.192	VV	0.024	258.64	0.156
13	3.803		0.0000	0.000	495.4	0.256	VV	0.063	131.57	0.079
14	3.870		0.0000	0.000	521.1	0.269	VB	0.035	251.01	0.151
15	3.928		0.0000	0.000	272.8	0.141	BV	0.024	192.22	0.116
16	3.964		0.0000	0.000	720.8	0.372	VV	0.060	201.51	0.122
17	4.104		0.0000	0.000	369.7	0.191	VV	0.028	223.41	0.135
18	4.128		0.0000	0.000	246.1	0.127	VB	0.026	160.80	0.097
19	4.242		0.0000	0.000	286.9	0.148	BV	0.030	157.63	0.095
20	4.307		0.0000	0.000	375.4	0.194	VB	0.023	267.66	0.161
21	4.492		0.0000	0.000	583.7	0.301	BB	0.046	213.65	0.129
22	4.573		0.0000	0.000	317.4	0.164	BB	0.019	284.66	0.172
23	4.644		0.0000	0.000	55.8	0.029	BV	0.015	61.58	0.037
24	4.693		0.0000	0.000	981.5	0.507	VB	0.039	420.29	0.253
25	4.825		0.0000	0.000	1004.2	0.519	BB	0.025	672.35	0.406
26	4.899		0.0000	0.000	88.5	0.046	BB	0.014	103.08	0.062
27	4.952		0.0000	0.000	99.1	0.051	BB	0.013	122.90	0.074
28	4.985		0.0000	0.000	5085.0	2.626	BB	0.017	4858.18	2.930
29	5.023		0.0000	0.000	182.7	0.094	BB	0.012	256.09	0.154
30	5.127		0.0000	0.000	184.7	0.095	BB	0.016	194.23	0.117
31	5.178		0.0000	0.000	8158.0	4.213	BV	0.019	7128.63	4.299
32	5.244		0.0000	0.000	2664.8	1.376	VV	0.022	2065.18	1.246
33	5.367		0.0000	0.000	3762.2	1.943	VB	0.021	2916.54	1.759
34	5.444		0.0000	0.000	1809.4	0.934	BB	0.018	1657.29	1.000
35	5.581		0.0000	0.000	24748.8	12.780	BB	0.017	24055.30	14.508
36	5.628		0.0000	0.000	21116.3	10.904	BB	0.018	19372.69	11.684
37	5.781		0.0000	0.000	21863.8	11.290	BB	0.018	19854.28	11.974
38	5.849		0.0000	0.000	5180.9	2.675	BB	0.017	5076.35	3.062
39	5.896		0.0000	0.000	356.6	0.184	BB	0.015	404.96	0.244
40	5.999		0.0000	0.000	9590.2	4.952	BB	0.021	7615.93	4.593
41	6.199		0.0000	0.000	65515.8	33.831	BB	0.021	52203.13	31.485
42	6.359	DCB	4.9080	99.743	13413.4	6.926	BB	0.019	11714.00	7.065
43	6.759		0.0000	0.000	59.2	0.031	BB	0.018	55.77	0.034
44	6.990		0.0000	0.000	190.5	0.098	BB	0.051	62.81	0.038
45	7.159		0.0000	0.000	134.9	0.070	BB	0.021	106.82	0.064

Total Area = 193656.1, Total Amount = 4.921, Total Height = 165805.6, Sample Units = PPB

ASX22 0694

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.25R

AR6810924C AAAR681AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.25R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 7:53:03 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.25R

AR6810924C AAAR681AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.25R



ASX22 0695

Instrument ID: CP06--S9147B Injected On: 11/6/2009 7:53:03 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.327	63	.03232	TCX	2.807	58	.01267	TCX
7.16	1609	5.07131	DCB	6.359	11714	4.90801	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.25A
Area File: C:\CPWIN\DATA1\1C6310B.25A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 8:00:40 PM
File Reported On: 11/6/2009 at 8:00:47 PM

REPLOTS

ASX22 8696

AR6810924C AAAR681AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.25R



AR6810924C AAAR681AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.25R



ASX22 8697

Sample Name: **AR6820924C AAAR682AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 08:12:28pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.26R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.144		0.0000	0.000	101.7	0.161	BB	0.033	51.82	0.134
2	3.786		0.0000	0.000	83.6	0.132	BB	0.033	42.67	0.111
3	4.075		0.0000	0.000	45.5	0.072	BB	0.016	47.72	0.124
4	4.293		0.0000	0.000	48.4	0.076	BB	0.017	48.02	0.125
5	4.324		0.0000	0.000	117.0	0.185	BB	0.038	51.69	0.134
6	4.430		0.0000	0.000	83.0	0.131	BB	0.032	42.78	0.111
7	4.701		0.0000	0.000	122.5	0.193	BB	0.036	55.97	0.145
8	4.876		0.0000	0.000	27.2	0.043	BB	0.016	28.87	0.075
9	5.143		0.0000	0.000	193.9	0.306	BB	0.056	58.02	0.150
10	5.444		0.0000	0.000	101.2	0.160	BB	0.029	58.26	0.151
11	5.539		0.0000	0.000	328.3	0.518	BB	0.038	145.03	0.376
12	5.687		0.0000	0.000	90.3	0.143	BB	0.025	61.04	0.158
13	5.755		0.0000	0.000	3074.8	4.853	BB	0.031	1648.37	4.276
14	5.935		0.0000	0.000	2861.3	4.516	BB	0.023	2062.42	5.350
15	5.998		0.0000	0.000	953.2	1.504	BB	0.026	618.60	1.605
16	6.151		0.0000	0.000	463.1	0.731	BB	0.016	472.47	1.226
17	6.177		0.0000	0.000	276.6	0.437	BB	0.015	304.78	0.791
18	6.353		0.0000	0.000	7903.6	12.474	BB	0.023	5793.00	15.028
19	6.410		0.0000	0.000	7045.1	11.119	BB	0.025	4650.85	12.065
20	6.542		0.0000	0.000	7024.6	11.087	BB	0.023	5074.27	13.163
21	6.604		0.0000	0.000	1611.1	2.543	BB	0.022	1218.10	3.160
22	6.774		0.0000	0.000	2955.1	4.664	BB	0.028	1765.22	4.579
23	6.988		0.0000	0.000	22884.5	36.117	BB	0.033	11618.26	30.139
24	7.159	DCB	8.2926	100.000	4965.8	7.837	BB	0.031	2630.99	6.825

Total Area = 63361.5, Total Amount = 8.293, Total Height = 38549.21, Sample Units = PPB

ASX22 0698

Sample Name: **AR6820924C AAAR682AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 08:12:28pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.26R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.612		0.0000	0.000	227.1	0.068	BV	0.035	107.35	0.038
2	2.701		0.0000	0.000	336.3	0.101	VB	0.024	237.24	0.085
3	2.927		0.0000	0.000	103.9	0.031	BB	0.026	66.42	0.024
4	3.308		0.0000	0.000	204.0	0.061	BB	0.028	122.62	0.044
5	3.389		0.0000	0.000	38.3	0.012	BB	0.014	45.32	0.016
6	3.507		0.0000	0.000	68.4	0.021	BV	0.016	72.90	0.026
7	3.530		0.0000	0.000	396.0	0.119	VV	0.014	469.66	0.167
8	3.580		0.0000	0.000	248.2	0.075	VB	0.022	192.41	0.069
9	3.700		0.0000	0.000	103.9	0.031	BB	0.014	124.22	0.044
10	3.735		0.0000	0.000	165.1	0.050	BV	0.015	189.71	0.068
11	3.762		0.0000	0.000	266.2	0.080	VB	0.019	235.71	0.084
12	3.870		0.0000	0.000	355.0	0.107	BB	0.025	238.01	0.085
13	3.947		0.0000	0.000	183.1	0.055	BV	0.034	90.70	0.032
14	3.964		0.0000	0.000	149.6	0.045	VB	0.015	166.02	0.059
15	4.078		0.0000	0.000	67.0	0.020	BV	0.012	90.38	0.032
16	4.103		0.0000	0.000	254.8	0.077	VV	0.016	266.34	0.095
17	4.128		0.0000	0.000	264.7	0.080	VB	0.022	199.57	0.071
18	4.206		0.0000	0.000	109.6	0.033	BV	0.020	92.96	0.033
19	4.244		0.0000	0.000	336.4	0.101	VV	0.034	165.28	0.059
20	4.306		0.0000	0.000	607.5	0.183	VB	0.036	284.41	0.101
21	4.454		0.0000	0.000	87.5	0.026	BV	0.021	70.44	0.025
22	4.480		0.0000	0.000	297.5	0.090	VB	0.032	156.16	0.056
23	4.574		0.0000	0.000	128.2	0.039	BB	0.019	109.70	0.039
24	4.645		0.0000	0.000	91.6	0.028	BB	0.016	95.59	0.034
25	4.693		0.0000	0.000	122.1	0.037	BB	0.015	137.94	0.049
26	4.824		0.0000	0.000	1392.5	0.420	BV	0.026	880.34	0.314
27	4.871		0.0000	0.000	323.3	0.097	VV	0.027	196.40	0.070
28	4.916		0.0000	0.000	278.1	0.084	VB	0.017	268.53	0.096
29	4.985		0.0000	0.000	8785.3	2.648	BB	0.018	8202.56	2.922
30	5.023		0.0000	0.000	240.4	0.072	BB	0.012	338.05	0.120
31	5.126		0.0000	0.000	358.9	0.108	BB	0.017	352.83	0.126
32	5.178		0.0000	0.000	14246.4	4.293	BV	0.020	12160.60	4.332
33	5.244		0.0000	0.000	4569.4	1.377	VV	0.021	3598.82	1.282
34	5.302		0.0000	0.000	366.7	0.111	VB	0.024	258.25	0.092
35	5.368		0.0000	0.000	5164.2	1.556	BB	0.021	4184.71	1.491
36	5.444		0.0000	0.000	3177.0	0.957	BB	0.018	2903.05	1.034
37	5.581		0.0000	0.000	42595.9	12.837	BB	0.017	41058.68	14.625
38	5.628		0.0000	0.000	38956.0	11.740	BB	0.019	34079.86	12.139
39	5.781		0.0000	0.000	39787.5	11.990	BB	0.018	35960.89	12.809
40	5.848		0.0000	0.000	9178.4	2.766	BB	0.017	8745.67	3.115
41	5.895		0.0000	0.000	448.6	0.135	BB	0.013	566.39	0.202
42	6.000		0.0000	0.000	16691.3	5.030	BB	0.022	12572.11	4.478
43	6.199		0.0000	0.000	116253.4	35.034	BB	0.022	90056.96	32.078
44	6.359	DCB	8.5059	100.000	23757.3	7.159	BB	0.020	20301.08	7.231
45	7.202		0.0000	0.000	47.2	0.014	BB	0.026	30.66	0.011

Total Area = 331829.8, Total Amount = 8.506, Total Height = 280743.5, Sample Units = PPB

ASX22 8699

LANCASTER LABORATORIES
File: C:\CPWIN\DATA1\1C6310.26R
AR6820924C AAAR682AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.26R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 8:04:58 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.26R
AR6820924C AAAR682AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.26R



ASX22 8789

Instrument ID: CP06--S9147B Injected On: 11/6/2009 8:04:58 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
7.159	2631	8.29261	DCB	6.359	20301	8.50589	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.26A
Area File: C:\CPWIN\DATA1\1C6310B.26A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 8:12:34 PM
File Reported On: 11/6/2009 at 8:12:43 PM

REPLOTS

ASX22 0701

AR6820924C AAAR682AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.26R



AR6820924C AAAR682AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.26R



ASX22 8782

Sample Name: **AR6830924C AAAR683AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 08:24:25pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.27R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.026		0.0000	0.000	21.9	0.018	BB	0.013	28.99	0.038
2	3.140		0.0000	0.000	189.2	0.151	BB	0.045	69.32	0.090
3	3.219		0.0000	0.000	295.5	0.236	BB	0.057	86.48	0.112
4	3.330	TCX	0.0437	0.264	71.0	0.057	BB	0.014	85.43	0.111
5	3.728		0.0000	0.000	44.3	0.035	BB	0.025	29.89	0.039
6	3.785		0.0000	0.000	200.6	0.160	BB	0.034	97.10	0.126
7	3.912		0.0000	0.000	98.1	0.078	BB	0.020	83.79	0.109
8	4.071		0.0000	0.000	70.3	0.056	BB	0.014	81.95	0.106
9	4.100		0.0000	0.000	121.8	0.097	BB	0.024	83.08	0.108
10	4.218		0.0000	0.000	81.8	0.065	BB	0.020	66.87	0.087
11	4.289		0.0000	0.000	57.1	0.046	BB	0.016	60.88	0.079
12	4.431		0.0000	0.000	128.8	0.103	BB	0.026	82.28	0.107
13	4.532		0.0000	0.000	69.0	0.055	BB	0.020	58.03	0.075
14	4.726		0.0000	0.000	154.9	0.124	BB	0.032	81.82	0.106
15	4.780		0.0000	0.000	236.4	0.189	BB	0.051	77.77	0.101
16	4.876		0.0000	0.000	60.5	0.048	BV	0.018	56.82	0.074
17	4.905		0.0000	0.000	500.7	0.400	VB	0.032	260.68	0.339
18	5.086		0.0000	0.000	83.7	0.067	BV	0.034	40.50	0.053
19	5.145		0.0000	0.000	157.4	0.126	VV	0.027	97.93	0.127
20	5.197		0.0000	0.000	109.8	0.088	VB	0.022	81.99	0.107
21	5.443		0.0000	0.000	250.2	0.200	BB	0.047	89.65	0.116
22	5.539		0.0000	0.000	584.9	0.467	BB	0.033	299.62	0.389
23	5.686		0.0000	0.000	132.2	0.106	BB	0.020	112.05	0.146
24	5.755		0.0000	0.000	5761.1	4.601	BB	0.031	3089.21	4.013
25	5.892		0.0000	0.000	83.0	0.066	BB	0.015	94.46	0.123
26	5.934		0.0000	0.000	5391.8	4.306	BB	0.023	3915.50	5.086
27	5.998		0.0000	0.000	1826.1	1.458	BB	0.025	1199.34	1.558
28	6.150		0.0000	0.000	892.2	0.712	BB	0.018	843.04	1.095
29	6.177		0.0000	0.000	498.6	0.398	BB	0.015	549.04	0.713
30	6.352		0.0000	0.000	15426.3	12.319	BB	0.022	11685.72	15.180
31	6.409		0.0000	0.000	13634.4	10.888	BB	0.024	9372.58	12.175
32	6.541		0.0000	0.000	13412.3	10.710	BB	0.022	10172.59	13.214
33	6.604		0.0000	0.000	3558.2	2.841	BB	0.025	2339.27	3.039
34	6.773		0.0000	0.000	6455.2	5.155	BB	0.031	3526.40	4.581
35	6.988		0.0000	0.000	44877.2	35.837	BB	0.033	22845.37	29.677
36	7.158	DCB	16.5029	99.736	9689.4	7.738	BB	0.031	5235.85	6.801

Total Area = 125225.8, Total Amount = 16.547, Total Height = 76981.26, Sample Units = PPB

ASX22 8783

Sample Name: **AR6830924C AAAR683AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 08:24:25pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.27R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.613		0.0000	0.000	414.2	0.064	BV	0.043	161.88	0.029
2	2.700		0.0000	0.000	758.0	0.118	VB	0.019	651.58	0.118
3	2.810	TCX	0.0138	0.085	111.4	0.017	BB	0.029	63.34	0.011
4	2.859		0.0000	0.000	106.8	0.017	BB	0.012	149.21	0.027
5	2.897		0.0000	0.000	56.3	0.009	BB	0.020	45.91	0.008
6	2.928		0.0000	0.000	178.0	0.028	BB	0.015	201.10	0.036
7	3.000		0.0000	0.000	50.2	0.008	BB	0.026	32.63	0.006
8	3.064		0.0000	0.000	20.0	0.003	BB	0.011	29.33	0.005
9	3.097		0.0000	0.000	337.1	0.052	BB	0.069	80.88	0.015
10	3.212		0.0000	0.000	548.4	0.085	BV	0.061	150.88	0.027
11	3.308		0.0000	0.000	828.6	0.129	VV	0.043	321.72	0.058
12	3.429		0.0000	0.000	464.8	0.072	VV	0.026	300.05	0.054
13	3.482		0.0000	0.000	389.4	0.061	VV	0.029	222.61	0.040
14	3.530		0.0000	0.000	771.9	0.120	VV	0.017	749.77	0.135
15	3.586		0.0000	0.000	698.8	0.109	VV	0.025	463.65	0.084
16	3.622		0.0000	0.000	428.0	0.067	VV	0.025	280.46	0.051
17	3.700		0.0000	0.000	334.5	0.052	VV	0.020	276.06	0.050
18	3.734		0.0000	0.000	342.1	0.053	VV	0.016	357.73	0.065
19	3.760		0.0000	0.000	478.9	0.075	VB	0.020	406.10	0.073
20	3.809		0.0000	0.000	130.0	0.020	BB	0.037	57.94	0.010
21	3.869		0.0000	0.000	268.2	0.042	BB	0.013	343.76	0.062
22	3.925		0.0000	0.000	389.2	0.061	BB	0.025	264.38	0.048
23	3.964		0.0000	0.000	146.5	0.023	BB	0.011	215.86	0.039
24	4.008		0.0000	0.000	121.6	0.019	BV	0.023	87.43	0.016
25	4.078		0.0000	0.000	169.1	0.026	VV	0.014	207.34	0.037
26	4.103		0.0000	0.000	459.4	0.071	VV	0.015	500.22	0.090
27	4.127		0.0000	0.000	701.1	0.109	VV	0.028	413.87	0.075
28	4.204		0.0000	0.000	211.1	0.033	VV	0.018	192.74	0.035
29	4.243		0.0000	0.000	1707.6	0.266	VV	0.033	850.77	0.153
30	4.305		0.0000	0.000	1127.5	0.175	VV	0.038	494.32	0.089
31	4.387		0.0000	0.000	351.5	0.055	VB	0.032	182.71	0.033
32	4.449		0.0000	0.000	232.4	0.036	BV	0.026	146.35	0.026
33	4.494		0.0000	0.000	742.6	0.116	VB	0.031	394.49	0.071
34	4.569		0.0000	0.000	379.6	0.059	BB	0.024	267.72	0.048
35	4.644		0.0000	0.000	183.3	0.029	BV	0.020	150.78	0.027
36	4.691		0.0000	0.000	459.0	0.071	VV	0.018	429.71	0.078
37	4.712		0.0000	0.000	501.6	0.078	VB	0.021	407.23	0.073
38	4.823		0.0000	0.000	2111.9	0.329	BV	0.023	1532.51	0.276
39	4.870		0.0000	0.000	480.1	0.075	VV	0.024	328.74	0.059
40	4.915		0.0000	0.000	596.9	0.093	VB	0.019	532.96	0.096
41	4.952		0.0000	0.000	45.2	0.007	BB	0.015	49.52	0.009
42	4.985		0.0000	0.000	16989.4	2.643	BB	0.017	16397.05	2.958
43	5.022		0.0000	0.000	447.8	0.070	BB	0.011	689.11	0.124
44	5.126		0.0000	0.000	668.4	0.104	BB	0.016	682.81	0.123
45	5.177		0.0000	0.000	26885.0	4.183	BV	0.019	23496.40	4.239
46	5.243		0.0000	0.000	8629.8	1.343	VV	0.021	6914.77	1.247
47	5.302		0.0000	0.000	508.3	0.079	VB	0.020	429.45	0.077
48	5.366		0.0000	0.000	10005.1	1.557	BB	0.020	8401.75	1.516
49	5.443		0.0000	0.000	6116.6	0.952	BB	0.018	5718.40	1.032
50	5.580		0.0000	0.000	83950.2	13.061	BB	0.017	80847.30	14.585
51	5.628		0.0000	0.000	71774.2	11.166	BB	0.018	66318.02	11.964
52	5.780		0.0000	0.000	77141.8	12.001	BB	0.018	69686.70	12.571
53	5.848		0.0000	0.000	16288.6	2.534	BB	0.016	17407.06	3.140
54	5.896		0.0000	0.000	1454.7	0.226	BB	0.019	1250.53	0.226
55	5.999		0.0000	0.000	32978.5	5.131	BB	0.022	25136.93	4.535
56	6.198		0.0000	0.000	223592.9	34.786	BB	0.021	178810.70	32.257
57	6.358	DCB	16.2710	99.915	45788.8	7.124	BB	0.020	38834.14	7.006
58	6.690		0.0000	0.000	101.9	0.016	BB	0.015	113.56	0.020
59	6.750		0.0000	0.000	305.8	0.048	BB	0.043	118.95	0.021
60	6.967		0.0000	0.000	136.6	0.021	BB	0.049	46.54	0.008
61	7.413		0.0000	0.000	177.0	0.028	BB	0.088	33.37	0.006

ASX22 8784

Total Area = 642774.3, Total Amount = 16.285, Total Height = 554329.9, Sample Units = PPB

ASX22. 0785

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.27R

AR6830924C AAAR683AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.27R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 8:16:55 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.27R

AR6830924C AAAR683AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.27R



ASX22 0786

Instrument ID: CP06--S9147B Injected On: 11/6/2009 8:16:55 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min; Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.33	85	.04367	TCX	2.81	63	.01384	TCX
7.158	5236	16.50288	DCB	6.358	38834	16.271	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.27A
Area File: C:\CPWIN\DATA1\1C6310B.27A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 8:24:32 PM
File Reported On: 11/6/2009 at 8:24:40 PM

REPLOTS

ASX22: 8787

AR6830924C AAAR683AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.27R



AR6830924C AAAR683AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.27R



ASX22 8788

Sample Name: **AR6840924C AAAR684AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 08:36:24pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.28R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.141		0.0000	0.000	74.0	0.025	BB	0.027	45.47	0.024
2	3.333		0.0000	0.000	48.9	0.017	BB	0.017	48.75	0.026
3	3.566		0.0000	0.000	76.9	0.026	BB	0.026	49.34	0.026
4	3.788		0.0000	0.000	128.3	0.043	BB	0.015	144.27	0.077
5	3.914		0.0000	0.000	161.0	0.055	BB	0.018	149.31	0.080
6	4.035		0.0000	0.000	41.4	0.014	BB	0.016	42.17	0.023
7	4.073		0.0000	0.000	73.2	0.025	BB	0.013	92.43	0.049
8	4.112		0.0000	0.000	314.7	0.107	BB	0.026	199.45	0.106
9	4.155		0.0000	0.000	129.1	0.044	BB	0.017	128.14	0.068
10	4.222		0.0000	0.000	142.8	0.048	BB	0.018	133.21	0.071
11	4.292		0.0000	0.000	145.5	0.049	BB	0.015	157.84	0.084
12	4.326		0.0000	0.000	267.9	0.091	BB	0.032	138.57	0.074
13	4.433		0.0000	0.000	161.2	0.055	BB	0.016	166.39	0.089
14	4.462		0.0000	0.000	45.6	0.015	BB	0.013	57.67	0.031
15	4.495		0.0000	0.000	46.5	0.016	BB	0.015	52.73	0.028
16	4.533		0.0000	0.000	170.8	0.058	BB	0.021	135.18	0.072
17	4.708		0.0000	0.000	182.3	0.062	BV	0.020	152.33	0.081
18	4.758		0.0000	0.000	437.8	0.148	VB	0.042	174.50	0.093
19	4.875		0.0000	0.000	120.8	0.041	BB	0.018	114.40	0.061
20	5.090		0.0000	0.000	63.9	0.022	BV	0.026	41.71	0.022
21	5.147		0.0000	0.000	199.6	0.068	VB	0.032	105.15	0.056
22	5.200		0.0000	0.000	64.1	0.022	BB	0.018	59.85	0.032
23	5.374		0.0000	0.000	114.6	0.039	BB	0.020	95.91	0.051
24	5.443		0.0000	0.000	136.7	0.046	BB	0.028	81.73	0.044
25	5.483		0.0000	0.000	169.2	0.057	BB	0.023	124.02	0.066
26	5.540		0.0000	0.000	773.9	0.262	BB	0.021	622.46	0.332
27	5.688		0.0000	0.000	348.5	0.118	BB	0.021	274.19	0.146
28	5.756		0.0000	0.000	13173.0	4.464	BB	0.030	7219.50	3.854
29	5.892		0.0000	0.000	162.5	0.055	BB	0.015	180.54	0.096
30	5.936		0.0000	0.000	12330.2	4.178	BB	0.023	8939.53	4.773
31	5.999		0.0000	0.000	3211.9	1.088	BB	0.022	2479.25	1.324
32	6.152		0.0000	0.000	2122.6	0.719	BB	0.018	2006.63	1.071
33	6.179		0.0000	0.000	1141.3	0.387	BB	0.014	1359.49	0.726
34	6.353		0.0000	0.000	36453.2	12.352	BB	0.022	27956.09	14.925
35	6.411		0.0000	0.000	32926.7	11.157	BB	0.024	22860.29	12.204
36	6.544		0.0000	0.000	32498.8	11.012	BB	0.022	24675.70	13.174
37	6.606		0.0000	0.000	7190.7	2.437	BB	0.021	5624.06	3.003
38	6.776		0.0000	0.000	15136.6	5.129	BB	0.030	8404.45	4.487
39	6.990		0.0000	0.000	111222.2	37.687	BB	0.031	59275.72	31.646
40	7.161	DCB	40.1654	100.000	22909.5	7.763	BB	0.030	12743.22	6.803

Total Area = 295118.3, Total Amount = 40.165, Total Height = 187311.6, Sample Units = PPB

ASX22 8789

Sample Name: **AR6840924C AAAR684AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 08:36:24pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.28R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.551		0.0000	0.000	50.2	0.003	BB	0.016	51.01	0.004
2	2.620		0.0000	0.000	254.2	0.017	BB	0.040	105.41	0.008
3	2.757		0.0000	0.000	98.5	0.006	BB	0.028	58.06	0.004
4	2.808	TCX	0.0159	0.039	149.7	0.010	BB	0.034	72.85	0.006
5	2.900		0.0000	0.000	53.6	0.004	BB	0.012	72.23	0.005
6	2.930		0.0000	0.000	102.1	0.007	BB	0.013	133.47	0.010
7	3.098		0.0000	0.000	167.9	0.011	BB	0.018	157.70	0.012
8	3.211		0.0000	0.000	80.8	0.005	BB	0.026	51.32	0.004
9	3.310		0.0000	0.000	671.7	0.044	BV	0.020	549.07	0.042
10	3.433		0.0000	0.000	471.3	0.031	VV	0.017	474.30	0.036
11	3.532		0.0000	0.000	954.1	0.062	VV	0.017	952.70	0.072
12	3.589		0.0000	0.000	1336.8	0.087	VV	0.023	954.26	0.072
13	3.625		0.0000	0.000	739.4	0.048	VV	0.024	516.06	0.039
14	3.702		0.0000	0.000	460.5	0.030	VV	0.015	518.54	0.039
15	3.737		0.0000	0.000	622.5	0.041	VV	0.014	732.41	0.056
16	3.762		0.0000	0.000	967.6	0.063	VB	0.019	832.37	0.063
17	3.871		0.0000	0.000	622.2	0.041	BB	0.013	816.05	0.062
18	3.895		0.0000	0.000	161.0	0.011	BB	0.009	307.22	0.023
19	3.920		0.0000	0.000	304.3	0.020	BV	0.017	296.36	0.022
20	3.966		0.0000	0.000	863.0	0.056	VV	0.021	698.39	0.053
21	4.011		0.0000	0.000	575.2	0.038	VV	0.031	307.32	0.023
22	4.080		0.0000	0.000	493.8	0.032	VV	0.015	535.26	0.041
23	4.105		0.0000	0.000	1142.9	0.075	VV	0.016	1219.67	0.092
24	4.130		0.0000	0.000	1642.8	0.107	VV	0.028	970.99	0.074
25	4.208		0.0000	0.000	550.0	0.036	VV	0.022	418.93	0.032
26	4.246		0.0000	0.000	984.1	0.064	VV	0.021	782.05	0.059
27	4.273		0.0000	0.000	781.0	0.051	VV	0.020	644.12	0.049
28	4.307		0.0000	0.000	925.5	0.060	VV	0.036	424.17	0.032
29	4.359		0.0000	0.000	573.5	0.037	VV	0.032	298.44	0.023
30	4.414		0.0000	0.000	142.8	0.009	VB	0.017	141.66	0.011
31	4.452		0.0000	0.000	248.0	0.016	BV	0.019	217.74	0.017
32	4.482		0.0000	0.000	979.5	0.064	VB	0.030	539.28	0.041
33	4.577		0.0000	0.000	421.7	0.028	BV	0.020	350.60	0.027
34	4.613		0.0000	0.000	48.4	0.003	VB	0.013	63.35	0.005
35	4.646		0.0000	0.000	426.0	0.028	BV	0.019	365.51	0.028
36	4.694		0.0000	0.000	496.4	0.032	VV	0.018	469.90	0.036
37	4.714		0.0000	0.000	1065.1	0.070	VB	0.022	822.62	0.062
38	4.825		0.0000	0.000	4069.9	0.266	BV	0.020	3464.74	0.263
39	4.873		0.0000	0.000	767.9	0.050	VV	0.022	589.91	0.045
40	4.917		0.0000	0.000	1150.3	0.075	VB	0.016	1215.03	0.092
41	4.953		0.0000	0.000	72.2	0.005	BB	0.012	97.44	0.007
42	4.987		0.0000	0.000	40422.4	2.643	BB	0.017	38985.94	2.957
43	5.025		0.0000	0.000	1314.3	0.086	BB	0.013	1678.87	0.127
44	5.128		0.0000	0.000	1788.9	0.117	BB	0.016	1831.53	0.139
45	5.179		0.0000	0.000	63766.4	4.169	BV	0.020	54202.29	4.111
46	5.244		0.0000	0.000	20192.6	1.320	VV	0.020	16628.95	1.261
47	5.304		0.0000	0.000	1436.4	0.094	VB	0.021	1152.38	0.087
48	5.368		0.0000	0.000	23623.9	1.544	BB	0.020	19856.81	1.506
49	5.445		0.0000	0.000	14511.9	0.949	BB	0.018	13190.48	1.000
50	5.583		0.0000	0.000	200336.3	13.097	BB	0.018	188446.40	14.291
51	5.630		0.0000	0.000	176489.3	11.538	BB	0.018	161474.30	12.246
52	5.782		0.0000	0.000	187127.3	12.233	BB	0.019	167729.30	12.720
53	5.850		0.0000	0.000	41692.8	2.726	BB	0.016	43021.10	3.263
54	5.898		0.0000	0.000	2607.8	0.170	BB	0.016	2758.83	0.209
55	6.001		0.0000	0.000	77817.3	5.087	BB	0.021	61860.93	4.691
56	6.127		0.0000	0.000	220.5	0.014	BB	0.013	284.02	0.022
57	6.200		0.0000	0.000	536937.3	35.101	BB	0.021	424807.90	32.217
58	6.360	DCB	41.1060	99.961	112441.1	7.351	BB	0.019	98108.00	7.440
59	6.691		0.0000	0.000	275.8	0.018	BB	0.016	286.32	0.022

ASX22 0718

Total Area = 1529691.0, Total Amount = 41.122, Total Height = 1318593.0, Sample Units = PPB

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.28R

AR6840924C AAAR684AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.28R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 8:28:54 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.28R

AR6840924C AAAR684AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.28R



ASX22 8711

Instrument ID: CP06--S9147B Injected On: 11/6/2009 8:28:54 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
.	0	.	TCX	2.808	73	.01592	TCX
7.161	12743	40.16536	DCB	6.36	98108	41.10597	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.28A
Area File: C:\CPWIN\DATA1\1C6310B.28A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 8:36:30 PM
File Reported On: 11/6/2009 at 8:36:39 PM

REPLOTS

ASX22 0712

AR6840924C AAAR684AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.28R



AR6840924C AAAR684AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.28R



ASX22 0713

Sample Name: **AR6850924C AAAR685AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 08:48:24pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.29R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.227		0.0000	0.000	135.6	0.023	BB	0.047	47.92	0.013
2	3.329	TCX	0.0262	0.033	64.8	0.011	BB	0.021	51.17	0.014
3	3.518		0.0000	0.000	37.6	0.006	BB	0.019	33.42	0.009
4	3.565		0.0000	0.000	91.8	0.016	BB	0.015	105.02	0.028
5	3.788		0.0000	0.000	563.3	0.097	BB	0.027	349.88	0.093
6	3.914		0.0000	0.000	264.5	0.046	BB	0.016	271.75	0.072
7	4.032		0.0000	0.000	52.7	0.009	BB	0.016	53.95	0.014
8	4.112		0.0000	0.000	859.8	0.148	BB	0.030	479.94	0.127
9	4.155		0.0000	0.000	266.3	0.046	BB	0.017	264.16	0.070
10	4.220		0.0000	0.000	288.8	0.050	BB	0.017	275.84	0.073
11	4.291		0.0000	0.000	293.9	0.051	BB	0.015	320.36	0.085
12	4.324		0.0000	0.000	505.4	0.087	BB	0.031	269.53	0.071
13	4.432		0.0000	0.000	317.5	0.055	BB	0.016	331.81	0.088
14	4.493		0.0000	0.000	98.9	0.017	BB	0.015	108.66	0.029
15	4.532		0.0000	0.000	313.1	0.054	BB	0.019	273.49	0.072
16	4.708		0.0000	0.000	348.8	0.060	BV	0.018	324.60	0.086
17	4.728		0.0000	0.000	427.3	0.074	VV	0.019	384.08	0.102
18	4.757		0.0000	0.000	754.3	0.130	VB	0.035	358.53	0.095
19	4.874		0.0000	0.000	267.5	0.046	BB	0.016	273.13	0.072
20	5.030		0.0000	0.000	73.5	0.013	BB	0.017	72.13	0.019
21	5.087		0.0000	0.000	153.3	0.026	BV	0.022	117.98	0.031
22	5.153		0.0000	0.000	383.7	0.066	VV	0.029	217.26	0.057
23	5.202		0.0000	0.000	168.4	0.029	VB	0.019	146.23	0.039
24	5.246		0.0000	0.000	82.7	0.014	BB	0.016	87.22	0.023
25	5.281		0.0000	0.000	81.9	0.014	BB	0.021	63.50	0.017
26	5.376		0.0000	0.000	164.9	0.028	BB	0.018	149.81	0.040
27	5.446		0.0000	0.000	86.5	0.015	BB	0.021	68.32	0.018
28	5.485		0.0000	0.000	284.8	0.049	BB	0.019	244.36	0.065
29	5.539		0.0000	0.000	1453.5	0.251	BB	0.019	1248.12	0.330
30	5.688		0.0000	0.000	634.3	0.110	BB	0.022	490.74	0.130
31	5.755		0.0000	0.000	24638.7	4.254	BB	0.029	13953.04	3.689
32	5.893		0.0000	0.000	332.7	0.057	BB	0.016	353.10	0.093
33	5.936		0.0000	0.000	23125.1	3.992	BB	0.023	16910.03	4.471
34	5.998		0.0000	0.000	5680.0	0.981	BB	0.021	4601.55	1.217
35	6.152		0.0000	0.000	4078.9	0.704	BB	0.017	4031.20	1.066
36	6.178		0.0000	0.000	2072.3	0.358	BB	0.015	2342.33	0.619
37	6.353		0.0000	0.000	72425.2	12.504	BB	0.021	57812.37	15.285
38	6.411		0.0000	0.000	65915.7	11.380	BB	0.024	45590.61	12.054
39	6.543		0.0000	0.000	64501.1	11.136	BB	0.021	50479.68	13.347
40	6.605		0.0000	0.000	15580.2	2.690	BB	0.023	11140.36	2.945
41	6.776		0.0000	0.000	28764.8	4.966	BB	0.029	16360.20	4.326
42	6.990		0.0000	0.000	218466.2	37.717	BB	0.030	122357.20	32.351
43	7.161	DCB	78.1931	99.967	44124.5	7.618	BB	0.030	24808.25	6.559

Total Area = 579224.9, Total Amount = 78.219, Total Height = 378222.8, Sample Units = PPB

ASX22 8714

Sample Name: **AR6850924C AAAR685AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 08:48:24pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.29R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.553		0.0000	0.000	12.3	0.000	BV	0.004	50.91	0.002
2	2.611		0.0000	0.000	230.3	0.008	VV	0.035	111.12	0.005
3	2.701		0.0000	0.000	468.3	0.016	VB	0.022	355.36	0.014
4	2.897		0.0000	0.000	45.5	0.002	BB	0.016	48.16	0.002
5	2.928		0.0000	0.000	84.2	0.003	BB	0.014	103.82	0.004
6	3.043		0.0000	0.000	35.8	0.001	BB	0.013	45.36	0.002
7	3.098		0.0000	0.000	281.6	0.010	BB	0.016	288.84	0.012
8	3.211		0.0000	0.000	111.0	0.004	BB	0.020	91.15	0.004
9	3.308		0.0000	0.000	1024.8	0.036	BV	0.018	929.32	0.038
10	3.360		0.0000	0.000	83.2	0.003	VB	0.014	98.37	0.004
11	3.432		0.0000	0.000	678.8	0.024	BB	0.014	822.72	0.033
12	3.512		0.0000	0.000	252.6	0.009	BV	0.015	290.11	0.012
13	3.531		0.0000	0.000	528.8	0.018	VV	0.014	616.92	0.025
14	3.587		0.0000	0.000	2382.9	0.083	VV	0.022	1767.75	0.072
15	3.624		0.0000	0.000	1323.4	0.046	VV	0.022	994.10	0.040
16	3.701		0.0000	0.000	909.8	0.032	VV	0.014	1072.73	0.043
17	3.736		0.0000	0.000	1285.1	0.045	VV	0.015	1443.13	0.058
18	3.761		0.0000	0.000	2028.3	0.071	VV	0.023	1499.53	0.061
19	3.870		0.0000	0.000	2159.1	0.076	VB	0.020	1782.16	0.072
20	3.919		0.0000	0.000	521.9	0.018	BV	0.016	537.66	0.022
21	3.965		0.0000	0.000	1640.2	0.057	VV	0.021	1303.51	0.053
22	4.010		0.0000	0.000	622.9	0.022	VV	0.019	555.89	0.023
23	4.057		0.0000	0.000	194.0	0.007	VV	0.021	150.82	0.006
24	4.080		0.0000	0.000	861.5	0.030	VV	0.014	999.30	0.040
25	4.103		0.0000	0.000	2171.8	0.076	VV	0.015	2405.15	0.097
26	4.129		0.0000	0.000	2686.8	0.094	VV	0.024	1837.61	0.074
27	4.207		0.0000	0.000	914.8	0.032	VV	0.022	702.40	0.028
28	4.245		0.0000	0.000	1614.6	0.056	VV	0.018	1466.58	0.059
29	4.273		0.0000	0.000	1586.3	0.055	VV	0.023	1158.03	0.047
30	4.307		0.0000	0.000	991.1	0.035	VV	0.034	489.88	0.020
31	4.358		0.0000	0.000	740.3	0.026	VV	0.027	459.57	0.019
32	4.416		0.0000	0.000	449.6	0.016	VV	0.017	442.78	0.018
33	4.452		0.0000	0.000	374.5	0.013	VV	0.017	361.79	0.015
34	4.481		0.0000	0.000	1818.8	0.064	VB	0.028	1066.32	0.043
35	4.577		0.0000	0.000	898.3	0.031	BV	0.022	696.00	0.028
36	4.611		0.0000	0.000	149.2	0.005	VV	0.018	137.35	0.006
37	4.645		0.0000	0.000	981.4	0.034	VV	0.021	794.97	0.032
38	4.693		0.0000	0.000	378.2	0.013	VV	0.016	391.12	0.016
39	4.714		0.0000	0.000	1980.4	0.069	VB	0.021	1565.39	0.063
40	4.824		0.0000	0.000	7399.5	0.259	BV	0.019	6610.67	0.268
41	4.873		0.0000	0.000	1455.4	0.051	VV	0.021	1141.40	0.046
42	4.918		0.0000	0.000	2366.6	0.083	VB	0.017	2334.15	0.095
43	4.953		0.0000	0.000	142.8	0.005	BB	0.013	189.17	0.008
44	4.986		0.0000	0.000	77171.9	2.699	BB	0.018	73453.27	2.975
45	5.024		0.0000	0.000	2431.6	0.085	BB	0.012	3290.82	0.133
46	5.127		0.0000	0.000	3483.5	0.122	BB	0.017	3515.94	0.142
47	5.178		0.0000	0.000	117422.2	4.107	BV	0.019	103528.80	4.193
48	5.244		0.0000	0.000	37789.6	1.322	VV	0.020	31160.65	1.262
49	5.303		0.0000	0.000	2427.1	0.085	VB	0.020	1982.00	0.080
50	5.368		0.0000	0.000	45528.4	1.592	BB	0.019	39085.36	1.583
51	5.444		0.0000	0.000	27716.9	0.969	BB	0.018	26062.92	1.056
52	5.582		0.0000	0.000	373848.7	13.074	BB	0.018	347559.00	14.077
53	5.629		0.0000	0.000	351623.3	12.297	BB	0.018	318826.30	12.913
54	5.782		0.0000	0.000	350009.7	12.241	BB	0.019	312836.00	12.670
55	5.849		0.0000	0.000	76467.1	2.674	BB	0.016	80651.32	3.267
56	5.897		0.0000	0.000	5217.2	0.182	BB	0.016	5354.82	0.217
57	6.000		0.0000	0.000	146303.6	5.117	BB	0.021	117034.90	4.740
58	6.126		0.0000	0.000	389.4	0.014	BB	0.012	524.21	0.021
59	6.200		0.0000	0.000	983313.0	34.389	BB	0.021	774479.50	31.368
60	6.359 DCB		79.0248	100.000	210427.6	7.359	BB	0.019	188609.20	7.639
61	6.523		0.0000	0.000	528.2	0.018	BB	0.023	384.17	0.016
62	6.690		0.0000	0.000	440.8	0.015	BB	0.016	467.70	0.019

ASX22 8715

Total Area = 2859407.0, Total Amount = 79.025, Total Height = 2469016.0, Sample Units = PPB

ASX22 0716

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.29R

AR6850924C AAAR685AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.29R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 8:40:54 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.29R

AR6850924C AAAR685AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.29R



Instrument ID: CP06--S9147B Injected On: 11/6/2009 8:40:54 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.329	51	.02616	TCX	.	0	.	TCX
7.161	24808	78.19311	DCB	6.359	188609	79.02479	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.29A
Area File: C:\CPWIN\DATA1\1C6310B.29A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 8:48:30 PM
File Reported On: 11/6/2009 at 8:48:39 PM

REPLOTS

ASX22 8718

AR6850924C AAAR685AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.29R



AR6850924C AAAR685AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.29R



ASX22 8719

Sample Name: **AR2130924B AAAR213AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 09:00:23pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.30R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.040		0.0000	0.000	104.0	0.075	BB	0.043	39.92	0.034
2	3.179		0.0000	0.000	60.3	0.044	BB	0.012	80.90	0.068
3	3.220		0.0000	0.000	1059.3	0.769	BB	0.017	1037.82	0.875
4	3.311	TCX	41.1371	49.948	77429.7	56.190	BB	0.016	80468.17	67.878
5	3.459		0.0000	0.000	3912.0	2.839	BB	0.017	3819.61	3.222
6	3.515		0.0000	0.000	3218.5	2.336	BB	0.015	3697.32	3.119
7	3.564		0.0000	0.000	8087.0	5.869	BB	0.015	8987.77	7.582
8	3.648		0.0000	0.000	269.2	0.195	BB	0.021	214.43	0.181
9	3.788		0.0000	0.000	313.2	0.227	BB	0.012	428.21	0.361
10	3.812		0.0000	0.000	369.8	0.268	BB	0.025	248.65	0.210
11	3.880		0.0000	0.000	565.6	0.410	BB	0.014	653.57	0.551
12	3.914		0.0000	0.000	514.3	0.373	BB	0.015	589.08	0.497
13	3.980		0.0000	0.000	197.2	0.143	BB	0.018	181.64	0.153
14	4.033		0.0000	0.000	62.4	0.045	BB	0.012	85.18	0.072
15	4.111		0.0000	0.000	1163.0	0.844	BB	0.027	723.37	0.610
16	4.155		0.0000	0.000	528.1	0.383	BB	0.017	506.77	0.427
17	4.222		0.0000	0.000	344.6	0.250	BB	0.017	334.27	0.282
18	4.291		0.0000	0.000	189.6	0.138	BB	0.016	194.48	0.164
19	4.323		0.0000	0.000	222.7	0.162	BB	0.038	96.41	0.081
20	4.432		0.0000	0.000	108.1	0.078	BB	0.015	121.35	0.102
21	4.534		0.0000	0.000	234.8	0.170	BB	0.034	115.86	0.098
22	4.709		0.0000	0.000	366.8	0.266	BB	0.025	240.47	0.203
23	4.758		0.0000	0.000	41.8	0.030	BB	0.013	53.17	0.045
24	4.875		0.0000	0.000	526.6	0.382	BB	0.036	240.48	0.203
25	5.090		0.0000	0.000	160.0	0.116	BB	0.030	88.84	0.075
26	5.153		0.0000	0.000	342.5	0.249	BB	0.037	156.26	0.132
27	5.281		0.0000	0.000	130.9	0.095	BB	0.026	83.43	0.070
28	5.373		0.0000	0.000	84.5	0.061	BB	0.020	70.10	0.059
29	5.440		0.0000	0.000	169.2	0.123	BB	0.034	82.49	0.070
30	5.509		0.0000	0.000	141.8	0.103	BB	0.036	66.15	0.056
31	5.649		0.0000	0.000	55.3	0.040	BB	0.026	34.81	0.029
32	5.700		0.0000	0.000	458.6	0.333	BB	0.094	80.96	0.068
33	5.933		0.0000	0.000	183.9	0.133	BB	0.048	64.01	0.054
34	6.033		0.0000	0.000	324.8	0.236	BB	0.061	88.89	0.075
35	6.150		0.0000	0.000	122.4	0.089	BB	0.028	72.08	0.061
36	6.349		0.0000	0.000	362.8	0.263	BB	0.025	242.75	0.205
37	6.407		0.0000	0.000	277.2	0.201	BB	0.023	197.26	0.166
38	6.540		0.0000	0.000	399.6	0.290	BB	0.035	192.36	0.162
39	6.767		0.0000	0.000	390.4	0.283	BB	0.045	143.45	0.121
40	6.984		0.0000	0.000	1433.2	1.040	BB	0.037	646.86	0.546
41	7.162	DCB	41.2225	50.052	32875.2	23.857	BB	0.042	13078.62	11.032

Total Area = 137800.9, Total Amount = 82.36, Total Height = 118548.2, Sample Units = PPB

ASX22 8728

Sample Name: **AR2130924B AAAR213AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 09:00:23pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.30R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.613		0.0000	0.000	285.9	0.075	BV	0.040	118.70	0.033
2	2.701		0.0000	0.000	569.3	0.149	VB	0.023	406.20	0.112
3	2.763		0.0000	0.000	1789.4	0.470	BB	0.014	2107.74	0.579
4	2.808	TCX	40.2490	49.788	172646.6	45.302	BB	0.016	184155.40	50.621
5	2.983		0.0000	0.000	9174.6	2.407	BB	0.014	10835.70	2.979
6	3.030		0.0000	0.000	636.6	0.167	BB	0.011	931.38	0.256
7	3.045		0.0000	0.000	3682.5	0.966	BB	0.011	5829.60	1.602
8	3.098		0.0000	0.000	19821.7	5.201	BB	0.014	24192.15	6.650
9	3.148		0.0000	0.000	397.7	0.104	BB	0.013	525.70	0.145
10	3.212		0.0000	0.000	203.2	0.053	BB	0.012	275.06	0.076
11	3.287		0.0000	0.000	102.1	0.027	BV	0.012	144.27	0.040
12	3.309		0.0000	0.000	3368.2	0.884	VV	0.029	1943.58	0.534
13	3.350		0.0000	0.000	1298.3	0.341	VV	0.023	959.02	0.264
14	3.406		0.0000	0.000	2298.7	0.603	VV	0.014	2748.72	0.756
15	3.432		0.0000	0.000	1925.5	0.505	VV	0.015	2122.33	0.583
16	3.468		0.0000	0.000	303.3	0.080	VV	0.018	286.03	0.079
17	3.495		0.0000	0.000	560.9	0.147	VV	0.015	629.71	0.173
18	3.514		0.0000	0.000	582.8	0.153	VV	0.012	780.49	0.215
19	3.530		0.0000	0.000	1137.7	0.299	VV	0.015	1270.74	0.349
20	3.589		0.0000	0.000	4207.6	1.104	VV	0.022	3222.27	0.886
21	3.625		0.0000	0.000	2618.3	0.687	VV	0.023	1921.01	0.528
22	3.700		0.0000	0.000	1461.8	0.384	VV	0.016	1533.35	0.421
23	3.736		0.0000	0.000	881.5	0.231	VV	0.016	933.46	0.257
24	3.761		0.0000	0.000	1433.5	0.376	VV	0.034	694.37	0.191
25	3.870		0.0000	0.000	756.8	0.199	VV	0.017	761.29	0.209
26	3.894		0.0000	0.000	302.4	0.079	VB	0.015	336.97	0.093
27	3.920		0.0000	0.000	296.2	0.078	BV	0.024	208.65	0.057
28	3.965		0.0000	0.000	893.0	0.234	VV	0.022	664.40	0.183
29	4.012		0.0000	0.000	472.5	0.124	VV	0.039	199.62	0.055
30	4.080		0.0000	0.000	274.1	0.072	VV	0.016	293.21	0.081
31	4.104		0.0000	0.000	1226.2	0.322	VV	0.016	1282.82	0.353
32	4.129		0.0000	0.000	691.3	0.181	VV	0.024	471.35	0.130
33	4.207		0.0000	0.000	889.4	0.233	VV	0.018	823.09	0.226
34	4.242		0.0000	0.000	1133.6	0.297	VV	0.039	481.19	0.132
35	4.306		0.0000	0.000	646.0	0.170	VV	0.032	337.20	0.093
36	4.359		0.0000	0.000	437.4	0.115	VV	0.024	302.43	0.083
37	4.416		0.0000	0.000	333.1	0.087	VV	0.017	320.33	0.088
38	4.452		0.0000	0.000	145.3	0.038	VV	0.016	151.55	0.042
39	4.480		0.0000	0.000	1151.5	0.302	VB	0.025	759.04	0.209
40	4.576		0.0000	0.000	461.8	0.121	BB	0.019	412.87	0.113
41	4.645		0.0000	0.000	451.2	0.118	BV	0.025	304.51	0.084
42	4.694		0.0000	0.000	324.0	0.085	VV	0.017	313.99	0.086
43	4.714		0.0000	0.000	455.9	0.120	VB	0.022	348.26	0.096
44	4.833		0.0000	0.000	1015.5	0.266	BV	0.039	437.50	0.120
45	4.872		0.0000	0.000	489.5	0.128	VV	0.027	301.14	0.083
46	4.954		0.0000	0.000	602.2	0.158	VV	0.022	447.76	0.123
47	4.985		0.0000	0.000	709.2	0.186	VB	0.034	349.57	0.096
48	5.178		0.0000	0.000	382.4	0.100	BV	0.019	344.41	0.095
49	5.244		0.0000	0.000	173.3	0.045	VB	0.029	101.15	0.028
50	5.368		0.0000	0.000	360.8	0.095	BB	0.025	237.04	0.065
51	5.445		0.0000	0.000	198.6	0.052	BB	0.034	96.97	0.027
52	5.581		0.0000	0.000	1534.2	0.403	BB	0.024	1045.77	0.287
53	5.628		0.0000	0.000	922.1	0.242	BB	0.019	812.37	0.223
54	5.723		0.0000	0.000	49.1	0.013	BB	0.018	45.73	0.013
55	5.782		0.0000	0.000	1152.0	0.302	BB	0.022	891.45	0.245
56	5.849		0.0000	0.000	385.8	0.101	BB	0.019	333.43	0.092
57	5.965		0.0000	0.000	224.6	0.059	BB	0.401	9.33	0.003
58	5.999		0.0000	0.000	527.9	0.139	BB	0.023	376.56	0.104
59	6.110		0.0000	0.000	89.2	0.023	BB	0.020	74.14	0.020
60	6.198		0.0000	0.000	5977.2	1.568	BB	0.029	3474.43	0.955
61	6.360	DCB	40.5918	50.212	123342.5	32.364	BB	0.021	96880.93	26.631
62	6.692		0.0000	0.000	45.1	0.012	BB	0.015	48.52	0.013

ASX22 8721

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.802		0.0000	0.000	46.3	0.012	BB	0.019	40.41	0.011
64	6.981		0.0000	0.000	57.6	0.015	BB	0.020	47.23	0.013
65	7.176		0.0000	0.000	90.0	0.024	BB	0.027	56.04	0.015

Total Area = 381104.8, Total Amount = 80.841, Total Height = 363791.6, Sample Units = PPB

ASX22 8722

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.30R

AR2130924B AAAR213AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.30R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 8:52:53 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.30R

AR2130924B AAAR213AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.30R



ASX22 0723

Instrument ID: CP06--S9147B Injected On: 11/6/2009 8:52:53 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Vol/Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	80468	41.13706	TCX	2.808	184155	40.24897	TCX
7.162	13079	41.2225	DCB	6.36	96881	40.59184	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.30A
Area File: C:\CPWIN\DATA1\1C6310B.30A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 9:00:30 PM
File Reported On: 11/6/2009 at 9:00:38 PM

REPLOTS

ASX22 8724

AR2130924B AAAR213AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.30R



AR2130924B AAAR213AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.30R



ASX22 0725

Sample Name: **AR32X0924B AAAR32XAA ICAL 0930999999**
Acquired from CP06–S9147A via port 1 on 11/6/09 09:12:20pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.31R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.099		0.0000	0.000	103.2	0.062	BB	0.038	45.62	0.033
2	3.220		0.0000	0.000	804.3	0.483	BB	0.021	628.59	0.455
3	3.312	TCX	40.1940	48.718	77339.6	46.436	BB	0.016	78623.55	56.885
4	3.459		0.0000	0.000	2531.1	1.520	BB	0.017	2483.02	1.796
5	3.515		0.0000	0.000	2596.8	1.559	BB	0.015	2882.93	2.086
6	3.565		0.0000	0.000	6537.2	3.925	BB	0.016	7001.09	5.065
7	3.662		0.0000	0.000	554.8	0.333	BB	0.019	480.71	0.348
8	3.788		0.0000	0.000	5443.5	3.268	BB	0.024	3768.80	2.727
9	3.846		0.0000	0.000	185.4	0.111	BB	0.011	269.92	0.195
10	3.880		0.0000	0.000	348.2	0.209	BB	0.014	406.37	0.294
11	3.914		0.0000	0.000	3183.1	1.911	BB	0.016	3391.40	2.454
12	3.981		0.0000	0.000	195.4	0.117	BB	0.022	144.95	0.105
13	4.033		0.0000	0.000	1170.9	0.703	BB	0.024	813.08	0.588
14	4.112		0.0000	0.000	7481.6	4.492	BB	0.028	4431.42	3.206
15	4.155		0.0000	0.000	2571.9	1.544	BB	0.017	2513.64	1.819
16	4.221		0.0000	0.000	2244.5	1.348	BB	0.017	2203.75	1.594
17	4.267		0.0000	0.000	76.0	0.046	BB	0.012	106.69	0.077
18	4.292		0.0000	0.000	1326.9	0.797	BB	0.015	1493.81	1.081
19	4.324		0.0000	0.000	2220.3	1.333	BB	0.028	1310.53	0.948
20	4.433		0.0000	0.000	1512.8	0.908	BB	0.016	1588.58	1.149
21	4.459		0.0000	0.000	282.6	0.170	BB	0.011	434.75	0.315
22	4.494		0.0000	0.000	410.0	0.246	BB	0.014	499.67	0.362
23	4.533		0.0000	0.000	1663.2	0.999	BB	0.020	1376.70	0.996
24	4.579		0.0000	0.000	432.6	0.260	BB	0.017	436.50	0.316
25	4.670		0.0000	0.000	68.7	0.041	BB	0.014	81.55	0.059
26	4.708		0.0000	0.000	1155.3	0.694	BV	0.017	1104.59	0.799
27	4.729		0.0000	0.000	1636.3	0.982	VV	0.020	1398.16	1.012
28	4.756		0.0000	0.000	2419.5	1.453	VV	0.029	1392.89	1.008
29	4.874		0.0000	0.000	3349.1	2.011	VV	0.048	1160.35	0.840
30	5.032		0.0000	0.000	228.4	0.137	VB	0.021	179.46	0.130
31	5.088		0.0000	0.000	314.7	0.189	BV	0.023	229.59	0.166
32	5.123		0.0000	0.000	181.5	0.109	VV	0.021	145.16	0.105
33	5.155		0.0000	0.000	523.1	0.314	VB	0.023	372.38	0.269
34	5.244		0.0000	0.000	155.4	0.093	BB	0.019	136.00	0.098
35	5.286		0.0000	0.000	61.0	0.037	BB	0.020	50.29	0.036
36	5.377		0.0000	0.000	209.0	0.125	BB	0.020	171.41	0.124
37	5.445		0.0000	0.000	60.0	0.036	BB	0.023	44.13	0.032
38	5.581		0.0000	0.000	324.9	0.195	BB	0.029	189.47	0.137
39	5.693		0.0000	0.000	86.6	0.052	BB	0.025	57.66	0.042
40	6.016		0.0000	0.000	604.7	0.363	BB	0.073	137.86	0.100
41	6.353		0.0000	0.000	191.5	0.115	BB	0.026	120.55	0.087
42	6.409		0.0000	0.000	112.6	0.068	BB	0.025	76.00	0.055
43	6.543		0.0000	0.000	95.3	0.057	BB	0.023	70.09	0.051
44	6.987		0.0000	0.000	797.0	0.479	BB	0.039	337.33	0.244
45	7.161	DCB	42.3095	51.282	32760.8	19.670	BB	0.041	13423.49	9.712

Total Area = 166551.3, Total Amount = 82.504, Total Height = 138214.5, Sample Units = PPB

ASX22 8726

Sample Name: **AR32X0924B AAAR32XAA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 09:12:20pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.31R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.608		0.0000	0.000	605.7	0.124	BV	0.038	266.80	0.059
2	2.701		0.0000	0.000	1093.5	0.223	VB	0.019	982.28	0.219
3	2.764		0.0000	0.000	875.0	0.179	BB	0.013	1091.14	0.243
4	2.808	TCX	36.8389	46.988	172718.9	35.281	BB	0.017	168553.10	37.510
5	2.983		0.0000	0.000	5903.2	1.206	BB	0.014	6888.41	1.533
6	3.030		0.0000	0.000	848.9	0.173	BB	0.011	1327.33	0.295
7	3.046		0.0000	0.000	2344.7	0.479	BB	0.010	3841.31	0.855
8	3.099		0.0000	0.000	15859.1	3.240	BB	0.014	18933.21	4.213
9	3.149		0.0000	0.000	353.9	0.072	BB	0.012	503.10	0.112
10	3.213		0.0000	0.000	992.0	0.203	BB	0.013	1237.48	0.275
11	3.309		0.0000	0.000	14200.5	2.901	BV	0.021	11476.20	2.554
12	3.361		0.0000	0.000	1596.8	0.326	VV	0.018	1497.33	0.333
13	3.382		0.0000	0.000	485.4	0.099	VV	0.013	628.27	0.140
14	3.406		0.0000	0.000	1423.6	0.291	VV	0.013	1842.53	0.410
15	3.432		0.0000	0.000	8961.2	1.831	VV	0.014	10560.06	2.350
16	3.468		0.0000	0.000	348.2	0.071	VV	0.014	407.63	0.091
17	3.495		0.0000	0.000	341.6	0.070	VV	0.012	479.37	0.107
18	3.514		0.0000	0.000	4127.3	0.843	VB	0.023	3048.71	0.678
19	3.588		0.0000	0.000	23916.8	4.885	BV	0.021	18596.81	4.139
20	3.625		0.0000	0.000	12174.9	2.487	VV	0.020	10080.90	2.243
21	3.701		0.0000	0.000	7878.3	1.609	VV	0.014	9212.28	2.050
22	3.736		0.0000	0.000	6554.4	1.339	VV	0.015	7253.14	1.614
23	3.761		0.0000	0.000	10001.8	2.043	VB	0.020	8233.75	1.832
24	3.871		0.0000	0.000	9123.8	1.864	BB	0.018	8225.63	1.831
25	3.917		0.0000	0.000	1498.3	0.306	BB	0.011	2229.91	0.496
26	3.949		0.0000	0.000	1270.3	0.259	BV	0.011	1856.29	0.413
27	3.966		0.0000	0.000	5385.5	1.100	VV	0.014	6206.49	1.381
28	4.010		0.0000	0.000	2064.7	0.422	VB	0.015	2249.23	0.501
29	4.057		0.0000	0.000	281.0	0.057	BV	0.013	365.40	0.081
30	4.080		0.0000	0.000	3147.3	0.643	VV	0.014	3734.62	0.831
31	4.104		0.0000	0.000	7155.8	1.462	VV	0.015	8046.37	1.791
32	4.129		0.0000	0.000	8400.6	1.716	VV	0.021	6526.01	1.452
33	4.207		0.0000	0.000	2267.2	0.463	VV	0.021	1824.68	0.406
34	4.244		0.0000	0.000	4782.2	0.977	VV	0.019	4130.99	0.919
35	4.273		0.0000	0.000	5511.5	1.126	VV	0.027	3427.02	0.763
36	4.358		0.0000	0.000	1292.6	0.264	VV	0.022	973.13	0.217
37	4.417		0.0000	0.000	1320.0	0.270	VV	0.017	1297.77	0.289
38	4.452		0.0000	0.000	926.2	0.189	VV	0.018	870.68	0.194
39	4.480		0.0000	0.000	2422.9	0.495	VB	0.020	2029.29	0.452
40	4.578		0.0000	0.000	1164.3	0.238	BV	0.021	924.58	0.206
41	4.645		0.0000	0.000	1302.8	0.266	VV	0.021	1021.32	0.227
42	4.693		0.0000	0.000	380.7	0.078	VV	0.017	374.43	0.083
43	4.715		0.0000	0.000	577.7	0.118	VB	0.027	351.84	0.078
44	4.833		0.0000	0.000	839.1	0.171	BV	0.032	437.32	0.097
45	4.872		0.0000	0.000	1663.0	0.340	VV	0.027	1038.15	0.231
46	4.954		0.0000	0.000	508.3	0.104	VV	0.022	378.56	0.084
47	4.986		0.0000	0.000	210.3	0.043	VB	0.020	178.52	0.040
48	5.106		0.0000	0.000	39.8	0.008	BB	0.022	30.41	0.007
49	5.179		0.0000	0.000	219.6	0.045	BV	0.018	208.85	0.046
50	5.249		0.0000	0.000	120.1	0.025	VB	0.023	86.06	0.019
51	5.368		0.0000	0.000	800.4	0.164	BB	0.040	330.71	0.074
52	5.582		0.0000	0.000	902.4	0.184	BB	0.034	445.96	0.099
53	5.628		0.0000	0.000	422.9	0.086	BB	0.018	385.29	0.086
54	5.781		0.0000	0.000	425.7	0.087	BB	0.021	338.32	0.075
55	5.848		0.0000	0.000	315.8	0.065	BB	0.017	300.75	0.067
56	5.957		0.0000	0.000	556.0	0.114	BB	0.360	25.74	0.006
57	6.001		0.0000	0.000	253.9	0.052	BB	0.023	184.24	0.041
58	6.199		0.0000	0.000	2882.2	0.589	BB	0.024	1961.28	0.436
59	6.360	DCB	41.5611	53.012	124978.4	25.529	BB	0.021	99194.39	22.075
60	6.746		0.0000	0.000	76.1	0.016	BB	0.025	51.26	0.011
61	6.985		0.0000	0.000	223.6	0.046	BB	0.057	65.22	0.015
62	7.170		0.0000	0.000	187.4	0.038	BB	0.038	82.41	0.018

ASX22 8727

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	7.326		0.0000	0.000	41.0	0.008	BB	0.030	23.13	0.005

Total Area = 489546.8, Total Amount = 78.4, Total Height = 449353.4, Sample Units = PPB

ASX22 8728

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.31R

AR32X0924B AAAR32XAA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.31R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 9:04:50 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.31R

AR32X0924B AAAR32XAA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.31R



ASX22 8729

Instrument ID: CP06--S9147B Injected On: 11/6/2009 9:04:50 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.312	78624	40.19405	TCX	2.808	168553	36.83893	TCX
7.161	13423	42.30949	DCB	6.36	99194	41.56115	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.31A
Area File: C:\CPWIN\DATA1\1C6310B.31A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 9:12:26 PM
File Reported On: 11/6/2009 at 9:12:35 PM

REPLOTS

ASX22: 0738

AR32X0924B AAAR32XAA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.31R



AR32X0924B AAAR32XAA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.31R



ASX22 8731

Sample Name: **AR4230924B AAAR423AA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 09:24:15pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.32R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.140		0.0000	0.000	87.7	0.045	BB	0.028	52.64	0.033
2	3.311	TCX	40.2603	50.897	76533.9	39.312	BB	0.016	78753.13	49.430
3	3.457		0.0000	0.000	1009.4	0.518	BB	0.017	977.49	0.614
4	3.515		0.0000	0.000	1920.8	0.987	BB	0.015	2147.43	1.348
5	3.565		0.0000	0.000	4689.1	2.409	BB	0.015	5115.20	3.211
6	3.661		0.0000	0.000	884.3	0.454	BB	0.017	850.71	0.534
7	3.788		0.0000	0.000	9495.9	4.878	BB	0.024	6686.30	4.197
8	3.846		0.0000	0.000	482.7	0.248	BB	0.014	555.06	0.348
9	3.914		0.0000	0.000	5684.6	2.920	BB	0.016	5908.77	3.709
10	3.980		0.0000	0.000	242.2	0.124	BB	0.028	143.79	0.090
11	4.033		0.0000	0.000	1985.6	1.020	BB	0.024	1367.94	0.859
12	4.112		0.0000	0.000	13113.1	6.736	BB	0.028	7828.05	4.913
13	4.155		0.0000	0.000	4359.8	2.239	BB	0.017	4313.31	2.707
14	4.221		0.0000	0.000	3966.1	2.037	BB	0.017	3967.34	2.490
15	4.266		0.0000	0.000	159.3	0.082	BB	0.012	215.00	0.135
16	4.292		0.0000	0.000	2600.8	1.336	BB	0.015	2930.93	1.840
17	4.323		0.0000	0.000	4218.4	2.167	BB	0.028	2518.93	1.581
18	4.433		0.0000	0.000	2948.8	1.515	BB	0.016	3045.02	1.911
19	4.459		0.0000	0.000	535.3	0.275	BB	0.011	825.77	0.518
20	4.493		0.0000	0.000	754.2	0.387	BB	0.014	872.53	0.548
21	4.532		0.0000	0.000	3123.7	1.604	BB	0.020	2582.97	1.621
22	4.578		0.0000	0.000	808.9	0.415	BB	0.016	836.52	0.525
23	4.669		0.0000	0.000	179.2	0.092	BB	0.016	191.59	0.120
24	4.708		0.0000	0.000	2317.2	1.190	BV	0.018	2156.20	1.353
25	4.729		0.0000	0.000	3133.7	1.610	VV	0.020	2637.07	1.655
26	4.756		0.0000	0.000	4564.0	2.344	VV	0.029	2667.34	1.674
27	4.825		0.0000	0.000	757.9	0.389	VV	0.027	475.64	0.299
28	4.874		0.0000	0.000	6552.8	3.366	VV	0.049	2213.33	1.389
29	5.031		0.0000	0.000	718.9	0.369	VV	0.026	465.71	0.292
30	5.088		0.0000	0.000	842.0	0.433	VV	0.023	604.75	0.380
31	5.122		0.0000	0.000	559.3	0.287	VV	0.022	415.29	0.261
32	5.154		0.0000	0.000	1428.5	0.734	VB	0.027	896.10	0.562
33	5.245		0.0000	0.000	297.4	0.153	BB	0.019	256.07	0.161
34	5.376		0.0000	0.000	515.1	0.265	BB	0.022	396.52	0.249
35	5.579		0.0000	0.000	743.4	0.382	BB	0.031	397.85	0.250
36	5.693		0.0000	0.000	308.6	0.159	BB	0.038	135.42	0.085
37	5.902		0.0000	0.000	138.6	0.071	BB	0.037	62.53	0.039
38	6.020		0.0000	0.000	539.1	0.277	BB	0.058	154.89	0.097
39	6.350		0.0000	0.000	347.8	0.179	BB	0.056	103.14	0.065
40	6.787		0.0000	0.000	90.3	0.046	BB	0.037	40.49	0.025
41	6.988		0.0000	0.000	535.0	0.275	BB	0.038	236.12	0.148
42	7.161	DCB	38.8405	49.103	30512.1	15.673	BB	0.041	12322.89	7.734

Total Area = 194685.3, Total Amount = 79.101, Total Height = 159323.8, Sample Units = PPB

ASX22: 8732

Sample Name: **AR4230924B AAAR423AA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 09:24:15pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.32R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.609		0.0000	0.000	556.7	0.094	BV	0.039	240.29	0.044
2	2.701		0.0000	0.000	308.9	0.052	VB	0.021	239.55	0.043
3	2.808	TCX	38.1105	50.539	168387.2	28.367	BB	0.016	174371.20	31.590
4	2.928		0.0000	0.000	219.4	0.037	BB	0.012	296.00	0.054
5	2.983		0.0000	0.000	2176.6	0.367	BB	0.014	2573.17	0.466
6	3.045		0.0000	0.000	4474.0	0.754	BB	0.024	3071.35	0.556
7	3.098		0.0000	0.000	11147.0	1.878	BB	0.014	13491.82	2.444
8	3.148		0.0000	0.000	365.5	0.062	BB	0.012	496.06	0.090
9	3.212		0.0000	0.000	1798.6	0.303	BB	0.013	2238.72	0.406
10	3.309		0.0000	0.000	24773.2	4.173	BV	0.021	20074.22	3.637
11	3.361		0.0000	0.000	2162.9	0.364	VV	0.014	2551.91	0.462
12	3.382		0.0000	0.000	877.9	0.148	VV	0.013	1136.55	0.206
13	3.405		0.0000	0.000	1026.2	0.173	VV	0.014	1254.23	0.227
14	3.432		0.0000	0.000	15500.7	2.611	VV	0.014	17964.35	3.255
15	3.467		0.0000	0.000	621.2	0.105	VV	0.015	681.57	0.123
16	3.495		0.0000	0.000	306.4	0.052	VV	0.012	423.81	0.077
17	3.513		0.0000	0.000	6790.6	1.144	VB	0.022	5093.81	0.923
18	3.588		0.0000	0.000	40803.7	6.874	BV	0.021	31829.12	5.766
19	3.624		0.0000	0.000	21998.7	3.706	VV	0.021	17272.79	3.129
20	3.701		0.0000	0.000	13964.6	2.352	VV	0.014	16327.67	2.958
21	3.736		0.0000	0.000	12672.7	2.135	VV	0.015	13789.89	2.498
22	3.761		0.0000	0.000	19171.5	3.230	VB	0.020	15676.33	2.840
23	3.871		0.0000	0.000	11401.1	1.921	BB	0.013	14792.73	2.680
24	3.895		0.0000	0.000	2522.5	0.425	BB	0.009	4507.00	0.817
25	3.917		0.0000	0.000	2817.6	0.475	BB	0.011	4191.42	0.759
26	3.949		0.0000	0.000	2534.7	0.427	BV	0.011	3692.01	0.669
27	3.966		0.0000	0.000	10265.4	1.729	VV	0.014	11810.41	2.140
28	4.011		0.0000	0.000	3974.3	0.670	VB	0.016	4172.91	0.756
29	4.057		0.0000	0.000	722.3	0.122	BV	0.015	796.33	0.144
30	4.080		0.0000	0.000	6060.3	1.021	VV	0.014	7231.08	1.310
31	4.104		0.0000	0.000	13807.4	2.326	VV	0.015	15567.34	2.820
32	4.129		0.0000	0.000	16046.0	2.703	VV	0.022	12184.69	2.207
33	4.207		0.0000	0.000	4726.6	0.796	VV	0.020	3906.04	0.708
34	4.244		0.0000	0.000	8621.1	1.452	VV	0.018	8052.86	1.459
35	4.273		0.0000	0.000	10988.1	1.851	VV	0.028	6550.63	1.187
36	4.359		0.0000	0.000	2700.8	0.455	VV	0.021	2101.79	0.381
37	4.417		0.0000	0.000	3081.8	0.519	VV	0.017	3110.16	0.563
38	4.452		0.0000	0.000	2085.1	0.351	VV	0.017	1996.36	0.362
39	4.480		0.0000	0.000	5131.0	0.864	VB	0.019	4488.44	0.813
40	4.578		0.0000	0.000	2165.5	0.365	BV	0.021	1742.67	0.316
41	4.611		0.0000	0.000	436.6	0.074	VV	0.017	434.34	0.079
42	4.645		0.0000	0.000	3137.5	0.529	VV	0.021	2489.48	0.451
43	4.692		0.0000	0.000	596.6	0.101	VV	0.019	529.16	0.096
44	4.714		0.0000	0.000	627.2	0.106	VV	0.020	531.66	0.096
45	4.740		0.0000	0.000	507.3	0.085	VB	0.023	368.85	0.067
46	4.821		0.0000	0.000	944.7	0.159	BV	0.032	489.07	0.089
47	4.872		0.0000	0.000	3586.3	0.604	VV	0.026	2312.89	0.419
48	4.955		0.0000	0.000	1979.9	0.334	VB	0.036	916.55	0.166
49	5.177		0.0000	0.000	59.3	0.010	BV	0.014	69.58	0.013
50	5.204		0.0000	0.000	407.9	0.069	VB	0.023	290.28	0.053
51	5.349		0.0000	0.000	77.8	0.013	BB	0.017	76.73	0.014
52	5.580		0.0000	0.000	1025.7	0.173	BB	0.050	344.18	0.062
53	5.626		0.0000	0.000	147.6	0.025	BB	0.016	150.36	0.027
54	5.781		0.0000	0.000	258.5	0.044	BB	0.021	200.71	0.036
55	5.848		0.0000	0.000	168.8	0.028	BB	0.026	107.34	0.019
56	5.998		0.0000	0.000	123.9	0.021	BB	0.027	76.76	0.014
57	6.200		0.0000	0.000	2039.4	0.344	BB	0.024	1446.00	0.262
58	6.360	DCB	37.2972	49.461	117360.7	19.771	BB	0.022	89017.55	16.127
59	6.803		0.0000	0.000	29.4	0.005	BB	0.014	34.41	0.006
60	7.176		0.0000	0.000	250.7	0.042	BB	0.054	77.64	0.014
61	7.426		0.0000	0.000	87.4	0.015	BB	0.064	22.81	0.004

ASX22 0733

Total Area = 593608.6, Total Amount = 75.408, Total Height = 551975.4, Sample Units = PPB

ASX22. 8734

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.32R

AR4230924B AAAR423AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.32R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 9:16:45 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.32R

AR4230924B AAAR423AA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.32R



ASX22 8735

Instrument ID: CP06--S9147B Injected On: 11/6/2009 9:16:45 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	78753	40.26029	TCX	2.808	174371	38.11053	TCX
7.161	12323	38.84053	DCB	6.36	89018	37.29718	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.32A
Area File: C:\CPWIN\DATA1\1C6310B.32A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 9:24:22 PM
File Reported On: 11/6/2009 at 9:24:30 PM

REPLOTS

ASX22 8736

AR4230924B AAAR423AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.32R



AR4230924B AAAR423AA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.32R



ASX22 8737

Sample Name: **AR16X0924A AAAR16XAA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 09:36:22pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.33R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.103		0.0000	0.000	116.2	0.106	BB	0.041	47.28	0.052
2	3.312	TCX	9.7250	49.140	18715.2	17.068	BB	0.016	19023.04	20.743
3	3.458		0.0000	0.000	1276.5	1.164	BB	0.017	1244.41	1.357
4	3.516		0.0000	0.000	1535.4	1.400	BB	0.015	1713.70	1.869
5	3.565		0.0000	0.000	5937.1	5.415	BB	0.016	6323.82	6.896
6	3.663		0.0000	0.000	1075.8	0.981	BB	0.016	1145.76	1.249
7	3.789		0.0000	0.000	12212.0	11.137	BB	0.024	8434.42	9.197
8	3.847		0.0000	0.000	1263.1	1.152	BB	0.024	865.94	0.944
9	3.915		0.0000	0.000	7271.7	6.632	BB	0.017	7302.19	7.962
10	3.995		0.0000	0.000	138.7	0.127	BB	0.014	162.07	0.177
11	4.034		0.0000	0.000	2584.5	2.357	BB	0.025	1755.98	1.915
12	4.113		0.0000	0.000	16182.2	14.758	BB	0.028	9803.80	10.690
13	4.156		0.0000	0.000	5319.1	4.851	BB	0.017	5240.11	5.714
14	4.222		0.0000	0.000	4877.8	4.449	BB	0.017	4811.98	5.247
15	4.267		0.0000	0.000	218.2	0.199	BB	0.012	303.41	0.331
16	4.293		0.0000	0.000	3121.2	2.847	BB	0.015	3446.07	3.758
17	4.324		0.0000	0.000	4928.5	4.495	BB	0.028	2960.01	3.228
18	4.433		0.0000	0.000	3458.1	3.154	BB	0.016	3622.15	3.950
19	4.460		0.0000	0.000	686.8	0.626	BB	0.012	960.10	1.047
20	4.495		0.0000	0.000	939.4	0.857	BB	0.014	1085.15	1.183
21	4.533		0.0000	0.000	3744.4	3.415	BB	0.020	3058.29	3.335
22	4.579		0.0000	0.000	965.2	0.880	BB	0.016	994.10	1.084
23	4.670		0.0000	0.000	145.7	0.133	BB	0.015	156.81	0.171
24	4.709		0.0000	0.000	1532.9	1.398	BV	0.019	1353.98	1.476
25	4.729		0.0000	0.000	1185.6	1.081	VV	0.018	1129.05	1.231
26	4.758		0.0000	0.000	1821.6	1.661	VV	0.035	872.09	0.951
27	4.873		0.0000	0.000	688.7	0.628	VB	0.037	307.05	0.335
28	6.023		0.0000	0.000	632.8	0.577	BB	0.073	143.99	0.157
29	6.264		0.0000	0.000	89.5	0.082	BB	0.046	32.31	0.035
30	6.354		0.0000	0.000	86.4	0.079	BB	0.027	53.54	0.058
31	6.770		0.0000	0.000	80.4	0.073	BB	0.035	38.15	0.042
32	6.990		0.0000	0.000	351.8	0.321	BB	0.047	124.81	0.136
33	7.161	DCB	10.0655	50.860	6467.4	5.898	BB	0.034	3193.48	3.482

Total Area = 109649.9, Total Amount = 19.791, Total Height = 91709.06, Sample Units = PPB

ASX22 0738

Sample Name: **AR16X0924A AAAR16XAA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 09:36:22pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.33R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.609		0.0000	0.000	451.4	0.124	BV	0.050	149.12	0.042
2	2.704		0.0000	0.000	251.8	0.069	VB	0.032	132.64	0.038
3	2.761		0.0000	0.000	104.4	0.029	BB	0.031	55.88	0.016
4	2.809	TCX	9.0153	48.294	37482.4	10.290	BB	0.015	41248.45	11.707
5	2.855		0.0000	0.000	1218.3	0.334	BB	0.014	1472.16	0.418
6	2.929		0.0000	0.000	117.2	0.032	BB	0.015	129.25	0.037
7	2.984		0.0000	0.000	2909.4	0.799	BB	0.014	3376.31	0.958
8	3.046		0.0000	0.000	3534.4	0.970	BB	0.016	3796.48	1.078
9	3.099		0.0000	0.000	14602.3	4.009	BB	0.014	16996.47	4.824
10	3.213		0.0000	0.000	2365.8	0.649	BB	0.014	2898.89	0.823
11	3.310		0.0000	0.000	31584.3	8.671	BV	0.020	26365.06	7.483
12	3.361		0.0000	0.000	2836.0	0.779	VV	0.014	3311.48	0.940
13	3.383		0.0000	0.000	1159.8	0.318	VV	0.013	1496.48	0.425
14	3.406		0.0000	0.000	1308.1	0.359	VV	0.013	1638.71	0.465
15	3.433		0.0000	0.000	20893.1	5.736	VV	0.015	23646.65	6.711
16	3.514		0.0000	0.000	8801.1	2.416	VB	0.021	6833.71	1.940
17	3.589		0.0000	0.000	53238.4	14.616	BV	0.022	41235.02	11.703
18	3.625		0.0000	0.000	28043.5	7.699	VV	0.021	21921.50	6.222
19	3.701		0.0000	0.000	18146.4	4.982	VV	0.014	21256.89	6.033
20	3.737		0.0000	0.000	16085.8	4.416	VV	0.015	17685.06	5.019
21	3.761		0.0000	0.000	24971.1	6.855	VB	0.021	19972.53	5.669
22	3.871		0.0000	0.000	14390.9	3.951	BB	0.013	18218.17	5.171
23	3.896		0.0000	0.000	3263.6	0.896	BB	0.009	5793.94	1.644
24	3.918		0.0000	0.000	3554.9	0.976	BB	0.011	5183.90	1.471
25	3.950		0.0000	0.000	3070.0	0.843	BV	0.011	4686.28	1.330
26	3.966		0.0000	0.000	12715.8	3.491	VV	0.015	14502.82	4.116
27	4.011		0.0000	0.000	4684.7	1.286	VB	0.015	5106.42	1.449
28	4.056		0.0000	0.000	460.3	0.126	BV	0.014	564.25	0.160
29	4.080		0.0000	0.000	2728.7	0.749	VV	0.014	3347.47	0.950
30	4.104		0.0000	0.000	7592.6	2.084	VV	0.015	8300.45	2.356
31	4.130		0.0000	0.000	5030.9	1.381	VV	0.023	3680.26	1.045
32	4.207		0.0000	0.000	372.5	0.102	VV	0.017	370.19	0.105
33	4.245		0.0000	0.000	414.9	0.114	VV	0.013	522.36	0.148
34	4.261		0.0000	0.000	1080.3	0.297	VV	0.022	835.08	0.237
35	4.306		0.0000	0.000	964.6	0.265	VB	0.055	293.28	0.083
36	4.450		0.0000	0.000	139.4	0.038	BB	0.031	73.95	0.021
37	4.694		0.0000	0.000	594.7	0.163	BB	0.026	379.65	0.108
38	4.837		0.0000	0.000	956.4	0.263	BB	0.042	377.57	0.107
39	5.178		0.0000	0.000	157.8	0.043	BB	0.054	48.44	0.014
40	5.547		0.0000	0.000	124.7	0.034	BB	0.016	129.41	0.037
41	5.583		0.0000	0.000	545.7	0.150	BB	0.046	195.93	0.056
42	5.781		0.0000	0.000	245.0	0.067	BB	0.025	166.45	0.047
43	5.916		0.0000	0.000	120.2	0.033	BB	0.034	59.09	0.017
44	6.001		0.0000	0.000	158.1	0.043	BB	0.033	79.03	0.022
45	6.201		0.0000	0.000	945.4	0.260	BB	0.025	635.78	0.180
46	6.360	DCB	9.6524	51.706	29386.3	8.068	BB	0.021	23037.41	6.539
47	6.755		0.0000	0.000	399.5	0.110	BB	0.077	85.93	0.024
48	7.300		0.0000	0.000	48.9	0.013	BB	0.021	39.46	0.011

Total Area = 364251.5, Total Amount = 18.668, Total Height = 352331.6, Sample Units = PPB

ASX22 8739

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.33R

AR16X0924A AAAR16XAA ICAL 0930999999

C:\CPWIN\DATA1\1C6310.33R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 9:28:52 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.33R

AR16X0924A AAAR16XAA ICAL 0930999999

C:\CPWIN\DATA1\1C6310B.33R



Instrument ID: CP06--S9147B Injected On: 11/6/2009 9:28:52 PM Column ID: RTX CLPII

ASX22 8748

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min; hold Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.312	19023	9.72499	TCX	2.809	41248	9.01525	TCX
7.161	3193	10.06552	DCB	6.36	23037	9.65237	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.33A
Area File: C:\CPWIN\DATA1\1C6310B.33A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 9:36:28 PM
File Reported On: 11/6/2009 at 9:36:37 PM

REPLOTS

ASX22 0741

AR16X0924A AAAR16XAA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.33R



AR16X0924A AAAR16XAA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.33R



ASX22 8742

Sample Name: **MD16X0924A AAMD16XAA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 09:48:21pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.34R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.311	TCX	1.0870	44.239	2138.6	13.242	BB	0.017	2126.25	18.205
2	3.459		0.0000	0.000	100.0	0.619	BB	0.019	85.70	0.734
3	3.514		0.0000	0.000	105.8	0.655	BB	0.015	118.00	1.010
4	3.565		0.0000	0.000	416.0	2.576	BB	0.015	458.81	3.928
5	3.661		0.0000	0.000	131.4	0.813	BB	0.021	103.46	0.886
6	3.787		0.0000	0.000	870.6	5.390	BB	0.022	649.96	5.565
7	3.847		0.0000	0.000	96.1	0.595	BB	0.023	71.10	0.609
8	3.914		0.0000	0.000	500.7	3.100	BB	0.016	524.89	4.494
9	4.034		0.0000	0.000	121.1	0.750	BB	0.019	104.69	0.896
10	4.112		0.0000	0.000	1073.8	6.649	BB	0.027	658.30	5.636
11	4.155		0.0000	0.000	382.1	2.366	BB	0.016	391.39	3.351
12	4.220		0.0000	0.000	338.3	2.095	BB	0.016	348.62	2.985
13	4.292		0.0000	0.000	236.5	1.464	BB	0.015	265.63	2.274
14	4.323		0.0000	0.000	415.2	2.571	BB	0.030	230.53	1.974
15	4.432		0.0000	0.000	232.8	1.441	BB	0.016	248.96	2.132
16	4.460		0.0000	0.000	53.9	0.334	BB	0.012	77.32	0.662
17	4.532		0.0000	0.000	235.4	1.458	BB	0.020	200.53	1.717
19	4.707		0.0000	0.000	356.5	2.207	VB	0.024	252.45	2.162
20	4.779		0.0000	0.000	206.8	1.281	BV	0.045	75.88	0.650
21	4.876		0.0000	0.000	263.0	1.629	VV	0.024	181.40	1.553
22	5.106		0.0000	0.000	84.5	0.523	VB	0.017	82.97	0.710
23	5.149		0.0000	0.000	137.7	0.852	BB	0.028	82.31	0.705
24	5.203		0.0000	0.000	397.8	2.463	BB	0.031	213.69	1.830
25	5.284		0.0000	0.000	820.3	5.079	BB	0.032	424.75	3.637
26	5.441		0.0000	0.000	120.3	0.745	BB	0.026	75.66	0.648
27	5.485		0.0000	0.000	574.5	3.557	BB	0.026	369.07	3.160
28	5.538		0.0000	0.000	103.8	0.643	BB	0.016	110.77	0.948
29	5.578		0.0000	0.000	280.6	1.737	BB	0.029	160.31	1.373
30	5.693		0.0000	0.000	532.4	3.297	BB	0.025	359.67	3.079
31	5.757		0.0000	0.000	341.2	2.112	BB	0.019	303.32	2.597
32	5.799		0.0000	0.000	232.2	1.438	BB	0.024	160.30	1.373
33	5.921		0.0000	0.000	536.3	3.321	BB	0.029	310.98	2.663
34	5.979		0.0000	0.000	84.3	0.522	BB	0.021	67.07	0.574
35	6.014		0.0000	0.000	222.0	1.374	BB	0.037	100.49	0.860
36	6.153		0.0000	0.000	1005.0	6.222	BB	0.030	564.55	4.834
37	6.357		0.0000	0.000	339.6	2.103	BB	0.020	282.88	2.422
38	6.407		0.0000	0.000	219.9	1.361	BB	0.031	120.07	1.028
39	6.625		0.0000	0.000	139.9	0.866	BB	0.039	60.09	0.514
40	6.775		0.0000	0.000	207.0	1.282	BB	0.027	129.91	1.112
41	6.990		0.0000	0.000	194.5	1.204	BB	0.035	92.08	0.788
42	7.159	DCB	1.3701	55.761	1302.4	8.064	BB	0.050	434.70	3.722

Total Area = 16150.5, Total Amount = 2.457, Total Height = 11679.51, Sample Units = PPB

ASX22 0743

Sample Name: **MD16X0924A AAMD16XAA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 09:48:21pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.34R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.604		0.0000	0.000	373.5	0.594	BV	0.047	131.07	0.241
2	2.702		0.0000	0.000	326.4	0.519	VB	0.033	162.78	0.300
3	2.808	TCX	1.0546	47.255	4422.4	7.029	BB	0.015	4825.13	8.880
4	2.928		0.0000	0.000	123.4	0.196	BB	0.016	131.02	0.241
5	2.982		0.0000	0.000	207.6	0.330	BB	0.015	229.43	0.422
6	3.045		0.0000	0.000	219.5	0.349	BB	0.016	227.48	0.419
7	3.098		0.0000	0.000	916.0	1.456	BB	0.013	1152.35	2.121
8	3.212		0.0000	0.000	158.2	0.251	BB	0.014	189.62	0.349
9	3.309		0.0000	0.000	2151.1	3.419	BV	0.020	1822.32	3.354
10	3.361		0.0000	0.000	171.5	0.273	VV	0.014	209.74	0.386
11	3.406		0.0000	0.000	80.6	0.128	VV	0.012	116.23	0.214
12	3.432		0.0000	0.000	1483.0	2.357	VV	0.016	1579.63	2.907
13	3.514		0.0000	0.000	356.0	0.566	VV	0.013	456.64	0.840
14	3.530		0.0000	0.000	592.8	0.942	VV	0.015	670.57	1.234
15	3.588		0.0000	0.000	3515.0	5.586	VV	0.022	2679.77	4.932
16	3.624		0.0000	0.000	2055.8	3.267	VV	0.023	1503.67	2.767
17	3.701		0.0000	0.000	1252.6	1.991	VV	0.015	1401.93	2.580
18	3.736		0.0000	0.000	1101.9	1.751	VV	0.014	1268.31	2.334
19	3.761		0.0000	0.000	1926.3	3.061	VV	0.023	1408.25	2.592
20	3.870		0.0000	0.000	1717.1	2.729	VB	0.020	1406.24	2.588
21	3.918		0.0000	0.000	397.3	0.631	BV	0.017	395.58	0.728
22	3.949		0.0000	0.000	285.8	0.454	VV	0.013	371.90	0.684
23	3.965		0.0000	0.000	951.3	1.512	VV	0.017	954.57	1.757
24	4.010		0.0000	0.000	405.1	0.644	VB	0.019	348.49	0.641
25	4.079		0.0000	0.000	145.6	0.231	BV	0.013	189.24	0.348
26	4.104		0.0000	0.000	1258.1	2.000	VB	0.018	1172.19	2.157
27	4.206		0.0000	0.000	654.6	1.040	BB	0.015	739.98	1.362
28	4.306		0.0000	0.000	1106.0	1.758	BB	0.058	317.56	0.584
29	4.463		0.0000	0.000	350.3	0.557	BV	0.017	347.01	0.639
30	4.493		0.0000	0.000	985.9	1.567	VB	0.016	1009.06	1.857
31	4.522		0.0000	0.000	314.8	0.500	BB	0.015	359.74	0.662
32	4.574		0.0000	0.000	2238.2	3.557	BB	0.018	2082.23	3.832
33	4.671		0.0000	0.000	339.8	0.540	BV	0.016	356.10	0.655
34	4.714		0.0000	0.000	2954.0	4.695	VB	0.023	2165.26	3.985
35	4.822		0.0000	0.000	2922.1	4.644	BV	0.024	2055.78	3.783
36	4.878		0.0000	0.000	283.8	0.451	VB	0.018	258.67	0.476
37	4.916		0.0000	0.000	384.0	0.610	BB	0.015	426.60	0.785
38	4.953		0.0000	0.000	2035.5	3.235	BV	0.021	1646.57	3.030
39	4.986		0.0000	0.000	1834.7	2.916	VB	0.018	1732.39	3.188
40	5.025		0.0000	0.000	727.2	1.156	BB	0.016	762.79	1.404
41	5.127		0.0000	0.000	329.4	0.524	BB	0.017	314.56	0.579
42	5.180		0.0000	0.000	2193.1	3.485	BV	0.020	1821.58	3.352
43	5.252		0.0000	0.000	998.0	1.586	VV	0.019	883.06	1.625
44	5.288		0.0000	0.000	398.3	0.633	VV	0.017	385.36	0.709
45	5.313		0.0000	0.000	284.3	0.452	VB	0.019	252.37	0.464
46	5.368		0.0000	0.000	4183.0	6.648	BB	0.022	3235.54	5.955
47	5.445		0.0000	0.000	146.0	0.232	BB	0.018	137.35	0.253
48	5.582		0.0000	0.000	1212.7	1.927	BB	0.022	913.82	1.682
49	5.627		0.0000	0.000	2799.2	4.449	BB	0.021	2176.71	4.006
50	5.782		0.0000	0.000	179.0	0.285	BB	0.020	145.90	0.269
51	5.898		0.0000	0.000	637.0	1.012	BB	0.025	416.78	0.767
52	6.001		0.0000	0.000	1230.0	1.955	BB	0.022	922.02	1.697
53	6.200		0.0000	0.000	891.1	1.416	BB	0.025	596.10	1.097
54	6.360	DCB	1.1771	52.745	3633.8	5.775	BB	0.022	2809.40	5.170
55	6.757		0.0000	0.000	80.5	0.128	BB	0.022	61.57	0.113

Total Area = 62920.5, Total Amount = 2.232, Total Height = 54336.0, Sample Units = PPB

ASX22 8744

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.34R
MD16X0924A AAMD16XAA ICAL 0930999999
------ C:\CPWIN\DATA1\1C6310.34R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 9:40:51 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.34R
MD16X0924A AAMD16XAA ICAL 0930999999
------ C:\CPWIN\DATA1\1C6310B.34R



Instrument ID: CP06--S9147B Injected On: 11/6/2009 9:40:51 PM Column ID: RTX CLPII

ASX22 0745

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	2126	1.08699	TCX	2.808	4825	1.05458	TCX
7.159	435	1.37012	DCB	6.36	2809	1.1771	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.34A
Area File: C:\CPWIN\DATA1\1C6310B.34A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 9:48:28 PM
File Reported On: 11/6/2009 at 9:48:36 PM

REPLOTS

ASX22 0746

MD16X0924A AAMD16XAA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.34R



MD16X0924A AAMD16XAA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.34R



ASX22 8747

Sample Name: **MD62X0924B AAMD62XAA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 10:00:21pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.35R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.100		0.0000	0.000	67.4	0.633	BB	0.028	40.68	0.664
2	3.141		0.0000	0.000	256.6	2.409	BB	0.051	84.23	1.375
3	3.244		0.0000	0.000	401.3	3.768	BB	0.073	91.59	1.495
4	3.310	TCX	0.0755	11.417	199.3	1.871	BB	0.022	147.75	2.411
5	3.643		0.0000	0.000	96.0	0.902	BB	0.034	46.73	0.763
6	4.073		0.0000	0.000	363.1	3.409	BB	0.048	126.93	2.072
7	4.707		0.0000	0.000	200.1	1.878	BB	0.026	127.22	2.076
8	4.778		0.0000	0.000	274.1	2.573	BV	0.057	79.84	1.303
9	4.877		0.0000	0.000	300.8	2.824	VB	0.053	93.74	1.530
10	5.108		0.0000	0.000	95.7	0.899	BB	0.021	74.97	1.224
11	5.202		0.0000	0.000	446.4	4.191	BB	0.030	248.33	4.053
12	5.283		0.0000	0.000	551.2	5.175	BB	0.023	397.61	6.489
13	5.436		0.0000	0.000	131.7	1.236	BB	0.025	86.43	1.411
14	5.484		0.0000	0.000	484.4	4.547	BB	0.022	359.77	5.872
15	5.539		0.0000	0.000	217.3	2.040	BB	0.018	196.99	3.215
16	5.579		0.0000	0.000	210.1	1.973	BB	0.028	124.58	2.033
17	5.690		0.0000	0.000	371.7	3.489	BB	0.021	290.48	4.741
18	5.755		0.0000	0.000	590.1	5.540	BB	0.019	517.13	8.440
19	5.797		0.0000	0.000	241.0	2.262	BB	0.021	195.21	3.186
20	5.920		0.0000	0.000	570.9	5.359	BB	0.023	416.15	6.792
21	5.977		0.0000	0.000	672.6	6.315	BB	0.069	162.96	2.660
22	6.152		0.0000	0.000	1660.2	15.586	BB	0.035	789.80	12.890
23	6.355		0.0000	0.000	609.8	5.725	BB	0.022	467.47	7.630
24	6.407		0.0000	0.000	395.9	3.716	BB	0.027	248.74	4.060
25	6.623		0.0000	0.000	123.1	1.155	BB	0.018	112.68	1.839
26	6.775		0.0000	0.000	475.7	4.466	BB	0.031	259.79	4.240
27	6.987		0.0000	0.000	281.5	2.643	BB	0.031	153.40	2.504
28	7.160	DCB	0.5861	88.583	363.7	3.414	BB	0.033	185.95	3.035

Total Area = 10651.5, Total Amount = 0.662, Total Height = 6127.15, Sample Units = PPB

ASX22 8748

Sample Name: **MD62X0924B AAMD62XAA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 10:00:21pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.35R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.606		0.0000	0.000	235.0	0.458	BV	0.040	98.48	0.248
2	2.700		0.0000	0.000	948.3	1.848	VB	0.017	926.23	2.337
3	2.807	TCX	0.0738	12.782	287.5	0.560	BB	0.014	337.62	0.852
4	2.927		0.0000	0.000	161.1	0.314	BB	0.023	117.42	0.296
5	3.066		0.0000	0.000	54.3	0.106	BB	0.018	51.02	0.129
6	3.310		0.0000	0.000	128.2	0.250	BB	0.019	112.01	0.283
7	3.385		0.0000	0.000	161.5	0.315	BB	0.017	161.10	0.406
8	3.530		0.0000	0.000	730.5	1.423	BV	0.013	914.88	2.308
9	3.577		0.0000	0.000	290.7	0.566	VV	0.033	145.16	0.366
10	3.622		0.0000	0.000	499.5	0.973	VV	0.079	105.98	0.267
11	3.735		0.0000	0.000	190.0	0.370	VV	0.019	165.10	0.417
12	3.762		0.0000	0.000	316.5	0.617	VV	0.025	213.12	0.538
13	3.803		0.0000	0.000	410.7	0.800	VV	0.055	124.86	0.315
14	3.868		0.0000	0.000	522.2	1.017	VB	0.042	206.45	0.521
15	3.930		0.0000	0.000	688.3	1.341	BV	0.060	191.23	0.483
16	4.043		0.0000	0.000	419.6	0.817	VV	0.035	199.99	0.505
17	4.103		0.0000	0.000	1282.5	2.499	VV	0.038	555.26	1.401
18	4.206		0.0000	0.000	1209.6	2.357	VV	0.040	504.73	1.274
19	4.307		0.0000	0.000	1476.0	2.876	VB	0.065	378.01	0.954
20	4.462		0.0000	0.000	265.2	0.517	BV	0.015	298.64	0.754
21	4.493		0.0000	0.000	1046.1	2.038	VB	0.015	1195.03	3.015
22	4.520		0.0000	0.000	245.0	0.477	BB	0.016	252.48	0.637
23	4.574		0.0000	0.000	2158.0	4.205	BV	0.018	2002.15	5.052
24	4.670		0.0000	0.000	196.8	0.383	VV	0.015	222.85	0.562
25	4.714		0.0000	0.000	2857.5	5.567	VB	0.023	2073.17	5.231
26	4.822		0.0000	0.000	3045.0	5.933	BV	0.023	2233.23	5.635
27	4.877		0.0000	0.000	399.2	0.778	VV	0.017	396.32	1.000
28	4.916		0.0000	0.000	525.5	1.024	VB	0.017	530.04	1.337
29	4.952		0.0000	0.000	1187.9	2.315	BV	0.018	1093.12	2.758
30	4.985		0.0000	0.000	2999.6	5.844	VB	0.017	2876.17	7.257
31	5.024		0.0000	0.000	1083.5	2.111	BB	0.016	1098.16	2.771
32	5.126		0.0000	0.000	605.7	1.180	BB	0.018	563.24	1.421
33	5.178		0.0000	0.000	3428.2	6.679	BV	0.020	2798.93	7.062
34	5.251		0.0000	0.000	1564.8	3.049	VV	0.021	1265.74	3.194
35	5.287		0.0000	0.000	471.6	0.919	VV	0.019	419.42	1.058
36	5.312		0.0000	0.000	399.5	0.778	VB	0.019	344.18	0.868
37	5.368		0.0000	0.000	5834.2	11.367	BB	0.021	4555.60	11.495
38	5.443		0.0000	0.000	390.0	0.760	BB	0.016	396.63	1.001
39	5.581		0.0000	0.000	2511.3	4.893	BB	0.019	2234.07	5.637
40	5.628		0.0000	0.000	2549.2	4.967	BB	0.021	2047.36	5.166
41	5.782		0.0000	0.000	287.0	0.559	BB	0.017	276.31	0.697
42	5.850		0.0000	0.000	202.8	0.395	BB	0.020	167.35	0.422
43	5.897		0.0000	0.000	1028.3	2.004	BB	0.023	739.98	1.867
44	5.999		0.0000	0.000	2378.1	4.633	BB	0.023	1703.29	4.298
45	6.199		0.0000	0.000	1241.9	2.420	BB	0.020	1043.72	2.634
46	6.359	DCB	0.5035	87.218	1758.1	3.425	BB	0.024	1201.71	3.032
47	6.865		0.0000	0.000	541.6	1.055	BB	0.147	61.44	0.155
48	7.343		0.0000	0.000	112.0	0.218	BB	0.056	33.26	0.084

Total Area = 51325.4, Total Amount = 0.577, Total Height = 39632.26, Sample Units = PPB

ASX22 0759

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.35R
MD62X0924B AAMD62XAA ICAL 0930999999
——— C:\CPWIN\DATA1\1C6310.35R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 9:52:51 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.35R
MD62X0924B AAMD62XAA ICAL 0930999999
——— C:\CPWIN\DATA1\1C6310B.35R



Instrument ID: CP06--S9147B Injected On: 11/6/2009 9:52:51 PM Column ID: RTX CLPII

ASX22 0750

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.31	148	.07553	TCX	2.807	338	.07379	TCX
7.16	186	.58609	DCB	6.359	1202	.5035	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.35A
Area File: C:\CPWIN\DATA1\1C6310B.35A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 10:00:28 PM
File Reported On: 11/6/2009 at 10:00:37 PM

REPLOTS

8SX22 8751

MD62X0924B AAMD62XAA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.35R



MD62X0924B AAMD62XAA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.35R



ASX22 0752

Sample Name: **MD68X0924C AAMD68XAA ICAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 10:12:18pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.36R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.140		0.0000	0.000	381.8	3.112	BB	0.045	142.88	2.272
2	3.328	TCX	0.0330	2.425	50.9	0.415	BB	0.013	64.57	1.027
3	4.073		0.0000	0.000	221.8	1.808	BB	0.018	205.56	3.269
4	4.430		0.0000	0.000	659.6	5.377	BB	0.212	51.88	0.825
5	4.781		0.0000	0.000	340.0	2.771	BB	0.059	95.57	1.520
6	5.143		0.0000	0.000	117.9	0.961	BB	0.027	72.91	1.159
7	5.444		0.0000	0.000	390.6	3.184	BB	0.057	114.78	1.825
8	5.549		0.0000	0.000	72.0	0.587	BB	0.028	43.36	0.690
9	5.754		0.0000	0.000	299.0	2.437	BB	0.025	202.27	3.216
10	5.936		0.0000	0.000	407.6	3.322	BB	0.024	286.06	4.549
11	6.001		0.0000	0.000	682.3	5.562	BB	0.064	177.83	2.828
12	6.151		0.0000	0.000	70.1	0.571	BB	0.015	76.13	1.211
13	6.352		0.0000	0.000	1136.5	9.264	BB	0.022	860.48	13.683
14	6.409		0.0000	0.000	1068.5	8.709	BB	0.026	689.14	10.958
15	6.542		0.0000	0.000	985.0	8.029	BB	0.023	715.21	11.373
16	6.603		0.0000	0.000	390.3	3.182	BB	0.036	182.17	2.897
17	6.774		0.0000	0.000	842.3	6.866	BB	0.050	280.89	4.466
18	6.989		0.0000	0.000	3319.5	27.058	BB	0.034	1605.57	25.531
19	7.158	DCB	1.3284	97.575	832.4	6.785	BB	0.033	421.44	6.702

Total Area = 12268.1, Total Amount = 1.361, Total Height = 6288.72, Sample Units = PPB

ASX22 8753

Sample Name: **MD68X0924C AAMD68XAA ICAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 10:12:18pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.36R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.609		0.0000	0.000	559.2	1.047	BV	0.053	176.23	0.432
2	2.700		0.0000	0.000	360.7	0.675	VB	0.024	247.50	0.607
3	2.739		0.0000	0.000	353.4	0.662	BB	0.024	250.63	0.615
4	2.897		0.0000	0.000	68.7	0.129	BB	0.013	87.39	0.214
5	2.929		0.0000	0.000	136.4	0.255	BB	0.016	138.77	0.340
6	3.094		0.0000	0.000	254.6	0.477	BV	0.062	67.99	0.167
7	3.386		0.0000	0.000	115.3	0.216	VV	0.020	96.14	0.236
8	3.429		0.0000	0.000	207.0	0.388	VV	0.042	82.38	0.202
9	3.530		0.0000	0.000	1331.5	2.493	VV	0.015	1518.70	3.726
10	3.614		0.0000	0.000	103.2	0.193	VB	0.022	76.81	0.188
11	3.760		0.0000	0.000	227.7	0.426	BB	0.023	161.91	0.397
12	3.887		0.0000	0.000	310.3	0.581	BB	0.027	189.88	0.466
13	3.932		0.0000	0.000	898.0	1.682	BV	0.058	257.34	0.631
14	4.035		0.0000	0.000	759.0	1.421	VV	0.065	196.10	0.481
15	4.121		0.0000	0.000	592.4	1.109	VV	0.049	200.07	0.491
16	4.270		0.0000	0.000	81.1	0.152	VV	0.015	89.24	0.219
17	4.306		0.0000	0.000	550.8	1.032	VB	0.032	286.45	0.703
18	4.486		0.0000	0.000	1331.4	2.493	BB	0.105	211.86	0.520
19	4.693		0.0000	0.000	496.4	0.930	BB	0.024	351.39	0.862
20	4.834		0.0000	0.000	983.2	1.841	BV	0.038	426.71	1.047
21	4.954		0.0000	0.000	66.6	0.125	VV	0.019	59.78	0.147
22	4.985		0.0000	0.000	1632.4	3.057	VB	0.024	1156.73	2.838
23	5.178		0.0000	0.000	1857.6	3.479	BV	0.019	1602.08	3.931
24	5.243		0.0000	0.000	528.3	0.989	VB	0.020	446.46	1.095
25	5.367		0.0000	0.000	782.9	1.466	BB	0.023	571.05	1.401
26	5.444		0.0000	0.000	500.7	0.938	BB	0.023	366.20	0.898
27	5.581		0.0000	0.000	5947.1	11.137	BB	0.018	5389.55	13.223
28	5.629		0.0000	0.000	4437.5	8.310	BB	0.018	4049.81	9.936
29	5.780		0.0000	0.000	5115.1	9.579	BB	0.019	4487.47	11.010
30	5.849		0.0000	0.000	1290.4	2.417	BB	0.016	1313.47	3.223
31	5.943		0.0000	0.000	370.3	0.693	BB	0.135	45.76	0.112
32	5.999		0.0000	0.000	2165.2	4.055	BB	0.022	1658.26	4.069
33	6.199		0.0000	0.000	15132.7	28.338	BB	0.022	11458.87	28.115
34	6.358	DCB	1.2249	100.000	3591.5	6.726	BB	0.020	2923.49	7.173
35	6.630		0.0000	0.000	99.1	0.186	BB	0.042	39.75	0.098
36	6.754		0.0000	0.000	162.2	0.304	BB	0.036	75.59	0.185

Total Area = 53399.8, Total Amount = 1.225, Total Height = 40757.81, Sample Units = PPB

ASX22 0754

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.36R
MD68X0924C AAMD68XAA ICAL 0930999999
——— C:\CPWIN\DATA1\1C6310.36R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 10:04:48 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.36R
MD68X0924C AAMD68XAA ICAL 0930999999
——— C:\CPWIN\DATA1\1C6310B.36R



ASX22 8755

Instrument ID: CP06--S9147B Injected On: 11/6/2009 10:04:48 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.328	65	.03301	TCX	.	0	.	TCX
7.158	421	1.32835	DCB	6.358	2923	1.2249	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.36A
Area File: C:\CPWIN\DATA1\1C6310B.36A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 10:12:24 PM
File Reported On: 11/6/2009 at 10:12:33 PM

REPLOTS

ASX22 0756

MD68X0924C AAMD68XAA ICAL 0930999999
C:\CPWIN\DATA1\1C6310.36R



MD68X0924C AAMD68XAA ICAL 0930999999
C:\CPWIN\DATA1\1C6310B.36R



ASX22 0757

Sample Name: **IC16X0924A AAIC16XAA CCAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 10:24:15pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.37R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.172		0.0000	0.000	72.9	0.027	BB	0.015	81.26	0.037
2	3.223		0.0000	0.000	400.3	0.148	BB	0.055	121.28	0.055
3	3.310	TCX	40.3354	51.409	74657.1	27.524	BB	0.016	78900.06	35.716
4	3.457		0.0000	0.000	1083.7	0.400	BB	0.016	1106.74	0.501
5	3.514		0.0000	0.000	2007.9	0.740	BB	0.014	2309.59	1.045
6	3.563		0.0000	0.000	5606.8	2.067	BB	0.015	6189.50	2.802
7	3.661		0.0000	0.000	1167.4	0.430	BB	0.017	1118.34	0.506
8	3.787		0.0000	0.000	4765.2	1.757	BB	0.013	6248.36	2.828
9	3.805		0.0000	0.000	1192.0	0.439	BB	0.008	2507.31	1.135
10	3.845		0.0000	0.000	1259.7	0.464	BB	0.025	829.63	0.376
11	3.914		0.0000	0.000	7372.8	2.718	BB	0.016	7713.22	3.492
12	3.992		0.0000	0.000	229.0	0.084	BB	0.022	171.42	0.078
13	4.032		0.0000	0.000	2600.5	0.959	BB	0.025	1768.65	0.801
14	4.111		0.0000	0.000	17056.6	6.288	BB	0.028	10281.75	4.654
15	4.154		0.0000	0.000	5557.6	2.049	BB	0.017	5429.19	2.458
16	4.220		0.0000	0.000	5047.5	1.861	BB	0.017	4856.87	2.199
17	4.265		0.0000	0.000	228.3	0.084	BB	0.013	299.15	0.135
18	4.291		0.0000	0.000	3443.3	1.269	BB	0.015	3839.89	1.738
19	4.323		0.0000	0.000	5344.6	1.970	BB	0.028	3174.74	1.437
20	4.431		0.0000	0.000	3585.2	1.322	BB	0.016	3697.72	1.674
21	4.459		0.0000	0.000	604.8	0.223	BB	0.010	963.60	0.436
22	4.493		0.0000	0.000	883.2	0.326	BB	0.014	1039.66	0.471
23	4.531		0.0000	0.000	3173.8	1.170	BB	0.020	2628.05	1.190
24	4.577		0.0000	0.000	715.6	0.264	BB	0.016	723.25	0.327
25	4.707		0.0000	0.000	4589.3	1.692	BV	0.023	3351.09	1.517
26	4.756		0.0000	0.000	994.5	0.367	VV	0.031	526.35	0.238
27	4.827		0.0000	0.000	562.3	0.207	VV	0.022	426.82	0.193
28	4.875		0.0000	0.000	5055.4	1.864	VV	0.029	2955.89	1.338
29	5.031		0.0000	0.000	174.6	0.064	VB	0.023	124.93	0.057
30	5.106		0.0000	0.000	2235.1	0.824	BV	0.026	1415.34	0.641
31	5.153		0.0000	0.000	1395.0	0.514	VB	0.020	1174.96	0.532
32	5.202		0.0000	0.000	3881.3	1.431	BB	0.023	2785.30	1.261
33	5.239		0.0000	0.000	277.2	0.102	BB	0.011	416.43	0.189
34	5.282		0.0000	0.000	8459.2	3.119	BB	0.024	5916.74	2.678
35	5.372		0.0000	0.000	682.9	0.252	BB	0.019	604.56	0.274
36	5.431		0.0000	0.000	1011.3	0.373	BB	0.020	856.49	0.388
37	5.484		0.0000	0.000	8830.0	3.255	BB	0.028	5301.42	2.400
38	5.538		0.0000	0.000	1079.1	0.398	BB	0.016	1151.83	0.521
39	5.577		0.0000	0.000	3730.4	1.375	BB	0.025	2438.63	1.104
40	5.656		0.0000	0.000	308.0	0.114	BB	0.017	305.19	0.138
41	5.692		0.0000	0.000	8041.3	2.965	BB	0.024	5515.94	2.497
42	5.755		0.0000	0.000	3718.5	1.371	BB	0.019	3222.50	1.459
43	5.797		0.0000	0.000	2093.9	0.772	BB	0.019	1818.96	0.823
44	5.919		0.0000	0.000	6452.9	2.379	BB	0.028	3895.86	1.764
45	5.976		0.0000	0.000	1615.9	0.596	BB	0.018	1488.99	0.674
46	6.015		0.0000	0.000	878.5	0.324	BB	0.018	802.87	0.363
47	6.096		0.0000	0.000	1001.4	0.369	BB	0.019	857.40	0.388
48	6.151		0.0000	0.000	14302.7	5.273	BB	0.033	7267.09	3.290
49	6.357		0.0000	0.000	4671.5	1.722	BB	0.021	3710.09	1.679
50	6.405		0.0000	0.000	1851.9	0.683	BB	0.026	1204.68	0.545
51	6.542		0.0000	0.000	192.3	0.071	BB	0.019	168.34	0.076
52	6.625		0.0000	0.000	1469.8	0.542	BB	0.030	829.58	0.376
53	6.775		0.0000	0.000	3216.4	1.186	BB	0.032	1651.94	0.748
54	6.988		0.0000	0.000	1152.6	0.425	BB	0.030	629.93	0.285
55	7.160	DCB	38.1238	48.591	29264.2	10.789	BB	0.040	12095.49	5.475

Total Area = 271244.9, Total Amount = 78.459, Total Height = 220910.8, Sample Units = PPB

ASX22 8758

Sample Name: **IC16X0924A AAIC16XAA CCAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 10:24:15pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.37R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.614		0.0000	0.000	562.5	0.055	BV	0.057	165.35	0.018
2	2.700		0.0000	0.000	938.1	0.092	VB	0.024	650.67	0.069
3	2.807	TCX	38.6833	51.351	164135.8	16.124	BB	0.015	176992.10	18.881
4	2.927		0.0000	0.000	93.1	0.009	BB	0.012	132.30	0.014
5	2.982		0.0000	0.000	2597.9	0.255	BB	0.014	3146.44	0.336
6	3.044		0.0000	0.000	4888.6	0.480	BB	0.022	3680.45	0.393
7	3.097		0.0000	0.000	13446.9	1.321	BB	0.014	16485.12	1.759
8	3.147		0.0000	0.000	360.1	0.035	BB	0.012	498.08	0.053
9	3.211		0.0000	0.000	2373.0	0.233	BB	0.014	2903.20	0.310
10	3.308		0.0000	0.000	31679.2	3.112	BV	0.020	26877.50	2.867
11	3.360		0.0000	0.000	3848.3	0.378	VV	0.020	3275.10	0.349
12	3.404		0.0000	0.000	1174.4	0.115	VV	0.012	1570.45	0.168
13	3.431		0.0000	0.000	20087.3	1.973	VV	0.014	23396.84	2.496
14	3.466		0.0000	0.000	737.1	0.072	VV	0.014	871.44	0.093
15	3.494		0.0000	0.000	330.1	0.032	VV	0.012	461.85	0.049
16	3.513		0.0000	0.000	8324.1	0.818	VB	0.020	6856.07	0.731
17	3.587		0.0000	0.000	53140.7	5.220	BV	0.021	41400.11	4.417
18	3.624		0.0000	0.000	28324.2	2.782	VV	0.021	22124.15	2.360
19	3.700		0.0000	0.000	18219.7	1.790	VV	0.014	21110.93	2.252
20	3.736		0.0000	0.000	15929.6	1.565	VV	0.014	18591.74	1.983
21	3.760		0.0000	0.000	25618.0	2.517	VB	0.021	20376.71	2.174
22	3.870		0.0000	0.000	16701.7	1.641	BV	0.014	19941.88	2.127
23	3.894		0.0000	0.000	5250.3	0.516	VB	0.012	7307.40	0.780
24	3.917		0.0000	0.000	3136.4	0.308	BB	0.011	4633.04	0.494
25	3.948		0.0000	0.000	2989.0	0.294	BV	0.012	4109.55	0.438
26	3.965		0.0000	0.000	9998.3	0.982	VV	0.014	11659.42	1.244
27	4.010		0.0000	0.000	3179.6	0.312	VB	0.015	3646.57	0.389
28	4.053		0.0000	0.000	159.4	0.016	BB	0.013	210.35	0.022
29	4.079		0.0000	0.000	1119.5	0.110	BV	0.012	1515.03	0.162
30	4.103		0.0000	0.000	16564.2	1.627	VV	0.018	15729.26	1.678
31	4.169		0.0000	0.000	1918.1	0.188	VV	0.015	2114.95	0.226
32	4.206		0.0000	0.000	14270.5	1.402	VV	0.017	13973.48	1.491
33	4.259		0.0000	0.000	1828.2	0.180	VV	0.021	1422.68	0.152
34	4.305		0.0000	0.000	634.7	0.062	VV	0.032	335.34	0.036
35	4.357		0.0000	0.000	557.0	0.055	VB	0.017	560.90	0.060
36	4.416		0.0000	0.000	1970.1	0.194	BB	0.015	2261.48	0.241
37	4.463		0.0000	0.000	4717.4	0.463	BV	0.014	5525.45	0.589
38	4.492		0.0000	0.000	16241.1	1.595	VB	0.018	15006.77	1.601
39	4.521		0.0000	0.000	5687.5	0.559	BB	0.017	5708.68	0.609
40	4.574		0.0000	0.000	33472.1	3.288	BV	0.018	31294.67	3.338
41	4.644		0.0000	0.000	1367.8	0.134	VV	0.015	1526.00	0.163
42	4.670		0.0000	0.000	6847.6	0.673	VV	0.017	6880.24	0.734
43	4.713		0.0000	0.000	39819.3	3.912	VB	0.020	33900.85	3.617
44	4.821		0.0000	0.000	37429.7	3.677	BV	0.020	31157.03	3.324
45	4.878		0.0000	0.000	3616.4	0.355	VB	0.017	3628.98	0.387
46	4.915		0.0000	0.000	6521.5	0.641	BB	0.016	6875.79	0.734
47	4.952		0.0000	0.000	35892.5	3.526	BV	0.020	29493.20	3.146
48	4.984		0.0000	0.000	24201.6	2.377	VB	0.018	22992.86	2.453
49	5.024		0.0000	0.000	10237.3	1.006	BB	0.015	11050.16	1.179
50	5.127		0.0000	0.000	4739.2	0.466	BB	0.019	4250.66	0.453
51	5.179		0.0000	0.000	29877.8	2.935	BV	0.020	24385.27	2.601
52	5.252		0.0000	0.000	13551.0	1.331	VV	0.019	12122.20	1.293
53	5.287		0.0000	0.000	5970.4	0.586	VV	0.017	5915.43	0.631
54	5.313		0.0000	0.000	4248.1	0.417	VV	0.017	4139.96	0.442
55	5.367		0.0000	0.000	58763.7	5.773	VB	0.022	45268.83	4.829
56	5.443		0.0000	0.000	1367.6	0.134	BB	0.014	1596.77	0.170
57	5.558		0.0000	0.000	623.5	0.061	BB	0.024	440.35	0.047
58	5.581		0.0000	0.000	10280.2	1.010	BB	0.016	10464.71	1.116
59	5.626		0.0000	0.000	42459.3	4.171	BB	0.022	31546.74	3.365
60	5.781		0.0000	0.000	962.6	0.095	BB	0.016	1023.75	0.109
61	5.820		0.0000	0.000	454.1	0.045	BB	0.015	494.08	0.053
62	5.848		0.0000	0.000	361.9	0.036	BB	0.015	406.24	0.043

ASX22 8759

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.897		0.0000	0.000	6426.7	0.631	BB	0.020	5423.63	0.579
64	6.000		0.0000	0.000	16161.3	1.588	BB	0.023	11598.22	1.237
65	6.199		0.0000	0.000	5829.6	0.573	BB	0.020	4784.61	0.510
66	6.359	DCB	36.6480	48.649	112724.2	11.073	BB	0.021	87468.19	9.331
67	7.435		0.0000	0.000	36.5	0.004	BB	0.021	29.17	0.003

Total Area = 1017975.0, Total Amount = 75.331, Total Height = 937387.4, Sample Units = PPB

ASX22 E760

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.37R
IC16X0924A AAIC16XAA CCAL 0930999999
C:\CPWIN\DATA1\1C6310.37R



Instrument ID: CP06–S9147A Injected On: 11/6/2009 10:16:45 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.37R
IC16X0924A AAIC16XAA CCAL 0930999999
C:\CPWIN\DATA1\1C6310B.37R



ASX22 0761

Instrument ID: CP06–S9147B Injected On: 11/6/2009 10:16:45 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Vol/Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.31	78900	40.33541	TCX	2.807	176992	38.68335	TCX
7.16	12095	38.12377	DCB	6.359	87468	36.64803	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.37A
Area File: C:\CPWIN\DATA1\1C6310B.37A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 10:24:20 PM
File Reported On: 11/6/2009 at 10:24:29 PM

REPLOTS

ASX22 8762

IC16X0924A AAIC16XAA CCAL 0930999999
C:\CPWIN\DATA1\1C6310.37R



IC16X0924A AAIC16XAA CCAL 0930999999
C:\CPWIN\DATA1\1C6310B.37R



ASX22 8763

Sample Name: **IC48X0924B AAIC48XAA CCAL 0930999999**
Acquired from CP06–S9147A via port 1 on 11/6/09 10:36:19pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.38R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.260		0.0000	0.000	105.8	0.050	BB	0.050	35.29	0.021
2	3.312 TCX		42.7712	50.880	79473.3	37.790	BB	0.016	83664.73	49.180
3	3.514		0.0000	0.000	1132.1	0.538	BB	0.015	1271.13	0.747
4	3.564		0.0000	0.000	565.2	0.269	BB	0.015	628.86	0.370
5	3.659		0.0000	0.000	472.9	0.225	BB	0.027	296.90	0.175
6	3.788		0.0000	0.000	5063.3	2.408	BB	0.021	4055.07	2.384
7	3.846		0.0000	0.000	241.6	0.115	BB	0.017	233.66	0.137
8	3.914		0.0000	0.000	2666.5	1.268	BB	0.016	2807.11	1.650
9	3.980		0.0000	0.000	192.3	0.091	BB	0.020	159.72	0.094
10	4.033		0.0000	0.000	615.2	0.293	BB	0.019	534.58	0.314
11	4.112		0.0000	0.000	9328.2	4.436	BB	0.032	4840.22	2.845
12	4.155		0.0000	0.000	2075.3	0.987	BB	0.017	2046.32	1.203
13	4.220		0.0000	0.000	2999.0	1.426	BB	0.017	2942.85	1.730
14	4.266		0.0000	0.000	238.4	0.113	BB	0.013	315.04	0.185
15	4.292		0.0000	0.000	4778.1	2.272	BB	0.016	5129.76	3.015
16	4.323		0.0000	0.000	6193.2	2.945	BB	0.027	3857.32	2.267
17	4.432		0.0000	0.000	5135.9	2.442	BB	0.016	5258.85	3.091
18	4.460		0.0000	0.000	809.2	0.385	BB	0.012	1155.55	0.679
19	4.494		0.0000	0.000	1308.1	0.622	BB	0.015	1475.36	0.867
20	4.532		0.0000	0.000	4462.7	2.122	BB	0.020	3807.19	2.238
21	4.578		0.0000	0.000	1346.5	0.640	BB	0.018	1229.97	0.723
22	4.668		0.0000	0.000	323.4	0.154	BB	0.017	324.95	0.191
23	4.708		0.0000	0.000	4446.5	2.114	BV	0.018	4049.33	2.380
24	4.729		0.0000	0.000	5711.3	2.716	VV	0.020	4783.89	2.812
25	4.756		0.0000	0.000	8042.4	3.824	VV	0.030	4454.12	2.618
26	4.827		0.0000	0.000	1493.6	0.710	VV	0.027	912.29	0.536
27	4.874		0.0000	0.000	6359.2	3.024	VV	0.023	4538.61	2.668
28	4.899		0.0000	0.000	6527.6	3.104	VV	0.045	2413.00	1.418
29	5.031		0.0000	0.000	1573.4	0.748	VV	0.026	1027.56	0.604
30	5.088		0.0000	0.000	2105.2	1.001	VV	0.024	1492.80	0.877
31	5.122		0.0000	0.000	1383.7	0.658	VV	0.023	1021.00	0.600
32	5.154		0.0000	0.000	3743.5	1.780	VB	0.027	2280.14	1.340
33	5.245		0.0000	0.000	707.9	0.337	BB	0.018	638.90	0.376
34	5.282		0.0000	0.000	172.3	0.082	BB	0.019	149.55	0.088
35	5.335		0.0000	0.000	152.4	0.072	BB	0.017	149.73	0.088
36	5.377		0.0000	0.000	1414.1	0.672	BB	0.025	925.66	0.544
37	5.485		0.0000	0.000	262.0	0.125	BB	0.035	124.43	0.073
38	5.580		0.0000	0.000	1816.3	0.864	BB	0.030	1022.57	0.601
39	5.694		0.0000	0.000	579.8	0.276	BB	0.031	309.82	0.182
40	5.904		0.0000	0.000	231.1	0.110	BB	0.031	122.84	0.072
41	6.016		0.0000	0.000	231.3	0.110	BB	0.039	99.06	0.058
42	6.153		0.0000	0.000	92.2	0.044	BB	0.020	76.12	0.045
43	6.358		0.0000	0.000	287.8	0.137	BB	0.048	100.82	0.059
44	6.610		0.0000	0.000	47.1	0.022	BB	0.022	36.10	0.021
45	6.786		0.0000	0.000	134.5	0.064	BB	0.039	57.30	0.034
46	6.988		0.0000	0.000	266.7	0.127	BB	0.027	164.12	0.096
47	7.161 DCB		41.2915	49.120	32996.9	15.690	BB	0.042	13100.52	7.701

Total Area = 210305.0, Total Amount = 84.063, Total Height = 170120.8, Sample Units = PPB

ASX22 8764

Sample Name: **IC48X0924B AAIC48XAA CCAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/6/09 10:36:19pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.38R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.607		0.0000	0.000	438.5	0.064	BV	0.049	148.90	0.024
2	2.701		0.0000	0.000	346.2	0.050	VV	0.031	188.04	0.030
3	2.746		0.0000	0.000	147.3	0.021	VB	0.024	102.11	0.016
4	2.808	TCX	41.7434	50.239	173702.2	25.211	BB	0.015	190992.90	30.418
5	2.981		0.0000	0.000	178.9	0.026	BB	0.014	212.99	0.034
6	3.030		0.0000	0.000	2532.4	0.368	BB	0.014	2928.90	0.466
7	3.098		0.0000	0.000	1149.2	0.167	BB	0.013	1519.23	0.242
8	3.148		0.0000	0.000	484.3	0.070	BB	0.014	595.30	0.095
9	3.212		0.0000	0.000	589.6	0.086	BB	0.014	720.54	0.115
10	3.308		0.0000	0.000	13079.6	1.898	BV	0.018	12295.60	1.958
11	3.361		0.0000	0.000	922.1	0.134	VB	0.017	917.78	0.146
12	3.404		0.0000	0.000	161.1	0.023	BV	0.014	186.67	0.030
13	3.431		0.0000	0.000	7583.3	1.101	VV	0.015	8710.37	1.387
14	3.495		0.0000	0.000	366.4	0.053	VV	0.011	547.37	0.087
15	3.513		0.0000	0.000	2505.8	0.364	VB	0.020	2111.52	0.336
16	3.581		0.0000	0.000	41600.2	6.038	BV	0.034	20637.12	3.287
17	3.702		0.0000	0.000	10131.2	1.470	VV	0.014	11711.72	1.865
18	3.736		0.0000	0.000	22040.3	3.199	VV	0.014	25397.53	4.045
19	3.761		0.0000	0.000	29383.0	4.265	VB	0.020	24550.85	3.910
20	3.871		0.0000	0.000	28269.6	4.103	BB	0.018	26861.23	4.278
21	3.917		0.0000	0.000	4774.3	0.693	BB	0.012	6856.67	1.092
22	3.950		0.0000	0.000	2894.4	0.420	BV	0.011	4290.08	0.683
23	3.966		0.0000	0.000	15468.4	2.245	VV	0.015	17687.82	2.817
24	4.010		0.0000	0.000	5933.2	0.861	VB	0.015	6518.07	1.038
25	4.055		0.0000	0.000	1298.4	0.188	BV	0.015	1469.16	0.234
26	4.079		0.0000	0.000	10315.4	1.497	VV	0.015	11838.38	1.885
27	4.104		0.0000	0.000	27877.4	4.046	VV	0.015	31116.95	4.956
28	4.129		0.0000	0.000	30549.8	4.434	VV	0.025	20091.42	3.200
29	4.206		0.0000	0.000	9717.3	1.410	VV	0.016	9941.80	1.583
30	4.243		0.0000	0.000	15322.4	2.224	VV	0.018	14581.16	2.322
31	4.272		0.0000	0.000	19054.4	2.766	VV	0.030	10610.07	1.690
32	4.321		0.0000	0.000	3025.2	0.439	VV	0.022	2243.62	0.357
33	4.358		0.0000	0.000	6736.2	0.978	VV	0.021	5412.64	0.862
34	4.417		0.0000	0.000	8482.2	1.231	VV	0.017	8182.32	1.303
35	4.452		0.0000	0.000	5333.8	0.774	VV	0.017	5356.86	0.853
36	4.480		0.0000	0.000	14747.2	2.140	VB	0.019	12776.14	2.035
37	4.578		0.0000	0.000	5564.3	0.808	BV	0.020	4680.47	0.745
38	4.610		0.0000	0.000	1170.6	0.170	VV	0.018	1104.54	0.176
39	4.645		0.0000	0.000	9121.2	1.324	VV	0.024	6442.14	1.026
40	4.714		0.0000	0.000	1598.5	0.232	VV	0.020	1363.43	0.217
41	4.740		0.0000	0.000	1530.1	0.222	VB	0.028	911.43	0.145
42	4.821		0.0000	0.000	1795.2	0.261	BV	0.023	1292.32	0.206
43	4.872		0.0000	0.000	9510.3	1.380	VV	0.027	5830.60	0.929
44	4.954		0.0000	0.000	6272.4	0.910	VB	0.045	2312.71	0.368
45	5.146		0.0000	0.000	280.3	0.041	BV	0.026	179.22	0.029
46	5.180		0.0000	0.000	388.3	0.056	VV	0.017	371.86	0.059
47	5.203		0.0000	0.000	1192.6	0.173	VV	0.026	770.12	0.123
48	5.286		0.0000	0.000	274.5	0.040	VV	0.025	186.47	0.030
49	5.347		0.0000	0.000	467.2	0.068	VV	0.019	404.89	0.064
50	5.368		0.0000	0.000	1407.2	0.204	VB	0.033	703.50	0.112
51	5.449		0.0000	0.000	301.3	0.044	BB	0.041	121.62	0.019
52	5.551		0.0000	0.000	110.0	0.016	BB	0.022	82.66	0.013
53	5.581		0.0000	0.000	161.5	0.023	BB	0.017	156.95	0.025
54	5.626		0.0000	0.000	533.2	0.077	BB	0.024	366.36	0.058
55	5.782		0.0000	0.000	212.3	0.031	BB	0.031	112.52	0.018
56	5.903		0.0000	0.000	273.3	0.040	BB	0.044	104.27	0.017
57	5.999		0.0000	0.000	114.3	0.017	BB	0.017	114.94	0.018
58	6.200		0.0000	0.000	1499.1	0.218	BB	0.022	1119.69	0.178
59	6.360	DCB	41.3464	49.761	127703.2	18.534	BB	0.022	98681.95	15.716
60	6.809		0.0000	0.000	112.0	0.016	BB	0.038	49.15	0.008
61	6.984		0.0000	0.000	58.4	0.008	BB	0.022	43.67	0.007
62	7.170		0.0000	0.000	212.6	0.031	BB	0.050	71.48	0.011

ASX22 8765

Total Area = 689005.2, Total Amount = 83.09, Total Height = 627888.6, Sample Units = PPB

ASX22 8766

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.38R

IC48X0924B AAIC48XAA CCAL 0930999999

C:\CPWIN\DATA1\1C6310.38R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 10:28:49 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.38R

IC48X0924B AAIC48XAA CCAL 0930999999

C:\CPWIN\DATA1\1C6310B.38R



ASX22 0767

Instrument ID: CP06--S9147B Injected On: 11/6/2009 10:28:49 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.312	83665	42.77121	TCX	2.808	190993	41.74337	TCX
7.161	13101	41.29153	DCB	6.36	98682	41.34644	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.38A
Area File: C:\CPWIN\DATA1\1C6310B.38A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 10:36:26 PM
File Reported On: 11/6/2009 at 10:36:34 PM

REPLOTS

ASX22 0768

IC48X0924B AAIC48XAA CCAL 0930999999
C:\CPWIN\DATA1\1C6310.38R



IC48X0924B AAIC48XAA CCAL 0930999999
C:\CPWIN\DATA1\1C6310B.38R



ASX22 8769

Sample Name: **IC54X0924B AAIC54XAA CCAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/6/09 10:48:18pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.39R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.135		0.0000	0.000	79.0	0.033	BB	0.024	56.03	0.030
2	3.243		0.0000	0.000	314.9	0.132	BB	0.070	74.73	0.040
3	3.310	TCX	43.4174	51.189	80577.7	33.855	BB	0.016	84928.67	45.124
4	3.512		0.0000	0.000	1048.0	0.440	BB	0.014	1226.75	0.652
5	3.563		0.0000	0.000	137.6	0.058	BB	0.015	153.89	0.082
6	3.648		0.0000	0.000	233.2	0.098	BB	0.017	225.09	0.120
7	3.787		0.0000	0.000	324.7	0.136	BB	0.022	250.87	0.133
8	3.913		0.0000	0.000	195.2	0.082	BB	0.015	215.08	0.114
9	3.979		0.0000	0.000	171.8	0.072	BB	0.018	162.44	0.086
10	4.032		0.0000	0.000	62.5	0.026	BB	0.019	55.43	0.029
11	4.099		0.0000	0.000	782.5	0.329	BB	0.036	363.58	0.193
12	4.156		0.0000	0.000	123.2	0.052	BB	0.016	128.34	0.068
13	4.219		0.0000	0.000	224.7	0.094	BB	0.020	188.31	0.100
14	4.291		0.0000	0.000	5733.0	2.409	BB	0.016	5910.04	3.140
15	4.321		0.0000	0.000	975.7	0.410	BB	0.014	1183.50	0.629
16	4.431		0.0000	0.000	4037.2	1.696	BB	0.020	3358.13	1.784
17	4.493		0.0000	0.000	174.1	0.073	BB	0.013	216.62	0.115
18	4.531		0.0000	0.000	1224.9	0.515	BB	0.017	1196.42	0.636
19	4.576		0.0000	0.000	276.5	0.116	BB	0.019	239.96	0.127
20	4.666		0.0000	0.000	185.2	0.078	BB	0.016	197.68	0.105
21	4.707		0.0000	0.000	14529.8	6.105	BV	0.028	8771.16	4.660
22	4.757		0.0000	0.000	4935.3	2.074	VV	0.029	2792.99	1.484
23	4.828		0.0000	0.000	2711.9	1.139	VV	0.022	2036.28	1.082
24	4.874		0.0000	0.000	14515.9	6.099	VV	0.023	10321.81	5.484
25	4.913		0.0000	0.000	7970.6	3.349	VV	0.032	4173.28	2.217
26	4.983		0.0000	0.000	1118.2	0.470	VV	0.022	831.91	0.442
27	5.032		0.0000	0.000	3200.1	1.345	VB	0.021	2557.11	1.359
28	5.087		0.0000	0.000	7434.1	3.123	BV	0.024	5217.75	2.772
29	5.120		0.0000	0.000	2160.9	0.908	VV	0.019	1927.78	1.024
30	5.154		0.0000	0.000	11902.7	5.001	VV	0.022	8869.57	4.713
31	5.200		0.0000	0.000	1534.3	0.645	VB	0.024	1058.43	0.562
32	5.243		0.0000	0.000	915.6	0.385	BB	0.015	1017.24	0.540
33	5.282		0.0000	0.000	4154.1	1.745	BB	0.021	3242.43	1.723
34	5.329		0.0000	0.000	256.0	0.108	BB	0.015	281.65	0.150
35	5.375		0.0000	0.000	5607.0	2.356	BB	0.022	4267.03	2.267
36	5.429		0.0000	0.000	451.0	0.189	BB	0.015	490.46	0.261
37	5.484		0.0000	0.000	1689.1	0.710	BB	0.016	1771.19	0.941
38	5.504		0.0000	0.000	932.4	0.392	BB	0.013	1163.05	0.618
39	5.578		0.0000	0.000	4013.7	1.686	BB	0.021	3227.77	1.715
40	5.619		0.0000	0.000	198.8	0.084	BB	0.011	299.09	0.159
41	5.646		0.0000	0.000	380.9	0.160	BB	0.020	318.21	0.169
42	5.693		0.0000	0.000	11255.8	4.729	BB	0.032	5861.16	3.114
43	5.903		0.0000	0.000	3106.3	1.305	BB	0.030	1718.96	0.913
44	5.975		0.0000	0.000	123.8	0.052	BB	0.016	130.99	0.070
45	6.096		0.0000	0.000	1033.2	0.434	BB	0.022	774.69	0.412
46	6.150		0.0000	0.000	891.5	0.375	BB	0.021	715.75	0.380
47	6.357		0.0000	0.000	1681.4	0.706	BB	0.038	743.60	0.395
48	6.987		0.0000	0.000	364.3	0.153	BB	0.038	161.74	0.086
49	7.159	DCB	41.4005	48.811	32057.5	13.469	BB	0.041	13135.10	6.979

Total Area = 238007.6, Total Amount = 84.818, Total Height = 188209.7, Sample Units = PPB

ASX22 6778

Sample Name: **IC54X0924B AAIC54XAA CCAL 0930999999**
Acquired from CP06–S9147B via port 2 on 11/6/09 10:48:18pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.39R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.612		0.0000	0.000	872.8	0.105	BV	0.064	226.11	0.028
2	2.699		0.0000	0.000	1066.8	0.129	VB	0.020	903.30	0.112
3	2.807	TCX	43.2911	51.013	171450.4	20.662	BB	0.014	198074.20	24.523
4	2.976		0.0000	0.000	58.9	0.007	BB	0.012	78.63	0.010
5	3.028		0.0000	0.000	2399.3	0.289	BB	0.013	2979.07	0.369
6	3.097		0.0000	0.000	282.1	0.034	BB	0.012	381.52	0.047
7	3.147		0.0000	0.000	467.7	0.056	BB	0.013	597.78	0.074
8	3.307		0.0000	0.000	907.0	0.109	BV	0.019	777.35	0.096
9	3.400		0.0000	0.000	75.2	0.009	VV	0.014	89.92	0.011
10	3.431		0.0000	0.000	627.3	0.076	VV	0.016	637.83	0.079
11	3.494		0.0000	0.000	626.0	0.075	VV	0.017	601.22	0.074
12	3.529		0.0000	0.000	450.5	0.054	VV	0.014	525.06	0.065
13	3.578		0.0000	0.000	1784.3	0.215	VV	0.021	1403.62	0.174
14	3.610		0.0000	0.000	1191.2	0.144	VB	0.025	787.27	0.097
15	3.700		0.0000	0.000	535.0	0.064	BB	0.013	689.04	0.085
16	3.734		0.0000	0.000	20594.9	2.482	BV	0.014	25065.05	3.103
17	3.759		0.0000	0.000	8469.2	1.021	VB	0.018	8058.53	0.998
18	3.869		0.0000	0.000	12922.0	1.557	BB	0.015	14576.15	1.805
19	3.917		0.0000	0.000	875.3	0.105	BB	0.014	1043.96	0.129
20	3.964		0.0000	0.000	4522.8	0.545	BV	0.015	5040.01	0.624
21	4.009		0.0000	0.000	988.6	0.119	VB	0.014	1146.57	0.142
22	4.050		0.0000	0.000	830.2	0.100	BV	0.014	960.22	0.119
23	4.079		0.0000	0.000	3431.7	0.414	VV	0.014	4193.80	0.519
24	4.102		0.0000	0.000	50365.9	6.070	VV	0.017	49171.69	6.088
25	4.150		0.0000	0.000	7295.8	0.879	VV	0.015	8032.06	0.994
26	4.168		0.0000	0.000	10192.8	1.228	VV	0.016	10536.47	1.304
27	4.205		0.0000	0.000	38062.5	4.587	VV	0.016	40658.39	5.034
28	4.239		0.0000	0.000	19784.0	2.384	VV	0.017	19159.77	2.372
29	4.259		0.0000	0.000	19841.0	2.391	VV	0.019	17346.09	2.148
30	4.321		0.0000	0.000	3716.0	0.448	VV	0.018	3446.51	0.427
31	4.357		0.0000	0.000	13938.3	1.680	VB	0.016	14117.73	1.748
32	4.415		0.0000	0.000	25055.1	3.019	BV	0.016	26083.37	3.229
33	4.451		0.0000	0.000	9134.5	1.101	VV	0.018	8463.16	1.048
34	4.478		0.0000	0.000	48396.7	5.832	VB	0.017	46890.07	5.805
35	4.519		0.0000	0.000	3042.2	0.367	BB	0.014	3558.12	0.441
36	4.574		0.0000	0.000	26430.6	3.185	BV	0.018	24226.35	2.999
37	4.609		0.0000	0.000	3586.1	0.432	VV	0.018	3334.32	0.413
38	4.643		0.0000	0.000	28006.0	3.375	VV	0.018	26021.04	3.222
39	4.668		0.0000	0.000	10407.8	1.254	VV	0.018	9590.75	1.187
40	4.713		0.0000	0.000	29010.9	3.496	VB	0.022	21903.12	2.712
41	4.818		0.0000	0.000	22449.5	2.705	BV	0.020	19032.23	2.356
42	4.871		0.0000	0.000	19369.2	2.334	VV	0.019	16947.40	2.098
43	4.911		0.0000	0.000	5721.6	0.690	VB	0.017	5613.89	0.695
44	4.952		0.0000	0.000	41713.6	5.027	BB	0.022	32162.34	3.982
45	5.000		0.0000	0.000	998.8	0.120	BB	0.010	1730.45	0.214
46	5.085		0.0000	0.000	368.6	0.044	BB	0.017	360.92	0.045
47	5.140		0.0000	0.000	1554.8	0.187	BV	0.018	1403.84	0.174
48	5.179		0.0000	0.000	2367.7	0.285	VV	0.016	2456.42	0.304
49	5.202		0.0000	0.000	12708.3	1.532	VV	0.020	10838.55	1.342
50	5.251		0.0000	0.000	1785.5	0.215	VV	0.018	1672.81	0.207
51	5.286		0.0000	0.000	989.2	0.119	VV	0.015	1101.51	0.136
52	5.313		0.0000	0.000	275.4	0.033	VB	0.013	354.65	0.044
53	5.345		0.0000	0.000	2372.5	0.286	BB	0.013	2962.64	0.367
54	5.366		0.0000	0.000	3049.8	0.368	BB	0.014	3524.07	0.436
55	5.580		0.0000	0.000	165.7	0.020	BB	0.014	196.85	0.024
56	5.623		0.0000	0.000	6401.0	0.771	BB	0.022	4769.69	0.591
57	5.820		0.0000	0.000	380.8	0.046	BB	0.033	193.37	0.024
58	5.898		0.0000	0.000	236.6	0.029	BB	0.025	158.55	0.020
59	5.999		0.0000	0.000	217.4	0.026	BB	0.016	231.68	0.029
60	6.198		0.0000	0.000	1472.1	0.177	BB	0.022	1128.67	0.140
61	6.358	DCB	41.5719	48.987	123059.9	14.830	BB	0.021	99220.03	12.284
62	6.700		0.0000	0.000	98.0	0.012	BB	0.026	63.88	0.008

ASX22 0771

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.986		0.0000	0.000	144.8	0.017	BB	0.034	71.26	0.009
64	7.044		0.0000	0.000	39.7	0.005	BB	0.019	34.10	0.004
65	7.166		0.0000	0.000	105.2	0.013	BB	0.025	69.66	0.009
66	7.314		0.0000	0.000	52.4	0.006	BB	0.019	46.44	0.006

Total Area = 829793.3, Total Amount = 84.863, Total Height = 807721.1, Sample Units = PPB

ASX22 0772

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.39R
IC54X0924B AAIC54XAA CCAL 0930999999
C:\CPWIN\DATA1\1C6310.39R



Instrument ID: CP06--S9147A Injected On: 11/6/2009 10:40:48 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.39R
IC54X0924B AAIC54XAA CCAL 0930999999
C:\CPWIN\DATA1\1C6310B.39R



Instrument ID: CP06--S9147B Injected On: 11/6/2009 10:40:48 PM Column ID: RTX CLPII

ASX22 8773

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.31	84929	43.41737	TCX	2.807	198074	43.29105	TCX
7.159	13135	41.40052	DCB	6.358	99220	41.57189	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.39A
Area File: C:\CPWIN\DATA1\1C6310B.39A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 10:48:24 PM
File Reported On: 11/6/2009 at 10:48:33 PM

REPLOTS

ASX22 8774

IC54X0924B AAIC54XAA CCAL 0930999999
C:\CPWIN\DATA1\1C6310.39R



IC54X0924B AAIC54XAA CCAL 0930999999
C:\CPWIN\DATA1\1C6310B.39R



ASX22 0775

Sample Name: **AR1640924A FSAR164FS CCAL 0930999999**
Acquired from CP06--S9147A via port 1 on 11/7/09 09:35:51am by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.93R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.029		0.0000	0.000	38.8	0.006	BB	0.013	49.24	0.010
2	3.102		0.0000	0.000	124.6	0.020	BB	0.038	54.44	0.011
3	3.226		0.0000	0.000	79.5	0.012	BB	0.026	50.92	0.010
4	3.246		0.0000	0.000	146.8	0.023	BB	0.026	95.62	0.019
5	3.309	TCX	64.3244	49.334	114760.5	18.029	BB	0.015	125824.80	25.290
6	3.456		0.0000	0.000	3080.7	0.484	BB	0.017	3012.56	0.606
7	3.513		0.0000	0.000	4499.0	0.707	BB	0.014	5176.64	1.040
8	3.563		0.0000	0.000	14387.1	2.260	BB	0.015	15766.28	3.169
9	3.659		0.0000	0.000	2739.9	0.430	BB	0.017	2684.24	0.540
10	3.786		0.0000	0.000	30376.5	4.772	BB	0.024	21266.05	4.274
11	3.844		0.0000	0.000	3209.5	0.504	BB	0.025	2123.30	0.427
12	3.912		0.0000	0.000	18295.1	2.874	BB	0.016	18930.77	3.805
13	3.991		0.0000	0.000	403.7	0.063	BB	0.016	415.33	0.083
14	4.031		0.0000	0.000	6835.0	1.074	BB	0.025	4618.69	0.928
15	4.110		0.0000	0.000	43081.3	6.768	BB	0.028	25940.79	5.214
16	4.153		0.0000	0.000	13965.7	2.194	BB	0.017	13839.45	2.782
17	4.219		0.0000	0.000	12911.3	2.028	BB	0.017	12751.22	2.563
18	4.264		0.0000	0.000	534.0	0.084	BB	0.012	747.70	0.150
19	4.290		0.0000	0.000	8448.5	1.327	BB	0.015	9426.38	1.895
20	4.321		0.0000	0.000	12978.0	2.039	BB	0.027	7872.76	1.582
21	4.430		0.0000	0.000	8941.5	1.405	BB	0.016	9333.30	1.876
22	4.457		0.0000	0.000	1733.4	0.272	BB	0.012	2376.31	0.478
23	4.492		0.0000	0.000	2425.5	0.381	BB	0.015	2746.93	0.552
24	4.529		0.0000	0.000	9906.6	1.556	BB	0.020	8265.27	1.661
25	4.576		0.0000	0.000	2479.6	0.390	BB	0.016	2583.66	0.519
26	4.666		0.0000	0.000	269.8	0.042	BB	0.015	310.05	0.062
27	4.706		0.0000	0.000	14686.3	2.307	BV	0.027	9109.16	1.831
28	4.754		0.0000	0.000	4697.6	0.738	VV	0.029	2682.10	0.539
29	4.825		0.0000	0.000	1484.7	0.233	VV	0.024	1030.29	0.207
30	4.874		0.0000	0.000	12149.9	1.909	VV	0.030	6660.21	1.339
31	5.029		0.0000	0.000	232.9	0.037	VB	0.019	201.20	0.040
32	5.105		0.0000	0.000	5351.5	0.841	BV	0.025	3586.97	0.721
33	5.153		0.0000	0.000	2916.3	0.458	VB	0.020	2475.59	0.498
34	5.201		0.0000	0.000	10879.2	1.709	BB	0.023	7848.44	1.577
35	5.239		0.0000	0.000	583.1	0.092	BB	0.010	947.16	0.190
36	5.282		0.0000	0.000	22760.8	3.576	BB	0.024	15583.02	3.132
37	5.371		0.0000	0.000	1473.3	0.231	BB	0.019	1295.07	0.260
38	5.429		0.0000	0.000	2452.0	0.385	BB	0.020	2047.40	0.412
39	5.483		0.0000	0.000	23675.3	3.719	BB	0.028	14175.17	2.849
40	5.537		0.0000	0.000	3387.6	0.532	BB	0.016	3587.83	0.721
41	5.576		0.0000	0.000	9777.6	1.536	BB	0.026	6355.77	1.277
42	5.654		0.0000	0.000	615.4	0.097	BB	0.016	656.08	0.132
43	5.691		0.0000	0.000	21041.3	3.306	BB	0.024	14699.38	2.954
44	5.753		0.0000	0.000	12262.8	1.927	BB	0.019	10903.19	2.191
45	5.796		0.0000	0.000	6576.6	1.033	BB	0.019	5793.36	1.164
46	5.918		0.0000	0.000	19442.5	3.054	BB	0.027	12028.97	2.418
47	5.976		0.0000	0.000	5126.2	0.805	BB	0.018	4667.49	0.938
48	6.013		0.0000	0.000	1699.9	0.267	BB	0.016	1770.35	0.356
49	6.095		0.0000	0.000	2763.1	0.434	BB	0.020	2304.93	0.463
50	6.151		0.0000	0.000	45518.1	7.151	BB	0.031	24292.28	4.883
51	6.355		0.0000	0.000	14290.5	2.245	BB	0.022	11046.47	2.220
52	6.403		0.0000	0.000	7011.2	1.101	BB	0.028	4182.40	0.841
53	6.539		0.0000	0.000	321.1	0.050	BB	0.019	281.87	0.057
54	6.625		0.0000	0.000	4798.7	0.754	BB	0.029	2760.87	0.555
55	6.774		0.0000	0.000	11576.8	1.819	BB	0.034	5688.79	1.143
56	6.987		0.0000	0.000	3127.9	0.491	BB	0.032	1641.81	0.330
57	7.160	DCB	66.0602	50.666	47190.2	7.414	BB	0.038	20958.86	4.213

Total Area = 636521.9, Total Amount = 130.385, Total Height = 497525.3, Sample Units = PPB

ASX22 8776

Sample Name: **AR1640924A FSAR164FS CCAL 0930999999**
Acquired from CP06--S9147B via port 2 on 11/7/09 09:35:51am by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.93R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.607		0.0000	0.000	413.5	0.017	BV	0.052	133.53	0.006
2	2.700		0.0000	0.000	368.1	0.016	VB	0.024	257.59	0.011
3	2.761		0.0000	0.000	52.9	0.002	BB	0.010	89.89	0.004
4	2.807	TCX	65.2752	52.805	232194.7	9.785	BB	0.013	298660.60	12.931
5	2.852		0.0000	0.000	2308.8	0.097	BB	0.011	3499.99	0.152
6	2.981		0.0000	0.000	7010.7	0.295	BB	0.013	9034.75	0.391
7	3.029		0.0000	0.000	1325.0	0.056	BB	0.013	1729.72	0.075
8	3.043		0.0000	0.000	4508.4	0.190	BB	0.010	7727.17	0.335
9	3.096		0.0000	0.000	33002.0	1.391	BB	0.013	43146.81	1.868
10	3.146		0.0000	0.000	614.5	0.026	BB	0.011	892.48	0.039
11	3.210		0.0000	0.000	5571.2	0.235	BB	0.012	7523.16	0.326
12	3.307		0.0000	0.000	73622.2	3.102	BV	0.019	63160.52	2.735
13	3.359		0.0000	0.000	7044.6	0.297	VV	0.014	8322.36	0.360
14	3.380		0.0000	0.000	2809.9	0.118	VV	0.012	3862.61	0.167
15	3.403		0.0000	0.000	3333.9	0.140	VV	0.013	4421.56	0.191
16	3.430		0.0000	0.000	48156.5	2.029	VV	0.014	57185.39	2.476
17	3.466		0.0000	0.000	1675.0	0.071	VV	0.013	2168.85	0.094
18	3.512		0.0000	0.000	21491.6	0.906	VB	0.021	17247.77	0.747
19	3.586		0.0000	0.000	132208.8	5.571	BV	0.021	105358.10	4.562
20	3.623		0.0000	0.000	68451.3	2.885	VV	0.020	56373.80	2.441
21	3.699		0.0000	0.000	46036.3	1.940	VV	0.014	53209.48	2.304
22	3.734		0.0000	0.000	39811.2	1.678	VV	0.014	45879.55	1.986
23	3.759		0.0000	0.000	61881.8	2.608	VB	0.021	50045.87	2.167
24	3.869		0.0000	0.000	42350.0	1.785	BV	0.014	49109.40	2.126
25	3.893		0.0000	0.000	11937.5	0.503	VB	0.011	17653.53	0.764
26	3.915		0.0000	0.000	9047.4	0.381	BB	0.011	13743.93	0.595
27	3.947		0.0000	0.000	9969.6	0.420	BV	0.013	12481.40	0.540
28	3.963		0.0000	0.000	31630.8	1.333	VV	0.014	37848.39	1.639
29	4.008		0.0000	0.000	12184.3	0.513	VB	0.015	13535.78	0.586
30	4.053		0.0000	0.000	1275.1	0.054	BV	0.014	1509.42	0.065
31	4.078		0.0000	0.000	7435.1	0.313	VV	0.014	9109.16	0.394
32	4.102		0.0000	0.000	41453.9	1.747	VV	0.015	46708.69	2.022
33	4.127		0.0000	0.000	10054.6	0.424	VV	0.016	10674.21	0.462
34	4.148		0.0000	0.000	3372.3	0.142	VV	0.014	3918.38	0.170
35	4.167		0.0000	0.000	4861.3	0.205	VV	0.016	5004.66	0.217
36	4.204		0.0000	0.000	30157.3	1.271	VV	0.016	31721.62	1.373
37	4.241		0.0000	0.000	2591.2	0.109	VV	0.015	2887.68	0.125
38	4.258		0.0000	0.000	6376.2	0.269	VV	0.026	4085.83	0.177
39	4.356		0.0000	0.000	1001.2	0.042	VB	0.017	999.37	0.043
40	4.414		0.0000	0.000	4729.5	0.199	BB	0.015	5242.97	0.227
41	4.461		0.0000	0.000	12888.2	0.543	BV	0.015	14469.96	0.626
42	4.491		0.0000	0.000	42258.2	1.781	VB	0.017	42437.09	1.837
43	4.519		0.0000	0.000	14124.7	0.595	BB	0.016	14643.79	0.634
44	4.573		0.0000	0.000	85224.0	3.591	BV	0.017	84738.17	3.669
45	4.642		0.0000	0.000	2580.8	0.109	VV	0.015	2862.52	0.124
46	4.669		0.0000	0.000	16542.9	0.697	VV	0.016	17151.55	0.743
47	4.712		0.0000	0.000	102165.5	4.305	VB	0.019	89599.30	3.879
48	4.820		0.0000	0.000	96076.1	4.049	BV	0.019	83181.34	3.601
49	4.876		0.0000	0.000	9814.1	0.414	VB	0.016	9962.93	0.431
50	4.913		0.0000	0.000	18740.1	0.790	BB	0.015	20288.80	0.878
51	4.950		0.0000	0.000	93137.2	3.925	BV	0.020	76282.55	3.303
52	4.983		0.0000	0.000	71328.9	3.006	VB	0.017	70997.59	3.074
53	5.023		0.0000	0.000	32765.6	1.381	BB	0.016	34799.40	1.507
54	5.124		0.0000	0.000	15303.4	0.645	BB	0.018	14348.00	0.621
55	5.177		0.0000	0.000	84705.9	3.569	BV	0.020	72078.13	3.121
56	5.250		0.0000	0.000	39503.8	1.665	VV	0.018	35670.27	1.544
57	5.286		0.0000	0.000	17376.1	0.732	VV	0.017	16719.62	0.724
58	5.311		0.0000	0.000	11784.5	0.497	VV	0.017	11668.00	0.505
59	5.366		0.0000	0.000	171174.5	7.213	VB	0.020	143786.70	6.225
60	5.442		0.0000	0.000	5432.9	0.229	BB	0.015	6154.61	0.266
61	5.556		0.0000	0.000	687.8	0.029	BB	0.019	600.67	0.026
62	5.580		0.0000	0.000	34835.1	1.468	BB	0.016	35658.69	1.544

ASX22 8777

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.625		0.0000	0.000	122594.0	5.166	BB	0.022	94025.44	4.071
64	5.779		0.0000	0.000	1831.8	0.077	BB	0.016	1964.21	0.085
65	5.819		0.0000	0.000	1208.8	0.051	BB	0.015	1356.24	0.059
66	5.847		0.0000	0.000	1118.4	0.047	BB	0.014	1365.39	0.059
67	5.895		0.0000	0.000	20344.4	0.857	BB	0.019	17922.24	0.776
68	5.998		0.0000	0.000	50104.0	2.111	BB	0.022	37906.93	1.641
69	6.197		0.0000	0.000	12801.7	0.539	BB	0.019	11182.30	0.484
70	6.357	DCB	58.3397	47.195	159290.8	6.712	BB	0.019	139240.00	6.029
71	6.526		0.0000	0.000	185.2	0.008	BB	0.021	146.98	0.006
72	6.797		0.0000	0.000	115.7	0.005	BB	0.020	98.64	0.004
73	6.986		0.0000	0.000	272.0	0.011	BB	0.081	55.87	0.002
74	7.173		0.0000	0.000	258.6	0.011	BB	0.042	102.09	0.004
75	7.440		0.0000	0.000	161.3	0.007	BB	0.048	55.75	0.002

Total Area = 2373057.0, Total Amount = 123.615, Total Height = 2309688.0, Sample Units = PPB

ASX22 8778

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.93R
AR1640924A FSAR164FS CCAL 0930999999
C:\CPWIN\DATA1\1C6310.93R



Instrument ID: CP06–S9147A Injected On: 11/7/2009 9:28:21 AM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.93R
AR1640924A FSAR164FS CCAL 0930999999
C:\CPWIN\DATA1\1C6310B.93R



ASX22 8779

Instrument ID: CP06–S9147B Injected On: 11/7/2009 9:28:21 AM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Valid Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.309	125825	64.32436	TCX	2.807	298661	65.27521	TCX
7.16	20959	66.06022	DCB	6.357	139240	58.33972	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.93A
Area File: C:\CPWIN\DATA1\1C6310B.93A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/7/2009 9:48:34 AM
File Reported On: 11/7/2009 at 9:48:43 AM

REPLOTS

ASX22 0788

AR1640924A FSAR164FS CCAL 0930999999
C:\CPWIN\DATA1\1C6310.93R



AR1640924A FSAR164FS CCAL 0930999999
C:\CPWIN\DATA1\1C6310B.93R



ASX22 2781

Sample Name: **IBLKX0924B FJPIBLKFJ PIBLK0930999999**
Acquired from CP06--S9147A via port 1 on 11/7/09 09:47:52am by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\1C6310.94R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.146		0.0000	0.000	257.2	0.410	BB	0.048	89.45	0.172
2	3.247		0.0000	0.000	457.6	0.729	BB	0.062	122.42	0.235
3	3.311	TCX	221.2513	49.982	41852.1	66.648	BB	0.016	43278.95	83.088
4	3.513		0.0000	0.000	604.2	0.962	BB	0.016	635.49	1.220
5	3.649		0.0000	0.000	183.1	0.292	BB	0.023	133.83	0.257
6	3.791		0.0000	0.000	54.8	0.087	BB	0.025	36.14	0.069
7	3.843		0.0000	0.000	145.6	0.232	BB	0.053	45.53	0.087
8	3.982		0.0000	0.000	98.8	0.157	BB	0.018	89.05	0.171
9	4.071		0.0000	0.000	354.1	0.564	BB	0.064	92.06	0.177
10	4.774		0.0000	0.000	258.8	0.412	BV	0.053	80.87	0.155
11	4.876		0.0000	0.000	167.3	0.266	VB	0.074	37.66	0.072
12	5.142		0.0000	0.000	136.3	0.217	BB	0.022	101.55	0.195
13	5.443		0.0000	0.000	100.8	0.160	BB	0.031	53.54	0.103
14	6.014		0.0000	0.000	473.1	0.753	BB	0.067	117.31	0.225
15	6.354		0.0000	0.000	288.9	0.460	BB	0.066	72.83	0.140
16	6.989		0.0000	0.000	178.6	0.284	BB	0.039	76.47	0.147
17	7.160	DCB	221.4151	50.018	17184.8	27.366	BB	0.041	7024.81	13.486

Total Area = 62796.0, Total Amount = 442.666, Total Height = 52087.97, Sample Units = PPB

ASX22 8782

Sample Name: **IBLKX0924B FJPIBLKFJ PIBLK0930999999**
Acquired from CP06--S9147B via port 2 on 11/7/09 09:47:52am by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\1C6310B.94R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.613		0.0000	0.000	390.1	0.248	BV	0.052	125.30	0.082
2	2.701		0.0000	0.000	1024.7	0.653	VB	0.019	886.57	0.583
3	2.808	TCX	213.7868	52.401	88472.8	56.345	BB	0.015	97816.16	64.297
4	2.929		0.0000	0.000	116.7	0.074	BB	0.014	134.56	0.088
5	3.030		0.0000	0.000	1322.2	0.842	BB	0.015	1462.61	0.961
6	3.148		0.0000	0.000	266.2	0.170	BV	0.015	300.82	0.198
7	3.309		0.0000	0.000	279.6	0.178	VB	0.035	132.02	0.087
8	3.433		0.0000	0.000	118.3	0.075	BV	0.019	105.52	0.069
9	3.495		0.0000	0.000	266.0	0.169	VV	0.015	297.31	0.195
10	3.531		0.0000	0.000	488.8	0.311	VV	0.016	510.45	0.336
11	3.587		0.0000	0.000	254.2	0.162	VV	0.025	167.51	0.110
12	3.621		0.0000	0.000	207.6	0.132	VV	0.032	107.67	0.071
13	3.736		0.0000	0.000	108.1	0.069	VV	0.018	98.31	0.065
14	3.762		0.0000	0.000	189.7	0.121	VB	0.023	138.03	0.091
15	3.869		0.0000	0.000	69.1	0.044	BB	0.016	69.87	0.046
16	3.931		0.0000	0.000	97.7	0.062	BB	0.020	82.69	0.054
17	4.104		0.0000	0.000	189.2	0.120	BV	0.035	88.95	0.058
18	4.209		0.0000	0.000	526.5	0.335	VV	0.081	108.08	0.071
19	4.307		0.0000	0.000	1393.1	0.887	VB	0.063	369.91	0.243
20	4.460		0.0000	0.000	136.8	0.087	BB	0.041	56.24	0.037
21	4.575		0.0000	0.000	177.1	0.113	BB	0.025	119.87	0.079
22	4.693		0.0000	0.000	943.6	0.601	BV	0.031	510.63	0.336
23	4.835		0.0000	0.000	1048.4	0.668	VV	0.044	396.22	0.260
24	4.949		0.0000	0.000	146.5	0.093	VV	0.022	113.28	0.074
25	4.984		0.0000	0.000	112.1	0.071	VB	0.018	104.74	0.069
26	5.105		0.0000	0.000	111.5	0.071	BB	0.041	44.84	0.029
27	5.180		0.0000	0.000	163.0	0.104	BV	0.023	118.98	0.078
28	5.286		0.0000	0.000	139.8	0.089	VV	0.043	54.01	0.036
29	5.368		0.0000	0.000	307.1	0.196	VB	0.023	221.44	0.146
30	5.547		0.0000	0.000	151.1	0.096	BB	0.020	126.65	0.083
31	5.651		0.0000	0.000	208.1	0.133	BB	0.033	104.19	0.068
32	5.785		0.0000	0.000	232.4	0.148	BB	0.054	71.79	0.047
33	6.001		0.0000	0.000	62.7	0.040	BB	0.015	68.47	0.045
34	6.199		0.0000	0.000	534.1	0.340	BB	0.018	499.34	0.328
35	6.359	DCB	194.1918	47.599	56500.8	35.983	BB	0.020	46347.94	30.466
36	6.711		0.0000	0.000	35.5	0.023	BB	0.014	41.92	0.028
37	6.746		0.0000	0.000	55.4	0.035	BB	0.021	43.31	0.028
38	6.863		0.0000	0.000	132.9	0.085	BB	0.037	60.65	0.040
39	7.253		0.0000	0.000	40.2	0.026	BB	0.027	24.83	0.016

Total Area = 157019.4, Total Amount = 407.979, Total Height = 152131.7, Sample Units = PPB

ASX22 8783

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\1C6310.94R

IBLKX0924B FJPIBLKFJ PIBLK0930999999

——— C:\CPWIN\DATA1\1C6310.94R



Instrument ID: CP06--S9147A Injected On: 11/7/2009 9:40:22 AM Column ID: RTX CLP

File: C:\CPWIN\DATA1\1C6310B.94R

IBLKX0924B FJPIBLKFJ PIBLK0930999999

——— C:\CPWIN\DATA1\1C6310B.94R



Instrument ID: CP06--S9147B Injected On: 11/7/2009 9:40:22 AM Column ID: RTX CLPII

ASX22 8784

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Valid Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1000 Dilution Factor: 10
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	43279	.22125	TCX	2.808	97816	.21379	TCX
7.16	7025	.22142	DCB	6.359	46348	.19419	DCB

Files:
Area File: C:\CPWIN\DATA1\1C6310.94A
Area File: C:\CPWIN\DATA1\1C6310B.94A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/7/2009 9:50:54 AM
File Reported On: 11/7/2009 at 9:51:04 AM

REPLOTS

ASX22 8785

IBLKX0924B FJPIBLKFJ PIBLK0930999999
C:\CPWIN\DATA1\1C6310.94R



IBLKX0924B FJPIBLKFJ PIBLK0930999999
C:\CPWIN\DATA1\1C6310B.94R



ASX22 0786

Sample Name: **AR1630924A PMAR163PM CCAL 0931099999**
Acquired from CP06--S9147A via port 1 on 11/7/09 12:08:01pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\2C6310.11R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.216		0.0000	0.000	214.0	0.074	BB	0.048	73.84	0.033
2	3.311	TCX	38.5230	48.403	74964.6	26.030	BB	0.017	75354.84	33.742
3	3.456		0.0000	0.000	1229.1	0.427	BB	0.018	1151.85	0.516
4	3.514		0.0000	0.000	2141.5	0.744	BB	0.015	2352.60	1.053
5	3.563		0.0000	0.000	5818.6	2.020	BB	0.016	6205.01	2.778
6	3.661		0.0000	0.000	1209.4	0.420	BB	0.018	1104.43	0.495
7	3.787		0.0000	0.000	11983.6	4.161	BB	0.024	8377.07	3.751
8	3.845		0.0000	0.000	1202.2	0.417	BB	0.025	804.48	0.360
9	3.913		0.0000	0.000	7105.9	2.467	BB	0.016	7229.24	3.237
10	3.993		0.0000	0.000	186.0	0.065	BB	0.020	156.85	0.070
11	4.032		0.0000	0.000	2394.6	0.831	BB	0.024	1666.70	0.746
12	4.111		0.0000	0.000	16047.1	5.572	BB	0.028	9578.95	4.289
13	4.154		0.0000	0.000	5342.1	1.855	BB	0.017	5304.08	2.375
14	4.219		0.0000	0.000	4794.2	1.665	BB	0.017	4748.58	2.126
15	4.265		0.0000	0.000	184.5	0.064	BB	0.012	264.95	0.119
16	4.291		0.0000	0.000	3308.6	1.149	BB	0.015	3732.40	1.671
17	4.322		0.0000	0.000	5152.8	1.789	BB	0.028	3048.17	1.365
18	4.431		0.0000	0.000	3501.1	1.216	BB	0.016	3650.95	1.635
19	4.458		0.0000	0.000	735.7	0.255	BB	0.011	1087.79	0.487
20	4.493		0.0000	0.000	947.0	0.329	BB	0.014	1092.88	0.489
21	4.530		0.0000	0.000	3789.9	1.316	BB	0.020	3158.38	1.414
22	4.576		0.0000	0.000	1088.0	0.378	BB	0.018	1014.86	0.454
23	4.666		0.0000	0.000	143.4	0.050	BB	0.015	162.27	0.073
24	4.707		0.0000	0.000	5892.7	2.046	BV	0.027	3682.77	1.649
25	4.755		0.0000	0.000	2024.0	0.703	VV	0.031	1073.72	0.481
26	4.826		0.0000	0.000	671.6	0.233	VV	0.025	448.47	0.201
27	4.875		0.0000	0.000	5175.0	1.797	VV	0.031	2758.58	1.235
28	5.026		0.0000	0.000	158.6	0.055	VB	0.023	115.76	0.052
29	5.106		0.0000	0.000	2130.3	0.740	BV	0.024	1463.42	0.655
30	5.153		0.0000	0.000	1130.2	0.392	VB	0.020	932.97	0.418
31	5.202		0.0000	0.000	4179.3	1.451	BB	0.023	3002.04	1.344
32	5.239		0.0000	0.000	256.8	0.089	BB	0.011	381.42	0.171
33	5.282		0.0000	0.000	8808.5	3.059	BB	0.024	6049.92	2.709
34	5.372		0.0000	0.000	626.3	0.217	BB	0.019	540.95	0.242
35	5.430		0.0000	0.000	979.4	0.340	BB	0.020	824.19	0.369
36	5.484		0.0000	0.000	8928.5	3.100	BB	0.027	5464.37	2.447
37	5.537		0.0000	0.000	1421.9	0.494	BB	0.016	1486.16	0.665
38	5.577		0.0000	0.000	3671.9	1.275	BB	0.026	2393.36	1.072
39	5.656		0.0000	0.000	233.3	0.081	BB	0.015	254.42	0.114
40	5.692		0.0000	0.000	7854.5	2.727	BB	0.024	5500.13	2.463
41	5.754		0.0000	0.000	4617.6	1.603	BB	0.020	3930.98	1.760
42	5.797		0.0000	0.000	2400.5	0.834	BB	0.019	2103.35	0.942
43	5.919		0.0000	0.000	7378.0	2.562	BB	0.027	4568.60	2.046
44	5.976		0.0000	0.000	1909.4	0.663	BB	0.019	1716.35	0.769
45	6.015		0.0000	0.000	679.4	0.236	BB	0.016	687.93	0.308
46	6.096		0.0000	0.000	1006.6	0.350	BB	0.020	850.17	0.381
47	6.151		0.0000	0.000	17298.0	6.006	BB	0.032	8946.87	4.006
48	6.357		0.0000	0.000	5301.0	1.841	BB	0.021	4228.56	1.893
49	6.405		0.0000	0.000	2656.0	0.922	BB	0.028	1590.82	0.712
50	6.625		0.0000	0.000	1929.2	0.670	BB	0.031	1044.47	0.468
51	6.775		0.0000	0.000	4714.6	1.637	BB	0.035	2231.79	0.999
52	6.989		0.0000	0.000	1366.2	0.474	BB	0.032	707.52	0.317
53	7.159	DCB	41.0644	51.597	29108.1	10.107	BB	0.037	13028.44	5.834

Total Area = 287991.1, Total Amount = 79.587, Total Height = 223329.7, Sample Units = PPB

ASX22 8787

Sample Name: **AR1630924A PMAR163PM CCAL 0931099999**
Acquired from CP06--S9147B via port 2 on 11/7/09 12:08:01pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\2C6310B.11R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.584		0.0000	0.000	65.2	0.006	BV	0.016	67.09	0.007
2	2.610		0.0000	0.000	526.9	0.050	VV	0.056	157.75	0.016
3	2.701		0.0000	0.000	462.2	0.044	VB	0.026	294.86	0.030
4	2.807	TCX	37.9344	52.078	170546.0	16.091	BB	0.016	173565.30	17.732
5	2.982		0.0000	0.000	2834.0	0.267	BB	0.014	3305.77	0.338
6	3.044		0.0000	0.000	5189.3	0.490	BB	0.022	3942.51	0.403
7	3.097		0.0000	0.000	14174.1	1.337	BB	0.014	17041.32	1.741
8	3.147		0.0000	0.000	376.0	0.035	BB	0.012	518.00	0.053
9	3.211		0.0000	0.000	2362.6	0.223	BB	0.013	2954.31	0.302
10	3.308		0.0000	0.000	31855.2	3.006	BV	0.020	26266.93	2.684
11	3.360		0.0000	0.000	4068.4	0.384	VV	0.020	3364.14	0.344
12	3.404		0.0000	0.000	1189.6	0.112	VV	0.012	1616.26	0.165
13	3.431		0.0000	0.000	20306.0	1.916	VV	0.014	23632.61	2.414
14	3.466		0.0000	0.000	700.0	0.066	VV	0.014	847.82	0.087
15	3.513		0.0000	0.000	9426.3	0.889	VB	0.023	6806.12	0.695
16	3.587		0.0000	0.000	53240.2	5.023	BV	0.021	41629.31	4.253
17	3.623		0.0000	0.000	28831.9	2.720	VV	0.022	22150.99	2.263
18	3.700		0.0000	0.000	18545.5	1.750	VV	0.015	20974.79	2.143
19	3.735		0.0000	0.000	16122.3	1.521	VV	0.014	18565.39	1.897
20	3.760		0.0000	0.000	25742.9	2.429	VB	0.021	20211.67	2.065
21	3.870		0.0000	0.000	17267.5	1.629	BV	0.014	20305.33	2.074
22	3.894		0.0000	0.000	5338.7	0.504	VB	0.012	7407.15	0.757
23	3.917		0.0000	0.000	3618.8	0.341	BB	0.011	5464.07	0.558
24	3.948		0.0000	0.000	3620.3	0.342	BV	0.012	4862.25	0.497
25	3.964		0.0000	0.000	12722.2	1.200	VV	0.014	14907.49	1.523
26	4.009		0.0000	0.000	4940.1	0.466	VB	0.015	5485.24	0.560
27	4.054		0.0000	0.000	483.2	0.046	BV	0.014	564.47	0.058
28	4.079		0.0000	0.000	2824.6	0.267	VV	0.013	3568.23	0.365
29	4.103		0.0000	0.000	24318.6	2.294	VV	0.021	19116.01	1.953
30	4.205		0.0000	0.000	12146.9	1.146	VV	0.016	12719.17	1.299
31	4.241		0.0000	0.000	1326.9	0.125	VV	0.018	1249.38	0.128
32	4.259		0.0000	0.000	2111.9	0.199	VV	0.021	1700.08	0.174
33	4.304		0.0000	0.000	644.9	0.061	VV	0.030	353.76	0.036
34	4.357		0.0000	0.000	458.4	0.043	VB	0.017	449.02	0.046
35	4.416		0.0000	0.000	1646.3	0.155	BB	0.015	1840.94	0.188
36	4.462		0.0000	0.000	5060.1	0.477	BV	0.015	5772.11	0.590
37	4.492		0.0000	0.000	16395.1	1.547	VB	0.018	15523.77	1.586
38	4.521		0.0000	0.000	5641.6	0.532	BB	0.017	5667.00	0.579
39	4.574		0.0000	0.000	34517.2	3.257	BV	0.017	33145.96	3.386
40	4.643		0.0000	0.000	1026.9	0.097	VV	0.015	1156.39	0.118
41	4.670		0.0000	0.000	6502.4	0.614	VV	0.016	6613.67	0.676
42	4.713		0.0000	0.000	40856.8	3.855	VB	0.019	35799.46	3.657
43	4.821		0.0000	0.000	38528.6	3.635	BV	0.020	32697.55	3.341
44	4.877		0.0000	0.000	4206.5	0.397	VB	0.017	4115.83	0.420
45	4.915		0.0000	0.000	7137.0	0.673	BB	0.015	7790.08	0.796
46	4.951		0.0000	0.000	35049.8	3.307	BV	0.020	29438.62	3.008
47	4.984		0.0000	0.000	28345.9	2.674	VB	0.017	27315.90	2.791
48	5.024		0.0000	0.000	12539.7	1.183	BB	0.016	13411.30	1.370
49	5.126		0.0000	0.000	5872.0	0.554	BB	0.018	5315.36	0.543
50	5.178		0.0000	0.000	33472.0	3.158	BV	0.020	28230.29	2.884
51	5.251		0.0000	0.000	15365.6	1.450	VV	0.018	14321.65	1.463
52	5.287		0.0000	0.000	6720.2	0.634	VV	0.017	6492.12	0.663
53	5.312		0.0000	0.000	4654.2	0.439	VB	0.017	4462.48	0.456
54	5.367		0.0000	0.000	65653.3	6.194	BB	0.020	53916.71	5.508
55	5.442		0.0000	0.000	2091.2	0.197	BB	0.014	2444.73	0.250
56	5.557		0.0000	0.000	167.4	0.016	BB	0.012	223.31	0.023
57	5.581		0.0000	0.000	13355.9	1.260	BB	0.016	13754.70	1.405
58	5.626		0.0000	0.000	44117.9	4.163	BB	0.022	32946.86	3.366
59	5.780		0.0000	0.000	674.8	0.064	BB	0.016	719.95	0.074
60	5.820		0.0000	0.000	418.4	0.039	BB	0.015	477.27	0.049
61	5.848		0.0000	0.000	455.4	0.043	BB	0.015	515.19	0.053
62	5.896		0.0000	0.000	7688.8	0.725	BB	0.019	6572.29	0.671

ASX22 0789

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.999		0.0000	0.000	18931.9	1.786	BB	0.022	14181.53	1.449
64	6.198		0.0000	0.000	5097.8	0.481	BB	0.019	4412.03	0.451
65	6.359	DCB	34.9066	47.922	99102.8	9.350	BB	0.020	83311.85	8.512
66	6.693		0.0000	0.000	61.6	0.006	BB	0.018	56.67	0.006
67	6.740		0.0000	0.000	86.4	0.008	BB	0.027	53.65	0.005
68	7.150		0.0000	0.000	14.4	0.001	BB	0.008	30.47	0.003
69	7.373		0.0000	0.000	37.1	0.003	BB	0.029	21.34	0.002

Total Area = 1059881.0, Total Amount = 72.841, Total Height = 978813.5, Sample Units = PPB

ASX22 8789

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\2C6310.11R

AR1630924A PMAR163PM CCAL 0931099999

C:\CPWIN\DATA1\2C6310.11R



Instrument ID: CP06--S9147A Injected On: 11/7/2009 12:00:31 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\2C6310B.11R

AR1630924A PMAR163PM CCAL 0931099999

C:\CPWIN\DATA1\2C6310B.11R



ASX22 8798

Instrument ID: CP06--S9147B Injected On: 11/7/2009 12:00:31 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	75355	38.52301	TCX	2.807	173565	37.9344	TCX
7.159	13028	41.06435	DCB	6.359	83312	34.90657	DCB

Files:
Area File: C:\CPWIN\DATA1\2C6310.11A
Area File: C:\CPWIN\DATA1\2C6310B.11A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/7/2009 12:08:06 PM
File Reported On: 11/7/2009 at 12:08:15 PM

REPLOTS

ASX22 8791

AR1630924A PMAR163PM CCAL 0931099999
C:\CPWIN\DATA1\2C6310.11R



AR1630924A PMAR163PM CCAL 0931099999
C:\CPWIN\DATA1\2C6310B.11R



ASX22 8792

Sample Name: **IBLKX0924B FUPIBLKFU PIBLK0931099999**
Acquired from CP06--S9147A via port 1 on 11/7/09 12:20:00pm by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\2C6310.12R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.237		0.0000	0.000	244.4	0.420	BB	0.057	72.04	0.147
2	3.312	TCX	209.6640	50.309	40736.7	70.078	BB	0.017	41012.35	83.759
3	3.515		0.0000	0.000	562.5	0.968	BB	0.016	599.85	1.225
4	3.650		0.0000	0.000	154.4	0.266	BB	0.019	137.69	0.281
5	3.981		0.0000	0.000	81.9	0.141	BB	0.020	68.86	0.141
6	4.160		0.0000	0.000	61.0	0.105	BB	0.024	42.99	0.088
7	4.289		0.0000	0.000	44.8	0.077	BB	0.021	35.42	0.072
8	4.780		0.0000	0.000	125.8	0.216	BB	0.030	69.59	0.142
9	5.146		0.0000	0.000	121.1	0.208	BB	0.029	69.01	0.141
10	5.447		0.0000	0.000	239.6	0.412	BB	0.055	72.14	0.147
11	5.750		0.0000	0.000	140.7	0.242	BB	0.049	47.39	0.097
12	6.361		0.0000	0.000	229.4	0.395	BB	0.050	76.30	0.156
13	6.993		0.0000	0.000	332.7	0.572	BB	0.061	90.63	0.185
14	7.162	DCB	207.0864	49.691	15055.7	25.900	BB	0.038	6570.21	13.418

Total Area = 58130.7, Total Amount = 416.75, Total Height = 48964.47, Sample Units = PPB

ASX22 8793

Sample Name: **IBLKX0924B FUPIBLKFU PIBLK0931099999**
Acquired from CP06--S9147B via port 2 on 11/7/09 12:20:00pm by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\2C6310B.12R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.616		0.0000	0.000	535.2	0.354	BV	0.049	180.97	0.126
2	2.703		0.0000	0.000	640.3	0.423	VB	0.026	413.64	0.289
3	2.761		0.0000	0.000	129.3	0.085	BB	0.024	88.23	0.062
4	2.809	TCX	200.3361	51.622	88223.0	58.277	BB	0.016	91661.89	64.060
5	3.031		0.0000	0.000	1341.8	0.886	BB	0.016	1413.49	0.988
6	3.150		0.0000	0.000	365.9	0.242	BV	0.021	290.71	0.203
7	3.313		0.0000	0.000	173.4	0.115	VB	0.035	82.42	0.058
8	3.433		0.0000	0.000	132.1	0.087	BV	0.031	70.87	0.050
9	3.496		0.0000	0.000	232.7	0.154	VV	0.014	277.18	0.194
10	3.532		0.0000	0.000	329.9	0.218	VB	0.013	421.70	0.295
11	3.699		0.0000	0.000	91.4	0.060	BV	0.022	67.78	0.047
12	3.760		0.0000	0.000	189.4	0.125	VV	0.026	119.13	0.083
13	3.871		0.0000	0.000	224.7	0.148	VB	0.035	105.89	0.074
14	4.105		0.0000	0.000	132.6	0.088	BV	0.040	55.57	0.039
15	4.207		0.0000	0.000	293.5	0.194	VV	0.047	104.12	0.073
16	4.312		0.0000	0.000	836.9	0.553	VB	0.051	273.16	0.191
17	4.487		0.0000	0.000	225.2	0.149	BB	0.045	82.50	0.058
18	4.575		0.0000	0.000	166.1	0.110	BB	0.019	149.26	0.104
19	4.695		0.0000	0.000	134.7	0.089	BB	0.014	160.01	0.112
20	4.839		0.0000	0.000	841.2	0.556	BB	0.041	338.02	0.236
21	4.955		0.0000	0.000	152.4	0.101	BB	0.019	135.04	0.094
22	5.025		0.0000	0.000	63.2	0.042	BB	0.016	65.57	0.046
23	5.180		0.0000	0.000	324.7	0.214	BB	0.030	180.09	0.126
24	5.253		0.0000	0.000	110.4	0.073	BV	0.023	81.09	0.057
25	5.290		0.0000	0.000	88.8	0.059	VB	0.025	58.81	0.041
26	5.369		0.0000	0.000	805.7	0.532	BB	0.038	351.28	0.246
27	5.546		0.0000	0.000	88.4	0.058	BB	0.019	76.02	0.053
28	5.584		0.0000	0.000	83.2	0.055	BB	0.019	73.87	0.052
29	5.629		0.0000	0.000	199.1	0.132	BB	0.019	174.73	0.122
30	6.003		0.0000	0.000	151.4	0.100	BB	0.021	118.09	0.083
31	6.201		0.0000	0.000	576.8	0.381	BB	0.021	452.32	0.316
32	6.361	DCB	187.7471	48.378	53167.2	35.120	BB	0.020	44809.78	31.316
33	6.753		0.0000	0.000	67.5	0.045	BB	0.024	46.54	0.033
34	6.947		0.0000	0.000	57.6	0.038	BB	0.034	28.40	0.020
35	7.232		0.0000	0.000	144.2	0.095	BB	0.063	38.22	0.027
36	7.477		0.0000	0.000	66.8	0.044	BB	0.027	41.36	0.029

Total Area = 151386.6, Total Amount = 388.083, Total Height = 143087.8, Sample Units = PPB

ASX22 8794

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\2C6310.12R

IBLKX0924B FUPIBLKFU PIBLK0931099999

——— C:\CPWIN\DATA1\2C6310.12R



Instrument ID: CP06--S9147A Injected On: 11/7/2009 12:12:30 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\2C6310B.12R

IBLKX0924B FUPIBLKFU PIBLK0931099999

——— C:\CPWIN\DATA1\2C6310B.12R



ASX22 8795

Instrument ID: CP06--S9147B Injected On: 11/7/2009 12:12:30 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1000 Dilution Factor: 10
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.312	41012	.20966	TCX	2.809	91662	.20034	TCX
7.162	6570	.20709	DCB	6.361	44810	.18775	DCB

Files:
Area File: C:\CPWIN\DATA1\2C6310.12A
Area File: C:\CPWIN\DATA1\2C6310B.12A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/7/2009 12:20:06 PM
File Reported On: 11/7/2009 at 12:20:15 PM

REPLOTS

ASX22 0796

IBLKX0924B FUPIBLKFU PIBLK0931099999
C:\CPWIN\DATA1\2C6310.12R



IBLKX0924B FUPIBLKFU PIBLK0931099999
C:\CPWIN\DATA1\2C6310B.12R



ASX22 0797

Sample Name: **AR1630924A QDAR163QD CCAL 0931299999**
Acquired from CP06--S9147A via port 1 on 11/9/09 07:32:20pm by 1151
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\3C6310.45R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.142		0.0000	0.000	224.7	0.071	BB	0.053	70.02	0.028
2	3.256		0.0000	0.000	405.6	0.128	BB	0.077	87.44	0.035
3	3.313	TCX	45.0053	50.943	80577.3	25.524	BB	0.015	88034.77	35.240
4	3.459		0.0000	0.000	1318.8	0.418	BB	0.017	1310.30	0.525
5	3.516		0.0000	0.000	2310.8	0.732	BB	0.014	2678.62	1.072
6	3.566		0.0000	0.000	6273.7	1.987	BB	0.015	6949.21	2.782
7	3.663		0.0000	0.000	1360.2	0.431	BB	0.018	1231.04	0.493
8	3.789		0.0000	0.000	13109.5	4.153	BB	0.024	9285.96	3.717
9	3.847		0.0000	0.000	1352.1	0.428	BB	0.023	959.70	0.384
10	3.916		0.0000	0.000	7836.3	2.482	BB	0.016	8148.68	3.262
11	3.995		0.0000	0.000	284.9	0.090	BB	0.020	232.81	0.093
12	4.034		0.0000	0.000	2567.5	0.813	BB	0.024	1757.18	0.703
13	4.113		0.0000	0.000	16818.1	5.327	BB	0.028	10028.50	4.014
14	4.157		0.0000	0.000	5549.2	1.758	BB	0.017	5416.47	2.168
15	4.222		0.0000	0.000	5148.5	1.631	BB	0.017	4989.86	1.997
16	4.267		0.0000	0.000	245.0	0.078	BB	0.012	348.41	0.139
17	4.293		0.0000	0.000	3512.2	1.113	BB	0.015	3842.48	1.538
18	4.325		0.0000	0.000	5454.8	1.728	BB	0.028	3251.24	1.301
19	4.434		0.0000	0.000	3715.0	1.177	BB	0.016	3814.95	1.527
20	4.461		0.0000	0.000	719.8	0.228	BB	0.012	996.80	0.399
21	4.495		0.0000	0.000	999.2	0.317	BB	0.014	1159.27	0.464
22	4.533		0.0000	0.000	3988.2	1.263	BB	0.021	3192.61	1.278
23	4.580		0.0000	0.000	1069.2	0.339	BB	0.016	1111.08	0.445
24	4.668		0.0000	0.000	164.0	0.052	BB	0.019	146.02	0.058
25	4.710		0.0000	0.000	6444.7	2.041	BV	0.027	3992.85	1.598
26	4.758		0.0000	0.000	2109.2	0.668	VV	0.032	1113.36	0.446
27	4.830		0.0000	0.000	792.9	0.251	VV	0.025	519.10	0.208
28	4.878		0.0000	0.000	5723.2	1.813	VB	0.032	2957.88	1.184
29	5.107		0.0000	0.000	2544.6	0.806	BV	0.025	1706.91	0.683
30	5.156		0.0000	0.000	1252.7	0.397	VB	0.020	1044.14	0.418
31	5.204		0.0000	0.000	7410.4	2.347	BB	0.031	3968.82	1.589
32	5.285		0.0000	0.000	10190.8	3.228	BB	0.024	7029.78	2.814
33	5.375		0.0000	0.000	668.7	0.212	BB	0.019	590.77	0.236
34	5.434		0.0000	0.000	1167.7	0.370	BB	0.020	972.03	0.389
35	5.487		0.0000	0.000	10030.4	3.177	BB	0.028	6062.80	2.427
36	5.540		0.0000	0.000	1606.3	0.509	BB	0.016	1716.02	0.687
37	5.580		0.0000	0.000	4318.7	1.368	BB	0.026	2769.41	1.109
38	5.660		0.0000	0.000	313.4	0.099	BB	0.018	290.01	0.116
39	5.694		0.0000	0.000	8680.4	2.750	BB	0.025	5843.17	2.339
40	5.757		0.0000	0.000	5217.4	1.653	BB	0.020	4444.59	1.779
41	5.799		0.0000	0.000	2713.1	0.859	BB	0.020	2311.38	0.925
42	5.921		0.0000	0.000	8658.8	2.743	BB	0.028	5215.35	2.088
43	5.978		0.0000	0.000	2338.0	0.741	BB	0.019	2087.00	0.835
44	6.017		0.0000	0.000	769.0	0.244	BB	0.015	839.31	0.336
45	6.098		0.0000	0.000	1181.8	0.374	BB	0.020	979.73	0.392
46	6.154		0.0000	0.000	19316.8	6.119	BB	0.034	9599.76	3.843
47	6.359		0.0000	0.000	5976.8	1.893	BB	0.021	4643.69	1.859
48	6.407		0.0000	0.000	3006.3	0.952	BB	0.028	1817.68	0.728
49	6.545		0.0000	0.000	139.7	0.044	BB	0.016	144.60	0.058
50	6.628		0.0000	0.000	1983.9	0.628	BB	0.029	1138.73	0.456
51	6.777		0.0000	0.000	4656.8	1.475	BB	0.032	2460.46	0.985
52	6.992		0.0000	0.000	1441.6	0.457	BB	0.032	762.50	0.305
53	7.163	DCB	43.3397	49.057	30037.2	9.515	BB	0.036	13750.35	5.504

Total Area = 315695.8, Total Amount = 88.345, Total Height = 249815.5, Sample Units = PPB

ASX22 0799

Sample Name: **AR1630924A QDAR163QD CCAL 0931299999**
Acquired from CP06--S9147B via port 2 on 11/9/09 07:32:20pm by 1151
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\3C6310B.45R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.603		0.0000	0.000	256.4	0.023	BV	0.032	133.43	0.012
2	2.702		0.0000	0.000	882.7	0.079	VB	0.016	907.83	0.084
3	2.810 TCX		45.3828	53.755	161461.6	14.484	BB	0.013	207644.90	19.100
4	2.854		0.0000	0.000	831.5	0.075	BB	0.011	1269.49	0.117
5	2.900		0.0000	0.000	163.9	0.015	BB	0.011	248.09	0.023
6	2.984		0.0000	0.000	2936.0	0.263	BB	0.013	3768.45	0.347
7	3.031		0.0000	0.000	1026.7	0.092	BB	0.009	1845.56	0.170
8	3.046		0.0000	0.000	1625.0	0.146	BB	0.009	3078.97	0.283
9	3.099		0.0000	0.000	14541.3	1.304	BB	0.013	18829.99	1.732
10	3.150		0.0000	0.000	460.5	0.041	BB	0.013	607.62	0.056
11	3.213		0.0000	0.000	2448.7	0.220	BB	0.013	3229.90	0.297
12	3.310		0.0000	0.000	33057.6	2.965	BV	0.019	28913.30	2.660
13	3.361		0.0000	0.000	3027.3	0.272	VV	0.014	3642.52	0.335
14	3.383		0.0000	0.000	1158.6	0.104	VV	0.012	1609.28	0.148
15	3.406		0.0000	0.000	1243.8	0.112	VV	0.013	1653.75	0.152
16	3.433		0.0000	0.000	21903.0	1.965	VV	0.014	25687.22	2.363
17	3.497		0.0000	0.000	276.0	0.025	VV	0.010	461.62	0.042
18	3.514		0.0000	0.000	9880.3	0.886	VB	0.023	7125.22	0.655
19	3.589		0.0000	0.000	55840.5	5.009	BV	0.021	43740.93	4.023
20	3.626		0.0000	0.000	29971.6	2.689	VV	0.021	23656.64	2.176
21	3.702		0.0000	0.000	19280.9	1.730	VV	0.014	22317.31	2.053
22	3.737		0.0000	0.000	17956.5	1.611	VV	0.015	20358.20	1.873
23	3.762		0.0000	0.000	26999.6	2.422	VB	0.021	21850.03	2.010
24	3.872		0.0000	0.000	15832.8	1.420	BB	0.013	19559.18	1.799
25	3.896		0.0000	0.000	3645.4	0.327	BB	0.010	6385.96	0.587
26	3.919		0.0000	0.000	3877.8	0.348	BB	0.011	5796.60	0.533
27	3.950		0.0000	0.000	4226.9	0.379	BV	0.013	5273.67	0.485
28	3.966		0.0000	0.000	12784.3	1.147	VV	0.014	15157.70	1.394
29	4.011		0.0000	0.000	5174.3	0.464	VB	0.015	5659.36	0.521
30	4.056		0.0000	0.000	479.7	0.043	BV	0.013	604.73	0.056
31	4.081		0.0000	0.000	2832.8	0.254	VV	0.014	3421.70	0.315
32	4.104		0.0000	0.000	17674.6	1.586	VV	0.015	20064.04	1.846
33	4.130		0.0000	0.000	5574.2	0.500	VV	0.022	4152.03	0.382
34	4.170		0.0000	0.000	2111.6	0.189	VV	0.016	2174.71	0.200
35	4.208		0.0000	0.000	12868.7	1.154	VV	0.016	13317.46	1.225
36	4.244		0.0000	0.000	1121.8	0.101	VV	0.016	1198.76	0.110
37	4.261		0.0000	0.000	2288.1	0.205	VV	0.021	1819.08	0.167
38	4.306		0.0000	0.000	713.1	0.064	VV	0.034	354.32	0.033
39	4.360		0.0000	0.000	427.8	0.038	VB	0.016	448.28	0.041
40	4.418		0.0000	0.000	1801.4	0.162	BB	0.015	2051.96	0.189
41	4.464		0.0000	0.000	5693.9	0.511	BV	0.014	6545.38	0.602
42	4.494		0.0000	0.000	17666.7	1.585	VB	0.017	17792.65	1.637
43	4.523		0.0000	0.000	6033.2	0.541	BB	0.016	6264.74	0.576
44	4.576		0.0000	0.000	37528.0	3.366	BV	0.017	36178.58	3.328
45	4.646		0.0000	0.000	851.2	0.076	VV	0.014	986.50	0.091
46	4.672		0.0000	0.000	6897.6	0.619	VV	0.016	7112.44	0.654
47	4.716		0.0000	0.000	43770.0	3.926	VB	0.020	36813.89	3.386
48	4.823		0.0000	0.000	41585.5	3.730	BV	0.020	35387.31	3.255
49	4.879		0.0000	0.000	4316.3	0.387	VB	0.016	4383.13	0.403
50	4.917		0.0000	0.000	7856.5	0.705	BB	0.016	8429.10	0.775
51	4.954		0.0000	0.000	36432.2	3.268	BV	0.019	31748.98	2.920
52	4.986		0.0000	0.000	32538.5	2.919	VB	0.018	30640.76	2.818
53	5.026		0.0000	0.000	13547.7	1.215	BB	0.015	14591.26	1.342
54	5.128		0.0000	0.000	6350.8	0.570	BB	0.018	6028.65	0.555
55	5.180		0.0000	0.000	36458.3	3.271	BV	0.019	31401.12	2.888
56	5.253		0.0000	0.000	16944.7	1.520	VV	0.018	15298.70	1.407
57	5.289		0.0000	0.000	7428.5	0.666	VV	0.018	7066.77	0.650
58	5.314		0.0000	0.000	4857.2	0.436	VB	0.017	4805.51	0.442
59	5.369		0.0000	0.000	71261.8	6.393	BB	0.021	56749.35	5.220
60	5.444		0.0000	0.000	2227.9	0.200	BB	0.015	2472.24	0.227
61	5.560		0.0000	0.000	212.3	0.019	BB	0.016	223.26	0.021
62	5.583		0.0000	0.000	14638.5	1.313	BB	0.016	15257.59	1.403

ASX22 8799

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.628		0.0000	0.000	51073.8	4.582	BB	0.022	38241.43	3.518
64	5.783		0.0000	0.000	699.1	0.063	BB	0.015	791.29	0.073
65	5.821		0.0000	0.000	381.7	0.034	BB	0.015	434.39	0.040
66	5.850		0.0000	0.000	432.4	0.039	BB	0.013	540.31	0.050
67	5.898		0.0000	0.000	8533.5	0.766	BB	0.020	7182.76	0.661
68	6.002		0.0000	0.000	20894.3	1.874	BB	0.022	15522.96	1.428
69	6.200		0.0000	0.000	5656.9	0.507	BB	0.019	5012.88	0.461
70	6.361	DCB	39.0426	46.245	111278.8	9.982	BB	0.020	93183.43	8.571
71	7.428		0.0000	0.000	48.0	0.004	BB	0.027	29.22	0.003

Total Area = 1114761.0, Total Amount = 84.425, Total Height = 1087136.0, Sample Units = PPB

ASX22 0008

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3C6310.45R
AR1630924A QDAR163QD CCAL 0931299999
——— C:\CPWIN\DATA1\3C6310.45R



Instrument ID: CP06--S9147A Injected On: 11/9/2009 7:24:50 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\3C6310B.45R
AR1630924A QDAR163QD CCAL 0931299999
——— C:\CPWIN\DATA1\3C6310B.45R



ASX22 9881

Instrument ID: CP06–S9147B Injected On: 11/9/2009 7:24:50 PM Column ID: RTX CLPII

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 1151

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.313	88035	45.00527	TCX	2.81	207645	45.38283	TCX
7.163	13750	43.33972	DCB	6.361	93183	39.04263	DCB

Files:
Area File: C:\CPWIN\DATA1\3C6310.45A
Area File: C:\CPWIN\DATA1\3C6310B.45A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/9/2009 7:32:26 PM
File Reported On: 11/9/2009 at 7:32:35 PM

REPLOTS

ASX22 7802

AR1630924A QDAR163QD CCAL 0931299999
C:\CPWIN\DATA1\3C6310.45R



AR1630924A QDAR163QD CCAL 0931299999
C:\CPWIN\DATA1\3C6310B.45R



ASX22 0003

Sample Name: **IBLKX0924B GHPIBLKGH PIBLK0931299999**
Acquired from CP06--S9147A via port 1 on 11/9/09 07:44:20pm by 1151
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\3C6310.46R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.143		0.0000	0.000	203.4	0.342	BB	0.038	89.95	0.175
2	3.249		0.0000	0.000	339.1	0.570	BB	0.059	95.90	0.186
3	3.311	TCX	220.3133	49.663	42318.0	71.145	BB	0.016	43095.47	83.711
4	3.513		0.0000	0.000	547.7	0.921	BB	0.015	598.72	1.163
5	3.648		0.0000	0.000	114.3	0.192	BB	0.013	142.42	0.277
6	3.899		0.0000	0.000	18.6	0.031	BB	0.007	44.35	0.086
7	3.977		0.0000	0.000	126.6	0.213	BB	0.026	82.68	0.161
8	5.373		0.0000	0.000	66.5	0.112	BB	0.016	70.47	0.137
9	5.974		0.0000	0.000	42.6	0.072	BB	0.018	38.95	0.076
10	6.462		0.0000	0.000	241.7	0.406	BB	0.081	49.90	0.097
11	6.992		0.0000	0.000	227.1	0.382	BB	0.043	87.87	0.171
12	7.161	DCB	223.3073	50.337	15235.9	25.615	BB	0.036	7084.85	13.762

Total Area = 59481.4, Total Amount = 443.621, Total Height = 51481.54, Sample Units = PPB

ASX22 0004

Sample Name: **IBLKX0924B GHPIBLKGH PIBLK0931299999**
Acquired from CP06--S9147B via port 2 on 11/9/09 07:44:20pm by 1151
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\3C6310B.46R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.604		0.0000	0.000	368.1	0.241	BB	0.043	141.86	0.093
2	2.700		0.0000	0.000	722.4	0.472	BB	0.017	698.19	0.456
3	2.808	TCX	217.8155	52.169	89408.1	58.415	BB	0.015	99659.45	65.119
4	3.030		0.0000	0.000	1372.7	0.897	BB	0.015	1492.71	0.975
5	3.149		0.0000	0.000	309.5	0.202	BB	0.017	309.21	0.202
6	3.400		0.0000	0.000	192.7	0.126	BB	0.045	72.13	0.047
7	3.496		0.0000	0.000	241.2	0.158	BV	0.014	295.78	0.193
8	3.531		0.0000	0.000	767.1	0.501	VV	0.026	497.81	0.325
9	3.759		0.0000	0.000	97.3	0.064	VB	0.021	78.03	0.051
10	3.933		0.0000	0.000	448.8	0.293	BB	0.043	174.80	0.114
11	4.121		0.0000	0.000	69.4	0.045	BB	0.012	95.54	0.062
12	4.310		0.0000	0.000	1060.0	0.693	BB	0.066	268.06	0.175
13	4.469		0.0000	0.000	190.3	0.124	BB	0.049	64.83	0.042
14	4.695		0.0000	0.000	589.6	0.385	BB	0.029	340.20	0.222
15	4.836		0.0000	0.000	1028.1	0.672	BB	0.050	343.64	0.225
16	5.546		0.0000	0.000	340.6	0.223	BB	0.027	210.82	0.138
17	5.687		0.0000	0.000	109.6	0.072	BB	0.037	48.87	0.032
18	5.880		0.0000	0.000	67.2	0.044	BB	0.024	45.75	0.030
19	6.013		0.0000	0.000	47.9	0.031	BB	0.019	41.47	0.027
20	6.200		0.0000	0.000	604.7	0.395	BB	0.022	458.72	0.300
21	6.359	DCB	199.7055	47.831	54938.8	35.894	BB	0.019	47663.91	31.144
22	7.469		0.0000	0.000	82.8	0.054	BB	0.034	41.01	0.027

Total Area = 153057.0, Total Amount = 417.521, Total Height = 153042.8, Sample Units = PPB

ASX22. 0885

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3C6310.46R

IBLKX0924B GHPIBLKGH PIBLK0931299999

C:\CPWIN\DATA1\3C6310.46R



Instrument ID: CP06--S9147A Injected On: 11/9/2009 7:36:50 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\3C6310B.46R

IBLKX0924B GHPIBLKGH PIBLK0931299999

C:\CPWIN\DATA1\3C6310B.46R



Instrument ID: CP06--S9147B Injected On: 11/9/2009 7:36:50 PM Column ID: RTX CLPII

ASX22 0886

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min; Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1000 Dilution Factor: 10
Analyst: 1151

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	43095	.22031	TCX	2.808	99659	.21782	TCX
7.161	7085	.22331	DCB	6.359	47664	.19971	DCB

Files:
Area File: C:\CPWIN\DATA1\3C6310.46A
Area File: C:\CPWIN\DATA1\3C6310B.46A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/9/2009 7:44:26 PM
File Reported On: 11/9/2009 at 7:44:34 PM

REPLOTS

ASX22: 8887

IBLKX0924B GHPIBLKGH PIBLK0931299999
C:\CPWIN\DATA1\3C6310.46R



IBLKX0924B GHPIBLKGH PIBLK0931299999
C:\CPWIN\DATA1\3C6310B.46R



ASX22 0088

Sample Name: **AR1620924A OTAR162OT CCAL 0931299999**
Acquired from CP06--S9147A via port 1 on 11/9/09 09:56:05pm by 1151
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\3C6310.57R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.103		0.0000	0.000	584.1	0.450	BB	0.082	118.41	0.108
2	3.222		0.0000	0.000	482.2	0.372	BB	0.049	162.68	0.148
3	3.313	TCX	23.7908	61.707	41475.9	31.980	BB	0.015	46537.11	42.411
4	3.458		0.0000	0.000	649.1	0.501	BB	0.016	663.65	0.605
5	3.516		0.0000	0.000	1200.0	0.925	BB	0.014	1385.31	1.262
6	3.566		0.0000	0.000	3409.4	2.629	BB	0.015	3798.16	3.461
7	3.663		0.0000	0.000	597.6	0.461	BB	0.016	628.10	0.572
8	3.790		0.0000	0.000	6716.5	5.179	BB	0.023	4888.23	4.455
9	3.847		0.0000	0.000	681.5	0.526	BB	0.024	470.73	0.429
10	3.916		0.0000	0.000	3839.0	2.960	BB	0.017	3745.43	3.413
11	4.035		0.0000	0.000	948.4	0.731	BB	0.023	684.11	0.623
12	4.113		0.0000	0.000	6092.5	4.698	BB	0.027	3711.07	3.382
13	4.157		0.0000	0.000	2002.7	1.544	BB	0.017	1933.13	1.762
14	4.223		0.0000	0.000	1770.7	1.365	BB	0.017	1709.38	1.558
15	4.268		0.0000	0.000	87.7	0.068	BB	0.010	139.70	0.127
16	4.293		0.0000	0.000	1204.4	0.929	BB	0.015	1303.36	1.188
17	4.325		0.0000	0.000	1897.2	1.463	BB	0.029	1093.85	0.997
18	4.434		0.0000	0.000	1288.2	0.993	BB	0.016	1342.31	1.223
19	4.462		0.0000	0.000	287.8	0.222	BB	0.012	413.95	0.377
20	4.494		0.0000	0.000	357.6	0.276	BB	0.014	426.09	0.388
21	4.533		0.0000	0.000	1478.3	1.140	BB	0.021	1185.29	1.080
22	4.579		0.0000	0.000	531.3	0.410	BB	0.021	420.77	0.383
23	4.711		0.0000	0.000	2643.5	2.038	BV	0.028	1591.97	1.451
24	4.759		0.0000	0.000	898.1	0.692	VV	0.031	485.50	0.442
25	4.828		0.0000	0.000	301.2	0.232	VV	0.025	200.04	0.182
26	4.879		0.0000	0.000	2081.0	1.605	VB	0.030	1167.47	1.064
27	5.032		0.0000	0.000	84.9	0.065	BB	0.025	57.75	0.053
28	5.108		0.0000	0.000	1129.6	0.871	BV	0.027	707.74	0.645
29	5.156		0.0000	0.000	487.0	0.375	VB	0.020	406.83	0.371
30	5.206		0.0000	0.000	2755.0	2.124	BB	0.031	1464.75	1.335
31	5.285		0.0000	0.000	3701.7	2.854	BB	0.023	2645.73	2.411
32	5.377		0.0000	0.000	311.7	0.240	BB	0.022	240.24	0.219
33	5.435		0.0000	0.000	420.1	0.324	BB	0.019	366.23	0.334
34	5.487		0.0000	0.000	3622.3	2.793	BB	0.027	2263.54	2.063
35	5.540		0.0000	0.000	565.0	0.436	BB	0.016	591.31	0.539
36	5.582		0.0000	0.000	1431.9	1.104	BB	0.025	941.67	0.858
37	5.658		0.0000	0.000	99.1	0.076	BB	0.014	119.77	0.109
38	5.695		0.0000	0.000	3259.7	2.513	BB	0.025	2211.98	2.016
39	5.758		0.0000	0.000	1937.7	1.494	BB	0.019	1696.82	1.546
40	5.801		0.0000	0.000	1011.4	0.780	BB	0.020	848.29	0.773
41	5.922		0.0000	0.000	3151.4	2.430	BB	0.028	1891.59	1.724
42	5.980		0.0000	0.000	825.0	0.636	BB	0.019	707.13	0.644
43	6.019		0.0000	0.000	222.4	0.171	BB	0.013	288.71	0.263
44	6.099		0.0000	0.000	399.1	0.308	BB	0.019	348.09	0.317
45	6.155		0.0000	0.000	6199.0	4.780	BB	0.031	3304.01	3.011
46	6.360		0.0000	0.000	2212.0	1.706	BB	0.021	1720.86	1.568
47	6.410		0.0000	0.000	981.3	0.757	BB	0.026	620.11	0.565
48	6.631		0.0000	0.000	601.9	0.464	BB	0.028	359.59	0.328
49	6.778		0.0000	0.000	1040.2	0.802	BB	0.022	771.38	0.703
50	6.991		0.0000	0.000	619.5	0.478	BB	0.039	263.76	0.240
51	7.165	DCB	14.7635	38.293	9117.6	7.030	BB	0.032	4684.01	4.269

Total Area = 129692.3, Total Amount = 38.554, Total Height = 109727.7, Sample Units = PPB

ASX22 0809

Sample Name: **AR1620924A OTAR162OT CCAL 0931299999**
Acquired from CP06--S9147B via port 2 on 11/9/09 09:56:05pm by 1151
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\3C6310B.57R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.603		0.0000	0.000	567.9	0.140	BV	0.046	204.06	0.049
2	2.701		0.0000	0.000	1042.5	0.257	VB	0.016	1076.28	0.257
3	2.811	TCX	22.9918	73.059	78598.0	19.394	BB	0.012	105196.70	25.127
4	2.855		0.0000	0.000	500.6	0.124	BB	0.012	696.63	0.166
5	2.929		0.0000	0.000	131.8	0.033	BB	0.013	174.06	0.042
6	2.984		0.0000	0.000	1445.9	0.357	BB	0.013	1869.99	0.447
7	3.047		0.0000	0.000	2714.4	0.670	BB	0.021	2165.97	0.517
8	3.099		0.0000	0.000	7190.8	1.774	BB	0.013	9579.01	2.288
9	3.150		0.0000	0.000	185.3	0.046	BB	0.012	254.38	0.061
10	3.213		0.0000	0.000	1113.3	0.275	BB	0.012	1539.07	0.368
11	3.310		0.0000	0.000	15832.5	3.907	BV	0.018	14390.37	3.437
12	3.362		0.0000	0.000	1371.0	0.338	VV	0.014	1655.34	0.395
13	3.383		0.0000	0.000	542.4	0.134	VV	0.012	726.46	0.174
14	3.406		0.0000	0.000	623.7	0.154	VV	0.013	782.88	0.187
15	3.433		0.0000	0.000	10419.1	2.571	VB	0.014	12464.83	2.977
16	3.497		0.0000	0.000	141.0	0.035	BV	0.010	232.78	0.056
17	3.515		0.0000	0.000	4478.9	1.105	VV	0.022	3453.12	0.825
18	3.589		0.0000	0.000	26703.2	6.589	VV	0.021	20871.92	4.985
19	3.626		0.0000	0.000	13909.2	3.432	VV	0.020	11372.97	2.717
20	3.703		0.0000	0.000	9221.4	2.275	VV	0.014	10636.65	2.541
21	3.737		0.0000	0.000	8088.4	1.996	VV	0.014	9380.60	2.241
22	3.763		0.0000	0.000	12307.9	3.037	VB	0.020	10356.91	2.474
23	3.872		0.0000	0.000	7412.3	1.829	BB	0.013	9343.94	2.232
24	3.896		0.0000	0.000	1669.7	0.412	BB	0.010	2917.11	0.697
25	3.919		0.0000	0.000	1747.9	0.431	BB	0.011	2647.68	0.632
26	3.950		0.0000	0.000	1843.6	0.455	BV	0.013	2387.20	0.570
27	3.967		0.0000	0.000	5858.2	1.446	VV	0.014	6971.06	1.665
28	4.011		0.0000	0.000	2223.7	0.549	VB	0.015	2472.14	0.590
29	4.056		0.0000	0.000	167.6	0.041	BV	0.013	214.38	0.051
30	4.082		0.0000	0.000	1118.0	0.276	VV	0.013	1439.48	0.344
31	4.105		0.0000	0.000	8757.4	2.161	VV	0.018	8104.01	1.936
32	4.171		0.0000	0.000	547.1	0.135	VB	0.014	673.00	0.161
33	4.208		0.0000	0.000	4282.7	1.057	BV	0.015	4755.69	1.136
34	4.261		0.0000	0.000	699.1	0.173	VB	0.024	479.28	0.114
35	4.305		0.0000	0.000	246.4	0.061	BV	0.027	154.63	0.037
36	4.360		0.0000	0.000	459.1	0.113	VV	0.036	211.13	0.050
37	4.418		0.0000	0.000	643.7	0.159	VB	0.015	721.69	0.172
38	4.464		0.0000	0.000	1801.5	0.445	BV	0.014	2093.38	0.500
39	4.495		0.0000	0.000	5889.0	1.453	VB	0.017	5810.12	1.388
40	4.524		0.0000	0.000	2123.0	0.524	BB	0.016	2157.56	0.515
41	4.576		0.0000	0.000	11637.4	2.872	BV	0.016	11835.96	2.827
42	4.647		0.0000	0.000	323.5	0.080	VV	0.015	357.50	0.085
43	4.673		0.0000	0.000	2526.2	0.623	VV	0.018	2358.86	0.563
44	4.716		0.0000	0.000	13676.3	3.375	VB	0.018	12579.47	3.005
45	4.824		0.0000	0.000	12734.4	3.142	BV	0.019	11179.77	2.670
46	4.880		0.0000	0.000	1320.6	0.326	VB	0.016	1371.37	0.328
47	4.918		0.0000	0.000	2508.6	0.619	BV	0.015	2734.79	0.653
48	4.954		0.0000	0.000	11789.6	2.909	VV	0.020	9848.91	2.352
49	4.987		0.0000	0.000	9275.5	2.289	VB	0.017	9117.89	2.178
50	5.027		0.0000	0.000	4167.1	1.028	BB	0.015	4529.92	1.082
51	5.129		0.0000	0.000	1864.2	0.460	BV	0.018	1753.56	0.419
52	5.181		0.0000	0.000	10442.4	2.577	VB	0.019	9238.21	2.207
53	5.254		0.0000	0.000	4913.9	1.213	BV	0.017	4726.57	1.129
54	5.289		0.0000	0.000	2142.3	0.529	VV	0.017	2074.71	0.496
55	5.316		0.0000	0.000	1558.2	0.384	VV	0.017	1496.96	0.358
56	5.370		0.0000	0.000	20432.4	5.042	VB	0.020	17394.86	4.155
57	5.445		0.0000	0.000	599.3	0.148	BB	0.014	725.67	0.173
58	5.541		0.0000	0.000	754.2	0.186	BB	0.018	680.44	0.163
59	5.583		0.0000	0.000	3702.7	0.914	BB	0.016	3932.46	0.939
60	5.629		0.0000	0.000	13158.9	3.247	BB	0.021	10548.16	2.520
61	5.782		0.0000	0.000	164.5	0.041	BB	0.016	170.82	0.041
62	5.823		0.0000	0.000	138.7	0.034	BB	0.017	139.24	0.033

ASX22 8818

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.899		0.0000	0.000	2095.2	0.517	BB	0.019	1861.03	0.445
64	6.002		0.0000	0.000	5016.5	1.238	BB	0.022	3828.64	0.914
65	6.202		0.0000	0.000	1179.7	0.291	BB	0.018	1097.58	0.262
66	6.361	DCB	8.4783	26.941	21952.1	5.417	BB	0.018	20235.15	4.833
67	6.591		0.0000	0.000	77.6	0.019	BB	0.036	35.73	0.009
68	6.670		0.0000	0.000	158.1	0.039	BB	0.037	71.02	0.017
69	6.916		0.0000	0.000	258.1	0.064	BB	0.096	44.69	0.011
70	7.166		0.0000	0.000	74.7	0.018	BB	0.022	56.11	0.013

Total Area = 405262.3, Total Amount = 31.47, Total Height = 418660.5, Sample Units = PPB

ASX22 0811

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3C6310.57R

AR1620924A OTAR162OT CCAL 0931299999

C:\CPWIN\DATA1\3C6310.57R



Instrument ID: CP06--S9147A Injected On: 11/9/2009 9:48:35 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\3C6310B.57R

AR1620924A OTAR162OT CCAL 0931299999

C:\CPWIN\DATA1\3C6310B.57R



Instrument ID: CP06--S9147B Injected On: 11/9/2009 9:48:35 PM Column ID: RTX CLPII

ASX22 0812

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min\Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1 Dilution Factor: 1
Analyst: 1151

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.313	46537	23.79078	TCX	2.811	105197	22.99177	TCX
7.165	4684	14.76352	DCB	6.361	20235	8.47826	DCB

Files:
Area File: C:\CPWIN\DATA1\3C6310.57A
Area File: C:\CPWIN\DATA1\3C6310B.57A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/9/2009 9:56:12 PM
File Reported On: 11/9/2009 at 9:56:20 PM

REPLOTS

ASX22 F813

AR1620924A OTAR162OT CCAL 0931299999
C:\CPWIN\DATA1\3C6310.57R



AR1620924A OTAR162OT CCAL 0931299999
C:\CPWIN\DATA1\3C6310B.57R



ASX22 8814

Sample Name: **IBLKX0924B GIPIBLKGI PIBLK0931299999**
Acquired from CP06–S9147A via port 1 on 11/9/09 10:08:08pm by 1151
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\3C6310.58R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.167		0.0000	0.000	91.9	0.171	BB	0.022	69.79	0.134
2	3.226		0.0000	0.000	131.8	0.246	BB	0.036	61.04	0.118
3	3.312	TCX	234.0531	61.802	42086.8	78.493	BB	0.015	45783.10	88.197
4	3.514		0.0000	0.000	506.7	0.945	BB	0.014	600.12	1.156
5	3.649		0.0000	0.000	150.6	0.281	BB	0.017	148.05	0.285
6	3.979		0.0000	0.000	135.6	0.253	BB	0.019	117.28	0.226
7	4.757		0.0000	0.000	308.7	0.576	BV	0.062	82.89	0.160
8	4.875		0.0000	0.000	1151.8	2.148	VV	0.138	139.18	0.268
9	5.140		0.0000	0.000	219.5	0.409	VB	0.030	123.17	0.237
10	5.994		0.0000	0.000	129.8	0.242	BB	0.026	84.42	0.163
11	6.180		0.0000	0.000	47.5	0.089	BB	0.033	24.03	0.046
12	6.959		0.0000	0.000	116.9	0.218	BB	0.065	29.80	0.057
13	6.990		0.0000	0.000	68.5	0.128	BB	0.020	57.61	0.111
14	7.162	DCB	144.6612	38.198	8472.3	15.801	BB	0.031	4589.65	8.842

Total Area = 53618.5, Total Amount = 378.714, Total Height = 51910.13, Sample Units = PPB

03X22 0815

Sample Name: **IBLKX0924B GIPIBLKGI PIBLK0931299999**
Acquired from CP06--S9147B via port 2 on 11/9/09 10:08:08pm by 1151
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\3C6310B.58R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.599		0.0000	0.000	687.5	0.611	BV	0.046	246.44	0.193
2	2.699		0.0000	0.000	580.7	0.516	VB	0.021	469.07	0.368
3	2.808	TCX	222.3897	72.441	81059.6	72.016	BB	0.013	101752.30	79.732
4	2.929		0.0000	0.000	235.1	0.209	BB	0.032	121.31	0.095
5	3.030		0.0000	0.000	1108.0	0.984	BB	0.012	1495.77	1.172
6	3.065		0.0000	0.000	172.1	0.153	BB	0.022	130.56	0.102
7	3.149		0.0000	0.000	226.7	0.201	BB	0.013	294.14	0.230
8	3.397		0.0000	0.000	68.9	0.061	BB	0.017	69.62	0.055
9	3.496		0.0000	0.000	386.5	0.343	BV	0.018	360.26	0.282
10	3.530		0.0000	0.000	1543.4	1.371	VB	0.065	396.65	0.311
11	3.941		0.0000	0.000	636.6	0.566	BB	0.053	200.55	0.157
12	4.182		0.0000	0.000	179.1	0.159	BB	0.052	57.49	0.045
13	4.305		0.0000	0.000	322.8	0.287	BB	0.027	202.92	0.159
14	4.692		0.0000	0.000	574.8	0.511	BB	0.027	357.36	0.280
15	4.832		0.0000	0.000	717.0	0.637	BB	0.043	278.30	0.218
16	5.388		0.0000	0.000	778.3	0.692	BB	0.086	150.23	0.118
17	5.542		0.0000	0.000	644.8	0.573	BB	0.024	455.27	0.357
18	5.682		0.0000	0.000	85.1	0.076	BB	0.036	39.01	0.031
19	6.083		0.0000	0.000	79.7	0.071	BB	0.027	48.62	0.038
20	6.200		0.0000	0.000	307.7	0.273	BB	0.024	215.55	0.169
21	6.359	DCB	84.6035	27.559	22040.5	19.582	BB	0.018	20192.41	15.823
22	7.054		0.0000	0.000	48.3	0.043	BB	0.020	39.24	0.031
23	7.168		0.0000	0.000	74.3	0.066	BB	0.027	45.20	0.035

Total Area = 112557.5, Total Amount = 306.993, Total Height = 127618.3, Sample Units = PPB

ASX22 6816

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3C6310.58R

IBLKX0924B GIPIBLKGI PIBLK0931299999

C:\CPWIN\DATA1\3C6310.58R



Instrument ID: CP06--S9147A Injected On: 11/9/2009 10:00:38 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\3C6310B.58R

IBLKX0924B GIPIBLKGI PIBLK0931299999

C:\CPWIN\DATA1\3C6310B.58R



Instrument ID: CP06--S9147B Injected On: 11/9/2009 10:00:38 PM Column ID: RTX CLPII

ASX22 0817

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 1000 Dilution Factor: 10
Analyst: 1151

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.312	45783	.23405	TCX	2.808	101752	.22239	TCX
7.162	4590	.14466	DCB	6.359	20192	.0846	DCB

Files:
Area File: C:\CPWIN\DATA1\3C6310.58A
Area File: C:\CPWIN\DATA1\3C6310B.58A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/9/2009 10:08:14 PM
File Reported On: 11/9/2009 at 10:08:23 PM

REPLOTS

ASX22 0018

IBLKX0924B GIPIBLKGI PIBLK0931299999
C:\CPWIN\DATA1\3C6310.58R



IBLKX0924B GIPIBLKGI PIBLK0931299999
C:\CPWIN\DATA1\3C6310B.58R



ASXZZ 8819

Raw QC Data

Multiple Component Data Summary

Sample Name: BLANKA 11/3/09 RI A PBLK65 BLK Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1000 ml Total Volume: 10 ml Analyst: 2131 SDG: State:
Analyses: 00173

Analysis Report (A)

Injected on	Nov 06, 2009 00:07:24
Instrument	H5088A
Result file	3P7307.42R
Calibration file	1P7307
Method file	PESTD

%SSR(TCX)	106.0%	Conc:	0.325525
%SSR(DCB)	69.1%	Conc:	0.210802

Analysis Report (B)

Injected on	Nov 06, 2009 00:07:24
Instrument	H5088B
Result file	3P7307B.42R
Calibration file	1P7307B
Method file	PESTDB

%SSR(TCX)	105.5%	Conc:	0.323885
%SSR(DCB)	82.8%	Conc:	0.252607

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<0.5	<0.1			4	15	
Aroclor-1221			<0.5	<0.1			3	5	
Aroclor-1232			<0.5	<0.1			4	10	
Aroclor-1242			<0.5	<0.1			4	30	
Aroclor-1248			<0.5	<0.1			4	30	
Aroclor-1254			<0.5	<0.1			4	40	
Aroclor-1260			<0.5	<0.1			4	40	

Units: ug/l

%Difference = High - Low divided by the Average times 100

Reviewed By: [illegible] Date: 11-4-09

Verified By: [illegible] Date: 11/9/09

ASX22 8821

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: BLANKA 11/3/09 RI A PBLK65 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1000 ml Total Volume: 10 ml Analyst: 2280 SDG: State:
Analyses: 00173 00516 06030 07394 10227

Analysis Report (A)

Injected on : NOV 06, 2009 00:07:24
Instrument : CP07–H5088A
Result file : 3P7307.42R
Calibration file : 1P7307.CAL
Method file : PESTD.MET
%SSR(TCX) : 106% Conc.: 0.325525
%SSR(DCB) : 69.1% Conc.: 0.210802

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					5	48.93	
3.35	3.37	3.41	88.995895	0.008509			1
3.83	3.86	3.89	450.149811	0.035497			2
4.10	4.10	4.16	453.924683	0.016696			3
4.24	4.27	4.30	423.972717	0.037041			5
4.62	4.68	4.68	271.740845	0.033312			6
Height Summation:			1688.783951				
Amount Avg CF:			0.026211	Linear:			
Aroclor-1221					2	138.57	
3.28	3.29	3.34	2630.936035	0.600328			2
3.35	3.37	3.41	88.995895	0.006123			3
Height Summation:			2719.931930				
Amount Avg CF:			0.303226	Linear:			
Aroclor-1232					5	59.35	
3.35	3.37	3.41	88.995895	0.00725			1
3.83	3.86	3.89	450.149811	0.079778			2
4.10	4.10	4.16	453.924683	0.034882			3
4.23	4.27	4.29	423.972717	0.078982			5
4.62	4.68	4.68	271.740845	0.076709			6
Height Summation:			1688.783951				
Amount Avg CF:			0.05552	Linear:			
Aroclor-1242					5	49.60	
3.35	3.37	3.41	88.995895	0.009952			1
3.83	3.86	3.89	450.149811	0.043442			2
4.09	4.10	4.15	453.924683	0.019303			3
4.23	4.27	4.29	423.972717	0.043185			5
4.62	4.68	4.68	271.740845	0.038014			6
Height Summation:			1688.783951				
Amount Avg CF:			0.030779	Linear:			
Aroclor-1248					2	9.88	
4.10	4.10	4.16	453.924683	0.030275			1
4.62	4.68	4.68	271.740845	0.026321			4
Height Summation:			725.665528				
Amount Avg CF:			0.028298	Linear:			
Aroclor-1254					3	97.66	
+ 4.83	4.83	4.89	1096.98877	0.091376			1
4.83	4.87	4.89	4580.540527	0.381547			1
5.34	5.38	5.40	3827.986816	0.161563			4
5.85	5.90	5.91	226.980698	0.018561			5
Height Summation:			8635.508041				
Amount Avg CF:			0.187224	Linear:			
Aroclor-1260					2	98.35	
6.05	6.09	6.11	661.011108	0.048112			3
6.80	6.84	6.86	127.270279	0.008642			6
Height Summation:			788.281387				
Amount Avg CF:			0.028377	Linear:			
Aroclor-1262					2	92.45	
6.05	6.09	6.11	661.011108	0.028819			3
6.79	6.84	6.85	127.270279	0.006034			6
Height Summation:			788.281387				
Amount Avg CF:			0.017427	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 00:07:24
Instrument : CP07–H5088B
Result file : 3P7307B.42R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Alincar

%SSR(TCX) : 105.5% Conc.: 0.323885
%SSR(DCB) : 82.8% Conc.: 0.252607

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					2	100.84	
2.98	3.01	3.04	87.377968	0.005104			2
3.14	3.15	3.20	359.28714	0.030468			3
Height Summation:			446.665108				
Amount Avg CF:			0.017786	Linear:			
Aroclor-1221					1		
2.61	2.62	2.67	1735.274902	0.611551			2
Height Summation:			1735.274902				
Amount Avg CF:			0.611551	Linear:			
Aroclor-1232					2	99.73	
2.98	3.01	3.04	87.377968	0.011258			2
3.14	3.15	3.20	359.28714	0.065124			3
Height Summation:			446.665108				
Amount Avg CF:			0.038191	Linear:			
Aroclor-1242					2	100.96	
2.98	3.01	3.04	87.377968	0.006265			2
3.14	3.15	3.20	359.28714	0.037527			3
Height Summation:			446.665108				
Amount Avg CF:			0.021896	Linear:			
Aroclor-1248					3	81.66	
3.14	3.15	3.20	359.28714	0.085243			1
4.26	4.30	4.32	843.295166	0.082114			5
4.55	4.60	4.61	1913.739014	0.307565			6
Height Summation:			3116.321320				
Amount Avg CF:			0.158307	Linear:			
Aroclor-1254					3	56.42	
+ 4.08	4.10	4.14	85.625145	0.004588			1
4.08	4.13	4.14	3052.045166	0.163522			1
4.55	4.60	4.61	1913.739014	0.091996			2
4.66	4.69	4.72	588.472351	0.049902			3
Height Summation:			5554.256531				
Amount Avg CF:			0.101807	Linear:			
Aroclor-1260					1		
4.66	4.69	4.72	588.472351	0.031348			1
Height Summation:			588.472351				
Amount Avg CF:			0.031348	Linear:			

Lmn

ASX22: 0822

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:07:13

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: BLANKA 11/3/09 RI A PBLK65 Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1000 ml Total Volume: 10 ml Analyst: 2280 SDG: State:
Analyses: 00173 00516 06030 07394 10227

Analysis Report (A)

Injected on : NOV 06, 2009 00:07:24
Instrument : CP07--H5088A
Result file : 3P7307.42R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1268					3	158.60	
6.05	6.09	6.11	661.011108	0.072836			1
6.79	6.84	6.85	127.270279	0.003044			3
6.87	6.91	6.93	47.195663	0.001307			4

Height Summation: **835.477050**
Amount Avg CF: **0.025729** Linear:

Analysis Report (B)

Injected on : NOV 06, 2009 00:07:24
Instrument : CP07--H5088B
Result file : 3P7307B.42R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			0.5	0.1		38.30	4	15	
Aroclor-1221			0.5	0.1		** 67.41	3	5	
Aroclor-1232			0.5	0.1		36.98	4	10	
			0.5	0.2					
Aroclor-1242			0.5	0.1		33.73	4	30	
Aroclor-1248			0.5	0.1		** 139.34	4	30	
Aroclor-1254			0.5	0.1		** 59.11	4	40	
Aroclor-1260			0.5	0.1		9.95	4	40	
Aroclor-1262			0.5	0.1			4	40	
			0.5	0.2					
Aroclor-1268			0.5	0.1			4	40	

Units: ug/L

ASX22 8823

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:07:13

Sample Name: **BLANKA 11/3/09 RI AABPBLK65 BLK 093060018A 00173**
Acquired from CP07--H5088A via port 3 on 11/6/09 12:16:32am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.42R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.579		0.0000	0.000	4019.0	0.687	BB	0.051	1323.31	0.503
2	1.803		0.0000	0.000	23818.8	4.069	BV	0.167	2370.18	0.900
3	1.934		0.0000	0.000	3343.6	0.571	VV	0.060	932.48	0.354
4	2.031		0.0000	0.000	6657.9	1.138	VV	0.058	1903.79	0.723
5	2.077		0.0000	0.000	2903.5	0.496	VV	0.034	1413.62	0.537
6	2.131		0.0000	0.000	740.0	0.126	VB	0.030	415.33	0.158
7	2.176		0.0000	0.000	1031.0	0.176	BB	0.023	733.06	0.278
8	2.289		0.0000	0.000	672.6	0.115	BV	0.032	349.95	0.133
9	2.343		0.0000	0.000	1017.7	0.174	VB	0.056	302.89	0.115
10	2.504		0.0000	0.000	1696.4	0.290	BV	0.057	492.39	0.187
11	2.567		0.0000	0.000	1840.6	0.314	VB	0.059	519.11	0.197
12	2.708		0.0000	0.000	133.5	0.023	BB	0.021	108.00	0.041
13	2.770		0.0000	0.000	1160.0	0.198	BB	0.040	488.78	0.186
14	2.840		0.0000	0.000	491.6	0.084	BB	0.054	150.43	0.057
15	2.940		0.0000	0.000	6480.7	1.107	BB	0.023	4644.92	1.764
16	3.000	TCX	325.5253	60.695	248192.5	42.404	BB	0.027	152377.10	57.873
17	3.176		0.0000	0.000	686.4	0.117	BB	0.023	499.13	0.190
18	3.289		0.0000	0.000	4254.2	0.727	BB	0.027	2630.94	0.999
19	3.366		0.0000	0.000	220.1	0.038	BB	0.041	89.00	0.034
20	3.474		0.0000	0.000	993.3	0.170	BB	0.035	474.68	0.180
21	3.557		0.0000	0.000	392.4	0.067	BB	0.054	120.12	0.046
22	3.702		0.0000	0.000	555.1	0.095	BB	0.123	75.32	0.029
23	3.770		0.0000	0.000	452.5	0.077	BB	0.057	133.37	0.051
24	3.857		0.0000	0.000	1099.0	0.188	BB	0.041	450.15	0.171
25	3.939		0.0000	0.000	845.8	0.145	BB	0.026	540.15	0.205
26	4.035		0.0000	0.000	227.7	0.039	BB	0.041	92.44	0.035
27	4.100		0.0000	0.000	863.3	0.147	BB	0.032	453.92	0.172
28	4.267		0.0000	0.000	1503.1	0.257	BB	0.059	423.97	0.161
29	4.315		0.0000	0.000	557.5	0.095	BB	0.034	274.15	0.104
30	4.437		0.0000	0.000	9051.5	1.546	BB	0.038	3990.30	1.516
31	4.676		0.0000	0.000	787.4	0.135	BB	0.048	271.74	0.103
32	4.739		0.0000	0.000	2507.3	0.428	BB	0.022	1903.83	0.723
33	4.829		0.0000	0.000	6749.2	1.153	BB	0.103	1096.99	0.417
34	4.867		0.0000	0.000	6438.2	1.100	BB	0.023	4580.54	1.740
35	4.913		0.0000	0.000	4005.8	0.684	BB	0.020	3279.19	1.245
36	4.996		0.0000	0.000	20859.4	3.564	BB	0.082	4219.69	1.603
37	5.176		0.0000	0.000	300.7	0.051	BB	0.018	280.19	0.106
38	5.236		0.0000	0.000	935.2	0.160	BB	0.029	546.86	0.208
39	5.382		0.0000	0.000	7280.6	1.244	BB	0.032	3827.99	1.454
40	5.477		0.0000	0.000	1459.5	0.249	BB	0.064	381.15	0.145
41	5.597		0.0000	0.000	1563.4	0.267	BB	0.114	227.69	0.086
42	5.823		0.0000	0.000	7760.1	1.326	BB	0.366	353.46	0.134
43	5.903		0.0000	0.000	512.2	0.088	BB	0.038	226.98	0.086
44	6.094		0.0000	0.000	1770.1	0.302	BB	0.045	661.01	0.251
45	6.394		0.0000	0.000	48871.5	8.350	BB	0.136	5994.21	2.277
46	6.839		0.0000	0.000	793.1	0.135	BB	0.104	127.27	0.048
47	6.910		0.0000	0.000	119.6	0.020	BB	0.042	47.20	0.018
48	7.190		0.0000	0.000	21811.2	3.726	BB	0.237	1534.83	0.583
49	7.586		0.0000	0.000	4893.1	0.836	BB	0.153	533.34	0.203
50	7.732		0.0000	0.000	1242.9	0.212	BB	0.037	566.07	0.215
51	7.977	DCB	210.8018	39.305	113191.5	19.339	BB	0.036	52223.76	19.835
52	8.244		0.0000	0.000	286.0	0.049	BB	0.039	123.45	0.047
53	8.378		0.0000	0.000	3759.6	0.642	BB	0.052	1216.14	0.462
54	8.613		0.0000	0.000	1014.1	0.173	BB	0.103	164.78	0.063

ASX22. 0824

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	9.003		0.0000	0.000	345.0	0.059	BB	0.068	84.63	0.032
56	9.086		0.0000	0.000	146.3	0.025	BB	0.051	47.96	0.018

Total Area = 585304.0, Total Amount = 536.327, Total Height = 263293.9, Sample Units = ug/l

ASX22 8025

Sample Name: **BLANKA 11/3/09 RI AABPBLK65 BLK 093060018A 00173**
Acquired from CP07–H5088B via port 4 on 11/6/09 12:16:32am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.42R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.609		0.0000	0.000	8637.1	1.113	BV	0.078	1837.92	0.684
2	1.707		0.0000	0.000	2319.7	0.299	VV	0.049	783.19	0.292
3	1.880		0.0000	0.000	1236.1	0.159	VB	0.246	83.80	0.031
4	1.956		0.0000	0.000	511.3	0.066	BB	0.042	204.52	0.076
5	2.091		0.0000	0.000	32.4	0.004	BB	0.013	40.09	0.015
6	2.180		0.0000	0.000	1359.1	0.175	BB	0.068	330.70	0.123
7	2.293		0.0000	0.000	55.5	0.007	BB	0.019	47.47	0.018
8	2.342	TCX	323.8846	56.182	313514.9	40.407	BB	0.037	140463.50	52.281
9	2.495		0.0000	0.000	2773.1	0.357	BB	0.049	949.58	0.353
10	2.623		0.0000	0.000	3550.1	0.458	BB	0.034	1735.27	0.646
11	2.774		0.0000	0.000	535.7	0.069	BB	0.035	253.44	0.094
12	2.866		0.0000	0.000	227.6	0.029	BB	0.034	112.40	0.042
13	3.013		0.0000	0.000	157.9	0.020	BB	0.030	87.38	0.033
14	3.117		0.0000	0.000	155.4	0.020	BB	0.079	32.70	0.012
15	3.153		0.0000	0.000	1378.2	0.178	BB	0.064	359.29	0.134
16	3.271		0.0000	0.000	619.8	0.080	BB	0.035	294.43	0.110
17	3.348		0.0000	0.000	35480.4	4.573	BB	1.062	556.99	0.207
18	3.831		0.0000	0.000	2553.4	0.329	BB	0.385	110.61	0.041
19	3.880		0.0000	0.000	10647.1	1.372	BB	0.036	4997.34	1.860
20	4.098		0.0000	0.000	10303.1	1.328	BB	2.005	85.63	0.032
21	4.133		0.0000	0.000	8190.8	1.056	BB	0.045	3052.05	1.136
22	4.300		0.0000	0.000	9303.8	1.199	BB	0.184	843.30	0.314
23	4.546		0.0000	0.000	20545.1	2.648	BB	0.104	3299.50	1.228
24	4.602		0.0000	0.000	2655.2	0.342	BB	0.023	1913.74	0.712
25	4.687		0.0000	0.000	2858.5	0.368	BB	0.081	588.47	0.219
26	4.827		0.0000	0.000	24103.4	3.107	BB	0.103	3896.03	1.450
27	5.213		0.0000	0.000	10739.7	1.384	BB	0.219	818.88	0.305
28	5.717		0.0000	0.000	507.0	0.065	BB	0.040	208.80	0.078
29	5.799		0.0000	0.000	33863.7	4.364	BB	0.055	10236.42	3.810
30	6.294		0.0000	0.000	884.6	0.114	BB	0.042	351.33	0.131
31	6.571		0.0000	0.000	522.9	0.067	BB	0.359	24.31	0.009
32	6.623		0.0000	0.000	2583.4	0.333	BB	0.044	988.47	0.368
33	6.800	DCB	252.6066	43.818	198881.8	25.633	BB	0.038	87205.26	32.458
34	7.258		0.0000	0.000	106.4	0.014	BB	0.023	75.51	0.028
35	7.356		0.0000	0.000	58236.7	7.506	BB	0.827	1173.10	0.437
36	8.513		0.0000	0.000	3225.3	0.416	BB	0.110	487.30	0.181
37	8.647		0.0000	0.000	2639.9	0.340	BB	0.307	143.21	0.053

Total Area = 775895.9, Total Amount = 576.491, Total Height = 268671.9, Sample Units = ug/l

ASX22 0026

BLANKA 11/3/09 RI AABPBLK65 BLK 093060018A 00173

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\3P7307.42R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 12:07:23 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.42R



Instrument ID: CP07--H5088B Injected On: 11/6/2009 12:07:23 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1000
Analyst: 2280

Dilution Factor: 10

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.	152377	.32553	TCX	2.342	140463	.32388	TCX
7.977	52224	.2108	DCB	6.8	87205	.25261	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.42A
Area File: C:\CPWIN\DATA1\3P7307B.42A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 12:16:40 AM
File Reported On: 11/6/2009 at 12:16:52 AM

10X REPLOTS

ASX22 0828

C:\CPWIN\DATA1\3P7307.42R



C:\CPWIN\DATA1\3P7307B.42R



ASX22 0829

Multiple Component Data Summary

Sample Name: LCSA 11/3/09 RI A LCSBG LCS Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1000 ml Total Volume: 10 ml Analyst: 2131 SDG: State:
Analyses: 00173

Analysis Report (A)

Injected on Nov 06, 2009 00:20:23
Instrument H5088A
Result file 3P7307.43R
Calibration file 1P7307
Method file PESTD

%SSR(TCX) 110.3% Conc: 0.338812
%SSR(DCB) 52.4% Conc: 0.159935

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016						**2.22**	
3.35	3.38	3.41	48462.91015	4.633786			1
3.83	3.86	3.89	57259.26171	4.515223			2
4.10	4.13	4.16	131380.0625	4.83242			3
4.15	4.18	4.21	56949.58203	4.711842			4
4.24	4.27	4.30	53438.59765	4.6687			5
4.62	4.65	4.68	38364.5	4.702973			6
Height summation:			**385854.914062**				
Concentration		CF:	**4.677491**	L:			
Aroclor-1260						**7.23**	
5.74	5.77	5.80	77361.70312	3.78928			1
5.98	6.01	6.04	76277.42187	4.392757			2
6.05	6.08	6.11	52737.24609	3.838535			3
6.25	6.28	6.31	58131.86718	3.868654			4
6.55	6.58	6.61	116081.5859	3.620623			5
6.80	6.83	6.86	53811.16015	3.653907			6
Height summation:			**434400.984376**				
Concentration		CF:	**3.860626**	L:			

Analysis Report (B)

Injected on Nov 06, 2009 00:20:23
Instrument H5088B
Result file 3P7307B.43R
Calibration file 1P7307B
Method file PESTDB

%SSR(TCX) 109.0% Conc: 0.334784
%SSR(DCB) 66.1% Conc: 0.201677

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016						**11.66**	
2.69	2.72	2.74	37979.01562	4.924231			1
2.98	3.01	3.04	74011.97656	4.323215			2
3.14	3.17	3.20	56674.42968	4.806013			3
3.37	3.40	3.43	92337.60156	5.259652			4
3.42	3.45	3.48	20884.83984	6.102432			5
3.77	3.80	3.83	49217.62890	5.031819			6
Height summation:			**331105.492189**				
Concentration		CF:	**5.07456**	L:			
Aroclor-1260						**5.77**	
4.66	4.69	4.72	95739.82812	5.099999			1
4.84	4.87	4.90	107570.3203	5.417653			2
4.96	4.99	5.02	79385.15625	5.401652			3
5.36	5.39	5.42	77222.46093	4.778316			4
5.60	5.63	5.66	164617.8906	5.1257			5
5.89	5.92	5.95	102449.9453	4.730408			6
Height summation:			**626985.601564**				
Concentration		CF:	**5.092288**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016	B	5.07456	0.5	0.1		8.14	4	15	
Aroclor-1221			<0.5	<0.1			3	5	
Aroclor-1232			<0.5	<0.1			4	10	
Aroclor-1242			<0.5	<0.1	E		4	30	
Aroclor-1248			<0.5	<0.1			4	30	
Aroclor-1254			<0.5	<0.1			4	40	
Aroclor-1260	B	5.092288	0.5	0.1		27.51	4	40	

Units: ug/l

%Difference = High - Low divided by the Average times 100

Reviewed By: James Bailhart Date: 11-6-09
Verified By: [illegible] Date: 11/9/09

ASX22: 0030

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **LCSA** 11/3/09 RI A LCSBG Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1000 ml Total Volume: 10 ml Analyst: 2280 SDG: State:
Analyses: 00173 00516 06030 07394 10227

Analysis Report (A)

Injected on : NOV 06, 2009 00:20:23
Instrument : CP07--H5088A
Result file : 3P7307.43R
Calibration file : 1P7307.CAL
Method file : PESTD.MET
%SSR(TCX) : 110.3% Conc.: 0.338812
%SSR(DCB) : 52.4% Conc.: 0.159935

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	2.22	
3.35	3.38	3.41	48462.91015	4.633786			1
3.83	3.86	3.89	57259.26171	4.515223			2
4.10	4.13	4.16	131380.0625	4.83242			3
4.15	4.18	4.21	56949.58203	4.711842			4
4.24	4.27	4.30	53438.59765	4.6687			5
4.62	4.65	4.68	38364.5	4.702973			6
Height Summation:			385854.914062				
Amount Avg CF:			4.677491	Linear:			
Aroclor-1221					3	35.73	
3.20	3.23	3.26	9672.710938	1.556976			1
E 3.28	3.31	3.34	11461.74902	2.615345			2
E 3.35	3.38	3.41	48462.91015	3.334107			3
Height Summation:			69597.370117				
Amount Avg CF:			2.502143	Linear:			
Aroclor-1232					6	28.08	
E 3.35	3.38	3.41	48462.91015	3.947938			1
E 3.83	3.86	3.89	57259.26171	10.147853			2
E 4.10	4.13	4.16	131380.0625	10.096063			3
E 4.15	4.18	4.21	56949.58203	9.845204			4
E 4.23	4.27	4.29	53438.59765	9.955029			5
E 4.62	4.65	4.68	38364.5	10.829781			6
Height Summation:			385854.914062				
Amount Avg CF:			9.136978	Linear:			
Aroclor-1242					6	1.46	
E 3.35	3.38	3.41	48462.91015	5.419392			1
E 3.83	3.86	3.89	57259.26171	5.525889			2
E 4.09	4.13	4.15	131380.0625	5.586956			3
E 4.15	4.18	4.21	56949.58203	5.509028			4
E 4.23	4.27	4.29	53438.59765	5.443198			5
E 4.62	4.65	4.68	38364.5	5.36679			6
Height Summation:			385854.914062				
Amount Avg CF:			5.475209	Linear:			
Aroclor-1248					6	66.52	
E 4.10	4.13	4.16	131380.0625	8.762526			1
4.33	4.36	4.39	33782.23828	3.237438			2
4.50	4.53	4.56	52621.5625	3.366901			3
4.62	4.65	4.68	38364.5	3.716014			4
5.02	5.06	5.08	35814.20312	2.864339			5
5.61	5.64	5.67	7009.011719	1.178877			6
Height Summation:			298971.578125				
Amount Avg CF:			3.854349	Linear:			
Aroclor-1254					6	58.86	
4.83	4.86	4.89	44384.90234	3.697149			1
5.03	5.06	5.09	35814.20312	1.973328			2
5.27	5.31	5.33	17654.77929	1.336362			3
5.34	5.38	5.40	15970.94921	0.674066			4
5.85	5.88	5.91	39829.66406	3.256974			5
5.98	6.01	6.04	76277.42187	4.709262			6
Height Summation:			229931.919923				
Amount Avg CF:			2.607857	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 00:20:23
Instrument : CP07--H5088B
Result file : 3P7307B.43R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET A linear
%SSR(TCX) : 109% Conc.: 0.334784
%SSR(DCB) : 66.1% Conc.: 0.201677

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	11.66	
2.69	2.72	2.75	37979.01562	4.924231			1
2.98	3.01	3.04	74011.97656	4.323215			2
3.14	3.17	3.20	56674.42968	4.806013			3
3.37	3.40	3.43	92337.60156	5.259652			4
3.42	3.45	3.48	20884.83984	6.102432			5
3.77	3.80	3.83	49217.62890	5.031819			6
Height Summation:			331105.492189				
Amount Avg CF:			5.07456	Linear:			
Aroclor-1221					3	35.92	
2.53	2.56	2.59	7274.504883	1.657251			1
E 2.61	2.65	2.67	7783.122559	2.742951			2
E 2.69	2.72	2.75	37979.01562	3.555921			3
Height Summation:			53036.643067				
Amount Avg CF:			2.652041	Linear:			
Aroclor-1232					6	32.35	
E 2.69	2.72	2.75	37979.01562	4.30284			1
E 2.98	3.01	3.04	74011.97656	9.535783			2
E 3.14	3.17	3.20	56674.42968	10.27269			3
E 3.38	3.40	3.44	92337.60156	12.767038			4
E 3.42	3.45	3.48	20884.83984	13.543995			5
E 3.77	3.80	3.83	49217.62890	12.378679			6
Height Summation:			331105.492189				
Amount Avg CF:			10.466838	Linear:			
Aroclor-1242					6	9.08	
E 2.68	2.72	2.74	37979.01562	5.931904			1
E 2.98	3.01	3.04	74011.97656	5.306279			2
E 3.14	3.17	3.20	56674.42968	5.919619			3
E 3.37	3.40	3.43	92337.60156	6.805292			4
E 3.42	3.45	3.48	20884.83984	6.699227			5
E 3.77	3.80	3.83	49217.62890	6.067636			6
Height Summation:			331105.492189				
Amount Avg CF:			6.12166	Linear:			
Aroclor-1248					5	62.01	
E 3.14	3.17	3.20	56674.42968	13.446274			1
3.37	3.40	3.43	92337.60156	7.260832			2
3.77	3.80	3.83	49217.62890	3.678871			3
3.89	3.92	3.95	16539.70507	2.641163			4
4.55	4.60	4.61	43930.56640	7.060272			6
Height Summation:			258699.931641				
Amount Avg CF:			6.817482	Linear:			
Aroclor-1254					6	87.60	
4.08	4.11	4.14	50808.75	2.722228			1
4.55	4.60	4.61	43930.56640	2.111811			2
4.66	4.69	4.72	95739.82812	8.118664			3
4.76	4.82	4.82	16131.99121	1.211094			4
E 4.85	4.87	4.91	107570.3203	12.22604			5
5.11	5.14	5.17	57420.15234	3.172531			6
Height Summation:			371601.608399				
Amount Avg CF:			4.927061	Linear:			

ASX22 8831

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:09:30

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: LCSA 11/3/09 RI A LCSBG Sample ID: AB **Batchnumber: 093060018A**
Sample Amount: 1000 ml Total Volume: 10 ml Analyst: 2280 SDG: State:
Analyses: 00173 00516 06030 07394 10227

Analysis Report (A)

Injected on : NOV 06, 2009 00:20:23
Instrument : CP07--H5088A
Result file : 3P7307.43R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	7.23	
5.74	5.77	5.80	77361.70312	3.78928			1
5.98	6.01	6.04	76277.42187	4.392757			2
6.05	6.08	6.11	52737.24609	3.838535			3
6.25	6.28	6.31	58131.86718	3.868654			4
6.55	6.58	6.61	116081.5859	3.620623			5
6.80	6.83	6.86	53811.16015	3.653907			6
Height Summation:			**434400.984376**				
Amount Avg CF:			**3.860626**	Linear:			
Aroclor-1262					6	25.94	
5.49	5.52	5.55	65341.17187	3.927253			1
5.73	5.77	5.79	77361.70312	4.463549			2
6.05	6.08	6.11	52737.24609	2.299294			3
6.25	6.28	6.31	58131.86718	3.257651			4
6.55	6.58	6.61	116081.5859	2.833092			5
6.79	6.83	6.85	53811.16015	2.551204			6
Height Summation:			**423464.734376**				
Amount Avg CF:			**3.222007**	Linear:			
Aroclor-1268					5	103.25	
6.05	6.08	6.11	52737.24609	5.811032			1
6.26	6.28	6.32	58131.86718	5.020136			2
6.79	6.83	6.85	53811.16015	1.287193			3
6.87	6.89	6.93	24584.95312	0.68085			4
7.03	7.06	7.09	1417.826416	0.040479			5
Height Summation:			**190683.052979**				
Amount Avg CF:			**2.567938**	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 00:20:23
Instrument : CP07--H5088B
Result file : 3P7307B.43R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	5.77	
4.66	4.69	4.72	95739.82812	5.099999			1
4.84	4.87	4.90	107570.3203	5.417653			2
4.96	4.99	5.02	79385.15625	5.401652			3
5.36	5.39	5.42	77222.46093	4.778316			4
5.60	5.63	5.66	164617.8906	5.1257			5
5.89	5.92	5.95	102449.9453	4.730408			6
Height Summation:			**626985.601564**				
Amount Avg CF:			**5.092288**	Linear:			
Aroclor-1262					6	35.10	
4.84	4.87	4.90	107570.3203	5.934446			1
5.36	5.39	5.42	77222.46093	3.419744			2
5.60	5.63	5.66	164617.8906	3.960996			3
5.84	5.87	5.90	31387.375	2.091424			4
5.89	5.92	5.95	102449.9453	3.891547			5
6.35	6.38	6.41	49444.48046	2.887994			6
Height Summation:			**532692.472658**				
Amount Avg CF:			**3.697692**	Linear:			
Aroclor-1268					6	103.33	
5.35	5.39	5.41	77222.46093	5.683714			1
5.84	5.87	5.90	31387.375	0.721119			2
5.90	5.92	5.96	102449.9453	2.683852			3
6.06	6.09	6.12	2156.882813	0.051765			4
6.14	6.17	6.20	3844.271484	0.393838			5
6.36	6.38	6.42	49444.48046	2.850999			6
Height Summation:			**266505.416017**				
Amount Avg CF:			**2.064215**	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			0.5	0.1		8.14	4	15	
Aroclor-1221			0.5	0.1	E	5.82	3	5	
Aroclor-1232			0.5	0.1	E	13.57	4	10	
			0.5	0.2	E				
Aroclor-1242			0.5	0.1	E	11.15	4	30	
Aroclor-1248			0.5	0.1	E	** 55.53	4	30	
Aroclor-1254			0.5	0.1		** 61.56	4	40	
Aroclor-1260			0.5	0.1		27.51	4	40	
Aroclor-1262			0.5	0.1		13.75	4	40	
			0.5	0.2					
Aroclor-1268			0.5	0.1		21.75	4	40	

Units: ug/L

ASX22 8832

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:09:30

Sample Name: **LCSA 11/3/09 RI AABLCSBG LCS 093060018A 00173**
Acquired from CP07--H5088A via port 3 on 11/6/09 12:29:31am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.43R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.579		0.0000	0.000	2556.3	0.078	BB	0.039	1079.81	0.059
2	1.710		0.0000	0.000	12341.8	0.376	BV	0.088	2329.01	0.127
3	1.819		0.0000	0.000	17531.9	0.533	VV	0.062	4743.81	0.258
4	1.942		0.0000	0.000	7220.0	0.220	VV	0.066	1834.08	0.100
5	1.981		0.0000	0.000	22659.8	0.689	VV	0.033	11453.65	0.623
6	2.027		0.0000	0.000	8103.3	0.247	VV	0.042	3245.47	0.177
7	2.078		0.0000	0.000	11029.2	0.336	VV	0.060	3059.29	0.166
8	2.176		0.0000	0.000	10478.7	0.319	VV	0.064	2736.86	0.149
9	2.290		0.0000	0.000	69061.7	2.101	VV	0.029	39788.07	2.164
10	2.451		0.0000	0.000	1920.6	0.058	VV	0.064	500.65	0.027
11	2.518		0.0000	0.000	3082.0	0.094	VV	0.045	1137.66	0.062
12	2.580		0.0000	0.000	710.5	0.022	VB	0.036	325.97	0.018
13	2.699		0.0000	0.000	263.9	0.008	BB	0.024	181.68	0.010
14	2.744		0.0000	0.000	2316.4	0.070	BB	0.032	1198.24	0.065
15	2.813		0.0000	0.000	121.3	0.004	BB	0.019	107.74	0.006
16	2.940		0.0000	0.000	7836.1	0.238	BB	0.023	5643.75	0.307
17	3.001	TCX	338.8116	67.933	264425.0	8.046	BB	0.028	157802.00	8.584
18	3.175		0.0000	0.000	617.7	0.019	BB	0.021	484.46	0.026
19	3.228		0.0000	0.000	15652.8	0.476	BB	0.027	9672.71	0.526
20	3.311		0.0000	0.000	20205.9	0.615	BB	0.029	11461.75	0.623
21	3.381		0.0000	0.000	78025.4	2.374	BB	0.027	48462.91	2.636
22	3.473		0.0000	0.000	100.5	0.003	BB	0.014	120.42	0.007
23	3.514		0.0000	0.000	11931.2	0.363	BB	0.026	7749.00	0.422
24	3.639		0.0000	0.000	560.1	0.017	BB	0.028	338.42	0.018
25	3.694		0.0000	0.000	156799.1	4.771	BB	0.040	65935.27	3.587
26	3.772		0.0000	0.000	2927.7	0.089	BB	0.016	3138.98	0.171
27	3.796		0.0000	0.000	2179.8	0.066	BB	0.013	2706.75	0.147
28	3.823		0.0000	0.000	1377.5	0.042	BB	0.012	1866.41	0.102
29	3.863		0.0000	0.000	90743.5	2.761	BB	0.026	57259.26	3.115
30	3.966		0.0000	0.000	8331.9	0.254	BB	0.028	4908.56	0.267
31	4.031		0.0000	0.000	52684.5	1.603	BB	0.043	20358.94	1.107
32	4.126		0.0000	0.000	261752.9	7.965	BB	0.033	131380.10	7.147
33	4.181		0.0000	0.000	81424.9	2.478	BB	0.024	56949.58	3.098
34	4.265		0.0000	0.000	85773.6	2.610	BB	0.027	53438.60	2.907
35	4.315		0.0000	0.000	2972.6	0.090	BB	0.016	3115.19	0.169
36	4.360		0.0000	0.000	39785.3	1.211	BB	0.020	33782.24	1.838
37	4.396		0.0000	0.000	70219.3	2.137	BB	0.034	34212.23	1.861
38	4.528		0.0000	0.000	114527.9	3.485	BB	0.036	52621.56	2.862
39	4.599		0.0000	0.000	10310.1	0.314	BB	0.017	10071.17	0.548
40	4.649		0.0000	0.000	64573.9	1.965	BB	0.028	38364.50	2.087
41	4.698		0.0000	0.000	18412.9	0.560	BB	0.029	10634.53	0.578
42	4.817		0.0000	0.000	5636.2	0.171	BB	0.036	2590.02	0.141
43	4.857		0.0000	0.000	76655.1	2.332	BB	0.029	44384.90	2.414
44	4.913		0.0000	0.000	10147.5	0.309	BB	0.016	10640.40	0.579
45	5.003		0.0000	0.000	13235.1	0.403	BB	0.036	6089.66	0.331
46	5.058		0.0000	0.000	63095.6	1.920	BB	0.029	35814.20	1.948
47	5.232		0.0000	0.000	1722.2	0.052	BB	0.026	1105.63	0.060
48	5.310		0.0000	0.000	32192.9	0.980	BB	0.030	17654.78	0.960
49	5.377		0.0000	0.000	23260.5	0.708	BB	0.024	15970.95	0.869
50	5.428		0.0000	0.000	82140.6	2.499	BB	0.037	37254.26	2.027
51	5.521		0.0000	0.000	103223.3	3.141	BB	0.026	65341.17	3.554
52	5.636		0.0000	0.000	16459.5	0.501	BB	0.039	7009.01	0.381
53	5.704		0.0000	0.000	13322.0	0.405	BB	0.024	9075.46	0.494
54	5.767		0.0000	0.000	130645.6	3.975	BB	0.028	77361.70	4.208

ASX22 0833

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.811		0.0000	0.000	9199.3	0.280	BB	0.018	8549.61	0.465
56	5.876		0.0000	0.000	67260.5	2.047	BB	0.028	39829.66	2.167
57	5.968		0.0000	0.000	4749.6	0.145	BB	0.021	3818.83	0.208
58	6.011		0.0000	0.000	131125.3	3.990	BB	0.029	76277.42	4.149
59	6.083		0.0000	0.000	75270.4	2.290	BB	0.024	52737.25	2.869
60	6.135		0.0000	0.000	36471.4	1.110	BB	0.024	25618.45	1.394
61	6.279		0.0000	0.000	121084.0	3.684	BB	0.035	58131.87	3.162
62	6.352		0.0000	0.000	34967.8	1.064	BB	0.025	23488.78	1.278
63	6.399		0.0000	0.000	15025.2	0.457	BB	0.022	11454.37	0.623
64	6.445		0.0000	0.000	1289.2	0.039	BB	0.019	1135.76	0.062
65	6.511		0.0000	0.000	22969.1	0.699	BB	0.026	14592.89	0.794
66	6.578		0.0000	0.000	203361.0	6.188	BB	0.029	116081.60	6.315
67	6.832		0.0000	0.000	93654.4	2.850	BB	0.029	53811.16	2.927
68	6.892		0.0000	0.000	52780.0	1.606	BB	0.036	24584.95	1.337
69	7.057		0.0000	0.000	2361.6	0.072	BB	0.028	1417.83	0.077
70	7.182		0.0000	0.000	35806.9	1.090	BB	0.041	14704.02	0.800
71	7.415		0.0000	0.000	57120.4	1.738	BB	0.031	30940.21	1.683
72	7.488		0.0000	0.000	2264.8	0.069	BB	0.027	1421.85	0.077
73	7.586		0.0000	0.000	1010.5	0.031	BB	0.082	205.97	0.011
74	7.731		0.0000	0.000	18336.5	0.558	BB	0.034	9090.95	0.495
75	7.976	DCB	159.9350	32.067	90374.5	2.750	BB	0.036	41921.95	2.280
76	8.247		0.0000	0.000	432.7	0.013	BB	0.040	179.10	0.010
77	8.313		0.0000	0.000	114.3	0.003	BB	0.032	58.72	0.003
78	8.375		0.0000	0.000	4154.6	0.126	BB	0.039	1778.46	0.097

Total Area = 3286469.0, Total Amount = 498.747, Total Height = 1838325.0, Sample Units = ug/l

ASX22. 0834

Sample Name: **LCSA 11/3/09 RI AABLCSBG LCS 093060018A 00173**
Acquired from CP07--H5088B via port 4 on 11/6/09 12:29:31am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.43R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.539		0.0000	0.000	20588.6	0.508	BV	0.051	6701.91	0.363
2	1.700		0.0000	0.000	425.4	0.011	VB	0.028	248.96	0.013
3	1.769		0.0000	0.000	68047.2	1.680	BB	0.035	32184.31	1.743
4	1.961		0.0000	0.000	1829.0	0.045	BB	0.043	715.81	0.039
5	2.077		0.0000	0.000	97.5	0.002	BB	0.031	53.17	0.003
6	2.119		0.0000	0.000	1859.6	0.046	BB	0.030	1038.79	0.056
7	2.253		0.0000	0.000	202.0	0.005	BB	0.026	127.42	0.007
8	2.342	TCX	334.7836	62.406	337363.0	8.329	BB	0.039	145190.20	7.865
9	2.494		0.0000	0.000	4424.5	0.109	BB	0.034	2173.50	0.118
10	2.563		0.0000	0.000	13962.2	0.345	BB	0.032	7274.50	0.394
11	2.645		0.0000	0.000	15843.5	0.391	BB	0.034	7783.12	0.422
12	2.715		0.0000	0.000	94104.6	2.323	BB	0.041	37979.02	2.057
13	2.860		0.0000	0.000	15845.4	0.391	BB	0.037	7224.98	0.391
14	3.007		0.0000	0.000	165017.9	4.074	BB	0.037	74011.98	4.009
15	3.067		0.0000	0.000	5260.0	0.130	BB	0.027	3288.94	0.178
16	3.170		0.0000	0.000	138193.9	3.412	BB	0.041	56674.43	3.070
17	3.310		0.0000	0.000	53179.6	1.313	BB	0.042	20939.61	1.134
18	3.403		0.0000	0.000	181286.6	4.476	BB	0.033	92337.60	5.002
19	3.448		0.0000	0.000	33263.3	0.821	BB	0.027	20884.84	1.131
20	3.557		0.0000	0.000	68947.0	1.702	BB	0.030	38474.40	2.084
21	3.618		0.0000	0.000	13140.5	0.324	BB	0.023	9409.55	0.510
22	3.643		0.0000	0.000	57897.1	1.429	BB	0.038	25212.84	1.366
23	3.797		0.0000	0.000	146118.9	3.608	BB	0.049	49217.63	2.666
24	3.883		0.0000	0.000	2489.8	0.061	BB	0.012	3443.80	0.187
25	3.924		0.0000	0.000	26531.2	0.655	BB	0.027	16539.71	0.896
26	3.972		0.0000	0.000	10306.5	0.254	BB	0.024	7174.32	0.389
27	4.049		0.0000	0.000	495.8	0.012	BB	0.023	366.42	0.020
28	4.108		0.0000	0.000	150994.6	3.728	BB	0.050	50808.75	2.752
29	4.200		0.0000	0.000	1891.8	0.047	BB	0.018	1734.80	0.094
30	4.250		0.0000	0.000	98731.0	2.438	BB	0.045	36254.79	1.964
31	4.504		0.0000	0.000	4258.7	0.105	BB	0.024	2971.55	0.161
32	4.541		0.0000	0.000	17345.9	0.428	BB	0.023	12578.19	0.681
33	4.598		0.0000	0.000	131691.1	3.251	BB	0.050	43930.57	2.380
34	4.692		0.0000	0.000	188250.2	4.648	BB	0.033	95739.83	5.186
35	4.821		0.0000	0.000	31548.6	0.779	BB	0.033	16131.99	0.874
36	4.872		0.0000	0.000	214088.8	5.286	BB	0.033	107570.30	5.827
37	4.985		0.0000	0.000	252143.8	6.225	BB	0.053	79385.16	4.300
38	5.099		0.0000	0.000	15788.8	0.390	BB	0.023	11573.65	0.627
39	5.141		0.0000	0.000	78178.4	1.930	BB	0.023	57420.15	3.110
40	5.175		0.0000	0.000	44133.1	1.090	BB	0.019	38944.18	2.110
41	5.219		0.0000	0.000	28659.1	0.708	BB	0.022	21954.67	1.189
42	5.336		0.0000	0.000	16579.3	0.409	BB	0.025	11032.61	0.598
43	5.389		0.0000	0.000	156881.3	3.873	BB	0.034	77222.46	4.183
44	5.473		0.0000	0.000	41051.1	1.014	BB	0.025	27851.88	1.509
45	5.509		0.0000	0.000	8303.1	0.205	BB	0.019	7198.81	0.390
46	5.561		0.0000	0.000	9031.7	0.223	BB	0.021	7079.15	0.383
47	5.626		0.0000	0.000	368828.8	9.106	BB	0.037	164617.90	8.917
48	5.798		0.0000	0.000	16279.8	0.402	BB	0.027	10075.08	0.546
49	5.867		0.0000	0.000	45984.1	1.135	BB	0.024	31387.38	1.700
50	5.921		0.0000	0.000	223856.9	5.527	BB	0.036	102450.00	5.550
51	6.088		0.0000	0.000	4228.5	0.104	BB	0.033	2156.88	0.117
52	6.173		0.0000	0.000	7022.9	0.173	BB	0.030	3844.27	0.208
53	6.238		0.0000	0.000	43808.1	1.082	BB	0.036	20555.35	1.113
54	6.382		0.0000	0.000	118550.4	2.927	BB	0.040	49444.48	2.678

ASX22 8835

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	6.621		0.0000	0.000	33708.3	0.832	BB	0.037	15020.74	0.814
56	6.800	DCB	201.6767	37.594	161364.2	3.984	BB	0.039	69623.16	3.772
57	7.253		0.0000	0.000	154.7	0.004	BB	0.027	95.07	0.005
58	7.356		0.0000	0.000	11496.0	0.284	BB	0.126	1523.07	0.083
59	7.723		0.0000	0.000	26980.5	0.666	BB	3.063	146.81	0.008
60	8.507		0.0000	0.000	21800.4	0.538	BB	0.352	1033.60	0.056

Total Area = 4050335.0, Total Amount = 536.46, Total Height = 1846029.0, Sample Units = ug/l

0

ASX22 0836

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\3P7307.43R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 12:20:22 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.43R



ASX22 0037

Instrument ID: CP07--H5088B Injected On: 11/6/2009 12:20:22 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1000
Analyst: 2280

Dilution Factor: 10

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.001	157802	.33881	TCX	2.342	145190	.33478	TCX
7.976	41922	.15993	DCB	6.8	69623	.20168	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.43A
Area File: C:\CPWIN\DATA1\3P7307B.43A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 12:29:38 AM
File Reported On: 11/6/2009 at 12:29:51 AM

10X REPLOTS

ASX22 9838

C:\CPWIN\DATA1\3P7307.43R



C:\CPWIN\DATA1\3P7307B.43R



ASX22 8839

Multiple Component Data Summary

Sample Name: 5822065 MS RI AF DF20 CARS2MS MS Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 1020 ml Total Volume: 200 ml Analyst: 2131 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on Nov 06, 2009 01:12:28
Instrument H5088A
Result file 3P7307.47R
Calibration file 1P7307
Method file PESTD

%SSR(TCX) * 0.0% Conc: 0
%SSR(DCB) * 0.0% Conc: 0

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016						**46.26**	
3.35	3.38	3.41	2707.552734	5.076135			1
3.83	3.86	3.89	5808.385254	8.980876			2
4.10	4.13	4.16	17004.23242:	12.263715			3
4.15	4.18	4.21	7108.855957	11.532666			4
4.24	4.27	4.30	4860.603027	8.32647			5
4.62	4.65	4.68	8349.301758	20.068869			6
Height summation:			**45838.931152**				
Concentration		CF:	**11.041455**	L:			
Aroclor-1248						**22.29**	
4.10	4.13	4.16	17004.23242:	22.237539			1
4.33	4.36	4.39	12172.32324:	22.872643			2
4.50	4.53	4.56	11842.29687!	14.857042			3
Height summation:			**41018.852539**				
Concentration		CF:	**19.989075**	L:			
Aroclor-1254						**13.86**	
5.03	5.06	5.09	57018.32421!	61.601065			2
5.27	5.30	5.33	31728.42187!	47.091223			3
5.34	5.37	5.40	71357.24218!	59.0527			4
Height summation:			**160103.988282**				
Concentration		CF:	**55.914996**	L:			
Aroclor-1260						**14.56**	
6.25	6.27	6.31	41126.85156:	53.666209			4
6.55	6.58	6.61	66328.23437!	40.564695			5
6.80	6.83	6.86	33387.13671!	44.452288			6
Height summation:			**140842.222657**				
Concentration		CF:	**46.227731**	L:			

Analysis Report (B)

Injected on Nov 06, 2009 01:12:28
Instrument H5088B
Result file 3P7307B.47R
Calibration file 1P7307B
Method file PESTDB

%SSR(TCX) 98.0% Conc: 0.294908
%SSR(DCB) 86.8% Conc: 0.259407

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016						**41.48**	
2.69	2.74	2.74	7057.592773	17.942424			1
2.98	3.01	3.04	6997.712402	8.014765			2
3.14	3.17	3.20	5461.371094	9.080909			3
3.37	3.40	3.43	10357.27832	11.567882			4
3.42	3.46	3.48	1554.839233	8.908141			5
3.77	3.80	3.83	10187.22070:	20.421604			6
Height summation:			**41616.014525**				
Concentration		CF:	**12.655954**	L:			
Aroclor-1248						**30.75**	
3.14	3.17	3.20	5461.371094	25.406585			1
3.37	3.40	3.43	10357.27832	15.969205			2
3.77	3.80	3.83	10187.22070:	14.930674			3
Height summation:			**26005.870117**				
Concentration		CF:	**18.768821**	L:			
Aroclor-1254						**40.19**	
4.55	4.59	4.61	68860.21093!	64.906232			2
4.66	4.69	4.72	66709.75	110.920302			3
4.76	4.79	4.82	103923.84375	152.980097			4
Height summation:			**239493.804688**				
Concentration		CF:	**109.60221**	L:			
Aroclor-1260						**8.28**	
5.36	5.41	5.42	43885.21484	53.245056			4
5.60	5.62	5.66	102760.47656	62.738199			5
5.89	5.92	5.95	63017.60546!	57.053013			6
Height summation:			**209663.296876**				
Concentration		CF:	**57.678756**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016	B	12.655954	9.8039	1.9608		13.63	4	15	
Aroclor-1221			<9.8039	<1.9608			3	5	
Aroclor-1232			<9.8039	<1.9608			4	10	
Aroclor-1242			<9.8039	<1.9608			4	30	
Aroclor-1248	A	19.989075	9.8039	1.9608		6.30	4	30	
Aroclor-1254	A	55.914996	9.8039	1.9608		64.87	4	40	

ASX22 0848

Multiple Component Data Summary

Sample Name: 5822065 MS RI AF DF20 CARS2MS MS Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 1020 ml Total Volume: 200 ml Analyst: 2131 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on	Nov 06, 2009 01:12:28
Instrument	H5088A
Result file	3P7307.47R
Calibration file	1P7307
Method file	PESTD

Analysis Report (B)

Injected on	Nov 06, 2009 01:12:28
Instrument	H5088B
Result file	3P7307B.47R
Calibration file	1P7307B
Method file	PESTDB

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1260	B	57.678756	9.8039	1.9608		22.04	4	40	

Units: ug/l

%Difference = High - Low divided by the Average times 100

Reviewed By: [signature] Date: 11-6-09

Verified By: [signature] Date: 11/9/09

ASX22 0841

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822065** MS RI AF DF20 CARS2MS Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 1020 ml Total Volume: 200 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173 00516 06030 07394 10227

Analysis Report (A)

Injected on : NOV 06, 2009 01:12:28
Instrument : CP07--H5088A
Result file : 3P7307.47R
Calibration file : 1P7307.CAL
Method file : PESTD.MET
%SSR(TCX) : *-159.2% Conc.: -0.479143
%SSR(DCB) : *-233.8% Conc.: -0.698877

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	46.26	
3.35	3.38	3.41	2707.552734	5.076135			1
3.83	3.86	3.89	5808.385254	8.980876			2
4.10	4.13	4.16	17004.23242:	12.263715			3
4.15	4.18	4.21	7108.855957	11.532666			4
4.24	4.27	4.30	4860.603027	8.32647			5
4.62	4.65	4.68	8349.301758	20.068869			6
Height Summation:			45838.931152				
Amount Avg CF:			11.041455	Linear:			
Aroclor-1221					3	76.43	
3.20	3.23	3.26	408.769073	1.290154			1
3.28	3.28	3.34	1704.426758	7.625815			2
3.35	3.38	3.41	2707.552734	3.652387			3
Height Summation:			4820.748565				
Amount Avg CF:			4.189452	Linear:			
Aroclor-1232					6	59.25	
3.35	3.38	3.41	2707.552734	4.324815			1
3.83	3.86	3.89	5808.385254	20.184299			2
4.10	4.13	4.16	17004.23242:	25.621788			3
4.15	4.18	4.21	7108.855957	24.097041			4
4.23	4.27	4.29	4860.603027	17.75446			5
E 4.62	4.65	4.68	8349.301758	46.213634			6
Height Summation:			45838.931152				
Amount Avg CF:			23.032673	Linear:			
Aroclor-1242					6	44.56	
3.35	3.38	3.41	2707.552734	5.936737			1
3.83	3.86	3.89	5808.385254	10.991112			2
4.09	4.13	4.15	17004.23242:	14.178577			3
4.15	4.18	4.21	7108.855957	13.483853			4
4.23	4.27	4.29	4860.603027	9.707761			5
4.62	4.65	4.68	8349.301758	22.90156			6
Height Summation:			45838.931152				
Amount Avg CF:			12.8666	Linear:			
Aroclor-1248					6	144.99	
4.10	4.13	4.16	17004.23242:	22.237539			1
4.33	4.36	4.39	12172.32324:	22.872643			2
4.50	4.53	4.56	11842.29687!	14.857042			3
+ 4.50	4.56	4.56	2535.299805	3.180722			3
4.62	4.65	4.68	8349.301758	15.857247			4
5.02	5.06	5.08	57018.32421!	89.415615			5
E 5.61	5.64	5.67	90587.45312!	298.751084			6
Height Summation:			196973.931641				
Amount Avg CF:			77.331862	Linear:			
Aroclor-1254					6	50.02	
4.83	4.86	4.89	26306.96875	42.966733			1
+ 4.83	4.89	4.89	15074.22656:	24.620483			1
5.03	5.06	5.09	57018.32421!	61.601065			2
5.27	5.30	5.33	31728.42187!	47.091223			3
5.34	5.37	5.40	71357.24218!	59.0527			4
5.85	5.88	5.91	84758.33593!	135.900132			5
5.98	6.01	6.04	92605.92968!	112.105123			6
Height Summation:			363775.222658				
Amount Avg CF:			76.452829	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 01:12:28
Instrument : CP07--H5088B
Result file : 3P7307B.47R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET A unclear
%SSR(TCX) : 98% Conc.: 0.294908
%SSR(DCB) : 86.8% Conc.: 0.259407

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	41.48	
2.69	2.74	2.75	7057.592773	17.942424			1
2.98	3.01	3.04	6997.712402	8.014765			2
3.14	3.17	3.20	5461.371094	9.080909			3
3.37	3.40	3.43	10357.27832	11.567882			4
3.42	3.46	3.48	1554.839233	8.908141			5
3.77	3.80	3.83	10187.22070:	20.421604			6
Height Summation:			41616.014525				
Amount Avg CF:			12.655954	Linear:			
Aroclor-1221					1		
2.69	2.74	2.75	7057.592773	12.956713			3
Height Summation:			7057.592773				
Amount Avg CF:			12.956713	Linear:			
Aroclor-1232					6	51.72	
2.69	2.74	2.75	7057.592773	15.678261			1
2.98	3.01	3.04	6997.712402	17.67829			2
3.14	3.17	3.20	5461.371094	19.410134			3
3.38	3.40	3.44	10357.27832	28.079349			4
3.42	3.46	3.48	1554.839233	19.771105			5
E 3.77	3.80	3.83	10187.22070:	50.238789			6
Height Summation:			41616.014525				
Amount Avg CF:			25.142655	Linear:			
Aroclor-1242					6	41.67	
2.68	2.74	2.74	7057.592773	21.614084			1
2.98	3.01	3.04	6997.712402	9.837256			2
3.14	3.17	3.20	5461.371094	11.185054			3
3.37	3.40	3.43	10357.27832	14.967306			4
3.42	3.46	3.48	1554.839233	9.779324			5
3.77	3.80	3.83	10187.22070:	24.625461			6
Height Summation:			41616.014525				
Amount Avg CF:			15.334748	Linear:			
Aroclor-1248					6	132.61	
3.14	3.17	3.20	5461.371094	25.406585			1
3.37	3.40	3.43	10357.27832	15.969205			2
3.77	3.80	3.83	10187.22070:	14.930674			3
3.89	3.92	3.95	3681.585449	11.527423			4
4.26	4.29	4.32	42091.97656:	80.365352			5
E 4.55	4.59	4.61	68860.21093!	216.996541			6
Height Summation:			140639.643067				
Amount Avg CF:			60.865963	Linear:			
Aroclor-1254					6	52.38	
4.08	4.11	4.14	28854.29296!	30.312812			1
+ 4.08	4.14	4.14	9131.191406	9.592752			1
4.55	4.59	4.61	68860.21093!	64.906232			2
4.66	4.69	4.72	66709.75	110.920302			3
4.76	4.79	4.82	103923.8437!	152.980097			4
E 4.85	4.87	4.91	87839.53125	195.755138			5
5.11	5.14	5.17	120490.375	130.534034			6
Height Summation:			476678.003907				
Amount Avg CF:			114.234769	Linear:			

ASX22 0842

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:16:56

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822065** MS RI AF DF20 CARS2MS Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 1020 ml Total Volume: 200 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173 00516 06030 07394 10227

Analysis Report (A)

Injected on : NOV 06, 2009 01:12:28
Instrument : CP07--H5088A
Result file : 3P7307.47R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	46.27	
5.74	5.77	5.80	72449.953125	69.582263			1
5.98	6.01	6.04	92605.929688	104.570658			2
6.05	6.08	6.11	21910.49804	31.270161			3
6.25	6.27	6.31	41126.85156	53.666209			4
6.55	6.58	6.61	66328.234375	40.564695			5
6.80	6.83	6.86	33387.136715	44.452288			6
Height Summation:			**327808.603517**				
Amount Avg CF:			**57.351046**	Linear:			
Aroclor-1262					6	51.49	
5.49	5.52	5.55	39100.730465	46.08045			1
5.73	5.77	5.79	72449.953125	81.963833			2
6.05	6.08	6.11	21910.49804	18.730919			3
6.25	6.27	6.31	41126.85156	45.190334			4
6.55	6.58	6.61	66328.234375	31.741367			5
6.79	6.83	6.85	33387.136715	31.037145			6
Height Summation:			**274303.404298**				
Amount Avg CF:			**42.457341**	Linear:			
Aroclor-1268					5	105.12	
6.05	6.08	6.11	21910.49804	47.33887			1
6.26	6.27	6.32	41126.85156	69.639634			2
6.79	6.83	6.85	33387.136715	15.659581			3
6.87	6.89	6.93	13002.30664	7.060454			4
7.03	7.06	7.09	550.520264	0.308184			5
Height Summation:			**109977.313234**				
Amount Avg CF:			**28.001345**	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 01:12:28
Instrument : CP07--H5088B
Result file : 3P7307B.47R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	18.40	
4.66	4.69	4.72	66709.75	69.678142			1
4.84	4.87	4.90	87839.53125	86.743825			2
4.96	4.98	5.02	45729.59375	61.011883			3
5.36	5.41	5.42	43885.21484	53.245056			4
5.60	5.62	5.66	102760.47656	62.738199			5
5.89	5.92	5.95	63017.605465	57.053013			6
Height Summation:			**409942.171876**				
Amount Avg CF:			**65.078353**	Linear:			
Aroclor-1262					6	57.30	
4.84	4.87	4.90	87839.53125	95.018362			1
5.36	5.41	5.42	43885.21484	38.106403			2
5.60	5.62	5.66	102760.47656	48.482312			3
5.84	5.87	5.90	14123.39746	18.452534			4
5.89	5.92	5.95	63017.605465	46.935584			5
6.35	6.38	6.41	26006.47656	29.784466			6
Height Summation:			**337632.702150**				
Amount Avg CF:			**46.129944**	Linear:			
Aroclor-1268					6	104.19	
5.35	5.41	5.41	43885.21484	63.333957			1
5.84	5.87	5.90	14123.39746	6.362398			2
5.90	5.92	5.96	63017.605465	32.369696			3
6.06	6.09	6.12	893.38269	0.420417			4
6.14	6.17	6.20	2753.904297	5.532005			5
6.36	6.38	6.42	26006.47656	29.402931			6
Height Summation:			**150679.981324**				
Amount Avg CF:			**22.903567**	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			9.8039	1.9608		13.63	4	15	
Aroclor-1221			9.8039	1.9608		**102.26	3	5	
Aroclor-1232			9.8039	1.9608	E	8.76	4	10	
			9.8039	3.9216	E				
Aroclor-1242			9.8039	1.9608		17.50	4	30	
Aroclor-1248			9.8039	1.9608	E	23.83	4	30	
Aroclor-1254			9.8039	1.9608		39.63	4	40	
Aroclor-1260			9.8039	1.9608		12.62	4	40	
Aroclor-1262			9.8039	1.9608		8.29	4	40	
			9.8039	3.9216					
Aroclor-1268			9.8039	1.9608		20.03	4	40	

Units: ug/L

ASX22 0843

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:16:56

Sample Name: **5822065MS RI AF DF20ACCARS2MS MS 093060018A 00173**
Acquired from CP07--H5088A via port 3 on 11/6/09 01:21:36am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.47R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.511		0.0000	0.000	10208.5	0.484	BB	0.023	7501.57	0.582
2	1.660		0.0000	0.000	89951.8	4.261	BV	0.033	45845.69	3.560
3	1.769		0.0000	0.000	8332.9	0.395	VV	0.064	2175.63	0.169
4	1.881		0.0000	0.000	22789.8	1.079	VV	0.074	5150.68	0.400
5	2.062		0.0000	0.000	81844.8	3.877	VV	0.036	38174.18	2.964
6	2.180		0.0000	0.000	7013.4	0.332	VV	0.055	2126.19	0.165
7	2.279		0.0000	0.000	13908.5	0.659	VB	0.032	7224.59	0.561
8	2.482		0.0000	0.000	1839.3	0.087	BV	0.052	589.82	0.046
9	2.574		0.0000	0.000	2705.6	0.128	VB	0.053	857.99	0.067
10	2.628		0.0000	0.000	10094.6	0.478	BB	0.021	7987.15	0.620
11	2.746		0.0000	0.000	13516.8	0.640	BB	0.027	8308.57	0.645
12	2.855		0.0000	0.000	1657.6	0.079	BB	0.042	662.68	0.051
13	2.953		0.0000	0.000	169.3	0.008	BB	0.019	148.79	0.012
14	3.002	TCX	-488.7260	40.674	14381.4	0.681	BB	0.025	9484.30	0.736
15	3.054		0.0000	0.000	5672.8	0.269	BB	0.028	3382.81	0.263
16	3.185		0.0000	0.000	616.6	0.029	BB	0.022	458.64	0.036
17	3.227		0.0000	0.000	452.1	0.021	BB	0.018	408.77	0.032
18	3.269		0.0000	0.000	2207.6	0.105	BB	0.037	984.92	0.076
19	3.282		0.0000	0.000	1269.5	0.060	BB	0.012	1704.43	0.132
20	3.345		0.0000	0.000	2941.6	0.139	BB	0.018	2789.16	0.217
21	3.381		0.0000	0.000	3977.5	0.188	BB	0.024	2707.55	0.210
22	3.459		0.0000	0.000	2959.5	0.140	BB	0.027	1806.06	0.140
23	3.514		0.0000	0.000	1728.0	0.082	BB	0.023	1265.38	0.098
24	3.598		0.0000	0.000	368.7	0.017	BB	0.026	238.05	0.018
25	3.641		0.0000	0.000	136.5	0.006	BB	0.018	125.68	0.010
26	3.695		0.0000	0.000	14186.8	0.672	BB	0.041	5774.63	0.448
27	3.776		0.0000	0.000	2931.0	0.139	BB	0.030	1632.52	0.127
28	3.826		0.0000	0.000	141.9	0.007	BB	0.015	152.94	0.012
29	3.864		0.0000	0.000	9052.0	0.429	BB	0.026	5808.39	0.451
30	3.928		0.0000	0.000	181.4	0.009	BB	0.024	124.56	0.010
31	3.970		0.0000	0.000	367.7	0.017	BB	0.028	221.09	0.017
32	4.046		0.0000	0.000	11554.4	0.547	BB	0.035	5520.72	0.429
33	4.129		0.0000	0.000	34202.0	1.620	BB	0.034	17004.23	1.320
34	4.185		0.0000	0.000	10530.2	0.499	BB	0.025	7108.86	0.552
35	4.266		0.0000	0.000	8019.6	0.380	BB	0.027	4860.60	0.377
36	4.316		0.0000	0.000	1079.1	0.051	BB	0.018	984.81	0.076
37	4.360		0.0000	0.000	14336.1	0.679	BB	0.020	12172.32	0.945
38	4.394		0.0000	0.000	8589.1	0.407	BB	0.017	8657.85	0.672
39	4.424		0.0000	0.000	11515.6	0.545	BB	0.021	9273.65	0.720
40	4.526		0.0000	0.000	15304.1	0.725	BB	0.022	11842.30	0.920
41	4.555		0.0000	0.000	1920.8	0.091	BB	0.013	2535.30	0.197
42	4.599		0.0000	0.000	8577.0	0.406	BB	0.021	6967.31	0.541
43	4.648		0.0000	0.000	12316.7	0.583	BB	0.025	8349.30	0.648
44	4.699		0.0000	0.000	12082.1	0.572	BB	0.055	3694.58	0.287
45	4.813		0.0000	0.000	7102.8	0.336	BB	0.036	3324.72	0.258
46	4.857		0.0000	0.000	29793.0	1.411	BB	0.019	26306.97	2.043
47	4.885		0.0000	0.000	12990.8	0.615	BB	0.014	15074.23	1.170
48	4.913		0.0000	0.000	26141.9	1.238	BB	0.016	26703.86	2.073
49	5.002		0.0000	0.000	13553.2	0.642	BB	0.023	9769.37	0.759
50	5.056		0.0000	0.000	87760.5	4.157	BB	0.026	57018.32	4.427
51	5.097		0.0000	0.000	34639.0	1.641	BB	0.020	28790.97	2.236
52	5.149		0.0000	0.000	371.1	0.018	BB	0.011	552.41	0.043
53	5.176		0.0000	0.000	6521.8	0.309	BB	0.019	5730.05	0.445
54	5.231		0.0000	0.000	37383.0	1.771	BB	0.025	24440.89	1.898

ASX22 0844

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.299		0.0000	0.000	44934.6	2.128	BB	0.024	31728.42	2.464
56	5.332		0.0000	0.000	7848.1	0.372	BB	0.013	9801.21	0.761
57	5.374		0.0000	0.000	100077.6	4.740	BB	0.023	71357.24	5.541
58	5.429		0.0000	0.000	18279.2	0.866	BB	0.026	11887.39	0.923
59	5.474		0.0000	0.000	14267.4	0.676	BB	0.020	11801.10	0.916
60	5.521		0.0000	0.000	63080.6	2.988	BB	0.027	39100.73	3.036
61	5.588		0.0000	0.000	9090.0	0.431	BB	0.019	7941.05	0.617
62	5.638		0.0000	0.000	141385.1	6.697	BB	0.026	90587.45	7.034
63	5.701		0.0000	0.000	34315.6	1.625	BB	0.024	23879.59	1.854
64	5.767		0.0000	0.000	151451.3	7.173	BB	0.035	72449.95	5.626
65	5.876		0.0000	0.000	127319.9	6.030	BB	0.025	84758.34	6.581
66	5.920		0.0000	0.000	1474.4	0.070	BB	0.012	2105.13	0.163
67	5.965		0.0000	0.000	6745.1	0.319	BB	0.023	4941.03	0.384
68	6.011		0.0000	0.000	162013.7	7.674	BB	0.029	92605.93	7.191
69	6.082		0.0000	0.000	28299.6	1.340	BB	0.022	21910.50	1.701
70	6.135		0.0000	0.000	22404.7	1.061	BB	0.025	14921.79	1.159
71	6.274		0.0000	0.000	102334.7	4.847	BB	0.041	41126.85	3.193
72	6.352		0.0000	0.000	15283.2	0.724	BB	0.022	11586.27	0.900
73	6.398		0.0000	0.000	17534.7	0.831	BB	0.028	10290.05	0.799
74	6.512		0.0000	0.000	36101.0	1.710	BB	0.027	22423.37	1.741
75	6.578		0.0000	0.000	108918.1	5.159	BB	0.027	66328.23	5.150
76	6.833		0.0000	0.000	56357.9	2.669	BB	0.028	33387.14	2.592
77	6.888		0.0000	0.000	28166.1	1.334	BB	0.036	13002.31	1.010
78	6.984		0.0000	0.000	208.4	0.010	BB	0.026	131.93	0.010
79	7.059		0.0000	0.000	920.5	0.044	BB	0.028	550.52	0.043
80	7.181		0.0000	0.000	22490.9	1.065	BB	0.045	8260.64	0.641
81	7.415		0.0000	0.000	29851.6	1.414	BB	0.030	16486.73	1.280
82	7.489		0.0000	0.000	2279.6	0.108	BB	0.048	793.01	0.062
83	7.587		0.0000	0.000	833.6	0.039	BB	0.089	156.96	0.012
84	7.730		0.0000	0.000	8674.2	0.411	BB	0.034	4231.99	0.329
85	7.821		0.0000	0.000	724.2	0.034	BB	0.027	445.40	0.035
86	7.883		0.0000	0.000	58.8	0.003	BB	0.034	28.47	0.002
87	7.977	DCB	-712.8540	59.326	5097.3	0.241	BB	0.037	2312.50	0.180
88	8.147		0.0000	0.000	130.2	0.006	BB	0.030	72.02	0.006
89	8.246		0.0000	0.000	329.9	0.016	BB	0.044	125.12	0.010
90	8.317		0.0000	0.000	111.4	0.005	BB	0.032	57.38	0.004
91	8.375		0.0000	0.000	4097.5	0.194	BB	0.044	1550.88	0.120
92	9.091		0.0000	0.000	3901.7	0.185	BB	0.265	245.56	0.019

Total Area = 2111282.0, Total Amount = -1201.58, Total Height = 1287877.0, Sample Units = ug/l

ASX22 8845

Sample Name: **5822065MS RI AF DF20ACCARS2MS MS 093060018A 00173**
Acquired from CP07--H5088B via port 4 on 11/6/09 01:21:36am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.47R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.567		0.0000	0.000	616.2	0.023	BV	0.024	432.24	0.034
2	1.621		0.0000	0.000	85882.0	3.189	VB	0.043	33559.36	2.658
3	1.817		0.0000	0.000	12462.5	0.463	BB	0.036	5783.13	0.458
4	1.956		0.0000	0.000	137.0	0.005	BB	0.024	94.15	0.007
5	2.054		0.0000	0.000	32382.6	1.202	BB	0.036	14855.73	1.176
6	2.167		0.0000	0.000	356.1	0.013	BB	0.042	142.63	0.011
7	2.188		0.0000	0.000	1331.6	0.049	BB	0.027	818.81	0.065
8	2.250		0.0000	0.000	404.8	0.015	BB	0.020	344.27	0.027
9	2.283		0.0000	0.000	512.3	0.019	BB	0.022	386.86	0.031
10	2.343	TCX	300.8060	53.202	12133.9	0.451	BB	0.031	6522.73	0.517
11	2.434		0.0000	0.000	20725.2	0.770	BB	0.043	7973.45	0.631
12	2.487		0.0000	0.000	950.9	0.035	BB	0.052	304.38	0.024
13	2.599		0.0000	0.000	3533.8	0.131	BB	0.045	1297.43	0.103
14	2.742		0.0000	0.000	18262.7	0.678	BB	0.043	7057.59	0.559
15	2.861		0.0000	0.000	3158.8	0.117	BB	0.039	1345.90	0.107
16	3.009		0.0000	0.000	15949.3	0.592	BB	0.038	6997.71	0.554
17	3.072		0.0000	0.000	1506.2	0.056	BB	0.029	851.34	0.067
18	3.169		0.0000	0.000	16228.6	0.603	BB	0.050	5461.37	0.433
19	3.319		0.0000	0.000	11845.7	0.440	BB	0.049	4003.98	0.317
20	3.405		0.0000	0.000	24617.4	0.914	BB	0.040	10357.28	0.820
21	3.459		0.0000	0.000	2113.1	0.078	BB	0.023	1554.84	0.123
22	3.559		0.0000	0.000	5714.3	0.212	BB	0.028	3377.61	0.267
23	3.617		0.0000	0.000	7596.1	0.282	BB	0.024	5186.99	0.411
24	3.663		0.0000	0.000	36618.4	1.360	BB	0.042	14494.78	1.148
25	3.796		0.0000	0.000	23641.2	0.878	BB	0.039	10187.22	0.807
26	3.855		0.0000	0.000	12574.9	0.467	BB	0.043	4909.43	0.389
27	3.924		0.0000	0.000	5604.6	0.208	BB	0.025	3681.59	0.292
28	3.973		0.0000	0.000	3007.3	0.112	BB	0.027	1873.51	0.148
29	4.045		0.0000	0.000	1606.6	0.060	BB	0.039	681.80	0.054
30	4.112		0.0000	0.000	48624.6	1.805	BB	0.028	28854.29	2.285
31	4.138		0.0000	0.000	14112.9	0.524	BB	0.026	9131.19	0.723
32	4.198		0.0000	0.000	4190.2	0.156	BB	0.019	3601.92	0.285
33	4.250		0.0000	0.000	48550.5	1.803	BB	0.022	36498.23	2.890
34	4.287		0.0000	0.000	73890.6	2.744	BB	0.029	42091.98	3.333
35	4.384		0.0000	0.000	3172.2	0.118	BB	0.020	2603.35	0.206
36	4.431		0.0000	0.000	41051.5	1.524	BB	0.031	21751.58	1.723
37	4.503		0.0000	0.000	37834.5	1.405	BB	0.022	28083.33	2.224
38	4.537		0.0000	0.000	12062.1	0.448	BB	0.017	12020.57	0.952
39	4.591		0.0000	0.000	144786.2	5.376	BB	0.035	68860.21	5.453
40	4.635		0.0000	0.000	24782.4	0.920	BB	0.019	21610.93	1.711
41	4.692		0.0000	0.000	115258.5	4.280	BB	0.029	66709.75	5.283
42	4.744		0.0000	0.000	1956.3	0.073	BB	0.014	2310.75	0.183
43	4.790		0.0000	0.000	230618.5	8.563	BB	0.037	103923.80	8.230
44	4.872		0.0000	0.000	178618.0	6.632	BB	0.034	87839.53	6.956
45	4.984		0.0000	0.000	116395.0	4.322	BB	0.042	45729.59	3.622
46	5.045		0.0000	0.000	116989.0	4.344	BB	0.027	71407.43	5.655
47	5.091		0.0000	0.000	4632.8	0.172	BB	0.017	4642.75	0.368
48	5.141		0.0000	0.000	310198.1	11.518	BB	0.043	120490.40	9.542
49	5.217		0.0000	0.000	9780.9	0.363	BB	0.021	7882.38	0.624
50	5.334		0.0000	0.000	5016.0	0.186	BB	0.022	3860.79	0.306
51	5.407		0.0000	0.000	130274.9	4.837	BB	0.049	43885.21	3.475
52	5.471		0.0000	0.000	16086.1	0.597	BB	0.022	12093.02	0.958
53	5.510		0.0000	0.000	5870.2	0.218	BB	0.020	4927.83	0.390
54	5.560		0.0000	0.000	3018.6	0.112	BB	0.018	2847.32	0.225

ASX22 8846

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.625		0.0000	0.000	260442.9	9.670	BB	0.042	102760.50	8.138
56	5.798		0.0000	0.000	30447.2	1.131	BB	0.029	17267.09	1.367
57	5.867		0.0000	0.000	20721.6	0.769	BB	0.024	14123.40	1.118
58	5.921		0.0000	0.000	140470.4	5.216	BB	0.037	63017.61	4.991
59	6.087		0.0000	0.000	1690.1	0.063	BB	0.032	893.38	0.071
60	6.172		0.0000	0.000	5086.8	0.189	BB	0.031	2753.90	0.218
61	6.237		0.0000	0.000	23737.4	0.881	BB	0.038	10417.86	0.825
62	6.382		0.0000	0.000	64014.2	2.377	BB	0.041	26006.48	2.060
63	6.621		0.0000	0.000	16016.2	0.595	BB	0.038	7074.85	0.560
64	6.804	DCB	264.5948	46.798	17323.0	0.643	BB	0.063	4567.19	0.362
65	7.352		0.0000	0.000	4076.3	0.151	BB	0.055	1242.37	0.098
66	7.454		0.0000	0.000	42045.3	1.561	BB	0.198	3546.24	0.281
67	8.510		0.0000	0.000	4404.0	0.164	BB	0.113	647.35	0.051
68	8.650		0.0000	0.000	3470.1	0.129	BB	0.294	196.78	0.016

Total Area = 2693190.0, Total Amount = 565.401, Total Height = 1262711.0, Sample Units = ug/l

ASX22 0847

5822065MS RI AF DF20ACCARS2MS MS 093060018A 00173

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\3P7307.47R

C:\CPWIN\DATA1\3P7307.47R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 1:12:27 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.47R

C:\CPWIN\DATA1\3P7307B.47R



Instrument ID: CP07--H5088B Injected On: 11/6/2009 1:12:27 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1020
Analyst: 2280

Dilution Factor: 200

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.002	9484	-.47914	TCX	2.343	6523	.29491	TCX
7.977	2312	-.69888	DCB	6.804	4567	.25941	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.47A
Area File: C:\CPWIN\DATA1\3P7307B.47A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 1:21:44 AM
File Reported On: 11/6/2009 at 1:21:57 AM

10X REPLOTS

ASX22 0849

C:\CPWIN\DATA1\3P7307.47R



C:\CPWIN\DATA1\3P7307B.47R



ASX22 0858

Multiple Component Data Summary

Sample Name: 5822066 MSD RI AF DF2 CARS2MSC MSD Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 1041 ml Total Volume: 200 ml Analyst: 2131 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on Nov 06, 2009 01:25:29
Instrument H5088A
Result file 3P7307.48R
Calibration file 1P7307
Method file PESTD

%SSR(TCX) * 0.0% Conc: 0
%SSR(DCB) * 0.0% Conc: 0

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016						**46.71**	
3.35	3.38	3.41	2445.943848	4.493163			1
3.83	3.86	3.89	5205.755371	7.886722			2
4.10	4.13	4.16	14569.97460	10.296112			3
4.15	4.18	4.21	5975.90625	9.499116			4
4.24	4.27	4.30	4400.979492	7.387025			5
4.62	4.65	4.68	7494.260254	17.650254			6
Height summation:			**40092.819824**				
Concentration		CF:	**9.535399**	L:			
Aroclor-1248						**10.03**	
4.10	4.13	4.16	14569.97460	18.669726			1
4.33	4.36	4.39	11163.74707	20.554282			2
4.50	4.53	4.56	13674.09570	16.8091			3
Height summation:			**39407.817382**				
Concentration		CF:	**18.677703**	L:			
Aroclor-1254						**14.62**	
5.03	5.06	5.09	54078.46484	57.246319			2
5.27	5.30	5.33	29965.40820	43.577383			3
5.34	5.37	5.40	69580.32812	56.420589			4
Height summation:			**153624.201172**				
Concentration		CF:	**52.414764**	L:			
Aroclor-1260						**13.35**	
6.25	6.28	6.31	38428.92968	49.134121			4
6.55	6.58	6.61	63503.33203	38.053599			5
6.80	6.83	6.86	31576.45703	41.193414			6
Height summation:			**133508.71875**				
Concentration		CF:	**42.793711**	L:			

Analysis Report (B)

Injected on Nov 06, 2009 01:25:29
Instrument H5088B
Result file 3P7307B.48R
Calibration file 1P7307B
Method file PESTDB

%SSR(TCX) 93.0% Conc: 0.274445
%SSR(DCB) 80.1% Conc: 0.23464

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016						**41.23**	
2.69	2.74	2.74	7157.365723	17.829007			1
2.98	3.01	3.04	6384.763672	7.165211			2
3.14	3.17	3.20	4879.429688	7.949616			3
3.37	3.40	3.43	12402.72265	13.572964			4
3.77	3.80	3.83	9522.288086	18.703586			6
Height summation:			**40346.569825**				
Concentration		CF:	**13.044077**	L:			
Aroclor-1248						**23.64**	
3.14	3.17	3.20	4879.429688	22.241451			1
3.37	3.40	3.43	12402.72265	18.737175			2
3.77	3.80	3.83	9522.288086	13.674595			3
Height summation:			**26804.44043**				
Concentration		CF:	**18.21774**	L:			
Aroclor-1254						**34.22**	
4.55	4.59	4.61	73893.19531	68.245173			2
4.66	4.69	4.72	61524.58203	100.235104			3
4.76	4.79	4.82	95747.64062	138.101137			4
Height summation:			**231165.417969**				
Concentration		CF:	**102.193805**	L:			
Aroclor-1260						**8.99**	
5.36	5.41	5.42	40922.57812	48.648952			4
5.60	5.63	5.66	96905.55468	57.970101			5
5.89	5.92	5.95	58332.98046	51.746419			6
Height summation:			**196161.113282**				
Concentration		CF:	**52.788491**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016	B	13.044077	9.6061	1.9212		31.08	4	15	
Aroclor-1221			<9.6061	<1.9212			3	5	
Aroclor-1232			<9.6061	<1.9212			4	10	
Aroclor-1242			<9.6061	<1.9212			4	30	
Aroclor-1248	A	18.677703	9.6061	1.9212		2.49	4	30	
Aroclor-1254	A	52.414764	9.6061	1.9212		64.39	4	40	

ASX22 0851

Multiple Component Data Summary

Sample Name: 5822066 MSD RI AF DF2 CARS2MSD MSD Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 1041 ml Total Volume: 200 ml Analyst: 2131 SDG: ASX22 State: DE
Analyses: 00173

Analysis Report (A)

Injected on	Nov 06, 2009 01:25:29
Instrument	H5088A
Result file	3P7307.48R
Calibration file	1P7307
Method file	PESTD

Analysis Report (B)

Injected on	Nov 06, 2009 01:25:29
Instrument	H5088B
Result file	3P7307B.48R
Calibration file	1P7307B
Method file	PESTDB

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1260	B	52.788491	9.6061	1.9212		20.91	4	40	

Units: ug/l

%Difference = High - Low divided by the Average times 100

Reviewed By: [illegible] Date: 11-6-09

Verified By: [illegible] Date: 11/9/09

ASX22 0852

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822066** MSD RI AF DF2 CARS2MSD Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 1041 ml Total Volume: 200 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173 00516 06030 07394 10227

Analysis Report (A)

Injected on : NOV 06, 2009 01:25:29
Instrument : CP07--H5088A
Result file : 3P7307.48R
Calibration file : 1P7307.CAL
Method file : PESTD.MET
%SSR(TCX) : *-166.8% Conc.: -0.491882
%SSR(DCB) : *-235.5% Conc.: -0.68999

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	46.71	
3.35	3.38	3.41	2445.943848	4.493163			1
3.83	3.86	3.89	5205.755371	7.886722			2
4.10	4.13	4.16	14569.97460	10.296112			3
4.15	4.18	4.21	5975.90625	9.499116			4
4.24	4.27	4.30	4400.979492	7.387025			5
4.62	4.65	4.68	7494.260254	17.650254			6
Height Summation:			40092.819824				
Amount Avg CF:			9.535399	Linear:			
Aroclor-1221					2	58.79	
3.20	3.23	3.26	431.443604	1.334249			1
3.35	3.38	3.41	2445.943848	3.232926			3
Height Summation:			2877.387452				
Amount Avg CF:			2.283588	Linear:			
Aroclor-1232					6	60.06	
3.35	3.38	3.41	2445.943848	3.828129			1
3.83	3.86	3.89	5205.755371	17.725213			2
4.10	4.13	4.16	14569.97460	21.511002			3
4.15	4.18	4.21	5975.90625	19.848021			4
4.23	4.27	4.29	4400.979492	15.751289			5
E 4.62	4.65	4.68	7494.260254	40.644162			6
Height Summation:			40092.819824				
Amount Avg CF:			19.884636	Linear:			
Aroclor-1242					6	44.97	
3.35	3.38	3.41	2445.943848	5.254928			1
3.83	3.86	3.89	5205.755371	9.652047			2
4.09	4.13	4.15	14569.97460	11.903751			3
4.15	4.18	4.21	5975.90625	11.106252			4
4.23	4.27	4.29	4400.979492	8.61247			5
4.62	4.65	4.68	7494.260254	20.141561			6
Height Summation:			40092.819824				
Amount Avg CF:			11.111835	Linear:			
Aroclor-1248					6	144.32	
4.10	4.13	4.16	14569.97460	18.669726			1
4.33	4.36	4.39	11163.74707	20.554282			2
4.50	4.53	4.56	13674.09570	16.8091			3
4.62	4.65	4.68	7494.260254	13.946198			4
5.02	5.06	5.08	54078.46484	83.094584			5
E 5.61	5.64	5.67	84654.59375	273.552956			6
Height Summation:			185635.136230				
Amount Avg CF:			71.104474	Linear:			
Aroclor-1254					6	51.28	
4.83	4.86	4.89	23966.51953	38.354466			1
+ 4.83	4.89	4.89	13320.67578	21.317547			1
5.03	5.06	5.09	54078.46484	67.246819			2
5.27	5.30	5.33	29965.40820	43.577383			3
5.34	5.37	5.40	69580.32812	56.420589			4
5.85	5.88	5.91	82510.85937	129.627758			5
5.98	6.01	6.04	87275.42968	103.520914			6
Height Summation:			347377.009766				
Amount Avg CF:			71.457905	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 01:25:29
Instrument : CP07--H5088B
Result file : 3P7307B.48R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET A Unear
%SSR(TCX) : 93% Conc.: 0.274445
%SSR(DCB) : 80.1% Conc.: 0.23464

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					5	41.23	
2.69	2.74	2.75	7157.365723	17.829007			1
2.98	3.01	3.04	6384.763672	7.165211			2
3.14	3.17	3.20	4879.429688	7.949616			3
3.37	3.40	3.43	12402.72265	13.572964			4
3.77	3.80	3.83	9522.288086	18.703586			6
Height Summation:			40346.569825				
Amount Avg CF:			13.044077	Linear:			
Aroclor-1221					1		
2.69	2.74	2.75	7157.365723	12.874812			3
Height Summation:			7157.365723				
Amount Avg CF:			12.874812	Linear:			
Aroclor-1232					5	53.45	
2.69	2.74	2.75	7157.365723	15.579157			1
2.98	3.01	3.04	6384.763672	15.804415			2
3.14	3.17	3.20	4879.429688	16.992033			3
3.38	3.40	3.44	12402.72265	32.946392			4
E 3.77	3.80	3.83	9522.288086	46.012328			6
Height Summation:			40346.569825				
Amount Avg CF:			25.466865	Linear:			
Aroclor-1242					5	40.15	
2.68	2.74	2.74	7157.365723	21.477458			1
2.98	3.01	3.04	6384.763672	8.79452			2
3.14	3.17	3.20	4879.429688	9.791628			3
3.37	3.40	3.43	12402.72265	17.561615			4
3.77	3.80	3.83	9522.288086	22.553784			6
Height Summation:			40346.569825				
Amount Avg CF:			16.035801	Linear:			
Aroclor-1248					6	140.45	
3.14	3.17	3.20	4879.429688	22.241451			1
3.37	3.40	3.43	12402.72265	18.737175			2
3.77	3.80	3.83	9522.288086	13.674595			3
3.89	3.92	3.95	3495.210449	10.723094			4
4.26	4.29	4.32	37041.11718	69.295183			5
E 4.55	4.59	4.61	73893.19531	228.159392			6
Height Summation:			141233.963380				
Amount Avg CF:			60.471815	Linear:			
Aroclor-1254					6	39.27	
4.08	4.11	4.14	60719.90625	62.502347			1
4.55	4.59	4.61	73893.19531	68.246173			2
4.66	4.69	4.72	61524.58203	100.235104			3
4.76	4.79	4.82	95747.64062	138.101137			4
E 4.85	4.87	4.91	81182.02343	177.268869			5
5.11	5.14	5.17	112603.2031	119.528534			6
Height Summation:			485670.550782				
Amount Avg CF:			110.980194	Linear:			

ASX22 0853

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:19:23

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822066** MSD RI AF DF2 CARS2MSD Sample ID: AC **Batchnumber: 093060018A**
Sample Amount: 1041 ml Total Volume: 200 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 00173 00516 06030 07394 10227

Analysis Report (A)

Injected on : NOV 06, 2009 01:25:29
Instrument : CP07--H5088A
Result file : 3P7307.48R
Calibration file : 1P7307.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	46.02	
5.74	5.77	5.80	68474.39843	64.437415			1
5.98	6.01	6.04	87275.42968	96.563385			2
6.05	6.08	6.11	20748.95898	29.01507			3
6.25	6.28	6.31	38428.92968	46.134121			4
6.55	6.58	6.61	63503.33203	88.053599			5
6.80	6.83	6.86	31576.45703	41.193414			6
Height Summation:			**310007.505860**				
Amount Avg CF:			**53.066167**	Linear:			
Aroclor-1262					6	51.28	
5.49	5.52	5.55	37450.51953	43.245321			1
5.73	5.77	5.79	68474.39843	75.903503			2
6.05	6.08	6.11	20748.95898	17.380113			3
6.25	6.28	6.31	38428.92968	41.374029			4
6.55	6.58	6.61	63503.33203	29.776465			5
6.79	6.83	6.85	31576.45703	28.761758			6
Height Summation:			**260182.595703**				
Amount Avg CF:			**39.406865**	Linear:			
Aroclor-1268					5	105.44	
6.05	6.08	6.11	20748.95898	43.924962			1
6.26	6.28	6.32	38428.92968	63.758596			2
6.79	6.83	6.85	31576.45703	14.51155			3
6.87	6.88	6.93	11231.60546	5.975902			4
7.03	7.06	7.09	625.03186	0.342838			5
Height Summation:			**102610.983032**				
Amount Avg CF:			**25.70277**	Linear:			

Analysis Report (B)

Injected on : NOV 06, 2009 01:25:29
Instrument : CP07--H5088B
Result file : 3P7307B.48R
Calibration file : 1P7307B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	17.99	
4.66	4.69	4.72	61524.58203	62.965892			1
4.84	4.87	4.90	81182.02343	78.552113			2
4.96	4.98	5.02	42667.74218	55.778418			3
5.36	5.41	5.42	40922.57812	48.648952			4
5.60	5.63	5.66	96905.55468	57.970101			5
5.89	5.92	5.95	58332.98046	51.746419			6
Height Summation:			**381535.460939**				
Amount Avg CF:			**59.276983**	Linear:			
Aroclor-1262					6	57.10	
4.84	4.87	4.90	81182.02343	86.045239			1
5.36	5.41	5.42	40922.57812	34.817065			2
5.60	5.63	5.66	96905.55468	44.79766			3
5.84	5.87	5.90	12841.72753	16.439543			4
5.89	5.92	5.95	58332.98046	42.570028			5
6.35	6.38	6.41	24230.51171	27.190697			6
Height Summation:			**314415.375978**				
Amount Avg CF:			**41.976705**	Linear:			
Aroclor-1268					6	104.69	
5.35	5.41	5.41	40922.57812	57.866981			1
5.84	5.87	5.90	12841.72753	5.668323			2
5.90	5.92	5.96	58332.98046	29.358937			3
6.06	6.09	6.12	801.124329	0.369396			4
6.14	6.17	6.20	2518.891357	4.957841			5
6.36	6.38	6.42	24230.51171	26.842387			6
Height Summation:			**139647.813538**				
Amount Avg CF:			**20.843978**	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			9.6061	1.9212		31.08	4	15	
Aroclor-1221			9.6061	1.9212		** 139.74	3	5	
Aroclor-1232			9.6061	1.9212	E	24.62	4	10	
			9.6061	3.8425	E				
Aroclor-1242			9.6061	1.9212		36.28	4	30	
Aroclor-1248			9.6061	1.9212	E	16.16	4	30	
Aroclor-1254			9.6061	1.9212		** 43.33	4	40	
Aroclor-1260			9.6061	1.9212		11.06	4	40	
Aroclor-1262			9.6061	1.9212		6.32	4	40	
			9.6061	3.8425					
Aroclor-1268			9.6061	1.9212		20.88	4	40	

Units: ug/L

ASX22 0054

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/6/09 06:19:23

Sample Name: **5822066MSD RI AF DF2ACCARS2MSD MSD 093060018A 00173**
Acquired from CP07--H5088A via port 3 on 11/6/09 01:34:37am by 2280
RTX-CLPI 30M x .32MM x .5UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307.48R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1P7307.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.511		0.0000	0.000	9672.0	0.488	BB	0.023	7085.85	0.593
2	1.660		0.0000	0.000	69544.9	3.512	BV	0.033	35602.50	2.979
3	1.769		0.0000	0.000	6300.8	0.318	VV	0.063	1674.42	0.140
4	1.880		0.0000	0.000	16285.3	0.822	VV	0.063	4325.32	0.362
5	2.062		0.0000	0.000	74407.0	3.757	VV	0.036	34191.79	2.861
6	2.178		0.0000	0.000	6498.4	0.328	VV	0.045	2398.92	0.201
7	2.279		0.0000	0.000	9752.5	0.492	VV	0.026	6318.11	0.529
8	2.314		0.0000	0.000	2042.6	0.103	VB	0.024	1441.70	0.121
9	2.446		0.0000	0.000	472.5	0.024	BV	0.037	211.42	0.018
10	2.481		0.0000	0.000	545.0	0.028	VV	0.026	351.73	0.029
11	2.574		0.0000	0.000	2368.2	0.120	VB	0.049	807.97	0.068
12	2.628		0.0000	0.000	8787.4	0.444	BB	0.021	6955.94	0.582
13	2.746		0.0000	0.000	11981.8	0.605	BB	0.027	7320.21	0.612
14	2.854		0.0000	0.000	2376.4	0.120	BB	0.046	868.01	0.073
15	2.954		0.0000	0.000	176.2	0.009	BB	0.019	151.14	0.013
16	3.002	TCX	-512.0491	41.619	13617.3	0.688	BB	0.025	9008.14	0.754
17	3.054		0.0000	0.000	7921.3	0.400	BB	0.030	4356.58	0.365
18	3.184		0.0000	0.000	492.7	0.025	BB	0.022	367.95	0.031
19	3.226		0.0000	0.000	478.4	0.024	BB	0.018	431.44	0.036
20	3.270		0.0000	0.000	1981.9	0.100	BB	0.037	902.58	0.076
21	3.345		0.0000	0.000	3019.4	0.152	BB	0.018	2849.98	0.238
22	3.381		0.0000	0.000	3508.1	0.177	BB	0.024	2445.94	0.205
23	3.459		0.0000	0.000	2655.1	0.134	BB	0.027	1611.59	0.135
24	3.514		0.0000	0.000	1829.3	0.092	BB	0.023	1305.09	0.109
25	3.598		0.0000	0.000	289.5	0.015	BB	0.025	193.04	0.016
26	3.641		0.0000	0.000	349.8	0.018	BB	0.021	281.76	0.024
27	3.695		0.0000	0.000	12699.1	0.641	BB	0.040	5258.36	0.440
28	3.780		0.0000	0.000	2710.3	0.137	BB	0.030	1520.67	0.127
29	3.826		0.0000	0.000	143.6	0.007	BB	0.016	147.71	0.012
30	3.864		0.0000	0.000	8301.4	0.419	BB	0.027	5205.76	0.436
31	3.931		0.0000	0.000	299.6	0.015	BB	0.027	188.40	0.016
32	3.973		0.0000	0.000	169.9	0.009	BB	0.024	119.30	0.010
33	4.045		0.0000	0.000	9167.0	0.463	BB	0.035	4331.26	0.362
34	4.129		0.0000	0.000	29337.0	1.481	BB	0.034	14569.97	1.219
35	4.184		0.0000	0.000	8722.9	0.440	BB	0.024	5975.91	0.500
36	4.266		0.0000	0.000	7216.5	0.364	BB	0.027	4400.98	0.368
37	4.316		0.0000	0.000	788.3	0.040	BB	0.018	728.42	0.061
38	4.360		0.0000	0.000	13164.6	0.665	BB	0.020	11163.75	0.934
39	4.394		0.0000	0.000	6682.2	0.337	BB	0.016	6790.50	0.568
40	4.425		0.0000	0.000	9343.4	0.472	BB	0.022	6987.39	0.585
41	4.527		0.0000	0.000	27758.5	1.402	BB	0.034	13674.10	1.144
42	4.600		0.0000	0.000	6322.1	0.319	BB	0.020	5171.09	0.433
43	4.648		0.0000	0.000	11070.3	0.559	BB	0.025	7494.26	0.627
44	4.699		0.0000	0.000	8261.9	0.417	BB	0.048	2858.54	0.239
45	4.813		0.0000	0.000	5099.7	0.258	BB	0.035	2431.37	0.203
46	4.857		0.0000	0.000	27539.2	1.391	BB	0.019	23966.52	2.005
47	4.885		0.0000	0.000	11395.5	0.575	BB	0.014	13320.68	1.115
48	4.913		0.0000	0.000	21691.7	1.095	BB	0.017	21432.92	1.793
49	5.003		0.0000	0.000	11076.6	0.559	BB	0.022	8525.51	0.713
50	5.056		0.0000	0.000	84551.7	4.269	BB	0.026	54078.46	4.525
51	5.098		0.0000	0.000	29045.2	1.467	BB	0.020	24360.79	2.038
52	5.149		0.0000	0.000	242.2	0.012	BB	0.010	401.10	0.034
53	5.177		0.0000	0.000	8052.4	0.407	BB	0.019	6932.44	0.580
54	5.232		0.0000	0.000	35070.2	1.771	BB	0.025	23222.36	1.943

ASX22 0855

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.299		0.0000	0.000	43035.2	2.173	BB	0.024	29965.41	2.507
56	5.332		0.0000	0.000	7871.4	0.397	BB	0.014	9696.48	0.811
57	5.375		0.0000	0.000	98503.7	4.974	BB	0.024	69580.33	5.822
58	5.429		0.0000	0.000	17390.4	0.878	BB	0.025	11431.66	0.956
59	5.475		0.0000	0.000	13504.1	0.682	BB	0.020	11181.07	0.936
60	5.521		0.0000	0.000	61069.0	3.084	BB	0.027	37450.52	3.134
61	5.588		0.0000	0.000	8424.5	0.425	BB	0.020	7082.05	0.593
62	5.638		0.0000	0.000	132172.2	6.674	BB	0.026	84654.59	7.083
63	5.702		0.0000	0.000	34896.2	1.762	BB	0.024	24187.94	2.024
64	5.767		0.0000	0.000	143073.3	7.224	BB	0.035	68474.40	5.729
65	5.876		0.0000	0.000	135556.2	6.845	BB	0.027	82510.86	6.904
66	5.965		0.0000	0.000	6270.9	0.317	BB	0.023	4511.27	0.377
67	6.012		0.0000	0.000	151609.9	7.655	BB	0.029	87275.43	7.302
68	6.083		0.0000	0.000	27444.7	1.386	BB	0.022	20748.96	1.736
69	6.136		0.0000	0.000	21094.2	1.065	BB	0.025	14001.74	1.172
70	6.276		0.0000	0.000	96294.3	4.862	BB	0.042	38428.93	3.215
71	6.352		0.0000	0.000	13714.3	0.692	BB	0.023	10118.07	0.847
72	6.397		0.0000	0.000	27014.9	1.364	BB	0.031	14458.76	1.210
73	6.512		0.0000	0.000	33405.4	1.687	BB	0.027	20956.31	1.753
74	6.578		0.0000	0.000	103081.4	5.205	BB	0.027	63503.33	5.313
75	6.833		0.0000	0.000	52832.2	2.668	BB	0.028	31576.46	2.642
76	6.884		0.0000	0.000	26528.7	1.340	BB	0.039	11231.61	0.940
77	6.983		0.0000	0.000	264.9	0.013	BB	0.029	153.96	0.013
78	7.059		0.0000	0.000	1100.1	0.056	BB	0.029	625.03	0.052
79	7.181		0.0000	0.000	21264.7	1.074	BB	0.046	7702.60	0.644
80	7.416		0.0000	0.000	28518.0	1.440	BB	0.031	15419.51	1.290
81	7.489		0.0000	0.000	501.4	0.025	BB	0.024	351.59	0.029
82	7.581		0.0000	0.000	615.5	0.031	BB	0.083	123.98	0.010
83	7.731		0.0000	0.000	8375.3	0.423	BB	0.035	4006.45	0.335
84	7.821		0.0000	0.000	645.2	0.033	BB	0.027	403.21	0.034
85	7.977	DCB	-718.2800	58.381	5538.4	0.280	BB	0.041	2257.55	0.189
86	8.147		0.0000	0.000	261.9	0.013	BB	0.040	108.74	0.009
87	8.243		0.0000	0.000	323.2	0.016	BB	0.039	137.21	0.011
88	8.320		0.0000	0.000	88.0	0.004	BB	0.034	43.00	0.004
89	8.374		0.0000	0.000	4665.6	0.236	BB	0.040	1954.57	0.164
90	8.459		0.0000	0.000	198.6	0.010	BB	0.026	128.54	0.011

Total Area = 1980425.0, Total Amount = -1230.329, Total Height = 1195160.0, Sample Units = ug/l

6SX22 0856

Sample Name: **5822066MSD RI AF DF2ACCARS2MSD MSD 093060018A 00173**
Acquired from CP07--H5088B via port 4 on 11/6/09 01:34:37am by 2280
RTX-CLPII 30M x .32MM x .25UM
140 - 300 @ 10C/min
Data File: **C:\CPWIN\DATA1\3P7307B.48R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1P7307B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	1.568		0.0000	0.000	1380.3	0.051	BV	0.025	910.59	0.075
2	1.620		0.0000	0.000	78205.6	2.915	VB	0.043	30187.64	2.484
3	1.784		0.0000	0.000	374.5	0.014	BV	0.017	361.19	0.030
4	1.817		0.0000	0.000	10765.2	0.401	VB	0.034	5207.61	0.429
5	2.054		0.0000	0.000	28295.4	1.055	BB	0.036	12973.72	1.068
6	2.167		0.0000	0.000	353.2	0.013	BB	0.039	149.06	0.012
7	2.188		0.0000	0.000	1243.0	0.046	BB	0.026	781.83	0.064
8	2.251		0.0000	0.000	515.1	0.019	BB	0.021	412.24	0.034
9	2.283		0.0000	0.000	630.1	0.023	BB	0.023	466.10	0.038
10	2.343	TCX	285.6975	53.909	11350.9	0.423	BB	0.031	6195.12	0.510
11	2.433		0.0000	0.000	18762.9	0.699	BB	0.044	7165.19	0.590
12	2.491		0.0000	0.000	2197.0	0.082	BB	0.047	777.43	0.064
13	2.599		0.0000	0.000	3227.7	0.120	BB	0.047	1153.27	0.095
14	2.743		0.0000	0.000	18156.8	0.677	BB	0.042	7157.37	0.589
15	2.862		0.0000	0.000	3274.5	0.122	BB	0.040	1359.33	0.112
16	3.009		0.0000	0.000	14763.5	0.550	BB	0.039	6384.76	0.525
17	3.073		0.0000	0.000	1100.9	0.041	BB	0.029	630.15	0.052
18	3.170		0.0000	0.000	14314.9	0.534	BB	0.049	4879.43	0.402
19	3.271		0.0000	0.000	459.0	0.017	BB	0.019	395.12	0.033
20	3.320		0.0000	0.000	7066.5	0.263	BB	0.042	2773.99	0.228
21	3.404		0.0000	0.000	53528.2	1.995	BB	0.072	12402.72	1.021
22	3.560		0.0000	0.000	6095.1	0.227	BB	0.028	3607.29	0.297
23	3.617		0.0000	0.000	6674.6	0.249	BB	0.024	4691.68	0.386
24	3.661		0.0000	0.000	25178.5	0.938	BB	0.043	9829.05	0.809
25	3.796		0.0000	0.000	21517.7	0.802	BB	0.038	9522.29	0.784
26	3.857		0.0000	0.000	9285.9	0.346	BB	0.044	3538.42	0.291
27	3.923		0.0000	0.000	5369.2	0.200	BB	0.026	3495.21	0.288
28	3.973		0.0000	0.000	2278.5	0.085	BB	0.026	1454.76	0.120
29	4.047		0.0000	0.000	1224.9	0.046	BB	0.053	385.07	0.032
30	4.112		0.0000	0.000	218845.6	8.157	BB	0.060	60719.91	4.997
31	4.199		0.0000	0.000	3748.6	0.140	BB	0.017	3690.44	0.304
32	4.250		0.0000	0.000	43633.1	1.626	BB	0.022	33387.01	2.747
33	4.287		0.0000	0.000	64981.5	2.422	BB	0.029	37041.12	3.048
34	4.384		0.0000	0.000	2661.3	0.099	BB	0.020	2236.50	0.184
35	4.431		0.0000	0.000	37798.3	1.409	BB	0.031	20246.75	1.666
36	4.504		0.0000	0.000	35166.1	1.311	BB	0.022	26148.28	2.152
37	4.537		0.0000	0.000	11103.7	0.414	BB	0.017	11085.40	0.912
38	4.593		0.0000	0.000	153178.7	5.709	BB	0.035	73893.20	6.081
39	4.635		0.0000	0.000	22408.0	0.835	BB	0.019	19815.48	1.631
40	4.692		0.0000	0.000	106773.7	3.980	BB	0.029	61524.58	5.063
41	4.743		0.0000	0.000	1476.3	0.055	BB	0.013	1881.05	0.155
42	4.790		0.0000	0.000	215692.3	8.039	BB	0.038	95747.64	7.879
43	4.872		0.0000	0.000	164609.2	6.135	BB	0.034	81182.02	6.681
44	4.985		0.0000	0.000	108388.3	4.040	BB	0.042	42667.74	3.511
45	5.045		0.0000	0.000	108566.6	4.047	BB	0.027	66291.75	5.455
46	5.091		0.0000	0.000	4158.7	0.155	BB	0.017	4094.58	0.337
47	5.141		0.0000	0.000	290292.3	10.820	BB	0.043	112603.20	9.266
48	5.217		0.0000	0.000	9919.1	0.370	BB	0.022	7529.86	0.620
49	5.335		0.0000	0.000	4609.7	0.172	BB	0.021	3587.02	0.295
50	5.408		0.0000	0.000	122486.6	4.565	BB	0.050	40922.58	3.368
51	5.472		0.0000	0.000	15104.5	0.563	BB	0.022	11263.03	0.927
52	5.510		0.0000	0.000	5596.6	0.209	BB	0.020	4596.62	0.378
53	5.560		0.0000	0.000	2865.3	0.107	BB	0.018	2695.60	0.222
54	5.625		0.0000	0.000	244492.0	9.113	BB	0.042	96905.55	7.974

ASX22 0857

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
55	5.799		0.0000	0.000	53950.9	2.011	BB	0.031	29111.27	2.396
56	5.867		0.0000	0.000	18397.1	0.686	BB	0.024	12841.73	1.057
57	5.921		0.0000	0.000	130755.5	4.874	BB	0.037	58332.98	4.800
58	6.087		0.0000	0.000	1478.5	0.055	BB	0.031	801.12	0.066
59	6.173		0.0000	0.000	4656.9	0.174	BB	0.031	2518.89	0.207
60	6.237		0.0000	0.000	22545.8	0.840	BB	0.039	9744.70	0.802
61	6.382		0.0000	0.000	58963.7	2.198	BB	0.041	24230.51	1.994
62	6.621		0.0000	0.000	15139.8	0.564	BB	0.038	6580.24	0.541
63	6.804	DCB	244.2606	46.091	11665.0	0.435	BB	0.046	4216.20	0.347
64	7.001		0.0000	0.000	1328.6	0.050	BB	0.061	364.61	0.030
65	7.350		0.0000	0.000	5312.8	0.198	BB	0.055	1611.11	0.133
66	7.454		0.0000	0.000	6582.5	0.245	BB	0.037	2954.02	0.243
67	8.513		0.0000	0.000	4831.7	0.180	BB	0.116	693.69	0.057
68	8.662		0.0000	0.000	1174.6	0.044	BB	0.099	197.90	0.016

Total Area = 2682930.0, Total Amount = 529.958, Total Height = 1215204.0, Sample Units = ug/l

ASX22 0858

LANCASTER LABORATORIES

File: C:\CPWIN\DATA1\3P7307.48R

C:\CPWIN\DATA1\3P7307.48R



Instrument ID: CP07--H5088A Injected On: 11/6/2009 1:25:28 AM Column ID: RTX-CLPI 30M x .32MM x .5UM

File: C:\CPWIN\DATA1\3P7307B.48R

C:\CPWIN\DATA1\3P7307B.48R



ASX22 0859

Instrument ID: CP07--H5088B Injected On: 11/6/2009 1:25:28 AM Column ID: RTX-CLPII 30M x .32MM x .25UM

Oven Parameters: 140 - 300 @ 10C/min

Volume Inj: 1

Detector A Parameters:
Threshold: 2 Width: 0.01
Calibration Type: External

Area Reject: 0
Quantitation: Height

Detector B Parameters:
Threshold: 0 Width: 0.03
Calibration Type: External

Area Reject: 0
Quantiation: Height

Sample Weight: 1041
Analyst: 2280

Dilution Factor: 200

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.002	9008	-.49188	TCX	2.343	6195	.27445	TCX
7.977	2258	-.68999	DCB	6.804	4216	.23464	DCB

Files:
Area File: C:\CPWIN\DATA1\3P7307.48A
Area File: C:\CPWIN\DATA1\3P7307B.48A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1P7307.CAL
Calibration File B: C:\CPWIN\DATA1\1P7307B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/6/2009 1:34:44 AM
File Reported On: 11/6/2009 at 1:34:57 AM

10X REPLOTS

ASX22 0860

C:\CPWIN\DATA1\3P7307.48R



C:\CPWIN\DATA1\3P7307B.48R



ASX22 8861

Multiple Component Data Summary

Sample Name: BLANKA 11/4/09ACF PBLK8D BLK Sample ID:AB **Batchnumber: 093070028A**
Sample Amount:30 g Total Volume: 10 ml Analyst: 1151 SDG: State:
Analyses: 01216

Analysis Report (A)

Injected on	Nov 07, 2009 10:24:40
Instrument	S9147A
Result file	2C6310.03R
Calibration file	1C6310
Method file	PESTD

%SSR(TCX)	97.9%	Conc:	10.02148
%SSR(DCB)	106.4%	Conc:	10.81407

Analysis Report (B)

Injected on	Nov 07, 2009 10:24:40
Instrument	S9147B
Result file	2C6310B.03R
Calibration file	1C6310B
Method file	PESTDB

%SSR(TCX)	87.8%	Conc:	8.983605
%SSR(DCB)	91.4%	Conc:	9.29414

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<17	<3.3			4	40	
Aroclor-1221			<17	<4			3	5	
Aroclor-1232			<17	<5.3			4	10	
Aroclor-1242			<17	<3.3			4	30	
Aroclor-1248			<17	<3.3			4	40	
Aroclor-1254			<17	<3.3			4	40	
Aroclor-1260			<17	<3.3			4	40	

Units: ug/kg

%Difference = High - Low divided by the Average times 100

Reviewed By: Andrea Carey Date: 11-10-09

Verified By: Mark [illegible] Date: 11/11/09

ASX22 0862

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: BLANKA 11/4/09ACF PBLK8D Sample ID: AB **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10 ml Analyst: 2280 SDG: State:
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on : NOV 07, 2009 10:24:40
Instrument : CP06–S9147A
Result file : 2C6310.03R
Calibration file : 1C6310.CAL
Method file : PESTD.MET
%SSR(TCX) : 97.9% Conc.: 10.021488
%SSR(DCB) : 106.4% Conc.: 10.814075

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1221					1		
+ 3.49	3.49	3.55	120.225357	2.16798			2
3.49	3.51	3.55	693.495972	12.505563			2
Height Summation:			**693.495972**				
Amount Avg CF:			**12.505563**	Linear:			
Aroclor-1248					1		
4.70	4.71	4.76	108.437897	1.385465			6
Height Summation:			**108.437897**				
Amount Avg CF:			**1.385465**	Linear:			
Aroclor-1254					2	**21.81**	
4.68	4.71	4.74	108.437897	1.00675			2
5.13	5.15	5.19	82.372246	0.737769			4
Height Summation:			**190.810143**				
Amount Avg CF:			**0.87226**	Linear:			
Aroclor-1262					1		
5.73	5.77	5.79	67.850677	0.636898			3
Height Summation:			**67.850677**				
Amount Avg CF:			**0.636898**	Linear:			
Aroclor-1268					2	**3.31**	
6.51	6.53	6.57	38.614517	0.248009			4
6.96	6.99	7.02	95.315315	0.25989			6
Height Summation:			**133.929832**				
Amount Avg CF:			**0.25395**	Linear:			

Analysis Report (B)

Injected on : NOV 07, 2009 10:24:40
Instrument : CP06–S9147B
Result file : 2C6310B.03R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET
%SSR(TCX) : 87.8% Conc.: 8.983605
%SSR(DCB) : 91.4% Conc.: 9.29414

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					4	**162.46**	
3.07	3.09	3.13	145.478226	0.560426			1
+ 3.07	3.12	3.13	87.436852	0.336833			1
+ 3.40	3.42	3.46	393.526154	1.117092			2
3.40	3.45	3.46	3553.284912	10.08661			2
3.56	3.56	3.62	363.923279	0.580308			3
3.67	3.69	3.73	166.815765	0.512009			5
Height Summation:			**4229.502182**				
Amount Avg CF:			**2.934838**	Linear:			
Aroclor-1221					3	**156.76**	
2.95	2.96	3.01	163.007034	1.002873			1
3.02	3.03	3.08	1810.431763	20.702479			2
3.07	3.09	3.13	145.478226	0.400899			3
+ 3.07	3.12	3.13	87.436852	0.240953			3
Height Summation:			**2118.917023**				
Amount Avg CF:			**7.36875**	Linear:			
Aroclor-1232					4	**168.41**	
3.07	3.09	3.13	145.478226	0.512256			1
+ 3.07	3.12	3.13	87.436852	0.307882			1
+ 3.40	3.42	3.46	393.526154	2.484382			2
3.40	3.45	3.46	3553.284912	22.432354			2
3.56	3.56	3.62	363.923279	1.304595			3
3.67	3.69	3.73	166.815765	1.207235			5
Height Summation:			**4229.502182**				
Amount Avg CF:			**6.36411**	Linear:			
Aroclor-1242					4	**162.44**	
3.07	3.09	3.13	145.478226	0.718837			1
+ 3.07	3.12	3.13	87.436852	0.432043			1
+ 3.40	3.42	3.46	393.526154	1.460425			2
3.40	3.45	3.46	3553.284912	13.186688			2
3.56	3.56	3.62	363.923279	0.762247			3
3.67	3.69	3.73	166.815765	0.681103			5
Height Summation:			**4229.502182**				
Amount Avg CF:			**3.837219**	Linear:			
Aroclor-1248					4	**40.47**	
3.28	3.28	3.34	160.936859	0.869445			1
3.56	3.56	3.62	363.923279	1.142996			2
3.71	3.74	3.77	267.190613	0.684467			3
4.45	4.46	4.51	82.570206	0.398553			6
Height Summation:			**874.620957**				
Amount Avg CF:			**0.773865**	Linear:			
Aroclor-1254					4	**83.27**	
4.18	4.20	4.24	169.454895	0.335545			2
4.45	4.46	4.51	82.570206	0.139781			3
4.79	4.83	4.85	367.416962	1.491219			5
4.92	4.96	4.98	462.061157	1.052529			6
Height Summation:			**1081.503220**				
Amount Avg CF:			**0.754769**	Linear:			

ASX22 ■963

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/9/09 05:32:50

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: BLANKA 11/4/09ACF PBLK8D Sample ID: AB **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10 ml Analyst: 2280 SDG: State:
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on : NOV 07, 2009 10:24:40
Instrument : CP06--S9147A
Result file : 2C6310.03R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Analysis Report (B)

Injected on : NOV 07, 2009 10:24:40
Instrument : CP06--S9147B
Result file : 2C6310B.03R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					4	42.62	
4.55	4.58	4.61	203.41774	0.413706			1
4.68	4.69	4.74	252.66658	0.48099			2
4.79	4.83	4.85	367.416962	0.764851			3
4.92	4.96	4.98	462.061157	1.041868			4
Height Summation:			**1285.562439**				
Amount Avg CF:			**0.675354**	Linear:			
Aroclor-1262					2	18.90	
4.68	4.69	4.74	252.66658	0.540684			1
4.95	4.96	5.01	462.061157	0.707485			2
Height Summation:			**714.727737**				
Amount Avg CF:			**0.624085**	Linear:			
Aroclor-1268					1		
6.17	6.20	6.23	641.920471	0.239731			6
Height Summation:			**641.920471**				
Amount Avg CF:			**0.239731**	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			17	3.3			4	40	
Aroclor-1221			17	3.3		** 51.69	3	5	
Aroclor-1232			17	3.3			4	10	
Aroclor-1242			17	3.3			4	30	
Aroclor-1248			17	3.3		** 56.65	4	40	
Aroclor-1254			17	3.3		14.44	4	40	
Aroclor-1260			17	3.3			4	40	
Aroclor-1262			17	3.3		2.03	4	40	
Aroclor-1268			17	3.3		5.76	4	40	

Units: _____

ASX22 9864

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/9/09 05:32:50

Sample Name: **BLANKA 11/4/09ACF ABPBLK8D BLK 093070028A 01216**
Acquired from CP06--S9147A via port 1 on 11/7/09 10:32:10am by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\2C6310.03R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.021		0.0000	0.000	102.9	0.112	BB	0.020	84.19	0.112
2	3.082		0.0000	0.000	189.4	0.206	BB	0.018	171.37	0.228
3	3.119		0.0000	0.000	1619.1	1.762	BB	0.023	1177.91	1.567
4	3.249		0.0000	0.000	256.4	0.279	BB	0.019	222.70	0.296
5	3.311	TCX	300.6447	48.098	59954.0	65.263	BB	0.017	58809.07	78.243
6	3.493		0.0000	0.000	84.6	0.092	BB	0.012	120.23	0.160
7	3.513		0.0000	0.000	548.0	0.597	BB	0.013	693.50	0.923
8	3.647		0.0000	0.000	200.2	0.218	BB	0.019	173.36	0.231
9	3.739		0.0000	0.000	53.1	0.058	BB	0.026	33.75	0.045
10	3.862		0.0000	0.000	692.9	0.754	BB	0.020	570.73	0.759
11	4.000		0.0000	0.000	1195.7	1.302	BB	0.019	1054.59	1.403
12	4.075		0.0000	0.000	1181.3	1.286	BB	0.020	967.65	1.287
13	4.343		0.0000	0.000	108.7	0.118	BB	0.028	64.06	0.085
14	4.707		0.0000	0.000	463.4	0.504	BB	0.071	108.44	0.144
15	5.031		0.0000	0.000	54.5	0.059	BB	0.034	26.77	0.036
16	5.148		0.0000	0.000	128.2	0.140	BB	0.026	82.37	0.110
17	5.443		0.0000	0.000	187.3	0.204	BB	0.039	79.14	0.105
18	5.766		0.0000	0.000	139.8	0.152	BB	0.034	67.85	0.090
19	5.860		0.0000	0.000	60.1	0.065	BB	0.027	37.03	0.049
20	6.007		0.0000	0.000	230.0	0.250	BB	0.050	76.22	0.101
21	6.277		0.0000	0.000	59.1	0.064	BB	0.026	38.42	0.051
22	6.457		0.0000	0.000	282.6	0.308	BB	0.113	41.73	0.056
23	6.527		0.0000	0.000	39.5	0.043	BB	0.017	38.61	0.051
24	6.907		0.0000	0.000	128.1	0.139	BB	0.062	34.20	0.045
25	6.993		0.0000	0.000	154.2	0.168	BB	0.027	95.32	0.127
26	7.161	DCB	324.4223	51.902	23752.6	25.856	BB	0.038	10292.91	13.694

Total Area = 91865.7, Total Amount = 625.067, Total Height = 75162.12, Sample Units = PPB

ASX22 0865

Sample Name: **BLANKA 11/4/09ACF ABPBLK8D BLK 093070028A 01216**
Acquired from CP06--S9147B via port 2 on 11/7/09 10:32:10am by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\2C6310B.03R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.538		0.0000	0.000	227.9	0.098	BB	0.022	174.93	0.082
2	2.586		0.0000	0.000	224.8	0.097	BV	0.016	238.81	0.112
3	2.623		0.0000	0.000	2582.1	1.109	VV	0.019	2290.87	1.078
4	2.651		0.0000	0.000	1709.0	0.734	VB	0.017	1661.91	0.782
5	2.725		0.0000	0.000	127.5	0.055	BB	0.020	106.09	0.050
6	2.807	TCX	269.5081	49.151	128467.2	55.183	BB	0.017	123311.00	58.004
7	2.932		0.0000	0.000	415.5	0.178	BB	0.013	536.64	0.252
8	2.958		0.0000	0.000	173.4	0.074	BB	0.018	163.01	0.077
9	3.029		0.0000	0.000	1646.5	0.707	BB	0.015	1810.43	0.852
10	3.090		0.0000	0.000	101.4	0.044	BB	0.012	145.48	0.068
11	3.117		0.0000	0.000	55.9	0.024	BB	0.011	87.44	0.041
12	3.147		0.0000	0.000	401.5	0.172	BB	0.017	398.14	0.187
13	3.192		0.0000	0.000	144.3	0.062	BB	0.021	112.11	0.053
14	3.283		0.0000	0.000	190.8	0.082	BV	0.020	160.94	0.076
15	3.349		0.0000	0.000	1691.8	0.727	VV	0.014	2062.13	0.970
16	3.392		0.0000	0.000	212.6	0.091	VV	0.022	160.97	0.076
17	3.421		0.0000	0.000	503.5	0.216	VV	0.021	393.53	0.185
18	3.449		0.0000	0.000	3263.1	1.402	VV	0.015	3553.28	1.671
19	3.494		0.0000	0.000	740.2	0.318	VV	0.022	553.59	0.260
20	3.524		0.0000	0.000	2814.9	1.209	VV	0.014	3349.32	1.575
21	3.558		0.0000	0.000	652.0	0.280	VV	0.030	363.92	0.171
22	3.665		0.0000	0.000	139.7	0.060	VV	0.019	123.48	0.058
23	3.694		0.0000	0.000	245.4	0.105	VV	0.025	166.82	0.078
24	3.744		0.0000	0.000	410.5	0.176	VB	0.026	267.19	0.126
25	3.810		0.0000	0.000	177.3	0.076	BB	0.026	112.54	0.053
26	3.927		0.0000	0.000	118.5	0.051	BB	0.017	115.61	0.054
27	3.958		0.0000	0.000	208.8	0.090	BB	0.041	85.65	0.040
28	4.053		0.0000	0.000	630.7	0.271	BV	0.031	343.65	0.162
29	4.153		0.0000	0.000	105.3	0.045	VV	0.026	68.78	0.032
30	4.203		0.0000	0.000	475.1	0.204	VB	0.047	169.45	0.080
31	4.307		0.0000	0.000	368.6	0.158	BB	0.022	279.09	0.131
32	4.460		0.0000	0.000	274.8	0.118	BB	0.055	82.57	0.039
33	4.582		0.0000	0.000	202.0	0.087	BB	0.017	203.42	0.096
34	4.692		0.0000	0.000	500.6	0.215	BV	0.033	252.67	0.119
35	4.835		0.0000	0.000	906.6	0.389	VV	0.041	367.42	0.173
36	4.964		0.0000	0.000	853.2	0.366	VB	0.031	462.06	0.217
37	5.417		0.0000	0.000	266.2	0.114	BB	0.027	167.31	0.079
38	5.541		0.0000	0.000	420.1	0.180	BB	0.025	274.57	0.129
39	6.150		0.0000	0.000	140.4	0.060	BB	0.071	32.94	0.015
40	6.199		0.0000	0.000	700.6	0.301	BB	0.018	641.92	0.302
41	6.317		0.0000	0.000	38.5	0.017	BB	0.022	29.30	0.014
42	6.359	DCB	278.8242	50.849	78799.4	33.848	BB	0.020	66547.24	31.303
43	6.690		0.0000	0.000	139.3	0.060	BB	0.045	51.05	0.024
44	7.164		0.0000	0.000	154.7	0.066	BB	0.039	66.48	0.031
45	7.403		0.0000	0.000	178.3	0.077	BB	0.068	43.55	0.020

Total Area = 232800.6, Total Amount = 548.332, Total Height = 212589.2, Sample Units = PPB

ASX22. 0866

Lancaster Laboratories

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\2C6310.03R

BLANKA 11/4/09ACF ABPBLK8D BLK 093070028A 01216

C:\CPWIN\DATA1\2C6310.03R



Instrument ID: CP06--S9147A Injected On: 11/7/2009 10:24:40 AM Column ID: RTX CLP

File: C:\CPWIN\DATA1\2C6310B.03R

BLANKA 11/4/09ACF ABPBLK8D BLK 093070028A 01216

C:\CPWIN\DATA1\2C6310B.03R



ASX22 0867

Instrument ID: CP06--S9147B Injected On: 11/7/2009 10:24:40 AM Column ID: RTX CLPII

Printed on 11/7/2009 10:32:25 AM

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Volume Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 30 Dilution Factor: 10
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	58809	10.02149	TCX	2.807	123311	8.98361	TCX
7.161	10293	10.81408	DCB	6.359	66547	9.29414	DCB

Files:
Area File: C:\CPWIN\DATA1\2C6310.03A
Area File: C:\CPWIN\DATA1\2C6310B.03A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/7/2009 10:32:16 AM
File Reported On: 11/7/2009 at 10:32:24 AM

REPLOTS

ASX22. 0869

BLANKA 11/4/09ACF ABPBLK8D BLK 093070028A 01216
C:\CPWIN\DATA1\2C6310.03R



BLANKA 11/4/09ACF ABPBLK8D BLK 093070028A 01216
C:\CPWIN\DATA1\2C6310B.03R



ASX22 0869

Multiple Component Data Summary

Sample Name: LCSA 11/4/09ACF LCSDS LCS Sample ID:AB **Batchnumber: 093070028A**
Sample Amount:30 g Total Volume: 10 ml Analyst: 1151 SDG: State:
Analyses: 01216

Analysis Report (A)

Injected on Nov 07, 2009 10:36:39
Instrument S9147A
Result file 2C6310.04R
Calibration file 1C6310
Method file PESTD

%SSR(TCX) 107.9% Conc: 11.04460
%SSR(DCB) 123.0% Conc: 12.50299

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016						**3.51**	
3.54	3.56	3.60	14560.816406	150.68255			1
3.88	3.91	3.94	17433.164063	152.513114			2
4.08	4.11	4.14	23675.308594	154.15246			3
4.13	4.15	4.19	12759.310547	148.279178			4
4.19	4.22	4.25	11902.207031	154.710745			5
4.40	4.43	4.46	9384.029297	164.00198			6
Height summation:			**89714.835938**				
Concentration		CF:	**154.056671**	L:			
Aroclor-1260						**2.57**	
5.25	5.28	5.31	16783.384766	179.797193			1
5.45	5.48	5.51	15086.407227	177.102935			2
5.66	5.69	5.72	15204.579102	181.142751			3
5.89	5.92	5.95	12902.864258	189.986605			4
6.12	6.15	6.18	24151.800781	184.472477			5
6.33	6.36	6.39	11945.039063	186.418698			6
Height summation:			**96074.075197**				
Concentration		CF:	**183.153443**	L:			

Analysis Report (B)

Injected on Nov 07, 2009 10:36:39
Instrument S9147B
Result file 2C6310B.04R
Calibration file 1C6310B
Method file PESTDB

%SSR(TCX) 103.8% Conc: 10.61909
%SSR(DCB) 105.8% Conc: 10.75311

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016						**2.69**	
3.07	3.10	3.13	37732.046875	145.355155			1
3.40	3.43	3.46	51583.359375	146.428239			2
3.56	3.59	3.62	95563.023438	152.383805			3
3.59	3.62	3.65	51073.308594	149.825802			4
3.67	3.70	3.73	47617.351563	146.152201			5
3.84	3.87	3.90	46934.6875	155.415137			6
Height summation:			**330503.777345**				
Concentration		CF:	**149.260057**	L:			
Aroclor-1260						**2.66**	
4.55	4.57	4.61	82274.617188	167.32829			1
4.68	4.71	4.74	88323.632813	168.137683			2
4.79	4.82	4.85	84038.820313	174.943326			3
4.92	4.95	4.98	75625.09375	170.521505			4
5.34	5.37	5.40	146384	178.381808			5
5.60	5.63	5.66	95602.671875	176.428015			6
Height summation:			**572248.835939**				
Concentration		CF:	**172.623438**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016	A	154.056671	17	3.3		3.16	4	40	
Aroclor-1221			<17	<4			3	5	
Aroclor-1232			<17	<5.3			4	10	
Aroclor-1242			<17	<3.3	E Begins		4	30	
Aroclor-1248			<17	<3.3		11-10-09	4	40	
Aroclor-1254			<17	<3.3			4	40	
Aroclor-1260	A	183.153443	17	3.3		5.92	4	40	

Units: ug/kg

%Difference = High - Low divided by the Average times 100

Reviewed By: Andrea Carey Date: 11-10-09
Verified By: [illegible] Date: 11/11/09

ASX22 0878

11/10/2009

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **LCSA** 11/4/09ACF LCSDS Sample ID: AB **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10 ml Analyst: 2280 SDG: State:
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on : NOV 07, 2009 10:36:39
Instrument : CP06–S9147A
Result file : 2C6310.04R
Calibration file : 1C6310.CAL
Method file : PESTD.MET
%SSR(TCX) : 107.9% Conc.: 11.044606
%SSR(DCB) : 123% Conc.: 12.502991

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	3.51	
3.54	3.56	3.60	14560.816406	150.68255			1
3.88	3.91	3.94	17433.164063	152.513114			2
4.08	4.11	4.14	23675.308594	154.15246			3
4.13	4.15	4.19	12759.310547	148.279178			4
4.19	4.22	4.25	11902.207031	154.710745			5
4.40	4.43	4.46	9384.029297	164.00198			6
+ 4.40	4.46	4.46	2274.038574	39.742718			6
Height Summation:			89714.835938				
Amount Avg CF:			154.056671	Linear:			
Aroclor-1221					3	38.21	
3.43	3.46	3.49	2845.437744	49.658599			1
+ 3.49	3.49	3.55	82.518837	1.488032			2
E 3.49	3.51	3.55	4044.151367	72.926722			2
E 3.53	3.56	3.59	14560.816406	108.001902			3
Height Summation:			21450.405517				
Amount Avg CF:			76.862408	Linear:			
Aroclor-1232					6	28.52	
E 3.54	3.56	3.60	14560.816406	138.654634			1
E 3.88	3.91	3.94	17433.164063	342.733983			2
E 4.08	4.11	4.14	23675.308594	356.207156			3
E 4.13	4.15	4.19	12759.310547	338.353502			4
E 4.19	4.22	4.25	11902.207031	360.018361			5
E 4.40	4.43	4.46	9384.029297	393.707963			6
E+ 4.40	4.46	4.46	2274.038574	95.407534			6
Height Summation:			89714.835938				
Amount Avg CF:			321.6126	Linear:			
Aroclor-1242					6	2.68	
E 3.54	3.56	3.60	14560.816406	189.779295			1
E 3.88	3.91	3.94	17433.164063	196.684877			2
E 4.08	4.11	4.14	23675.308594	201.629268			3
E 4.13	4.15	4.19	12759.310547	197.222514			4
E 4.19	4.22	4.25	11902.207031	200.020285			5
E 4.40	4.43	4.46	9384.029297	205.452201			6
+ 4.40	4.46	4.46	2274.038574	49.78738			6
Height Summation:			89714.835938				
Amount Avg CF:			198.46474	Linear:			
Aroclor-1248					6	62.70	
E 3.88	3.91	3.94	17433.164063	401.016831			1
4.19	4.22	4.25	11902.207031	251.59505			2
+ 4.26	4.27	4.32	592.069946	6.959715			3
4.26	4.29	4.32	9042.320313	106.291454			3
4.40	4.43	4.46	9384.029297	111.601439			4
+ 4.40	4.46	4.46	2274.038574	27.044457			4
4.50	4.53	4.56	8361.603516	131.62946			5
4.70	4.71	4.76	9695.290039	123.87264			6
+ 4.70	4.75	4.76	2679.597412	34.236088			6
Height Summation:			65818.614259				
Amount Avg CF:			187.667812	Linear:			

Analysis Report (B)

Injected on : NOV 07, 2009 10:36:39
Instrument : CP06–S9147B
Result file : 2C6310B.04R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET
%SSR(TCX) : 103.8% Conc.: 10.619094
%SSR(DCB) : 105.8% Conc.: 10.753113

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	2.69	
3.07	3.10	3.13	37732.046875	145.355155			1
+ 3.40	3.40	3.46	3773.45459	10.7116			2
3.40	3.43	3.46	51583.359375	146.428239			2
3.56	3.59	3.62	95563.023438	152.383805			3
3.59	3.62	3.65	51073.308594	149.825802			4
3.67	3.70	3.73	47617.351563	146.152201			5
3.84	3.87	3.90	46934.6875	155.415137			6
+ 3.84	3.89	3.90	15663.847656	51.867801			6
Height Summation:			330503.777345				
Amount Avg CF:			149.260057	Linear:			
Aroclor-1221					3	35.47	
2.95	2.98	3.01	7926.403809	48.765866			1
+ 3.02	3.03	3.08	843.242432	9.642566			2
E 3.02	3.04	3.08	7257.440918	82.989605			2
E 3.07	3.10	3.13	37732.046875	103.979406			3
Height Summation:			52915.891602				
Amount Avg CF:			78.578292	Linear:			
Aroclor-1232					6	28.65	
E 3.07	3.10	3.13	37732.046875	132.861659			1
+ 3.40	3.40	3.46	3773.45459	23.822314			2
E 3.40	3.43	3.46	51583.359375	325.652521			2
E 3.56	3.59	3.62	95563.023438	342.575051			3
E 3.60	3.62	3.66	51073.308594	337.752925			4
E 3.67	3.70	3.73	47617.351563	344.603789			5
E 3.84	3.87	3.90	46934.6875	380.376753			6
E+ 3.84	3.89	3.90	15663.847656	126.945844			6
Height Summation:			330503.777345				
Amount Avg CF:			310.637116	Linear:			
Aroclor-1242					6	4.37	
E 3.07	3.10	3.13	37732.046875	186.44158			1
+ 3.40	3.40	3.46	3773.45459	14.003765			2
E 3.40	3.43	3.46	51583.359375	191.432344			2
E 3.56	3.59	3.62	95563.023438	200.159233			3
E 3.59	3.62	3.65	51073.308594	197.121938			4
E 3.67	3.70	3.73	47617.351563	194.420021			5
E 3.84	3.87	3.90	46934.6875	211.517553			6
E+ 3.84	3.89	3.90	15663.847656	70.59126			6
Height Summation:			330503.777345				
Amount Avg CF:			196.848778	Linear:			
Aroclor-1248					6	65.70	
E 3.28	3.31	3.34	56327.132813	304.301566			1
E 3.56	3.59	3.62	95563.023438	300.140591			2
+ 3.71	3.73	3.77	42318.972656	108.409282			3
3.71	3.76	3.77	46203.542969	118.360457			3
3.84	3.87	3.90	46934.6875	111.966791			4
+ 3.84	3.89	3.90	15663.847656	37.367475			4
+ 4.21	4.24	4.27	2457.125244	10.027347			5
4.21	4.26	4.27	3408.31958	13.909101			5
+ 4.45	4.46	4.51	14126.086914	68.184293			6
4.45	4.49	4.51	42028.113281	202.862774			6
Height Summation:			290464.819581				
Amount Avg CF:			175.25688	Linear:			

ASX22 8871

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/9/09 05:34:19

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **LCSA** 11/4/09ACF LCSDS Sample ID: AB **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10 ml Analyst: 2280 SDG: State:
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on : NOV 07, 2009 10:36:39
Instrument : CP06--S9147A
Result file : 2C6310.04R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					7	70.84	
+ 4.26	4.27	4.32	592.069946	8.063931			1
4.26	4.29	4.32	9042.320313	123.155462			1
4.26	4.32	4.32	7563.362793	103.012214			1
4.68	4.71	4.74	9695.290039	90.012237			2
4.85	4.88	4.91	7096.395508	55.181496			3
5.13	5.15	5.19	2581.519043	23.121449			4
5.35	5.37	5.41	1344.983521	24.584499			5
5.66	5.69	5.72	15204.579102	201.581128			6
Height Summation:			52528.450319				
Amount Avg CF:			88.664069	Linear:			
Aroclor-1260					6	2.57	
5.25	5.28	5.31	16783.384766	179.797193			1
5.45	5.48	5.51	15086.407227	177.102935			2
5.66	5.69	5.72	15204.579102	181.142751			3
5.89	5.92	5.95	12902.864258	189.986605			4
6.12	6.15	6.18	24151.800781	184.472477			5
6.33	6.36	6.39	11945.039063	186.418698			6
Height Summation:			96074.075197				
Amount Avg CF:			183.153443	Linear:			
Aroclor-1262					6	23.55	
5.25	5.28	5.31	16783.384766	197.575244			1
5.45	5.48	5.51	15086.407227	194.504704			2
5.73	5.75	5.79	11487.595703	107.831336			3
5.89	5.92	5.95	12902.864258	156.990924			4
6.12	6.15	6.18	24151.800781	143.364419			5
6.32	6.36	6.38	11945.039063	125.547214			6
Height Summation:			92357.091798				
Amount Avg CF:			154.302307	Linear:			
Aroclor-1268					6	114.20	
5.90	5.92	5.96	12902.864258	216.760143			1
6.32	6.36	6.38	11945.039063	65.777226			2
6.38	6.41	6.44	4710.437012	32.445585			3
6.51	6.54	6.57	283.335449	1.819773			4
6.74	6.77	6.80	5849.986816	108.299396			5
6.96	6.99	7.02	1560.083862	4.25378			6
Height Summation:			37251.746460				
Amount Avg CF:			71.559317	Linear:			

Analysis Report (B)

Injected on : NOV 07, 2009 10:36:39
Instrument : CP06--S9147B
Result file : 2C6310B.04R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					7	99.39	
+ 4.07	4.08	4.13	8736.235352	14.197744			1
4.07	4.10	4.13	44758.398438	72.739374			1
4.07	4.13	4.13	9919.626953	16.12094			1
4.18	4.21	4.24	30987.908203	61.360493			2
+ 4.45	4.46	4.51	14126.086914	23.91371			3
4.45	4.49	4.51	42028.113281	71.148375			3
+ 4.62	4.64	4.68	3061.686035	9.111103			4
4.62	4.67	4.68	17270.957031	51.395692			4
E 4.79	4.82	4.85	84038.820313	341.084636			5
4.92	4.95	4.98	75625.09375	172.266414			6
Height Summation:			304628.917969				
Amount Avg CF:			112.302275	Linear:			
Aroclor-1260					6	2.66	
4.55	4.57	4.61	82274.617188	167.32829			1
4.68	4.71	4.74	88323.632813	168.137683			2
4.79	4.82	4.85	84038.820313	174.943326			3
4.92	4.95	4.98	75625.09375	170.521505			4
5.34	5.37	5.40	146384	178.381808			5
5.60	5.63	5.66	95602.671875	176.428015			6
Height Summation:			572248.835939				
Amount Avg CF:			172.623438	Linear:			
Aroclor-1262					6	24.07	
4.68	4.71	4.74	88323.632813	189.004614			1
4.95	4.98	5.01	70890.726563	108.544282			2
5.15	5.18	5.21	72723.359375	118.260441			3
5.34	5.37	5.40	146384	138.476677			4
5.60	5.63	5.66	95602.671875	133.531043			5
5.97	6.00	6.03	38581.273438	100.798875			6
Height Summation:			512505.664064				
Amount Avg CF:			131.435989	Linear:			
Aroclor-1268					6	107.34	
5.15	5.18	5.21	72723.359375	204.186298			1
+ 5.55	5.56	5.61	726.44635	0.599267			2
5.55	5.58	5.61	36164.148438	29.832853			2
5.60	5.63	5.66	95602.671875	93.638657			3
5.75	5.78	5.81	1868.775513	1.776316			4
5.97	6.00	6.03	38581.273438	99.773341			5
6.17	6.20	6.23	10856.868164	4.054597			6
Height Summation:			255797.096803				
Amount Avg CF:			72.210344	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			17	3.3		3.16	4	40	
Aroclor-1221			17	3.3	E	2.21	3	5	
Aroclor-1232			17	3.3	E	3.47	4	10	
Aroclor-1242			17	3.3	E	0.82	4	30	
Aroclor-1248			17	3.3	E	6.84	4	40	
Aroclor-1254			17	3.3		23.52	4	40	
Aroclor-1260			17	3.3		5.92	4	40	
Aroclor-1262			17	3.3		16.01	4	40	

ASX22 8872

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/9/09 05:34:19

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **LCSA** 11/4/09ACF LCSDS Sample ID: AB **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10 ml Analyst: 2280 SDG: State:
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on	: NOV 07, 2009 10:36:39
Instrument	: CP06–S9147A
Result file	: 2C6310.04R
Calibration file	: 1C6310.CAL
Method file	: PESTD.MET

Analysis Report (B)

Injected on	: NOV 07, 2009 10:36:39
Instrument	: CP06–S9147B
Result file	: 2C6310B.04R
Calibration file	: 1C6310B.CAL
Method file	: PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1268			17	3.3		0.91	4	40	

Units: _____

ASX22 0873

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/9/09 05:34:19

Lancaster Laboratories

Sample Name: **LCSA 11/4/09ACF ABLCSDS LCS 093070028A 01216**
Acquired from CP06--S9147A via port 1 on 11/7/09 10:44:09am by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\2C6310.04R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.018		0.0000	0.000	58.1	0.011	BB	0.010	94.82	0.022
2	3.075		0.0000	0.000	87.0	0.016	BB	0.013	114.64	0.027
3	3.114		0.0000	0.000	1656.7	0.305	BB	0.024	1174.22	0.274
4	3.203		0.0000	0.000	44.1	0.008	BB	0.013	57.56	0.013
5	3.247		0.0000	0.000	263.8	0.049	BB	0.015	294.83	0.069
6	3.310	TCX	331.3382	46.903	62028.3	11.417	BB	0.016	64813.04	15.142
7	3.456		0.0000	0.000	2886.7	0.531	BB	0.017	2845.44	0.665
8	3.493		0.0000	0.000	53.7	0.010	BB	0.011	82.52	0.019
9	3.514		0.0000	0.000	3398.9	0.626	BB	0.014	4044.15	0.945
10	3.563		0.0000	0.000	13388.9	2.464	BB	0.015	14560.82	3.402
11	3.660		0.0000	0.000	2336.4	0.430	BB	0.015	2536.52	0.593
12	3.787		0.0000	0.000	10756.1	1.980	BB	0.013	13795.19	3.223
13	3.805		0.0000	0.000	2690.6	0.495	BB	0.008	5820.19	1.360
14	3.845		0.0000	0.000	912.4	0.168	BB	0.011	1369.77	0.320
15	3.862		0.0000	0.000	605.0	0.111	BB	0.016	636.82	0.149
16	3.913		0.0000	0.000	17182.5	3.163	BB	0.016	17433.16	4.073
17	3.997		0.0000	0.000	1237.6	0.228	BB	0.015	1403.10	0.328
18	4.032		0.0000	0.000	5522.7	1.017	BB	0.023	3963.29	0.926
19	4.111		0.0000	0.000	39213.1	7.218	BB	0.028	23675.31	5.531
20	4.154		0.0000	0.000	13197.3	2.429	BB	0.017	12759.31	2.981
21	4.220		0.0000	0.000	12253.1	2.255	BB	0.017	11902.21	2.781
22	4.265		0.0000	0.000	454.9	0.084	BB	0.013	592.07	0.138
23	4.291		0.0000	0.000	8123.6	1.495	BB	0.015	9042.32	2.113
24	4.322		0.0000	0.000	12601.2	2.319	BB	0.028	7563.36	1.767
25	4.431		0.0000	0.000	8719.7	1.605	BB	0.015	9384.03	2.192
26	4.458		0.0000	0.000	1747.8	0.322	BB	0.013	2274.04	0.531
27	4.493		0.0000	0.000	2311.1	0.425	BB	0.014	2676.50	0.625
28	4.530		0.0000	0.000	9905.3	1.823	BB	0.020	8361.60	1.953
29	4.577		0.0000	0.000	2557.1	0.471	BB	0.018	2386.45	0.558
30	4.668		0.0000	0.000	336.0	0.062	BB	0.015	373.77	0.087
31	4.707		0.0000	0.000	15191.7	2.796	BV	0.026	9695.29	2.265
32	4.754		0.0000	0.000	4603.1	0.847	VV	0.029	2679.60	0.626
33	4.826		0.0000	0.000	1586.4	0.292	VV	0.024	1103.56	0.258
34	4.876		0.0000	0.000	12320.1	2.268	VV	0.029	7096.40	1.658
35	5.029		0.0000	0.000	274.9	0.051	VB	0.021	217.20	0.051
36	5.106		0.0000	0.000	5703.7	1.050	BV	0.025	3780.22	0.883
37	5.153		0.0000	0.000	3008.9	0.554	VB	0.019	2581.52	0.603
38	5.202		0.0000	0.000	11464.5	2.110	BB	0.023	8274.69	1.933
39	5.239		0.0000	0.000	720.8	0.133	BB	0.011	1049.32	0.245
40	5.282		0.0000	0.000	23334.1	4.295	BB	0.023	16783.38	3.921
41	5.373		0.0000	0.000	1493.8	0.275	BB	0.019	1344.98	0.314
42	5.430		0.0000	0.000	2676.3	0.493	BB	0.019	2296.93	0.537
43	5.483		0.0000	0.000	25047.3	4.610	BB	0.028	15086.41	3.525
44	5.537		0.0000	0.000	3756.1	0.691	BB	0.016	3952.78	0.923
45	5.577		0.0000	0.000	10462.0	1.926	BB	0.025	7108.92	1.661
46	5.655		0.0000	0.000	722.0	0.133	BB	0.017	688.06	0.161
47	5.692		0.0000	0.000	22069.9	4.062	BB	0.024	15204.58	3.552
48	5.754		0.0000	0.000	13181.9	2.426	BB	0.019	11487.60	2.684
49	5.796		0.0000	0.000	6644.3	1.223	BB	0.018	6130.55	1.432
50	5.918		0.0000	0.000	20319.4	3.740	BB	0.026	12902.86	3.014
51	5.975		0.0000	0.000	5352.2	0.985	BB	0.019	4799.82	1.121
52	6.014		0.0000	0.000	1774.5	0.327	BB	0.016	1900.76	0.444
53	6.096		0.0000	0.000	2925.6	0.538	BB	0.019	2544.06	0.594
54	6.150		0.0000	0.000	45434.2	8.363	BB	0.031	24151.80	5.643
55	6.356		0.0000	0.000	14894.8	2.742	BB	0.021	11945.04	2.791
56	6.405		0.0000	0.000	7336.7	1.350	BB	0.026	4710.44	1.100
57	6.539		0.0000	0.000	326.3	0.060	BB	0.019	283.34	0.066
58	6.624		0.0000	0.000	4907.9	0.903	BB	0.028	2890.80	0.675
59	6.774		0.0000	0.000	11318.9	2.083	BB	0.032	5849.99	1.367
60	6.987		0.0000	0.000	2767.9	0.509	BB	0.030	1560.08	0.364
61	7.159	DCB	375.0897	53.097	25129.8	4.626	BB	0.035	11900.43	2.780

ASX22: 6874

Printed on 11/7/2009 10:44:14 AM

Total Area = 543279.6, Total Amount = 706.428, Total Height = 428032.5, Sample Units = PPB

ASX22 0875

Lancaster Laboratories

Sample Name: **LCSA 11/4/09ACF ABLCSDS LCS 093070028A 01216**
Acquired from CP06--S9147B via port 2 on 11/7/09 10:44:09am by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\2C6310B.04R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.526		0.0000	0.000	244.2	0.012	BB	0.023	177.87	0.009
2	2.585		0.0000	0.000	292.9	0.014	BV	0.017	293.77	0.014
3	2.623		0.0000	0.000	1946.5	0.092	VV	0.015	2179.08	0.107
4	2.650		0.0000	0.000	1333.0	0.063	VB	0.013	1662.08	0.082
5	2.725		0.0000	0.000	86.8	0.004	BV	0.013	110.61	0.005
6	2.760		0.0000	0.000	84.7	0.004	VB	0.008	170.51	0.008
7	2.782		0.0000	0.000	168.1	0.008	BB	0.008	340.88	0.017
8	2.807	TCX	318.5728	49.686	122940.3	5.814	BB	0.014	145760.00	7.186
9	2.852		0.0000	0.000	2117.5	0.100	BB	0.011	3083.16	0.152
10	2.931		0.0000	0.000	392.0	0.019	BB	0.011	575.60	0.028
11	2.981		0.0000	0.000	6369.4	0.301	BB	0.013	7926.40	0.391
12	3.028		0.0000	0.000	489.0	0.023	BB	0.010	843.24	0.042
13	3.044		0.0000	0.000	4943.8	0.234	BB	0.011	7257.44	0.358
14	3.097		0.0000	0.000	30477.8	1.441	BB	0.013	37732.05	1.860
15	3.146		0.0000	0.000	192.5	0.009	BB	0.009	350.03	0.017
16	3.211		0.0000	0.000	5066.0	0.240	BB	0.013	6392.11	0.315
17	3.307		0.0000	0.000	67913.9	3.212	BV	0.020	56327.13	2.777
18	3.359		0.0000	0.000	8268.6	0.391	VV	0.018	7545.31	0.372
19	3.380		0.0000	0.000	2748.4	0.130	VV	0.013	3466.75	0.171
20	3.404		0.0000	0.000	2788.3	0.132	VV	0.012	3773.45	0.186
21	3.431		0.0000	0.000	48047.3	2.272	VB	0.016	51583.36	2.543
22	3.512		0.0000	0.000	21822.2	1.032	BV	0.023	15640.61	0.771
23	3.586		0.0000	0.000	121966.4	5.768	VV	0.021	95563.02	4.711
24	3.623		0.0000	0.000	62230.6	2.943	VV	0.020	51073.31	2.518
25	3.699		0.0000	0.000	41212.2	1.949	VV	0.014	47617.35	2.348
26	3.734		0.0000	0.000	36375.1	1.720	VV	0.014	42318.97	2.086
27	3.759		0.0000	0.000	56863.2	2.689	VB	0.021	46203.54	2.278
28	3.869		0.0000	0.000	39188.8	1.853	BV	0.014	46934.69	2.314
29	3.893		0.0000	0.000	10663.5	0.504	VB	0.011	15663.85	0.772
30	3.916		0.0000	0.000	8645.1	0.409	BB	0.011	12952.32	0.639
31	3.948		0.0000	0.000	8759.9	0.414	BV	0.013	11530.22	0.568
32	3.964		0.0000	0.000	29340.0	1.388	VV	0.014	34927.36	1.722
33	4.009		0.0000	0.000	10780.5	0.510	VB	0.015	12158.92	0.599
34	4.054		0.0000	0.000	1614.6	0.076	BV	0.014	1879.60	0.093
35	4.078		0.0000	0.000	7118.1	0.337	VV	0.014	8736.24	0.431
36	4.102		0.0000	0.000	39085.9	1.848	VV	0.015	44758.40	2.207
37	4.128		0.0000	0.000	11976.9	0.566	VV	0.020	9919.63	0.489
38	4.168		0.0000	0.000	4242.4	0.201	VV	0.015	4618.58	0.228
39	4.205		0.0000	0.000	28322.0	1.339	VV	0.015	30987.91	1.528
40	4.241		0.0000	0.000	2484.3	0.117	VV	0.017	2457.13	0.121
41	4.259		0.0000	0.000	3619.7	0.171	VB	0.018	3408.32	0.168
42	4.357		0.0000	0.000	894.9	0.042	BB	0.015	1015.01	0.050
43	4.415		0.0000	0.000	4671.3	0.221	BB	0.015	5216.92	0.257
44	4.461		0.0000	0.000	12375.8	0.585	BV	0.015	14126.09	0.696
45	4.491		0.0000	0.000	42098.4	1.991	VB	0.017	42028.11	2.072
46	4.520		0.0000	0.000	14588.3	0.690	BB	0.016	14870.27	0.733
47	4.573		0.0000	0.000	82972.6	3.924	BV	0.017	82274.62	4.056
48	4.643		0.0000	0.000	2677.2	0.127	VV	0.015	3061.69	0.151
49	4.669		0.0000	0.000	16342.6	0.773	VV	0.016	17270.96	0.851
50	4.712		0.0000	0.000	100034.4	4.731	VB	0.019	88323.63	4.354
51	4.820		0.0000	0.000	93244.8	4.410	BV	0.018	84038.82	4.143
52	4.877		0.0000	0.000	9690.9	0.458	VB	0.016	9911.37	0.489
53	4.914		0.0000	0.000	19112.1	0.904	BV	0.016	20338.12	1.003
54	4.951		0.0000	0.000	91967.9	4.349	VV	0.020	75625.09	3.728
55	4.984		0.0000	0.000	71617.1	3.387	VB	0.017	70890.73	3.495
56	5.024		0.0000	0.000	33328.4	1.576	BB	0.015	36071.36	1.778
57	5.125		0.0000	0.000	15220.5	0.720	BB	0.018	14206.03	0.700
58	5.178		0.0000	0.000	82292.8	3.892	BV	0.019	72723.36	3.585
59	5.251		0.0000	0.000	38266.7	1.810	VV	0.017	36642.23	1.806
60	5.286		0.0000	0.000	16907.8	0.800	VV	0.017	16689.48	0.823
61	5.312		0.0000	0.000	12147.0	0.574	VV	0.017	12099.79	0.597
62	5.366		0.0000	0.000	168694.0	7.978	VB	0.019	146384.00	7.217

ASX22 0876

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.442		0.0000	0.000	5484.0	0.259	BB	0.015	6140.36	0.303
64	5.557		0.0000	0.000	643.8	0.030	BB	0.015	726.45	0.036
65	5.580		0.0000	0.000	34478.2	1.631	BB	0.016	36164.15	1.783
66	5.625		0.0000	0.000	119924.4	5.672	BB	0.021	95602.67	4.713
67	5.780		0.0000	0.000	1716.0	0.081	BB	0.015	1868.78	0.092
68	5.819		0.0000	0.000	1252.4	0.059	BB	0.015	1427.59	0.070
69	5.847		0.0000	0.000	1109.9	0.052	BB	0.014	1312.97	0.065
70	5.896		0.0000	0.000	19818.9	0.937	BB	0.019	17822.61	0.879
71	5.998		0.0000	0.000	49051.4	2.320	BB	0.021	38581.27	1.902
72	6.198		0.0000	0.000	11794.1	0.558	BB	0.018	10856.87	0.535
73	6.358	DCB	322.5934	50.314	86312.8	4.082	BB	0.019	76993.68	3.796
74	6.878		0.0000	0.000	67.5	0.003	BB	0.030	38.11	0.002
75	7.013		0.0000	0.000	50.1	0.002	BB	0.019	43.37	0.002
76	7.057		0.0000	0.000	306.5	0.014	BB	0.081	63.36	0.003
77	7.289		0.0000	0.000	93.5	0.004	BB	0.033	47.93	0.002

Total Area = 2114472.0, Total Amount = 641.166, Total Height = 2028399.0, Sample Units = PPB

ASX22: 8877

Lancaster Laboratories

File: C:\CPWIN\DATA1\2C6310.04R

LANCASTER LABORATORIES
LCSA 11/4/09ACF ABLCSDS LCS 093070028A 01216
C:\CPWIN\DATA1\2C6310.04R



Instrument ID: CP06--S9147A Injected On: 11/7/2009 10:36:39 AM Column ID: RTX CLP

File: C:\CPWIN\DATA1\2C6310B.04R

LCSA 11/4/09ACF ABLCSDS LCS 093070028A 01216
C:\CPWIN\DATA1\2C6310B.04R



Instrument ID: CP06--S9147B Injected On: 11/7/2009 10:36:39 AM Column ID: RTX CLPII

Printed on 11/7/2009 10:44:24 AM

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 30 Dilution Factor: 10
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.31	64813	11.04461	TCX	2.807	145760	10.61909	TCX
7.159	11900	12.50299	DCB	6.358	76994	10.75311	DCB

Files:
Area File: C:\CPWIN\DATA1\2C6310.04A
Area File: C:\CPWIN\DATA1\2C6310B.04A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/7/2009 10:44:14 AM
File Reported On: 11/7/2009 at 10:44:23 AM

REPLOTS

ASX22 0879

LCSA 11/4/09ACF ABLCSDS LCS 093070028A 01216
C:\CPWIN\DATA1\2C6310.04R



LCSA 11/4/09ACF ABLCSDS LCS 093070028A 01216
C:\CPWIN\DATA1\2C6310B.04R



ASX22 0888

Multiple Component Data Summary

Sample Name: 5822071 MSACF DF1000 STEA3MS MS Sample ID:AD **Batchnumber: 093070028A**
Sample Amount:30 g Total Volume: 10000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on Nov 07, 2009 11:48:35
Instrument S9147A
Result file 2C6310.10R
Calibration file 1C6310
Method file PESTD

%SSR(TCX) * 193.4% Conc: 19.79280
%SSR(DCB) * 0.0% Conc: 0

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**30.47**	
4.68	4.71	4.74	5787.859863	53735.185915			2
4.85	4.87	4.91	10544.015625	81990.152682			3
5.13	5.15	5.19	11357.291016	101721.901722			4
Height summation:			**27689.166504**				
Concentration		CF:	**79149.080106**	L:			
Aroclor-1260						**7.61**	
5.89	5.92	5.95	5485.69043	80773.359926			4
6.12	6.15	6.18	9105.746094	69550.074196			5
6.33	6.36	6.39	4716.222656	73603.115281			6
Height summation:			**19307.65918**				
Concentration		CF:	**74642.183134**	L:			

Analysis Report (B)

Injected on Nov 07, 2009 11:48:35
Instrument S9147B
Result file 2C6310B.10R
Calibration file 1C6310B
Method file PESTDB

%SSR(TCX) * 238.5% Conc: 24.41074
%SSR(DCB) * 1880.3% Conc: 191.1365

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**19.90**	
4.07	4.10	4.13	34566.769531	56176.388169			1
4.18	4.21	4.24	34796.472656	68901.996095			2
4.45	4.48	4.51	49542.332031	83869.013451			3
Height summation:			**118905.574218**				
Concentration		CF:	**69649.132572**	L:			
Aroclor-1260						**17.44**	
4.79	4.82	4.85	34247.808594	71293.546686			3
5.34	5.37	5.40	44200.605469	53862.333987			5
5.60	5.62	5.66	28666.335938	52901.709344			6
Height summation:			**107114.750001**				
Concentration		CF:	**59352.530006**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<17000	<3300			4	40	
Aroclor-1221			<17000	<4000			3	5	
Aroclor-1232			<17000	<5300			4	10	
Aroclor-1242			<17000	<3300			4	30	
Aroclor-1248			<17000	<3300			4	40	
Aroclor-1254	A	79149.080106	17000	3300		12.77	4	40	
Aroclor-1260	A	74642.183134	17000	3300		22.82	4	40	

Units: ug/kg

%Difference = High - Low divided by the Average times 100

Reviewed By: Andrea Cary Date: 11.1009

Verified By: [illegible] Date: 11/11/09

ASX22 0881

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822071** MSACF DF1000 STEA3MS Sample ID: AD **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10000 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on : NOV 07, 2009 11:48:35
Instrument : CP06--S9147A
Result file : 2C6310.10R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

diluted out

%SSR(TCX) : *193.4% Conc.: 19.792803
%SSR(DCB) : *0% Conc.: 0

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	160.66	
3.54	3.56	3.60	212.336029	2197.358536			1
3.88	3.91	3.94	167.280014	1463.440362			2
4.08	4.11	4.14	361.437592	2353.358724			3
4.13	4.16	4.19	126.235291	1467.012271			4
4.19	4.22	4.25	155.359375	2019.43762			5
4.40	4.43	4.46	1348.891724	23574.192558			6
Height Summation:			2371.540025				
Amount Avg CF:			5512.466679	Linear:			
Aroclor-1221					2	48.32	
3.43	3.45	3.49	44.281357	772.798551			1
3.53	3.56	3.59	212.336029	1574.95942			3
Height Summation:			256.617386				
Amount Avg CF:			1173.878986	Linear:			
Aroclor-1232					6	171.93	
3.54	3.56	3.60	212.336029	2021.959044			1
3.88	3.91	3.94	167.280014	3288.705672			2
4.08	4.11	4.14	361.437592	5438.013872			3
4.13	4.16	4.19	126.235291	3347.528268			4
4.19	4.22	4.25	155.359375	4699.315638			5
4.40	4.43	4.46	1348.891724	56592.898007			6
Height Summation:			2371.540025				
Amount Avg CF:			12564.73675	Linear:			
Aroclor-1242					6	158.69	
3.54	3.56	3.60	212.336029	2767.494676			1
3.88	3.91	3.94	167.280014	1887.290732			2
4.08	4.11	4.14	361.437592	3078.160382			3
4.13	4.16	4.19	126.235291	1951.237205			4
4.19	4.22	4.25	155.359375	2610.862533			5
4.40	4.43	4.46	1348.891724	29532.385857			6
Height Summation:			2371.540025				
Amount Avg CF:			6971.238564	Linear:			
Aroclor-1248					7	105.22	
3.88	3.91	3.94	167.280014	3847.959027			1
4.19	4.22	4.25	155.359375	3284.067368			2
4.26	4.29	4.32	1715.544556	20166.032561			3
+ 4.26	4.32	4.32	529.152893	6220.132513			3
4.40	4.43	4.46	1348.891724	16041.963675			4
4.50	4.53	4.56	774.731873	12195.930864			5
4.70	4.71	4.76	5787.859863	73949.049239			6
4.70	4.73	4.76	5084.635742	64964.250646			6
+ 4.70	4.75	4.76	3720.336914	47533.178783			6
Height Summation:			15034.303147				
Amount Avg CF:			27778.464769	Linear:			

Analysis Report (B)

Injected on : NOV 07, 2009 11:48:35
Instrument : CP06--S9147B
Result file : 2C6310B.10R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

%SSR(TCX) : *238.5% Conc.: 24.410742
%SSR(DCB) : *1880.3% Conc.: 191.136551

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	149.71	
+ 3.07	3.10	3.13	87.882454	338.549555			1
3.07	3.12	3.13	371.917969	1432.739497			1
3.40	3.43	3.46	443.461212	1258.840936			2
3.56	3.59	3.62	1322.143921	2108.276974			3
3.59	3.62	3.65	447.907471	1313.95631			4
3.67	3.70	3.73	633.745789	1945.159457			5
3.84	3.87	3.90	5059.392578	16753.199643			6
+ 3.84	3.89	3.90	896.521851	2968.658653			6
Height Summation:			8278.568940				
Amount Avg CF:			4135.362136	Linear:			
Aroclor-1221					1		
+ 3.07	3.10	3.13	87.882454	242.180484			3
3.07	3.12	3.13	371.917969	1024.90622			3
Height Summation:			371.917969				
Amount Avg CF:			1024.90622	Linear:			
Aroclor-1232					6	161.53	
+ 3.07	3.10	3.13	87.882454	309.450709			1
3.07	3.12	3.13	371.917969	1309.59337			1
3.40	3.43	3.46	443.461212	2799.628864			2
3.56	3.59	3.62	1322.143921	4739.631557			3
3.60	3.62	3.66	447.907471	2962.057144			4
3.67	3.70	3.73	633.745789	4586.378557			5
3.84	3.87	3.90	5059.392578	41003.262647			6
+ 3.84	3.89	3.90	896.521851	7265.757768			6
Height Summation:			8278.568940				
Amount Avg CF:			9566.75869	Linear:			
Aroclor-1242					6	152.09	
+ 3.07	3.10	3.13	87.882454	434.244757			1
3.07	3.12	3.13	371.917969	1837.720966			1
3.40	3.43	3.46	443.461212	1645.740414			2
3.56	3.59	3.62	1322.143921	2769.264761			3
3.59	3.62	3.65	447.907471	1728.738382			4
3.67	3.70	3.73	633.745789	2587.562424			5
3.84	3.87	3.90	5059.392578	22800.840839			6
+ 3.84	3.89	3.90	896.521851	4040.297668			6
Height Summation:			8278.568940				
Amount Avg CF:			5561.644831	Linear:			
Aroclor-1248					6	149.94	
3.28	3.31	3.34	369.136047	1994.219688			1
3.56	3.59	3.62	1322.143921	4152.537706			2
3.71	3.73	3.77	6774.937988	17355.481918			3
+ 3.71	3.76	3.77	3274.94458	8389.48512			3
3.84	3.87	3.90	5059.392578	12069.622357			4
+ 3.84	3.89	3.90	896.521851	2138.731085			4
4.21	4.24	4.27	25141.337891	102599.949605			5
+ 4.45	4.45	4.51	14539.563477	70180.075931			6
4.45	4.48	4.51	49542.332031	239132.669232			6
Height Summation:			88209.280456				
Amount Avg CF:			62884.080084	Linear:			

ASX22 0882

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/9/09 05:41:40

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822071** MSACF DF1000 STEA3MS Sample ID: AD **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10000 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on : NOV 07, 2009 11:48:35
Instrument : CP06–S9147A
Result file : 2C6310.10R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					6	65.17	
4.26	4.29	4.32	1715.544556	23365.53834			1
+ 4.26	4.32	4.32	529.152893	7207.007341			1
4.68	4.71	4.74	5787.859863	53735.185915 /			2
+ 4.68	4.73	4.74	5084.635742	47206.368739			2
4.85	4.87	4.91	10544.015625	81990.152682 /			3
5.13	5.15	5.19	11357.291016	101721.901722 /			4
5.35	5.37	5.41	10531.96875	192510.295456			5
5.66	5.69	5.72	14092.176758	186832.984093			6
Height Summation:			54028.856568				
Amount Avg CF:			106692.676368	Linear:			
Aroclor-1260					6	40.20	
5.25	5.28	5.31	6697.268555	71746.558028			1
5.45	5.48	5.51	9397.589844	110320.549819			2
5.66	5.69	5.72	14092.176758	167889.926469			3
5.89	5.92	5.95	5485.69043	80773.359926 /			4
6.12	6.15	6.18	9105.746094	69550.074196 /			5
6.33	6.36	6.39	4716.222656	73603.115281 /			6
Height Summation:			49494.694337				
Amount Avg CF:			95647.263953	Linear:			
Aroclor-1262					6	47.94	
5.25	5.28	5.31	6697.268555	78840.739792			1
5.45	5.48	5.51	9397.589844	121160.419426			2
5.73	5.75	5.79	2981.392822	27985.627195			3
5.89	5.92	5.95	5485.69043	66745.149936			4
6.12	6.15	6.18	9105.746094	54051.45594			5
6.32	6.36	6.38	4716.222656	49569.416567			6
Height Summation:			38383.910401				
Amount Avg CF:			66392.134809	Linear:			
Aroclor-1268					6	123.86	
5.90	5.92	5.96	5485.69043	92156.207876			1
6.32	6.36	6.38	4716.222656	25970.617891			2
6.38	6.40	6.44	1455.674194	10026.713078			3
6.51	6.54	6.57	159.658493	1025.435702			4
6.74	6.77	6.80	2001.164673	37047.079298			5
6.96	6.99	7.02	435.305084	1186.918161			6
Height Summation:			14253.715530				
Amount Avg CF:			27902.162001	Linear:			

Analysis Report (B)

Injected on : NOV 07, 2009 11:48:35
Instrument : CP06–S9147B
Result file : 2C6310B.10R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					7	53.06	
+ 4.07	4.08	4.13	7199.945313	11701.033339			1
4.07	4.10	4.13	34566.769531	56176.388169/			1
4.07	4.13	4.13	12468.730469	20263.630426			1
4.18	4.21	4.24	34796.472656	68901.996095/			2
4.45	4.48	4.51	49542.332031	83869.013451/			3
4.62	4.64	4.68	51046.425781	151906.254277			4
4.79	4.82	4.85	34247.808594	139000.063291			5
4.92	4.95	4.98	60778.6875	138447.783011			6
+ 4.92	4.98	4.98	16862.208984	38410.428828			6
Height Summation:			277447.226562				
Amount Avg CF:			94080.732674	Linear:			
Aroclor-1260					6	39.12	
4.55	4.57	4.61	36508.167969	74249.501356			1
4.68	4.71	4.74	46198.261719	87945.529818			2
4.79	4.82	4.85	34247.808594	71293.546686 /			3
4.92	4.95	4.98	60778.6875	137045.426873			4
+ 5.34	5.34	5.40	11250.198242	13709.358248			5
5.34	5.37	5.40	44200.605469	53862.333987 /			5
5.60	5.62	5.66	28666.335938	52901.709344 /			6
Height Summation:			250599.867189				
Amount Avg CF:			79549.674677	Linear:			
Aroclor-1262					6	64.10	
4.68	4.71	4.74	46198.261719	98860.116486			1
4.95	4.98	5.01	16862.208984	25818.558484			2
5.15	5.18	5.21	17606.621094	28631.333754			3
+ 5.15	5.20	5.21	17378.492188	28260.357699			3
+ 5.34	5.34	5.40	11250.198242	10642.488722			4
5.34	5.37	5.40	44200.605469	41812.991657			4
5.60	5.62	5.66	28666.335938	40039.108331			5
5.97	6.00	6.03	10113.441406	26422.754519			6
Height Summation:			163647.474610				
Amount Avg CF:			43597.477205	Linear:			
Aroclor-1268					6	104.18	
5.15	5.18	5.21	17606.621094	49434.333199			1
+ 5.15	5.20	5.21	17378.492188	48793.812778			1
+ 5.55	5.56	5.61	210.286545	173.47146			2
5.55	5.58	5.61	8169.849121	6739.545115			2
5.60	5.62	5.66	28666.335938	28077.428717			3
5.75	5.78	5.81	450.556427	428.264652			4
5.97	6.00	6.03	10113.441406	26153.927769			5
6.17	6.20	6.23	2409.951172	900.018461			6
Height Summation:			67416.755158				
Amount Avg CF:			18622.252986	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			17000	3300		28.55	4	40	
Aroclor-1221			17000	3300		13.55	3	5	
Aroclor-1232			17000	3300		27.09	4	10	
Aroclor-1242			17000	3300		22.49	4	30	
Aroclor-1248			17000	3300		** 77.44	4	40	
Aroclor-1254			17000	3300		12.56	4	40	
Aroclor-1260			17000	3300		18.38	4	40	~~ASX22 8883~~

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/9/09 05:41:40

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822071** MSACF DF1000 STEA3MS Sample ID: AD **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10000 ml Analyst: 2280 SDG: ASX22 State: DE
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on : NOV 07, 2009 11:48:35
Instrument : CP06--S9147A
Result file : 2C6310.10R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Analysis Report (B)

Injected on : NOV 07, 2009 11:48:35
Instrument : CP06--S9147B
Result file : 2C6310B.10R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1262			17000	3300		** 41.45	4	40	
Aroclor-1268			17000	3300		39.89	4	40	

Units: _____

ASX22 0884

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/9/09 05:41:40

Sample Name: **5822071MSACF DF1000 ADSTEA3MS MS 093070028A 01216**
Acquired from CP06--S9147A via port 1 on 11/7/09 11:56:05am by 2280
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\2C6310.10R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.026		0.0000	0.000	305.0	0.144	BB	0.017	297.18	0.188
2	3.096		0.0000	0.000	145.0	0.068	BB	0.036	68.00	0.043
3	3.309 TCX		593.7841	100.000	181.9	0.086	BB	0.026	116.15	0.074
4	3.450		0.0000	0.000	57.9	0.027	BB	0.022	44.28	0.028
5	3.563		0.0000	0.000	187.3	0.088	BB	0.015	212.34	0.134
6	3.617		0.0000	0.000	137.3	0.065	BB	0.016	141.62	0.090
7	3.787		0.0000	0.000	197.8	0.093	BB	0.022	152.22	0.096
8	3.913		0.0000	0.000	166.1	0.078	BB	0.017	167.28	0.106
9	4.046		0.0000	0.000	49.1	0.023	BB	0.026	32.00	0.020
10	4.075		0.0000	0.000	84.0	0.040	BB	0.015	95.03	0.060
11	4.111		0.0000	0.000	584.0	0.275	BB	0.027	361.44	0.229
12	4.155		0.0000	0.000	127.7	0.060	BB	0.017	126.24	0.080
13	4.219		0.0000	0.000	174.1	0.082	BB	0.019	155.36	0.098
14	4.290		0.0000	0.000	1614.0	0.760	BB	0.016	1715.54	1.086
15	4.321		0.0000	0.000	429.4	0.202	BB	0.014	529.15	0.335
16	4.431		0.0000	0.000	1839.1	0.866	BB	0.023	1348.89	0.854
17	4.493		0.0000	0.000	223.0	0.105	BB	0.015	252.61	0.160
18	4.532		0.0000	0.000	863.4	0.407	BB	0.019	774.73	0.490
19	4.578		0.0000	0.000	300.5	0.142	BB	0.022	225.82	0.143
20	4.667		0.0000	0.000	199.5	0.094	BB	0.017	193.71	0.123
21	4.707		0.0000	0.000	6543.2	3.081	BV	0.019	5787.86	3.663
22	4.727		0.0000	0.000	6081.1	2.864	VV	0.020	5084.64	3.218
23	4.755		0.0000	0.000	6236.5	2.937	VV	0.028	3720.34	2.355
24	4.827		0.0000	0.000	2397.4	1.129	VV	0.023	1728.32	1.094
25	4.875		0.0000	0.000	14908.3	7.021	VV	0.024	10544.02	6.674
26	4.913		0.0000	0.000	10133.3	4.772	VV	0.029	5728.18	3.626
27	4.983		0.0000	0.000	1011.6	0.476	VV	0.022	766.51	0.485
28	5.031		0.0000	0.000	4018.0	1.892	VB	0.020	3369.06	2.132
29	5.087		0.0000	0.000	8408.9	3.960	BV	0.024	5887.44	3.726
30	5.121		0.0000	0.000	4432.0	2.087	VV	0.021	3528.33	2.233
31	5.153		0.0000	0.000	14663.9	6.905	VV	0.022	11357.29	7.189
32	5.201		0.0000	0.000	2984.5	1.405	VB	0.023	2169.48	1.373
33	5.243		0.0000	0.000	1673.6	0.788	BB	0.015	1855.72	1.175
34	5.283		0.0000	0.000	8374.7	3.944	BB	0.021	6697.27	4.239
35	5.332		0.0000	0.000	767.1	0.361	BB	0.015	870.91	0.551
36	5.375		0.0000	0.000	13391.0	6.306	BB	0.021	10531.97	6.666
37	5.430		0.0000	0.000	848.9	0.400	BB	0.015	966.97	0.612
38	5.483		0.0000	0.000	15222.9	7.169	BB	0.027	9397.59	5.948
39	5.536		0.0000	0.000	490.7	0.231	BB	0.012	692.09	0.438
40	5.578		0.0000	0.000	12000.6	5.651	BB	0.020	9926.19	6.283
41	5.619		0.0000	0.000	232.1	0.109	BB	0.008	461.46	0.292
42	5.651		0.0000	0.000	773.8	0.364	BB	0.020	641.85	0.406
43	5.692		0.0000	0.000	19852.7	9.349	BB	0.023	14092.18	8.920
44	5.753		0.0000	0.000	2784.3	1.311	BB	0.016	2981.39	1.887
45	5.797		0.0000	0.000	2369.7	1.116	BB	0.018	2150.35	1.361
46	5.917		0.0000	0.000	11384.4	5.361	BB	0.035	5485.69	3.472
47	5.976		0.0000	0.000	1711.8	0.806	BB	0.018	1599.13	1.012
48	6.015		0.0000	0.000	1008.3	0.475	BB	0.016	1071.68	0.678
49	6.057		0.0000	0.000	103.5	0.049	BB	0.012	141.08	0.089
50	6.096		0.0000	0.000	3230.5	1.521	BB	0.021	2567.00	1.625
51	6.151		0.0000	0.000	12551.2	5.911	BB	0.023	9105.75	5.764
52	6.356		0.0000	0.000	5771.9	2.718	BB	0.020	4716.22	2.985
53	6.403		0.0000	0.000	2534.6	1.194	BB	0.029	1455.67	0.921
54	6.535		0.0000	0.000	238.4	0.112	BB	0.025	159.66	0.101
55	6.625		0.0000	0.000	1806.6	0.851	BB	0.028	1072.97	0.679
56	6.774		0.0000	0.000	2624.6	1.236	BB	0.022	2001.16	1.267
57	6.906		0.0000	0.000	88.0	0.041	BB	0.043	33.89	0.021
58	6.989		0.0000	0.000	562.3	0.265	BB	0.022	435.31	0.276
59	7.029		0.0000	0.000	174.4	0.082	BB	0.019	151.45	0.096
60	7.425		0.0000	0.000	96.5	0.045	BB	0.036	44.83	0.028

ASX22 8095

Total Area = 212353.7, Total Amount = 593.784, Total Height = 157988.5, Sample Units = PPB

Lancaster Laboratories

Sample Name: **5822071MSACF DF1000 ADSTEA3MS MS 093070028A 01216**
Acquired from CP06--S9147B via port 2 on 11/7/09 11:56:05am by 2280
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\2C6310B.10R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.552		0.0000	0.000	622.1	0.078	BB	0.014	734.86	0.098
2	2.617		0.0000	0.000	463.6	0.058	BB	0.071	108.42	0.014
3	2.807 TCX		732.3223	11.325	329.4	0.041	BB	0.016	335.07	0.045
4	2.897		0.0000	0.000	148.6	0.019	BB	0.027	93.46	0.012
5	3.097		0.0000	0.000	63.0	0.008	BB	0.012	87.88	0.012
6	3.120		0.0000	0.000	228.0	0.028	BB	0.010	371.92	0.050
7	3.140		0.0000	0.000	304.3	0.038	BB	0.011	477.16	0.064
8	3.308		0.0000	0.000	452.8	0.056	BV	0.020	369.14	0.049
9	3.359		0.0000	0.000	54.7	0.007	VB	0.014	67.27	0.009
10	3.431		0.0000	0.000	354.3	0.044	BB	0.013	443.46	0.059
11	3.513		0.0000	0.000	75.0	0.009	BV	0.009	134.70	0.018
12	3.530		0.0000	0.000	863.5	0.108	VV	0.014	1063.92	0.142
13	3.587		0.0000	0.000	1597.2	0.199	VV	0.020	1322.14	0.176
14	3.623		0.0000	0.000	773.9	0.096	VB	0.029	447.91	0.060
15	3.700		0.0000	0.000	497.1	0.062	BB	0.013	633.75	0.084
16	3.735		0.0000	0.000	5466.5	0.681	BV	0.013	6774.94	0.902
17	3.759		0.0000	0.000	3561.2	0.444	VB	0.018	3274.94	0.436
18	3.869		0.0000	0.000	4090.2	0.510	BV	0.013	5059.39	0.674
19	3.893		0.0000	0.000	608.0	0.076	VB	0.011	896.52	0.119
20	3.917		0.0000	0.000	642.0	0.080	BB	0.012	911.33	0.121
21	3.964		0.0000	0.000	2827.1	0.352	BV	0.015	3114.06	0.415
22	4.009		0.0000	0.000	853.2	0.106	VB	0.014	989.13	0.132
23	4.053		0.0000	0.000	733.8	0.091	BV	0.017	728.98	0.097
24	4.078		0.0000	0.000	5881.4	0.733	VV	0.014	7199.95	0.959
25	4.102		0.0000	0.000	30816.3	3.840	VV	0.015	34566.77	4.603
26	4.128		0.0000	0.000	11341.4	1.413	VV	0.015	12468.73	1.661
27	4.149		0.0000	0.000	4191.5	0.522	VV	0.014	4979.56	0.663
28	4.168		0.0000	0.000	6639.2	0.827	VV	0.016	6835.81	0.910
29	4.205		0.0000	0.000	31989.1	3.986	VV	0.015	34796.47	4.634
30	4.239		0.0000	0.000	39316.8	4.899	VV	0.026	25141.34	3.348
31	4.321		0.0000	0.000	2981.6	0.372	VV	0.017	2946.62	0.392
32	4.357		0.0000	0.000	13985.2	1.743	VB	0.016	14714.02	1.960
33	4.416		0.0000	0.000	25308.9	3.154	BV	0.016	27208.35	3.623
34	4.450		0.0000	0.000	14065.7	1.753	VV	0.016	14539.56	1.936
35	4.479		0.0000	0.000	52013.7	6.481	VB	0.017	49542.33	6.598
36	4.520		0.0000	0.000	5033.7	0.627	BB	0.016	5281.62	0.703
37	4.574		0.0000	0.000	37968.6	4.731	BV	0.017	36508.17	4.862
38	4.610		0.0000	0.000	6563.9	0.818	VV	0.017	6613.82	0.881
39	4.643		0.0000	0.000	64219.4	8.002	VV	0.021	51046.43	6.798
40	4.712		0.0000	0.000	53381.3	6.652	VB	0.019	46198.26	6.152
41	4.819		0.0000	0.000	36173.2	4.507	BB	0.018	34247.81	4.561
42	4.870		0.0000	0.000	41780.1	5.206	BV	0.018	39011.74	5.195
43	4.912		0.0000	0.000	10102.0	1.259	VB	0.017	9940.15	1.324
44	4.951		0.0000	0.000	73088.6	9.107	BV	0.020	60778.69	8.094
45	4.983		0.0000	0.000	17951.4	2.237	VB	0.018	16862.21	2.246
46	5.023		0.0000	0.000	7383.0	0.920	BB	0.013	9118.77	1.214
47	5.086		0.0000	0.000	632.5	0.079	BB	0.017	618.89	0.082
48	5.127		0.0000	0.000	5903.0	0.736	BV	0.027	3634.73	0.484
49	5.177		0.0000	0.000	17813.0	2.220	VV	0.017	17606.62	2.345
50	5.201		0.0000	0.000	17002.8	2.119	VV	0.016	17378.49	2.314
51	5.251		0.0000	0.000	9293.3	1.158	VV	0.017	9239.51	1.230
52	5.286		0.0000	0.000	5056.8	0.630	VV	0.016	5186.18	0.691
53	5.312		0.0000	0.000	3274.4	0.408	VV	0.016	3376.64	0.450
54	5.344		0.0000	0.000	9800.4	1.221	VV	0.015	11250.20	1.498
55	5.366		0.0000	0.000	50009.6	6.231	VB	0.019	44200.61	5.886
56	5.442		0.0000	0.000	1221.8	0.152	BB	0.014	1464.54	0.195
57	5.556		0.0000	0.000	180.2	0.022	BB	0.014	210.29	0.028
58	5.580		0.0000	0.000	7629.9	0.951	BB	0.016	8169.85	1.088
59	5.624		0.0000	0.000	33159.6	4.132	BB	0.019	28666.34	3.818
60	5.700		0.0000	0.000	156.5	0.020	BB	0.013	206.56	0.028
61	5.780		0.0000	0.000	434.3	0.054	BB	0.016	450.56	0.060
62	5.819		0.0000	0.000	735.2	0.092	BB	0.015	811.31	0.108

ASX22 18886

Printed on 11/7/2009 11:56:33 AM

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	5.847		0.0000	0.000	215.5	0.027	BB	0.013	284.45	0.038
64	5.896		0.0000	0.000	4658.1	0.580	BB	0.017	4586.67	0.611
65	5.998		0.0000	0.000	11864.9	1.478	BB	0.020	10113.44	1.347
66	6.198		0.0000	0.000	2789.7	0.348	BB	0.019	2409.95	0.321
67	6.355	DCB	5734.0970	88.675	1714.2	0.214	BB	0.021	1368.56	0.182
68	6.772		0.0000	0.000	805.2	0.100	BB	0.032	425.14	0.057
69	6.894		0.0000	0.000	47.7	0.006	BB	0.017	46.87	0.006
70	6.972		0.0000	0.000	61.0	0.008	BB	0.032	31.39	0.004
71	7.041		0.0000	0.000	156.3	0.019	BB	0.022	120.35	0.016
72	7.259		0.0000	0.000	134.6	0.017	BB	0.048	46.45	0.006

Total Area = 802535.6, Total Amount = 6466.419, Total Height = 750892.1, Sample Units = PPB

ASX22 8887

Lancaster Laboratories

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\2C6310.10R

5822071MSACF DF1000 ADSTEA3MS MS 093070028A 01216

C:\CPWIN\DATA1\2C6310.10R



Instrument ID: CP06--S9147A Injected On: 11/7/2009 11:48:35 AM Column ID: RTX CLP

File: C:\CPWIN\DATA1\2C6310B.10R

5822071MSACF DF1000 ADSTEA3MS MS 093070028A 01216

C:\CPWIN\DATA1\2C6310B.10R



ASX22 0889

Instrument ID: CP06--S9147B Injected On: 11/7/2009 11:48:35 AM Column ID: RTX CLPII

Printed on 11/7/2009 11:56:36 AM

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 30 Dilution Factor: 10000
Analyst: 2280

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.309	116	19.7928	TCX	2.807	335	24.41074	TCX
.	0	.	DCB	6.355	1369	191.1366	DCB

Files:
Area File: C:\CPWIN\DATA1\2C6310.10A
Area File: C:\CPWIN\DATA1\2C6310B.10A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/7/2009 11:56:18 AM
File Reported On: 11/7/2009 at 11:56:35 AM

REPLOTS

ASX22 0889

5822071MSACF DF1000 ADSTEA3MS MS 093070028A 01216
C:\CPWIN\DATA1\2C6310.10R



5822071MSACF DF1000 ADSTEA3MS MS 093070028A 01216
C:\CPWIN\DATA1\2C6310B.10R



ASX22 0890

Multiple Component Data Summary

Sample Name: 5822072 MSD RI ACF DF STEA3MSD MSD Sample ID:AD **Batchnumber: 093070028A**
Sample Amount:30 g Total Volume: 10000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216

Analysis Report (A)

Injected on Nov 09, 2009 19:48:45
Instrument S9147A
Result file 3C6310.47R
Calibration file 1C6310
Method file PESTD

%SSR(TCX) * 239.0% Conc: 24.46653
%SSR(DCB) * 457.6% Conc: 46.51936

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**29.49**	
4.68	4.71	4.74	5460.931152	50699.940507			2
4.85	4.88	4.91	10243.083008	79650.103872			3
5.13	5.16	5.19	10492.989258	93980.758313			4
Height summation:			**26197.003418**				
Concentration		CF:	**74776.934231**	L:			
Aroclor-1260						**9.75**	
5.89	5.92	5.95	4940.236816	72741.896676			4
6.12	6.15	6.18	7834.095703	59837.154669			5
6.33	6.36	6.39	4231.111328	66032.288456			6
Height summation:			**17005.443847**				
Concentration		CF:	**66203.779934**	L:			

Analysis Report (B)

Injected on Nov 09, 2009 19:48:45
Instrument S9147B
Result file 3C6310B.47R
Calibration file 1C6310B
Method file PESTDB

%SSR(TCX) * 187.9% Conc: 19.23131
%SSR(DCB) * 1046.1% Conc: 106.3393

Min	RT	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254						**19.65**	
4.07	4.10	4.13	31179.992188	50672.347113			1
4.18	4.21	4.24	31505.873047	62386.137903			2
4.45	4.48	4.51	44501.667969	75335.795399			3
Height summation:			**107187.533204**				
Concentration		CF:	**62798.093472**	L:			
Aroclor-1260						**17.19**	
4.79	4.82	4.85	30634.625	63771.994686			3
5.34	5.37	5.40	39237.855469	47814.785653			5
5.60	5.63	5.66	26038.226563	48051.718101			6
Height summation:			**95910.707032**				
Concentration		CF:	**53212.832813**	L:			

Summary Report

Compound Name	Column	Amount Found	LOQ	MDL	Qualifiers	%Difference	No. of Hits Required	Max %RSD	Comments
Aroclor-1016			<17000	<3300			4	40	
Aroclor-1221			<17000	<4000			3	5	
Aroclor-1232			<17000	<5300			4	10	
Aroclor-1242			<17000	<3300			4	30	
Aroclor-1248			<17000	<3300			4	40	
Aroclor-1254	A	74776.934231	17000	3300		17.41	4	40	
Aroclor-1260	A	66203.779934	17000	3300		21.76	4	40	

Units: ug/kg

%Difference = High - Low divided by the Average times 100

Reviewed By: Andrea Carey Date: 11-10-09
Verified By: [illegible] Date: 11/11/09

ASX22 0891

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: 5822072 MSD RI ACF DF STEA3MSD Sample ID: AD **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on : NOV 09, 2009 19:48:45
Instrument : CP06--S9147A
Result file : 3C6310.47R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

%SSR(TCX) : *239% Conc.: 24.466536
%SSR(DCB) : *457.6% Conc.: 46.51936

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	143.17	
3.54	3.56	3.60	253.425934	2622.577252			1
3.88	3.92	3.94	154.1604	1348.664112			2
4.08	4.12	4.14	384.667694	2504.612396			3
4.13	4.16	4.19	156.220245	1815.475012			4
4.19	4.22	4.25	168.149429	2185.689037			5
4.40	4.43	4.46	1126.684204	19690.735665			6
Height Summation:			2243.307906				
Amount Avg CF:			5027.958912	Linear:			
Aroclor-1221					1		
3.53	3.56	3.59	253.425934	1879.735455			3
Height Summation:			253.425934				
Amount Avg CF:			1879.735455	Linear:			
Aroclor-1232					6	156.55	
3.54	3.56	3.60	253.425934	2413.235576			1
3.88	3.92	3.94	154.1604	3030.775582			2
4.08	4.12	4.14	384.667694	5787.522666			3
4.13	4.16	4.19	156.220245	4142.674224			4
4.19	4.22	4.25	168.149429	5086.189625			5
4.40	4.43	4.46	1126.684204	47270.157499			6
Height Summation:			2243.307906				
Amount Avg CF:			11288.425862	Linear:			
Aroclor-1242					6	140.99	
3.54	3.56	3.60	253.425934	3303.042476			1
3.88	3.92	3.94	154.1604	1739.272296			2
4.08	4.12	4.14	384.667694	3275.998075			3
4.13	4.16	4.19	156.220245	2414.718989			4
4.19	4.22	4.25	168.149429	2825.803361			5
4.40	4.43	4.46	1126.684204	24667.415523			6
Height Summation:			2243.307906				
Amount Avg CF:			6371.041787	Linear:			
Aroclor-1248					6	124.19	
3.88	3.92	3.94	154.1604	3546.167223			1
4.19	4.22	4.25	168.149429	3554.43019			2
4.26	4.29	4.32	1566.866211	18418.335402			3
4.40	4.43	4.46	1126.684204	13399.316455			4
4.50	4.53	4.56	770.093872	12122.918843			5
4.70	4.71	4.76	5460.931152	69772.018709			6
+ 4.70	4.73	4.76	4813.338867	61498.004669			6
Height Summation:			9246.885268				
Amount Avg CF:			20135.531137	Linear:			
Aroclor-1254					6	62.91	
4.26	4.29	4.32	1566.866211	21340.554752			1
4.68	4.71	4.74	5460.931152	50699.940507 ✓			2
+ 4.68	4.73	4.74	4813.338867	44687.615977			2
4.85	4.88	4.91	10243.083008	79650.103872 ✓			3
5.13	5.16	5.19	10492.989258	93980.758313 ✓			4
5.35	5.38	5.41	9444.018555	172624.021726			5
5.66	5.69	5.72	12628.880859	167432.720804			6
Height Summation:			49836.769043				
Amount Avg CF:			97621.349996	Linear:			

Analysis Report (B)

Injected on : NOV 09, 2009 19:48:45
Instrument : CP06--S9147B
Result file : 3C6310B.47R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

%SSR(TCX) : *187.9% Conc.: 19.231312
%SSR(DCB) : *1046.1% Conc.: 106.33934

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1016					6	142.99	
+ 3.07	3.10	3.13	92.267212	355.440957			1
3.07	3.12	3.13	299.409943	1153.416847			1
3.40	3.43	3.46	422.926666	1200.550095			2
3.56	3.59	3.62	1293.497437	2062.597588			3
3.59	3.63	3.65	482.552948	1415.590343			4
3.67	3.70	3.73	615.709106	1889.799366			5
3.84	3.87	3.90	4368.020508	14463.854796			6
+ 3.84	3.90	3.90	668.207397	2212.639512			6
Height Summation:			7482.116608				
Amount Avg CF:			3697.63484	Linear:			
Aroclor-1221					1		
+ 3.07	3.10	3.13	92.267212	254.263701			3
3.07	3.12	3.13	299.409943	825.093538			3
Height Summation:			299.409943				
Amount Avg CF:			825.093538	Linear:			
Aroclor-1232					6	154.17	
+ 3.07	3.10	3.13	92.267212	324.890269			1
3.07	3.12	3.13	299.409943	1054.278924			1
3.40	3.43	3.46	422.926666	2669.991578			2
3.56	3.59	3.62	1293.497437	4636.939424			3
3.60	3.63	3.66	482.552948	3191.171167			4
3.67	3.70	3.73	615.709106	4455.848212			5
3.84	3.87	3.90	4368.020508	35400.117578			6
+ 3.84	3.90	3.90	668.207397	5415.409652			6
Height Summation:			7482.116608				
Amount Avg CF:			8568.057814	Linear:			
Aroclor-1242					6	145.40	
+ 3.07	3.10	3.13	92.267212	455.910722			1
3.07	3.12	3.13	299.409943	1479.444328			1
3.40	3.43	3.46	422.926666	1569.534128			2
3.56	3.59	3.62	1293.497437	2709.263956			3
3.59	3.63	3.65	482.552948	1862.455655			4
3.67	3.70	3.73	615.709106	2513.919263			5
3.84	3.87	3.90	4368.020508	19685.078564			6
+ 3.84	3.90	3.90	668.207397	3011.367525			6
Height Summation:			7482.116608				
Amount Avg CF:			4969.949316	Linear:			
Aroclor-1248					6	150.34	
3.28	3.31	3.34	322.81546	1743.977461			1
3.56	3.59	3.62	1293.497437	4062.565954			2
3.71	3.74	3.77	6012.682129	15402.797214			3
+ 3.71	3.76	3.77	2902.472168	7435.315765			3
3.84	3.87	3.90	4368.020508	10420.293971			4
+ 3.84	3.90	3.90	668.207397	1594.067037			4
4.21	4.24	4.27	22415.181641	91474.706585			5
+ 4.45	4.45	4.51	12841.144531	61982.087678			6
4.45	4.48	4.51	44501.667969	214802.214802			6
Height Summation:			78913.865144				
Amount Avg CF:			56317.759331	Linear:			

ASX22 0092

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:48:46

Lancaster Laboratories Multiple Component Peak Data Report

Sample Name: **5822072** MSD RI ACF DF STEA3MSD Sample ID: AD **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on : NOV 09, 2009 19:48:45
Instrument : CP06--S9147A
Result file : 3C6310.47R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1260					6	40.40	
5.25	5.28	5.31	6169.211426	66089.582929			1
5.45	5.49	5.51	8671.551758	101797.415466			2
5.66	5.69	5.72	12628.880859	150456.662247			3
5.89	5.92	5.95	4940.236816	72741.896676 /			4
6.12	6.15	6.18	7834.095703	59837.154669 /			5
6.33	6.36	6.39	4231.111328	66032.288456 /			6
Height Summation:			44475.087890				
Amount Avg CF:			86159.166741	Linear:			
Aroclor-1262					6	49.81	
5.25	5.28	5.31	6169.211426	72624.412291			1
5.45	5.49	5.51	8671.551758	111799.819476			2
5.73	5.76	5.79	2677.591553	25133.916749			3
5.89	5.92	5.95	4940.236816	60108.540795			4
6.12	6.15	6.18	7834.095703	46502.974533			5
6.32	6.36	6.38	4231.111328	44470.699383			6
Height Summation:			34523.798584				
Amount Avg CF:			60106.727205	Linear:			
Aroclor-1268					5	106.72	
5.90	5.92	5.96	4940.236816	82992.924369			1
6.32	6.36	6.38	4231.111328	23299.276469			2
6.38	6.40	6.44	1346.681274	9275.967657			3
6.74	6.78	6.80	1837.599609	34019.038688			5
6.96	6.99	7.02	391.118896	1066.438545			6
Height Summation:			12746.747923				
Amount Avg CF:			30130.729146	Linear:			

Analysis Report (B)

Injected on : NOV 09, 2009 19:48:45
Instrument : CP06--S9147B
Result file : 3C6310B.47R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Min	R.T.	Max	Height	Amount	Pks	%RSD	Peak
Aroclor-1254					7	52.64	
+ 4.07	4.08	4.13	6694.534668	10879.662266			1
4.07	4.10	4.13	31179.992188	50672.347113 /			1
4.07	4.13	4.13	11257.30957	18294.882649			1
4.18	4.21	4.24	31505.873047	62386.137903 /			2
+ 4.45	4.45	4.51	12841.144531	21738.462427			3
4.45	4.48	4.51	44501.667969	75335.795399 /			3
4.62	4.65	4.68	45129.121094	134297.272323			4
4.79	4.82	4.85	30634.625	124335.39513			5
4.92	4.95	4.98	55033.714844	125361.308781			6
Height Summation:			249242.303712				
Amount Avg CF:			84383.305614	Linear:			
Aroclor-1260					6	39.47	
4.55	4.58	4.61	33223.292969	67568.795522			1
4.68	4.71	4.74	41605.902344	79203.264128			2
4.79	4.82	4.85	30634.625	63771.994686 /			3
4.92	4.95	4.98	55033.714844	124091.507293			4
+ 5.34	5.35	5.40	10138.666992	12354.85944			5
5.34	5.37	5.40	39237.855469	47814.785653 /			5
5.60	5.63	5.66	26038.226563	48051.718101 /			6
Height Summation:			225773.617189				
Amount Avg CF:			71750.344231	Linear:			
Aroclor-1262					6	63.03	
4.68	4.71	4.74	41605.902344	89032.881307			1
4.95	4.99	5.01	15854.745117	24275.980949			2
5.15	5.18	5.21	16345.600586	26580.701845			3
+ 5.15	5.20	5.21	15697.878906	25527.39721			3
+ 5.34	5.35	5.40	10138.666992	9590.999803			4
5.34	5.37	5.40	39237.855469	37118.317859			4
5.60	5.63	5.66	26038.226563	36368.351238			5
5.97	6.00	6.03	9175.452148	23972.128772			6
Height Summation:			148257.782227				
Amount Avg CF:			39558.060328	Linear:			
Aroclor-1268					6	104.97	
5.15	5.18	5.21	16345.600586	45893.749936			1
+ 5.15	5.20	5.21	15697.878906	44075.133566			1
+ 5.55	5.56	5.61	143.383026	118.28081			2
5.55	5.58	5.61	7634.914551	6298.262098			2
5.60	5.63	5.66	26038.226563	25503.309939			3
5.75	5.78	5.81	353.913208	336.402963			4
5.97	6.00	6.03	9175.452148	23728.234841			5
6.17	6.20	6.23	2032.657715	759.114744			6
Height Summation:			61580.764771				
Amount Avg CF:			17086.51242	Linear:			

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1016			17000	3300		30.49	4	40	
Aroclor-1221			17000	3300		** 77.98	3	5	
Aroclor-1232			17000	3300		27.40	4	10	
Aroclor-1242			17000	3300		24.71	4	30	
Aroclor-1248			17000	3300		** 94.65	4	40	
Aroclor-1254			17000	3300		14.55	4	40	
Aroclor-1260			17000	3300		18.25	4	40	~~ASX22 8893~~

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:48:46

Lancaster Laboratories-Multiple Component Peak Data Report

Sample Name: **5822072** MSD RI ACF DF STEA3MSD Sample ID: AD **Batchnumber: 093070028A**
Sample Amount: 30 g Total Volume: 10000 ml Analyst: 1151 SDG: ASX22 State: DE
Analyses: 01216 05108 08899 10225

Analysis Report (A)

Injected on : NOV 09, 2009 19:48:45
Instrument : CP06--S9147A
Result file : 3C6310.47R
Calibration file : 1C6310.CAL
Method file : PESTD.MET

Analysis Report (B)

Injected on : NOV 09, 2009 19:48:45
Instrument : CP06--S9147B
Result file : 3C6310B.47R
Calibration file : 1C6310B.CAL
Method file : PESTDB.MET

Summary Report

Compound Name	Column	Higher Amount Found	LOQ	MDL	Qualifiers	%Difference	No of Hits Required	Max %RSD	Comments
Aroclor-1262			17000	3300		** 41.24	4	40	
Aroclor-1268			17000	3300		** 55.25	4	40	

Units: _____

ASX22 8894

*Peak found within more than one window
+Duplicate Peak in window - not included in average

Printed on: 11/10/09 05:48:46

Sample Name: **5822072MSD RI ACF DFADSTEA3MSD MSD 093070028A 01216**
Acquired from CP06--S9147A via port 1 on 11/9/09 07:56:15pm by 1151
RTX CLP
110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min hold 1
Data File: **C:\CPWIN\DATA1\3C6310.47R**
Method File: C:\CPWIN\DATA1\PESTD.MET
Calibration File: C:\CPWIN\DATA1\1C6310.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	3.029		0.0000	0.000	246.3	0.124	BB	0.021	198.14	0.137
2	3.311	TCX	733.9961	34.467	294.9	0.148	BB	0.034	143.58	0.099
3	3.565		0.0000	0.000	227.5	0.114	BB	0.015	253.43	0.175
4	3.623		0.0000	0.000	134.2	0.067	BB	0.021	108.02	0.074
5	3.788		0.0000	0.000	177.4	0.089	BB	0.027	108.91	0.075
6	3.916		0.0000	0.000	139.8	0.070	BB	0.015	154.16	0.106
7	4.115		0.0000	0.000	647.7	0.325	BB	0.028	384.67	0.265
8	4.157		0.0000	0.000	158.3	0.080	BB	0.017	156.22	0.108
9	4.224		0.0000	0.000	234.9	0.118	BB	0.023	168.15	0.116
10	4.293		0.0000	0.000	1500.3	0.754	BB	0.016	1566.87	1.080
11	4.324		0.0000	0.000	403.6	0.203	BB	0.014	496.14	0.342
12	4.349		0.0000	0.000	100.0	0.050	BB	0.009	185.79	0.128
13	4.433		0.0000	0.000	1086.4	0.546	BB	0.016	1126.68	0.777
14	4.496		0.0000	0.000	215.3	0.108	BB	0.015	240.23	0.166
15	4.534		0.0000	0.000	880.1	0.442	BB	0.019	770.09	0.531
16	4.580		0.0000	0.000	250.6	0.126	BB	0.018	233.41	0.161
17	4.670		0.0000	0.000	202.0	0.101	BB	0.016	211.70	0.146
18	4.710		0.0000	0.000	6368.1	3.200	BV	0.019	5460.93	3.765
19	4.730		0.0000	0.000	5735.0	2.882	VV	0.020	4813.34	3.319
20	4.758		0.0000	0.000	5833.3	2.931	VV	0.027	3548.77	2.447
21	4.830		0.0000	0.000	2310.8	1.161	VV	0.024	1633.78	1.126
22	4.877		0.0000	0.000	14121.9	7.096	VV	0.023	10243.08	7.062
23	4.916		0.0000	0.000	9229.3	4.637	VV	0.028	5397.79	3.722
24	4.986		0.0000	0.000	898.8	0.452	VV	0.021	703.66	0.485
25	5.033		0.0000	0.000	3783.5	1.901	VB	0.020	3101.94	2.139
26	5.090		0.0000	0.000	8013.8	4.027	BV	0.024	5511.37	3.800
27	5.123		0.0000	0.000	4086.7	2.053	VV	0.020	3347.75	2.308
28	5.155		0.0000	0.000	13593.6	6.830	VV	0.022	10492.99	7.235
29	5.204		0.0000	0.000	2752.9	1.383	VB	0.023	1976.02	1.362
30	5.246		0.0000	0.000	1542.0	0.775	BB	0.015	1690.54	1.166
31	5.285		0.0000	0.000	7833.8	3.936	BB	0.021	6169.21	4.253
32	5.334		0.0000	0.000	789.3	0.397	BB	0.016	806.87	0.556
33	5.378		0.0000	0.000	12162.1	6.111	BB	0.021	9444.02	6.511
34	5.433		0.0000	0.000	807.0	0.406	BB	0.016	848.66	0.585
35	5.487		0.0000	0.000	14327.5	7.199	BB	0.028	8671.55	5.979
36	5.539		0.0000	0.000	563.7	0.283	BB	0.012	767.06	0.529
37	5.580		0.0000	0.000	13212.9	6.639	BB	0.024	9299.83	6.412
38	5.651		0.0000	0.000	643.3	0.323	BB	0.021	511.23	0.352
39	5.695		0.0000	0.000	18007.4	9.048	BB	0.024	12628.88	8.707
40	5.756		0.0000	0.000	2550.4	1.282	BB	0.016	2677.59	1.846
41	5.799		0.0000	0.000	2393.1	1.202	BB	0.019	2049.92	1.413
42	5.919		0.0000	0.000	10557.1	5.305	BB	0.036	4940.24	3.406
43	5.978		0.0000	0.000	1666.5	0.837	BB	0.017	1592.10	1.098
44	6.017		0.0000	0.000	936.8	0.471	BB	0.017	901.18	0.621
45	6.059		0.0000	0.000	113.7	0.057	BB	0.012	152.04	0.105
46	6.098		0.0000	0.000	2992.9	1.504	BB	0.020	2437.42	1.681
47	6.154		0.0000	0.000	11422.4	5.739	BB	0.024	7834.10	5.401
48	6.359		0.0000	0.000	5295.4	2.661	BB	0.021	4231.11	2.917
49	6.405		0.0000	0.000	2252.7	1.132	BB	0.028	1346.68	0.928
50	6.628		0.0000	0.000	1569.9	0.789	BB	0.027	952.45	0.657
51	6.777		0.0000	0.000	3088.3	1.552	BB	0.028	1837.60	1.267
52	6.990		0.0000	0.000	495.3	0.249	BB	0.021	391.12	0.270
53	7.033		0.0000	0.000	82.4	0.041	BB	0.018	77.47	0.053
54	7.150	DCB	1395.5810	65.533	83.0	0.042	BB	0.031	44.28	0.031

Total Area = 199016.4, Total Amount = 2129.577, Total Height = 145040.8, Sample Units = PPB

ASX22 8895

Lancaster Laboratories

Sample Name: **5822072MSD RI ACF DFADSTEA3MSD MSD 093070028A 01216**
Acquired from CP06--S9147B via port 2 on 11/9/09 07:56:15pm by 1151
RTX CLPII
110 to 250 @40/min; to 280 @20/min; to 330 @30/min hold 1.5
Data File: **C:\CPWIN\DATA1\3C6310B.47R**
Method File: C:\CPWIN\DATA1\PESTDB.MET
Calibration File: C:\CPWIN\DATA1\1C6310B.CAL

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
1	2.553		0.0000	0.000	415.8	0.057	BV	0.016	430.33	0.064
2	2.605		0.0000	0.000	214.2	0.029	VB	0.051	70.32	0.010
3	2.810 TCX		576.9393	15.315	333.3	0.046	BB	0.021	263.97	0.039
4	2.929		0.0000	0.000	247.9	0.034	BB	0.027	155.64	0.023
5	3.101		0.0000	0.000	78.5	0.011	BB	0.014	92.27	0.014
6	3.122		0.0000	0.000	205.0	0.028	BV	0.011	299.41	0.044
7	3.142		0.0000	0.000	564.1	0.078	VB	0.013	732.99	0.108
8	3.311		0.0000	0.000	524.4	0.072	BB	0.027	322.82	0.048
9	3.433		0.0000	0.000	530.0	0.073	BV	0.021	422.93	0.063
10	3.530		0.0000	0.000	499.7	0.069	VV	0.027	306.45	0.045
11	3.589		0.0000	0.000	1563.8	0.215	VV	0.020	1293.50	0.191
12	3.626		0.0000	0.000	811.7	0.112	VB	0.028	482.55	0.071
13	3.702		0.0000	0.000	504.9	0.069	BV	0.014	615.71	0.091
14	3.737		0.0000	0.000	4726.2	0.649	VV	0.013	6012.68	0.889
15	3.762		0.0000	0.000	3526.5	0.484	VB	0.020	2902.47	0.429
16	3.871		0.0000	0.000	3312.3	0.455	BB	0.013	4368.02	0.646
17	3.896		0.0000	0.000	347.5	0.048	BB	0.009	668.21	0.099
18	3.919		0.0000	0.000	635.2	0.087	BV	0.013	812.52	0.120
19	3.967		0.0000	0.000	2449.0	0.336	VV	0.014	2859.03	0.423
20	4.011		0.0000	0.000	738.1	0.101	VB	0.014	855.30	0.126
21	4.055		0.0000	0.000	692.3	0.095	BV	0.017	698.81	0.103
22	4.081		0.0000	0.000	5644.4	0.775	VV	0.014	6694.53	0.990
23	4.104		0.0000	0.000	27980.2	3.844	VV	0.015	31179.99	4.610
24	4.130		0.0000	0.000	10206.2	1.402	VV	0.015	11257.31	1.665
25	4.152		0.0000	0.000	3836.2	0.527	VV	0.014	4553.38	0.673
26	4.171		0.0000	0.000	5873.1	0.807	VV	0.016	6110.30	0.903
27	4.207		0.0000	0.000	28670.0	3.939	VV	0.015	31505.87	4.659
28	4.241		0.0000	0.000	35136.7	4.827	VV	0.026	22415.18	3.314
29	4.324		0.0000	0.000	2884.4	0.396	VV	0.018	2659.45	0.393
30	4.359		0.0000	0.000	12596.7	1.731	VB	0.016	13315.54	1.969
31	4.418		0.0000	0.000	22091.0	3.035	BV	0.016	23752.81	3.512
32	4.453		0.0000	0.000	13013.9	1.788	VV	0.017	12841.14	1.899
33	4.481		0.0000	0.000	46430.1	6.379	VB	0.017	44501.67	6.580
34	4.522		0.0000	0.000	4393.1	0.604	BB	0.015	4823.14	0.713
35	4.576		0.0000	0.000	34648.9	4.760	BV	0.017	33223.29	4.913
36	4.612		0.0000	0.000	5888.6	0.809	VV	0.017	5734.80	0.848
37	4.645		0.0000	0.000	58385.6	8.022	VV	0.022	45129.12	6.673
38	4.715		0.0000	0.000	49055.3	6.740	VB	0.020	41605.90	6.152
39	4.821		0.0000	0.000	33100.2	4.548	BB	0.018	30634.63	4.530
40	4.873		0.0000	0.000	37617.6	5.168	BV	0.018	35795.72	5.293
41	4.915		0.0000	0.000	9729.1	1.337	VB	0.018	9190.95	1.359
42	4.954		0.0000	0.000	66241.7	9.101	BV	0.020	55033.71	8.137
43	4.986		0.0000	0.000	15955.5	2.192	VB	0.017	15854.75	2.344
44	5.026		0.0000	0.000	6635.4	0.912	BB	0.013	8489.59	1.255
45	5.088		0.0000	0.000	619.6	0.085	BB	0.019	555.89	0.082
46	5.129		0.0000	0.000	5391.4	0.741	BV	0.027	3290.25	0.487
47	5.180		0.0000	0.000	15664.4	2.152	VV	0.016	16345.60	2.417
48	5.203		0.0000	0.000	16301.5	2.240	VV	0.017	15697.88	2.321
49	5.253		0.0000	0.000	8650.3	1.188	VV	0.017	8465.11	1.252
50	5.288		0.0000	0.000	4655.6	0.640	VV	0.016	4944.74	0.731
51	5.315		0.0000	0.000	2936.6	0.403	VV	0.016	3113.53	0.460
52	5.347		0.0000	0.000	9164.5	1.259	VV	0.015	10138.67	1.499
53	5.368		0.0000	0.000	45262.0	6.218	VB	0.019	39237.86	5.802
54	5.444		0.0000	0.000	1141.9	0.157	BB	0.014	1323.64	0.196
55	5.542		0.0000	0.000	214.4	0.029	BB	0.011	319.52	0.047
56	5.557		0.0000	0.000	68.7	0.009	BB	0.008	143.38	0.021
57	5.582		0.0000	0.000	7094.8	0.975	BB	0.015	7634.91	1.129
58	5.626		0.0000	0.000	31363.1	4.309	BB	0.020	26038.23	3.850
59	5.783		0.0000	0.000	314.4	0.043	BB	0.015	353.91	0.052
60	5.821		0.0000	0.000	613.6	0.084	BB	0.015	684.48	0.101
61	5.850		0.0000	0.000	227.5	0.031	BB	0.012	310.07	0.046
62	5.897		0.0000	0.000	4477.3	0.615	BB	0.018	4234.46	0.626

ASX22 8896

PK#	Ret Time	Name	Amount	Amount%	Area	Area%	Type	Width	Height	Height%
63	6.000		0.0000	0.000	10742.4	1.476	BB	0.020	9175.45	1.357
64	6.200		0.0000	0.000	2351.6	0.323	BB	0.019	2032.66	0.301
65	6.357	DCB	3190.1800	84.685	802.0	0.110	BB	0.018	761.40	0.113
66	6.530		0.0000	0.000	62.3	0.009	BB	0.027	38.94	0.006
67	6.772		0.0000	0.000	503.2	0.069	BB	0.024	349.55	0.052
68	7.047		0.0000	0.000	233.8	0.032	BB	0.042	92.75	0.014
69	7.340		0.0000	0.000	72.5	0.010	BB	0.025	47.64	0.007

Total Area = 727863.8, Total Amount = 3767.12, Total Height = 676299.4, Sample Units = PPB

ASX22:18897

Lancaster Laboratories

LANCASTER LABORATORIES File: C:\CPWIN\DATA1\3C6310.47R

5822072MSD RI ACF DFADSTEA3MSD MSD 093070028A 01216

C:\CPWIN\DATA1\3C6310.47R



Instrument ID: CP06–S9147A Injected On: 11/9/2009 7:48:45 PM Column ID: RTX CLP

File: C:\CPWIN\DATA1\3C6310B.47R

5822072MSD RI ACF DFADSTEA3MSD MSD 093070028A 01216

C:\CPWIN\DATA1\3C6310B.47R



ASX22 8898

Instrument ID: CP06–S9147B Injected On: 11/9/2009 7:48:45 PM Column ID: RTX CLPII

Printed on 11/9/2009 7:56:30 PM

Oven Parameters: 110 to 250 @ 40/min; to 280 @ 20/min; to 330 @ 30/min Multiple Inj: 1

Detector A Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantitation: Height

Detector B Parameters:
Threshold: 1 Width: 0.008 Area Reject: 0
Calibration Type: External Quantiation: Height

Sample Weight: 30 Dilution Factor: 10000
Analyst: 1151

RT A	Height A	Amount A	Compound A	RT B	Height B	Amount B	Compound B
3.311	144	24.46654	TCX	2.81	264	19.23131	TCX
7.15	44	46.51936	DCB	6.357	761	106.3393	DCB

Files:
Area File: C:\CPWIN\DATA1\3C6310.47A
Area File: C:\CPWIN\DATA1\3C6310B.47A
Method A: C:\CPWIN\DATA1\PESTD.MET
Method B: C:\CPWIN\DATA1\PESTDB.MET
Calibration File A: C:\CPWIN\DATA1\1C6310.CAL
Calibration File B: C:\CPWIN\DATA1\1C6310B.CAL
Format A: C:\CPWIN\DATA1\PESTD.FMTA
Format B: C:\CPWIN\DATA1\PESTD.FMTB
Area File Created On: 11/9/2009 7:56:22 PM
File Reported On: 11/9/2009 at 7:56:30 PM

REPLOTS

ASX22 0899

5822072MSD RI ACF DFADSTEA3MSD MSD 093070028A 01216
C:\CPWIN\DATA1\3C6310.47R



5822072MSD RI ACF DFADSTEA3MSD MSD 093070028A 01216
C:\CPWIN\DATA1\3C6310B.47R



ASX22 8988

Extraction/Distillation/Digestion Logs

ASX22 0901

Organic Extraction Batchlog

093060018A

Reviewed By: MB 11 Start Date: 11/3/09 Start Time: 4:00 Am

Tech 1: RK2133 Tech 2:

Prep Group # 11 PCBs - waters Dept: 24 Prep Analysis # 00817 Water Sample Pest. Extraction

QC	Sample Code	Amt (ml)	SS/IS Sol.	Amt (mL)	MS Sol.	Amt (mL)	FV (mL)	pH	pH	BC	Comments
BLANKA	PBLK65	1000	~~SS0927524A~~	1.0			10.0	6	NA	NA	DI H_2O
LCSA	LCSBG	↓	SS0927524A	↓	MS0926824A	1.0	↓	↓	↓	↓	↓
5822065MS (8)	CARS2MS	1020	SS0927524A	↓	MS0926824A	↓	↓	7	↓	30A	tan w/ sed cent 1X
5822066MSD (8)	CARS2MSD	1041	SS0927524A	↓	MS0926824A	↓	↓	↓	↓	↓	↓ ↓

Solvent Used	Lot No.
Hexane	H17E39
Methylene Chloride	095634
Sodium Sulfate	09307A

Spike Solutions: **Witness:** SLM 1785

~~SS0927524A~~ SW846 SURROGATE STANDARD
MS0926824A PCB SPIKE
SS0930124A

SS0930124A
(5) RK2133 11/3/09

(8) foamy during conc.

	Sample #	Sample Code	Amt (ml)	SS/IS Sol.	Amt (mL)	FV (mL)	pH	pH	BC	Comments	Analyses	Due Date	Pr
1	5822062	EBS28	1060	~~SS0927524A~~	1.0	10.0	7	NA	30A	clear	173	11/11/2009	N
2	5822063 (8)	CARS1	805	SS0927524A	↓	↓	↓	↓	↓	black w/ sed cent 2X	173	11/11/2009	N
3	5822064 bkg (8)	CARS2	893	SS0927524A	↓	↓	↓	↓	↓	tan w/ sed cent 1X	173	11/11/2009	N
4	5822067 (8)	DPW29	1018	SS0927524A	↓	↓	↓	↓	↓	↓	173	11/11/2009	N
5	5822068 (8)	STEA1	1041	SS0927524A	↓	↓	↓	↓	↓	tan; slightly foamy cent 1X	173	11/11/2009	N
6	5822078	BB-1	1030	SS0927524A	↓	↓	↓	↓	↓	orange w/ some sed	173	11/11/2009	N
7	5822079	EBW29	1056	SS0927524A	↓	↓	↓	↓	↓	clear	173	11/11/2009	N
8	5822080	EBS29	1044	SS0927524A	↓	↓	↓	↓	↓	↓	173	11/11/2009	N
9	5822187	FE101	1046	SS0927524A	↓	↓	↓	↓	31A	brown w/ some sed	173	11/11/2009	N
10	5822188	FE102	1051	SS0927524A	↓	↓	↓	↓	↓	orange	173	11/11/2009	N
11	5822189	FE103	1057	SS0927524A	↓	↓	↓	↓	↓	↓	173	11/11/2009	N
12	5822190	FE201	1055	SS0927524A	↓	↓	↓	↓	↓	clear	173	11/11/2009	N
13	5822191	FE202	1046	SS0927524A	↓	↓	↓	↓	↓	orange	173	11/11/2009	N
14	5822192	FEM4-	1060	SS0927524A	↓	↓	↓	↓	↓	dark orange	173	11/11/2009	N
15	5822193	FEM5-	1059	SS0927524A	↓	↓	↓	↓	↓	light tan	173	11/11/2009	N
16													
17													
18													
19													
20													

RK2133 11/3/09

ASX22

Rack ID:		Work Station: H.D.3
Internal Standard		Balance # 15661

S-bath ID	1	98 °C	S-bath ID	°C	N-Evap	°C	M-vap	°C

093060018A

Documented temps are NIST corrected.

DF = Dilution Factor FV = Final Volume

Prep-Process Worksheet

Acid Cleanup

Prep Analysis # 00817 Water Sample Pest. Extraction

Prep Group # 11 PCBs - waters

Verified: [illegible]
Start Date: 11/3/09
Start Time: 0315
Tech 1: RK2133
Tech 2:

BATCH NO. 093060018A

	Sample #	Aliquot (ml)	Final Volume (ml)	D.F. Allq.	D.F. F.V.	Comments	Analyses
1	BLANKA	10.0	10.0				
2	LCSA						
3	5822065MS						173
4	5822066MSD						173
5	5822062						173
6	5822063						173
7	5822064						173
8	5822067						173
9	5822068						173
10	5822078						173
11	5822079						173
12	5822080						173
13	5822187						173
14	5822188						173
15	5822189						173
16	5822190						173
17	5822191						173
18	5822192						173
19	5822193						173
20							
21							
22							
23							
24							

RK2133
11/3/09

Additional Comment:

DF = Dilution Factor FV = Final Volume

page 1 of 1

Solvent Used	Lot No.	Solvent Used	Lot No.
H_2SO_4	091646		

ASX22. 8983

Prep-Process Worksheet

Florisil

Prep Analysis # 00817 Water Sample Pest. Extraction

Prep Group # 11 PCBs - waters

Verified: [illegible]
Start Date: 11/3/09
Start Time: 0800
Tech 1: RK2133
Tech 2:

BATCH NO. 093060018A

	Sample #	Aliquot (ml)	Final Volume (ml)	D.F. Aliq.	D.F. F.V.	Comments	Analyses
1	BLANKA	2.0	2.0				
2	LCSA						
3	5822065MS						173
4	5822066MSD						173
5	5822062						173
6	5822063						173
7	5822064						173
8	5822067						173
9	5822068						173
10	5822078						173
11	5822079						173
12	5822080						173
13	5822187						173
14	5822188						173
15	5822189						173
16	5822190						173
17	5822191						173
18	5822192						173
19	5822193						173
20							
21							
22							
23							
24							

3 RK2133
11/3/09

Additional Comment:

DF = Dilution Factor FV = Final Volume

page 1 of 1

Solvent Used	Lot No.	Solvent Used	Lot No.
Florisil	805021-BQ		

S-Evap/bath	°C	S-Evap/bath	°C	N-Evap	°C

The documented temperatures are NIST corrected.

ASX22 8984

Organic Extraction Batchlog

093070028A

Reviewed By: Man/[illegible] Start Date: 11-4-09 Start Time: 9:30

Tech 1: KF1703 Tech 2:

Prep Group # 101 PCB soils Dept: 24 Prep Analysis # 00819 Solid Sample Pesticide Extract

QC	Sample Code	Amt (g)	SS/IS Sol.	Amt (mL)	MS Sol.	Amt (mL)	FV (mL)	pH	pH	BC	Comments
BLANKA	PBLK8D	30	~~SS0927524A~~ *	1.0			10	NA	NA	–	Na_2SO_4
LCSA	LCSDS	30	SS0927524A		MS0926824A	1.0				–	
5822071MS	STEA3MS	30	SS0927524A		MS0926824A					36A	moist, clumpy brown soil
5822072MSD	STEA3MSD	30	~~SS0927524A~~ *		MS0926824A						

Solvent Used	Lot No.
1:1 Methylene Chloride/Aceton	185103009A
Filter Paper	K11671186
Hexane	H17E39
Sodium Sulfate	09308A

Spike Solutions: **Witness:** [illegible]

*~~SS0927524A~~ SW846 SURROGATE STANDARD
MS0926824A PCB SPIKE

*SS0930124A @KF1703 11-4-09

	Sample #	Sample Code	Amt (g)	SS/IS Sol.	Amt (mL)	FV (mL)	pH	pH	BC	Comments	Analyses	Due Date	Pr
1	5814908	WOO54	30	~~SS0927524A~~ *	1.0	10	NA	NA	20A	brown soil	1216	11/12/2009	P
2	5822069	STEA2	30	SS0927524A					36A	brown soil w/ clumps + small yellow + orange pieces	1216	11/11/2009	N
3	5822070 bkg	STEA3	30	SS0927524A						moist, clumpy brown soil	1216	11/11/2009	N
4	5822073	DPS29	30	SS0927524A							1216	11/11/2009	N
5	5822074	ELEC3	30	SS0927524A						brown soil w/ twigs	1216	11/11/2009	N
6	5822075	WHEE1	30	SS0927524A						wet, brown, gritty soil	1216	11/11/2009	N
7	5822076	WHEE2	30	SS0927524A		25				spongy black soil w/ roots	1216	11/11/2009	N
8	5822077	WHEE3	30	~~SS0927524A~~ *		10				brown soil w/ hard pieces	1216	11/11/2009	N
9													
10													
11													
12													
13													
14													
15													
16													
17													
18													
19													
20													

KF1703 11-4-09

Rack ID:		Work Station: SON	HOOD 3
Internal Standard		Balance # 5311	

S-bath ID 6 96 °C | S-bath ID — °C | N-Evap — °C | M-vap — °C

093070028A

Documented temps are NIST corrected.

DF = Dilution Factor FV = Final Volume

Prep-Process Worksheet

Acid Cleanup

Prep Analysis # 00819 Solid Sample Pesticide Extract

Prep Group # 101 PCB soils

Verified: Ma 1860

Start Date: 11-4-09

Start Time: 1400

Tech 1: KF1703

Tech 2:

BATCH NO. 093070028A

	Sample #	Aliquot (ml)	Final Volume (ml)	D.F. Aliq.	D.F. F.V.	Comments	Analyses
1	BLANKA	10	10				
2	LCSA						1216
3	5822071MS						1216
4	5822072MSD						1216
5	5814908						1216
6	5822069						1216
7	5822070						1216
8	5822073						1216
9	5822074						1216
10	5822075						1216
11	5822076						1216
12	5822077						
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							
23							
24							

KF1703 11-4-09

Additional Comment:

DF = Dilution Factor FV = Final Volume

Solvent Used	Lot No.	Solvent Used	Lot No.
		H_2SO_4	093330

ASX22 0906

Prep-Process Worksheet

Florisil

Prep Analysis # 00819 Solid Sample Pesticide Extract

Prep Group # 101 PCB soils

Verified: MA 1806
Start Date: 11-4-09
Start Time: 1500
Tech 1: KF1703
Tech 2:

BATCH NO. 093070028A

	Sample #	Aliquot (ml)	Final Volume (ml)	D.F. Aliq.	D.F. F.V.	Comments	Analyses
1	BLANKA	2	2				
2	LCSA						
3	5822071MS						1216
4	5822072MSD						1216
5	5814908						1216
6	5822069						1216
7	5822070						1216
8	5822073						1216
9	5822074						1216
10	5822075						1216
11	5822076						1216
12	5822077						1216
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							
23							
24							

KF1703
11-4-09

Additional Comment:

DF = Dilution Factor FV = Final Volume

page 1 of 1

Solvent Used	Lot No.	Solvent Used	Lot No.
		2	
Florisil	805021-B6		
S-Evap/bath	°C	S-Evap/bath	°C
N-Evap	°C		

The documented temperatures are NIST corrected.

ASX22: 8987

Prep-Process Worksheet

Copper

Prep Analysis # 00819 Solid Sample Pesticide Extract

Prep Group # 101 PCB soils

Verified: Na 1856

Start Date: 11-4-09

Start Time: 15:15

Tech 1: KF1703

Tech 2:

BATCH NO. 093070028A

	Sample #	Aliquot (ml)	Final Volume (ml)	D.F. Aliq.	D.F. F.V.	Comments	Analyses
1	BLANKA	2	2				
2	LCSA						
3	5822071MS						1216
4	5822072MSD						1216
5	5814908						1216
6	5822069						1216
7	5822070						1216
8	5822073						1216
9	5822074						1216
10	5822075						1216
11	5822076						1216
12	5822077						1216
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							
23							
24							

KF1703 11-4-09

Additional Comment:

DF = Dilution Factor FV = Final Volume

page 1 of 1

Solvent Used	Lot No.	Solvent Used	Lot No.
		Copper	9020110-AS

ASX22 5988

Moisture Data

Lancaster Laboratories

CLIENT: Stantec
SDG: ASX22

MOISTURE

SAMPLE NUMBERS:

Sample #	Sample Code
5822069	STEA2
5822070	STEA3BKG
5822071	STEA3MS
5822072	STEA3MSD
5822073	DPS29FD
5822074	ELEC3
5822075	WHEE1
5822076	WHEE2
5822077	WHEE3

Laboratory Compliance Quality Control

Analysis Name	LCS %REC	LCSD %REC	LCS/LCSD Limits	RPD	RPD Max
Batch number: 09307820007B	Sample number(s): 5822069, 5822073-5822077				
Moisture	100		99-101		
Batch number: 09308820003A	Sample number(s): 5822070-5822072				
Moisture	100		99-101		
Moisture	100		99-101		
Moisture Duplicate	100		99-101		

Sample Matrix Quality Control

Analysis Name	BKG Conc	DUP Conc	RPD	RPD Max
Batch number: 09307820007B	Sample number(s): 5822069, 5822073-5822077			
Moisture	25.8	35.3	31*	15
Batch number: 09308820003A	Sample number(s): 5822070-5822072			
Moisture	35.1	32.2	9	15
Moisture Duplicate	35.1	32.2	9	15
Moisture	35.1	32.2	9	15

* - Outside of specification

(1) - The result for one or both determinations was less than five times the LOQ.

ASX22 8910

Lancaster Laboratories

MOISTURE

CLIENT: Stantec
SDG: ASX22

Moisture Data Report

Batch #: 09307820007

Sample ID	Batch ID	Analysis#	Tare Wt	Sample Wt	Dry Wt	%Moisture	Analysis Date (Emp#)	Verified Date (Emp#)
LCS 89.5% Std.			1.1096	5.0489	1.6468	89.36	11/ 3/09 (1201/SWF)	11/ 4/09 (0236/CW)
5822069	B	00111	1.0980	8.3539	9.2173	2.81	11/ 3/09 (1201/SWF)	11/ 4/09 (0236/CW)
5822073FD	B	00111	1.0992	6.9535	6.0862	28.28	11/ 3/09 (1201/SWF)	11/ 4/09 (0236/CW)
5822074	B	00111	1.0980	5.3203	6.2429	3.30	11/ 3/09 (1201/SWF)	11/ 4/09 (0236/CW)
5822075	B	00111	1.1103	8.9787	7.5464	28.32	11/ 3/09 (1201/SWF)	11/ 4/09 (0236/CW)
5822076	B	00111	1.0970	6.6938	6.8150	14.58	11/ 3/09 (1201/SWF)	11/ 4/09 (0236/CW)
5822077	B	00111	1.0951	9.4470	10.0909	4.78	11/ 3/09 (1201/SWF)	11/ 4/09 (0236/CW)

Batch #: 09308820003

Sample ID	Batch ID	Analysis#	Tare Wt	Sample Wt	Dry Wt	%Moisture	Analysis Date (Emp#)	Verified Date (Emp#)
LCS 89.5% Std.			1.1067	5.0374	1.6550	89.12	11/ 4/09 (1201/SWF)	11/ 5/09 (0236/CW)
5822070BKG	A	00111	1.0808	6.3841	5.2214	35.14	11/ 4/09 (1201/SWF)	11/ 5/09 (0236/CW)
5822071MS	A	00118				35.14	11/ 4/09 (1201/SWF)	11/ 5/09 (0236/CW)
5822072MSD	A	00118				35.14	11/ 4/09 (1201/SWF)	11/ 5/09 (0236/CW)
5822072MSD	A	00121	1.1237	6.8983	5.8008	32.20	11/ 4/09 (1201/SWF)	11/ 5/09 (0236/CW)

ASX22 0911