ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Soil

Laboratory Reference #: MWI 8781

Sampled : 03-31-97
Received: 03-31-97
Analyzed: 04-01-97
Reported: 04-01-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	Client Sample Number	Sample Result (mg/kg)
97031109	B37-RE4-SP34	110
97031110	B37-RE4-SP35	N.D.
97031111	B37-GS-112-5.5'	15
97031112	B37-GS-113-5.5'	14
97031113	B37-GS-114-5.5'	13
97031114	B37-GS-115-5.5'	28
97031115	B37-GS-116-8'	N.D.
97031116	B37-RE4-SP36	77
Detection Limit:		8.0

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP35

Laboratory Sample Number: 97031110
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-02-97
Reported: 04-14-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	150	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	26	<---
Cobalt	6010	80	8000	0.5	7.1	<---
Copper	6010	25	2500	0.1	15	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	9.1	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	26	<---
Zinc	6010	250	5000	0.1	59	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

'ontgomery Watson
.TTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP36

Laboratory Sample Number: 97031116
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-02-97
Reported: 04-14-97

CCR - METALS

Analyte	EPA Method	STLC Limits mg/l	TTLC Limits mg/kg	Detection Limit mg/kg	Analysis Result mg/kg	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	100	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	22	<---
'obalt	6010	80	8000	0.5	8.3	<---
.opper	6010	25	2500	0.1	21	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	11	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	25	<---
Zinc	6010	250	5000	0.1	42	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP35
Laboratory Sample Number: 97031110
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-07-97
Reported: 04-14-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
1,1-Dichloroethene	75-35-4	2.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP35

Laboratory Sample Number: 97031110

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/kg)	SAMPLE RESULT (ug/kg)
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
n-Butylbenzene	104-51-8	2.5	N.D.
1,2-Dichlorobenzene	95-50-1	2.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*103*
Toluene-d8	*95*
4-Bromofluorobenbzene	*101*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP36
Laboratory Sample Number: 97031116
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-07-97
Reported: 04-14-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
1,1-Dichloroethene	75-35-4	2.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ample Description: Soil, B37-RE4-SP36

Laboratory Sample Number: 97031116

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
n-Butylbenzene	104-51-8	2.5	N.D.
1,2-Dichlorobenzene	95-50-1	2.5	N.D.
-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
aboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*110*
Toluene-d8	*91*
4-Bromofluorobenbzene	*91*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP35
Laboratory Sample #: 97031110
Laboratory Reference #: MWI 8781

Sampled : 03-31-97
Received: 03-31-97
Analyzed: 04-07-97
Reported: 04-14-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
Acenaphthene	83-32-9	100	N.D.	
Acenaphthylene	208-96-8	100	N.D.	
Aniline	62-53-3	100	N.D.	
Anthracene	120-12-7	100	N.D.	
Benzidine	92-87-5	500	N.D.	
Benzoic Acid	65-85-0	250	N.D.	
Benzo(a)anthracene	56-55-3	100	N.D.	
Benzo(b)fluoranthene	205-99-2	250	N.D.	
Benzo(k)fluoranthene	207-08-9	250	N.D.	
Benzo(g,h,i)perylene	191-24-2	250	N.D.	
Benzo(a)pyrene	50-32-8	250	N.D.	
Benzyl alcohol	100-51-6	100	N.D.	
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.	
Bis(2-chloroethyl)ether	111-44-4	100	N.D.	
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.	
Bis(2-ethylhexyl)phthalate	117-81-7	100	190	<---
4-Bromophenyl phenyl ether	101-55-3	100	N.D.	
Butyl benzyl phthalate	85-68-7	100	N.D.	
4-Chloroaniline	106-47-8	100	N.D.	
2-Chloronaphthalene	91-58-7	100	N.D.	
4-Chloro-3-methylphenol	59-50-7	100	N.D.	
2-Chlorophenol	95-57-8	100	N.D.	
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.	
Chrysene	218-0109	100	N.D.	
Dibenz(a,h)anthracene	53-70-3	100	N.D.	
Dibenzofuran	132-64-9	100	N.D.	
Di-n-butyl phthalate	84-74-2	250	N.D.	
1,3-Dichlorobenzene	541-73-1	100	N.D.	
1,4-Dichlorobenzene	106-46-7	100	N.D.	
1,2-Dichlorobenzene	95-50-1	100	N.D.	
3,3'-Dichlorobenzidine	91-94-1	100	N.D.	
2,4-Dichlorophenol	120-83-2	100	N.D.	
Diethyl phthalate	84-66-2	100	N.D.	
2,4-Dimethylphenol	105-67-9	100	N.D.	
Dimethyl phthalate	131-11-3	100	N.D.	
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.	
2,4-Dinitrophenol	51-28-5	100	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP35
Laboratory Sample #: 97031110

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
2,4-Dinitrotoluene	121-14-2	250	N.D.
2,6-Dinitrotoluene	606-20-2	250	N.D.
Di-n-octyl phthalate	117-84-0	250	N.D.
Fluoranthene	206-44-0	100	N.D.
Fluorene	86-73-7	100	N.D.
Hexachlorobenzene	118-74-1	100	N.D.
Hexachlorobutadiene	87-68-3	100	N.D.
Hexachlorocyclopentadiene	77-47-4	100	N.D.
Hexachloroethane	67-72-1	100	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.
Isophorone	78-59-1	100	N.D.
2-Methylnaphthalene	91-57-6	100	N.D.
2-Methylphenol	95-48-7	100	N.D.
4-Methylphenol	106-44-5	100	N.D.
Naphthalene	91-20-3	100	N.D.
2-Nitroaniline	88-74-4	250	N.D.
3-Nitroaniline	99-09-2	250	N.D.
4-Nitroaniline	100-01-6	250	N.D.
Nitrobenzene	98-95-3	100	N.D.
2-Nitrophenol	88-75-5	100	N.D.
4-Nitrophenol	100-02-7	100	N.D.
N-Nitrosodiphenylamine	86-30-6	100	N.D.
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.
N-Nitrosodimethylamine	62-75-9	100	N.D.
Pentachlorophenol	87-86-5	250	N.D.
Phenanthrene	85-01-8	100	N.D.
Phenol	108-95-2	100	N.D.
Pyrene	129-00-0	100	N.D.
1,2,4-Trichlorobenzene	120-82-1	100	N.D.
2,4,5-Trichlorophenol	95-95-4	100	N.D.
2,4,6-Trichlorophenol	88-06-2	100	N.D.

Analytes reported as N.D. were not present above the stated limit of detectic

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP36
Laboratory Sample #: 97031116
Laboratory Reference #: MWI 8781

Sampled : 03-31-97
Received: 03-31-97
Analyzed: 04-07-97
Reported: 04-14-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
Benzo(k)fluoranthene	207-08-9	250	N.D.
Benzo(g,h,i)perylene	191-24-2	250	N.D.
Benzo(a)pyrene	50-32-8	250	N.D.
Benzyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
Dimethyl phthalate	131-11-3	100	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.
2,4-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP36
Laboratory Sample #: 97031116

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
2,4-Dinitrotoluene	121-14-2	250	N.D.
2,6-Dinitrotoluene	606-20-2	250	N.D.
Di-n-octyl phthalate	117-84-0	250	N.D.
Fluoranthene	206-44-0	100	N.D.
Fluorene	86-73-7	100	N.D.
Hexachlorobenzene	118-74-1	100	N.D.
Hexachlorobutadiene	87-68-3	100	N.D.
Hexachlorocyclopentadiene	77-47-4	100	N.D.
Hexachloroethane	67-72-1	100	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.
Isophorone	78-59-1	100	N.D.
2-Methylnaphthalene	91-57-6	100	N.D.
·Methylphenol	95-48-7	100	N.D.
''ethylphenol	106-44-5	100	N.D.
nthalene	91-20-3	100	N.D.
-Nitroaniline	88-74-4	250	N.D.
3-Nitroaniline	99-09-2	250	N.D.
4-Nitroaniline	100-01-6	250	N.D.
Nitrobenzene	98-95-3	100	N.D.
2-Nitrophenol	88-75-5	100	N.D.
4-Nitrophenol	100-02-7	100	N.D.
N-Nitrosodiphenylamine	86-30-6	100	N.D.
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.
N-Nitrosodimethylamine	62-75-9	100	N.D.
Pentachlorophenol	87-86-5	250	N.D.
Phenanthrene	85-01-8	100	N.D.
Phenol	108-95-2	100	N.D.
Pyrene	129-00-0	100	N.D.
1,2,4-Trichlorobenzene	120-82-1	100	N.D.
2,4,5-Trichlorophenol	95-95-4	100	N.D.
2,4,6-Trichlorophenol	88-06-2	100	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

: Noorani
oratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Water

Laboratory Reference #: MWI 8781

Sampled : 03-31-97
Received: 03-31-97
Analyzed: 04-02-97
Reported: 04-14-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/l)*
97031117	Equip Blank	N.D.
97031118	Rinstate Blank	N.D.
Detection Limit:		0.5

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97031117
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-01-97
Reported: 04-14-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97031118
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-01-97
Reported: 04-14-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss'
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Equip Blank
Laboratory Sample Number: 97031117
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-03-97
Reported: 04-14-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97031117

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*106*
Toluene-d8	*93*
4-Bromofluorobenzene	*107*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss,
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample Number: 97031118
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-12-97
Reported: 04-14-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97031118

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*104*
Toluene-d8	*94*
4-Bromofluorobenzene	*108*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Trip Blank
Laboratory Sample Number: 97031119
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-12-97
Reported: 04-14-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Trip Blank

Laboratory Sample Number: 97031119

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*107*
Toluene-d8	*91*
4-Bromofluorobenzene	*107*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
:TN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank
Laboratory Sample #: 97031117
Laboratory Reference #: MWI 8781

Sampled : 03-31-97
Received: 03-31-97
Analyzed: 04-04-97
Reported: 04-14-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
⁻enzo(k)fluoranthene	207-08-9	25	N.D.
·nzo(g,h,i)perylene	191-24-2	25	N.D.
zo(a)pyrene	50-32-8	25	N.D.
nzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
·methyl phthalate	131-11-3	5.0	N.D.
-Dinitro-2-methylphenol	534-52-1	50	N.D.
-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Equip Blank
Laboratory Sample #: 97031117

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
4-Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97031118
Laboratory Reference #: MWI 8781

Sampled : 03-31-97
Received: 03-31-97
Analyzed: 04-04-97
Reported: 04-14-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
nzo(k)fluoranthene	207-08-9	25	N.D.
zo(g,h,i)perylene	191-24-2	25	N.D.
o(a)pyrene	50-32-8	25	N.D.
nzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
ethyl phthalate	131-11-3	5.0	N.D.
Dinitro-2-methylphenol	534-52-1	50	N.D.
Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97031118

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
4-Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97031117
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-08-97
Reported: 04-14-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97031118
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-08-97
Reported: 04-14-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97031117
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-08-97
Reported: 04-14-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97031118
Laboratory Reference #: MWI 8781

Sampled: 03-31-97
Received: 03-31-97
Analyzed: 04-08-97
Reported: 04-14-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/01/97
Laboratory Sample No : 97030931
Laboratory Reference No : MWI 8781

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	100	100	100	100	100	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/02/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8781

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	2.5	2.3	2.2	92	88	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/01/97
Laboratory Sample No : 97031017
Laboratory Reference No : MWI 8781

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	3.000	2.800	3.000	93	100	7
Arsenic	0.00	1.000	0.960	0.920	96	92	4
Barium	0.00	0.100	0.097	0.095	97	95	2
Beryllium	0.00	0.100	0.094	0.089	94	89	5
Cadmium	0.00	0.100	0.098	0.102	98	102	4
Chromium (VI)	0.00	0.500	0.430	0.410	86	82	5
Chromium (Total)	0.00	0.100	0.098	0.095	98	95	3
Cobalt	0.00	0.100	0.099	0.096	99	96	3
Copper	0.00	0.100	0.102	0.098	102	98	4
Lead	0.00	1.000	0.910	0.800	91	80	13
Mercury	0.000	0.020	0.022	0.020	110	100	10
Molybdenum	0.00	1.000	1.150	1.060	115	106	8
Nickel	0.00	0.500	0.520	0.490	104	98	6
Selenium	0.00	1.000	1.090	1.050	109	105	4
Silver	0.00	0.500	0.570	0.570	114	114	0
Thallium	0.00	3.000	2.600	3.000	87	100	14
Vanadium	0.00	0.500	0.470	0.450	94	90	4
Zinc	0.00	0.100	0.098	0.088	98	88	11

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/02/97
Laboratory Sample No : 97031116
Laboratory Reference No : MWI 8781

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.0	30.0	30.7	30.1	102	100	2
Arsenic	0.0	10.0	10.0	9.2	100	92	8
Barium	10.0	10.0	21.8	20.9	118	109	4
Beryllium	0.00	1.00	0.99	1.05	99	105	6
Cadmium	0.00	1.00	0.99	1.00	99	100	1
Chromium (Total)	2.2	5.0	7.6	7.3	108	102	4
Chromium (VI)	0.00	1.00	1.05	0.92	105	92	13
Cobalt	0.83	1.00	1.80	1.70	97	87	6
Copper	2.1	5.0	7.5	7.4	108	106	1
Lead	0.00	10.00	9.40	9.70	94	97	3
Mercury	0.000	0.020	0.022	0.019	110	95	15
Molybdenum	0.0	10.0	9.1	9.2	91	92	1
Nickel	1.1	5.0	5.6	5.4	90	86	4
Selenium	0.0	10.0	9.9	9.3	99	93	6
Silver	0.0	5.0	4.3	4.3	86	86	0
Thallium	0.0	30.0	25.4	27.0	85	90	6
Vanadium	2.5	5.0	7.4	7.2	98	94	3
Zinc	4.2	5.0	8.7	8.7	90	90	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/07/97
Laboratory Sample No : 97041110
Laboratory Reference No : MWI 8781

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	54	52	108	104	4
1,1-Dichloroethene	0.0	50	53	52	106	104	2
Trichloroethene	0.0	50	54	52	108	104	4
Toluene	0.0	50	55	54	110	108	2
Chlorobenzene	0.0	50	48	50	96	100	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/12/97
Laboratory Sample No : 97031118
Laboratory Reference No : MWI 8781

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	20	17	19	85	95	11
1,1-Dichloroethene	0.0	20	15	17	75	85	13
Trichloroethene	0.0	20	17	18	85	90	6
Toluene	0.0	20	18	20	90	100	11
Chlorobenzene	0.0	20	18	19	90	95	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/04/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8781

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	41	39	82	78	5
n-Nitroso-di-n-propylamine	0.0	50	45	42	90	84	7
1,2,4-Trichlorobenzene	0.0	50	40	40	80	80	0
Acenaphthene	0.0	50	47	45	94	90	4
Pyrene	0.0	50	44	45	88	90	2
Pentachlorophenol	0.0	100	82	83	82	83	1
4-Chloro-3-Methylphenol	0.0	100	70	69	70	69	1
2-Chlorophenol	0.0	100	81	79	81	79	3
Phenol	0.0	100	31	31	31	31	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/07/97
Laboratory Sample No : 97031115
Laboratory Reference No : MWI 8781

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	42	44	84	88	5
n-Nitroso-di-n-propylamine	0.0	50	44	47	88	94	7
1,2,4-Trichlorobenzene	0.0	50	42	44	84	88	5
Acenaphthene	0.0	50	45	47	90	94	4
Pyrene	0.0	50	44	45	88	90	2
Pentachlorophenol	0.0	100	77	82	77	82	6
4-Chloro-3-Methylphenol	0.0	100	69	72	69	72	4
2-Chlorophenol	0.0	100	76	79	76	79	4
Phenol	0.0	100	71	74	71	74	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Organochlorine Pesticides (EPA 8080)

Date of Analysis : 04/08/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8781

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
4,4'-DDT	0.0	1.00	0.73	0.69	73	69	6

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : PCB 'S (EPA 8080)

Date of Analysis : 04/08/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8781

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
PCB-1260	0.0	50	45	41	90	82	9

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532
Tustin, CA 92680
(714) 832-0064, Fax (714) 832-0067

A[illegible] Request and Chain [illegible] Custody Record

Lab Job No. _____
Page 1 of _____

24 hr. TAT on Soil TRPH only
REQUIRED TAT: Standard TAT otherwise

CUSTOMER INFORMATION	PROJECT INFORMATION
COMPANY: Montgomery Watson	PROJECT NAME: McDonnell Douglas C.C. Facility
SEND REPORT TO: Fred Strauss	NUMBER: 1706035.01090010
ADDRESS: 250 N. Madison Ave. Pasadena, CA 91101	LOCATION: BEDG 37 Area
	ADDRESS: 19503 S. Normandie Ave. Torrance CA
PHONE (818) 568-6582 FAX (818) 796-5441	SAMPLED BY: JAM / IV

ANALYSIS/METHOD REQUEST

SAMPLE ID	NO. OF CONTAINERS	SAMPLE DATE	SAMPLE TIME	SAMPLE MATRIX	CONTAINER TYPE	PRES.	TRPH 418.1	7.1622 metals	VOCs (8260)	SVOCs (8270)	PCBs (8080)	Pesticides (8080)			REMARKS/PRECAUTIONS
B37-RE4-SP34	2	3/31/97	0840	Soil	2" x 6" SS Sleeve	ICE	X	X	X	X					
B37-RE4-SP35	↓	↓	1050	↓	↓	↓	X	X	X	X					
B37-GS-112-5.5'	↓	↓	1125	↓	↓	↓	X	X							
B37-GS-113-5.5'	↓	↓	1135	↓	↓	↓	X	X							
B37-GS-114-5.5'	↓	↓	1140	↓	↓	↓	X	X							
B37-GS-115-5.5'	↓	↓	1145	↓	↓	↓	X	X							
B37-GS-116-8'	↓	↓	1155	↓	↓	↓	X	X							
B37-RE4-SP36	↓	↓	1350	↓	↓	↓	X	X	X	X					
Equipment Blank	7	3/31/97	1615	Water	4 Amber 1 Amber 2 Vials	[illegible] HCl	X	X	X	X	X	X			
Rinsate Blank	7	↓	1630	↓	↓	↓	X	X	X	X	X	X			
Trip Blank	2	↓	—	↓	2 vials	HCl			X						

Total No. of Samples: 11

Method of Shipment: Courier (Orange Coast Analytical)

Relinquished By:	Date/Time:	Received By:	Date/Time:
John Mann	3/31/97 16:45		
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received For Lab By: [signature]	Date/Time: 03-31-97 4:45 P.M.

Reporting Format: (check)
NORMAL _____ S.D. HMMD _____
RWQCB _____ OTHER _____

Sample Integrity: (check)
intact _____ on ice _____

All samples remain the property of the client who is responsible for disposal. A disposal fee may be imposed if client fails to pickup samples.

BOE-C6-0143587

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Soil

Laboratory Reference #: MWI 8785

Sampled : 04-01-97
Received: 04-01-97
Analyzed: 04-02-97
Reported: 04-02-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/kg)*
97040012	B37-GS-117-5'	N.D.
97040013	B37-GS-118-6.5'	N.D.
97040014	B37-GS-119-6.5'	N.D.
97040015	B37-GS-120-6.5'	N.D.
97040016	B37-GS-121-6.5'	N.D.
97040017	B37-GS-122-6.5'	N.D.
97040018	B37-GS-123-6.5'	N.D.
97040019	B37-GS-124-6.5'	N.D.
97040020	B37-GS-125-6.5'	N.D.
97040021	B37-GS-126-6.5'	N.D.
97040022	B37-GS-127-6.5'	N.D.
97040023	B37-GS-128-6.5'	N.D.
97040024	B37-GS-129-7'	N.D.
97040025	B37-GS-130-9'	N.D.
97040026	B37-GS-131-9'	N.D.
97040027	B37-RE4-SP37	88
97040028	B37-RE4-SP38	N.D.
97040029	B37-RE4-SP39	340
Detection Limit:		8.0

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

ark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP37

Laboratory Sample Number: 97040027
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-02-97
Reported: 04-15-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	140	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	32	<---
Cobalt	6010	80	8000	0.5	8.4	<---
Copper	6010	25	2500	0.1	22	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	11	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	36	<---
Zinc	6010	250	5000	0.1	53	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP38

Laboratory Sample Number: 97040028
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-02-97
Reported: 04-15-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	99	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	27	<---
Cobalt	6010	80	8000	0.5	6.7	<---
Copper	6010	25	2500	0.1	12	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	9.6	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	28	<---
Zinc	6010	250	5000	0.1	42	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP39

Laboratory Sample Number: 97040029
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-02-97
Reported: 04-15-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	110	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	26	<---
Cobalt	6010	80	8000	0.5	6.4	<---
Copper	6010	25	2500	0.1	13	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	11	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	26	<---
Zinc	6010	250	5000	0.1	35	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP37
Laboratory Sample Number: 97040027
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-09-97
Reported: 04-15-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
1,1-Dichloroethene	75-35-4	2.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.
,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP37

Laboratory Sample Number: 97040027

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
n-Butylbenzene	104-51-8	2.5	N.D.
1,2-Dichlorobenzene	95-50-1	2.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*102*
Toluene-d8	*89*
4-Bromofluorobenbzene	*101*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP38
Laboratory Sample Number: 97040028
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-09-97
Reported: 04-15-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
1,1-Dichloroethene	75-35-4	2.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP38

Laboratory Sample Number: 97040028

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/kg)	SAMPLE RESULT (ug/kg)
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
n-Butylbenzene	104-51-8	2.5	N.D.
1,2-Dichlorobenzene	95-50-1	2.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*99*
Toluene-d8	*89*
4-Bromofluorobenbzene	*97*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP39
Laboratory Sample Number: 97040029
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-10-97
Reported: 04-15-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*	
Benzene	71-43-2	2.5	N.D.	
Bromodichloromethane	75-27-4	2.5	N.D.	
Bromoform	75-25-2	2.5	N.D.	
Bromomethane	74-83-9	2.5	N.D.	
Carbon Disulfide	75-15-0	5.0	N.D.	
Carbon tetrachloride	56-23-5	2.5	N.D.	
Chlorobenzene	108-90-7	2.5	N.D.	
Chlorodibromomethane	124-48-1	2.5	N.D.	
Chloroethane	75-00-3	2.5	N.D.	
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.	
Chloroform	67-66-3	2.5	N.D.	
Chloromethane	74-87-3	2.5	N.D.	
1,1-Dichloroethane	75-35-3	2.5	N.D.	
1,2-Dichloroethane	107-06-2	2.5	N.D.	
1,1-Dichloroethene	75-35-4	2.5	N.D.	
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.	
.,2-Dichloropropane	78-87-5	2.5	N.D.	
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.	
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.	
Ethylbenzene	100-41-4	2.5	N.D.	
Methylene chloride	75-09-2	5.0	N.D.	
Styrene	100-42-5	2.5	N.D.	
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.	
Tetrachloroethene	127-18-4	2.5	N.D.	
Toluene	108-88-3	2.5	N.D.	
1,1,1-Trichloroethane	71-55-6	2.5	N.D.	
1,1,2-Trichloroethane	79-00-5	2.5	N.D.	
Trichloroethene	79-01-6	2.5	8.6	<---
Trichlorofluoromethane	75-69-4	5.0	N.D.	
Vinyl acetate	108-05-4	5.0	N.D.	
Vinyl chloride	75-01-4	2.5	N.D.	
Total Xylenes	1330-20-7	2.5	N.D.	
Dichlorofluoromethane	75-71-8	2.5	N.D.	
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.	
2,2-Dichloropropane	590-20-7	2.5	N.D.	
Bromochloromethane	74-97-5	2.5	N.D.	
1,1-Dichloropropene	563-58-6	2.5	N.D.	
1,2-Dichloropropane	78-87-5	2.5	N.D.	
Dibromomethane	74-95-3	2.5	N.D.	
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.	
1,1,2-Trichloroethane	79-00-5	2.5	N.D.	
1,2-Dibromoethane	106-93-4	2.5	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP39

Laboratory Sample Number: 97040029

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*	
1,3-Dichloropropane	142-28-9	2.5	N.D.	
Dibromochloromethane	124-48-1	2.5	N.D.	
Isopropylbenzene	98-82-8	2.5	N.D.	
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.	
1,2,3-Trichloropropane	96-18-4	2.5	N.D.	
Bromobenzene	108-86-1	2.5	N.D.	
n-Propylbenzene	103-65-1	2.5	4.0	<---
2-Chlorotoluene	95-49-8	2.5	N.D.	
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.	
4-Chlorotoluene	106-43-4	2.5	N.D.	
tert-Butylbenzene	98-06-6	2.5	N.D.	
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.	
sec-Butylbenzene	135-98-8	2.5	N.D.	
4-Isopropyltoluene	99-87-6	2.5	N.D.	
1,3-Dichlorobenzene	541-73-1	2.5	N.D.	
1,4-Dichlorobenzene	106-46-7	2.5	N.D.	
n-Butylbenzene	104-51-8	2.5	6.0	<---
1,2-Dichlorobenzene	95-50-1	2.5	N.D.	
1-2-Dibromo-2-CPA	96-12-8	5.0	N.D.	
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.	
Hexachlorobutadiene	87-68-3	2.5	N.D.	
Naphthalene	91-20-3	2.5	11	<---
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*93*
Toluene-d8	*91*
4-Bromofluorobenbzene	*103*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
TN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP37
Laboratory Sample #: 97040027
Laboratory Reference #: MWI 8785

Sampled : 04-01-97
Received: 04-01-97
Analyzed: 04-07-97
Reported: 04-15-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
nzo(k)fluoranthene	207-08-9	250	N.D.
nzo(g,h,i)perylene	191-24-2	250	N.D.
zo(a)pyrene	50-32-8	250	N.D.
nzyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
methyl phthalate	131-11-3	100	N.D.
-Dinitro-2-methylphenol	534-52-1	100	N.D.
-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP37
Laboratory Sample #: 97040027

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	150	<---
Fluorene	86-73-7	100	N.D.	
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.	
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	N.D.	
2-Methylphenol	95-48-7	100	N.D.	
4-Methylphenol	106-44-5	100	N.D.	
Naphthalene	91-20-3	100	N.D.	
2-Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	N.D.	
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	120	<---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
.TN: Mr. Fred Strauss
.50 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP38
Laboratory Sample #: 97040028
Laboratory Reference #: MWI 8785

Sampled : 04-01-97
Received: 04-01-97
Analyzed: 04-07-97
Reported: 04-15-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
Benzo(k)fluoranthene	207-08-9	250	N.D.
nzo(g,h,i)perylene	191-24-2	250	N.D.
zo(a)pyrene	50-32-8	250	N.D.
.zyl alcohol	100-51-6	100	N.D.
bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
Dimethyl phthalate	131-11-3	100	N.D.
-Dinitro-2-methylphenol	534-52-1	100	N.D.
Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP38
Laboratory Sample #: 97040028

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
2,4-Dinitrotoluene	121-14-2	250	N.D.
2,6-Dinitrotoluene	606-20-2	250	N.D.
Di-n-octyl phthalate	117-84-0	250	N.D.
Fluoranthene	206-44-0	100	N.D.
Fluorene	86-73-7	100	N.D.
Hexachlorobenzene	118-74-1	100	N.D.
Hexachlorobutadiene	87-68-3	100	N.D.
Hexachlorocyclopentadiene	77-47-4	100	N.D.
Hexachloroethane	67-72-1	100	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.
Isophorone	78-59-1	100	N.D.
2-Methylnaphthalene	91-57-6	100	N.D.
2-Methylphenol	95-48-7	100	N.D.
4-Methylphenol	106-44-5	100	N.D.
Naphthalene	91-20-3	100	N.D.
2-Nitroaniline	88-74-4	250	N.D.
3-Nitroaniline	99-09-2	250	N.D.
4-Nitroaniline	100-01-6	250	N.D.
Nitrobenzene	98-95-3	100	N.D.
2-Nitrophenol	88-75-5	100	N.D.
4-Nitrophenol	100-02-7	100	N.D.
N-Nitrosodiphenylamine	86-30-6	100	N.D.
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.
N-Nitrosodimethylamine	62-75-9	100	N.D.
Pentachlorophenol	87-86-5	250	N.D.
Phenanthrene	85-01-8	100	N.D.
Phenol	108-95-2	100	N.D.
Pyrene	129-00-0	100	N.D.
1,2,4-Trichlorobenzene	120-82-1	100	N.D.
2,4,5-Trichlorophenol	95-95-4	100	N.D.
2,4,6-Trichlorophenol	88-06-2	100	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sampled : 04-01-97
Received: 04-01-97
Analyzed: 04-07-97
Reported: 04-15-97

Sample Description: Soil, B37-RE4-SP39
Laboratory Sample #: 97040029
Laboratory Reference #: MWI 8785

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
Acenaphthene	83-32-9	100	130	<---
Acenaphthylene	208-96-8	100	N.D.	
Aniline	62-53-3	100	N.D.	
Anthracene	120-12-7	100	150	<---
Benzidine	92-87-5	500	N.D.	
Benzoic Acid	65-85-0	250	N.D.	
Benzo(a)anthracene	56-55-3	100	130	<---
Benzo(b)fluoranthene	205-99-2	250	N.D.	
Benzo(k)fluoranthene	207-08-9	250	N.D.	
Benzo(g,h,i)perylene	191-24-2	250	N.D.	
Benzo(a)pyrene	50-32-8	250	N.D.	
Benzyl alcohol	100-51-6	100	N.D.	
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.	
Bis(2-chloroethyl)ether	111-44-4	100	N.D.	
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.	
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.	
4-Bromophenyl phenyl ether	101-55-3	100	N.D.	
Butyl benzyl phthalate	85-68-7	100	N.D.	
4-Chloroaniline	106-47-8	100	N.D.	
2-Chloronaphthalene	91-58-7	100	N.D.	
4-Chloro-3-methylphenol	59-50-7	100	N.D.	
2-Chlorophenol	95-57-8	100	N.D.	
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.	
Chrysene	218-0109	100	180	<---
Dibenz(a,h)anthracene	53-70-3	100	N.D.	
Dibenzofuran	132-64-9	100	N.D.	
Di-n-butyl phthalate	84-74-2	250	N.D.	
1,3-Dichlorobenzene	541-73-1	100	N.D.	
1,4-Dichlorobenzene	106-46-7	100	N.D.	
1,2-Dichlorobenzene	95-50-1	100	N.D.	
3,3'-Dichlorobenzidine	91-94-1	100	N.D.	
2,4-Dichlorophenol	120-83-2	100	N.D.	
Diethyl phthalate	84-66-2	100	N.D.	
2,4-Dimethylphenol	105-67-9	100	N.D.	
Dimethyl phthalate	131-11-3	100	N.D.	
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.	
2,4-Dinitrophenol	51-28-5	100	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP39
Laboratory Sample #: 97040029

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	120	<---
Fluorene	86-73-7	100	N.D.	
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.	
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	N.D.	
2-Methylphenol	95-48-7	100	N.D.	
4-Methylphenol	106-44-5	100	N.D.	
Naphthalene	91-20-3	100	N.D.	
2-Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	320	<---
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	290	<---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Water

Laboratory Reference #: MWI 8785

Sampled : 04-01-97
Received: 04-01-97
Analyzed: 04-02-97
Reported: 04-16-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/l)*
97040030	Equip Blank	N.D.
97040031	Rinstate Blank	N.D.
Detection Limit:		0.5

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040030
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-03-97
Reported: 04-15-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040031
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-03-97
Reported: 04-15-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Equip Blank
Laboratory Sample Number: 97040030
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-10-97
Reported: 04-15-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040030

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*95*
Toluene-d8	*90*
4-Bromofluorobenzene	*108*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample Number: 97040031
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-10-97
Reported: 04-15-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040031

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*109*
Toluene-d8	*91*
4-Bromofluorobenzene	*112*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Trip Blank
Laboratory Sample Number: 97040032
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-12-97
Reported: 04-15-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Trip Blank

Laboratory Sample Number: 97040032

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*101*
Toluene-d8	*93*
4-Bromofluorobenzene	*107*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040030
Laboratory Reference #: MWI 8785

Sampled : 04-01-97
Received: 04-01-97
Analyzed: 04-04-97
Reported: 04-15-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040030

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
4-Methylphenol	106-44-5	5.0	N.D.
hthalene	91-20-3	50	N.D.
Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

k Noorani
boratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040031
Laboratory Reference #: MWI 8785

Sampled : 04-01-97
Received: 04-01-97
Analyzed: 04-04-97
Reported: 04-15-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040031

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
Methylphenol	106-44-5	5.0	N.D.
ıthalene	91-20-3	50	N.D.
Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

. Noorani
ıboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040030
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-09-97
Reported: 04-15-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040031
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-09-97
Reported: 04-15-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040030
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-09-97
Reported: 04-15-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040031
Laboratory Reference #: MWI 8785

Sampled: 04-01-97
Received: 04-01-97
Analyzed: 04-09-97
Reported: 04-15-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Organochlorine Pesticides (EPA 8080)

Date of Analysis : 04/08/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8785

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
4,4'-DDT	0.0	1.00	0.73	0.69	73	69	6

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : PCB 'S (EPA 8080)

Date of Analysis : 04/08/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8785

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
PCB-1260	0.0	50	45	41	90	82	9

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/08/97
Laboratory Sample No : 97040012
Laboratory Reference No : MWI 8785

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	50	50	100	100	0
1,1-Dichloroethene	0.0	50	49	44	98	88	11
Trichloroethene	0.0	50	50	49	100	98	2
Toluene	0.0	50	52	52	104	104	0
Chlorobenzene	0.0	50	49	50	98	100	2

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/09/97
Laboratory Sample No : 97040025
Laboratory Reference No : MWI 8785

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	51	51	102	102	0
1,1-Dichloroethene	0.0	50	46	46	92	92	0
Trichloroethene	0.0	50	52	52	104	104	0
Toluene	0.0	50	54	52	108	104	4
Chlorobenzene	0.0	50	52	51	104	102	2

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/11/97
Laboratory Sample No : 97031118
Laboratory Reference No : MWI 8785

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	20	17	19	85	95	11
1,1-Dichloroethene	0.0	20	15	17	75	85	13
Trichloroethene	0.0	20	17	18	85	90	6
Toluene	0.0	20	18	20.5	90	103	13
Chlorobenzene	0.0	20	18	19	90	95	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/04/97
Laboratory Sample No : 97040013
Laboratory Reference No : MWI 8785

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	52	45	104	90	14
n-Nitroso-di-n-propylamine	0.0	50	54	49	108	98	10
1,2,4-Trichlorobenzene	0.0	50	53	45	106	90	16
Acenaphthene	0.0	50	57	48	114	96	17
Pyrene	0.0	50	56	47	112	94	17
Pentachlorophenol	0.0	100	94	86	94	86	9
4-Chloro-3-Methylphenol	0.0	100	83	76	83	76	9
2-Chlorophenol	0.0	100	91	82	91	82	10
Phenol	0.0	100	84	76	84	76	10

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/05/97
Laboratory Sample No : 97040015
Laboratory Reference No : MWI 8785

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	42	39	84	78	7
n-Nitroso-di-n-propylamine	0.0	50	35	36	70	72	3
1,2,4-Trichlorobenzene	0.0	50	43	40	86	80	7
Acenaphthene	0.0	50	46	43	92	86	7
Pyrene	0.0	50	44	40	88	80	10
Pentachlorophenol	0.0	100	79	79	79	79	0
4-Chloro-3-Methylphenol	0.0	100	70	69	70	69	1
2-Chlorophenol	0.0	100	75	73	75	73	3
Phenol	0.0	100	69	67	69	67	3

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/04/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8785

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	41	39	82	78	5
n-Nitroso-di-n-propylamine	0.0	50	45	42	90	84	7
1,2,4-Trichlorobenzene	0.0	50	40	40	80	80	0
Acenaphthene	0.0	50	47	45	94	90	4
Pyrene	0.0	50	44	45	88	90	2
Pentachlorophenol	0.0	100	82	83	82	83	1
4-Chloro-3-Methylphenol	0.0	100	70	69	70	69	1
2-Chlorophenol	0.0	100	81	79	81	79	3
Phenol	0.0	100	31	31	31	31	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/02/97
Laboratory Sample No : 97031103
Laboratory Reference No : MWI 8785

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	100	88	89	88	89	1

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/02/97
Laboratory Sample No : 97040024
Laboratory Reference No : MWI 8785

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	100	79	76	79	76	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/02/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8785

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	2.5	2.3	2.2	92	88	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/03/97
Laboratory Sample No : 97040030
Laboratory Reference No : MWI 8785

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	3.000	3.400	3.300	113	110	3
Arsenic	0.00	1.000	1.190	1.110	119	111	7
Barium	0.00	0.100	0.086	0.096	86	96	11
Beryllium	0.00	0.100	0.088	0.102	88	102	15
Cadmium	0.00	0.100	0.104	0.098	104	98	6
Chromium (VI)	0.00	0.500	0.440	0.470	88	94	7
Chromium (Total)	0.00	0.100	0.095	0.104	95	104	9
Cobalt	0.00	0.100	0.120	0.112	120	112	7
Copper	0.00	0.100	0.099	0.098	99	98	1
Lead	0.00	1.000	0.890	1.030	89	103	15
Mercury	0.000	0.020	0.022	0.020	110	100	10
Molybdenum	0.00	1.000	1.140	1.180	114	118	3
Nickel	0.00	0.500	0.490	0.520	98	104	6
Selenium	0.00	1.000	1.090	1.050	109	105	4
Silver	0.00	0.500	0.510	0.510	102	102	0
Thallium	0.00	3.000	2.900	3.100	97	103	7
Vanadium	0.00	0.500	0.570	0.500	114	100	13
Zinc	0.00	0.100	0.103	0.110	103	110	7

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/02/97
Laboratory Sample No : 97031116
Laboratory Reference No : MWI 8785

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.0	30.0	30.7	30.1	102	100	2
Arsenic	0.0	10.0	10.0	9.2	100	92	8
Barium	10.0	10.0	21.8	20.9	118	109	4
Beryllium	0.00	1.00	0.99	1.05	99	105	6
Cadmium	0.00	1.00	0.99	1.00	99	100	1
Chromium (Total)	2.2	5.0	7.6	7.3	108	102	4
Chromium (VI)	0.00	1.00	1.05	0.92	105	92	13
Cobalt	0.83	1.00	1.80	1.70	97	87	6
Copper	2.1	5.0	7.5	7.4	108	106	1
Lead	0.00	10.00	9.40	9.70	94	97	3
Mercury	0.000	0.020	0.022	0.019	110	95	15
Molybdenum	0.0	10.0	9.1	9.2	91	92	1
Nickel	1.1	5.0	5.6	5.4	90	86	4
Selenium	0.0	10.0	9.9	9.3	99	93	6
Silver	0.0	5.0	4.3	4.3	86	86	0
Thallium	0.0	30.0	25.4	27.0	85	90	6
Vanadium	2.5	5.0	7.4	7.2	98	94	3
Zinc	4.2	5.0	8.7	8.7	90	90	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/02/97
Laboratory Sample No : 97040028
Laboratory Reference No : MWI 8785

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.0	30.0	30.2	30.7	101	102	2
Arsenic	0.0	10.0	8.2	8.4	82	84	2
Barium	9.9	10.0	18.7	19.5	88	96	4
Beryllium	0.00	1.00	1.07	1.02	107	102	5
Cadmium	0.00	1.00	0.98	0.94	98	94	4
Chromium (Total)	2.7	5.0	7.5	7.7	96	100	3
Chromium (VI)	0.00	1.00	0.89	0.98	89	98	10
Cobalt	0.67	1.00	1.60	1.57	93	90	2
Copper	1.2	1.0	2.2	2.3	100	110	4
Lead	0.00	10.00	9.00	8.90	90	89	1
Mercury	0.000	0.020	0.018	0.020	90	100	11
Molybdenum	0.0	10.0	9.5	9.6	95	96	1
Nickel	0.96	5.0	5.6	5.3	93	87	6
Selenium	0.0	10.0	10.2	9.3	102	93	9
Silver	0.0	5.0	4.9	4.9	98	98	0
Thallium	0.0	30.0	26.8	25.9	89	86	3
Vanadium	2.8	5.0	7.9	7.6	102	96	4
Zinc	4.2	5.0	8.4	8.8	84	92	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532
Tustin, CA 92680
(714) 832-0064, Fax (714) 832-0067

Analysis Request and Chain of Custody Record

Lab Job No: _____
Page 1 of 2

24hr TAT on soil TRPH only
REQUIRED TAT: Standard TAT otherwise.

CUSTOMER INFORMATION	PROJECT INFORMATION
COMPANY: Montgomery Watson.	PROJECT NAME: McDonnell Douglas C-6 facility
SEND REPORT TO: Fred Strauss	NUMBER: 1206035.01090010
ADDRESS: 250 N. Madison Ave. Pasadena, CA 91101	LOCATION: Bldg 37 Area
	ADDRESS: 19503 S. Normandie Ave Torrance, CA.
PHONE: (818) 568-6582 FAX: (818) 796-5941	SAMPLED BY: JAH / IV

ANALYSIS/METHOD REQUEST

SAMPLE ID	NO. OF CONTAINERS	SAMPLE DATE	SAMPLE TIME	SAMPLE MATRIX	CONTAINER TYPE	PRES.	TRPH 418.1	Title 22 Metals	VOCs 8260	SVOCs 8270	PCBs 8080	Pesticides 8080			REMARKS/PRECAUTIONS
B37-GS-117-5'	2	4/01/97	8:22	Soil	2"x6" SS sleeve	Ice	X	X							
B37-GS-118-6.5'	2	4/01/97	9:50	Soil	2"x6" SS sleeve		X	X							
B37-GS-119-6.5'			10:37				X	X							
B37-GS-120-6.5'			1040				X	X							
B37-GS-121-6.5'			1045				X	X							
B37-GS-122-6.5'			1109				X	X							
B37-GS-123-6.5'			1116				X	X							
B37-GS-124-6.5'			1124				X	X							
B37-GS-125-6.5'			1125				X	X							
B37-GS-126-6.5'			1136				X	X							
B37-GS-127-6.5'			1141				X	X							
B37-GS-128-6.5'			1148				X	X							
B37-GS-129-7'			1322				X	X							
B37-GS-130-9'			1330				X	X							

Total No. of Samples: 14 (this page)

Method of Shipment: Orange Coast Courier

Relinquished By:	Date/Time:	Received By:	Date/Time:
Ivan Vargas	4/1/97 - 16:52		
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received For Lab By: [signature]	Date/Time: 04-01-97 4:53PM

Reporting Format: (check)
NORMAL _____ S.D. HMMD _____
RWQCB _____ OTHER _____

Sample Integrity: (check)
intact _____ on ice _____

All samples remain the property of the client who is responsible f[illegible]sal. A disposal fee may be imposed if client fails to pickup samples.

BOE-C6-0143635

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532
Tustin, CA 92680
(714) 832-0064, Fax (714) 832-0067

A Request and Chai. of Custody Record

Lab Job No 1
Page 2 of 2

24hr TAT ON SOIL TRPH;
REQUIRED TAT: Standard TAT otherwise.

CUSTOMER INFORMATION		PROJECT INFORMATION	
COMPANY:	Montgomery Watson	PROJECT NAME:	McDonnell Douglas
SEND REPORT TO:	Fred Strauss	NUMBER:	120603V-01090010
ADDRESS:	250 N Madison Ave. Pasadena, CA 91101	LOCATION:	C-6 facility - Bldg 37
		ADDRESS:	19503 S Normandie Ave, Torrance
PHONE:	(818) 568-6582 FAX: (818) 796-5941	SAMPLED BY:	JM/IV

ANALYSIS/METHOD REQUEST

SAMPLE ID	NO. OF CONTAINERS	SAMPLE DATE	SAMPLE TIME	SAMPLE MATRIX	CONTAINER TYPE	PRES.	TRPH 418.1	Title 22 metals	VOCs 8260	SVOCs 8270	PCBs 8080	Pesticides 8080			REMARKS/PRECAUTIONS
B37-GS-131-9'	2	4/1/97	1336	Soil	2"x6" Sleeve	ICE	X	X	-						
B37-RE4-SP37	↓	↓	1418	↓	↓	↓	X	X	X	X					
B37-RE4-SP38	↓	↓	1430	↓	↓	↓	X	X	X	X					
B37-RE4-SP39	↓	↓	1445	↓	↓	↓	X	X	X	X					
Equipment Blank	7	4/1/97	1530	Water	4 Amber Poly 2 Vols	Hno; HCl	X	X	X	X	X	X			
Rinsate Blank	7	↓	1545	↓	↓	↓	X	X	X	X	X	X			
Trip Blank	2	↓	—	↓	2 Vols	HCl			X						

Total No. of Samples: 7 (this page) - 21 total shipment

Method of Shipment: Orange Coast Courier (2 Ice chest).

Relinquished By:	Date/Time:	Received By:	Date/Time:
Ivan Vargas	4/1/97 - 16:52		
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received For Lab By: [signature]	Date/Time: 04-01-97 4:53PM

Reporting Format: (check)
NORMAL _____ S.D. HMMD _____
RWQCB _____ OTHER _____

Sample Integrity: (check)
Intact _____ on Ice _____

All samples remain the property of the client who is responsible for disposal. A disposal fee may be imposed if client fails to pickup samples.

BOE-C6-0143636

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Soil

Laboratory Reference #: MWI 8787

Sampled : 04-02-97
Received: 04-02-97
Analyzed: 04-03-97
Reported: 04-03-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/kg)*
97040059	B37-RE4-SP40	2,400
Detection Limit:		8.0

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP40

Laboratory Sample Number: 97040059
Laboratory Reference #: MWI 8787

Sampled: 04-02-97
Received: 04-02-97
Analyzed: 04-03-97
Reported: 04-14-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	130	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	27	<---
Cobalt	6010	80	8000	0.5	8.2	<---
Copper	6010	25	2500	0.1	15	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	12	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	32	<---
Zinc	6010	250	5000	0.1	52	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

'ontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP40
Laboratory Sample Number: 97040059
Laboratory Reference #: MWI 8787

Sampled: 04-02-97
Received: 04-02-97
Analyzed: 04-07-97
Reported: 04-14-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/kg)***	***SAMPLE RESULT (ug/kg)***	
Benzene	71-43-2	2.5	N.D.	
Bromodichloromethane	75-27-4	2.5	N.D.	
Bromoform	75-25-2	2.5	N.D.	
Bromomethane	74-83-9	2.5	N.D.	
Carbon Disulfide	75-15-0	5.0	N.D.	
Carbon tetrachloride	56-23-5	2.5	N.D.	
Chlorobenzene	108-90-7	2.5	N.D.	
Chlorodibromomethane	124-48-1	2.5	N.D.	
Chloroethane	75-00-3	2.5	N.D.	
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.	
Chloroform	67-66-3	2.5	N.D.	
Chloromethane	74-87-3	2.5	N.D.	
1,1-Dichloroethane	75-35-3	2.5	N.D.	
1,2-Dichloroethane	107-06-2	2.5	N.D.	
1,1-Dichloroethene	75-35-4	2.5	N.D.	
'rans 1,2-Dichloroethene	156-60-5	2.5	N.D.	
,2-Dichloropropane	78-87-5	2.5	N.D.	
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.	
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.	
Ethylbenzene	100-41-4	2.5	14	<---
Methylene chloride	75-09-2	5.0	N.D.	
Styrene	100-42-5	2.5	N.D.	
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.	
Tetrachloroethene	127-18-4	2.5	N.D.	
Toluene	108-88-3	2.5	N.D.	
1,1,1-Trichloroethane	71-55-6	2.5	N.D.	
1,1,2-Trichloroethane	79-00-5	2.5	N.D.	
Trichloroethene	79-01-6	2.5	N.D.	
Trichlorofluoromethane	75-69-4	5.0	N.D.	
Vinyl acetate	108-05-4	5.0	N.D.	
Vinyl chloride	75-01-4	2.5	N.D.	
Total Xylenes	1330-20-7	2.5	34	<---
Dichlorofluoromethane	75-71-8	2.5	N.D.	
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.	
2,2-Dichloropropane	590-20-7	2.5	N.D.	
Bromochloromethane	74-97-5	2.5	N.D.	
1,1-Dichloropropene	563-58-6	2.5	N.D.	
1,2-Dichloropropane	78-87-5	2.5	N.D.	
Dibromomethane	74-95-3	2.5	N.D.	
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.	
1,1,2-Trichloroethane	79-00-5	2.5	N.D.	
1,2-Dibromoethane	106-93-4	2.5	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP40

Laboratory Sample Number: 97040059

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*	
1,3-Dichloropropane	142-28-9	2.5	N.D.	
Dibromochloromethane	124-48-1	2.5	N.D.	
Isopropylbenzene	98-82-8	2.5	13	<---
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.	
1,2,3-Trichloropropane	96-18-4	2.5	N.D.	
Bromobenzene	108-86-1	2.5	N.D.	
n-Propylbenzene	103-65-1	2.5	23	<---
2-Chlorotoluene	95-49-8	2.5	N.D.	
1,3,5-Trimethylbenzene	108-67-8	2.5	28	<---
4-Chlorotoluene	106-43-4	2.5	N.D.	
tert-Butylbenzene	98-06-6	2.5	14	<---
1,2,4-Trimethylbenzene	95-63-6	2.5	130	<---
sec-Butylbenzene	135-98-8	2.5	20	<---
4-Isopropyltoluene	99-87-6	2.5	19	<---
1,3-Dichlorobenzene	541-73-1	2.5	N.D.	
1,4-Dichlorobenzene	106-46-7	2.5	N.D.	
n-Butylbenzene	104-51-8	2.5	19	<---
1,2-Dichlorobenzene	95-50-1	2.5	N.D.	
1-2-Dibromo-2-CPA	96-12-8	5.0	N.D.	
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.	
Hexachlorobutadiene	87-68-3	2.5	N.D.	
Naphthalene	91-20-3	2.5	210	<---
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*115*
Toluene-d8	*91*
4-Bromofluorobenbzene	*90*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
N: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP40
Laboratory Sample #: 97040059
Laboratory Reference #: MWI 8787

Sampled : 04-02-97
Received: 04-02-97
Analyzed: 04-14-97
Reported: 04-14-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
Benzo(k)fluoranthene	207-08-9	250	N.D.
nzo(g,h,i)perylene	191-24-2	250	N.D.
zo(a)pyrene	50-32-8	250	N.D.
nzyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
methyl phthalate	131-11-3	100	N.D.
-Dinitro-2-methylphenol	534-52-1	100	N.D.
-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP40
Laboratory Sample #: 97040059

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	N.D.	
Fluorene	86-73-7	100	130	<---
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.	
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	480	<---
2-Methylphenol	95-48-7	100	N.D.	
4-Methylphenol	106-44-5	100	N.D.	
Naphthalene	91-20-3	100	130	<--
-Nitroaniline	88-74-4	250	N.D.	
-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	300	<---
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	N.D.	
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP40

Laboratory Sample Number: 97040059
Laboratory Reference #: MWI 8787

Sampled: 04-02-97
Received: 04-02-97
Analyzed: 04-09-97
Reported: 04-14-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULTS (ug/kg)*
PCB-1016	12674-11-2	20	N.D.
PCB-1221	11104-28-2	20	N.D.
PCB-1232	11141-16-5	20	N.D.
PCB-1242	53469-21-9	20	N.D.
PCB-1248	12672-29-6	20	N.D.
PCB-1254	11097-69-1	20	N.D.
PCB-1260	11096-82-5	20	170

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Water

Laboratory Reference #: MWI 8787

Sampled : 04-02-97
Received: 04-02-97
Analyzed: 04-04-97
Reported: 04-14-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/l)*
97040060	Rinsate Blank	N.D.
97040061	Equip Blank	N.D.
Detection Limit:		0.5

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

'ontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040060
Laboratory Reference #: MWI 8787

Sampled: 04-02-97
Received: 04-02-97
Analyzed: 04-03-97
Reported: 04-14-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040061
Laboratory Reference #: MWI 8787

Sampled: 04-02-97
Received: 04-02-97
Analyzed: 04-03-97
Reported: 04-14-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample Number: 97040060
Laboratory Reference #: MWI 8787

Sampled: 04-02-97
Received: 04-02-97
Analyzed: 04-10-97
Reported: 04-14-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040060

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*104*
Toluene-d8	*92*
4-Bromofluorobenzene	*109*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Equip Blank
Laboratory Sample Number: 97040061
Laboratory Reference #: MWI 8787

Sampled: 04-02-97
Received: 04-02-97
Analyzed: 04-10-97
Reported: 04-14-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040061

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*101*
Toluene-d8	*94*
4-Bromofluorobenzene	*111*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Trip Blank
Laboratory Sample Number: 97040062
Laboratory Reference #: MWI 8787

Sampled: 04-02-97
Received: 04-02-97
Analyzed: 04-10-97
Reported: 04-14-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/l)***	***SAMPLE RESULT (ug/l)***
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Trip Blank

Laboratory Sample Number: 97040062

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*108*
Toluene-d8	*92*
4-Bromofluorobenzene	*109*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
.TN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sampled : 04-02-97
Sample Description: Water, Rinsate Blank *Received:* 04-02-97
Laboratory Sample #: 97040060 *Analyzed:* 04-04-97
Laboratory Reference #: MWI 8787 *Reported:* 04-14-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
enzo(g,h,i)perylene	191-24-2	25	N.D.
zo(a)pyrene	50-32-8	25	N.D.
nzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
imethyl phthalate	131-11-3	5.0	N.D.
6-Dinitro-2-methylphenol	534-52-1	50	N.D.
-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040060

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
4-Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
.TN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040061
Laboratory Reference #: MWI 8787

Sampled : 04-02-97
Received: 04-02-97
Analyzed: 04-04-97
Reported: 04-14-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
enzo(k)fluoranthene	207-08-9	25	N.D.
enzo(g,h,i)perylene	191-24-2	25	N.D.
zo(a)pyrene	50-32-8	25	N.D.
enzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
imethyl phthalate	131-11-3	5.0	N.D.
-Dinitro-2-methylphenol	534-52-1	50	N.D.
.-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040061

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
4-Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040597
Laboratory Reference #: MWI 8895

Sampled : 04-11-97
Received: 04-11-97
Analyzed: 04-16-97
Reported: 04-25-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	CAS NUMBER	DETECTION LIMIT ug/l	SAMPLE RESULTS ug/l
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
[illegible]zo(b)fluoranthene	205-99-2	25	N.D.
[illegible](k)fluoranthene	207-08-9	25	N.D.
[illegible](g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
[illegible] Dimethylphenol	105-67-9	5.0	N.D.
[illegible]hyl phthalate	131-11-3	5.0	N.D.
[illegible]-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040597

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
4-Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040596
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-16-97
Reported: 04-25-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040597
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-16-97
Reported: 04-25-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/22/97
Laboratory Sample No : 97040691
Laboratory Reference No : MWI 8895

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	46	46	92	92	0
n-Nitroso-di-n-propylamine	0.0	50	49	47	98	94	4
1,2,4-Trichlorobenzene	0.0	50	48	47	96	94	2
Acenaphthene	0.0	50	52	51	104	102	2
[illegible]rene	0.0	50	46	35	92	70	27
Pentachlorophenol	0.0	100	77	78	77	78	1
4-Chloro-3-Methylphenol	0.0	100	73	70	73	70	4
2-Chlorophenol	0.0	100	80	79	80	79	1
Phenol	0.0	100	70	70	70	70	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/16/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8895

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	38	44	76	88	15
n-Nitroso-di-n-propylamine	0.0	50	47	34	94	68	32
1,2,4-Trichlorobenzene	0.0	50	38	44	76	88	15
Acenaphthene	0.0	50	47	50	94	100	6
Pyrene	0.0	50	43	46	86	92	7
Pentachlorophenol	0.0	100	78	74	78	74	5
4-Chloro-3-Methylphenol	0.0	100	60	62	60	62	3
2-Chlorophenol	0.0	100	79	81	79	81	3
Phenol	0.0	100	22	24	22	24	9

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/22/97
Laboratory Sample No : 97040609
Laboratory Reference No : MWI 8895

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	20	18	17	90	85	6
1,1-Dichloroethene	0.0	20	17	17	85	85	0
Trichloroethene	0.0	20	19	19	95	95	0
Toluene	0.0	20	18	18	90	90	0
Chlorobenzene	0.0	20	16	17	80	85	6

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/21/97
Laboratory Sample No : 97040717
Laboratory Reference No : MWI 8895

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	56	51	112	102	9
1,1-Dichloroethene	0.0	50	42	39	84	78	7
Trichloroethene	0.0	50	65	62	130	124	5
Toluene	0.0	50	59	56	118	112	5
Chlorobenzene	0.0	50	53	51	106	102	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040596
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-16-97
Reported: 04-25-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040597
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-16-97
Reported: 04-25-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040060
Laboratory Reference #: MWI 8787

Sampled: 04-02-97
Received: 04-02-97
Analyzed: 04-09-97
Reported: 04-14-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040061
Laboratory Reference #: MWI 8787

Sampled: 04-02-97
Received: 04-02-97
Analyzed: 04-09-97
Reported: 04-14-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040060
Laboratory Reference #: MWI 8787

Sampled: 04-02-97
Received: 04-02-97
Analyzed: 04-09-97
Reported: 04-14-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040061
Laboratory Reference #: MWI 8787

Sampled: 04-02-97
Received: 04-02-97
Analyzed: 04-09-97
Reported: 04-14-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : PCB 'S (EPA 8080)

Date of Analysis : 04/08/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8787

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
PCB-1260	0.0	50	45	41	90	82	9

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Organochlorine Pesticides (EPA 8080)

Date of Analysis : 04/08/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8787

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
4,4'-DDT	0.0	1.00	0.73	0.69	73	69	6

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : PCB 'S (EPA 8080)

Date of Analysis : 04/09/97
Laboratory Sample No : 97040067
Laboratory Reference No : MWI 8787

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
PCB-1260	0.0	500	230	240	46	48	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/03/97
Laboratory Sample No : 97040030
Laboratory Reference No : MWI 8787

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	3.000	3.400	3.300	113	110	3
Arsenic	0.00	1.000	1.190	1.110	119	111	7
Barium	0.00	0.100	0.086	0.096	86	96	11
Beryllium	0.00	0.100	0.088	0.102	88	102	15
Cadmium	0.00	0.100	0.104	0.098	104	98	6
Chromium (VI)	0.00	0.500	0.440	0.470	88	94	7
Chromium (Total)	0.00	0.100	0.095	0.104	95	104	9
Cobalt	0.00	0.100	0.120	0.112	120	112	7
Copper	0.00	0.100	0.099	0.098	99	98	1
Lead	0.00	1.000	0.890	1.030	89	103	15
Mercury	0.000	0.020	0.022	0.020	110	100	10
Molybdenum	0.00	1.000	1.140	1.180	114	118	3
Nickel	0.00	0.500	0.490	0.520	98	104	6
Selenium	0.00	1.000	1.090	1.050	109	105	4
Silver	0.00	0.500	0.510	0.510	102	102	0
Thallium	0.00	3.000	2.900	3.100	97	103	7
Vanadium	0.00	0.500	0.570	0.500	114	100	13
Zinc	0.00	0.100	0.103	0.110	103	110	7

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/03/97
Laboratory Sample No : 97040045
Laboratory Reference No : MWI 8787

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	30.00	27.00	25.00	90	83	8
Arsenic	9.00	10.00	19.40	20.40	104	114	5
Barium	11.00	10.00	19.00	19.00	80	80	0
Beryllium	0.00	1.00	0.98	0.96	98	96	2
Cadmium	0.00	1.00	1.12	1.08	112	108	4
Chromium (Total)	2.10	5.00	6.50	6.80	88	94	5
Chromium (VI)	0.00	1.00	0.85	0.98	85	98	14
Cobalt	1.00	1.00	1.99	1.87	99	87	6
Copper	1.10	1.00	2.20	2.05	110	95	7
Lead	0.00	10.00	9.70	10.00	97	100	3
Mercury	0.000	0.020	0.021	0.018	105	90	15
Molybdenum	0.00	10.00	9.50	9.90	95	99	4
Nickel	1.00	5.00	5.10	5.50	82	90	8
Selenium	0.00	10.00	11.30	11.00	113	110	3
Silver	0.00	5.00	5.20	5.40	104	108	4
Thallium	0.00	30.00	29.00	27.00	97	90	7
Vanadium	2.60	5.00	7.30	6.90	94	86	6
Zinc	3.70	5.00	8.30	8.70	92	100	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/04/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8787

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	41	39	82	78	5
n-Nitroso-di-n-propylamine	0.0	50	45	42	90	84	7
1,2,4-Trichlorobenzene	0.0	50	40	40	80	80	0
Acenaphthene	0.0	50	47	45	94	90	4
Pyrene	0.0	50	44	45	88	90	2
Pentachlorophenol	0.0	100	82	83	82	83	1
4-Chloro-3-Methylphenol	0.0	100	70	69	70	69	1
2-Chlorophenol	0.0	100	81	79	81	79	3
Phenol	0.0	100	31	31	31	31	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/14/97
Laboratory Sample No : 97040418
Laboratory Reference No : MWI 8787

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	45	45	90	90	0
n-Nitroso-di-n-propylamine	0.0	50	45	38	90	76	17
1,2,4-Trichlorobenzene	0.0	50	43	45	86	90	5
Acenaphthene	0.0	50	51	49	102	98	4
Pyrene	0.0	50	47	46	94	92	2
Pentachlorophenol	0.0	100	83	87	83	87	5
4-Chloro-3-Methylphenol	0.0	100	75	71	75	71	5
2-Chlorophenol	0.0	100	84	83	84	83	1
Phenol	0.0	100	77	74	77	74	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/07/97
Laboratory Sample No : 97041110
Laboratory Reference No : MWI 8787

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	54	52	108	104	4
1,1-Dichloroethene	0.0	50	53	52	106	104	2
Trichloroethene	0.0	50	54	52	108	104	4
Toluene	0.0	50	55	54	110	108	2
Chlorobenzene	0.0	50	48	50	96	100	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/03/97
Laboratory Sample No : 97040017
Laboratory Reference No : MWI 8787

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	100	84	81	84	81	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/04/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8787

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	2.5	2.0	2.0	80	80	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/11/97
Laboratory Sample No : 97031118
Laboratory Reference No : MWI 8787

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	20	17	19	85	95	11
1,1-Dichloroethene	0.0	20	15	17	75	85	13
Trichloroethene	0.0	20	17	18	85	90	6
Toluene	0.0	20	18	20.5	90	103	13
Chlorobenzene	0.0	20	18	19	90	95	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532
Tustin, CA 92680
(714) 832-0064, Fax (714) 832-0067

Analysis Request and Chain of Custody Record

Lab Job No: _____
Page 1 of 1

REQUIRED TAT: 24 hr. TAT on Soil TRPH only Standard TAT Otherwise.

CUSTOMER INFORMATION

COMPANY: Montgomery Watson
SEND REPORT TO: Fred Strauss
ADDRESS: 250 N. Madison Ave. Pasadena, CA 91101
PHONE: (818) 568-6582 FAX: (818) 796

PROJECT INFORMATION

PROJECT NAME: Ma Donnell Douglas C-6 Facility
NUMBER: 12.06035.01090010
LOCATION: Bldg. 37 Area
ADDRESS: 19503 S. Normandie Ave. Torrance, CA
SAMPLED BY: AWN/IV

ANALYSIS/METHOD REQUEST

SAMPLE ID	NO. OF CONTAINERS	SAMPLE DATE	SAMPLE TIME	SAMPLE MATRIX	CONTAINER TYPE	PRES.	TRPH 418.1	Title 22 Metals	VOCs (8260)	SVOCs (8270)	PCBs (8080)	Pesticides (8080)			REMARKS/PRECAUTIONS
B37-RE4-SP40	2	4/2/97	1100	Soil	2"x6" SS sleeve	ICE	X	X	X	X					
Rinsate Blank	7	↓	1230	Water	Amber 1 poly 2 VOAs	HNO3 HCl	X	X	X	X	X	X			
Equipment Blank	7	↓	1700	↓	↓	↓	X	X	X	X	X	X			
Trip Blank	2	↓	—	↓	2 VOAs	HCl			X						

Total No. of Samples: 4

Method of Shipment: Via Orange Coast Courier (1 cooler)

Relinquished By:	Date/Time:	Received By:	Date/Time:
Adam W. Venis	4/2/97 1705		
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received For Lab By:	Date/Time: 4/2/97 17:05

Reporting Format: (check)
NORMAL _____ S.D. HMMD _____
RWQCB _____ OTHER _____

Sample Integrity: (check)
intact _____ on ice _____

All samples remain the property of the client who is responsible . sposal. A disposal fee may be imposed if client fails to pickup samples.

BOE-C6-0143682

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Soil

Laboratory Reference #: MWI 8883

Sampled : 04-09-97
Received: 04-09-97
Analyzed: 04-10-97
Reported: 04-10-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/kg)*
97040411	B37-RE4-SP41	360
97040412	B37-RE4-SP42	120
Detection Limit:		8.0

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP41

Laboratory Sample Number: 97040411
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-11-97
Reported: 04-24-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	130	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	41	<---
Cobalt	6010	80	8000	0.5	9.1	<---
Copper	6010	25	2500	0.1	71	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	17	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	35	<---
Zinc	6010	250	5000	0.1	120	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP42

Laboratory Sample Number: 97040412
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-11-97
Reported: 04-24-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	120	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	32	<---
Cobalt	6010	80	8000	0.5	8.7	<---
Copper	6010	25	2500	0.1	16	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	13	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	35	<---
Zinc	6010	250	5000	0.1	57	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

ark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP41
Laboratory Sample Number: 97040411
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-21-97
Reported: 04-24-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
1,1-Dichloroethene	75-35-4	2.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP41

Laboratory Sample Number: 97040411

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
n-Butylbenzene	104-51-8	2.5	N.D.
1,2-Dichlorobenzene	95-50-1	2.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*119*
Toluene-d8	*87*
4-Bromofluorobenbzene	*105*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP42
Laboratory Sample Number: 97040412
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-21-97
Reported: 04-24-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
1,1-Dichloroethene	75-35-4	2.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP42

Laboratory Sample Number: 97040412

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
4-Dichlorobenzene	106-46-7	2.5	N.D.
Butylbenzene	104-51-8	2.5	N.D.
1,2-Dichlorobenzene	95-50-1	2.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*111*
Toluene-d8	*92*
4-Bromofluorobenbzene	*109*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP41
Laboratory Sample #: 97040411
Laboratory Reference #: MWI 8883

Sampled : 04-09-97
Received: 04-09-97
Analyzed: 04-14-97
Reported: 04-24-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	490 <---
Benzo(b)fluoranthene	205-99-2	250	1,900 <---
Benzo(k)fluoranthene	207-08-9	250	490 <---
Benzo(g,h,i)perylene	191-24-2	250	1,100 <---
Benzo(a)pyrene	50-32-8	250	950 <---
Benzyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	2,000 <---
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
Dimethyl phthalate	131-11-3	100	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.
2,4-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040414
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-16-97
Reported: 04-24-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040413
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-16-97
Reported: 04-24-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Rinstate Blank
Laboratory Sample #: 97040414

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
ethylphenol	95-48-7	5.0	N.D.
Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040414
Laboratory Reference #: MWI 8883

Sampled : 04-09-97
Received: 04-09-97
Analyzed: 04-16-97
Reported: 04-24-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040413

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
Methylphenol	95-48-7	5.0	N.D.
Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040413
Laboratory Reference #: MWI 8883

Sampled : 04-09-97
Received: 04-09-97
Analyzed: 04-16-97
Reported: 04-24-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Trip Blank

Laboratory Sample Number: 97040415

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*113*
Toluene-d8	*98*
4-Bromofluorobenzene	*98*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Trip Blank
Laboratory Sample Number: 97040415
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-22-97
Reported: 04-24-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040414

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*117*
Toluene-d8	*101*
4-Bromofluorobenzene	*109*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample Number: 97040414
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-22-97
Reported: 04-24-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040413

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*115*
Toluene-d8	*101*
4-Bromofluorobenzene	*105*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Equip Blank
Laboratory Sample Number: 97040413
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-22-97
Reported: 04-24-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040414
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-11-97
Reported: 04-24-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Mark Noorani
aboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040413
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-11-97
Reported: 04-24-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Water

Laboratory Reference #: MWI 8883

Sampled : 04-09-97
Received: 04-09-97
Analyzed: 04-10-97
Reported: 04-24-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/l)*
97040413	Equip Blank	N.D.
97040414	Rinsate Blank	N.D.
Detection Limit:		0.5

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP42
Laboratory Sample #: 97040412

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
2,4-Dinitrotoluene	121-14-2	250	N.D.
2,6-Dinitrotoluene	606-20-2	250	N.D.
Di-n-octyl phthalate	117-84-0	250	N.D.
Fluoranthene	206-44-0	100	N.D.
Fluorene	86-73-7	100	N.D.
Hexachlorobenzene	118-74-1	100	N.D.
Hexachlorobutadiene	87-68-3	100	N.D.
Hexachlorocyclopentadiene	77-47-4	100	N.D.
Hexachloroethane	67-72-1	100	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.
Isophorone	78-59-1	100	N.D.
2-Methylnaphthalene	91-57-6	100	N.D.
2-Methylphenol	95-48-7	100	N.D.
4-Methylphenol	106-44-5	100	N.D.
Naphthalene	91-20-3	100	N.D.
2-Nitroaniline	88-74-4	250	N.D.
3-Nitroaniline	99-09-2	250	N.D.
4-Nitroaniline	100-01-6	250	N.D.
Nitrobenzene	98-95-3	100	N.D.
2-Nitrophenol	88-75-5	100	N.D.
4-Nitrophenol	100-02-7	100	N.D.
N-Nitrosodiphenylamine	86-30-6	100	N.D.
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.
N-Nitrosodimethylamine	62-75-9	100	N.D.
Pentachlorophenol	87-86-5	250	N.D.
Phenanthrene	85-01-8	100	N.D.
Phenol	108-95-2	100	N.D.
Pyrene	129-00-0	100	N.D.
1,2,4-Trichlorobenzene	120-82-1	100	N.D.
2,4,5-Trichlorophenol	95-95-4	100	N.D.
2,4,6-Trichlorophenol	88-06-2	100	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP42
Laboratory Sample #: 97040412
Laboratory Reference #: MWI 8883

Sampled : 04-09-97
Received: 04-09-97
Analyzed: 04-14-97
Reported: 04-24-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
ızo(k)fluoranthene	207-08-9	250	N.D.
ızo(g,h,i)perylene	191-24-2	250	N.D.
Benzo(a)pyrene	50-32-8	250	N.D.
Benzyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2 ⁴-Dimethylphenol	105-67-9	100	N.D.
I ıthyl phthalate	131-11-3	100	N.D.
4,ɔ-Dinitro-2-methylphenol	534-52-1	100	N.D.
2,4-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP41
Laboratory Sample #: 97040411

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	N.D.	
Fluorene	86-73-7	100	N.D.	
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	800	<---
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	N.D.	
2-Methylphenol	95-48-7	100	N.D.	
4-Methylphenol	106-44-5	100	N.D.	
Naphthalene	91-20-3	100	N.D.	
2-Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	N.D.	
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	180	<---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040413
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-16-97
Reported: 04-24-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040414
Laboratory Reference #: MWI 8883

Sampled: 04-09-97
Received: 04-09-97
Analyzed: 04-16-97
Reported: 04-24-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/10/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8883

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	2.5	2.2	2.2	88	88	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/10/97
Laboratory Sample No : 97040342
Laboratory Reference No : MWI 8883

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0	100	91	101	91	101	10

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/11/97
Laboratory Sample No : 97040139
Laboratory Reference No : MWI 8883

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	3.000	2.800	3.300	93	110	16
Arsenic	0.00	1.000	1.180	1.170	118	117	1
Barium	0.00	0.100	0.095	0.096	95	96	1
Beryllium	0.00	0.100	0.113	0.120	113	120	6
Cadmium	0.00	0.100	0.113	0.117	113	117	3
Chromium (VI)	0.00	0.500	0.440	0.490	88	98	11
Chromium (Total)	0.00	0.100	0.105	0.112	105	112	6
Cobalt	0.00	0.100	0.117	0.102	117	102	14
Copper	0.00	0.100	0.096	0.109	96	109	13
Lead	0.00	1.000	0.980	1.010	98	101	3
Mercury	0.000	0.020	0.021	0.023	105	115	9
Molybdenum	0.00	1.000	0.940	1.100	94	110	16
Nickel	0.00	0.500	0.580	0.560	116	112	4
Selenium	0.00	1.000	0.980	1.010	98	101	3
Silver	0.00	0.500	0.530	0.550	106	110	4
Thallium	0.00	3.000	3.300	3.400	110	113	3
Vanadium	0.00	0.500	0.590	0.600	118	120	2
Zinc	0.00	0.100	0.120	0.120	120	120	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

RANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/12/97
Laboratory Sample No : 97040513
Laboratory Reference No : MWI 8883

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	30.00	27.00	26.00	90	87	4
Arsenic	0.00	10.00	8.00	9.30	80	93	15
Barium	13.00	10.00	23.00	22.00	100	90	4
Beryllium	0.00	1.00	1.01	1.03	101	103	2
Cadmium	0.00	1.00	0.87	0.89	87	89	2
Chromium (Total)	2.50	5.00	7.40	7.70	98	104	4
Chromium (VI)	0.00	1.00	0.92	0.86	92	86	7
Cobalt	0.79	1.00	1.72	1.72	93	93	0
Copper	1.70	1.00	2.65	2.80	95	110	6
Lead	0.00	10.00	9.40	9.30	94	93	1
Mercury	0.000	0.020	0.018	0.016	90	80	12
Molybdenum	0.00	10.00	8.50	8.40	85	84	1
Nickel	1.40	5.00	6.20	6.20	96	96	0
Selenium	0.00	10.00	10.20	8.90	102	89	14
Silver	0.00	5.00	4.40	4.20	88	84	5
Thallium	0.00	30.00	28.00	27.00	93	90	4
Vanadium	3.10	5.00	8.20	8.30	102	104	1
Zinc	4.70	5.00	9.90	9.30	104	92	6

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Organochlorine Pesticides (EPA 8080)

Date of Analysis : 04/16/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8883

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
4,4'-DDT	0.0	1.00	0.92	0.78	92	78	16

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : PCB 'S (EPA 8080)

Date of Analysis : 04/16/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8883

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
PCB-1260	0.0	50	25	29	50	58	15

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/16/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8883

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	38	44	76	88	15
n-Nitroso-di-n-propylamine	0.0	50	47	34	94	68	32
1,2,4-Trichlorobenzene	0.0	50	38	44	76	88	15
Acenaphthene	0.0	50	47	50	94	100	6
Pyrene	0.0	50	43	46	86	92	7
Pentachlorophenol	0.0	100	78	74	78	74	5
4-Chloro-3-Methylphenol	0.0	100	60	62	60	62	3
2-Chlorophenol	0.0	100	79	81	79	81	3
Phenol	0.0	100	22	24	22	24	9

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/14/97
Laboratory Sample No : 97040433
Laboratory Reference No : MWI 8883

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	45	49	90	98	9
n-Nitroso-di-n-propylamine	0.0	50	44	47	88	94	7
1,2,4-Trichlorobenzene	0.0	50	45	49	90	98	9
Acenaphthene	0.0	50	49	53	98	106	8
Pyrene	0.0	50	46	49	92	98	6
Pentachlorophenol	0.0	100	78	88	78	88	12
4-Chloro-3-Methylphenol	0.0	100	70	80	70	80	13
2-Chlorophenol	0.0	100	83	91	83	91	9
Phenol	0.0	100	74	82	74	82	10

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/22/97
Laboratory Sample No : 97040609
Laboratory Reference No : MWI 8883

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	20	18	17	90	85	6
1,1-Dichloroethene	0.0	20	17	17	85	85	0
Trichloroethene	0.0	20	19	19	95	95	0
Toluene	0.0	20	18	18	90	90	0
Chlorobenzene	0.0	20	16	17	80	85	6

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/21/97
Laboratory Sample No : 97040717
Laboratory Reference No : MWI 8883

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	56	51	112	102	9
1,1-Dichloroethene	0.0	50	42	39	84	78	7
Trichloroethene	0.0	50	65	62	130	124	5
Toluene	0.0	50	59	56	118	112	5
Chlorobenzene	0.0	50	53	51	106	102	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532
Tustin, CA 92680
(714) 832-0064, Fax (714) 832-0067

Analys[illegible] [illegible]st and Chain of Custody Record

Lab Job No: _____
Page 1

REQUIRED TAT: 24hr TAT on Soil TRPH, STANDARD TAT Otherwise

CUSTOMER INFORMATION		PROJECT INFORMATION	
COMPANY:	Montgomery Watson	PROJECT NAME:	McDonnell Douglas C-6 Facility
SEND REPORT TO:	Fred Strauss	NUMBER:	1206035.01090010
ADDRESS:	250 N. Madison Ave. Pasadena, CA 91101	LOCATION:	Bldg. 37 Area
		ADDRESS:	19503 S. Normandie Ave. Torrance, CA
PHONE:	(818) 568-6582	SAMPLED BY:	AWW/IV
FAX:	(818) 796-5941		

SAMPLE ID	NO. OF CONTAINERS	SAMPLE DATE	SAMPLE TIME	SAMPLE MATRIX	CONTAINER TYPE	PRES.	TRPH 418.1	Title 22 Metals	VOCs 8260	SVOCs 8270	PCBs 8080	Pesticides 8080			REMARKS/PRECAUTIONS
B37-RE4-SP41	2	4/9/97	1345	Soil	2" x 6" SS sleeves	ICE	X	X	X	X					
B37-RE4-SP42	2	↓	1530	↓	↓	↓	X	X	X	X					
Equipment Blank	7	4/9/97	1530	Water	Amber & Poly bottles	HNO3 HCl	X	X	X	X	X	X			
Rinsate Blank	7	↓	1545	↓	↓	↓	X	X	X	X	X	X			
Trip Blank	2	↓	—	↓	2 VOAs	HCl			X						

Total No. of Samples: 5

Method of Shipment: One cooler via Orange Coast Courier

Relinquished By:	Date/Time:	Received By:	Date/Time:
Adam W. Nonig	4/9/97 1605		
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received For Lab By:	Date/Time: 4/9/97 16:05

Reporting Format: (check)
NORMAL _____ S.D. HMMD _____
RWQCB _____ OTHER _____

Sample Integrity: (check)
intact _____ on ice _____

All samples remain the property of the client who is responsible for disposal. A disposal fee may be imposed if client fails to pickup samples.

BOE-C6-0143721

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Soil

Laboratory Reference #: MWI 8895

Sampled : 04-11-97
Received: 04-11-97
Analyzed: 04-14-97
Reported: 04-14-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/kg)*
97040594	B37-RE4-SP43	180
97040595	B37-RE4-SP44	140
Detection Limit:		8.0

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP43

Laboratory Sample Number: 97040594
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-17-97
Reported: 04-25-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	110	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	30	<---
Cobalt	6010	80	8000	0.5	5.9	<---
Copper	6010	25	2500	0.1	14	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	11	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	25	<---
Zinc	6010	250	5000	0.1	40	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP44

Laboratory Sample Number: 97040595
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-17-97
Reported: 04-25-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	130	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	27	<---
Cobalt	6010	80	8000	0.5	7.0	<---
Copper	6010	25	2500	0.1	18	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	12	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	27	<---
Zinc	6010	250	5000	0.1	57	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP43
Laboratory Sample Number: 97040594
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-21-97
Reported: 04-25-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
1,1-Dichloroethene	75-35-4	2.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood; Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP43

Laboratory Sample Number: 97040594

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
Butylbenzene	104-51-8	2.5	N.D.
2-Dichlorobenzene	95-50-1	2.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

rk Noorani
aboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*105*
Toluene-d8	*89*
4-Bromofluorobenbzene	*112*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP44
Laboratory Sample Number: 97040595
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-21-97
Reported: 04-25-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
Benzene	71-43-2	100	N.D.
Bromodichloromethane	75-27-4	100	N.D.
Bromoform	75-25-2	100	N.D.
Bromomethane	74-83-9	100	N.D.
Carbon Disulfide	75-15-0	200	N.D.
Carbon tetrachloride	56-23-5	100	N.D.
Chlorobenzene	108-90-7	100	N.D.
Chlorodibromomethane	124-48-1	100	N.D.
Chloroethane	75-00-3	100	N.D.
2-Chloroethyl vinyl ether	110-75-8	200	N.D.
Chloroform	67-66-3	100	N.D.
Chloromethane	74-87-3	100	N.D.
1,1-Dichloroethane	75-35-3	100	N.D.
1,2-Dichloroethane	107-06-2	100	N.D.
1,1-Dichloroethene	75-35-4	100	N.D.
Trans 1,2-Dichloroethene	156-60-5	100	N.D.
1,2-Dichloropropane	78-87-5	100	N.D.
cis-1,3-Dichloropropene	10061-01-5	100	N.D.
trans-1,3-Dichloropropene	10061-02-6	100	N.D.
Ethylbenzene	100-41-4	100	N.D.
Methylene chloride	75-09-2	200	N.D.
Styrene	100-42-5	100	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	100	N.D.
Tetrachloroethene	127-18-4	100	N.D.
Toluene	108-88-3	100	N.D.
1,1,1-Trichloroethane	71-55-6	100	N.D.
1,1,2-Trichloroethane	79-00-5	100	N.D.
Trichloroethene	79-01-6	100	N.D.
Trichlorofluoromethane	75-69-4	200	N.D.
Vinyl acetate	108-05-4	200	N.D.
Vinyl chloride	75-01-4	100	N.D.
Total Xylenes	1330-20-7	100	N.D.
Dichlorofluoromethane	75-71-8	100	N.D.
cis-1-2,-Dichloroethane	156-59-4	100	N.D.
2,2-Dichloropropane	590-20-7	100	N.D.
Bromochloromethane	74-97-5	100	N.D.
1,1-Dichloropropene	563-58-6	100	N.D.
1,2-Dichloropropane	78-87-5	100	N.D.
Dibromomethane	74-95-3	100	N.D.
trans-1,3-Dichloropropane	10061-02-6	100	N.D.
1,1,2-Trichloroethane	79-00-5	100	N.D.
1,2-Dibromoethane	106-93-4	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP44

Laboratory Sample Number: 97040595

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*	
1,3-Dichloropropane	142-28-9	100	N.D.	
Dibromochloromethane	124-48-1	100	N.D.	
Isopropylbenzene	98-82-8	100	N.D.	
1,1,2,2-Tetrachloroethane	79-34-5	100	N.D.	
1,2,3-Trichloropropane	96-18-4	100	N.D.	
Bromobenzene	108-86-1	100	N.D.	
n-Propylbenzene	103-65-1	100	N.D.	
2-Chlorotoluene	95-49-8	100	N.D.	
1,3,5-Trimethylbenzene	108-67-8	100	N.D.	
4-Chlorotoluene	106-43-4	100	N.D.	
tert-Butylbenzene	98-06-6	100	N.D.	
1,2,4-Trimethylbenzene	95-63-6	100	230	<---
sec-Butylbenzene	135-98-8	100	N.D.	
4-Isopropyltoluene	99-87-6	100	N.D.	
1,3-Dichlorobenzene	541-73-1	100	N.D.	
1,4-Dichlorobenzene	106-46-7	100	N.D.	
n-Butylbenzene	104-51-8	100	100	<---
1,2-Dichlorobenzene	95-50-1	100	N.D.	
1-2-Dibromo-2-CPA	96-12-8	200	N.D.	
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Naphthalene	91-20-3	100	N.D.	
1,2,3-Trichlorobenzene	87-61-6	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*115*
Toluene-d8	*91*
4-Bromofluorobenbzene	*123**

*Matrix Interference

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP43
Laboratory Sample #: 97040594
Laboratory Reference #: MWI 8895

Sampled : 04-11-97
Received: 04-11-97
Analyzed: 04-21-97
Reported: 04-25-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
Acenaphthene	83-32-9	100	N.D.	
Acenaphthylene	208-96-8	100	N.D.	
Aniline	62-53-3	100	N.D.	
Anthracene	120-12-7	100	N.D.	
Benzidine	92-87-5	500	N.D.	
Benzoic Acid	65-85-0	250	N.D.	
Benzo(a)anthracene	56-55-3	100	220	<---
Benzo(b)fluoranthene	205-99-2	250	N.D.	
Benzo(k)fluoranthene	207-08-9	250	N.D.	
Benzo(g,h,i)perylene	191-24-2	250	N.D.	
Benzo(a)pyrene	50-32-8	250	N.D.	
Benzyl alcohol	100-51-6	100	N.D.	
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.	
Bis(2-chloroethyl)ether	111-44-4	100	N.D.	
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.	
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.	
4-Bromophenyl phenyl ether	101-55-3	100	N.D.	
Butyl benzyl phthalate	85-68-7	100	N.D.	
4-Chloroaniline	106-47-8	100	N.D.	
2-Chloronaphthalene	91-58-7	100	N.D.	
4-Chloro-3-methylphenol	59-50-7	100	N.D.	
2-Chlorophenol	95-57-8	100	N.D.	
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.	
Chrysene	218-0109	100	280	<---
Dibenz(a,h)anthracene	53-70-3	100	N.D.	
Dibenzofuran	132-64-9	100	N.D.	
Di-n-butyl phthalate	84-74-2	250	N.D.	
1,3-Dichlorobenzene	541-73-1	100	N.D.	
1,4-Dichlorobenzene	106-46-7	100	N.D.	
1,2-Dichlorobenzene	95-50-1	100	N.D.	
3,3'-Dichlorobenzidine	91-94-1	100	N.D.	
2,4-Dichlorophenol	120-83-2	100	N.D.	
Diethyl phthalate	84-66-2	100	N.D.	
2,4-Dimethylphenol	105-67-9	100	N.D.	
Dimethyl phthalate	131-11-3	100	N.D.	
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.	
2,4-Dinitrophenol	51-28-5	100	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP43
Laboratory Sample #: 97040594

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	370	<---
Fluorene	86-73-7	100	N.D.	
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
I[illegible]eno(1,2,3-cd)pyrene	193-39-5	250	N.D.	
I[illegible]phorone	78-59-1	100	N.D.	
2[illegible]ethylnaphthalene	91-57-6	100	N.D.	
2[illegible]thylphenol	95-48-7	100	N.D.	
4-methylphenol	106-44-5	100	N.D.	
Naphthalene	91-20-3	100	N.D.	
2-Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	110	<---
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	210	<---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

[illegible]NGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sampled : 04-11-97
Received: 04-11-97
Analyzed: 04-22-97
Reported: 04-25-97

Sample Description: Soil, B37-RE4-SP44
Laboratory Sample #: 97040595
Laboratory Reference #: MWI 8895

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	500 <---
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	700 <---
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	5,300 <---
Benzo(b)fluoranthene	205-99-2	250	7,100 <---
Benzo(k)fluoranthene	207-08-9	250	2,900 <---
Benzo(g,h,i)perylene	191-24-2	250	4,200 <---
Benzo(a)pyrene	50-32-8	250	5,700 <---
Benzyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	150 <---
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	5,800 <---
Dibenz(a,h)anthracene	53-70-3	100	1,100 <---
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
Dimethyl phthalate	131-11-3	100	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.
2,4-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP44
Laboratory Sample #: 97040595

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
2,4-Dinitrotoluene	121-14-2	250	N.D.
2,6-Dinitrotoluene	606-20-2	250	N.D.
Di-n-octyl phthalate	117-84-0	250	N.D.
Fluoranthene	206-44-0	100	9,300 <---
Fluorene	86-73-7	100	240 <---
Hexachlorobenzene	118-74-1	100	N.D.
Hexachlorobutadiene	87-68-3	100	N.D.
Hexachlorocyclopentadiene	77-47-4	100	N.D.
Hexachloroethane	67-72-1	100	N.D.
eno(1,2,3-cd)pyrene	193-39-5	250	4,600 <---
phorone	78-59-1	100	N.D.
ethylnaphthalene	91-57-6	100	120 <---
2 thylphenol	95-48-7	100	N.D.
4-methylphenol	106-44-5	100	N.D.
Naphthalene	91-20-3	100	280 <---
2-Nitroaniline	88-74-4	250	N.D.
3-Nitroaniline	99-09-2	250	N.D.
4-Nitroaniline	100-01-6	250	N.D.
Nitrobenzene	98-95-3	100	N.D.
2-Nitrophenol	88-75-5	100	N.D.
4-Nitrophenol	100-02-7	100	N.D.
N-Nitrosodiphenylamine	86-30-6	100	N.D.
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.
N-Nitrosodimethylamine	62-75-9	100	N.D.
Pentachlorophenol	87-86-5	250	N.D.
Phenanthrene	85-01-8	100	2,600 <---
Phenol	108-95-2	100	N.D.
Pyrene	129-00-0	100	7,200 <---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.
2,4,5-Trichlorophenol	95-95-4	100	N.D.
2,4,6-Trichlorophenol	88-06-2	100	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

NGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Water

Laboratory Reference #: MWI 8895

Sampled : 04-11-97
Received: 04-11-97
Analyzed: 04-18-97
Reported: 04-25-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/l)*
97040596	Equip Blank	N.D.
97040597	Rinsate Blank	N.D.

Detection Limit: 0.5

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040596
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-22-97
Reported: 04-25-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
boratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040597
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-22-97
Reported: 04-25-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Equip Blank
Laboratory Sample Number: 97040596
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-22-97
Reported: 04-25-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040596

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*107*
Toluene-d8	*96*
4-Bromofluorobenzene	*98*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample Number: 97040597
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-22-97
Reported: 04-25-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040597

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*111*
Toluene-d8	*99*
4-Bromofluorobenzene	*99*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Trip Blank
Laboratory Sample Number: 97040598
Laboratory Reference #: MWI 8895

Sampled: 04-11-97
Received: 04-11-97
Analyzed: 04-22-97
Reported: 04-25-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Trip Blank

Laboratory Sample Number: 97040598

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*107*
Toluene-d8	*90*
4-Bromofluorobenzene	*101*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040596
Laboratory Reference #: MWI 8895

Sampled : 04-11-97
Received: 04-11-97
Analyzed: 04-16-97
Reported: 04-25-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040596

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
4-Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : PCB 'S (EPA 8080)

Date of Analysis : 04/16/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8895

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
PCB-1260	0.0	50	25	29	50	58	15

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
S	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

RANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Organochlorine Pesticides (EPA 8080)

Date of Analysis : 04/16/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8895

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
4,4'-DDT	0.0	1.00	0.92	0.78	92	78	16

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/17/97
Laboratory Sample No : 97040636
Laboratory Reference No : MWI 8895

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	30.00	30.20	28.80	101	96	5
Arsenic	0.00	10.00	8.20	8.00	82	80	2
Barium	12.00	10.00	22.00	21.00	100	90	5
Beryllium	0.00	1.00	1.03	0.92	103	92	11
Cadmium	0.00	1.00	1.00	1.05	100	105	5
Chromium (Total)	2.60	5.00	7.00	6.60	88	80	6
Chromium (VI)	0.00	1.00	0.87	0.81	87	81	7
Cobalt	0.69	1.00	1.51	1.64	82	95	8
Copper	1.50	1.00	2.30	2.30	80	80	0
ad	0.00	10.00	9.90	10.20	99	102	3
Mercury	0.000	0.020	0.019	0.021	95	105	10
Molybdenum	0.00	10.00	10.80	10.90	108	109	1
Nickel	1.10	5.00	5.70	5.60	92	90	2
Selenium	0.00	10.00	10.50	9.90	105	99	6
Silver	0.00	5.00	5.20	5.20	104	104	0
Thallium	0.00	30.00	32.00	33.00	107	110	3
Vanadium	2.90	5.00	7.50	7.60	92	94	1
Zinc	4.70	5.00	8.80	8.70	82	80	1

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

RANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/22/97
Laboratory Sample No : 97040596
Laboratory Reference No : MWI 8893

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	3.000	2.420	2.700	81	90	11
Arsenic	0.00	1.000	0.800	0.920	80	92	14
Barium	0.00	0.100	0.105	0.104	105	104	1
Beryllium	0.00	0.100	0.085	0.087	85	87	2
Cadmium	0.00	0.100	0.089	0.097	89	97	9
Chromium (VI)	0.00	0.500	0.450	0.490	90	98	9
Chromium (Total)	0.00	0.100	0.087	0.091	87	91	4
Cobalt	0.00	0.100	0.109	0.107	109	107	2
Copper	0.00	0.100	0.080	0.089	80	89	11
Lead	0.00	1.000	1.040	1.010	104	101	3
Mercury	0.000	0.020	0.020	0.022	100	110	10
Molybdenum	0.00	1.000	0.920	1.000	92	100	8
Nickel	0.00	0.500	0.500	0.480	100	96	4
Selenium	0.00	1.000	0.860	0.950	86	95	10
Silver	0.00	0.500	0.440	0.480	88	96	9
Thallium	0.00	3.000	2.460	2.650	82	88	7
Vanadium	0.00	0.500	0.450	0.470	90	94	4
Zinc	0.00	0.100	0.109	0.109	109	109	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/14/97
Laboratory Sample No : 97040511
Laboratory Reference No : MWI 8895

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	100	106	106	106	106	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
.S	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

RANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/18/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8895

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	5.0	3.8	4.0	76	80	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532
Tustin, CA 92680
(714) 832-0064, Fax (714) 832-0067

Anal[illegible] [illegible]quest and Chain o[illegible] [illegible]ustody Record

Lab Job No: [illegible]
Page [illegible] 1

REQUIRED TAT: 24 hr TAT on SOIL TRPH only Standard TAT Otherwise

CUSTOMER INFORMATION		PROJECT INFORMATION	
COMPANY:	Montgomery Watson	PROJECT NAME:	McDonnell Douglas
SEND REPORT TO:	Fred Strauss	NUMBER:	1206035.01090010
ADDRESS:	2150 N Madison Ave	LOCATION:	C-6 facility Bldg 37
	Pasadena, CA	ADDRESS:	19503 S Normandie Ave
			Torrance, CA
PHONE:	(818)568-6582 FAX: (818)796-5941	SAMPLED BY:	AN / JV

SAMPLE ID	NO. OF CONTAINERS	SAMPLE DATE	SAMPLE TIME	SAMPLE MATRIX	CONTAINER TYPE	PRES.	TRPH 418.1	Title 22 Metals	VOCs 8260	SVOCs 8270	PCBs 8080	Pesticides 8080			REMARKS/PRECAUTIONS
B37-RE4-SP43	2	4/11/97	1100	SOIL	2"x6" SS Sleeve	ICE	X	X	X	X					
B37-RE4-SP44	2	4/11/97	1520	SOIL	↓	↓	X	X	X	X					
Equipment Blank	7	4/11/97	1530	Water	4 amber 1 poly 2 VOAs	HNO3 HCl	X	X	X	X	X	X			
Rinsate Blank	7	4/11/97	1545	Water	↓	↓	X	X	X	X	X	X			
Trip Blank	2	4/11/97	—	Water	2 VOAs	HCl			X						

Total No. of Samples: 5

Method of Shipment: 1 cooler via Orange Coast Courier

Relinquished By:	Date/Time:	Received By:	Date/Time:
Adam W. Vonis	4/11/97 1615		
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received For Lab By:	Date/Time:
		[illegible]	4/11/97 16:15

Reporting Format: (check)
NORMAL _____ S.D. HMMD _____
RWQCB _____ OTHER _____

Sample Integrity: (check)
intact _____ on ice _____

All samples remain the property of the client who is responsible for disposal. A disposal fee may be imposed if client fails to pickup samples.

BOE-C6-0143750

Date/Time:	Received For Lab By:	Date/Time:	Sample Integrity: (check)
	[illegible]	4/11/97 16:15	intact _____ on ice _____

All samples remain the property of the client who is responsible for disposal. A disposal fee may be imposed if client fails to pickup samples.



ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
.N: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sampled : 05-09-97
Sample Description: Soil, B37-RE4-SP44A *Received:* 05-09-97
Laboratory Sample #: 97050208 *Analyzed:* 05-10-97
Laboratory Reference #: MWI 9037 *Reported:* 05-12-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
nzo(k)fluoranthene	207-08-9	250	N.D.
nzo(g,h,i)perylene	191-24-2	250	N.D.
zo(a)pyrene	50-32-8	250	N.D.
zyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
methyl phthalate	131-11-3	100	N.D.
-Dinitro-2-methylphenol	534-52-1	100	N.D.
Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP44A
Laboratory Sample #: 97050208

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
2,4-Dinitrotoluene	121-14-2	250	N.D.
2,6-Dinitrotoluene	606-20-2	250	N.D.
Di-n-octyl phthalate	117-84-0	250	N.D.
Fluoranthene	206-44-0	100	N.D.
Fluorene	86-73-7	100	N.D.
Hexachlorobenzene	118-74-1	100	N.D.
Hexachlorobutadiene	87-68-3	100	N.D.
Hexachlorocyclopentadiene	77-47-4	100	N.D.
Hexachloroethane	67-72-1	100	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.
Isophorone	78-59-1	100	N.D.
2-Methylnaphthalene	91-57-6	100	N.D.
2-Methylphenol	95-48-7	100	N.D.
4-Methylphenol	106-44-5	100	N.D.
Naphthalene	91-20-3	100	N.D.
2-Nitroaniline	88-74-4	250	N.D.
3-Nitroaniline	99-09-2	250	N.D.
4-Nitroaniline	100-01-6	250	N.D.
Nitrobenzene	98-95-3	100	N.D.
2-Nitrophenol	88-75-5	100	N.D.
4-Nitrophenol	100-02-7	100	N.D.
N-Nitrosodiphenylamine	86-30-6	100	N.D.
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.
N-Nitrosodimethylamine	62-75-9	100	N.D.
Pentachlorophenol	87-86-5	250	N.D.
Phenanthrene	85-01-8	100	N.D.
Phenol	108-95-2	100	N.D.
Pyrene	129-00-0	100	N.D.
1,2,4-Trichlorobenzene	120-82-1	100	N.D.
2,4,5-Trichlorophenol	95-95-4	100	N.D.
2,4,6-Trichlorophenol	88-06-2	100	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
.TN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP44B
Laboratory Sample #: 97050209
Laboratory Reference #: MWI 9037

Sampled : 05-09-97
Received: 05-09-97
Analyzed: 05-10-97
Reported: 05-12-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
Acenaphthene	83-32-9	100	N.D.	
Acenaphthylene	208-96-8	100	N.D.	
Aniline	62-53-3	100	N.D.	
Anthracene	120-12-7	100	N.D.	
Benzidine	92-87-5	500	N.D.	
Benzoic Acid	65-85-0	250	N.D.	
Benzo(a)anthracene	56-55-3	100	260	<---
Benzo(b)fluoranthene	205-99-2	250	340	<---
nzo(k)fluoranthene	207-08-9	250	N.D.	
nzo(g,h,i)perylene	191-24-2	250	N.D.	
zo(a)pyrene	50-32-8	250	300	<---
nzyl alcohol	100-51-6	100	N.D.	
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.	
Bis(2-chloroethyl)ether	111-44-4	100	N.D.	
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.	
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.	
4-Bromophenyl phenyl ether	101-55-3	100	N.D.	
Butyl benzyl phthalate	85-68-7	100	N.D.	
4-Chloroaniline	106-47-8	100	N.D.	
2-Chloronaphthalene	91-58-7	100	N.D.	
4-Chloro-3-methylphenol	59-50-7	100	N.D.	
2-Chlorophenol	95-57-8	100	N.D.	
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.	
Chrysene	218-0109	100	380	<---
Dibenz(a,h)anthracene	53-70-3	100	N.D.	
Dibenzofuran	132-64-9	100	N.D.	
Di-n-butyl phthalate	84-74-2	250	N.D.	
1,3-Dichlorobenzene	541-73-1	100	N.D.	
1,4-Dichlorobenzene	106-46-7	100	N.D.	
1,2-Dichlorobenzene	95-50-1	100	N.D.	
3,3'-Dichlorobenzidine	91-94-1	100	N.D.	
2,4-Dichlorophenol	120-83-2	100	N.D.	
Diethyl phthalate	84-66-2	100	N.D.	
2,4-Dimethylphenol	105-67-9	100	N.D.	
methyl phthalate	131-11-3	100	N.D.	
-Dinitro-2-methylphenol	534-52-1	100	N.D.	
-Dinitrophenol	51-28-5	100	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP44B
Laboratory Sample #: 97050209

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	480	<---
Fluorene	86-73-7	100	N.D.	
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.	
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	N.D.	
2-Methylphenol	95-48-7	100	N.D.	
4-Methylphenol	106-44-5	100	N.D.	
Naphthalene	91-20-3	100	N.D.	
2-Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	210	<---
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	470	<---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
..N: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sampled : 05-09-97
Sample Description: Soil, B37-RE4-SP44C *Received:* 05-09-97
Laboratory Sample #: 97050210 *Analyzed:* 05-10-97
Laboratory Reference #: MWI 9037 *Reported:* 05-12-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
Acenaphthene	83-32-9	100	N.D.	
Acenaphthylene	208-96-8	100	N.D.	
Aniline	62-53-3	100	N.D.	
Anthracene	120-12-7	100	N.D.	
Benzidine	92-87-5	500	N.D.	
Benzoic Acid	65-85-0	250	N.D.	
Benzo(a)anthracene	56-55-3	100	520	<---
Benzo(b)fluoranthene	205-99-2	250	820	<---
..nzo(k)fluoranthene	207-08-9	250	300	<---
..nzo(g,h,i)perylene	191-24-2	250	480	<---
..zo(a)pyrene	50-32-8	250	610	<---
..zyl alcohol	100-51-6	100	N.D.	
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.	
Bis(2-chloroethyl)ether	111-44-4	100	N.D.	
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.	
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.	
4-Bromophenyl phenyl ether	101-55-3	100	N.D.	
Butyl benzyl phthalate	85-68-7	100	N.D.	
4-Chloroaniline	106-47-8	100	N.D.	
2-Chloronaphthalene	91-58-7	100	N.D.	
4-Chloro-3-methylphenol	59-50-7	100	N.D.	
2-Chlorophenol	95-57-8	100	N.D.	
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.	
Chrysene	218-0109	100	680	<---
Dibenz(a,h)anthracene	53-70-3	100	110	<---
Dibenzofuran	132-64-9	100	N.D.	
Di-n-butyl phthalate	84-74-2	250	N.D.	
1,3-Dichlorobenzene	541-73-1	100	N.D.	
1,4-Dichlorobenzene	106-46-7	100	N.D.	
1,2-Dichlorobenzene	95-50-1	100	N.D.	
3,3'-Dichlorobenzidine	91-94-1	100	N.D.	
2,4-Dichlorophenol	120-83-2	100	N.D.	
Diethyl phthalate	84-66-2	100	N.D.	
2,4-Dimethylphenol	105-67-9	100	N.D.	
..methyl phthalate	131-11-3	100	N.D.	
..-Dinitro-2-methylphenol	534-52-1	100	N.D.	
..Dinitrophenol	51-28-5	100	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP44C
Laboratory Sample #: 97050210

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	870	<---
Fluorene	86-73-7	100	N.D.	
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	520	<---
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	N.D.	
2-Methylphenol	95-48-7	100	N.D.	
4-Methylphenol	106-44-5	100	N.D.	
Naphthalene	91-20-3	100	N.D.	
2-Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	260	<---
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	840	<---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP44D
Laboratory Sample #: 97050211
Laboratory Reference #: MWI 9037

Sampled : 05-09-97
Received: 05-09-97
Analyzed: 05-10-97
Reported: 05-12-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
enzo(k)fluoranthene	207-08-9	250	N.D.
enzo(g,h,i)perylene	191-24-2	250	N.D.
zo(a)pyrene	50-32-8	250	N.D.
enzyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
imethyl phthalate	131-11-3	100	N.D.
-Dinitro-2-methylphenol	534-52-1	100	N.D.
-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP44D
Laboratory Sample #: 97050211

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
2,4-Dinitrotoluene	121-14-2	250	N.D.
2,6-Dinitrotoluene	606-20-2	250	N.D.
Di-n-octyl phthalate	117-84-0	250	N.D.
Fluoranthene	206-44-0	100	N.D.
Fluorene	86-73-7	100	N.D.
Hexachlorobenzene	118-74-1	100	N.D.
Hexachlorobutadiene	87-68-3	100	N.D.
Hexachlorocyclopentadiene	77-47-4	100	N.D.
Hexachloroethane	67-72-1	100	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.
Isophorone	78-59-1	100	N.D.
2-Methylnaphthalene	91-57-6	100	N.D.
2-Methylphenol	95-48-7	100	N.D.
4-Methylphenol	106-44-5	100	N.D.
Naphthalene	91-20-3	100	N.D.
2-Nitroaniline	88-74-4	250	N.D.
3-Nitroaniline	99-09-2	250	N.D.
4-Nitroaniline	100-01-6	250	N.D.
Nitrobenzene	98-95-3	100	N.D.
2-Nitrophenol	88-75-5	100	N.D.
4-Nitrophenol	100-02-7	100	N.D.
N-Nitrosodiphenylamine	86-30-6	100	N.D.
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.
N-Nitrosodimethylamine	62-75-9	100	N.D.
Pentachlorophenol	87-86-5	250	N.D.
Phenanthrene	85-01-8	100	N.D.
Phenol	108-95-2	100	N.D.
Pyrene	129-00-0	100	N.D.
1,2,4-Trichlorobenzene	120-82-1	100	N.D.
2,4,5-Trichlorophenol	95-95-4	100	N.D.
2,4,6-Trichlorophenol	88-06-2	100	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

'ontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Water

Laboratory Reference #: MWI 9037

Sampled : 05-09-97
Received: 05-09-97
Analyzed: 05-14-97
Reported: 05-16-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	Client Sample Number	Sample Result (mg/l)
97050213	Equip Blank	N.D.
97050214	Rinsate Blank	N.D.
Detection Limit:		0.5

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

ark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97050214
Laboratory Reference #: MWI 9037

Sampled: 05-09-97
Received: 05-09-97
Analyzed: 05-14-97
Reported: 05-16-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97050213
Laboratory Reference #: MWI 9037

Sampled: 05-09-97
Received: 05-09-97
Analyzed: 05-14-97
Reported: 05-16-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Equip Blank
Laboratory Sample Number: 97050213
Laboratory Reference #: MWI 9037

Sampled: 05-09-97
Received: 05-09-97
Analyzed: 05-13-97
Reported: 05-16-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97050213

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*97*
Toluene-d8	*99*
4-Bromofluorobenzene	*100*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample Number: 97050214
Laboratory Reference #: MWI 9037

Sampled: 05-09-97
Received: 05-09-97
Analyzed: 05-13-97
Reported: 05-16-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97050214

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*103*
Toluene-d8	*101*
4-Bromofluorobenzene	*103*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss'
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Trip Blank
Laboratory Sample Number: 97050215
Laboratory Reference #: MWI 9037

Sampled: 05-09-97
Received: 05-09-97
Analyzed: 05-13-97
Reported: 05-16-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Trip Blank

Laboratory Sample Number: 97050215

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*103*
Toluene-d8	*103*
4-Bromofluorobenzene	*103*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank
Laboratory Sample #: 97050213
Laboratory Reference #: MWI 9037

Sampled : 05-09-97
Received: 05-09-97
Analyzed: 05-13-97
Reported: 05-16-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Equip Blank
Laboratory Sample #: 97050213

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
Methylphenol	95-48-7	5.0	N.D.
Methylphenol	106-44-5	5.0	N.D.
hthalene	91-20-3	50	N.D.
Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

: Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97050214
Laboratory Reference #: MWI 9037

Sampled : 05-09-97
Received: 05-09-97
Analyzed: 05-13-97
Reported: 05-16-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97050214

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
-Methylphenol	95-48-7	5.0	N.D.
-Methylphenol	106-44-5	5.0	N.D.
ıthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97050213
Laboratory Reference #: MWI 9037

Sampled: 05-09-97
Received: 05-09-97
Analyzed: 05-15-97
Reported: 05-16-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97050214
Laboratory Reference #: MWI 9037

Sampled: 05-09-97
Received: 05-09-97
Analyzed: 05-15-97
Reported: 05-16-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97050213
Laboratory Reference #: MWI 9037

Sampled: 05-09-97
Received: 05-09-97
Analyzed: 05-15-97
Reported: 05-16-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97050214
Laboratory Reference #: MWI 9037

Sampled: 05-09-97
Received: 05-09-97
Analyzed: 05-15-97
Reported: 05-16-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 05/12/97
Laboratory Sample No : 97050207
Laboratory Reference No : MWI 9037

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.0	30.0	29.0	28.0	97	93	4
Arsenic	0.0	10.0	11.0	11.8	110	118	7
Barium	22.0	10.0	32.0	32.0	100	100	0
Beryllium	0.00	1.00	0.98	0.97	98	97	1
Cadmium	0.00	1.00	1.05	1.03	105	103	2
Chromium (Total)	3.1	5.0	7.7	7.7	92	92	0
Chromium (VI)	0.00	1.00	1.05	0.92	105	92	13
Cobalt	0.97	1.00	1.85	1.95	88	98	5
Copper	0.84	1.00	1.77	1.80	93	96	2
Lead	0.0	10.0	9.5	9.5	95	95	0
Mercury	0.000	0.020	0.018	0.019	90	95	5
Molybdenum	0.0	10.0	8.8	8.5	88	85	3
Nickel	0.80	5.00	5.40	5.40	92	92	0
Selenium	0.0	10.0	9.0	9.2	90	92	2
Silver	0.0	5.0	4.1	3.9	82	78	5
Thallium	0.0	30.0	27.0	27.0	90	90	0
Vanadium	2.9	5.0	7.8	7.8	98	98	0
Zinc	5.4	5.0	9.6	9.9	84	90	3

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 05/14/97
Laboratory Sample No : 97050046
Laboratory Reference No : MWI 9037

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	3.00	2.90	3.10	97	103	7
Arsenic	0.00	1.00	0.90	0.98	90	98	9
Barium	0.00	0.100	0.093	0.101	93	101	8
Beryllium	0.00	0.100	0.098	0.094	98	94	4
Cadmium	0.00	0.100	0.112	0.117	112	117	4
Chromium (VI)	0.00	0.50	0.47	0.50	94	100	6
Chromium (Total)	0.00	0.100	0.110	0.114	110	114	4
Cobalt	0.00	0.100	0.109	0.117	109	117	7
Copper	0.00	0.100	0.115	0.106	115	106	8
Lead	0.00	1.00	1.07	1.14	107	114	6
ercury	0.000	0.020	0.018	0.017	90	85	6
olybdenum	0.00	1.00	1.05	0.92	105	92	13
Nickel	0.00	0.50	0.55	0.53	110	106	4
Selenium	0.00	1.00	0.93	0.94	93	94	1
Silver	0.00	1.00	0.92	0.91	92	91	1
Thallium	0.00	3.00	2.90	3.10	97	103	7
Vanadium	0.00	0.50	0.52	0.53	104	106	2
Zinc	0.00	0.100	0.120	0.116	120	116	3

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : PCB 'S (EPA 8080)

Date of Analysis : 05/15/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 9037

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
PCB-1260	0.0	20	11	10	55	50	10

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Organochlorine Pesticides (EPA 8080)

Date of Analysis : 05/15/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 9037

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
4,4'-DDT	0.0	1.00	0.43	0.55	43	55	24

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 05/13/97
Laboratory Sample No : 97050215
Laboratory Reference No : MWI 9037

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	20	18	18	90	90	0
1,1-Dichloroethene	0.0	20	16	16	80	80	0
Trichloroethene	0.0	20	17	17	85	85	0
Toluene	0.0	20	19	18	95	90	5
Chlorobenzene	0.0	20	18	17	90	85	6

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 05/13/97
Laboratory Sample No : OCA 200
Laboratory Reference No : MWI 9037

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	51	50	102	100	2
1,1-Dichloroethene	0.0	50	46	46	92	92	0
Trichloroethene	0.0	50	49	49	98	98	0
Toluene	0.0	50	51	49	102	98	4
Chlorobenzene	0.0	50	49	50	98	100	2

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 05/12/97
Laboratory Sample No : 97050175
Laboratory Reference No : MWI 9037

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0	100	102	102	102	102	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 05/14/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 9037

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	2.5	2.4	2.4	96	96	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 05/10/97
Laboratory Sample No : OCA 200
Laboratory Reference No : MWI 9037

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	44	42	88	84	5
n-Nitroso-di-n-propylamine	0.0	50	41	40	82	80	2
1,2,4-Trichlorobenzene	0.0	50	44	42	88	84	5
Acenaphthene	0.0	50	47	45	94	90	4
Pyrene	0.0	50	40	38	80	76	5
Pentachlorophenol	0.0	100	72	72	72	72	0
4-Chloro-3-Methylphenol	0.0	100	70	70	70	70	0
2-Chlorophenol	0.0	100	84	80	84	80	5
Phenol	0.0	100	72	71	72	71	1

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 05/13/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 9037

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	40	37	80	74	8
n-Nitroso-di-n-propylamine	0.0	50	51	45	102	90	13
1,2,4-Trichlorobenzene	0.0	50	39	40	78	80	3
Acenaphthene	0.0	50	46	45	92	90	2
Pyrene	0.0	50	44	43	88	86	2
entachlorophenol	0.0	100	88	82	88	82	7
4-Chloro-3-Methylphenol	0.0	100	79	79	79	79	0
2-Chlorophenol	0.0	100	81	77	81	77	5
Phenol	0.0	100	35	32	35	32	9

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 05/14/97
Laboratory Sample No : 97050184
Laboratory Reference No : MWI 9037

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	45	45	90	90	0
n-Nitroso-di-n-propylamine	0.0	50	56	55	112	110	2
1,2,4-Trichlorobenzene	0.0	50	45	44	90	88	2
Acenaphthene	0.0	50	47	46	94	92	2
Pyrene	0.0	50	43	43	86	86	0
Pentachlorophenol	0.0	100	85	92	85	92	8
4-Chloro-3-Methylphenol	0.0	100	80	82	80	82	2
2-Chlorophenol	0.0	100	80	81	80	81	1
Phenol	0.0	100	73	75	73	75	3

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532
Tustin, CA 92680
(714) 832-0064, Fax (714) 832-0067

Analysis Request and Chain of Custody Record

Lab Job No: _____
Page 1 of 1

REQUIRED TAT: 24 hr TAT on Soil TRPH and 24 hr ASAP on Soil SVOCs specified [illegible] Standard TAT [illegible]

CUSTOMER INFORMATION		PROJECT INFORMATION	
COMPANY:	Montgomery Watson	PROJECT NAME:	McDonnell Douglas
SEND REPORT TO:	Fred Strauss	NUMBER:	1206038.01090010
ADDRESS:	250 N. Madison Ave.	LOCATION:	C-6 Facility - Bldg. 37 Area
	Pasadena, CA 91101	ADDRESS:	19503 S. Normandie Ave.
			Los Angeles, CA
PHONE:	(818) 568-6582 FAX: (818) 796-5944	SAMPLED BY:	AWN/IV/JAM

SAMPLE ID	NO. OF CONTAINERS	SAMPLE DATE	SAMPLE TIME	SAMPLE MATRIX	CONTAINER TYPE	PRES	TRPH 418.1	Title 22 Metals	VOCs 8260	SVOCs 8270	PCBs 8080	Pesticides 8080			REMARKS/PRECAUTIONS
B37-GS-184-1.5'	2	5/9/97	0915	Soil	3"x6" SS Sleeve	ICE	X	X							
B37-GS-185-6'	2		0930				X	X							
B37-GS-170A-5.5'	2		1015				X	X							
B37-RE4-SP44A	2		1100							X					24 hr or ASAP TAT
B37-RE4-SP44B	2		1110							X					
B37-RE4-SP44C	2		1120							X					
B37-RE4-SP44D	2		1130							X					
B37-RE4-SP53	2		1300				X	X	X	X					
Equipment Blank	7		1415	Water	4 amber 1 Poly 2 VOAs	HNO3 HCl	X	X	X	X	X	X			
Rinsate Blank	7		1430				X	X	X	X	X	X			
Trip Blank	2		—		2 VOAs	HCl			X						

Total No. of Samples: 11

Method of Shipment: 2 coolers via Orange Coast Courier

Relinquished By:	Date/Time:	Received By:	Date/Time:
Adam W. Vonia	5/9/97 1630		
		Received For Lab By: [signature]	05-09-97 4:26 PM

Reporting Format: (check)
NORMAL _____ S.D. HMMD _____
RWQCB _____ OTHER _____

Sample Integrity: (check)
intact _____ on ice _____

All samples remain the property of the client who is responsible for disposal. A disposal fee may be imposed if client fails to pickup samples.

BOE-C6-0143788

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Soil

Laboratory Reference #: MWI 8900

Sampled : 04-14-97
Received: 04-14-97
Analyzed: 04-15-97
Reported: 04-15-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/kg)*
97040608	B37-RE4-SP45	20
Detection Limit:		8.0

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP45

Laboratory Sample Number: 97040608
Laboratory Reference #: MWI 8900

Sampled: 04-14-97
Received: 04-14-97
Analyzed: 04-17-97
Reported: 04-25-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	110	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	22	<---
Cobalt	6010	80	8000	0.5	6.2	<---
Copper	6010	25	2500	0.1	14	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	8.8	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	23	<---
Zinc	6010	250	5000	0.1	40	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

fontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP45
Laboratory Sample Number: 97040608
Laboratory Reference #: MWI 8900

Sampled: 04-14-97
Received: 04-14-97
Analyzed: 04-23-97
Reported: 04-25-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
',1-Dichloroethene	75-35-4	2.5	N.D.
rans 1,2-Dichloroethene	156-60-5	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP45

Laboratory Sample Number: 97040608

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
n-Butylbenzene	104-51-8	2.5	N.D.
1,2-Dichlorobenzene	95-50-1	2.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*109*
Toluene-d8	*96*
4-Bromofluorobenbzene	*77*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ıontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP45
Laboratory Sample #: 97040608
Laboratory Reference #: MWI 8900

Sampled : 04-14-97
Received: 04-14-97
Analyzed: 04-21-97
Reported: 04-25-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
Acenaphthene	83-32-9	100	N.D.	
Acenaphthylene	208-96-8	100	N.D.	
Aniline	62-53-3	100	N.D.	
Anthracene	120-12-7	100	N.D.	
Benzidine	92-87-5	500	N.D.	
Benzoic Acid	65-85-0	250	N.D.	
Benzo(a)anthracene	56-55-3	100	690	<---
ɜnzo(b)fluoranthene	205-99-2	250	790	<---
·ɔnzo(k)fluoranthene	207-08-9	250	350	<---
zo(g,h,i)perylene	191-24-2	250	600	<---
ɜenzo(a)pyrene	50-32-8	250	670	<---
Benzyl alcohol	100-51-6	100	N.D.	
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.	
Bis(2-chloroethyl)ether	111-44-4	100	N.D.	
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.	
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.	
4-Bromophenyl phenyl ether	101-55-3	100	N.D.	
Butyl benzyl phthalate	85-68-7	100	N.D.	
4-Chloroaniline	106-47-8	100	N.D.	
2-Chloronaphthalene	91-58-7	100	N.D.	
4-Chloro-3-methylphenol	59-50-7	100	N.D.	
2-Chlorophenol	95-57-8	100	N.D.	
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.	
Chrysene	218-0109	100	790	<---
Dibenz(a,h)anthracene	53-70-3	100	N.D.	
Dibenzofuran	132-64-9	100	N.D.	
Di-n-butyl phthalate	84-74-2	250	N.D.	
1,3-Dichlorobenzene	541-73-1	100	N.D.	
1,4-Dichlorobenzene	106-46-7	100	N.D.	
1,2-Dichlorobenzene	95-50-1	100	N.D.	
3,3'-Dichlorobenzidine	91-94-1	100	N.D.	
2,4-Dichlorophenol	120-83-2	100	N.D.	
Diethyl phthalate	84-66-2	100	N.D.	
4-Dimethylphenol	105-67-9	100	N.D.	
ethyl phthalate	131-11-3	100	N.D.	
-Dinitro-2-methylphenol	534-52-1	100	N.D.	
2,4-Dinitrophenol	51-28-5	100	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP45
Laboratory Sample #: 97040608

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	990	<---
Fluorene	86-73-7	100	N.D.	
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	610	<---
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	N.D.	
2-Methylphenol	95-48-7	100	N.D.	
4-Methylphenol	106-44-5	100	N.D.	
Naphthalene	91-20-3	100	N.D.	
2-Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	300	<---
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	880	<---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Water

Laboratory Reference #: MWI 8900

Sampled : 04-14-97
Received: 04-14-97
Analyzed: 04-18-97
Reported: 04-25-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/l)*
97040609	Equip Blank	N.D.
97040610	Rinsate Blank	N.D.
Detection Limit:		0.5

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040609
Laboratory Reference #: MWI 8900

Sampled: 04-14-97
Received: 04-14-97
Analyzed: 04-22-97
Reported: 04-25-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040610
Laboratory Reference #: MWI 8900

Sampled: 04-14-97
Received: 04-14-97
Analyzed: 04-22-97
Reported: 04-25-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Equip Blank
Laboratory Sample Number: 97040609
Laboratory Reference #: MWI 8900

Sampled: 04-14-97
Received: 04-14-97
Analyzed: 04-22-97
Reported: 04-25-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/l)***	***SAMPLE RESULT (ug/l)***
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040609

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*98*
Toluene-d8	*101*
4-Bromofluorobenzene	*99*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample Number: 97040610
Laboratory Reference #: MWI 8900

Sampled: 04-14-97
Received: 04-14-97
Analyzed: 04-22-97
Reported: 04-25-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040610

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*103*
Toluene-d8	*104*
4-Bromofluorobenzene	*103*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Trip Blank
Laboratory Sample Number: 97040611
Laboratory Reference #: MWI 8900

Sampled: 04-14-97
Received: 04-14-97
Analyzed: 04-22-97
Reported: 04-25-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Trip Blank

Laboratory Sample Number: 97040611

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*107*
Toluene-d8	*98*
4-Bromofluorobenzene	*97*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040609
Laboratory Reference #: MWI 8900

Sampled : 04-14-97
Received: 04-14-97
Analyzed: 04-16-97
Reported: 04-25-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040609

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
-Methylphenol	95-48-7	5.0	N.D.
Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040610
Laboratory Reference #: MWI 8900

Sampled : 04-14-97
Received: 04-14-97
Analyzed: 04-16-97
Reported: 04-25-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040610

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
-Methylnaphthalene	91-57-6	5.0	N.D.
-Methylphenol	95-48-7	5.0	N.D.
ethylphenol	106-44-5	5.0	N.D.
aphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040609
Laboratory Reference #: MWI 8900

Sampled: 04-14-97
Received: 04-14-97
Analyzed: 04-22-97
Reported: 04-25-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040610
Laboratory Reference #: MWI 8900

Sampled: 04-14-97
Received: 04-14-97
Analyzed: 04-22-97
Reported: 04-25-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040609
Laboratory Reference #: MWI 8900

Sampled: 04-14-97
Received: 04-14-97
Analyzed: 04-21-97
Reported: 04-25-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040610
Laboratory Reference #: MWI 8900

Sampled: 04-14-97
Received: 04-14-97
Analyzed: 04-21-97
Reported: 04-25-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/22/97
Laboratory Sample No : 97040691
Laboratory Reference No : MWI 8900

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	46	46	92	92	0
n-Nitroso-di-n-propylamine	0.0	50	49	47	98	94	4
1,2,4-Trichlorobenzene	0.0	50	48	47	96	94	2
Acenaphthene	0.0	50	52	51	104	102	2
Pyrene	0.0	50	46	35	92	70	27
Pentachlorophenol	0.0	100	77	78	77	78	1
4-Chloro-3-Methylphenol	0.0	100	73	70	73	70	4
2-Chlorophenol	0.0	100	80	79	80	79	1
Phenol	0.0	100	70	70	70	70	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/16/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8900

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	38	44	76	88	15
n-Nitroso-di-n-propylamine	0.0	50	47	34	94	68	32
1,2,4-Trichlorobenzene	0.0	50	38	44	76	88	15
Acenaphthene	0.0	50	47	50	94	100	6
Pyrene	0.0	50	43	46	86	92	7
Pentachlorophenol	0.0	100	78	74	78	74	5
4-Chloro-3-Methylphenol	0.0	100	60	62	60	62	3
2-Chlorophenol	0.0	100	79	81	79	81	3
Phenol	0.0	100	22	24	22	24	9

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/15/97
Laboratory Sample No : 97040511
Laboratory Reference No : MWI 8900

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	100	99	95	99	95	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : PCB 'S (EPA 8080)

Date of Analysis : 04/21/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8900

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
PCB-1260	0.0	20	9.6	7.5	48	38	25

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/18/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8900

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	5.0	3.8	4.0	76	80	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Organochlorine Pesticides (EPA 8080)

Date of Analysis : 04/22/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8900

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
4,4'-DDT	0.0	1.00	0.95	1.00	95	100	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/22/97
Laboratory Sample No : 97040609
Laboratory Reference No : MWI 8900

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	20	18	17	90	85	6
1,1-Dichloroethene	0.0	20	17	17	85	85	0
Trichloroethene	0.0	20	19	19	95	95	0
Toluene	0.0	20	18	18	90	90	0
Chlorobenzene	0.0	20	16	17	80	85	6

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/23/97
Laboratory Sample No : 97040747
Laboratory Reference No : MWI 8900

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	46	44	92	88	4
1,1-Dichloroethene	0.0	50	45	43	90	86	5
Trichloroethene	0.0	50	49	48	98	96	2
Toluene	0.0	50	48	44	96	88	9
Chlorobenzene	0.0	50	45	42	90	84	7

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/17/97
Laboratory Sample No : 97040636
Laboratory Reference No :MWI 8900

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	30.00	30.20	28.80	101	96	5
Arsenic	0.00	10.00	8.20	8.00	82	80	2
Barium	12.00	10.00	22.00	21.00	100	90	5
Beryllium	0.00	1.00	1.03	0.92	103	92	11
Cadmium	0.00	1.00	1.00	1.05	100	105	5
Chromium (Total)	2.60	5.00	7.00	6.60	88	80	6
Chromium (VI)	0.00	1.00	0.87	0.81	87	81	7
Cobalt	0.69	1.00	1.51	1.64	82	95	8
Copper	1.50	1.00	2.30	2.30	80	80	0
Lead	0.00	10.00	9.90	10.20	99	102	3
Mercury	0.000	0.020	0.019	0.021	95	105	10
Molybdenum	0.00	10.00	10.80	10.90	108	109	1
Nickel	1.10	5.00	5.70	5.60	92	90	2
Selenium	0.00	10.00	10.50	9.90	105	99	6
Silver	0.00	5.00	5.20	5.20	104	104	0
Thallium	0.00	30.00	32.00	33.00	107	110	3
Vanadium	2.90	5.00	7.50	7.60	92	94	1
Zinc	4.70	5.00	8.80	8.70	82	80	1

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/22/97
Laboratory Sample No : 97040596
Laboratory Reference No : MWI 8900

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	3.000	2.420	2.700	81	90	11
Arsenic	0.00	1.000	0.800	0.920	80	92	14
Barium	0.00	0.100	0.105	0.104	105	104	1
Beryllium	0.00	0.100	0.085	0.087	85	87	2
Cadmium	0.00	0.100	0.089	0.097	89	97	9
Chromium (VI)	0.00	0.500	0.450	0.490	90	98	9
Chromium (Total)	0.00	0.100	0.087	0.091	87	91	4
Cobalt	0.00	0.100	0.109	0.107	109	107	2
Copper	0.00	0.100	0.080	0.089	80	89	11
Lead	0.00	1.000	1.040	1.010	104	101	3
Mercury	0.000	0.020	0.020	0.022	100	110	10
Molybdenum	0.00	1.000	0.920	1.000	92	100	8
Nickel	0.00	0.500	0.500	0.480	100	96	4
Selenium	0.00	1.000	0.860	0.950	86	95	10
Silver	0.00	0.500	0.440	0.480	88	96	9
Thallium	0.00	3.000	2.460	2.650	82	88	7
Vanadium	0.00	0.500	0.450	0.470	90	94	4
Zinc	0.00	0.100	0.109	0.109	109	109	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532
Tustin, CA 92680
(714) 832-0064, Fax (714) 832-0067

Analysis Request and Chain of Custody Record

Lab Job No: _____
Page 1 of 1

REQUIRED TAT: 24 hr TAT on SOIL TRPH only Standard TAT Otherwise

CUSTOMER INFORMATION		PROJECT INFORMATION	
COMPANY:	Montgomery Watson	PROJECT NAME:	Mc Donnell Douglas
SEND REPORT TO:	Fred Strauss	NUMBER:	1206035.01090010
ADDRESS:	250 N. Madison Ave. Pasadena, CA	LOCATION:	C-6 Facility Bldg. 37
		ADDRESS:	19503 S. Normandie Ave. Los Angeles, CA
PHONE:	(818) 568-6582	FAX:	(818) 796-5941
		SAMPLED BY:	AWN/IV

SAMPLE ID	NO. OF CONTAINERS	SAMPLE DATE	SAMPLE TIME	SAMPLE MATRIX	CONTAINER TYPE	PRES.	TRPH 418.1	Title 22 Metals	VOCs 8260	SVOCs 8270	PCBs 8080	Pesticides 8080	REMARKS/PRECAUTIONS
B37-RE4-SP45	2	4/14/97	1130	SOIL	2"x6" SS sleeve	ICE	X	X	X	X			
Equipment Blank	7		1500	Water	4 amber 1 poly 2 VOAs	HNO3 HCl	X	X	X	X	X	X	
Rinsate Blank	7		1515				X	X	X	X	X	X	
Trip Blank	2		—		2 VOAs	HCl			X				

Total No. of Samples: 4

Method of Shipment: 1 cooler via Orange Coast Courier

Relinquished By:	Date/Time:	Received By:	Date/Time:
Adam W. Nomi	4/14/97 1645		
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received For Lab By:	Date/Time:
		[signature]	4/14/97 16:45

Reporting Format: (check)
NORMAL _____ S.D. HMMD _____
RWQCB _____ OTHER _____

Sample Integrity: (check)
Intact _____ on ice _____

All samples remain the property of the client who is responsible for ... A disposal fee may be imposed if client fails to pickup samples.

BOE-C6-0143822

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Soil

Laboratory Reference #: MWI 8908

Sampled : 04-15-97
Received: 04-15-97
Analyzed: 04-16-97
Reported: 04-16-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/kg)*
97040681	B37-RE4-SP46	85
97040682	B37-RE4-SP47	N.D.
Detection Limit:		8.0

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP46

Laboratory Sample Number: 97040681
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-17-97
Reported: 04-28-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	110	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	22	<---
Cobalt	6010	80	8000	0.5	5.9	<---
Copper	6010	25	2500	0.1	14	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	9.2	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	23	<---
Zinc	6010	250	5000	0.1	44	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP47

Laboratory Sample Number: 97040682
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-17-97
Reported: 04-28-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	100	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	24	<---
Cobalt	6010	80	8000	0.5	6.7	<---
Copper	6010	25	2500	0.1	12	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	9.2	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	24	<---
Zinc	6010	250	5000	0.1	36	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP46
Laboratory Sample Number: 97040681
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-23-97
Reported: 04-28-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
1,1-Dichloroethene	75-35-4	2.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ample Description: Soil, B3.7-RE4-SP46

Laboratory Sample Number: 97040681

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
n-Butylbenzene	104-51-8	2.5	N.D.
.2-Dichlorobenzene	95-50-1	2.5	N.D.
-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Mark Noorani
boratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*109*
Toluene-d8	*99*
4-Bromofluorobenbzene	*83*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP47
Laboratory Sample Number: 97040682
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-23-97
Reported: 04-28-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
1,1-Dichloroethene	75-35-4	2.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ample Description: Soil, B37-RE4-SP47

Laboratory Sample Number: 97040682

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
n-Butylbenzene	104-51-8	2.5	N.D.
1,2-Dichlorobenzene	95-50-1	2.5	N.D.
-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
aboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*108*
Toluene-d8	*103*
4-Bromofluorobenbzene	*80*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP46
Laboratory Sample #: 97040681
Laboratory Reference #: MWI 8908

Sampled : 04-15-97
Received: 04-15-97
Analyzed: 04-22-97
Reported: 04-28-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
Benzo(k)fluoranthene	207-08-9	250	N.D.
Benzo(g,h,i)perylene	191-24-2	250	N.D.
Benzo(a)pyrene	50-32-8	250	N.D.
Benzyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
Dimethyl phthalate	131-11-3	100	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.
2,4-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP46
Laboratory Sample #: 97040681

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	150	<---
Fluorene	86-73-7	100	N.D.	
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.	
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	N.D.	
-Methylphenol	95-48-7	100	N.D.	
Methylphenol	106-44-5	100	N.D.	
ıthalene	91-20-3	100	N.D.	
Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	140	<---
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	200	<---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

k Noorani
ɔratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP47
Laboratory Sample #: 97040682
Laboratory Reference #: MWI 8908

Sampled : 04-15-97
Received: 04-15-97
Analyzed: 04-22-97
Reported: 04-28-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
Benzo(k)fluoranthene	207-08-9	250	N.D.
Benzo(g,h,i)perylene	191-24-2	250	N.D.
Benzo(a)pyrene	50-32-8	250	N.D.
Benzyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
Dimethyl phthalate	131-11-3	100	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.
2,4-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP47
Laboratory Sample #: 97040682

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
2,4-Dinitrotoluene	121-14-2	250	N.D.
2,6-Dinitrotoluene	606-20-2	250	N.D.
Di-n-octyl phthalate	117-84-0	250	N.D.
Fluoranthene	206-44-0	100	N.D.
Fluorene	86-73-7	100	N.D.
Hexachlorobenzene	118-74-1	100	N.D.
Hexachlorobutadiene	87-68-3	100	N.D.
Hexachlorocyclopentadiene	77-47-4	100	N.D.
Hexachloroethane	67-72-1	100	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.
Isophorone	78-59-1	100	N.D.
2-Methylnaphthalene	91-57-6	100	N.D.
-Methylphenol	95-48-7	100	N.D.
Methylphenol	106-44-5	100	N.D.
hthalene	91-20-3	100	N.D.
Nitroaniline	88-74-4	250	N.D.
3-Nitroaniline	99-09-2	250	N.D.
4-Nitroaniline	100-01-6	250	N.D.
Nitrobenzene	98-95-3	100	N.D.
2-Nitrophenol	88-75-5	100	N.D.
4-Nitrophenol	100-02-7	100	N.D.
N-Nitrosodiphenylamine	86-30-6	100	N.D.
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.
N-Nitrosodimethylamine	62-75-9	100	N.D.
Pentachlorophenol	87-86-5	250	N.D.
Phenanthrene	85-01-8	100	N.D.
Phenol	108-95-2	100	N.D.
Pyrene	129-00-0	100	N.D.-
1,2,4-Trichlorobenzene	120-82-1	100	N.D.
2,4,5-Trichlorophenol	95-95-4	100	N.D.
2,4,6-Trichlorophenol	88-06-2	100	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

k Noorani
oratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Water

Laboratory Reference #: MWI 8908

Sampled : 04-15-97
Received: 04-15-97
Analyzed: 04-18-97
Reported: 04-28-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/l)*
97040683	Equip Blank	N.D.
97040684	Rinsate Blank	N.D.
Detection Limit:		0.5

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Equip Blank
Laboratory Sample Number: 97040683
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-22-97
Reported: 04-28-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040683

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*109*
Toluene-d8	*97*
4-Bromofluorobenzene	*99*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss,
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample Number: 97040684
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-22-97
Reported: 04-28-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040684

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*105*
Toluene-d8	*101*
4-Bromofluorobenzene	*102*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Trip Blank
Laboratory Sample Number: 97040685
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-22-97
Reported: 04-28-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Trip Blank

Laboratory Sample Number: 97040685

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*107*
Toluene-d8	*98*
4-Bromofluorobenzene	*96*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
ATTN: Mr. Fred Strauss
50 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040683
Laboratory Reference #: MWI 8908

Sampled : 04-15-97
Received: 04-15-97
Analyzed: 04-23-97
Reported: 04-28-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
zo(a)pyrene	50-32-8	25	N.D.
nzyl alcohol	100-51-6	50	N.D.
is(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4 -Dinitro-2-methylphenol	534-52-1	50	N.D.
-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040683

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
4-Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
.TN: Mr. Fred Strauss
50 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sampled : 04-15-97
Sample Description: Water, Rinsate Blank *Received:* 04-15-97
Laboratory Sample #: 97040684 *Analyzed:* 04-23-97
Laboratory Reference #: MWI 8908 *Reported:* 04-28-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
zo(a)pyrene	50-32-8	25	N.D.
.nzyl alcohol	100-51-6	50	N.D.
is(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4 -Dinitro-2-methylphenol	534-52-1	50	N.D.
-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040684

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
4-Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040683
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-22-97
Reported: 04-28-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040684
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-22-97
Reported: 04-28-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040683
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-21-97
Reported: 04-28-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040684
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-21-97
Reported: 04-28-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040683
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-22-97
Reported: 04-28-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040684
Laboratory Reference #: MWI 8908

Sampled: 04-15-97
Received: 04-15-97
Analyzed: 04-22-97
Reported: 04-28-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Organochlorine Pesticides (EPA 8080)

Date of Analysis : 04/22/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8908

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
4,4'-DDT	0.0	1.00	0.95	1.00	95	100	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : PCB 'S (EPA 8080)

Date of Analysis : 04/21/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8908

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
PCB-1260	0.0	20	9.6	7.5	48	38	25

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/22/97
Laboratory Sample No : 97040609
Laboratory Reference No : MWI 8908

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	20	18	17	90	85	6
1,1-Dichloroethene	0.0	20	17	17	85	85	0
Trichloroethene	0.0	20	19	19	95	95	0
Toluene	0.0	20	18	18	90	90	0
Chlorobenzene	0.0	20	16	17	80	85	6

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/23/97
Laboratory Sample No : 97040747
Laboratory Reference No : MWI 8908

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	46	44	92	88	4
1,1-Dichloroethene	0.0	50	45	43	90	86	5
Trichloroethene	0.0	50	49	49	98	98	0
Toluene	0.0	50	48	43	54	86	11
Chlorobenzene	0.0	50	45	42	51	84	7

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/22/97
Laboratory Sample No : 97040596
Laboratory Reference No : MWI 8908

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	3.000	2.420	2.700	81	90	11
Arsenic	0.00	1.000	0.800	0.920	80	92	14
Barium	0.00	0.100	0.105	0.104	105	104	1
Beryllium	0.00	0.100	0.085	0.087	85	87	2
Cadmium	0.00	0.100	0.089	0.097	89	97	9
Chromium (VI)	0.00	0.500	0.450	0.490	90	98	9
Chromium (Total)	0.00	0.100	0.087	0.091	87	91	4
Cobalt	0.00	0.100	0.109	0.107	109	107	2
Copper	0.00	0.100	0.080	0.089	80	89	11
Lead	0.00	1.000	1.040	1.010	104	101	3
Mercury	0.000	0.020	0.020	0.022	100	110	10
Molybdenum	0.00	1.000	0.920	1.000	92	100	8
Nickel	0.00	0.500	0.500	0.480	100	96	4
Selenium	0.00	1.000	0.860	0.950	86	95	10
Silver	0.00	0.500	0.440	0.480	88	96	9
Thallium	0.00	3.000	2.460	2.650	82	88	7
Vanadium	0.00	0.500	0.450	0.470	90	94	4
Zinc	0.00	0.100	0.109	0.109	109	109	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/17/97
Laboratory Sample No : 97040693
Laboratory Reference No : MWI 8908

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	30.00	29.00	29.00	97	97	0
Arsenic	0.00	10.00	8.40	8.10	84	81	4
Barium	18.00	10.00	26.20	26.40	82	84	1
Beryllium	0.00	1.00	1.02	0.99	102	99	3
Cadmium	0.00	1.00	0.97	0.97	97	97	0
Chromium (Total)	2.50	5.00	7.60	7.00	102	90	8
Chromium (VI)	0.00	1.00	1.03	0.90	103	90	13
Cobalt	0.71	1.00	1.72	1.87	101	116	8
Copper	1.30	1.00	2.20	2.35	90	105	7
Lead	0.00	10.00	10.80	9.60	108	96	12
Mercury	0.000	0.020	0.022	0.021	110	105	5
Molybdenum	0.00	10.00	9.00	8.60	90	86	5
Nickel	1.40	5.00	6.30	5.70	98	86	10
Selenium	0.00	10.00	9.50	9.20	95	92	3
Silver	0.00	5.00	4.90	4.80	98	96	2
Thallium	0.00	30.00	25.00	27.00	83	90	8
Vanadium	3.10	5.00	7.70	8.10	92	100	5
Zinc	4.30	5.00	10.00	9.40	114	102	6

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/18/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8908

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	5.0	3.8	4.0	76	80	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/16/97
Laboratory Sample No : 97040511
Laboratory Reference No : MWI 8908

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	100	95	96	95	96	1

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/22/97
Laboratory Sample No : 97040691
Laboratory Reference No : MWI 8908

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	46	46	92	92	0
n-Nitroso-di-n-propylamine	0.0	50	49	47	98	94	4
1,2,4-Trichlorobenzene	0.0	50	48	47	96	94	2
Acenaphthene	0.0	50	52	51	104	102	2
Pyrene	0.0	50	46	35	92	70	27
Pentachlorophenol	0.0	100	77	78	77	78	1
4-Chloro-3-Methylphenol	0.0	100	73	70	73	70	4
2-Chlorophenol	0.0	100	80	79	80	79	1
Phenol	0.0	100	70	70	70	70	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/22/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8908

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	37	43	74	86	15
n-Nitroso-di-n-propylamine	0.0	50	47	50	94	100	6
1,2,4-Trichlorobenzene	0.0	50	39	45	78	90	14
Acenaphthene	0.0	50	47	51	94	102	8
Pyrene	0.0	50	40	46	80	92	14
Pentachlorophenol	0.0	100	80	80	80	80	0
4-Chloro-3-Methylphenol	0.0	100	65	67	65	67	3
2-Chlorophenol	0.0	100	80	83	80	83	4
Phenol	0.0	100	23	26	23	26	12

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532
Tustin, CA 92680
(714) 832-0064, Fax (714) 832-0067

Analysis Request and Chain of Custody Record

Lab Job No: _____
Page 1 of 1

REQUIRED TAT: 24 hr TAT on Soil TRPH only! Standard TAT otherwise

CUSTOMER INFORMATION

COMPANY: Montgomery Watson
SEND REPORT TO: Fred Strauss
ADDRESS: 250 N. Madison Ave. Pasadena, CA
PHONE: (818) 568-6582 FAX: (818) 796-5941

PROJECT INFORMATION

PROJECT NAME: Mc Donnell Douglas
NUMBER: 1206035.01090010
LOCATION: C-6 Facility - Bldg 37 Area
ADDRESS: 19503 S. Normandie Ave. Los Angeles, CA
SAMPLED BY: AWN/EV

SAMPLE ID	NO. OF CONTAINERS	SAMPLE DATE	SAMPLE TIME	SAMPLE MATRIX	CONTAINER TYPE	PRES.	TRPH 418.1	Title 22 Metals	VOCs 8260	SVOCs 8270	PCBs 8080	Pesticides 8080			REMARKS/PRECAUTIONS
B37-RE4-SP46	2	4/15/97	0940	Soil	2"x6" SS Sleeve	ICE	X	X	X	X					
B37-RE4-SP47	2	↓	1300	↓	↓	↓	X	X	X	X					
Equipment Blank	7	↓	1530	Water	Amber 1 poly 2 VOAs	HNO_3 HCl	X	X	X	X	X	X			
Rinsate Blank	7	↓	1545	↓	↓	↓	X	X	X	X	X	X			
Trip Blank	2	↓	—	↓	2 VOAs	HCl			X						

Total No. of Samples: 5

Method of Shipment: 1 cooler via Orange Coast Courier

Relinquished By:	Date/Time:	Received By:	Date/Time:
Adam W. Vernia	4/15/97 1625		
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received For Lab By: [signature]	Date/Time: 4/15/97 16:25

Reporting Format: (check)
NORMAL _____ S.D. HMMD _____
RWQCB _____ OTHER _____

Sample Integrity: (check)
Intact _____ on Ice _____

All samples remain the property of the client who is responsible for disposal. A disposal fee may be imposed if client fails to pickup samples.

BOE-C6-0143861

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Descript : Soil

Laboratory Reference #: MWI 8916

Sampled : 04-16-97
Received: 04-16-97
Analyzed: 04-17-97
Reported: 04-17-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	Client Sample Number	Sample Result (mg/kg)
9704074	B37-RE4-SP48	23
9704074	B37-RE4-SP49	1,700
9704074	B37-GS-134-5.5'	N.D.
9704074	B37-GS-135-5.5'	N.D.
Detection Limit:		8.0

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP48

Laboratory Sample Number: 97040744
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 05-01-97
Reported: 05-05-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	100	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	33	<---
Cobalt	6010	80	8000	0.5	7.6	<---
Copper	6010	25	2500	0.1	12	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	10	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	29	<---
Zinc	6010	250	5000	0.1	48	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP49

Laboratory Sample Number: 97040745
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 05-01-97
Reported: 05-05-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	110	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	N.D.	
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	22	<---
Cobalt	6010	80	8000	0.5	6.9	<---
Copper	6010	25	2500	0.1	11	<---
Lead	6010	5.0	1000	1.0	N.D.	
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	9.9	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	25	<---
Zinc	6010	250	5000	0.1	41	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP48
Laboratory Sample Number: 97040744
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 04-23-97
Reported: 05-05-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/kg)***	***SAMPLE RESULT (ug/kg)***
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
1,1-Dichloroethene	75-35-4	2.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP48

Laboratory Sample Number: 97040744

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
n-Butylbenzene	104-51-8	2.5	N.D.
',2-Dichlorobenzene	95-50-1	2.5	N.D.
,-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
aboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*105*
Toluene-d8	*99*
4-Bromofluorobenbzene	*90*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP49
Laboratory Sample Number: 97040745
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 04-23-97
Reported: 05-05-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/kg)	SAMPLE RESULT (ug/kg)	
Benzene	71-43-2	50	N.D.	
Bromodichloromethane	75-27-4	50	N.D.	
Bromoform	75-25-2	50	N.D.	
Bromomethane	74-83-9	50	N.D.	
Carbon Disulfide	75-15-0	100	N.D.	
Carbon tetrachloride	56-23-5	50	N.D.	
Chlorobenzene	108-90-7	50	N.D.	
Chlorodibromomethane	124-48-1	50	N.D.	
Chloroethane	75-00-3	50	N.D.	
2-Chloroethyl vinyl ether	110-75-8	100	N.D.	
Chloroform	67-66-3	50	N.D.	
Chloromethane	74-87-3	50	N.D.	
1,1-Dichloroethane	75-35-3	50	N.D.	
1,2-Dichloroethane	107-06-2	50	N.D.	
1,1-Dichloroethene	75-35-4	50	N.D.	
Trans 1,2-Dichloroethene	156-60-5	50	N.D.	
1,2-Dichloropropane	78-87-5	50	N.D.	
cis-1,3-Dichloropropene	10061-01-5	50	N.D.	
trans-1,3-Dichloropropene	10061-02-6	50	N.D.	
Ethylbenzene	100-41-4	50	N.D.	
Methylene chloride	75-09-2	100	N.D.	
Styrene	100-42-5	50	N.D.	
1,1,2,2-Tetrachloroethane	79-34-5	50	N.D.	
Tetrachloroethene	127-18-4	50	N.D.	
Toluene	108-88-3	50	N.D.	
1,1,1-Trichloroethane	71-55-6	50	N.D.	
1,1,2-Trichloroethane	79-00-5	50	N.D.	
Trichloroethene	79-01-6	50	N.D.	
Trichlorofluoromethane	75-69-4	100	N.D.	
Vinyl acetate	108-05-4	100	N.D.	
Vinyl chloride	75-01-4	50	N.D.	
Total Xylenes	1330-20-7	50	120	<---
Dichlorofluoromethane	75-71-8	50	N.D.	
cis-1-2,-Dichloroethane	156-59-4	50	N.D.	
2,2-Dichloropropane	590-20-7	50	N.D.	
Bromochloromethane	74-97-5	50	N.D.	
1,1-Dichloropropene	563-58-6	50	N.D.	
1,2-Dichloropropane	78-87-5	50	N.D.	
Dibromomethane	74-95-3	50	N.D.	
trans-1,3-Dichloropropane	10061-02-6	50	N.D.	
1,1,2-Trichloroethane	79-00-5	50	N.D.	
1,2-Dibromoethane	106-93-4	50	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ımple Description: Soil, B37-RE4-SP49

Laboratory Sample Number: 97040745

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/kg)	SAMPLE RESULT (ug/kg)	
1,3-Dichloropropane	142-28-9	50	N.D.	
Dibromochloromethane	124-48-1	50	N.D.	
Isopropylbenzene	98-82-8	50	N.D.	
1,1,2,2-Tetrachloroethane	79-34-5	50	N.D.	
1,2,3-Trichloropropane	96-18-4	50	N.D.	
Bromobenzene	108-86-1	50	N.D.	
n-Propylbenzene	103-65-1	50	58	<---
2-Chlorotoluene	95-49-8	50	N.D.	
1,3,5-Trimethylbenzene	108-67-8	50	98	<---
4-Chlorotoluene	106-43-4	50	N.D.	
tert-Butylbenzene	98-06-6	50	N.D.	
1,2,4-Trimethylbenzene	95-63-6	50	N.D.	
sec-Butylbenzene	135-98-8	50	N.D.	
4-Isopropyltoluene	99-87-6	50	N.D.	
1,3-Dichlorobenzene	541-73-1	50	N.D.	
1,4-Dichlorobenzene	106-46-7	50	N.D.	
n-Butylbenzene	104-51-8	50	74	<---
,2-Dichlorobenzene	95-50-1	50	N.D.	
-2-Dibromo-2-CPA	96-12-8	100	N.D.	
1,2,4-Trichlorobenzene	120-82-1	50	130	<---
Hexachlorobutadiene	87-68-3	50	N.D.	
Naphthalene	91-20-3	50	280	<---
1,2,3-Trichlorobenzene	87-61-6	50	170	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
ıboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*119*
Toluene-d8	*97*
4-Bromofluorobenbzene	*109*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP48
Laboratory Sample #: 97040744
Laboratory Reference #: MWI 8916

Sampled : 04-16-97
Received: 04-16-97
Analyzed: 04-22-97
Reported: 05-05-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	CAS NUMBER	DETECTION LIMIT ug/kg	SAMPLE RESULTS ug/kg
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
Benzo(k)fluoranthene	207-08-9	250	N.D.
Benzo(g,h,i)perylene	191-24-2	250	N.D.
Benzo(a)pyrene	50-32-8	250	N.D.
Benzyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
Dimethyl phthalate	131-11-3	100	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.
2,4-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP48
Laboratory Sample #: 97040744

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	110	<---
Fluorene	86-73-7	100	N.D.	
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.	
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	N.D.	
2-Methylphenol	95-48-7	100	N.D.	
ethylphenol	106-44-5	100	N.D.	
phthalene	91-20-3	100	N.D.	
-Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	N.D.	
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	N.D.	
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

k Noorani
aboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP49
Laboratory Sample #: 97040745
Laboratory Reference #: MWI 8916

Sampled : 04-16-97
Received: 04-16-97
Analyzed: 04-22-97
Reported: 05-05-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
Acenaphthene	83-32-9	100	160	<---
Acenaphthylene	208-96-8	100	N.D.	
Aniline	62-53-3	100	N.D.	
Anthracene	120-12-7	100	N.D.	
Benzidine	92-87-5	500	N.D.	
Benzoic Acid	65-85-0	250	N.D.	
Benzo(a)anthracene	56-55-3	100	N.D.	
Benzo(b)fluoranthene	205-99-2	250	270	<---
Benzo(k)fluoranthene	207-08-9	250	N.D.	
Benzo(g,h,i)perylene	191-24-2	250	N.D.	
Benzo(a)pyrene	50-32-8	250	N.D.	
Benzyl alcohol	100-51-6	100	N.D.	
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.	
Bis(2-chloroethyl)ether	111-44-4	100	N.D.	
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.	
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.	
4-Bromophenyl phenyl ether	101-55-3	100	N.D.	
Butyl benzyl phthalate	85-68-7	100	N.D.	
4-Chloroaniline	106-47-8	100	N.D.	
2-Chloronaphthalene	91-58-7	100	N.D.	
4-Chloro-3-methylphenol	59-50-7	100	N.D.	
2-Chlorophenol	95-57-8	100	N.D.	
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.	
Chrysene	218-0109	100	350	<---
Dibenz(a,h)anthracene	53-70-3	100	N.D.	
Dibenzofuran	132-64-9	100	N.D.	
Di-n-butyl phthalate	84-74-2	250	N.D.	
1,3-Dichlorobenzene	541-73-1	100	N.D.	
1,4-Dichlorobenzene	106-46-7	100	N.D.	
1,2-Dichlorobenzene	95-50-1	100	N.D.	
3,3'-Dichlorobenzidine	91-94-1	100	N.D.	
2,4-Dichlorophenol	120-83-2	100	N.D.	
Diethyl phthalate	84-66-2	100	N.D.	
2,4-Dimethylphenol	105-67-9	100	N.D.	
Dimethyl phthalate	131-11-3	100	N.D.	
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.	
2,4-Dinitrophenol	51-28-5	100	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP49
Laboratory Sample #: 97040745

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	330	<---
Fluorene	86-73-7	100	610	<---
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.	
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	540	<---
2-Methylphenol	95-48-7	100	N.D.	
ethylphenol	106-44-5	100	N.D.	
phthalene	91-20-3	100	310	<---
-Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	1,500	<---
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	200	<---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

c Noorani
boratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Water

Laboratory Reference #: MWI 8916

Sampled : 04-16-97
Received: 04-16-97
Analyzed: 04-22-97
Reported: 05-05-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/l)*
97040748	Equip Blank	N.D.
97040749	Rinsate Blank	N.D.
Detection Limit:		0.5

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040748
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 04-22-97
Reported: 05-05-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040749
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 04-22-97
Reported: 05-05-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Equip Blank
Laboratory Sample Number: 97040748
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 04-22-97
Reported: 05-05-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/l)***	***SAMPLE RESULT (ug/l)***
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040748

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*102*
Toluene-d8	*99*
4-Bromofluorobenzene	*95*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample Number: 97040749
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 04-22-97
Reported: 05-05-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040749

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*114*
Toluene-d8	*99*
4-Bromofluorobenzene	*100*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Trip Blank
Laboratory Sample Number: 97040750
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 04-22-97
Reported: 05-05-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Trip Blank

Laboratory Sample Number: 97040750

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*106*
Toluene-d8	*100*
4-Bromofluorobenzene	*92*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ıtgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040748
Laboratory Reference #: MWI 8916

Sampled : 04-16-97
Received: 04-16-97
Analyzed: 04-23-97
Reported: 05-05-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
zo(g,h,i)perylene	191-24-2	25	N.D.
ızo(a)pyrene	50-32-8	25	N.D.
ınzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
·Dinitro-2-methylphenol	534-52-1	50	N.D.
4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040748

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
4-Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ıtgomery Watson
:TN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040749
Laboratory Reference #: MWI 8916

Sampled : 04-16-97
Received: 04-16-97
Analyzed: 04-23-97
Reported: 05-05-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
[illegible]nzo(g,h,i)perylene	191-24-2	25	N.D.
[illegible]zo(a)pyrene	50-32-8	25	N.D.
[illegible]nzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
[illegible]-Dinitro-2-methylphenol	534-52-1	50	N.D.
[illegible]-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040749

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
4-Methylphenol	106-44-5	5.0	N.D.
Naphthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040748
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 04-21-97
Reported: 05-05-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040749
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 04-21-97
Reported: 05-05-97

PCB'S (EPA 8080)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULTS (ug/l)
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040748
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 04-22-97
Reported: 05-05-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040749
Laboratory Reference #: MWI 8916

Sampled: 04-16-97
Received: 04-16-97
Analyzed: 04-22-97
Reported: 05-05-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Organochlorine Pesticides (EPA 8080)

Date of Analysis : 04/22/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8916

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
4,4'-DDT	0.0	1.00	0.95	1.00	95	100	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : PCB 'S (EPA 8080)

Date of Analysis : 04/21/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8916

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
PCB-1260	0.0	20	9.6	7.5	48	38	25

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/22/97
Laboratory Sample No : 97040691
Laboratory Reference No : MWI 8916

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	46	46	92	92	0
n-Nitroso-di-n-propylamine	0.0	50	49	47	98	94	4
1,2,4-Trichlorobenzene	0.0	50	48	47	96	94	2
Acenaphthene	0.0	50	52	51	104	102	2
Pyrene	0.0	50	46	35	92	70	27
Pentachlorophenol	0.0	100	77	78	77	78	1
4-Chloro-3-Methylphenol	0.0	100	73	70	73	70	4
2-Chlorophenol	0.0	100	80	79	80	79	1
Phenol	0.0	100	70	70	70	70	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/22/97
Laboratory Sample No : 97040700
Laboratory Reference No : MWI 8916

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	43	43	86	86	0
n-Nitroso-di-n-propylamine	0.0	50	47	35	94	70	29
1,2,4-Trichlorobenzene	0.0	50	44	46	88	92	4
Acenaphthene	0.0	50	47	47	94	94	0
Pyrene	0.0	50	44	43	88	86	2
Pentachlorophenol	0.0	100	74	71	74	71	4
4-Chloro-3-Methylphenol	0.0	100	64	64	64	64	0
2-Chlorophenol	0.0	100	79	79	79	79	0
Phenol	0.0	100	72	71	72	71	1

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/23/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8916

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	37	43	74	86	15
n-Nitroso-di-n-propylamine	0.0	50	47	50	94	100	6
1,2,4-Trichlorobenzene	0.0	50	39	45	78	90	14
Acenaphthene	0.0	50	47	51	94	102	8
Pyrene	0.0	50	40	46	80	92	14
Pentachlorophenol	0.0	100	80	80	80	80	0
4-Chloro-3-Methylphenol	0.0	100	65	67	65	67	3
2-Chlorophenol	0.0	100	80	83	80	83	4
Phenol	0.0	100	23	26	23	26	12

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/22/97
Laboratory Sample No : 97040609
Laboratory Reference No : MWI 8916

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	20	18	17	90	85	6
1,1-Dichloroethene	0.0	20	17	17	85	85	0
Trichloroethene	0.0	20	19	19	95	95	0
Toluene	0.0	20	18	18	90	90	0
Chlorobenzene	0.0	20	16	17	80	85	6

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/23/97
Laboratory Sample No : 97040747
Laboratory Reference No : MWI 8916

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	46	44	92	88	4
1,1-Dichloroethene	0.0	50	45	43	90	86	5
Trichloroethene	0.0	50	49	48	98	96	2
Toluene	0.0	50	48	44	54	88	9
Chlorobenzene	0.0	50	45	42	51	84	7

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/22/97
Laboratory Sample No : 97040596
Laboratory Reference No : MWI 8916

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	3.000	2.420	2.700	81	90	11
Arsenic	0.00	1.000	0.800	0.920	80	92	14
Barium	0.00	0.100	0.105	0.104	105	104	1
Beryllium	0.00	0.100	0.085	0.087	85	87	2
Cadmium	0.00	0.100	0.089	0.097	89	97	9
Chromium (VI)	0.00	0.500	0.450	0.490	90	98	9
Chromium (Total)	0.00	0.100	0.087	0.091	87	91	4
Cobalt	0.00	0.100	0.109	0.107	109	107	2
Copper	0.00	0.100	0.080	0.089	80	89	11
Lead	0.00	1.000	1.040	1.010	104	101	3
Mercury	0.000	0.020	0.020	0.022	100	110	10
Molybdenum	0.00	1.000	0.920	1.000	92	100	8
Nickel	0.00	0.500	0.500	0.480	100	96	4
Selenium	0.00	1.000	0.860	0.950	86	95	10
Silver	0.00	0.500	0.440	0.480	88	96	9
Thallium	0.00	3.000	2.460	2.650	82	88	7
Vanadium	0.00	0.500	0.450	0.470	90	94	4
Zinc	0.00	0.100	0.109	0.109	109	109	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 05/01/97
Laboratory Sample No : 97040939
Laboratory Reference No : MWI 8916

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	30.00	29.00	29.00	97	97	0
Arsenic	0.00	10.00	10.70	11.00	107	110	3
Barium	7.60	10.00	18.40	18.90	108	113	3
Beryllium	0.00	1.00	0.93	0.99	93	99	6
Cadmium	0.00	1.00	1.06	1.15	106	115	8
Chromium (Total)	1.80	5.00	6.90	6.90	102	102	0
Chromium (VI)	0.00	1.00	0.87	0.95	87	95	9
Cobalt	0.54	1.00	1.55	1.62	101	108	4
Copper	0.48	1.00	1.38	1.52	90	104	10
Lead	0.00	10.00	9.90	10.10	99	101	2
Mercury	0.000	0.020	0.017	0.019	85	95	11
Molybdenum	0.00	10.00	10.20	10.00	102	100	2
Nickel	0.68	5.00	5.50	5.70	96	100	4
Selenium	0.00	10.00	10.40	10.10	104	101	3
Silver	0.00	5.00	4.70	5.00	94	100	6
Thallium	0.00	30.00	26.00	30.00	87	100	14
Vanadium	2.00	5.00	6.90	7.00	98	100	1
Zinc	3.30	5.00	8.10	8.70	96	108	7

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/22/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8916

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	2.5	2.2	2.1	88	84	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/17/97
Laboratory Sample No : 97040746
Laboratory Reference No : MWI 8916

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	100	101	99	101	99	2

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532
Tustin, CA 92680
(714) 832-0064, Fax (714) 832-0067

Analysis Request and Chain of Custody Record

Lab Job No: _____
Page 1 of 1

REQUIRED TAT: 24 hr TAT on Soil TRPH on Kaj; Standard TAT otherwise

CUSTOMER INFORMATION	PROJECT INFORMATION
COMPANY: Montgomery Watson	PROJECT NAME: Mc Donnell Douglas
SEND REPORT TO: Fred Strauss	NUMBER: 1206035.01090010
ADDRESS: 250 N. Madison Ave. Pasadena, CA 91101	LOCATION: C-6 Facility - Bldg. 37 Area
	ADDRESS: 19503 S. Normandie Ave. Los Angeles, CA
PHONE: (818) 568-6582 FAX: (818) 796-5941	SAMPLED BY: AWN/IV

ANALYSIS/METHOD REQUEST

SAMPLE ID	NO. OF CONTAINERS	SAMPLE DATE	SAMPLE TIME	SAMPLE MATRIX	CONTAINER TYPE	PRES.	TRPH 418.1	Title 22 Metals	VOCs 8260	SVOCs 8270	PCBs 8080	Pesticides 8080			REMARKS/PRECAUTIONS
B37-RE4-SP48	2	4/16/97	0850	Soil	2"x6" SS sleeve	ICE	X	X	X	X					
B37-RE4-SP49	2		1145				X	X	X	X					
B37-GS-134-5.5'	2		1328				X	X							
B37-GS-135-5.5'	2		1337				X	X							
Equipment Blank	7		1530	Water	Amber 1 poly 2 VOAs	HNO3 HCl	X	X	X	X	X	X			
Rinsate Blank	7		1545				X	X	X	X	X	X			
Trip Blank	2		—		2 VOAs	HCl			X						

Total No. of Samples: 7

Method of Shipment: 1 cooler via Orange Coast Courier

Relinquished By:	Date/Time:	Received By:	Date/Time:
Adam W. Nomi	4/16/97 1700		
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received For Lab By:	Date/Time:
		[signature]	4/16/97 17:00

Reporting Format: (check)
NORMAL _____ S.D. HMMD _____
RWQCB _____ OTHER _____

Sample Integrity: (check)
Intact _____ on Ice _____

All samples remain the property of the client who is responsible for disposal. A disposal fee may be imposed if client fails to pickup samples.

BOE-C6-0143901

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Soil

Laboratory Reference #: MWI 8924

Sampled : 04-17-97
Received: 04-17-97
Analyzed: 04-18-97
Reported: 04-18-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/kg)*
97040816	B37-RE4-SP50	320
Detection Limit:		8.0

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP50

Laboratory Sample Number: 97040816
Laboratory Reference #: MWI 8924

Sampled: 04-17-97
Received: 04-17-97
Analyzed: 04-21-97
Reported: 04-28-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	100	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	8.0	<---
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	63	<---
Cobalt	6010	80	8000	0.5	8.5	<---
Copper	6010	25	2500	0.1	29	<---
Lead	6010	5.0	1000	1.0	17	<---
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	15	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	32	<---
Zinc	6010	250	5000	0.1	99	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP50
Laboratory Sample Number: 97040816
Laboratory Reference #: MWI 8924

Sampled: 04-17-97
Received: 04-17-97
Analyzed: 04-25-97
Reported: 04-28-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
',1-Dichloroethene	75-35-4	2.5	N.D.
rans 1,2-Dichloroethene	156-60-5	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP50

Laboratory Sample Number: 97040816

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/kg)*	*SAMPLE RESULT (ug/kg)*
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
n-Butylbenzene	104-51-8	2.5	N.D.
1,2-Dichlorobenzene	95-50-1	2.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*113*
Toluene-d8	*95*
4-Bromofluorobenbzene	*98*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP50
Laboratory Sample #: 97040816

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	150	<---
Fluorene	86-73-7	100	N.D.	
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.	
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	N.D.	
2-Methylphenol	95-48-7	100	N.D.	
'ethylphenol	106-44-5	100	N.D.	
phthalene	91-20-3	100	N.D.	
2-Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	N.D.	
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	170	<---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

: Noorani
aboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP34A
Laboratory Sample #: 97040817
Laboratory Reference #: MWI 8924

Sampled : 04-17-97
Received: 04-17-97
Analyzed: 04-18-97
Reported: 04-18-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
Benzo(k)fluoranthene	207-08-9	250	N.D.
Benzo(g,h,i)perylene	191-24-2	250	N.D.
Benzo(a)pyrene	50-32-8	250	N.D.
Benzyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	120 <---
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
Dimethyl phthalate	131-11-3	100	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.
2,4-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Water

Laboratory Reference #: MWI 8924

Sampled : 04-17-97
Received: 04-17-97
Analyzed: 04-22-97
Reported: 04-28-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/l)*
97040822	Equip Blank	N.D.
97040823	Rinsate Blank	N.D.
Detection Limit:		0.5

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040822
Laboratory Reference #: MWI 8924

Sampled: 04-17-97
Received: 04-17-97
Analyzed: 04-22-97
Reported: 04-28-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP50
Laboratory Sample #: 97040816

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	150	<---
Fluorene	86-73-7	100	N.D.	
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.	
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	N.D.	
2-Methylphenol	95-48-7	100	N.D.	
'ethylphenol	106-44-5	100	N.D.	
phthalene	91-20-3	100	N.D.	
2-Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	N.D.	
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	170	<---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

: Noorani
aboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP34A
Laboratory Sample #: 97040817
Laboratory Reference #: MWI 8924

Sampled : 04-17-97
Received: 04-17-97
Analyzed: 04-18-97
Reported: 04-18-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
Benzo(k)fluoranthene	207-08-9	250	N.D.
Benzo(g,h,i)perylene	191-24-2	250	N.D.
Benzo(a)pyrene	50-32-8	250	N.D.
Benzyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	120 <---
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
Dimethyl phthalate	131-11-3	100	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.
2,4-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040823
Laboratory Reference #: MWI 8924

Sampled: 04-17-97
Received: 04-17-97
Analyzed: 04-22-97
Reported: 04-28-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Equip Blank
Laboratory Sample Number: 97040822
Laboratory Reference #: MWI 8924

Sampled: 04-17-97
Received: 04-17-97
Analyzed: 04-24-97
Reported: 04-28-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040822

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*109*
Toluene-d8	*101*
4-Bromofluorobenzene	*108*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample Number: 97040823
Laboratory Reference #: MWI 8924

Sampled: 04-17-97
Received: 04-17-97
Analyzed: 04-24-97
Reported: 04-28-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040823

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*107*
Toluene-d8	*100*
4-Bromofluorobenzene	*109*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Trip Blank
Laboratory Sample Number: 97040824
Laboratory Reference #: MWI 8924

Sampled: 04-17-97
Received: 04-17-97
Analyzed: 04-24-97
Reported: 04-28-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Trip Blank

Laboratory Sample Number: 97040824

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULT (ug/l)*
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*113*
Toluene-d8	*98*
4-Bromofluorobenzene	*106*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040822
Laboratory Reference #: MWI 8924

Sampled : 04-17-97
Received: 04-17-97
Analyzed: 04-23-97
Reported: 04-28-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040822

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
ethylphenol	106-44-5	5.0	N.D.
hthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

k Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040823
Laboratory Reference #: MWI 8924

Sampled : 04-17-97
Received: 04-17-97
Analyzed: 04-23-97
Reported: 04-28-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040823

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
'ethylphenol	106-44-5	5.0	N.D.
phthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

ark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040822
Laboratory Reference #: MWI 8924

Sampled: 04-17-97
Received: 04-17-97
Analyzed: 04-22-97
Reported: 04-28-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040823
Laboratory Reference #: MWI 8924

Sampled: 04-17-97
Received: 04-17-97
Analyzed: 04-22-97
Reported: 04-28-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040822
Laboratory Reference #: MWI 8924

Sampled: 04-17-97
Received: 04-17-97
Analyzed: 04-21-97
Reported: 04-28-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040823
Laboratory Reference #: MWI 8924

Sampled: 04-17-97
Received: 04-17-97
Analyzed: 04-21-97
Reported: 04-28-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP50
Laboratory Sample #: 97040816
Laboratory Reference #: MWI 8924

Sampled : 04-17-97
Received: 04-17-97
Analyzed: 04-18-97
Reported: 04-18-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
Acenaphthene	83-32-9	100	N.D.	
Acenaphthylene	208-96-8	100	N.D.	
Aniline	62-53-3	100	N.D.	
Anthracene	120-12-7	100	N.D.	
Benzidine	92-87-5	500	N.D.	
Benzoic Acid	65-85-0	250	N.D.	
Benzo(a)anthracene	56-55-3	100	110	<---
Benzo(b)fluoranthene	205-99-2	250	N.D.	
Benzo(k)fluoranthene	207-08-9	250	N.D.	
Benzo(g,h,i)perylene	191-24-2	250	N.D.	
Benzo(a)pyrene	50-32-8	250	N.D.	
Benzyl alcohol	100-51-6	100	N.D.	
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.	
Bis(2-chloroethyl)ether	111-44-4	100	N.D.	
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.	
Bis(2-ethylhexyl)phthalate	117-81-7	100	410	<---
4-Bromophenyl phenyl ether	101-55-3	100	N.D.	
Butyl benzyl phthalate	85-68-7	100	N.D.	
4-Chloroaniline	106-47-8	100	N.D.	
2-Chloronaphthalene	91-58-7	100	N.D.	
4-Chloro-3-methylphenol	59-50-7	100	N.D.	
2-Chlorophenol	95-57-8	100	N.D.	
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.	
Chrysene	218-0109	100	170	<---
Dibenz(a,h)anthracene	53-70-3	100	N.D.	
Dibenzofuran	132-64-9	100	N.D.	
Di-n-butyl phthalate	84-74-2	250	N.D.	
1,3-Dichlorobenzene	541-73-1	100	N.D.	
1,4-Dichlorobenzene	106-46-7	100	N.D.	
1,2-Dichlorobenzene	95-50-1	100	N.D.	
3,3'-Dichlorobenzidine	91-94-1	100	N.D.	
2,4-Dichlorophenol	120-83-2	100	N.D.	
Diethyl phthalate	84-66-2	100	N.D.	
2,4-Dimethylphenol	105-67-9	100	N.D.	
Dimethyl phthalate	131-11-3	100	N.D.	
4,6-Dinitro-2-methylphenol	534-52-1	100	N.D.	
2,4-Dinitrophenol	51-28-5	100	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/22/97
Laboratory Sample No : 97040596
Laboratory Reference No : MWI 8924

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	3.000	2.420	2.700	81	90	11
Arsenic	0.00	1.000	0.800	0.920	80	92	14
Barium	0.00	0.100	0.105	0.104	105	104	1
Beryllium	0.00	0.100	0.085	0.087	85	87	2
Cadmium	0.00	0.100	0.089	0.097	89	97	9
Chromium (VI)	0.00	0.500	0.450	0.490	90	98	9
Chromium (Total)	0.00	0.100	0.087	0.091	87	91	4
Cobalt	0.00	0.100	0.109	0.107	109	107	2
Copper	0.00	0.100	0.080	0.089	80	89	11
Lead	0.00	1.000	1.040	1.010	104	101	3
Mercury	0.000	0.020	0.020	0.022	100	110	10
Molybdenum	0.00	1.000	0.920	1.000	92	100	8
Nickel	0.00	0.500	0.500	0.480	100	96	4
Selenium	0.00	1.000	0.860	0.950	86	95	10
Silver	0.00	0.500	0.440	0.480	88	96	9
Thallium	0.00	3.000	2.460	2.650	82	88	7
Vanadium	0.00	0.500	0.450	0.470	90	94	4
Zinc	0.00	0.100	0.109	0.109	109	109	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/21/97
Laboratory Sample No : 97040833
Laboratory Reference No : MWI 8924

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	30.00	28.00	28.00	93	93	0
Arsenic	0.00	10.00	10.80	10.80	108	108	0
Barium	14.00	10.00	24.00	25.00	100	110	4
Beryllium	0.00	1.00	1.10	1.14	110	114	4
Cadmium	0.00	1.00	0.92	0.94	92	94	2
Chromium (Total)	2.90	5.00	7.70	7.90	96	100	3
Chromium (VI)	0.00	1.00	0.88	0.91	88	91	3
Cobalt	0.80	1.00	1.88	1.87	108	107	1
Copper	1.40	1.00	2.30	2.45	90	105	6
Lead	0.00	10.00	9.00	9.60	90	96	6
Mercury	0.000	0.020	0.024	0.023	120	115	4
Molybdenum	0.00	10.00	10.40	9.80	104	98	6
Nickel	1.30	5.00	6.40	5.90	102	92	8
Selenium	0.00	10.00	9.30	9.90	93	99	6
Silver	0.00	5.00	5.00	4.50	100	90	11
Thallium	0.00	30.00	28.00	27.00	93	90	4
Vanadium	3.80	5.00	8.90	8.80	102	100	1
Zinc	4.70	5.00	9.80	9.50	102	96	3

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/25/97
Laboratory Sample No : 97040851
Laboratory Reference No : MWI 8924

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	45	47	90	94	4
1,1-Dichloroethene	53	50	110	99	114	92	11
Trichloroethene	6.5	50	62	65	111	117	5
Toluene	0.0	50	48	49	54	98	2
Chlorobenzene	0.0	50	47	47	51	94	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/24/97
Laboratory Sample No : 97040824
Laboratory Reference No : MWI 8924

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	20	17	16	85	80	6
1,1-Dichloroethene	0.0	20	15	14	75	70	7
Trichloroethene	0.0	20	19	17	95	85	11
Toluene	0.0	20	18	16	90	80	12
Chlorobenzene	0.0	20	18	16	90	80	12

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/16/97
Laboratory Sample No : 97040492
Laboratory Reference No : MWI 8924

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	46	45	92	90	2
n-Nitroso-di-n-propylamine	0.0	50	49	49	98	98	0
1,2,4-Trichlorobenzene	0.0	50	48	47	96	94	2
Acenaphthene	0.0	50	51	50	102	100	2
Pyrene	0.0	50	38	33	76	66	14
Pentachlorophenol	0.0	100	77	82	77	82	6
4-Chloro-3-Methylphenol	0.0	100	71	72	71	72	1
2-Chlorophenol	0.0	100	79	77	79	77	3
Phenol	0.0	100	70	69	70	69	1

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/18/97
Laboratory Sample No : 97040628
Laboratory Reference No : MWI 8924

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	100	97	107	97	107	10

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : PCB 'S (EPA 8080)

Date of Analysis : 04/21/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8924

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
PCB-1260	0.0	20	9.6	7.5	48	38	25

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/23/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8924

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	37	43	74	86	15
n-Nitroso-di-n-propylamine	0.0	50	47	50	94	100	6
1,2,4-Trichlorobenzene	0.0	50	39	45	78	90	14
Acenaphthene	0.0	50	47	51	94	102	8
Pyrene	0.0	50	40	46	80	92	14
Pentachlorophenol	0.0	100	80	80	80	80	0
4-Chloro-3-Methylphenol	0.0	100	65	67	65	67	3
2-Chlorophenol	0.0	100	80	83	80	83	4
Phenol	0.0	100	23	26	23	26	12

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Organochlorine Pesticides (EPA 8080)

Date of Analysis : 04/22/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8924

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
4,4'-DDT	0.0	1.00	0.95	1.00	95	100	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/22/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8924

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	2.5	2.2	2.1	88	84	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532
Tustin, CA 92680
(714) 832-0064, Fax (714) 832-0067

Analysis Request and Chain of Custody Record

Lab Job No: _____
Page 1 of 1

REQUIRED TAT: 24 hr TAT on Soil TRPH; 24 hr TAT (or ASAP) on B37-RE4-SP34 ~~STANDARD TAT OTHERWISE~~

CUSTOMER INFORMATION	PROJECT INFORMATION
COMPANY: Montgomery Watson	PROJECT NAME: McDonnell Douglas
SEND REPORT TO: Fred Strauss	NUMBER: 1206035.01090010
ADDRESS: 250 N. Madison Ave. Pasadena, CA 91101	LOCATION: C-6 Facility, Bldg. 37 Area
	ADDRESS: 19503 S. Normandie Ave. Los Angeles, CA
PHONE: (818) 568-6582 FAX (818) 796-5941	SAMPLED BY: AWN/IV

SAMPLE ID	NO. OF CONTAINERS	SAMPLE DATE	SAMPLE TIME	SAMPLE MATRIX	CONTAINER TYPE	PRES.	TRPH 418.1	Title 22 Metals	VOCs 8260	SVOCs 8270	PCBs 8080	Pesticides 8080	REMARKS/PRECAUTIONS
B37-RE4-SP50	2	4/17/97	0900	SOIL	2"x6" ss sleeve	ICE	X	X	X	X			24 hr TAT on TRPH only.
B37-RE4-SP34A	2		1310							X			24 hr TAT or ASAP (48 hr)
B37-RE4-SP34B	2		1315							X			
B37-RE4-SP34C	2		1330							X			
B37-RE4-SP34D	2		1345							X			
B37-RE4-SP34E	2		1405							X			
Equipment Blank	7		1545	Water	4 amber 1 poly 2 vials		X	X	X	X	X	X	Standard TAT
Rinsate Blank	7		1600				X	X	X	X	X	X	
Trip Blank	2		—		2 VOAs				X				

Total No. of Samples: 9

Method of Shipment: 1 cooler via Orange Coast Courier

Relinquished By:	Date/Time:	Received By:	Date/Time:
Adam W. Nomi	4/17/97 1630		
Relinquished By:	Date/Time:	Received By:	Date/Time:
Relinquished By:	Date/Time:	Received For Lab By:	Date/Time: 4/17/97 16:30

Reporting Format: (check)
NORMAL _____ S.D. HMMD _____
RWQCB _____ OTHER _____

Sample Integrity: (check)
Intact _____ on Ice ✓

All samples remain the property of the client who is responsible for disposal. A disposal fee may be imposed if client fails to pickup samples.

BOE-C6-0143938

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Soil, B37-RE4-SP50

Laboratory Sample Number: 97040816
Laboratory Reference #: MWI 8924A

Sampled: 04-17-97
Received: 04-17-97
Analyzed: 05-06-97
Reported: 05-07-97

STLC METALS

Analyte	*EPA Method*	*Detection Limit ppm*	*Analysis Result ppm*
Chromium	6010	0.1	2.8

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Chromium (EPA 6010)

Date of Analysis : 05/06/97
Laboratory Sample No : 97040816
Laboratory Reference No : MWI 8924A

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Chromium	2.8	1.0	3.68	3.60	88	80	2

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Soil

Laboratory Reference #: MWI 8931

Sampled : 04-18-97
Received: 04-18-97
Analyzed: 04-21-97
Reported: 04-21-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/kg)*
97040917	B37-RE4-SP51	17
97040918	B37-GS-136-6'	N.D.
97040919	B37-GS-137-6'	N.D.
97040920	B37-GS-138-6'	N.D.
97040921	B37-GS-139-6'	N.D.
97040922	B37-GS-140-6'	N.D.
97040923	B37-GS-141-6'	N.D.
Detection Limit:		8.0

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP51

Laboratory Sample Number: 97040917
Laboratory Reference #: MWI 8931

Sampled: 04-18-97
Received: 04-18-97
Analyzed: 05-01-97
Reported: 05-05-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	110	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	4.8	<---
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	41	<---
Cobalt	6010	80	8000	0.5	7.4	<---
Copper	6010	25	2500	0.1	17	<---
Lead	6010	5.0	1000	1.0	7.2	<---
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	12	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	27	<---
Zinc	6010	250	5000	0.1	68	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP51
Laboratory Sample Number: 97040917
Laboratory Reference #: MWI 8931

Sampled: 04-18-97
Received: 04-18-97
Analyzed: 04-25-97
Reported: 05-05-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/kg)***	***SAMPLE RESULT (ug/kg)***
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
1,1-Dichloroethene	75-35-4	2.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.
,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-RE4-SP51

Laboratory Sample Number: 97040917

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/kg)	SAMPLE RESULT (ug/kg)
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
n-Butylbenzene	104-51-8	2.5	N.D.
1,2-Dichlorobenzene	95-50-1	2.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*108*
Toluene-d8	*100*
4-Bromofluorobenbzene	*101*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
..N: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP51
Laboratory Sample #: 97040917
Laboratory Reference #: MWI 8931

Sampled : 04-18-97
Received: 04-18-97
Analyzed: 04-25-97
Reported: 05-05-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*
Acenaphthene	83-32-9	100	N.D.
Acenaphthylene	208-96-8	100	N.D.
Aniline	62-53-3	100	N.D.
Anthracene	120-12-7	100	N.D.
Benzidine	92-87-5	500	N.D.
Benzoic Acid	65-85-0	250	N.D.
Benzo(a)anthracene	56-55-3	100	N.D.
Benzo(b)fluoranthene	205-99-2	250	N.D.
Benzo(k)fluoranthene	207-08-9	250	N.D.
enzo(g,h,i)perylene	191-24-2	250	N.D.
zo(a)pyrene	50-32-8	250	N.D.
.zyl alcohol	100-51-6	100	N.D.
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.
Bis(2-chloroethyl)ether	111-44-4	100	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.
4-Bromophenyl phenyl ether	101-55-3	100	N.D.
Butyl benzyl phthalate	85-68-7	100	N.D.
4-Chloroaniline	106-47-8	100	N.D.
2-Chloronaphthalene	91-58-7	100	N.D.
4-Chloro-3-methylphenol	59-50-7	100	N.D.
2-Chlorophenol	95-57-8	100	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.
Chrysene	218-0109	100	N.D.
Dibenz(a,h)anthracene	53-70-3	100	N.D.
Dibenzofuran	132-64-9	100	N.D.
Di-n-butyl phthalate	84-74-2	250	N.D.
1,3-Dichlorobenzene	541-73-1	100	N.D.
1,4-Dichlorobenzene	106-46-7	100	N.D.
1,2-Dichlorobenzene	95-50-1	100	N.D.
3,3'-Dichlorobenzidine	91-94-1	100	N.D.
2,4-Dichlorophenol	120-83-2	100	N.D.
Diethyl phthalate	84-66-2	100	N.D.
2,4-Dimethylphenol	105-67-9	100	N.D.
imethyl phthalate	131-11-3	100	N.D.
6-Dinitro-2-methylphenol	534-52-1	100	N.D.
-Dinitrophenol	51-28-5	100	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP51
Laboratory Sample #: 97040917

ANALYTE	CAS NUMBER	DETECTION LIMIT ug/kg	SAMPLE RESULTS ug/kg
2,4-Dinitrotoluene	121-14-2	250	N.D.
2,6-Dinitrotoluene	606-20-2	250	N.D.
Di-n-octyl phthalate	117-84-0	250	N.D.
Fluoranthene	206-44-0	100	N.D.
Fluorene	86-73-7	100	N.D.
Hexachlorobenzene	118-74-1	100	N.D.
Hexachlorobutadiene	87-68-3	100	N.D.
Hexachlorocyclopentadiene	77-47-4	100	N.D.
Hexachloroethane	67-72-1	100	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.
Isophorone	78-59-1	100	N.D.
2-Methylnaphthalene	91-57-6	100	N.D.
2-Methylphenol	95-48-7	100	N.D.
4-Methylphenol	106-44-5	100	N.D.
Naphthalene	91-20-3	100	N.D.
2-Nitroaniline	88-74-4	250	N.D.
3-Nitroaniline	99-09-2	250	N.D.
4-Nitroaniline	100-01-6	250	N.D.
Nitrobenzene	98-95-3	100	N.D.
2-Nitrophenol	88-75-5	100	N.D.
4-Nitrophenol	100-02-7	100	N.D.
N-Nitrosodiphenylamine	86-30-6	100	N.D.
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.
N-Nitrosodimethylamine	62-75-9	100	N.D.
Pentachlorophenol	87-86-5	250	N.D.
Phenanthrene	85-01-8	100	N.D.
Phenol	108-95-2	100	N.D.
Pyrene	129-00-0	100	N.D.
1,2,4-Trichlorobenzene	120-82-1	100	N.D.
2,4,5-Trichlorophenol	95-95-4	100	N.D.
2,4,6-Trichlorophenol	88-06-2	100	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Water

Laboratory Reference #: MWI 8931

Sampled : 04-18-97
Received: 04-18-97
Analyzed: 04-22-97
Reported: 05-05-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/l)*
97040924	Equip Blank	N.D.
97040925	Rinsate Blank	N.D.
Detection Limit:		0.5

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040924
Laboratory Reference #: MWI 8931

Sampled: 04-18-97
Received: 04-18-97
Analyzed: 04-22-97
Reported: 05-05-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Iontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040925
Laboratory Reference #: MWI 8931

Sampled: 04-18-97
Received: 04-18-97
Analyzed: 04-22-97
Reported: 05-05-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Equip Blank
Laboratory Sample Number: 97040924
Laboratory Reference #: MWI 8931

Sampled: 04-18-97
Received: 04-18-97
Analyzed: 04-24-97
Reported: 05-05-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/l)***	***SAMPLE RESULT (ug/l)***
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040924

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*109*
Toluene-d8	*97*
4-Bromofluorobenzene	*110*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss,
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample Number: 97040925
Laboratory Reference #: MWI 8931

Sampled: 04-18-97
Received: 04-18-97
Analyzed: 04-24-97
Reported: 05-05-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/l)***	***SAMPLE RESULT (ug/l)***
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040925

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*112*
Toluene-d8	*98*
4-Bromofluorobenzene	*109*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Trip Blank
Laboratory Sample Number: 97040926
Laboratory Reference #: MWI 8931

Sampled: 04-18-97
Received: 04-18-97
Analyzed: 04-24-97
Reported: 05-05-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/l)***	***SAMPLE RESULT (ug/l)***
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Trip Blank

Laboratory Sample Number: 97040926

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*117*
Toluene-d8	*95*
4-Bromofluorobenzene	*109*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040924
Laboratory Reference #: MWI 8931

Sampled : 04-18-97
Received: 04-18-97
Analyzed: 04-28-97
Reported: 05-05-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Equip Blank
Laboratory Sample #: 97040924

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
-Methylphenol	95-48-7	5.0	N.D.
Methylphenol	106-44-5	5.0	N.D.
hthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040925
Laboratory Reference #: MWI 8931

Sampled : 04-18-97
Received: 04-18-97
Analyzed: 04-28-97
Reported: 05-05-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*	
Acenaphthene	83-32-9	5.0	N.D.	
Acenaphthylene	208-96-8	5.0	N.D.	
Aniline	62-53-3	5.0	N.D.	
Anthracene	120-12-7	5.0	N.D.	
Benzidine	92-87-5	5.0	N.D.	
Benzoic Acid	65-85-0	50	N.D.	
Benzo(a)anthracene	56-55-3	5.0	N.D.	
Benzo(b)fluoranthene	205-99-2	25	N.D.	
Benzo(k)fluoranthene	207-08-9	25	N.D.	
Benzo(g,h,i)perylene	191-24-2	25	N.D.	
Benzo(a)pyrene	50-32-8	25	N.D.	
Benzyl alcohol	100-51-6	50	N.D.	
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.	
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.	
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.	
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	9.0	<---
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.	
Butyl benzyl phthalate	85-68-7	5.0	N.D.	
4-Chloroaniline	106-47-8	5.0	N.D.	
2-Chloronaphthalene	91-58-7	5.0	N.D.	
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.	
2-Chlorophenol	95-57-8	5.0	N.D.	
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.	
Chrysene	218-0109	5.0	N.D.	
Dibenz(a,h)anthracene	53-70-3	25	N.D.	
Dibenzofuran	132-64-9	5.0	N.D.	
Di-N-butyl phthalate	84-74-2	5.0	N.D.	
1,3-Dichlorobenzene	541-73-1	5.0	N.D.	
1,4-Dichlorobenzene	106-46-7	5.0	N.D.	
1,2-Dichlorobenzene	95-50-1	5.0	N.D.	
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.	
2,4-Dichlorophenol	120-83-2	5.0	N.D.	
Diethyl phthalate	84-66-2	5.0	N.D.	
2,4-Dimethylphenol	105-67-9	5.0	N.D.	
Dimethyl phthalate	131-11-3	5.0	N.D.	
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.	
2,4-Dinitrophenol	51-28-5	50	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97040925

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
2-Methylnaphthalene	91-57-6	5.0	N.D.
2-Methylphenol	95-48-7	5.0	N.D.
Methylphenol	106-44-5	5.0	N.D.
hthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

M. K Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Stráuss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040924
Laboratory Reference #: MWI 8931

Sampled: 04-18-97
Received: 04-18-97
Analyzed: 04-22-97
Reported: 05-05-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040925
Laboratory Reference #: MWI 8931

Sampled: 04-18-97
Received: 04-18-97
Analyzed: 04-22-97
Reported: 05-05-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97040924
Laboratory Reference #: MWI 8931

Sampled: 04-18-97
Received: 04-18-97
Analyzed: 04-21-97
Reported: 05-05-97

PCB'S (EPA 8080)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULTS (ug/l)
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97040925
Laboratory Reference #: MWI 8931

Sampled: 04-18-97
Received: 04-18-97
Analyzed: 04-21-97
Reported: 05-05-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/22/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8931

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	2.5	2.2	2.1	88	84	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/21/97
Laboratory Sample No : 97040921
Laboratory Reference No : MWI 8931

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	100	82	91	82	91	10

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 04/22/97
Laboratory Sample No : 97040596
Laboratory Reference No : MWI 8931

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	3.000	2.420	2.700	81	90	11
Arsenic	0.00	1.000	0.800	0.920	80	92	14
Barium	0.00	0.100	0.105	0.104	105	104	1
Beryllium	0.00	0.100	0.085	0.087	85	87	2
Cadmium	0.00	0.100	0.089	0.097	89	97	9
Chromium (VI)	0.00	0.500	0.450	0.490	90	98	9
Chromium (Total)	0.00	0.100	0.087	0.091	87	91	4
Cobalt	0.00	0.100	0.109	0.107	109	107	2
Copper	0.00	0.100	0.080	0.089	80	89	11
Lead	0.00	1.000	1.040	1.010	104	101	3
Mercury	0.000	0.020	0.020	0.022	100	110	10
Molybdenum	0.00	1.000	0.920	1.000	92	100	8
Nickel	0.00	0.500	0.500	0.480	100	96	4
Selenium	0.00	1.000	0.860	0.950	86	95	10
Silver	0.00	0.500	0.440	0.480	88	96	9
Thallium	0.00	3.000	2.460	2.650	82	88	7
Vanadium	0.00	0.500	0.450	0.470	90	94	4
Zinc	0.00	0.100	0.109	0.109	109	109	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 05/01/97
Laboratory Sample No : 97040939
Laboratory Reference No : MWI 8931

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	30.00	29.00	29.00	97	97	0
Arsenic	0.00	10.00	10.70	11.00	107	110	3
Barium	7.60	10.00	18.40	18.90	108	113	3
Beryllium	0.00	1.00	0.93	0.99	93	99	6
Cadmium	0.00	1.00	1.06	1.15	106	115	8
Chromium (Total)	1.80	5.00	6.90	6.90	102	102	0
Chromium (VI)	0.00	1.00	0.87	0.95	87	95	9
Cobalt	0.54	1.00	1.55	1.62	101	108	4
Copper	0.48	1.00	1.38	1.52	90	104	10
Lead	0.00	10.00	9.90	10.10	99	101	2
Mercury	0.000	0.020	0.017	0.019	85	95	11
Molybdenum	0.00	10.00	10.20	10.00	102	100	2
Nickel	0.68	5.00	5.50	5.70	96	100	4
Selenium	0.00	10.00	10.40	10.10	104	101	3
Silver	0.00	5.00	4.70	5.00	94	100	6
Thallium	0.00	30.00	26.00	30.00	87	100	14
Vanadium	2.00	5.00	6.90	7.00	98	100	1
Zinc	3.30	5.00	8.10	8.70	96	108	7

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS + MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/24/97
Laboratory Sample No : 97040824
Laboratory Reference No : MWI 8931

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	20	17	16	85	80	6
1,1-Dichloroethene	0.0	20	15	14	75	70	7
Trichloroethene	0.0	20	19	17	95	85	11
Toluene	0.0	20	18	16	90	80	12
Chlorobenzene	0.0	20	18	16	90	80	12

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/25/97
Laboratory Sample No : 97040920
Laboratory Reference No : MWI 8931

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	43	48	86	96	11
1,1-Dichloroethene	0.0	50	42	44	84	88	5
Trichloroethene	0.0	50	51	56	102	112	9
Toluene	0.0	50	45	49	90	98	9
Chlorobenzene	0.0	50	46	46	92	92	0

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/28/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8931

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	39	39	78	78	0
n-Nitroso-di-n-propylamine	0.0	50	42	49	84	98	15
1,2,4-Trichlorobenzene	0.0	50	39	39	78	78	0
Acenaphthene	0.0	50	47	46	94	92	2
Pyrene	0.0	50	43	41	86	82	5
Pentachlorophenol	0.0	100	72	76	72	76	5
4-Chloro-3-Methylphenol	0.0	100	56	60	56	60	7
2-Chlorophenol	0.0	100	77	79	77	79	3
Phenol	0.0	100	25	27	25	27	8

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/24/97
Laboratory Sample No : 97040783
Laboratory Reference No : MWI 8931

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	43	45	86	90	5
n-Nitroso-di-n-propylamine	0.0	50	48	49	96	98	2
1,2,4-Trichlorobenzene	0.0	50	42	44	84	88	5
Acenaphthene	0.0	50	46	47	92	94	2
Pyrene	0.0	50	30	32	60	64	6
`entachlorophenol	0.0	100	70	78	70	78	11
4-Chloro-3-Methylphenol	0.0	100	65	68	65	68	5
2-Chlorophenol	0.0	100	75	79	75	79	5
Phenol	0.0	100	65	69	65	69	6

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : PCB 'S (EPA 8080)

Date of Analysis : 04/21/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8931

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
PCB-1260	0.0	20.0	9.6	7.5	48	38	25

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Organochlorine Pesticides (EPA 8080)

Date of Analysis : 04/22/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8931

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
4-4' -DDT	0.00	1.00	0.95	1.00	95	100	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532
Tustin, CA 92680
(714) 832-0064, Fax (714) 832-0067

Analy[illegible] quest and Chain of Custody Record

Lab Job No: _____
Page 1 of 1

REQUIRED TAT: 24 hr TAT on Soil TRPH only; Standard TAT Otherwise

CUSTOMER INFORMATION

COMPANY: Montgomery Watson
SEND REPORT TO: Fred Strauss
ADDRESS: 250 N. Madison Ave. Pasadena, CA 91101
PHONE: (818) 568-6582 FAX: (818) 796-5941

PROJECT INFORMATION

PROJECT NAME: Mc Donnell Douglas
NUMBER: 1206035.01090010
LOCATION: C-6 Facility - Bldg. 37 Area
ADDRESS: 19503 S. Normandie Ave. Los Angeles, CA
SAMPLED BY: AWN/IV

ANALYSIS/METHOD REQUEST

SAMPLE ID	NO. OF CONTAINERS	SAMPLE DATE	SAMPLE TIME	SAMPLE MATRIX	CONTAINER TYPE	PRES.	TRPH 418.1	Title 22 Metals	VOCs 8260	SVOCs 8270	PCBs 8080	Pesticides 8080			REMARKS/PRECAUTIONS
B37-RE4-SP51	2	4/18/97	1000	SOIL	2"x6" SS Sleeve	ICE	X	X	X	X					
B37-GS-136-6'	2	↓	1005	↓	↓	↓	X	X							
B37-GS-137-6'	2	↓	1036	↓	↓	↓	X	X							
B37-GS-138-6'	2	↓	1100	↓	↓	↓	X	X							
B37-GS-139-6'	2	↓	1128	↓	↓	↓	X	X							
B37-GS-140-6'	2	↓	1152	↓	↓	↓	X	X							
B37-GS-141-6'	2	↓	1243	↓	↓	↓	X	X							
Equipment Blank	7	↓	1535	Water	Harbor L Poly 2 VOAs	HNO_3 HCl	X	X	X	X	X	X			
Rinsate Blank	7	↓	1548	↓	↓	↓	X	X	X	X	X	X			
Trip Blank	2	↓		↓	2 VOAs	HCl			X						

Total No. of Samples: 10

Method of Shipment: Cooler Via Orange Coast Courier

Relinquished By: Adam W. Young Date/Time: 4/18/97 1625

Received By: Date/Time:

Relinquished By: Date/Time:

Received By: Date/Time:

Relinquished By: Date/Time:

Received For Lab By: [illegible] Date/Time: 16:25

Reporting Format: (check)
NORMAL _____ S.D. HMMD _____
RWQCB _____ OTHER _____

Sample Integrity: (check)
Intact _____ on ice _____

All samples remain the property of the client who is responsible for disposal. A disposal fee may be imposed if client fails to pickup samples.

BOE-C6-0143974

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Soil

Laboratory Reference #: MWI 8973

Sampled : 04-24-97
Received: 04-24-97
Analyzed: 04-25-97
Reported: 04-25-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/kg)*
97041296	B37-RE4-SP52	76
Detection Limit:		8.0

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP52

Laboratory Sample Number: 97041296
Laboratory Reference #: MWI 8973

Sampled: 04-24-97
Received: 04-24-97
Analyzed: 05-06-97
Reported: 05-07-97

CCR - METALS

Analyte	*EPA Method*	*STLC Limits mg/l*	*TTLC Limits mg/kg*	*Detection Limit mg/kg*	*Analysis Result mg/kg*	
Antimony	6010	15	500	5.0	N.D.	
Arsenic	6010	5.0	500	1.0	N.D.	
Barium	6010	100	10000	0.1	91	<---
Beryllium	6010	0.75	75	0.1	N.D.	
Cadmium	6010	1.0	100	0.1	2.8	<---
Chromium (VI)	7196	5.0	500	0.5	N.D.	
Chromium Total	6010	560	2500	0.05	35	<---
Cobalt	6010	80	8000	0.5	6.8	<---
Copper	6010	25	2500	0.1	16	<---
Lead	6010	5.0	1000	1.0	34	<---
Mercury	7471	0.2	20	0.01	N.D.	
Molybdenum	6010	350	3500	0.5	N.D.	
Nickel	6010	20	2000	0.5	12	<---
Selenium	6010	1.0	100	1.0	N.D.	
Silver	6010	5.0	500	0.1	N.D.	
Thallium	6010	7.0	700	5.0	N.D.	
Vanadium	6010	24	2400	0.5	26	<---
Zinc	6010	250	5000	0.1	60	<---

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP52
Laboratory Sample Number: 97041296
Laboratory Reference #: MWI 8973

Sampled: 04-24-97
Received: 04-24-97
Analyzed: 04-29-97
Reported: 05-07-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/kg)***	***SAMPLE RESULT (ug/kg)***
Benzene	71-43-2	2.5	N.D.
Bromodichloromethane	75-27-4	2.5	N.D.
Bromoform	75-25-2	2.5	N.D.
Bromomethane	74-83-9	2.5	N.D.
Carbon Disulfide	75-15-0	5.0	N.D.
Carbon tetrachloride	56-23-5	2.5	N.D.
Chlorobenzene	108-90-7	2.5	N.D.
Chlorodibromomethane	124-48-1	2.5	N.D.
Chloroethane	75-00-3	2.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	5.0	N.D.
Chloroform	67-66-3	2.5	N.D.
Chloromethane	74-87-3	2.5	N.D.
1,1-Dichloroethane	75-35-3	2.5	N.D.
1,2-Dichloroethane	107-06-2	2.5	N.D.
1,1-Dichloroethene	75-35-4	2.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	2.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	2.5	N.D.
Ethylbenzene	100-41-4	2.5	N.D.
Methylene chloride	75-09-2	5.0	N.D.
Styrene	100-42-5	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
Tetrachloroethene	127-18-4	2.5	N.D.
Toluene	108-88-3	2.5	N.D.
1,1,1-Trichloroethane	71-55-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
Trichloroethene	79-01-6	2.5	N.D.
Trichlorofluoromethane	75-69-4	5.0	N.D.
Vinyl acetate	108-05-4	5.0	N.D.
Vinyl chloride	75-01-4	2.5	N.D.
Total Xylenes	1330-20-7	2.5	N.D.
Dichlorofluoromethane	75-71-8	2.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	2.5	N.D.
2,2-Dichloropropane	590-20-7	2.5	N.D.
Bromochloromethane	74-97-5	2.5	N.D.
1,1-Dichloropropene	563-58-6	2.5	N.D.
1,2-Dichloropropane	78-87-5	2.5	N.D.
Dibromomethane	74-95-3	2.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	2.5	N.D.
1,1,2-Trichloroethane	79-00-5	2.5	N.D.
1,2-Dibromoethane	106-93-4	2.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Soil, B37-Re4-SP52

Laboratory Sample Number: 97041296

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/kg)	SAMPLE RESULT (ug/kg)
1,3-Dichloropropane	142-28-9	2.5	N.D.
Dibromochloromethane	124-48-1	2.5	N.D.
Isopropylbenzene	98-82-8	2.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	2.5	N.D.
1,2,3-Trichloropropane	96-18-4	2.5	N.D.
Bromobenzene	108-86-1	2.5	N.D.
n-Propylbenzene	103-65-1	2.5	N.D.
2-Chlorotoluene	95-49-8	2.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	2.5	N.D.
4-Chlorotoluene	106-43-4	2.5	N.D.
tert-Butylbenzene	98-06-6	2.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	2.5	N.D.
sec-Butylbenzene	135-98-8	2.5	N.D.
4-Isopropyltoluene	99-87-6	2.5	N.D.
1,3-Dichlorobenzene	541-73-1	2.5	N.D.
1,4-Dichlorobenzene	106-46-7	2.5	N.D.
n-Butylbenzene	104-51-8	2.5	N.D.
1,2-Dichlorobenzene	95-50-1	2.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	2.5	N.D.
Hexachlorobutadiene	87-68-3	2.5	N.D.
Naphthalene	91-20-3	2.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	2.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*113*
Toluene-d8	*98*
4-Bromofluorobenbzene	*98*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

tgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Soil, B37-RE4-SP52
Laboratory Sample #: 97041296
Laboratory Reference #: MWI 8973

Sampled : 04-24-97
Received: 04-24-97
Analyzed: 04-30-97
Reported: 05-06-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	CAS NUMBER	DETECTION LIMIT ug/kg	SAMPLE RESULTS ug/kg	
Acenaphthene	83-32-9	100	N.D.	
Acenaphthylene	208-96-8	100	N.D.	
Aniline	62-53-3	100	N.D.	
Anthracene	120-12-7	100	N.D.	
Benzidine	92-87-5	500	N.D.	
Benzoic Acid	65-85-0	250	N.D.	
Benzo(a)anthracene	56-55-3	100	570	<---
enzo(b)fluoranthene	205-99-2	250	740	<---
enzo(k)fluoranthene	207-08-9	250	300	<---
nzo(g,h,i)perylene	191-24-2	250	N.D.	
zo(a)pyrene	50-32-8	250	590	<---
benzyl alcohol	100-51-6	100	N.D.	
Bis(2-chloroethoxy)methane	111-91-1	100	N.D.	
Bis(2-chloroethyl)ether	111-44-4	100	N.D.	
Bis(2-chloroisopropyl)ether	39638-32-9	100	N.D.	
Bis(2-ethylhexyl)phthalate	117-81-7	100	N.D.	
4-Bromophenyl phenyl ether	101-55-3	100	N.D.	
Butyl benzyl phthalate	85-68-7	100	N.D.	
4-Chloroaniline	106-47-8	100	N.D.	
2-Chloronaphthalene	91-58-7	100	N.D.	
4-Chloro-3-methylphenol	59-50-7	100	N.D.	
2-Chlorophenol	95-57-8	100	N.D.	
4-Chlorophenyl phenyl ether	7005-72-3	100	N.D.	
Chrysene	218-0109	100	600	<---
Dibenz(a,h)anthracene	53-70-3	100	N.D.	
Dibenzofuran	132-64-9	100	N.D.	
Di-n-butyl phthalate	84-74-2	250	N.D.	
1,3-Dichlorobenzene	541-73-1	100	N.D.	
1,4-Dichlorobenzene	106-46-7	100	N.D.	
1,2-Dichlorobenzene	95-50-1	100	N.D.	
3,3'-Dichlorobenzidine	91-94-1	100	N.D.	
2,4-Dichlorophenol	120-83-2	100	N.D.	
Diethyl phthalate	84-66-2	100	N.D.	
4-Dimethylphenol	105-67-9	100	N.D.	
imethyl phthalate	131-11-3	100	N.D.	
6-Dinitro-2-methylphenol	534-52-1	100	N.D.	
4-Dinitrophenol	51-28-5	100	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Soil, B37-RE4-SP52
Laboratory Sample #: 97041296

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/kg*	*SAMPLE RESULTS ug/kg*	
2,4-Dinitrotoluene	121-14-2	250	N.D.	
2,6-Dinitrotoluene	606-20-2	250	N.D.	
Di-n-octyl phthalate	117-84-0	250	N.D.	
Fluoranthene	206-44-0	100	660	<---
Fluorene	86-73-7	100	N.D.	
Hexachlorobenzene	118-74-1	100	N.D.	
Hexachlorobutadiene	87-68-3	100	N.D.	
Hexachlorocyclopentadiene	77-47-4	100	N.D.	
Hexachloroethane	67-72-1	100	N.D.	
Indeno(1,2,3-cd)pyrene	193-39-5	250	N.D.	
Isophorone	78-59-1	100	N.D.	
2-Methylnaphthalene	91-57-6	100	N.D.	
2-Methylphenol	95-48-7	100	N.D.	
4-Methylphenol	106-44-5	100	N.D.	
Naphthalene	91-20-3	100	N.D.	
2-Nitroaniline	88-74-4	250	N.D.	
3-Nitroaniline	99-09-2	250	N.D.	
4-Nitroaniline	100-01-6	250	N.D.	
Nitrobenzene	98-95-3	100	N.D.	
2-Nitrophenol	88-75-5	100	N.D.	
4-Nitrophenol	100-02-7	100	N.D.	
N-Nitrosodiphenylamine	86-30-6	100	N.D.	
N-Nitroso-di-n-propylamine	621-64-7	100	N.D.	
N-Nitrosodimethylamine	62-75-9	100	N.D.	
Pentachlorophenol	87-86-5	250	N.D.	
Phenanthrene	85-01-8	100	220	<---
Phenol	108-95-2	100	N.D.	
Pyrene	129-00-0	100	550	<---
1,2,4-Trichlorobenzene	120-82-1	100	N.D.	
2,4,5-Trichlorophenol	95-95-4	100	N.D.	
2,4,6-Trichlorophenol	88-06-2	100	N.D.	

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

ontgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Analysis Method: 418.1

Sample Description: Water

Laboratory Reference #: MWI 8973

Sampled : 04-24-97
Received: 04-24-97
Analyzed: 04-29-97
Reported: 05-07-97

TOTAL RECOVERABLE PETROLEUM HYDROCARBONS

Laboratory Sample Number	*Client Sample Number*	*Sample Result (mg/l)*
97041297	Equip Blank	N.D.
97041298	Rinsate Blank	N.D.
Detection Limit:		0.5

Analyte reported as N.D. was not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97041297
Laboratory Reference #: MWI 8973

Sampled: 04-24-97
Received: 04-24-97
Analyzed: 05-02-97
Reported: 05-06-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97041298
Laboratory Reference #: MWI 8973

Sampled: 04-24-97
Received: 04-24-97
Analyzed: 05-02-97
Reported: 05-06-97

CCR - METALS

Analyte	*EPA Method*	*Detection Limit mg/l*	*Analysis Result mg/l*
Antimony	6010	0.5	N.D.
Arsenic	6010	0.1	N.D.
Barium	6010	0.01	N.D.
Beryllium	6010	0.01	N.D.
Cadmium	6010	0.01	N.D.
Chromium (VI)	7196	0.01	N.D.
Chromium (Total)	6010	0.01	N.D.
Cobalt	6010	0.05	N.D.
Copper	6010	0.01	N.D.
Lead	6010	0.1	N.D.
Mercury	7471	0.002	N.D.
Molybdenum	6010	0.1	N.D.
Nickel	6010	0.05	N.D.
Selenium	6010	0.1	N.D.
Silver	6010	0.05	N.D.
Thallium	6010	0.5	N.D.
Vanadium	6010	0.1	N.D.
Zinc	6010	0.01	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Equip Blank
Laboratory Sample Number: 97041297
Laboratory Reference #: MWI 8973

Sampled: 04-24-97
Received: 04-24-97
Analyzed: 05-06-97
Reported: 05-07-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/l)***	***SAMPLE RESULT (ug/l)***
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97041297

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*115*
Toluene-d8	*98*
4-Bromofluorobenzene	*112*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss.
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample Number: 97041298
Laboratory Reference #: MWI 8973

Sampled: 04-24-97
Received: 04-24-97
Analyzed: 05-06-97
Reported: 05-07-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/l)***	***SAMPLE RESULT (ug/l)***
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97041298

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*114*
Toluene-d8	*99*
4-Bromofluorobenzene	*114*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss.
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035-01090010

Sample Description: Water, Trip Blank
Laboratory Sample Number: 97041299
Laboratory Reference #: MWI 8973

Sampled: 04-24-97
Received: 04-24-97
Analyzed: 05-06-97
Reported: 05-07-97

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	***CAS NUMBER***	***DETECTION LIMIT (ug/l)***	***SAMPLE RESULT (ug/l)***
Benzene	71-43-2	0.5	N.D.
Bromodichloromethane	75-27-4	0.5	N.D.
Bromoform	75-25-2	0.5	N.D.
Bromomethane	74-83-9	1.0	N.D.
Carbon Disulfide	75-15-0	0.5	N.D.
Carbon tetrachloride	56-23-5	0.5	N.D.
Chlorobenzene	108-90-7	0.5	N.D.
Chlorodibromomethane	124-48-1	0.5	N.D.
Chloroethane	75-00-3	0.5	N.D.
2-Chloroethyl vinyl ether	110-75-8	1.0	N.D.
Chloroform	67-66-3	0.5	N.D.
Chloromethane	74-87-3	0.5	N.D.
1,1-Dichloroethane	75-35-3	0.5	N.D.
1,2-Dichloroethane	107-06-2	0.5	N.D.
1,1-Dichloroethene	75-35-4	0.5	N.D.
Trans 1,2-Dichloroethene	156-60-5	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
cis-1,3-Dichloropropene	10061-01-5	0.5	N.D.
trans-1,3-Dichloropropene	10061-02-6	0.5	N.D.
Ethylbenzene	100-41-4	0.5	N.D.
Methylene chloride	75-09-2	2.5	N.D.
Styrene	100-42-5	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
Tetrachloroethene	127-18-4	0.5	N.D.
Toluene	108-88-3	0.5	N.D.
1,1,1-Trichloroethane	71-55-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
Trichloroethene	79-01-6	0.5	N.D.
Trichlorofluoromethane	75-69-4	0.5	N.D.
Vinyl acetate	108-05-4	1.0	N.D.
Vinyl chloride	75-01-4	0.5	N.D.
Total Xylenes	1330-20-7	1.0	N.D.
Dichlorofluoromethane	75-71-8	0.5	N.D.
cis-1-2,-Dichloroethane	156-59-4	0.5	N.D.
2,2-Dichloropropane	590-20-7	0.5	N.D.
Bromochloromethane	74-97-5	0.5	N.D.
1,1-Dichloropropene	563-58-6	0.5	N.D.
1,2-Dichloropropane	78-87-5	0.5	N.D.
Dibromomethane	74-95-3	0.5	N.D.
trans-1,3-Dichloropropane	10061-02-6	0.5	N.D.
1,1,2-Trichloroethane	79-00-5	0.5	N.D.
1,2-Dibromoethane	106-93-4	0.5	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Sample Description: Water, Trip Blank

Laboratory Sample Number: 97041299

Volatile Organics by GC/MS (EPA 8260)

ANALYTE	CAS NUMBER	DETECTION LIMIT (ug/l)	SAMPLE RESULT (ug/l)
1,3-Dichloropropane	142-28-9	0.5	N.D.
Dibromochloromethane	124-48-1	0.5	N.D.
Isopropylbenzene	98-82-8	0.5	N.D.
1,1,2,2-Tetrachloroethane	79-34-5	0.5	N.D.
1,2,3-Trichloropropane	96-18-4	0.5	N.D.
Bromobenzene	108-86-1	0.5	N.D.
n-Propylbenzene	103-65-1	0.5	N.D.
2-Chlorotoluene	95-49-8	0.5	N.D.
1,3,5-Trimethylbenzene	108-67-8	0.5	N.D.
4-Chlorotoluene	106-43-4	0.5	N.D.
tert-Butylbenzene	98-06-6	0.5	N.D.
1,2,4-Trimethylbenzene	95-63-6	0.5	N.D.
sec-Butylbenzene	135-98-8	0.5	N.D.
4-Isopropyltoluene	99-87-6	0.5	N.D.
1,3-Dichlorobenzene	541-73-1	0.5	N.D.
1,4-Dichlorobenzene	106-46-7	0.5	N.D.
n-Butylbenzene	104-51-8	0.5	N.D.
1,2-Dichlorobenzene	95-50-1	0.5	N.D.
1-2-Dibromo-2-CPA	96-12-8	1.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	0.5	N.D.
Hexachlorobutadiene	87-68-3	0.5	N.D.
Naphthalene	91-20-3	0.5	N.D.
1,2,3-Trichlorobenzene	87-61-6	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

Surrogate Recoveries	*%*
Dibromofluoromethane	*110*
Toluene-d8	*98*
4-Bromofluorobenzene	*114*

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank
Laboratory Sample #: 97041297
Laboratory Reference #: MWI 8973

Sampled : 04-24-97
Received: 04-24-97
Analyzed: 04-28-97
Reported: 05-07-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
Acenaphthene	83-32-9	5.0	N.D.
Acenaphthylene	208-96-8	5.0	N.D.
Aniline	62-53-3	5.0	N.D.
Anthracene	120-12-7	5.0	N.D.
Benzidine	92-87-5	5.0	N.D.
Benzoic Acid	65-85-0	50	N.D.
Benzo(a)anthracene	56-55-3	5.0	N.D.
Benzo(b)fluoranthene	205-99-2	25	N.D.
Benzo(k)fluoranthene	207-08-9	25	N.D.
Benzo(g,h,i)perylene	191-24-2	25	N.D.
Benzo(a)pyrene	50-32-8	25	N.D.
Benzyl alcohol	100-51-6	50	N.D.
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	N.D.
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.
Butyl benzyl phthalate	85-68-7	5.0	N.D.
4-Chloroaniline	106-47-8	5.0	N.D.
2-Chloronaphthalene	91-58-7	5.0	N.D.
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.
2-Chlorophenol	95-57-8	5.0	N.D.
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.
Chrysene	218-0109	5.0	N.D.
Dibenz(a,h)anthracene	53-70-3	25	N.D.
Dibenzofuran	132-64-9	5.0	N.D.
Di-N-butyl phthalate	84-74-2	5.0	N.D.
1,3-Dichlorobenzene	541-73-1	5.0	N.D.
1,4-Dichlorobenzene	106-46-7	5.0	N.D.
1,2-Dichlorobenzene	95-50-1	5.0	N.D.
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.
2,4-Dichlorophenol	120-83-2	5.0	N.D.
Diethyl phthalate	84-66-2	5.0	N.D.
2,4-Dimethylphenol	105-67-9	5.0	N.D.
Dimethyl phthalate	131-11-3	5.0	N.D.
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.
2,4-Dinitrophenol	51-28-5	50	N.D.

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Equip Blank
Laboratory Sample #: 97041297

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
-Methylnaphthalene	91-57-6	5.0	N.D.
-Methylphenol	95-48-7	5.0	N.D.
Methylphenol	106-44-5	5.0	N.D.
hthalene	91-20-3	50	N.D.
2-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

k Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97041298
Laboratory Reference #: MWI 8973

Sampled : 04-24-97
Received: 04-24-97
Analyzed: 04-28-97
Reported: 05-07-97

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*	
Acenaphthene	83-32-9	5.0	N.D.	
Acenaphthylene	208-96-8	5.0	N.D.	
Aniline	62-53-3	5.0	N.D.	
Anthracene	120-12-7	5.0	N.D.	
Benzidine	92-87-5	5.0	N.D.	
Benzoic Acid	65-85-0	50	N.D.	
Benzo(a)anthracene	56-55-3	5.0	N.D.	
Benzo(b)fluoranthene	205-99-2	25	N.D.	
Benzo(k)fluoranthene	207-08-9	25	N.D.	
Benzo(g,h,i)perylene	191-24-2	25	N.D.	
Benzo(a)pyrene	50-32-8	25	N.D.	
Benzyl alcohol	100-51-6	50	N.D.	
Bis(2-chloroethoxy)methane	111-91-1	5.0	N.D.	
Bis(2-chloroethyl)ether	111-44-4	5.0	N.D.	
Bis(2-chloroisopropyl)ether	39638-32-9	5.0	N.D.	
Bis(2-ethylhexyl)phthalate	117-81-7	3.0	6.4	<---
4-Bromophenyl phenyl ether	101-55-3	5.0	N.D.	
Butyl benzyl phthalate	85-68-7	5.0	N.D.	
4-Chloroaniline	106-47-8	5.0	N.D.	
2-Chloronaphthalene	91-58-7	5.0	N.D.	
4-Chloro-3-methylphenol	59-50-7	5.0	N.D.	
2-Chlorophenol	95-57-8	5.0	N.D.	
4-Chlorophenyl phenyl ether	7005-72-3	5.0	N.D.	
Chrysene	218-0109	5.0	N.D.	
Dibenz(a,h)anthracene	53-70-3	25	N.D.	
Dibenzofuran	132-64-9	5.0	N.D.	
Di-N-butyl phthalate	84-74-2	5.0	N.D.	
1,3-Dichlorobenzene	541-73-1	5.0	N.D.	
1,4-Dichlorobenzene	106-46-7	5.0	N.D.	
1,2-Dichlorobenzene	95-50-1	5.0	N.D.	
3,3-Dichlorobenzidine	91-94-1	5.0	N.D.	
2,4-Dichlorophenol	120-83-2	5.0	N.D.	
Diethyl phthalate	84-66-2	5.0	N.D.	
2,4-Dimethylphenol	105-67-9	5.0	N.D.	
Dimethyl phthalate	131-11-3	5.0	N.D.	
4,6-Dinitro-2-methylphenol	534-52-1	50	N.D.	
2,4-Dinitrophenol	51-28-5	50	N.D.	

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

SEMI-VOLATILE ORGANICS by GC/MS (EPA 8270) (continued)

Sample Description: Water, Rinsate Blank
Laboratory Sample #: 97041298

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT ug/l*	*SAMPLE RESULTS ug/l*
2,4-Dinitrotoluene	121-14-2	5.0	N.D.
2,6-Dinitrotoluene	606-20-2	5.0	N.D.
Di-N-octyl phthalate	117-84-0	25	N.D.
Fluoranthene	206-44-0	5.0	N.D.
Fluorene	86-73-7	5.0	N.D.
Hexachlorobenzene	118-74-1	5.0	N.D.
Hexachlorobutadiene	87-68-3	5.0	N.D.
Hexachlorocyclopentadiene	77-47-4	5.0	N.D.
Hexachloroethane	67-72-1	5.0	N.D.
Indeno(1,2,3-cd)pyrene	193-39-5	25	N.D.
Isophorone	78-59-1	5.0	N.D.
^-Methylnaphthalene	91-57-6	5.0	N.D.
-Methylphenol	95-48-7	5.0	N.D.
-Methylphenol	106-44-5	5.0	N.D.
›hthalene	91-20-3	50	N.D.
₂-Nitroaniline	88-74-4	50	N.D.
3-Nitroaniline	99-09-2	50	N.D.
4-Nitroaniline	100-01-6	5.0	N.D.
Nitrobenzene	98-95-3	5.0	N.D.
2-Nitrophenol	88-75-5	5.0	N.D.
4-Nitrophenol	100-02-7	50	N.D.
N-Nitrosodiphenylamine	86-30-6	5.0	N.D.
N-Nitroso-di-N-propylamine	621-64-7	5.0	N.D.
N-Nitrosodimethylamine	62-75-9	5.0	N.D.
Pentachlorophenol	87-86-5	50	N.D.
Phenanthrene	85-01-8	5.0	N.D.
Phenol	108-95-2	5.0	N.D.
Pyrene	129-00-0	5.0	N.D.
1,2,4-Trichlorobenzene	120-82-1	5.0	N.D.
2,4,5-Trichlorophenol	95-95-4	5.0	N.D.
2,4,6-Trichlorophenol	88-06-2	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

k Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Straüss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97041297
Laboratory Reference #: MWI 8938

Sampled: 04-24-97
Received: 04-24-97
Analyzed: 04-28-97
Reported: 05-07-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97041298
Laboratory Reference #: MWI 8938

Sampled: 04-24-97
Received: 04-24-97
Analyzed: 04-28-97
Reported: 05-07-97

PCB'S (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
PCB-1016	12674-11-2	5.0	N.D.
PCB-1221	11104-28-2	5.0	N.D.
PCB-1232	11141-16-5	5.0	N.D.
PCB-1242	53469-21-9	5.0	N.D.
PCB-1248	12672-29-6	5.0	N.D.
PCB-1254	11097-69-1	5.0	N.D.
PCB-1260	11096-82-5	5.0	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Equip Blank

Laboratory Sample Number: 97041297
Laboratory Reference #: MWI 8973

Sampled: 04-24-97
Received: 04-24-97
Analyzed: 04-28-97
Reported: 05-07-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

Montgomery Watson
ATTN: Mr. Fred Strauss
250 N. Madison Ave.
Pasadena, CA 91101

Client Project ID: McDonnell Douglas
Client Project #: 1206035.01090010

Sample Description: Water, Rinsate Blank

Laboratory Sample Number: 97041298
Laboratory Reference #: MWI 8973

Sampled: 04-24-97
Received: 04-24-97
Analyzed: 04-28-97
Reported: 05-07-97

ORGANOCHLORINE PESTICIDES (EPA 8080)

ANALYTE	*CAS NUMBER*	*DETECTION LIMIT (ug/l)*	*SAMPLE RESULTS (ug/l)*
Aldrin	309-00-2	0.1	N.D.
alpha-BHC	319-84-6	0.2	N.D.
beta-BHC	319-85-7	0.2	N.D.
delta-BHC	319-86-8	0.2	N.D.
gamma-BHC (Lindane)	58-89-9	0.2	N.D.
Chlordane	57-74-9	0.2	N.D.
4,4'-DDD	72-54-8	0.5	N.D.
4,4'-DDE	72-55-9	0.1	N.D.
4,4'-DDT	50-29-3	0.1	N.D.
Dieldrin	60-57-1	0.5	N.D.
Endosulfan I	959-98-8	0.5	N.D.
Endosulfan II	33213-65-9	0.5	N.D.
Endosulfan sulfate	1031-07-8	0.5	N.D.
Endrin	72-20-8	0.02	N.D.
Endrin aldehyde	7421-93-4	0.2	N.D.
Heptachlor	76-44-8	0.1	N.D.
Heptachlor epoxide	1024-57-3	0.2	N.D.
Methoxychlor	72-43-5	9.0	N.D.
Toxaphene	8001-35-2	0.5	N.D.

Analytes reported as N.D. were not present above the stated limit of detection.

ORANGE COAST ANALYTICAL

Mark Noorani

Mark Noorani
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/25/97
Laboratory Sample No : 97041014
Laboratory Reference No : MWI 8973

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	100	116	118	116	118	2

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Total Recoverable Petroleum Hydrocarbons (EPA 418.1)

Date of Analysis : 04/29/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8973

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Hydrocarbons	0.0	2.5	2.0	2.2	80	88	10

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/28/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8973

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	39	39	78	78	0
n-Nitroso-di-n-propylamine	0.0	50	42	49	84	98	15
1,2,4-Trichlorobenzene	0.0	50	39	39	78	78	0
Acenaphthene	0.0	50	47	46	94	92	2
Pyrene	0.0	50	43	41	86	82	5
Pentachlorophenol	0.0	100	72	76	72	76	5
4-Chloro-3-Methylphenol	0.0	100	56	60	56	60	7
2-Chlorophenol	0.0	100	77	79	77	79	3
Phenol	0.0	100	25	27	25	27	8

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

Mark Noorani

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Semi-Volatile Organics by GC/MS (EPA 8270)

Date of Analysis : 04/30/97
Laboratory Sample No : 97041323
Laboratory Reference No : MWI 8973

Analyte	R1 (ng)	SP (ng)	MS (ng)	MSD (ng)	PR1 %	PR2 %	RPD %
1,4-Dichlorobenzene	0.0	50	44	45	88	90	2
n-Nitroso-di-n-propylamine	0.0	50	50	53	100	106	6
1,2,4-Trichlorobenzene	0.0	50	42	44	84	88	5
Acenaphthene	0.0	50	48	49	96	98	2
Pyrene	0.0	50	24	25	48	50	4
Pentachlorophenol	0.0	100	72	80	72	80	11
4-Chloro-3-Methylphenol	0.0	100	75	78	75	78	4
2-Chlorophenol	0.0	100	82	83	82	83	1
Phenol	0.0	100	70	73	70	73	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 05/06/97
Laboratory Sample No : 97041299
Laboratory Reference No : MWI 8973

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	20	17	17	85	85	0
1,1-Dichloroethene	0.0	20	14	14	70	70	0
Trichloroethene	0.0	20	18	19	90	95	5
Toluene	0.0	20	18	18	90	90	0
Chlorobenzene	0.0	20	16	18	80	90	12

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Volatile Organics by GC/MS (EPA 8260)

Date of Analysis : 04/29/97
Laboratory Sample No : 97041020
Laboratory Reference No : MWI 8973

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
Benzene	0.0	50	45	42	90	84	7
1,1-Dichloroethene	0.0	50	44	39	88	78	12
Trichloroethene	0.0	50	52	47	104	94	10
Toluene	0.0	50	47	44	94	88	7
Chlorobenzene	0.0	50	44	42	88	84	5

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 05/06/97
Laboratory Sample No : 97041306
Laboratory Reference No : MWI 8973

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.00	30.00	26.00	26.00	87	87	0
Arsenic	0.00	10.00	9.10	8.20	91	82	10
Barium	3.60	10.00	11.80	12.40	82	88	5
Beryllium	0.00	1.00	0.99	0.93	99	93	6
Cadmium	0.00	1.00	1.04	1.04	104	104	0
Chromium (Total)	2.50	5.00	7.40	7.50	98	100	1
Chromium (VI)	0.00	1.00	0.82	0.95	82	95	15
Cobalt	0.44	1.00	1.39	1.32	95	88	5
Copper	0.80	1.00	1.70	1.61	90	81	5
Lead	0.00	10.00	8.60	8.70	86	87	1
Mercury	0.000	0.020	0.017	0.019	85	95	11
Molybdenum	0.00	10.00	8.30	8.40	83	84	1
Nickel	1.10	5.00	5.70	5.80	92	94	2
Selenium	0.00	10.00	9.00	8.50	90	85	6
Silver	0.00	5.00	4.80	4.90	96	98	2
Thallium	0.00	30.00	25.00	26.00	83	87	4
Vanadium	1.90	5.00	6.20	6.20	86	86	0
Zinc	2.60	5.00	7.70	7.10	102	90	8

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Metals

Date of Analysis : 05/02/97
Laboratory Sample No : 97041297
Laboratory Reference No : MWI 8973

Analyte	R1 (ppm)	SP (ppm)	MS (ppm)	MSD (ppm)	PR1 %	PR2 %	RPD %
Antimony	0.000	3.000	3.000	3.600	100	120	18
Arsenic	0.000	1.000	0.910	0.850	91	85	7
Barium	0.000	0.100	0.103	0.113	103	113	9
Beryllium	0.000	0.100	0.086	0.085	86	85	1
Cadmium	0.000	0.100	0.100	0.096	100	96	4
Chromium (Total)	0.000	0.100	0.088	0.084	88	84	5
Chromium (VI)	0.000	0.500	0.440	0.410	88	82	7
Cobalt	0.000	0.100	0.120	0.105	120	105	13
Copper	0.000	0.100	0.088	0.089	88	89	1
Lead	0.000	1.000	1.100	1.130	110	113	3
Mercury	0.000	0.020	0.023	0.020	115	100	14
Molybdenum	0.000	1.000	1.100	0.960	110	96	14
Nickel	0.000	0.500	0.470	0.460	94	92	2
Selenium	0.000	1.000	1.030	0.940	103	94	9
Silver	0.000	0.500	0.460	0.480	92	96	4
Thallium	0.000	3.000	2.700	2.800	90	93	4
Vanadium	0.000	0.500	0.440	0.440	88	88	0
Zinc	0.000	0.100	0.115	0.120	115	120	4

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director

ORANGE COAST ANALYTICAL, INC.
3002 Dow, Suite 532, Tustin, CA 92780 (714) 832-0064 Fax (714) 832-0067
4620 E. Elwood, Suite 4, Phoenix, AZ 85040 (602) 736-0960 Fax (602) 736-0970

QC DATA REPORT

Analysis : Organochlorine Pesticides (EPA 8080)

Date of Analysis : 04/28/97
Laboratory Sample No : OCA 100
Laboratory Reference No : MWI 8973

Analyte	R1 (ppb)	SP (ppb)	MS (ppb)	MSD (ppb)	PR1 %	PR2 %	RPD %
4,4'-DDT	0.0	1.0	1.1	1.0	110	100	10

Definition of Terms :

R1	Results Of First Analysis
SP	Spike Concentration Added to Sample
MS	Matrix Spike Results
MSD	Matrix Spike Duplicate Results
PR1	Percent Recovery Of MS: {(MS-R1) / SP} x100
PR2	Percent Recovery Of MSD: {(MSD-R1) / SP} x 100
RPD	Relative Percent Difference: {(MS-MSD) / (MS+MSD)} x 100 x 2

ORANGE COAST ANALYTICAL

MARK NOORANI
Laboratory Director