LABORATORY CONTROL SAMPLE DATA REPORT

GC/MS Volatiles

Client Lot #...: E2F010157 **Work Order #...:** E2FFC1AC **Matrix.....:** WATER
LCS Lot-Sample#: E2F040000-393
Prep Date.....: 06/03/02 **Analysis Date..:** 06/03/02
Prep Batch #...: 2155393 **Analysis Time..:** 20:25
Dilution Factor: 1 **Instrument ID..:** MSC
Analyst ID.....: 004648

PARAMETER	SPIKE AMOUNT	MEASURED AMOUNT	UNITS	PERCENT RECOVERY	METHOD
1,1-Dichloroethene	**10.0**	**9.65**	**ug/L**	**96**	**SW846 8260B**
Benzene	**10.0**	**9.51**	**ug/L**	**95**	**SW846 8260B**
Trichloroethene	**10.0**	**10.7**	**ug/L**	**107**	**SW846 8260B**
Toluene	**10.0**	**9.89**	**ug/L**	**99**	**SW846 8260B**
Chlorobenzene	**10.0**	**9.66**	**ug/L**	**97**	**SW846 8260B**

SURROGATE	PERCENT RECOVERY	RECOVERY LIMITS
Bromofluorobenzene	104	(75 - 130)
1,2-Dichloroethane-d4	100	(65 - 135)
Toluene-d8	106	(80 - 130)

NOTE(S):

Calculations are performed before rounding to avoid round-off errors in calculated results.

Bold print denotes control parameters