TAIT ENVIRONMENTAL

Client Sample ID: MWD021_081202_0001

GC/MS Volatiles

Lot-Sample #...: E2H120198-005 **Work Order #...:** E6CN92AA **Matrix.....:** WATER

PARAMETER	RESULT	REPORTING LIMIT	UNITS	MDL
Hexachlorobutadiene	ND	5.0	ug/L	1.5
2-Hexanone	ND	25	ug/L	15
Isopropylbenzene	ND	5.0	ug/L	1.5
p-Isopropyltoluene	ND	5.0	ug/L	1.5
Methylene chloride	ND	5.0	ug/L	1.5
4-Methyl-2-pentanone	ND	25	ug/L	15
Methyl tert-butyl ether	ND	5.0	ug/L	2.5
n-Propylbenzene	ND	5.0	ug/L	2.0
Styrene	ND	5.0	ug/L	1.5
1,1,1,2-Tetrachloroethane	ND	5.0	ug/L	1.5
1,1,2,2-Tetrachloroethane	ND	5.0	ug/L	2.0
Tetrachloroethene	ND	5.0	ug/L	1.5
Toluene	ND	5.0	ug/L	1.5
1,2,3-Trichlorobenzene	ND	5.0	ug/L	2.0
1,2,4-Trichloro-benzene	ND	5.0	ug/L	1.5
1,1,1-Trichloroethane	ND	5.0	ug/L	1.0
1,1,2-Trichloroethane	ND	5.0	ug/L	1.5
Trichloroethene	**1200 E**	**5.0**	**ug/L**	**1.5**
Trichlorofluoromethane	ND	10	ug/L	1.5
1,2,3-Trichloropropane	ND	5.0	ug/L	2.0
1,1,2-Trichlorotrifluoro-ethane	ND	5.0	ug/L	2.0
1,2,4-Trimethylbenzene	ND	5.0	ug/L	1.5
1,3,5-Trimethylbenzene	ND	5.0	ug/L	1.0
Xylenes (total)	ND	5.0	ug/L	4.0
Acrolein	ND	100	ug/L	60
Acrylonitrile	ND	100	ug/L	50
Vinyl acetate	ND	25	ug/L	10
Tetrahydrofuran	ND	50	ug/L	10
2-Chloroethyl vinyl ether	ND	25	ug/L	10

SURROGATE	PERCENT RECOVERY	RECOVERY LIMITS
Bromofluorobenzene	94	(75 - 130)
1,2-Dichloroethane-d4	111	(65 - 135)
Toluene-d8	92	(80 - 130)

NOTE(S):

E Estimated result. Result concentration exceeds the calibration range.