LABORATORY CONTROL SAMPLE DATA REPORT

GC/MS Volatiles

Client Lot #...: E2F280352 **Work Order #...:** E32PL1AC **Matrix.....:** WATER
LCS Lot-Sample#: E2F290000-147
Prep Date.....: 06/28/02 **Analysis Date..:** 06/28/02
Prep Batch #...: 2180147 **Analysis Time..:** 21:55
Dilution Factor: 1 **Instrument ID..:** MSH
Analyst ID.....: 004648

PARAMETER	SPIKE AMOUNT	MEASURED AMOUNT	UNITS	PERCENT RECOVERY	METHOD
Benzene	**10.0**	**11.8**	**ug/L**	**118**	**SW846 8260B**
Chlorobenzene	**10.0**	**10.6**	**ug/L**	**106**	**SW846 8260B**
1,1-Dichloroethene	**10.0**	**11.6**	**ug/L**	**116**	**SW846 8260B**
Toluene	**10.0**	**11.9**	**ug/L**	**119**	**SW846 8260B**
Trichloroethene	**10.0**	**11.8**	**ug/L**	**118**	**SW846 8260B**

SURROGATE	PERCENT RECOVERY	RECOVERY LIMITS
Bromofluorobenzene	111	(75 - 130)
1,2-Dichloroethane-d4	100	(65 - 135)
Toluene-d8	120	(80 - 130)

NOTE(S):

Calculations are performed before rounding to avoid round-off errors in calculated results.

Bold print denotes control parameters