MATRIX SPIKE SAMPLE DATA REPORT

DISSOLVED Metals

Client Lot #...: E2H200306 **Matrix.....:** WATER
Date Sampled...: 08/20/02 10:45 Date Received..: 08/20/02 12:30

PARAMETER	SAMPLE AMOUNT	SPIKE AMT	MEASRD AMOUNT	UNITS	PERCNT RECVRY	RPD	METHOD	PREPARATION-ANALYSIS DATE	WORK ORDER #
Chromium									
	ND	0.200	0.204	mg/L	102		SW846 6010B	08/22-08/23/02	E60VA1CG
	ND	0.200	0.208	mg/L	104	1.5	SW846 6010B	08/22-08/23/02	E60VA1CH
			Dilution Factor: 1						
			Analysis Time..: 13:07		Instrument ID..: M01			Analyst ID.....: 021088	
			MS Run #.....: 2234111						
Cobalt									
	ND	0.500	0.510	mg/L	102		SW846 6010B	08/22-08/23/02	E60VA1CE
	ND	0.500	0.515	mg/L	103	1.1	SW846 6010B	08/22-08/23/02	E60VA1CF
			Dilution Factor: 1						
			Analysis Time..: 13:07		Instrument ID..: M01			Analyst ID.....: 021088	
			MS Run #.....: 2234111						
Copper									
	ND	0.250	0.264	mg/L	106		SW846 6010B	08/22-08/23/02	E60VA1CJ
	ND	0.250	0.269	mg/L	107	1.8	SW846 6010B	08/22-08/23/02	E60VA1CK
			Dilution Factor: 1						
			Analysis Time..: 13:07		Instrument ID..: M01			Analyst ID.....: 021088	
			MS Run #.....: 2234111						
Lead									
	ND	0.500	0.524	mg/L	105		SW846 6010B	08/22-08/23/02	E60VA1CQ
	ND	0.500	0.525	mg/L	105	0.25	SW846 6010B	08/22-08/23/02	E60VA1CR
			Dilution Factor: 1						
			Analysis Time..: 13:07		Instrument ID..: M01			Analyst ID.....: 021088	
			MS Run #.....: 2234111						
Molybdenum									
	0.010	1.00	1.04	mg/L	103		SW846 6010B	08/22-08/23/02	E60VA1CL
	0.010	1.00	1.06	mg/L	105	1.6	SW846 6010B	08/22-08/23/02	E60VA1CM
			Dilution Factor: 1						
			Analysis Time..: 13:07		Instrument ID..: M01			Analyst ID.....: 021088	
			MS Run #.....: 2234111						
Nickel									
	ND	0.500	0.496	mg/L	99		SW846 6010B	08/22-08/23/02	E60VA1CN
	ND	0.500	0.501	mg/L	100	0.82	SW846 6010B	08/22-08/23/02	E60VA1CP
			Dilution Factor: 1						
			Analysis Time..: 13:07		Instrument ID..: M01			Analyst ID.....: 021088	
			MS Run #.....: 2234111						

(Continued on next page)