Site: Shiloh Church Road Site	Contamination Control Survey	Date: 9/23/19

Sample Number	Sample Location Description	removable alpha (cts/100cm^2)	removable alpha (dpm/100cm^2)	removable beta (cts/100cm^2)	removable beta (dpm/100cm^2)	Pass/Fail	Followup/Resolution
1	Blue Portajohn Handle	8	3	199	4	PASS	
2	Decon Sign-In Binder	7	2	182	-4	PASS	
3	Decon Folding Chair (White)	4	0	173	-9	PASS	
4	Decon Trash Can	9	4	187	-2	PASS	
5	ERG 44-10	2	-1	161	-14	PASS	
6	START Drink Cooler	2	-1	201	3	PASS	
7	Dane's Computer Mouse	2	-1	154	-19	PASS	
8	START Rav4 Drivers Side Floormat	3	-1	188	-3	PASS	
9	44-10 Probe on Ratemeter ERT-734	4	0	191	-1	PASS	
10	ER Truck (1447-64) Steering Wheel	1	-2	169	-12	PASS	
11	Command Area Portajohn Handle	0	-3	191	-1	PASS	
12	Command Area Table	1	-2	185	-4	PASS	
13	START Rav4 (LBY 4467) Passenger Handle	3	-1	177	-8	PASS	
14	START Chevy Equinox (LCE 2787) Driver Side Handle	1	-2	171	-11	PASS	
15	Handwash Station Soap Dispenser	3	-1	186	-4	PASS	
16	START Trailer Counter	1	-2	205	5	PASS	
17	Ludlum 2241 Handle (S/N 143886)	5	1	180	-7	PASS	
18	ER Truck (1330-77) Driver Side Handle	6	1	190	-2	PASS	
19	Generator Handle	4	0	219	12	PASS	
20	ER Table	6	1	185	-4	PASS	
21	ER Large Blue Drink Cooler	1	-2	188	-3	PASS	
22	Decon Scissor Handles	5	1	164	-14	PASS	
23	Decon Small Red Drink Cooler	3	-1	188	-3	PASS	

	Smear Counter			Time (min)	alpha	beta
	Alpha	Beta	Background Count	120	98	4660
Make/Model	Ludlum 3030		Sample Count	5		
Serial Number	187982		Pass/Fail Criteria (counts)		37	621
Cal Due Date	2/19/2020		Minimum Detectable Activity (dpm)		12	47
Efficiency	0 3244	0.4264	Surveyed By (print)	[redacted]		
Action Level (dpm)	20	200				

Action levels based on radium-226 (alpha) and strontium-90 (beta). Efficiencies used are for Th-230 (alpha) and Sr-90 (beta). Analysis on samples 1 through 5 conducted on 9/23/19, 6 through 23 conducted on 9/24/19.

QC Check Passed ☑

Initials [redacted]