Residential Well Sampling QA/QC Work Plan (Rev01)

Dimock Residential Groundwater Site

Dimock, Susquehanna County, Pennsylvania
TDD No: TL01-11-12-001
Contract No: EP-S3-10-04

February 28, 2012

EPA Region III

START IV - West

Superfund Technical Assessment and Response Team

Submitted to: Richard Fetzer, On-Scene Coordinator
United States Environmental Protection Agency, Region III
1650 Arch Street
Philadelphia, Pennsylvania 19103

Sampling QA/QC Work Plan (Rev01)
Dimock Residential Groundwater Site
Dimock, Susquehanna County, Pennsylvania

Prepared for:

U.S. Environmental Protection Agency
Region III
Philadelphia, PA

EPA Contract No.	:	EP-S3-10-04
TDD No.	:	TL01-11-12-001
EPA Work Assigner	:	Richard Fetzer
Date Prepared	:	January 9, 2012
Revision date	:	February 28, 2012
Prepared by	:	TechLaw, Inc.

TABLE OF CONTENTS

Section **Page**

1.0 INTRODUCTION 1
2.0 SITE DESCRIPTION 1
3.0 BACKGROUND 2
4.0 PROJECT ORGANIZATION AND RESPONSIBILITIES 2
5.0 PROJECT DESCRIPTION 2
5.1 Objective 2
5.2 Scope of Work 2
6.0 DATA USE OBJECTIVES 2
7.0 SAMPLING APPROACH AND ANALYTICAL PARAMETERS 3
7.1 Residential Home Wells 3
7.2 Surface Water Sampling 5
7.3 Sample Identification Numbers 6
7.3.1 CLP Sample Numbers 6
7.3.2 Station Location Numbers 6
7.3.3 DAS Sample Numbers 6
7.4 Sampling Equipment and Decontamination 6
7.5 Investigation Derived Wastes 6
8.0 QUALITY ASSURANCE/QUALITY CONTROL (QA/QC) PROCEDURES 7
8.1 Quality Control of Field Activities 7
8.2 Sample Documentation 7
8.2.1 Field Logbook 7
8.2.2 Sample Labels/Tags 7
8.2.3 Chain of Custody Record 9
8.2.4 Custody Seals 9
8.3 Sample Packaging, Storage, and Shipping 9
8.4 Field QC Samples 10
8.5 Laboratory QC 10
8.6 Data Validation 10
9.0 SCHEDULE OF ACTIVITIES 11
10.0 DELIVERABLES 11
11.0 REFERENCES 12

TABLES:
Table 1 - Field and QA/QC Sampling Summary
Table 2 - Sample Analytical Requirements Summary

Appendix A - Deliverable Requirements for Level IV Data Package

1.0 INTRODUCTION

On December 19th, 2011, EPA Region III On-Scene Coordinator (OSC) Richard Fetzer tasked TechLaw, Inc. (TechLaw) Superfund Technical Assessment and Response Team (START) to perform a removal site evaluation at the Dimock Residential Groundwater Site (site) located at or near Pennsylvania (PA) Route 29 in Dimock, Susquehanna County, Pennsylvania. The purpose of the assessment is to provide information to EPA to assist in determining if residential home wells have been impacted by nearby gas well installation and development activities. Sampling activities will include the collection of residential home well groundwater samples and surface water samples. These sampling activities will be conducted under Technical Direction Document (TDD) No. TL01-11-12-001, START Contract No. EP-S3-10-04. Revisions have been made to the Sampling QA/QC Work Plan to address changes in sample bottle ware requirements and minor changes to sampling procedures. These changes were implemented starting on January 30, 2012.

2.0 SITE DESCRIPTION

The Dimock Residential Groundwater Site is located in the rural community of Dimock Township in northeastern Pennsylvania (pop. 1,497 – 2010 Census). Degradation of drinking water and surface water quality from contamination claimed to be associated with Marcellus shale drilling and hydraulic fracturing (a.k.a. fracking) operations has been reported by local private well owners. Privately owned wells constitute the primary source of drinking water for residents in the area. Drilling and production activities involving deep shale gas extraction is prevalent throughout Susquehanna County.

The site includes affected and potentially affected media, namely ground water and surface waters, in the rural area surrounding the intersection of State Route 29 and County Route 2024 in Dimock Township. The coordinates for this location are 41.746411 north latitude, and 75.898498 west longitude. Surface waters in the area enter tributaries of Burdick Creek located east/southeast from the site. Burdick Creek flows to Meshoppen Creek also located east/southeast from the site. Meshoppen Creek flows southwest and confluences the Susquehanna River at Meshoppen, PA. Surface water impoundments and/or ponds and lakes are observed in aerial photos to be present near the site. Topographic relief in the vicinity of the site is approximately 400 feet ranging from approximately 1,100 feet to 1,500 feet above mean sea level (amsl). The site is located within the glaciated low plateaus section of the Appalachian Plateaus Province. Surficial bedrock is comprised of the Devonian Catskill Formation having sandstone, siltstone, shale, mudstone and conglomerate lithology.

The Pennsylvania Ground Water Information System database (PAGWIS) identifies 44 ground water withdrawal wells within a 2-mile radius of the site, although more wells are likely to be present. Most of the wells recorded in the PAGWIS are used for domestic purposes. Depths of 19 of these wells are recorded ranging from 125 to 700 feet deep with a median depth of about 250 feet. Yields from 42 of the wells are recorded as ranging from 1-50 gallons per minute (gpm) with a median yield of 13.7 gpm.

3.0 BACKGROUND

Since 2009, the site has received widespread publicity beginning with reports of methane migration into local domestic water supplies following Marcellus Shale drilling operations in the area. Ground water sampling activities have also identified the presence of other organic and inorganic contaminants in the private-use wells which may potentially be associated with nonconventional deep shale drilling activities. The origin of the contaminants has not been fully determined.

4.0 PROJECT ORGANIZATION AND RESPONSIBILITIES

EPA On-Scene Coordinator Richard Fetzer will provide overall direction to TechLaw (START) staff concerning project sampling requirements, objectives, and schedule. The START Site Leader is the primary point of contact with the EPA OSC. The Site Leader is responsible for the development and completion of the Sampling QA/QC Work Plan, project team organization, and supervision of all project tasks, including reports and deliverables.

5.0 PROJECT DESCRIPTION

5.1 Objectives

The objective of the sampling activity is:

- To assess for the presence and origin of substances that may present a threat to the health of persons ingesting, contacting or engaging in typical residential or recreational uses of groundwater or surface water. The analytical methods selected are based in part on contaminants that may be present due to the natural gas exploration, drilling or hydraulic fracturing activities located in the region.

5.2 Scope of Work

The scope of work includes collection of approximately forty to sixty residential home well samples in the vicinity of Dimock Township. Tap water samples will be collected at homes where access has been granted to EPA officials by property owners. Additionally, it is anticipated that as many as twelve surface water samples may be collected from nearby water bodies.

6.0 DATA USE OBJECTIVES

The following data quality objectives apply to this project:

Program Area	Sampling Objective	Data Type
Removal	Determine presence/concentration of contamination	Definitive

7.0 SAMPLING APPROACH AND ANALYTICAL PARAMETERS

Table 1, Field QC and Sampling Summary and Table 2, Sample Analytical Requirements Summary, include a summary of the numbers of samples, matrices, analytical parameters/methods, quality control (QC) samples, sample preservation, holding times, and containers. Samples will be collected using certified pre-cleaned sample bottles.

7.1 Residential Home Wells

Residential well samples will be collected in accordance with the EPA Environmental Response Team (ERT) Standard Operating Procedure (SOP) No. 2007 (ERT, 1995). Samples are anticipated to be collected from a valve closest to the well head (wellhead sample) and from the kitchen faucet (tap sample) within each home. Inspection of the water system may be required to identify the appropriate valve sampling location and to determine if it is downstream or upstream of any treatment apparatus. The water samples will be collected simultaneously from the valve closest to the wellhead and the kitchen faucet, after purging procedures are completed and water quality parameters are stabilized as described below.

Wellhead Sample

1. For the sample closest to the wellhead, the spigot will first be examined to determine if it is equipped with an aerator. If the spigot is equipped with an aerator, it shall be removed before purging. A garden hose will be connected to the spigot to direct the purge water away from the home. The spigot will be allowed to purge for a target time of 1 hour. However, if one of the following conditions exists, different purging procedures will be followed: 1) the well is currently in routine use; 2) the well has a history of being pumped dry; 3) the pump is in poor operating condition. If any of these three conditions exist, the well will only be purged for approximately 15 minutes and then water quality parameter monitoring will begin following procedures presented below. The volume of the purge water will be measured periodically using a stop watch and a large graduated cylinder or equivalent container. Once the target time of 1 hour, or alternately 15 minutes, has been reached, water quality parameters will be recorded using an YSI 556 water quality meter or equivalent that is equipped with data logging capability and flow-through cell. The flow-through cell will be connected to the spigot using a dedicated, clean adapter and flexible tubing. Additionally, water quality parameters will be measured and recorded on field data sheets at approximately 3-5 minute intervals (in addition to instrument data logging) to determine when parameters stabilize. Stabilization will be achieved after all parameters have stabilized for three consecutive readings using the following criteria:

pH $\pm$ 0.1 unit
Specific Conductance $\pm$ 3%
Dissolved Oxygen $\pm$ 10%
Oxidation Reduction Potential $\pm$ 10 mV
Temperature $\pm$ 3%

These criteria are initial guidelines; professional judgment in the field will be used to

determine on a well-by-well basis when stabilization occurs. After parameter stabilization is achieved, purging at the spigot nearest the wellhead will be stopped until purging procedures for the residential plumbing are completed as described under the Tap Water section below. The samples from the spigot nearest the wellhead and the kitchen faucet (tap sample) will be collected simultaneously.

2. When purging and stabilization procedures as described in this section and the Tap Sample section are achieved, sample collection will commence simultaneously at both the spigot nearest the wellhead and the kitchen faucet. A dissolved gas sample may be collected first in the sequence of samples if the water is found to be effervescent. The water quality instrument, flow-through cell and tubing will be removed from the spigot. A new clean length of tubing will be attached to the adapter. The sample container will be submerged in a new, clean and dedicated plastic bucket containing the sample media in order to prevent exposure of the sample to the atmosphere. (See attached Isotech procedure titled "Collection of Ground Water Samples from Domestic and Municipal Water Wells for Dissolved Gas Analysis").

3. Once the dissolved gas sample is collected, the pre-made adapters will be removed and a 1-L HDPE container will be filled to perform field screening. **Note: if space limitations require continued use of the adaptor and tubing to collect samples, then Teflon® tubing will be attached to the adaptor and used to collect the remaining sample bottles.** Field measurements will consist of turbidity, alkalinity, ferrous iron, and dissolved sulfide. Turbidity (Standard Method 180.1) will be measured using a HACH 2100Q portable turbidimeter (or equivalent instrument). Alkalinity will be measured by titrating ground water with 1.6N H2S04 to the bromcresol green-methyl red endpoint using a HACH titrator (HACH method 8203, equivalent to Standard Method 23208 for alkalinity). Ferrous iron will be measured using the 1, 10 phenanthroline colorimetric method (HACH DR890 spectrometer, HACH method 8146, equivalent to Standard Method 3500-Fe B for wastewater). Dissolved sulfide will be measured using the methylene blue colorimetric method (HACH DR890 spectrometer; HACH method 8131, equivalent to Standard Method 4500-S2-D for wastewater).

4. The VOC, SVOC and remaining organic sample volumes, respectively, will be collected in sequence by directly filling the sample containers from the spigot. After collection of the organics samples, volumes for all the remaining parameters included in Table 1 shall be collected directly from the spigot. Sample volumes for bacteria analysis should be collected last, but if necessary may be moved forward in the order of collection as long as it is taken after all organic samples are collected.

5. For the total metals analysis, a 1-L HDPE container will be filled. An additional 1-L HDPE container will be filled for dissolved metals analysis. This sample will be filtered using a peristaltic pump and an in-line high-capacity (0.45 micron) filter. At least 100 ml of water will be allowed to pass through the filter before the sample is collected.

6. The sample for bacteria analysis will be collected by first using an alcohol swab/wipe to disinfect the sampling port or spigot. Two applications of alcohol will be applied to the spigot, with the first application removing the gross contamination and the second application for final cleaning of the spigot surface. The alcohol will be allowed to evaporate from the spigot surface before sampling proceeds. The spigot will be turned on to flush any residue from the spigot surface before collection of the sample occurs. During sample collection, care will be made to not touch the mouth of the bottle to the spigot.

Tap Sample

1. After the required purging and parameter stabilization is achieved at the spigot nearest the wellhead as described in the preceding section, purging of the internal building plumbing/piping system will be accomplished from the kitchen faucet (tap). The faucet will first be examined to determine if it is equipped with an aerator, which will be removed if present. The faucet will be turned on and allowed to run for approximately 15 minutes to flush any water from within the indoor plumbing. After the 15-minute purge is completed, sampling will begin with simultaneous sample collection from the spigot nearest the wellhead and from the kitchen faucet.

2. Sample volumes will be collected in similar sequence and in general accordance with the procedures outlined for well head samples stated above, except for the field screening sample. Samples for dissolved gas analysis will not be collected at the tap.

3. Towards the end of sample collection at each residence, an additional 1-L HDPE container will be filled to perform field screening for turbidity, alkalinity, ferrous iron, and dissolved sulfide in accordance with the procedure noted in item 3 above. This will be conducted in order to evaluate changes in water quality across the sampling period.

All samples will be placed on ice after collection and will be brought to the sample management trailer to be prepared for shipment to approved laboratories.

Analytical services will be coordinated through EPA and include using the EPA Regional Laboratory, the EPA Contract Laboratory Program (CLP) and/or Tier IV subcontracted laboratory services through TechLaw.

7.2 Surface Water Sampling

Up to twelve surface water samples may be collected from locations near the site. The surface water samples will be collected in accordance with ERT SOP No. 2013, utilizing the direct method (ERT, 1994). Surface water sample media will be collected directly into laboratory certified pre-cleaned sample bottles as specified in Table 2. All samples will be placed on ice after collection. The analyses to be conducted on the surface water samples are summarized in Table 1.

Analytical services will be coordinated through EPA and include using the EPA Regional Laboratory, the EPA Contract Laboratory Program (CLP) and/or Tier IV subcontracted laboratory services through TechLaw.

7.3 Sample Identification Numbers

7.3.1 CLP Sample Numbers

No samples are scheduled to be analyzed by CLP laboratories.

7.3.2 Station Location Numbers

Sample Station Location numbers will be assigned by the sampling team to correspond with the location and the type of sample collected. The sample station location No. format will be as follows:

XX##-PFz, where XX is:

HW	= a ground water sample from a Dimock residential well
SW	= a surface water sample
TB	= trip blank
FB	= field blank
EB	= equipment blank
F	= indicates a sample is filtered for metal analysis
P	= indicates a post filtration sample
z	= indicates a duplicate sample

and where ## is:

##	= the unique identifier for each residential well sampled. This identifier will be related in a separate database to the specific residence being sampled.

7.3.3 DAS Sample Numbers

Station location numbers will be used in lieu of DAS sample numbers for this project.

7.4 Sampling Equipment and Decontamination

Dedicated, disposable sampling equipment will be used by TechLaw whenever possible.

7.5 Investigation Derived Wastes

TechLaw field team members will make every effort to minimize the generation of investigation-derived wastes (IDW) throughout the field event. Purge water for residential home well samples will be discharged in accordance with the Groundwater

Monitoring and Maintenance Manual (PADEP, 2001). Disposable personal protective clothing and/or any sampling equipment generated during field activities will be bagged in opaque plastic garbage bags, and disposed of appropriately.

8.0 QUALITY ASSURANCE/QUALITY CONTROL (QA/QC) PROCEDURES

8.1 Quality Control of Field Activities

The START Site Leader is responsible for ensuring that sample quality and integrity are maintained in accordance with the QA/QC Procedures, and that the sample labeling and documentation is performed as described in Section 8.2 of this sampling plan.

8.2 Sample Documentation

All sample documentation will be completed legibly using indelible black or blue ink. Any corrections or revisions will be made by lining through the incorrect entry and by initialing and dating the error.

At each sampling location GPS coordinates will be collected using a GPS unit. These coordinates are to be recorded on the field data sheet or in the field logbook. The field data sheet will be used to document pertinent field screening measurements and sample information. Photographs of each house and each sampling spigot/faucet will be collected. The date and time stamp option will be selected (if available) on each camera. The data logging option (if available) will be selected on each water quality instrument used.

8.2.1 Field Logbook

The use of field logbooks by START for site documentation will be consistent with TechLaw SOP 03-01-04, Maintaining a Field Logbook (TechLaw, 2011a). The field logbook is a descriptive notebook detailing site activities and observations so that an accurate account of field procedures can be reconstructed. All entries will be dated and signed by the individual making the entries, and include (at a minimum) the following:

1. Site name and project number.
2. Name(s) of personnel on site.
3. Dates and times of all entries (military time preferred).
4. Descriptions of all site activities, including site entry and exit times.
5. Noteworthy events and discussions.
6. Weather conditions.
7. Site observations.
8. Identification and description of samples and locations.
9. Subcontractor information and names of on-site personnel.
10. Date and time of sample collections, along with chain of custody information.
11. Record of photographs.
12. Site sketches.

8.2.2 Sample Labels/Tags QC

Sample labels and tags must clearly identify the particular sample. Required information for sample labels and tags is presented in *Contract Laboratory Program Guidance for Field Samplers*, EPA Publication 540-R-09-03, Final (January 2011) and is provided below.

Sample bottle labels must include the following information:

1. Sample number, as applicable;
2. Case No.;
3. Preservative(s);
4. Analysis/fraction.

Additional information may be included on the label, such as the Station Location (Sampler-assigned sample No.), date and time collected, etc.

Sample tags must include the following information:
1. Sample number, as applicable;
2. Case No.;
3. Station No. and/or Station Location No. (assigned by sampler);
4. Date sample was collected (month, day, and year);
5. Time sample was collected (in military time);
6. Preservative, if any (specify "None" if sample is not preserved);
7. Type of sample (grab or composite);
8. Analysis/fraction requested;
9. Sampler's names/signature(s);

Sample labels will be securely affixed to the sample container. Tie-on sample tags will be properly secured around the neck of the container.

For this project, TechLaw was given an exemption by the EPA laboratories on affixing sample tags on each container. In lieu of affixing sample tags to each sample container, a sample bottle label will be adhered to each bottle. The information which is required on the sample tag will be included on the sample bottle label.

8.2.3 Chain of Custody Record QC

Proper chain of custody will be maintained from the time the sample is collected to its final deposition. Every transfer of custody will be noted and signed. When samples (or groups of samples) are not under direct control of the individual responsible for them, they will be stored in a locked container sealed with a Custody Seal.

The Chain of Custody record/EPA Traffic Report (COC/TR) will include (at minimum) the following information:

1. Sample number, as applicable;
2. Case No.;
3. Sample matrices;
4. Specify sample type (grab or composite);
5. Analyses requested;
6. Laboratory turnaround time (TAT) [*Note: This does not include the TAT for data validation. If preliminary results (PR) are required, this must be specified on the COC.*)
7. Preservative(s);
8. Station location identifier (sampler assigned sample No.);
9. Date and time sample collected;
10. Field QC information (identify trip/field/blanks only as "Field QC");
11. Specify samples to be used for laboratory QC (MS/MSD);
12. Name(s) and signature(s) of sampler(s);
13. Signature(s) of any individual(s) with control over samples;
14. Carrier, air bill No., and date of the shipment.

8.2.4 Custody seals QC

Custody seals will be used on all shipping containers used to ship samples. Custody Seals demonstrate that a container has not been tampered with or opened. The individual shipping the sample(s) will sign and date the seal, affixing it in such a manner that the container cannot be opened without breaking the seal. The name of this individual, along with a description of the sample packaging, will be noted in the field logbook. EPA Region III does not require custody seals on individual sample containers and has specifically directed samplers not to use custody seals on individual sample containers containing samples for volatile organics analysis (VOA).

8.3 Sample Packaging, Storage, and Shipping

In accordance with TechLaw SOP 04-02-01, Packaging and Shipping Samples – Environmental Procedures (TechLaw, 2011b), and *Contract Laboratory Program Guidance for Field Samplers*, sample containers will be labeled and shipped with a label and sample tag affixed to each container. Samples will be placed in plastic zipping bags. Bagged containers will be placed in appropriate transport containers and the containers will be packed with appropriate absorbent material and bubble wrap. All sample/traffic

reports/COC documents will be affixed to the underside of each transport container lid. The lid will be sealed with shipping tape and custody seals affixed to the transport container. Transport containers will be labeled with the origin and destination locations. Regulations for packaging, marking, labeling, and shipping of hazardous materials and wastes are promulgated by the U.S. Department of Transportation (DOT). Air carriers which transport hazardous materials, in particular, Federal Express, require compliance with the current International Air Transport Association (IATA) Regulations, which apply to the shipment and transport of hazardous materials by air carrier. TechLaw will follow IATA regulations to ensure compliance.

8.4 Field QC Samples

Field QC will consist of one field duplicate for every ten field samples, or one per matrix if fewer than ten are collected. Duplicate samples will be documented on the Field Data Sheet and on the Traffic Report (TR)/COC. The field duplicate will test the reproducibility of sampling procedures and analytical procedures. A trip blank will be collected and included in all coolers shipped that contain samples for VOC, dissolved gas, and GRO analyses. Trip blanks will be prepared for each sample team each morning and will accompany the sampling team throughout the morning sample collection event. A second set of trip blanks will accompany each sampling team for the afternoon sample collection event. A field blank will be collected to ensure the cleanliness of sample containers and to ensure that no cross-contamination has occurred during sample collection, preservation, and shipment, as well as in the laboratory. One field blank will be collected each day that sampling is conducted. The field blank sample will include a sample collected each day from the in-line filter which is used to collect samples for the dissolved metals analysis. This field blank sample for the in-line filter will also be considered an equipment blank. An equipment blank will also be collected for any new piece of sampling equipment to be used, on a one-time basis.

8.5 Laboratory QC

Laboratory QC will be in accordance with the applicable method requirements. TechLaw will designate one sample per matrix in each batch/sample delivery group (SDG) submitted to the laboratory to be used for a matrix spike/matrix spike duplicate (MS/MSD) or appropriate QC sample required by the method. A batch/SDG is defined as up to 20 samples of a specific matrix submitted for a specific case; or all the samples of a specific matrix (up to a maximum of 20) received by the laboratory for a specific case within a 7 calendar day period (3 calendar day period for 7 day TAT), with the period beginning with the receipt of the last sample in the SDG.

8.6 Data Validation

Analytical data generated by the EPA OASQA laboratory and other EPA laboratories will be reviewed and validated in accordance with OASQA standard procedures and/or standard procedures for the applicable EPA laboratories. Other analytical data for organic analyses generated under this Sampling QA/QC Work Plan, including Tier IV data, will be evaluated in accordance with EPA *Region III Modifications to National*

Functional Guidelines for Organic Data Review Multi-Media, Multi-Concentration (OLMO1.0-OLMO1.9) (September 1994) to Data Validation Level M2, and in accordance with *EPA Region III Modifications to the Laboratory Data Validation Functional Guidelines for Evaluating Inorganics Analyses* (April 1993) at the IM2 Level. Validation for the analytical services subcontract arranged through TechLaw will be requested through EPA and conducted by the EPA ESAT contractor. Analytical data packages from EPA Regional Laboratories and NAREL will undergo a QA review by the EPA Environmental Response Team's (ERT) Scientific, Engineering, Response, & Analytical Services (SERAS) contractor.

9.0 SCHEDULE OF ACTIVITIES

The schedule for the site is projected as follows:

Task Description	Start Date	End Date
Mobilize to area	1/23/12	1/23/12
Sample collection; sample packaging; sample shipping to laboratory	01/23/12	TBD

10.0 DELIVERABLES

The following deliverables will be provided under this project:

Analytical Data

Three types of reports will be prepared under this project:

Preliminary – Initial laboratory data report consisting primarily of analytical test results. The data in this report will not have undergone internal laboratory QA review and will not include the full QA/QC documentation required in the final report.

Final – Final analytical data report provided by the respective laboratories. This report will contain the final analytical results following completion of the internal laboratory QA review. The Final data package will also include all necessary documentation required for data validation (Appendix A).

Validated – Final data that has undergone data validation review in accordance with the guidance documents referenced in Section 8.6.

- Expedited preliminary data turnaround time (5 days) will be provided on the following list of compounds/tests:

coliform bacteria	aluminum
bis(2-ethylhexyl) phthalate (DEHP)	arsenic
ethylene glycol	lithium
2-methoxyethanol (Ethylene glycol monomethyl ether)	manganese
methane	sodium
2,2'oxybisethanol (diethylene glycol)	iron
triethylene glycol	

Preliminary reports are submitted by each laboratory within 15 business days (or sooner, with the exception of Isotech) after receipt of a batch of samples at the laboratory. A batch is defined in Section 8.5. Each laboratory will then submit a final report within 21 days of receipt (28 days for Isotech). The final report is to include all results for all parameters associated with the assigned analyses for that particular laboratory. A Level IV (CLP-like) data package, or equivalent, will be included in the final report. Requirements for a Level IV data package are included in Appendix A.

Each laboratory will also provide an electronic data deliverable (EDD), where applicable. The EDD is to be submitted with the final report. EDD files will be uploaded to SCRIBE.

11.0 REFERENCES

EPA, 2011. U.S. Environmental Protection Agency, *Contract Laboratory Program (CLP) Guidance for Field Samplers, Final*, Office of Solid Waste and Emergency Response (OSWER) publication EPA540-R-07-006, Washington, D.C. January.

ERT, 1994. U.S. Environmental Protection Agency Environmental Response Team. Standard Operating Procedure for Surface Water Sampling, SOP# 2013. January 26.

ERT, 1995. U.S. Environmental Protection Agency Environmental Response Team. Standard Operating Procedure for Groundwater Well Sampling, SOP# 2007. January 26.

Isotech, 2011. Isotech Laboratories, Inc., Collection of Ground Water Samples from Domestic and Municipal Water Wells for Dissolved Gas Analysis, Website Accessed December 2011: < http://www.isotechlabs.com/customersupport/samplingprocedures/DGbottle.pdf>

PADEP, 2001. Pennsylvania Department of Environmental Protection, *Groundwater Monitoring Guidance Manual,* Document number 383-3000-001, dated January 1st, 1999, revised December 1st, 2001.

TechLaw, 2011a. TechLaw, Inc., Standard Operating Procedures, *Field Documentation Procedures - Maintaining a Field Logbook, 03-01-04*, Chantilly, VA. March.

TechLaw, 2011b. Standard Operating Procedures, *Packaging and Shipping Samples-Environmental Procedures, 04-02-01,* Chantilly, VA. March.

TABLE 1 - 01/12/12 (Rev01 - 2/3/12)
FIELD AND QC SAMPLING SUMMARY
DIMOCK RESIDENTIAL GROUNDWATER SITE
DIMOCK, SUSQUEHANNA COUNTY, PENNSYLVANIA

Parameter/Method	Matrix	Field Samples	QC Sample Summary					Total Field and QA/QC Analyses (not including MS/MSD)
			Dup	Trip[1,2] Blanks	Field[1,3] Blanks	Equip[1] Blanks	MS/MSD	
EPA R2 Lab								
Methylene Blue Active Substances (MBAS) (SM 5540C)	drinking water	120	12	-	20	-	-	152
EPA R3 Lab								
Anions: Chloride, Bromide, Fluoride, Orthophosphorus as P, Sulfate as SO4 (300.0)	drinking water	120	12	-	20	-	16	152
Glycols incl. 2-Butoxyethanol (Modified 8321)	drinking water	120	12	-	20	-	8	152
Metals Dissolved: Al, As, B, Ba, Be, Ca, Cd, Co, Cr, Cu, Fe, Hg, K, Li, Mg, Mn, Na, Ni, Sb, Se, Sn, Sr,Ti, Tl, U, V, Zn (200.7/200.8/245.1)	Filtered drinking water	120	12	-	-	20	16	152
Metals: Al, As, B, Ba, Be, Ca, Cd, Co, Cr, Cu, Fe, Hg, K, Li, Mg, Mn, Na, Ni, Sb, Se, Sn, Sr,Ti, Tl, U, V, Zn (200.7/200.8/245.1)	drinking water	120	12	-	20	-	16	152
Semi-Volatiles (TCL plus TICs) (OLC03.2)	drinking water	120	12	-	20	-	8	152
Solids, Total Dissolved (TDS) (SM 2540C)	drinking water	120	12	-	20	-	16	152
Solids, Total Suspended (TSS) (SM 2540D)	drinking water	120	12	-	20	-	16	152
Volatiles + Acrylonitrile (TCL + TICs) (OLC03.2)	drinking water	120	12	80	20	-	12	232
Wet Chemistry: - Phosphorus, Total (365.4); - Nitrate/Nitrite (353.2); - Nitrogen; Total (353.2)	drinking water	120	12	-	20	-	16	152
Oil & Grease (HEM) (1664A)	drinking water	120	12	-	20	-	-	152
Alcohols: Ethanol, methanol, 1-propanol, 1-butanol, 2-butanol (8015D)	drinking water	120	12	-	20	-	8	152
EPA R9 Lab								
Dissolved Gases, Methane, Ethane, Ethene, Propane, Butane (RSK-175, or equiv - EPA R9 SOP 325)	drinking water	120	12	80	20	-	12	232
DRO (8015M, or equiv-EPA R9 SOP 385)	drinking water	120	12	-	20	-	16	152
GRO (8015M, or equiv-EPA R9 SOP 380)	drinking water	120	12	80	20	-	12	232
NAREL								
Alpha Spec (Th-228, Th-230, Th-232) (DOE HASL 300)	drinking water	120	12	-	20	-	-	152
Alpha Spec (U-234, U-235, U-236, U-238) (DOE HASL 300)	drinking water	120	12	-	20	-	-	152
Gamma Spec Bi-212, Bi-214, K-40, Ra-226, Ra-228, Th-232, Th-234, U-234, U-235, U-238 (901.1)	drinking water	120	12	-	20	-	-	152
Gross Alpha/Beta (900.0)	drinking water	120	12	-	20	-	-	152
Ra-226 (903.1)	drinking water	120	12	-	20	-	-	152
Ra-228 (904.0)	drinking water	120	12	-	20	-	-	152
Tier IV - Pace Analytical								
Ethylene Glycol (8015M)	drinking water	120	12	-	20	-	4	152
Tier IV - Test America								
Glycols (8015M)	drinking water	36	4	-	-	-	2	40
Tier IV - Northeastern Environmental Lab								
Bacteria (fecal & total coliform, HPC) (SM 9222B; SM 9215B w/R2A medium)	drinking water	120	12	-	20	-	-	152
Tier IV - Isotech								
Isotech - d13C and d2H of methane; - Complete compositional analysis of headspace gas; - Stable isotopes of water (O,H)	drinking water	18	2	-	-	-	-	20

Notes:

1. This QA sample will be an aqueous matrix.
2. Trip blank samples will be collected at a rate of one VOA, one RSK-175, and one GRO per residence. One trip blank will be included in every cooler containing samples for these analyses.
3. Field blank samples will be collected at a rate of 1 per day

Key:

Bkgd = Background
MS/MSD = Matrix Spike/Matrix Spike Duplicate
CRQL = Contract-Required Quantitation limit.
QA/QC = Quality assurance/quality control
Dup = Duplicate

TABLE 2 - 01/12/12 (Rev01 - 2/3/12)
SAMPLE ANALYTICAL REQUIREMENTS SUMMARY
DIMOCK RESIDENTIAL GROUNDWATER SITE
DIMOCK, SUSQUEHANNA COUNTY, PENNSYLVANIA

Analytical parameter and Method	Matrix	Sample Preservation	Holding Time (Days)	Sample Container(s)			
				Qty	Vol (ml)	Bottle Type	Comments
EPA R2 Lab							
Methylene Blue Active Substances (MBAS) (SM 5540C)	drinking water	Ice, 4°C	2	1	500	HDPE	
EPA R3 Lab							
Anions: Chloride, Bromide, Fluoride, Orthophosphorus as P, Sulfate as SO4 (300.0)	drinking water	Ice, 6°C	28	1	500	HDPE	
Glycols incl. 2-Butoxyethanol (Modified 8321)	drinking water	Ice, 6°C	7	1	40	Glass Vial	No Headspace
Metals Dissolved: Al, As, B, Ba, Be, Ca, Cd, Co, Cr, Cu, Fe, Hg, K, Li, Mg, Mn, Na, Ni, Sb, Se, Sn, Sr,Ti, Tl, U, V, Zn (200.7/200.8/245.1)	Filtered drinking water	pH<2 with HNO3 and cool with ice, 4°C	180	1	500	HDPE	
Metals: Al, As, B, Ba, Be, Ca, Cd, Co, Cr, Cu, Fe, Hg, K, Li, Mg, Mn, Na, Ni, Sb, Se, Sn, Sr,Ti, Tl, U, V, Zn (200.7/200.8/245.1)	drinking water	pH<2 with HNO3 and cool with ice, 4°C	180	1	500	HDPE	
Semi-Volatiles (TCL plus TICs) (OLC03.2)	drinking water	Ice, 6°C	7	1	1000	Amber Glass	Teflon Lined Lids
Solids, Total Dissolved (TDS) (SM 2540C)	drinking water	Ice, 6°C	7	1	500	HDPE	
Solids, Total Suspended (TSS) (SM 2540D)	drinking water	Ice, 6°C	7	1	500	HDPE	
Volatiles + Acrylonitrile (TCL + TICs) (OLC03.2)	drinking water	2 drops of 1:1 HCl, pH<2, Ice, 6°C	14	3	40	Glass Vial	Teflon Lined Lids No Headspace
Wet Chemistry: - Phosphorus, Total (365.4); - Nitrate/Nitrite (353.2); - Nitrogen; Total (353.2)	drinking water	pH<2, H2SO4, and cool with ice, 4°C	28	1	500	HDPE	
Oil & Grease (HEM) (1664A)	drinking water	pH<2, H2SO4, and cool with ice, 4°C	28	3	1000	WM Amber Glass	Teflon Lined Lids
Alcohols: Ethanol, methanol, 1-propanol, 1-butanol, 2-butanol (8015D)	drinking water	Ice, 6°C	7	2	40	Glass Vial	Teflon Lined Lids No Headspace
EPA R9 Lab							
Dissolved Gases, Methane, Ethane, Ethene, Propane, Butane (RSK-175, or equiv - EPA R9 SOP 325)	drinking water	pH<2 with HCl and cool with ice, 4°C	7	2	40	Glass Vial	
DRO (8015M, or equiv-EPA R9 SOP 385)	drinking water	Ice, 4°C	7 (1)	2	1000	Amber Glass	Teflon Lined Lids
GRO (8015M, or equiv-EPA R9 SOP 380)	drinking water	pH<2 with HCl and cool with ice, 4°C	14	2	40	Glass Vial	Teflon Lined Lids No Headspace
NAREL							
Alpha Spec (Th-228, Th-230, Th-232) (DOE HASL 300)	drinking water	pH<2 with HNO3 and cool with ice, 4°C	180	1	1000	HDPE	
Alpha Spec (U-234, U-235, U-236, U-238) (DOE HASL 300)	drinking water	pH<2 with HNO3 and cool with ice, 4°C	180	1	1000	HDPE	
Gamma Spec Bi-212, Bi-214, K-40, Ra-226, Ra-228, Th-232, Th-234, U-234, U-235, U-238 (901.1)	drinking water	pH<2 with HNO3 and cool with ice, 4°C	180	1	1000	HDPE	
Gross Alpha/Beta (900.0)	drinking water	pH<2 with HNO3 and cool with ice, 4°C	180	1	1000	HDPE	
Ra-226 (903.1)	drinking water	pH<2 with HNO3 and cool with ice, 4°C	180	1	1000	HDPE	
Ra-228 (904.0)	drinking water	pH<2 with HNO3 and cool with ice, 4°C	180	1	1000	HDPE	
Tier IV - Pace Analytical							
Ethylene Glycol (8015M)	drinking water	Ice, 4°C	14	2	40	Glass Vial	Teflon Lined Lids No Headspace
Tier IV - Test America							
Glycols (8015M)	drinking water	Ice, 4°C	14	2	40	Glass Vial	Teflon Lined Lids No Headspace
Tier IV - Northeastern Environmental Lab							
Bacteria (fecal & total coliform, HPC) (SM 9222B; SM 9215B w/R2A medium)	drinking water	Ice, 4°C (.008% Na2S2O3 if residual Cl- present)	1	1	125	Pre- Sterilized Poly	
Tier IV - Isotech							
Isotech - d13C and d2H of methane; - Complete compositional analysis of headspace gas; - Stable isotopes of water (O,H)	drinking water	Ice, 4°C, biocide pill in sample container	180	1	1000	HDPE	

KEY:

°C = degrees Celsius	HN03 = Nitric Acid	QL = Quantitation Limit
CLP = Contract Lab Program	HPC = Heterotrophic Plate Count	TCL = Target Compound List
D2H = delta of deuterium	ml = milliliter	TICs = Tentatively Identified Compounds
H2SO4 = Sulfuric Acid	Na2S2O3 = Sodium Thiosulfate	ug/L = micrograms per liter
HDPE = High density polyethylene	pH = potential Hydrogen	(1) Days to extract

ISOTECH®
ISOTECH LABORATORIES INC
1308 Parkland Court Champaign, IL 61821 • (877) 362-4190 • www.isotechlabs.com

Collection of Ground Water Samples from Domestic and Municipal Water Wells for Dissolved Gas Analysis

These instructions are based on sampling protocol created by Anthony Gorody, adopted by the Colorado Oil and Gas Conservation Commission, and are reproduced here with their permission.

The basic technique is to fill a white 5 gallon bucket with source water and then fill the 1 liter sample collection bottle fully immersed in the bucket.

When sampling from a pressurized water system, it is recommended to use an outdoor spigot or other source which bypasses any water treatment systems (i.e. water softeners, etc.).

To collect a sample for isotopic and chromatographic analysis from water that is not effervescent, using 1L bottle with septum cap:

After purging the well, fill the 5 gallon bucket with water. Attach a nozzle and 12" length of ¼ inch diameter tubing to the end of the 5/8 inch hose connected to a faucet. Make sure that the flow rates through the tubing are low. Remove the cap of the 1 L bottle and fill it with water. Once the bottle filled, immerse it in the 5 gallon bucket full of water, keeping the tubing at the bottom of the bottle. Place the bottle at the bottom of the bucket under a head of water, and keep water flowing at a low rate until another 2 volumes of water have been displaced from the bottle. Then slowly lift the tubing out of the bottle and immediately cap it under water. No air should be allowed into the 1 L bottle. When finished, tape the cap to the bottle around the neck, pack the bottle upside down in ice, and ship it overnight.

To collect a headspace gas sample from an effervescent water well:

Fill the bottle with water. Submerge the bottle into the 5 gallon bucket filled with well water and invert it. Insert the ¼ inch tubing into the bottle, increase the flow rate to 2-3 gpm and allow the bubbling gases to displace water in a headspace until 1/4 to 1/2 of the water in the bottle has been displaced. Seal the container under water with the septum and screw cap, tighten it securely. When finished, tape the cap to the bottle around the neck, pack the bottle upside down in ice, and ship it overnight.

Please note Isotech's receiving hours of **Monday thru Friday** 8:00 am to 4:30 pm.
Ship samples to:

Isotech Laboratories, Inc.
1308 Parkland Court
Champaign, IL 61821

These instructions have been provided to simplify the collection of samples for dissolved gas analysis. Although we try to foresee and avoid problems in the field, it is never possible to predict every situation. If you encounter any difficulties, or if any additions or changes in these instructions would be beneficial, please let us know. Isotech Laboratories, Inc. makes no warrantee as to the applicability and/or safety of the procedures described herein.

DIM0076430 DIM0076450

Appendix A
Deliverable Requirements for Level IV Data Packages

The following summary forms are required to meet the deliverable requirements for both the Full Data Package (Level IV CLP-like) and the Results Only Data Package. The CLP equivalent form number is listed in [ ] after the deliverable as an example of the information that is required. Use of the CLP form is not required. All deliverables must include a Case Narrative describing any analytical problems and observations. As part of the Case Narrative, the laboratory must provide a table cross referencing the TechLaw and Lab sample numbers. The deliverable must include copies of all correspondence (both verbal and written) with TechLaw personnel concerning the project. All deliverables must include Chain-of-Custody documentation. Full Data Packages will include all raw instrument data. Results Only Data Packages will include the summary forms only.

DELIVERABLE REQUIREMENTS FOR VOLATILE ORGANICS

CLP form number is listed in [ ] after the deliverable.

*1. Sample Results Summary [1A] TICs (if requested) [1E] (also used for method blank, MS and MSD results)
*2. System Monitoring Compound Recoveries [2A, 2B]
*3. Matrix Spike/Matrix Spike Duplicate Summary [3A, 3B]
*4. Method Blank Summary [4A]
5. Instrument Performance Check (Bromofluorobenzene, BFB tuning) [5A]
6. Initial Calibration Data [6A]
7. Continuing Calibration Data [7A]
8. Internal Standard Area and Retention Time Summary [8A]
9. Target compound identification
10. Quantitation limits.
11. System performance
12. Raw instrument data for all samples, method blanks, matrix spikes, matrix spike duplicates, instrument tunings and calibrations.

DELIVERABLE REQUIREMENTS FOR SEMI-VOLATILE ORGANICS

CLP form number is listed in [ ] after the deliverable.

*1. Sample Results Summary [1B, 1C] TICs (if requested) [1F] (also used for method blank, MS and MSD results)
*2. Surrogate Recoveries [2C, 2D]
*3. Matrix Spike/Matrix Spike Duplicate Summary [3C, 3D]
*4. Method Blank Summary [4B]
5. Instrument Performance Check (DFTPP tuning) [5B]
6. Initial Calibration Data [6B, 6C]
7. Continuing Calibration Data [7B, 7C]
8. Internal Standard Area and Retention Time Summary [8B, 8C]
#9. Raw instrument data for all samples, method blanks, matrix spikes, matrix spike duplicates, instrument tunings and calibrations

Appendix A
Deliverable Requirements for Level IV Data Packages

DELIVERABLE REQUIREMENTS FOR METALS AND CYANIDE

CLP form number is listed in [ ] after the deliverable.

*1. Analysis Data Sheet [1] for all samples
2. Initial and Continuing Calibration Verification [2A]
3. RDL standard for AA and ICP [2B]
*4. Blanks [3]
5. Interference Check Sample [4]
*6. Matrix Spike Sample Recovery [5A]
7. Post Digestion Spike Recovery [5B]
*8. Duplicates [6]
9. Laboratory Control Sample (if applicable) [7]
10. Standard Addition Results (if applicable) [8]
11. ICP Serial Dilution [9]
12. Instrument Detection Limits [10]
13. CP Interelement Correction Factors [11A, 11B]
14. ICP Linear Ranges [12]
15. Sample preparation log [13]
16. Analysis run log [14]
17. Raw instrument data for all samples, method blanks, matrix spikes, matrix spike duplicates, instrument calibrations and control samples.