Appendix C
Laboratory Analytical Reports

EXCEPTION SUMMARY FOR LABORATORY DATA

QUALITY ASSURANCE REVIEW

1. DATA SUMMARY

The laboratory data quality assurance review and data validation of 163 soil and nine water samples from the East Bay Habitat for Humanity, Inc. – 10800 Edes Avenue Brownfields site, Oakland, California has been completed.

Analyses were performed following protocols described in *Test Methods for Evaluating Solid Waste, Physical/ Chemical Methods* (SW846; USEPA 1996), *Method for the Determination of Asbestos in Bulk Building Materials* (USEPA 1993), and *Methods for the Chemical Analysis of Waters and Wastes* (MCAWW; USEPA 1983).

Samples were analyzed by Severn Trent Laboratories of Pleasanton, California, as laboratory project groups **720-381-1**, **720-381-2**, **720-401-1**, and **720-413-1** for the following constituents.

- Volatile organic compounds (VOC) in water were prepared following USEPA SW846 Method 5030B (purge-and-trap extraction) and analyzed following SW846 Method 8260B (gas chromatography – mass spectral detection [GC-MS]; low-level).

- Semivolatile organic compounds (SVOC) in soil were prepared following SW846 Method 3550B (ultrasonic extraction) and analyzed following SW846 Method 8270C (GC-MS/selective ion monitoring [GC-MS/SIM]).

- Semivolatile organic compounds (SVOC) in water were prepared following SW846 Method 3510C (separatory funnel liquid-liquid extraction) and analyzed following Method 8270C (GC-MS/SIM).

- Diesel-range organic compounds (*non-halogenated* DRO-extended to include motor oil) in soil were prepared following SW846 Method 3550B (ultrasonic extraction) and analyzed following SW846 Method 8015B (gas chromatography – flame ionization detection [GC-FID]).

- Diesel-range organic compounds (*non-halogenated* DRO-extended to include motor oil) in water were prepared following SW846 Method 3510C (separatory funnel liquid-liquid extraction) and analyzed following SW846 Method 8015B (GC–FID).

- Organochlorine pesticides (pesticides) in soil were prepared following SW846 Method 3550B (ultrasonic extraction) and analyzed following SW846 Method 8081A (gas chromatography – electron capture detection [GC-ECD]).

- Organochlorine pesticides (pesticides) in water were prepared following SW846 Method 3510C (separatory funnel liquid-liquid extraction) and analyzed following SW846 Method 8081A (GC-ECD).

- Total recoverable metals (arsenic, barium, cadmium, chromium, lead, selenium, silver) in soil were prepared following SW846 Method 3050B (acid digestion of sediments, sludges, and soils) and analyzed following SW846 Method 6010B (inductively coupled plasma – atomic emission spectroscopy [ICP-AES]).

- Total recoverable metals (arsenic, barium, cadmium, chromium, lead, selenium, silver) in water were prepared following SW846 Method 3010A (acid digestion of aqueous samples and extracts) and analyzed following SW846 Method 6010B (ICP-AES).

- Mercury in soil was prepared and analyzed following SW846 Method 7471A (cold-vapor atomic absorption spectroscopy [CVAAS]).

- Mercury in water was prepared and analyzed following SW846 Method 7470A (CVAAS).

- Percent moisture was determined following USEPA MCAWW Method 160.3 (total residue by gravimetry).

A subset of samples was analyzed by Forensic Analytical of Hayward, California, as laboratory project groups **2595-543** and **2595-544** for the following constituent.

- Asbestos in soil following USEPA Method 600/R-93/116 (visual area estimation by polarized light microscopy [PLM]).

Data were evaluated following the *Guidance for Evaluating Performance-Based Chemical Data* (USACE 2005), *National Functional Guidelines for Organic Data Review* (EPA 1999), *National Functional Guidelines for Inorganic Date Review* (EPA 2004), and specific criteria listed in the analytical methods.

This is an exception summary. All laboratory quality assurance results (holding time, blank sample analysis, laboratory control standard analysis, surrogate spike analysis) supplied to WESTON for the project analyses met acceptance criteria with the following exceptions.

LABORATORY GROUP 720-381-1

Metals

- Recovery of barium from the matrix spike duplicate (MSD) sample in laboratory soil preparation batch **720-1636** was less than the lower control limit. Recoveries of barium from the matrix spike (MS) sample and the laboratory control sample (LCS) met acceptance criteria. Barium results from all associated samples were qualified as estimated concentrations, unknown bias (J).

- Recoveries of mercury from the MS and MSD samples in laboratory soil preparation batch **720-1666** were each less than 10-percent. The parent sample concentration was comparable to the concentration of spike added. Recovery of mercury from the LCS met

acceptance criteria. Detected mercury results from all associated samples were qualified as estimated concentrations (J), possible low bias**.** Non-detected mercury results from all associated samples were rejected for use (R).

- Recovery of barium from the MSD sample in laboratory soil preparation batch **720-1734** was greater than the upper control limit. Recoveries of barium from the MS sample and the LCS met acceptance criteria. Detected barium results from all associated samples were qualified as estimated concentrations, unknown bias (J)**.** Non-detected barium results were not qualified.

- Recovery of lead from the MSD sample in laboratory soil preparation batch **720-1734** was less than the lower control limit. Recoveries of lead from the MS sample and the LCS met acceptance criteria. Lead results from all associated samples were qualified as estimated concentrations, unknown bias (J).

- Recovery of mercury from the MSD sample in laboratory soil preparation batch **720-1737** was greater than the upper control limit. Recoveries of mercury from the MS sample and the LCS met acceptance criteria. Detected mercury results from all associated samples were qualified as estimated concentrations, unknown bias (J)**.** Non-detected mercury results were not qualified.

- Recovery of lead from the MSD sample in laboratory soil preparation batch **720-1773** was less than the lower control limit. Recoveries of lead from the MS sample and the LCS met acceptance criteria. Lead results from all associated samples were qualified as estimated concentrations, unknown bias (J).

LABORATORY GROUP 720-413-1

- Recoveries of barium from the MS and MSD samples in laboratory soil preparation batch **720-1909** were less than the lower control limit. Recovery of barium from the LCS met acceptance criteria. Barium results from all associated samples were qualified as estimated concentrations (J), possible low bias.

- Recoveries of cadmium, lead, and selenium from the MSD sample in laboratory soil preparation batch **720-1909** were less than their respective lower control limits. Recoveries of these metals from the MS sample and the LCS met acceptance criteria. Cadmium, lead, and selenium results from all associated samples were qualified as estimated concentrations, unknown bias (J).

- Recovery of mercury from the MSD sample in laboratory soil preparation batch **720-17912** was greater than the upper control limit. The relative percent difference (RPD) between MS and MSD recoveries of mercury exceeded the acceptance criterion. Recoveries of mercury from the MS sample and the LCS met acceptance criteria. Detected mercury results from all associated samples were qualified as estimated concentrations (J), possible high bias**.** Non-detected mercury results were not qualified.

1.1 OTHER DATA QUALIFICATION

Apparent diesel-range detections could not be evaluated to determine their origin (i.e., anthropogenic versus naturally-occurring), since chromatograms and quantitation reports were not provided by the laboratory.

All analytical results below the laboratory quantitation limits (RL) have been reported by the laboratory as not-detected at the RL concentrations.

No other data were qualified for this review.

2. DATA QUALIFIERS

Any data qualifiers applied by the laboratory have been removed from the data summary sheets and superceded by data validation qualifiers as follow.

The following qualifiers were used to modify the data quality and usefulness of individual analytical results.

<	-	The analyte was not detected at the given quantitation (reporting) limit.
J	-	The analyte was positively identified and detected; however, the concentration is an estimated value because quality control criteria were not met.
R	-	Data are rejected due to significant exceedance(s) of quality control criteria. The analyte may or may not be present. The associated analyte result is unusable for any purpose.

3. DATA ASSESSMENT

Data review was performed by an experienced quality assurance chemist independent of the analytical laboratory and not directly involved in the project.

This is to certify that I have examined the analytical data and, based on the information provided to me by the laboratory, and in my professional judgment, the data as qualified are ACCEPTABLE for uses specified in the project QAPP.

Rejected analyte results may not be used for any purpose.

_____ January 4, 2006_____

R. Paul Swift, Ph.D. Date

Chief Chemist

4. REFERENCES

United States Army Corps of Engineers (USACE). 2005. *Guidance for Evaluating Performance-Based Chemical Data*, Engineer Manual EM200-1-10, Department of the Army, United States Army Corps of Engineers, Washington, D.C. June.

USEPA. 2004. *Contract Laboratory Program National Functional Guidelines for Inorganic Data Review*, EPA 540/R-04/004. Office of Emergency and Remedial Response, United States Environmental Protection Agency, Washington, D.C. October.

USEPA. 1999. *Contract Laboratory Program National Functional Guidelines for Organic Data Review*, EPA 540/R-99/08. Office of Emergency and Remedial Response, United States Environmental Protection Agency, Washington, D.C. October.

USEPA 1996. *Test Methods for Evaluating Solid Waste, Physical/ Chemical Methods SW-846* (through update III), Office of Solid Waste and Emergency Response, United States Environmental Protection Agency. December.

USEPA. 1993. *Method for the Determination of Asbestos in Bulk Building Materials*, EPA 600/R-93-116. July.

USEPA. 1983. *Methods for the Chemical Analysis of Waters and Wastes (including updates)*, Office of Research and Development, United States Environmental Protection Agency, Washington, D.C.

ANALYTICAL REPORT

Job Number: 720-381-1

Job Description: EBHH EDES AVE OAKLAND

For:

Weston Solutions,Inc
1575 Treat Blvd Suite 212
Walnut Creek, CA 94598

Attention: Mr. Tom Fortner

Surinder Sidhu

Surinder Sidhu
Project Manager I
ssidhu@stl-inc.com
11/30/2005

Case Narrative

Non Conformance Summary for job: 720-J381-1

Client: Weston Solutions,Inc
Date: 11/30/2005

Semi Volatiles GC Analysis

Effected Samples

720-381-39

MS/MSD recovery outside of control limits.

ms/msd required for sub #381.

Semi Volatiles MS Analysis

Effected Login

720-381

RL elevated due to matrix problems

The RL for samples (56, 100,) is elevated due to the following matrix problem:Sample would not concentrate to less than 5mls. Therefore RL raised by a factor of 5x

METHOD SUMMARY

Client: Weston Solutions,Inc

Job Number: 720-381-1

Description	Lab Location	Method	Preparation Method
Matrix: Solid			
Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)	STL-SF	SW846 8270C	
Ultrasonic Extraction	STL-SF		SW846 3550B
Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)	STL-SF	SW846 8015B	
Ultrasonic Extraction	STL-SF		SW846 3550B
Organochlorine Pesticides by Gas Chromatography	STL-SF	SW846 8081A	
Ultrasonic Extraction	STL-SF		SW846 3550B
Inductively Coupled Plasma - Atomic Emission Spectrometry	STL-SF	SW846 6010B	
Acid Digestion of Sediments, Sludges, and Soils	STL-SF		SW846 3050B
Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)	STL-SF	SW846 7471A	
Mercury in Solid or Semi-Solid Waste (Manual	STL-SF		SW846 7471A
Percent Moisture	STL-SF	EPA 160.3	
Asbestos	STL-SF	EPA	

LAB REFERENCES:

STL-SF = STL-San Francisco

METHOD REFERENCES:

SW846 - "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

EPA - US Environmental Protection Agency

SAMPLE SUMMARY

Client: Weston Solutions,Inc

Job Number: 720-381-1

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
720-381-1	EBHH-1A	Solid	11/09/2005 0900	11/09/2005 1730
720-381-2	EBHH-1B	Solid	11/09/2005 0925	11/09/2005 1730
720-381-3	EBHH-1C	Solid	11/09/2005 0940	11/09/2005 1730
720-381-5	EBHH-2B	Solid	11/09/2005 0945	11/09/2005 1730
720-381-6	EBHH-2C	Solid	11/09/2005 0951	11/09/2005 1730
720-381-8	EBHH-2A	Solid	11/09/2005 1102	11/09/2005 1730
720-381-9	EBHH-3B	Solid	11/09/2005 1000	11/09/2005 1730
720-381-10	EBHH-3C	Solid	11/09/2005 1000	11/09/2005 1730
720-381-12	EBHH-3A	Solid	11/09/2005 1105	11/09/2005 1730
720-381-13	EBHH-4B	Solid	11/09/2005 1015	11/09/2005 1730
720-381-14	EBHH-4C	Solid	11/09/2005 1020	11/09/2005 1730
720-381-16	EBHH-4A	Solid	11/09/2005 1110	11/09/2005 1730
720-381-17	EBHH-5B	Solid	11/09/2005 1030	11/09/2005 1730
720-381-18	EBHH-5C	Solid	11/09/2005 1036	11/09/2005 1730
720-381-20	EBHH-5A	Solid	11/09/2005 1115	11/09/2005 1730
720-381-21	EBHH-6B	Solid	11/09/2005 1045	11/09/2005 1730
720-381-22	EBHH-6C	Solid	11/09/2005 1045	11/09/2005 1730
720-381-24	EBHH-6A	Solid	11/09/2005 1118	11/09/2005 1730
720-381-25	EBHH-7B	Solid	11/09/2005 1125	11/09/2005 1730
720-381-26	EBHH-7C	Solid	11/09/2005 1125	11/09/2005 1730
720-381-28	EBHH-7A	Solid	11/09/2005 1145	11/09/2005 1730
720-381-29	EBHH-8B	Solid	11/09/2005 1135	11/09/2005 1730
720-381-30	EBHH-8C	Solid	11/09/2005 1135	11/09/2005 1730
720-381-32	EBHH-8A	Solid	11/09/2005 1150	11/09/2005 1730
720-381-33	EBHH-9B	Solid	11/09/2005 1146	11/09/2005 1730
720-381-34	EBHH-9C	Solid	11/09/2005 1146	11/09/2005 1730
720-381-36	EBHH-9A	Solid	11/09/2005 1155	11/09/2005 1730
720-381-37	EBHH-10B	Solid	11/09/2005 1203	11/09/2005 1730
720-381-38	EBHH-10C	Solid	11/09/2005 1203	11/09/2005 1730
720-381-39	EBHH-10D	Solid	11/09/2005 1203	11/09/2005 1730
720-381-40	EBHH-10A	Solid	11/09/2005 1340	11/09/2005 1730
720-381-41	EBHH-11B	Solid	11/09/2005 1208	11/09/2005 1730
720-381-42	EBHH-11C	Solid	11/09/2005 1215	11/09/2005 1730
720-381-44	EBHH-11A	Solid	11/09/2005 1243	11/09/2005 1730
720-381-45	EBHH-12B	Solid	11/09/2005 1223	11/09/2005 1730
720-381-46	EBHH-12C	Solid	11/09/2005 1223	11/09/2005 1730
720-381-48	EBHH-12A	Solid	11/09/2005 1248	11/09/2005 1730
720-381-49	EBHH-13B	Solid	11/09/2005 1250	11/09/2005 1730
720-381-50	EBHH-13C	Solid	11/09/2005 1252	11/09/2005 1730
720-381-52	EBHH-13A	Solid	11/09/2005 1350	11/09/2005 1730
720-381-53	EBHH-14B	Solid	11/09/2005 1301	11/09/2005 1730
720-381-54	EBHH-14C	Solid	11/09/2005 1309	11/09/2005 1730
720-381-56	EBHH-14A	Solid	11/09/2005 1445	11/09/2005 1730
720-381-57	EBHH-15B	Solid	11/09/2005 1341	11/09/2005 1730
720-381-58	EBHH-15C	Solid	11/09/2005 1346	11/09/2005 1730
720-381-60	EBHH-15A	Solid	11/09/2005 1450	11/09/2005 1730

SAMPLE SUMMARY

Client: Weston Solutions,Inc

Job Number: 720-381-1

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
720-381-61	EBHH-16B	Solid	11/09/2005 1358	11/09/2005 1730
720-381-62	EBHH-16C	Solid	11/09/2005 1402	11/09/2005 1730
720-381-64	EBHH-16A	Solid	11/09/2005 1455	11/09/2005 1730
720-381-65	EBHH-17B	Solid	11/09/2005 1408	11/09/2005 1730
720-381-66	EBHH-17C	Solid	11/09/2005 1411	11/09/2005 1730
720-381-68	EBHH-17A	Solid	11/09/2005 1500	11/09/2005 1730
720-381-69	EBHH-18B	Solid	11/09/2005 1422	11/09/2005 1730
720-381-70	EBHH-18C	Solid	11/09/2005 1422	11/09/2005 1730
720-381-71	EBHH-18D	Solid	11/09/2005 1422	11/09/2005 1730
720-381-72	EBHH-18A	Solid	11/09/2005 1505	11/09/2005 1730
720-381-73	EBHH-19B	Solid	11/09/2005 1430	11/09/2005 1730
720-381-74	EBHH-19C	Solid	11/09/2005 1430	11/09/2005 1730
720-381-75	EBHH-19D	Solid	11/09/2005 1430	11/09/2005 1730
720-381-76	EBHH-19A	Solid	11/09/2005 1530	11/09/2005 1730
720-381-77	EBHH-20B	Solid	11/09/2005 1438	11/09/2005 1730
720-381-78	EBHH-20C	Solid	11/09/2005 1438	11/09/2005 1730
720-381-80	EBHH-20A	Solid	11/09/2005 1535	11/09/2005 1730
720-381-81	EBHH-21B	Solid	11/09/2005 1450	11/09/2005 1730
720-381-82	EBHH-21C	Solid	11/09/2005 1450	11/09/2005 1730
720-381-84	EBHH-21A	Solid	11/09/2005 1540	11/09/2005 1730
720-381-85	EBHH-22B	Solid	11/09/2005 1510	11/09/2005 1730
720-381-86	EBHH-22C	Solid	11/09/2005 1510	11/09/2005 1730
720-381-88	EBHH-22A	Solid	11/09/2005 1545	11/09/2005 1730
720-381-89	EBHH-23B	Solid	11/09/2005 1515	11/09/2005 1730
720-381-90	EBHH-23C	Solid	11/09/2005 1520	11/09/2005 1730
720-381-92	EBHH-23A	Solid	11/09/2005 1550	11/09/2005 1730
720-381-93	EBHH-24B	Solid	11/09/2005 1530	11/09/2005 1730
720-381-94	EBHH-24C	Solid	11/09/2005 1530	11/09/2005 1730
720-381-96	EBHH-24A	Solid	11/09/2005 1615	11/09/2005 1730
720-381-97	EBHH-26B	Solid	11/09/2005 1540	11/09/2005 1730
720-381-98	EBHH-26C	Solid	11/09/2005 1548	11/09/2005 1730
720-381-100	EBHH-26A	Solid	11/09/2005 1630	11/09/2005 1730
720-381-101	EBHH-27B	Solid	11/09/2005 1600	11/09/2005 1730
720-381-102	EBHH-27C	Solid	11/09/2005 1605	11/09/2005 1730
720-381-104	EBHH-27A	Solid	11/09/2005 1650	11/09/2005 1730
720-381-105	EBHH-25B	Solid	11/09/2005 1620	11/09/2005 1730
720-381-106	EBHH-51A	Solid	11/09/2005 1240	11/09/2005 1730
720-381-107	EBHH-51B	Solid	11/09/2005 1501	11/09/2005 1730
720-381-108	EBHH-51C	Solid	11/09/2005 1501	11/09/2005 1730
720-381-109	EBHH-52A	Solid	11/09/2005 1600	11/09/2005 1730
720-381-110	EBHH-52B	Solid	11/09/2005 1600	11/09/2005 1730
720-381-111	EBHH-52C	Solid	11/09/2005 1600	11/09/2005 1730
720-381-112	EBHH-53A	Solid	11/09/2005 1700	11/09/2005 1730
720-381-113	EBHH-53B	Solid	11/09/2005 1515	11/09/2005 1730
720-381-114	EBHH-54A	Solid	11/09/2005 1630	11/09/2005 1730
720-381-115	EBHH-54B	Solid	11/09/2005 1600	11/09/2005 1730

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-1A**

Lab Sample ID:	720-381-1			Date Sampled:	11/09/2005 0900
Client Matrix:	Solid	% Moisture:	9.5	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-1821		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1667		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.21 g
Date Analyzed:	11/11/2005 1846			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1215			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		55
Acenaphthene		ND		55
Acenaphthylene		ND		55
Fluorene		ND		55
Phenanthrene		ND		55
Anthracene		ND		55
Benzo[a]anthracene		ND		55
Chrysene		ND		55
Benzo[a]pyrene		ND		55
Benzo[b]fluoranthene		ND		55
Benzo[k]fluoranthene		ND		55
Benzo[g,h,i]perylene		ND		55
Indeno[1,2,3-cd]pyrene		ND		55
Fluoranthene		ND		55
Pyrene		ND		55
Dibenz(a,h)anthracene		ND		55

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl		
Terphenyl-d14		

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-1B**

Lab Sample ID:	720-381-2			Date Sampled:	11/09/2005 0925
Client Matrix:	Solid	% Moisture:	7.8	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-1821		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1667		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.30 g
Date Analyzed:	11/11/2005 1913			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1215			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		54
Acenaphthene		ND		54
Acenaphthylene		ND		54
Fluorene		ND		54
Phenanthrene		ND		54
Anthracene		ND		54
Benzo[a]anthracene		ND		54
Chrysene		54		54
Benzo[a]pyrene		ND		54
Benzo[b]fluoranthene		ND		54
Benzo[k]fluoranthene		ND		54
Benzo[g,h,i]perylene		ND		54
Indeno[1,2,3-cd]pyrene		ND		54
Fluoranthene		ND		54
Pyrene		67		54
Dibenz(a,h)anthracene		ND		54

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl		
Terphenyl-d14		

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-4C**

Lab Sample ID:	720-381-14			Date Sampled:	11/09/2005 1020
Client Matrix:	Solid	% Moisture:	12.7	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-1821		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1667		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.01 g
Date Analyzed:	11/11/2005 1941			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1215			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.7
Acenaphthene		ND		5.7
Acenaphthylene		ND		5.7
Fluorene		ND		5.7
Phenanthrene		ND		5.7
Anthracene		ND		5.7
Benzo[a]anthracene		7.4		5.7
Chrysene		ND		5.7
Benzo[a]pyrene		ND		5.7
Benzo[b]fluoranthene		9.2		5.7
Benzo[k]fluoranthene		ND		5.7
Benzo[g,h,i]perylene		6.0		5.7
Indeno[1,2,3-cd]pyrene		ND		5.7
Fluoranthene		6.1		5.7
Pyrene		8.0		5.7
Dibenz(a,h)anthracene		ND		5.7

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	67	30 - 115
Terphenyl-d14	76	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-4A**

Lab Sample ID:	720-381-16			Date Sampled:	11/09/2005 1110
Client Matrix:	Solid	% Moisture:	9.5	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-1821		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1667		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.15 g
Date Analyzed:	11/11/2005 2009			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1215			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		55
Acenaphthene		ND		55
Acenaphthylene		ND		55
Fluorene		ND		55
Phenanthrene		ND		55
Anthracene		ND		55
Benzo[a]anthracene		71		55
Chrysene		67		55
Benzo[a]pyrene		72		55
Benzo[b]fluoranthene		73		55
Benzo[k]fluoranthene		ND		55
Benzo[g,h,i]perylene		ND		55
Indeno[1,2,3-cd]pyrene		ND		55
Fluoranthene		91		55
Pyrene		87		55
Dibenz(a,h)anthracene		ND		55

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	51	30 - 115
Terphenyl-d14	78	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-10D**

Lab Sample ID:	720-381-39			Date Sampled:	11/09/2005 1203
Client Matrix:	Solid	% Moisture:	11.7	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-1821		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1667		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.02 g
Date Analyzed:	11/11/2005 2037			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1215			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.7
Acenaphthene		ND		5.7
Acenaphthylene		ND		5.7
Fluorene		ND		5.7
Phenanthrene		ND		5.7
Anthracene		ND		5.7
Benzo[a]anthracene		ND		5.7
Chrysene		ND		5.7
Benzo[a]pyrene		ND		5.7
Benzo[b]fluoranthene		ND		5.7
Benzo[k]fluoranthene		ND		5.7
Benzo[g,h,i]perylene		ND		5.7
Indeno[1,2,3-cd]pyrene		ND		5.7
Fluoranthene		ND		5.7
Pyrene		ND		5.7
Dibenz(a,h)anthracene		ND		5.7

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	67	30 - 115
Terphenyl-d14	84	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-14B**

Lab Sample ID:	720-381-53			Date Sampled:	11/09/2005 1301
Client Matrix:	Solid	% Moisture:	7.5	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-1821		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1667		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.29 g
Date Analyzed:	11/11/2005 2105			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1215			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.4
Acenaphthene		ND		5.4
Acenaphthylene		7.0		5.4
Fluorene		ND		5.4
Phenanthrene		41		5.4
Anthracene		ND		5.4
Benzo[a]anthracene		16		5.4
Chrysene		41		5.4
Benzo[a]pyrene		27		5.4
Benzo[b]fluoranthene		52		5.4
Benzo[k]fluoranthene		21		5.4
Benzo[g,h,i]perylene		31		5.4
Indeno[1,2,3-cd]pyrene		29		5.4
Fluoranthene		57		5.4
Pyrene		59		5.4
Dibenz(a,h)anthracene		5.4		5.4

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	85	30 - 115
Terphenyl-d14	87	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-14A**

Lab Sample ID: 720-381-56
Client Matrix: Solid

Date Sampled: 11/09/2005 1445
Date Received: 11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/11/2005 2132
Date Prepared: 11/11/2005 1215

Analysis Batch: 720-1821
Prep Batch: 720-1667

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.19 g
Final Weight/Volume: 5 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		25
Acenaphthene		ND		25
Acenaphthylene		31		25
Fluorene		ND		25
Phenanthrene		110		25
Anthracene		35		25
Benzo[a]anthracene		150		25
Chrysene		240		25
Benzo[a]pyrene		270		25
Benzo[b]fluoranthene		380		25
Benzo[k]fluoranthene		140		25
Benzo[g,h,i]perylene		280		25
Indeno[1,2,3-cd]pyrene		250		25
Fluoranthene		220		25
Pyrene		260		25
Dibenz(a,h)anthracene		67		25

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	68	30 - 115
Terphenyl-d14	101	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-18D**

Lab Sample ID:	720-381-71			Date Sampled:	11/09/2005 1422
Client Matrix:	Solid	% Moisture:	12.3	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2227		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1830		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.03 g
Date Analyzed:	11/23/2005 2227			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.7
Acenaphthene		ND		5.7
Acenaphthylene		ND		5.7
Fluorene		ND		5.7
Phenanthrene		ND		5.7
Anthracene		ND		5.7
Benzo[a]anthracene		ND		5.7
Chrysene		ND		5.7
Benzo[a]pyrene		ND		5.7
Benzo[b]fluoranthene		ND		5.7
Benzo[k]fluoranthene		ND		5.7
Benzo[g,h,i]perylene		ND		5.7
Indeno[1,2,3-cd]pyrene		ND		5.7
Fluoranthene		ND		5.7
Pyrene		ND		5.7
Dibenz(a,h)anthracene		ND		5.7

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	74	30 - 115
Terphenyl-d14	88	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-22C**

Lab Sample ID: 720-381-86
Client Matrix: Solid

Date Sampled: 11/09/2005 1510
Date Received: 11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/16/2005 1059
Date Prepared: 11/14/2005 1006

Analysis Batch: 720-1851
Prep Batch: 720-1711

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.04 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.0
Acenaphthene		ND		5.0
Acenaphthylene		ND		5.0
Fluorene		ND		5.0
Phenanthrene		ND		5.0
Anthracene		ND		5.0
Benzo[a]anthracene		ND		5.0
Chrysene		ND		5.0
Benzo[a]pyrene		ND		5.0
Benzo[b]fluoranthene		ND		5.0
Benzo[k]fluoranthene		ND		5.0
Benzo[g,h,i]perylene		ND		5.0
Indeno[1,2,3-cd]pyrene		ND		5.0
Fluoranthene		ND		5.0
Pyrene		ND		5.0
Dibenz(a,h)anthracene		ND		5.0

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	59	30 - 115
Terphenyl-d14	66	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-22A**

Lab Sample ID:	720-381-88			Date Sampled:	11/09/2005 1545
Client Matrix:	Solid	% Moisture:	11.5	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-1821		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1667		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.18 g
Date Analyzed:	11/11/2005 2256			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1215			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		56
Acenaphthene		ND		56
Acenaphthylene		ND		56
Fluorene		ND		56
Phenanthrene		ND		56
Anthracene		ND		56
Benzo[a]anthracene		ND		56
Chrysene		ND		56
Benzo[a]pyrene		ND		56
Benzo[b]fluoranthene		74		56
Benzo[k]fluoranthene		ND		56
Benzo[g,h,i]perylene		ND		56
Indeno[1,2,3-cd]pyrene		ND		56
Fluoranthene		ND		56
Pyrene		66		56
Dibenz(a,h)anthracene		ND		56

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	57	30 - 115
Terphenyl-d14	81	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-26C**

Lab Sample ID:	720-381-98			Date Sampled:	11/09/2005 1548
Client Matrix:	Solid	% Moisture:	12.8	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-1821		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1667		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.01 g
Date Analyzed:	11/11/2005 2324			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1215			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.7
Acenaphthene		ND		5.7
Acenaphthylene		ND		5.7
Fluorene		ND		5.7
Phenanthrene		ND		5.7
Anthracene		ND		5.7
Benzo[a]anthracene		ND		5.7
Chrysene		ND		5.7
Benzo[a]pyrene		ND		5.7
Benzo[b]fluoranthene		ND		5.7
Benzo[k]fluoranthene		ND		5.7
Benzo[g,h,i]perylene		ND		5.7
Indeno[1,2,3-cd]pyrene		ND		5.7
Fluoranthene		ND		5.7
Pyrene		ND		5.7
Dibenz(a,h)anthracene		ND		5.7

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	66	30 - 115
Terphenyl-d14	82	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-26A**

Lab Sample ID:	720-381-100			Date Sampled:	11/09/2005 1630
Client Matrix:	Solid	% Moisture:	7.7	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-1821		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1667		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.08 g
Date Analyzed:	11/11/2005 2351			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1215			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		270
Acenaphthene		ND		270
Acenaphthylene		ND		270
Fluorene		ND		270
Phenanthrene		ND		270
Anthracene		ND		270
Benzo[a]anthracene		ND		270
Chrysene		ND		270
Benzo[a]pyrene		ND		270
Benzo[b]fluoranthene		ND		270
Benzo[k]fluoranthene		ND		270
Benzo[g,h,i]perylene		ND		270
Indeno[1,2,3-cd]pyrene		ND		270
Fluoranthene		ND		270
Pyrene		ND		270
Dibenz(a,h)anthracene		ND		270

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	62	30 - 115
Terphenyl-d14	90	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-25B**

Lab Sample ID:	720-381-105			Date Sampled:	11/09/2005 1620
Client Matrix:	Solid	% Moisture:	10.3	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-1821		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1667		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.15 g
Date Analyzed:	11/12/2005 0019			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1215			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.5
Acenaphthene		ND		5.5
Acenaphthylene		24		5.5
Fluorene		ND		5.5
Phenanthrene		43		5.5
Anthracene		16		5.5
Benzo[a]anthracene		65		5.5
Chrysene		94		5.5
Benzo[a]pyrene		76		5.5
Benzo[b]fluoranthene		120		5.5
Benzo[k]fluoranthene		40		5.5
Benzo[g,h,i]perylene		54		5.5
Indeno[1,2,3-cd]pyrene		47		5.5
Fluoranthene		150		5.5
Pyrene		130		5.5
Dibenz(a,h)anthracene		14		5.5

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	66	30 - 115
Terphenyl-d14	72	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-1A**

Lab Sample ID:	720-381-1			Date Sampled:	11/09/2005 0900
Client Matrix:	Solid	% Moisture:	9.5	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1682		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1624		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.14 g
Date Analyzed:	11/15/2005 2328			Final Weight/Volume:	5 mL
Date Prepared:	11/10/2005 1200			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		16		1.1
Motor Oil Range Organics [C24-C36]		110		55

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	81	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-1B**

Lab Sample ID:	720-381-2			Date Sampled:	11/09/2005 0925
Client Matrix:	Solid	% Moisture:	7.8	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1682		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1624		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.01 g
Date Analyzed:	11/12/2005 0116			Final Weight/Volume:	5 mL
Date Prepared:	11/10/2005 1200			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		27		1.1
Motor Oil Range Organics [C24-C36]		150		54

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	85	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-4A**

Lab Sample ID:	720-381-16			Date Sampled:	11/09/2005 1110
Client Matrix:	Solid	% Moisture:	9.5	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1682		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1624		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.32 g
Date Analyzed:	11/15/2005 2328			Final Weight/Volume:	5 mL
Date Prepared:	11/10/2005 1200			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		91		1.1
Motor Oil Range Organics [C24-C36]		420		55

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	67	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-10D**

Lab Sample ID:	720-381-39			Date Sampled:	11/09/2005 1203
Client Matrix:	Solid	% Moisture:	11.7	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1806		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1758		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.02 g
Date Analyzed:	11/15/2005 1736			Final Weight/Volume:	5 mL
Date Prepared:	11/15/2005 1005			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		ND		1.1
Motor Oil Range Organics [C24-C36]		ND		57

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	72	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-14B**

Lab Sample ID:	720-381-53			Date Sampled:	11/09/2005 1301
Client Matrix:	Solid	% Moisture:	7.5	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1682		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1624		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.23 g
Date Analyzed:	11/11/2005 1952			Final Weight/Volume:	5 mL
Date Prepared:	11/10/2005 1200			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		18		1.1
Motor Oil Range Organics [C24-C36]		84		54

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	80	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-14A**

Lab Sample ID:	720-381-56			Date Sampled:	11/09/2005 1445
Client Matrix:	Solid	% Moisture:	8.8	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1682		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1624		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.39 g
Date Analyzed:	11/11/2005 2234			Final Weight/Volume:	5 mL
Date Prepared:	11/10/2005 1200			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		20		0.99
Motor Oil Range Organics [C24-C36]		110		49

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	80	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-18D**

Lab Sample ID:	720-381-71			Date Sampled:	11/09/2005 1422
Client Matrix:	Solid	% Moisture:	12.3	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1987		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1829		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.17 g
Date Analyzed:	11/16/2005 1356			Final Weight/Volume:	5 mL
Date Prepared:	11/16/2005 1152			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		ND		1.1
Motor Oil Range Organics [C24-C36]		ND		57

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	80	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-22C**

Lab Sample ID:	720-381-86			Date Sampled:	11/09/2005 1510
Client Matrix:	Solid	% Moisture:	7.4	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1682		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1624		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.47 g
Date Analyzed:	11/11/2005 1858			Final Weight/Volume:	5 mL
Date Prepared:	11/10/2005 1200			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		ND		1.1
Motor Oil Range Organics [C24-C36]		ND		53

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	73	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-22A**

Lab Sample ID:	720-381-88			Date Sampled:	11/09/2005 1545
Client Matrix:	Solid	% Moisture:	11.5	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1682		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1624		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.42 g
Date Analyzed:	11/12/2005 0022			Final Weight/Volume:	5 mL
Date Prepared:	11/10/2005 1200			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		35		1.1
Motor Oil Range Organics [C24-C36]		200		56

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	72	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-26C**

Lab Sample ID:	720-381-98			Date Sampled:	11/09/2005 1548
Client Matrix:	Solid	% Moisture:	12.8	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1682		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1624		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.36 g
Date Analyzed:	11/11/2005 1925			Final Weight/Volume:	5 mL
Date Prepared:	11/10/2005 1200			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		1.3		1.1
Motor Oil Range Organics [C24-C36]		ND		57

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	68	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-26A**

Lab Sample ID:	720-381-100			Date Sampled:	11/09/2005 1630
Client Matrix:	Solid	% Moisture:	7.7	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1682		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1624		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.29 g
Date Analyzed:	11/12/2005 0022			Final Weight/Volume:	5 mL
Date Prepared:	11/10/2005 1200			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		25		1.1
Motor Oil Range Organics [C24-C36]		170		54

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	69	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-25B**

Lab Sample ID:	720-381-105			Date Sampled:	11/09/2005 1620
Client Matrix:	Solid	% Moisture:	10.3	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1682		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1624		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.02 g
Date Analyzed:	11/11/2005 2019			Final Weight/Volume:	5 mL
Date Prepared:	11/10/2005 1200			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		2.8		1.1
Motor Oil Range Organics [C24-C36]		ND		56

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	75	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-1B**

Lab Sample ID:	720-381-2			Date Sampled:	11/09/2005 0925
Client Matrix:	Solid	% Moisture:	7.8	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.30 g
Date Analyzed:	11/17/2005 1301			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Chlordane (technical)		220		54

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-1B**

Lab Sample ID:	720-381-2			Date Sampled:	11/09/2005 0925
Client Matrix:	Solid	% Moisture:	7.8	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.30 g
Date Analyzed:	11/15/2005 0226			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		11
Dieldrin		ND		11
Endrin aldehyde		ND		11
Endrin		ND		11
Endrin ketone		ND		11
Heptachlor		ND		11
Heptachlor epoxide		ND		11
4,4'-DDT		110		11
4,4'-DDE		330		11
4,4'-DDD		ND		11
Endosulfan I		ND		11
Endosulfan II		ND		11
alpha-BHC		ND		11
beta-BHC		ND		11
gamma-BHC (Lindane)		ND		11
delta-BHC		ND		11
Endosulfan sulfate		ND		11
Methoxychlor		ND		11
Toxaphene		ND		540
alpha-Chlordane		110		11
gamma-Chlordane		85		11

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	94	50 - 125
DCB Decachlorobiphenyl	92	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-1C**

Lab Sample ID:	720-381-3			Date Sampled:	11/09/2005 0940
Client Matrix:	Solid	% Moisture:	11.8	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.15 g
Date Analyzed:	11/15/2005 0255			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		11
Dieldrin		ND		11
Endrin aldehyde		ND		11
Endrin		ND		11
Endrin ketone		ND		11
Heptachlor		ND		11
Heptachlor epoxide		ND		11
4,4'-DDT		ND		11
4,4'-DDE		ND		11
4,4'-DDD		ND		11
Endosulfan I		ND		11
Endosulfan II		ND		11
alpha-BHC		ND		11
beta-BHC		ND		11
gamma-BHC (Lindane)		ND		11
delta-BHC		ND		11
Endosulfan sulfate		ND		11
Methoxychlor		ND		11
Toxaphene		ND		560
Chlordane (technical)		ND		280
alpha-Chlordane		ND		11
gamma-Chlordane		ND		11

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	105	50 - 125
DCB Decachlorobiphenyl	98	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-3B**

Lab Sample ID:	720-381-9			Date Sampled:	11/09/2005 1000
Client Matrix:	Solid	% Moisture:	8.3	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	10			Initial Weight/Volume:	30.21 g
Date Analyzed:	11/15/2005 0353			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		22
Dieldrin		ND		22
Endrin aldehyde		ND		22
Endrin		ND		22
Endrin ketone		ND		22
Heptachlor		ND		22
Heptachlor epoxide		ND		22
4,4'-DDT		ND		22
4,4'-DDE		ND		22
4,4'-DDD		ND		22
Endosulfan I		ND		22
Endosulfan II		ND		22
alpha-BHC		ND		22
beta-BHC		ND		22
gamma-BHC (Lindane)		ND		22
delta-BHC		ND		22
Endosulfan sulfate		ND		22
Methoxychlor		ND		22
Toxaphene		ND		1100
Chlordane (technical)		ND		540
alpha-Chlordane		ND		22
gamma-Chlordane		ND		22

Surrogate	%Rec		Acceptance Limits
Tetrachloro-m-xylene	0	D	50 - 125
DCB Decachlorobiphenyl	0	D	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-3C**

Lab Sample ID:	720-381-10			Date Sampled:	11/09/2005 1000
Client Matrix:	Solid	% Moisture:	15.8	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.14 g
Date Analyzed:	11/11/2005 1114			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.4
Dieldrin		ND		2.4
Endrin aldehyde		ND		2.4
Endrin		ND		2.4
Endrin ketone		ND		2.4
Heptachlor		ND		2.4
Heptachlor epoxide		ND		2.4
4,4'-DDT		11		2.4
4,4'-DDE		14		2.4
4,4'-DDD		ND		2.4
Endosulfan I		ND		2.4
Endosulfan II		ND		2.4
alpha-BHC		ND		2.4
beta-BHC		ND		2.4
gamma-BHC (Lindane)		ND		2.4
delta-BHC		ND		2.4
Endosulfan sulfate		ND		2.4
Methoxychlor		ND		2.4
Toxaphene		ND		120
Chlordane (technical)		ND		59
alpha-Chlordane		ND		2.4
gamma-Chlordane		ND		2.4

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	79	50 - 125
DCB Decachlorobiphenyl	92	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-4C**

Lab Sample ID:	720-381-14			Date Sampled:	11/09/2005 1020
Client Matrix:	Solid	% Moisture:	12.7	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.11 g
Date Analyzed:	11/11/2005 1143			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.3
Dieldrin		ND		2.3
Endrin aldehyde		ND		2.3
Endrin		ND		2.3
Endrin ketone		ND		2.3
Heptachlor		ND		2.3
Heptachlor epoxide		ND		2.3
4,4'-DDT		3.0		2.3
4,4'-DDE		9.8		2.3
4,4'-DDD		ND		2.3
Endosulfan I		ND		2.3
Endosulfan II		ND		2.3
alpha-BHC		ND		2.3
beta-BHC		ND		2.3
gamma-BHC (Lindane)		ND		2.3
delta-BHC		ND		2.3
Endosulfan sulfate		ND		2.3
Methoxychlor		ND		2.3
Toxaphene		ND		110
Chlordane (technical)		ND		57
alpha-Chlordane		ND		2.3
gamma-Chlordane		ND		2.3

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	85	50 - 125
DCB Decachlorobiphenyl	94	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-5B**

Lab Sample ID:	720-381-17			Date Sampled:	11/09/2005 1030
Client Matrix:	Solid	% Moisture:	12.2	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.20 g
Date Analyzed:	11/15/2005 0324			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		11
Dieldrin		ND		11
Endrin aldehyde		ND		11
Endrin		ND		11
Endrin ketone		ND		11
Heptachlor		ND		11
Heptachlor epoxide		ND		11
4,4'-DDT		ND		11
4,4'-DDE		16		11
4,4'-DDD		ND		11
Endosulfan I		ND		11
Endosulfan II		ND		11
alpha-BHC		ND		11
beta-BHC		ND		11
gamma-BHC (Lindane)		ND		11
delta-BHC		ND		11
Endosulfan sulfate		ND		11
Methoxychlor		ND		11
Toxaphene		ND		570
Chlordane (technical)		ND		280
alpha-Chlordane		ND		11
gamma-Chlordane		ND		11

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	105	50 - 125
DCB Decachlorobiphenyl	102	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-5C**

Lab Sample ID:	720-381-18			Date Sampled:	11/09/2005 1036
Client Matrix:	Solid	% Moisture:	15.5	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.17 g
Date Analyzed:	11/11/2005 1212			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.4
Dieldrin		ND		2.4
Endrin aldehyde		ND		2.4
Endrin		ND		2.4
Endrin ketone		ND		2.4
Heptachlor		ND		2.4
Heptachlor epoxide		ND		2.4
4,4'-DDT		ND		2.4
4,4'-DDE		ND		2.4
4,4'-DDD		ND		2.4
Endosulfan I		ND		2.4
Endosulfan II		ND		2.4
alpha-BHC		ND		2.4
beta-BHC		ND		2.4
gamma-BHC (Lindane)		ND		2.4
delta-BHC		ND		2.4
Endosulfan sulfate		ND		2.4
Methoxychlor		ND		2.4
Toxaphene		ND		120
Chlordane (technical)		ND		59
alpha-Chlordane		ND		2.4
gamma-Chlordane		ND		2.4

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	79	50 - 125
DCB Decachlorobiphenyl	90	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-7B**

Lab Sample ID: 720-381-25
Client Matrix: Solid
% Moisture: 15.3
Date Sampled: 11/09/2005 1125
Date Received: 11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method: 8081A
Preparation: 3550B
Dilution: 5.0
Date Analyzed: 11/15/2005 0422
Date Prepared: 11/10/2005 1454

Analysis Batch: 720-1801
Prep Batch: 720-1635

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.20 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		12
Dieldrin		ND		12
Endrin aldehyde		ND		12
Endrin		ND		12
Endrin ketone		ND		12
Heptachlor		ND		12
Heptachlor epoxide		ND		12
4,4'-DDT		93		12
4,4'-DDE		190		12
4,4'-DDD		35		12
Endosulfan I		ND		12
Endosulfan II		ND		12
alpha-BHC		ND		12
beta-BHC		ND		12
gamma-BHC (Lindane)		ND		12
delta-BHC		ND		12
Endosulfan sulfate		ND		12
Methoxychlor		ND		12
Toxaphene		ND		590
alpha-Chlordane		40		12
gamma-Chlordane		26		12

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	107	50 - 125
DCB Decachlorobiphenyl	87	46 - 142

Method: 8081A
Preparation: 3550B
Dilution: 5.0
Date Analyzed: 11/17/2005 1324
Date Prepared: 11/10/2005 1454

Analysis Batch: 720-1801
Prep Batch: 720-1635

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.20 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Chlordane (technical)		770		290

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-7C**

Lab Sample ID:	720-381-26			Date Sampled:	11/09/2005 1125
Client Matrix:	Solid	% Moisture:	13.5	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.09 g
Date Analyzed:	11/15/2005 0450			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.3
Dieldrin		ND		2.3
Endrin aldehyde		ND		2.3
Endrin		ND		2.3
Endrin ketone		ND		2.3
Heptachlor		ND		2.3
Heptachlor epoxide		ND		2.3
4,4'-DDT		ND		2.3
4,4'-DDE		ND		2.3
4,4'-DDD		ND		2.3
Endosulfan I		ND		2.3
Endosulfan II		ND		2.3
alpha-BHC		ND		2.3
beta-BHC		ND		2.3
gamma-BHC (Lindane)		ND		2.3
delta-BHC		ND		2.3
Endosulfan sulfate		ND		2.3
Methoxychlor		ND		2.3
Toxaphene		ND		120
Chlordane (technical)		ND		58
alpha-Chlordane		ND		2.3
gamma-Chlordane		ND		2.3

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	91	50 - 125
DCB Decachlorobiphenyl	93	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-9B**

Lab Sample ID:	720-381-33			Date Sampled:	11/09/2005 1146
Client Matrix:	Solid	% Moisture:	12.1	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.05 g
Date Analyzed:	11/15/2005 0519			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		11
Dieldrin		ND		11
Endrin aldehyde		ND		11
Endrin		ND		11
Endrin ketone		ND		11
Heptachlor		ND		11
Heptachlor epoxide		ND		11
4,4'-DDT		ND		11
4,4'-DDE		30		11
4,4'-DDD		ND		11
Endosulfan I		ND		11
Endosulfan II		ND		11
alpha-BHC		ND		11
beta-BHC		ND		11
gamma-BHC (Lindane)		ND		11
delta-BHC		ND		11
Endosulfan sulfate		ND		11
Methoxychlor		ND		11
Toxaphene		ND		570
Chlordane (technical)		ND		280
alpha-Chlordane		ND		11
gamma-Chlordane		ND		11

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	104	50 - 125
DCB Decachlorobiphenyl	75	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-9C**

Lab Sample ID: 720-381-34
Client Matrix: Solid
% Moisture: 12.8
Date Sampled: 11/09/2005 1146
Date Received: 11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method: 8081A
Preparation: 3550B
Dilution: 10
Date Analyzed: 11/15/2005 0744
Date Prepared: 11/10/2005 1454

Analysis Batch: 720-1801
Prep Batch: 720-1635

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.23 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		23
Dieldrin		ND		23
Endrin aldehyde		ND		23
Endrin		ND		23
Endrin ketone		ND		23
Heptachlor		ND		23
Heptachlor epoxide		ND		23
4,4'-DDT		83		23
4,4'-DDE		120		23
4,4'-DDD		ND		23
Endosulfan I		ND		23
Endosulfan II		ND		23
alpha-BHC		ND		23
beta-BHC		ND		23
gamma-BHC (Lindane)		ND		23
delta-BHC		ND		23
Endosulfan sulfate		ND		23
Methoxychlor		ND		23
Toxaphene		ND		1100
alpha-Chlordane		33		23
gamma-Chlordane		36		23

Surrogate	%Rec		Acceptance Limits
Tetrachloro-m-xylene	0	D	50 - 125
DCB Decachlorobiphenyl	0	D	46 - 142

Method: 8081A
Preparation: 3550B
Dilution: 10
Date Analyzed: 11/17/2005 1346
Date Prepared: 11/10/2005 1454

Analysis Batch: 720-1801
Prep Batch: 720-1635

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.23 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Chlordane (technical)		680		570

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-11B**

Lab Sample ID:	720-381-41			Date Sampled:	11/09/2005 1208
Client Matrix:	Solid	% Moisture:	10.3	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.05 g
Date Analyzed:	11/11/2005 1310			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.2
Dieldrin		ND		2.2
Endrin aldehyde		ND		2.2
Endrin		ND		2.2
Endrin ketone		ND		2.2
Heptachlor		ND		2.2
Heptachlor epoxide		ND		2.2
4,4'-DDT		ND		2.2
4,4'-DDE		4.8		2.2
4,4'-DDD		ND		2.2
Endosulfan I		ND		2.2
Endosulfan II		ND		2.2
alpha-BHC		ND		2.2
beta-BHC		ND		2.2
gamma-BHC (Lindane)		ND		2.2
delta-BHC		ND		2.2
Endosulfan sulfate		ND		2.2
Methoxychlor		ND		2.2
Toxaphene		ND		110
Chlordane (technical)		ND		56
alpha-Chlordane		ND		2.2
gamma-Chlordane		ND		2.2

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	84	50 - 125
DCB Decachlorobiphenyl	93	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-11C**

Lab Sample ID:	720-381-42			Date Sampled:	11/09/2005 1215
Client Matrix:	Solid	% Moisture:	11.6	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.07 g
Date Analyzed:	11/11/2005 1339			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.3
Dieldrin		ND		2.3
Endrin aldehyde		ND		2.3
Endrin		ND		2.3
Endrin ketone		ND		2.3
Heptachlor		ND		2.3
Heptachlor epoxide		ND		2.3
4,4'-DDT		ND		2.3
4,4'-DDE		ND		2.3
4,4'-DDD		ND		2.3
Endosulfan I		ND		2.3
Endosulfan II		ND		2.3
alpha-BHC		ND		2.3
beta-BHC		ND		2.3
gamma-BHC (Lindane)		ND		2.3
delta-BHC		ND		2.3
Endosulfan sulfate		ND		2.3
Methoxychlor		ND		2.3
Toxaphene		ND		110
Chlordane (technical)		ND		56
alpha-Chlordane		ND		2.3
gamma-Chlordane		ND		2.3

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	83	50 - 125
DCB Decachlorobiphenyl	90	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-13B**

Lab Sample ID:	720-381-49			Date Sampled:	11/09/2005 1250
Client Matrix:	Solid	% Moisture:	7.8	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.31 g
Date Analyzed:	11/15/2005 0548			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		11
Dieldrin		ND		11
Endrin aldehyde		ND		11
Endrin		ND		11
Endrin ketone		ND		11
Heptachlor		ND		11
Heptachlor epoxide		ND		11
4,4'-DDT		21		11
4,4'-DDE		49		11
4,4'-DDD		ND		11
Endosulfan I		ND		11
Endosulfan II		ND		11
alpha-BHC		ND		11
beta-BHC		ND		11
gamma-BHC (Lindane)		ND		11
delta-BHC		ND		11
Endosulfan sulfate		ND		11
Methoxychlor		ND		11
Toxaphene		ND		540
Chlordane (technical)		ND		270
alpha-Chlordane		ND		11
gamma-Chlordane		ND		11

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	98	50 - 125
DCB Decachlorobiphenyl	74	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-13C**

Lab Sample ID:	720-381-50			Date Sampled:	11/09/2005 1252
Client Matrix:	Solid	% Moisture:	10.2	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.17 g
Date Analyzed:	11/11/2005 1407			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.2
Dieldrin		ND		2.2
Endrin aldehyde		ND		2.2
Endrin		ND		2.2
Endrin ketone		ND		2.2
Heptachlor		ND		2.2
Heptachlor epoxide		ND		2.2
4,4'-DDT		ND		2.2
4,4'-DDE		7.7		2.2
4,4'-DDD		ND		2.2
Endosulfan I		ND		2.2
Endosulfan II		ND		2.2
alpha-BHC		ND		2.2
beta-BHC		ND		2.2
gamma-BHC (Lindane)		ND		2.2
delta-BHC		ND		2.2
Endosulfan sulfate		ND		2.2
Methoxychlor		ND		2.2
Toxaphene		ND		110
Chlordane (technical)		ND		55
alpha-Chlordane		ND		2.2
gamma-Chlordane		ND		2.2

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	85	50 - 125
DCB Decachlorobiphenyl	94	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-15B**

Lab Sample ID:	720-381-57			Date Sampled:	11/09/2005 1341
Client Matrix:	Solid	% Moisture:	12.6	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.12 g
Date Analyzed:	11/11/2005 1436			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	SECONDARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.3
Dieldrin		ND		2.3
Endrin aldehyde		ND		2.3
Endrin		ND		2.3
Endrin ketone		ND		2.3
Heptachlor		ND		2.3
Heptachlor epoxide		ND		2.3
4,4'-DDT		ND		2.3
4,4'-DDE		ND		2.3
4,4'-DDD		ND		2.3
Endosulfan I		ND		2.3
Endosulfan II		ND		2.3
alpha-BHC		ND		2.3
beta-BHC		ND		2.3
gamma-BHC (Lindane)		ND		2.3
delta-BHC		ND		2.3
Endosulfan sulfate		ND		2.3
Methoxychlor		ND		2.3
Toxaphene		ND		110
Chlordane (technical)		ND		57
alpha-Chlordane		ND		2.3
gamma-Chlordane		ND		2.3

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	87	50 - 125
DCB Decachlorobiphenyl	93	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-15C**

Lab Sample ID:	720-381-58			Date Sampled:	11/09/2005 1346
Client Matrix:	Solid	% Moisture:	12.3	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.20 g
Date Analyzed:	11/17/2005 0321			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	SECONDARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.3
Dieldrin		ND		2.3
Endrin aldehyde		ND		2.3
Endrin		ND		2.3
Endrin ketone		ND		2.3
Heptachlor		ND		2.3
Heptachlor epoxide		ND		2.3
4,4'-DDT		ND		2.3
4,4'-DDE		ND		2.3
4,4'-DDD		ND		2.3
Endosulfan I		ND		2.3
Endosulfan II		ND		2.3
alpha-BHC		ND		2.3
beta-BHC		ND		2.3
gamma-BHC (Lindane)		ND		2.3
delta-BHC		ND		2.3
Endosulfan sulfate		ND		2.3
Methoxychlor		ND		2.3
Toxaphene		ND		110
Chlordane (technical)		ND		57
alpha-Chlordane		ND		2.3
gamma-Chlordane		ND		2.3

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	90	50 - 125
DCB Decachlorobiphenyl	91	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-17B**

Lab Sample ID:	720-381-65			Date Sampled:	11/09/2005 1408
Client Matrix:	Solid	% Moisture:	9.2	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.27 g
Date Analyzed:	11/15/2005 0617			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	SECONDARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		11
Dieldrin		ND		11
Endrin aldehyde		ND		11
Endrin		ND		11
Endrin ketone		ND		11
Heptachlor		ND		11
Heptachlor epoxide		ND		11
4,4'-DDT		17		11
4,4'-DDE		58		11
4,4'-DDD		ND		11
Endosulfan I		ND		11
Endosulfan II		ND		11
alpha-BHC		ND		11
beta-BHC		ND		11
gamma-BHC (Lindane)		ND		11
delta-BHC		ND		11
Endosulfan sulfate		ND		11
Methoxychlor		ND		11
Toxaphene		ND		550
Chlordane (technical)		ND		270
alpha-Chlordane		ND		11
gamma-Chlordane		ND		11

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	112	50 - 125
DCB Decachlorobiphenyl	76	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-17C**

Lab Sample ID:	720-381-66			Date Sampled:	11/09/2005 1411
Client Matrix:	Solid	% Moisture:	6.3	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.21 g
Date Analyzed:	11/15/2005 0646			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	SECONDARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		11
Dieldrin		ND		11
Endrin aldehyde		ND		11
Endrin		ND		11
Endrin ketone		ND		11
Heptachlor		ND		11
Heptachlor epoxide		ND		11
4,4'-DDT		ND		11
4,4'-DDE		ND		11
4,4'-DDD		ND		11
Endosulfan I		ND		11
Endosulfan II		ND		11
alpha-BHC		ND		11
beta-BHC		ND		11
gamma-BHC (Lindane)		ND		11
delta-BHC		ND		11
Endosulfan sulfate		ND		11
Methoxychlor		ND		11
Toxaphene		ND		530
Chlordane (technical)		ND		270
alpha-Chlordane		ND		11
gamma-Chlordane		ND		11

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	110	50 - 125
DCB Decachlorobiphenyl	91	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-19B**

Lab Sample ID:	720-381-73			Date Sampled:	11/09/2005 1430
Client Matrix:	Solid	% Moisture:	13.9	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1801		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1635		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.22 g
Date Analyzed:	11/15/2005 0715			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1454			Injection Volume:	
				Column ID:	SECONDARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		12
Dieldrin		ND		12
Endrin aldehyde		ND		12
Endrin		ND		12
Endrin ketone		ND		12
Heptachlor		ND		12
Heptachlor epoxide		ND		12
4,4'-DDT		19		12
4,4'-DDE		120		12
4,4'-DDD		49		12
Endosulfan I		ND		12
Endosulfan II		ND		12
alpha-BHC		ND		12
beta-BHC		ND		12
gamma-BHC (Lindane)		ND		12
delta-BHC		ND		12
Endosulfan sulfate		ND		12
Methoxychlor		ND		12
Toxaphene		ND		580
Chlordane (technical)		ND		290
alpha-Chlordane		ND		12
gamma-Chlordane		ND		12

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	103	50 - 125
DCB Decachlorobiphenyl	89	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-19C**

Lab Sample ID:	720-381-74			Date Sampled:	11/09/2005 1430
Client Matrix:	Solid	% Moisture:	12.0	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1940		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1646		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.25 g
Date Analyzed:	11/15/2005 1003			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1810			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.3
Dieldrin		ND		2.3
Endrin aldehyde		ND		2.3
Endrin		ND		2.3
Endrin ketone		ND		2.3
Heptachlor		ND		2.3
Heptachlor epoxide		ND		2.3
4,4'-DDT		ND		2.3
4,4'-DDE		ND		2.3
4,4'-DDD		ND		2.3
Endosulfan I		ND		2.3
Endosulfan II		ND		2.3
alpha-BHC		ND		2.3
beta-BHC		ND		2.3
gamma-BHC (Lindane)		ND		2.3
delta-BHC		ND		2.3
Endosulfan sulfate		ND		2.3
Methoxychlor		ND		2.3
Toxaphene		ND		110
Chlordane (technical)		ND		56
alpha-Chlordane		ND		2.3
gamma-Chlordane		ND		2.3

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	81	50 - 125
DCB Decachlorobiphenyl	87	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-21B**

Lab Sample ID:	720-381-81			Date Sampled:	11/09/2005 1450
Client Matrix:	Solid	% Moisture:	12.7	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1940		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1646		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.20 g
Date Analyzed:	11/16/2005 1745			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1810			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		11
Dieldrin		ND		11
Endrin aldehyde		ND		11
Endrin		ND		11
Endrin ketone		ND		11
Heptachlor		ND		11
Heptachlor epoxide		ND		11
4,4'-DDT		55		11
4,4'-DDE		260		11
4,4'-DDD		ND		11
Endosulfan I		ND		11
Endosulfan II		ND		11
alpha-BHC		ND		11
beta-BHC		ND		11
gamma-BHC (Lindane)		ND		11
delta-BHC		ND		11
Endosulfan sulfate		ND		11
Methoxychlor		ND		11
Toxaphene		ND		570
Chlordane (technical)		ND		280
alpha-Chlordane		ND		11
gamma-Chlordane		ND		11

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	99	50 - 125
DCB Decachlorobiphenyl	103	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-21C**

Lab Sample ID:	720-381-82			Date Sampled:	11/09/2005 1450
Client Matrix:	Solid	% Moisture:	9.0	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1940		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1646		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.29 g
Date Analyzed:	11/16/2005 1814			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1810			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.2
Dieldrin		ND		2.2
Endrin aldehyde		ND		2.2
Endrin		ND		2.2
Endrin ketone		ND		2.2
Heptachlor		ND		2.2
Heptachlor epoxide		ND		2.2
4,4'-DDT		ND		2.2
4,4'-DDE		2.6		2.2
4,4'-DDD		ND		2.2
Endosulfan I		ND		2.2
Endosulfan II		ND		2.2
alpha-BHC		ND		2.2
beta-BHC		ND		2.2
gamma-BHC (Lindane)		ND		2.2
delta-BHC		ND		2.2
Endosulfan sulfate		ND		2.2
Methoxychlor		ND		2.2
Toxaphene		ND		110
Chlordane (technical)		ND		54
alpha-Chlordane		ND		2.2
gamma-Chlordane		ND		2.2

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	102	50 - 125
DCB Decachlorobiphenyl	105	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-23B**

Lab Sample ID:	720-381-89			Date Sampled:	11/09/2005 1515
Client Matrix:	Solid	% Moisture:	11.5	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1940		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1646		Lab File ID:	N/A
Dilution:	10			Initial Weight/Volume:	30.21 g
Date Analyzed:	11/16/2005 1843			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1810			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		22
Dieldrin		ND		22
Endrin aldehyde		ND		22
Endrin		ND		22
Endrin ketone		ND		22
Heptachlor		ND		22
Heptachlor epoxide		ND		22
4,4'-DDT		63		22
4,4'-DDE		190		22
4,4'-DDD		ND		22
Endosulfan I		ND		22
Endosulfan II		ND		22
alpha-BHC		ND		22
beta-BHC		ND		22
gamma-BHC (Lindane)		ND		22
delta-BHC		ND		22
Endosulfan sulfate		ND		22
Methoxychlor		ND		22
Toxaphene		ND		1100
Chlordane (technical)		ND		560
alpha-Chlordane		23		22
gamma-Chlordane		ND		22

Surrogate	%Rec		Acceptance Limits
Tetrachloro-m-xylene	0	*	50 - 125
DCB Decachlorobiphenyl	0	*	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-23C**

Lab Sample ID:	720-381-90			Date Sampled:	11/09/2005 1520
Client Matrix:	Solid	% Moisture:	11.6	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1940		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1646		Lab File ID:	N/A
Dilution:	10			Initial Weight/Volume:	30.17 g
Date Analyzed:	11/16/2005 1912			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1810			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		22
Dieldrin		ND		22
Endrin aldehyde		ND		22
Endrin		ND		22
Endrin ketone		ND		22
Heptachlor		ND		22
Heptachlor epoxide		ND		22
4,4'-DDT		140		22
4,4'-DDE		98		22
4,4'-DDD		55		22
Endosulfan I		ND		22
Endosulfan II		ND		22
alpha-BHC		ND		22
beta-BHC		ND		22
gamma-BHC (Lindane)		ND		22
delta-BHC		ND		22
Endosulfan sulfate		ND		22
Methoxychlor		ND		22
Toxaphene		ND		1100
Chlordane (technical)		ND		560
alpha-Chlordane		ND		22
gamma-Chlordane		ND		22

Surrogate	%Rec		Acceptance Limits
Tetrachloro-m-xylene	0	*	50 - 125
DCB Decachlorobiphenyl	0	*	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-27B**

Lab Sample ID:	720-381-101			Date Sampled:	11/09/2005 1600
Client Matrix:	Solid	% Moisture:	10.5	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1940		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1646		Lab File ID:	N/A
Dilution:	10			Initial Weight/Volume:	30.21 g
Date Analyzed:	11/16/2005 1840			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1810			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		22
Dieldrin		ND		22
Endrin aldehyde		ND		22
Endrin		ND		22
Endrin ketone		ND		22
Heptachlor		ND		22
Heptachlor epoxide		ND		22
4,4'-DDT		200		22
4,4'-DDE		360		22
4,4'-DDD		ND		22
Endosulfan I		ND		22
Endosulfan II		ND		22
alpha-BHC		ND		22
beta-BHC		ND		22
gamma-BHC (Lindane)		ND		22
delta-BHC		ND		22
Endosulfan sulfate		ND		22
Methoxychlor		ND		22
Toxaphene		ND		1100
Chlordane (technical)		ND		550
alpha-Chlordane		ND		22
gamma-Chlordane		ND		22

Surrogate	%Rec		Acceptance Limits
Tetrachloro-m-xylene	0	*	50 - 125
DCB Decachlorobiphenyl	0	*	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-27C**

Lab Sample ID:	720-381-102			Date Sampled:	11/09/2005 1605
Client Matrix:	Solid	% Moisture:	12.7	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1940		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1646		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.16 g
Date Analyzed:	11/16/2005 2215			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1810			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		11
Dieldrin		ND		11
Endrin aldehyde		ND		11
Endrin		ND		11
Endrin ketone		ND		11
Heptachlor		ND		11
Heptachlor epoxide		ND		11
4,4'-DDT		ND		11
4,4'-DDE		ND		11
4,4'-DDD		ND		11
Endosulfan I		ND		11
Endosulfan II		ND		11
alpha-BHC		ND		11
beta-BHC		ND		11
gamma-BHC (Lindane)		ND		11
delta-BHC		ND		11
Endosulfan sulfate		ND		11
Methoxychlor		ND		11
Toxaphene		ND		570
Chlordane (technical)		ND		280
alpha-Chlordane		ND		11
gamma-Chlordane		ND		11

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	89	50 - 125
DCB Decachlorobiphenyl	82	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-25B**

Lab Sample ID:	720-381-105			Date Sampled:	11/09/2005 1620
Client Matrix:	Solid	% Moisture:	10.3	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1940		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1646		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.10 g
Date Analyzed:	11/16/2005 2244			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1810			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.2
Dieldrin		ND		2.2
Endrin aldehyde		ND		2.2
Endrin		ND		2.2
Endrin ketone		ND		2.2
Heptachlor		ND		2.2
Heptachlor epoxide		ND		2.2
4,4'-DDT		ND		2.2
4,4'-DDE		ND		2.2
4,4'-DDD		ND		2.2
Endosulfan I		ND		2.2
Endosulfan II		ND		2.2
alpha-BHC		ND		2.2
beta-BHC		ND		2.2
gamma-BHC (Lindane)		ND		2.2
delta-BHC		ND		2.2
Endosulfan sulfate		ND		2.2
Methoxychlor		ND		2.2
Toxaphene		ND		110
Chlordane (technical)		ND		56
alpha-Chlordane		ND		2.2
gamma-Chlordane		ND		2.2

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	95	50 - 125
DCB Decachlorobiphenyl	93	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-51B**

Lab Sample ID:	720-381-107			Date Sampled:	11/09/2005 1501
Client Matrix:	Solid	% Moisture:	9.7	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1940		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1646		Lab File ID:	N/A
Dilution:	10			Initial Weight/Volume:	30.17 g
Date Analyzed:	11/16/2005 2313			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1810			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		22
Dieldrin		ND		22
Endrin aldehyde		ND		22
Endrin		ND		22
Endrin ketone		ND		22
Heptachlor		ND		22
Heptachlor epoxide		ND		22
4,4'-DDT		ND		22
4,4'-DDE		ND		22
4,4'-DDD		ND		22
Endosulfan I		ND		22
Endosulfan II		ND		22
alpha-BHC		ND		22
beta-BHC		ND		22
gamma-BHC (Lindane)		ND		22
delta-BHC		ND		22
Endosulfan sulfate		ND		22
Methoxychlor		ND		22
Toxaphene		ND		1100
Chlordane (technical)		ND		550
alpha-Chlordane		ND		22
gamma-Chlordane		ND		22

Surrogate	%Rec		Acceptance Limits
Tetrachloro-m-xylene	0	*	50 - 125
DCB Decachlorobiphenyl	0	*	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-51C**

Lab Sample ID: 720-381-108
Client Matrix: Solid
% Moisture: 14.8
Date Sampled: 11/09/2005 1501
Date Received: 11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method: 8081A
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/16/2005 2009
Date Prepared: 11/10/2005 1810

Analysis Batch: 720-1940
Prep Batch: 720-1646

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.25 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.3
Dieldrin		ND		2.3
Endrin aldehyde		ND		2.3
Endrin		ND		2.3
Endrin ketone		ND		2.3
Heptachlor		ND		2.3
Heptachlor epoxide		ND		2.3
4,4'-DDT		9.8		2.3
4,4'-DDE		10		2.3
4,4'-DDD		ND		2.3
Endosulfan I		ND		2.3
Endosulfan II		ND		2.3
alpha-BHC		ND		2.3
beta-BHC		ND		2.3
gamma-BHC (Lindane)		ND		2.3
delta-BHC		ND		2.3
Endosulfan sulfate		ND		2.3
Methoxychlor		ND		2.3
Toxaphene		ND		120
Chlordane (technical)		ND		58
alpha-Chlordane		ND		2.3
gamma-Chlordane		ND		2.3

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	92	50 - 125

Method: 8081A
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/16/2005 2009
Date Prepared: 11/10/2005 1810

Analysis Batch: 720-1940
Prep Batch: 720-1646

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.25 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: SECONDARY

Surrogate	%Rec	Acceptance Limits
DCB Decachlorobiphenyl	92	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-52B**

Lab Sample ID: 720-381-110
Client Matrix: Solid
% Moisture: 8.5
Date Sampled: 11/09/2005 1600
Date Received: 11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method: 8081A
Preparation: 3550B
Dilution: 10
Date Analyzed: 11/16/2005 2038
Date Prepared: 11/10/2005 1810

Analysis Batch: 720-1940
Prep Batch: 720-1646

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.12 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		22
Dieldrin		ND		22
Endrin aldehyde		ND		22
Endrin		ND		22
Endrin ketone		ND		22
Heptachlor		ND		22
Heptachlor epoxide		ND		22
4,4'-DDT		36		22
4,4'-DDE		83		22
4,4'-DDD		ND		22
Endosulfan I		ND		22
Endosulfan II		ND		22
alpha-BHC		ND		22
beta-BHC		ND		22
gamma-BHC (Lindane)		ND		22
delta-BHC		ND		22
Endosulfan sulfate		ND		22
Methoxychlor		ND		22
Toxaphene		ND		1100
Chlordane (technical)		ND		540
alpha-Chlordane		ND		22
gamma-Chlordane		ND		22

Surrogate	%Rec		Acceptance Limits
Tetrachloro-m-xylene	0	*	50 - 125

Method: 8081A
Preparation: 3550B
Dilution: 10
Date Analyzed: 11/16/2005 2038
Date Prepared: 11/10/2005 1810

Analysis Batch: 720-1940
Prep Batch: 720-1646

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.12 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: SECONDARY

Surrogate	%Rec		Acceptance Limits
DCB Decachlorobiphenyl	0	*	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-52C**

Lab Sample ID:	720-381-111			Date Sampled:	11/09/2005 1600
Client Matrix:	Solid	% Moisture:	9.4	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1940		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1646		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.15 g
Date Analyzed:	11/16/2005 2107			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1810			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.2
Dieldrin		ND		2.2
Endrin aldehyde		ND		2.2
Endrin		ND		2.2
Endrin ketone		ND		2.2
Heptachlor		ND		2.2
Heptachlor epoxide		ND		2.2
4,4'-DDT		ND		2.2
4,4'-DDE		7.9		2.2
4,4'-DDD		ND		2.2
Endosulfan I		ND		2.2
Endosulfan II		ND		2.2
alpha-BHC		ND		2.2
beta-BHC		ND		2.2
gamma-BHC (Lindane)		ND		2.2
delta-BHC		ND		2.2
Endosulfan sulfate		ND		2.2
Methoxychlor		ND		2.2
Toxaphene		ND		110
Chlordane (technical)		ND		55
alpha-Chlordane		ND		2.2
gamma-Chlordane		ND		2.2

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	97	50 - 125
DCB Decachlorobiphenyl	95	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: **EBHH-53B**

Lab Sample ID:	720-381-113			Date Sampled:	11/09/2005 1515
Client Matrix:	Solid	% Moisture:	11.2	Date Received:	11/09/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1940		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1646		Lab File ID:	N/A
Dilution:	10			Initial Weight/Volume:	30.19 g
Date Analyzed:	11/16/2005 2136			Final Weight/Volume:	10 mL
Date Prepared:	11/10/2005 1810			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		22
Dieldrin		ND		22
Endrin aldehyde		ND		22
Endrin		ND		22
Endrin ketone		ND		22
Heptachlor		ND		22
Heptachlor epoxide		ND		22
4,4'-DDT		57		22
4,4'-DDE		180		22
4,4'-DDD		ND		22
Endosulfan I		ND		22
Endosulfan II		ND		22
alpha-BHC		ND		22
beta-BHC		ND		22
gamma-BHC (Lindane)		ND		22
delta-BHC		ND		22
Endosulfan sulfate		ND		22
Methoxychlor		ND		22
Toxaphene		ND		1100
Chlordane (technical)		ND		560
alpha-Chlordane		ND		22
gamma-Chlordane		ND		22

Surrogate	%Rec		Acceptance Limits
Tetrachloro-m-xylene	0	D	50 - 125
DCB Decachlorobiphenyl	0	D	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-1A

Lab Sample ID:	720-381-1	Date Sampled:	11/09/2005 0900
Client Matrix:	Solid % Moisture: 9.5	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/11/2005 1207		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		14		1.1
Barium		100		1.1
Cadmium		1.9		0.55
Chromium		38		1.1
Lead		38		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 0905		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0043	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-1B

Lab Sample ID:	720-381-2		Date Sampled:	11/09/2005 0925
Client Matrix:	Solid	% Moisture: 7.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/11/2005 1219		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		4.6		1.1
Barium		97		1.1
Cadmium		1.3		0.54
Chromium		35		1.1
Lead		70		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 0906		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0013	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-1C

Lab Sample ID:	720-381-3	Date Sampled:	11/09/2005 0940
Client Matrix:	Solid % Moisture: 11.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683		Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/11/2005 1222			Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		4.6		1.1
Barium		150		1.1
Cadmium		1.5		0.57
Chromium		36		1.1
Lead		8.3		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760		Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 0907			Final Weight/Volume:	
Date Prepared:	11/11/2005 1159				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-2B

Lab Sample ID:	720-381-5		Date Sampled:	11/09/2005 0945
Client Matrix:	Solid	% Moisture: 14.5	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/11/2005 1226		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.2
Arsenic		5.3		1.2
Barium		180		1.2
Cadmium		1.8		0.58
Chromium		41		1.2
Lead		48		1.2
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/15/2005 0908		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0018	*	0.0012

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-2C

Lab Sample ID:	720-381-6		Date Sampled:	11/09/2005 0951
Client Matrix:	Solid	% Moisture: 15.3	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/11/2005 1237		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.2
Arsenic		5.8		1.2
Barium		150		1.2
Cadmium		1.6		0.58
Chromium		40		1.2
Lead		18		1.2
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0910		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0013	*	0.0012

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-2A

Lab Sample ID:	720-381-8		Date Sampled:	11/09/2005 1102
Client Matrix:	Solid	% Moisture: 9.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/11/2005 1240		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		6.7		1.1
Barium		320		1.1
Cadmium		1.2		0.55
Chromium		22		1.1
Lead		130		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/15/2005 0911		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0047		0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-3B

Lab Sample ID:	720-381-9	Date Sampled:	11/09/2005 1000
Client Matrix:	Solid % Moisture: 8.3	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/11/2005 1244		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		1.9		1.1
Arsenic		6.6		1.1
Barium		120		1.1
Cadmium		2.3		0.53
Chromium		33		1.1
Lead		110		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 0912		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0028	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-3C

Lab Sample ID:	720-381-10		Date Sampled:	11/09/2005 1000
Client Matrix:	Solid	% Moisture: 15.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/11/2005 1248		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.7		1.1
Barium		160		1.1
Cadmium		1.8		0.57
Chromium		37		1.1
Lead		13		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0916		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.0012

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-3A

Lab Sample ID:	720-381-12	Date Sampled:	11/09/2005 1105
Client Matrix:	Solid % Moisture: 10.1	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683		Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.03 g
Date Analyzed:	11/11/2005 1252			Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.4		1.1
Barium		230		1.1
Cadmium		0.88		0.54
Chromium		18		1.1
Lead		100		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760		Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0917			Final Weight/Volume:	
Date Prepared:	11/11/2005 1159				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0041	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-4B

Lab Sample ID:	720-381-13		Date Sampled:	11/09/2005 1015
Client Matrix:	Solid	% Moisture: 12.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/11/2005 1256		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.5		1.1
Barium		190		1.1
Cadmium		1.9		0.57
Chromium		39		1.1
Lead		170		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0918		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0020	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-4C

Lab Sample ID:	720-381-14		Date Sampled:	11/09/2005 1020
Client Matrix:	Solid	% Moisture: 12.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/11/2005 1300		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		7.2		1.1
Barium		170		1.1
Cadmium		2.0		0.56
Chromium		38		1.1
Lead		30		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0919		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0019	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-4A

Lab Sample ID:	720-381-16		Date Sampled:	11/09/2005 1110
Client Matrix:	Solid	% Moisture: 9.5	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/11/2005 1304		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		8.7		1.1
Barium		190		1.1
Cadmium		1.0		0.55
Chromium		21		1.1
Lead		120		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/15/2005 0921		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0038	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-5B

Lab Sample ID:	720-381-17	Date Sampled:	11/09/2005 1030
Client Matrix:	Solid % Moisture: 12.2	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/11/2005 1307		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		4.8		1.1
Barium		170		1.1
Cadmium		1.7		0.57
Chromium		33		1.1
Lead		130		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0922		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0014	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-5C

Lab Sample ID:	720-381-18		Date Sampled:	11/09/2005 1036
Client Matrix:	Solid	% Moisture: 15.5	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/11/2005 1211		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.2
Arsenic		5.8		1.2
Barium		180		1.2
Cadmium		1.6		0.59
Chromium		39		1.2
Lead		7.4		1.2
Selenium		ND		2.4

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0923		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0012	*	0.0012

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-5A

Lab Sample ID:	720-381-20		Date Sampled:	11/09/2005 1115
Client Matrix:	Solid	% Moisture: 10.3	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/11/2005 1322		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.4		1.1
Barium		140		1.1
Cadmium		1.9		0.54
Chromium		22		1.1
Lead		130		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 0924		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0027	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-6B

Lab Sample ID:	720-381-21		Date Sampled:	11/09/2005 1045
Client Matrix:	Solid	% Moisture: 9.4	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/11/2005 1215		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		8.9		1.1
Barium		530		1.1
Cadmium		2.8		0.55
Chromium		43		1.1
Lead		700		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0930		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0034	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-6C

Lab Sample ID:	720-381-22		Date Sampled:	11/09/2005 1045
Client Matrix:	Solid	% Moisture: 9.1	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/11/2005 1333		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.8		1.1
Barium		150		1.1
Cadmium		2.2		0.55
Chromium		29		1.1
Lead		78		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0932		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0012	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-6A

Lab Sample ID:	720-381-24		Date Sampled:	11/09/2005 1118
Client Matrix:	Solid	% Moisture: 12.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/11/2005 1337		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.2		1.1
Barium		150		1.1
Cadmium		1.5		0.57
Chromium		37		1.1
Lead		9.2		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0933		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0032	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-7B

Lab Sample ID:	720-381-25		Date Sampled:	11/09/2005 1125
Client Matrix:	Solid	% Moisture: 15.3	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1808	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1318		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.2
Silver		ND		1.2
Arsenic		6.0		1.2
Arsenic		6.0		1.2
Barium		190		1.2
Barium		190		1.2
Cadmium		1.6		0.59
Cadmium		1.6		0.59
Chromium		40		1.2
Chromium		40		1.2
Lead		49		1.2
Lead		49		1.2
Selenium		ND		2.4
Selenium		ND		2.4

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 1405		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.058

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-7C

Lab Sample ID:	720-381-26	Date Sampled:	11/09/2005 1125
Client Matrix:	Solid % Moisture: 13.5	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1808		Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1322			Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.2
Silver		ND		1.2
Arsenic		4.9		1.2
Arsenic		4.9		1.2
Barium		180		1.2
Barium		180		1.2
Cadmium		1.2		0.58
Cadmium		1.2		0.58
Chromium		31		1.2
Chromium		31		1.2
Lead		5.9		1.2
Lead		5.9		1.2
Selenium		ND		2.3
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807		Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1406			Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.058

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-7A

Lab Sample ID:	720-381-28		Date Sampled:	11/09/2005 1145
Client Matrix:	Solid	% Moisture: 9.0	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1326		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Silver		ND		1.1
Arsenic		5.6		1.1
Arsenic		5.6		1.1
Barium		240		1.1
Barium		240		1.1
Cadmium		0.79		0.55
Cadmium		0.79		0.55
Chromium		20		1.1
Chromium		20		1.1
Lead		110		1.1
Lead		110		1.1
Selenium		ND		2.2
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1407		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.28		0.055

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-8B

Lab Sample ID:	720-381-29		Date Sampled:	11/09/2005 1135
Client Matrix:	Solid	% Moisture: 15.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1330		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.2
Silver		ND		1.2
Arsenic		7.3		1.2
Arsenic		7.3		1.2
Barium		190		1.2
Barium		190		1.2
Cadmium		1.8		0.59
Cadmium		1.8		0.59
Chromium		42		1.2
Chromium		42		1.2
Lead		75		1.2
Lead		75		1.2
Selenium		ND		2.4
Selenium		ND		2.4

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 1408		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.075		0.058

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-8C

Lab Sample ID:	720-381-30		Date Sampled:	11/09/2005 1135
Client Matrix:	Solid	% Moisture: 12.0	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/15/2005 1334		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Silver		ND		1.1
Arsenic		5.3		1.1
Arsenic		5.3		1.1
Barium		140		1.1
Barium		140		1.1
Cadmium		1.2		0.55
Cadmium		1.2		0.55
Chromium		30		1.1
Chromium		30		1.1
Lead		5.8		1.1
Lead		5.8		1.1
Selenium		ND		2.2
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/15/2005 1409		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.055

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-8A

Lab Sample ID:	720-381-32		Date Sampled:	11/09/2005 1150
Client Matrix:	Solid	% Moisture: 10.2	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/15/2005 1338		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Silver		ND		1.1
Arsenic		5.5		1.1
Arsenic		5.5		1.1
Barium		140		1.1
Barium		140		1.1
Cadmium		0.76		0.55
Cadmium		0.76		0.55
Chromium		20		1.1
Chromium		20		1.1
Lead		88		1.1
Lead		88		1.1
Selenium		ND		2.2
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1413		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.26		0.056

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-9B

Lab Sample ID:	720-381-33		Date Sampled:	11/09/2005 1146
Client Matrix:	Solid	% Moisture: 12.1	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/15/2005 1355		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Silver		ND		1.1
Arsenic		6.5		1.1
Arsenic		6.5		1.1
Barium		160		1.1
Barium		160		1.1
Cadmium		2.2		0.56
Cadmium		2.2		0.56
Chromium		38		1.1
Chromium		38		1.1
Lead		230		1.1
Lead		230		1.1
Selenium		ND		2.2
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1414		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.14		0.057

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-9C

Lab Sample ID:	720-381-34		Date Sampled:	11/09/2005 1146
Client Matrix:	Solid	% Moisture: 12.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1808	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1359		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Silver		ND		1.1
Arsenic		6.4		1.1
Arsenic		6.4		1.1
Barium		170		1.1
Barium		170		1.1
Cadmium		1.5		0.57
Cadmium		1.5		0.57
Chromium		39		1.1
Chromium		39		1.1
Lead		33		1.1
Lead		33		1.1
Selenium		ND		2.3
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 1415		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.071		0.057

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-9A

Lab Sample ID:	720-381-36		Date Sampled:	11/09/2005 1155
Client Matrix:	Solid	% Moisture: 7.2	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1808	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 1403		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Silver		ND		1.1
Arsenic		12		1.1
Arsenic		12		1.1
Barium		120		1.1
Barium		120		1.1
Cadmium		1.3		0.53
Cadmium		1.3		0.53
Chromium		17		1.1
Chromium		17		1.1
Lead		320		1.1
Lead		320		1.1
Selenium		ND		2.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1417		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.083		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-10B

Lab Sample ID:	720-381-37	Date Sampled:	11/09/2005 1203
Client Matrix:	Solid % Moisture: 10.1	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1808		Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1407			Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Silver		ND		1.1
Arsenic		7.5		1.1
Arsenic		7.5		1.1
Barium		180		1.1
Barium		180		1.1
Cadmium		1.8		0.56
Cadmium		1.8		0.56
Chromium		29		1.1
Chromium		29		1.1
Lead		130		1.1
Lead		130		1.1
Selenium		ND		2.2
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807		Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 1418			Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.18		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-10C

Lab Sample ID:	720-381-38		Date Sampled:	11/09/2005 1203
Client Matrix:	Solid	% Moisture: 8.9	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1808	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/15/2005 1411		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Silver		ND		1.1
Arsenic		4.3		1.1
Arsenic		4.3		1.1
Barium		120		1.1
Barium		120		1.1
Cadmium		1.2		0.54
Cadmium		1.2		0.54
Chromium		28		1.1
Chromium		28		1.1
Lead		7.7		1.1
Lead		7.7		1.1
Selenium		ND		2.2
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1419		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.055

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-10A

Lab Sample ID:	720-381-40		Date Sampled:	11/09/2005 1340
Client Matrix:	Solid	% Moisture: 12.4	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1808	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1415		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Silver		ND		1.0
Arsenic		6.6		1.1
Arsenic		5.8		1.0
Barium		140		1.1
Barium		130		1.0
Cadmium		2.6		0.57
Cadmium		2.3		0.50
Chromium		35		1.1
Chromium		30		1.0
Lead		120		1.1
Lead		100		1.0
Selenium		ND		2.3
Selenium		ND		2.0

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 1420		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.14		0.055

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-11B

Lab Sample ID:	720-381-41		Date Sampled:	11/09/2005 1208
Client Matrix:	Solid	% Moisture: 10.3	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1808	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/15/2005 1426		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		7.0		1.1
Barium		170		1.1
Cadmium		1.7		0.53
Chromium		34		1.1
Lead		190		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 1424		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.11		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-11C

Lab Sample ID:	720-381-42		Date Sampled:	11/09/2005 1215
Client Matrix:	Solid	% Moisture: 11.6	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1810	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 1430		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.6		1.1
Barium		180		1.1
Cadmium		1.4		0.55
Chromium		37		1.1
Lead		11		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/15/2005 1428		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-11A

Lab Sample ID:	720-381-44		Date Sampled:	11/09/2005 1243
Client Matrix:	Solid	% Moisture: 7.3	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1810	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 1856		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		7.0		1.1
Barium		220		1.1
Cadmium		0.80		0.53
Chromium		19		1.1
Lead		100		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1431		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.11		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-12B

Lab Sample ID:	720-381-45		Date Sampled:	11/09/2005 1223
Client Matrix:	Solid	% Moisture: 12.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1810	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1900		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		6.0		1.1
Barium		150		1.1
Cadmium		1.5		0.57
Chromium		39		1.1
Lead		14		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1432		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.061		0.057

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-12C

Lab Sample ID:	720-381-46		Date Sampled:	11/09/2005 1223
Client Matrix:	Solid	% Moisture: 11.2	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1810	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 1904		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.3		1.1
Barium		160		1.1
Cadmium		1.4		0.56
Chromium		32		1.1
Lead		7.1		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 1434		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.055

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-12A

Lab Sample ID:	720-381-48		Date Sampled:	11/09/2005 1248
Client Matrix:	Solid	% Moisture: 6.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1951	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1734	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/17/2005 1940		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.7		1.1
Barium		150		1.1
Cadmium		0.92		0.53
Chromium		18		1.1
Lead		77		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1807	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1743	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1435		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1700			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.10		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-13B

Lab Sample ID:	720-381-49		Date Sampled:	11/09/2005 1250
Client Matrix:	Solid	% Moisture: 7.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/15/2005 0825		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.0
Arsenic		5.1		1.0
Barium		130		1.0
Cadmium		2.2		0.52
Chromium		39		1.0
Lead		77		1.0
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 0939		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.12		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-13C

Lab Sample ID:	720-381-50		Date Sampled:	11/09/2005 1252
Client Matrix:	Solid	% Moisture: 10.2	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/15/2005 0829		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.6		1.1
Barium		120		1.1
Cadmium		1.1		0.53
Chromium		32		1.1
Lead		6.2		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 0940		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-13A

Lab Sample ID:	720-381-52		Date Sampled:	11/09/2005 1350
Client Matrix:	Solid	% Moisture: 14.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 0833		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		20		1.1
Barium		130		1.1
Cadmium		1.8		0.57
Chromium		27		1.1
Lead		590		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/15/2005 0941		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.19		0.056

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-14B

Lab Sample ID:	720-381-53		Date Sampled:	11/09/2005 1301
Client Matrix:	Solid	% Moisture: 7.5	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 0837		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.2		1.1
Barium		130		1.1
Cadmium		1.2		0.54
Chromium		30		1.1
Lead		69		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0945		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.084		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-14C

Lab Sample ID:	720-381-54		Date Sampled:	11/09/2005 1309
Client Matrix:	Solid	% Moisture: 6.4	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 0841		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		4.3		1.1
Barium		110		1.1
Cadmium		1.1		0.53
Chromium		32		1.1
Lead		18		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0946		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.053

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-14A

Lab Sample ID:	720-381-56		Date Sampled:	11/09/2005 1445
Client Matrix:	Solid	% Moisture: 8.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 0845		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		0.97
Arsenic		5.2		0.97
Barium		140		0.97
Cadmium		2.0		0.49
Chromium		29		0.97
Lead		110		0.97
Selenium		ND		1.9

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/15/2005 0947		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.11		0.048

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-15B

Lab Sample ID:	720-381-57		Date Sampled:	11/09/2005 1341
Client Matrix:	Solid	% Moisture: 12.6	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/15/2005 0848		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		3.9		1.1
Barium		110		1.1
Cadmium		1.2		0.55
Chromium		29		1.1
Lead		7.3		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 0949		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.056

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-15C

Lab Sample ID:	720-381-58		Date Sampled:	11/09/2005 1346
Client Matrix:	Solid	% Moisture: 12.3	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/15/2005 0859		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		4.3		1.1
Barium		180		1.1
Cadmium		1.2		0.55
Chromium		36		1.1
Lead		6.8		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 0950		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.055

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-15A

Lab Sample ID:	720-381-60		Date Sampled:	11/09/2005 1450
Client Matrix:	Solid	% Moisture: 22.3	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/15/2005 0910		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.3
Arsenic		9.2		1.3
Barium		220		1.3
Cadmium		1.4		0.63
Chromium		31		1.3
Lead		170		1.3
Selenium		ND		2.5

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/15/2005 0954		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.19		0.061

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-16B

Lab Sample ID:	720-381-61		Date Sampled:	11/09/2005 1358
Client Matrix:	Solid	% Moisture: 6.9	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0914		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.3		1.1
Barium		170		1.1
Cadmium		1.2		0.54
Chromium		32		1.1
Lead		15		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/15/2005 0955		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.066		0.052

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-16C

Lab Sample ID:	720-381-62		Date Sampled:	11/09/2005 1402
Client Matrix:	Solid	% Moisture: 10.5	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 0918		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		4.5		1.1
Barium		120		1.1
Cadmium		1.0		0.55
Chromium		29		1.1
Lead		5.5		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/15/2005 0956		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.055

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-16A

Lab Sample ID:	720-381-64		Date Sampled:	11/09/2005 1455
Client Matrix:	Solid	% Moisture: 11.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/15/2005 0922		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.3		1.1
Barium		140		1.1
Cadmium		1.1		0.54
Chromium		20		1.1
Lead		55		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/15/2005 1000		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.30		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-17B

Lab Sample ID:	720-381-65		Date Sampled:	11/09/2005 1408
Client Matrix:	Solid	% Moisture: 9.2	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/15/2005 0926		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.3		1.1
Barium		140		1.1
Cadmium		1.2		0.54
Chromium		31		1.1
Lead		17		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 1001		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.15		0.055

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-17C

Lab Sample ID:	720-381-66		Date Sampled:	11/09/2005 1411
Client Matrix:	Solid	% Moisture: 6.3	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0932		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		4.7		1.1
Barium		99		1.1
Cadmium		1.0		0.53
Chromium		29		1.1
Lead		13		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 1002		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.055		0.052

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-17A

Lab Sample ID:	720-381-68		Date Sampled:	11/09/2005 1500
Client Matrix:	Solid	% Moisture: 8.4	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 0935		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		7.8		1.1
Barium		190		1.1
Cadmium		0.99		0.53
Chromium		22		1.1
Lead		170		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/15/2005 1003		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.15		0.052

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-18B

Lab Sample ID:	720-381-69		Date Sampled:	11/09/2005 1422
Client Matrix:	Solid	% Moisture: 8.2	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0946		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		4.7		1.1
Barium		130		1.1
Cadmium		1.2		0.54
Chromium		33		1.1
Lead		12		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 1005		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.057		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-18C

Lab Sample ID:	720-381-70		Date Sampled:	11/09/2005 1422
Client Matrix:	Solid	% Moisture: 7.6	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 0950		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		7.0		1.1
Barium		170		1.1
Cadmium		1.6		0.54
Chromium		37		1.1
Lead		55		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 1006		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.098		0.053

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-18A

Lab Sample ID:	720-381-72		Date Sampled:	11/09/2005 1505
Client Matrix:	Solid	% Moisture: 12.0	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 0954		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		9.0		1.1
Barium		180		1.1
Cadmium		2.2		0.56
Chromium		39		1.1
Lead		150		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/15/2005 1007		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.25		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-19B

Lab Sample ID:	720-381-73		Date Sampled:	11/09/2005 1430
Client Matrix:	Solid	% Moisture: 13.9	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 0958		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.0		1.1
Barium		170		1.1
Cadmium		1.6		0.56
Chromium		38		1.1
Lead		110		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 1008		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.15		0.056

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-19C

Lab Sample ID:	720-381-74		Date Sampled:	11/09/2005 1430
Client Matrix:	Solid	% Moisture: 12.0	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1762	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1731	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/15/2005 1002		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		3.8		1.1
Barium		130		1.1
Cadmium		1.1		0.56
Chromium		31		1.1
Lead		4.9		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1737	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 1009		Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1554			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.056

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-19A

Lab Sample ID:	720-381-76		Date Sampled:	11/09/2005 1530
Client Matrix:	Solid	% Moisture: 5.9	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 1024		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.3		1.1
Barium		140		1.1
Cadmium		0.99		0.53
Chromium		23		1.1
Lead		86		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/16/2005 0746		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.12		0.051

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-20B

Lab Sample ID:	720-381-77		Date Sampled:	11/09/2005 1438
Client Matrix:	Solid	% Moisture: 13.1	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1027		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		6.7		1.1
Barium		170		1.1
Cadmium		1.9		0.57
Chromium		56		1.1
Lead		57		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/16/2005 0747		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.13		0.056

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-20C

Lab Sample ID:	720-381-78		Date Sampled:	11/09/2005 1438
Client Matrix:	Solid	% Moisture: 6.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1030		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		4.9		1.1
Barium		96		1.1
Cadmium		0.97		0.53
Chromium		29		1.1
Lead		4.5		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/16/2005 0748		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.052

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-20A

Lab Sample ID:	720-381-80		Date Sampled:	11/09/2005 1535
Client Matrix:	Solid	% Moisture: 7.9	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1040		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.7		1.1
Barium		180		1.1
Cadmium		0.90		0.54
Chromium		26		1.1
Lead		120		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/16/2005 0752		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.23		0.053

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-21B

Lab Sample ID:	720-381-81		Date Sampled:	11/09/2005 1450
Client Matrix:	Solid	% Moisture: 12.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1683	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1636	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/11/2005 1311		Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		6.1		1.1
Barium		170		1.1
Cadmium		1.6		0.56
Chromium		37		1.1
Lead		27		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1760	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1666	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 0934		Final Weight/Volume:	
Date Prepared:	11/11/2005 1159			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.0035	*	0.0011

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-21C

Lab Sample ID:	720-381-82		Date Sampled:	11/09/2005 1450
Client Matrix:	Solid	% Moisture: 9.0	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/16/2005 1044		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.0		1.1
Barium		120		1.1
Cadmium		1.2		0.53
Chromium		33		1.1
Lead		6.0		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 0753		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.055

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-21A

Lab Sample ID:	720-381-84		Date Sampled:	11/09/2005 1540
Client Matrix:	Solid	% Moisture: 5.9	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 1047		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		9.4		1.1
Barium		180		1.1
Cadmium		0.98		0.53
Chromium		27		1.1
Lead		240		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/16/2005 0757		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.098		0.051

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-22B

Lab Sample ID:	720-381-85		Date Sampled:	11/09/2005 1510
Client Matrix:	Solid	% Moisture: 8.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 1050		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		7.3		1.1
Barium		200		1.1
Cadmium		1.9		0.55
Chromium		41		1.1
Lead		61		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/16/2005 0758		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.076		0.053

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-22C

Lab Sample ID:	720-381-86		Date Sampled:	11/09/2005 1510
Client Matrix:	Solid	% Moisture: 7.4	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/16/2005 1059		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.0
Arsenic		4.8		1.0
Barium		120		1.0
Cadmium		1.1		0.52
Chromium		31		1.0
Lead		7.1		1.0
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 0759		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.088		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-22A

Lab Sample ID:	720-381-88		Date Sampled:	11/09/2005 1545
Client Matrix:	Solid	% Moisture: 11.5	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 1103		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		11		1.1
Barium		240		1.1
Cadmium		3.7		0.56
Chromium		46		1.1
Lead		260		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 0800		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.30		0.056

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-23B

Lab Sample ID:	720-381-89		Date Sampled:	11/09/2005 1515
Client Matrix:	Solid	% Moisture: 11.5	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 1106		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		11		1.1
Barium		170		1.1
Cadmium		1.8		0.57
Chromium		47		1.1
Lead		140		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 0802		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.086		0.057

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-23C

Lab Sample ID:	720-381-90		Date Sampled:	11/09/2005 1520
Client Matrix:	Solid	% Moisture: 11.6	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 1109		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		6.5		1.1
Barium		130		1.1
Cadmium		1.3		0.57
Chromium		36		1.1
Lead		6.8		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 0803		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.057

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-23A

Lab Sample ID:	720-381-92		Date Sampled:	11/09/2005 1550
Client Matrix:	Solid	% Moisture: 4.5	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/16/2005 1112		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.0
Arsenic		6.4		1.0
Barium		260		1.0
Cadmium		0.85		0.50
Chromium		23		1.0
Lead		350		1.0
Selenium		ND		2.0

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 0804		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.43		0.052

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-24B

Lab Sample ID:	720-381-93		Date Sampled:	11/09/2005 1530
Client Matrix:	Solid	% Moisture: 15.6	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1116		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.2
Arsenic		15		1.2
Barium		210		1.2
Cadmium		2.0		0.59
Chromium		47		1.2
Lead		110		1.2
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 0805		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.13		0.059

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-24C

Lab Sample ID:	720-381-94		Date Sampled:	11/09/2005 1530
Client Matrix:	Solid	% Moisture: 15.9	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1119		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.2
Arsenic		16		1.2
Barium		190		1.2
Cadmium		1.8		0.59
Chromium		47		1.2
Lead		70		1.2
Selenium		ND		2.4

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/16/2005 0807		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.079		0.058

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-24A

Lab Sample ID:	720-381-96		Date Sampled:	11/09/2005 1615
Client Matrix:	Solid	% Moisture: 6.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/16/2005 1129		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.0
Arsenic		14		1.0
Barium		170		1.0
Cadmium		1.9		0.52
Chromium		44		1.0
Lead		64		1.0
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 0813		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.14		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-26B

Lab Sample ID:	720-381-97		Date Sampled:	11/09/2005 1540
Client Matrix:	Solid	% Moisture: 10.1	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1138		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		6.7		1.1
Barium		190		1.1
Cadmium		1.9		0.55
Chromium		45		1.1
Lead		180		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/16/2005 0814		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.097		0.053

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-26C

Lab Sample ID:	720-381-98		Date Sampled:	11/09/2005 1548
Client Matrix:	Solid	% Moisture: 12.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1142		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		7.7		1.1
Barium		170		1.1
Cadmium		1.5		0.57
Chromium		42		1.1
Lead		11		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 0815		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.057

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-26A

Lab Sample ID:	720-381-100		Date Sampled:	11/09/2005 1630
Client Matrix:	Solid	% Moisture: 7.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1773	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 1145		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		7.5		1.1
Barium		130		1.1
Cadmium		1.1		0.54
Chromium		29		1.1
Lead		160		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1778	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 0817		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1319			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.16		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-27B

Lab Sample ID:	720-381-101	Date Sampled:	11/09/2005 1600
Client Matrix:	Solid % Moisture: 10.5	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1158		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		9.7		1.1
Barium		180		1.1
Cadmium		2.3		0.55
Chromium		50		1.1
Lead		150		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/16/2005 0821		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.11		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-27C

Lab Sample ID:	720-381-102	Date Sampled:	11/09/2005 1605
Client Matrix:	Solid % Moisture: 12.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/16/2005 1201		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		8.1		1.1
Barium		140		1.1
Cadmium		1.5		0.55
Chromium		43		1.1
Lead		8.9		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 0822		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.057

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-27A

Lab Sample ID:	720-381-104	Date Sampled:	11/09/2005 1650
Client Matrix:	Solid % Moisture: 7.9	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850		Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1204			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.6		1.1
Barium		150		1.1
Cadmium		1.1		0.54
Chromium		22		1.1
Lead		100		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822		Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 0826			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.19		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-25B

Lab Sample ID:	720-381-105	Date Sampled:	11/09/2005 1620
Client Matrix:	Solid % Moisture: 10.3	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850		Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.02 g
Date Analyzed:	11/16/2005 1208			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		7.7		1.1
Barium		170		1.1
Cadmium		1.7		0.55
Chromium		43		1.1
Lead		100		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822		Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 0827			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.13		0.055

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-51A

Lab Sample ID:	720-381-106		Date Sampled:	11/09/2005 1240
Client Matrix:	Solid	% Moisture: 14.2	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1224		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.2
Arsenic		4.5		1.2
Barium		300		1.2
Cadmium		0.98		0.58
Chromium		17		1.2
Lead		100		1.2
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/16/2005 0831		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.18		0.056

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-51B

Lab Sample ID:	720-381-107	Date Sampled:	11/09/2005 1501
Client Matrix:	Solid % Moisture: 9.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/16/2005 1228		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		4.5		1.1
Arsenic		9.1		1.1
Barium		100		1.1
Cadmium		4.8		0.54
Chromium		37		1.1
Lead		87		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 0832		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.11		0.055

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-51C

Lab Sample ID:	720-381-108	Date Sampled:	11/09/2005 1501
Client Matrix:	Solid % Moisture: 14.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1231		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.2
Arsenic		4.6		1.2
Barium		130		1.2
Cadmium		1.3		0.58
Chromium		30		1.2
Lead		14		1.2
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 0833		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.058

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-52A

Lab Sample ID:	720-381-109	Date Sampled:	11/09/2005 1600
Client Matrix:	Solid % Moisture: 8.9	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/16/2005 1234		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		13		1.1
Barium		130		1.1
Cadmium		1.5		0.53
Chromium		23		1.1
Lead		420		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/16/2005 0835		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.13		0.053

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-52B

Lab Sample ID:	720-381-110	Date Sampled:	11/09/2005 1600
Client Matrix:	Solid % Moisture: 8.5	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850		Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.02 g
Date Analyzed:	11/16/2005 1238			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.4		1.1
Barium		130		1.1
Cadmium		2.4		0.54
Chromium		30		1.1
Lead		71		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822		Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.04 g
Date Analyzed:	11/16/2005 0836			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.11		0.053

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-52C

Lab Sample ID:	720-381-111	Date Sampled:	11/09/2005 1600
Client Matrix:	Solid % Moisture: 9.4	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850		Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.04 g
Date Analyzed:	11/16/2005 1241			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.5		1.1
Barium		120		1.1
Cadmium		1.1		0.53
Chromium		31		1.1
Lead		6.7		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822		Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 0837			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.055

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-53A

Lab Sample ID:	720-381-112	Date Sampled:	11/09/2005 1700
Client Matrix:	Solid % Moisture: 6.8	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850		Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.05 g
Date Analyzed:	11/16/2005 1244			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.0
Arsenic		6.8		1.0
Barium		300		1.0
Cadmium		0.92		0.51
Chromium		25		1.0
Lead		410		1.0
Selenium		ND		2.0

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822		Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 0916			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.41		0.053

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-53B

Lab Sample ID:	720-381-113	Date Sampled:	11/09/2005 1515
Client Matrix:	Solid % Moisture: 11.2	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850		Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.02 g
Date Analyzed:	11/16/2005 1247			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		12		1.1
Barium		140		1.1
Cadmium		1.6		0.55
Chromium		35		1.1
Lead		120		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822		Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 0918			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.11		0.056

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-54A

Lab Sample ID:	720-381-114	Date Sampled:	11/09/2005 1630
Client Matrix:	Solid % Moisture: 7.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850		Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.02 g
Date Analyzed:	11/16/2005 1321			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.6		1.1
Barium		190		1.1
Cadmium		1.1		0.53
Chromium		25		1.1
Lead		240		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822		Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 0919			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538				

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.13		0.054

Client: Weston Solutions,Inc

Job Number: 720-381-1

Client Sample ID: EBHH-54B

Lab Sample ID:	720-381-115	Date Sampled:	11/09/2005 1600
Client Matrix:	Solid % Moisture: 9.7	Date Received:	11/09/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1850	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1788	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1325		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		8.4		1.1
Barium		190		1.1
Cadmium		2.3		0.55
Chromium		48		1.1
Lead		160		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1822	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1803	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/16/2005 0920		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1538			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.13		0.053

General Chemistry

Client Sample ID: EBHH-1A

Lab Sample ID: 720-381-1
Client Matrix: Solid

Date Sampled: 11/09/2005 0900
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.5		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-1B

Lab Sample ID: 720-381-2
Client Matrix: Solid

Date Sampled: 11/09/2005 0925
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.8		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-1C

Lab Sample ID: 720-381-3
Client Matrix: Solid

Date Sampled: 11/09/2005 0940
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-2B

Lab Sample ID: 720-381-5
Client Matrix: Solid

Date Sampled: 11/09/2005 0945
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	15		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-2C

Lab Sample ID: 720-381-6
Client Matrix: Solid

Date Sampled: 11/09/2005 0951
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-2C

Lab Sample ID: 720-381-6
Client Matrix: Solid

Date Sampled: 11/09/2005 0951
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	15		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-2A

Lab Sample ID: 720-381-8
Client Matrix: Solid

Date Sampled: 11/09/2005 1102
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.7		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-3B

Lab Sample ID: 720-381-9
Client Matrix: Solid

Date Sampled: 11/09/2005 1000
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	8.3		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-3C

Lab Sample ID: 720-381-10
Client Matrix: Solid

Date Sampled: 11/09/2005 1000
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	16		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-3A

Lab Sample ID: 720-381-12
Client Matrix: Solid

Date Sampled: 11/09/2005 1105
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-3A

Lab Sample ID: 720-381-12
Client Matrix: Solid

Date Sampled: 11/09/2005 1105
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	10		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-4B

Lab Sample ID: 720-381-13
Client Matrix: Solid

Date Sampled: 11/09/2005 1015
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	13		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-4C

Lab Sample ID: 720-381-14
Client Matrix: Solid

Date Sampled: 11/09/2005 1020
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	13		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-4A

Lab Sample ID: 720-381-16
Client Matrix: Solid

Date Sampled: 11/09/2005 1110
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.5		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-5B

Lab Sample ID: 720-381-17
Client Matrix: Solid

Date Sampled: 11/09/2005 1030
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-5B

Lab Sample ID: 720-381-17
Client Matrix: Solid

Date Sampled: 11/09/2005 1030
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-5C

Lab Sample ID: 720-381-18
Client Matrix: Solid

Date Sampled: 11/09/2005 1036
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	16		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-5A

Lab Sample ID: 720-381-20
Client Matrix: Solid

Date Sampled: 11/09/2005 1115
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	10		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-6B

Lab Sample ID: 720-381-21
Client Matrix: Solid

Date Sampled: 11/09/2005 1045
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.4		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-6C

Lab Sample ID: 720-381-22
Client Matrix: Solid

Date Sampled: 11/09/2005 1045
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-6C

Lab Sample ID: 720-381-22
Client Matrix: Solid

Date Sampled: 11/09/2005 1045
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.1		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-6A

Lab Sample ID: 720-381-24
Client Matrix: Solid

Date Sampled: 11/09/2005 1118
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	13		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-7B

Lab Sample ID: 720-381-25
Client Matrix: Solid

Date Sampled: 11/09/2005 1125
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	15		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-7C

Lab Sample ID: 720-381-26
Client Matrix: Solid

Date Sampled: 11/09/2005 1125
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	14		%	0.010	1.0	160.3

Anly Batch: 720-1791 Date Analyzed 11/14/2005 0301

Client Sample ID: EBHH-7A

Lab Sample ID: 720-381-28
Client Matrix: Solid

Date Sampled: 11/09/2005 1145
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-7A

Lab Sample ID: 720-381-28
Client Matrix: Solid

Date Sampled: 11/09/2005 1145
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.0		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-8B

Lab Sample ID: 720-381-29
Client Matrix: Solid

Date Sampled: 11/09/2005 1135
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	16		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-8C

Lab Sample ID: 720-381-30
Client Matrix: Solid

Date Sampled: 11/09/2005 1135
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-8A

Lab Sample ID: 720-381-32
Client Matrix: Solid

Date Sampled: 11/09/2005 1150
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	10		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-9B

Lab Sample ID: 720-381-33
Client Matrix: Solid

Date Sampled: 11/09/2005 1146
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-9B

Lab Sample ID: 720-381-33
Client Matrix: Solid

Date Sampled: 11/09/2005 1146
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-9C

Lab Sample ID: 720-381-34
Client Matrix: Solid

Date Sampled: 11/09/2005 1146
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	13		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-9A

Lab Sample ID: 720-381-36
Client Matrix: Solid

Date Sampled: 11/09/2005 1155
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.2		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-10B

Lab Sample ID: 720-381-37
Client Matrix: Solid

Date Sampled: 11/09/2005 1203
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	10		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-10C

Lab Sample ID: 720-381-38
Client Matrix: Solid

Date Sampled: 11/09/2005 1203
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-10C

Lab Sample ID: 720-381-38
Client Matrix: Solid

Date Sampled: 11/09/2005 1203
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	8.9		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-10D

Lab Sample ID: 720-381-39
Client Matrix: Solid

Date Sampled: 11/09/2005 1203
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-10A

Lab Sample ID: 720-381-40
Client Matrix: Solid

Date Sampled: 11/09/2005 1340
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-11B

Lab Sample ID: 720-381-41
Client Matrix: Solid

Date Sampled: 11/09/2005 1208
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	10		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-11C

Lab Sample ID: 720-381-42
Client Matrix: Solid

Date Sampled: 11/09/2005 1215
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-11C

Lab Sample ID: 720-381-42
Client Matrix: Solid

Date Sampled: 11/09/2005 1215
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-11A

Lab Sample ID: 720-381-44
Client Matrix: Solid

Date Sampled: 11/09/2005 1243
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.3		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-12B

Lab Sample ID: 720-381-45
Client Matrix: Solid

Date Sampled: 11/09/2005 1223
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	13		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-12C

Lab Sample ID: 720-381-46
Client Matrix: Solid

Date Sampled: 11/09/2005 1223
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	11		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-12A

Lab Sample ID: 720-381-48
Client Matrix: Solid

Date Sampled: 11/09/2005 1248
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-12A

Lab Sample ID: 720-381-48
Client Matrix: Solid

Date Sampled: 11/09/2005 1248
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	6.7		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-13B

Lab Sample ID: 720-381-49
Client Matrix: Solid

Date Sampled: 11/09/2005 1250
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.8		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-13C

Lab Sample ID: 720-381-50
Client Matrix: Solid

Date Sampled: 11/09/2005 1252
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	10		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-13A

Lab Sample ID: 720-381-52
Client Matrix: Solid

Date Sampled: 11/09/2005 1350
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	15		%	0.010	1.0	160.3

Anly Batch: 720-1777 Date Analyzed 11/14/2005 1500

Client Sample ID: EBHH-14B

Lab Sample ID: 720-381-53
Client Matrix: Solid

Date Sampled: 11/09/2005 1301
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-14B

Lab Sample ID: 720-381-53
Client Matrix: Solid

Date Sampled: 11/09/2005 1301
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.5		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-14C

Lab Sample ID: 720-381-54
Client Matrix: Solid

Date Sampled: 11/09/2005 1309
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	6.4		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-14A

Lab Sample ID: 720-381-56
Client Matrix: Solid

Date Sampled: 11/09/2005 1445
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	Dil	Method
Percent Moisture	0.0000088		% by Wt	1.0	160.3

Anly Batch: 720-2301 Date Analyzed 11/28/2005 1540

Client Sample ID: EBHH-15B

Lab Sample ID: 720-381-57
Client Matrix: Solid

Date Sampled: 11/09/2005 1341
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	13		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-15C

Lab Sample ID: 720-381-58
Client Matrix: Solid

Date Sampled: 11/09/2005 1346
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-15C

Lab Sample ID: 720-381-58
Client Matrix: Solid

Date Sampled: 11/09/2005 1346
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-15A

Lab Sample ID: 720-381-60
Client Matrix: Solid

Date Sampled: 11/09/2005 1450
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	22		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-16B

Lab Sample ID: 720-381-61
Client Matrix: Solid

Date Sampled: 11/09/2005 1358
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	6.9		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-16C

Lab Sample ID: 720-381-62
Client Matrix: Solid

Date Sampled: 11/09/2005 1402
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	11		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-16A

Lab Sample ID: 720-381-64
Client Matrix: Solid

Date Sampled: 11/09/2005 1455
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-16A

Lab Sample ID: 720-381-64
Client Matrix: Solid

Date Sampled: 11/09/2005 1455
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-17B

Lab Sample ID: 720-381-65
Client Matrix: Solid

Date Sampled: 11/09/2005 1408
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.2		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-17C

Lab Sample ID: 720-381-66
Client Matrix: Solid

Date Sampled: 11/09/2005 1411
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	6.3		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-17A

Lab Sample ID: 720-381-68
Client Matrix: Solid

Date Sampled: 11/09/2005 1500
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	8.4		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-18B

Lab Sample ID: 720-381-69
Client Matrix: Solid

Date Sampled: 11/09/2005 1422
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-18B

Lab Sample ID: 720-381-69
Client Matrix: Solid

Date Sampled: 11/09/2005 1422
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	8.2		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-18C

Lab Sample ID: 720-381-70
Client Matrix: Solid

Date Sampled: 11/09/2005 1422
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.6		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-18D

Lab Sample ID: 720-381-71
Client Matrix: Solid

Date Sampled: 11/09/2005 1422
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	Dil	Method
Percent Moisture	12		%	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-18A

Lab Sample ID: 720-381-72
Client Matrix: Solid

Date Sampled: 11/09/2005 1505
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-19B

Lab Sample ID: 720-381-73
Client Matrix: Solid

Date Sampled: 11/09/2005 1430
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-19B

Lab Sample ID: 720-381-73
Client Matrix: Solid

Date Sampled: 11/09/2005 1430
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	14		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-19C

Lab Sample ID: 720-381-74
Client Matrix: Solid

Date Sampled: 11/09/2005 1430
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-19D

Lab Sample ID: 720-381-75
Client Matrix: Solid

Date Sampled: 11/09/2005 1430
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	15		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-19A

Lab Sample ID: 720-381-76
Client Matrix: Solid

Date Sampled: 11/09/2005 1530
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	5.9		%	0.010	1.0	160.3

Anly Batch: 720-1969 Date Analyzed 11/17/2005 1430

Client Sample ID: EBHH-20B

Lab Sample ID: 720-381-77
Client Matrix: Solid

Date Sampled: 11/09/2005 1438
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-20B

Lab Sample ID: 720-381-77
Client Matrix: Solid

Date Sampled: 11/09/2005 1438
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	14		%	0.010	1.0	160.3
	Anly Batch: 720-1969	Date Analyzed	11/17/2005 1430			
Percent Moisture	13		%	0.010	1.0	160.3
	Anly Batch: 720-1969	Date Analyzed	11/17/2005 1430			

Client Sample ID: EBHH-20C

Lab Sample ID: 720-381-78
Client Matrix: Solid

Date Sampled: 11/09/2005 1438
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	6.8		%	0.010	1.0	160.3
	Anly Batch: 720-1980	Date Analyzed	11/17/2005 1500			

Client Sample ID: EBHH-20A

Lab Sample ID: 720-381-80
Client Matrix: Solid

Date Sampled: 11/09/2005 1535
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.9		%	0.010	1.0	160.3
	Anly Batch: 720-1980	Date Analyzed	11/17/2005 1500			

Client Sample ID: EBHH-21B

Lab Sample ID: 720-381-81
Client Matrix: Solid

Date Sampled: 11/09/2005 1450
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	13		%	0.010	1.0	160.3
	Anly Batch: 720-1980	Date Analyzed	11/17/2005 1500			

General Chemistry

Client Sample ID: EBHH-21C

Lab Sample ID: 720-381-82
Client Matrix: Solid

Date Sampled: 11/09/2005 1450
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.0		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-21A

Lab Sample ID: 720-381-84
Client Matrix: Solid

Date Sampled: 11/09/2005 1540
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	5.9		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-22B

Lab Sample ID: 720-381-85
Client Matrix: Solid

Date Sampled: 11/09/2005 1510
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	8.7		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-22C

Lab Sample ID: 720-381-86
Client Matrix: Solid

Date Sampled: 11/09/2005 1510
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.4		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-22A

Lab Sample ID: 720-381-88
Client Matrix: Solid

Date Sampled: 11/09/2005 1545
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-22A

Lab Sample ID: 720-381-88
Client Matrix: Solid

Date Sampled: 11/09/2005 1545
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	11		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-23B

Lab Sample ID: 720-381-89
Client Matrix: Solid

Date Sampled: 11/09/2005 1515
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-23C

Lab Sample ID: 720-381-90
Client Matrix: Solid

Date Sampled: 11/09/2005 1520
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-23A

Lab Sample ID: 720-381-92
Client Matrix: Solid

Date Sampled: 11/09/2005 1550
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	4.5		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-24B

Lab Sample ID: 720-381-93
Client Matrix: Solid

Date Sampled: 11/09/2005 1530
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-24B

Lab Sample ID: 720-381-93
Client Matrix: Solid

Date Sampled: 11/09/2005 1530
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	16		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-24C

Lab Sample ID: 720-381-94
Client Matrix: Solid

Date Sampled: 11/09/2005 1530
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	16		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-24A

Lab Sample ID: 720-381-96
Client Matrix: Solid

Date Sampled: 11/09/2005 1615
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	6.7		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-26B

Lab Sample ID: 720-381-97
Client Matrix: Solid

Date Sampled: 11/09/2005 1540
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	10		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-26C

Lab Sample ID: 720-381-98
Client Matrix: Solid

Date Sampled: 11/09/2005 1548
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-26C

Lab Sample ID: 720-381-98
Client Matrix: Solid

Date Sampled: 11/09/2005 1548
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	13		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-26A

Lab Sample ID: 720-381-100
Client Matrix: Solid

Date Sampled: 11/09/2005 1630
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.7		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-27B

Lab Sample ID: 720-381-101
Client Matrix: Solid

Date Sampled: 11/09/2005 1600
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	10		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-27C

Lab Sample ID: 720-381-102
Client Matrix: Solid

Date Sampled: 11/09/2005 1605
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	13		%	0.010	1.0	160.3

Anly Batch: 720-1980 Date Analyzed 11/17/2005 1500

Client Sample ID: EBHH-27A

Lab Sample ID: 720-381-104
Client Matrix: Solid

Date Sampled: 11/09/2005 1650
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-27A

Lab Sample ID: 720-381-104
Client Matrix: Solid

Date Sampled: 11/09/2005 1650
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	8.9		%	0.010	1.0	160.3
	Anly Batch: 720-1980	Date Analyzed	11/17/2005 1500			
Percent Moisture	7.9		%	0.010	1.0	160.3
	Anly Batch: 720-1980	Date Analyzed	11/17/2005 1500			

Client Sample ID: EBHH-25B

Lab Sample ID: 720-381-105
Client Matrix: Solid

Date Sampled: 11/09/2005 1620
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	10		%	0.010	1.0	160.3
	Anly Batch: 720-1982	Date Analyzed	11/17/2005 1530			

Client Sample ID: EBHH-51A

Lab Sample ID: 720-381-106
Client Matrix: Solid

Date Sampled: 11/09/2005 1240
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	14		%	0.010	1.0	160.3
	Anly Batch: 720-1982	Date Analyzed	11/17/2005 1530			

Client Sample ID: EBHH-51B

Lab Sample ID: 720-381-107
Client Matrix: Solid

Date Sampled: 11/09/2005 1501
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.7		%	0.010	1.0	160.3
	Anly Batch: 720-1982	Date Analyzed	11/17/2005 1530			

General Chemistry

Client Sample ID: EBHH-51C

Lab Sample ID: 720-381-108
Client Matrix: Solid

Date Sampled: 11/09/2005 1501
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	15		%	0.010	1.0	160.3

Anly Batch: 720-1982 Date Analyzed 11/17/2005 1530

Client Sample ID: EBHH-52A

Lab Sample ID: 720-381-109
Client Matrix: Solid

Date Sampled: 11/09/2005 1600
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	8.9		%	0.010	1.0	160.3

Anly Batch: 720-1982 Date Analyzed 11/17/2005 1530

Client Sample ID: EBHH-52B

Lab Sample ID: 720-381-110
Client Matrix: Solid

Date Sampled: 11/09/2005 1600
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	8.5		%	0.010	1.0	160.3

Anly Batch: 720-1982 Date Analyzed 11/17/2005 1530

Client Sample ID: EBHH-52C

Lab Sample ID: 720-381-111
Client Matrix: Solid

Date Sampled: 11/09/2005 1600
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.4		%	0.010	1.0	160.3

Anly Batch: 720-1982 Date Analyzed 11/17/2005 1530

Client Sample ID: EBHH-53A

Lab Sample ID: 720-381-112
Client Matrix: Solid

Date Sampled: 11/09/2005 1700
Date Received: 11/09/2005 1730

General Chemistry

Client Sample ID: EBHH-53A

Lab Sample ID: 720-381-112
Client Matrix: Solid

Date Sampled: 11/09/2005 1700
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	6.8		%	0.010	1.0	160.3

Anly Batch: 720-1982 Date Analyzed 11/17/2005 1530

Client Sample ID: EBHH-53B

Lab Sample ID: 720-381-113
Client Matrix: Solid

Date Sampled: 11/09/2005 1515
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	11		%	0.010	1.0	160.3

Anly Batch: 720-1982 Date Analyzed 11/17/2005 1530

Client Sample ID: EBHH-54A

Lab Sample ID: 720-381-114
Client Matrix: Solid

Date Sampled: 11/09/2005 1630
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.7		%	0.010	1.0	160.3

Anly Batch: 720-1982 Date Analyzed 11/17/2005 1530

Client Sample ID: EBHH-54B

Lab Sample ID: 720-381-115
Client Matrix: Solid

Date Sampled: 11/09/2005 1600
Date Received: 11/09/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.4		%	0.010	1.0	160.3

Anly Batch: 720-1982 Date Analyzed 11/17/2005 1530

Percent Moisture	9.7		%	0.010	1.0	160.3

Anly Batch: 720-1982 Date Analyzed 11/17/2005 1530

DATA REPORTING QUALIFIERS

Client: Weston Solutions,Inc

Job Number: 720-381-1

Lab Section	Qualifier	Description
GC Semi VOA		
	*	LCS, LCSD, MS, MSD, MD, or Surrogate exceeds the control limits
	D	Surrogate or matrix spike recoveries were not obtained because the extract was diluted for analysis; also compounds analyzed at a dilution will be flagged with a D.
Metals		
	*	LCS, LCSD, MS, MSD, MD, or Surrogate exceeds the control limits
	N	MS, MSD: Spike recovery exceeds upper or lower control limits.

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC/MS Semi VOA				
Prep Batch: 720-1667				
LCS 720-1667/15-A	Lab Control Spike	Solid	3550B	
LCSD 720-1667/16-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1667/14-A	Method Blank	Solid	3550B	
720-381-1	EBHH-1A	Solid	3550B	
720-381-2	EBHH-1B	Solid	3550B	
720-381-14	EBHH-4C	Solid	3550B	
720-381-16	EBHH-4A	Solid	3550B	
720-381-39	EBHH-10D	Solid	3550B	
720-381-53	EBHH-14B	Solid	3550B	
720-381-56	EBHH-14A	Solid	3550B	
720-381-88	EBHH-22A	Solid	3550B	
720-381-98	EBHH-26C	Solid	3550B	
720-381-100	EBHH-26A	Solid	3550B	
720-381-105	EBHH-25B	Solid	3550B	
720-381-105MS	Matrix Spike	Solid	3550B	
720-381-105MSD	Matrix Spike Duplicate	Solid	3550B	
Prep Batch: 720-1711				
LCS 720-1711/2-A	Lab Control Spike	Solid	3550B	
LCSD 720-1711/3-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1711/1-A	Method Blank	Solid	3550B	
720-381-86	EBHH-22C	Solid	3550B	
720-416-A-1-B MS	Matrix Spike	Solid	3550B	
720-416-A-1-C MSD	Matrix Spike Duplicate	Solid	3550B	
Prep Batch: 720-1830				
LCS 720-1830/3-A	Lab Control Spike	Solid	3550B	
LCSD 720-1830/4-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1830/2-A	Method Blank	Solid	3550B	
720-381-71	EBHH-18D	Solid	3550B	
720-381-71MS	Matrix Spike	Solid	3550B	
720-381-71MSD	Matrix Spike Duplicate	Solid	3550B	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC/MS Semi VOA				
Analysis Batch:720-1821				
LCS 720-1667/15-A	Lab Control Spike	Solid	8270C	720-1667
LCSD 720-1667/16-A	Lab Control Spike Duplicate	Solid	8270C	720-1667
MB 720-1667/14-A	Method Blank	Solid	8270C	720-1667
720-381-1	EBHH-1A	Solid	8270C	720-1667
720-381-2	EBHH-1B	Solid	8270C	720-1667
720-381-14	EBHH-4C	Solid	8270C	720-1667
720-381-16	EBHH-4A	Solid	8270C	720-1667
720-381-39	EBHH-10D	Solid	8270C	720-1667
720-381-53	EBHH-14B	Solid	8270C	720-1667
720-381-56	EBHH-14A	Solid	8270C	720-1667
720-381-88	EBHH-22A	Solid	8270C	720-1667
720-381-98	EBHH-26C	Solid	8270C	720-1667
720-381-100	EBHH-26A	Solid	8270C	720-1667
720-381-105	EBHH-25B	Solid	8270C	720-1667
720-381-105MS	Matrix Spike	Solid	8270C	720-1667
720-381-105MSD	Matrix Spike Duplicate	Solid	8270C	720-1667
Analysis Batch:720-1849				
LCS 720-1711/2-A	Lab Control Spike	Solid	8270C	720-1711
LCSD 720-1711/3-A	Lab Control Spike Duplicate	Solid	8270C	720-1711
MB 720-1711/1-A	Method Blank	Solid	8270C	720-1711
Analysis Batch:720-1851				
720-381-86	EBHH-22C	Solid	8270C	720-1711
720-416-A-1-B MS	Matrix Spike	Solid	8270C	720-1711
720-416-A-1-C MSD	Matrix Spike Duplicate	Solid	8270C	720-1711
Analysis Batch:720-2227				
LCS 720-1830/3-A	Lab Control Spike	Solid	8270C	720-1830
LCSD 720-1830/4-A	Lab Control Spike Duplicate	Solid	8270C	720-1830
MB 720-1830/2-A	Method Blank	Solid	8270C	720-1830
720-381-71	EBHH-18D	Solid	8270C	720-1830
720-381-71MS	Matrix Spike	Solid	8270C	720-1830
720-381-71MSD	Matrix Spike Duplicate	Solid	8270C	720-1830

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC Semi VOA				
Prep Batch: 720-1624				
LCS 720-1624/3-A	Lab Control Spike	Solid	3550B	
LCSD 720-1624/4-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1624/2-A	Method Blank	Solid	3550B	
720-381-1	EBHH-1A	Solid	3550B	
720-381-2	EBHH-1B	Solid	3550B	
720-381-16	EBHH-4A	Solid	3550B	
720-381-53	EBHH-14B	Solid	3550B	
720-381-56	EBHH-14A	Solid	3550B	
720-381-86	EBHH-22C	Solid	3550B	
720-381-88	EBHH-22A	Solid	3550B	
720-381-98	EBHH-26C	Solid	3550B	
720-381-100	EBHH-26A	Solid	3550B	
720-381-105	EBHH-25B	Solid	3550B	
Prep Batch: 720-1635				
LCS 720-1635/22-B	Lab Control Spike	Solid	3550B	
LCSD 720-1635/23-B	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1635/21-B	Method Blank	Solid	3550B	
720-381-2	EBHH-1B	Solid	3550B	
720-381-3	EBHH-1C	Solid	3550B	
720-381-9	EBHH-3B	Solid	3550B	
720-381-10	EBHH-3C	Solid	3550B	
720-381-14	EBHH-4C	Solid	3550B	
720-381-17	EBHH-5B	Solid	3550B	
720-381-18	EBHH-5C	Solid	3550B	
720-381-25	EBHH-7B	Solid	3550B	
720-381-26	EBHH-7C	Solid	3550B	
720-381-33	EBHH-9B	Solid	3550B	
720-381-34	EBHH-9C	Solid	3550B	
720-381-41	EBHH-11B	Solid	3550B	
720-381-42	EBHH-11C	Solid	3550B	
720-381-42MS	Matrix Spike	Solid	3550B	
720-381-42MSD	Matrix Spike Duplicate	Solid	3550B	
720-381-49	EBHH-13B	Solid	3550B	
720-381-50	EBHH-13C	Solid	3550B	
720-381-57	EBHH-15B	Solid	3550B	
720-381-58	EBHH-15C	Solid	3550B	
720-381-65	EBHH-17B	Solid	3550B	
720-381-66	EBHH-17C	Solid	3550B	
720-381-73	EBHH-19B	Solid	3550B	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC Semi VOA				
Prep Batch: 720-1646				
LCS 720-1646/16-A	Lab Control Spike	Solid	3550B	
LCSD 720-1646/17-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1646/15-A	Method Blank	Solid	3550B	
720-381-74	EBHH-19C	Solid	3550B	
720-381-81	EBHH-21B	Solid	3550B	
720-381-82	EBHH-21C	Solid	3550B	
720-381-89	EBHH-23B	Solid	3550B	
720-381-90	EBHH-23C	Solid	3550B	
720-381-101	EBHH-27B	Solid	3550B	
720-381-102	EBHH-27C	Solid	3550B	
720-381-105	EBHH-25B	Solid	3550B	
720-381-105MS	Matrix Spike	Solid	3550B	
720-381-105MSD	Matrix Spike Duplicate	Solid	3550B	
720-381-107	EBHH-51B	Solid	3550B	
720-381-108	EBHH-51C	Solid	3550B	
720-381-110	EBHH-52B	Solid	3550B	
720-381-111	EBHH-52C	Solid	3550B	
720-381-113	EBHH-53B	Solid	3550B	
Prep Batch: 720-1758				
LCS 720-1758/2-A	Lab Control Spike	Solid	3550B	
LCSD 720-1758/3-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1758/1-A	Method Blank	Solid	3550B	
720-381-39	EBHH-10D	Solid	3550B	
720-381-39MS	Matrix Spike	Solid	3550B	
720-381-39MSD	Matrix Spike Duplicate	Solid	3550B	
Prep Batch: 720-1829				
LCS 720-1829/3-A	Lab Control Spike	Solid	3550B	
LCSD 720-1829/4-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1829/2-A	Method Blank	Solid	3550B	
720-381-71	EBHH-18D	Solid	3550B	
720-401-A-28-E MS	Matrix Spike	Solid	3550B	
720-401-A-28-F MSD	Matrix Spike Duplicate	Solid	3550B	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC Semi VOA				
Analysis Batch:720-1682				
LCS 720-1624/3-A	Lab Control Spike	Solid	8015B	720-1624
LCSD 720-1624/4-A	Lab Control Spike Duplicate	Solid	8015B	720-1624
MB 720-1624/2-A	Method Blank	Solid	8015B	720-1624
720-381-1	EBHH-1A	Solid	8015B	720-1624
720-381-2	EBHH-1B	Solid	8015B	720-1624
720-381-16	EBHH-4A	Solid	8015B	720-1624
720-381-53	EBHH-14B	Solid	8015B	720-1624
720-381-56	EBHH-14A	Solid	8015B	720-1624
720-381-86	EBHH-22C	Solid	8015B	720-1624
720-381-88	EBHH-22A	Solid	8015B	720-1624
720-381-98	EBHH-26C	Solid	8015B	720-1624
720-381-100	EBHH-26A	Solid	8015B	720-1624
720-381-105	EBHH-25B	Solid	8015B	720-1624
Analysis Batch:720-1801				
LCS 720-1635/22-B	Lab Control Spike	Solid	8081A	720-1635
LCSD 720-1635/23-B	Lab Control Spike Duplicate	Solid	8081A	720-1635
MB 720-1635/21-B	Method Blank	Solid	8081A	720-1635
720-381-2	EBHH-1B	Solid	8081A	720-1635
720-381-3	EBHH-1C	Solid	8081A	720-1635
720-381-9	EBHH-3B	Solid	8081A	720-1635
720-381-10	EBHH-3C	Solid	8081A	720-1635
720-381-14	EBHH-4C	Solid	8081A	720-1635
720-381-17	EBHH-5B	Solid	8081A	720-1635
720-381-18	EBHH-5C	Solid	8081A	720-1635
720-381-25	EBHH-7B	Solid	8081A	720-1635
720-381-26	EBHH-7C	Solid	8081A	720-1635
720-381-33	EBHH-9B	Solid	8081A	720-1635
720-381-34	EBHH-9C	Solid	8081A	720-1635
720-381-41	EBHH-11B	Solid	8081A	720-1635
720-381-42	EBHH-11C	Solid	8081A	720-1635
720-381-42MS	Matrix Spike	Solid	8081A	720-1635
720-381-42MSD	Matrix Spike Duplicate	Solid	8081A	720-1635
720-381-49	EBHH-13B	Solid	8081A	720-1635
720-381-50	EBHH-13C	Solid	8081A	720-1635
720-381-57	EBHH-15B	Solid	8081A	720-1635
720-381-58	EBHH-15C	Solid	8081A	720-1635
720-381-65	EBHH-17B	Solid	8081A	720-1635
720-381-66	EBHH-17C	Solid	8081A	720-1635
720-381-73	EBHH-19B	Solid	8081A	720-1635

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC Semi VOA				
Analysis Batch:720-1940				
LCS 720-1646/16-A	Lab Control Spike	Solid	8081A	720-1646
LCSD 720-1646/17-A	Lab Control Spike Duplicate	Solid	8081A	720-1646
MB 720-1646/15-A	Method Blank	Solid	8081A	720-1646
720-381-74	EBHH-19C	Solid	8081A	720-1646
720-381-81	EBHH-21B	Solid	8081A	720-1646
720-381-82	EBHH-21C	Solid	8081A	720-1646
720-381-89	EBHH-23B	Solid	8081A	720-1646
720-381-90	EBHH-23C	Solid	8081A	720-1646
720-381-101	EBHH-27B	Solid	8081A	720-1646
720-381-102	EBHH-27C	Solid	8081A	720-1646
720-381-105	EBHH-25B	Solid	8081A	720-1646
720-381-105MS	Matrix Spike	Solid	8081A	720-1646
720-381-105MSD	Matrix Spike Duplicate	Solid	8081A	720-1646
720-381-107	EBHH-51B	Solid	8081A	720-1646
720-381-108	EBHH-51C	Solid	8081A	720-1646
720-381-110	EBHH-52B	Solid	8081A	720-1646
720-381-111	EBHH-52C	Solid	8081A	720-1646
720-381-113	EBHH-53B	Solid	8081A	720-1646
Analysis Batch:720-1806				
LCS 720-1758/2-A	Lab Control Spike	Solid	8015B	720-1758
LCSD 720-1758/3-A	Lab Control Spike Duplicate	Solid	8015B	720-1758
MB 720-1758/1-A	Method Blank	Solid	8015B	720-1758
720-381-39	EBHH-10D	Solid	8015B	720-1758
720-381-39MS	Matrix Spike	Solid	8015B	720-1758
720-381-39MSD	Matrix Spike Duplicate	Solid	8015B	720-1758
Analysis Batch:720-1987				
LCS 720-1829/3-A	Lab Control Spike	Solid	8015B	720-1829
LCSD 720-1829/4-A	Lab Control Spike Duplicate	Solid	8015B	720-1829
MB 720-1829/2-A	Method Blank	Solid	8015B	720-1829
720-381-71	EBHH-18D	Solid	8015B	720-1829
720-401-A-28-E MS	Matrix Spike	Solid	8015B	720-1829
720-401-A-28-F MSD	Matrix Spike Duplicate	Solid	8015B	720-1829

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Prep Batch: 720-1636				
LCS 720-1636/2-A	Lab Control Spike	Solid	3050B	
LCSD 720-1636/3-A	Lab Control Spike Duplicate	Solid	3050B	
MB 720-1636/1-A	Method Blank	Solid	3050B	
720-381-1	EBHH-1A	Solid	3050B	
720-381-2	EBHH-1B	Solid	3050B	
720-381-3	EBHH-1C	Solid	3050B	
720-381-5	EBHH-2B	Solid	3050B	
720-381-6	EBHH-2C	Solid	3050B	
720-381-8	EBHH-2A	Solid	3050B	
720-381-9	EBHH-3B	Solid	3050B	
720-381-10	EBHH-3C	Solid	3050B	
720-381-12	EBHH-3A	Solid	3050B	
720-381-13	EBHH-4B	Solid	3050B	
720-381-14	EBHH-4C	Solid	3050B	
720-381-16	EBHH-4A	Solid	3050B	
720-381-17	EBHH-5B	Solid	3050B	
720-381-18	EBHH-5C	Solid	3050B	
720-381-20	EBHH-5A	Solid	3050B	
720-381-20MS	Matrix Spike	Solid	3050B	
720-381-20MSD	Matrix Spike Duplicate	Solid	3050B	
720-381-21	EBHH-6B	Solid	3050B	
720-381-22	EBHH-6C	Solid	3050B	
720-381-24	EBHH-6A	Solid	3050B	
720-381-81	EBHH-21B	Solid	3050B	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Prep Batch: 720-1666				
LCS 720-1666/2-A	Lab Control Spike	Solid	7471A	
LCSD 720-1666/3-A	Lab Control Spike Duplicate	Solid	7471A	
MB 720-1666/1-A	Method Blank	Solid	7471A	
720-381-1	EBHH-1A	Solid	7471A	
720-381-2	EBHH-1B	Solid	7471A	
720-381-3	EBHH-1C	Solid	7471A	
720-381-5	EBHH-2B	Solid	7471A	
720-381-6	EBHH-2C	Solid	7471A	
720-381-8	EBHH-2A	Solid	7471A	
720-381-9	EBHH-3B	Solid	7471A	
720-381-10	EBHH-3C	Solid	7471A	
720-381-12	EBHH-3A	Solid	7471A	
720-381-13	EBHH-4B	Solid	7471A	
720-381-14	EBHH-4C	Solid	7471A	
720-381-16	EBHH-4A	Solid	7471A	
720-381-17	EBHH-5B	Solid	7471A	
720-381-18	EBHH-5C	Solid	7471A	
720-381-20	EBHH-5A	Solid	7471A	
720-381-20MS	Matrix Spike	Solid	7471A	
720-381-20MSD	Matrix Spike Duplicate	Solid	7471A	
720-381-21	EBHH-6B	Solid	7471A	
720-381-22	EBHH-6C	Solid	7471A	
720-381-24	EBHH-6A	Solid	7471A	
720-381-81	EBHH-21B	Solid	7471A	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Prep Batch: 720-1731				
LCS 720-1731/2-A	Lab Control Spike	Solid	3050B	
LCSD 720-1731/3-A	Lab Control Spike Duplicate	Solid	3050B	
MB 720-1731/1-A	Method Blank	Solid	3050B	
720-381-49	EBHH-13B	Solid	3050B	
720-381-50	EBHH-13C	Solid	3050B	
720-381-52	EBHH-13A	Solid	3050B	
720-381-53	EBHH-14B	Solid	3050B	
720-381-54	EBHH-14C	Solid	3050B	
720-381-56	EBHH-14A	Solid	3050B	
720-381-57	EBHH-15B	Solid	3050B	
720-381-58	EBHH-15C	Solid	3050B	
720-381-58MS	Matrix Spike	Solid	3050B	
720-381-58MSD	Matrix Spike Duplicate	Solid	3050B	
720-381-60	EBHH-15A	Solid	3050B	
720-381-61	EBHH-16B	Solid	3050B	
720-381-62	EBHH-16C	Solid	3050B	
720-381-64	EBHH-16A	Solid	3050B	
720-381-65	EBHH-17B	Solid	3050B	
720-381-66	EBHH-17C	Solid	3050B	
720-381-68	EBHH-17A	Solid	3050B	
720-381-69	EBHH-18B	Solid	3050B	
720-381-70	EBHH-18C	Solid	3050B	
720-381-72	EBHH-18A	Solid	3050B	
720-381-73	EBHH-19B	Solid	3050B	
720-381-74	EBHH-19C	Solid	3050B	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Prep Batch: 720-1734				
LCS 720-1734/2-A	Lab Control Spike	Solid	3050B	
LCSD 720-1734/3-A	Lab Control Spike Duplicate	Solid	3050B	
MB 720-1734/1-A	Method Blank	Solid	3050B	
720-381-25	EBHH-7B	Solid	3050B	
720-381-26	EBHH-7C	Solid	3050B	
720-381-28	EBHH-7A	Solid	3050B	
720-381-29	EBHH-8B	Solid	3050B	
720-381-30	EBHH-8C	Solid	3050B	
720-381-32	EBHH-8A	Solid	3050B	
720-381-33	EBHH-9B	Solid	3050B	
720-381-34	EBHH-9C	Solid	3050B	
720-381-36	EBHH-9A	Solid	3050B	
720-381-37	EBHH-10B	Solid	3050B	
720-381-38	EBHH-10C	Solid	3050B	
720-381-40	EBHH-10A	Solid	3050B	
720-381-40MS	Matrix Spike	Solid	3050B	
720-381-40MSD	Matrix Spike Duplicate	Solid	3050B	
720-381-41	EBHH-11B	Solid	3050B	
720-381-A-42-B MSMS	Matrix Spike	Solid	3050B	
720-381-A-42-C MSDMSD	Matrix Spike Duplicate	Solid	3050B	
720-381-42	EBHH-11C	Solid	3050B	
720-381-42MS	Matrix Spike	Solid	3050B	
720-381-42MSD	Matrix Spike Duplicate	Solid	3050B	
720-381-44	EBHH-11A	Solid	3050B	
720-381-45	EBHH-12B	Solid	3050B	
720-381-46	EBHH-12C	Solid	3050B	
720-381-48	EBHH-12A	Solid	3050B	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Prep Batch: 720-1737				
LCS 720-1737/2-A	Lab Control Spike	Solid	7471A	
LCSD 720-1737/3-A	Lab Control Spike Duplicate	Solid	7471A	
MB 720-1737/1-A	Method Blank	Solid	7471A	
720-381-49	EBHH-13B	Solid	7471A	
720-381-50	EBHH-13C	Solid	7471A	
720-381-52	EBHH-13A	Solid	7471A	
720-381-53	EBHH-14B	Solid	7471A	
720-381-54	EBHH-14C	Solid	7471A	
720-381-56	EBHH-14A	Solid	7471A	
720-381-57	EBHH-15B	Solid	7471A	
720-381-58	EBHH-15C	Solid	7471A	
720-381-58MS	Matrix Spike	Solid	7471A	
720-381-58MSD	Matrix Spike Duplicate	Solid	7471A	
720-381-60	EBHH-15A	Solid	7471A	
720-381-61	EBHH-16B	Solid	7471A	
720-381-62	EBHH-16C	Solid	7471A	
720-381-64	EBHH-16A	Solid	7471A	
720-381-65	EBHH-17B	Solid	7471A	
720-381-66	EBHH-17C	Solid	7471A	
720-381-68	EBHH-17A	Solid	7471A	
720-381-69	EBHH-18B	Solid	7471A	
720-381-70	EBHH-18C	Solid	7471A	
720-381-72	EBHH-18A	Solid	7471A	
720-381-73	EBHH-19B	Solid	7471A	
720-381-74	EBHH-19C	Solid	7471A	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Prep Batch: 720-1743				
LCS 720-1743/2-A	Lab Control Spike	Solid	7471A	
LCSD 720-1743/3-A	Lab Control Spike Duplicate	Solid	7471A	
MB 720-1743/1-A	Method Blank	Solid	7471A	
720-381-25	EBHH-7B	Solid	7471A	
720-381-26	EBHH-7C	Solid	7471A	
720-381-28	EBHH-7A	Solid	7471A	
720-381-29	EBHH-8B	Solid	7471A	
720-381-30	EBHH-8C	Solid	7471A	
720-381-32	EBHH-8A	Solid	7471A	
720-381-33	EBHH-9B	Solid	7471A	
720-381-34	EBHH-9C	Solid	7471A	
720-381-36	EBHH-9A	Solid	7471A	
720-381-37	EBHH-10B	Solid	7471A	
720-381-38	EBHH-10C	Solid	7471A	
720-381-40	EBHH-10A	Solid	7471A	
720-381-40MS	Matrix Spike	Solid	7471A	
720-381-40MSD	Matrix Spike Duplicate	Solid	7471A	
720-381-41	EBHH-11B	Solid	7471A	
720-381-42	EBHH-11C	Solid	7471A	
720-381-42MS	Matrix Spike	Solid	7471A	
720-381-42MSD	Matrix Spike Duplicate	Solid	7471A	
720-381-44	EBHH-11A	Solid	7471A	
720-381-45	EBHH-12B	Solid	7471A	
720-381-46	EBHH-12C	Solid	7471A	
720-381-48	EBHH-12A	Solid	7471A	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Prep Batch: 720-1773				
LCS 720-1773/2-B	Lab Control Spike	Solid	3050B	
LCSD 720-1773/3-A	Lab Control Spike Duplicate	Solid	3050B	
MB 720-1773/1-A	Method Blank	Solid	3050B	
720-381-76	EBHH-19A	Solid	3050B	
720-381-77	EBHH-20B	Solid	3050B	
720-381-78	EBHH-20C	Solid	3050B	
720-381-78MS	Matrix Spike	Solid	3050B	
720-381-78MSD	Matrix Spike Duplicate	Solid	3050B	
720-381-80	EBHH-20A	Solid	3050B	
720-381-82	EBHH-21C	Solid	3050B	
720-381-84	EBHH-21A	Solid	3050B	
720-381-85	EBHH-22B	Solid	3050B	
720-381-86	EBHH-22C	Solid	3050B	
720-381-88	EBHH-22A	Solid	3050B	
720-381-89	EBHH-23B	Solid	3050B	
720-381-90	EBHH-23C	Solid	3050B	
720-381-92	EBHH-23A	Solid	3050B	
720-381-93	EBHH-24B	Solid	3050B	
720-381-94	EBHH-24C	Solid	3050B	
720-381-94MS	Matrix Spike	Solid	3050B	
720-381-94MSD	Matrix Spike Duplicate	Solid	3050B	
720-381-96	EBHH-24A	Solid	3050B	
720-381-97	EBHH-26B	Solid	3050B	
720-381-98	EBHH-26C	Solid	3050B	
720-381-100	EBHH-26A	Solid	3050B	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Prep Batch: 720-1778				
LCS 720-1778/2-A	Lab Control Spike	Solid	7471A	
LCSD 720-1778/3-A	Lab Control Spike Duplicate	Solid	7471A	
MB 720-1778/1-A	Method Blank	Solid	7471A	
720-381-76	EBHH-19A	Solid	7471A	
720-381-77	EBHH-20B	Solid	7471A	
720-381-78	EBHH-20C	Solid	7471A	
720-381-78MS	Matrix Spike	Solid	7471A	
720-381-78MSD	Matrix Spike Duplicate	Solid	7471A	
720-381-80	EBHH-20A	Solid	7471A	
720-381-82	EBHH-21C	Solid	7471A	
720-381-84	EBHH-21A	Solid	7471A	
720-381-85	EBHH-22B	Solid	7471A	
720-381-86	EBHH-22C	Solid	7471A	
720-381-88	EBHH-22A	Solid	7471A	
720-381-89	EBHH-23B	Solid	7471A	
720-381-90	EBHH-23C	Solid	7471A	
720-381-92	EBHH-23A	Solid	7471A	
720-381-93	EBHH-24B	Solid	7471A	
720-381-94	EBHH-24C	Solid	7471A	
720-381-94MS	Matrix Spike	Solid	7471A	
720-381-94MSD	Matrix Spike Duplicate	Solid	7471A	
720-381-96	EBHH-24A	Solid	7471A	
720-381-97	EBHH-26B	Solid	7471A	
720-381-98	EBHH-26C	Solid	7471A	
720-381-100	EBHH-26A	Solid	7471A	
Prep Batch: 720-1788				
LCS 720-1788/2-A	Lab Control Spike	Solid	3050B	
LCSD 720-1788/3-A	Lab Control Spike Duplicate	Solid	3050B	
MB 720-1788/1-A	Method Blank	Solid	3050B	
720-381-101	EBHH-27B	Solid	3050B	
720-381-102	EBHH-27C	Solid	3050B	
720-381-104	EBHH-27A	Solid	3050B	
720-381-105	EBHH-25B	Solid	3050B	
720-381-105MS	Matrix Spike	Solid	3050B	
720-381-105MSD	Matrix Spike Duplicate	Solid	3050B	
720-381-106	EBHH-51A	Solid	3050B	
720-381-107	EBHH-51B	Solid	3050B	
720-381-108	EBHH-51C	Solid	3050B	
720-381-109	EBHH-52A	Solid	3050B	
720-381-110	EBHH-52B	Solid	3050B	
720-381-111	EBHH-52C	Solid	3050B	
720-381-112	EBHH-53A	Solid	3050B	
720-381-113	EBHH-53B	Solid	3050B	
720-381-114	EBHH-54A	Solid	3050B	
720-381-115	EBHH-54B	Solid	3050B	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Prep Batch: 720-1803				
LCS 720-1803/2-A	Lab Control Spike	Solid	7471A	
LCSD 720-1803/3-A	Lab Control Spike Duplicate	Solid	7471A	
MB 720-1803/1-A	Method Blank	Solid	7471A	
720-381-101	EBHH-27B	Solid	7471A	
720-381-102	EBHH-27C	Solid	7471A	
720-381-104	EBHH-27A	Solid	7471A	
720-381-105	EBHH-25B	Solid	7471A	
720-381-105MS	Matrix Spike	Solid	7471A	
720-381-105MSD	Matrix Spike Duplicate	Solid	7471A	
720-381-106	EBHH-51A	Solid	7471A	
720-381-107	EBHH-51B	Solid	7471A	
720-381-108	EBHH-51C	Solid	7471A	
720-381-109	EBHH-52A	Solid	7471A	
720-381-110	EBHH-52B	Solid	7471A	
720-381-111	EBHH-52C	Solid	7471A	
720-381-112	EBHH-53A	Solid	7471A	
720-381-113	EBHH-53B	Solid	7471A	
720-381-114	EBHH-54A	Solid	7471A	
720-381-115	EBHH-54B	Solid	7471A	
Analysis Batch:720-1683				
LCS 720-1636/2-A	Lab Control Spike	Solid	6010B	720-1636
LCSD 720-1636/3-A	Lab Control Spike Duplicate	Solid	6010B	720-1636
MB 720-1636/1-A	Method Blank	Solid	6010B	720-1636
720-381-1	EBHH-1A	Solid	6010B	720-1636
720-381-2	EBHH-1B	Solid	6010B	720-1636
720-381-3	EBHH-1C	Solid	6010B	720-1636
720-381-5	EBHH-2B	Solid	6010B	720-1636
720-381-6	EBHH-2C	Solid	6010B	720-1636
720-381-8	EBHH-2A	Solid	6010B	720-1636
720-381-9	EBHH-3B	Solid	6010B	720-1636
720-381-10	EBHH-3C	Solid	6010B	720-1636
720-381-12	EBHH-3A	Solid	6010B	720-1636
720-381-13	EBHH-4B	Solid	6010B	720-1636
720-381-14	EBHH-4C	Solid	6010B	720-1636
720-381-16	EBHH-4A	Solid	6010B	720-1636
720-381-17	EBHH-5B	Solid	6010B	720-1636
720-381-18	EBHH-5C	Solid	6010B	720-1636
720-381-20	EBHH-5A	Solid	6010B	720-1636
720-381-20MS	Matrix Spike	Solid	6010B	720-1636
720-381-20MSD	Matrix Spike Duplicate	Solid	6010B	720-1636
720-381-21	EBHH-6B	Solid	6010B	720-1636
720-381-22	EBHH-6C	Solid	6010B	720-1636
720-381-24	EBHH-6A	Solid	6010B	720-1636
720-381-81	EBHH-21B	Solid	6010B	720-1636

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Analysis Batch:720-1760				
LCS 720-1666/2-A	Lab Control Spike	Solid	7471A	720-1666
LCSD 720-1666/3-A	Lab Control Spike Duplicate	Solid	7471A	720-1666
MB 720-1666/1-A	Method Blank	Solid	7471A	720-1666
720-381-1	EBHH-1A	Solid	7471A	720-1666
720-381-2	EBHH-1B	Solid	7471A	720-1666
720-381-3	EBHH-1C	Solid	7471A	720-1666
720-381-5	EBHH-2B	Solid	7471A	720-1666
720-381-6	EBHH-2C	Solid	7471A	720-1666
720-381-8	EBHH-2A	Solid	7471A	720-1666
720-381-9	EBHH-3B	Solid	7471A	720-1666
720-381-10	EBHH-3C	Solid	7471A	720-1666
720-381-12	EBHH-3A	Solid	7471A	720-1666
720-381-13	EBHH-4B	Solid	7471A	720-1666
720-381-14	EBHH-4C	Solid	7471A	720-1666
720-381-16	EBHH-4A	Solid	7471A	720-1666
720-381-17	EBHH-5B	Solid	7471A	720-1666
720-381-18	EBHH-5C	Solid	7471A	720-1666
720-381-20	EBHH-5A	Solid	7471A	720-1666
720-381-20MS	Matrix Spike	Solid	7471A	720-1666
720-381-20MSD	Matrix Spike Duplicate	Solid	7471A	720-1666
720-381-21	EBHH-6B	Solid	7471A	720-1666
720-381-22	EBHH-6C	Solid	7471A	720-1666
720-381-24	EBHH-6A	Solid	7471A	720-1666
720-381-81	EBHH-21B	Solid	7471A	720-1666

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Analysis Batch:720-1762				
LCS 720-1731/2-A	Lab Control Spike	Solid	6010B	720-1731
LCSD 720-1731/3-A	Lab Control Spike Duplicate	Solid	6010B	720-1731
MB 720-1731/1-A	Method Blank	Solid	6010B	720-1731
720-381-49	EBHH-13B	Solid	6010B	720-1731
720-381-50	EBHH-13C	Solid	6010B	720-1731
720-381-52	EBHH-13A	Solid	6010B	720-1731
720-381-53	EBHH-14B	Solid	6010B	720-1731
720-381-54	EBHH-14C	Solid	6010B	720-1731
720-381-56	EBHH-14A	Solid	6010B	720-1731
720-381-57	EBHH-15B	Solid	6010B	720-1731
720-381-58	EBHH-15C	Solid	6010B	720-1731
720-381-58MS	Matrix Spike	Solid	6010B	720-1731
720-381-58MSD	Matrix Spike Duplicate	Solid	6010B	720-1731
720-381-60	EBHH-15A	Solid	6010B	720-1731
720-381-61	EBHH-16B	Solid	6010B	720-1731
720-381-62	EBHH-16C	Solid	6010B	720-1731
720-381-64	EBHH-16A	Solid	6010B	720-1731
720-381-65	EBHH-17B	Solid	6010B	720-1731
720-381-66	EBHH-17C	Solid	6010B	720-1731
720-381-68	EBHH-17A	Solid	6010B	720-1731
720-381-69	EBHH-18B	Solid	6010B	720-1731
720-381-70	EBHH-18C	Solid	6010B	720-1731
720-381-72	EBHH-18A	Solid	6010B	720-1731
720-381-73	EBHH-19B	Solid	6010B	720-1731
720-381-74	EBHH-19C	Solid	6010B	720-1731
Analysis Batch:720-1762				
LCS 720-1734/2-A	Lab Control Spike	Solid	6010B	720-1734
LCSD 720-1734/3-A	Lab Control Spike Duplicate	Solid	6010B	720-1734
MB 720-1734/1-A	Method Blank	Solid	6010B	720-1734
720-381-25	EBHH-7B	Solid	6010B	720-1734
720-381-26	EBHH-7C	Solid	6010B	720-1734
720-381-28	EBHH-7A	Solid	6010B	720-1734
720-381-29	EBHH-8B	Solid	6010B	720-1734
720-381-30	EBHH-8C	Solid	6010B	720-1734
720-381-32	EBHH-8A	Solid	6010B	720-1734
720-381-33	EBHH-9B	Solid	6010B	720-1734
720-381-34	EBHH-9C	Solid	6010B	720-1734
720-381-36	EBHH-9A	Solid	6010B	720-1734
720-381-37	EBHH-10B	Solid	6010B	720-1734
720-381-38	EBHH-10C	Solid	6010B	720-1734
720-381-40	EBHH-10A	Solid	6010B	720-1734
720-381-40MS	Matrix Spike	Solid	6010B	720-1734
720-381-40MSD	Matrix Spike Duplicate	Solid	6010B	720-1734
720-381-A-42-B MSMS	Matrix Spike	Solid	6010B	720-1734
720-381-A-42-C MSDMSD	Matrix Spike Duplicate	Solid	6010B	720-1734

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Analysis Batch:720-1808				
LCS 720-1734/2-A	Lab Control Spike	Solid	6010B	720-1734
LCSD 720-1734/3-A	Lab Control Spike Duplicate	Solid	6010B	720-1734
MB 720-1734/1-A	Method Blank	Solid	6010B	720-1734
720-381-25	EBHH-7B	Solid	6010B	720-1734
720-381-26	EBHH-7C	Solid	6010B	720-1734
720-381-28	EBHH-7A	Solid	6010B	720-1734
720-381-29	EBHH-8B	Solid	6010B	720-1734
720-381-30	EBHH-8C	Solid	6010B	720-1734
720-381-32	EBHH-8A	Solid	6010B	720-1734
720-381-33	EBHH-9B	Solid	6010B	720-1734
720-381-34	EBHH-9C	Solid	6010B	720-1734
720-381-36	EBHH-9A	Solid	6010B	720-1734
720-381-37	EBHH-10B	Solid	6010B	720-1734
720-381-38	EBHH-10C	Solid	6010B	720-1734
720-381-40	EBHH-10A	Solid	6010B	720-1734
720-381-40MS	Matrix Spike	Solid	6010B	720-1734
720-381-40MSD	Matrix Spike Duplicate	Solid	6010B	720-1734
720-381-41	EBHH-11B	Solid	6010B	720-1734
Analysis Batch:720-1810				
720-381-42	EBHH-11C	Solid	6010B	720-1734
720-381-42MS	Matrix Spike	Solid	6010B	720-1734
720-381-42MSD	Matrix Spike Duplicate	Solid	6010B	720-1734
720-381-44	EBHH-11A	Solid	6010B	720-1734
720-381-45	EBHH-12B	Solid	6010B	720-1734
720-381-46	EBHH-12C	Solid	6010B	720-1734
Analysis Batch:720-1951				
720-381-48	EBHH-12A	Solid	6010B	720-1734

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Analysis Batch:720-1760				
LCS 720-1737/2-A	Lab Control Spike	Solid	7471A	720-1737
LCSD 720-1737/3-A	Lab Control Spike Duplicate	Solid	7471A	720-1737
MB 720-1737/1-A	Method Blank	Solid	7471A	720-1737
720-381-49	EBHH-13B	Solid	7471A	720-1737
720-381-50	EBHH-13C	Solid	7471A	720-1737
720-381-52	EBHH-13A	Solid	7471A	720-1737
720-381-53	EBHH-14B	Solid	7471A	720-1737
720-381-54	EBHH-14C	Solid	7471A	720-1737
720-381-56	EBHH-14A	Solid	7471A	720-1737
720-381-57	EBHH-15B	Solid	7471A	720-1737
720-381-58	EBHH-15C	Solid	7471A	720-1737
720-381-58MS	Matrix Spike	Solid	7471A	720-1737
720-381-58MSD	Matrix Spike Duplicate	Solid	7471A	720-1737
720-381-60	EBHH-15A	Solid	7471A	720-1737
720-381-61	EBHH-16B	Solid	7471A	720-1737
720-381-62	EBHH-16C	Solid	7471A	720-1737
720-381-64	EBHH-16A	Solid	7471A	720-1737
720-381-65	EBHH-17B	Solid	7471A	720-1737
720-381-66	EBHH-17C	Solid	7471A	720-1737
720-381-68	EBHH-17A	Solid	7471A	720-1737
720-381-69	EBHH-18B	Solid	7471A	720-1737
720-381-70	EBHH-18C	Solid	7471A	720-1737
720-381-72	EBHH-18A	Solid	7471A	720-1737
720-381-73	EBHH-19B	Solid	7471A	720-1737
720-381-74	EBHH-19C	Solid	7471A	720-1737

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Analysis Batch:720-1807				
LCS 720-1743/2-A	Lab Control Spike	Solid	7471A	720-1743
LCSD 720-1743/3-A	Lab Control Spike Duplicate	Solid	7471A	720-1743
MB 720-1743/1-A	Method Blank	Solid	7471A	720-1743
720-381-25	EBHH-7B	Solid	7471A	720-1743
720-381-26	EBHH-7C	Solid	7471A	720-1743
720-381-28	EBHH-7A	Solid	7471A	720-1743
720-381-29	EBHH-8B	Solid	7471A	720-1743
720-381-30	EBHH-8C	Solid	7471A	720-1743
720-381-32	EBHH-8A	Solid	7471A	720-1743
720-381-33	EBHH-9B	Solid	7471A	720-1743
720-381-34	EBHH-9C	Solid	7471A	720-1743
720-381-36	EBHH-9A	Solid	7471A	720-1743
720-381-37	EBHH-10B	Solid	7471A	720-1743
720-381-38	EBHH-10C	Solid	7471A	720-1743
720-381-40	EBHH-10A	Solid	7471A	720-1743
720-381-40MS	Matrix Spike	Solid	7471A	720-1743
720-381-40MSD	Matrix Spike Duplicate	Solid	7471A	720-1743
720-381-41	EBHH-11B	Solid	7471A	720-1743
720-381-42	EBHH-11C	Solid	7471A	720-1743
720-381-42MS	Matrix Spike	Solid	7471A	720-1743
720-381-42MSD	Matrix Spike Duplicate	Solid	7471A	720-1743
720-381-44	EBHH-11A	Solid	7471A	720-1743
720-381-45	EBHH-12B	Solid	7471A	720-1743
720-381-46	EBHH-12C	Solid	7471A	720-1743
720-381-48	EBHH-12A	Solid	7471A	720-1743

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Analysis Batch:720-1850				
LCS 720-1773/2-B	Lab Control Spike	Solid	6010B	720-1773
LCSD 720-1773/3-A	Lab Control Spike Duplicate	Solid	6010B	720-1773
MB 720-1773/1-A	Method Blank	Solid	6010B	720-1773
720-381-76	EBHH-19A	Solid	6010B	720-1773
720-381-77	EBHH-20B	Solid	6010B	720-1773
720-381-78	EBHH-20C	Solid	6010B	720-1773
720-381-78MS	Matrix Spike	Solid	6010B	720-1773
720-381-78MSD	Matrix Spike Duplicate	Solid	6010B	720-1773
720-381-80	EBHH-20A	Solid	6010B	720-1773
720-381-82	EBHH-21C	Solid	6010B	720-1773
720-381-84	EBHH-21A	Solid	6010B	720-1773
720-381-85	EBHH-22B	Solid	6010B	720-1773
720-381-86	EBHH-22C	Solid	6010B	720-1773
720-381-88	EBHH-22A	Solid	6010B	720-1773
720-381-89	EBHH-23B	Solid	6010B	720-1773
720-381-90	EBHH-23C	Solid	6010B	720-1773
720-381-92	EBHH-23A	Solid	6010B	720-1773
720-381-93	EBHH-24B	Solid	6010B	720-1773
720-381-94	EBHH-24C	Solid	6010B	720-1773
720-381-94MS	Matrix Spike	Solid	6010B	720-1773
720-381-94MSD	Matrix Spike Duplicate	Solid	6010B	720-1773
720-381-96	EBHH-24A	Solid	6010B	720-1773
720-381-97	EBHH-26B	Solid	6010B	720-1773
720-381-98	EBHH-26C	Solid	6010B	720-1773
720-381-100	EBHH-26A	Solid	6010B	720-1773

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Analysis Batch:720-1822				
LCS 720-1778/2-A	Lab Control Spike	Solid	7471A	720-1778
LCSD 720-1778/3-A	Lab Control Spike Duplicate	Solid	7471A	720-1778
MB 720-1778/1-A	Method Blank	Solid	7471A	720-1778
720-381-76	EBHH-19A	Solid	7471A	720-1778
720-381-77	EBHH-20B	Solid	7471A	720-1778
720-381-78	EBHH-20C	Solid	7471A	720-1778
720-381-78MS	Matrix Spike	Solid	7471A	720-1778
720-381-78MSD	Matrix Spike Duplicate	Solid	7471A	720-1778
720-381-80	EBHH-20A	Solid	7471A	720-1778
720-381-82	EBHH-21C	Solid	7471A	720-1778
720-381-84	EBHH-21A	Solid	7471A	720-1778
720-381-85	EBHH-22B	Solid	7471A	720-1778
720-381-86	EBHH-22C	Solid	7471A	720-1778
720-381-88	EBHH-22A	Solid	7471A	720-1778
720-381-89	EBHH-23B	Solid	7471A	720-1778
720-381-90	EBHH-23C	Solid	7471A	720-1778
720-381-92	EBHH-23A	Solid	7471A	720-1778
720-381-93	EBHH-24B	Solid	7471A	720-1778
720-381-94	EBHH-24C	Solid	7471A	720-1778
720-381-94MS	Matrix Spike	Solid	7471A	720-1778
720-381-94MSD	Matrix Spike Duplicate	Solid	7471A	720-1778
720-381-96	EBHH-24A	Solid	7471A	720-1778
720-381-97	EBHH-26B	Solid	7471A	720-1778
720-381-98	EBHH-26C	Solid	7471A	720-1778
720-381-100	EBHH-26A	Solid	7471A	720-1778
Analysis Batch:720-1850				
LCS 720-1788/2-A	Lab Control Spike	Solid	6010B	720-1788
LCSD 720-1788/3-A	Lab Control Spike Duplicate	Solid	6010B	720-1788
MB 720-1788/1-A	Method Blank	Solid	6010B	720-1788
720-381-101	EBHH-27B	Solid	6010B	720-1788
720-381-102	EBHH-27C	Solid	6010B	720-1788
720-381-104	EBHH-27A	Solid	6010B	720-1788
720-381-105	EBHH-25B	Solid	6010B	720-1788
720-381-105MS	Matrix Spike	Solid	6010B	720-1788
720-381-105MSD	Matrix Spike Duplicate	Solid	6010B	720-1788
720-381-106	EBHH-51A	Solid	6010B	720-1788
720-381-107	EBHH-51B	Solid	6010B	720-1788
720-381-108	EBHH-51C	Solid	6010B	720-1788
720-381-109	EBHH-52A	Solid	6010B	720-1788
720-381-110	EBHH-52B	Solid	6010B	720-1788
720-381-111	EBHH-52C	Solid	6010B	720-1788
720-381-112	EBHH-53A	Solid	6010B	720-1788
720-381-113	EBHH-53B	Solid	6010B	720-1788
720-381-114	EBHH-54A	Solid	6010B	720-1788
720-381-115	EBHH-54B	Solid	6010B	720-1788

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Analysis Batch:720-1822				
LCS 720-1803/2-A	Lab Control Spike	Solid	7471A	720-1803
LCSD 720-1803/3-A	Lab Control Spike Duplicate	Solid	7471A	720-1803
MB 720-1803/1-A	Method Blank	Solid	7471A	720-1803
720-381-101	EBHH-27B	Solid	7471A	720-1803
720-381-102	EBHH-27C	Solid	7471A	720-1803
720-381-104	EBHH-27A	Solid	7471A	720-1803
720-381-105	EBHH-25B	Solid	7471A	720-1803
720-381-105MS	Matrix Spike	Solid	7471A	720-1803
720-381-105MSD	Matrix Spike Duplicate	Solid	7471A	720-1803
720-381-106	EBHH-51A	Solid	7471A	720-1803
720-381-107	EBHH-51B	Solid	7471A	720-1803
720-381-108	EBHH-51C	Solid	7471A	720-1803
720-381-109	EBHH-52A	Solid	7471A	720-1803
720-381-110	EBHH-52B	Solid	7471A	720-1803
720-381-111	EBHH-52C	Solid	7471A	720-1803
720-381-112	EBHH-53A	Solid	7471A	720-1803
720-381-113	EBHH-53B	Solid	7471A	720-1803
720-381-114	EBHH-54A	Solid	7471A	720-1803
720-381-115	EBHH-54B	Solid	7471A	720-1803

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
General Chemistry				
Analysis Batch:720-1777				
MB 720-1777/1	Method Blank	Solid	160.3	
720-381-28	EBHH-7A	Solid	160.3	
720-381-29	EBHH-8B	Solid	160.3	
720-381-30	EBHH-8C	Solid	160.3	
720-381-32	EBHH-8A	Solid	160.3	
720-381-33	EBHH-9B	Solid	160.3	
720-381-34	EBHH-9C	Solid	160.3	
720-381-36	EBHH-9A	Solid	160.3	
720-381-37	EBHH-10B	Solid	160.3	
720-381-38	EBHH-10C	Solid	160.3	
720-381-39	EBHH-10D	Solid	160.3	
720-381-40	EBHH-10A	Solid	160.3	
720-381-41	EBHH-11B	Solid	160.3	
720-381-42	EBHH-11C	Solid	160.3	
720-381-44	EBHH-11A	Solid	160.3	
720-381-45	EBHH-12B	Solid	160.3	
720-381-46	EBHH-12C	Solid	160.3	
720-381-48	EBHH-12A	Solid	160.3	
720-381-49	EBHH-13B	Solid	160.3	
720-381-50	EBHH-13C	Solid	160.3	
720-381-52	EBHH-13A	Solid	160.3	
720-381-52DU	Duplicate	Solid	160.3	
Analysis Batch:720-1791				
720-381-1	EBHH-1A	Solid	160.3	
720-381-2	EBHH-1B	Solid	160.3	
720-381-3	EBHH-1C	Solid	160.3	
720-381-5	EBHH-2B	Solid	160.3	
720-381-6	EBHH-2C	Solid	160.3	
720-381-8	EBHH-2A	Solid	160.3	
720-381-9	EBHH-3B	Solid	160.3	
720-381-10	EBHH-3C	Solid	160.3	
720-381-12	EBHH-3A	Solid	160.3	
720-381-13	EBHH-4B	Solid	160.3	
720-381-14	EBHH-4C	Solid	160.3	
720-381-16	EBHH-4A	Solid	160.3	
720-381-17	EBHH-5B	Solid	160.3	
720-381-18	EBHH-5C	Solid	160.3	
720-381-20	EBHH-5A	Solid	160.3	
720-381-21	EBHH-6B	Solid	160.3	
720-381-22	EBHH-6C	Solid	160.3	
720-381-24	EBHH-6A	Solid	160.3	
720-381-25	EBHH-7B	Solid	160.3	
720-381-26	EBHH-7C	Solid	160.3	
720-381-26DU	Duplicate	Solid	160.3	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
General Chemistry				
Analysis Batch:720-1969				
MB 720-1969/1	Method Blank	Solid	160.3	
720-381-53	EBHH-14B	Solid	160.3	
720-381-54	EBHH-14C	Solid	160.3	
720-381-57	EBHH-15B	Solid	160.3	
720-381-58	EBHH-15C	Solid	160.3	
720-381-60	EBHH-15A	Solid	160.3	
720-381-61	EBHH-16B	Solid	160.3	
720-381-62	EBHH-16C	Solid	160.3	
720-381-64	EBHH-16A	Solid	160.3	
720-381-65	EBHH-17B	Solid	160.3	
720-381-66	EBHH-17C	Solid	160.3	
720-381-68	EBHH-17A	Solid	160.3	
720-381-69	EBHH-18B	Solid	160.3	
720-381-70	EBHH-18C	Solid	160.3	
720-381-71	EBHH-18D	Solid	160.3	
720-381-72	EBHH-18A	Solid	160.3	
720-381-73	EBHH-19B	Solid	160.3	
720-381-74	EBHH-19C	Solid	160.3	
720-381-75	EBHH-19D	Solid	160.3	
720-381-76	EBHH-19A	Solid	160.3	
720-381-77	EBHH-20B	Solid	160.3	
Analysis Batch:720-1980				
MB 720-1980/1	Method Blank	Solid	160.3	
720-381-78	EBHH-20C	Solid	160.3	
720-381-80	EBHH-20A	Solid	160.3	
720-381-81	EBHH-21B	Solid	160.3	
720-381-82	EBHH-21C	Solid	160.3	
720-381-84	EBHH-21A	Solid	160.3	
720-381-85	EBHH-22B	Solid	160.3	
720-381-86	EBHH-22C	Solid	160.3	
720-381-88	EBHH-22A	Solid	160.3	
720-381-89	EBHH-23B	Solid	160.3	
720-381-90	EBHH-23C	Solid	160.3	
720-381-92	EBHH-23A	Solid	160.3	
720-381-93	EBHH-24B	Solid	160.3	
720-381-94	EBHH-24C	Solid	160.3	
720-381-96	EBHH-24A	Solid	160.3	
720-381-97	EBHH-26B	Solid	160.3	
720-381-98	EBHH-26C	Solid	160.3	
720-381-100	EBHH-26A	Solid	160.3	
720-381-101	EBHH-27B	Solid	160.3	
720-381-102	EBHH-27C	Solid	160.3	
720-381-104	EBHH-27A	Solid	160.3	

Client: Weston Solutions,Inc

Job Number: 720-381-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
General Chemistry				
Analysis Batch:720-1982				
MB 720-1982/1	Method Blank	Solid	160.3	
720-381-105	EBHH-25B	Solid	160.3	
720-381-106	EBHH-51A	Solid	160.3	
720-381-107	EBHH-51B	Solid	160.3	
720-381-108	EBHH-51C	Solid	160.3	
720-381-109	EBHH-52A	Solid	160.3	
720-381-110	EBHH-52B	Solid	160.3	
720-381-111	EBHH-52C	Solid	160.3	
720-381-112	EBHH-53A	Solid	160.3	
720-381-113	EBHH-53B	Solid	160.3	
720-381-114	EBHH-54A	Solid	160.3	
720-381-115	EBHH-54B	Solid	160.3	
Analysis Batch:720-2301				
MB 720-2301/1	Method Blank	Solid	160.3	
720-381-56	EBHH-14A	Solid	160.3	
720-381-56DU	Duplicate	Solid	160.3	

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1667

Method: 8270C
Preparation: 3550B

Lab Sample ID: MB 720-1667/14-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/11/2005 1627
Date Prepared: 11/11/2005 1215

Analysis Batch: 720-1821
Prep Batch: 720-1667
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.12 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result	Qual	RL
Naphthalene	ND		5.0
Acenaphthene	ND		5.0
Acenaphthylene	ND		5.0
Fluorene	ND		5.0
Phenanthrene	ND		5.0
Anthracene	ND		5.0
Benzo[a]anthracene	ND		5.0
Chrysene	ND		5.0
Benzo[a]pyrene	ND		5.0
Benzo[b]fluoranthene	ND		5.0
Benzo[k]fluoranthene	ND		5.0
Benzo[g,h,i]perylene	ND		5.0
Indeno[1,2,3-cd]pyrene	ND		5.0
Fluoranthene	ND		5.0
Pyrene	ND		5.0
Dibenz(a,h)anthracene	ND		5.0

Surrogate	% Rec	Acceptance Limits
2-Fluorobiphenyl	68	30 - 115
Terphenyl-d14	87	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1667

Method: 8270C
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1667/15-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/11/2005 1654
Date Prepared: 11/11/2005 1215

Analysis Batch: 720-1821
Prep Batch: 720-1667
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\111105
Initial Weight/Volume: 30.08 g
Final Weight/Volume: 1 mL
Injection Volume:

LCSD Lab Sample ID: LCSD 720-1667/16-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/11/2005 1722
Date Prepared: 11/11/2005 1215

Analysis Batch: 720-1821
Prep Batch: 720-1667
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\111105\l
Initial Weight/Volume: 30.18 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Naphthalene	74	76	21 - 133	2	35		
Acenaphthene	70	75	47 - 145	6	35		
Acenaphthylene	76	80	33 - 145	5	35		
Fluorene	81	82	59 - 121	0	35		
Phenanthrene	74	76	10 - 130	1	35		
Anthracene	79	77	27 - 133	4	35		
Benzo[a]anthracene	75	74	33 - 143	1	35		
Chrysene	84	83	17 - 168	1	35		
Benzo[a]pyrene	76	77	17 - 163	1	35		
Benzo[b]fluoranthene	75	76	24 - 159	1	35		
Benzo[k]fluoranthene	88	94	11 - 162	6	35		
Benzo[g,h,i]perylene	81	86	9 - 219	5	35		
Indeno[1,2,3-cd]pyrene	77	80	9 - 171	4	35		
Fluoranthene	79	74	26 - 137	7	35		
Pyrene	85	89	52 - 115	4	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
2-Fluorobiphenyl	70	75	30 - 115
Terphenyl-d14	85	90	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1667

Method: 8270C
Preparation: 3550B

MS Lab Sample ID:	720-381-105	Analysis Batch:	720-1821	Instrument ID:	Sat 2K2
Client Matrix:	Solid	Prep Batch:	720-1667	Lab File ID:	c:\saturnws\data\200511\
Dilution:	1.0			Initial Weight/Volume:	30.13 g
Date Analyzed:	11/11/2005 1750			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1215			Injection Volume:	

MSD Lab Sample ID:	720-381-105	Analysis Batch:	720-1821	Instrument ID:	Sat 2K2
Client Matrix:	Solid	Prep Batch:	720-1667	Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.08 g
Date Analyzed:	11/11/2005 1818			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1215			Injection Volume:	

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Naphthalene	71	71	21 - 133	1	35		
Acenaphthene	71	70	47 - 145	0	35		
Acenaphthylene	75	68	33 - 145	8	35		
Fluorene	81	75	59 - 121	8	35		
Phenanthrene	76	73	10 - 130	3	35		
Anthracene	78	76	27 - 133	3	35		
Benzo[a]anthracene	66	65	33 - 143	1	35		
Chrysene	70	66	17 - 168	4	35		
Benzo[a]pyrene	64	63	17 - 163	1	35		
Benzo[b]fluoranthene	54	59	24 - 159	6	35		
Benzo[k]fluoranthene	84	80	11 - 162	4	35		
Benzo[g,h,i]perylene	70	68	9 - 219	2	35		
Indeno[1,2,3-cd]pyrene	66	69	9 - 171	4	35		
Fluoranthene	57	58	26 - 137	0	35		
Pyrene	69	67	52 - 115	2	35		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
2-Fluorobiphenyl	70	72	30 - 115
Terphenyl-d14	84	80	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1711

Method: 8270C
Preparation: 3550B

Lab Sample ID: MB 720-1711/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0914
Date Prepared: 11/14/2005 1006

Analysis Batch: 720-1849
Prep Batch: 720-1711
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.09 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result	Qual	RL
Naphthalene	ND		5.0
Acenaphthene	ND		5.0
Acenaphthylene	ND		5.0
Fluorene	ND		5.0
Phenanthrene	ND		5.0
Anthracene	ND		5.0
Benzo[a]anthracene	ND		5.0
Chrysene	ND		5.0
Benzo[a]pyrene	ND		5.0
Benzo[b]fluoranthene	ND		5.0
Benzo[k]fluoranthene	ND		5.0
Benzo[g,h,i]perylene	ND		5.0
Indeno[1,2,3-cd]pyrene	ND		5.0
Fluoranthene	ND		5.0
Pyrene	ND		5.0
Dibenz(a,h)anthracene	ND		5.0

Surrogate	% Rec	Acceptance Limits
2-Fluorobiphenyl	68	30 - 115
Terphenyl-d14	78	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1711

Method: 8270C
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1711/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0942
Date Prepared: 11/14/2005 1006

Analysis Batch: 720-1849
Prep Batch: 720-1711
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\111505
Initial Weight/Volume: 30.08 g
Final Weight/Volume: 1 mL
Injection Volume:

LCSD Lab Sample ID: LCSD 720-1711/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1010
Date Prepared: 11/14/2005 1006

Analysis Batch: 720-1849
Prep Batch: 720-1711
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\111505\l
Initial Weight/Volume: 30.40 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Naphthalene	65	67	21 - 133	2	35		
Acenaphthene	58	64	47 - 145	9	35		
Acenaphthylene	61	66	33 - 145	7	35		
Fluorene	64	69	59 - 121	8	35		
Phenanthrene	64	67	10 - 130	4	35		
Anthracene	62	66	27 - 133	6	35		
Benzo[a]anthracene	70	74	33 - 143	4	35		
Chrysene	70	64	17 - 168	11	35		
Benzo[a]pyrene	66	63	17 - 163	5	35		
Benzo[b]fluoranthene	69	71	24 - 159	1	35		
Benzo[k]fluoranthene	72	72	11 - 162	1	35		
Benzo[g,h,i]perylene	72	68	9 - 219	8	35		
Indeno[1,2,3-cd]pyrene	68	66	9 - 171	4	35		
Fluoranthene	64	66	26 - 137	1	35		
Pyrene	75	71	52 - 115	6	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
2-Fluorobiphenyl	64	71	30 - 115
Terphenyl-d14	77	73	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1711

Method: 8270C
Preparation: 3550B

MS Lab Sample ID:	720-416-A-1-B MS	Analysis Batch: 720-1851	Instrument ID: Sat 2K2
Client Matrix:	Solid	Prep Batch: 720-1711	Lab File ID: c:\saturnws\data\200511\
Dilution:	1.0		Initial Weight/Volume: 30.40 g
Date Analyzed:	11/16/2005 1605		Final Weight/Volume: 1 mL
Date Prepared:	11/14/2005 1006		Injection Volume:

MSD Lab Sample ID:	720-416-A-1-C MSD	Analysis Batch: 720-1851	Instrument ID: Sat 2K2
Client Matrix:	Solid	Prep Batch: 720-1711	Lab File ID: c:\saturnws\data\200511\11
Dilution:	1.0		Initial Weight/Volume: 30.30 g
Date Analyzed:	11/16/2005 1633		Final Weight/Volume: 1 mL
Date Prepared:	11/14/2005 1006		Injection Volume:

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Naphthalene	64	66	21 - 133	3	35		
Acenaphthene	61	64	47 - 145	4	35		
Acenaphthylene	63	67	33 - 145	7	35		
Fluorene	63	65	59 - 121	4	35		
Phenanthrene	66	62	10 - 130	5	35		
Anthracene	65	65	27 - 133	0	35		
Benzo[a]anthracene	60	63	33 - 143	6	35		
Chrysene	64	65	17 - 168	1	35		
Benzo[a]pyrene	64	61	17 - 163	4	35		
Benzo[b]fluoranthene	66	66	24 - 159	1	35		
Benzo[k]fluoranthene	73	69	11 - 162	5	35		
Benzo[g,h,i]perylene	68	67	9 - 219	2	35		
Indeno[1,2,3-cd]pyrene	67	66	9 - 171	1	35		
Fluoranthene	63	63	26 - 137	2	35		
Pyrene	68	71	52 - 115	5	35		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
2-Fluorobiphenyl	67	68	30 - 115
Terphenyl-d14	69	73	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1830

Method: 8270C
Preparation: 3550B

Lab Sample ID: MB 720-1830/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/23/2005 2103
Date Prepared: 11/16/2005 1218

Analysis Batch: 720-2227
Prep Batch: 720-1830
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.09 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result	Qual	RL
Naphthalene	ND		5.0
Acenaphthene	ND		5.0
Acenaphthylene	ND		5.0
Fluorene	ND		5.0
Phenanthrene	ND		5.0
Anthracene	ND		5.0
Benzo[a]anthracene	ND		5.0
Chrysene	ND		5.0
Benzo[a]pyrene	ND		5.0
Benzo[b]fluoranthene	ND		5.0
Benzo[k]fluoranthene	ND		5.0
Benzo[g,h,i]perylene	ND		5.0
Indeno[1,2,3-cd]pyrene	ND		5.0
Fluoranthene	ND		5.0
Pyrene	ND		5.0
Dibenz(a,h)anthracene	ND		5.0

Surrogate	% Rec	Acceptance Limits
2-Fluorobiphenyl	76	30 - 115
Terphenyl-d14	87	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1830

Method: 8270C
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1830/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/23/2005 2131
Date Prepared: 11/16/2005 1218

Analysis Batch: 720-2227
Prep Batch: 720-1830
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\112305
Initial Weight/Volume: 30.14 g
Final Weight/Volume: 1 mL
Injection Volume:

LCSD Lab Sample ID: LCSD 720-1830/4-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/23/2005 2159
Date Prepared: 11/16/2005 1218

Analysis Batch: 720-2227
Prep Batch: 720-1830
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\112305\l
Initial Weight/Volume: 30.19 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Naphthalene	68	69	21 - 133	1	35		
Acenaphthene	72	74	47 - 145	3	35		
Acenaphthylene	66	66	33 - 145	1	35		
Fluorene	74	73	59 - 121	1	35		
Phenanthrene	77	72	10 - 130	6	35		
Anthracene	77	71	27 - 133	8	35		
Benzo[a]anthracene	91	86	33 - 143	6	35		
Chrysene	81	74	17 - 168	9	35		
Benzo[a]pyrene	85	81	17 - 163	5	35		
Benzo[b]fluoranthene	91	86	24 - 159	6	35		
Benzo[k]fluoranthene	91	90	11 - 162	1	35		
Benzo[g,h,i]perylene	86	85	9 - 219	1	35		
Indeno[1,2,3-cd]pyrene	95	124	9 - 171	26	35		
Fluoranthene	84	76	26 - 137	10	35		
Pyrene	84	83	52 - 115	2	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
2-Fluorobiphenyl	73	72	30 - 115
Terphenyl-d14	92	84	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1830

Method: 8270C
Preparation: 3550B

MS Lab Sample ID:	720-381-71	Analysis Batch:	720-2227	Instrument ID:	Sat 2K2
Client Matrix:	Solid	Prep Batch:	720-1830	Lab File ID:	c:\saturnws\data\200511\
Dilution:	1.0			Initial Weight/Volume:	30.00 g
Date Analyzed:	11/23/2005 2255			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

MSD Lab Sample ID:	720-381-71	Analysis Batch:	720-2227	Instrument ID:	Sat 2K2
Client Matrix:	Solid	Prep Batch:	720-1830	Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.15 g
Date Analyzed:	11/23/2005 2322			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Naphthalene	68	67	21 - 133	2	35		
Acenaphthene	70	75	47 - 145	6	35		
Acenaphthylene	64	66	33 - 145	1	35		
Fluorene	73	73	59 - 121	1	35		
Phenanthrene	68	69	10 - 130	0	35		
Anthracene	70	72	27 - 133	3	35		
Benzo[a]anthracene	74	80	33 - 143	7	35		
Chrysene	66	71	17 - 168	6	35		
Benzo[a]pyrene	70	75	17 - 163	6	35		
Benzo[b]fluoranthene	73	81	24 - 159	10	35		
Benzo[k]fluoranthene	80	84	11 - 162	5	35		
Benzo[g,h,i]perylene	67	73	9 - 219	8	35		
Indeno[1,2,3-cd]pyrene	68	72	9 - 171	5	35		
Fluoranthene	72	76	26 - 137	5	35		
Pyrene	73	77	52 - 115	5	35		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
2-Fluorobiphenyl	69	71	30 - 115
Terphenyl-d14	78	82	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1624

Method: 8015B
Preparation: 3550B

Lab Sample ID: MB 720-1624/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/10/2005 1414
Date Prepared: 11/10/2005 1200

Analysis Batch: 720-1682
Prep Batch: 720-1624
Units: mg/Kg

Instrument ID: HP DRO5
Lab File ID: N/A
Initial Weight/Volume: 30.16 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Diesel Range Organics [C10-C28]	ND		0.99
Motor Oil Range Organics [C24-C36]	ND		50

Surrogate	% Rec	Acceptance Limits
o-Terphenyl	67	60 - 130

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1624

Method: 8015B
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1624/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/10/2005 1441
Date Prepared: 11/10/2005 1200

Analysis Batch: 720-1682
Prep Batch: 720-1624
Units: mg/Kg

Instrument ID: HP DRO5
Lab File ID: N/A
Initial Weight/Volume: 30.15 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1624/4-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/10/2005 1508
Date Prepared: 11/10/2005 1200

Analysis Batch: 720-1682
Prep Batch: 720-1624
Units: mg/Kg

Instrument ID: HP DRO5
Lab File ID: N/A
Initial Weight/Volume: 30.21 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Diesel Range Organics [C10-C28]	78	85	60 - 130	9	30		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
o-Terphenyl	68	73	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1758

Method: 8015B
Preparation: 3550B

Lab Sample ID: MB 720-1758/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1448
Date Prepared: 11/15/2005 1005

Analysis Batch: 720-1806
Prep Batch: 720-1758
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.25 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Diesel Range Organics [C10-C28]	ND		0.99
Motor Oil Range Organics [C24-C36]	ND		50

Surrogate	% Rec	Acceptance Limits
o-Terphenyl	81	60 - 130

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1758

Method: 8015B
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1758/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1448
Date Prepared: 11/15/2005 1005

Analysis Batch: 720-1806
Prep Batch: 720-1758
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.29 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1758/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1516
Date Prepared: 11/15/2005 1005

Analysis Batch: 720-1806
Prep Batch: 720-1758
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.19 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Diesel Range Organics [C10-C28]	91	94	60 - 130	3	30		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
o-Terphenyl	77	77	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1758

Method: 8015B
Preparation: 3550B

MS Lab Sample ID:	720-381-39	Analysis Batch: 720-1806	Instrument ID: HP DRO3
Client Matrix:	Solid	Prep Batch: 720-1758	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 30.26 g
Date Analyzed:	11/15/2005 1116		Final Weight/Volume: 5 mL
Date Prepared:	11/15/2005 1005		Injection Volume:
			Column ID: PRIMARY

MSD Lab Sample ID:	720-381-39	Analysis Batch: 720-1806	Instrument ID: HP DRO3
Client Matrix:	Solid	Prep Batch: 720-1758	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 30.21 g
Date Analyzed:	11/15/2005 1139		Final Weight/Volume: 5 mL
Date Prepared:	11/15/2005 1005		Injection Volume:
			Column ID: PRIMARY

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Diesel Range Organics [C10-C28]	84	84	60 - 130	0	30		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
o-Terphenyl	82	83	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1829

Method: 8015B
Preparation: 3550B

Lab Sample ID: MB 720-1829/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 0025
Date Prepared: 11/16/2005 1152

Analysis Batch: 720-1987
Prep Batch: 720-1829
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.28 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Diesel Range Organics [C10-C28]	ND		0.99
Motor Oil Range Organics [C24-C36]	ND		50

Surrogate	% Rec	Acceptance Limits
o-Terphenyl	76	60 - 130

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1829

Method: 8015B
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1829/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 2330
Date Prepared: 11/16/2005 1152

Analysis Batch: 720-1987
Prep Batch: 720-1829
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.12 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1829/4-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 2357
Date Prepared: 11/16/2005 1152

Analysis Batch: 720-1987
Prep Batch: 720-1829
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.25 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Diesel Range Organics [C10-C28]	80	85	60 - 130	6	30		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
o-Terphenyl	90	92	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1829

Method: 8015B
Preparation: 3550B

MS Lab Sample ID:	720-401-A-28-E MS	Analysis Batch: 720-1987	Instrument ID: HP DRO3
Client Matrix:	Solid	Prep Batch: 720-1829	Lab File ID: N/A
Dilution:	2.0		Initial Weight/Volume: 30.23 g
Date Analyzed:	11/17/2005 1332		Final Weight/Volume: 5 mL
Date Prepared:	11/16/2005 1152		Injection Volume:
			Column ID: PRIMARY

MSD Lab Sample ID:	720-401-A-28-F MSD	Analysis Batch: 720-1987	Instrument ID: HP DRO3
Client Matrix:	Solid	Prep Batch: 720-1829	Lab File ID: N/A
Dilution:	2.0		Initial Weight/Volume: 30.28 g
Date Analyzed:	11/17/2005 1359		Final Weight/Volume: 5 mL
Date Prepared:	11/16/2005 1152		Injection Volume:
			Column ID: PRIMARY

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Diesel Range Organics [C10-C28]	65	77	60 - 130	6	30		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
o-Terphenyl	87	89	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1635

Method: 8081A
Preparation: 3550B

Lab Sample ID: MB 720-1635/21-B
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/11/2005 0948
Date Prepared: 11/10/2005 1454

Analysis Batch: 720-1801
Prep Batch: 720-1635
Units: ug/Kg

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.16 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Aldrin	ND		2.0
Dieldrin	ND		2.0
Endrin aldehyde	ND		2.0
Endrin	ND		2.0
Endrin ketone	ND		2.0
Heptachlor	ND		2.0
Heptachlor epoxide	ND		2.0
4,4'-DDT	ND		2.0
4,4'-DDE	ND		2.0
4,4'-DDD	ND		2.0
Endosulfan I	ND		2.0
Endosulfan II	ND		2.0
alpha-BHC	ND		2.0
beta-BHC	ND		2.0
gamma-BHC (Lindane)	ND		2.0
delta-BHC	ND		2.0
Endosulfan sulfate	ND		2.0
Methoxychlor	ND		2.0
Toxaphene	ND		99
Chlordane (technical)	ND		50
alpha-Chlordane	ND		2.0
gamma-Chlordane	ND		2.0

Surrogate	% Rec	Acceptance Limits
Tetrachloro-m-xylene	80	50 - 125
DCB Decachlorobiphenyl	89	46 - 142

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1635

Method: 8081A
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1635/22-B
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/11/2005 1016
Date Prepared: 11/10/2005 1454

Analysis Batch: 720-1801
Prep Batch: 720-1635
Units: ug/Kg

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.05 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1635/23-B
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/11/2005 1045
Date Prepared: 11/10/2005 1454

Analysis Batch: 720-1801
Prep Batch: 720-1635
Units: ug/Kg

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.25 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Aldrin	84	94	37 - 136	10	35		
Dieldrin	83	91	58 - 135	9	35		
Endrin	88	97	58 - 134	9	35		
Heptachlor	78	85	40 - 136	8	35		
4,4'-DDT	89	98	55 - 132	9	35		
gamma-BHC (Lindane)	82	89	37 - 137	8	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
Tetrachloro-m-xylene	88	99	50 - 125
DCB Decachlorobiphenyl	100	110	46 - 142

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1635

Method: 8081A
Preparation: 3550B

MS Lab Sample ID:	720-381-42	Analysis Batch: 720-1801	Instrument ID: Varian Pest 2
Client Matrix:	Solid	Prep Batch: 720-1635	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 30.07 g
Date Analyzed:	11/17/2005 1529		Final Weight/Volume: 10 mL
Date Prepared:	11/10/2005 1454		Injection Volume:
			Column ID: PRIMARY

MSD Lab Sample ID:	720-381-42	Analysis Batch: 720-1801	Instrument ID: Varian Pest 2
Client Matrix:	Solid	Prep Batch: 720-1635	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 30.09 g
Date Analyzed:	11/17/2005 1529		Final Weight/Volume: 10 mL
Date Prepared:	11/10/2005 1454		Injection Volume:
			Column ID: PRIMARY

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Aldrin	92	85	37 - 136	8	35		
Dieldrin	92	87	58 - 135	7	35		
Endrin	97	91	58 - 134	6	35		
Heptachlor	93	86	40 - 136	7	35		
4,4'-DDT	96	91	55 - 132	6	35		
gamma-BHC (Lindane)	92	86	37 - 137	7	35		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
Tetrachloro-m-xylene	92	86	50 - 125
DCB Decachlorobiphenyl	97	88	46 - 142

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1646

Method: 8081A
Preparation: 3550B

Lab Sample ID: MB 720-1646/15-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0912
Date Prepared: 11/10/2005 1810

Analysis Batch: 720-1940
Prep Batch: 720-1646
Units: ug/Kg

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.16 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Aldrin	ND		2.0
Dieldrin	ND		2.0
Endrin aldehyde	ND		2.0
Endrin	ND		2.0
Endrin ketone	ND		2.0
Heptachlor	ND		2.0
Heptachlor epoxide	ND		2.0
4,4'-DDT	ND		2.0
4,4'-DDE	ND		2.0
4,4'-DDD	ND		2.0
Endosulfan I	ND		2.0
Endosulfan II	ND		2.0
alpha-BHC	ND		2.0
beta-BHC	ND		2.0
gamma-BHC (Lindane)	ND		2.0
delta-BHC	ND		2.0
Endosulfan sulfate	ND		2.0
Methoxychlor	ND		2.0
Toxaphene	ND		99
Chlordane (technical)	ND		50
alpha-Chlordane	ND		2.0
gamma-Chlordane	ND		2.0

Surrogate	% Rec	Acceptance Limits
Tetrachloro-m-xylene	87	50 - 125
DCB Decachlorobiphenyl	92	46 - 142

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1646

Method: 8081A
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1646/16-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0741
Date Prepared: 11/10/2005 1810

Analysis Batch: 720-1940
Prep Batch: 720-1646
Units: ug/Kg

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.11 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1646/17-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0825
Date Prepared: 11/10/2005 1810

Analysis Batch: 720-1940
Prep Batch: 720-1646
Units: ug/Kg

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.13 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Aldrin	85	101	37 - 136	17	35		
Dieldrin	85	96	58 - 135	11	35		
Endrin	92	102	58 - 134	11	35		
Heptachlor	84	93	40 - 136	10	35		
4,4'-DDT	93	103	55 - 132	10	35		
gamma-BHC (Lindane)	88	96	37 - 137	8	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
Tetrachloro-m-xylene	91	93	50 - 125
DCB Decachlorobiphenyl	94	100	46 - 142

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1646

Method: 8081A
Preparation: 3550B

MS Lab Sample ID:	720-381-105	Analysis Batch: 720-1940	Instrument ID: Varian Pest 2
Client Matrix:	Solid	Prep Batch: 720-1646	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 30.26 g
Date Analyzed:	11/17/2005 0447		Final Weight/Volume: 10 mL
Date Prepared:	11/10/2005 1810		Injection Volume:
			Column ID: PRIMARY

MSD Lab Sample ID:	720-381-105	Analysis Batch: 720-1940	Instrument ID: Varian Pest 2
Client Matrix:	Solid	Prep Batch: 720-1646	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 30.05 g
Date Analyzed:	11/17/2005 0516		Final Weight/Volume: 10 mL
Date Prepared:	11/10/2005 1810		Injection Volume:
			Column ID: PRIMARY

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Aldrin	93	94	37 - 136	3	35		
Dieldrin	91	94	58 - 135	4	35		
Endrin	94	98	58 - 134	4	35		
Heptachlor	94	95	40 - 136	2	35		
4,4'-DDT	99	103	55 - 132	5	35		
gamma-BHC (Lindane)	92	96	37 - 137	5	35		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
Tetrachloro-m-xylene	95	97	50 - 125
DCB Decachlorobiphenyl	92	94	46 - 142

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1636

Method: 6010B
Preparation: 3050B

Lab Sample ID: MB 720-1636/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/11/2005 1152
Date Prepared: 11/10/2005 1457

Analysis Batch: 720-1683
Prep Batch: 720-1636
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Silver	ND		1.0
Arsenic	ND		1.0
Barium	ND		1.0
Cadmium	ND		0.50
Chromium	ND		1.0
Lead	ND		1.0
Selenium	ND		2.0

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1636

Method: 6010B
Preparation: 3050B

LCS Lab Sample ID: LCS 720-1636/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/11/2005 1155
Date Prepared: 11/10/2005 1457

Analysis Batch: 720-1683
Prep Batch: 720-1636
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1636/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/11/2005 1159
Date Prepared: 11/10/2005 1457

Analysis Batch: 720-1683
Prep Batch: 720-1636
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Silver	107	103	80 - 120	4	20		
Arsenic	110	106	80 - 120	4	20		
Barium	108	103	80 - 120	4	20		
Cadmium	107	103	80 - 120	4	20		
Chromium	108	103	80 - 120	4	20		
Lead	105	101	80 - 120	4	20		
Selenium	110	106	80 - 120	4	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1636

Method: 6010B
Preparation: 3050B

MS Lab Sample ID:	720-381-20	Analysis Batch:	720-1683	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1636	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/11/2005 1326			Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457				

MSD Lab Sample ID:	720-381-20	Analysis Batch:	720-1683	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1636	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.01 g
Date Analyzed:	11/11/2005 1330			Final Weight/Volume:	50 mL
Date Prepared:	11/10/2005 1457				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Silver	96	96	75 - 125	1	20		
Arsenic	96	97	75 - 125	0	20		
Barium	82	71	75 - 125	6	20		N
Cadmium	87	87	75 - 125	1	20		
Chromium	97	95	75 - 125	3	20		
Lead	81	117	75 - 125	16	20		
Selenium	92	93	75 - 125	1	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1731

Method: 6010B
Preparation: 3050B

Lab Sample ID: MB 720-1731/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0814
Date Prepared: 11/14/2005 1454

Analysis Batch: 720-1762
Prep Batch: 720-1731
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Silver	ND		1.0
Arsenic	ND		1.0
Barium	ND		1.0
Cadmium	ND		0.50
Chromium	ND		1.0
Lead	ND		1.0
Selenium	ND		2.0

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1731

Method: 6010B
Preparation: 3050B

LCS Lab Sample ID: LCS 720-1731/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0817
Date Prepared: 11/14/2005 1454

Analysis Batch: 720-1762
Prep Batch: 720-1731
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1731/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0821
Date Prepared: 11/14/2005 1454

Analysis Batch: 720-1762
Prep Batch: 720-1731
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Silver	94	96	80 - 120	2	20		
Arsenic	96	99	80 - 120	3	20		
Barium	96	99	80 - 120	3	20		
Cadmium	95	97	80 - 120	2	20		
Chromium	95	97	80 - 120	3	20		
Lead	93	96	80 - 120	2	20		
Selenium	96	99	80 - 120	3	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1731

Method: 6010B
Preparation: 3050B

MS Lab Sample ID:	720-381-58	Analysis Batch:	720-1762	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1731	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.04 g
Date Analyzed:	11/15/2005 0902			Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454				

MSD Lab Sample ID:	720-381-58	Analysis Batch:	720-1762	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1731	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.03 g
Date Analyzed:	11/15/2005 0906			Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1454				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Silver	88	87	75 - 125	0	20		
Arsenic	89	87	75 - 125	1	20		
Barium	89	80	75 - 125	3	20		
Cadmium	84	83	75 - 125	1	20		
Chromium	88	87	75 - 125	0	20		
Lead	84	82	75 - 125	2	20		
Selenium	86	85	75 - 125	0	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1734

Method: 6010B
Preparation: 3050B

Lab Sample ID: MB 720-1734/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1304
Date Prepared: 11/14/2005 1539

Analysis Batch: 720-1808
Prep Batch: 720-1734
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Silver	ND		1.0
Silver	ND		1.0
Arsenic	ND		1.0
Arsenic	ND		1.0
Barium	ND		1.0
Barium	ND		1.0
Cadmium	ND		0.50
Cadmium	ND		0.50
Chromium	ND		1.0
Chromium	ND		1.0
Lead	ND		1.0
Lead	ND		1.0
Selenium	ND		2.0
Selenium	ND		2.0

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1734

Method: 6010B
Preparation: 3050B

LCS Lab Sample ID: LCS 720-1734/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1307
Date Prepared: 11/14/2005 1539

Analysis Batch: 720-1762
Prep Batch: 720-1734
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1734/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1311
Date Prepared: 11/14/2005 1539

Analysis Batch: 720-1762
Prep Batch: 720-1734
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Silver	94	97	80 - 120	3	20		
Silver	94	97	80 - 120	3	20		
Arsenic	96	100	80 - 120	4	20		
Arsenic	96	100	80 - 120	4	20		
Barium	97	100	80 - 120	3	20		
Barium	97	100	80 - 120	3	20		
Cadmium	94	97	80 - 120	3	20		
Cadmium	94	97	80 - 120	3	20		
Chromium	95	98	80 - 120	4	20		
Chromium	95	98	80 - 120	4	20		
Lead	93	96	80 - 120	3	20		
Lead	93	96	80 - 120	3	20		
Selenium	97	100	80 - 120	4	20		
Selenium	97	100	80 - 120	4	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1734

Method: 6010B
Preparation: 3050B

MS Lab Sample ID: 720-381-40
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1418
Date Prepared: 11/14/2005 1539

Analysis Batch: 720-1762
Prep Batch: 720-1734

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1.03 g
Final Weight/Volume: 50 mL

MSD Lab Sample ID: 720-381-40
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1422
Date Prepared: 11/14/2005 1539

Analysis Batch: 720-1762
Prep Batch: 720-1734

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1.02 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Silver	99	97	75 - 125	1	20		
Silver	86	85	75 - 125	1	20		
Arsenic	103	100	75 - 125	1	20		
Arsenic	86	84	75 - 125	1	20		
Barium	118	130	75 - 125	5	20		*
Barium	80	90	75 - 125	5	20		
Cadmium	91	92	75 - 125	1	20		
Cadmium	78	78	75 - 125	1	20		
Chromium	95	103	75 - 125	7	20		
Chromium	77	85	75 - 125	7	20		
Lead	86	118	75 - 125	16	20		
Lead	58	85	75 - 125	16	20	N	
Selenium	101	98	75 - 125	2	20		
Selenium	85	82	75 - 125	2	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1734

Method: 6010B
Preparation: 3050B

MS Lab Sample ID:	720-381-42	Analysis Batch:	720-1810	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1734	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/15/2005 1849			Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539				

MSD Lab Sample ID:	720-381-42	Analysis Batch:	720-1810	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1734	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.01 g
Date Analyzed:	11/15/2005 1853			Final Weight/Volume:	50 mL
Date Prepared:	11/14/2005 1539				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Silver	95	91	75 - 125	5	20		
Arsenic	96	96	75 - 125	1	20		
Barium	92	96	75 - 125	1	20		
Cadmium	91	90	75 - 125	2	20		
Chromium	98	94	75 - 125	4	20		
Lead	90	90	75 - 125	1	20		
Selenium	94	95	75 - 125	1	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1773

Method: 6010B
Preparation: 3050B

Lab Sample ID: MB 720-1773/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 1432
Date Prepared: 11/15/2005 1245

Analysis Batch: 720-1850
Prep Batch: 720-1773
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Silver	ND		1.0
Arsenic	ND		1.0
Barium	ND		1.0
Cadmium	ND		0.50
Chromium	ND		1.0
Lead	ND		1.0
Selenium	ND		2.0

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1773

Method: 6010B
Preparation: 3050B

LCS Lab Sample ID: LCS 720-1773/2-B
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 1435
Date Prepared: 11/15/2005 1245

Analysis Batch: 720-1850
Prep Batch: 720-1773
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1773/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 1021
Date Prepared: 11/15/2005 1245

Analysis Batch: 720-1850
Prep Batch: 720-1773
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Silver	97	99	80 - 120	2	20		
Arsenic	99	102	80 - 120	3	20		
Barium	99	100	80 - 120	1	20		
Cadmium	97	100	80 - 120	3	20		
Chromium	98	100	80 - 120	2	20		
Lead	96	98	80 - 120	3	20		
Selenium	100	104	80 - 120	4	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1773

Method: 6010B
Preparation: 3050B

MS Lab Sample ID:	720-381-78	Analysis Batch:	720-1850	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1773	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 1033			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245				

MSD Lab Sample ID:	720-381-78	Analysis Batch:	720-1850	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1773	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.05 g
Date Analyzed:	11/16/2005 1037			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Silver	87	88	75 - 125	4	20		
Arsenic	91	87	75 - 125	9	20		
Barium	79	76	75 - 125	4	20		
Cadmium	88	84	75 - 125	10	20		
Chromium	88	83	75 - 125	8	20		
Lead	86	82	75 - 125	9	20		
Selenium	89	85	75 - 125	10	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1773

Method: 6010B
Preparation: 3050B

MS Lab Sample ID:	720-381-94	Analysis Batch:	720-1850	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1773	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 1122			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245				

MSD Lab Sample ID:	720-381-94	Analysis Batch:	720-1850	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1773	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/16/2005 1126			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1245				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Silver	84	83	75 - 125	2	20		
Arsenic	84	81	75 - 125	3	20		
Barium	82	75	75 - 125	3	20		
Cadmium	79	78	75 - 125	1	20		
Chromium	83	79	75 - 125	4	20		
Lead	80	59	75 - 125	17	20		N
Selenium	82	80	75 - 125	2	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1788

Method: 6010B
Preparation: 3050B

Lab Sample ID: MB 720-1788/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 1148
Date Prepared: 11/15/2005 1414

Analysis Batch: 720-1850
Prep Batch: 720-1788
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Silver	ND		1.0
Arsenic	ND		1.0
Barium	ND		1.0
Cadmium	ND		0.50
Chromium	ND		1.0
Lead	ND		1.0
Selenium	ND		2.0

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1788

Method: 6010B
Preparation: 3050B

LCS Lab Sample ID: LCS 720-1788/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 1151
Date Prepared: 11/15/2005 1414

Analysis Batch: 720-1850
Prep Batch: 720-1788
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1788/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 1155
Date Prepared: 11/15/2005 1414

Analysis Batch: 720-1850
Prep Batch: 720-1788
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Silver	100	96	80 - 120	4	20		
Arsenic	102	98	80 - 120	5	20		
Barium	102	98	80 - 120	4	20		
Cadmium	101	97	80 - 120	4	20		
Chromium	102	97	80 - 120	5	20		
Lead	99	95	80 - 120	4	20		
Selenium	106	100	80 - 120	5	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1788

Method: 6010B
Preparation: 3050B

MS Lab Sample ID:	720-381-105	Analysis Batch:	720-1850	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1788	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.05 g
Date Analyzed:	11/16/2005 1217			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414				

MSD Lab Sample ID:	720-381-105	Analysis Batch:	720-1850	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1788	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.01 g
Date Analyzed:	11/16/2005 1221			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 1414				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Silver	82	81	75 - 125	3	20		
Arsenic	81	82	75 - 125	5	20		
Barium	79	89	75 - 125	5	20		
Cadmium	78	78	75 - 125	5	20		
Chromium	78	81	75 - 125	4	20		
Lead	79	93	75 - 125	10	20		
Selenium	77	77	75 - 125	4	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1666

Method: 7471A
Preparation: 7471A

Lab Sample ID: MB 720-1666/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0901
Date Prepared: 11/11/2005 1159

Analysis Batch: 720-1760
Prep Batch: 720-1666
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume:

Analyte	Result	Qual	RL
Mercury	ND		0.0010

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1666

Method: 7471A
Preparation: 7471A

LCS Lab Sample ID: LCS 720-1666/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0902
Date Prepared: 11/11/2005 1159

Analysis Batch: 720-1760
Prep Batch: 720-1666
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume:

LCSD Lab Sample ID: LCSD 720-1666/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0904
Date Prepared: 11/11/2005 1159

Analysis Batch: 720-1760
Prep Batch: 720-1666
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume:

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Mercury	2	2	85 - 115	0	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1666

Method: 7471A
Preparation: 7471A

MS Lab Sample ID:	720-381-20	Analysis Batch: 720-1760	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1666	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.00 g
Date Analyzed:	11/15/2005 0926		Final Weight/Volume:
Date Prepared:	11/11/2005 1159		

MSD Lab Sample ID:	720-381-20	Analysis Batch: 720-1760	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1666	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.01 g
Date Analyzed:	11/15/2005 0927		Final Weight/Volume:
Date Prepared:	11/11/2005 1159		

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	2	2	85 - 115	4	20	*	*

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1737

Method: 7471A
Preparation: 7471A

Lab Sample ID: MB 720-1737/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0935
Date Prepared: 11/14/2005 1554

Analysis Batch: 720-1760
Prep Batch: 720-1737
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Mercury	ND		0.050

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1737

Method: 7471A
Preparation: 7471A

LCS Lab Sample ID: LCS 720-1737/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0937
Date Prepared: 11/14/2005 1554

Analysis Batch: 720-1760
Prep Batch: 720-1737
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1737/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0938
Date Prepared: 11/14/2005 1554

Analysis Batch: 720-1760
Prep Batch: 720-1737
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Mercury	104	104	85 - 115	0	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1737

Method: 7471A
Preparation: 7471A

MS Lab Sample ID:	720-381-58	Analysis Batch: 720-1760	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1737	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.00 g
Date Analyzed:	11/15/2005 0951		Final Weight/Volume: 50 mL
Date Prepared:	11/14/2005 1554		

MSD Lab Sample ID:	720-381-58	Analysis Batch: 720-1760	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1737	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.00 g
Date Analyzed:	11/15/2005 0952		Final Weight/Volume: 50 mL
Date Prepared:	11/14/2005 1554		

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	110	117	85 - 115	5	20		*

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1743

Method: 7471A
Preparation: 7471A

Lab Sample ID: MB 720-1743/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1358
Date Prepared: 11/14/2005 1700

Analysis Batch: 720-1807
Prep Batch: 720-1743
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Mercury	ND		0.050

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1743

Method: 7471A
Preparation: 7471A

LCS Lab Sample ID: LCS 720-1743/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1400
Date Prepared: 11/14/2005 1700

Analysis Batch: 720-1807
Prep Batch: 720-1743
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1743/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1401
Date Prepared: 11/14/2005 1700

Analysis Batch: 720-1807
Prep Batch: 720-1743
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Mercury	100	101	85 - 115	1	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1743

Method: 7471A
Preparation: 7471A

MS Lab Sample ID:	720-381-40	Analysis Batch: 720-1807	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1743	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.04 g
Date Analyzed:	11/15/2005 1422		Final Weight/Volume: 50 mL
Date Prepared:	11/14/2005 1700		

MSD Lab Sample ID:	720-381-40	Analysis Batch: 720-1807	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1743	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.00 g
Date Analyzed:	11/15/2005 1423		Final Weight/Volume: 50 mL
Date Prepared:	11/14/2005 1700		

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	94	97	85 - 115	6	20		

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1743

Method: 7471A
Preparation: 7471A

MS Lab Sample ID:	720-381-42	Analysis Batch: 720-1807	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1743	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.03 g
Date Analyzed:	11/15/2005 1429		Final Weight/Volume: 50 mL
Date Prepared:	11/14/2005 1700		

MSD Lab Sample ID:	720-381-42	Analysis Batch: 720-1807	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1743	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.00 g
Date Analyzed:	11/15/2005 1430		Final Weight/Volume: 50 mL
Date Prepared:	11/14/2005 1700		

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	93	98	85 - 115	8	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1778

Method: 7471A
Preparation: 7471A

Lab Sample ID: MB 720-1778/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 0742
Date Prepared: 11/15/2005 1319

Analysis Batch: 720-1822
Prep Batch: 720-1778
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Mercury	ND		0.050

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1778

Method: 7471A
Preparation: 7471A

LCS Lab Sample ID: LCS 720-1778/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 0743
Date Prepared: 11/15/2005 1319

Analysis Batch: 720-1822
Prep Batch: 720-1778
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1778/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 0745
Date Prepared: 11/15/2005 1319

Analysis Batch: 720-1822
Prep Batch: 720-1778
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Mercury	103	96	85 - 115	6	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1778

Method: 7471A
Preparation: 7471A

MS Lab Sample ID:	720-381-78	Analysis Batch: 720-1822	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1778	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.00 g
Date Analyzed:	11/16/2005 0750		Final Weight/Volume: 50 mL
Date Prepared:	11/15/2005 1319		

MSD Lab Sample ID:	720-381-78	Analysis Batch: 720-1822	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1778	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.02 g
Date Analyzed:	11/16/2005 0751		Final Weight/Volume: 50 mL
Date Prepared:	11/15/2005 1319		

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	102	104	85 - 115	0	20		

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1778

Method: 7471A
Preparation: 7471A

MS Lab Sample ID:	720-381-94	Analysis Batch: 720-1822	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1778	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.01 g
Date Analyzed:	11/16/2005 0808		Final Weight/Volume: 50 mL
Date Prepared:	11/15/2005 1319		

MSD Lab Sample ID:	720-381-94	Analysis Batch: 720-1822	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1778	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.01 g
Date Analyzed:	11/16/2005 0812		Final Weight/Volume: 50 mL
Date Prepared:	11/15/2005 1319		

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	101	107	85 - 115	5	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1803

Method: 7471A
Preparation: 7471A

Lab Sample ID: MB 720-1803/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 0818
Date Prepared: 11/15/2005 1538

Analysis Batch: 720-1822
Prep Batch: 720-1803
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Mercury	ND		0.050

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1803

Method: 7471A
Preparation: 7471A

LCS Lab Sample ID: LCS 720-1803/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 0819
Date Prepared: 11/15/2005 1538

Analysis Batch: 720-1822
Prep Batch: 720-1803
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1803/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 0820
Date Prepared: 11/15/2005 1538

Analysis Batch: 720-1822
Prep Batch: 720-1803
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Mercury	104	102	85 - 115	1	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1803

Method: 7471A
Preparation: 7471A

MS Lab Sample ID:	720-381-105	Analysis Batch: 720-1822	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1803	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.03 g
Date Analyzed:	11/16/2005 0829		Final Weight/Volume: 50 mL
Date Prepared:	11/15/2005 1538		

MSD Lab Sample ID:	720-381-105	Analysis Batch: 720-1822	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1803	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.02 g
Date Analyzed:	11/16/2005 0830		Final Weight/Volume: 50 mL
Date Prepared:	11/15/2005 1538		

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	100	97	85 - 115	2	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1777

Method: 160.3
Preparation: N/A

Lab Sample ID: MB 720-1777/1
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/14/2005 1500
Date Prepared: N/A

Analysis Batch: 720-1777
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Result	Qual	RL
Percent Moisture	ND		0.010

Matrix Duplicate - Batch: 720-1777

Method: 160.3
Preparation: N/A

Lab Sample ID: 720-381-52
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/14/2005 1500
Date Prepared: N/A

Analysis Batch: 720-1777
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Sample Result/Qual	Result	RPD	Limit	Qual
Percent Moisture	15	15			

Calculations are performed before rounding to avoid round-off errors in calculated results.

Quality Control Results

Client: Weston Solutions,Inc

Job Number: 720-381-1

Matrix Duplicate - Batch: 720-1791

Method: 160.3
Preparation: N/A

Lab Sample ID: 720-381-26
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/14/2005 0301
Date Prepared: N/A

Analysis Batch: 720-1791
Prep Batch: N/A
Units: %

Instrument ID: Sartorius
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Sample Result/Qual	Result	RPD	Limit	Qual
Percent Moisture	14	13			

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1969

Method: 160.3
Preparation: N/A

Lab Sample ID: MB 720-1969/1
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 1430
Date Prepared: N/A

Analysis Batch: 720-1969
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Result	Qual	RL
Percent Moisture	ND		0.010

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1980

Method: 160.3
Preparation: N/A

Lab Sample ID: MB 720-1980/1
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 1500
Date Prepared: N/A

Analysis Batch: 720-1980
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Result	Qual	RL
Percent Moisture	ND		0.010

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-1982

Method: 160.3
Preparation: N/A

Lab Sample ID: MB 720-1982/1
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 1530
Date Prepared: N/A

Analysis Batch: 720-1982
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Result	Qual	RL
Percent Moisture	ND		0.010

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-1

Method Blank - Batch: 720-2301

Method: 160.3
Preparation: N/A

Lab Sample ID: MB 720-2301/1
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/28/2005 1540
Date Prepared: N/A

Analysis Batch: 720-2301
Prep Batch: N/A
Units: % by Wt

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Result	Qual	NONE
Percent Moisture	0.00010		

Matrix Duplicate - Batch: 720-2301

Method: 160.3
Preparation: N/A

Lab Sample ID: 720-381-56
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/28/2005 1540
Date Prepared: N/A

Analysis Batch: 720-2301
Prep Batch: N/A
Units: % by Wt

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Sample Result/Qual	Result	RPD	Limit	Qual
Percent Moisture	0.0000088	0.0000091			

Calculations are performed before rounding to avoid round-off errors in calculated results.

ONMENTAL PROTECTION AGENCY
Office of Enforcement

720-381

CHAIN OF CUSTODY RECORD

REGION 9
75 Hawthorne Street
San Francisco, California 94105

PROJ. NO.

PROJECT NAME: EBHH Edes Ave.

SAMPLERS: (Signature) M Smith

pg 1 of 8

REMARKS: 114405

STA. NO.	DATE	TIME	COMP.	GRAB	STATION LOCATION	NO. OF CONTAINERS	Asbestos	RCRA Metals	Pesticides	PAHs	TPH-d,o	HOLD
	11/9/05	0900		X	EBHH-1A	5	X	X		X	X	
		0925		X	EBHH-1B	4		X	X	X	X	
		0940		X	EBHH-1C	2		X	X			
		0940		X	EBHH-1D	2						X
		0945		X	EBHH-2B	1		X				
		0951		X	EBHH-2C	1		X				
		0951		X	EBHH-2D	1						X
		1102		X	EBHH-2A	2	X	X				
		1000		X	EBHH-3B	2		X	X			
		1000		X	EBHH-3C	2		X	X			
		1005		X	EBHH-3D	2						X
		1105		X	EBHH-3A	2	X	X				
		1015		X	EBHH-4B	1		X				
		1020		X	EBHH-4C			X	X	X		
		1020		X	EBHH-4D	1						X

Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
Mindy Smith M Smith	11/9/05 1730				
Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
Relinquished by: (Signature)	Date / Time	Received for Laboratory by: (Signature)	Date / Time	Remarks	
		[illegible]	11/9/05 17:30	TEMP 5°C	

Distribution: Original Accompanies Shipment; Copy to Coordinator Field Files

9 29076

ENVIRONMENTAL PROTECTION AGENCY
Office of Enforcement

720-381

CHAIN OF CUSTODY RECORD

REGION 9
75 Hawthorne Street
San Francisco, California 94105

PROJ. NO.

PROJECT NAME: EBHH Eden Ave.

SAMPLERS: (Signature) [signature]

pg 2 of 8

STA. NO.	DATE	TIME	COMP.	GRAB	STATION LOCATION	NO. OF CONTAINERS	Asbestos	RCRA Metals	Pesticides	PAHs	TPH-d,o	HOLD	REMARKS
													114405
	11/9/05	1110		X	EBHH-4A								
	11/9/05	1030		X	EBHH-5B	2		X	X				
		1030		X	EBHH-5C	2		X	X				
		1030		X	EBHH-5D	2						X	
		1115		X	EBHH-5A	2	X	X					MS/MSD RCRA Metals
		1045		X	EBHH-6B	1		X					
		1045		X	EBHH-6C	1		X					
		1055		X	EBHH-6D	1						X	
		1118		X	EBHH-~~7A~~ 6A	2	X	X					
		1125		X	EBHH-7B	2		X	X				
		1125		X	EBHH-7C	2		X	X				
		1130		X	EBHH-7D	2						X	
		1145		X	EBHH-7A	2	X	X					
		1135		X	EBHH-8B	1		X					
		1135		X	EBHH-8C	1		X					

Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
M. Smith [signature]	11/9/05 1730				
Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
Relinquished by: (Signature)	Date / Time	Received for Laboratory by: (Signature) [signature]	Date / Time 11/9/05 1730	Remarks Temp 5°C	

Distribution: Original Accompanies Shipment; Copy to Coordinator Field Files

9 29077

ENVIRONMENTAL PROTECTION AGENCY
Office of Enforcement

720-381

CHAIN OF CUSTODY RECORD

REGION 9
75 Hawthorne Street
San Francisco, California 94105

PROJ. NO.

PROJECT NAME: EBHH Eden Ave

SAMPLERS: (Signature)

Pg 3 of 8

STA. NO.	DATE	TIME	COMP.	GRAB	STATION LOCATION	NO. OF CONTAINERS	Asbestos	RCRA Metals	Pesticides	PCBs	TPH-d/o	HDLD	REMARKS
													114405
	11/9/05	1137		X	EBHH - 8D	1						X	
		1150		X	EBHH - 8A	2	X	X					
		1146		X	EBHH - 9B	2		X	X				
		1146		X	EBHH - 9C	2		X	X				
		1148		X	EBHH - 9D	2						X	
		1155		X	EBHH - 9A	2	X	X					
		1203		X	EBHH - 10B	1		X					
		1203		X	EBHH - 10C	1		X					
		1203		X	EBHH - 10D	4		H		X	X		
		1340		X	EBHH - 10A	3	X	X					MS/MSD RCRA Metals
		1208		X	EBHH - 11B	2		X	X				
		1215		X	EBHH - 11C	3		X	X				MS/MSD Pesticides
		1215		X	EBHH - 11D	2						X	
		1243		X	EBHH - 11A	2	X	X					
		1223		X	EBHH - 23	1		X					

Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
[illegible]	11/9/05 1730				
Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
Relinquished by: (Signature)	Date / Time	Received for Laboratory by: (Signature)	Date / Time	Remarks	
		[illegible]	11/9/05 17:30	Temp 5°C	

Distribution: Original Accompanies Shipment; Copy to Coordinator Field Files

9 29079

ENVIRONMENTAL PROTECTION AGENCY
Office of Enforcement

720-381

CHAIN OF CUSTODY RECORD

REGION 9
75 Hawthorne Street
San Francisco, California 94105-3901

PROJ. NO.

PROJECT NAME: EBHH Edes Ave

SAMPLERS: (Signature) [signature] [signature]

Pg 4 of 8

114405

STA. NO.	DATE	TIME	COMP.	GRAB	STATION LOCATION	NO. OF CONTAINERS	Asbestos	RCRA Metals	Pesticides	PAHs	TPH-d,o	HOLD	REMARKS
	11/9/05	1223		X	EBHH - 12C	1		X					
		1228		X	EBHH - 12D	1						X	
		1248		X	EBHH - 12A	2	X	X					
		1250		X	EBHH - 13B	2		X	X				
		1252		X	EBHH - 13C	2		X	X				
		1252		X	EBHH - 13D	2						X	
		1350		X	EBHH - 13A	2	X	X					
		1301		X	EBHH - 14B	4		X		X	X		
		1309		X	EBHH - 14C	1		X					
		1309		X	EBHH - 14D	1						X	
		1445		X	EBHH - 14A	5	X	X		X	X		
		1341		X	EBHH - 15B	2		X	X				
		1346		X	EBHH - 15C	3		X	X				MS/MSD RCRA Metals
		1346		X	EBHH - 15D	2						X	
		1450		X	EBHH - 15A	2	X	X					

Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
[signature]	11/9/05 1730				
Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
Relinquished by: (Signature)	Date / Time	Received for Laboratory by: (Signature) [signature]	Date / Time 11/9/05 17:30	Remarks Temp. 5°C	

Distribution: Original Accompanies Shipment; Copy to Coordinator Field Files

9 17019

ENVIRONMENTAL PROTECTION AGENCY
Office of Enforcement

720-381

CHAIN OF CUSTODY RECORD

REGION 9
75 Hawthorne Street
San Francisco, California 94105-3901

PROJ. NO.

PROJECT NAME: E.BHH Eden Ave

SAMPLERS: (Signature)

Pg 5 of 8

STA. NO.	DATE	TIME	COMP.	GRAB	STATION LOCATION	NO. OF CONTAINERS	Asbestos	RCRA Metals	Pesticides	PAHs	TPH-do	Hold	REMARKS
													114405
	11/9/05	1358		X	EBHH-16B	1		X					
		1402		X	EBHH-16C	1		X					
		1402		X	EBHH-16D	1					X		
		1455		X	EBHH-16A	2	X	X					
		1408		X	EBHH-17B	2		X	X				
		1411		X	EBHH-17C	2		X	X				
		1411		X	EBHH-17D	2					X		
		1500		X	EBHH-17A	2	X	X					
		1422		X	EBHH-18B	1		X					
		1422		X	EBHH-18C	1		X					
		1422		X	EBHH-18D	1	X				X		
		1505		X	EBHH-18A	2	X	X					
		1430		X	EBHH-19B	2		X	X				
		1430		X	EBHH-19C	2		X	X				
		1430		X	EBHH-19D	3		H	X	X			Hold samples for RCRA Metals

Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
	11/9/05 1730				
Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
Relinquished by: (Signature)	Date / Time	Received for Laboratory by: (Signature)	Date / Time	Remarks	
			11/9/05 1730	Temp. 5°C	

Distribution: Original Accompanies Shipment; Copy to Coordinator Field Files

9 17020

ENVIRONMENTAL PROTECTION AGENCY
Office of Enforcement

720-381

CHAIN OF CUSTODY RECORD

REGION 9
75 Hawthorne Street
San Francisco, California 94105-3901

PROJ. NO.

PROJECT NAME: EBHH Eden Ave

SAMPLERS: (Signature) [illegible] M Smith

Pg. 6 of 8

STA. NO.	DATE	TIME	COMP.	GRAB	STATION LOCATION	NO. OF CONTAINERS	Asbestos	RCRA Metals	Pesticides	PAHs	TPH-do	HD'D	REMARKS
													114405
	11/9/05	1530		X	EBHH-19A	2	X	X					
		1438		X	EBHH-20B	1		X					
		1438		X	EBHH-20C	2		X					MS/MSD RCRA Metals
		1438		X	EBHH-20D	1						X	
		1535		X	EBHH-20A	2	X	X					
		1450		X	EBHH-21B	2		X	X				
		1450		X	EBHH-21C	2		X	X				
		1450		X	EBHH-21D	2						X	
		1540		X	EBHH-21A	2	X	X					
		1510		X	EBHH-22B	1		X					
		1510		X	EBHH-22C	3		X		X	X		
		1510		X	EBHH-22D	1						X	
		1545		X	EBHH-22A	5	X	X		X	X		
		1515		X	EBHH-23B	2		X	X				
		1520		X	EBHH-23C	2		X	X				

Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
M. Smith [illegible]	11/9/05 1730				
Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
Relinquished by: (Signature)	Date / Time	Received for Laboratory by: (Signature) [illegible]	Date / Time	Remarks	
			11/9/05 1730	Temp. 5°C	

Distribution: Original Accompanies Shipment; Copy to Coordinator Field Files

9 17021

ENVIRONMENTAL PROTECTION AGENCY
Office of Enforcement

720-381

CHAIN OF CUSTODY RECORD

REGION 9
75 Hawthorne Street
San Francisco, California 94105-3901

PROJ. NO.

PROJECT NAME: EBHH Edes Ave

SAMPLERS: (Signature) M Smith

Pg. 7 of 8

STA. NO.	DATE	TIME	COMP.	GRAB	STATION LOCATION	NO. OF CONTAINERS	Asbestos	RCRA Metals	Pesticides	PAHs	TPH-d	HCLD	REMARKS
													114405
	11/9/05	1520		X	EBHH-23D	2						X	
		1550		X	EBHH-23A	2	X	X					
		1530		X	EBHH-24B	1		X					
		1530		X	EBHH-24C	2		X					MS/MSD RCRA Metals
		1530		X	EBHH-24D	1		~~X~~ ms				X	
		1615		X	EBHH-24A	2	X	X					
		1540		X	EBHH-26B	1		X					
		1548		X	EBHH-26C	3		X		X	X		
		1548		X	EBHH-26D	1						X	
		1630		X	EBHH-26A	5	X	X		X	X		
		1600		X	EBHH-27B	2		X	X				
		1605		X	EBHH-27C	2		X	X				
		1605		X	EBHH-27D	2						X	
		1650		X	EBHH-27A	2	X	X					
		1620		X	EBHH-25B	5	X	X	X	X	X		MS/MSD Pesticides

Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
M. Smith M Smith	11/9/05 1730				
Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
Relinquished by: (Signature)	Date / Time	Received for Laboratory by: (Signature)	Date / Time	Remarks	
			11/9/05 1730	TEMP. 5°C	

Distribution: Original Accompanies Shipment; Copy to Coordinator Field Files

9 17022

ENVIRONMENTAL PROTECTION AGENCY
Office of Enforcement

REGION 9
75 Hawthorne Street
San Francisco, California 94105-3901

CHAIN OF CUSTODY RECORD 720-381

PROJ. NO.

PROJECT NAME: EBHH Eckes Ave

SAMPLERS: (Signature)

Pg. 8 of 8

STA. NO.	DATE	TIME	COMP.	GRAB	STATION LOCATION	NO. OF CONTAINERS	Asbestos	RCRA Metals	Pesticides	PAHs	TPH-d,o	Hold	REMARKS
													114405
	11/9/05	1240		X	EBHH-51A	2	X	X					
		1501		X	EBHH-51B	2		X	X				
		1501		X	EBHH-51C	2		X	X				
		1600		X	EBHH-52A	2	X	X					
		1600		X	EBHH-52B	2		X	X				
		1600		X	EBHH-52C	2		X	X				
		1700		X	EBHH-53A	2	X	X					
		1515		X	EBHH-53B	2		X	X				
		1630		X	EBHH-54A	2	X	X					
		1600		X	EBHH-54B	2		X	X				

Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
R. Smith	11/9/05 1730				
Relinquished by: (Signature)	Date / Time	Received by: (Signature)	Relinquished by: (Signature)	Date / Time	Received by: (Signature)
Relinquished by: (Signature)	Date / Time	Received for Laboratory by: (Signature)	Date / Time	Remarks	
			11/9/05 17:30	Temp. 5°C	

Distribution: Original Accompanies Shipment; Copy to Coordinator Field Files

9 17023

LOGIN SAMPLE RECEIPT CHECK LIST

Client: Weston Solutions,Inc

Job Number: 720-381-1

Login Number: 381

Question	T/F/NA	Comment
Radioactivity either was not measured or, if measured, is at or below background	NA	
The cooler's custody seal, if present.	NA	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	True	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present	True	
Samples do not require splitting or compositing	True	

2595-544

19 3

STL San Francisco
1220 Quarry Lane
Pleasanton, CA 94566
Phone (925) 484-1019 Fax (925) 484-1096

Chain of Custody

SEVERN TRENT STL
Severn Trent Laboratories, Inc.

Client Contact	Project Manager: Sidhu, Surinder	Site Contact:	Date:
Shipping/Receiving@ Block Environmental	Tel/Fax:	Lab Contact:	Carrier No:
2451 Estand Way, [illegible]	Analysis Due Date		
Pleasant Hill, CA 94523	11/15/2025		
Phone: 925-682-7200	Shipped with Order No.: 720-49		
Fax			
Project Name: EBHH EDES AVE. OAKLAND			
Site			
PO #			

ICOC No: 720-49
Job Number: 720-381-1

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	Fill Cont	Filtered Sample	SUBCONTRACT/Asbestos 600/R	Special Instructions/Note
720-381-1	11/9/05	9:00		Solid	1		X	
720-381-8	11/9/05	11:02		Solid	1		X	
720-381-12	11/9/05	11:05		Solid	1		X	
720-381-16	11/9/05	11:10		Solid	1		X	
720-381-20	11/9/05	11:15		Solid	1		X	
720-381-24	11/9/05	11:18		Solid	1		X	
720-381-28	11/9/05	11:45		Solid	1		X	
720-381-32	11/9/05	11:50		Solid	1		X	
720-381-36	11/9/05	11:55		Solid	1		X	
720-381-40	11/9/05	13:40		Solid	1		X	
720-381-44	11/9/05	12:43		Solid	1		X	
720-381-48	11/9/05	12:48		Solid	1		X	

Possible Hazard Identification
☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison ☐ Unk

Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)
☐ Return To Client ☐ Disposal By Lab ☐ Archive For Months

Special Instructions/QC Requirements:

Relinquished by:	Company	Date/Time	Received by: [signature] 11/13	Company: FST
Relinquished by:	Company	Date/Time:	Received by:	Company
Relinquished by:	Company:	Date/Time:	Received by:	Company

Comments:

Page 1 of 3

STL San Francisco
1220 Quarry Lane
Pleasanton, CA 94566
Phone (925) 484-1919 Fax (925) 484-1096

Forensic

2595-544

Chain of Custody

SEVERN TRENT STL
Severn Trent Laboratories, Inc.

Client Contact: Shipping/Receiving@ Block Environmental 2451 Estand Way, Pleasant Hill, CA 94523 Phone: 925-682-7200 Fax:	Project Manager: Sidhu, Surinder Tel/Fax:	Site Contact: Lab Contact:	Date: Carrier No:
Project Name: EBHH EDES AVE OAKLAND Site: PO #:	Analysis Due Date 11/16/2005 Shipped with Order No: 720-40		

ICOC No. 720-48
Job Number 720-381-1

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	# of Cont.	SUBCONTRACT: Asbestos 600/R	Special Instructions/Note
720-381-52	11/9/05	13:50		Solid	1		
720-381-56	11/9/05	14:45		Solid	1		
720-381-60	11/9/05	14:50		Solid	1		
720-381-64	11/9/05	14:55		Solid	1		
720-381-68	11/9/05	15:00		Solid	1		
720-381-72	11/9/05	15:05		Solid	1		
720-381-76	11/9/05	15:30		Solid	1		
720-381-80	11/9/05	15:35		Solid	1		
720-381-84	11/9/05	15:40		Solid	1		
720-381-88	11/9/05	15:45		Solid	1		
720-381-92	11/9/05	16:50		Solid	1		
720-381-96	11/9/05	16:15		Solid	1		

Possible Hazard Identification
☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison ☐ Unkn

Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)
☐ Return To Client ☐ Disposal By Lab ☐ Archive For ____ Months

Special Instructions/QC Requirements:

Relinquished by: Joan Mullen	Company: STL SF	Date/Time: 11/10/05 1055	Received by: [signature]	Company:
Relinquished by: [signature]	Company: STL-SF	Date/Time: 11/11/05 1520	Received by: Donna Conca	Company: Fan
Relinquished by:	Company:	Date/Time:	Received by:	Company:

Comments:

Page 2 of 3

STL San Francisco
1220 Quarry Lane
Pleasanton, CA 94566
Phone (925) 484-1919 Fax (925) 484-1096

Forensic

Chain of Custody

SEVERN TRENT STL
Severn Trent Laboratories, Inc.

Client Contact Shipping/Receiving@ Block Environmental 2451 Estand Way Pleasant Hill, CA 94523 Phone: 925-682-7200 Fax: Project Name: FBHH EDES AVE OAKLAND Site PO #	Project Manager: Sidhu, Surinder Tel/Fax: Analysis Due Date 11/16/2005 Shipped with Order No. 720-49	Site Contact: Lab Contact:	Date: Carrier No:

ICOC No: 720-48
Job Number: 720-381-1

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	# of Cont.	SUBCONTRACT: Asbestos 600/R	Special Instructions/Note:
720-381-100	11/9/05	16:30		Solid	1		
720-381-104	11/9/05	16:50		Solid	1		
720-381-106	11/9/05	12:40		Solid	1		
720-381-109	11/9/05	15:00		Solid	1		
720-381-112	11/9/05	17:00		Solid	1		
720-381-114	11/9/05	16:30		Solid	1		

Possible Hazard Identification
☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison ☐ Unk

Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)
☐ Return To Client ☐ Disposal By Lab ☐ Archive For _____ Months

Special Instructions/QC Requirements:

Relinquished by: [signature]	Company: STL SF	Date/Time: 11/11/05 1014	Received by: [signature]	Company: STL SF
Relinquished by: [signature]	Company: STL-SF	Date/Time: 1525	Received by: [signature]	Company: [illegible]
Relinquished by:	Company:	Date/Time:	Received by:	Company:

Comments:

Page 3 of 3

Forensic Analytical

Final Report

Bulk Asbestos Analysis

(EPA Method 600/R-93-116, Visual Area Estimation)

STL San Francisco
Surinder Sidhu

1220 Quarry Ln
Pleasanton, CA 94566-4756

Client ID: 2595
Report Number: B078973
Date Received: 11/13/05
Date Analyzed: 11/21/05
Date Printed: 11/21/05
First Reported: 11/21/05

Job ID/Site: 720-381-1 - EBHH Edes Ave., Oakland

Date(s) Collected: 11/09/2005

FASI Job ID: 2595-544
Total Samples Submitted: 30
Total Samples Analyzed: 30

Sample ID	Lab Number	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer
720-381-1	10470555						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-8	10470556						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-12	10470557						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-16	10470558						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-20	10470559						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-24	10470560						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-28	10470561						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-32	10470562						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					

Forensic Analytical

Report Number: B078973
Client Name: STL San Francisco **Date Printed:** 11/21/05

Sample ID	Lab Number	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer
720-381-36	10470563						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-40	10470564						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-44	10470565						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-48	10470566						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-52	10470567						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-56	10470568						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-60	10470569						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-64	10470570						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-68	10470571						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-72	10470572						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					

Forensic Analytical

Report Number: B078973
Client Name: STL San Francisco **Date Printed:** 11/21/05

Sample ID	Lab Number	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer
720-381-76	10470573						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-80	10470574						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-84	10470575						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-88	10470576						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-92	10470577						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-96	10470578						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-100	10470579						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-104	10470580						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-106	10470581						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-109	10470582						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					

Forensic Analytical

Report Number: B078973
Client Name: STL San Francisco **Date Printed:** 11/21/05

Sample ID	Lab Number	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer
720-381-112	10470583						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-381-114	10470584						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					

James Flores, Laboratory Supervisor, Hayward Laboratory

Note: Limit of Quantification ('LOQ') = 1%. 'Trace' denotes the presence of asbestos below the LOQ. 'ND' = 'None Detected'.

Analytical results and reports are generated by Forensic Analytical at the request of and for the exclusive use of the person or entity (client) named on such report. Results, reports or copies of same will not be released by Forensic Analytical to any third party without prior written request from client. This report applies only to the sample(s) tested. Supporting laboratory documentation is available upon request. This report must not be reproduced except in full, unless approved by Forensic Analytical. The client is solely responsible for the use and interpretation of test results and reports requested from Forensic Analytical. This report must not be used by the client to claim product endorsement by NVLAP or any other agency of the U.S. Government. Forensic Analytical is not able to assess the degree of hazard resulting from materials analyzed. Forensic Analytical reserves the right to dispose of all samples after a period of thirty (30) days, according to all state and federal guidelines, unless otherwise specified. All samples were received in acceptable condition unless otherwise noted.

3777 Depot Road, Suite 409, Hayward, California 94545-2761 Telephone: 510/887-8828 800/827-FASI Fax: 510/887-4218

ANALYTICAL REPORT

Job Number: 720-381-2

Job Description: EBHH EDES AVE OAKLAND

For:

Weston Solutions,Inc
1575 Treat Blvd Suite 212
Walnut Creek, CA 94598

Attention: Mr. Tom Fortner

Surinder Sidhu
Project Manager I
ssidhu@stl-inc.com
11/30/2005

METHOD SUMMARY

Client: Weston Solutions,Inc

Job Number: 720-381-2

Description	Lab Location	Method	Preparation Method
Matrix: Solid			
Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)	STL-SF	SW846 8270C	
Ultrasonic Extraction	STL-SF		SW846 3550B
Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)	STL-SF	SW846 8015B	
Ultrasonic Extraction	STL-SF		SW846 3550B

LAB REFERENCES:

STL-SF = STL-San Francisco

METHOD REFERENCES:

SW846 - "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

SAMPLE SUMMARY

Client: Weston Solutions,Inc

Job Number: 720-381-2

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
720-381-14	EBHH-4C	Solid	11/09/2005 1020	11/09/2005 1730
720-381-40	EBHH-10A	Solid	11/09/2005 1340	11/09/2005 1730
720-381-72	EBHH-18A	Solid	11/09/2005 1505	11/09/2005 1730

Client: Weston Solutions,Inc

Job Number: 720-381-2

Client Sample ID: **EBHH-10A**

Lab Sample ID:	720-381-40			Date Sampled:	11/09/2005 1340
Client Matrix:	Solid	% Moisture:	12.4	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2253		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-2189		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.13 g
Date Analyzed:	11/26/2005 2146			Final Weight/Volume:	1 mL
Date Prepared:	11/23/2005 1328			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		57
Acenaphthene		ND		57
Acenaphthylene		ND		57
Fluorene		ND		57
Phenanthrene		ND		57
Anthracene		ND		57
Benzo[a]anthracene		67		57
Chrysene		100		57
Benzo[a]pyrene		88		57
Benzo[b]fluoranthene		130		57
Benzo[k]fluoranthene		ND		57
Benzo[g,h,i]perylene		97		57
Indeno[1,2,3-cd]pyrene		100		57
Fluoranthene		160		57
Pyrene		150		57
Dibenz(a,h)anthracene		ND		57

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	75	30 - 115
Terphenyl-d14	78	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-2

Client Sample ID: **EBHH-18A**

Lab Sample ID:	720-381-72			Date Sampled:	11/09/2005 1505
Client Matrix:	Solid	% Moisture:	12.0	Date Received:	11/09/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2253		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-2189		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.10 g
Date Analyzed:	11/26/2005 2214			Final Weight/Volume:	1 mL
Date Prepared:	11/23/2005 1328			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		57
Acenaphthene		ND		57
Acenaphthylene		ND		57
Fluorene		ND		57
Phenanthrene		93		57
Anthracene		ND		57
Benzo[a]anthracene		62		57
Chrysene		85		57
Benzo[a]pyrene		82		57
Benzo[b]fluoranthene		120		57
Benzo[k]fluoranthene		ND		57
Benzo[g,h,i]perylene		94		57
Indeno[1,2,3-cd]pyrene		120		57
Fluoranthene		130		57
Pyrene		130		57
Dibenz(a,h)anthracene		ND		57

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	57	30 - 115
Terphenyl-d14	81	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-381-2

Client Sample ID: **EBHH-4C**

Lab Sample ID:	720-381-14			Date Sampled:	11/09/2005 1020
Client Matrix:	Solid	% Moisture:	12.7	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-2299		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-2179		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.01 g
Date Analyzed:	11/24/2005 1316			Final Weight/Volume:	5 mL
Date Prepared:	11/23/2005 1229			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		ND		1.1
Motor Oil Range Organics [C24-C36]		ND		57

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	78	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-2

Client Sample ID: **EBHH-10A**

Lab Sample ID:	720-381-40			Date Sampled:	11/09/2005 1340
Client Matrix:	Solid	% Moisture:	12.4	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-2299		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-2179		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.05 g
Date Analyzed:	11/24/2005 1033			Final Weight/Volume:	5 mL
Date Prepared:	11/23/2005 1229			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		24		1.1
Motor Oil Range Organics [C24-C36]		130		57

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	87	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-381-2

Client Sample ID: **EBHH-18A**

Lab Sample ID:	720-381-72			Date Sampled:	11/09/2005 1505
Client Matrix:	Solid	% Moisture:	12.0	Date Received:	11/09/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-2299		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-2179		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.11 g
Date Analyzed:	11/24/2005 1154			Final Weight/Volume:	5 mL
Date Prepared:	11/23/2005 1229			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		25		1.1
Motor Oil Range Organics [C24-C36]		180		57

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	85	60 - 130

DATA REPORTING QUALIFIERS

Lab Section	Qualifier	Description

Client: Weston Solutions,Inc

Job Number: 720-381-2

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC/MS Semi VOA				
Prep Batch: 720-2189				
LCS 720-2189/2-A	Lab Control Spike	Solid	3550B	
LCSD 720-2189/3-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-2189/1-A	Method Blank	Solid	3550B	
720-381-40	EBHH-10A	Solid	3550B	
720-381-72	EBHH-18A	Solid	3550B	
720-608-A-1-B MS	Matrix Spike	Solid	3550B	
720-608-A-1-C MSD	Matrix Spike Duplicate	Solid	3550B	
Analysis Batch:720-2253				
LCS 720-2189/2-A	Lab Control Spike	Solid	8270C	720-2189
LCSD 720-2189/3-A	Lab Control Spike Duplicate	Solid	8270C	720-2189
MB 720-2189/1-A	Method Blank	Solid	8270C	720-2189
720-381-40	EBHH-10A	Solid	8270C	720-2189
720-381-72	EBHH-18A	Solid	8270C	720-2189
720-608-A-1-B MS	Matrix Spike	Solid	8270C	720-2189
720-608-A-1-C MSD	Matrix Spike Duplicate	Solid	8270C	720-2189
GC Semi VOA				
Prep Batch: 720-2179				
LCS 720-2179/2-A	Lab Control Spike	Solid	3550B	
LCSD 720-2179/3-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-2179/1-A	Method Blank	Solid	3550B	
720-381-14	EBHH-4C	Solid	3550B	
720-381-40	EBHH-10A	Solid	3550B	
720-381-72	EBHH-18A	Solid	3550B	
720-595-G-1-B MS	Matrix Spike	Solid	3550B	
720-595-G-1-C MSD	Matrix Spike Duplicate	Solid	3550B	
Analysis Batch:720-2299				
LCS 720-2179/2-A	Lab Control Spike	Solid	8015B	720-2179
LCSD 720-2179/3-A	Lab Control Spike Duplicate	Solid	8015B	720-2179
MB 720-2179/1-A	Method Blank	Solid	8015B	720-2179
720-381-14	EBHH-4C	Solid	8015B	720-2179
720-381-40	EBHH-10A	Solid	8015B	720-2179
720-381-72	EBHH-18A	Solid	8015B	720-2179
720-595-G-1-B MS	Matrix Spike	Solid	8015B	720-2179
720-595-G-1-C MSD	Matrix Spike Duplicate	Solid	8015B	720-2179

Client: Weston Solutions,Inc

Job Number: 720-381-2

Method Blank - Batch: 720-2189

Method: 8270C
Preparation: 3550B

Lab Sample ID: MB 720-2189/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/26/2005 1831
Date Prepared: 11/23/2005 1328

Analysis Batch: 720-2253
Prep Batch: 720-2189
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.20 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result	Qual	RL
Naphthalene	ND		5.0
Acenaphthene	ND		5.0
Acenaphthylene	ND		5.0
Fluorene	ND		5.0
Phenanthrene	ND		5.0
Anthracene	ND		5.0
Benzo[a]anthracene	ND		5.0
Chrysene	ND		5.0
Benzo[a]pyrene	ND		5.0
Benzo[b]fluoranthene	ND		5.0
Benzo[k]fluoranthene	ND		5.0
Benzo[g,h,i]perylene	ND		5.0
Indeno[1,2,3-cd]pyrene	ND		5.0
Fluoranthene	ND		5.0
Pyrene	ND		5.0
Dibenz(a,h)anthracene	ND		5.0

Surrogate	% Rec	Acceptance Limits
2-Fluorobiphenyl	84	30 - 115
Terphenyl-d14	88	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-2

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-2189

Method: 8270C
Preparation: 3550B

LCS Lab Sample ID: LCS 720-2189/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/26/2005 1859
Date Prepared: 11/23/2005 1328

Analysis Batch: 720-2253
Prep Batch: 720-2189
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\112605
Initial Weight/Volume: 30.10 g
Final Weight/Volume: 1 mL
Injection Volume:

LCSD Lab Sample ID: LCSD 720-2189/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/26/2005 1927
Date Prepared: 11/23/2005 1328

Analysis Batch: 720-2253
Prep Batch: 720-2189
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\112605\l
Initial Weight/Volume: 30.16 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Naphthalene	66	72	21 - 133	9	35		
Acenaphthene	65	70	47 - 145	6	35		
Acenaphthylene	72	76	33 - 145	5	35		
Fluorene	76	81	59 - 121	6	35		
Phenanthrene	80	83	10 - 130	4	35		
Anthracene	75	70	27 - 133	7	35		
Benzo[a]anthracene	85	89	33 - 143	4	35		
Chrysene	73	75	17 - 168	3	35		
Benzo[a]pyrene	77	82	17 - 163	6	35		
Benzo[b]fluoranthene	82	86	24 - 159	6	35		
Benzo[k]fluoranthene	79	83	11 - 162	5	35		
Benzo[g,h,i]perylene	76	80	9 - 219	5	35		
Indeno[1,2,3-cd]pyrene	85	90	9 - 171	6	35		
Fluoranthene	79	81	26 - 137	3	35		
Pyrene	76	81	52 - 115	6	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
2-Fluorobiphenyl	74	81	30 - 115
Terphenyl-d14	83	89	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-2

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-2189

Method: 8270C
Preparation: 3550B

MS Lab Sample ID:	720-608-A-1-B MS	Analysis Batch: 720-2253	Instrument ID: Sat 2K2
Client Matrix:	Solid	Prep Batch: 720-2189	Lab File ID: c:\saturnws\data\200511\
Dilution:	1.0		Initial Weight/Volume: 30.03 g
Date Analyzed:	11/26/2005 2050		Final Weight/Volume: 1 mL
Date Prepared:	11/23/2005 1328		Injection Volume:

MSD Lab Sample ID:	720-608-A-1-C MSD	Analysis Batch: 720-2253	Instrument ID: Sat 2K2
Client Matrix:	Solid	Prep Batch: 720-2189	Lab File ID: c:\saturnws\data\200511\11
Dilution:	1.0		Initial Weight/Volume: 30.15 g
Date Analyzed:	11/26/2005 2118		Final Weight/Volume: 1 mL
Date Prepared:	11/23/2005 1328		Injection Volume:

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Naphthalene	99	117	21 - 133	16	35		
Acenaphthene	61	68	47 - 145	10	35		
Acenaphthylene	60	71	33 - 145	16	35		
Fluorene	69	75	59 - 121	8	35		
Phenanthrene	73	77	10 - 130	4	35		
Anthracene	67	70	27 - 133	4	35		
Benzo[a]anthracene	73	82	33 - 143	11	35		
Chrysene	63	70	17 - 168	10	35		
Benzo[a]pyrene	72	78	17 - 163	8	35		
Benzo[b]fluoranthene	73	79	24 - 159	7	35		
Benzo[k]fluoranthene	70	75	11 - 162	7	35		
Benzo[g,h,i]perylene	64	73	9 - 219	13	35		
Indeno[1,2,3-cd]pyrene	75	86	9 - 171	13	35		
Fluoranthene	70	73	26 - 137	3	35		
Pyrene	71	76	52 - 115	6	35		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
2-Fluorobiphenyl	64	74	30 - 115
Terphenyl-d14	75	79	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-2

Method Blank - Batch: 720-2179

Method: 8015B
Preparation: 3550B

Lab Sample ID: MB 720-2179/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/23/2005 1843
Date Prepared: 11/23/2005 1229

Analysis Batch: 720-2299
Prep Batch: 720-2179
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.41 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Diesel Range Organics [C10-C28]	ND		0.99
Motor Oil Range Organics [C24-C36]	ND		49

Surrogate	% Rec	Acceptance Limits
o-Terphenyl	71	60 - 130

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-2179

Method: 8015B
Preparation: 3550B

LCS Lab Sample ID: LCS 720-2179/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/23/2005 1910
Date Prepared: 11/23/2005 1229

Analysis Batch: 720-2299
Prep Batch: 720-2179
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.20 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-2179/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/23/2005 1937
Date Prepared: 11/23/2005 1229

Analysis Batch: 720-2299
Prep Batch: 720-2179
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.23 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Diesel Range Organics [C10-C28]	91	87	60 - 130	5	30		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
o-Terphenyl	78	73	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-381-2

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-2179

Method: 8015B
Preparation: 3550B

MS Lab Sample ID:	720-595-G-1-B MS	Analysis Batch: 720-2299	Instrument ID: HP DRO3
Client Matrix:	Solid	Prep Batch: 720-2179	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 30.03 g
Date Analyzed:	11/23/2005 2153		Final Weight/Volume: 5 mL
Date Prepared:	11/23/2005 1229		Injection Volume:
			Column ID: PRIMARY

MSD Lab Sample ID:	720-595-G-1-C MSD	Analysis Batch: 720-2299	Instrument ID: HP DRO3
Client Matrix:	Solid	Prep Batch: 720-2179	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 30.31 g
Date Analyzed:	11/23/2005 2220		Final Weight/Volume: 5 mL
Date Prepared:	11/23/2005 1229		Injection Volume:
			Column ID: PRIMARY

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Diesel Range Organics [C10-C28]	69	72	60 - 130	4	30		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
o-Terphenyl	67	69	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

LOGIN SAMPLE RECEIPT CHECK LIST

Client: Weston Solutions,Inc

Job Number: 720-381-2

Login Number: 381

Question	T/F/NA	Comment
Radioactivity either was not measured or, if measured, is at or below background	NA	
The cooler's custody seal, if present.	NA	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	True	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present	True	
Samples do not require splitting or compositing	True	

Forensic Analytical

Final Report

Bulk Asbestos Analysis

(EPA Method 600/R-93-116, Visual Area Estimation)

STL San Francisco
Surinder Sidhu

1220 Quarry Ln
Pleasanton, CA 94566-4756

Client ID: 2595
Report Number: B078913
Date Received: 11/11/05
Date Analyzed: 11/18/05
Date Printed: 11/18/05
First Reported: 11/18/05

Job ID/Site: 720-401-1

Date(s) Collected: 11/10/2005

FASI Job ID: 2595-543
Total Samples Submitted: 10
Total Samples Analyzed: 10

Sample ID	Lab Number	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer
720-401-7	10469916						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-401-9	10469917						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-401-13	10469918						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-401-15	10469919						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-401-17	10469920						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-401-21	10469921						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-401-22	10469922						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-401-25	10469923						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					

Forensic Analytical

Report Number: B078913
Client Name: STL San Francisco **Date Printed:** 11/18/05

Sample ID	Lab Number	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer
720-401-28	10469924						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					
720-401-31	10469925						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components: Cellulose (Trace)		**Asbestos (ND)**					

James Flores, Laboratory Supervisor, Hayward Laboratory

Note: Limit of Quantification ('LOQ') = 1%. 'Trace' denotes the presence of asbestos below the LOQ. 'ND' = 'None Detected'.

Analytical results and reports are generated by Forensic Analytical at the request of and for the exclusive use of the person or entity (client) named on such report. Results, reports or copies of same will not be released by Forensic Analytical to any third party without prior written request from client. This report applies only to the sample(s) tested. Supporting laboratory documentation is available upon request. This report must not be reproduced except in full, unless approved by Forensic Analytical. The client is solely responsible for the use and interpretation of test results and reports requested from Forensic Analytical. This report must not be used by the client to claim product endorsement by NVLAP or any other agency of the U.S. Government. Forensic Analytical is not able to assess the degree of hazard resulting from materials analyzed. Forensic Analytical reserves the right to dispose of all samples after a period of thirty (30) days, according to all state and federal guidelines, unless otherwise specified. All samples were received in acceptable condition unless otherwise noted.

3777 Depot Road, Suite 409, Hayward, California 94545-2761 Telephone: 510/887-8828 800/827-FASI Fax: 510/887-4218

ANALYTICAL REPORT

Job Number: 720-401-1

Job Description: EBHH EDES AVE OAKLAND

For:

Weston Solutions,Inc
1575 Treat Blvd Suite 212
Use the other WESTON SOLUTIONS, CA 94598

Attention: Mr. Tom Fortner

Surinder Sidhu

Surinder Sidhu
Project Manager I
ssidhu@stl-inc.com
11/30/2005

METHOD SUMMARY

Client: Weston Solutions,Inc

Job Number: 720-401-1

Description	Lab Location	Method	Preparation Method
Matrix: Solid			
Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)	STL-SF	SW846 8270C	
Ultrasonic Extraction	STL-SF		SW846 3550B
Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)	STL-SF	SW846 8015B	
Ultrasonic Extraction	STL-SF		SW846 3550B
Organochlorine Pesticides by Gas Chromatography	STL-SF	SW846 8081A	
Ultrasonic Extraction	STL-SF		SW846 3550B
Inductively Coupled Plasma - Atomic Emission Spectrometry	STL-SF	SW846 6010B	
Acid Digestion of Sediments, Sludges, and Soils	STL-SF		SW846 3050B
Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)	STL-SF	SW846 7471A	
Mercury in Solid or Semi-Solid Waste (Manual	STL-SF		SW846 7471A
Percent Moisture	STL-SF	EPA 160.3	
Asbestos	STL-SF	EPA	
Matrix: Water			
Volatile Organic Compounds by GC/MS (Low Level)	STL-SF	SW846 8260B	
Purge-and-Trap	STL-SF		SW846 5030B
Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)	STL-SF	SW846 8270C	
Separatory Funnel Liquid-Liquid Extraction	STL-SF		SW846 3510C
Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)	STL-SF	SW846 8015B	
Separatory Funnel Liquid-Liquid Extraction	STL-SF		SW846 3510C
Organochlorine Pesticides by Gas Chromatography	STL-SF	SW846 8081A	
Separatory Funnel Liquid-Liquid Extraction	STL-SF		SW846 3510C
Inductively Coupled Plasma - Atomic Emission Spectrometry	STL-SF	SW846 6010B	
Acid Digestion of Aqueous Samples and Extracts	STL-SF		SW846 3010A
Mercury in Liquid Waste (Manual Cold Vapor Technique)	STL-SF	SW846 7470A	
Mercury in Liquid Waste (Manual Cold Vapor	STL-SF		SW846 7470A

LAB REFERENCES:

STL-SF = STL-San Francisco

METHOD SUMMARY

Client: Weston Solutions,Inc

Job Number: 720-401-1

METHOD REFERENCES:

EPA - US Environmental Protection Agency

SW846 - "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986
And Its Updates.

SAMPLE SUMMARY

Client: Weston Solutions,Inc

Job Number: 720-401-1

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
720-401-1	EBHH-EB-1	Water	11/10/2005 0726	11/10/2005 1730
720-401-2	EBHH-25C	Solid	11/10/2005 0734	11/10/2005 1730
720-401-4	EBHH-28B	Solid	11/10/2005 0730	11/10/2005 1730
720-401-5	EBHH-28C	Solid	11/10/2005 0730	11/10/2005 1730
720-401-6	EBHH-28D	Solid	11/10/2005 0730	11/10/2005 1730
720-401-7	EBHH-28A	Solid	11/10/2005 1005	11/10/2005 1730
720-401-8	EBHH-29B	Solid	11/10/2005 0745	11/10/2005 1730
720-401-9	EBHH-29A	Solid	11/10/2005 1010	11/10/2005 1730
720-401-10	EBHH-30C	Solid	11/10/2005 0800	11/10/2005 1730
720-401-11	EBHH-30D	Solid	11/10/2005 0800	11/10/2005 1730
720-401-12	EBHH-31D	Solid	11/10/2005 0815	11/10/2005 1730
720-401-13	EBHH-31A	Solid	11/10/2005 1025	11/10/2005 1730
720-401-14	EBHH-33C	Solid	11/10/2005 0835	11/10/2005 1730
720-401-15	EBHH-33A	Solid	11/10/2005 1035	11/10/2005 1730
720-401-16	EBHH-35C	Solid	11/10/2005 0910	11/10/2005 1730
720-401-17	EBHH-35A	Solid	11/10/2005 1045	11/10/2005 1730
720-401-18	EBHH-34D	Solid	11/10/2005 0845	11/10/2005 1730
720-401-19	EBHH-36C	Solid	11/10/2005 0925	11/10/2005 1730
720-401-20	EBHH-37B	Solid	11/10/2005 1048	11/10/2005 1730
720-401-21	EBHH-37A	Solid	11/10/2005 1330	11/10/2005 1730
720-401-22	EBHH-25A	Solid	11/10/2005 0955	11/10/2005 1730
720-401-23	EBHH-38B	Solid	11/10/2005 1100	11/10/2005 1730
720-401-24	EBHH-39B	Solid	11/10/2005 1114	11/10/2005 1730
720-401-25	EBHH-39A	Solid	11/10/2005 1345	11/10/2005 1730
720-401-26	EBHH-40C	Solid	11/10/2005 1124	11/10/2005 1730
720-401-27	EBHH-41B	Solid	11/10/2005 1134	11/10/2005 1730
720-401-28	EBHH-41A	Solid	11/10/2005 1355	11/10/2005 1730
720-401-29	EBHH-42D	Solid	11/10/2005 1150	11/10/2005 1730
720-401-30	EBHH-43B	Solid	11/10/2005 1201	11/10/2005 1730
720-401-31	EBHH-43A	Solid	11/10/2005 1410	11/10/2005 1730
720-401-32	EBHH-44B	Solid	11/10/2005 1218	11/10/2005 1730
720-401-33	EBHH-55A	Solid	11/10/2005 1400	11/10/2005 1730
720-401-34	EBHH-56A	Solid	11/10/2005 1430	11/10/2005 1730
720-401-35	EBHH-GW-2	Water	11/10/2005 1328	11/10/2005 1730
720-401-36	EBHH-GW-1	Water	11/10/2005 1446	11/10/2005 1730

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-GW-1**

Lab Sample ID: 720-401-36
Client Matrix: Water

Date Sampled: 11/10/2005 1446
Date Received: 11/10/2005 1730

8260B Volatile Organic Compounds by GC/MS (Low Level)

Method: 8260B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/19/2005 1538
Date Prepared: 11/19/2005 1538

Analysis Batch: 720-2028

Instrument ID: Varian 3900F
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 40 mL
Final Weight/Volume: 40 mL

Analyte	Result (ug/L)	Qualifier	RL
1,1-Dichloroethene	ND		0.50
1,1-Dichloroethane	ND		0.50
Dichlorodifluoromethane	ND		0.50
Vinyl chloride	ND		0.50
Chloroethane	ND		1.0
Trichlorofluoromethane	ND		1.0
Methylene Chloride	ND		5.0
trans-1,2-Dichloroethene	ND		0.50
cis-1,2-Dichloroethene	ND		0.50
Chloroform	ND		1.0
1,1,1-Trichloroethane	ND		0.50
Carbon tetrachloride	ND		0.50
1,2-Dichloroethane	ND		0.50
Trichloroethene	1.5		0.50
1,2-Dichloropropane	ND		0.50
Dichlorobromomethane	ND		0.50
trans-1,3-Dichloropropene	ND		0.50
cis-1,3-Dichloropropene	ND		0.50
1,1,2-Trichloroethane	ND		0.50
Tetrachloroethene	0.87		0.50
Chlorodibromomethane	ND		0.50
Chlorobenzene	ND		0.50
Bromoform	ND		1.0
1,1,2,2-Tetrachloroethane	ND		0.50
1,3-Dichlorobenzene	ND		0.50
1,4-Dichlorobenzene	ND		0.50
1,2-Dichlorobenzene	ND		0.50
Chloromethane	1.0		1.0
Bromomethane	ND		1.0
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		0.50
EDB	ND		0.50
1,2,4-Trichlorobenzene	ND		1.0

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-EB-1**

Lab Sample ID: 720-401-1
Client Matrix: Water

Date Sampled: 11/10/2005 0726
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/15/2005 1516
Date Prepared: 11/11/2005 1325

Analysis Batch: 720-1849
Prep Batch: 720-1673

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 860 mL
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result (ug/L)	Qualifier	RL
Naphthalene	ND		0.12
Acenaphthene	ND		0.12
Acenaphthylene	ND		0.12
Fluorene	ND		0.12
Phenanthrene	ND		0.12
Anthracene	ND		0.12
Benzo[a]anthracene	ND		0.23
Chrysene	ND		0.12
Benzo[a]pyrene	ND		0.12
Benzo[b]fluoranthene	ND		0.12
Benzo[k]fluoranthene	ND		0.12
Benzo[g,h,i]perylene	ND		0.12
Indeno[1,2,3-cd]pyrene	ND		0.12
Fluoranthene	ND		0.12
Pyrene	ND		0.12
Dibenz(a,h)anthracene	ND		1.2

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	69	43 - 116
Terphenyl-d14	83	33 - 141

Analytical Data

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-28D**

Lab Sample ID: 720-401-6
Client Matrix: Solid

Date Sampled: 11/10/2005 0730
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/25/2005 1716
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2238
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.21 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.0
Acenaphthene		ND		5.0
Acenaphthylene		ND		5.0
Fluorene		ND		5.0
Phenanthrene		ND		5.0
Anthracene		ND		5.0
Benzo[a]anthracene		ND		5.0
Chrysene		ND		5.0
Benzo[a]pyrene		ND		5.0
Benzo[b]fluoranthene		ND		5.0
Benzo[k]fluoranthene		ND		5.0
Benzo[g,h,i]perylene		ND		5.0
Indeno[1,2,3-cd]pyrene		ND		5.0
Fluoranthene		ND		5.0
Pyrene		ND		5.0
Dibenz(a,h)anthracene		ND		5.0

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	72	30 - 115
Terphenyl-d14	86	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-28A**

Lab Sample ID: 720-401-7
Client Matrix: Solid

Date Sampled: 11/10/2005 1005
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 10
Date Analyzed: 11/25/2005 1743
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2238
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.21 g
Final Weight/Volume: 5 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		250
Acenaphthene		ND		250
Acenaphthylene		ND		250
Fluorene		ND		250
Phenanthrene		ND		250
Anthracene		ND		250
Benzo[a]anthracene		ND		250
Chrysene		ND		250
Benzo[a]pyrene		ND		250
Benzo[b]fluoranthene		ND		250
Benzo[k]fluoranthene		ND		250
Benzo[g,h,i]perylene		ND		250
Indeno[1,2,3-cd]pyrene		ND		250
Fluoranthene		ND		250
Pyrene		ND		250
Dibenz(a,h)anthracene		ND		250

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	70	30 - 115
Terphenyl-d14	87	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-29B**

Lab Sample ID: 720-401-8
Client Matrix: Solid

Date Sampled: 11/10/2005 0745
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/25/2005 1811
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2238
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.11 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		8.0		5.0
Acenaphthene		ND		5.0
Acenaphthylene		32		5.0
Fluorene		8.7		5.0
Phenanthrene		190		5.0
Anthracene		18		5.0
Benzo[a]anthracene		120		5.0
Chrysene		160		5.0
Benzo[a]pyrene		170		5.0
Benzo[b]fluoranthene		210		5.0
Benzo[k]fluoranthene		75		5.0
Benzo[g,h,i]perylene		130		5.0
Indeno[1,2,3-cd]pyrene		200		5.0
Fluoranthene		250		5.0
Pyrene		260		5.0
Dibenz(a,h)anthracene		35		5.0

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	74	30 - 115
Terphenyl-d14	84	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-29A**

Lab Sample ID:	720-401-9			Date Sampled:	11/10/2005 1010
Client Matrix:	Solid	% Moisture:	7.9	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2238		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1676		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.12 g
Date Analyzed:	11/25/2005 1839			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1442			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		27
Acenaphthene		ND		27
Acenaphthylene		ND		27
Fluorene		ND		27
Phenanthrene		68		27
Anthracene		ND		27
Benzo[a]anthracene		77		27
Chrysene		84		27
Benzo[a]pyrene		110		27
Benzo[b]fluoranthene		120		27
Benzo[k]fluoranthene		40		27
Benzo[g,h,i]perylene		92		27
Indeno[1,2,3-cd]pyrene		150		27
Fluoranthene		120		27
Pyrene		140		27
Dibenz(a,h)anthracene		ND		27

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	75	30 - 115
Terphenyl-d14	92	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-30C**

Lab Sample ID: 720-401-10
Client Matrix: Solid

Date Sampled: 11/10/2005 0800
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 10
Date Analyzed: 11/25/2005 1907
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2238
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.14 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		50
Acenaphthene		ND		50
Acenaphthylene		ND		50
Fluorene		ND		50
Phenanthrene		ND		50
Anthracene		ND		50
Benzo[a]anthracene		ND		50
Chrysene		ND		50
Benzo[a]pyrene		ND		50
Benzo[b]fluoranthene		ND		50
Benzo[k]fluoranthene		ND		50
Benzo[g,h,i]perylene		ND		50
Indeno[1,2,3-cd]pyrene		ND		50
Fluoranthene		ND		50
Pyrene		ND		50
Dibenz(a,h)anthracene		ND		50

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	63	30 - 115
Terphenyl-d14	68	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-30D**

Lab Sample ID: 720-401-11
Client Matrix: Solid

Date Sampled: 11/10/2005 0800
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/25/2005 1934
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2238
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.17 g
Final Weight/Volume: 5 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		25
Acenaphthene		ND		25
Acenaphthylene		ND		25
Fluorene		ND		25
Phenanthrene		ND		25
Anthracene		ND		25
Benzo[a]anthracene		ND		25
Chrysene		ND		25
Benzo[a]pyrene		ND		25
Benzo[b]fluoranthene		ND		25
Benzo[k]fluoranthene		ND		25
Benzo[g,h,i]perylene		ND		25
Indeno[1,2,3-cd]pyrene		ND		25
Fluoranthene		ND		25
Pyrene		ND		25
Dibenz(a,h)anthracene		ND		25

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	77	30 - 115
Terphenyl-d14	90	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-31D**

Lab Sample ID: 720-401-12
Client Matrix: Solid

Date Sampled: 11/10/2005 0815
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/25/2005 2002
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2238
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.05 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.0
Acenaphthene		ND		5.0
Acenaphthylene		ND		5.0
Fluorene		ND		5.0
Phenanthrene		5.1		5.0
Anthracene		ND		5.0
Benzo[a]anthracene		ND		5.0
Chrysene		ND		5.0
Benzo[a]pyrene		ND		5.0
Benzo[b]fluoranthene		ND		5.0
Benzo[k]fluoranthene		ND		5.0
Benzo[g,h,i]perylene		ND		5.0
Indeno[1,2,3-cd]pyrene		ND		5.0
Fluoranthene		7.4		5.0
Pyrene		6.1		5.0
Dibenz(a,h)anthracene		ND		5.0

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	70	30 - 115
Terphenyl-d14	77	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-31A**

Lab Sample ID: 720-401-13
Client Matrix: Solid

Date Sampled: 11/10/2005 1025
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 10
Date Analyzed: 11/25/2005 2030
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2238
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.21 g
Final Weight/Volume: 5 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		250
Acenaphthene		ND		250
Acenaphthylene		ND		250
Fluorene		ND		250
Phenanthrene		290		250
Anthracene		ND		250
Benzo[a]anthracene		250		250
Chrysene		ND		250
Benzo[a]pyrene		ND		250
Benzo[b]fluoranthene		ND		250
Benzo[k]fluoranthene		ND		250
Benzo[g,h,i]perylene		ND		250
Indeno[1,2,3-cd]pyrene		310		250
Fluoranthene		330		250
Pyrene		350		250
Dibenz(a,h)anthracene		ND		250

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	62	30 - 115
Terphenyl-d14	85	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-33C**

Lab Sample ID: 720-401-14
Client Matrix: Solid

Date Sampled: 11/10/2005 0835
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/25/2005 2058
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2238
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.03 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.0
Acenaphthene		ND		5.0
Acenaphthylene		ND		5.0
Fluorene		ND		5.0
Phenanthrene		ND		5.0
Anthracene		ND		5.0
Benzo[a]anthracene		ND		5.0
Chrysene		ND		5.0
Benzo[a]pyrene		ND		5.0
Benzo[b]fluoranthene		ND		5.0
Benzo[k]fluoranthene		ND		5.0
Benzo[g,h,i]perylene		ND		5.0
Indeno[1,2,3-cd]pyrene		ND		5.0
Fluoranthene		ND		5.0
Pyrene		ND		5.0
Dibenz(a,h)anthracene		ND		5.0

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	67	30 - 115
Terphenyl-d14	78	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-33A**

Lab Sample ID: 720-401-15
Client Matrix: Solid

Date Sampled: 11/10/2005 1035
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/26/2005 1231
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2253
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.00 g
Final Weight/Volume: 5 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		25
Acenaphthene		ND		25
Acenaphthylene		ND		25
Fluorene		ND		25
Phenanthrene		84		25
Anthracene		29		25
Benzo[a]anthracene		89		25
Chrysene		110		25
Benzo[a]pyrene		99		25
Benzo[b]fluoranthene		110		25
Benzo[k]fluoranthene		52		25
Benzo[g,h,i]perylene		100		25
Indeno[1,2,3-cd]pyrene		70		25
Fluoranthene		140		25
Pyrene		160		25
Dibenz(a,h)anthracene		33		25

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	91	30 - 115
Terphenyl-d14	99	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-35C**

Lab Sample ID: 720-401-16
Client Matrix: Solid

Date Sampled: 11/10/2005 0910
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/26/2005 1258
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2253
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.20 g
Final Weight/Volume: 5 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		25
Acenaphthene		ND		25
Acenaphthylene		ND		25
Fluorene		ND		25
Phenanthrene		100		25
Anthracene		29		25
Benzo[a]anthracene		110		25
Chrysene		120		25
Benzo[a]pyrene		83		25
Benzo[b]fluoranthene		130		25
Benzo[k]fluoranthene		47		25
Benzo[g,h,i]perylene		70		25
Indeno[1,2,3-cd]pyrene		110		25
Fluoranthene		210		25
Pyrene		190		25
Dibenz(a,h)anthracene		50		25

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	83	30 - 115
Terphenyl-d14	83	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-35A**

Lab Sample ID: 720-401-17
Client Matrix: Solid

Date Sampled: 11/10/2005 1045
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/26/2005 1326
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2253
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.27 g
Final Weight/Volume: 5 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		25
Acenaphthene		33		25
Acenaphthylene		68		25
Fluorene		ND		25
Phenanthrene		250		25
Anthracene		87		25
Benzo[a]anthracene		330		25
Chrysene		350		25
Benzo[a]pyrene		450		25
Benzo[b]fluoranthene		470		25
Benzo[k]fluoranthene		170		25
Benzo[g,h,i]perylene		270		25
Indeno[1,2,3-cd]pyrene		430		25
Fluoranthene		460		25
Pyrene		530		25
Dibenz(a,h)anthracene		180		25

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	90	30 - 115
Terphenyl-d14	99	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-34D**

Lab Sample ID: 720-401-18
Client Matrix: Solid

Date Sampled: 11/10/2005 0845
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/26/2005 1354
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2253
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.09 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.0
Acenaphthene		ND		5.0
Acenaphthylene		ND		5.0
Fluorene		ND		5.0
Phenanthrene		ND		5.0
Anthracene		ND		5.0
Benzo[a]anthracene		ND		5.0
Chrysene		ND		5.0
Benzo[a]pyrene		ND		5.0
Benzo[b]fluoranthene		ND		5.0
Benzo[k]fluoranthene		ND		5.0
Benzo[g,h,i]perylene		ND		5.0
Indeno[1,2,3-cd]pyrene		ND		5.0
Fluoranthene		ND		5.0
Pyrene		ND		5.0
Dibenz(a,h)anthracene		ND		5.0

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	73	30 - 115
Terphenyl-d14	74	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-36C**

Lab Sample ID: 720-401-19
Client Matrix: Solid

Date Sampled: 11/10/2005 0925
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 1.0
Date Analyzed: 11/26/2005 1421
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2253
Prep Batch: 720-1676

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.15 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.0
Acenaphthene		ND		5.0
Acenaphthylene		8.2		5.0
Fluorene		ND		5.0
Phenanthrene		31		5.0
Anthracene		5.6		5.0
Benzo[a]anthracene		20		5.0
Chrysene		26		5.0
Benzo[a]pyrene		21		5.0
Benzo[b]fluoranthene		29		5.0
Benzo[k]fluoranthene		9.8		5.0
Benzo[g,h,i]perylene		21		5.0
Indeno[1,2,3-cd]pyrene		22		5.0
Fluoranthene		50		5.0
Pyrene		46		5.0
Dibenz(a,h)anthracene		ND		5.0

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	72	30 - 115
Terphenyl-d14	81	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-37B**

Lab Sample ID:	720-401-20			Date Sampled:	11/10/2005 1048
Client Matrix:	Solid	% Moisture:	5.1	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2253		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1676		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.27 g
Date Analyzed:	11/26/2005 1449			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1442			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.2
Acenaphthene		ND		5.2
Acenaphthylene		ND		5.2
Fluorene		ND		5.2
Phenanthrene		ND		5.2
Anthracene		ND		5.2
Benzo[a]anthracene		ND		5.2
Chrysene		ND		5.2
Benzo[a]pyrene		7.2		5.2
Benzo[b]fluoranthene		9.2		5.2
Benzo[k]fluoranthene		ND		5.2
Benzo[g,h,i]perylene		9.0		5.2
Indeno[1,2,3-cd]pyrene		ND		5.2
Fluoranthene		ND		5.2
Pyrene		ND		5.2
Dibenz(a,h)anthracene		ND		5.2

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	66	30 - 115
Terphenyl-d14	74	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-37A**

Lab Sample ID:	720-401-21			Date Sampled:	11/10/2005 1330
Client Matrix:	Solid	% Moisture:	6.4	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2253		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1676		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.06 g
Date Analyzed:	11/26/2005 1517			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1442			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		27
Acenaphthene		ND		27
Acenaphthylene		ND		27
Fluorene		ND		27
Phenanthrene		35		27
Anthracene		ND		27
Benzo[a]anthracene		34		27
Chrysene		58		27
Benzo[a]pyrene		79		27
Benzo[b]fluoranthene		80		27
Benzo[k]fluoranthene		ND		27
Benzo[g,h,i]perylene		92		27
Indeno[1,2,3-cd]pyrene		46		27
Fluoranthene		57		27
Pyrene		71		27
Dibenz(a,h)anthracene		ND		27

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	90	30 - 115
Terphenyl-d14	101	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-25A**

Lab Sample ID:	720-401-22			Date Sampled:	11/10/2005 0955
Client Matrix:	Solid	% Moisture:	10.9	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2258		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1676		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	5.0			Initial Weight/Volume:	30.11 g
Date Analyzed:	11/27/2005 0903			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1442			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		140
Acenaphthene		220		140
Acenaphthylene		ND		140
Fluorene		360		140
Phenanthrene		7600		140
Anthracene		860		140
Benzo[a]anthracene		1600		140
Chrysene		2200		140
Benzo[a]pyrene		1600		140
Benzo[b]fluoranthene		2700		140
Benzo[k]fluoranthene		920		140
Benzo[g,h,i]perylene		990		140
Indeno[1,2,3-cd]pyrene		1500		140
Fluoranthene		7100		140
Pyrene		4900		140
Dibenz(a,h)anthracene		320		140

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	53	30 - 115
Terphenyl-d14	90	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-38B**

Lab Sample ID:	720-401-23			Date Sampled:	11/10/2005 1100
Client Matrix:	Solid	% Moisture:	14.9	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2258		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1676		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.15 g
Date Analyzed:	11/27/2005 0835			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1442			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		29
Acenaphthene		ND		29
Acenaphthylene		ND		29
Fluorene		ND		29
Phenanthrene		53		29
Anthracene		ND		29
Benzo[a]anthracene		ND		29
Chrysene		47		29
Benzo[a]pyrene		31		29
Benzo[b]fluoranthene		59		29
Benzo[k]fluoranthene		ND		29
Benzo[g,h,i]perylene		41		29
Indeno[1,2,3-cd]pyrene		40		29
Fluoranthene		68		29
Pyrene		81		29
Dibenz(a,h)anthracene		ND		29

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	76	30 - 115
Terphenyl-d14	88	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-39B**

Lab Sample ID:	720-401-24			Date Sampled:	11/10/2005 1114
Client Matrix:	Solid	% Moisture:	4.2	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2253		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1676		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.20 g
Date Analyzed:	11/26/2005 1640			Final Weight/Volume:	1 mL
Date Prepared:	11/11/2005 1442			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.2
Acenaphthene		ND		5.2
Acenaphthylene		ND		5.2
Fluorene		ND		5.2
Phenanthrene		ND		5.2
Anthracene		ND		5.2
Benzo[a]anthracene		ND		5.2
Chrysene		ND		5.2
Benzo[a]pyrene		ND		5.2
Benzo[b]fluoranthene		ND		5.2
Benzo[k]fluoranthene		ND		5.2
Benzo[g,h,i]perylene		ND		5.2
Indeno[1,2,3-cd]pyrene		ND		5.2
Fluoranthene		ND		5.2
Pyrene		ND		5.2
Dibenz(a,h)anthracene		ND		5.2

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	79	30 - 115
Terphenyl-d14	83	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-39A**

Lab Sample ID:	720-401-25			Date Sampled:	11/10/2005 1345
Client Matrix:	Solid	% Moisture:	7.5	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2253		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1676		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.13 g
Date Analyzed:	11/26/2005 1708			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1442			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		27
Acenaphthene		ND		27
Acenaphthylene		ND		27
Fluorene		ND		27
Phenanthrene		66		27
Anthracene		ND		27
Benzo[a]anthracene		98		27
Chrysene		150		27
Benzo[a]pyrene		150		27
Benzo[b]fluoranthene		240		27
Benzo[k]fluoranthene		72		27
Benzo[g,h,i]perylene		140		27
Indeno[1,2,3-cd]pyrene		190		27
Fluoranthene		130		27
Pyrene		150		27
Dibenz(a,h)anthracene		72		27

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	85	30 - 115
Terphenyl-d14	99	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-40C**

Lab Sample ID:	720-401-26			Date Sampled:	11/10/2005 1124
Client Matrix:	Solid	% Moisture:	13.5	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2238		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1830		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.10 g
Date Analyzed:	11/25/2005 1046			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.8
Acenaphthene		ND		5.8
Acenaphthylene		ND		5.8
Fluorene		ND		5.8
Phenanthrene		ND		5.8
Anthracene		ND		5.8
Benzo[a]anthracene		ND		5.8
Chrysene		ND		5.8
Benzo[a]pyrene		ND		5.8
Benzo[b]fluoranthene		ND		5.8
Benzo[k]fluoranthene		ND		5.8
Benzo[g,h,i]perylene		ND		5.8
Indeno[1,2,3-cd]pyrene		ND		5.8
Fluoranthene		ND		5.8
Pyrene		ND		5.8
Dibenz(a,h)anthracene		ND		5.8

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	75	30 - 115
Terphenyl-d14	94	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-41B**

Lab Sample ID:	720-401-27			Date Sampled:	11/10/2005 1134
Client Matrix:	Solid	% Moisture:	3.7	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2238		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1830		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.08 g
Date Analyzed:	11/25/2005 1114			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.2
Acenaphthene		ND		5.2
Acenaphthylene		ND		5.2
Fluorene		ND		5.2
Phenanthrene		42		5.2
Anthracene		6.2		5.2
Benzo[a]anthracene		45		5.2
Chrysene		50		5.2
Benzo[a]pyrene		55		5.2
Benzo[b]fluoranthene		77		5.2
Benzo[k]fluoranthene		30		5.2
Benzo[g,h,i]perylene		55		5.2
Indeno[1,2,3-cd]pyrene		72		5.2
Fluoranthene		70		5.2
Pyrene		60		5.2
Dibenz(a,h)anthracene		19		5.2

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	75	30 - 115
Terphenyl-d14	90	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-41A**

Lab Sample ID:	720-401-28			Date Sampled:	11/10/2005 1355
Client Matrix:	Solid	% Moisture:	9.0	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2238		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1830		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.24 g
Date Analyzed:	11/25/2005 1142			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		54
Acenaphthene		ND		54
Acenaphthylene		ND		54
Fluorene		ND		54
Phenanthrene		440		54
Anthracene		100		54
Benzo[a]anthracene		380		54
Chrysene		430		54
Benzo[a]pyrene		360		54
Benzo[b]fluoranthene		420		54
Benzo[k]fluoranthene		160		54
Benzo[g,h,i]perylene		200		54
Indeno[1,2,3-cd]pyrene		300		54
Fluoranthene		500		54
Pyrene		690		54
Dibenz(a,h)anthracene		ND		54

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	70	30 - 115
Terphenyl-d14	76	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-42D**

Lab Sample ID:	720-401-29			Date Sampled:	11/10/2005 1150
Client Matrix:	Solid	% Moisture:	15.4	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2238		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1830		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.06 g
Date Analyzed:	11/25/2005 1209			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.9
Acenaphthene		ND		5.9
Acenaphthylene		ND		5.9
Fluorene		ND		5.9
Phenanthrene		ND		5.9
Anthracene		ND		5.9
Benzo[a]anthracene		ND		5.9
Chrysene		ND		5.9
Benzo[a]pyrene		ND		5.9
Benzo[b]fluoranthene		ND		5.9
Benzo[k]fluoranthene		ND		5.9
Benzo[g,h,i]perylene		ND		5.9
Indeno[1,2,3-cd]pyrene		ND		5.9
Fluoranthene		ND		5.9
Pyrene		ND		5.9
Dibenz(a,h)anthracene		ND		5.9

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	65	30 - 115
Terphenyl-d14	82	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-43B**

Lab Sample ID:	720-401-30			Date Sampled:	11/10/2005 1201
Client Matrix:	Solid	% Moisture:	3.5	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2238		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1830		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.05 g
Date Analyzed:	11/25/2005 1237			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.2
Acenaphthene		6.8		5.2
Acenaphthylene		ND		5.2
Fluorene		6.4		5.2
Phenanthrene		66		5.2
Anthracene		13		5.2
Benzo[a]anthracene		77		5.2
Chrysene		77		5.2
Benzo[a]pyrene		110		5.2
Benzo[b]fluoranthene		140		5.2
Benzo[k]fluoranthene		46		5.2
Benzo[g,h,i]perylene		98		5.2
Indeno[1,2,3-cd]pyrene		140		5.2
Fluoranthene		100		5.2
Pyrene		110		5.2
Dibenz(a,h)anthracene		20		5.2

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	74	30 - 115
Terphenyl-d14	88	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-43A**

Lab Sample ID:	720-401-31			Date Sampled:	11/10/2005 1410
Client Matrix:	Solid	% Moisture:	8.3	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2238		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1830		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.15 g
Date Analyzed:	11/25/2005 1305			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		54
Acenaphthene		ND		54
Acenaphthylene		ND		54
Fluorene		ND		54
Phenanthrene		540		54
Anthracene		130		54
Benzo[a]anthracene		380		54
Chrysene		400		54
Benzo[a]pyrene		300		54
Benzo[b]fluoranthene		350		54
Benzo[k]fluoranthene		130		54
Benzo[g,h,i]perylene		190		54
Indeno[1,2,3-cd]pyrene		270		54
Fluoranthene		470		54
Pyrene		660		54
Dibenz(a,h)anthracene		77		54

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	70	30 - 115
Terphenyl-d14	77	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-44B**

Lab Sample ID:	720-401-32			Date Sampled:	11/10/2005 1218
Client Matrix:	Solid	% Moisture:	7.9	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2238		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1830		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.35 g
Date Analyzed:	11/25/2005 1333			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		54
Acenaphthene		ND		54
Acenaphthylene		ND		54
Fluorene		ND		54
Phenanthrene		ND		54
Anthracene		ND		54
Benzo[a]anthracene		ND		54
Chrysene		ND		54
Benzo[a]pyrene		ND		54
Benzo[b]fluoranthene		ND		54
Benzo[k]fluoranthene		ND		54
Benzo[g,h,i]perylene		ND		54
Indeno[1,2,3-cd]pyrene		ND		54
Fluoranthene		ND		54
Pyrene		ND		54
Dibenz(a,h)anthracene		ND		54

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	73	30 - 115
Terphenyl-d14	79	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-55A**

Lab Sample ID:	720-401-33			Date Sampled:	11/10/2005 1400
Client Matrix:	Solid	% Moisture:	7.6	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2238		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1830		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.46 g
Date Analyzed:	11/25/2005 1401			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		53
Acenaphthene		ND		53
Acenaphthylene		ND		53
Fluorene		ND		53
Phenanthrene		93		53
Anthracene		ND		53
Benzo[a]anthracene		100		53
Chrysene		90		53
Benzo[a]pyrene		110		53
Benzo[b]fluoranthene		130		53
Benzo[k]fluoranthene		ND		53
Benzo[g,h,i]perylene		64		53
Indeno[1,2,3-cd]pyrene		81		53
Fluoranthene		140		53
Pyrene		170		53
Dibenz(a,h)anthracene		ND		53

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	73	30 - 115
Terphenyl-d14	82	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-56A**

Lab Sample ID:	720-401-34			Date Sampled:	11/10/2005 1430
Client Matrix:	Solid	% Moisture:	7.7	Date Received:	11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2238		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1830		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.03 g
Date Analyzed:	11/25/2005 1429			Final Weight/Volume:	5 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		270
Acenaphthene		ND		270
Acenaphthylene		ND		270
Fluorene		ND		270
Phenanthrene		1100		270
Anthracene		ND		270
Benzo[a]anthracene		780		270
Chrysene		790		270
Benzo[a]pyrene		690		270
Benzo[b]fluoranthene		700		270
Benzo[k]fluoranthene		ND		270
Benzo[g,h,i]perylene		420		270
Indeno[1,2,3-cd]pyrene		420		270
Fluoranthene		980		270
Pyrene		1500		270
Dibenz(a,h)anthracene		ND		270

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	70	30 - 115
Terphenyl-d14	87	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-GW-2**

Lab Sample ID: 720-401-35
Client Matrix: Water

Date Sampled: 11/10/2005 1328
Date Received: 11/10/2005 1730

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/15/2005 1544
Date Prepared: 11/11/2005 1325

Analysis Batch: 720-1849
Prep Batch: 720-1673

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 920 mL
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result (ug/L)	Qualifier	RL
Naphthalene	ND		0.11
Acenaphthene	ND		0.11
Acenaphthylene	ND		0.11
Fluorene	ND		0.11
Phenanthrene	ND		0.11
Anthracene	ND		0.11
Benzo[a]anthracene	ND		0.22
Chrysene	ND		0.11
Benzo[a]pyrene	ND		0.11
Benzo[b]fluoranthene	ND		0.11
Benzo[k]fluoranthene	ND		0.11
Benzo[g,h,i]perylene	ND		0.11
Indeno[1,2,3-cd]pyrene	ND		0.11
Fluoranthene	ND		0.11
Pyrene	ND		0.11
Dibenz(a,h)anthracene	ND		1.1

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	57	43 - 116
Terphenyl-d14	71	33 - 141

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-EB-1**

Lab Sample ID: 720-401-1
Client Matrix: Water

Date Sampled: 11/10/2005 0726
Date Received: 11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method: 8015B
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/13/2005 1750
Date Prepared: 11/11/2005 1237

Analysis Batch: 720-1785
Prep Batch: 720-1670

Instrument ID: HP DRO5
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Diesel Range Organics [C10-C28]	ND		50
Motor Oil Range Organics [C24-C36]	ND		500

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	70	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-28D**

Lab Sample ID:	720-401-6			Date Sampled:	11/10/2005 0730
Client Matrix:	Solid	% Moisture:	6.6	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.08 g
Date Analyzed:	11/13/2005 2315			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		2.6		1.1
Motor Oil Range Organics [C24-C36]		ND		53

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	66	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-28A**

Lab Sample ID:	720-401-7			Date Sampled:	11/10/2005 1005
Client Matrix:	Solid	% Moisture:	10.9	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.35 g
Date Analyzed:	11/14/2005 1522			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		70		1.1
Motor Oil Range Organics [C24-C36]		410		55

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	79	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-29B**

Lab Sample ID:	720-401-8			Date Sampled:	11/10/2005 0745
Client Matrix:	Solid	% Moisture:	7.5	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.29 g
Date Analyzed:	11/14/2005 1113			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		38		1.1
Motor Oil Range Organics [C24-C36]		230		54

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	86	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-29A**

Lab Sample ID:	720-401-9			Date Sampled:	11/10/2005 1010
Client Matrix:	Solid	% Moisture:	7.9	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.07 g
Date Analyzed:	11/14/2005 1335			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		17		1.1
Motor Oil Range Organics [C24-C36]		140		54

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	80	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-30C**

Lab Sample ID:	720-401-10			Date Sampled:	11/10/2005 0800
Client Matrix:	Solid	% Moisture:	14.1	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.04 g
Date Analyzed:	11/14/2005 1208			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		41		1.2
Motor Oil Range Organics [C24-C36]		220		58

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	86	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-30D**

Lab Sample ID:	720-401-11			Date Sampled:	11/10/2005 0800
Client Matrix:	Solid	% Moisture:	12.9	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.03 g
Date Analyzed:	11/14/2005 1113			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		41		1.1
Motor Oil Range Organics [C24-C36]		200		57

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	74	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-31D**

Lab Sample ID:	720-401-12			Date Sampled:	11/10/2005 0815
Client Matrix:	Solid	% Moisture:	14.0	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.29 g
Date Analyzed:	11/15/2005 1421			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		3.3		1.2
Motor Oil Range Organics [C24-C36]		ND		58

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	84	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-31A**

Lab Sample ID:	720-401-13			Date Sampled:	11/10/2005 1025
Client Matrix:	Solid	% Moisture:	8.2	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.13 g
Date Analyzed:	11/14/2005 1616			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		33		1.1
Motor Oil Range Organics [C24-C36]		220		54

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	84	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-33C**

Lab Sample ID:	720-401-14			Date Sampled:	11/10/2005 0835
Client Matrix:	Solid	% Moisture:	9.2	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.21 g
Date Analyzed:	11/13/2005 2342			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		2.9		1.1
Motor Oil Range Organics [C24-C36]		ND		55

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	66	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-33A**

Lab Sample ID:	720-401-15			Date Sampled:	11/10/2005 1035
Client Matrix:	Solid	% Moisture:	7.0	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.08 g
Date Analyzed:	11/14/2005 1714			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		30		1.1
Motor Oil Range Organics [C24-C36]		190		54

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	82	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-35C**

Lab Sample ID:	720-401-16			Date Sampled:	11/10/2005 0910
Client Matrix:	Solid	% Moisture:	4.4	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.22 g
Date Analyzed:	11/14/2005 1522			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		40		1.0
Motor Oil Range Organics [C24-C36]		290		52

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	76	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-35A**

Lab Sample ID:	720-401-17			Date Sampled:	11/10/2005 1045
Client Matrix:	Solid	% Moisture:	6.6	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.12 g
Date Analyzed:	11/14/2005 1436			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		93		1.1
Motor Oil Range Organics [C24-C36]		480		53

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	81	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-34D**

Lab Sample ID:	720-401-18			Date Sampled:	11/10/2005 0845
Client Matrix:	Solid	% Moisture:	12.6	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.08 g
Date Analyzed:	11/14/2005 1451			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		5.9		1.1
Motor Oil Range Organics [C24-C36]		ND		57

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	78	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-36C**

Lab Sample ID:	720-401-19			Date Sampled:	11/10/2005 0925
Client Matrix:	Solid	% Moisture:	8.8	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.27 g
Date Analyzed:	11/14/2005 1808			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		16		1.1
Motor Oil Range Organics [C24-C36]		70		54

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	73	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-37B**

Lab Sample ID:	720-401-20			Date Sampled:	11/10/2005 1048
Client Matrix:	Solid	% Moisture:	5.1	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.22 g
Date Analyzed:	11/14/2005 1335			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		5.9		1.0
Motor Oil Range Organics [C24-C36]		ND		52

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	70	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-37A**

Lab Sample ID:	720-401-21			Date Sampled:	11/10/2005 1330
Client Matrix:	Solid	% Moisture:	6.4	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.32 g
Date Analyzed:	11/14/2005 1616			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		120		1.1
Motor Oil Range Organics [C24-C36]		440		53

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	70	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-25A**

Lab Sample ID:	720-401-22			Date Sampled:	11/10/2005 0955
Client Matrix:	Solid	% Moisture:	10.9	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.20 g
Date Analyzed:	11/15/2005 1331			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		190		5.6
Motor Oil Range Organics [C24-C36]		520		280

Surrogate	%Rec		Acceptance Limits
o-Terphenyl	0	*	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-38B**

Lab Sample ID:	720-401-23			Date Sampled:	11/10/2005 1100
Client Matrix:	Solid	% Moisture:	14.9	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.23 g
Date Analyzed:	11/14/2005 1509			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		23		1.2
Motor Oil Range Organics [C24-C36]		120		58

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	71	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-39B**

Lab Sample ID:	720-401-24			Date Sampled:	11/10/2005 1114
Client Matrix:	Solid	% Moisture:	4.2	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.02 g
Date Analyzed:	11/14/2005 1424			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		4.1		1.0
Motor Oil Range Organics [C24-C36]		ND		52

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	75	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-39A**

Lab Sample ID:	720-401-25			Date Sampled:	11/10/2005 1345
Client Matrix:	Solid	% Moisture:	7.5	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1796		Instrument ID:	HP DRO5
Preparation:	3550B	Prep Batch: 720-1672		Lab File ID:	N/A
Dilution:	2.0			Initial Weight/Volume:	30.06 g
Date Analyzed:	11/14/2005 1637			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1318			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		65		2.2
Motor Oil Range Organics [C24-C36]		320		110

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	80	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-40C**

Lab Sample ID:	720-401-26			Date Sampled:	11/10/2005 1124
Client Matrix:	Solid	% Moisture:	13.5	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1987		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1829		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.07 g
Date Analyzed:	11/17/2005 1142			Final Weight/Volume:	5 mL
Date Prepared:	11/16/2005 1152			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		1.2		1.2
Motor Oil Range Organics [C24-C36]		ND		58

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	71	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-41B**

Lab Sample ID:	720-401-27			Date Sampled:	11/10/2005 1134
Client Matrix:	Solid	% Moisture:	3.7	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1987		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1829		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.25 g
Date Analyzed:	11/17/2005 1237			Final Weight/Volume:	5 mL
Date Prepared:	11/16/2005 1152			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		41		1.0
Motor Oil Range Organics [C24-C36]		100		51

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	79	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-41A**

Lab Sample ID:	720-401-28			Date Sampled:	11/10/2005 1355
Client Matrix:	Solid	% Moisture:	9.0	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1987		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1829		Lab File ID:	N/A
Dilution:	2.0			Initial Weight/Volume:	30.34 g
Date Analyzed:	11/17/2005 1304			Final Weight/Volume:	5 mL
Date Prepared:	11/16/2005 1152			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		50		2.2
Motor Oil Range Organics [C24-C36]		270		110

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	81	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-42D**

Lab Sample ID:	720-401-29			Date Sampled:	11/10/2005 1150
Client Matrix:	Solid	% Moisture:	15.4	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1987		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1829		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.12 g
Date Analyzed:	11/17/2005 1209			Final Weight/Volume:	5 mL
Date Prepared:	11/16/2005 1152			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		1.2		1.2
Motor Oil Range Organics [C24-C36]		ND		59

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	63	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-43B**

Lab Sample ID:	720-401-30			Date Sampled:	11/10/2005 1201
Client Matrix:	Solid	% Moisture:	3.5	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1987		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1829		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.04 g
Date Analyzed:	11/17/2005 1304			Final Weight/Volume:	5 mL
Date Prepared:	11/16/2005 1152			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		6.6		1.0
Motor Oil Range Organics [C24-C36]		ND		52

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	76	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-43A**

Lab Sample ID:	720-401-31			Date Sampled:	11/10/2005 1410
Client Matrix:	Solid	% Moisture:	8.3	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1987		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1829		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.32 g
Date Analyzed:	11/17/2005 1644			Final Weight/Volume:	5 mL
Date Prepared:	11/16/2005 1152			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		42		1.1
Motor Oil Range Organics [C24-C36]		210		54

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	76	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-44B**

Lab Sample ID:	720-401-32			Date Sampled:	11/10/2005 1218
Client Matrix:	Solid	% Moisture:	7.9	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1987		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1829		Lab File ID:	N/A
Dilution:	2.0			Initial Weight/Volume:	30.36 g
Date Analyzed:	11/17/2005 1738			Final Weight/Volume:	5 mL
Date Prepared:	11/16/2005 1152			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		59		2.1
Motor Oil Range Organics [C24-C36]		360		110

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	75	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-55A**

Lab Sample ID:	720-401-33			Date Sampled:	11/10/2005 1400
Client Matrix:	Solid	% Moisture:	7.6	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1987		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1829		Lab File ID:	N/A
Dilution:	2.0			Initial Weight/Volume:	30.04 g
Date Analyzed:	11/17/2005 1833			Final Weight/Volume:	5 mL
Date Prepared:	11/16/2005 1152			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		53		2.2
Motor Oil Range Organics [C24-C36]		320		110

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	79	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-56A**

Lab Sample ID:	720-401-34			Date Sampled:	11/10/2005 1430
Client Matrix:	Solid	% Moisture:	7.7	Date Received:	11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-1987		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1829		Lab File ID:	N/A
Dilution:	2.0			Initial Weight/Volume:	30.08 g
Date Analyzed:	11/17/2005 1927			Final Weight/Volume:	5 mL
Date Prepared:	11/16/2005 1152			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		74		2.2
Motor Oil Range Organics [C24-C36]		370		110

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	67	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-GW-2**

Lab Sample ID: 720-401-35
Client Matrix: Water

Date Sampled: 11/10/2005 1328
Date Received: 11/10/2005 1730

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method: 8015B
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/13/2005 1938
Date Prepared: 11/11/2005 1237

Analysis Batch: 720-1785
Prep Batch: 720-1670

Instrument ID: HP DRO5
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Diesel Range Organics [C10-C28]	ND		50
Motor Oil Range Organics [C24-C36]	ND		500

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	78	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-EB-1**

Lab Sample ID: 720-401-1
Client Matrix: Water

Date Sampled: 11/10/2005 0726
Date Received: 11/10/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method: 8081A
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/16/2005 0332
Date Prepared: 11/11/2005 1219

Analysis Batch: 720-1910
Prep Batch: 720-1668

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 830 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Aldrin	ND		0.072
Dieldrin	ND		0.072
Endrin aldehyde	ND		0.072
Endrin	ND		0.072
Endrin ketone	ND		0.072
Heptachlor	ND		0.072
Heptachlor epoxide	ND		0.072
4,4'-DDT	ND		0.072
4,4'-DDE	ND		0.072
4,4'-DDD	ND		0.072
Endosulfan I	ND		0.072
Endosulfan II	ND		0.072
alpha-BHC	ND		0.072
beta-BHC	ND		0.072
gamma-BHC (Lindane)	ND		0.072
delta-BHC	ND		0.072
Endosulfan sulfate	ND		0.072
Methoxychlor	ND		0.072
Toxaphene	ND		1.2
Chlordane (technical)	ND		1.2
alpha-Chlordane	ND		0.072
gamma-Chlordane	ND		0.072

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	101	62 - 123
DCB Decachlorobiphenyl	94	56 - 136

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: **EBHH-25C**

Lab Sample ID:	720-401-2			Date Sampled:	11/10/2005 0734
Client Matrix:	Solid	% Moisture:	12.3	Date Received:	11/10/2005 1730

8081A Organochlorine Pesticides by Gas Chromatography

Method:	8081A	Analysis Batch: 720-1975		Instrument ID:	Varian Pest 2
Preparation:	3550B	Prep Batch: 720-1679		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.18 g
Date Analyzed:	11/16/2005 0303			Final Weight/Volume:	10 mL
Date Prepared:	11/11/2005 1452			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Aldrin		ND		2.3
Dieldrin		ND		2.3
Endrin aldehyde		ND		2.3
Endrin		ND		2.3
Endrin ketone		ND		2.3
Heptachlor		ND		2.3
Heptachlor epoxide		ND		2.3
4,4'-DDT		ND		2.3
4,4'-DDE		ND		2.3
4,4'-DDD		ND		2.3
Endosulfan I		ND		2.3
Endosulfan II		ND		2.3
alpha-BHC		ND		2.3
beta-BHC		ND		2.3
gamma-BHC (Lindane)		ND		2.3
delta-BHC		ND		2.3
Endosulfan sulfate		ND		2.3
Methoxychlor		ND		2.3
Toxaphene		ND		110
Chlordane (technical)		ND		57
alpha-Chlordane		ND		2.3
gamma-Chlordane		ND		2.3

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	85	50 - 125
DCB Decachlorobiphenyl	88	46 - 142

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-EB-1

Lab Sample ID:	720-401-1	Date Sampled:	11/10/2005 0726
Client Matrix:	Water	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3010A	Prep Batch: 720-1816	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	50 mL
Date Analyzed:	11/17/2005 0851		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 0856			

Analyte	Result (mg/L)	Qualifier	RL
Silver	ND		0.0050
Arsenic	ND		0.0050
Barium	ND		0.0050
Cadmium	ND		0.0020
Chromium	ND		0.0050
Lead	ND		0.0050
Selenium	ND		0.0050

7470A Mercury in Liquid Waste (Manual Cold Vapor Technique)

Method:	7470A	Analysis Batch: 720-1793	Instrument ID:	FIMS 100
Preparation:	7470A	Prep Batch: 720-1754	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	25 mL
Date Analyzed:	11/15/2005 1315		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 0859			

Analyte	Result (mg/L)	Qualifier	RL
Mercury	ND		0.00020

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-25C

Lab Sample ID:	720-401-2		Date Sampled:	11/10/2005 0734
Client Matrix:	Solid	% Moisture: 12.3	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/17/2005 1229		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.0		1.1
Barium		140		1.1
Cadmium		1.3		0.57
Chromium		36		1.1
Lead		7.4		1.1
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/17/2005 1258		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.057

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-28B

Lab Sample ID:	720-401-4		Date Sampled:	11/10/2005 0730
Client Matrix:	Solid	% Moisture: 14.1	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/17/2005 1233		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.2
Arsenic		6.7		1.2
Barium		190		1.2
Cadmium		1.8		0.58
Chromium		47		1.2
Lead		9.0		1.2
Selenium		ND		2.3

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/17/2005 1259		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.057

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-28C

Lab Sample ID:	720-401-5		Date Sampled:	11/10/2005 0730
Client Matrix:	Solid	% Moisture: 11.1	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/17/2005 1237		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		4.6		1.1
Barium		120		1.1
Cadmium		1.2		0.56
Chromium		33		1.1
Lead		5.6		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/17/2005 1300		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.055

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-28A

Lab Sample ID:	720-401-7		Date Sampled:	11/10/2005 1005
Client Matrix:	Solid	% Moisture: 10.9	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/17/2005 1240		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.6		1.1
Barium		210		1.1
Cadmium		1.7		0.55
Chromium		36		1.1
Lead		390		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/17/2005 1301		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.10		0.055

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-29A

Lab Sample ID:	720-401-9		Date Sampled:	11/10/2005 1010
Client Matrix:	Solid	% Moisture: 7.9	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/17/2005 1251		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.0
Arsenic		9.0		1.0
Barium		110		1.0
Cadmium		1.6		0.52
Chromium		27		1.0
Lead		56		1.0
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/17/2005 1302		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.26		0.053

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-31A

Lab Sample ID:	720-401-13		Date Sampled:	11/10/2005 1025
Client Matrix:	Solid	% Moisture: 8.2	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/17/2005 1303		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		14		1.1
Barium		120		1.1
Cadmium		1.7		0.54
Chromium		28		1.1
Lead		74		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/17/2005 1309		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.27		0.052

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-33A

Lab Sample ID:	720-401-15		Date Sampled:	11/10/2005 1035
Client Matrix:	Solid	% Moisture: 7.0	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/17/2005 1307		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		11		1.1
Barium		120		1.1
Cadmium		1.7		0.53
Chromium		33		1.1
Lead		63		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/17/2005 1310		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.23		0.053

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-35A

Lab Sample ID:	720-401-17		Date Sampled:	11/10/2005 1045
Client Matrix:	Solid	% Moisture: 6.6	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/17/2005 1310		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.0
Arsenic		12		1.0
Barium		140		1.0
Cadmium		1.9		0.52
Chromium		26		1.0
Lead		460		1.0
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/17/2005 1311		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.65		0.053

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-37A

Lab Sample ID:	720-401-21		Date Sampled:	11/10/2005 1330
Client Matrix:	Solid	% Moisture: 6.4	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/17/2005 1314		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.0
Arsenic		4.4		1.0
Barium		130		1.0
Cadmium		3.3		0.52
Chromium		120		1.0
Lead		38		1.0
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/17/2005 1312		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		ND		0.051

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-25A

Lab Sample ID:	720-401-22		Date Sampled:	11/10/2005 0955
Client Matrix:	Solid	% Moisture: 10.9	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/17/2005 1318		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		10		1.1
Barium		190		1.1
Cadmium		2.1		0.55
Chromium		36		1.1
Lead		190		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/17/2005 1313		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.13		0.055

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-39A

Lab Sample ID:	720-401-25		Date Sampled:	11/10/2005 1345
Client Matrix:	Solid	% Moisture: 7.5	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.04 g
Date Analyzed:	11/17/2005 1322		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.0
Arsenic		9.8		1.0
Barium		160		1.0
Cadmium		4.2		0.52
Chromium		92		1.0
Lead		91		1.0
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.05 g
Date Analyzed:	11/17/2005 1315		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.078		0.051

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-41A

Lab Sample ID:	720-401-28		Date Sampled:	11/10/2005 1355
Client Matrix:	Solid	% Moisture: 9.0	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/17/2005 1325		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		7.0		1.1
Barium		130		1.1
Cadmium		4.3		0.55
Chromium		95		1.1
Lead		70		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/17/2005 1316		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.092		0.054

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-43A

Lab Sample ID:	720-401-31		Date Sampled:	11/10/2005 1410
Client Matrix:	Solid	% Moisture: 8.3	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/17/2005 1407		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.6		1.1
Barium		120		1.1
Cadmium		2.0		0.53
Chromium		51		1.1
Lead		82		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/17/2005 1320		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.17		0.055

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-55A

Lab Sample ID:	720-401-33		Date Sampled:	11/10/2005 1400
Client Matrix:	Solid	% Moisture: 7.6	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/17/2005 1411		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		10		1.1
Barium		120		1.1
Cadmium		2.0		0.53
Chromium		28		1.1
Lead		70		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1922	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1825	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/17/2005 1323		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1112			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.37		0.054

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-56A

Lab Sample ID:	720-401-34		Date Sampled:	11/10/2005 1430
Client Matrix:	Solid	% Moisture: 7.7	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1824	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.02 g
Date Analyzed:	11/17/2005 1415		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		5.3		1.1
Barium		150		1.1
Cadmium		1.8		0.53
Chromium		68		1.1
Lead		89		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-2039	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-2022	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/21/2005 1027		Final Weight/Volume:	50 mL
Date Prepared:	11/21/2005 0740			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		1.8		0.054

Client: Weston Solutions,Inc

Job Number: 720-401-1

Client Sample ID: EBHH-GW-1

Lab Sample ID:	720-401-36	Date Sampled:	11/10/2005 1446
Client Matrix:	Water	Date Received:	11/10/2005 1730

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3010A	Prep Batch: 720-1816	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	50 mL
Date Analyzed:	11/17/2005 0854		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 0856			

Analyte	Result (mg/L)	Qualifier	RL
Silver	ND		0.0050
Arsenic	ND		0.0050
Barium	0.089		0.0050
Cadmium	ND		0.0020
Chromium	ND		0.0050
Lead	ND		0.0050
Selenium	ND		0.0050

7470A Mercury in Liquid Waste (Manual Cold Vapor Technique)

Method:	7470A	Analysis Batch: 720-1793	Instrument ID:	FIMS 100
Preparation:	7470A	Prep Batch: 720-1754	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	25 mL
Date Analyzed:	11/15/2005 1317		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 0859			

Analyte	Result (mg/L)	Qualifier	RL
Mercury	ND		0.00020

General Chemistry

Client Sample ID: EBHH-25C

Lab Sample ID: 720-401-2
Client Matrix: Solid

Date Sampled: 11/10/2005 0734
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	12		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-28B

Lab Sample ID: 720-401-4
Client Matrix: Solid

Date Sampled: 11/10/2005 0730
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	14		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-28C

Lab Sample ID: 720-401-5
Client Matrix: Solid

Date Sampled: 11/10/2005 0730
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	11		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-28D

Lab Sample ID: 720-401-6
Client Matrix: Solid

Date Sampled: 11/10/2005 0730
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	6.6		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-28A

Lab Sample ID: 720-401-7
Client Matrix: Solid

Date Sampled: 11/10/2005 1005
Date Received: 11/10/2005 1730

General Chemistry

Client Sample ID: EBHH-28A

Lab Sample ID: 720-401-7
Client Matrix: Solid

Date Sampled: 11/10/2005 1005
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	11		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-29B

Lab Sample ID: 720-401-8
Client Matrix: Solid

Date Sampled: 11/10/2005 0745
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.5		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-29A

Lab Sample ID: 720-401-9
Client Matrix: Solid

Date Sampled: 11/10/2005 1010
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.9		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-30C

Lab Sample ID: 720-401-10
Client Matrix: Solid

Date Sampled: 11/10/2005 0800
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	14		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-30D

Lab Sample ID: 720-401-11
Client Matrix: Solid

Date Sampled: 11/10/2005 0800
Date Received: 11/10/2005 1730

General Chemistry

Client Sample ID: EBHH-30D

Lab Sample ID: 720-401-11
Client Matrix: Solid

Date Sampled: 11/10/2005 0800
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	13		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-31D

Lab Sample ID: 720-401-12
Client Matrix: Solid

Date Sampled: 11/10/2005 0815
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	14		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-31A

Lab Sample ID: 720-401-13
Client Matrix: Solid

Date Sampled: 11/10/2005 1025
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	8.2		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-33C

Lab Sample ID: 720-401-14
Client Matrix: Solid

Date Sampled: 11/10/2005 0835
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.2		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-33A

Lab Sample ID: 720-401-15
Client Matrix: Solid

Date Sampled: 11/10/2005 1035
Date Received: 11/10/2005 1730

General Chemistry

Client Sample ID: EBHH-33A

Lab Sample ID: 720-401-15
Client Matrix: Solid

Date Sampled: 11/10/2005 1035
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.0		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-35C

Lab Sample ID: 720-401-16
Client Matrix: Solid

Date Sampled: 11/10/2005 0910
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	4.4		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-35A

Lab Sample ID: 720-401-17
Client Matrix: Solid

Date Sampled: 11/10/2005 1045
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	6.6		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-34D

Lab Sample ID: 720-401-18
Client Matrix: Solid

Date Sampled: 11/10/2005 0845
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	13		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-36C

Lab Sample ID: 720-401-19
Client Matrix: Solid

Date Sampled: 11/10/2005 0925
Date Received: 11/10/2005 1730

General Chemistry

Client Sample ID: **EBHH-36C**

Lab Sample ID: 720-401-19
Client Matrix: Solid

Date Sampled: 11/10/2005 0925
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	8.8		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: **EBHH-37B**

Lab Sample ID: 720-401-20
Client Matrix: Solid

Date Sampled: 11/10/2005 1048
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	5.1		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: **EBHH-37A**

Lab Sample ID: 720-401-21
Client Matrix: Solid

Date Sampled: 11/10/2005 1330
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	6.4		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: **EBHH-25A**

Lab Sample ID: 720-401-22
Client Matrix: Solid

Date Sampled: 11/10/2005 0955
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	11		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: **EBHH-38B**

Lab Sample ID: 720-401-23
Client Matrix: Solid

Date Sampled: 11/10/2005 1100
Date Received: 11/10/2005 1730

General Chemistry

Client Sample ID: EBHH-38B

Lab Sample ID: 720-401-23
Client Matrix: Solid

Date Sampled: 11/10/2005 1100
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	15		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-39B

Lab Sample ID: 720-401-24
Client Matrix: Solid

Date Sampled: 11/10/2005 1114
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	4.2		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-39A

Lab Sample ID: 720-401-25
Client Matrix: Solid

Date Sampled: 11/10/2005 1345
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.5		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-40C

Lab Sample ID: 720-401-26
Client Matrix: Solid

Date Sampled: 11/10/2005 1124
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	13		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-41B

Lab Sample ID: 720-401-27
Client Matrix: Solid

Date Sampled: 11/10/2005 1134
Date Received: 11/10/2005 1730

General Chemistry

Client Sample ID: EBHH-41B

Lab Sample ID: 720-401-27
Client Matrix: Solid

Date Sampled: 11/10/2005 1134
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	3.7		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-41A

Lab Sample ID: 720-401-28
Client Matrix: Solid

Date Sampled: 11/10/2005 1355
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.0		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-42D

Lab Sample ID: 720-401-29
Client Matrix: Solid

Date Sampled: 11/10/2005 1150
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	15		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-43B

Lab Sample ID: 720-401-30
Client Matrix: Solid

Date Sampled: 11/10/2005 1201
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	3.5		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-43A

Lab Sample ID: 720-401-31
Client Matrix: Solid

Date Sampled: 11/10/2005 1410
Date Received: 11/10/2005 1730

General Chemistry

Client Sample ID: **EBHH-43A**

Lab Sample ID: 720-401-31
Client Matrix: Solid

Date Sampled: 11/10/2005 1410
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	8.3		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: **EBHH-44B**

Lab Sample ID: 720-401-32
Client Matrix: Solid

Date Sampled: 11/10/2005 1218
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.9		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: **EBHH-55A**

Lab Sample ID: 720-401-33
Client Matrix: Solid

Date Sampled: 11/10/2005 1400
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.6		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: **EBHH-56A**

Lab Sample ID: 720-401-34
Client Matrix: Solid

Date Sampled: 11/10/2005 1430
Date Received: 11/10/2005 1730

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.7		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

DATA REPORTING QUALIFIERS

Client: Weston Solutions,Inc

Job Number: 720-401-1

Lab Section	Qualifier	Description
GC/MS Semi VOA		
	*	LCS, LCSD, MS, MSD, MD, or Surrogate exceeds the control limits
GC Semi VOA		
	*	LCS, LCSD, MS, MSD, MD, or Surrogate exceeds the control limits
Metals		
	N	MS, MSD: Spike recovery exceeds upper or lower control limits.

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC/MS VOA				
Analysis Batch:720-2028				
LCS 720-2028/17	Lab Control Spike	Water	8260B	
MB 720-2028/18	Method Blank	Water	8260B	
720-401-36	EBHH-GW-1	Water	8260B	
720-524-B-5 MS	Matrix Spike	Water	8260B	
720-524-B-5 MSD	Matrix Spike Duplicate	Water	8260B	

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC/MS Semi VOA				
Prep Batch: 720-1673				
LCS 720-1673/2-A	Lab Control Spike	Water	3510C	
LCSD 720-1673/3-A	Lab Control Spike Duplicate	Water	3510C	
MB 720-1673/1-A	Method Blank	Water	3510C	
720-401-1	EBHH-EB-1	Water	3510C	
720-401-35	EBHH-GW-2	Water	3510C	
Prep Batch: 720-1676				
LCS 720-1676/22-A	Lab Control Spike	Solid	3550B	
LCSD 720-1676/23-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1676/21-A	Method Blank	Solid	3550B	
720-401-6	EBHH-28D	Solid	3550B	
720-401-7	EBHH-28A	Solid	3550B	
720-401-8	EBHH-29B	Solid	3550B	
720-401-9	EBHH-29A	Solid	3550B	
720-401-9MS	Matrix Spike	Solid	3550B	
720-401-9MSD	Matrix Spike Duplicate	Solid	3550B	
720-401-10	EBHH-30C	Solid	3550B	
720-401-11	EBHH-30D	Solid	3550B	
720-401-12	EBHH-31D	Solid	3550B	
720-401-13	EBHH-31A	Solid	3550B	
720-401-14	EBHH-33C	Solid	3550B	
720-401-15	EBHH-33A	Solid	3550B	
720-401-16	EBHH-35C	Solid	3550B	
720-401-17	EBHH-35A	Solid	3550B	
720-401-18	EBHH-34D	Solid	3550B	
720-401-19	EBHH-36C	Solid	3550B	
720-401-20	EBHH-37B	Solid	3550B	
720-401-21	EBHH-37A	Solid	3550B	
720-401-22	EBHH-25A	Solid	3550B	
720-401-23	EBHH-38B	Solid	3550B	
720-401-24	EBHH-39B	Solid	3550B	
720-401-25	EBHH-39A	Solid	3550B	

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC/MS Semi VOA				
Prep Batch: 720-1830				
LCS 720-1830/3-A	Lab Control Spike	Solid	3550B	
LCSD 720-1830/4-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1830/2-A	Method Blank	Solid	3550B	
720-381-C-71-D MS	Matrix Spike	Solid	3550B	
720-381-C-71-E MSD	Matrix Spike Duplicate	Solid	3550B	
720-401-26	EBHH-40C	Solid	3550B	
720-401-27	EBHH-41B	Solid	3550B	
720-401-28	EBHH-41A	Solid	3550B	
720-401-28MS	Matrix Spike	Solid	3550B	
720-401-28MSD	Matrix Spike Duplicate	Solid	3550B	
720-401-29	EBHH-42D	Solid	3550B	
720-401-30	EBHH-43B	Solid	3550B	
720-401-31	EBHH-43A	Solid	3550B	
720-401-32	EBHH-44B	Solid	3550B	
720-401-33	EBHH-55A	Solid	3550B	
720-401-34	EBHH-56A	Solid	3550B	
Analysis Batch:720-1849				
LCS 720-1673/2-A	Lab Control Spike	Water	8270C	720-1673
LCSD 720-1673/3-A	Lab Control Spike Duplicate	Water	8270C	720-1673
MB 720-1673/1-A	Method Blank	Water	8270C	720-1673
720-401-1	EBHH-EB-1	Water	8270C	720-1673
720-401-35	EBHH-GW-2	Water	8270C	720-1673
Analysis Batch:720-2238				
LCS 720-1676/22-A	Lab Control Spike	Solid	8270C	720-1676
LCSD 720-1676/23-A	Lab Control Spike Duplicate	Solid	8270C	720-1676
MB 720-1676/21-A	Method Blank	Solid	8270C	720-1676
720-401-6	EBHH-28D	Solid	8270C	720-1676
720-401-7	EBHH-28A	Solid	8270C	720-1676
720-401-8	EBHH-29B	Solid	8270C	720-1676
720-401-9	EBHH-29A	Solid	8270C	720-1676
720-401-10	EBHH-30C	Solid	8270C	720-1676
720-401-11	EBHH-30D	Solid	8270C	720-1676
720-401-12	EBHH-31D	Solid	8270C	720-1676
720-401-13	EBHH-31A	Solid	8270C	720-1676
720-401-14	EBHH-33C	Solid	8270C	720-1676

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC/MS Semi VOA				
Analysis Batch:720-2253				
720-401-9MS	Matrix Spike	Solid	8270C	720-1676
720-401-9MSD	Matrix Spike Duplicate	Solid	8270C	720-1676
720-401-15	EBHH-33A	Solid	8270C	720-1676
720-401-16	EBHH-35C	Solid	8270C	720-1676
720-401-17	EBHH-35A	Solid	8270C	720-1676
720-401-18	EBHH-34D	Solid	8270C	720-1676
720-401-19	EBHH-36C	Solid	8270C	720-1676
720-401-20	EBHH-37B	Solid	8270C	720-1676
720-401-21	EBHH-37A	Solid	8270C	720-1676
720-401-24	EBHH-39B	Solid	8270C	720-1676
720-401-25	EBHH-39A	Solid	8270C	720-1676
Analysis Batch:720-2258				
720-401-22	EBHH-25A	Solid	8270C	720-1676
720-401-23	EBHH-38B	Solid	8270C	720-1676
Analysis Batch:720-2227				
LCS 720-1830/3-A	Lab Control Spike	Solid	8270C	720-1830
LCSD 720-1830/4-A	Lab Control Spike Duplicate	Solid	8270C	720-1830
MB 720-1830/2-A	Method Blank	Solid	8270C	720-1830
720-381-C-71-D MS	Matrix Spike	Solid	8270C	720-1830
720-381-C-71-E MSD	Matrix Spike Duplicate	Solid	8270C	720-1830
Analysis Batch:720-2238				
720-401-26	EBHH-40C	Solid	8270C	720-1830
720-401-27	EBHH-41B	Solid	8270C	720-1830
720-401-28	EBHH-41A	Solid	8270C	720-1830
720-401-28MS	Matrix Spike	Solid	8270C	720-1830
720-401-28MSD	Matrix Spike Duplicate	Solid	8270C	720-1830
720-401-29	EBHH-42D	Solid	8270C	720-1830
720-401-30	EBHH-43B	Solid	8270C	720-1830
720-401-31	EBHH-43A	Solid	8270C	720-1830
720-401-32	EBHH-44B	Solid	8270C	720-1830
720-401-33	EBHH-55A	Solid	8270C	720-1830
720-401-34	EBHH-56A	Solid	8270C	720-1830

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC Semi VOA				
Prep Batch: 720-1668				
LCS 720-1668/3-A	Lab Control Spike	Water	3510C	
LCSD 720-1668/4-A	Lab Control Spike Duplicate	Water	3510C	
MB 720-1668/2-A	Method Blank	Water	3510C	
720-401-1	EBHH-EB-1	Water	3510C	
Prep Batch: 720-1670				
LCS 720-1670/2-A	Lab Control Spike	Water	3510C	
LCSD 720-1670/3-A	Lab Control Spike Duplicate	Water	3510C	
MB 720-1670/1-A	Method Blank	Water	3510C	
720-401-1	EBHH-EB-1	Water	3510C	
720-401-35	EBHH-GW-2	Water	3510C	
Prep Batch: 720-1672				
LCS 720-1672/22-A	Lab Control Spike	Solid	3550B	
LCSD 720-1672/23-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1672/21-A	Method Blank	Solid	3550B	
720-401-6	EBHH-28D	Solid	3550B	
720-401-7	EBHH-28A	Solid	3550B	
720-401-8	EBHH-29B	Solid	3550B	
720-401-9	EBHH-29A	Solid	3550B	
720-401-10	EBHH-30C	Solid	3550B	
720-401-11	EBHH-30D	Solid	3550B	
720-401-12	EBHH-31D	Solid	3550B	
720-401-13	EBHH-31A	Solid	3550B	
720-401-14	EBHH-33C	Solid	3550B	
720-401-15	EBHH-33A	Solid	3550B	
720-401-16	EBHH-35C	Solid	3550B	
720-401-17	EBHH-35A	Solid	3550B	
720-401-18	EBHH-34D	Solid	3550B	
720-401-19	EBHH-36C	Solid	3550B	
720-401-20	EBHH-37B	Solid	3550B	
720-401-21	EBHH-37A	Solid	3550B	
720-401-22	EBHH-25A	Solid	3550B	
720-401-23	EBHH-38B	Solid	3550B	
720-401-24	EBHH-39B	Solid	3550B	
720-401-25	EBHH-39A	Solid	3550B	
Prep Batch: 720-1679				
LCS 720-1679/3-A	Lab Control Spike	Solid	3550B	
LCSD 720-1679/4-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1679/2-A	Method Blank	Solid	3550B	
720-401-2	EBHH-25C	Solid	3550B	
720-401-2MS	Matrix Spike	Solid	3550B	
720-401-2MSD	Matrix Spike Duplicate	Solid	3550B	

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC Semi VOA				
Prep Batch: 720-1829				
LCS 720-1829/3-A	Lab Control Spike	Solid	3550B	
LCSD 720-1829/4-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1829/2-A	Method Blank	Solid	3550B	
720-401-26	EBHH-40C	Solid	3550B	
720-401-27	EBHH-41B	Solid	3550B	
720-401-28	EBHH-41A	Solid	3550B	
720-401-28MS	Matrix Spike	Solid	3550B	
720-401-28MSD	Matrix Spike Duplicate	Solid	3550B	
720-401-29	EBHH-42D	Solid	3550B	
720-401-30	EBHH-43B	Solid	3550B	
720-401-31	EBHH-43A	Solid	3550B	
720-401-32	EBHH-44B	Solid	3550B	
720-401-33	EBHH-55A	Solid	3550B	
720-401-34	EBHH-56A	Solid	3550B	
Analysis Batch:720-1910				
LCS 720-1668/3-A	Lab Control Spike	Water	8081A	720-1668
LCSD 720-1668/4-A	Lab Control Spike Duplicate	Water	8081A	720-1668
MB 720-1668/2-A	Method Blank	Water	8081A	720-1668
720-401-1	EBHH-EB-1	Water	8081A	720-1668
Analysis Batch:720-1785				
LCS 720-1670/2-A	Lab Control Spike	Water	8015B	720-1670
LCSD 720-1670/3-A	Lab Control Spike Duplicate	Water	8015B	720-1670
MB 720-1670/1-A	Method Blank	Water	8015B	720-1670
720-401-1	EBHH-EB-1	Water	8015B	720-1670
720-401-35	EBHH-GW-2	Water	8015B	720-1670

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC Semi VOA				
Analysis Batch:720-1796				
LCS 720-1672/22-A	Lab Control Spike	Solid	8015B	720-1672
LCSD 720-1672/23-A	Lab Control Spike Duplicate	Solid	8015B	720-1672
MB 720-1672/21-A	Method Blank	Solid	8015B	720-1672
720-401-6	EBHH-28D	Solid	8015B	720-1672
720-401-7	EBHH-28A	Solid	8015B	720-1672
720-401-8	EBHH-29B	Solid	8015B	720-1672
720-401-9	EBHH-29A	Solid	8015B	720-1672
720-401-10	EBHH-30C	Solid	8015B	720-1672
720-401-11	EBHH-30D	Solid	8015B	720-1672
720-401-12	EBHH-31D	Solid	8015B	720-1672
720-401-13	EBHH-31A	Solid	8015B	720-1672
720-401-14	EBHH-33C	Solid	8015B	720-1672
720-401-15	EBHH-33A	Solid	8015B	720-1672
720-401-16	EBHH-35C	Solid	8015B	720-1672
720-401-17	EBHH-35A	Solid	8015B	720-1672
720-401-18	EBHH-34D	Solid	8015B	720-1672
720-401-19	EBHH-36C	Solid	8015B	720-1672
720-401-20	EBHH-37B	Solid	8015B	720-1672
720-401-21	EBHH-37A	Solid	8015B	720-1672
720-401-22	EBHH-25A	Solid	8015B	720-1672
720-401-23	EBHH-38B	Solid	8015B	720-1672
720-401-24	EBHH-39B	Solid	8015B	720-1672
720-401-25	EBHH-39A	Solid	8015B	720-1672
Analysis Batch:720-1975				
LCS 720-1679/3-A	Lab Control Spike	Solid	8081A	720-1679
LCSD 720-1679/4-A	Lab Control Spike Duplicate	Solid	8081A	720-1679
MB 720-1679/2-A	Method Blank	Solid	8081A	720-1679
720-401-2	EBHH-25C	Solid	8081A	720-1679
720-401-2MS	Matrix Spike	Solid	8081A	720-1679
720-401-2MSD	Matrix Spike Duplicate	Solid	8081A	720-1679

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC Semi VOA				
Analysis Batch:720-1987				
LCS 720-1829/3-A	Lab Control Spike	Solid	8015B	720-1829
LCSD 720-1829/4-A	Lab Control Spike Duplicate	Solid	8015B	720-1829
MB 720-1829/2-A	Method Blank	Solid	8015B	720-1829
720-401-26	EBHH-40C	Solid	8015B	720-1829
720-401-27	EBHH-41B	Solid	8015B	720-1829
720-401-28	EBHH-41A	Solid	8015B	720-1829
720-401-28MS	Matrix Spike	Solid	8015B	720-1829
720-401-28MSD	Matrix Spike Duplicate	Solid	8015B	720-1829
720-401-29	EBHH-42D	Solid	8015B	720-1829
720-401-30	EBHH-43B	Solid	8015B	720-1829
720-401-31	EBHH-43A	Solid	8015B	720-1829
720-401-32	EBHH-44B	Solid	8015B	720-1829
720-401-33	EBHH-55A	Solid	8015B	720-1829
720-401-34	EBHH-56A	Solid	8015B	720-1829

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Prep Batch: 720-1754				
LCS 720-1754/2-A	Lab Control Spike	Water	7470A	
LCSD 720-1754/3-A	Lab Control Spike Duplicate	Water	7470A	
MB 720-1754/1-A	Method Blank	Water	7470A	
720-401-1	EBHH-EB-1	Water	7470A	
720-401-36	EBHH-GW-1	Water	7470A	
720-413-M-9-B MS	Matrix Spike	Water	7470A	
720-413-M-9-C MSD	Matrix Spike Duplicate	Water	7470A	
Prep Batch: 720-1816				
LCS 720-1816/2-A	Lab Control Spike	Water	3010A	
LCSD 720-1816/3-A	Lab Control Spike Duplicate	Water	3010A	
MB 720-1816/1-A	Method Blank	Water	3010A	
720-401-1	EBHH-EB-1	Water	3010A	
720-401-36	EBHH-GW-1	Water	3010A	
720-413-M-9-H MS	Matrix Spike	Water	3010A	
720-413-M-9-I MSD	Matrix Spike Duplicate	Water	3010A	
Prep Batch: 720-1824				
LCS 720-1824/2-A	Lab Control Spike	Solid	3050B	
LCSD 720-1824/3-A	Lab Control Spike Duplicate	Solid	3050B	
MB 720-1824/1-A	Method Blank	Solid	3050B	
720-401-2	EBHH-25C	Solid	3050B	
720-401-4	EBHH-28B	Solid	3050B	
720-401-5	EBHH-28C	Solid	3050B	
720-401-7	EBHH-28A	Solid	3050B	
720-401-9	EBHH-29A	Solid	3050B	
720-401-9MS	Matrix Spike	Solid	3050B	
720-401-9MSD	Matrix Spike Duplicate	Solid	3050B	
720-401-13	EBHH-31A	Solid	3050B	
720-401-15	EBHH-33A	Solid	3050B	
720-401-17	EBHH-35A	Solid	3050B	
720-401-21	EBHH-37A	Solid	3050B	
720-401-22	EBHH-25A	Solid	3050B	
720-401-25	EBHH-39A	Solid	3050B	
720-401-28	EBHH-41A	Solid	3050B	
720-401-28MS	Matrix Spike	Solid	3050B	
720-401-28MSD	Matrix Spike Duplicate	Solid	3050B	
720-401-31	EBHH-43A	Solid	3050B	
720-401-33	EBHH-55A	Solid	3050B	
720-401-34	EBHH-56A	Solid	3050B	

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Prep Batch: 720-1825				
LCS 720-1825/2-A	Lab Control Spike	Solid	7471A	
LCSD 720-1825/3-A	Lab Control Spike Duplicate	Solid	7471A	
MB 720-1825/1-A	Method Blank	Solid	7471A	
720-401-2	EBHH-25C	Solid	7471A	
720-401-4	EBHH-28B	Solid	7471A	
720-401-5	EBHH-28C	Solid	7471A	
720-401-7	EBHH-28A	Solid	7471A	
720-401-9	EBHH-29A	Solid	7471A	
720-401-9MS	Matrix Spike	Solid	7471A	
720-401-9MSD	Matrix Spike Duplicate	Solid	7471A	
720-401-13	EBHH-31A	Solid	7471A	
720-401-15	EBHH-33A	Solid	7471A	
720-401-17	EBHH-35A	Solid	7471A	
720-401-21	EBHH-37A	Solid	7471A	
720-401-22	EBHH-25A	Solid	7471A	
720-401-25	EBHH-39A	Solid	7471A	
720-401-28	EBHH-41A	Solid	7471A	
720-401-28MS	Matrix Spike	Solid	7471A	
720-401-28MSD	Matrix Spike Duplicate	Solid	7471A	
720-401-31	EBHH-43A	Solid	7471A	
720-401-33	EBHH-55A	Solid	7471A	
Prep Batch: 720-2022				
LCS 720-2022/2-A	Lab Control Spike	Solid	7471A	
LCSD 720-2022/3-A	Lab Control Spike Duplicate	Solid	7471A	
MB 720-2022/1-A	Method Blank	Solid	7471A	
720-401-34	EBHH-56A	Solid	7471A	
720-401-34MS	Matrix Spike	Solid	7471A	
720-401-34MSD	Matrix Spike Duplicate	Solid	7471A	
Analysis Batch:720-1793				
LCS 720-1754/2-A	Lab Control Spike	Water	7470A	720-1754
LCSD 720-1754/3-A	Lab Control Spike Duplicate	Water	7470A	720-1754
MB 720-1754/1-A	Method Blank	Water	7470A	720-1754
720-401-1	EBHH-EB-1	Water	7470A	720-1754
720-401-36	EBHH-GW-1	Water	7470A	720-1754
720-413-M-9-B MS	Matrix Spike	Water	7470A	720-1754
720-413-M-9-C MSD	Matrix Spike Duplicate	Water	7470A	720-1754

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Analysis Batch:720-1924				
LCS 720-1816/2-A	Lab Control Spike	Water	6010B	720-1816
LCSD 720-1816/3-A	Lab Control Spike Duplicate	Water	6010B	720-1816
MB 720-1816/1-A	Method Blank	Water	6010B	720-1816
720-401-1	EBHH-EB-1	Water	6010B	720-1816
720-401-36	EBHH-GW-1	Water	6010B	720-1816
720-413-M-9-H MS	Matrix Spike	Water	6010B	720-1816
720-413-M-9-I MSD	Matrix Spike Duplicate	Water	6010B	720-1816
Analysis Batch:720-1924				
LCS 720-1824/2-A	Lab Control Spike	Solid	6010B	720-1824
LCSD 720-1824/3-A	Lab Control Spike Duplicate	Solid	6010B	720-1824
MB 720-1824/1-A	Method Blank	Solid	6010B	720-1824
720-401-2	EBHH-25C	Solid	6010B	720-1824
720-401-4	EBHH-28B	Solid	6010B	720-1824
720-401-5	EBHH-28C	Solid	6010B	720-1824
720-401-7	EBHH-28A	Solid	6010B	720-1824
720-401-9	EBHH-29A	Solid	6010B	720-1824
720-401-9MS	Matrix Spike	Solid	6010B	720-1824
720-401-9MSD	Matrix Spike Duplicate	Solid	6010B	720-1824
720-401-13	EBHH-31A	Solid	6010B	720-1824
720-401-15	EBHH-33A	Solid	6010B	720-1824
720-401-17	EBHH-35A	Solid	6010B	720-1824
720-401-21	EBHH-37A	Solid	6010B	720-1824
720-401-22	EBHH-25A	Solid	6010B	720-1824
720-401-25	EBHH-39A	Solid	6010B	720-1824
720-401-28	EBHH-41A	Solid	6010B	720-1824
720-401-28MS	Matrix Spike	Solid	6010B	720-1824
720-401-28MSD	Matrix Spike Duplicate	Solid	6010B	720-1824
720-401-31	EBHH-43A	Solid	6010B	720-1824
720-401-33	EBHH-55A	Solid	6010B	720-1824
720-401-34	EBHH-56A	Solid	6010B	720-1824

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Analysis Batch:720-1922				
LCS 720-1825/2-A	Lab Control Spike	Solid	7471A	720-1825
LCSD 720-1825/3-A	Lab Control Spike Duplicate	Solid	7471A	720-1825
MB 720-1825/1-A	Method Blank	Solid	7471A	720-1825
720-401-2	EBHH-25C	Solid	7471A	720-1825
720-401-4	EBHH-28B	Solid	7471A	720-1825
720-401-5	EBHH-28C	Solid	7471A	720-1825
720-401-7	EBHH-28A	Solid	7471A	720-1825
720-401-9	EBHH-29A	Solid	7471A	720-1825
720-401-9MS	Matrix Spike	Solid	7471A	720-1825
720-401-9MSD	Matrix Spike Duplicate	Solid	7471A	720-1825
720-401-13	EBHH-31A	Solid	7471A	720-1825
720-401-15	EBHH-33A	Solid	7471A	720-1825
720-401-17	EBHH-35A	Solid	7471A	720-1825
720-401-21	EBHH-37A	Solid	7471A	720-1825
720-401-22	EBHH-25A	Solid	7471A	720-1825
720-401-25	EBHH-39A	Solid	7471A	720-1825
720-401-28	EBHH-41A	Solid	7471A	720-1825
720-401-28MS	Matrix Spike	Solid	7471A	720-1825
720-401-28MSD	Matrix Spike Duplicate	Solid	7471A	720-1825
720-401-31	EBHH-43A	Solid	7471A	720-1825
720-401-33	EBHH-55A	Solid	7471A	720-1825
Analysis Batch:720-2039				
LCS 720-2022/2-A	Lab Control Spike	Solid	7471A	720-2022
LCSD 720-2022/3-A	Lab Control Spike Duplicate	Solid	7471A	720-2022
MB 720-2022/1-A	Method Blank	Solid	7471A	720-2022
720-401-34	EBHH-56A	Solid	7471A	720-2022
720-401-34MS	Matrix Spike	Solid	7471A	720-2022
720-401-34MSD	Matrix Spike Duplicate	Solid	7471A	720-2022

Client: Weston Solutions,Inc

Job Number: 720-401-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
General Chemistry				
Analysis Batch:720-1872				
MB 720-1872/1	Method Blank	Solid	160.3	
720-401-2	EBHH-25C	Solid	160.3	
720-401-4	EBHH-28B	Solid	160.3	
720-401-5	EBHH-28C	Solid	160.3	
720-401-6	EBHH-28D	Solid	160.3	
720-401-7	EBHH-28A	Solid	160.3	
720-401-8	EBHH-29B	Solid	160.3	
720-401-9	EBHH-29A	Solid	160.3	
720-401-10	EBHH-30C	Solid	160.3	
720-401-11	EBHH-30D	Solid	160.3	
720-401-12	EBHH-31D	Solid	160.3	
720-401-13	EBHH-31A	Solid	160.3	
720-401-14	EBHH-33C	Solid	160.3	
720-401-15	EBHH-33A	Solid	160.3	
720-401-16	EBHH-35C	Solid	160.3	
720-401-17	EBHH-35A	Solid	160.3	
720-401-18	EBHH-34D	Solid	160.3	
720-401-19	EBHH-36C	Solid	160.3	
720-413-A-4 DU	Duplicate	Solid	160.3	
Analysis Batch:720-2055				
MB 720-2055/1	Method Blank	Solid	160.3	
720-401-20	EBHH-37B	Solid	160.3	
720-401-21	EBHH-37A	Solid	160.3	
720-401-22	EBHH-25A	Solid	160.3	
720-401-23	EBHH-38B	Solid	160.3	
720-401-24	EBHH-39B	Solid	160.3	
720-401-25	EBHH-39A	Solid	160.3	
720-401-26	EBHH-40C	Solid	160.3	
720-401-27	EBHH-41B	Solid	160.3	
720-401-28	EBHH-41A	Solid	160.3	
720-401-29	EBHH-42D	Solid	160.3	
720-401-30	EBHH-43B	Solid	160.3	
720-401-31	EBHH-43A	Solid	160.3	
720-401-32	EBHH-44B	Solid	160.3	
720-401-33	EBHH-55A	Solid	160.3	
720-401-34	EBHH-56A	Solid	160.3	

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-2028

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 720-2028/18
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/19/2005 0929
Date Prepared: 11/19/2005 0929

Analysis Batch: 720-2028
Prep Batch: N/A
Units: ug/L

Instrument ID: Varian 3900F
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 40 mL
Final Weight/Volume: 40 mL

Analyte	Result	Qual	RL
1,1-Dichloroethene	ND		0.50
1,1-Dichloroethane	ND		0.50
Dichlorodifluoromethane	ND		0.50
Vinyl chloride	ND		0.50
Chloroethane	ND		1.0
Trichlorofluoromethane	ND		1.0
Methylene Chloride	ND		5.0
trans-1,2-Dichloroethene	ND		0.50
cis-1,2-Dichloroethene	ND		0.50
Chloroform	ND		1.0
1,1,1-Trichloroethane	ND		0.50
Carbon tetrachloride	ND		0.50
1,2-Dichloroethane	ND		0.50
Trichloroethene	ND		0.50
1,2-Dichloropropane	ND		0.50
Dichlorobromomethane	ND		0.50
trans-1,3-Dichloropropene	ND		0.50
cis-1,3-Dichloropropene	ND		0.50
1,1,2-Trichloroethane	ND		0.50
Tetrachloroethene	ND		0.50
Chlorodibromomethane	ND		0.50
Chlorobenzene	ND		0.50
Bromoform	ND		1.0
1,1,2,2-Tetrachloroethane	ND		0.50
1,3-Dichlorobenzene	ND		0.50
1,4-Dichlorobenzene	ND		0.50
1,2-Dichlorobenzene	ND		0.50
Chloromethane	ND		1.0
Bromomethane	ND		1.0
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		0.50
EDB	ND		0.50
1,2,4-Trichlorobenzene	ND		1.0

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Laboratory Control Sample - Batch: 720-2028

Method: 8260B
Preparation: 5030B

Lab Sample ID: LCS 720-2028/17
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/19/2005 0856
Date Prepared: 11/19/2005 0856

Analysis Batch: 720-2028
Prep Batch: N/A
Units: ug/L

Instrument ID: Varian 3900F
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 40 mL
Final Weight/Volume: 40 mL

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
1,1-Dichloroethene	20.0	19	95	65 - 125	
Trichloroethene	20.0	20	101	74 - 134	
Chlorobenzene	20.0	22	108	61 - 121	

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-2028

Method: 8260B
Preparation: 5030B

MS Lab Sample ID: 720-524-B-5 MS
Client Matrix: Water
Dilution: 200
Date Analyzed: 11/19/2005 1357
Date Prepared: 11/19/2005 1357

Analysis Batch: 720-2028
Prep Batch: N/A

Instrument ID: Varian 3900F
Lab File ID: c:\saturnws\data\200511\1
Initial Weight/Volume: 40 mL
Final Weight/Volume: 40 mL

MSD Lab Sample ID: 720-524-B-5 MSD
Client Matrix: Water
Dilution: 200
Date Analyzed: 11/19/2005 1431
Date Prepared: 11/19/2005 1431

Analysis Batch: 720-2028
Prep Batch: N/A

Instrument ID: Varian 3900F
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 40 mL
Final Weight/Volume: 40 mL

Analyte	% Rec. MS	% Rec. MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
1,1-Dichloroethene	99	106	65 - 125	7	20		
Trichloroethene	93	98	74 - 134	6	20		
Chlorobenzene	99	108	61 - 121	8	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1673

Method: 8270C
Preparation: 3510C

Lab Sample ID: MB 720-1673/1-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1257
Date Prepared: 11/11/2005 1325

Analysis Batch: 720-1849
Prep Batch: 720-1673
Units: ug/L

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 1000 mL
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result	Qual	RL
Naphthalene	ND		0.10
Acenaphthene	ND		0.10
Acenaphthylene	ND		0.10
Fluorene	ND		0.10
Phenanthrene	ND		0.10
Anthracene	ND		0.10
Benzo[a]anthracene	ND		0.20
Chrysene	ND		0.10
Benzo[a]pyrene	ND		0.10
Benzo[b]fluoranthene	ND		0.10
Benzo[k]fluoranthene	ND		0.10
Benzo[g,h,i]perylene	ND		0.10
Indeno[1,2,3-cd]pyrene	ND		0.10
Fluoranthene	ND		0.10
Pyrene	ND		0.10
Dibenz(a,h)anthracene	ND		1.0

Surrogate	% Rec	Acceptance Limits
2-Fluorobiphenyl	68	43 - 116
Terphenyl-d14	81	33 - 141

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1673

Method: 8270C
Preparation: 3510C

LCS Lab Sample ID: LCS 720-1673/2-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1325
Date Prepared: 11/11/2005 1325

Analysis Batch: 720-1849
Prep Batch: 720-1673
Units: ug/L

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\111505
Initial Weight/Volume: 1000 mL
Final Weight/Volume: 1 mL
Injection Volume:

LCSD Lab Sample ID: LCSD 720-1673/3-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1353
Date Prepared: 11/11/2005 1325

Analysis Batch: 720-1849
Prep Batch: 720-1673
Units: ug/L

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\111505\l
Initial Weight/Volume: 1000 mL
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Naphthalene	59	63	36 - 119	6	35		
Acenaphthene	60	63	56 - 118	4	30		
Acenaphthylene	66	66	54 - 126	1	35		
Fluorene	67	65	72 - 108	2	35		
Phenanthrene	67	69	44 - 125	3	35		
Anthracene	68	69	44 - 118	1	35		
Benzo[a]anthracene	76	74	42 - 133	2	35		
Chrysene	75	70	42 - 139	7	35		
Benzo[a]pyrene	73	70	32 - 148	4	35		
Benzo[b]fluoranthene	75	73	42 - 140	3	35		
Benzo[k]fluoranthene	81	78	26 - 145	4	35		
Benzo[g,h,i]perylene	78	76	10 - 140	3	35		
Indeno[1,2,3-cd]pyrene	74	73	10 - 150	1	35		
Fluoranthene	70	74	43 - 121	5	35		
Pyrene	70	70	52 - 115	1	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
2-Fluorobiphenyl	62	64	43 - 116
Terphenyl-d14	77	72	33 - 141

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1676

Method: 8270C
Preparation: 3550B

Lab Sample ID: MB 720-1676/21-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/25/2005 1552
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2238
Prep Batch: 720-1676
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.08 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result	Qual	RL
Naphthalene	ND		5.0
Acenaphthene	ND		5.0
Acenaphthylene	ND		5.0
Fluorene	ND		5.0
Phenanthrene	ND		5.0
Anthracene	ND		5.0
Benzo[a]anthracene	ND		5.0
Chrysene	ND		5.0
Benzo[a]pyrene	ND		5.0
Benzo[b]fluoranthene	ND		5.0
Benzo[k]fluoranthene	ND		5.0
Benzo[g,h,i]perylene	ND		5.0
Indeno[1,2,3-cd]pyrene	ND		5.0
Fluoranthene	ND		5.0
Pyrene	ND		5.0
Dibenz(a,h)anthracene	ND		5.0

Surrogate	% Rec	Acceptance Limits
2-Fluorobiphenyl	70	30 - 115
Terphenyl-d14	90	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1676

Method: 8270C
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1676/22-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/25/2005 1620
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2238
Prep Batch: 720-1676
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\112505
Initial Weight/Volume: 30.09 g
Final Weight/Volume: 1 mL
Injection Volume:

LCSD Lab Sample ID: LCSD 720-1676/23-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/25/2005 1648
Date Prepared: 11/11/2005 1442

Analysis Batch: 720-2238
Prep Batch: 720-1676
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\112505\l
Initial Weight/Volume: 30.09 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Naphthalene	67	63	21 - 133	5	35		
Acenaphthene	77	74	47 - 145	4	35		
Acenaphthylene	74	67	33 - 145	9	35		
Fluorene	82	74	59 - 121	10	35		
Phenanthrene	76	68	10 - 130	12	35		
Anthracene	82	75	27 - 133	9	35		
Benzo[a]anthracene	95	89	33 - 143	6	35		
Chrysene	75	70	17 - 168	6	35		
Benzo[a]pyrene	89	80	17 - 163	11	35		
Benzo[b]fluoranthene	91	80	24 - 159	12	35		
Benzo[k]fluoranthene	88	81	11 - 162	8	35		
Benzo[g,h,i]perylene	81	74	9 - 219	9	35		
Indeno[1,2,3-cd]pyrene	100	87	9 - 171	13	35		
Fluoranthene	83	76	26 - 137	10	35		
Pyrene	82	80	52 - 115	3	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
2-Fluorobiphenyl	75	68	30 - 115
Terphenyl-d14	91	85	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1676

Method: 8270C
Preparation: 3550B

MS Lab Sample ID:	720-401-9	Analysis Batch:	720-2253	Instrument ID:	Sat 2K2
Client Matrix:	Solid	Prep Batch:	720-1676	Lab File ID:	c:\saturnws\data\200511\
Dilution:	1.0			Initial Weight/Volume:	30.08 g
Date Analyzed:	11/26/2005 1736			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1442			Injection Volume:	

MSD Lab Sample ID:	720-401-9	Analysis Batch:	720-2253	Instrument ID:	Sat 2K2
Client Matrix:	Solid	Prep Batch:	720-1676	Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.06 g
Date Analyzed:	11/26/2005 1804			Final Weight/Volume:	5 mL
Date Prepared:	11/11/2005 1442			Injection Volume:	

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Naphthalene	74	72	21 - 133	3	35		
Acenaphthene	84	73	47 - 145	14	35		
Acenaphthylene	85	76	33 - 145	11	35		
Fluorene	89	80	59 - 121	11	35		
Phenanthrene	88	77	10 - 130	11	35		
Anthracene	81	81	27 - 133	0	35		
Benzo[a]anthracene	92	80	33 - 143	10	35		
Chrysene	76	73	17 - 168	3	35		
Benzo[a]pyrene	83	80	17 - 163	2	35		
Benzo[b]fluoranthene	86	73	24 - 159	11	35		
Benzo[k]fluoranthene	78	75	11 - 162	3	35		
Benzo[g,h,i]perylene	76	58	9 - 219	20	35		
Indeno[1,2,3-cd]pyrene	71	66	9 - 171	4	35		
Fluoranthene	75	71	26 - 137	3	35		
Pyrene	72	72	52 - 115	0	35		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
2-Fluorobiphenyl	91	81	30 - 115
Terphenyl-d14	87	88	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1830

Method: 8270C
Preparation: 3550B

Lab Sample ID: MB 720-1830/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/23/2005 2103
Date Prepared: 11/16/2005 1218

Analysis Batch: 720-2227
Prep Batch: 720-1830
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.09 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result	Qual	RL
Naphthalene	ND		5.0
Acenaphthene	ND		5.0
Acenaphthylene	ND		5.0
Fluorene	ND		5.0
Phenanthrene	ND		5.0
Anthracene	ND		5.0
Benzo[a]anthracene	ND		5.0
Chrysene	ND		5.0
Benzo[a]pyrene	ND		5.0
Benzo[b]fluoranthene	ND		5.0
Benzo[k]fluoranthene	ND		5.0
Benzo[g,h,i]perylene	ND		5.0
Indeno[1,2,3-cd]pyrene	ND		5.0
Fluoranthene	ND		5.0
Pyrene	ND		5.0
Dibenz(a,h)anthracene	ND		5.0

Surrogate	% Rec	Acceptance Limits
2-Fluorobiphenyl	76	30 - 115
Terphenyl-d14	87	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1830

Method: 8270C
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1830/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/23/2005 2131
Date Prepared: 11/16/2005 1218

Analysis Batch: 720-2227
Prep Batch: 720-1830
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\112305
Initial Weight/Volume: 30.14 g
Final Weight/Volume: 1 mL
Injection Volume:

LCSD Lab Sample ID: LCSD 720-1830/4-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/23/2005 2159
Date Prepared: 11/16/2005 1218

Analysis Batch: 720-2227
Prep Batch: 720-1830
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\112305\l
Initial Weight/Volume: 30.19 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Naphthalene	68	69	21 - 133	1	35		
Acenaphthene	72	74	47 - 145	3	35		
Acenaphthylene	66	66	33 - 145	1	35		
Fluorene	74	73	59 - 121	1	35		
Phenanthrene	77	72	10 - 130	6	35		
Anthracene	77	71	27 - 133	8	35		
Benzo[a]anthracene	91	86	33 - 143	6	35		
Chrysene	81	74	17 - 168	9	35		
Benzo[a]pyrene	85	81	17 - 163	5	35		
Benzo[b]fluoranthene	91	86	24 - 159	6	35		
Benzo[k]fluoranthene	91	90	11 - 162	1	35		
Benzo[g,h,i]perylene	86	85	9 - 219	1	35		
Indeno[1,2,3-cd]pyrene	95	124	9 - 171	26	35		
Fluoranthene	84	76	26 - 137	10	35		
Pyrene	84	83	52 - 115	2	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
2-Fluorobiphenyl	73	72	30 - 115
Terphenyl-d14	92	84	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1830

Method: 8270C
Preparation: 3550B

MS Lab Sample ID:	720-381-C-71-D MS	Analysis Batch: 720-2227	Instrument ID: Sat 2K2
Client Matrix:	Solid	Prep Batch: 720-1830	Lab File ID: c:\saturnws\data\200511\
Dilution:	1.0		Initial Weight/Volume: 30.00 g
Date Analyzed:	11/23/2005 2255		Final Weight/Volume: 1 mL
Date Prepared:	11/16/2005 1218		Injection Volume:

MSD Lab Sample ID:	720-381-C-71-E MSD	Analysis Batch: 720-2227	Instrument ID: Sat 2K2
Client Matrix:	Solid	Prep Batch: 720-1830	Lab File ID: c:\saturnws\data\200511\11
Dilution:	1.0		Initial Weight/Volume: 30.15 g
Date Analyzed:	11/23/2005 2322		Final Weight/Volume: 1 mL
Date Prepared:	11/16/2005 1218		Injection Volume:

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Naphthalene	68	67	21 - 133	2	35		
Acenaphthene	70	75	47 - 145	6	35		
Acenaphthylene	64	66	33 - 145	1	35		
Fluorene	73	73	59 - 121	1	35		
Phenanthrene	68	69	10 - 130	0	35		
Anthracene	70	72	27 - 133	3	35		
Benzo[a]anthracene	74	80	33 - 143	7	35		
Chrysene	66	71	17 - 168	6	35		
Benzo[a]pyrene	70	75	17 - 163	6	35		
Benzo[b]fluoranthene	73	81	24 - 159	10	35		
Benzo[k]fluoranthene	80	84	11 - 162	5	35		
Benzo[g,h,i]perylene	67	73	9 - 219	8	35		
Indeno[1,2,3-cd]pyrene	68	72	9 - 171	5	35		
Fluoranthene	72	76	26 - 137	5	35		
Pyrene	73	77	52 - 115	5	35		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
2-Fluorobiphenyl	69	71	30 - 115
Terphenyl-d14	78	82	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1830

Method: 8270C
Preparation: 3550B

MS Lab Sample ID:	720-401-28	Analysis Batch:	720-2238	Instrument ID:	Sat 2K2
Client Matrix:	Solid	Prep Batch:	720-1830	Lab File ID:	c:\saturnws\data\200511\
Dilution:	10			Initial Weight/Volume:	30.10 g
Date Analyzed:	11/25/2005 1456			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

MSD Lab Sample ID:	720-401-28	Analysis Batch:	720-2238	Instrument ID:	Sat 2K2
Client Matrix:	Solid	Prep Batch:	720-1830	Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.32 g
Date Analyzed:	11/25/2005 1524			Final Weight/Volume:	1 mL
Date Prepared:	11/16/2005 1218			Injection Volume:	

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Naphthalene	21	24	21 - 133	14	35		
Acenaphthene	35	44	47 - 145	19	35	*	*
Acenaphthylene	33	46	33 - 145	26	35		
Fluorene	31	45	59 - 121	28	35	*	*
Phenanthrene	-18	34	10 - 130	40	35	*	*
Anthracene	26	46	27 - 133	30	35	*	
Benzo[a]anthracene	4	23	33 - 143	15	35	*	*
Chrysene	-12	10	17 - 168	19	35	*	*
Benzo[a]pyrene	20	44	17 - 163	19	35		
Benzo[b]fluoranthene	-4	14	24 - 159	15	35	*	*
Benzo[k]fluoranthene	23	36	11 - 162	16	35		
Benzo[g,h,i]perylene	32	36	9 - 219	4	35		
Indeno[1,2,3-cd]pyrene	21	28	9 - 171	6	35		
Fluoranthene	-23	15	26 - 137	28	35	*	*
Pyrene	-39	-3	52 - 115	21	35	*	*

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
2-Fluorobiphenyl	31	42	30 - 115
Terphenyl-d14	45	52	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1670

Method: 8015B
Preparation: 3510C

Lab Sample ID: MB 720-1670/1-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/13/2005 1346
Date Prepared: 11/11/2005 1237

Analysis Batch: 720-1785
Prep Batch: 720-1670
Units: ug/L

Instrument ID: HP DRO5
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Diesel Range Organics [C10-C28]	ND		50
Motor Oil Range Organics [C24-C36]	ND		500

Surrogate	% Rec	Acceptance Limits
o-Terphenyl	70	60 - 130

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1670

Method: 8015B
Preparation: 3510C

LCS Lab Sample ID: LCS 720-1670/2-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/13/2005 1413
Date Prepared: 11/11/2005 1237

Analysis Batch: 720-1785
Prep Batch: 720-1670
Units: ug/L

Instrument ID: HP DRO5
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1670/3-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/13/2005 1440
Date Prepared: 11/11/2005 1237

Analysis Batch: 720-1785
Prep Batch: 720-1670
Units: ug/L

Instrument ID: HP DRO5
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Diesel Range Organics [C10-C28]	81	76	60 - 130	6	30		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
o-Terphenyl	79	75	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1672

Method: 8015B
Preparation: 3550B

Lab Sample ID: MB 720-1672/21-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0130
Date Prepared: 11/11/2005 1318

Analysis Batch: 720-1796
Prep Batch: 720-1672
Units: mg/Kg

Instrument ID: HP DRO5
Lab File ID: N/A
Initial Weight/Volume: 30.30 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Diesel Range Organics [C10-C28]	ND		0.99
Motor Oil Range Organics [C24-C36]	ND		50

Surrogate	% Rec	Acceptance Limits
o-Terphenyl	74	60 - 130

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1672

Method: 8015B
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1672/22-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0157
Date Prepared: 11/11/2005 1318

Analysis Batch: 720-1796
Prep Batch: 720-1672
Units: mg/Kg

Instrument ID: HP DRO5
Lab File ID: N/A
Initial Weight/Volume: 30.14 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1672/23-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 0224
Date Prepared: 11/11/2005 1318

Analysis Batch: 720-1796
Prep Batch: 720-1672
Units: mg/Kg

Instrument ID: HP DRO5
Lab File ID: N/A
Initial Weight/Volume: 30.07 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Diesel Range Organics [C10-C28]	86	86	60 - 130	1	30		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
o-Terphenyl	72	73	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1829

Method: 8015B
Preparation: 3550B

Lab Sample ID: MB 720-1829/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 0025
Date Prepared: 11/16/2005 1152

Analysis Batch: 720-1987
Prep Batch: 720-1829
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.28 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Diesel Range Organics [C10-C28]	ND		0.99
Motor Oil Range Organics [C24-C36]	ND		50

Surrogate	% Rec	Acceptance Limits
o-Terphenyl	76	60 - 130

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1829

Method: 8015B
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1829/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 2330
Date Prepared: 11/16/2005 1152

Analysis Batch: 720-1987
Prep Batch: 720-1829
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.12 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1829/4-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 2357
Date Prepared: 11/16/2005 1152

Analysis Batch: 720-1987
Prep Batch: 720-1829
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.25 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Diesel Range Organics [C10-C28]	80	85	60 - 130	6	30		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
o-Terphenyl	90	92	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1829

Method: 8015B
Preparation: 3550B

MS Lab Sample ID:	720-401-28	Analysis Batch: 720-1987	Instrument ID: HP DRO3
Client Matrix:	Solid	Prep Batch: 720-1829	Lab File ID: N/A
Dilution:	2.0		Initial Weight/Volume: 30.23 g
Date Analyzed:	11/17/2005 1332		Final Weight/Volume: 5 mL
Date Prepared:	11/16/2005 1152		Injection Volume:
			Column ID: PRIMARY

MSD Lab Sample ID:	720-401-28	Analysis Batch: 720-1987	Instrument ID: HP DRO3
Client Matrix:	Solid	Prep Batch: 720-1829	Lab File ID: N/A
Dilution:	2.0		Initial Weight/Volume: 30.28 g
Date Analyzed:	11/17/2005 1359		Final Weight/Volume: 5 mL
Date Prepared:	11/16/2005 1152		Injection Volume:
			Column ID: PRIMARY

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Diesel Range Organics [C10-C28]	65	77	60 - 130	6	30		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
o-Terphenyl	87	89	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1668

Method: 8081A
Preparation: 3510C

Lab Sample ID: MB 720-1668/2-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/16/2005 0137
Date Prepared: 11/11/2005 1219

Analysis Batch: 720-1910
Prep Batch: 720-1668
Units: ug/L

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 1000 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Aldrin	ND		0.060
Dieldrin	ND		0.060
Endrin aldehyde	ND		0.060
Endrin	ND		0.060
Endrin ketone	ND		0.060
Heptachlor	ND		0.060
Heptachlor epoxide	ND		0.060
4,4'-DDT	ND		0.060
4,4'-DDE	ND		0.060
4,4'-DDD	ND		0.060
Endosulfan I	ND		0.060
Endosulfan II	ND		0.060
alpha-BHC	ND		0.060
beta-BHC	ND		0.060
gamma-BHC (Lindane)	ND		0.060
delta-BHC	ND		0.060
Endosulfan sulfate	ND		0.060
Methoxychlor	ND		0.060
Toxaphene	ND		1.0
Chlordane (technical)	ND		1.0
alpha-Chlordane	ND		0.060
gamma-Chlordane	ND		0.060

Surrogate	% Rec	Acceptance Limits
Tetrachloro-m-xylene	99	62 - 123
DCB Decachlorobiphenyl	93	56 - 136

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1668

Method: 8081A
Preparation: 3510C

LCS Lab Sample ID: LCS 720-1668/3-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/16/2005 2204
Date Prepared: 11/11/2005 1219

Analysis Batch: 720-1910
Prep Batch: 720-1668
Units: ug/L

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 1000 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1668/4-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/16/2005 2233
Date Prepared: 11/11/2005 1219

Analysis Batch: 720-1910
Prep Batch: 720-1668
Units: ug/L

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 1000 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Aldrin	116	118	65 - 135	2	35		
Dieldrin	118	117	65 - 135	1	35		
Endrin	117	120	65 - 135	2	35		
Heptachlor	117	118	65 - 135	1	35		
4,4'-DDT	118	119	65 - 135	1	35		
gamma-BHC (Lindane)	119	120	65 - 135	1	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
Tetrachloro-m-xylene	105	107	62 - 123
DCB Decachlorobiphenyl	102	106	56 - 136

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1679

Method: 8081A
Preparation: 3550B

Lab Sample ID: MB 720-1679/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 2302
Date Prepared: 11/11/2005 1452

Analysis Batch: 720-1975
Prep Batch: 720-1679
Units: ug/Kg

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.13 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Aldrin	ND		2.0
Dieldrin	ND		2.0
Endrin aldehyde	ND		2.0
Endrin	ND		2.0
Endrin ketone	ND		2.0
Heptachlor	ND		2.0
Heptachlor epoxide	ND		2.0
4,4'-DDT	ND		2.0
4,4'-DDE	ND		2.0
4,4'-DDD	ND		2.0
Endosulfan I	ND		2.0
Endosulfan II	ND		2.0
alpha-BHC	ND		2.0
beta-BHC	ND		2.0
gamma-BHC (Lindane)	ND		2.0
delta-BHC	ND		2.0
Endosulfan sulfate	ND		2.0
Methoxychlor	ND		2.0
Toxaphene	ND		100
Chlordane (technical)	ND		50
alpha-Chlordane	ND		2.0
gamma-Chlordane	ND		2.0

Surrogate	% Rec	Acceptance Limits
Tetrachloro-m-xylene	93	50 - 125
DCB Decachlorobiphenyl	89	46 - 142

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1679

Method: 8081A
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1679/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/16/2005 2331
Date Prepared: 11/11/2005 1452

Analysis Batch: 720-1975
Prep Batch: 720-1679
Units: ug/Kg

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.15 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1679/4-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 0000
Date Prepared: 11/11/2005 1452

Analysis Batch: 720-1975
Prep Batch: 720-1679
Units: ug/Kg

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 30.02 g
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Aldrin	96	101	37 - 136	5	35		
Dieldrin	94	98	58 - 135	4	35		
Endrin	97	102	58 - 134	5	35		
Heptachlor	95	100	40 - 136	6	35		
4,4'-DDT	98	104	55 - 132	6	35		
gamma-BHC (Lindane)	92	97	37 - 137	5	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
Tetrachloro-m-xylene	90	91	50 - 125
DCB Decachlorobiphenyl	91	96	46 - 142

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1679

Method: 8081A
Preparation: 3550B

MS Lab Sample ID:	720-401-2	Analysis Batch: 720-1975	Instrument ID: Varian Pest 2
Client Matrix:	Solid	Prep Batch: 720-1679	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 30.23 g
Date Analyzed:	11/18/2005 0237		Final Weight/Volume: 10 mL
Date Prepared:	11/11/2005 1452		Injection Volume:
			Column ID: PRIMARY

MSD Lab Sample ID:	720-401-2	Analysis Batch: 720-1975	Instrument ID: Varian Pest 2
Client Matrix:	Solid	Prep Batch: 720-1679	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 30.11 g
Date Analyzed:	11/18/2005 0306		Final Weight/Volume: 10 mL
Date Prepared:	11/11/2005 1452		Injection Volume:
			Column ID: PRIMARY

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Aldrin	86	91	37 - 136	6	35		
Dieldrin	89	92	58 - 135	4	35		
Endrin	92	95	58 - 134	4	35		
Heptachlor	88	91	40 - 136	4	35		
4,4'-DDT	98	102	55 - 132	4	35		
gamma-BHC (Lindane)	87	91	37 - 137	5	35		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
Tetrachloro-m-xylene	85	89	50 - 125
DCB Decachlorobiphenyl	88	91	46 - 142

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1816

Method: 6010B
Preparation: 3010A

Lab Sample ID: MB 720-1816/1-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/17/2005 0841
Date Prepared: 11/16/2005 0856

Analysis Batch: 720-1924
Prep Batch: 720-1816
Units: mg/L

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Silver	ND		0.0050
Arsenic	ND		0.0050
Barium	ND		0.0050
Cadmium	ND		0.0020
Chromium	ND		0.0050
Lead	ND		0.0050
Selenium	ND		0.0050

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1816

Method: 6010B
Preparation: 3010A

LCS Lab Sample ID: LCS 720-1816/2-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/17/2005 0844
Date Prepared: 11/16/2005 0856

Analysis Batch: 720-1924
Prep Batch: 720-1816
Units: mg/L

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1816/3-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/17/2005 0848
Date Prepared: 11/16/2005 0856

Analysis Batch: 720-1924
Prep Batch: 720-1816
Units: mg/L

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Silver	97	93	80 - 120	4	20		
Arsenic	97	94	80 - 120	4	20		
Barium	99	95	80 - 120	4	20		
Cadmium	98	94	80 - 120	4	20		
Chromium	98	94	80 - 120	4	20		
Lead	98	94	80 - 120	4	20		
Selenium	97	94	80 - 120	4	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1816

Method: 6010B
Preparation: 3010A

MS Lab Sample ID:	720-413-M-9-H MS	Analysis Batch: 720-1924	Instrument ID: Varian ICP
Client Matrix:	Water	Prep Batch: 720-1816	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 50 mL
Date Analyzed:	11/17/2005 0918		Final Weight/Volume: 50 mL
Date Prepared:	11/16/2005 0856		

MSD Lab Sample ID:	720-413-M-9-I MSD	Analysis Batch: 720-1924	Instrument ID: Varian ICP
Client Matrix:	Water	Prep Batch: 720-1816	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 50 mL
Date Analyzed:	11/17/2005 0922		Final Weight/Volume: 50 mL
Date Prepared:	11/16/2005 0856		

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Silver	105	110	75 - 125	5	25		
Arsenic	106	112	75 - 125	6	25		
Barium	108	115	75 - 125	5	25		
Cadmium	103	109	75 - 125	5	25		
Chromium	105	111	75 - 125	5	25		
Lead	103	108	75 - 125	5	25		
Selenium	105	111	75 - 125	6	25		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1824

Method: 6010B
Preparation: 3050B

Lab Sample ID: MB 720-1824/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 1218
Date Prepared: 11/16/2005 1102

Analysis Batch: 720-1924
Prep Batch: 720-1824
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Silver	ND		1.0
Arsenic	ND		1.0
Barium	ND		1.0
Cadmium	ND		0.50
Chromium	ND		1.0
Lead	ND		1.0
Selenium	ND		2.0

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1824

Method: 6010B
Preparation: 3050B

LCS Lab Sample ID: LCS 720-1824/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 1221
Date Prepared: 11/16/2005 1102

Analysis Batch: 720-1924
Prep Batch: 720-1824
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1824/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 1225
Date Prepared: 11/16/2005 1102

Analysis Batch: 720-1924
Prep Batch: 720-1824
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Silver	105	104	80 - 120	0	20		
Arsenic	107	106	80 - 120	0	20		
Barium	108	108	80 - 120	0	20		
Cadmium	105	104	80 - 120	0	20		
Chromium	105	105	80 - 120	0	20		
Lead	103	102	80 - 120	0	20		
Selenium	108	108	80 - 120	0	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1824

Method: 6010B
Preparation: 3050B

MS Lab Sample ID:	720-401-9	Analysis Batch:	720-1924	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1824	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.03 g
Date Analyzed:	11/17/2005 1255			Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102				

MSD Lab Sample ID:	720-401-9	Analysis Batch:	720-1924	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1824	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.04 g
Date Analyzed:	11/17/2005 1259			Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Silver	93	85	75 - 125	10	20		
Arsenic	94	92	75 - 125	2	20		
Barium	84	84	75 - 125	0	20		
Cadmium	85	84	75 - 125	2	20		
Chromium	93	88	75 - 125	5	20		
Lead	85	101	75 - 125	10	20		
Selenium	87	86	75 - 125	2	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1824

Method: 6010B
Preparation: 3050B

MS Lab Sample ID:	720-401-28	Analysis Batch:	720-1924	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1824	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.01 g
Date Analyzed:	11/17/2005 1359			Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102				

MSD Lab Sample ID:	720-401-28	Analysis Batch:	720-1924	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1824	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.04 g
Date Analyzed:	11/17/2005 1403			Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 1102				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Silver	98	102	75 - 125	1	20		
Arsenic	96	99	75 - 125	1	20		
Barium	79	93	75 - 125	6	20		
Cadmium	90	93	75 - 125	1	20		
Chromium	77	77	75 - 125	1	20		
Lead	90	95	75 - 125	1	20		
Selenium	87	97	75 - 125	8	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1754

Method: 7470A
Preparation: 7470A

Lab Sample ID: MB 720-1754/1-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1309
Date Prepared: 11/15/2005 0859

Analysis Batch: 720-1793
Prep Batch: 720-1754
Units: mg/L

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 25 mL
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Mercury	ND		0.00020

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1754

Method: 7470A
Preparation: 7470A

LCS Lab Sample ID: LCS 720-1754/2-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1311
Date Prepared: 11/15/2005 0859

Analysis Batch: 720-1793
Prep Batch: 720-1754
Units: mg/L

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 25 mL
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1754/3-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1312
Date Prepared: 11/15/2005 0859

Analysis Batch: 720-1793
Prep Batch: 720-1754
Units: mg/L

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 25 mL
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Mercury	89	89	85 - 115	0	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1754

Method: 7470A
Preparation: 7470A

MS Lab Sample ID:	720-413-M-9-B MS	Analysis Batch: 720-1793	Instrument ID: FIMS 100
Client Matrix:	Water	Prep Batch: 720-1754	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 25 mL
Date Analyzed:	11/15/2005 1330		Final Weight/Volume: 50 mL
Date Prepared:	11/15/2005 0859		

MSD Lab Sample ID:	720-413-M-9-C MSD	Analysis Batch: 720-1793	Instrument ID: FIMS 100
Client Matrix:	Water	Prep Batch: 720-1754	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 25 mL
Date Analyzed:	11/15/2005 1331		Final Weight/Volume: 50 mL
Date Prepared:	11/15/2005 0859		

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	83	88	85 - 115	7	20	N	

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1825

Method: 7471A
Preparation: 7471A

Lab Sample ID: MB 720-1825/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 1254
Date Prepared: 11/16/2005 1112

Analysis Batch: 720-1922
Prep Batch: 720-1825
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Mercury	ND		0.050

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1825

Method: 7471A
Preparation: 7471A

LCS Lab Sample ID: LCS 720-1825/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 1255
Date Prepared: 11/16/2005 1112

Analysis Batch: 720-1922
Prep Batch: 720-1825
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1825/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 1256
Date Prepared: 11/16/2005 1112

Analysis Batch: 720-1922
Prep Batch: 720-1825
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Mercury	105	100	85 - 115	4	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1825

Method: 7471A
Preparation: 7471A

MS Lab Sample ID:	720-401-9	Analysis Batch: 720-1922	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1825	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.02 g
Date Analyzed:	11/17/2005 1304		Final Weight/Volume: 50 mL
Date Prepared:	11/16/2005 1112		

MSD Lab Sample ID:	720-401-9	Analysis Batch: 720-1922	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1825	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.01 g
Date Analyzed:	11/17/2005 1305		Final Weight/Volume: 50 mL
Date Prepared:	11/16/2005 1112		

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	104	97	85 - 115	4	20		

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1825

Method: 7471A
Preparation: 7471A

MS Lab Sample ID:	720-401-28	Analysis Batch: 720-1922	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1825	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.02 g
Date Analyzed:	11/17/2005 1317		Final Weight/Volume: 50 mL
Date Prepared:	11/16/2005 1112		

MSD Lab Sample ID:	720-401-28	Analysis Batch: 720-1922	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1825	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.05 g
Date Analyzed:	11/17/2005 1318		Final Weight/Volume: 50 mL
Date Prepared:	11/16/2005 1112		

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	96	100	85 - 115	0	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-2022

Method: 7471A
Preparation: 7471A

Lab Sample ID: MB 720-2022/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/21/2005 1022
Date Prepared: 11/21/2005 0740

Analysis Batch: 720-2039
Prep Batch: 720-2022
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Mercury	ND		0.050

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-2022

Method: 7471A
Preparation: 7471A

LCS Lab Sample ID: LCS 720-2022/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/21/2005 1024
Date Prepared: 11/21/2005 0740

Analysis Batch: 720-2039
Prep Batch: 720-2022
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-2022/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/21/2005 1025
Date Prepared: 11/21/2005 0740

Analysis Batch: 720-2039
Prep Batch: 720-2022
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Mercury	101	105	85 - 115	4	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-2022

Method: 7471A
Preparation: 7471A

MS Lab Sample ID:	720-401-34	Analysis Batch:	720-2039	Instrument ID:	FIMS 100
Client Matrix:	Solid	Prep Batch:	720-2022	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/21/2005 1028			Final Weight/Volume:	50 mL
Date Prepared:	11/21/2005 0740				

MSD Lab Sample ID:	720-401-34	Analysis Batch:	720-2039	Instrument ID:	FIMS 100
Client Matrix:	Solid	Prep Batch:	720-2022	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/21/2005 1030			Final Weight/Volume:	50 mL
Date Prepared:	11/21/2005 0740				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	-201	-199	85 - 115	1	20	N	N

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-1872

Method: 160.3
Preparation: N/A

Lab Sample ID: MB 720-1872/1
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1400
Date Prepared: N/A

Analysis Batch: 720-1872
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Result	Qual	RL
Percent Moisture	ND		0.010

Matrix Duplicate - Batch: 720-1872

Method: 160.3
Preparation: N/A

Lab Sample ID: 720-413-A-4 DU
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1400
Date Prepared: N/A

Analysis Batch: 720-1872
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Sample Result/Qual	Result	RPD	Limit	Qual
Percent Moisture	14	15			

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-401-1

Method Blank - Batch: 720-2055

Method: 160.3
Preparation: N/A

Lab Sample ID: MB 720-2055/1
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/18/2005 1340
Date Prepared: N/A

Analysis Batch: 720-2055
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Result	Qual	RL
Percent Moisture	ND		0.010

Calculations are performed before rounding to avoid round-off errors in calculated results.

720-401

SEVERN TRENT STL

STL San Francisco Chain of Custody

1220 Quarry Lane • Pleasanton CA 94566-4756
Phone: (925) 484-1919 • Fax: (925) 484-1096
Email: sflogin@stl-inc.com

Reference #: 114451

Date 11/10/05 Page 1 of 5

Report To

Attn:
Company: Weston
Address:
Phone: Email:
Bill To: Sampled By: Tom Fortney, Mindy Smith
Attn: Phone:

Analysis Request

Sample ID	Date	Time	Matrix	Preserv.	TPH EPA - ☐ 8015/8021 ☐ 8260B ☐ Gas w/ ☐ BTEX ☐ MTBE	Purgeable Aromatics BTEX EPA - ☐ 8021 ☐ 8260B	TEPH EPA 8015M* ☐ Silica Gel ☒ Diesel ☒ Motor Oil ☐ Other ___	Fuel Tests EPA 8260B: ☐ Gas ☐ BTEX ☐ Five Oxyenates ☐ DCA, EDB ☐ Ethanol	Purgeable Halocarbons (HVOCs) EPA 8021 by 8260B	Volatile Organics GC/MS (VOCs) ☐ EPA 8260B ☐ 624	Semivolatiles GC/MS ☐ EPA 8270 ☐ 625	Oil and Grease ☐ Petroleum (EPA 1664) ☐ Total	Pesticides ☒ EPA 8081 ☐ 608 PCBs ☐ EPA 8082 ☐ 608	PNAs by ☒ 8270 ☐ 8310	CAM17 Metals (EPA 6010/7470/7471)	Metals: ☐ Lead ☐ LUFT ☒ RCRA ☐ Other: ___	Low Level Metals by EPA 200.8/6020 (ICP-MS): ___	W.E.T (STLC) ☐ TCLP	Hexavalent Chromium ☐ pH (24h hold time for H_2O)	Spec Cond. ☐ Alkalinity ☐ TSS ☐ TDS ☐	Anions: ☐ Cl ☐ SO_4 ☐ NO_3 ☐ F ☐ Br ☐ NO_2 ☐ PO_4	Asbestos	HPLD		Number of Containers
5 EBHH-EB-1	11/10/05	0726	W				X						X	X		X								5
2 EBHH-25C	11/9/05	1734	S										X			X								2
2 EBHH-25D	11/9/05	1734											~~X~~			~~X~~						X		2
1 EBHH-28B	11/10/05	0730														X								1
EBHH-28C	11/10/05	0730														X								1
3 EBHH-28D	11/10/05	0730					X							X		H								3
5 EBHH-28A	11/10/05	1005					X						~~X~~	X		X					X			5
2 EBHH-29B	11/10/05	0745					X							X										2
~~EBHH-29C~~	~~11/10/05~~	~~0745~~																						
~~EBHH-29D~~	~~11/10/05~~	~~0745~~																						

MS
MSD

Project Info.

Project Name: EBHH Edes Ave.
Project#:
PO#:
Credit Card#:

Sample Receipt

of Containers:
Head Space:
Temp: 2°C
Conforms to record:

TAT: 5 Day | 72h | 48h | 24h | Other:

Report: ☐ Routine ☐ Level 3 ☐ Level 4 ☐ EDD ☐ State Tank Fund EDF ☐ Global ID _____
Special Instructions / Comments:

*STL SF reports 8015M from C_9-C_{24} (industry norm). Default for 8015B is C_{10}-C_{28}

1) Relinquished by:	2) Relinquished by:	3) Relinquished by:
Signature: M Smith Time: 1730	Signature: Time:	Signature: Time:
Printed Name: Mindy Smith Date: 11/10/05	Printed Name: Date:	Printed Name: Date:
Company: Weston	Company:	Company:

1) Received by:	2) Received by:	3) Received by:
Signature: SB Bull Time: 1730	Signature: Time:	Signature: Time:
Printed Name: T Bullock Date: 11/10/05	Printed Name: Date:	Printed Name: Date:
Company: STL-SF	Company:	Company:

Rev 06/04

SEVERN TRENT STL

STL San Francisco Chain of Custody

1220 Quarry Lane • Pleasanton CA 94566-4756
Phone: (925) 484-1919 Fax (925) 484-1096
Email: slogin@stl-inc.com

720-401

Reference #: 114451

Date 11/10/05 Page 2 of 5

Report To

Attn:
Company: Weston
Address:
Phone: Email:
Bill To: Sampled By: Jim Fertner, Mindy Smith
Attn: Phone:

Analysis Request

Sample ID	Date	Time	Matrix	Preserv.	TPH EPA - ☐ 8015/8021 ☐ 8260B ☐ Gas w/ ☐ BTEX ☐ MTBE	Purgeable Aromatics BTEX EPA - ☐ 8021 ☐ 8260B	TEPH EPA 8015M* ☐ Silica Gel ☒ Diesel ☒ Motor Oil ☐ Other ___	Fuel Tests EPA 8260B: ☐ Gas ☐ BTEX ☐ Five Oxyenates ☐ DCA, EDB ☐ Ethanol	Purgeable Halocarbons (HVOCs) EPA 8021 by 8260B	Volatile Organics GC/MS (VOCs) ☐ EPA 8260B ☐ 624	Semivolatiles GC/MS ☐ EPA 8270 ☐ 625	Oil and Grease ☐ Petroleum (EPA 1664) ☐ Total	Pesticides ☒ EPA 8081 ☐ 608 PCBs ☐ EPA 8082 ☐ 608	PNAs by ☒ 8270 ☐ 8310	CAM17 Metals (EPA 6010/7470/7471)	Metals: ☐ Lead ☐ LUFT ☒ RCRA ☐ Other:	Low Level Metals by EPA 200.8/6020 (ICP-MS):	W.E.T (STLC) ☐ TCLP	Hexavalent Chromium ☐ pH (24h hold time for H_2O)	Spec Cond. ☐ Alkalinity ☐ TSS ☐ TDS	Anions: ☐ Cl ☐ SO_4 ☐ NO_3 ☐ F ☐ Br ☐ NO_2 ☐ PO_4	Asbestos	HOLD	MS/MSD	Number of Containers
7 EBHH-29A	11/10/05	1010	S				X							X		X						X		2*	7
~~EBHH-30B~~		~~0800~~																							~~2~~
EBHH-30C		0800					X							X											2
EBHH-30D		0800					X							X											2
~~EBHH-30A~~																									~~2~~
EBHH-31D		0815					X							X											2
EBHH-31A		1025					X							X		X						X			2
EBHH-33C		0835					X							X											2
EBHH-33A		1035					X							X		X						X			2
EBHH-35C		0910					X							X											2

Project Info.

Project Name: EBHH Edes Ave.
Project#:
PO#:
Credit Card#:

Sample Receipt

# of Containers:
Head Space:
Temp: 2°C
Conforms to record:

TAT: 5 Day | 72h | 48h | 24h | Other:

Report: ☐ Routine ☐ Level 3 ☐ Level 4 ☐ EDD ☐ State Tank Fund EDF ☐ Global ID _____
Special Instructions / Comments:

*STL SF reports 8015M from C_9-C_{24} (industry norm). Default for 8015B is C_{10}-C_{28}

1) Relinquished by:	2) Relinquished by:	3) Relinquished by:
Signature: [signature] Time: 1730	Signature: Time:	Signature: Time:
Printed Name: Mindy Smith Date: 11/10/05	Printed Name: Date:	Printed Name: Date:
Company: Weston	Company:	Company:

1) Received by:	2) Received by:	3) Received by:
Signature: [signature] Time: 1730	Signature: Time:	Signature: Time:
Printed Name: T Bullock Date: 11/10/05	Printed Name: Date:	Printed Name: Date:
Company: STL-SF	Company:	Company:

Rev 08/04

SEVERN TRENT STL

STL San Francisco Chain of Custody

1220 Quarry Lane Pleasanton CA 94566-4756
Phone: (925) 484-1919 Fax: (925) 484-1096
Email: stlogin@stl-inc.com

120-401

Reference #: 11445

Date 11/10/05 Page 3 of 5

Report To

Attn:
Company: Weston
Address:
Phone: Email:
Bill To: Sampled By: Tom Fertig, Mindy Smith
Attn: Phone:

Analysis Request

Sample ID	Date	Time	Matrix	Preserv.	TPH EPA - ☐ 8015/8021 ☐ 8260B ☐ Gas w/ ☐ BTEX ☐ MTBE	Purgeable Aromatics BTEX EPA - ☐ 8021 ☐ 8260B	TEPH EPA 8015M* ☐ Silica Gel ☒ Diesel ☐ Motor Oil ☐ Other	Fuel Tests EPA 8260B: ☐ Gas ☐ BTEX ☐ Five Oxyenates ☐ DCA, EDB ☐ Ethanol	Purgeable Halocarbons (HVOCs) EPA 8021 by 8260B	Volatile Organics GC/MS (VOCs) ☐ EPA 8260B ☐ 624	Semivolatiles GC/MS ☐ EPA 8270 ☐ 625	Oil and Grease ☐ Petroleum (EPA 1664) ☐ Total	Pesticides ☒ EPA 8081 ☐ 608 PCBs ☐ EPA 8082 ☐ 608	PNAs by ☒ 8270 ☐ 8310	CAM17 Metals (EPA 6010/7470/7471)	Metals: ☐ Lead ☐ LUFT ☒ RCRA ☐ Other	Low Level Metals by EPA 200.8/6020 (ICP-MS):	☐ W.E.T (STLC) ☐ TCLP	☐ Hexavalent Chromium ☐ pH (24h hold time for H_2O)	☐ Spec Cond. ☐ TSS ☐ Alkalinity ☐ TDS	Anions: ☐ Cl ☐ SO_4 ☐ NO_3 ☐ F ☐ Br ☐ NO_2 ☐ PO_4	Asbestos	HOLD		Number of Containers
EBHH-35A	11/10/05	1045	S				X							X		X						X			5
EBHH-34D		0845					X							X											5
EBHH-36C		0925					X							X											5
EBHH-37B		1048					X							X											5
EBHH-32A		1330					X							X		X						X			5
EBHH-25A		0955					X							X		X						X			5
EBHH-38B		1100					X							X											5
~~EBHH-38A~~							X							X		X						X			5
EBHH-39B		1114	S				X							X											5
EBHH-39A		1345	S				X							X		X						X			5

Project Info.

Project Name: EBHH Edes Ave
Project#:
PO#:
Credit Card#:

Sample Receipt

of Containers:
Head Space:
Temp: 2°c
Conforms to record:

TAT: 5 Day | 72h | 48h | 24h | Other:

Report: ☐ Routine ☐ Level 3 ☐ Level 4 ☐ EDD ☐ State Tank Fund EDF ☐ Global ID _____
Special Instructions / Comments:

*STL SF reports 8015M from C_9-C_{24} (industry norm). Default for 8015B is C_{10}-C_{28}.

1) Relinquished by:	2) Relinquished by:	3) Relinquished by:
Signature: [signature] Time: 1730	Signature: Time:	Signature: Time:
Printed Name: Mindy Smith Date: 11/10/05	Printed Name: Date:	Printed Name: Date:
Company: Weston	Company:	Company:

1) Received by:	2) Received by:	3) Received by:
Signature: [signature] Time: 1730	Signature: Time:	Signature: Time:
Printed Name: T Bulbeck Date: 11/10/05	Printed Name: Date:	Printed Name: Date:
Company: STL-SF	Company:	Company:

Rev 06/04

SEVERN TRENT STL

STL San Francisco Chain of Custody

1220 Quarry Lane Pleasanton CA 94566-4756
Phone: (925) 484-1919 Fax: (925) 484-1096
Email: sflocsn@stl-inc.com

726-401

Reference #: 114451

Date 11/10/05 Page 4 of 5

Report To

Attn:
Company: Weston
Address:
Phone: Email:
Bill To: Sampled By: Tom Fortner, Mindy Smith
Attn: Phone:

Analysis Request

Sample ID	Date	Time	Matrix	Preserv.	TPH EPA - ☐ 8015/8021 ☐ 8260B ☐ Gas w/ ☐ BTEX ☐ MTBE	Purgeable Aromatics BTEX EPA - ☐ 8021 ☐ 8260B	TEPH EPA 8015M* ☐ Silica Gel ☒ Diesel ☒ Motor Oil ☐ Other ___	Fuel Tests EPA 8260B ☐ Gas ☐ BTEX ☐ Five Oxyenates ☐ DCA, EDB ☐ Ethanol	Purgeable Halocarbons (HVOCs) EPA 8021 by 8260B	Volatile Organics GC/MS (VOCs) ☐ EPA 8260B ☐ 624	Semivolatiles GC/MS ☐ EPA 8270 ☐ 625	Oil and Grease ☐ Petroleum (EPA 1664) ☐ Total	Pesticides ☒ EPA 8081 ☐ 608 PCBs ☐ EPA 8082 ☐ 608	PNAs by ☒ 8270 ☐ 8310	CAM17 Metals (EPA 6010/7470/7471)	Metals: ☐ Lead ☐ LUFT ☒ RCRA ☐ Other: ___	Low Level Metals by EPA 200.8/6020 (ICP-MS): ___	☐ W.E.T (STLC) ☐ TCLP	☐ Hexavalent Chromium ☐ pH (24h hold time for H_2O)	☐ Spec Cond. ☐ Alkalinity ☐ TSS ☐ TDS ☐	Anions: ☐ Cl ☐ SO_4 ☐ NO_3 ☐ F ☐ Br ☐ NO_2 ☐ PO_4	Asbestos	Hold	MS/MSD	Number of Containers
EBHH-40C	11/10/05	1124	S				X							X											2
EBHH-41B		1134					X							X											2
EBHH-41A		1355					X							X		X						X		X	6
EBHH-42D		1150					X							X											2
EBHH-43B		1201					X							X											2
EBHH-43A		1410					X							X		X						X			5
EBHH-44B		1218					X							X											2
EBHH-55A		1400					X							X		X									4
EBHH-56A		1430					X							X		X									4
EBHH-GW-2		1328	W				X							X											3

Project Info.

Project Name: EBHH Edes Ave.
Project#:
PO#:
Credit Card#:

Sample Receipt

of Containers:
Head Space:
Temp: 2°c
Conforms to record:

TAT: 5 Day | 72h | 48h | 24h | Other:

Report: ☐ Routine ☐ Level 3 ☐ Level 4 ☐ EDD ☐ State Tank Fund EDF ☐ Global ID _____
Special Instructions / Comments:

*STL SF reports 8015M from C_9-C_{24} (industry norm). Default for 8015B is C_{10}-C_{28}

1) Relinquished by:	2) Relinquished by:	3) Relinquished by:
Signature: [signature] Time: 1730	Signature: Time:	Signature: Time:
Printed Name: Mindy Smith Date: 11/10/05	Printed Name: Date:	Printed Name: Date:
Company: Weston	Company:	Company:

1) Received by:	2) Received by:	3) Received by:
Signature: [signature] Time: 1730	Signature: Time:	Signature: Time:
Printed Name: L. Bulleac Date: 11/10/05	Printed Name: Date:	Printed Name: Date:
Company: STL-ST	Company:	Company:

Rev 06/04

SEVERN TRENT STL

STL San Francisco Chain of Custody

1220 Quarry Lane • Pleasanton CA 94566-4756
Phone: (925) 484-1919 Fax: (925) 484-1096
Email: sflogin@stl-inc.com

720401

Reference #: 114457

Date 11/10/05 Page 5 of 5

Report To

Attn:
Company: Weston
Address:
Phone: Email: MindySmith
Bill To: Sampled By: Tom Fortner
Attn: Phone:

Analysis Request

Sample ID	Date	Time	Matrix	Preserv.	TPH EPA - ☐ 8015/8021 ☐ 8260B ☐ Gas w/ ☐ BTEX ☐ MTBE	Purgeable Aromatics BTEX EPA - ☐ 8021 ☐ 8260B	TEPH EPA 8015M* ☐ Silica Gel ☒ Diesel ☒ Motor Oil ☐ Other ___	Fuel Tests EPA 8260B: ☐ Gas ☐ BTEX ☐ Five Oxyenates ☐ DCA, EDB ☐ Ethanol	Purgeable Halocarbons (HVOCs) EPA 8021 by 8260B	Volatile Organics GC/MS (VOCs) ☐ EPA 8260B ☐ 624	Semivolatiles GC/MS ☐ EPA 8270 ☐ 625	Oil and Grease ☐ Petroleum (EPA 1664) ☐ Total	Pesticides ☒ EPA 8081 ☐ 608 PCBs ☐ EPA 8082 ☐ 608	PNAs by ☐ 8270 ☐ 8310	CAM17 Metals (EPA 6010/7470/7471)	Metals: ☐ Lead ☐ LUFT ☒ RCRA ☐ Other: ___	Low Level Metals by EPA 200.8/6020 (ICP-MS): ___	W.E.T (STLC) ☐ TCLP	Hexavalent Chromium ☐ pH (24h hold time for H_2O)	Spec Cond. ☐ Alkalinity ☐ TSS ☐ TDS	Anions: ☐ Cl ☐ SO_4 ☐ NO_3 ☐ F ☐ Br ☐ NO_2 ☐ PO_4	Asbestos	HOLD	MS/MSD	Number of Containers
EBHH-GW-1	11/10/05	1446	W						X							X									4

Project Info.

Project Name: EBHH Edes Ave
Project#:
PO#:
Credit Card#:

Sample Receipt

of Containers: 4
Head Space:
Temp: 2°c
Conforms to record:

TAT: 5 Day | 72h | 48h | 24h | Other:

Report: ☐ Routine ☐ Level 3 ☐ Level 4 ☐ EDD ☐ State Tank Fund EDF ☐ Global ID _____
Special Instructions / Comments:

*STL SF reports 8015M from C_9-C_{24} (industry norm). Default for 8015B is C_{10}-C_{28}

1) Relinquished by:	2) Relinquished by:	3) Relinquished by:
Signature: [signature] Time: 1730	Signature: Time:	Signature: Time:
Printed Name: Mindy Smith Date: 11/10/05	Printed Name: Date:	Printed Name: Date:
Company: Weston	Company:	Company:

1) Received by:	2) Received by:	3) Received by:
Signature: [signature] Time: 1730	Signature: Time:	Signature: Time:
Printed Name: T. Bullock Date: 11/10/05	Printed Name: Date:	Printed Name: Date:
Company: STL-SF	Company:	Company:

Rev 06/04

STL San Francisco

Sample Receipt Checklist

Submission #:2005- [redacted]

Checklist completed by:		DATE	
Courier: ☐ STL SF	Courier ☐ Fedex UPS Other		Client ☑

	Log-In Details	Yes	No	Comments
1	Custody seals intact on shipping container/samples		✓	
2	Chain of custody present?	✓		
3	Chain of custody signed when relinquished and received?	✓		☐ Picked-Up at Secure Location ☐ Client signed-off at time prior to pick-up
4	All samples checked when COC relinquished		✓	
5	Chain of custody agrees with sample labels?			
6	Samples in proper container/bottle?			
7	Sample containers intact?			
8	Sufficient sample volume for indicated test?			
9	All samples received within holding time?			

Cooler Temperature Compliance Check

Temperature Blank Reading	Cooler Sample Temperature #1	#2	#3	Average
2°c				

If no trip blank is submitted individual temperatures must be taken as per SOP →

2 coolers

Reason for Elevated Temperature	Samples with Temp > 6°C - Comments
☐ - Ice Melted ☐ Insufficient Ice ☐	
☐ Samp. in boxes ☐ Sampled < 4hr ☐ Ice not req.	

VOA Sample Inspection

Are bubbles present in any of the VOA vials ?	Sample #	Small	Med.	Large	Samples with broken, cracked or leaking containers
		O	O	O	
		☐	☐	☐	
		☐	☐	☐	
		☐	☐	☐	

Water - pH acceptable upon receipt?	Yes	No	Samples with Unacceptable pH
	☐	☐	

☐ pH adjusted– Preservative used: ☐ HNO_3 ☐ HCl ☐ H_2SO_4 ☐ NaOH ☐ ZnOAc –Lot #(s) _____

Comments:

Project Management [Routing for instruction of indicated discrepancy(ies)]

Project Manager: (initials) _____ Date: _____/_____/05 Client contacted: Yes ☐ No ☐

Summary of discussion:

Corrective Action (per PM/Client):

2005 Checklist Ver. 2.0

LOGIN SAMPLE RECEIPT CHECK LIST

Client: Weston Solutions,Inc

Job Number: 720-401-1

Login Number: 401

Question	T/F/NA	Comment
Radioactivity either was not measured or, if measured, is at or below background	NA	
The cooler's custody seal, if present.	NA	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	True	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present	True	
Samples do not require splitting or compositing	True	

2595-543

Forensic

STL San Francisco
1220 Quarry Lane
Pleasanton, CA 94566
Phone (925) 484-1919 Fax (925) 484-1096

Chain of Custody

SEVERN TRENT STL
Severn Trent Laboratories, Inc.

Client Contact: Shipping/Receiving@ Block Environmental 2461 Island Way Pleasant Hill CA 94523 Phone: 925 [illegible]-7200 Fax: Project Name: EBHn EDES AVE OAKLAND Site PO #:	Project Manager: Sidhu, Surinder Tel/Fax: Analysis Due Date: 11/17/2005 Shipped with Order No: [illegible]	Site Contact: Lab Contact:	Date: Carrier No:	COC No: 720-49 Job Number: 720-401-1

Sample Identification	Sample Date	Sample Time	Sample Type	Matrix	# of Cont.	Filtered Sample	SUBCONTRACT: Asbestos	Special Instructions/Note:
720-401-7	11/10/05	10:05		Solid	1		X	
720-401-9	11/10/05	10:10		Solid	1		X	
720-401-13	11/10/05	10:25		Solid	1			
720-401-15	11/10/05	10:35		Solid	1			
720-401-17	11/10/05	10:45		Solid	1			
720-401-21	11/10/05	13:30		Solid	1			
720-401-22	11/10/05	9:55		Solid	1			
720-401-25	11/10/05	13:45		Solid	1		X	
720-401-26	11/10/05	13:55		Solid	1			
720-401-31	11/10/05	14:10		Solid	1			

Possible Hazard Identification: ☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison ☐ Unk.

Sample Disposal (A fee may be assessed if samples are retained longer than 1 month): ☐ Return To Client ☐ Disposal By Lab ☐ Archive For ____ Months

Special Instructions/QC Requirements:

Relinquished by: [signature]	Company: STL SF	Date/Time: 11-11-05 1100	Received by: [signature]	Company: 11/11/05 1100
Relinquished by: [signature]	Company: 11/11/05 1525	Date/Time: 11/11/05 [illegible]	Received by: [signature]	Company: FedEx
Relinquished by:	Company:	Date/Time:	Received by:	Company:

Comments:

ANALYTICAL REPORT

Job Number: 720-413-1

Job Description: EBHH EDES AVE OAKLAND

For:

Weston Solutions,Inc
1575 Treat Blvd Suite 212
Use the other WESTON SOLUTIONS, CA 94598

Attention: Mr. Tom Fortner

Surinder Sidhu

Surinder Sidhu
Project Manager I
ssidhu@stl-inc.com
12/02/2005

Case Narrative

Non Conformance Summary for job: 720-J413-1

Client: Weston Solutions,Inc
Date: 12/2/2005

Semi Volatiles MS Analysis

Effected Login

720-413

MS/MSD could not be evaluated due to high analyte

Due to diluted out ,the MS/MSD could not be evaluated.

METHOD SUMMARY

Client: Weston Solutions,Inc

Job Number: 720-413-1

Description	Lab Location	Method	Preparation Method
Matrix: Solid			
Volatile Organic Compounds by GC/MS	STL-SF	SW846 8260B	
Closed System Purge & Trap/Laboratory	STL-SF		SW846 5035
Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)	STL-SF	SW846 8270C	
Ultrasonic Extraction	STL-SF		SW846 3550B
Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)	STL-SF	SW846 8015B	
Ultrasonic Extraction	STL-SF		SW846 3550B
Inductively Coupled Plasma - Atomic Emission Spectrometry	STL-SF	SW846 6010B	
Acid Digestion of Sediments, Sludges, and Soils	STL-SF		SW846 3050B
Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)	STL-SF	SW846 7471A	
Mercury in Solid or Semi-Solid Waste (Manual	STL-SF		SW846 7471A
Percent Moisture	STL-SF	EPA 160.3	
Asbestos	STL-SF	EPA	
Matrix: Water			
Volatile Organic Compounds by GC/MS	STL-SF	SW846 8260B	
Purge-and-Trap	STL-SF		SW846 5030B
Volatile Organic Compounds by GC/MS (Low Level)	STL-SF	SW846 8260B	
Purge-and-Trap	STL-SF		SW846 5030B
Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)	STL-SF	SW846 8270C	
Separatory Funnel Liquid-Liquid Extraction	STL-SF		SW846 3510C
Nonhalogenated Organics using GC/FID -Modified (Gasoline Range Organics)	STL-SF	SW846 8015B	
Purge-and-Trap	STL-SF		SW846 5030B
Aromatic and Halogenated VOCs by Gas Chromatography using PID or ECD	STL-SF	SW846 8021B	
Purge-and-Trap	STL-SF		SW846 5030B
Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)	STL-SF	SW846 8015B	
Separatory Funnel Liquid-Liquid Extraction	STL-SF		SW846 3510C
Organochlorine Pesticides by Gas Chromatography	STL-SF	SW846 8081A	
Separatory Funnel Liquid-Liquid Extraction	STL-SF		SW846 3510C
Inductively Coupled Plasma - Atomic Emission Spectrometry	STL-SF	SW846 6010B	
Acid Digestion of Aqueous Samples and Extracts	STL-SF		SW846 3010A
Mercury in Liquid Waste (Manual Cold Vapor Technique)	STL-SF	SW846 7470A	

METHOD SUMMARY

Client: Weston Solutions,Inc

Job Number: 720-413-1

Matrix: Water

Mercury in Liquid Waste (Manual Cold Vapor Technique)	STL-SF	SW846 7470A	
Mercury in Liquid Waste (Manual Cold Vapor	STL-SF		SW846 7470A

LAB REFERENCES:

STL-SF = STL-San Francisco

METHOD REFERENCES:

EPA - US Environmental Protection Agency

SW846 - "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986
And Its Updates.

SAMPLE SUMMARY

Client: Weston Solutions,Inc

Job Number: 720-413-1

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
720-413-1	EBHH-EB-2	Water	11/11/2005 0715	11/11/2005 1340
720-413-2	EBHH-EB-4	Water	11/11/2005 0720	11/11/2005 1340
720-413-3	EBHH-45B	Solid	11/11/2005 0806	11/11/2005 1340
720-413-4	EBHH-45D	Solid	11/11/2005 0806	11/11/2005 1340
720-413-5	EBHH-48B	Solid	11/11/2005 0806	11/11/2005 1340
720-413-6	EBHH-49C	Solid	11/11/2005 0825	11/11/2005 1340
720-413-7	EBHH-50C	Solid	11/11/2005 0830	11/11/2005 1340
720-413-8	EBHH-58C	Solid	11/11/2005 0835	11/11/2005 1340
720-413-9	EBHH-GW-4	Water	11/11/2005 1030	11/11/2005 1340
720-413-10	EBHH-GW-5	Water	11/11/2005 1100	11/11/2005 1340
720-413-11	EBHH-GW-3	Water	11/11/2005 1130	11/11/2005 1340
720-413-12	EBHH-EB3	Water	11/11/2005 1205	11/11/2005 1340
720-413-13	EBHH-50A	Solid	11/11/2005 1045	11/11/2005 1340
720-413-14	EBHH-49A	Solid	11/11/2005 1145	11/11/2005 1340
720-413-15	EBHH-57A	Solid	11/11/2005 1200	11/11/2005 1340

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-EB-4**

Lab Sample ID: 720-413-2
Client Matrix: Water

Date Sampled: 11/11/2005 0720
Date Received: 11/11/2005 1340

8260B Volatile Organic Compounds by GC/MS (Low Level)

Method: 8260B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/23/2005 2148
Date Prepared: 11/23/2005 2148

Analysis Batch: 720-2239

Instrument ID: Saturn 2K3
Lab File ID: d:\data\200511\112305\720-
Initial Weight/Volume: 40 mL
Final Weight/Volume: 40 mL

Analyte	Result (ug/L)	Qualifier	RL
1,1-Dichloroethene	ND		0.50
1,1-Dichloroethane	ND		0.50
Dichlorodifluoromethane	ND		0.50
Vinyl chloride	ND		0.50
Chloroethane	ND		1.0
Trichlorofluoromethane	ND		1.0
Methylene Chloride	ND		5.0
trans-1,2-Dichloroethene	ND		0.50
cis-1,2-Dichloroethene	ND		0.50
Chloroform	ND		1.0
1,1,1-Trichloroethane	ND		0.50
Carbon tetrachloride	ND		0.50
1,2-Dichloroethane	ND		0.50
Trichloroethene	ND		0.50
1,2-Dichloropropane	ND		0.50
Dichlorobromomethane	ND		0.50
trans-1,3-Dichloropropene	ND		0.50
cis-1,3-Dichloropropene	ND		0.50
1,1,2-Trichloroethane	ND		0.50
Tetrachloroethene	ND		0.50
Chlorodibromomethane	ND		0.50
Chlorobenzene	ND		0.50
Bromoform	ND		1.0
1,1,2,2-Tetrachloroethane	ND		0.50
1,3-Dichlorobenzene	ND		0.50
1,4-Dichlorobenzene	ND		0.50
1,2-Dichlorobenzene	ND		0.50
Chloromethane	ND		1.0
Bromomethane	ND		1.0
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		0.50
EDB	ND		0.50
1,2,4-Trichlorobenzene	ND		1.0

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-EB-4**

Lab Sample ID: 720-413-2
Client Matrix: Water

Date Sampled: 11/11/2005 0720
Date Received: 11/11/2005 1340

8260B Volatile Organic Compounds by GC/MS

Method: 8260B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/30/2005 2137
Date Prepared: 11/30/2005 2137

Analysis Batch: 720-2456

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\113005\720-
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Result (ug/L)	Qualifier	RL
Benzene	ND		0.50
Ethylbenzene	ND		0.50
Toluene	ND		0.50
MTBE	ND		0.50
Xylenes, Total	ND		1.0
Gasoline Range Organics (GRO)-C5-C12	ND		50

Surrogate	%Rec	Acceptance Limits
Toluene-d8	91	77 - 121
1,2-Dichloroethane-d4	85	73 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-45B**

Lab Sample ID:	720-413-3			Date Sampled:	11/11/2005 0806
Client Matrix:	Solid	% Moisture:	14.8	Date Received:	11/11/2005 1340

8260B Volatile Organic Compounds by GC/MS

Method:	8260B	Analysis Batch: 720-2427		Instrument ID:	Saturn 2100
Preparation:	N/A			Lab File ID:	d:\data\200511\112905\720-
Dilution:	1.0			Initial Weight/Volume:	10 mL
Date Analyzed:	11/30/2005 0512			Final Weight/Volume:	10 mL
Date Prepared:	N/A				

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Benzene		ND		2.9
Ethylbenzene		ND		2.9
Toluene		ND		2.9
Xylenes, Total		ND		5.9
Gasoline Range Organics (GRO)-C5-C12		ND		590

Surrogate	%Rec	Acceptance Limits
Toluene-d8	101	70 - 130
1,2-Dichloroethane-d4	101	60 - 140

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-45D**

Lab Sample ID:	720-413-4			Date Sampled:	11/11/2005 0806
Client Matrix:	Solid	% Moisture:	14.1	Date Received:	11/11/2005 1340

8260B Volatile Organic Compounds by GC/MS

Method:	8260B	Analysis Batch: 720-2416		Instrument ID:	Varian 3900E
Preparation:	5035	Prep Batch: 720-2423		Lab File ID:	c:\varianws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	6.13 g
Date Analyzed:	11/25/2005 2347			Final Weight/Volume:	6.13 g
Date Prepared:	11/11/2005 1145				

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Benzene		ND		2.9
Ethylbenzene		ND		2.9
Toluene		ND		2.9
Xylenes, Total		ND		5.8
Gasoline Range Organics (GRO)-C5-C12		ND		580

Surrogate	%Rec		Acceptance Limits
Toluene-d8	92		70 - 130
1,2-Dichloroethane-d4	162	*	60 - 140

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-48B**

Lab Sample ID:	720-413-5			Date Sampled:	11/11/2005 0806
Client Matrix:	Solid	% Moisture:	14.5	Date Received:	11/11/2005 1340

8260B Volatile Organic Compounds by GC/MS

Method:	8260B	Analysis Batch: 720-2362		Instrument ID:	Saturn 2100
Preparation:	N/A			Lab File ID:	d:\data\200511\112905\720-
Dilution:	1.0			Initial Weight/Volume:	10 mL
Date Analyzed:	11/30/2005 0538			Final Weight/Volume:	10 mL
Date Prepared:	N/A				

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Benzene		ND		2.9
Ethylbenzene		ND		2.9
Toluene		ND		2.9
Xylenes, Total		ND		5.9
Gasoline Range Organics (GRO)-C5-C12		ND		590

Surrogate	%Rec	Acceptance Limits
Toluene-d8	97	70 - 130
1,2-Dichloroethane-d4	107	60 - 140

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-49C**

Lab Sample ID:	720-413-6			Date Sampled:	11/11/2005 0825
Client Matrix:	Solid	% Moisture:	13.8	Date Received:	11/11/2005 1340

8260B Volatile Organic Compounds by GC/MS

Method:	8260B	Analysis Batch: 720-2417		Instrument ID:	Saturn 2100
Preparation:	N/A			Lab File ID:	d:\data\200511\112505\sa-s
Dilution:	1.0			Initial Weight/Volume:	6.16 mL
Date Analyzed:	11/25/2005 2322			Final Weight/Volume:	10 mL
Date Prepared:	N/A				

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Benzene		ND		4.7
Ethylbenzene		ND		4.7
Toluene		ND		4.7
Xylenes, Total		ND		9.4
Gasoline Range Organics (GRO)-C5-C12		ND		940

Surrogate	%Rec	Acceptance Limits
Toluene-d8	89	70 - 130
1,2-Dichloroethane-d4	125	60 - 140

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-58C**

Lab Sample ID:	720-413-8			Date Sampled:	11/11/2005 0835
Client Matrix:	Solid	% Moisture:	11.5	Date Received:	11/11/2005 1340

8260B Volatile Organic Compounds by GC/MS

Method:	8260B	Analysis Batch: 720-2417		Instrument ID:	Saturn 2100
Preparation:	N/A			Lab File ID:	d:\data\200511\112505\sa-s
Dilution:	1.0			Initial Weight/Volume:	6.14 mL
Date Analyzed:	11/25/2005 2348			Final Weight/Volume:	10 mL
Date Prepared:	N/A				

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Benzene		ND		4.6
Ethylbenzene		ND		4.6
Toluene		ND		4.6
Xylenes, Total		ND		9.2
Gasoline Range Organics (GRO)-C5-C12		ND		920

Surrogate	%Rec	Acceptance Limits
Toluene-d8	96	70 - 130
1,2-Dichloroethane-d4	125	60 - 140

Analytical Data

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-4**

Lab Sample ID: 720-413-9
Client Matrix: Water

Date Sampled: 11/11/2005 1030
Date Received: 11/11/2005 1340

8260B Volatile Organic Compounds by GC/MS (Low Level)

Method: 8260B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/23/2005 2221
Date Prepared: 11/23/2005 2221

Analysis Batch: 720-2239

Instrument ID: Saturn 2K3
Lab File ID: d:\data\200511\112305\720-
Initial Weight/Volume: 40 mL
Final Weight/Volume: 40 mL

Analyte	Result (ug/L)	Qualifier	RL
1,1-Dichloroethene	ND		0.50
1,1-Dichloroethane	ND		0.50
Dichlorodifluoromethane	ND		0.50
Vinyl chloride	ND		0.50
Chloroethane	ND		1.0
Trichlorofluoromethane	ND		1.0
Methylene Chloride	ND		5.0
trans-1,2-Dichloroethene	ND		0.50
cis-1,2-Dichloroethene	ND		0.50
Chloroform	ND		1.0
1,1,1-Trichloroethane	ND		0.50
Carbon tetrachloride	ND		0.50
1,2-Dichloroethane	ND		0.50
Trichloroethene	ND		0.50
1,2-Dichloropropane	ND		0.50
Dichlorobromomethane	ND		0.50
trans-1,3-Dichloropropene	ND		0.50
cis-1,3-Dichloropropene	ND		0.50
1,1,2-Trichloroethane	ND		0.50
Tetrachloroethene	ND		0.50
Chlorodibromomethane	ND		0.50
Chlorobenzene	ND		0.50
Bromoform	ND		1.0
1,1,2,2-Tetrachloroethane	ND		0.50
1,3-Dichlorobenzene	ND		0.50
1,4-Dichlorobenzene	ND		0.50
1,2-Dichlorobenzene	ND		0.50
Chloromethane	ND		1.0
Bromomethane	ND		1.0
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		0.50
EDB	ND		0.50
1,2,4-Trichlorobenzene	ND		1.0

Analytical Data

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-5**

Lab Sample ID: 720-413-10
Client Matrix: Water

Date Sampled: 11/11/2005 1100
Date Received: 11/11/2005 1340

8260B Volatile Organic Compounds by GC/MS (Low Level)

Method: 8260B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/23/2005 2255
Date Prepared: 11/23/2005 2255

Analysis Batch: 720-2239

Instrument ID: Saturn 2K3
Lab File ID: d:\data\200511\112305\720-
Initial Weight/Volume: 40 mL
Final Weight/Volume: 40 mL

Analyte	Result (ug/L)	Qualifier	RL
1,1-Dichloroethene	ND		0.50
1,1-Dichloroethane	ND		0.50
Dichlorodifluoromethane	ND		0.50
Vinyl chloride	ND		0.50
Chloroethane	ND		1.0
Trichlorofluoromethane	ND		1.0
Methylene Chloride	ND		5.0
trans-1,2-Dichloroethene	ND		0.50
cis-1,2-Dichloroethene	ND		0.50
Chloroform	ND		1.0
1,1,1-Trichloroethane	ND		0.50
Carbon tetrachloride	ND		0.50
1,2-Dichloroethane	ND		0.50
Trichloroethene	ND		0.50
1,2-Dichloropropane	ND		0.50
Dichlorobromomethane	ND		0.50
trans-1,3-Dichloropropene	ND		0.50
cis-1,3-Dichloropropene	ND		0.50
1,1,2-Trichloroethane	ND		0.50
Tetrachloroethene	ND		0.50
Chlorodibromomethane	ND		0.50
Chlorobenzene	ND		0.50
Bromoform	ND		1.0
1,1,2,2-Tetrachloroethane	ND		0.50
1,3-Dichlorobenzene	ND		0.50
1,4-Dichlorobenzene	ND		0.50
1,2-Dichlorobenzene	ND		0.50
Chloromethane	ND		1.0
Bromomethane	ND		1.0
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		0.50
EDB	ND		0.50
1,2,4-Trichlorobenzene	ND		1.0

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-EB-2**

Lab Sample ID: 720-413-1
Client Matrix: Water

Date Sampled: 11/11/2005 0715
Date Received: 11/11/2005 1340

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/23/2005 2350
Date Prepared: 11/15/2005 1510

Analysis Batch: 720-2227
Prep Batch: 720-1797

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 850 mL
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result (ug/L)	Qualifier	RL
Naphthalene	ND		0.12
Acenaphthene	ND		0.12
Acenaphthylene	ND		0.12
Fluorene	ND		0.12
Phenanthrene	ND		0.12
Anthracene	ND		0.12
Benzo[a]anthracene	ND		0.12
Chrysene	ND		0.12
Benzo[a]pyrene	ND		0.12
Benzo[b]fluoranthene	ND		0.12
Benzo[k]fluoranthene	ND		0.12
Benzo[g,h,i]perylene	ND		0.12
Indeno[1,2,3-cd]pyrene	ND		0.12
Fluoranthene	ND		0.12
Pyrene	ND		0.12
Dibenz(a,h)anthracene	ND		1.2

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	78	43 - 116
Terphenyl-d14	98	33 - 141

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-EB-4**

Lab Sample ID: 720-413-2
Client Matrix: Water

Date Sampled: 11/11/2005 0720
Date Received: 11/11/2005 1340

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/24/2005 0018
Date Prepared: 11/15/2005 1510

Analysis Batch: 720-2227
Prep Batch: 720-1797

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 900 mL
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result (ug/L)	Qualifier	RL
Naphthalene	ND		0.11
Acenaphthene	ND		0.11
Acenaphthylene	ND		0.11
Fluorene	ND		0.11
Phenanthrene	ND		0.11
Anthracene	ND		0.11
Benzo[a]anthracene	ND		0.11
Chrysene	ND		0.11
Benzo[a]pyrene	ND		0.11
Benzo[b]fluoranthene	ND		0.11
Benzo[k]fluoranthene	ND		0.11
Benzo[g,h,i]perylene	ND		0.11
Indeno[1,2,3-cd]pyrene	ND		0.11
Fluoranthene	ND		0.11
Pyrene	ND		0.11
Dibenz(a,h)anthracene	ND		1.1

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	70	43 - 116
Terphenyl-d14	86	33 - 141

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-49C**

Lab Sample ID:	720-413-6			Date Sampled:	11/11/2005 0825
Client Matrix:	Solid	% Moisture:	13.8	Date Received:	11/11/2005 1340

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2258		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1920		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.21 g
Date Analyzed:	11/27/2005 1407			Final Weight/Volume:	1 mL
Date Prepared:	11/17/2005 1413			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.8
Acenaphthene		ND		5.8
Acenaphthylene		ND		5.8
Fluorene		ND		5.8
Phenanthrene		ND		5.8
Anthracene		ND		5.8
Benzo[a]anthracene		ND		5.8
Chrysene		ND		5.8
Benzo[a]pyrene		ND		5.8
Benzo[b]fluoranthene		ND		5.8
Benzo[k]fluoranthene		ND		5.8
Benzo[g,h,i]perylene		ND		5.8
Indeno[1,2,3-cd]pyrene		ND		5.8
Fluoranthene		ND		5.8
Pyrene		ND		5.8
Dibenz(a,h)anthracene		ND		5.8

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	73	30 - 115
Terphenyl-d14	88	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-50C**

Lab Sample ID:	720-413-7			Date Sampled:	11/11/2005 0830
Client Matrix:	Solid	% Moisture:	14.0	Date Received:	11/11/2005 1340

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2258		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1920		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	1.0			Initial Weight/Volume:	30.04 g
Date Analyzed:	11/27/2005 1435			Final Weight/Volume:	1 mL
Date Prepared:	11/17/2005 1413			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		5.8
Acenaphthene		ND		5.8
Acenaphthylene		ND		5.8
Fluorene		ND		5.8
Phenanthrene		ND		5.8
Anthracene		ND		5.8
Benzo[a]anthracene		ND		5.8
Chrysene		ND		5.8
Benzo[a]pyrene		ND		5.8
Benzo[b]fluoranthene		ND		5.8
Benzo[k]fluoranthene		ND		5.8
Benzo[g,h,i]perylene		ND		5.8
Indeno[1,2,3-cd]pyrene		ND		5.8
Fluoranthene		ND		5.8
Pyrene		ND		5.8
Dibenz(a,h)anthracene		ND		5.8

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	62	30 - 115
Terphenyl-d14	72	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-4**

Lab Sample ID: 720-413-9
Client Matrix: Water

Date Sampled: 11/11/2005 1030
Date Received: 11/11/2005 1340

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/24/2005 0046
Date Prepared: 11/15/2005 1510

Analysis Batch: 720-2227
Prep Batch: 720-1797

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 880 mL
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result (ug/L)	Qualifier	RL
Naphthalene	ND		0.11
Acenaphthene	ND		0.11
Acenaphthylene	ND		0.11
Fluorene	ND		0.11
Phenanthrene	0.18		0.11
Anthracene	ND		0.11
Benzo[a]anthracene	ND		0.11
Chrysene	ND		0.11
Benzo[a]pyrene	ND		0.11
Benzo[b]fluoranthene	ND		0.11
Benzo[k]fluoranthene	ND		0.11
Benzo[g,h,i]perylene	ND		0.11
Indeno[1,2,3-cd]pyrene	ND		0.11
Fluoranthene	ND		0.11
Pyrene	ND		0.11
Dibenz(a,h)anthracene	ND		1.1

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	63	43 - 116
Terphenyl-d14	80	33 - 141

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-5**

Lab Sample ID: 720-413-10
Client Matrix: Water

Date Sampled: 11/11/2005 1100
Date Received: 11/11/2005 1340

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/24/2005 0114
Date Prepared: 11/15/2005 1510

Analysis Batch: 720-2227
Prep Batch: 720-1797

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 760 mL
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result (ug/L)	Qualifier	RL
Naphthalene	ND		0.13
Acenaphthene	ND		0.13
Acenaphthylene	ND		0.13
Fluorene	ND		0.13
Phenanthrene	0.18		0.13
Anthracene	ND		0.13
Benzo[a]anthracene	ND		0.13
Chrysene	ND		0.13
Benzo[a]pyrene	ND		0.13
Benzo[b]fluoranthene	ND		0.13
Benzo[k]fluoranthene	ND		0.13
Benzo[g,h,i]perylene	ND		0.13
Indeno[1,2,3-cd]pyrene	ND		0.13
Fluoranthene	ND		0.13
Pyrene	ND		0.13
Dibenz(a,h)anthracene	ND		1.3

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	72	43 - 116
Terphenyl-d14	84	33 - 141

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-EB3**

Lab Sample ID: 720-413-12
Client Matrix: Water

Date Sampled: 11/11/2005 1205
Date Received: 11/11/2005 1340

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/24/2005 0142
Date Prepared: 11/15/2005 1510

Analysis Batch: 720-2227
Prep Batch: 720-1797

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 880 mL
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result (ug/L)	Qualifier	RL
Naphthalene	ND		0.11
Acenaphthene	ND		0.11
Acenaphthylene	ND		0.11
Fluorene	ND		0.11
Phenanthrene	ND		0.11
Anthracene	ND		0.11
Benzo[a]anthracene	ND		0.11
Chrysene	ND		0.11
Benzo[a]pyrene	ND		0.11
Benzo[b]fluoranthene	ND		0.11
Benzo[k]fluoranthene	ND		0.11
Benzo[g,h,i]perylene	ND		0.11
Indeno[1,2,3-cd]pyrene	ND		0.11
Fluoranthene	ND		0.11
Pyrene	ND		0.11
Dibenz(a,h)anthracene	ND		1.1

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	77	43 - 116
Terphenyl-d14	92	33 - 141

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-50A**

Lab Sample ID:	720-413-13			Date Sampled:	11/11/2005 1045
Client Matrix:	Solid	% Moisture:	10.5	Date Received:	11/11/2005 1340

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method:	8270C	Analysis Batch: 720-2258		Instrument ID:	Sat 2K2
Preparation:	3550B	Prep Batch: 720-1920		Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.27 g
Date Analyzed:	11/27/2005 1503			Final Weight/Volume:	5 mL
Date Prepared:	11/17/2005 1413			Injection Volume:	

Analyte	DryWt Corrected: Y	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		280
Acenaphthene		ND		280
Acenaphthylene		ND		280
Fluorene		ND		280
Phenanthrene		ND		280
Anthracene		ND		280
Benzo[a]anthracene		ND		280
Chrysene		320		280
Benzo[a]pyrene		ND		280
Benzo[b]fluoranthene		380		280
Benzo[k]fluoranthene		ND		280
Benzo[g,h,i]perylene		ND		280
Indeno[1,2,3-cd]pyrene		ND		280
Fluoranthene		ND		280
Pyrene		ND		280
Dibenz(a,h)anthracene		ND		280

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	82	30 - 115
Terphenyl-d14	104	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-49A**

Lab Sample ID: 720-413-14
Client Matrix: Solid

Date Sampled: 11/11/2005 1145
Date Received: 11/11/2005 1340

8270C Semivolatile Organic Compounds by GC/MS (Selective Ion Monitoring)

Method: 8270C
Preparation: 3550B
Dilution: 10
Date Analyzed: 11/27/2005 1530
Date Prepared: 11/17/2005 1413

Analysis Batch: 720-2258
Prep Batch: 720-1920

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.20 g
Final Weight/Volume: 5 mL
Injection Volume:

Analyte	DryWt Corrected: N	Result (ug/Kg)	Qualifier	RL
Naphthalene		ND		250
Acenaphthene		ND		250
Acenaphthylene		ND		250
Fluorene		ND		250
Phenanthrene		ND		250
Anthracene		ND		250
Benzo[a]anthracene		ND		250
Chrysene		ND		250
Benzo[a]pyrene		ND		250
Benzo[b]fluoranthene		ND		250
Benzo[k]fluoranthene		ND		250
Benzo[g,h,i]perylene		ND		250
Indeno[1,2,3-cd]pyrene		ND		250
Fluoranthene		ND		250
Pyrene		ND		250
Dibenz(a,h)anthracene		ND		250

Surrogate	%Rec	Acceptance Limits
2-Fluorobiphenyl	60	30 - 115
Terphenyl-d14	76	18 - 137

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method:	8015B	Analysis Batch: 720-1869	Instrument ID:	PID/FID Gas/Btex
Preparation:	5030B		Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	10 mL
Date Analyzed:	11/15/2005 1041		Final Weight/Volume:	10 mL
Date Prepared:	11/15/2005 1041		Injection Volume:	
			Column ID:	PRIMARY

Surrogate	%Rec	Acceptance Limits
4-Bromofluorobenzene		

Method:	8015B	Analysis Batch: 720-1997	Instrument ID:	PID/FID Gas/Btex
Preparation:	5030B		Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	10 mL
Date Analyzed:	11/15/2005 1052		Final Weight/Volume:	10 mL
Date Prepared:	11/15/2005 1052		Injection Volume:	
			Column ID:	PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Gasoline Range Organics (GRO)-C5-C12	ND		50

Method:	8015B	Analysis Batch: 720-1997	Instrument ID:	PID/FID Gas/Btex
Preparation:	5030B		Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	10 mL
Date Analyzed:	11/15/2005 1052		Final Weight/Volume:	10 mL
Date Prepared:	11/15/2005 1052		Injection Volume:	
			Column ID:	SECONDARY

Surrogate	%Rec	Acceptance Limits
4-Bromofluorobenzene	90	50 - 150

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-5**

Lab Sample ID: 720-413-10
Client Matrix: Water

Date Sampled: 11/11/2005 1100
Date Received: 11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Gasoline Range Organics)

Method: 8015B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/15/2005 1126
Date Prepared: 11/15/2005 1126

Analysis Batch: 720-1997

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Gasoline Range Organics (GRO)-C5-C12	ND		50

Surrogate	%Rec	Acceptance Limits
4-Bromofluorobenzene	92	50 - 150

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-3**

Lab Sample ID: 720-413-11
Client Matrix: Water

Date Sampled: 11/11/2005 1130
Date Received: 11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Gasoline Range Organics)

Method: 8015B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/16/2005 0000
Date Prepared: 11/16/2005 0000

Analysis Batch: 720-1997

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Gasoline Range Organics (GRO)-C5-C12	ND		50

Surrogate	%Rec	Acceptance Limits
4-Bromofluorobenzene	90	50 - 150

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-EB3**

Lab Sample ID: 720-413-12
Client Matrix: Water

Date Sampled: 11/11/2005 1205
Date Received: 11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Gasoline Range Organics)

Method: 8015B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/16/2005 0034
Date Prepared: 11/16/2005 0034

Analysis Batch: 720-1997

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Gasoline Range Organics (GRO)-C5-C12	ND		50

Surrogate	%Rec	Acceptance Limits
4-Bromofluorobenzene	91	50 - 150

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-4**

Lab Sample ID: 720-413-9
Client Matrix: Water

Date Sampled: 11/11/2005 1030
Date Received: 11/11/2005 1340

8021B Aromatic and Halogenated VOCs by Gas Chromatography using PID or ECD

Method: 8021B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/15/2005 1052
Date Prepared: 11/15/2005 1052

Analysis Batch: 720-1991

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Benzene	ND		0.50
Toluene	ND		0.50
Ethylbenzene	ND		0.50
Xylenes, Total	ND		0.50

Surrogate	%Rec	Acceptance Limits
a,a,a-Trifluorotoluene (pid)	100	58 - 124
4-Bromofluorobenzene	95	50 - 150

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-5**

Lab Sample ID: 720-413-10
Client Matrix: Water

Date Sampled: 11/11/2005 1100
Date Received: 11/11/2005 1340

8021B Aromatic and Halogenated VOCs by Gas Chromatography using PID or ECD

Method: 8021B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/15/2005 1041
Date Prepared: 11/15/2005 1041

Analysis Batch: 720-1869

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: SECONDARY

Analyte	Result (ug/L)	Qualifier	RL
Benzene	ND		0.50
Toluene	ND		0.50
Ethylbenzene	ND		0.50
Xylenes, Total	ND		0.50

Surrogate	%Rec	Acceptance Limits
a,a,a-Trifluorotoluene (pid)		
4-Bromofluorobenzene		

Method: 8021B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/15/2005 1052
Date Prepared: 11/15/2005 1052

Analysis Batch: 720-1991

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Benzene	ND		0.50
Toluene	ND		0.50
Ethylbenzene	ND		0.50
Xylenes, Total	ND		0.50

Surrogate	%Rec	Acceptance Limits
a,a,a-Trifluorotoluene (pid)	102	58 - 124
4-Bromofluorobenzene	99	50 - 150

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-3**

Lab Sample ID: 720-413-11
Client Matrix: Water

Date Sampled: 11/11/2005 1130
Date Received: 11/11/2005 1340

8021B Aromatic and Halogenated VOCs by Gas Chromatography using PID or ECD

Method: 8021B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/15/2005 1041
Date Prepared: 11/15/2005 1041

Analysis Batch: 720-1869

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: SECONDARY

Analyte	Result (ug/L)	Qualifier	RL
Benzene	ND		0.50
Toluene	ND		0.50
Ethylbenzene	ND		0.50
Xylenes, Total	ND		0.50

Surrogate	%Rec	Acceptance Limits
a,a,a-Trifluorotoluene (pid)		
4-Bromofluorobenzene		

Method: 8021B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/15/2005 1052
Date Prepared: 11/15/2005 1052

Analysis Batch: 720-1991

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Benzene	ND		0.50
Toluene	ND		0.50
Ethylbenzene	ND		0.50
Xylenes, Total	ND		0.50

Surrogate	%Rec	Acceptance Limits
a,a,a-Trifluorotoluene (pid)	102	58 - 124
4-Bromofluorobenzene	95	50 - 150

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-EB3**

Lab Sample ID: 720-413-12
Client Matrix: Water

Date Sampled: 11/11/2005 1205
Date Received: 11/11/2005 1340

8021B Aromatic and Halogenated VOCs by Gas Chromatography using PID or ECD

Method: 8021B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/15/2005 1041
Date Prepared: 11/15/2005 1041

Analysis Batch: 720-1869

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: SECONDARY

Analyte	Result (ug/L)	Qualifier	RL
Benzene	ND		0.50
Toluene	ND		0.50
Ethylbenzene	ND		0.50
Xylenes, Total	ND		0.50

Surrogate	%Rec	Acceptance Limits
a,a,a-Trifluorotoluene (pid)		
4-Bromofluorobenzene		

Method: 8021B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 11/15/2005 1052
Date Prepared: 11/15/2005 1052

Analysis Batch: 720-1991

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Benzene	ND		0.50
Toluene	ND		0.50
Ethylbenzene	ND		0.50
Xylenes, Total	ND		0.50

Surrogate	%Rec	Acceptance Limits
a,a,a-Trifluorotoluene (pid)	101	58 - 124
4-Bromofluorobenzene	95	50 - 150

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-EB-2**

Lab Sample ID: 720-413-1
Client Matrix: Water

Date Sampled: 11/11/2005 0715
Date Received: 11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method: 8015B
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/16/2005 1830
Date Prepared: 11/15/2005 0848

Analysis Batch: 720-1935
Prep Batch: 720-1753

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Diesel Range Organics [C10-C28]	ND		50
Motor Oil Range Organics [C24-C36]	ND		500

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	78	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-EB-4**

Lab Sample ID: 720-413-2
Client Matrix: Water

Date Sampled: 11/11/2005 0720
Date Received: 11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method: 8015B
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/16/2005 1952
Date Prepared: 11/15/2005 0848

Analysis Batch: 720-1935
Prep Batch: 720-1753

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Diesel Range Organics [C10-C28]	ND		50
Motor Oil Range Organics [C24-C36]	ND		500

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	72	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-45B**

Lab Sample ID:	720-413-3			Date Sampled:	11/11/2005 0806
Client Matrix:	Solid	% Moisture:	14.8	Date Received:	11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-2006		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1919		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.32 g
Date Analyzed:	11/18/2005 0117			Final Weight/Volume:	5 mL
Date Prepared:	11/17/2005 1403			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		4.3		1.2
Motor Oil Range Organics [C24-C36]		ND		58

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	73	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-45D**

Lab Sample ID:	720-413-4			Date Sampled:	11/11/2005 0806
Client Matrix:	Solid	% Moisture:	14.1	Date Received:	11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-2006		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1919		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.36 g
Date Analyzed:	11/18/2005 0143			Final Weight/Volume:	5 mL
Date Prepared:	11/17/2005 1403			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		ND		1.2
Motor Oil Range Organics [C24-C36]		ND		58

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	66	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-48B**

Lab Sample ID:	720-413-5			Date Sampled:	11/11/2005 0806
Client Matrix:	Solid	% Moisture:	14.5	Date Received:	11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-2006		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1919		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.12 g
Date Analyzed:	11/18/2005 0210			Final Weight/Volume:	5 mL
Date Prepared:	11/17/2005 1403			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		1.7		1.2
Motor Oil Range Organics [C24-C36]		ND		58

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	65	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-49C**

Lab Sample ID:	720-413-6			Date Sampled:	11/11/2005 0825
Client Matrix:	Solid	% Moisture:	13.8	Date Received:	11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-2006		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1919		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.02 g
Date Analyzed:	11/18/2005 0236			Final Weight/Volume:	5 mL
Date Prepared:	11/17/2005 1403			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		ND		1.2
Motor Oil Range Organics [C24-C36]		ND		58

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	68	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-50C**

Lab Sample ID:	720-413-7			Date Sampled:	11/11/2005 0830
Client Matrix:	Solid	% Moisture:	14.0	Date Received:	11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-2006		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1919		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.41 g
Date Analyzed:	11/18/2005 0541			Final Weight/Volume:	5 mL
Date Prepared:	11/17/2005 1403			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		ND		1.1
Motor Oil Range Organics [C24-C36]		ND		57

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	75	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-58C**

Lab Sample ID:	720-413-8			Date Sampled:	11/11/2005 0835
Client Matrix:	Solid	% Moisture:	11.5	Date Received:	11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-2006		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1919		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	30.48 g
Date Analyzed:	11/18/2005 0515			Final Weight/Volume:	5 mL
Date Prepared:	11/17/2005 1403			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		1.8		1.1
Motor Oil Range Organics [C24-C36]		ND		56

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	71	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-4**

Lab Sample ID: 720-413-9
Client Matrix: Water

Date Sampled: 11/11/2005 1030
Date Received: 11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method: 8015B
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/16/2005 2019
Date Prepared: 11/15/2005 0848

Analysis Batch: 720-1935
Prep Batch: 720-1753

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Diesel Range Organics [C10-C28]	ND		50
Motor Oil Range Organics [C24-C36]	ND		500

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	75	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-5**

Lab Sample ID: 720-413-10
Client Matrix: Water

Date Sampled: 11/11/2005 1100
Date Received: 11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method: 8015B
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/16/2005 2047
Date Prepared: 11/15/2005 0848

Analysis Batch: 720-1935
Prep Batch: 720-1753

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Diesel Range Organics [C10-C28]	ND		50
Motor Oil Range Organics [C24-C36]	ND		500

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	78	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-3**

Lab Sample ID: 720-413-11
Client Matrix: Water

Date Sampled: 11/11/2005 1130
Date Received: 11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method: 8015B
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/16/2005 2236
Date Prepared: 11/15/2005 0848

Analysis Batch: 720-1935
Prep Batch: 720-1753

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Diesel Range Organics [C10-C28]	ND		50
Motor Oil Range Organics [C24-C36]	ND		500

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	77	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-EB3**

Lab Sample ID: 720-413-12
Client Matrix: Water

Date Sampled: 11/11/2005 1205
Date Received: 11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method: 8015B
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/16/2005 2303
Date Prepared: 11/15/2005 0848

Analysis Batch: 720-1935
Prep Batch: 720-1753

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Diesel Range Organics [C10-C28]	ND		50
Motor Oil Range Organics [C24-C36]	ND		500

Surrogate	%Rec	Acceptance Limits
o-Terphenyl	76	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-50A**

Lab Sample ID:	720-413-13			Date Sampled:	11/11/2005 1045
Client Matrix:	Solid	% Moisture:	10.5	Date Received:	11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-2006		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1919		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.35 g
Date Analyzed:	11/17/2005 2238			Final Weight/Volume:	5 mL
Date Prepared:	11/17/2005 1403			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		100		5.5
Motor Oil Range Organics [C24-C36]		800		280

Surrogate	%Rec		Acceptance Limits
o-Terphenyl	0	*	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-49A**

Lab Sample ID:	720-413-14			Date Sampled:	11/11/2005 1145
Client Matrix:	Solid	% Moisture:	5.7	Date Received:	11/11/2005 1340

8015B Nonhalogenated Organics using GC/FID -Modified (Diesel Range Organics)

Method:	8015B	Analysis Batch: 720-2006		Instrument ID:	HP DRO3
Preparation:	3550B	Prep Batch: 720-1919		Lab File ID:	N/A
Dilution:	5.0			Initial Weight/Volume:	30.21 g
Date Analyzed:	11/17/2005 2305			Final Weight/Volume:	5 mL
Date Prepared:	11/17/2005 1403			Injection Volume:	
				Column ID:	PRIMARY

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Diesel Range Organics [C10-C28]		71		5.3
Motor Oil Range Organics [C24-C36]		700		260

Surrogate	%Rec		Acceptance Limits
o-Terphenyl	0	*	60 - 130

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-EB-2**

Lab Sample ID: 720-413-1
Client Matrix: Water

Date Sampled: 11/11/2005 0715
Date Received: 11/11/2005 1340

8081A Organochlorine Pesticides by Gas Chromatography

Method: 8081A
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/17/2005 0740
Date Prepared: 11/15/2005 0827

Analysis Batch: 720-2045
Prep Batch: 720-1752

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 800 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Aldrin	ND		0.075
Dieldrin	ND		0.075
Endrin aldehyde	ND		0.075
Endrin	ND		0.075
Endrin ketone	ND		0.075
Heptachlor	ND		0.075
Heptachlor epoxide	ND		0.075
4,4'-DDT	ND		0.075
4,4'-DDE	ND		0.075
4,4'-DDD	ND		0.075
Endosulfan I	ND		0.075
Endosulfan II	ND		0.075
alpha-BHC	ND		0.075
beta-BHC	ND		0.075
gamma-BHC (Lindane)	ND		0.075
delta-BHC	ND		0.075
Endosulfan sulfate	ND		0.075
Methoxychlor	ND		0.075
Toxaphene	ND		1.3
Chlordane (technical)	ND		1.3
alpha-Chlordane	ND		0.075
gamma-Chlordane	ND		0.075

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	108	62 - 123
DCB Decachlorobiphenyl	86	56 - 136

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-EB-4**

Lab Sample ID: 720-413-2
Client Matrix: Water

Date Sampled: 11/11/2005 0720
Date Received: 11/11/2005 1340

8081A Organochlorine Pesticides by Gas Chromatography

Method: 8081A
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/17/2005 0809
Date Prepared: 11/15/2005 0827

Analysis Batch: 720-2045
Prep Batch: 720-1752

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 850 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Aldrin	ND		0.071
Dieldrin	ND		0.071
Endrin aldehyde	ND		0.071
Endrin	ND		0.071
Endrin ketone	ND		0.071
Heptachlor	ND		0.071
Heptachlor epoxide	ND		0.071
4,4'-DDT	ND		0.071
4,4'-DDE	ND		0.071
4,4'-DDD	ND		0.071
Endosulfan I	ND		0.071
Endosulfan II	ND		0.071
alpha-BHC	ND		0.071
beta-BHC	ND		0.071
gamma-BHC (Lindane)	ND		0.071
delta-BHC	ND		0.071
Endosulfan sulfate	ND		0.071
Methoxychlor	ND		0.071
Toxaphene	ND		1.2
Chlordane (technical)	ND		1.2
alpha-Chlordane	ND		0.071
gamma-Chlordane	ND		0.071

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	105	62 - 123
DCB Decachlorobiphenyl	61	56 - 136

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-4**

Lab Sample ID: 720-413-9
Client Matrix: Water

Date Sampled: 11/11/2005 1030
Date Received: 11/11/2005 1340

8081A Organochlorine Pesticides by Gas Chromatography

Method: 8081A
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/17/2005 0838
Date Prepared: 11/15/2005 0827

Analysis Batch: 720-2045
Prep Batch: 720-1752

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 870 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Aldrin	ND		0.069
Dieldrin	ND		0.069
Endrin aldehyde	ND		0.069
Endrin	ND		0.069
Endrin ketone	ND		0.069
Heptachlor	ND		0.069
Heptachlor epoxide	ND		0.069
4,4'-DDT	ND		0.069
4,4'-DDE	ND		0.069
4,4'-DDD	ND		0.069
Endosulfan I	ND		0.069
Endosulfan II	ND		0.069
alpha-BHC	ND		0.069
beta-BHC	ND		0.069
gamma-BHC (Lindane)	ND		0.069
delta-BHC	ND		0.069
Endosulfan sulfate	ND		0.069
Methoxychlor	ND		0.069
Toxaphene	ND		1.1
Chlordane (technical)	ND		1.1
alpha-Chlordane	ND		0.069
gamma-Chlordane	ND		0.069

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	107	62 - 123
DCB Decachlorobiphenyl	96	56 - 136

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: **EBHH-GW-5**

Lab Sample ID: 720-413-10
Client Matrix: Water

Date Sampled: 11/11/2005 1100
Date Received: 11/11/2005 1340

8081A Organochlorine Pesticides by Gas Chromatography

Method: 8081A
Preparation: 3510C
Dilution: 1.0
Date Analyzed: 11/21/2005 1011
Date Prepared: 11/18/2005 1443

Analysis Batch: 720-2116
Prep Batch: 720-1983

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 620 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result (ug/L)	Qualifier	RL
Aldrin	ND		0.097
Dieldrin	ND		0.097
Endrin aldehyde	ND		0.097
Endrin	ND		0.097
Endrin ketone	ND		0.097
Heptachlor	ND		0.097
Heptachlor epoxide	ND		0.097
4,4'-DDT	ND		0.097
4,4'-DDE	ND		0.097
4,4'-DDD	ND		0.097
Endosulfan I	ND		0.097
Endosulfan II	ND		0.097
alpha-BHC	ND		0.097
beta-BHC	ND		0.097
gamma-BHC (Lindane)	ND		0.097
delta-BHC	ND		0.097
Endosulfan sulfate	ND		0.097
Methoxychlor	ND		0.097
Toxaphene	ND		1.6
Chlordane (technical)	ND		1.6
alpha-Chlordane	ND		0.097
gamma-Chlordane	ND		0.097

Surrogate	%Rec	Acceptance Limits
Tetrachloro-m-xylene	97	62 - 123
DCB Decachlorobiphenyl	73	56 - 136

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: EBHH-EB-2

Lab Sample ID:	720-413-1	Date Sampled:	11/11/2005 0715
Client Matrix:	Water	Date Received:	11/11/2005 1340

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3010A	Prep Batch: 720-1816	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	50 mL
Date Analyzed:	11/17/2005 0909		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 0856			

Analyte	Result (mg/L)	Qualifier	RL
Silver	ND		0.0050
Arsenic	ND		0.0050
Barium	ND		0.0050
Cadmium	ND		0.0020
Chromium	ND		0.0050
Lead	ND		0.0050
Selenium	ND		0.0050

7470A Mercury in Liquid Waste (Manual Cold Vapor Technique)

Method:	7470A	Analysis Batch: 720-1793	Instrument ID:	FIMS 100
Preparation:	7470A	Prep Batch: 720-1754	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	25 mL
Date Analyzed:	11/15/2005 1326		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 0859			

Analyte	Result (mg/L)	Qualifier	RL
Mercury	ND		0.00020

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: EBHH-EB-4

Lab Sample ID:	720-413-2	Date Sampled:	11/11/2005 0720
Client Matrix:	Water	Date Received:	11/11/2005 1340

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3010A	Prep Batch: 720-1816	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	50 mL
Date Analyzed:	11/17/2005 0912		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 0856			

Analyte	Result (mg/L)	Qualifier	RL
Silver	ND		0.0050
Arsenic	ND		0.0050
Barium	ND		0.0050
Cadmium	ND		0.0020
Chromium	ND		0.0050
Lead	ND		0.0050
Selenium	ND		0.0050

7470A Mercury in Liquid Waste (Manual Cold Vapor Technique)

Method:	7470A	Analysis Batch: 720-1793	Instrument ID:	FIMS 100
Preparation:	7470A	Prep Batch: 720-1754	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	25 mL
Date Analyzed:	11/15/2005 1328		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 0859			

Analyte	Result (mg/L)	Qualifier	RL
Mercury	ND		0.00020

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: EBHH-GW-4

Lab Sample ID:	720-413-9	Date Sampled:	11/11/2005 1030
Client Matrix:	Water	Date Received:	11/11/2005 1340

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1924	Instrument ID:	Varian ICP
Preparation:	3010A	Prep Batch: 720-1816	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	50 mL
Date Analyzed:	11/17/2005 0915		Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 0856			

Analyte	Result (mg/L)	Qualifier	RL
Silver	ND		0.0050
Arsenic	ND		0.0050
Barium	0.14		0.0050
Cadmium	ND		0.0020
Chromium	ND		0.0050
Lead	ND		0.0050
Selenium	ND		0.0050

7470A Mercury in Liquid Waste (Manual Cold Vapor Technique)

Method:	7470A	Analysis Batch: 720-1793	Instrument ID:	FIMS 100
Preparation:	7470A	Prep Batch: 720-1754	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	25 mL
Date Analyzed:	11/15/2005 1329		Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 0859			

Analyte	Result (mg/L)	Qualifier	RL
Mercury	ND		0.00020

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: EBHH-GW-5

Lab Sample ID: 720-413-10
Client Matrix: Water

Date Sampled: 11/11/2005 1100
Date Received: 11/11/2005 1340

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method: 6010B
Preparation: 3010A
Dilution: 1.0
Date Analyzed: 11/17/2005 0926
Date Prepared: 11/16/2005 0856

Analysis Batch: 720-1924
Prep Batch: 720-1816

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 50 mL

Analyte	Result (mg/L)	Qualifier	RL
Silver	ND		0.0050
Arsenic	ND		0.0050
Barium	0.15		0.0050
Cadmium	ND		0.0020
Chromium	ND		0.0050
Lead	ND		0.0050
Selenium	ND		0.0050

7470A Mercury in Liquid Waste (Manual Cold Vapor Technique)

Method: 7470A
Preparation: 7470A
Dilution: 1.0
Date Analyzed: 11/15/2005 1333
Date Prepared: 11/15/2005 0859

Analysis Batch: 720-1793
Prep Batch: 720-1754

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 25 mL
Final Weight/Volume: 50 mL

Analyte	Result (mg/L)	Qualifier	RL
Mercury	ND		0.00020

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: EBHH-50A

Lab Sample ID:	720-413-13		Date Sampled:	11/11/2005 1045
Client Matrix:	Solid	% Moisture: 10.5	Date Received:	11/11/2005 1340

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1964	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1909	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/18/2005 0754		Final Weight/Volume:	50 mL
Date Prepared:	11/17/2005 1306			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		7.8		1.1
Barium		170		1.1
Cadmium		1.6		0.55
Chromium		41		1.1
Lead		100		1.1
Selenium		ND		2.2

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1952	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1912	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/18/2005 0725		Final Weight/Volume:	50 mL
Date Prepared:	11/17/2005 1318			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.16		0.056

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: EBHH-49A

Lab Sample ID:	720-413-14		Date Sampled:	11/11/2005 1145
Client Matrix:	Solid	% Moisture: 5.7	Date Received:	11/11/2005 1340

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1964	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1909	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/18/2005 0758		Final Weight/Volume:	50 mL
Date Prepared:	11/17/2005 1306			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.1
Arsenic		7.2		1.1
Barium		160		1.1
Cadmium		1.9		0.53
Chromium		26		1.1
Lead		50		1.1
Selenium		ND		2.1

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1952	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1912	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.01 g
Date Analyzed:	11/18/2005 0726		Final Weight/Volume:	50 mL
Date Prepared:	11/17/2005 1318			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.24		0.052

Client: Weston Solutions,Inc

Job Number: 720-413-1

Client Sample ID: EBHH-57A

Lab Sample ID:	720-413-15		Date Sampled:	11/11/2005 1200
Client Matrix:	Solid	% Moisture: 5.0	Date Received:	11/11/2005 1340

6010B Inductively Coupled Plasma - Atomic Emission Spectrometry

Method:	6010B	Analysis Batch: 720-1964	Instrument ID:	Varian ICP
Preparation:	3050B	Prep Batch: 720-1909	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.03 g
Date Analyzed:	11/18/2005 0816		Final Weight/Volume:	50 mL
Date Prepared:	11/17/2005 1306			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Silver		ND		1.0
Arsenic		7.0		1.0
Barium		150		1.0
Cadmium		1.6		0.51
Chromium		43		1.0
Lead		110		1.0
Selenium		ND		2.0

7471A Mercury in Solid or Semisolid Waste (Manual Cold Vapor Technique)

Method:	7471A	Analysis Batch: 720-1952	Instrument ID:	FIMS 100
Preparation:	7471A	Prep Batch: 720-1912	Lab File ID:	N/A
Dilution:	1.0		Initial Weight/Volume:	1.00 g
Date Analyzed:	11/18/2005 0730		Final Weight/Volume:	50 mL
Date Prepared:	11/17/2005 1318			

Analyte	DryWt Corrected: Y	Result (mg/Kg)	Qualifier	RL
Mercury		0.13		0.053

General Chemistry

Client Sample ID: EBHH-45B

Lab Sample ID: 720-413-3
Client Matrix: Solid

Date Sampled: 11/11/2005 0806
Date Received: 11/11/2005 1340

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	15		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-45D

Lab Sample ID: 720-413-4
Client Matrix: Solid

Date Sampled: 11/11/2005 0806
Date Received: 11/11/2005 1340

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	14		%	0.010	1.0	160.3

Anly Batch: 720-1872 Date Analyzed 11/15/2005 1400

Client Sample ID: EBHH-48B

Lab Sample ID: 720-413-5
Client Matrix: Solid

Date Sampled: 11/11/2005 0806
Date Received: 11/11/2005 1340

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	15		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-49C

Lab Sample ID: 720-413-6
Client Matrix: Solid

Date Sampled: 11/11/2005 0825
Date Received: 11/11/2005 1340

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	14		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-50C

Lab Sample ID: 720-413-7
Client Matrix: Solid

Date Sampled: 11/11/2005 0830
Date Received: 11/11/2005 1340

General Chemistry

Client Sample ID: EBHH-50C

Lab Sample ID: 720-413-7
Client Matrix: Solid

Date Sampled: 11/11/2005 0830
Date Received: 11/11/2005 1340

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	14		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-58C

Lab Sample ID: 720-413-8
Client Matrix: Solid

Date Sampled: 11/11/2005 0835
Date Received: 11/11/2005 1340

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	11		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-50A

Lab Sample ID: 720-413-13
Client Matrix: Solid

Date Sampled: 11/11/2005 1045
Date Received: 11/11/2005 1340

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	11		%	0.010	1.0	160.3

Anly Batch: 720-2055 Date Analyzed 11/18/2005 1340

Client Sample ID: EBHH-49A

Lab Sample ID: 720-413-14
Client Matrix: Solid

Date Sampled: 11/11/2005 1145
Date Received: 11/11/2005 1340

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	7.6		%	0.010	1.0	160.3

Anly Batch: 720-2403 Date Analyzed 11/18/2005 1400

Client Sample ID: EBHH-57A

Lab Sample ID: 720-413-15
Client Matrix: Solid

Date Sampled: 11/11/2005 1200
Date Received: 11/11/2005 1340

General Chemistry

Client Sample ID: **EBHH-57A**

Lab Sample ID: 720-413-15
Client Matrix: Solid

Date Sampled: 11/11/2005 1200
Date Received: 11/11/2005 1340

Analyte	Result	Qual	Units	RL	Dil	Method
Percent Moisture	9.6		%	0.010	1.0	160.3

Anly Batch: 720-2403 Date Analyzed 11/18/2005 1400

DATA REPORTING QUALIFIERS

Client: Weston Solutions,Inc

Job Number: 720-413-1

Lab Section	Qualifier	Description
GC/MS VOA		
	*	LCS, LCSD, MS, MSD, MD, or Surrogate exceeds the control limits
GC/MS Semi VOA		
	*	LCS, LCSD, MS, MSD, MD, or Surrogate exceeds the control limits
GC Semi VOA		
	*	LCS, LCSD, MS, MSD, MD, or Surrogate exceeds the control limits
	N	MS, MSD: Spike recovery exceeds upper or lower control limits.
	D	Surrogate or matrix spike recoveries were not obtained because the extract was diluted for analysis; also compounds analyzed at a dilution will be flagged with a D.
Metals		
	*	LCS, LCSD, MS, MSD, MD, or Surrogate exceeds the control limits
	N	MS, MSD: Spike recovery exceeds upper or lower control limits.

Client: Weston Solutions,Inc

Job Number: 720-413-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC/MS VOA				
Analysis Batch:720-2239				
LCS 720-2239/16	Lab Control Spike	Water	8260B	
MB 720-2239/17	Method Blank	Water	8260B	
720-413-2	EBHH-EB-4	Water	8260B	
720-413-9	EBHH-GW-4	Water	8260B	
720-413-10	EBHH-GW-5	Water	8260B	
720-518-B-5 MS	Matrix Spike	Water	8260B	
720-518-B-5 MSD	Matrix Spike Duplicate	Water	8260B	
Analysis Batch:720-2362				
LCS 720-2362/7	Lab Control Spike	Solid	8260B	
LCSD 720-2362/6	Lab Control Spike Duplicate	Solid	8260B	
MB 720-2362/8	Method Blank	Solid	8260B	
720-413-5	EBHH-48B	Solid	8260B	
720-527-B-7 MS	Matrix Spike	Solid	8260B	
720-527-B-7 MSD	Matrix Spike Duplicate	Solid	8260B	
Analysis Batch:720-2417				
LCS 720-2417/2	Lab Control Spike	Solid	8260B	
LCSD 720-2417/1	Lab Control Spike Duplicate	Solid	8260B	
MB 720-2417/3	Method Blank	Solid	8260B	
720-413-6	EBHH-49C	Solid	8260B	
720-413-8	EBHH-58C	Solid	8260B	
Prep Batch: 720-2423				
LCS 720-2423/3-A	Lab Control Spike	Solid	5035	
LCSD 720-2423/4-A	Lab Control Spike Duplicate	Solid	5035	
MB 720-2423/2-A	Method Blank	Solid	5035	
720-413-4	EBHH-45D	Solid	5035	
Analysis Batch:720-2427				
LCS 720-2427/3	Lab Control Spike	Solid	8260B	
LCSD 720-2427/2	Lab Control Spike Duplicate	Solid	8260B	
MB 720-2427/4	Method Blank	Solid	8260B	
720-413-3	EBHH-45B	Solid	8260B	
Analysis Batch:720-2456				
LCS 720-2456/13	Lab Control Spike	Water	8260B	
LCSD 720-2456/12	Lab Control Spike Duplicate	Water	8260B	
MB 720-2456/14	Method Blank	Water	8260B	
720-413-2	EBHH-EB-4	Water	8260B	
Analysis Batch:720-2416				
LCS 720-2423/3-A	Lab Control Spike	Solid	8260B	720-2423
LCSD 720-2423/4-A	Lab Control Spike Duplicate	Solid	8260B	720-2423
MB 720-2423/2-A	Method Blank	Solid	8260B	720-2423
720-413-4	EBHH-45D	Solid	8260B	720-2423

Client: Weston Solutions,Inc

Job Number: 720-413-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC/MS Semi VOA				
Prep Batch: 720-1797				
720-413-1	EBHH-EB-2	Water	3510C	
720-413-2	EBHH-EB-4	Water	3510C	
720-413-9	EBHH-GW-4	Water	3510C	
720-413-10	EBHH-GW-5	Water	3510C	
720-413-12	EBHH-EB3	Water	3510C	
Prep Batch: 720-1920				
LCS 720-1920/2-A	Lab Control Spike	Solid	3550B	
LCSD 720-1920/3-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1920/1-A	Method Blank	Solid	3550B	
720-413-6	EBHH-49C	Solid	3550B	
720-413-7	EBHH-50C	Solid	3550B	
720-413-13	EBHH-50A	Solid	3550B	
720-413-14	EBHH-49A	Solid	3550B	
720-413-14MS	Matrix Spike	Solid	3550B	
720-413-14MSD	Matrix Spike Duplicate	Solid	3550B	
Analysis Batch:720-2227				
720-413-1	EBHH-EB-2	Water	8270C	720-1797
720-413-2	EBHH-EB-4	Water	8270C	720-1797
720-413-9	EBHH-GW-4	Water	8270C	720-1797
720-413-10	EBHH-GW-5	Water	8270C	720-1797
720-413-12	EBHH-EB3	Water	8270C	720-1797
Analysis Batch:720-2258				
LCS 720-1920/2-A	Lab Control Spike	Solid	8270C	720-1920
LCSD 720-1920/3-A	Lab Control Spike Duplicate	Solid	8270C	720-1920
MB 720-1920/1-A	Method Blank	Solid	8270C	720-1920
720-413-6	EBHH-49C	Solid	8270C	720-1920
720-413-7	EBHH-50C	Solid	8270C	720-1920
720-413-13	EBHH-50A	Solid	8270C	720-1920
720-413-14	EBHH-49A	Solid	8270C	720-1920
720-413-14MS	Matrix Spike	Solid	8270C	720-1920
720-413-14MSD	Matrix Spike Duplicate	Solid	8270C	720-1920

Client: Weston Solutions,Inc

Job Number: 720-413-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC VOA				
Analysis Batch:720-1869				
LCS 720-1869/2	Lab Control Spike	Water	8015B	
MB 720-1869/1	Method Blank	Water	8015B	
720-413-A-9 MSMS	Matrix Spike	Water	8021B	
720-413-9	EBHH-GW-4	Water	8015B	
720-413-9MSD	Matrix Spike Duplicate	Water	8015B	
720-413-10	EBHH-GW-5	Water	8021B	
720-413-11	EBHH-GW-3	Water	8021B	
720-413-12	EBHH-EB3	Water	8021B	
Analysis Batch:720-1991				
LCS 720-1991/2	Lab Control Spike	Water	8021B	
MB 720-1991/1	Method Blank	Water	8021B	
720-413-9	EBHH-GW-4	Water	8021B	
720-413-9MS	Matrix Spike	Water	8021B	
720-413-9MSD	Matrix Spike Duplicate	Water	8021B	
720-413-10	EBHH-GW-5	Water	8021B	
720-413-11	EBHH-GW-3	Water	8021B	
720-413-12	EBHH-EB3	Water	8021B	
Analysis Batch:720-1997				
LCS 720-1997/2	Lab Control Spike	Water	8015B	
MB 720-1997/1	Method Blank	Water	8015B	
720-413-9	EBHH-GW-4	Water	8015B	
720-413-10	EBHH-GW-5	Water	8015B	
720-413-11	EBHH-GW-3	Water	8015B	
720-413-12	EBHH-EB3	Water	8015B	

Client: Weston Solutions,Inc

Job Number: 720-413-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC Semi VOA				
Prep Batch: 720-1752				
LCS 720-1752/1-A	Lab Control Spike	Water	3510C	
LCSD 720-1752/2-A	Lab Control Spike Duplicate	Water	3510C	
MB 720-1752/3-A	Method Blank	Water	3510C	
720-413-1	EBHH-EB-2	Water	3510C	
720-413-2	EBHH-EB-4	Water	3510C	
720-413-9	EBHH-GW-4	Water	3510C	
Prep Batch: 720-1753				
LCS 720-1753/1-A	Lab Control Spike	Water	3510C	
LCSD 720-1753/2-A	Lab Control Spike Duplicate	Water	3510C	
MB 720-1753/3-A	Method Blank	Water	3510C	
720-413-1	EBHH-EB-2	Water	3510C	
720-413-1MS	Matrix Spike	Water	3510C	
720-413-1MSD	Matrix Spike Duplicate	Water	3510C	
720-413-2	EBHH-EB-4	Water	3510C	
720-413-9	EBHH-GW-4	Water	3510C	
720-413-10	EBHH-GW-5	Water	3510C	
720-413-11	EBHH-GW-3	Water	3510C	
720-413-12	EBHH-EB3	Water	3510C	
Prep Batch: 720-1919				
LCS 720-1919/2-A	Lab Control Spike	Solid	3550B	
LCSD 720-1919/3-A	Lab Control Spike Duplicate	Solid	3550B	
MB 720-1919/1-A	Method Blank	Solid	3550B	
720-413-3	EBHH-45B	Solid	3550B	
720-413-4	EBHH-45D	Solid	3550B	
720-413-5	EBHH-48B	Solid	3550B	
720-413-6	EBHH-49C	Solid	3550B	
720-413-7	EBHH-50C	Solid	3550B	
720-413-8	EBHH-58C	Solid	3550B	
720-413-13	EBHH-50A	Solid	3550B	
720-413-14	EBHH-49A	Solid	3550B	
720-413-14MS	Matrix Spike	Solid	3550B	
720-413-14MSD	Matrix Spike Duplicate	Solid	3550B	
Prep Batch: 720-1983				
LCS 720-1983/2-A	Lab Control Spike	Water	3510C	
LCSD 720-1983/3-B	Lab Control Spike Duplicate	Water	3510C	
MB 720-1983/1-A	Method Blank	Water	3510C	
720-413-10	EBHH-GW-5	Water	3510C	

Client: Weston Solutions,Inc

Job Number: 720-413-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
GC Semi VOA				
Analysis Batch:720-2045				
LCS 720-1752/1-A	Lab Control Spike	Water	8081A	720-1752
LCSD 720-1752/2-A	Lab Control Spike Duplicate	Water	8081A	720-1752
MB 720-1752/3-A	Method Blank	Water	8081A	720-1752
720-413-1	EBHH-EB-2	Water	8081A	720-1752
720-413-2	EBHH-EB-4	Water	8081A	720-1752
720-413-9	EBHH-GW-4	Water	8081A	720-1752
Analysis Batch:720-1935				
LCS 720-1753/1-A	Lab Control Spike	Water	8015B	720-1753
LCSD 720-1753/2-A	Lab Control Spike Duplicate	Water	8015B	720-1753
MB 720-1753/3-A	Method Blank	Water	8015B	720-1753
720-413-1	EBHH-EB-2	Water	8015B	720-1753
720-413-1MS	Matrix Spike	Water	8015B	720-1753
720-413-1MSD	Matrix Spike Duplicate	Water	8015B	720-1753
720-413-2	EBHH-EB-4	Water	8015B	720-1753
720-413-9	EBHH-GW-4	Water	8015B	720-1753
720-413-10	EBHH-GW-5	Water	8015B	720-1753
720-413-11	EBHH-GW-3	Water	8015B	720-1753
720-413-12	EBHH-EB3	Water	8015B	720-1753
Analysis Batch:720-2006				
LCS 720-1919/2-A	Lab Control Spike	Solid	8015B	720-1919
LCSD 720-1919/3-A	Lab Control Spike Duplicate	Solid	8015B	720-1919
MB 720-1919/1-A	Method Blank	Solid	8015B	720-1919
720-413-3	EBHH-45B	Solid	8015B	720-1919
720-413-4	EBHH-45D	Solid	8015B	720-1919
720-413-5	EBHH-48B	Solid	8015B	720-1919
720-413-6	EBHH-49C	Solid	8015B	720-1919
720-413-7	EBHH-50C	Solid	8015B	720-1919
720-413-8	EBHH-58C	Solid	8015B	720-1919
720-413-13	EBHH-50A	Solid	8015B	720-1919
720-413-14	EBHH-49A	Solid	8015B	720-1919
720-413-14MS	Matrix Spike	Solid	8015B	720-1919
720-413-14MSD	Matrix Spike Duplicate	Solid	8015B	720-1919
Analysis Batch:720-2116				
LCS 720-1983/2-A	Lab Control Spike	Water	8081A	720-1983
LCSD 720-1983/3-B	Lab Control Spike Duplicate	Water	8081A	720-1983
MB 720-1983/1-A	Method Blank	Water	8081A	720-1983
720-413-10	EBHH-GW-5	Water	8081A	720-1983

Client: Weston Solutions,Inc

Job Number: 720-413-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Prep Batch: 720-1754				
LCS 720-1754/2-A	Lab Control Spike	Water	7470A	
LCSD 720-1754/3-A	Lab Control Spike Duplicate	Water	7470A	
MB 720-1754/1-A	Method Blank	Water	7470A	
720-413-1	EBHH-EB-2	Water	7470A	
720-413-2	EBHH-EB-4	Water	7470A	
720-413-9	EBHH-GW-4	Water	7470A	
720-413-9MS	Matrix Spike	Water	7470A	
720-413-9MSD	Matrix Spike Duplicate	Water	7470A	
720-413-10	EBHH-GW-5	Water	7470A	
Prep Batch: 720-1816				
LCS 720-1816/2-A	Lab Control Spike	Water	3010A	
LCSD 720-1816/3-A	Lab Control Spike Duplicate	Water	3010A	
MB 720-1816/1-A	Method Blank	Water	3010A	
720-413-1	EBHH-EB-2	Water	3010A	
720-413-2	EBHH-EB-4	Water	3010A	
720-413-9	EBHH-GW-4	Water	3010A	
720-413-9MS	Matrix Spike	Water	3010A	
720-413-9MSD	Matrix Spike Duplicate	Water	3010A	
720-413-10	EBHH-GW-5	Water	3010A	
Prep Batch: 720-1909				
LCS 720-1909/2-A	Lab Control Spike	Solid	3050B	
LCSD 720-1909/3-A	Lab Control Spike Duplicate	Solid	3050B	
MB 720-1909/1-A	Method Blank	Solid	3050B	
720-413-13	EBHH-50A	Solid	3050B	
720-413-14	EBHH-49A	Solid	3050B	
720-413-14MS	Matrix Spike	Solid	3050B	
720-413-14MSD	Matrix Spike Duplicate	Solid	3050B	
720-413-15	EBHH-57A	Solid	3050B	
Prep Batch: 720-1912				
LCS 720-1912/2-A	Lab Control Spike	Solid	7471A	
LCSD 720-1912/3-A	Lab Control Spike Duplicate	Solid	7471A	
MB 720-1912/1-A	Method Blank	Solid	7471A	
720-413-13	EBHH-50A	Solid	7471A	
720-413-14	EBHH-49A	Solid	7471A	
720-413-14MS	Matrix Spike	Solid	7471A	
720-413-14MSD	Matrix Spike Duplicate	Solid	7471A	
720-413-15	EBHH-57A	Solid	7471A	

Client: Weston Solutions,Inc

Job Number: 720-413-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
Metals				
Analysis Batch:720-1793				
LCS 720-1754/2-A	Lab Control Spike	Water	7470A	720-1754
LCSD 720-1754/3-A	Lab Control Spike Duplicate	Water	7470A	720-1754
MB 720-1754/1-A	Method Blank	Water	7470A	720-1754
720-413-1	EBHH-EB-2	Water	7470A	720-1754
720-413-2	EBHH-EB-4	Water	7470A	720-1754
720-413-9	EBHH-GW-4	Water	7470A	720-1754
720-413-9MS	Matrix Spike	Water	7470A	720-1754
720-413-9MSD	Matrix Spike Duplicate	Water	7470A	720-1754
720-413-10	EBHH-GW-5	Water	7470A	720-1754
Analysis Batch:720-1924				
LCS 720-1816/2-A	Lab Control Spike	Water	6010B	720-1816
LCSD 720-1816/3-A	Lab Control Spike Duplicate	Water	6010B	720-1816
MB 720-1816/1-A	Method Blank	Water	6010B	720-1816
720-413-1	EBHH-EB-2	Water	6010B	720-1816
720-413-2	EBHH-EB-4	Water	6010B	720-1816
720-413-9	EBHH-GW-4	Water	6010B	720-1816
720-413-9MS	Matrix Spike	Water	6010B	720-1816
720-413-9MSD	Matrix Spike Duplicate	Water	6010B	720-1816
720-413-10	EBHH-GW-5	Water	6010B	720-1816
Analysis Batch:720-1964				
LCS 720-1909/2-A	Lab Control Spike	Solid	6010B	720-1909
LCSD 720-1909/3-A	Lab Control Spike Duplicate	Solid	6010B	720-1909
MB 720-1909/1-A	Method Blank	Solid	6010B	720-1909
720-413-13	EBHH-50A	Solid	6010B	720-1909
720-413-14	EBHH-49A	Solid	6010B	720-1909
720-413-14MS	Matrix Spike	Solid	6010B	720-1909
720-413-14MSD	Matrix Spike Duplicate	Solid	6010B	720-1909
720-413-15	EBHH-57A	Solid	6010B	720-1909
Analysis Batch:720-1952				
LCS 720-1912/2-A	Lab Control Spike	Solid	7471A	720-1912
LCSD 720-1912/3-A	Lab Control Spike Duplicate	Solid	7471A	720-1912
MB 720-1912/1-A	Method Blank	Solid	7471A	720-1912
720-413-13	EBHH-50A	Solid	7471A	720-1912
720-413-14	EBHH-49A	Solid	7471A	720-1912
720-413-14MS	Matrix Spike	Solid	7471A	720-1912
720-413-14MSD	Matrix Spike Duplicate	Solid	7471A	720-1912
720-413-15	EBHH-57A	Solid	7471A	720-1912

Client: Weston Solutions,Inc

Job Number: 720-413-1

QC Association Summary

Lab Sample ID	Client Sample ID	Client Matrix	Method	Prep Batch
General Chemistry				
Analysis Batch:720-1872				
MB 720-1872/1	Method Blank	Solid	160.3	
720-413-3	EBHH-45B	Solid	160.3	
720-413-4	EBHH-45D	Solid	160.3	
720-413-4DU	Duplicate	Solid	160.3	
Analysis Batch:720-2055				
MB 720-2055/1	Method Blank	Solid	160.3	
720-413-5	EBHH-48B	Solid	160.3	
720-413-6	EBHH-49C	Solid	160.3	
720-413-7	EBHH-50C	Solid	160.3	
720-413-8	EBHH-58C	Solid	160.3	
720-413-13	EBHH-50A	Solid	160.3	
Analysis Batch:720-2060				
MB 720-2060/1	Method Blank	Solid	160.3	
Analysis Batch:720-2403				
MB 720-2403/1	Method Blank	Solid	160.3	
720-413-14	EBHH-49A	Solid	160.3	
720-413-15	EBHH-57A	Solid	160.3	
720-413-15DU	Duplicate	Solid	160.3	

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-2239

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 720-2239/17
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/23/2005 1508
Date Prepared: 11/23/2005 1508

Analysis Batch: 720-2239
Prep Batch: N/A
Units: ug/L

Instrument ID: Saturn 2K3
Lab File ID: d:\data\200511\112305\mb
Initial Weight/Volume: 40 mL
Final Weight/Volume: 40 mL

Analyte	Result	Qual	RL
1,1-Dichloroethene	ND		0.50
1,1-Dichloroethane	ND		0.50
Dichlorodifluoromethane	ND		0.50
Vinyl chloride	ND		0.50
Chloroethane	ND		1.0
Trichlorofluoromethane	ND		1.0
Methylene Chloride	ND		5.0
trans-1,2-Dichloroethene	ND		0.50
cis-1,2-Dichloroethene	ND		0.50
Chloroform	ND		1.0
1,1,1-Trichloroethane	ND		0.50
Carbon tetrachloride	ND		0.50
1,2-Dichloroethane	ND		0.50
Trichloroethene	ND		0.50
1,2-Dichloropropane	ND		0.50
Dichlorobromomethane	ND		0.50
trans-1,3-Dichloropropene	ND		0.50
cis-1,3-Dichloropropene	ND		0.50
1,1,2-Trichloroethane	ND		0.50
Tetrachloroethene	ND		0.50
Chlorodibromomethane	ND		0.50
Chlorobenzene	ND		0.50
Bromoform	ND		1.0
1,1,2,2-Tetrachloroethane	ND		0.50
1,3-Dichlorobenzene	ND		0.50
1,4-Dichlorobenzene	ND		0.50
1,2-Dichlorobenzene	ND		0.50
Chloromethane	ND		1.0
Bromomethane	ND		1.0
1,1,2-Trichloro-1,2,2-trifluoroethane	ND		0.50
EDB	ND		0.50
1,2,4-Trichlorobenzene	ND		1.0

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Laboratory Control Sample - Batch: 720-2239

Method: 8260B
Preparation: 5030B

Lab Sample ID: LCS 720-2239/16
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/23/2005 1434
Date Prepared: 11/23/2005 1434

Analysis Batch: 720-2239
Prep Batch: N/A
Units: ug/L

Instrument ID: Saturn 2K3
Lab File ID: d:\data\200511\112305\ls-v
Initial Weight/Volume: 40 mL
Final Weight/Volume: 40 mL

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
1,1-Dichloroethene	20.0	19	97	65 - 125	
Trichloroethene	20.0	18	89	74 - 134	
Chlorobenzene	20.0	20	102	61 - 121	

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-2239

Method: 8260B
Preparation: 5030B

MS Lab Sample ID: 720-518-B-5 MS
Client Matrix: Water
Dilution: 20
Date Analyzed: 11/23/2005 1648
Date Prepared: 11/23/2005 1648

Analysis Batch: 720-2239
Prep Batch: N/A

Instrument ID: Saturn 2K3
Lab File ID: d:\data\200511\112305\72
Initial Weight/Volume: 40 mL
Final Weight/Volume: 40 mL

MSD Lab Sample ID: 720-518-B-5 MSD
Client Matrix: Water
Dilution: 20
Date Analyzed: 11/23/2005 1721
Date Prepared: 11/23/2005 1721

Analysis Batch: 720-2239
Prep Batch: N/A

Instrument ID: Saturn 2K3
Lab File ID: d:\data\200511\112305\720
Initial Weight/Volume: 40 mL
Final Weight/Volume: 40 mL

Analyte	% Rec. MS	% Rec. MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
1,1-Dichloroethene	93	97	65 - 125	3	20		
Trichloroethene	89	90	74 - 134	1	20		
Chlorobenzene	103	103	61 - 121	0	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-2362

Method: 8260B
Preparation: N/A

Lab Sample ID: MB 720-2362/8
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/29/2005 0857
Date Prepared: N/A

Analysis Batch: 720-2362
Prep Batch: N/A
Units: ug/Kg

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\112905\mb
Initial Weight/Volume: 5 g
Final Weight/Volume: 10 mL

Analyte	Result	Qual	RL
Benzene	ND		5.0
Ethylbenzene	ND		5.0
Toluene	ND		5.0
Xylenes, Total	ND		10
Gasoline Range Organics (GRO)-C5-C12	ND		1000

Surrogate	% Rec	Acceptance Limits
Toluene-d8	102	70 - 130
1,2-Dichloroethane-d4	92	60 - 140

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-2362

Method: 8260B
Preparation: N/A

LCS Lab Sample ID: LCS 720-2362/7
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/29/2005 0804
Date Prepared: N/A

Analysis Batch: 720-2362
Prep Batch: N/A
Units: ug/Kg

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\112905\ls-so-5-1
Initial Weight/Volume: 5 g
Final Weight/Volume: 10 mL

LCSD Lab Sample ID: LCSD 720-2362/6
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/29/2005 0831
Date Prepared: N/A

Analysis Batch: 720-2362
Prep Batch: N/A
Units: ug/Kg

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\112905\ld-so-5-11
Initial Weight/Volume: 5 g
Final Weight/Volume: 10 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	92	94	69 - 129	2	20		
Toluene	98	101	70 - 130	3	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
Toluene-d8	101	105	70 - 130
1,2-Dichloroethane-d4	83	90	60 - 140

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
Toluene-d8	106	98	70 - 130
1,2-Dichloroethane-d4	81	88	60 - 140

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-2417

Method: 8260B
Preparation: N/A

Lab Sample ID: MB 720-2417/3
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/25/2005 2230
Date Prepared: N/A

Analysis Batch: 720-2417
Prep Batch: N/A
Units: ug/Kg

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\112505\mb
Initial Weight/Volume: 5 g
Final Weight/Volume: 10 mL

Analyte	Result	Qual	RL
Benzene	ND		5.0
Ethylbenzene	ND		5.0
Toluene	ND		5.0
Xylenes, Total	ND		10

Surrogate	% Rec	Acceptance Limits
Toluene-d8	95	70 - 130
1,2-Dichloroethane-d4	81	60 - 140

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-2417

Method: 8260B
Preparation: N/A

LCS Lab Sample ID: LCS 720-2417/2
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/25/2005 2138
Date Prepared: N/A

Analysis Batch: 720-2417
Prep Batch: N/A
Units: ug/Kg

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\112505\ls-so-5-1
Initial Weight/Volume: 5 g
Final Weight/Volume: 10 mL

LCSD Lab Sample ID: LCSD 720-2417/1
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/25/2005 2204
Date Prepared: N/A

Analysis Batch: 720-2417
Prep Batch: N/A
Units: ug/Kg

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\112505\ld-so-5-11
Initial Weight/Volume: 5 g
Final Weight/Volume: 10 mL

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	111	102	69 - 129	8	20		
Toluene	101	98	70 - 130	3	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
Toluene-d8	95	97	70 - 130
1,2-Dichloroethane-d4	87	79	60 - 140

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-2423

Method: 8260B
Preparation: 5035

Lab Sample ID: MB 720-2423/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/25/2005 2302
Date Prepared: 11/11/2005 1145

Analysis Batch: 720-2416
Prep Batch: 720-2423
Units: ug/Kg

Instrument ID: Varian 3900E
Lab File ID: c:\varianws\data\200511\11
Initial Weight/Volume: 5 g
Final Weight/Volume: 5 g

Analyte	Result	Qual	RL
Benzene	ND		2.5
Ethylbenzene	ND		2.5
Toluene	ND		2.5
Xylenes, Total	ND		5.0
Gasoline Range Organics (GRO)-C5-C12	ND		500

Surrogate	% Rec	Acceptance Limits
Toluene-d8	98	70 - 130
1,2-Dichloroethane-d4	110	60 - 140

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-2423

Method: 8260B
Preparation: 5035

LCS Lab Sample ID: LCS 720-2423/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/25/2005 2217
Date Prepared: 11/11/2005 1145

Analysis Batch: 720-2416
Prep Batch: 720-2423
Units: ug/Kg

Instrument ID: Varian 3900E
Lab File ID: c:\varianws\data\200511\112505
Initial Weight/Volume: 5 g
Final Weight/Volume: 5 g

LCSD Lab Sample ID: LCSD 720-2423/4-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/25/2005 2240
Date Prepared: 11/11/2005 1145

Analysis Batch: 720-2416
Prep Batch: 720-2423
Units: ug/Kg

Instrument ID: Varian 3900E
Lab File ID: c:\varianws\data\200511\112505\l
Initial Weight/Volume: 5 g
Final Weight/Volume: 5 g

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	103	106	69 - 129	3	20		
Toluene	104	107	70 - 130	3	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
Toluene-d8	100	100	70 - 130
1,2-Dichloroethane-d4	99	99	60 - 140

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-2427

Method: 8260B
Preparation: N/A

Lab Sample ID: MB 720-2427/4
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/29/2005 2049
Date Prepared: N/A

Analysis Batch: 720-2427
Prep Batch: N/A
Units: ug/Kg

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\112905\mb
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Result	Qual	RL
Benzene	ND		2.5
Ethylbenzene	ND		2.5
Toluene	ND		2.5
Xylenes, Total	ND		5.0
Gasoline Range Organics (GRO)-C5-C12	ND		500

Surrogate	% Rec	Acceptance Limits
Toluene-d8	96	70 - 130
1,2-Dichloroethane-d4	90	60 - 140

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-2427

Method: 8260B
Preparation: N/A

LCS Lab Sample ID: LCS 720-2427/3
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/29/2005 1957
Date Prepared: N/A

Analysis Batch: 720-2427
Prep Batch: N/A
Units: ug/Kg

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\112905\ls-so-5-1
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

LCSD Lab Sample ID: LCSD 720-2427/2
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/29/2005 2023
Date Prepared: N/A

Analysis Batch: 720-2427
Prep Batch: N/A
Units: ug/Kg

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\112905\ld-so-5-11
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	102	91	69 - 129	11	20		
Toluene	100	89	70 - 130	12	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
Toluene-d8	94	95	70 - 130
1,2-Dichloroethane-d4	79	88	60 - 140

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-2456

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 720-2456/14
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/30/2005 2111
Date Prepared: 11/30/2005 2111

Analysis Batch: 720-2456
Prep Batch: N/A
Units: ug/L

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\113005\mb
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Result	Qual	RL
Benzene	ND		0.50
Ethylbenzene	ND		0.50
Toluene	ND		0.50
MTBE	ND		0.50
Xylenes, Total	ND		1.0
Gasoline Range Organics (GRO)-C5-C12	ND		50

Surrogate	% Rec	Acceptance Limits
Toluene-d8	93	77 - 121
1,2-Dichloroethane-d4	81	73 - 130

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-2456

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID: LCS 720-2456/13
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/30/2005 2019
Date Prepared: 11/30/2005 2019

Analysis Batch: 720-2456
Prep Batch: N/A
Units: ug/L

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\113005\ls-wa-5-
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

LCSD Lab Sample ID: LCSD 720-2456/12
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/30/2005 2045
Date Prepared: 11/30/2005 2045

Analysis Batch: 720-2456
Prep Batch: N/A
Units: ug/L

Instrument ID: Saturn 2100
Lab File ID: d:\data\200511\113005\ld-wa-5-1
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	85	84	69 - 129	2	25		
Toluene	94	97	70 - 130	4	25		
MTBE	94	96	65 - 165	2	25		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
Toluene-d8	97	98	77 - 121
1,2-Dichloroethane-d4	83	80	73 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-1920

Method: 8270C
Preparation: 3550B

Lab Sample ID: MB 720-1920/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/27/2005 1244
Date Prepared: 11/17/2005 1413

Analysis Batch: 720-2258
Prep Batch: 720-1920
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\11
Initial Weight/Volume: 30.09 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	Result	Qual	RL
Naphthalene	ND		5.0
Acenaphthene	ND		5.0
Acenaphthylene	ND		5.0
Fluorene	ND		5.0
Phenanthrene	ND		5.0
Anthracene	ND		5.0
Benzo[a]anthracene	ND		5.0
Chrysene	ND		5.0
Benzo[a]pyrene	ND		5.0
Benzo[b]fluoranthene	ND		5.0
Benzo[k]fluoranthene	ND		5.0
Benzo[g,h,i]perylene	ND		5.0
Indeno[1,2,3-cd]pyrene	ND		5.0
Fluoranthene	ND		5.0
Pyrene	ND		5.0
Dibenz(a,h)anthracene	ND		5.0

Surrogate	% Rec	Acceptance Limits
2-Fluorobiphenyl	75	30 - 115
Terphenyl-d14	90	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1920

Method: 8270C
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1920/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/27/2005 1312
Date Prepared: 11/17/2005 1413

Analysis Batch: 720-2258
Prep Batch: 720-1920
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\112705
Initial Weight/Volume: 30.03 g
Final Weight/Volume: 1 mL
Injection Volume:

LCSD Lab Sample ID: LCSD 720-1920/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/27/2005 1340
Date Prepared: 11/17/2005 1413

Analysis Batch: 720-2258
Prep Batch: 720-1920
Units: ug/Kg

Instrument ID: Sat 2K2
Lab File ID: c:\saturnws\data\200511\112705\l
Initial Weight/Volume: 30.05 g
Final Weight/Volume: 1 mL
Injection Volume:

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Naphthalene	69	63	21 - 133	9	35		
Acenaphthene	67	68	47 - 145	1	35		
Acenaphthylene	72	68	33 - 145	6	35		
Fluorene	76	75	59 - 121	1	35		
Phenanthrene	74	74	10 - 130	1	35		
Anthracene	70	69	27 - 133	1	35		
Benzo[a]anthracene	85	81	33 - 143	5	35		
Chrysene	72	72	17 - 168	0	35		
Benzo[a]pyrene	76	78	17 - 163	2	35		
Benzo[b]fluoranthene	81	80	24 - 159	2	35		
Benzo[k]fluoranthene	82	79	11 - 162	4	35		
Benzo[g,h,i]perylene	73	72	9 - 219	0	35		
Indeno[1,2,3-cd]pyrene	84	82	9 - 171	3	35		
Fluoranthene	76	76	26 - 137	0	35		
Pyrene	76	78	52 - 115	2	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
2-Fluorobiphenyl	73	70	30 - 115
Terphenyl-d14	86	85	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1920

Method: 8270C
Preparation: 3550B

MS Lab Sample ID:	720-413-14	Analysis Batch:	720-2258	Instrument ID:	Sat 2K2
Client Matrix:	Solid	Prep Batch:	720-1920	Lab File ID:	c:\saturnws\data\200511\
Dilution:	10			Initial Weight/Volume:	30.16 g
Date Analyzed:	11/27/2005 1558			Final Weight/Volume:	5 mL
Date Prepared:	11/17/2005 1413			Injection Volume:	

MSD Lab Sample ID:	720-413-14	Analysis Batch:	720-2258	Instrument ID:	Sat 2K2
Client Matrix:	Solid	Prep Batch:	720-1920	Lab File ID:	c:\saturnws\data\200511\11
Dilution:	10			Initial Weight/Volume:	30.13 g
Date Analyzed:	11/27/2005 1626			Final Weight/Volume:	5 mL
Date Prepared:	11/17/2005 1413			Injection Volume:	

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Naphthalene	50	50	21 - 133	1	35		
Acenaphthene	81	80	47 - 145	2	35		
Acenaphthylene	74	68	33 - 145	9	35		
Fluorene	72	69	59 - 121	6	35		
Phenanthrene	90	66	10 - 130	31	35		
Anthracene	58	45	27 - 133	17	35		
Benzo[a]anthracene	84	72	33 - 143	17	35		
Chrysene	70	56	17 - 168	17	35		
Benzo[a]pyrene	92	70	17 - 163	28	35		
Benzo[b]fluoranthene	93	70	24 - 159	29	35		
Benzo[k]fluoranthene	73	76	11 - 162	3	35		
Benzo[g,h,i]perylene	108	74	9 - 219	37	35		*
Indeno[1,2,3-cd]pyrene	71	37	9 - 171	65	35		*
Fluoranthene	87	55	26 - 137	34	35		
Pyrene	78	56	52 - 115	23	35		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
2-Fluorobiphenyl	67	59	30 - 115
Terphenyl-d14	79	74	18 - 137

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Surrogate	% Rec	Acceptance Limits
4-Bromofluorobenzene		

Surrogate	% Rec	Acceptance Limits
4-Bromofluorobenzene	97	50 - 150

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-1997

Method: 8015B
Preparation: 5030B

Lab Sample ID: MB 720-1997/1
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1041
Date Prepared: 11/15/2005 1041

Analysis Batch: 720-1997
Prep Batch: N/A
Units: ug/L

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Gasoline Range Organics (GRO)-C5-C12	ND		50

Surrogate	% Rec	Acceptance Limits
4-Bromofluorobenzene	92	50 - 150

Laboratory Control Sample - Batch: 720-1997

Method: 8015B
Preparation: 5030B

Lab Sample ID: LCS 720-1997/2
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1149
Date Prepared: 11/15/2005 1149

Analysis Batch: 720-1997
Prep Batch: N/A
Units: ug/L

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Gasoline Range Organics (GRO)-C5-C12	250	290	116	75 - 125	

Surrogate	% Rec	Acceptance Limits
4-Bromofluorobenzene	93	50 - 150

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Matrix Spike - Batch: 720-1869

Method: 8021B
Preparation: 5030B

Lab Sample ID: 720-413-A-9 MS
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/16/2005 0540
Date Prepared: 11/16/2005 0540

Analysis Batch: 720-1869
Prep Batch: N/A
Units: ug/L

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Sample Result/Qual	Spike Amount	Result	% Rec.	Limit	Qual
Benzene	ND	50.5	55	109	65 - 135	
Toluene	ND	50.4	52	104	65 - 135	
Ethylbenzene	ND	50.3	53	105	65 - 135	
Xylenes, Total	ND	152	150	101	65 - 135	

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-1991

Method: 8021B
Preparation: 5030B

Lab Sample ID: MB 720-1991/1
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1041
Date Prepared: 11/15/2005 1041

Analysis Batch: 720-1991
Prep Batch: N/A
Units: ug/L

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Benzene	ND		0.50
Toluene	ND		0.50
Ethylbenzene	ND		0.50
Xylenes, Total	ND		0.50

Surrogate	% Rec	Acceptance Limits
a,a,a-Trifluorotoluene (pid)	106	58 - 124
4-Bromofluorobenzene	97	50 - 150

Laboratory Control Sample - Batch: 720-1991

Method: 8021B
Preparation: 5030B

Lab Sample ID: LCS 720-1991/2
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1115
Date Prepared: 11/15/2005 1115

Analysis Batch: 720-1991
Prep Batch: N/A
Units: ug/L

Instrument ID: PID/FID Gas/Btex
Lab File ID: N/A
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Benzene	50.5	56	110	77 - 123	
Toluene	50.4	55	109	78 - 122	
Ethylbenzene	50.3	55	109	70 - 130	
Xylenes, Total	152	160	108	75 - 125	

Surrogate	% Rec	Acceptance Limits
a,a,a-Trifluorotoluene (pid)	104	58 - 124
4-Bromofluorobenzene	97	50 - 150

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1991

Method: 8021B
Preparation: 5030B

MS Lab Sample ID:	720-413-9	Analysis Batch: 720-1991	Instrument ID: PID/FID Gas/Btex
Client Matrix:	Water	Prep Batch: N/A	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 10 mL
Date Analyzed:	11/16/2005 0540		Final Weight/Volume: 10 mL
Date Prepared:	11/16/2005 0540		Injection Volume:
			Column ID: PRIMARY

MSD Lab Sample ID:	720-413-9	Analysis Batch: 720-1991	Instrument ID: PID/FID Gas/Btex
Client Matrix:	Water	Prep Batch: N/A	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 10 mL
Date Analyzed:	11/16/2005 0614		Final Weight/Volume: 10 mL
Date Prepared:	11/16/2005 0614		Injection Volume:
			Column ID: PRIMARY

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	109	110	65 - 135	1	20		
Toluene	104	105	65 - 135	1	20		
Ethylbenzene	105	106	65 - 135	2	20		
Xylenes, Total	101	102	65 - 135	1	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
a,a,a-Trifluorotoluene (pid)	100	101	58 - 124
4-Bromofluorobenzene	98	98	50 - 150

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-1753

Method: 8015B
Preparation: 3510C

Lab Sample ID: MB 720-1753/3-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/16/2005 1103
Date Prepared: 11/15/2005 0848

Analysis Batch: 720-1935
Prep Batch: 720-1753
Units: ug/L

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Diesel Range Organics [C10-C28]	ND		50
Motor Oil Range Organics [C24-C36]	ND		500

Surrogate	% Rec	Acceptance Limits
o-Terphenyl	94	60 - 130

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1753

Method: 8015B
Preparation: 3510C

LCS Lab Sample ID: LCS 720-1753/1-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/16/2005 1111
Date Prepared: 11/15/2005 0848

Analysis Batch: 720-1935
Prep Batch: 720-1753
Units: ug/L

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1753/2-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/16/2005 1139
Date Prepared: 11/15/2005 0848

Analysis Batch: 720-1935
Prep Batch: 720-1753
Units: ug/L

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 250 mL
Final Weight/Volume: 1 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Diesel Range Organics [C10-C28]	88	78	60 - 130	11	30		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
o-Terphenyl	79	77	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1753

Method: 8015B
Preparation: 3510C

MS Lab Sample ID:	720-413-1	Analysis Batch: 720-1935	Instrument ID: HP DRO3
Client Matrix:	Water	Prep Batch: 720-1753	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 250 mL
Date Analyzed:	11/17/2005 1858		Final Weight/Volume: 1 mL
Date Prepared:	11/15/2005 0848		Injection Volume:
			Column ID: PRIMARY

MSD Lab Sample ID:	720-413-1	Analysis Batch: 720-1935	Instrument ID: HP DRO3
Client Matrix:	Water	Prep Batch: 720-1753	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 250 mL
Date Analyzed:	11/17/2005 1925		Final Weight/Volume: 1 mL
Date Prepared:	11/15/2005 0848		Injection Volume:
			Column ID: PRIMARY

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Diesel Range Organics [C10-C28]	95	92	60 - 130	3	30		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
o-Terphenyl	88	86	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-1919

Method: 8015B
Preparation: 3550B

Lab Sample ID: MB 720-1919/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 2238
Date Prepared: 11/17/2005 1403

Analysis Batch: 720-2006
Prep Batch: 720-1919
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.36 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Diesel Range Organics [C10-C28]	ND		0.99
Motor Oil Range Organics [C24-C36]	ND		49

Surrogate	% Rec	Acceptance Limits
o-Terphenyl	70	60 - 130

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1919

Method: 8015B
Preparation: 3550B

LCS Lab Sample ID: LCS 720-1919/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 2305
Date Prepared: 11/17/2005 1403

Analysis Batch: 720-2006
Prep Batch: 720-1919
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.43 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1919/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/17/2005 2333
Date Prepared: 11/17/2005 1403

Analysis Batch: 720-2006
Prep Batch: 720-1919
Units: mg/Kg

Instrument ID: HP DRO3
Lab File ID: N/A
Initial Weight/Volume: 30.18 g
Final Weight/Volume: 5 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Diesel Range Organics [C10-C28]	87	93	60 - 130	7	30		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
o-Terphenyl	72	77	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1919

Method: 8015B
Preparation: 3550B

MS Lab Sample ID:	720-413-14	Analysis Batch: 720-2006	Instrument ID: HP DRO3
Client Matrix:	Solid	Prep Batch: 720-1919	Lab File ID: N/A
Dilution:	5.0		Initial Weight/Volume: 30.47 g
Date Analyzed:	11/17/2005 2331		Final Weight/Volume: 5 mL
Date Prepared:	11/17/2005 1403		Injection Volume:
			Column ID: PRIMARY

MSD Lab Sample ID:	720-413-14	Analysis Batch: 720-2006	Instrument ID: HP DRO3
Client Matrix:	Solid	Prep Batch: 720-1919	Lab File ID: N/A
Dilution:	5.0		Initial Weight/Volume: 30.33 g
Date Analyzed:	11/17/2005 2357		Final Weight/Volume: 5 mL
Date Prepared:	11/17/2005 1403		Injection Volume:
			Column ID: PRIMARY

Analyte	% Rec. MS	% Rec. MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Diesel Range Organics [C10-C28]	58	47	60 - 130	5	30	N	N

Surrogate	MS % Rec		MSD % Rec		Acceptance Limits
o-Terphenyl	0	D	0	D	60 - 130

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-1752

Method: 8081A
Preparation: 3510C

Lab Sample ID: MB 720-1752/3-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/17/2005 0545
Date Prepared: 11/15/2005 0827

Analysis Batch: 720-2045
Prep Batch: 720-1752
Units: ug/L

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 1000 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Aldrin	ND		0.060
Dieldrin	ND		0.060
Endrin aldehyde	ND		0.060
Endrin	ND		0.060
Endrin ketone	ND		0.060
Heptachlor	ND		0.060
Heptachlor epoxide	ND		0.060
4,4'-DDT	ND		0.060
4,4'-DDE	ND		0.060
4,4'-DDD	ND		0.060
Endosulfan I	ND		0.060
Endosulfan II	ND		0.060
alpha-BHC	ND		0.060
beta-BHC	ND		0.060
gamma-BHC (Lindane)	ND		0.060
delta-BHC	ND		0.060
Endosulfan sulfate	ND		0.060
Methoxychlor	ND		0.060
Toxaphene	ND		1.0
Chlordane (technical)	ND		1.0
alpha-Chlordane	ND		0.060
gamma-Chlordane	ND		0.060

Surrogate	% Rec	Acceptance Limits
Tetrachloro-m-xylene	91	62 - 123
DCB Decachlorobiphenyl	92	56 - 136

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1752

Method: 8081A
Preparation: 3510C

LCS Lab Sample ID: LCS 720-1752/1-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/17/2005 0614
Date Prepared: 11/15/2005 0827

Analysis Batch: 720-2045
Prep Batch: 720-1752
Units: ug/L

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 1000 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1752/2-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/17/2005 0642
Date Prepared: 11/15/2005 0827

Analysis Batch: 720-2045
Prep Batch: 720-1752
Units: ug/L

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 1000 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Aldrin	120	118	65 - 135	1	35		
Dieldrin	121	122	65 - 135	0	35		
Endrin	125	125	65 - 135	0	35		
Heptachlor	119	119	65 - 135	1	35		
4,4'-DDT	126	127	65 - 135	0	35		
gamma-BHC (Lindane)	121	120	65 - 135	1	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
Tetrachloro-m-xylene	112	102	62 - 123
DCB Decachlorobiphenyl	109	115	56 - 136

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-1983

Method: 8081A
Preparation: 3510C

Lab Sample ID: MB 720-1983/1-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/21/2005 2045
Date Prepared: 11/18/2005 1443

Analysis Batch: 720-2116
Prep Batch: 720-1983
Units: ug/L

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 1000 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	Result	Qual	RL
Aldrin	ND		0.060
Dieldrin	ND		0.060
Endrin aldehyde	ND		0.060
Endrin	ND		0.060
Endrin ketone	ND		0.060
Heptachlor	ND		0.060
Heptachlor epoxide	ND		0.060
4,4'-DDT	ND		0.060
4,4'-DDE	ND		0.060
4,4'-DDD	ND		0.060
Endosulfan I	ND		0.060
Endosulfan II	ND		0.060
alpha-BHC	ND		0.060
beta-BHC	ND		0.060
gamma-BHC (Lindane)	ND		0.060
delta-BHC	ND		0.060
Endosulfan sulfate	ND		0.060
Methoxychlor	ND		0.060
Toxaphene	ND		1.0
Chlordane (technical)	ND		1.0
alpha-Chlordane	ND		0.060
gamma-Chlordane	ND		0.060

Surrogate	% Rec	Acceptance Limits
Tetrachloro-m-xylene	97	62 - 123
DCB Decachlorobiphenyl	99	56 - 136

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1983

Method: 8081A
Preparation: 3510C

LCS Lab Sample ID: LCS 720-1983/2-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/21/2005 0913
Date Prepared: 11/18/2005 1443

Analysis Batch: 720-2116
Prep Batch: 720-1983
Units: ug/L

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 1000 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

LCSD Lab Sample ID: LCSD 720-1983/3-B
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/21/2005 0942
Date Prepared: 11/18/2005 1443

Analysis Batch: 720-2116
Prep Batch: 720-1983
Units: ug/L

Instrument ID: Varian Pest 2
Lab File ID: N/A
Initial Weight/Volume: 1000 mL
Final Weight/Volume: 10 mL
Injection Volume:
Column ID: PRIMARY

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Aldrin	104	94	65 - 135	10	35		
Dieldrin	106	98	65 - 135	8	35		
Endrin	110	100	65 - 135	9	35		
Heptachlor	104	98	65 - 135	6	35		
4,4'-DDT	111	103	65 - 135	8	35		
gamma-BHC (Lindane)	1160	98	65 - 135	169	35		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
Tetrachloro-m-xylene	98	90	62 - 123
DCB Decachlorobiphenyl	103	95	56 - 136

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-1816

Method: 6010B
Preparation: 3010A

Lab Sample ID: MB 720-1816/1-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/17/2005 0841
Date Prepared: 11/16/2005 0856

Analysis Batch: 720-1924
Prep Batch: 720-1816
Units: mg/L

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Silver	ND		0.0050
Arsenic	ND		0.0050
Barium	ND		0.0050
Cadmium	ND		0.0020
Chromium	ND		0.0050
Lead	ND		0.0050
Selenium	ND		0.0050

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1816

Method: 6010B
Preparation: 3010A

LCS Lab Sample ID: LCS 720-1816/2-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/17/2005 0844
Date Prepared: 11/16/2005 0856

Analysis Batch: 720-1924
Prep Batch: 720-1816
Units: mg/L

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1816/3-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/17/2005 0848
Date Prepared: 11/16/2005 0856

Analysis Batch: 720-1924
Prep Batch: 720-1816
Units: mg/L

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 50 mL

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Silver	97	93	80 - 120	4	20		
Arsenic	97	94	80 - 120	4	20		
Barium	99	95	80 - 120	4	20		
Cadmium	98	94	80 - 120	4	20		
Chromium	98	94	80 - 120	4	20		
Lead	98	94	80 - 120	4	20		
Selenium	97	94	80 - 120	4	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1816

Method: 6010B
Preparation: 3010A

MS Lab Sample ID:	720-413-9	Analysis Batch:	720-1924	Instrument ID:	Varian ICP
Client Matrix:	Water	Prep Batch:	720-1816	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	11/17/2005 0918			Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 0856				

MSD Lab Sample ID:	720-413-9	Analysis Batch:	720-1924	Instrument ID:	Varian ICP
Client Matrix:	Water	Prep Batch:	720-1816	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	11/17/2005 0922			Final Weight/Volume:	50 mL
Date Prepared:	11/16/2005 0856				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Silver	105	110	75 - 125	5	25		
Arsenic	106	112	75 - 125	6	25		
Barium	108	115	75 - 125	5	25		
Cadmium	103	109	75 - 125	5	25		
Chromium	105	111	75 - 125	5	25		
Lead	103	108	75 - 125	5	25		
Selenium	105	111	75 - 125	6	25		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-1909

Method: 6010B
Preparation: 3050B

Lab Sample ID: MB 720-1909/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/18/2005 0743
Date Prepared: 11/17/2005 1306

Analysis Batch: 720-1964
Prep Batch: 720-1909
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Silver	ND		1.0
Arsenic	ND		1.0
Barium	ND		1.0
Cadmium	ND		0.50
Chromium	ND		1.0
Lead	ND		1.0
Selenium	ND		2.0

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1909

Method: 6010B
Preparation: 3050B

LCS Lab Sample ID: LCS 720-1909/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/18/2005 0746
Date Prepared: 11/17/2005 1306

Analysis Batch: 720-1964
Prep Batch: 720-1909
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1909/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/18/2005 0750
Date Prepared: 11/17/2005 1306

Analysis Batch: 720-1964
Prep Batch: 720-1909
Units: mg/Kg

Instrument ID: Varian ICP
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Silver	99	94	80 - 120	5	20		
Arsenic	103	97	80 - 120	5	20		
Barium	100	96	80 - 120	5	20		
Cadmium	100	95	80 - 120	5	20		
Chromium	99	94	80 - 120	5	20		
Lead	99	94	80 - 120	5	20		
Selenium	101	96	80 - 120	5	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1909

Method: 6010B
Preparation: 3050B

MS Lab Sample ID:	720-413-14	Analysis Batch:	720-1964	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1909	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/18/2005 0802			Final Weight/Volume:	50 mL
Date Prepared:	11/17/2005 1306				

MSD Lab Sample ID:	720-413-14	Analysis Batch:	720-1964	Instrument ID:	Varian ICP
Client Matrix:	Solid	Prep Batch:	720-1909	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	1.00 g
Date Analyzed:	11/18/2005 0812			Final Weight/Volume:	50 mL
Date Prepared:	11/17/2005 1306				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Silver	82	78	75 - 125	6	20		
Arsenic	81	76	75 - 125	6	20		
Barium	39	38	75 - 125	1	20	N	N
Cadmium	75	72	75 - 125	5	20		N
Chromium	78	75	75 - 125	3	20		
Lead	76	67	75 - 125	7	20		N
Selenium	77	73	75 - 125	6	20		N

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-1754

Method: 7470A
Preparation: 7470A

Lab Sample ID: MB 720-1754/1-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1309
Date Prepared: 11/15/2005 0859

Analysis Batch: 720-1793
Prep Batch: 720-1754
Units: mg/L

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 25 mL
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Mercury	ND		0.00020

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1754

Method: 7470A
Preparation: 7470A

LCS Lab Sample ID: LCS 720-1754/2-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1311
Date Prepared: 11/15/2005 0859

Analysis Batch: 720-1793
Prep Batch: 720-1754
Units: mg/L

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 25 mL
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1754/3-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 11/15/2005 1312
Date Prepared: 11/15/2005 0859

Analysis Batch: 720-1793
Prep Batch: 720-1754
Units: mg/L

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 25 mL
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Mercury	89	89	85 - 115	0	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1754

Method: 7470A
Preparation: 7470A

MS Lab Sample ID:	720-413-9	Analysis Batch:	720-1793	Instrument ID:	FIMS 100
Client Matrix:	Water	Prep Batch:	720-1754	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	25 mL
Date Analyzed:	11/15/2005 1330			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 0859				

MSD Lab Sample ID:	720-413-9	Analysis Batch:	720-1793	Instrument ID:	FIMS 100
Client Matrix:	Water	Prep Batch:	720-1754	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	25 mL
Date Analyzed:	11/15/2005 1331			Final Weight/Volume:	50 mL
Date Prepared:	11/15/2005 0859				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	83	88	85 - 115	7	20	N	

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-1912

Method: 7471A
Preparation: 7471A

Lab Sample ID: MB 720-1912/1-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/18/2005 0702
Date Prepared: 11/17/2005 1318

Analysis Batch: 720-1952
Prep Batch: 720-1912
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

Analyte	Result	Qual	RL
Mercury	ND		0.050

Laboratory Control/
Laboratory Control Duplicate Recovery Report - Batch: 720-1912

Method: 7471A
Preparation: 7471A

LCS Lab Sample ID: LCS 720-1912/2-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/18/2005 0703
Date Prepared: 11/17/2005 1318

Analysis Batch: 720-1952
Prep Batch: 720-1912
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

LCSD Lab Sample ID: LCSD 720-1912/3-A
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/18/2005 0704
Date Prepared: 11/17/2005 1318

Analysis Batch: 720-1952
Prep Batch: 720-1912
Units: mg/Kg

Instrument ID: FIMS 100
Lab File ID: N/A
Initial Weight/Volume: 1 g
Final Weight/Volume: 50 mL

	% Rec.						
Analyte	LCS	LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Mercury	107	106	85 - 115	0	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 720-1912

Method: 7471A
Preparation: 7471A

MS Lab Sample ID:	720-413-14	Analysis Batch: 720-1952	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1912	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.00 g
Date Analyzed:	11/18/2005 0728		Final Weight/Volume: 50 mL
Date Prepared:	11/17/2005 1318		

MSD Lab Sample ID:	720-413-14	Analysis Batch: 720-1952	Instrument ID: FIMS 100
Client Matrix:	Solid	Prep Batch: 720-1912	Lab File ID: N/A
Dilution:	1.0		Initial Weight/Volume: 1.01 g
Date Analyzed:	11/18/2005 0729		Final Weight/Volume: 50 mL
Date Prepared:	11/17/2005 1318		

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Mercury	98	157	85 - 115	33	20		N *

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-1872

Method: 160.3
Preparation: N/A

Lab Sample ID: MB 720-1872/1
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1400
Date Prepared: N/A

Analysis Batch: 720-1872
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Result	Qual	RL
Percent Moisture	ND		0.010

Matrix Duplicate - Batch: 720-1872

Method: 160.3
Preparation: N/A

Lab Sample ID: 720-413-4
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/15/2005 1400
Date Prepared: N/A

Analysis Batch: 720-1872
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Sample Result/Qual	Result	RPD	Limit	Qual
Percent Moisture	14	15			

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-2055

Method: 160.3
Preparation: N/A

Lab Sample ID: MB 720-2055/1
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/18/2005 1340
Date Prepared: N/A

Analysis Batch: 720-2055
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Result	Qual	RL
Percent Moisture	ND		0.010

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-2060

Method: 160.3
Preparation: N/A

Lab Sample ID: MB 720-2060/1
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/18/2005 1400
Date Prepared: N/A

Analysis Batch: 720-2060
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Result	Qual	RL
Percent Moisture	ND		0.010

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: Weston Solutions,Inc

Job Number: 720-413-1

Method Blank - Batch: 720-2403

Method: 160.3
Preparation: N/A

Lab Sample ID: MB 720-2403/1
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/18/2005 1400
Date Prepared: N/A

Analysis Batch: 720-2403
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Result	Qual	RL
Percent Moisture	ND		0.010

Matrix Duplicate - Batch: 720-2403

Method: 160.3
Preparation: N/A

Lab Sample ID: 720-413-15
Client Matrix: Solid
Dilution: 1.0
Date Analyzed: 11/18/2005 1400
Date Prepared: N/A

Analysis Batch: 720-2403
Prep Batch: N/A
Units: %

Instrument ID: No Equipment Assigned
Lab File ID: N/A
Initial Weight/Volume:
Final Weight/Volume:

Analyte	Sample Result/Qual	Result	RPD	Limit	Qual
Percent Moisture	9.6	8.0			

Calculations are performed before rounding to avoid round-off errors in calculated results.

LOGIN SAMPLE RECEIPT CHECK LIST

Client: Weston Solutions,Inc

Job Number: 720-413-1

Login Number: 413

Question	T/F/NA	Comment
Radioactivity either was not measured or, if measured, is at or below background	NA	
The cooler's custody seal, if present.	NA	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	True	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present	True	
Samples do not require splitting or compositing	True	

Forensic Analytical

Final Report

Bulk Asbestos Analysis

(EPA Method 600/R-93-116, Visual Area Estimation)

STL San Francisco
S.Sidhu

1220 Quarry Ln
Pleasanton, CA 94566-4756

Client ID: 2595
Report Number: B078897
Date Received: 11/14/05
Date Analyzed: 11/16/05
Date Printed: 11/16/05
First Reported: 11/16/05

Job ID/Site: 720-413

Date(s) Collected: 11/11/2005

FASI Job ID: 2595-545
Total Samples Submitted: 2
Total Samples Analyzed: 2

Sample ID	Lab Number	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer	Asbestos Type	Percent in Layer
EBHH-50A	10469775						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components:		**Asbestos (ND)**					
Cellulose (Trace)							
EBHH-49A	10469776						
Layer: Brown Soil			**ND**				
Total Composite Values of Fibrous Components:		**Asbestos (ND)**					
Cellulose (Trace)							

James Flores, Laboratory Supervisor, Hayward Laboratory

Note: Limit of Quantification ('LOQ') = 1%. 'Trace' denotes the presence of asbestos below the LOQ. 'ND' = 'None Detected'.

Analytical results and reports are generated by Forensic Analytical at the request of and for the exclusive use of the person or entity (client) named on such report. Results, reports or copies of same will not be released by Forensic Analytical to any third party without prior written request from client. This report applies only to the sample(s) tested. Supporting laboratory documentation is available upon request. This report must not be reproduced except in full, unless approved by Forensic Analytical. The client is solely responsible for the use and interpretation of test results and reports requested from Forensic Analytical. This report must not be used by the client to claim product endorsement by NVLAP or any other agency of the U.S. Government. Forensic Analytical is not able to assess the degree of hazard resulting from materials analyzed. Forensic Analytical reserves the right to dispose of all samples after a period of thirty (30) days, according to all state and federal guidelines, unless otherwise specified. All samples were received in acceptable condition unless otherwise noted.

3777 Depot Road, Suite 409, Hayward, California 94545-2761 Telephone: 510/887-8828 800/827-FASI Fax: 510/887-4218

2595-545

SEVERN TRENT STL Forensic

STL San Francisco Chain of Custody

1220 Quarry Lane • Pleasanton CA 94566-4756
Phone (925) 484-1919 • Fax: (925) 484-1096
Email: sflogin@stl-inc.com

Reference #: 720-413

Date 11-14-05 Page 1 of 1

Report To

Attn: S. Sidu
Company: STL SF
Address:
Phone: Email:
Bill To: Sampled By:
Attn: Phone:

Analysis Request

Sample ID	Date	Time	Matrix	Preserv.	Purgeable Aromatics BTEX EPA 8021	Fuel Trace EPA 8260B	Purgeable Halocarbons (PVOCs) EPA 8021 by 8260B	Volatile Organics (VOCs) EPA 8260B	Semivolatiles (SVOCs) EPA 8270	Oil and Grease (EPA 1664)	Pesticides PCBs	PNAs by	CAM17 Metals (EPA 6010/7470/7471)	Metals	Low Level Metals by EPA 200.8/6020 (ICP-MS)	WET (STLC) TCLP	Hexavalent Chromium 3H (24h Fold time for H2O)	Spec Cond TFS	Anions	Asbestos	Number of Containers
EBHH-50A	11-11	1145	S	–																X	1
EBHH-49A	11-11	1134	S	–																X	1

Project Info.

Project Name:
Project#:
PO#:
Credit Card#:

Sample Receipt

# of Containers: 2
Head Space:
Temp:
Conforms to record:

AT	5 Day	72h	48h	24h	Other

Report: ☐ Routine ☐ Level 3 ☐ Level 4 ☐ FOD ☐ State Tank Fund ECF ☐ Global ID ____
Special Instructions / Comments

720-413

1) Relinquished by: Joan Mullen 1000
Signature Time
Joan Mullen 11-14-05
Printed Name Date
STL SF
Company

2) Relinquished by: [signature] 1104
Signature Time
B Molina 11/14/05
Printed Name Date
STL-SF
Company

3) Relinquished by:
Signature Time
Printed Name Date
Company

1) Received by: [signature] 1000
Signature Time
B Molina 11/14/05
Printed Name Date
STL-SF
Company

2) Received by: Alessa Correa 1104
Signature Time
M. Coobert 11/14/05
Printed Name Date
FASI
Company

3) Received by:
Signature Time
Printed Name Date
Company

STL SF reports 8015M from Cy-Cy (industry norm). Default for 8015B is Cy-Cy