QUALITY ASSURANCE TECHNICAL SUPPORT LABORATORY
PES SCORING EVALUATION REPORT

PES:	SRS3037	**EPA Sample No.:**	GWL06071705	**Report Date:**	07/12/2017
Lab Name:	TestAmerica - Irvine	**Case No.:**	440-185985-1	**Lab Code:**	TAI
Contract:	NA	**Matrix:**	Water	**Lab Sample ID:**	440-185985-1
SDG No.:	NR	**Date Received:**	06/08/2017	**Lab File ID:**	NR
Units:	µg/L	**Date Analyzed:**	06/12/2017	**Dilution Factor:**	5.0
Method:	EPA Method 218.6				
Comments:	Scored by APTIM Personnel	**Project:**	Leviathan Mine – Ground Water Sampling		

CAS No.	Analyte	Laboratory Results		PES Evaluation	
		Concentration	Q		
18540-29-9	Hexavalent Chromium	24		Pass	Within Limits
****	END All Analytes	**********	****	******	**********

NR = Not Reported
NE = Not Evaluated
NA = Not Applicable

Document ID#: 100103-07122017-2

The Quality Assurance Technical Support (QATS) contract is operated by APTIM.
The QATS Program's Quality Management System is certified to the ISO 9001:2008 International Standard.