Report of Analysis

Client Sample ID:	BPOW5-6		
Lab Sample ID:	JC25914-1	Date Sampled:	08/15/16
Matrix:	AQ - Ground Water	Date Received:	08/16/16
Method:	EPA 524.2 REV 4.1	Percent Solids:	n/a
Project:	Northrop Grumman, OU2 Hydro, Bethpage, NY		

	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1 [a]	1B105286.D	1	08/17/16	BK	n/a	n/a	V1B4994
Run #2							

	Purge Volume
Run #1	5.0 ml
Run #2	

VOA OU2 Outpost List

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	5.0	0.91	ug/l	
78-93-3	2-Butanone	ND	5.0	0.57	ug/l	
71-43-2	Benzene	ND	0.50	0.057	ug/l	
75-27-4	Bromodichloromethane	ND	0.50	0.082	ug/l	
75-25-2	Bromoform	ND	0.50	0.046	ug/l	
74-83-9	Bromomethane	ND	0.50	0.077	ug/l	
75-15-0	Carbon disulfide	ND	0.50	0.028	ug/l	
108-90-7	Chlorobenzene	ND	0.50	0.027	ug/l	
75-00-3	Chloroethane	ND	0.50	0.037	ug/l	
67-66-3	Chloroform	ND	0.50	0.031	ug/l	
74-87-3	Chloromethane	ND	0.50	0.044	ug/l	
56-23-5	Carbon tetrachloride	ND	0.50	0.074	ug/l	
75-34-3	1,1-Dichloroethane	ND	0.50	0.039	ug/l	
75-35-4	1,1-Dichloroethylene	ND	0.50	0.054	ug/l	
107-06-2	1,2-Dichloroethane	ND	0.50	0.034	ug/l	
78-87-5	1,2-Dichloropropane	ND	0.50	0.082	ug/l	
124-48-1	Dibromochloromethane	ND	0.50	0.042	ug/l	
156-60-5	trans-1,2-Dichloroethylene	ND	0.50	0.039	ug/l	
156-59-2	cis-1,2-Dichloroethylene	ND	0.50	0.081	ug/l	
10061-02-6	trans-1,3-Dichloropropene	ND	0.50	0.033	ug/l	
10061-01-5	cis-1,3-Dichloropropene	ND	0.50	0.063	ug/l	
100-41-4	Ethylbenzene	ND	0.50	0.033	ug/l	
76-13-1	Freon 113	ND	1.0	0.10	ug/l	
591-78-6	2-Hexanone	ND	2.0	0.084	ug/l	
75-09-2	Methylene chloride	ND	0.50	0.047	ug/l	
108-10-1	4-Methyl-2-pentanone	ND	2.0	0.27	ug/l	
100-42-5	Styrene	ND	0.50	0.028	ug/l	
71-55-6	1,1,1-Trichloroethane	ND	0.50	0.050	ug/l	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	0.035	ug/l	
79-00-5	1,1,2-Trichloroethane	ND	0.50	0.052	ug/l	
127-18-4	Tetrachloroethylene	ND	0.50	0.091	ug/l	
108-88-3	Toluene	ND	0.50	0.044	ug/l	

ND = Not detected MDL = Method Detection Limit J = Indicates an estimated value
RL = Reporting Limit B = Indicates analyte found in associated method blank
E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	BPOW5-6		
Lab Sample ID:	JC25914-1	Date Sampled:	08/15/16
Matrix:	AQ - Ground Water	Date Received:	08/16/16
Method:	EPA 524.2 REV 4.1	Percent Solids:	n/a
Project:	Northrop Grumman, OU2 Hydro, Bethpage, NY		

4.1

4

VOA OU2 Outpost List

CAS No.	Compound	Result	RL	MDL	Units	Q
79-01-6	Trichloroethylene	ND	0.50	0.024	ug/l	
75-01-4	Vinyl chloride	ND	0.50	0.032	ug/l	
	m,p-Xylene	ND	0.50	0.13	ug/l	
95-47-6	o-Xylene	ND	0.50	0.029	ug/l	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
2199-69-1	1,2-Dichlorobenzene-d4	106%		78-114%
460-00-4	4-Bromofluorobenzene	98%		77-115%

CAS No.	Tentatively Identified Compounds	R.T.	Est. Conc.	Units	Q
	Total TIC, Volatile		0	ug/l	

(a) EPA 524.2 is not a certified method for non-potable water samples.

ND = Not detected MDL = Method Detection Limit
RL = Reporting Limit
E = Indicates value exceeds calibration range

J = Indicates an estimated value
B = Indicates analyte found in associated method blank
N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	BPOW5-5		
Lab Sample ID:	JC25914-2	Date Sampled:	08/15/16
Matrix:	AQ - Ground Water	Date Received:	08/16/16
Method:	EPA 524.2 REV 4.1	Percent Solids:	n/a
Project:	Northrop Grumman, OU2 Hydro, Bethpage, NY		

	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1 [a]	1B105287.D	1	08/17/16	BK	n/a	n/a	V1B4994
Run #2							

	Purge Volume
Run #1	5.0 ml
Run #2	

VOA OU2 Outpost List

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	5.0	0.91	ug/l	
78-93-3	2-Butanone	0.81	5.0	0.57	ug/l	J
71-43-2	Benzene	ND	0.50	0.057	ug/l	
75-27-4	Bromodichloromethane	ND	0.50	0.082	ug/l	
75-25-2	Bromoform	ND	0.50	0.046	ug/l	
74-83-9	Bromomethane	ND	0.50	0.077	ug/l	
75-15-0	Carbon disulfide	ND	0.50	0.028	ug/l	
108-90-7	Chlorobenzene	ND	0.50	0.027	ug/l	
75-00-3	Chloroethane	ND	0.50	0.037	ug/l	
67-66-3	Chloroform	ND	0.50	0.031	ug/l	
74-87-3	Chloromethane	ND	0.50	0.044	ug/l	
56-23-5	Carbon tetrachloride	ND	0.50	0.074	ug/l	
75-34-3	1,1-Dichloroethane	ND	0.50	0.039	ug/l	
75-35-4	1,1-Dichloroethylene	ND	0.50	0.054	ug/l	
107-06-2	1,2-Dichloroethane	ND	0.50	0.034	ug/l	
78-87-5	1,2-Dichloropropane	ND	0.50	0.082	ug/l	
124-48-1	Dibromochloromethane	ND	0.50	0.042	ug/l	
156-60-5	trans-1,2-Dichloroethylene	ND	0.50	0.039	ug/l	
156-59-2	cis-1,2-Dichloroethylene	ND	0.50	0.081	ug/l	
10061-02-6	trans-1,3-Dichloropropene	ND	0.50	0.033	ug/l	
10061-01-5	cis-1,3-Dichloropropene	ND	0.50	0.063	ug/l	
100-41-4	Ethylbenzene	ND	0.50	0.033	ug/l	
76-13-1	Freon 113	ND	1.0	0.10	ug/l	
591-78-6	2-Hexanone	ND	2.0	0.084	ug/l	
75-09-2	Methylene chloride	ND	0.50	0.047	ug/l	
108-10-1	4-Methyl-2-pentanone	ND	2.0	0.27	ug/l	
100-42-5	Styrene	ND	0.50	0.028	ug/l	
71-55-6	1,1,1-Trichloroethane	ND	0.50	0.050	ug/l	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	0.035	ug/l	
79-00-5	1,1,2-Trichloroethane	ND	0.50	0.052	ug/l	
127-18-4	Tetrachloroethylene	ND	0.50	0.091	ug/l	
108-88-3	Toluene	ND	0.50	0.044	ug/l	

ND = Not detected MDL = Method Detection Limit
RL = Reporting Limit
E = Indicates value exceeds calibration range

J = Indicates an estimated value
B = Indicates analyte found in associated method blank
N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	BPOW5-5		
Lab Sample ID:	JC25914-2	Date Sampled:	08/15/16
Matrix:	AQ - Ground Water	Date Received:	08/16/16
Method:	EPA 524.2 REV 4.1	Percent Solids:	n/a
Project:	Northrop Grumman, OU2 Hydro, Bethpage, NY		

VOA OU2 Outpost List

CAS No.	Compound	Result	RL	MDL	Units	Q
79-01-6	Trichloroethylene	ND	0.50	0.024	ug/l	
75-01-4	Vinyl chloride	ND	0.50	0.032	ug/l	
	m,p-Xylene	ND	0.50	0.13	ug/l	
95-47-6	o-Xylene	ND	0.50	0.029	ug/l	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
2199-69-1	1,2-Dichlorobenzene-d4	101%		78-114%
460-00-4	4-Bromofluorobenzene	93%		77-115%

CAS No.	Tentatively Identified Compounds	R.T.	Est. Conc.	Units	Q
	Total TIC, Volatile		0	ug/l	

(a) EPA 524.2 is not a certified method for non-potable water samples.

ND = Not detected MDL = Method Detection Limit
RL = Reporting Limit
E = Indicates value exceeds calibration range

J = Indicates an estimated value
B = Indicates analyte found in associated method blank
N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	TB081516MO1		
Lab Sample ID:	JC25914-3	Date Sampled:	08/15/16
Matrix:	AQ - Trip Blank Water	Date Received:	08/16/16
Method:	EPA 524.2 REV 4.1	Percent Solids:	n/a
Project:	Northrop Grumman, OU2 Hydro, Bethpage, NY		

	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1 [a]	1B105288.D	1	08/17/16	BK	n/a	n/a	V1B4994
Run #2							

	Purge Volume
Run #1	5.0 ml
Run #2	

VOA OU2 Outpost List

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	5.0	0.91	ug/l	
78-93-3	2-Butanone	ND	5.0	0.57	ug/l	
71-43-2	Benzene	ND	0.50	0.057	ug/l	
75-27-4	Bromodichloromethane	ND	0.50	0.082	ug/l	
75-25-2	Bromoform	ND	0.50	0.046	ug/l	
74-83-9	Bromomethane	ND	0.50	0.077	ug/l	
75-15-0	Carbon disulfide	ND	0.50	0.028	ug/l	
108-90-7	Chlorobenzene	ND	0.50	0.027	ug/l	
75-00-3	Chloroethane	ND	0.50	0.037	ug/l	
67-66-3	Chloroform	ND	0.50	0.031	ug/l	
74-87-3	Chloromethane	ND	0.50	0.044	ug/l	
56-23-5	Carbon tetrachloride	ND	0.50	0.074	ug/l	
75-34-3	1,1-Dichloroethane	ND	0.50	0.039	ug/l	
75-35-4	1,1-Dichloroethylene	ND	0.50	0.054	ug/l	
107-06-2	1,2-Dichloroethane	ND	0.50	0.034	ug/l	
78-87-5	1,2-Dichloropropane	ND	0.50	0.082	ug/l	
124-48-1	Dibromochloromethane	ND	0.50	0.042	ug/l	
156-60-5	trans-1,2-Dichloroethylene	ND	0.50	0.039	ug/l	
156-59-2	cis-1,2-Dichloroethylene	ND	0.50	0.081	ug/l	
10061-02-6	trans-1,3-Dichloropropene	ND	0.50	0.033	ug/l	
10061-01-5	cis-1,3-Dichloropropene	ND	0.50	0.063	ug/l	
100-41-4	Ethylbenzene	ND	0.50	0.033	ug/l	
76-13-1	Freon 113	ND	1.0	0.10	ug/l	
591-78-6	2-Hexanone	ND	2.0	0.084	ug/l	
75-09-2	Methylene chloride	ND	0.50	0.047	ug/l	
108-10-1	4-Methyl-2-pentanone	ND	2.0	0.27	ug/l	
100-42-5	Styrene	ND	0.50	0.028	ug/l	
71-55-6	1,1,1-Trichloroethane	ND	0.50	0.050	ug/l	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	0.035	ug/l	
79-00-5	1,1,2-Trichloroethane	ND	0.50	0.052	ug/l	
127-18-4	Tetrachloroethylene	ND	0.50	0.091	ug/l	
108-88-3	Toluene	ND	0.50	0.044	ug/l	

ND = Not detected MDL = Method Detection Limit
RL = Reporting Limit
E = Indicates value exceeds calibration range

J = Indicates an estimated value
B = Indicates analyte found in associated method blank
N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	TB081516MO1		
Lab Sample ID:	JC25914-3	Date Sampled:	08/15/16
Matrix:	AQ - Trip Blank Water	Date Received:	08/16/16
Method:	EPA 524.2 REV 4.1	Percent Solids:	n/a
Project:	Northrop Grumman, OU2 Hydro, Bethpage, NY		

4.3

4

VOA OU2 Outpost List

CAS No.	Compound	Result	RL	MDL	Units	Q
79-01-6	Trichloroethylene	ND	0.50	0.024	ug/l	
75-01-4	Vinyl chloride	ND	0.50	0.032	ug/l	
	m,p-Xylene	ND	0.50	0.13	ug/l	
95-47-6	o-Xylene	ND	0.50	0.029	ug/l	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
2199-69-1	1,2-Dichlorobenzene-d4	101%		78-114%
460-00-4	4-Bromofluorobenzene	96%		77-115%

CAS No.	Tentatively Identified Compounds	R.T.	Est. Conc.	Units	Q
	Total TIC, Volatile		0	ug/l	

(a) EPA 524.2 is not a certified method for non-potable water samples.

ND = Not detected MDL = Method Detection Limit
RL = Reporting Limit
E = Indicates value exceeds calibration range

J = Indicates an estimated value
B = Indicates analyte found in associated method blank
N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	BPOW5-4		
Lab Sample ID:	JC26191-1	**Date Sampled:**	08/19/16
Matrix:	AQ - Ground Water	**Date Received:**	08/19/16
Method:	EPA 524.2 REV 4.1	**Percent Solids:**	n/a
Project:	Northrop Grumman, OU2 Hydro, Bethpage, NY		

	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1 [a]	1B105375.D	1	08/24/16	BK	n/a	n/a	V1B4999
Run #2							

	Purge Volume
Run #1	5.0 ml
Run #2	

VOA OU2 Outpost List

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	5.0	0.91	ug/l	
78-93-3	2-Butanone	ND	5.0	0.57	ug/l	
71-43-2	Benzene	ND	0.50	0.057	ug/l	
75-27-4	Bromodichloromethane	ND	0.50	0.082	ug/l	
75-25-2	Bromoform	ND	0.50	0.046	ug/l	
74-83-9	Bromomethane	ND	0.50	0.077	ug/l	
75-15-0	Carbon disulfide	ND	0.50	0.028	ug/l	
108-90-7	Chlorobenzene	ND	0.50	0.027	ug/l	
75-00-3	Chloroethane	ND	0.50	0.037	ug/l	
67-66-3	Chloroform	ND	0.50	0.031	ug/l	
74-87-3	Chloromethane	ND	0.50	0.044	ug/l	
56-23-5	Carbon tetrachloride	ND	0.50	0.074	ug/l	
75-34-3	1,1-Dichloroethane	ND	0.50	0.039	ug/l	
75-35-4	1,1-Dichloroethylene	ND	0.50	0.054	ug/l	
107-06-2	1,2-Dichloroethane	ND	0.50	0.034	ug/l	
78-87-5	1,2-Dichloropropane	ND	0.50	0.082	ug/l	
124-48-1	Dibromochloromethane	ND	0.50	0.042	ug/l	
156-60-5	trans-1,2-Dichloroethylene	ND	0.50	0.039	ug/l	
156-59-2	cis-1,2-Dichloroethylene	ND	0.50	0.081	ug/l	
10061-02-6	trans-1,3-Dichloropropene	ND	0.50	0.033	ug/l	
10061-01-5	cis-1,3-Dichloropropene	ND	0.50	0.063	ug/l	
100-41-4	Ethylbenzene	ND	0.50	0.033	ug/l	
76-13-1	Freon 113	ND	1.0	0.10	ug/l	
591-78-6	2-Hexanone	ND	2.0	0.084	ug/l	
75-09-2	Methylene chloride	ND	0.50	0.047	ug/l	
108-10-1	4-Methyl-2-pentanone	ND	2.0	0.27	ug/l	
100-42-5	Styrene	ND	0.50	0.028	ug/l	
71-55-6	1,1,1-Trichloroethane	ND	0.50	0.050	ug/l	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	0.035	ug/l	
79-00-5	1,1,2-Trichloroethane	ND	0.50	0.052	ug/l	
127-18-4	Tetrachloroethylene	ND	0.50	0.091	ug/l	
108-88-3	Toluene	ND	0.50	0.044	ug/l	

ND = Not detected MDL = Method Detection Limit J = Indicates an estimated value
RL = Reporting Limit B = Indicates analyte found in associated method blank
E = Indicates value exceeds calibration range N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	BPOW5-4		
Lab Sample ID:	JC26191-1	**Date Sampled:**	08/19/16
Matrix:	AQ - Ground Water	**Date Received:**	08/19/16
Method:	EPA 524.2 REV 4.1	**Percent Solids:**	n/a
Project:	Northrop Grumman, OU2 Hydro, Bethpage, NY		

VOA OU2 Outpost List

CAS No.	Compound	Result	RL	MDL	Units	Q
79-01-6	Trichloroethylene	ND	0.50	0.024	ug/l	
75-01-4	Vinyl chloride	ND	0.50	0.032	ug/l	
	m,p-Xylene	ND	0.50	0.13	ug/l	
95-47-6	o-Xylene	ND	0.50	0.029	ug/l	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
2199-69-1	1,2-Dichlorobenzene-d4	102%		78-114%
460-00-4	4-Bromofluorobenzene	101%		77-115%

CAS No.	Tentatively Identified Compounds	R.T.	Est. Conc.	Units	Q
	Total TIC, Volatile		0	ug/l	

(a) EPA 524.2 is not a certified method for non-potable water samples.

ND = Not detected MDL = Method Detection Limit
RL = Reporting Limit
E = Indicates value exceeds calibration range

J = Indicates an estimated value
B = Indicates analyte found in associated method blank
N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	TB081916MD1		
Lab Sample ID:	JC26191-2	**Date Sampled:**	08/19/16
Matrix:	AQ - Trip Blank Water	**Date Received:**	08/19/16
Method:	EPA 524.2 REV 4.1	**Percent Solids:**	n/a
Project:	Northrop Grumman, OU2 Hydro, Bethpage, NY		

	File ID	DF	Analyzed	By	Prep Date	Prep Batch	Analytical Batch
Run #1 [a]	1B105376.D	1	08/24/16	BK	n/a	n/a	V1B4999
Run #2							

	Purge Volume
Run #1	5.0 ml
Run #2	

VOA OU2 Outpost List

CAS No.	Compound	Result	RL	MDL	Units	Q
67-64-1	Acetone	ND	5.0	0.91	ug/l	
78-93-3	2-Butanone	ND	5.0	0.57	ug/l	
71-43-2	Benzene	ND	0.50	0.057	ug/l	
75-27-4	Bromodichloromethane	ND	0.50	0.082	ug/l	
75-25-2	Bromoform	ND	0.50	0.046	ug/l	
74-83-9	Bromomethane	ND	0.50	0.077	ug/l	
75-15-0	Carbon disulfide	ND	0.50	0.028	ug/l	
108-90-7	Chlorobenzene	ND	0.50	0.027	ug/l	
75-00-3	Chloroethane	ND	0.50	0.037	ug/l	
67-66-3	Chloroform	ND	0.50	0.031	ug/l	
74-87-3	Chloromethane	ND	0.50	0.044	ug/l	
56-23-5	Carbon tetrachloride	ND	0.50	0.074	ug/l	
75-34-3	1,1-Dichloroethane	ND	0.50	0.039	ug/l	
75-35-4	1,1-Dichloroethylene	ND	0.50	0.054	ug/l	
107-06-2	1,2-Dichloroethane	ND	0.50	0.034	ug/l	
78-87-5	1,2-Dichloropropane	ND	0.50	0.082	ug/l	
124-48-1	Dibromochloromethane	ND	0.50	0.042	ug/l	
156-60-5	trans-1,2-Dichloroethylene	ND	0.50	0.039	ug/l	
156-59-2	cis-1,2-Dichloroethylene	ND	0.50	0.081	ug/l	
10061-02-6	trans-1,3-Dichloropropene	ND	0.50	0.033	ug/l	
10061-01-5	cis-1,3-Dichloropropene	ND	0.50	0.063	ug/l	
100-41-4	Ethylbenzene	ND	0.50	0.033	ug/l	
76-13-1	Freon 113	ND	1.0	0.10	ug/l	
591-78-6	2-Hexanone	ND	2.0	0.084	ug/l	
75-09-2	Methylene chloride	ND	0.50	0.047	ug/l	
108-10-1	4-Methyl-2-pentanone	ND	2.0	0.27	ug/l	
100-42-5	Styrene	ND	0.50	0.028	ug/l	
71-55-6	1,1,1-Trichloroethane	ND	0.50	0.050	ug/l	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	0.035	ug/l	
79-00-5	1,1,2-Trichloroethane	ND	0.50	0.052	ug/l	
127-18-4	Tetrachloroethylene	ND	0.50	0.091	ug/l	
108-88-3	Toluene	ND	0.50	0.044	ug/l	

ND = Not detected MDL = Method Detection Limit
RL = Reporting Limit
E = Indicates value exceeds calibration range

J = Indicates an estimated value
B = Indicates analyte found in associated method blank
N = Indicates presumptive evidence of a compound

Report of Analysis

Client Sample ID:	TB081916MD1		
Lab Sample ID:	JC26191-2	**Date Sampled:**	08/19/16
Matrix:	AQ - Trip Blank Water	**Date Received:**	08/19/16
Method:	EPA 524.2 REV 4.1	**Percent Solids:**	n/a
Project:	Northrop Grumman, OU2 Hydro, Bethpage, NY		

VOA OU2 Outpost List

CAS No.	Compound	Result	RL	MDL	Units	Q
79-01-6	Trichloroethylene	ND	0.50	0.024	ug/l	
75-01-4	Vinyl chloride	ND	0.50	0.032	ug/l	
	m,p-Xylene	ND	0.50	0.13	ug/l	
95-47-6	o-Xylene	ND	0.50	0.029	ug/l	

CAS No.	Surrogate Recoveries	Run# 1	Run# 2	Limits
2199-69-1	1,2-Dichlorobenzene-d4	101%		78-114%
460-00-4	4-Bromofluorobenzene	100%		77-115%

CAS No.	Tentatively Identified Compounds	R.T.	Est. Conc.	Units	Q
	Total TIC, Volatile		0	ug/l	

(a) EPA 524.2 is not a certified method for non-potable water samples.

ND = Not detected MDL = Method Detection Limit
RL = Reporting Limit
E = Indicates value exceeds calibration range

J = Indicates an estimated value
B = Indicates analyte found in associated method blank
N = Indicates presumptive evidence of a compound