ICF
ESAT Region 3
US Environmental Protection Agency Environmental Science Center
701 Mapes Road Ft. Meade, MD 20755-5350
Phone 410-305-3012

Date: October 5, 2018

To: Brandon McDonald
ESAT Region 3 Project Officer

From: Ex. 4 CBI
Validator

Ex. 4 CBI
Reviewer

Subject: Organic Data Validation (S4VM)
Procino Plating
R35365 C0B27

Overview

This data package consisted of five (5) Field Reagent Blanks (FRBs) and five (5) potable water samples analyzed for perfluorinated alkyl acid compounds by Liquid Chromatography/Mass Spectrometry (LC/MS). This sample set included a field duplicate pair.

Analyses were performed by Test America Sacramento (TAMC). The samples were submitted to the laboratory directly by the sampling contractor. The laboratory indicated analyses were performed according to EPA Method 537 utilizing a quantitative isotope dilution-internal standard technique.

Data were validated according to the National Functional Guidelines for Organic Superfund Methods Data Review and applicable USEPA Region 3 modifications. The validation report has been assigned the Superfund Data Validation Label S4VM (Stage_4_Validation_Manual).

The following validation narrative is an evaluation of laboratory reported data based on the electronic data package received by Region 3 on August 17, 2018.

The laboratory did not provide sufficient data to determine if branched chain isomers of PFOA were included in the reporting of this analyte in field samples. No data were qualified based on this finding.

Summary

A preservation outlier was identified that required the rejection of sample results.

Major Problem

The laboratory reported receipt of sample C0B27 at a pH of 9. Sample pH exceeding 7 may result in low analyte recovery. Reporting Limits (RLs) are unusable and have been qualified "R".

Notes

Analytes detected below RLs are estimated and have been qualified "J".

Accuracy and precision criteria were met by the laboratory in the initial and continuing calibration verification standard analyses associated with the samples in this Sample Data Group (SDG).

The method blank and FRBs associated with the samples in this SDG were free from contamination. No data were qualified based on blank contamination.

Percent recoveries for target analytes in the Laboratory Control Sample (LCS) analysis were within control limits. No field sample data were qualified based on LCS accuracy.

Percent recoveries and Relative Percent Differences (RPDs) for target analytes in the Matrix Spike/Matrix Spike Duplicate (MS/MSD) analyses of sample C0B39 were within control limits. No data were qualified based on MS/MSD precision or accuracy.

Field duplicate pair C0B37/C0B38 were comparable. No data were qualified based on field duplicate precision.

The samples in this SDG were found to be free of residual chlorine at the time of sample preparation by the laboratory. No data were qualified based on this finding.

Samples in this SDG were preserved and reported pH within the optimum range except for sample C0B27, as described in "Major problem".

Manual integrations were performed and identified by the laboratory. A subset of these was evaluated by the reviewer and found to be accurate and consistent. No action was taken by the reviewer based on manual integrations.

Glossary of Organic Data Qualifier Codes

Validation Qualifiers	In order of descending precedence. Only one of these qualifiers may apply to any result.
R	The data are unusable. The sample results are rejected due to serious deficiencies in meeting QC criteria. The analyte may or may not be present in the sample.
UJ	The analyte was analyzed for, but was not detected. The reported quantitation limit is approximate and may be inaccurate or imprecise.
U	The analyte was analyzed for, but was not detected above the level of the reported sample quantitation limit
J	The result is an estimated quantity. The associated numerical value is the approximate concentration of the analyte in the sample.
J+	The result is an estimated quantity, but the result may be biased high.
J-	The result is an estimated quantity, but the result may be biased low.

Additional Qualifiers	Additional qualifiers may be combined with other qualifiers.
N	The analyte has been “tentatively identified” or “presumptively” as present.
B	The result is presumed a blank contaminant. This qualifier is used for drinking water samples only.
C	The target Pesticide or Aroclor analyte identification has been confirmed by Gas Chromatography/Mass Spectrometry (GC/MS). This qualifier may be added to other qualifiers.
X	The target Pesticide or Aroclor analyte identification was not confirmed when GC/MS analysis was performed. This qualifier may be added to other qualifiers.