LEGEND
Technical Services, Inc.
www.legend-group.com

17631 North 25th Avenue • Phoenix, AZ 85023
P (602) 324-6100 • F (602) 324-6101

23 October 2013

John Traux
Yellow Jacket Drilling Service
3922 E University Dr. Suite#1
Phoenix, AZ 85034

RE: New source

Laboratory Work Order No.: 3101699

Legend Technical Services of Arizona, Inc. is pleased to provide the enclosed analytical results for the aforementioned project. These results relate only to the items tested. This cover letter and the accompanying pages represent the full report for these analyses and should only be reproduced in full. Samples for this project were received by the laboratory on 10/18/13 11:48.

The samples were processed in accordance with the Chain of Custody document and the results presented relate only to the samples tested. The Chain of Custody is considered part of this report.

All samples will be retained by LEGEND for 30 days from the date of this report and then discarded unless other arrangements are made.

This entire report was reviewed and approved for release by the undersigned. If you have any questions concerning this report, please feel free to contact me.

Sincerely,
LEGEND TECHNICAL SERVICES OF ARIZONA, INC.

DRAFT REPORT
DATA SUBJECT TO CHANGE
(602) 324-6100

This laboratory report is confidential and is intended for the sole use of LEGEND and it's client.

ED_005149_00023119-00001

ANALYTICAL REPORT FOR SAMPLES

Sample ID	Laboratory ID	Matrix	Type	Date Sampled	Date Received
DRAFT: Hamp Well #3 (Z1,975)	3101699-01	Drinking Water	Grab	10/17/13 15:00	10/18/13 11:48
DRAFT: Travel Blank	3101699-02	DI Water	Grab	10/18/13 00:00	10/18/13 11:48

Sample Condition Upon Receipt:

Temperature: 4.40 C

All samples were received in acceptable condition unless noted otherwise in the case narrative.

Case Narrative:

Holding Times: All holding times were met unless otherwise qualified.

QA/QC Criteria: All analyses met method requirements unless otherwise qualified.

Certifications: AZ(PHX)0004, AZ(TUC)OOO4, AIHA#102982, CDC ELITE Member.

Accreditation is applicable only to the test methods specified on each scope of accreditation held by LEGEND.

Comments: There were no problems encountered during the processing of the samples, unless otherwise noted. All samples were analyzed on a "wet" basis unless designated as "dry weight".

Client notified of the result via email by the micro analyst on 10/20/13 LP

Sample was turbid resulting in apparent color. In order to attempt to record the true color, the sample was centrifuged at 4500 rpm for 10 minutes. This resulted in a pellet at the bottom of the sample. The supernatant was poured off, but the sample remained very turbid. This centrifugation process was run again, resulting in another smaller pellet. The sample remained visibly turbid. The color analysis was run from the supernatant of the second centrifugation; however the result is still “apparent” color due to the turbidity. CL 10/22/13

DRAFT: Hamp Well #3 (Z1,975) (3101699-01) Drinking Water (Grab) Sampled: 10/17/13 15:00 Received: 10/18/13 11:48

Legend Technical Services of Arizona, Inc.

Analyte	Result	PQL	Units	Dilution	Batch	Prepared	Analyzed	Method	Notes
DRAFT: Microbiology									
Total Coliforms, MPN	>23	1.1	MPN/100 mL	1	B3J0611	10/18/13 14:15	10/18/13 14:15	SM 9221B	
DRAFT: Total Metals									
Antimony	<0.0050	0.0050	mg/L	10	B3J0604	10/21/13 16:40	10/22/13 16:43	EPA 200.8	D1
Arsenic	**0.0183**	0.0100	mg/L	10	B3J0604	10/21/13 16:40	10/22/13 16:43	EPA 200.8	D1
Cadmium	**0.0010**	0.0010	mg/L	10	B3J0604	10/21/13 16:40	10/22/13 16:43	EPA 200.8	D1
Lead	**0.0766**	0.0100	mg/L	10	B3J0604	10/21/13 16:40	10/22/13 16:43	EPA 200.8	D1
Selenium	<0.0200	0.0200	mg/L	10	B3J0604	10/21/13 16:40	10/22/13 16:43	EPA 200.8	D1
Silver	**0.0021**	0.0010	mg/L	10	B3J0604	10/21/13 16:40	10/22/13 16:43	EPA 200.8	D1
Thallium	<0.0050	0.0050	mg/L	10	B3J0604	10/21/13 16:40	10/22/13 16:43	EPA 200.8	D1
DRAFT: Inorganic Chemistry									
Color	**600**	1	Color Units	1	B3J0555	10/18/13 15:30	10/18/13 15:30	SM 2120 B	N1
Nitrate as N	<0.10	0.10	mg/L	1	[CALC]	10/18/13 14:21	10/18/13 14:21	Calculation	
Nitrite as N	<0.10	0.10	mg/L	1	B3J0621	10/18/13 14:21	10/18/13 14:21	SM 4500 NO2 B	
Threshold Odor Number	<1	1	T.O.N.	1	B3J0556	10/18/13 14:30	10/18/13 14:30	EPA 140.1	
pH	**9.4**		pH Units	1	B3J0582	10/20/13 14:20	10/20/13 14:20	SM 4500H B	H5
Temperature	**12.7**		°C	1	B3J0582	10/20/13 14:20	10/20/13 14:20	pH Temperature	H5
Total Dissolved Solids	**1580**	1	mg/L	1	B3J0580	10/20/13 13:50	10/20/13 13:50	SM 2540 C	

DRAFT: Total Metals - Quality Control

Legend Technical Services of Arizona, Inc.

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B3J0604 - EPA 200.8										
Blank (B3J0604-BLK1)				*Prepared: 10/21/13 Analyzed: 10/22/13*						
Antimony	<0.0005	0.0005	mg/L							
Arsenic	<0.0010	0.0010	mg/L							
Cadmium	<0.0001	0.0001	mg/L							
Lead	<0.0010	0.0010	mg/L							
Selenium	<0.0020	0.0020	mg/L							
Silver	<0.0001	0.0001	mg/L							
Thallium	<0.0005	0.0005	mg/L							
LCS (B3J0604-BS1)				*Prepared: 10/21/13 Analyzed: 10/22/13*						
Antimony	0.0276	0.0005	mg/L	0.0250		111	85-115			
Arsenic	0.0255	0.0010	mg/L	0.0250		102	85-115			
Cadmium	0.0264	0.0001	mg/L	0.0250		106	85-115			
Lead	0.0279	0.0010	mg/L	0.0250		112	85-115			
Selenium	0.0254	0.0020	mg/L	0.0250		102	85-115			
Silver	0.0271	0.0001	mg/L	0.0250		108	85-115			
Thallium	0.0267	0.0005	mg/L	0.0250		107	85-115			
LCS Dup (B3J0604-BSD1)				*Prepared: 10/21/13 Analyzed: 10/22/13*						
Antimony	0.0271	0.0005	mg/L	0.0250		108	85-115	2	20	
Arsenic	0.0245	0.0010	mg/L	0.0250		98	85-115	4	20	
Cadmium	0.0259	0.0001	mg/L	0.0250		104	85-115	2	20	
Lead	0.0272	0.0010	mg/L	0.0250		109	85-115	2	20	
Selenium	0.0250	0.0020	mg/L	0.0250		100	85-115	2	20	
Silver	0.0263	0.0001	mg/L	0.0250		105	85-115	3	20	
Thallium	0.0268	0.0005	mg/L	0.0250		107	85-115	0.3	20	
Matrix Spike (B3J0604-MS1)		**Source: 3101731-01**		*Prepared: 10/21/13 Analyzed: 10/22/13*						
Antimony	0.0283	0.0005	mg/L	0.0250	0.00003	113	70-130			
Arsenic	0.0317	0.0010	mg/L	0.0250	0.0078	95	70-130			
Cadmium	0.0256	0.0001	mg/L	0.0250	<0.0001	102	70-130			
Lead	0.0266	0.0010	mg/L	0.0250	0.0001	106	70-130			
Selenium	0.0253	0.0020	mg/L	0.0250	0.0019	94	70-130			
Silver	0.0227	0.0001	mg/L	0.0250	<0.0001	91	70-130			
Thallium	0.0246	0.0005	mg/L	0.0250	<0.0005	98	70-130			

DRAFT: Total Metals - Quality Control

Legend Technical Services of Arizona, Inc.

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B3J0604 - EPA 200.8										
Matrix Spike (B3J0604-MS2)		**Source: 3101731-09**		*Prepared: 10/21/13 Analyzed: 10/22/13*						
Antimony	0.0277	0.0005	mg/L	0.0250	0.00003	111	70-130			
Arsenic	0.0313	0.0010	mg/L	0.0250	0.0064	99	70-130			
Cadmium	0.0252	0.0001	mg/L	0.0250	0.00002	101	70-130			
Lead	0.0265	0.0010	mg/L	0.0250	0.00008	106	70-130			
Selenium	0.0246	0.0020	mg/L	0.0250	0.0008	95	70-130			
Silver	0.0236	0.0001	mg/L	0.0250	<0.0001	95	70-130			
Thallium	0.0247	0.0005	mg/L	0.0250	<0.0005	99	70-130			
Matrix Spike Dup (B3J0604-MSD1)		**Source: 3101731-01**		*Prepared: 10/21/13 Analyzed: 10/22/13*						
Antimony	0.0305	0.0005	mg/L	0.0250	0.00003	122	70-130	8	20	
Arsenic	0.0343	0.0010	mg/L	0.0250	0.0078	106	70-130	8	20	
Cadmium	0.0269	0.0001	mg/L	0.0250	<0.0001	108	70-130	5	20	
Lead	0.0284	0.0010	mg/L	0.0250	0.0001	113	70-130	7	20	
Selenium	0.0270	0.0020	mg/L	0.0250	0.0019	100	70-130	6	20	
Silver	0.0256	0.0001	mg/L	0.0250	<0.0001	102	70-130	12	20	
Thallium	0.0263	0.0005	mg/L	0.0250	<0.0005	105	70-130	7	20	
Matrix Spike Dup (B3J0604-MSD2)		**Source: 3101731-09**		*Prepared: 10/21/13 Analyzed: 10/22/13*						
Antimony	0.0283	0.0005	mg/L	0.0250	0.00003	113	70-130	2	20	
Arsenic	0.0313	0.0010	mg/L	0.0250	0.0064	99	70-130	0.1	20	
Cadmium	0.0256	0.0001	mg/L	0.0250	0.00002	102	70-130	2	20	
Lead	0.0268	0.0010	mg/L	0.0250	0.00008	107	70-130	1	20	
Selenium	0.0255	0.0020	mg/L	0.0250	0.0008	99	70-130	3	20	
Silver	0.0239	0.0001	mg/L	0.0250	<0.0001	96	70-130	1	20	
Thallium	0.0246	0.0005	mg/L	0.0250	<0.0005	98	70-130	0.3	20	

DRAFT: Inorganic Chemistry - Quality Control

Legend Technical Services of Arizona, Inc.

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B3J0555 - NO PREP										
Duplicate (B3J0555-DUP1)		**Source: 3101453-01**		*Prepared & Analyzed: 10/18/13*						
Color	30	1	Color Units		30			0	20	
Batch B3J0580 - NO PREP										
Blank (B3J0580-BLK1)				*Prepared & Analyzed: 10/20/13*						
Total Dissolved Solids	<1	1	mg/L							
Duplicate (B3J0580-DUP1)		**Source: 3101453-01**		*Prepared & Analyzed: 10/20/13*						
Total Dissolved Solids	255	1	mg/L		255			0	5	
Batch B3J0582 - NO PREP										
Duplicate (B3J0582-DUP1)		**Source: 3101699-01**		*Prepared & Analyzed: 10/20/13*						
pH	9.5		pH Units		9.4			0.2	10	H5
Batch B3J0621 - NO PREP										
Blank (B3J0621-BLK1)				*Prepared & Analyzed: 10/18/13*						
Nitrite as N	<0.10	0.10	mg/L							
LCS (B3J0621-BS1)				*Prepared & Analyzed: 10/18/13*						
Nitrite as N	0.196	0.10	mg/L	0.200		98	80-120			
LCS Dup (B3J0621-BSD1)				*Prepared & Analyzed: 10/18/13*						
Nitrite as N	0.195	0.10	mg/L	0.200		97	80-120	0.5	20	

DRAFT: Inorganic Chemistry - Quality Control

Legend Technical Services of Arizona, Inc.

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B3J0621 - NO PREP										
Matrix Spike (B3J0621-MS1)		**Source: 3101695-02**		*Prepared & Analyzed: 10/18/13*						
Nitrite as N	0.197	0.10	mg/L	0.200	<0.10	98	80-120			
Matrix Spike Dup (B3J0621-MSD1)		**Source: 3101695-02**		*Prepared & Analyzed: 10/18/13*						
Nitrite as N	0.197	0.10	mg/L	0.200	<0.10	98	80-120	0	20	

Notes and Definitions

N1	See case narrative.
H5	This test is specified to be performed in the field within 15 minutes of sampling; sample was received and analyzed past the regulatory holding time.
D1	Sample required dilution due to matrix.
BLK	Method Blank
LCS/Dup	Laboratory Control Sample/Laboratory Fortified Blank/Duplicate
MS/Dup	Matrix Spike/Duplicate
Dry	Sample results reported on a dry weight basis
RPD	Relative Percent Difference

Laboratory Sample ID: 3101699

CHAIN OF CUSTODY RECORD

LEGEND Technical Services, Inc.
www.legend-group.com

☑ 17631 N. 25th Avenue • Phoenix, AZ 85023 • (602) 324-6100 • Fax (602) 324-6101
☐ 4585 S. Palo Verde Rd, Ste 423 • Tucson, AZ 85706 • (520) 327-1234 • Fax (520) 327-0518

Page 1 of 1

Please Print Clearly

CLIENT INFORMATION

Client Name	Address	City	State	Zip	Phone	Fax Number or Email Address
YELLOW JACKET DRILLING SERVICES	POB 801	GILBERT	AZ	85299-0801	602-453-3252	Vern@yjdrilling.com Sfritz@shomaku.com

Project Name	Project Number/PWS	Contact	P.O. No.	
New Source	AZ 0110-13	Vern		Fax Results ☐ Email Results ☒ QC Report ☐ EDD ☐ Special Detection Limits ☐

SAMPLE TYPE CODES: DW=Drinking Water, WW=Wastewater, SW=Surface Water, G=Groundwater, O=Other, S=Soil/Solid, T=Travel Blank, F=Food, G=Sludge/Biosolids

TURN AROUND TIME: Laboratory Authorization Required for Rush
☒ Standard 15-20 Working Days for all others
☒ Other Rush for Arsenic (Laboratory Authorization Required)

REQUESTED ANALYSES

Client's	Date	Time	Sample Location/POE/DWR	Composite	Grab	Sample Type	Compliance	No. of Containers	pH ✓ (Lab Use Only)	Metals*	Asbestos	Cyanide	Nitrate, Nitrite	Langlier Index*	Total Coliform	Sulfate	Fluoride	Total Hardness	505, 515, 525	531, 547, 548	549, Dioxin	524, 504	Gross Alpha	Radium 226/22	Color, Odor	LAB NO.
Temperature Blank					X	DW		1	✓	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X		-01
Travel Blank (Do Not Open)					X	DI		3														X				-02
Z1,975	10/17/13	15:00	HAMP WELL # 3		X	G		31	✓																	--------

Arsenic Rush = 5 day (turbid)

TO ENSURE COMPLETION OF ANALYSIS, SAMPLES MUST BE RECEIVED AT LEAST 3 HOURS PRIOR TO THE HOLD TIME EXPIRATION

Comments / Special Instructions: *Metals= Sb, As, Ba, Be, Cd, Cr, Cu, Pb, Hg, Ni, Se, Na, Tl **Langlier Index= Ca, Alkalinity, TDS, pH ***Total Hardness= Ca, Mg

If sample source is chlorinated, different sample containers/preservatives must be used for dechlorination. List of tests required accompanies C.C. Drinking water standards

SAMPLE CONDITION UPON RECEIPT (Lab Use)		
No. of Containers	35	
Temperature	4.4 °C	
Custody Seals	Y	N
Seals Intact	Y	N
Preserved	Y	N

	RELINQUISHED BY		SAMPLES RECEIVED BY	
①	Sampler Signature: Sherry Fritz	Date 10/17/13	Signature: Marion E Phillips	Date 10/17/13
	Sampler Printed Name: SHERRY FRITZ	Time 15:40	Printed Name: Marion E Phillips	Time 1540
②	Signature: Marion E Phillips	Date 10/16/13	Signature: [signature]	Date 10/18/13
	Printed Name: Marion Phillips	Time 1146	Printed Name: K McFarlin	Time 1146
③	Signature	Date	Signature	Date
	Printed Name	Time	Printed Name	Time

WHITE-LAB YELLOW-CLIENT

FORM GEN-170 (05/06)

3101699

T = 1,100 to T = 2,000, every 15 minutes.
T = 2,000 to T = 3,000, every 30 minutes.

If the pump is shut off for any reason during the pumping portion of the test, it shall remain shut off for 24 hours and the complete test shall be rerun.

The pump must remain in place until all recovery measurements are completed.

Upon completion of the test, recorded measurements shall be turned over to Project Engineer for analysis.

2.14 WATER QUALITY TESTING

Near the end of the constant-rate test, the Contractor will collect and properly preserve a suite of water samples. The laboratory listed below will provide appropriate sample containers. Analyses for the chemical constituents listed below shall be performed at the Contractor's expense and the results given to the Project Engineer.

Water quality analysis and methods are based on the National Primary Drinking Water Regulations as found in the Code of Federal Regulations (CFR) Title 40 Part 141. A summary of the National Primary Drinking Water Regulations are attached as Appendix E.

The water quality testing laboratory shall be EPA certified. The water quality testing shall be performed by Legend Technical Services, Inc., 17631 North 25th Avenue, Phoenix, AZ 85023, Phone: (602) 324-6100 or approved equal. When contacting Legend Technical Services the Contractor shall request sample containers for a "new source approval" test for groundwater. As the well is a "new source", there is no need to test for disinfectants or disinfection byproducts.

The "new source approval" test from Legend Technical Services shall comply with the most recent version of the National Primary Drinking Water Regulations and shall test for the following general classes of contaminants:

1. Microorganisms
2. Organic Chemicals
3. Inorganic Chemicals
4. Radionuclides

3101699

The Contractor shall also sample and test for the following secondary parameters to determine water chemistry that may lead to customer complaints or operation and maintenance problems:

1. Alkalinity
2. Aluminum
3. Calcium
4. Chloride
5. ~~Color~~
6. Copper
7. Corrosivity
8. Flouride
9. Hardness as $CaCO_3$
10. Iron
11. Magnesium
12. Manganese
13. Nickel
14. ~~Odor~~
15. pH
16. Silver
17. Sodium
18. Sulfate
19. Total Dissolved Solids (TDS)
20. Total Organic Carbon (TOC)
21. Zinc
22. Silica

+ Potassium
+ bicarbonate

Potassium?

Sample analysis (both acquisition and payment) shall be the sole responsibility of the Contractor. The Project Engineer shall be present during water sampling and shall initial the custody seals of the sample containers.

O:\Sanitation Facilities Const\Eastern AZ\Hopi\E37 Hopi Arsenic Mitigation Project\Contracts\Agreements\CA 11-1201 - Project Manual\R1 - Originals\13 Well2-Specs.doc

Legend Technical Services of Arizona, Inc.

SUBCONTRACT ORDER

3101699

PO Number: 3101699
Sampled By: Vern
Compliance?: Yes No

SENDING LABORATORY:

Legend Technical Services of Arizona, Inc.
17631 North 25th Avenue
Phoenix, AZ 85023
Phone: 602-324-6100
Fax: 602-324-6101
Project Manager: Cory Lund

RECEIVING LABORATORY:

Radiation Safety Engineering
3245 N. Washington
Chandler, AZ 85225
Phone :(480) 897-9459
Fax: (480) 892-5446

Analysis	Due	Expires	Sample Type circle one	Sample Comments
Sample ID: 3101699-01	**Drinking W.**	**Sampled: 10/17/13 15:00**	**Grab Composite**	**N/A**
Radium 228	10/29/13 09:00	04/15/14 15:00		
Radium 226	10/29/13 09:00	04/15/14 15:00		
Gross Alpha, Measured	10/29/13 09:00	04/15/14 15:00		
Gross Alpha, Deviation	10/29/13 09:00	04/15/14 15:00		

Containers Supplied:
06_1000mL Plastic pH <2 w/ HNO3 (C)
06_1000mL Plastic pH <2 w/ HNO3 (D)
06_1000mL Plastic pH <2 w/ HNO3 (E)
06_1000mL Plastic pH <2 w/ HNO3 (F)

Released By (Print & Sign)	Date	Time	Received By (Print & Sign)	Date	Time
Lisa Parrish x L Parrish	10/19/13	13:34	Cecil Crawford Cecil Crawford	10-21-13	13:34
Cecil Crawford Cecil Crawford	10-21-13	14:41	[illegible]	10-21-13	14:41

Page 1 of 1

Legend Technical Services of Arizona, Inc. **PO Number:** 3101699

SUBCONTRACT ORDER
3101699

SENDING LABORATORY:

Legend Technical Services of Arizona, Inc.
17631 North 25th Avenue
Phoenix, AZ 85023
Phone: 602-324-6100
Fax: 602-324-6101
Project Manager: Cory Lund

RECEIVING LABORATORY:

Underwriter's Laboratories
110 S. Hill Street
South Bend, IN 46617
Phone :(800) 332-4345
Fax: (574) 233-8207

Analysis	Due	Expires	Sample Type circle one	Sample Comments
Sample ID: 3101699-01	**Drinking W:**	**Sampled:10/17/13 15:00**	**Grab Composite N/A**	
549-Subcontract	10/29/13 09:00	10/24/13 15:00		
548-Subcontract	10/29/13 09:00	10/24/13 15:00		
525-Subcontract	10/29/13 09:00	10/31/13 15:00		
515-Subcontract	10/29/13 09:00	10/31/13 15:00		
505-Subcontract	10/29/13 09:00	10/31/13 15:00		
504-Subcontract	10/29/13 09:00	10/31/13 15:00		

Containers Supplied:
11_125mL Amber Glass Bottle w/ Sodium Sulfite (H)
11_125mL Amber Glass Bottle w/ Sodium Sulfite (I)
20_500mL Amber Plastic w/ S.T. & H2SO4 (J)
07_1000mL Amber Glass pH <2 w/ SS & HCl (K)
07_1000mL Amber Glass pH <2 w/ SS & HCl (L)
10_40mL Sodium Thiosulfate Clear Vial Cool to 4° C (M)
10_40mL Sodium Thiosulfate Clear Vial Cool to 4° C (N)
10_40mL Sodium Thiosulfate Clear Vial Cool to 4° C (O)
10_40mL Sodium Thiosulfate Clear Vial Cool to 4° C (P)
10_40mL Sodium Thiosulfate Clear Vial Cool to 4° C (Q)
10_40mL Sodium Thiosulfate Clear Vial Cool to 4° C (R)

Released By (Print & Sign)	Date	Time	Received By (Print & Sign)	Date	Time
Lisa Parrish X [signature]	10/18/13	1680	Fed Ex X	10/18/13	
Fed Ex X					

Page 1 of 1

Legend Technical Services of Arizona, Inc. **PO Number:** 3101699

SUBCONTRACT ORDER
3101699

SENDING LABORATORY:
Legend Technical Services of Arizona, Inc.
17631 North 25th Avenue
Phoenix, AZ 85023
Phone: 602-324-6100
Fax: 602-324-6101
Project Manager: Cory Lund

RECEIVING LABORATORY:
Fiberquant Analytical Services
5025 S. 33rd Street
Phoenix, AZ 85040
Phone :(602) 276-6139
Fax: (602) 276-4558

Analysis	Due	Expires	Sample Type circle one	Sample Comments
Sample ID: 3101699-01	**Drinking W.**	**Sampled: 10/17/13 15:00**	**Grab Composite**	**N/A**
Asbestos	10/29/13 09:00	10/19/13 15:00		

Containers Supplied:
03_1000mL Plastic Cool to 4° C (T)

Lisa Parrish	X L Parrish	10/18/13	13:25	Cecil Crawford	X Cecil Crawford	10.18.13	13:25
Released By (Print & Sign)		Date	Time	Received By (Print & Sign)		Date	Time
Cecil Crawford	X Cecil Crawford	10/18/13	14:00	[signature]		10/18/13	1400
Released By (Print & Sign)		Date	Time	Received By (Print & Sign)		Date	Time

Review of Analysis Request (Initials) [initials]

Page 1 of 1