Attachment 5

Laboratory reports for NPDES monitoring

Laboratory Quality Control Reports

LTS ENVIRONMENTAL, INC.

Beta Offshore
111 W. Ocean Blvd., Suite 1240
Long Beach, Ca 90802

August 5, 2021

Attn: Diana Lang

Quarterly NPDES chlorine residuals on the non-contact cooling water outlet were as follows:

Sample Date / Time	**Location**	**Total Chlorine Residual** *(EPA Method 330.5)*
	Platform Ellen	End of Pipe
July 29, 2021 @ 0715 hrs	Non-Contact Cooling Water Outlet West Seawater Pump	0.09 mg/l
LTS Meter S/N: 041200088375		Method Blank < 0.05 mg/l (MDL)

Technician: Konrad Pala

S.G. Lawry
Environmental Specialist / LTS

704 Adirondack Ave. • Ventura, Ca 93003 • (805) 798-1213 • (805) 798-1211

LTS ENVIRONMENTAL, INC.

Beta Offshore
111 West Ocean Blvd.
Suite 1240
Long Beach, Ca 90802

September 8, 2021

Attn: Diana Lang

As part of the Annual Marine Sanitation Device (MSD) Inspection, and to ensure proper operation of the device, LTS Environmental performed an EPA-approved chlorine residual on the effluent. Results of this test are as follows:

Sample Date / Time	**Location**	**Total Chlorine Residual** *(EPA Method 330.5)*
August 27, 2021 @ 0730 hrs	Platform Ellen Sewage Effluent Omnipure 12MX	3.0 mg/l
LTS Meter S/N: 041200088375		Method Blank < 0.05 mg/l (MDL)

LTS Technician: Konrad Pala

S.G. Lawry
Environmental Specialist / LTS

704 Adirondack Ave. • Ventura, Ca 93003 • (805) 798-1213 • (805) 798-1211

LTS ENVIRONMENTAL, INC.

October 5, 2020

Quality Control

As part of the annual in-house quality control chlorine meter check and to ensure proper operation of the meters, LTS Environmental Inc. performed a total residual chlorine test with a known value obtained from Sigma-Aldrich Corporation. Results of this test are as follows:

Test Date	**Total Residual Chlorine**
October 1, 2020	*(EPA Method 330.5)*
LTS meter (SN 041200088375)	**1.87 mg/l**

Acceptance Limits	0.902 mg/l – 2.38 mg/l
Certified Value	1.64 mg/l ± 0.0574

Sigma-Aldrich Corporation test sample (QC1065-2ml. Lot LRAB5132)

LTS Lead Technician: Konrad Pala

Method Blank < 0.05 mg/l

S.G. Lawry
Environmental Specialist
President, LTS Environmental, Inc.

704 Adirondack Ave. • Ventura, Ca 93003 • (805) 798-1213 • (805) 798-1211