From:	Ellis Koch
To:	Eric J. Laning
Cc:	Wiedemer, Ashley
Subject:	Glen Isle: ASAP request- Demo Radiological Monitoring Plan Comments
Date:	Friday, May 09, 2014 11:13:48 AM

Eric, the EPA has provided comments (below) on the RPP. Please respond to them ASAP.

Let me know if any present a problem.

---------- Forwarded message ----------
From: **Wiedemer, Ashley** <Wiedemer.Ashley@epa.gov>
Date: Fri, May 9, 2014 at 11:09 AM
Subject: Demo Radiological Monitoring Plan Comments
To: Ellis Koch <ellis.koch@gmail.com>

EPA has reviewed the RPP and have the below comments. I realize the work has already begun, but the contractor should demonstrate the sampling and measurement methods actually are capable of detecting if a problem did exist.

1. Section 5 air monitoring: Sample collection is appropriate, however, It should be demonstrated in the plan that the stated counting method (44-9 probe) is capable of detecting airborne levels of concern for the primary contaminant (Th-232). At a minimum, the minimum detection limit of the counting method should be presented along with applicable standard for comparison.

2. Section 6: Action limits should be specified for each type of survey. The minimum detection limit of the counting method should be presented along with applicable standard for comparison.

Ashley Wiedemer
Remedial Project Manager
US Environmental Protection Agency Region 2
290 Broadway, 20th Floor
New York, NY 10007
Office: (212) 637-4263
Fax: (212) 637-3966

--
Ellis Koch
Consulting Director
Posillico Consulting
1750 New Highway
Farmingdale, NY 11735

O: 631-390-5755
C: 516-983-7333
fax: 516-740-0071