RECORD

NAVAL BASE SAN DIEGO
FORCE PROTECTION

INTOXIMETER EC/IR II

CALIBRATION/ACCURACY LOG

OPENED: 02 NOV 2015

CLOSED: _____

2

TEST DATE	TEST TIME	TEST TYPE	TECH NAME	TECH INFO	STANDARD LOT NUMBER	EXPECTED VALUE	RESULTS
02 NOV 15	1538	CAL	(b) (6) [redacted]	NEW INTOX PUT IN SERVICE, SERIAL # 012804 (b) (6) [redacted]	AG5Ø6602	.080	.079/PASS
10 NOV 15	0800	ACC			AG5Ø65Ø2	.Ø8Ø	.Ø8Ø/PASS
26 JAN 16	0800	ACC			AG5Ø65Ø2	.Ø8Ø	.Ø8Ø/PASS
02 FEB 16	0800	ACC			AG5Ø65Ø2	.Ø8Ø	.Ø8Ø/PASS
09 FEB 16	0800	ACC			AG5Ø65Ø2	Ø8Ø	.Ø79/PASS
16 FEB 16	0800	ACC			AG5Ø65Ø2	.Ø8Ø	.Ø78/PASS
23 FEB 16	0800	ACC			AG5Ø65Ø2	.Ø8Ø	.Ø79/PASS
08 MAR 16	0800	ACC			AG5Ø65Ø2	.Ø8Ø	.Ø79/PASS
15 MAR 16	Ø8ØØ	ACC			AG5Ø65Ø2	.Ø8Ø	.Ø79/PASS
22 MAR 16	Ø8ØØ	ACC			AG5Ø65Ø2	.Ø8Ø	.Ø79/PASS
29 MAR 16	Ø8ØØ	ACC			AG5Ø65Ø2	.Ø8Ø	.078/PASS
05 APR 16	Ø8ØØ	ACC			G5Ø65Ø2	.Ø8Ø	.079/PASS
12 APR 16	Ø8ØØ	ACC			G5Ø65Ø2	.Ø8Ø	.079/PASS
19 APR 16	Ø8ØØ	ACC			G5Ø65Ø2	.Ø8Ø	.078/PASS
26 APR 16	Ø8ØØ	ACC			G5Ø65Ø2	.080	.079/PASS
03 MAY 16	Ø8ØØ	ACC			G5Ø65Ø2	.Ø8Ø	.078/PASS
10 MAY 16	Ø8ØØ	ACC			G5Ø65Ø2	.080	.079/PASS
17 MAY 16	Ø8ØØ	ACC			AG5Ø65Ø2	.Ø8Ø	.078/PASS
24 MAY 16	Ø8ØØ	ACC			AG5Ø65Ø2	.Ø8Ø	.079/PASS
31 MAY 16	Ø8ØØ	ACC			AG5Ø65Ø2	.080	.078/PASS
7 JUN 16	0800	ACC			AG5Ø65Ø2	.080	.078/PASS
14 JUN 16	Ø8ØØ	ACC			AG5Ø65Ø2	.080	.079/PASS
21 JUN 16	Ø8ØØ	ACC			AG5Ø65Ø2	.Ø8Ø	.079/PASS
28 JUN 16	Ø8ØØ	ACC			G5Ø65Ø2	.Ø8Ø	.Ø79/PASS
05 JUL 16	Ø8ØØ	ACC			G5Ø65Ø2	.080	.Ø79/PASS
12 JUL 16	Ø8ØØ	ACC			G5Ø65Ø2	.Ø8Ø	.Ø79/PASS
19 JUL 16	Ø8ØØ	ACC			G5Ø65Ø2	.080	.Ø8Ø/PASS

26 JUL 16	0800	ACC	(b) (6) [redacted]		(b) (6) [redacted]			
02AUG16	0800	ACC				5065002	.080	.080/ PASS
09AUG16	0800	ACC				5065002	.080	.079/ PASS
16AUG16	0800	ACC				5065002	.080	.079 / PASS
23AUG16	0800	ACC				506502	.080	.080/ PASS
30AUG16	0800	ACC				506502	.080	.079/ PASS
06SEP16	0800	ACC				5506502	.080	.079/ PASS
13SEP16	0800	ACC				5506502	.080	.080/ PASS
20SEP16	0800	ACC				5065002	.080	.080/ PASS
02OCT16	0800	ACC				5065002	.080	.080/ PASS
25OCT16	0800	ACC				5065002	.080	.080/ PASS
01NOV16	0800	ACC				5065002	.080	.080/ PASS
08NOV16	0800	ACC				5065002	.080	.080/ PASS

TEST DATE	TEST TIME	TEST TYPE	TECH NAME	TECH INITIALS	STANDARD LOT NUMBER	EXPECTED VALUE	RESULTS	RE
15 NOV 16	0800	ACC	(b) (6) [redacted]	(b) (6) [redacted]	G506502	.080	.080/PASS	
22 NOV 16	0800	ACC			G506502	.080	.080/PASS	
29 NOV 16	0800	ACC			G506502	.080	.080/PASS	
13 DEC 16	0800	ACC			G506502	.080	.080/PASS	
20 DEC 16	0800	ACC			G506502	.080	.080/PASS	
27 DEC 16	0800	ACC			G506502	.080	.080/PASS	
3 JAN 17	0800	ACC			G506502	.080	.080/PASS	
10 JAN 17	0800	ACC			G506502	.080	.080/PASS	
17 JAN 17	0800	ACC			G506502	.080	.080/PASS	
24 JAN 17	0800	ACC			G506502	.080	.080/PASS	
31 JAN 17	0800	ACC			G506502	.080	.080/PASS	
7 FEB 17	0800	ACC			G506502	.080	.080/PASS	
14 FEB 17	0800	ACC			G506502	.080	.080/PASS	
21 FEB 17	0800	ACC			G506502	.080	.080/PASS	
28 FEB 17	0800	ACC			G506502	.080	.080/PASS	
25 MAY 2017	NEW MACHINE, SERIAL # 012792 PUT INTO SERVICE							
25 MAY 2017	0954	CALIBRATION	(b) (6) [redacted]	(b) (6) [redacted]	AG11202	.079	.079/PASS	
25 MAY 2017	1033	ACC			AG11202	.082	.078/PASS	
30 MAY 2017	0800	ACC			AG11202	.082	.079/PASS	