TestAmerica
ANALYTICAL TESTING CORPORATION

LABORATORY REPORT

Prepared For: Burns & McDonnell, San Francisco
393 E. Grand Avenue, Suite J
San Francisco, CA 94080
Attention: Gary Messerotes

Project: Weber Waters
40641

Sampled: 10/17/06
Received: 10/17/06
Issued: 11/01/06 14:35

NELAP #01108CA California ELAP#1197 CSDLAC #10256

The results listed within this Laboratory Report pertain only to the samples tested in the laboratory. The analyses contained in this report were performed in accordance with the applicable certifications as noted. All soil samples are reported on a wet weight basis unless otherwise noted in the report. This Laboratory Report is confidential and is intended for the sole use of TestAmerica and its client. This report shall not be reproduced, except in full, without written permission from TestAmerica. The Chain(s) of Custody, 2 pages, are included and are an integral part of this report.
This entire report was reviewed and approved for release.

SAMPLE CROSS REFERENCE

SUBCONTRACTED: Refer to the last page for specific subcontract laboratory information included in this report.

LABORATORY ID	CLIENT ID	MATRIX
IPJ1647-01	SW-1	Water
IPJ1647-02	SW-3	Water
IPJ1647-03	SW-2	Water
IPJ1647-04	Trip Blank	Water

Reviewed By:

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco
393 E. Grand Avenue, Suite J
San Francisco, CA 94080
Attention: Gary Messerotes

Project ID: Weber Waters
40641
Report Number: IPJ1647

Sampled: 10/17/06
Received: 10/17/06

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-01 (SW-1 - Water)								
Reporting Units: ug/l								
Acetone	EPA 8260B	6J23005	10	ND	1	10/23/2006	10/23/2006	
Benzene	EPA 8260B	6J23005	0.50	ND	1	10/23/2006	10/23/2006	
Bromobenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Bromochloromethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Bromodichloromethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Bromoform	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Bromomethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
n-Butylbenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
sec-Butylbenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
tert-Butylbenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Carbon tetrachloride	EPA 8260B	6J23005	0.50	ND	1	10/23/2006	10/23/2006	
Chlorobenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Chloroethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Chloroform	EPA 8260B	6J23005	1.0	**1.6**	1	10/23/2006	10/23/2006	
Chloromethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
2-Chlorotoluene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
4-Chlorotoluene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Dibromochloromethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,2-Dibromo-3-chloropropane	EPA 8260B	6J23005	5.0	ND	1	10/23/2006	10/23/2006	
1,2-Dibromoethane (EDB)	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Dibromomethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,2-Dichlorobenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,3-Dichlorobenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,4-Dichlorobenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Dichlorodifluoromethane	EPA 8260B	6J23005	5.0	ND	1	10/23/2006	10/23/2006	
1,1-Dichloroethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,2-Dichloroethane	EPA 8260B	6J23005	0.50	ND	1	10/23/2006	10/23/2006	
1,1-Dichloroethene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
cis-1,2-Dichloroethene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
trans-1,2-Dichloroethene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,2-Dichloropropane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,3-Dichloropropane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
2,2-Dichloropropane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,1-Dichloropropene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
cis-1,3-Dichloropropene	EPA 8260B	6J23005	0.50	ND	1	10/23/2006	10/23/2006	
trans-1,3-Dichloropropene	EPA 8260B	6J23005	0.50	ND	1	10/23/2006	10/23/2006	
Ethylbenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Hexachlorobutadiene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Isopropylbenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
p-Isopropyltoluene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Methylene chloride	EPA 8260B	6J23005	5.0	ND	1	10/23/2006	10/23/2006	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-01 (SW-1 - Water) - cont.								
Reporting Units: ug/l								
Naphthalene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
n-Propylbenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Styrene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,1,1,2-Tetrachloroethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,1,2,2-Tetrachloroethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Tetrachloroethene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Toluene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,2,3-Trichlorobenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,2,4-Trichlorobenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,1,1-Trichloroethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,1,2-Trichloroethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Trichloroethene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Trichlorofluoromethane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Trichlorotrifluoroethane (Freon 113)	EPA 8260B	6J23005	5.0	ND	1	10/23/2006	10/23/2006	
1,2,3-Trichloropropane	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,2,4-Trimethylbenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
1,3,5-Trimethylbenzene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Vinyl chloride	EPA 8260B	6J23005	0.50	ND	1	10/23/2006	10/23/2006	
o-Xylene	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
m,p-Xylenes	EPA 8260B	6J23005	1.0	ND	1	10/23/2006	10/23/2006	
Surrogate: Dibromofluoromethane (80-120%)				*99 %*				
Surrogate: Toluene-d8 (80-120%)				*110 %*				
Surrogate: 4-Bromofluorobenzene (80-120%)				*103 %*				

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-02 (SW-3 - Water)								
Reporting Units: ug/l								
Acetone	EPA 8260B	6J22009	50	ND	5	10/22/2006	10/22/2006	
Benzene	EPA 8260B	6J22009	2.5	ND	5	10/22/2006	10/22/2006	
Bromobenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Bromochloromethane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Bromodichloromethane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Bromoform	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Bromomethane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
n-Butylbenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
sec-Butylbenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
tert-Butylbenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Carbon tetrachloride	EPA 8260B	6J22009	2.5	ND	5	10/22/2006	10/22/2006	
Chlorobenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Chloroethane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Chloroform	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Chloromethane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
2-Chlorotoluene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
4-Chlorotoluene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Dibromochloromethane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,2-Dibromo-3-chloropropane	EPA 8260B	6J22009	25	ND	5	10/22/2006	10/22/2006	
1,2-Dibromoethane (EDB)	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Dibromomethane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,2-Dichlorobenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,3-Dichlorobenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,4-Dichlorobenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Dichlorodifluoromethane	EPA 8260B	6J22009	25	ND	5	10/22/2006	10/22/2006	
1,1-Dichloroethane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,2-Dichloroethane	EPA 8260B	6J22009	2.5	ND	5	10/22/2006	10/22/2006	
1,1-Dichloroethene	EPA 8260B	6J22009	5.0	**48**	5	10/22/2006	10/22/2006	
cis-1,2-Dichloroethene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
trans-1,2-Dichloroethene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,2-Dichloropropane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,3-Dichloropropane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
2,2-Dichloropropane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,1-Dichloropropene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
cis-1,3-Dichloropropene	EPA 8260B	6J22009	2.5	ND	5	10/22/2006	10/22/2006	
trans-1,3-Dichloropropene	EPA 8260B	6J22009	2.5	ND	5	10/22/2006	10/22/2006	
Ethylbenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Hexachlorobutadiene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Isopropylbenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
p-Isopropyltoluene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Methylene chloride	EPA 8260B	6J22009	25	ND	5	10/22/2006	10/22/2006	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-02 (SW-3 - Water) - cont.								
Reporting Units: ug/l								
Naphthalene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
n-Propylbenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Styrene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,1,1,2-Tetrachloroethane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,1,2,2-Tetrachloroethane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Tetrachloroethene	EPA 8260B	6J22009	5.0	**210**	5	10/22/2006	10/22/2006	
Toluene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,2,3-Trichlorobenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,2,4-Trichlorobenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,1,1-Trichloroethane	EPA 8260B	6J22009	5.0	**7.8**	5	10/22/2006	10/22/2006	
1,1,2-Trichloroethane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Trichloroethene	EPA 8260B	6J22009	5.0	**130**	5	10/22/2006	10/22/2006	
Trichlorofluoromethane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Trichlorotrifluoroethane (Freon 113)	EPA 8260B	6J22009	25	ND	5	10/22/2006	10/22/2006	
1,2,3-Trichloropropane	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,2,4-Trimethylbenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
1,3,5-Trimethylbenzene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Vinyl chloride	EPA 8260B	6J22009	2.5	ND	5	10/22/2006	10/22/2006	
o-Xylene	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
m,p-Xylenes	EPA 8260B	6J22009	5.0	ND	5	10/22/2006	10/22/2006	
Surrogate: Dibromofluoromethane (80-120%)				*115 %*				
Surrogate: Toluene-d8 (80-120%)				*102 %*				
Surrogate: 4-Bromofluorobenzene (80-120%)				*95 %*				

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco
393 E. Grand Avenue, Suite J
San Francisco, CA 94080
Attention: Gary Messerotes

Project ID: Weber Waters
40641
Report Number: IPJ1647

Sampled: 10/17/06
Received: 10/17/06

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-03 (SW-2 - Water)								
Reporting Units: ug/l								
Acetone	EPA 8260B	6J22009	10	ND	1	10/22/2006	10/22/2006	
Benzene	EPA 8260B	6J22009	0.50	ND	1	10/22/2006	10/22/2006	
Bromobenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Bromochloromethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Bromodichloromethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Bromoform	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Bromomethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
n-Butylbenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
sec-Butylbenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
tert-Butylbenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Carbon tetrachloride	EPA 8260B	6J22009	0.50	ND	1	10/22/2006	10/22/2006	
Chlorobenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Chloroethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Chloroform	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Chloromethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
2-Chlorotoluene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
4-Chlorotoluene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Dibromochloromethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,2-Dibromo-3-chloropropane	EPA 8260B	6J22009	5.0	ND	1	10/22/2006	10/22/2006	
1,2-Dibromoethane (EDB)	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Dibromomethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,2-Dichlorobenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,3-Dichlorobenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,4-Dichlorobenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Dichlorodifluoromethane	EPA 8260B	6J22009	5.0	ND	1	10/22/2006	10/22/2006	
1,1-Dichloroethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,2-Dichloroethane	EPA 8260B	6J22009	0.50	ND	1	10/22/2006	10/22/2006	
1,1-Dichloroethene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
cis-1,2-Dichloroethene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
trans-1,2-Dichloroethene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,2-Dichloropropane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,3-Dichloropropane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
2,2-Dichloropropane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,1-Dichloropropene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
cis-1,3-Dichloropropene	EPA 8260B	6J22009	0.50	ND	1	10/22/2006	10/22/2006	
trans-1,3-Dichloropropene	EPA 8260B	6J22009	0.50	ND	1	10/22/2006	10/22/2006	
Ethylbenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Hexachlorobutadiene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Isopropylbenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
p-Isopropyltoluene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Methylene chloride	EPA 8260B	6J22009	5.0	ND	1	10/22/2006	10/22/2006	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-03 (SW-2 - Water) - cont.								
Reporting Units: ug/l								
Naphthalene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
n-Propylbenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Styrene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,1,1,2-Tetrachloroethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,1,2,2-Tetrachloroethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Tetrachloroethene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Toluene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,2,3-Trichlorobenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,2,4-Trichlorobenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,1,1-Trichloroethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,1,2-Trichloroethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Trichloroethene	EPA 8260B	6J22009	1.0	**6.6**	1	10/22/2006	10/22/2006	
Trichlorofluoromethane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Trichlorotrifluoroethane (Freon 113)	EPA 8260B	6J22009	5.0	ND	1	10/22/2006	10/22/2006	
1,2,3-Trichloropropane	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,2,4-Trimethylbenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
1,3,5-Trimethylbenzene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Vinyl chloride	EPA 8260B	6J22009	0.50	ND	1	10/22/2006	10/22/2006	
o-Xylene	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
m,p-Xylenes	EPA 8260B	6J22009	1.0	ND	1	10/22/2006	10/22/2006	
Surrogate: Dibromofluoromethane (80-120%)				*116 %*				
Surrogate: Toluene-d8 (80-120%)				*101 %*				
Surrogate: 4-Bromofluorobenzene (80-120%)				*98 %*				

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco
393 E. Grand Avenue, Suite J
San Francisco, CA 94080
Attention: Gary Messerotes

Project ID: Weber Waters
40641
Report Number: IPJ1647

Sampled: 10/17/06
Received: 10/17/06

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-04 (Trip Blank - Water)								
Reporting Units: ug/l								
Acetone	EPA 8260B	6J26012	10	ND	1	10/26/2006	10/26/2006	C
Benzene	EPA 8260B	6J26012	0.50	ND	1	10/26/2006	10/26/2006	
Bromobenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Bromochloromethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Bromodichloromethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Bromoform	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Bromomethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
n-Butylbenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
sec-Butylbenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
tert-Butylbenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Carbon tetrachloride	EPA 8260B	6J26012	0.50	ND	1	10/26/2006	10/26/2006	
Chlorobenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Chloroethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Chloroform	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Chloromethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
2-Chlorotoluene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
4-Chlorotoluene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Dibromochloromethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,2-Dibromo-3-chloropropane	EPA 8260B	6J26012	5.0	ND	1	10/26/2006	10/26/2006	
1,2-Dibromoethane (EDB)	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Dibromomethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,2-Dichlorobenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,3-Dichlorobenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,4-Dichlorobenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Dichlorodifluoromethane	EPA 8260B	6J26012	5.0	ND	1	10/26/2006	10/26/2006	L
1,1-Dichloroethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,2-Dichloroethane	EPA 8260B	6J26012	0.50	ND	1	10/26/2006	10/26/2006	
1,1-Dichloroethene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
cis-1,2-Dichloroethene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
trans-1,2-Dichloroethene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,2-Dichloropropane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,3-Dichloropropane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
2,2-Dichloropropane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,1-Dichloropropene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
cis-1,3-Dichloropropene	EPA 8260B	6J26012	0.50	ND	1	10/26/2006	10/26/2006	
trans-1,3-Dichloropropene	EPA 8260B	6J26012	0.50	ND	1	10/26/2006	10/26/2006	
Ethylbenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Hexachlorobutadiene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Isopropylbenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
p-Isopropyltoluene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Methylene chloride	EPA 8260B	6J26012	5.0	ND	1	10/26/2006	10/26/2006	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-04 (Trip Blank - Water) - cont.								
Reporting Units: ug/l								
Naphthalene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
n-Propylbenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Styrene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,1,1,2-Tetrachloroethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,1,2,2-Tetrachloroethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Tetrachloroethene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Toluene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,2,3-Trichlorobenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,2,4-Trichlorobenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,1,1-Trichloroethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,1,2-Trichloroethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Trichloroethene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Trichlorofluoromethane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Trichlorotrifluoroethane (Freon 113)	EPA 8260B	6J26012	5.0	ND	1	10/26/2006	10/26/2006	
1,2,3-Trichloropropane	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,2,4-Trimethylbenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
1,3,5-Trimethylbenzene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Vinyl chloride	EPA 8260B	6J26012	0.50	ND	1	10/26/2006	10/26/2006	
o-Xylene	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
m,p-Xylenes	EPA 8260B	6J26012	1.0	ND	1	10/26/2006	10/26/2006	
Surrogate: Dibromofluoromethane (80-120%)				*104 %*				
Surrogate: Toluene-d8 (80-120%)				*112 %*				
Surrogate: 4-Bromofluorobenzene (80-120%)				*106 %*				

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

SEMI-VOLATILE ORGANICS BY GC/MS (EPA 3520C/8270C MOD)

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-01 (SW-1 - Water)								
Reporting Units: ug/l								
1,4-Dioxane	EPA 8270C	6J19073	0.50	ND	0.99	10/19/2006	10/20/2006	
Surrogate: 1,4-Dioxane-d8 (30-120%)				*66 %*				
Sample ID: IPJ1647-02 (SW-3 - Water)								
Reporting Units: ug/l								
1,4-Dioxane	EPA 8270C	6J19073	0.48	**0.58**	0.952	10/19/2006	10/20/2006	
Surrogate: 1,4-Dioxane-d8 (30-120%)				*53 %*				
Sample ID: IPJ1647-03 (SW-2 - Water)								
Reporting Units: ug/l								
1,4-Dioxane	EPA 8270C	6J19073	0.48	ND	0.952	10/19/2006	10/20/2006	
Surrogate: 1,4-Dioxane-d8 (30-120%)				*61 %*				

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

SEMI-VOL ORGANICS by GC/MS-CHEMICAL IONIZATION (EPA 3520C/1625C-CI MOD)

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-01 (SW-1 - Water)								
Reporting Units: ug/l								
N-Nitrosodimethylamine	EPA 1625C-CI Mod	6J19073	0.0020	ND	0.99	10/19/2006	10/20/2006	
Sample ID: IPJ1647-02 (SW-3 - Water)								
Reporting Units: ug/l								
N-Nitrosodimethylamine	EPA 1625C-CI Mod	6J19073	0.0019	ND	0.952	10/19/2006	10/20/2006	
Sample ID: IPJ1647-03 (SW-2 - Water)								
Reporting Units: ug/l								
N-Nitrosodimethylamine	EPA 1625C-CI Mod	6J19073	0.0019	ND	0.952	10/19/2006	10/20/2006	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco
393 E. Grand Avenue, Suite J
San Francisco, CA 94080
Attention: Gary Messerotes

Project ID: Weber Waters
40641
Report Number: IPJ1647

Sampled: 10/17/06
Received: 10/17/06

METALS

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-01 (SW-1 - Water)								
Reporting Units: mg/l								
Aluminum	EPA 6010B	6J19085	0.050	**0.21**	1	10/19/2006	10/22/2006	
Antimony	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Arsenic	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Barium	EPA 6010B	6J19085	0.010	**0.085**	1	10/19/2006	10/22/2006	
Beryllium	EPA 6010B	6J19085	0.0040	ND	1	10/19/2006	10/22/2006	
Cadmium	EPA 6010B	6J19085	0.0050	ND	1	10/19/2006	10/22/2006	
Calcium	EPA 6010B	6J19085	0.10	**60**	1	10/19/2006	10/22/2006	
Chromium	EPA 6010B	6J19085	0.0050	**0.023**	1	10/19/2006	10/22/2006	
Cobalt	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Copper	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Iron	EPA 6010B	6J19085	0.040	**1.5**	1	10/19/2006	10/22/2006	
Lead	EPA 6010B	6J19085	0.0050	**0.0073**	1	10/19/2006	10/22/2006	
Magnesium	EPA 6010B	6J19085	0.020	**15**	1	10/19/2006	10/24/2006	
Manganese	EPA 6010B	6J19085	0.020	**0.11**	1	10/19/2006	10/23/2006	
Mercury	EPA 7470A	6J24066	0.00020	ND	1	10/24/2006	10/24/2006	
Molybdenum	EPA 6010B	6J19085	0.020	ND	1	10/19/2006	10/22/2006	
Nickel	EPA 6010B	6J19085	0.010	**0.012**	1	10/19/2006	10/22/2006	
Potassium	EPA 6010B	6J19085	0.50	**2.8**	1	10/19/2006	10/22/2006	
Selenium	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Silver	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Sodium	EPA 6010B	6J19085	0.50	**92**	1	10/19/2006	10/23/2006	
Thallium	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Vanadium	EPA 6010B	6J19085	0.010	**0.016**	1	10/19/2006	10/22/2006	
Zinc	EPA 6010B	6J19085	0.020	**0.13**	1	10/19/2006	10/22/2006	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco
393 E. Grand Avenue, Suite J
San Francisco, CA 94080
Attention: Gary Messerotes

Project ID: Weber Waters
40641
Report Number: IPJ1647

Sampled: 10/17/06
Received: 10/17/06

METALS

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-02 (SW-3 - Water)								
Reporting Units: mg/l								
Aluminum	EPA 6010B	6J19085	0.050	**0.79**	1	10/19/2006	10/22/2006	
Antimony	EPA 6010B	6J19085	0.010	**0.019**	1	10/19/2006	10/22/2006	
Arsenic	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Barium	EPA 6010B	6J19085	0.010	**0.14**	1	10/19/2006	10/22/2006	
Beryllium	EPA 6010B	6J19085	0.0040	ND	1	10/19/2006	10/22/2006	
Cadmium	EPA 6010B	6J19085	0.0050	ND	1	10/19/2006	10/22/2006	
Calcium	EPA 6010B	6J19085	0.10	**85**	1	10/19/2006	10/22/2006	
Chromium	EPA 6010B	6J19085	0.0050	**0.30**	1	10/19/2006	10/22/2006	
Cobalt	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Copper	EPA 6010B	6J19085	0.010	**0.011**	1	10/19/2006	10/22/2006	
Iron	EPA 6010B	6J19085	0.040	**2.0**	1	10/19/2006	10/22/2006	
Lead	EPA 6010B	6J19085	0.0050	ND	1	10/19/2006	10/22/2006	
Magnesium	EPA 6010B	6J19085	0.020	**31**	1	10/19/2006	10/24/2006	
Manganese	EPA 6010B	6J19085	0.020	**0.13**	1	10/19/2006	10/23/2006	
Mercury	EPA 7470A	6J24066	0.00020	ND	1	10/24/2006	10/24/2006	
Molybdenum	EPA 6010B	6J19085	0.020	ND	1	10/19/2006	10/22/2006	
Nickel	EPA 6010B	6J19085	0.010	**0.048**	1	10/19/2006	10/22/2006	
Potassium	EPA 6010B	6J19085	0.50	**4.9**	1	10/19/2006	10/22/2006	
Selenium	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Silver	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Sodium	EPA 6010B	6J19085	0.50	**49**	1	10/19/2006	10/23/2006	
Thallium	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Vanadium	EPA 6010B	6J19085	0.010	**0.012**	1	10/19/2006	10/22/2006	
Zinc	EPA 6010B	6J19085	0.020	**0.22**	1	10/19/2006	10/22/2006	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco
393 E. Grand Avenue, Suite J
San Francisco, CA 94080
Attention: Gary Messerotes

Project ID: Weber Waters
40641
Report Number: IPJ1647

Sampled: 10/17/06
Received: 10/17/06

METALS

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-03 (SW-2 - Water)								
Reporting Units: mg/l								
Aluminum	EPA 6010B	6J19085	0.050	**0.39**	1	10/19/2006	10/22/2006	
Antimony	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Arsenic	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Barium	EPA 6010B	6J19085	0.010	**0.091**	1	10/19/2006	10/22/2006	
Beryllium	EPA 6010B	6J19085	0.0040	ND	1	10/19/2006	10/22/2006	
Cadmium	EPA 6010B	6J19085	0.0050	ND	1	10/19/2006	10/22/2006	
Calcium	EPA 6010B	6J19085	0.10	**63**	1	10/19/2006	10/22/2006	
Chromium	EPA 6010B	6J19085	0.0050	**0.014**	1	10/19/2006	10/22/2006	
Cobalt	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Copper	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Iron	EPA 6010B	6J19085	0.040	**0.46**	1	10/19/2006	10/22/2006	
Lead	EPA 6010B	6J19085	0.0050	ND	1	10/19/2006	10/22/2006	
Magnesium	EPA 6010B	6J19085	0.020	**21**	1	10/19/2006	10/24/2006	
Manganese	EPA 6010B	6J19085	0.020	**0.030**	1	10/19/2006	10/23/2006	
Mercury	EPA 7470A	6J24066	0.00020	ND	1	10/24/2006	10/24/2006	
Molybdenum	EPA 6010B	6J19085	0.020	ND	1	10/19/2006	10/22/2006	
Nickel	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Potassium	EPA 6010B	6J19085	0.50	**4.0**	1	10/19/2006	10/22/2006	
Selenium	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Silver	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Sodium	EPA 6010B	6J19085	0.50	**60**	1	10/19/2006	10/23/2006	
Thallium	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Vanadium	EPA 6010B	6J19085	0.010	ND	1	10/19/2006	10/22/2006	
Zinc	EPA 6010B	6J19085	0.020	**0.049**	1	10/19/2006	10/22/2006	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

DISSOLVED METALS

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-01 (SW-1 - Water)								
Reporting Units: mg/l								
Iron	EPA 6010B-Diss	6J18110	0.040	ND	1	10/18/2006	10/22/2006	
Manganese	EPA 6010B-Diss	6J18110	0.020	ND	1	10/18/2006	10/21/2006	
Sample ID: IPJ1647-02 (SW-3 - Water)								
Reporting Units: mg/l								
Iron	EPA 6010B-Diss	6J18110	0.040	ND	1	10/18/2006	10/21/2006	
Manganese	EPA 6010B-Diss	6J18110	0.020	ND	1	10/18/2006	10/21/2006	
Sample ID: IPJ1647-03 (SW-2 - Water)								
Reporting Units: mg/l								
Iron	EPA 6010B-Diss	6J18110	0.040	ND	1	10/18/2006	10/21/2006	
Manganese	EPA 6010B-Diss	6J18110	0.020	ND	1	10/18/2006	10/21/2006	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco
393 E. Grand Avenue, Suite J
San Francisco, CA 94080
Attention: Gary Messerotes

Project ID: Weber Waters
40641
Report Number: IPJ1647

Sampled: 10/17/06
Received: 10/17/06

INORGANICS

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-01 (SW-1 - Water)								
Reporting Units: mg/l								
Chloride	EPA 300.0	6J17033	2.5	**51**	5	10/17/2006	10/18/2006	
Chromium VI	EPA 7199	6J17132	0.0020	**0.0049**	1	10/17/2006	10/17/2006	
Dissolved Oxygen	EPA 360.1	6J17135	1.0	**5.0**	1	10/17/2006	10/17/2006	
Fluoride	EPA 300.0	6J17032	0.50	ND	1	10/17/2006	10/17/2006	
Nitrate-N	EPA 300.0	6J17033	0.55	**8.4**	5	10/17/2006	10/18/2006	
Nitrite-N	EPA 300.0	6J17032	0.15	ND	1	10/17/2006	10/17/2006	
Sulfate	EPA 300.0	6J17033	2.5	**76**	5	10/17/2006	10/18/2006	
Sulfide	EPA 9034	6J24096	1.0	ND	1	10/24/2006	10/24/2006	
Sample ID: IPJ1647-01 (SW-1 - Water)								
Reporting Units: ug/l								
Perchlorate	EPA 314.0	6J20100	4.0	ND	1	10/20/2006	10/21/2006	
Sample ID: IPJ1647-02 (SW-3 - Water)								
Reporting Units: mg/l								
Chloride	EPA 300.0	6J17033	2.5	**45**	5	10/17/2006	10/18/2006	
Chromium VI	EPA 7199	6J17132	0.020	**0.25**	10	10/17/2006	10/17/2006	
Dissolved Oxygen	EPA 360.1	6J17135	1.0	**5.1**	1	10/17/2006	10/17/2006	
Fluoride	EPA 300.0	6J17032	0.50	ND	1	10/17/2006	10/17/2006	
Nitrate-N	EPA 300.0	6J17033	0.55	**15**	5	10/17/2006	10/18/2006	
Nitrite-N	EPA 300.0	6J17032	0.15	ND	1	10/17/2006	10/17/2006	
Sulfate	EPA 300.0	6J17033	2.5	**81**	5	10/17/2006	10/18/2006	
Sulfide	EPA 9034	6J24096	1.0	ND	1	10/24/2006	10/24/2006	
Sample ID: IPJ1647-02 (SW-3 - Water)								
Reporting Units: ug/l								
Perchlorate	EPA 314.0	6J20100	4.0	ND	1	10/20/2006	10/21/2006	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

INORGANICS

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-03 (SW-2 - Water)								
Reporting Units: mg/l								
Chloride	EPA 300.0	6J17033	2.5	**43**	5	10/17/2006	10/18/2006	
Chromium VI	EPA 7199	6J17132	0.0020	**0.010**	1	10/17/2006	10/17/2006	
Dissolved Oxygen	EPA 360.1	6J17135	1.0	**4.6**	1	10/17/2006	10/17/2006	
Fluoride	EPA 300.0	6J17032	0.50	ND	1	10/17/2006	10/17/2006	
Nitrate-N	EPA 300.0	6J17033	0.55	**8.1**	5	10/17/2006	10/18/2006	
Nitrite-N	EPA 300.0	6J17032	0.15	ND	1	10/17/2006	10/17/2006	
Sulfate	EPA 300.0	6J17033	2.5	**78**	5	10/17/2006	10/18/2006	
Sulfide	EPA 9034	6J24096	1.0	ND	1	10/24/2006	10/24/2006	
Sample ID: IPJ1647-03 (SW-2 - Water)								
Reporting Units: ug/l								
Perchlorate	EPA 314.0	6J20100	4.0	ND	1	10/20/2006	10/21/2006	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

CDHS SRL 524 MODIFIED METHOD FOR 1,2,3-TRICHLOROPROPANE PT/GCMS

Analyte	Method	Batch	Reporting Limit	Sample Result	Dilution Factor	Date Extracted	Date Analyzed	Data Qualifiers
Sample ID: IPJ1647-01 (SW-1 - Water)								
Reporting Units: ng/L								
1,2,3-Trichloropropane (123-TCP)	SRL 524M-TCP	C6J1909	5.0	ND	1	10/19/2006	10/19/2006	
Sample ID: IPJ1647-02 (SW-3 - Water)								
Reporting Units: ng/L								
1,2,3-Trichloropropane (123-TCP)	SRL 524M-TCP	C6J1909	5.0	ND	1	10/19/2006	10/19/2006	
Sample ID: IPJ1647-03 (SW-2 - Water)								
Reporting Units: ng/L								
1,2,3-Trichloropropane (123-TCP)	SRL 524M-TCP	C6J1909	5.0	ND	1	10/19/2006	10/19/2006	
Sample ID: IPJ1647-04 (Trip Blank - Water)								
Reporting Units: ng/L								
1,2,3-Trichloropropane (123-TCP)	SRL 524M-TCP	C6J1909	5.0	ND	1	10/19/2006	10/19/2006	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco
393 E. Grand Avenue, Suite J
San Francisco, CA 94080
Attention: Gary Messerotes

Project ID: Weber Waters
40641
Report Number: IPJ1647

Sampled: 10/17/06
Received: 10/17/06

SHORT HOLD TIME DETAIL REPORT

	Hold Time (in days)	Date/Time Sampled	Date/Time Received	Date/Time Extracted	Date/Time Analyzed
Sample ID: SW-1 (IPJ1647-01) - Water					
EPA 300.0	2	10/17/2006 09:00	10/17/2006 18:30	10/17/2006 23:30	10/18/2006 12:54
Nitrite-N				10/17/2006 20:00	10/17/2006 21:35
EPA 360.1	1	10/17/2006 09:00	10/17/2006 18:30	10/17/2006 21:27	10/17/2006 22:40
EPA 7199	1	10/17/2006 09:00	10/17/2006 18:30	10/17/2006 21:01	10/17/2006 22:15
Sample ID: SW-3 (IPJ1647-02) - Water					
EPA 300.0	2	10/17/2006 11:40	10/17/2006 18:30	10/17/2006 23:30	10/18/2006 13:13
Nitrite-N				10/17/2006 20:00	10/17/2006 21:48
EPA 360.1	1	10/17/2006 11:40	10/17/2006 18:30	10/17/2006 21:27	10/17/2006 22:40
EPA 7199	1	10/17/2006 11:40	10/17/2006 18:30	10/17/2006 21:01	10/17/2006 23:33
Sample ID: SW-2 (IPJ1647-03) - Water					
EPA 300.0	2	10/17/2006 13:45	10/17/2006 18:30	10/17/2006 23:30	10/18/2006 13:31
Nitrite-N				10/17/2006 20:00	10/17/2006 22:00
EPA 360.1	1	10/17/2006 13:45	10/17/2006 18:30	10/17/2006 21:27	10/17/2006 22:40
EPA 7199	1	10/17/2006 13:45	10/17/2006 18:30	10/17/2006 21:01	10/17/2006 22:37

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco
393 E. Grand Avenue, Suite J
San Francisco, CA 94080
Attention: Gary Messerotes

Project ID: Weber Waters
40641
Report Number: IPJ1647

Sampled: 10/17/06
Received: 10/17/06

POTENTIAL STLC / TCLP / TTLC LIMITS EXCEEDANCE

Analyte	Units	Sample Result	STLC Max. Limit mg/L (ppm)	TTLC Max. Limit mg/Kg (ppm)	TCLP Max. Limit mg/L (ppm)
IPJ1647-01 (SW-1 - Water) EPA 6010B					
Aluminum	mg/l	0.21			
Antimony	mg/l	ND	15	500	
Arsenic	mg/l	ND	5.0	500	5.0
Barium	mg/l	0.085	100	10000	100
Beryllium	mg/l	ND	0.75	75	
Cadmium	mg/l	ND	1.0	100	1.0
Calcium	mg/l	60			
Chromium	mg/l	0.023	5.0	2500	5.0
Cobalt	mg/l	ND	80	8000	
Copper	mg/l	ND	25	2500	
Iron	mg/l	1.5			
Lead	mg/l	0.0073	5.0	1000	5.0
Magnesium	mg/l	15			
Manganese	mg/l	0.11			
Mercury	mg/l	ND	0.20	20	0.20
Molybdenum	mg/l	ND	350	3500	
Nickel	mg/l	0.012	20	2000	
Potassium	mg/l	2.8			
Selenium	mg/l	ND	1.0	100	1.0
Silver	mg/l	ND	5.0	500	5.0
Sodium	mg/l	92			
Thallium	mg/l	ND	7.0	700	
Vanadium	mg/l	0.016	24	2400	
Zinc	mg/l	0.13	250	5000	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco
393 E. Grand Avenue, Suite J
San Francisco, CA 94080
Attention: Gary Messerotes

Project ID: Weber Waters
40641
Report Number: IPJ1647

Sampled: 10/17/06
Received: 10/17/06

POTENTIAL STLC / TCLP / TTLC LIMITS EXCEEDANCE

Analyte	Units	Sample Result	STLC Max. Limit mg/L (ppm)	TTLC Max. Limit mg/Kg (ppm)	TCLP Max. Limit mg/L (ppm)
IPJ1647-02 (SW-3 - Water) EPA 6010B					
Aluminum	mg/l	0.79			
Antimony	mg/l	0.019	15	500	
Arsenic	mg/l	ND	5.0	500	5.0
Barium	mg/l	0.14	100	10000	100
Beryllium	mg/l	ND	0.75	75	
Cadmium	mg/l	ND	1.0	100	1.0
Calcium	mg/l	85			
Chromium	mg/l	0.30	5.0	2500	5.0
Cobalt	mg/l	ND	80	8000	
Copper	mg/l	0.011	25	2500	
Iron	mg/l	2.0			
Lead	mg/l	ND	5.0	1000	5.0
Magnesium	mg/l	31			
Manganese	mg/l	0.13			
Mercury	mg/l	ND	0.20	20	0.20
Molybdenum	mg/l	ND	350	3500	
Nickel	mg/l	0.048	20	2000	
Potassium	mg/l	4.9			
Selenium	mg/l	ND	1.0	100	1.0
Silver	mg/l	ND	5.0	500	5.0
Sodium	mg/l	49			
Thallium	mg/l	ND	7.0	700	
Vanadium	mg/l	0.012	24	2400	
Zinc	mg/l	0.22	250	5000	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco
393 E. Grand Avenue, Suite J
San Francisco, CA 94080
Attention: Gary Messerotes

Project ID: Weber Waters
40641
Report Number: IPJ1647

Sampled: 10/17/06
Received: 10/17/06

POTENTIAL STLC / TCLP / TTLC LIMITS EXCEEDANCE

Analyte	Units	Sample Result	STLC Max. Limit mg/L (ppm)	TTLC Max. Limit mg/Kg (ppm)	TCLP Max. Limit mg/L (ppm)
IPJ1647-03 (SW-2 - Water) EPA 6010B					
Aluminum	mg/l	0.39			
Antimony	mg/l	ND	15	500	
Arsenic	mg/l	ND	5.0	500	5.0
Barium	mg/l	0.091	100	10000	100
Beryllium	mg/l	ND	0.75	75	
Cadmium	mg/l	ND	1.0	100	1.0
Calcium	mg/l	63			
Chromium	mg/l	0.014	5.0	2500	5.0
Cobalt	mg/l	ND	80	8000	
Copper	mg/l	ND	25	2500	
Iron	mg/l	0.46			
Lead	mg/l	ND	5.0	1000	5.0
Magnesium	mg/l	21			
Manganese	mg/l	0.030			
Mercury	mg/l	ND	0.20	20	0.20
Molybdenum	mg/l	ND	350	3500	
Nickel	mg/l	ND	20	2000	
Potassium	mg/l	4.0			
Selenium	mg/l	ND	1.0	100	1.0
Silver	mg/l	ND	5.0	500	5.0
Sodium	mg/l	60			
Thallium	mg/l	ND	7.0	700	
Vanadium	mg/l	ND	24	2400	
Zinc	mg/l	0.049	250	5000	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

POTENTIAL STLC / TCLP / TTLC LIMITS EXCEEDANCE

Analyte	Units	Sample Result	STLC Max. Limit mg/L (ppm)	TTLC Max. Limit mg/Kg (ppm)	TCLP Max. Limit mg/L (ppm)
IPJ1647-01 (SW-1 - Water) EPA 6010B-Diss					
Iron	mg/l	ND			
Manganese	mg/l	ND			
IPJ1647-02 (SW-3 - Water) EPA 6010B-Diss					
Iron	mg/l	ND			
Manganese	mg/l	ND			
IPJ1647-03 (SW-2 - Water) EPA 6010B-Diss					
Iron	mg/l	ND			
Manganese	mg/l	ND			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J22009 Extracted: 10/22/06										
Blank Analyzed: 10/22/2006 (6J22009-BLK1)										
Acetone	ND	10	ug/l							
Benzene	ND	0.50	ug/l							
Bromobenzene	ND	1.0	ug/l							
Bromochloromethane	ND	1.0	ug/l							
Bromodichloromethane	ND	1.0	ug/l							
Bromoform	ND	1.0	ug/l							
Bromomethane	ND	1.0	ug/l							
n-Butylbenzene	ND	1.0	ug/l							
sec-Butylbenzene	ND	1.0	ug/l							
tert-Butylbenzene	ND	1.0	ug/l							
Carbon tetrachloride	ND	0.50	ug/l							
Chlorobenzene	ND	1.0	ug/l							
Chloroethane	ND	1.0	ug/l							
Chloroform	ND	1.0	ug/l							
Chloromethane	ND	1.0	ug/l							
2-Chlorotoluene	ND	1.0	ug/l							
4-Chlorotoluene	ND	1.0	ug/l							
Dibromochloromethane	ND	1.0	ug/l							
1,2-Dibromo-3-chloropropane	ND	5.0	ug/l							
1,2-Dibromoethane (EDB)	ND	1.0	ug/l							
Dibromomethane	ND	1.0	ug/l							
1,2-Dichlorobenzene	ND	1.0	ug/l							
1,3-Dichlorobenzene	ND	1.0	ug/l							
1,4-Dichlorobenzene	ND	1.0	ug/l							
Dichlorodifluoromethane	ND	5.0	ug/l							
1,1-Dichloroethane	ND	1.0	ug/l							
1,2-Dichloroethane	ND	0.50	ug/l							
1,1-Dichloroethene	ND	1.0	ug/l							
cis-1,2-Dichloroethene	ND	1.0	ug/l							
trans-1,2-Dichloroethene	ND	1.0	ug/l							
1,2-Dichloropropane	ND	1.0	ug/l							
1,3-Dichloropropane	ND	1.0	ug/l							
2,2-Dichloropropane	ND	1.0	ug/l							
1,1-Dichloropropene	ND	1.0	ug/l							
cis-1,3-Dichloropropene	ND	0.50	ug/l							

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J22009 Extracted: 10/22/06										
Blank Analyzed: 10/22/2006 (6J22009-BLK1)										
trans-1,3-Dichloropropene	ND	0.50	ug/l							
Ethylbenzene	ND	1.0	ug/l							
Hexachlorobutadiene	ND	1.0	ug/l							
Isopropylbenzene	ND	1.0	ug/l							
p-Isopropyltoluene	ND	1.0	ug/l							
Methylene chloride	ND	5.0	ug/l							
Naphthalene	ND	1.0	ug/l							
n-Propylbenzene	ND	1.0	ug/l							
Styrene	ND	1.0	ug/l							
1,1,1,2-Tetrachloroethane	ND	1.0	ug/l							
1,1,2,2-Tetrachloroethane	ND	1.0	ug/l							
Tetrachloroethene	ND	1.0	ug/l							
Toluene	ND	1.0	ug/l							
1,2,3-Trichlorobenzene	ND	1.0	ug/l							
1,2,4-Trichlorobenzene	ND	1.0	ug/l							
1,1,1-Trichloroethane	ND	1.0	ug/l							
1,1,2-Trichloroethane	ND	1.0	ug/l							
Trichloroethene	ND	1.0	ug/l							
Trichlorofluoromethane	ND	1.0	ug/l							
Trichlorotrifluoroethane (Freon 113)	ND	5.0	ug/l							
1,2,3-Trichloropropane	ND	1.0	ug/l							
1,2,4-Trimethylbenzene	ND	1.0	ug/l							
1,3,5-Trimethylbenzene	ND	1.0	ug/l							
Vinyl chloride	ND	0.50	ug/l							
o-Xylene	ND	1.0	ug/l							
m,p-Xylenes	ND	1.0	ug/l							
Surrogate: Dibromofluoromethane	*26.4*		*ug/l*	*25.0*		*106*	*80-120*			
Surrogate: Toluene-d8	*24.7*		*ug/l*	*25.0*		*99*	*80-120*			
Surrogate: 4-Bromofluorobenzene	*23.7*		*ug/l*	*25.0*		*95*	*80-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J22009 Extracted: 10/22/06										
LCS Analyzed: 10/22/2006 (6J22009-BS1)										
Acetone	17.5	10	ug/l	25.0		70	30-140			
Benzene	23.4	0.50	ug/l	25.0		94	70-120			
Bromobenzene	23.1	1.0	ug/l	25.0		92	75-120			
Bromochloromethane	25.1	1.0	ug/l	25.0		100	70-130			
Bromodichloromethane	24.4	1.0	ug/l	25.0		98	70-135			
Bromoform	23.1	1.0	ug/l	25.0		92	55-130			
Bromomethane	31.7	1.0	ug/l	25.0		127	65-140			
n-Butylbenzene	24.9	1.0	ug/l	25.0		100	70-130			
sec-Butylbenzene	24.9	1.0	ug/l	25.0		100	70-125			
tert-Butylbenzene	25.1	1.0	ug/l	25.0		100	70-125			
Carbon tetrachloride	24.6	0.50	ug/l	25.0		98	65-140			
Chlorobenzene	22.9	1.0	ug/l	25.0		92	75-120			
Chloroethane	28.9	1.0	ug/l	25.0		116	60-140			
Chloroform	24.7	1.0	ug/l	25.0		99	70-130			
Chloromethane	32.4	1.0	ug/l	25.0		130	50-140			
2-Chlorotoluene	23.8	1.0	ug/l	25.0		95	70-125			
4-Chlorotoluene	24.1	1.0	ug/l	25.0		96	75-125			
Dibromochloromethane	25.8	1.0	ug/l	25.0		103	70-140			
1,2-Dibromo-3-chloropropane	21.3	5.0	ug/l	25.0		85	50-135			
1,2-Dibromoethane (EDB)	23.6	1.0	ug/l	25.0		94	75-125			
Dibromomethane	23.0	1.0	ug/l	25.0		92	70-125			
1,2-Dichlorobenzene	23.2	1.0	ug/l	25.0		93	75-120			
1,3-Dichlorobenzene	22.8	1.0	ug/l	25.0		91	75-120			
1,4-Dichlorobenzene	22.5	1.0	ug/l	25.0		90	75-120			
Dichlorodifluoromethane	33.2	5.0	ug/l	25.0		133	35-155			
1,1-Dichloroethane	24.6	1.0	ug/l	25.0		98	70-125			
1,2-Dichloroethane	24.8	0.50	ug/l	25.0		99	60-140			
1,1-Dichloroethene	24.4	1.0	ug/l	25.0		98	70-125			
cis-1,2-Dichloroethene	25.0	1.0	ug/l	25.0		100	70-125			
trans-1,2-Dichloroethene	25.6	1.0	ug/l	25.0		102	70-125			
1,2-Dichloropropane	24.2	1.0	ug/l	25.0		97	70-125			
1,3-Dichloropropane	23.5	1.0	ug/l	25.0		94	70-120			
2,2-Dichloropropane	24.6	1.0	ug/l	25.0		98	65-140			
1,1-Dichloropropene	23.8	1.0	ug/l	25.0		95	75-130			
cis-1,3-Dichloropropene	22.7	0.50	ug/l	25.0		91	75-125			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J22009 Extracted: 10/22/06										
LCS Analyzed: 10/22/2006 (6J22009-BS1)										
trans-1,3-Dichloropropene	22.5	0.50	ug/l	25.0		90	70-125			
Ethylbenzene	25.2	1.0	ug/l	25.0		101	75-125			
Hexachlorobutadiene	21.3	1.0	ug/l	25.0		85	65-135			
Isopropylbenzene	28.3	1.0	ug/l	25.0		113	75-130			
p-Isopropyltoluene	25.2	1.0	ug/l	25.0		101	75-125			
Methylene chloride	23.8	5.0	ug/l	25.0		95	55-130			
Naphthalene	22.1	1.0	ug/l	25.0		88	55-135			
n-Propylbenzene	25.6	1.0	ug/l	25.0		102	75-130			
Styrene	25.4	1.0	ug/l	25.0		102	75-130			
1,1,1,2-Tetrachloroethane	24.0	1.0	ug/l	25.0		96	70-130			
1,1,2,2-Tetrachloroethane	22.4	1.0	ug/l	25.0		90	55-130			
Tetrachloroethene	22.0	1.0	ug/l	25.0		88	70-125			
Toluene	24.1	1.0	ug/l	25.0		96	70-120			
1,2,3-Trichlorobenzene	22.3	1.0	ug/l	25.0		89	65-125			
1,2,4-Trichlorobenzene	22.8	1.0	ug/l	25.0		91	70-135			
1,1,1-Trichloroethane	25.0	1.0	ug/l	25.0		100	65-135			
1,1,2-Trichloroethane	23.3	1.0	ug/l	25.0		93	70-125			
Trichloroethene	23.7	1.0	ug/l	25.0		95	70-125			
Trichlorofluoromethane	28.0	1.0	ug/l	25.0		112	65-145			
1,2,3-Trichloropropane	21.6	1.0	ug/l	25.0		86	60-130			
1,2,4-Trimethylbenzene	25.0	1.0	ug/l	25.0		100	75-125			
1,3,5-Trimethylbenzene	25.3	1.0	ug/l	25.0		101	75-125			
Vinyl chloride	30.6	0.50	ug/l	25.0		122	55-135			
o-Xylene	25.0	1.0	ug/l	25.0		100	75-125			
m,p-Xylenes	50.2	1.0	ug/l	50.0		100	75-125			
Surrogate: Dibromofluoromethane	*26.9*		*ug/l*	*25.0*		*108*	*80-120*			
Surrogate: Toluene-d8	*25.3*		*ug/l*	*25.0*		*101*	*80-120*			
Surrogate: 4-Bromofluorobenzene	*25.5*		*ug/l*	*25.0*		*102*	*80-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J22009 Extracted: 10/22/06										
Matrix Spike Analyzed: 10/22/2006 (6J22009-MS1)					**Source: IPJ1629-04**					
Acetone	15.9	10	ug/l	25.0	ND	64	20-150			
Benzene	23.4	0.50	ug/l	25.0	ND	94	65-125			
Bromobenzene	22.4	1.0	ug/l	25.0	ND	90	70-125			
Bromochloromethane	26.1	1.0	ug/l	25.0	ND	104	65-135			
Bromodichloromethane	25.5	1.0	ug/l	25.0	ND	102	70-135			
Bromoform	25.3	1.0	ug/l	25.0	ND	101	55-135			
Bromomethane	33.2	1.0	ug/l	25.0	ND	133	55-145			
n-Butylbenzene	24.8	1.0	ug/l	25.0	ND	99	65-135			
sec-Butylbenzene	24.2	1.0	ug/l	25.0	ND	97	70-125			
tert-Butylbenzene	24.4	1.0	ug/l	25.0	ND	98	65-130			
Carbon tetrachloride	25.6	0.50	ug/l	25.0	ND	102	65-140			
Chlorobenzene	23.5	1.0	ug/l	25.0	ND	94	75-125			
Chloroethane	31.4	1.0	ug/l	25.0	ND	126	55-140			
Chloroform	26.6	1.0	ug/l	25.0	ND	106	65-135			
Chloromethane	32.9	1.0	ug/l	25.0	ND	132	45-145			
2-Chlorotoluene	22.8	1.0	ug/l	25.0	ND	91	65-135			
4-Chlorotoluene	23.3	1.0	ug/l	25.0	ND	93	70-135			
Dibromochloromethane	27.6	1.0	ug/l	25.0	ND	110	65-140			
1,2-Dibromo-3-chloropropane	24.6	5.0	ug/l	25.0	ND	98	45-145			
1,2-Dibromoethane (EDB)	25.4	1.0	ug/l	25.0	ND	102	70-130			
Dibromomethane	24.3	1.0	ug/l	25.0	ND	97	65-135			
1,2-Dichlorobenzene	23.8	1.0	ug/l	25.0	ND	95	75-125			
1,3-Dichlorobenzene	23.0	1.0	ug/l	25.0	ND	92	75-125			
1,4-Dichlorobenzene	22.9	1.0	ug/l	25.0	ND	92	75-125			
Dichlorodifluoromethane	35.4	5.0	ug/l	25.0	ND	142	25-155			
1,1-Dichloroethane	26.1	1.0	ug/l	25.0	ND	104	65-130			
1,2-Dichloroethane	26.6	0.50	ug/l	25.0	ND	106	60-140			
1,1-Dichloroethene	25.2	1.0	ug/l	25.0	ND	101	60-130			
cis-1,2-Dichloroethene	26.1	1.0	ug/l	25.0	ND	104	65-130			
trans-1,2-Dichloroethene	26.0	1.0	ug/l	25.0	ND	104	65-130			
1,2-Dichloropropane	24.3	1.0	ug/l	25.0	ND	97	65-130			
1,3-Dichloropropane	24.8	1.0	ug/l	25.0	ND	99	65-135			
2,2-Dichloropropane	27.1	1.0	ug/l	25.0	ND	108	60-145			
1,1-Dichloropropene	23.8	1.0	ug/l	25.0	ND	95	70-135			
cis-1,3-Dichloropropene	23.0	0.50	ug/l	25.0	ND	92	70-130			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J22009 Extracted: 10/22/06										
Matrix Spike Analyzed: 10/22/2006 (6J22009-MS1)					**Source: IPJ1629-04**					
trans-1,3-Dichloropropene	23.9	0.50	ug/l	25.0	ND	96	65-135			
Ethylbenzene	25.8	1.0	ug/l	25.0	ND	103	65-130			
Hexachlorobutadiene	21.4	1.0	ug/l	25.0	ND	86	60-135			
Isopropylbenzene	26.4	1.0	ug/l	25.0	ND	106	70-135			
p-Isopropyltoluene	24.8	1.0	ug/l	25.0	ND	99	65-130			
Methylene chloride	25.2	5.0	ug/l	25.0	ND	101	50-135			
Naphthalene	25.1	1.0	ug/l	25.0	ND	100	50-140			
n-Propylbenzene	24.3	1.0	ug/l	25.0	ND	97	70-135			
Styrene	26.7	1.0	ug/l	25.0	ND	107	50-145			
1,1,1,2-Tetrachloroethane	25.1	1.0	ug/l	25.0	ND	100	65-140			
1,1,2,2-Tetrachloroethane	23.6	1.0	ug/l	25.0	ND	94	55-135			
Tetrachloroethene	22.3	1.0	ug/l	25.0	0.62	87	65-130			
Toluene	24.1	1.0	ug/l	25.0	ND	96	70-125			
1,2,3-Trichlorobenzene	24.3	1.0	ug/l	25.0	ND	97	60-135			
1,2,4-Trichlorobenzene	24.2	1.0	ug/l	25.0	ND	97	65-135			
1,1,1-Trichloroethane	27.6	1.0	ug/l	25.0	0.41	109	65-140			
1,1,2-Trichloroethane	24.2	1.0	ug/l	25.0	ND	97	65-130			
Trichloroethene	23.1	1.0	ug/l	25.0	ND	92	65-125			
Trichlorofluoromethane	30.5	1.0	ug/l	25.0	ND	122	60-145			
1,2,3-Trichloropropane	22.2	1.0	ug/l	25.0	ND	89	55-135			
1,2,4-Trimethylbenzene	24.8	1.0	ug/l	25.0	ND	99	55-135			
1,3,5-Trimethylbenzene	24.5	1.0	ug/l	25.0	ND	98	70-130			
Vinyl chloride	33.8	0.50	ug/l	25.0	ND	135	45-140			
o-Xylene	26.5	1.0	ug/l	25.0	ND	106	65-125			
m,p-Xylenes	51.6	1.0	ug/l	50.0	ND	103	65-130			
Surrogate: Dibromofluoromethane	*28.6*		*ug/l*	*25.0*		*114*	*80-120*			
Surrogate: Toluene-d8	*25.3*		*ug/l*	*25.0*		*101*	*80-120*			
Surrogate: 4-Bromofluorobenzene	*27.3*		*ug/l*	*25.0*		*109*	*80-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J22009 Extracted: 10/22/06										
Matrix Spike Dup Analyzed: 10/22/2006 (6J22009-MSD1)					**Source: IPJ1629-04**					
Acetone	14.8	10	ug/l	25.0	ND	59	20-150	7	35	
Benzene	23.7	0.50	ug/l	25.0	ND	95	65-125	1	20	
Bromobenzene	22.8	1.0	ug/l	25.0	ND	91	70-125	2	20	
Bromochloromethane	26.3	1.0	ug/l	25.0	ND	105	65-135	1	25	
Bromodichloromethane	25.1	1.0	ug/l	25.0	ND	100	70-135	2	20	
Bromoform	24.9	1.0	ug/l	25.0	ND	100	55-135	2	25	
Bromomethane	33.6	1.0	ug/l	25.0	ND	134	55-145	1	25	
n-Butylbenzene	25.2	1.0	ug/l	25.0	ND	101	65-135	2	20	
sec-Butylbenzene	24.6	1.0	ug/l	25.0	ND	98	70-125	2	20	
tert-Butylbenzene	24.7	1.0	ug/l	25.0	ND	99	65-130	1	20	
Carbon tetrachloride	25.0	0.50	ug/l	25.0	ND	100	65-140	2	25	
Chlorobenzene	23.6	1.0	ug/l	25.0	ND	94	75-125	0	20	
Chloroethane	30.7	1.0	ug/l	25.0	ND	123	55-140	2	25	
Chloroform	25.6	1.0	ug/l	25.0	ND	102	65-135	4	20	
Chloromethane	34.0	1.0	ug/l	25.0	ND	136	45-145	3	25	
2-Chlorotoluene	23.2	1.0	ug/l	25.0	ND	93	65-135	2	20	
4-Chlorotoluene	24.0	1.0	ug/l	25.0	ND	96	70-135	3	20	
Dibromochloromethane	26.6	1.0	ug/l	25.0	ND	106	65-140	4	25	
1,2-Dibromo-3-chloropropane	23.9	5.0	ug/l	25.0	ND	96	45-145	3	30	
1,2-Dibromoethane (EDB)	25.0	1.0	ug/l	25.0	ND	100	70-130	2	25	
Dibromomethane	24.3	1.0	ug/l	25.0	ND	97	65-135	0	25	
1,2-Dichlorobenzene	23.9	1.0	ug/l	25.0	ND	96	75-125	0	20	
1,3-Dichlorobenzene	23.7	1.0	ug/l	25.0	ND	95	75-125	3	20	
1,4-Dichlorobenzene	23.3	1.0	ug/l	25.0	ND	93	75-125	2	20	
Dichlorodifluoromethane	35.6	5.0	ug/l	25.0	ND	142	25-155	1	30	
1,1-Dichloroethane	25.4	1.0	ug/l	25.0	ND	102	65-130	3	20	
1,2-Dichloroethane	25.4	0.50	ug/l	25.0	ND	102	60-140	5	20	
1,1-Dichloroethene	25.2	1.0	ug/l	25.0	ND	101	60-130	0	20	
cis-1,2-Dichloroethene	25.9	1.0	ug/l	25.0	ND	104	65-130	1	20	
trans-1,2-Dichloroethene	26.3	1.0	ug/l	25.0	ND	105	65-130	1	20	
1,2-Dichloropropane	24.8	1.0	ug/l	25.0	ND	99	65-130	2	20	
1,3-Dichloropropane	24.0	1.0	ug/l	25.0	ND	96	65-135	3	25	
2,2-Dichloropropane	26.3	1.0	ug/l	25.0	ND	105	60-145	3	25	
1,1-Dichloropropene	23.8	1.0	ug/l	25.0	ND	95	70-135	0	20	
cis-1,3-Dichloropropene	23.2	0.50	ug/l	25.0	ND	93	70-130	1	20	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J22009 Extracted: 10/22/06										
Matrix Spike Dup Analyzed: 10/22/2006 (6J22009-MSD1)					**Source: IPJ1629-04**					
trans-1,3-Dichloropropene	23.7	0.50	ug/l	25.0	ND	95	65-135	1	25	
Ethylbenzene	25.4	1.0	ug/l	25.0	ND	102	65-130	2	20	
Hexachlorobutadiene	22.6	1.0	ug/l	25.0	ND	90	60-135	5	20	
Isopropylbenzene	26.8	1.0	ug/l	25.0	ND	107	70-135	2	20	
p-Isopropyltoluene	25.3	1.0	ug/l	25.0	ND	101	65-130	2	20	
Methylene chloride	25.5	5.0	ug/l	25.0	ND	102	50-135	1	20	
Naphthalene	25.8	1.0	ug/l	25.0	ND	103	50-140	3	30	
n-Propylbenzene	24.8	1.0	ug/l	25.0	ND	99	70-135	2	20	
Styrene	26.4	1.0	ug/l	25.0	ND	106	50-145	1	30	
1,1,1,2-Tetrachloroethane	24.7	1.0	ug/l	25.0	ND	99	65-140	2	20	
1,1,2,2-Tetrachloroethane	23.6	1.0	ug/l	25.0	ND	94	55-135	0	30	
Tetrachloroethene	22.5	1.0	ug/l	25.0	0.62	88	65-130	1	20	
Toluene	24.2	1.0	ug/l	25.0	ND	97	70-125	0	20	
1,2,3-Trichlorobenzene	25.6	1.0	ug/l	25.0	ND	102	60-135	5	20	
1,2,4-Trichlorobenzene	25.6	1.0	ug/l	25.0	ND	102	65-135	6	20	
1,1,1-Trichloroethane	26.2	1.0	ug/l	25.0	0.41	103	65-140	5	20	
1,1,2-Trichloroethane	24.7	1.0	ug/l	25.0	ND	99	65-130	2	25	
Trichloroethene	23.9	1.0	ug/l	25.0	ND	96	65-125	3	20	
Trichlorofluoromethane	29.0	1.0	ug/l	25.0	ND	116	60-145	5	25	
1,2,3-Trichloropropane	22.6	1.0	ug/l	25.0	ND	90	55-135	2	30	
1,2,4-Trimethylbenzene	25.2	1.0	ug/l	25.0	ND	101	55-135	2	25	
1,3,5-Trimethylbenzene	25.0	1.0	ug/l	25.0	ND	100	70-130	2	20	
Vinyl chloride	32.6	0.50	ug/l	25.0	ND	130	45-140	4	30	
o-Xylene	26.1	1.0	ug/l	25.0	ND	104	65-125	2	20	
m,p-Xylenes	51.2	1.0	ug/l	50.0	ND	102	65-130	1	25	
Surrogate: Dibromofluoromethane	*27.2*		*ug/l*	*25.0*		*109*	*80-120*			
Surrogate: Toluene-d8	*25.1*		*ug/l*	*25.0*		*100*	*80-120*			
Surrogate: 4-Bromofluorobenzene	*26.1*		*ug/l*	*25.0*		*104*	*80-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J23005 Extracted: 10/23/06										
Blank Analyzed: 10/23/2006 (6J23005-BLK1)										
Acetone	ND	10	ug/l							
Benzene	ND	0.50	ug/l							
Bromobenzene	ND	1.0	ug/l							
Bromochloromethane	ND	1.0	ug/l							
Bromodichloromethane	ND	1.0	ug/l							
Bromoform	ND	1.0	ug/l							
Bromomethane	ND	1.0	ug/l							
n-Butylbenzene	ND	1.0	ug/l							
sec-Butylbenzene	ND	1.0	ug/l							
tert-Butylbenzene	ND	1.0	ug/l							
Carbon tetrachloride	ND	0.50	ug/l							
Chlorobenzene	ND	1.0	ug/l							
Chloroethane	ND	1.0	ug/l							
Chloroform	ND	1.0	ug/l							
Chloromethane	ND	1.0	ug/l							
2-Chlorotoluene	ND	1.0	ug/l							
4-Chlorotoluene	ND	1.0	ug/l							
Dibromochloromethane	ND	1.0	ug/l							
1,2-Dibromo-3-chloropropane	ND	5.0	ug/l							
1,2-Dibromoethane (EDB)	ND	1.0	ug/l							
Dibromomethane	ND	1.0	ug/l							
1,2-Dichlorobenzene	ND	1.0	ug/l							
1,3-Dichlorobenzene	ND	1.0	ug/l							
1,4-Dichlorobenzene	ND	1.0	ug/l							
Dichlorodifluoromethane	ND	5.0	ug/l							
1,1-Dichloroethane	ND	1.0	ug/l							
1,2-Dichloroethane	ND	0.50	ug/l							
1,1-Dichloroethene	ND	1.0	ug/l							
cis-1,2-Dichloroethene	ND	1.0	ug/l							
trans-1,2-Dichloroethene	ND	1.0	ug/l							
1,2-Dichloropropane	ND	1.0	ug/l							
1,3-Dichloropropane	ND	1.0	ug/l							
2,2-Dichloropropane	ND	1.0	ug/l							
1,1-Dichloropropene	ND	1.0	ug/l							
cis-1,3-Dichloropropene	ND	0.50	ug/l							

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J23005 Extracted: 10/23/06										
Blank Analyzed: 10/23/2006 (6J23005-BLK1)										
trans-1,3-Dichloropropene	ND	0.50	ug/l							
Ethylbenzene	ND	1.0	ug/l							
Hexachlorobutadiene	ND	1.0	ug/l							
Isopropylbenzene	ND	1.0	ug/l							
p-Isopropyltoluene	ND	1.0	ug/l							
Methylene chloride	ND	5.0	ug/l							
Naphthalene	ND	1.0	ug/l							
n-Propylbenzene	ND	1.0	ug/l							
Styrene	ND	1.0	ug/l							
1,1,1,2-Tetrachloroethane	ND	1.0	ug/l							
1,1,2,2-Tetrachloroethane	ND	1.0	ug/l							
Tetrachloroethene	ND	1.0	ug/l							
Toluene	ND	1.0	ug/l							
1,2,3-Trichlorobenzene	ND	1.0	ug/l							
1,2,4-Trichlorobenzene	ND	1.0	ug/l							
1,1,1-Trichloroethane	ND	1.0	ug/l							
1,1,2-Trichloroethane	ND	1.0	ug/l							
Trichloroethene	ND	1.0	ug/l							
Trichlorofluoromethane	ND	1.0	ug/l							
Trichlorotrifluoroethane (Freon 113)	ND	5.0	ug/l							
1,2,3-Trichloropropane	ND	1.0	ug/l							
1,2,4-Trimethylbenzene	ND	1.0	ug/l							
1,3,5-Trimethylbenzene	ND	1.0	ug/l							
Vinyl chloride	ND	0.50	ug/l							
o-Xylene	ND	1.0	ug/l							
m,p-Xylenes	ND	1.0	ug/l							
Surrogate: Dibromofluoromethane	*23.4*		*ug/l*	*25.0*		*94*	*80-120*			
Surrogate: Toluene-d8	*27.3*		*ug/l*	*25.0*		*109*	*80-120*			
Surrogate: 4-Bromofluorobenzene	*25.9*		*ug/l*	*25.0*		*104*	*80-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J23005 Extracted: 10/23/06										
LCS Analyzed: 10/23/2006 (6J23005-BS1)										
Acetone	17.8	10	ug/l	25.0		71	30-140			
Benzene	22.8	0.50	ug/l	25.0		91	70-120			
Bromobenzene	23.2	1.0	ug/l	25.0		93	75-120			
Bromochloromethane	21.2	1.0	ug/l	25.0		85	70-130			
Bromodichloromethane	22.2	1.0	ug/l	25.0		89	70-135			
Bromoform	20.4	1.0	ug/l	25.0		82	55-130			
Bromomethane	26.3	1.0	ug/l	25.0		105	65-140			
n-Butylbenzene	25.4	1.0	ug/l	25.0		102	70-130			
sec-Butylbenzene	24.7	1.0	ug/l	25.0		99	70-125			
tert-Butylbenzene	24.3	1.0	ug/l	25.0		97	70-125			
Carbon tetrachloride	22.5	0.50	ug/l	25.0		90	65-140			
Chlorobenzene	22.6	1.0	ug/l	25.0		90	75-120			
Chloroethane	23.7	1.0	ug/l	25.0		95	60-140			
Chloroform	20.7	1.0	ug/l	25.0		83	70-130			
Chloromethane	26.0	1.0	ug/l	25.0		104	50-140			
2-Chlorotoluene	23.4	1.0	ug/l	25.0		94	70-125			
4-Chlorotoluene	23.8	1.0	ug/l	25.0		95	75-125			
Dibromochloromethane	23.7	1.0	ug/l	25.0		95	70-140			
1,2-Dibromo-3-chloropropane	24.1	5.0	ug/l	25.0		96	50-135			
1,2-Dibromoethane (EDB)	23.4	1.0	ug/l	25.0		94	75-125			
Dibromomethane	22.9	1.0	ug/l	25.0		92	70-125			
1,2-Dichlorobenzene	22.8	1.0	ug/l	25.0		91	75-120			
1,3-Dichlorobenzene	23.3	1.0	ug/l	25.0		93	75-120			
1,4-Dichlorobenzene	22.7	1.0	ug/l	25.0		91	75-120			
Dichlorodifluoromethane	31.1	5.0	ug/l	25.0		124	35-155			
1,1-Dichloroethane	21.5	1.0	ug/l	25.0		86	70-125			
1,2-Dichloroethane	22.0	0.50	ug/l	25.0		88	60-140			
1,1-Dichloroethene	23.7	1.0	ug/l	25.0		95	70-125			
cis-1,2-Dichloroethene	22.2	1.0	ug/l	25.0		89	70-125			
trans-1,2-Dichloroethene	23.3	1.0	ug/l	25.0		93	70-125			
1,2-Dichloropropane	23.0	1.0	ug/l	25.0		92	70-125			
1,3-Dichloropropane	23.4	1.0	ug/l	25.0		94	70-120			
2,2-Dichloropropane	23.1	1.0	ug/l	25.0		92	65-140			
1,1-Dichloropropene	24.4	1.0	ug/l	25.0		98	75-130			
cis-1,3-Dichloropropene	21.6	0.50	ug/l	25.0		86	75-125			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J23005 Extracted: 10/23/06										
LCS Analyzed: 10/23/2006 (6J23005-BS1)										
trans-1,3-Dichloropropene	21.6	0.50	ug/l	25.0		86	70-125			
Ethylbenzene	24.0	1.0	ug/l	25.0		96	75-125			
Hexachlorobutadiene	23.7	1.0	ug/l	25.0		95	65-135			
Isopropylbenzene	27.6	1.0	ug/l	25.0		110	75-130			
p-Isopropyltoluene	24.6	1.0	ug/l	25.0		98	75-125			
Methylene chloride	20.2	5.0	ug/l	25.0		81	55-130			
Naphthalene	22.6	1.0	ug/l	25.0		90	55-135			
n-Propylbenzene	25.6	1.0	ug/l	25.0		102	75-130			
Styrene	24.2	1.0	ug/l	25.0		97	75-130			
1,1,1,2-Tetrachloroethane	22.6	1.0	ug/l	25.0		90	70-130			
1,1,2,2-Tetrachloroethane	25.0	1.0	ug/l	25.0		100	55-130			
Tetrachloroethene	23.8	1.0	ug/l	25.0		95	70-125			
Toluene	23.4	1.0	ug/l	25.0		94	70-120			
1,2,3-Trichlorobenzene	24.4	1.0	ug/l	25.0		98	65-125			
1,2,4-Trichlorobenzene	25.0	1.0	ug/l	25.0		100	70-135			
1,1,1-Trichloroethane	21.5	1.0	ug/l	25.0		86	65-135			
1,1,2-Trichloroethane	23.2	1.0	ug/l	25.0		93	70-125			
Trichloroethene	23.8	1.0	ug/l	25.0		95	70-125			
Trichlorofluoromethane	24.8	1.0	ug/l	25.0		99	65-145			
1,2,3-Trichloropropane	25.5	1.0	ug/l	25.0		102	60-130			
1,2,4-Trimethylbenzene	24.0	1.0	ug/l	25.0		96	75-125			
1,3,5-Trimethylbenzene	24.6	1.0	ug/l	25.0		98	75-125			
Vinyl chloride	20.4	0.50	ug/l	25.0		82	55-135			
o-Xylene	23.5	1.0	ug/l	25.0		94	75-125			
m,p-Xylenes	48.8	1.0	ug/l	50.0		98	75-125			
Surrogate: Dibromofluoromethane	*24.4*		*ug/l*	*25.0*		*98*	*80-120*			
Surrogate: Toluene-d8	*27.4*		*ug/l*	*25.0*		*110*	*80-120*			
Surrogate: 4-Bromofluorobenzene	*26.1*		*ug/l*	*25.0*		*104*	*80-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J23005 Extracted: 10/23/06										
Matrix Spike Analyzed: 10/23/2006 (6J23005-MS1)					**Source: IPJ1792-01**					
Acetone	25.5	10	ug/l	25.0	ND	102	20-150			
Benzene	22.8	0.50	ug/l	25.0	ND	91	65-125			
Bromobenzene	22.8	1.0	ug/l	25.0	ND	91	70-125			
Bromochloromethane	20.8	1.0	ug/l	25.0	ND	83	65-135			
Bromodichloromethane	22.5	1.0	ug/l	25.0	ND	90	70-135			
Bromoform	19.7	1.0	ug/l	25.0	ND	79	55-135			
Bromomethane	26.4	1.0	ug/l	25.0	ND	106	55-145			
n-Butylbenzene	25.7	1.0	ug/l	25.0	ND	103	65-135			
sec-Butylbenzene	24.4	1.0	ug/l	25.0	ND	98	70-125			
tert-Butylbenzene	23.9	1.0	ug/l	25.0	ND	96	65-130			
Carbon tetrachloride	22.6	0.50	ug/l	25.0	ND	90	65-140			
Chlorobenzene	23.0	1.0	ug/l	25.0	ND	92	75-125			
Chloroethane	23.7	1.0	ug/l	25.0	ND	95	55-140			
Chloroform	20.6	1.0	ug/l	25.0	ND	82	65-135			
Chloromethane	25.2	1.0	ug/l	25.0	ND	101	45-145			
2-Chlorotoluene	23.3	1.0	ug/l	25.0	ND	93	65-135			
4-Chlorotoluene	23.6	1.0	ug/l	25.0	ND	94	70-135			
Dibromochloromethane	23.5	1.0	ug/l	25.0	ND	94	65-140			
1,2-Dibromo-3-chloropropane	21.2	5.0	ug/l	25.0	ND	85	45-145			
1,2-Dibromoethane (EDB)	22.4	1.0	ug/l	25.0	ND	90	70-130			
Dibromomethane	21.8	1.0	ug/l	25.0	ND	87	65-135			
1,2-Dichlorobenzene	22.8	1.0	ug/l	25.0	ND	91	75-125			
1,3-Dichlorobenzene	23.3	1.0	ug/l	25.0	ND	93	75-125			
1,4-Dichlorobenzene	22.8	1.0	ug/l	25.0	ND	91	75-125			
Dichlorodifluoromethane	29.5	5.0	ug/l	25.0	ND	118	25-155			
1,1-Dichloroethane	21.2	1.0	ug/l	25.0	ND	85	65-130			
1,2-Dichloroethane	21.8	0.50	ug/l	25.0	ND	87	60-140			
1,1-Dichloroethene	22.3	1.0	ug/l	25.0	ND	89	60-130			
cis-1,2-Dichloroethene	22.3	1.0	ug/l	25.0	0.41	88	65-130			
trans-1,2-Dichloroethene	22.9	1.0	ug/l	25.0	ND	92	65-130			
1,2-Dichloropropane	23.3	1.0	ug/l	25.0	ND	93	65-130			
1,3-Dichloropropane	22.7	1.0	ug/l	25.0	ND	91	65-135			
2,2-Dichloropropane	23.0	1.0	ug/l	25.0	ND	92	60-145			
1,1-Dichloropropene	24.0	1.0	ug/l	25.0	ND	96	70-135			
cis-1,3-Dichloropropene	21.5	0.50	ug/l	25.0	ND	86	70-130			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J23005 Extracted: 10/23/06										
Matrix Spike Analyzed: 10/23/2006 (6J23005-MS1)					**Source: IPJ1792-01**					
trans-1,3-Dichloropropene	21.3	0.50	ug/l	25.0	ND	85	65-135			
Ethylbenzene	24.2	1.0	ug/l	25.0	ND	97	65-130			
Hexachlorobutadiene	23.5	1.0	ug/l	25.0	ND	94	60-135			
Isopropylbenzene	27.2	1.0	ug/l	25.0	ND	109	70-135			
p-Isopropyltoluene	24.6	1.0	ug/l	25.0	ND	98	65-130			
Methylene chloride	19.7	5.0	ug/l	25.0	ND	79	50-135			
Naphthalene	21.1	1.0	ug/l	25.0	ND	84	50-140			
n-Propylbenzene	25.4	1.0	ug/l	25.0	ND	102	70-135			
Styrene	24.5	1.0	ug/l	25.0	ND	98	50-145			
1,1,1,2-Tetrachloroethane	22.8	1.0	ug/l	25.0	ND	91	65-140			
1,1,2,2-Tetrachloroethane	22.7	1.0	ug/l	25.0	ND	91	55-135			
Tetrachloroethene	26.2	1.0	ug/l	25.0	2.9	93	65-130			
Toluene	23.4	1.0	ug/l	25.0	ND	94	70-125			
1,2,3-Trichlorobenzene	24.4	1.0	ug/l	25.0	ND	98	60-135			
1,2,4-Trichlorobenzene	25.8	1.0	ug/l	25.0	ND	103	65-135			
1,1,1-Trichloroethane	21.4	1.0	ug/l	25.0	ND	86	65-140			
1,1,2-Trichloroethane	22.4	1.0	ug/l	25.0	ND	90	65-130			
Trichloroethene	66.3	1.0	ug/l	25.0	46	81	65-125			
Trichlorofluoromethane	22.6	1.0	ug/l	25.0	ND	90	60-145			
1,2,3-Trichloropropane	22.5	1.0	ug/l	25.0	ND	90	55-135			
1,2,4-Trimethylbenzene	23.9	1.0	ug/l	25.0	ND	96	55-135			
1,3,5-Trimethylbenzene	24.2	1.0	ug/l	25.0	ND	97	70-130			
Vinyl chloride	20.8	0.50	ug/l	25.0	ND	83	45-140			
o-Xylene	23.7	1.0	ug/l	25.0	ND	95	65-125			
m,p-Xylenes	49.1	1.0	ug/l	50.0	ND	98	65-130			
Surrogate: Dibromofluoromethane	*24.0*		*ug/l*	*25.0*		*96*	*80-120*			
Surrogate: Toluene-d8	*27.6*		*ug/l*	*25.0*		*110*	*80-120*			
Surrogate: 4-Bromofluorobenzene	*26.3*		*ug/l*	*25.0*		*105*	*80-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J23005 Extracted: 10/23/06										
Matrix Spike Dup Analyzed: 10/23/2006 (6J23005-MSD1)					**Source: IPJ1792-01**					
Acetone	26.3	10	ug/l	25.0	ND	105	20-150	3	35	
Benzene	22.6	0.50	ug/l	25.0	ND	90	65-125	1	20	
Bromobenzene	22.8	1.0	ug/l	25.0	ND	91	70-125	0	20	
Bromochloromethane	20.7	1.0	ug/l	25.0	ND	83	65-135	1	25	
Bromodichloromethane	22.0	1.0	ug/l	25.0	ND	88	70-135	2	20	
Bromoform	18.5	1.0	ug/l	25.0	ND	74	55-135	6	25	
Bromomethane	26.9	1.0	ug/l	25.0	ND	108	55-145	2	25	
n-Butylbenzene	26.0	1.0	ug/l	25.0	ND	104	65-135	1	20	
sec-Butylbenzene	24.7	1.0	ug/l	25.0	ND	99	70-125	1	20	
tert-Butylbenzene	24.2	1.0	ug/l	25.0	ND	97	65-130	1	20	
Carbon tetrachloride	22.5	0.50	ug/l	25.0	ND	90	65-140	0	25	
Chlorobenzene	22.5	1.0	ug/l	25.0	ND	90	75-125	2	20	
Chloroethane	24.1	1.0	ug/l	25.0	ND	96	55-140	2	25	
Chloroform	20.8	1.0	ug/l	25.0	ND	83	65-135	1	20	
Chloromethane	25.4	1.0	ug/l	25.0	ND	102	45-145	1	25	
2-Chlorotoluene	23.5	1.0	ug/l	25.0	ND	94	65-135	1	20	
4-Chlorotoluene	24.0	1.0	ug/l	25.0	ND	96	70-135	2	20	
Dibromochloromethane	22.5	1.0	ug/l	25.0	ND	90	65-140	4	25	
1,2-Dibromo-3-chloropropane	19.2	5.0	ug/l	25.0	ND	77	45-145	10	30	
1,2-Dibromoethane (EDB)	21.0	1.0	ug/l	25.0	ND	84	70-130	6	25	
Dibromomethane	20.9	1.0	ug/l	25.0	ND	84	65-135	4	25	
1,2-Dichlorobenzene	22.8	1.0	ug/l	25.0	ND	91	75-125	0	20	
1,3-Dichlorobenzene	23.6	1.0	ug/l	25.0	ND	94	75-125	1	20	
1,4-Dichlorobenzene	23.0	1.0	ug/l	25.0	ND	92	75-125	1	20	
Dichlorodifluoromethane	29.7	5.0	ug/l	25.0	ND	119	25-155	1	30	
1,1-Dichloroethane	21.5	1.0	ug/l	25.0	ND	86	65-130	1	20	
1,2-Dichloroethane	21.0	0.50	ug/l	25.0	ND	84	60-140	4	20	
1,1-Dichloroethene	22.4	1.0	ug/l	25.0	ND	90	60-130	0	20	
cis-1,2-Dichloroethene	22.3	1.0	ug/l	25.0	0.41	88	65-130	0	20	
trans-1,2-Dichloroethene	23.2	1.0	ug/l	25.0	ND	93	65-130	1	20	
1,2-Dichloropropane	22.9	1.0	ug/l	25.0	ND	92	65-130	2	20	
1,3-Dichloropropane	21.5	1.0	ug/l	25.0	ND	86	65-135	5	25	
2,2-Dichloropropane	23.4	1.0	ug/l	25.0	ND	94	60-145	2	25	
1,1-Dichloropropene	23.7	1.0	ug/l	25.0	ND	95	70-135	1	20	
cis-1,3-Dichloropropene	21.0	0.50	ug/l	25.0	ND	84	70-130	2	20	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J23005 Extracted: 10/23/06										
Matrix Spike Dup Analyzed: 10/23/2006 (6J23005-MSD1)					**Source: IPJ1792-01**					
trans-1,3-Dichloropropene	20.7	0.50	ug/l	25.0	ND	83	65-135	3	25	
Ethylbenzene	23.7	1.0	ug/l	25.0	ND	95	65-130	2	20	
Hexachlorobutadiene	24.0	1.0	ug/l	25.0	ND	96	60-135	2	20	
Isopropylbenzene	27.6	1.0	ug/l	25.0	ND	110	70-135	1	20	
p-Isopropyltoluene	24.8	1.0	ug/l	25.0	ND	99	65-130	1	20	
Methylene chloride	19.8	5.0	ug/l	25.0	ND	79	50-135	1	20	
Naphthalene	20.4	1.0	ug/l	25.0	ND	82	50-140	3	30	
n-Propylbenzene	25.6	1.0	ug/l	25.0	ND	102	70-135	1	20	
Styrene	23.8	1.0	ug/l	25.0	ND	95	50-145	3	30	
1,1,1,2-Tetrachloroethane	22.3	1.0	ug/l	25.0	ND	89	65-140	2	20	
1,1,2,2-Tetrachloroethane	21.2	1.0	ug/l	25.0	ND	85	55-135	7	30	
Tetrachloroethene	26.1	1.0	ug/l	25.0	2.9	93	65-130	0	20	
Toluene	23.1	1.0	ug/l	25.0	ND	92	70-125	1	20	
1,2,3-Trichlorobenzene	24.7	1.0	ug/l	25.0	ND	99	60-135	1	20	
1,2,4-Trichlorobenzene	25.8	1.0	ug/l	25.0	ND	103	65-135	0	20	
1,1,1-Trichloroethane	21.7	1.0	ug/l	25.0	ND	87	65-140	1	20	
1,1,2-Trichloroethane	21.3	1.0	ug/l	25.0	ND	85	65-130	5	25	
Trichloroethene	64.1	1.0	ug/l	25.0	46	72	65-125	3	20	
Trichlorofluoromethane	23.1	1.0	ug/l	25.0	ND	92	60-145	2	25	
1,2,3-Trichloropropane	21.3	1.0	ug/l	25.0	ND	85	55-135	5	30	
1,2,4-Trimethylbenzene	24.1	1.0	ug/l	25.0	ND	96	55-135	1	25	
1,3,5-Trimethylbenzene	24.7	1.0	ug/l	25.0	ND	99	70-130	2	20	
Vinyl chloride	21.7	0.50	ug/l	25.0	ND	87	45-140	4	30	
o-Xylene	23.2	1.0	ug/l	25.0	ND	93	65-125	2	20	
m,p-Xylenes	48.2	1.0	ug/l	50.0	ND	96	65-130	2	25	
Surrogate: Dibromofluoromethane	*24.1*		*ug/l*	*25.0*		*96*	*80-120*			
Surrogate: Toluene-d8	*27.6*		*ug/l*	*25.0*		*110*	*80-120*			
Surrogate: 4-Bromofluorobenzene	*26.0*		*ug/l*	*25.0*		*104*	*80-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J26012 Extracted: 10/26/06										
Blank Analyzed: 10/26/2006 (6J26012-BLK1)										
Acetone	ND	10	ug/l							
Benzene	ND	0.50	ug/l							
Bromobenzene	ND	1.0	ug/l							
Bromochloromethane	ND	1.0	ug/l							
Bromodichloromethane	ND	1.0	ug/l							
Bromoform	ND	1.0	ug/l							
Bromomethane	ND	1.0	ug/l							
n-Butylbenzene	ND	1.0	ug/l							
sec-Butylbenzene	ND	1.0	ug/l							
tert-Butylbenzene	ND	1.0	ug/l							
Carbon tetrachloride	ND	0.50	ug/l							
Chlorobenzene	ND	1.0	ug/l							
Chloroethane	ND	1.0	ug/l							
Chloroform	ND	1.0	ug/l							
Chloromethane	ND	1.0	ug/l							
2-Chlorotoluene	ND	1.0	ug/l							
4-Chlorotoluene	ND	1.0	ug/l							
Dibromochloromethane	ND	1.0	ug/l							
1,2-Dibromo-3-chloropropane	ND	5.0	ug/l							
1,2-Dibromoethane (EDB)	ND	1.0	ug/l							
Dibromomethane	ND	1.0	ug/l							
1,2-Dichlorobenzene	ND	1.0	ug/l							
1,3-Dichlorobenzene	ND	1.0	ug/l							
1,4-Dichlorobenzene	ND	1.0	ug/l							
Dichlorodifluoromethane	ND	5.0	ug/l							
1,1-Dichloroethane	ND	1.0	ug/l							
1,2-Dichloroethane	ND	0.50	ug/l							
1,1-Dichloroethene	ND	1.0	ug/l							
cis-1,2-Dichloroethene	ND	1.0	ug/l							
trans-1,2-Dichloroethene	ND	1.0	ug/l							
1,2-Dichloropropane	ND	1.0	ug/l							
1,3-Dichloropropane	ND	1.0	ug/l							
2,2-Dichloropropane	ND	1.0	ug/l							
1,1-Dichloropropene	ND	1.0	ug/l							
cis-1,3-Dichloropropene	ND	0.50	ug/l							

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J26012 Extracted: 10/26/06										
Blank Analyzed: 10/26/2006 (6J26012-BLK1)										
trans-1,3-Dichloropropene	ND	0.50	ug/l							
Ethylbenzene	ND	1.0	ug/l							
Hexachlorobutadiene	ND	1.0	ug/l							
Isopropylbenzene	ND	1.0	ug/l							
p-Isopropyltoluene	ND	1.0	ug/l							
Methylene chloride	ND	5.0	ug/l							
Naphthalene	ND	1.0	ug/l							
n-Propylbenzene	ND	1.0	ug/l							
Styrene	ND	1.0	ug/l							
1,1,1,2-Tetrachloroethane	ND	1.0	ug/l							
1,1,2,2-Tetrachloroethane	ND	1.0	ug/l							
Tetrachloroethene	ND	1.0	ug/l							
Toluene	ND	1.0	ug/l							
1,2,3-Trichlorobenzene	ND	1.0	ug/l							
1,2,4-Trichlorobenzene	ND	1.0	ug/l							
1,1,1-Trichloroethane	ND	1.0	ug/l							
1,1,2-Trichloroethane	ND	1.0	ug/l							
Trichloroethene	ND	1.0	ug/l							
Trichlorofluoromethane	ND	1.0	ug/l							
Trichlorotrifluoroethane (Freon 113)	ND	5.0	ug/l							
1,2,3-Trichloropropane	ND	1.0	ug/l							
1,2,4-Trimethylbenzene	ND	1.0	ug/l							
1,3,5-Trimethylbenzene	ND	1.0	ug/l							
Vinyl chloride	ND	0.50	ug/l							
o-Xylene	ND	1.0	ug/l							
m,p-Xylenes	ND	1.0	ug/l							
Surrogate: Dibromofluoromethane	*25.2*		*ug/l*	*25.0*		*101*	*80-120*			
Surrogate: Toluene-d8	*27.9*		*ug/l*	*25.0*		*112*	*80-120*			
Surrogate: 4-Bromofluorobenzene	*26.3*		*ug/l*	*25.0*		*105*	*80-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J26012 Extracted: 10/26/06										
LCS Analyzed: 10/26/2006 (6J26012-BS1)										
Acetone	26.8	10	ug/l	25.0		107	30-140			
Benzene	23.2	0.50	ug/l	25.0		93	70-120			
Bromobenzene	22.6	1.0	ug/l	25.0		90	75-120			
Bromochloromethane	23.1	1.0	ug/l	25.0		92	70-130			
Bromodichloromethane	26.8	1.0	ug/l	25.0		107	70-135			
Bromoform	21.5	1.0	ug/l	25.0		86	55-130			
Bromomethane	30.0	1.0	ug/l	25.0		120	65-140			
n-Butylbenzene	26.5	1.0	ug/l	25.0		106	70-130			
sec-Butylbenzene	25.1	1.0	ug/l	25.0		100	70-125			
tert-Butylbenzene	25.0	1.0	ug/l	25.0		100	70-125			
Carbon tetrachloride	28.2	0.50	ug/l	25.0		113	65-140			
Chlorobenzene	23.5	1.0	ug/l	25.0		94	75-120			
Chloroethane	26.1	1.0	ug/l	25.0		104	60-140			
Chloroform	25.4	1.0	ug/l	25.0		102	70-130			
Chloromethane	30.8	1.0	ug/l	25.0		123	50-140			
2-Chlorotoluene	24.5	1.0	ug/l	25.0		98	70-125			
4-Chlorotoluene	25.1	1.0	ug/l	25.0		100	75-125			
Dibromochloromethane	26.5	1.0	ug/l	25.0		106	70-140			
1,2-Dibromo-3-chloropropane	24.1	5.0	ug/l	25.0		96	50-135			
1,2-Dibromoethane (EDB)	23.3	1.0	ug/l	25.0		93	75-125			
Dibromomethane	24.7	1.0	ug/l	25.0		99	70-125			
1,2-Dichlorobenzene	23.9	1.0	ug/l	25.0		96	75-120			
1,3-Dichlorobenzene	24.2	1.0	ug/l	25.0		97	75-120			
1,4-Dichlorobenzene	23.9	1.0	ug/l	25.0		96	75-120			
Dichlorodifluoromethane	48.2	5.0	ug/l	25.0		193	35-155			*L*
1,1-Dichloroethane	24.4	1.0	ug/l	25.0		98	70-125			
1,2-Dichloroethane	27.9	0.50	ug/l	25.0		112	60-140			
1,1-Dichloroethene	24.6	1.0	ug/l	25.0		98	70-125			
cis-1,2-Dichloroethene	23.8	1.0	ug/l	25.0		95	70-125			
trans-1,2-Dichloroethene	24.0	1.0	ug/l	25.0		96	70-125			
1,2-Dichloropropane	23.4	1.0	ug/l	25.0		94	70-125			
1,3-Dichloropropane	24.2	1.0	ug/l	25.0		97	70-120			
2,2-Dichloropropane	29.3	1.0	ug/l	25.0		117	65-140			
1,1-Dichloropropene	25.7	1.0	ug/l	25.0		103	75-130			
cis-1,3-Dichloropropene	22.7	0.50	ug/l	25.0		91	75-125			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J26012 Extracted: 10/26/06										
LCS Analyzed: 10/26/2006 (6J26012-BS1)										
trans-1,3-Dichloropropene	24.0	0.50	ug/l	25.0		96	70-125			
Ethylbenzene	25.2	1.0	ug/l	25.0		101	75-125			
Hexachlorobutadiene	26.0	1.0	ug/l	25.0		104	65-135			
Isopropylbenzene	27.6	1.0	ug/l	25.0		110	75-130			
p-Isopropyltoluene	25.6	1.0	ug/l	25.0		102	75-125			
Methylene chloride	22.5	5.0	ug/l	25.0		90	55-130			
Naphthalene	23.0	1.0	ug/l	25.0		92	55-135			
n-Propylbenzene	25.7	1.0	ug/l	25.0		103	75-130			
Styrene	24.9	1.0	ug/l	25.0		100	75-130			
1,1,1,2-Tetrachloroethane	25.7	1.0	ug/l	25.0		103	70-130			
1,1,2,2-Tetrachloroethane	22.7	1.0	ug/l	25.0		91	55-130			
Tetrachloroethene	24.3	1.0	ug/l	25.0		97	70-125			
Toluene	24.1	1.0	ug/l	25.0		96	70-120			
1,2,3-Trichlorobenzene	26.6	1.0	ug/l	25.0		106	65-125			
1,2,4-Trichlorobenzene	27.4	1.0	ug/l	25.0		110	70-135			
1,1,1-Trichloroethane	27.3	1.0	ug/l	25.0		109	65-135			
1,1,2-Trichloroethane	23.7	1.0	ug/l	25.0		95	70-125			
Trichloroethene	24.5	1.0	ug/l	25.0		98	70-125			
Trichlorofluoromethane	32.6	1.0	ug/l	25.0		130	65-145			
1,2,3-Trichloropropane	23.8	1.0	ug/l	25.0		95	60-130			
1,2,4-Trimethylbenzene	25.1	1.0	ug/l	25.0		100	75-125			
1,3,5-Trimethylbenzene	25.4	1.0	ug/l	25.0		102	75-125			
Vinyl chloride	29.0	0.50	ug/l	25.0		116	55-135			
o-Xylene	24.8	1.0	ug/l	25.0		99	75-125			
m,p-Xylenes	50.0	1.0	ug/l	50.0		100	75-125			
Surrogate: Dibromofluoromethane	*28.3*		*ug/l*	*25.0*		*113*	*80-120*			
Surrogate: Toluene-d8	*28.0*		*ug/l*	*25.0*		*112*	*80-120*			
Surrogate: 4-Bromofluorobenzene	*27.9*		*ug/l*	*25.0*		*112*	*80-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J26012 Extracted: 10/26/06										
Matrix Spike Analyzed: 10/26/2006 (6J26012-MS1)					**Source: IPJ1633-02**					
Acetone	18.7	10	ug/l	25.0	ND	75	20-150			
Benzene	24.2	0.50	ug/l	25.0	ND	97	65-125			
Bromobenzene	23.6	1.0	ug/l	25.0	ND	94	70-125			
Bromochloromethane	24.3	1.0	ug/l	25.0	ND	97	65-135			
Bromodichloromethane	28.5	1.0	ug/l	25.0	ND	114	70-135			
Bromoform	23.5	1.0	ug/l	25.0	ND	94	55-135			
Bromomethane	30.2	1.0	ug/l	25.0	ND	121	55-145			
n-Butylbenzene	27.3	1.0	ug/l	25.0	ND	109	65-135			
sec-Butylbenzene	25.4	1.0	ug/l	25.0	ND	102	70-125			
tert-Butylbenzene	25.2	1.0	ug/l	25.0	ND	101	65-130			
Carbon tetrachloride	31.4	0.50	ug/l	25.0	2.4	116	65-140			
Chlorobenzene	24.3	1.0	ug/l	25.0	ND	97	75-125			
Chloroethane	26.5	1.0	ug/l	25.0	ND	106	55-140			
Chloroform	29.6	1.0	ug/l	25.0	3.3	105	65-135			
Chloromethane	30.0	1.0	ug/l	25.0	ND	120	45-145			
2-Chlorotoluene	24.8	1.0	ug/l	25.0	ND	99	65-135			
4-Chlorotoluene	25.6	1.0	ug/l	25.0	ND	102	70-135			
Dibromochloromethane	28.7	1.0	ug/l	25.0	ND	115	65-140			
1,2-Dibromo-3-chloropropane	26.6	5.0	ug/l	25.0	ND	106	45-145			
1,2-Dibromoethane (EDB)	25.6	1.0	ug/l	25.0	ND	102	70-130			
Dibromomethane	27.2	1.0	ug/l	25.0	ND	109	65-135			
1,2-Dichlorobenzene	24.8	1.0	ug/l	25.0	ND	99	75-125			
1,3-Dichlorobenzene	25.1	1.0	ug/l	25.0	ND	100	75-125			
1,4-Dichlorobenzene	24.6	1.0	ug/l	25.0	ND	98	75-125			
Dichlorodifluoromethane	48.8	5.0	ug/l	25.0	ND	195	25-155			*M7*
1,1-Dichloroethane	25.0	1.0	ug/l	25.0	ND	100	65-130			
1,2-Dichloroethane	30.6	0.50	ug/l	25.0	ND	122	60-140			
1,1-Dichloroethene	28.3	1.0	ug/l	25.0	3.4	100	60-130			
cis-1,2-Dichloroethene	25.2	1.0	ug/l	25.0	0.73	98	65-130			
trans-1,2-Dichloroethene	24.4	1.0	ug/l	25.0	ND	98	65-130			
1,2-Dichloropropane	24.8	1.0	ug/l	25.0	ND	99	65-130			
1,3-Dichloropropane	26.1	1.0	ug/l	25.0	ND	104	65-135			
2,2-Dichloropropane	29.0	1.0	ug/l	25.0	ND	116	60-145			
1,1-Dichloropropene	26.3	1.0	ug/l	25.0	ND	105	70-135			
cis-1,3-Dichloropropene	24.5	0.50	ug/l	25.0	ND	98	70-130			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J26012 Extracted: 10/26/06										
Matrix Spike Analyzed: 10/26/2006 (6J26012-MS1)					**Source: IPJ1633-02**					
trans-1,3-Dichloropropene	26.0	0.50	ug/l	25.0	ND	104	65-135			
Ethylbenzene	26.2	1.0	ug/l	25.0	ND	105	65-130			
Hexachlorobutadiene	26.4	1.0	ug/l	25.0	ND	106	60-135			
Isopropylbenzene	27.6	1.0	ug/l	25.0	ND	110	70-135			
p-Isopropyltoluene	26.0	1.0	ug/l	25.0	ND	104	65-130			
Methylene chloride	23.0	5.0	ug/l	25.0	ND	92	50-135			
Naphthalene	24.3	1.0	ug/l	25.0	ND	97	50-140			
n-Propylbenzene	26.2	1.0	ug/l	25.0	ND	105	70-135			
Styrene	26.3	1.0	ug/l	25.0	ND	105	50-145			
1,1,1,2-Tetrachloroethane	26.9	1.0	ug/l	25.0	ND	108	65-140			
1,1,2,2-Tetrachloroethane	24.8	1.0	ug/l	25.0	ND	99	55-135			
Tetrachloroethene	32.9	1.0	ug/l	25.0	8.2	99	65-130			
Toluene	25.2	1.0	ug/l	25.0	ND	101	70-125			
1,2,3-Trichlorobenzene	28.1	1.0	ug/l	25.0	ND	112	60-135			
1,2,4-Trichlorobenzene	28.8	1.0	ug/l	25.0	ND	115	65-135			
1,1,1-Trichloroethane	27.2	1.0	ug/l	25.0	ND	109	65-140			
1,1,2-Trichloroethane	26.1	1.0	ug/l	25.0	ND	104	65-130			
Trichloroethene	30.1	1.0	ug/l	25.0	4.5	102	65-125			
Trichlorofluoromethane	34.7	1.0	ug/l	25.0	1.8	132	60-145			
1,2,3-Trichloropropane	25.6	1.0	ug/l	25.0	ND	102	55-135			
1,2,4-Trimethylbenzene	25.6	1.0	ug/l	25.0	ND	102	55-135			
1,3,5-Trimethylbenzene	25.6	1.0	ug/l	25.0	ND	102	70-130			
Vinyl chloride	28.8	0.50	ug/l	25.0	ND	115	45-140			
o-Xylene	25.5	1.0	ug/l	25.0	ND	102	65-125			
m,p-Xylenes	51.8	1.0	ug/l	50.0	ND	104	65-130			
Surrogate: Dibromofluoromethane	*27.7*		*ug/l*	*25.0*		*111*	*80-120*			
Surrogate: Toluene-d8	*28.4*		*ug/l*	*25.0*		*114*	*80-120*			
Surrogate: 4-Bromofluorobenzene	*29.1*		*ug/l*	*25.0*		*116*	*80-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J26012 Extracted: 10/26/06										
Matrix Spike Dup Analyzed: 10/26/2006 (6J26012-MSD1)					**Source: IPJ1633-02**					
Acetone	18.4	10	ug/l	25.0	ND	74	20-150	2	35	
Benzene	25.6	0.50	ug/l	25.0	ND	102	65-125	6	20	
Bromobenzene	25.1	1.0	ug/l	25.0	ND	100	70-125	6	20	
Bromochloromethane	25.8	1.0	ug/l	25.0	ND	103	65-135	6	25	
Bromodichloromethane	29.0	1.0	ug/l	25.0	ND	116	70-135	2	20	
Bromoform	23.5	1.0	ug/l	25.0	ND	94	55-135	0	25	
Bromomethane	33.4	1.0	ug/l	25.0	ND	134	55-145	10	25	
n-Butylbenzene	28.8	1.0	ug/l	25.0	ND	115	65-135	5	20	
sec-Butylbenzene	27.4	1.0	ug/l	25.0	ND	110	70-125	8	20	
tert-Butylbenzene	27.3	1.0	ug/l	25.0	ND	109	65-130	8	20	
Carbon tetrachloride	33.1	0.50	ug/l	25.0	2.4	123	65-140	5	25	
Chlorobenzene	25.9	1.0	ug/l	25.0	ND	104	75-125	6	20	
Chloroethane	30.2	1.0	ug/l	25.0	ND	121	55-140	13	25	
Chloroform	31.4	1.0	ug/l	25.0	3.3	112	65-135	6	20	
Chloromethane	35.3	1.0	ug/l	25.0	ND	141	45-145	16	25	
2-Chlorotoluene	26.7	1.0	ug/l	25.0	ND	107	65-135	7	20	
4-Chlorotoluene	27.4	1.0	ug/l	25.0	ND	110	70-135	7	20	
Dibromochloromethane	28.6	1.0	ug/l	25.0	ND	114	65-140	0	25	
1,2-Dibromo-3-chloropropane	26.0	5.0	ug/l	25.0	ND	104	45-145	2	30	
1,2-Dibromoethane (EDB)	25.5	1.0	ug/l	25.0	ND	102	70-130	0	25	
Dibromomethane	26.5	1.0	ug/l	25.0	ND	106	65-135	3	25	
1,2-Dichlorobenzene	25.9	1.0	ug/l	25.0	ND	104	75-125	4	20	
1,3-Dichlorobenzene	26.4	1.0	ug/l	25.0	ND	106	75-125	5	20	
1,4-Dichlorobenzene	25.4	1.0	ug/l	25.0	ND	102	75-125	3	20	
Dichlorodifluoromethane	54.0	5.0	ug/l	25.0	ND	216	25-155	10	30	*M7*
1,1-Dichloroethane	27.3	1.0	ug/l	25.0	ND	109	65-130	9	20	
1,2-Dichloroethane	30.3	0.50	ug/l	25.0	ND	121	60-140	1	20	
1,1-Dichloroethene	31.1	1.0	ug/l	25.0	3.4	111	60-130	9	20	
cis-1,2-Dichloroethene	27.7	1.0	ug/l	25.0	0.73	108	65-130	9	20	
trans-1,2-Dichloroethene	27.4	1.0	ug/l	25.0	ND	110	65-130	12	20	
1,2-Dichloropropane	26.0	1.0	ug/l	25.0	ND	104	65-130	5	20	
1,3-Dichloropropane	26.4	1.0	ug/l	25.0	ND	106	65-135	1	25	
2,2-Dichloropropane	31.9	1.0	ug/l	25.0	ND	128	60-145	10	25	
1,1-Dichloropropene	28.1	1.0	ug/l	25.0	ND	112	70-135	7	20	
cis-1,3-Dichloropropene	24.9	0.50	ug/l	25.0	ND	100	70-130	2	20	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

VOLATILE ORGANICS by GC/MS (EPA 5030B/8260B)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J26012 Extracted: 10/26/06										
Matrix Spike Dup Analyzed: 10/26/2006 (6J26012-MSD1)					**Source: IPJ1633-02**					
trans-1,3-Dichloropropene	26.3	0.50	ug/l	25.0	ND	105	65-135	1	25	
Ethylbenzene	27.6	1.0	ug/l	25.0	ND	110	65-130	5	20	
Hexachlorobutadiene	27.3	1.0	ug/l	25.0	ND	109	60-135	3	20	
Isopropylbenzene	30.3	1.0	ug/l	25.0	ND	121	70-135	9	20	
p-Isopropyltoluene	27.8	1.0	ug/l	25.0	ND	111	65-130	7	20	
Methylene chloride	25.1	5.0	ug/l	25.0	ND	100	50-135	9	20	
Naphthalene	24.0	1.0	ug/l	25.0	ND	96	50-140	1	30	
n-Propylbenzene	28.1	1.0	ug/l	25.0	ND	112	70-135	7	20	
Styrene	27.4	1.0	ug/l	25.0	ND	110	50-145	4	30	
1,1,1,2-Tetrachloroethane	28.0	1.0	ug/l	25.0	ND	112	65-140	4	20	
1,1,2,2-Tetrachloroethane	25.1	1.0	ug/l	25.0	ND	100	55-135	1	30	
Tetrachloroethene	34.6	1.0	ug/l	25.0	8.2	106	65-130	5	20	
Toluene	26.7	1.0	ug/l	25.0	ND	107	70-125	6	20	
1,2,3-Trichlorobenzene	28.1	1.0	ug/l	25.0	ND	112	60-135	0	20	
1,2,4-Trichlorobenzene	28.8	1.0	ug/l	25.0	ND	115	65-135	0	20	
1,1,1-Trichloroethane	29.6	1.0	ug/l	25.0	ND	118	65-140	8	20	
1,1,2-Trichloroethane	25.9	1.0	ug/l	25.0	ND	104	65-130	1	25	
Trichloroethene	31.3	1.0	ug/l	25.0	4.5	107	65-125	4	20	
Trichlorofluoromethane	36.9	1.0	ug/l	25.0	1.8	140	60-145	6	25	
1,2,3-Trichloropropane	25.6	1.0	ug/l	25.0	ND	102	55-135	0	30	
1,2,4-Trimethylbenzene	27.2	1.0	ug/l	25.0	ND	109	55-135	6	25	
1,3,5-Trimethylbenzene	27.5	1.0	ug/l	25.0	ND	110	70-130	7	20	
Vinyl chloride	33.6	0.50	ug/l	25.0	ND	134	45-140	15	30	
o-Xylene	27.0	1.0	ug/l	25.0	ND	108	65-125	6	20	
m,p-Xylenes	54.6	1.0	ug/l	50.0	ND	109	65-130	5	25	
Surrogate: Dibromofluoromethane	*28.4*		*ug/l*	*25.0*		*114*	*80-120*			
Surrogate: Toluene-d8	*27.9*		*ug/l*	*25.0*		*112*	*80-120*			
Surrogate: 4-Bromofluorobenzene	*28.3*		*ug/l*	*25.0*		*113*	*80-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

SEMI-VOLATILE ORGANICS BY GC/MS (EPA 3520C/8270C MOD)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J19073 Extracted: 10/19/06										
Blank Analyzed: 10/20/2006 (6J19073-BLK1)										
1,4-Dioxane	ND	0.50	ug/l							
Surrogate: 1,4-Dioxane-d8	*1.24*		*ug/l*	*2.00*		*62*	*30-120*			
LCS Analyzed: 10/20/2006 (6J19073-BS1)										**M-NR1**
1,4-Dioxane	1.11	0.50	ug/l	2.00		56	35-120			
Surrogate: 1,4-Dioxane-d8	*1.34*		*ug/l*	*2.00*		*67*	*30-120*			
LCS Dup Analyzed: 10/20/2006 (6J19073-BSD1)										
1,4-Dioxane	1.06	0.50	ug/l	2.00		53	35-120	5	25	
Surrogate: 1,4-Dioxane-d8	*1.29*		*ug/l*	*2.00*		*64*	*30-120*			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

SEMI-VOL ORGANICS by GC/MS-CHEMICAL IONIZATION (EPA 3520C/1625C-CI MOD)

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J19073 Extracted: 10/19/06										
Blank Analyzed: 10/20/2006 (6J19073-BLK1)										
N-Nitrosodimethylamine	ND	0.0020	ug/l							
LCS Analyzed: 10/20/2006 (6J19073-BS1)										**M-NR1**
N-Nitrosodimethylamine	0.00961	0.0020	ug/l	0.0100		96	70-130			
LCS Analyzed: 10/20/2006 (6J19073-BS2)										
N-Nitrosodimethylamine	0.00236	0.0020	ug/l	0.00200		118	70-130			
LCS Dup Analyzed: 10/20/2006 (6J19073-BSD1)										
N-Nitrosodimethylamine	0.00969	0.0020	ug/l	0.0100		97	70-130	1	20	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

METALS

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J19085 Extracted: 10/19/06										
Blank Analyzed: 10/22/2006-10/24/2006 (6J19085-BLK1)										
Aluminum	ND	0.050	mg/l							
Antimony	ND	0.010	mg/l							
Arsenic	ND	0.010	mg/l							
Barium	ND	0.010	mg/l							
Beryllium	ND	0.0040	mg/l							
Cadmium	ND	0.0050	mg/l							
Calcium	ND	0.10	mg/l							
Chromium	ND	0.0050	mg/l							
Cobalt	ND	0.010	mg/l							
Copper	ND	0.010	mg/l							
Iron	ND	0.040	mg/l							
Lead	ND	0.0050	mg/l							
Magnesium	ND	0.020	mg/l							
Manganese	ND	0.020	mg/l							
Molybdenum	ND	0.020	mg/l							
Nickel	ND	0.010	mg/l							
Potassium	ND	0.50	mg/l							
Selenium	ND	0.010	mg/l							
Silver	ND	0.010	mg/l							
Sodium	ND	0.50	mg/l							
Thallium	ND	0.010	mg/l							
Vanadium	ND	0.010	mg/l							
Zinc	ND	0.020	mg/l							
LCS Analyzed: 10/22/2006-10/24/2006 (6J19085-BS1)										
Aluminum	0.941	0.050	mg/l	1.00		94	80-120			
Antimony	0.967	0.010	mg/l	1.00		97	80-120			
Arsenic	0.986	0.010	mg/l	1.00		99	80-120			
Barium	0.949	0.010	mg/l	1.00		95	80-120			
Beryllium	0.990	0.0040	mg/l	1.00		99	80-120			
Cadmium	0.950	0.0050	mg/l	1.00		95	80-120			
Calcium	4.96	0.10	mg/l	5.00		99	80-120			
Chromium	0.981	0.0050	mg/l	1.00		98	80-120			
Cobalt	1.01	0.010	mg/l	1.00		101	80-120			
Copper	0.992	0.010	mg/l	1.00		99	80-120			
Iron	0.962	0.040	mg/l	1.00		96	80-120			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

METALS

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J19085 Extracted: 10/19/06										
LCS Analyzed: 10/22/2006-10/24/2006 (6J19085-BS1)										
Lead	0.970	0.0050	mg/l	1.00		97	80-120			
Magnesium	5.00	0.020	mg/l	5.00		100	80-120			
Manganese	0.963	0.020	mg/l	1.00		96	80-120			
Molybdenum	0.956	0.020	mg/l	1.00		96	80-120			
Nickel	0.963	0.010	mg/l	1.00		96	80-120			
Potassium	9.52	0.50	mg/l	10.0		95	80-120			
Selenium	0.944	0.010	mg/l	1.00		94	80-120			
Silver	0.517	0.010	mg/l	0.500		103	80-120			
Sodium	9.46	0.50	mg/l	10.0		95	80-120			
Thallium	0.967	0.010	mg/l	1.00		97	80-120			
Vanadium	0.989	0.010	mg/l	1.00		99	80-120			
Zinc	0.979	0.020	mg/l	1.00		98	80-120			
Matrix Spike Analyzed: 10/22/2006-10/24/2006 (6J19085-MS1)					**Source: IPJ1844-01**					
Aluminum	0.819	0.050	mg/l	1.00	ND	82	75-125			
Antimony	0.923	0.010	mg/l	1.00	ND	92	75-125			
Arsenic	0.964	0.010	mg/l	1.00	ND	96	75-125			
Barium	0.930	0.010	mg/l	1.00	0.016	91	75-125			
Beryllium	0.973	0.0040	mg/l	1.00	ND	97	75-125			
Cadmium	0.892	0.0050	mg/l	1.00	ND	89	75-125			
Calcium	282	0.10	mg/l	5.00	280	40	75-125			*M-HA*
Chromium	0.937	0.0050	mg/l	1.00	ND	94	75-125			
Cobalt	0.936	0.010	mg/l	1.00	ND	94	75-125			
Copper	0.989	0.010	mg/l	1.00	ND	99	75-125			
Iron	0.953	0.040	mg/l	1.00	0.046	91	75-125			
Lead	0.910	0.0050	mg/l	1.00	ND	91	75-125			
Magnesium	187	0.020	mg/l	5.00	180	140	75-125			*M-HA*
Manganese	1.15	0.040	mg/l	1.00	0.094	106	75-125			
Molybdenum	0.930	0.020	mg/l	1.00	0.0055	92	75-125			
Nickel	0.904	0.010	mg/l	1.00	0.0024	90	75-125			
Potassium	18.1	0.50	mg/l	10.0	7.1	110	75-125			
Selenium	0.893	0.010	mg/l	1.00	ND	89	75-125			
Silver	0.528	0.010	mg/l	0.500	ND	106	75-125			
Sodium	656	1.0	mg/l	10.0	610	460	75-125			*M-HA*
Thallium	0.860	0.010	mg/l	1.00	ND	86	75-125			
Vanadium	0.968	0.010	mg/l	1.00	ND	97	75-125			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

METALS

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J19085 Extracted: 10/19/06										
Matrix Spike Analyzed: 10/22/2006-10/24/2006 (6J19085-MS1)					**Source: IPJ1844-01**					
Zinc	0.926	0.020	mg/l	1.00	ND	93	75-125			
Matrix Spike Dup Analyzed: 10/22/2006-10/24/2006 (6J19085-MSD1)					**Source: IPJ1844-01**					
Aluminum	0.843	0.050	mg/l	1.00	ND	84	75-125	3	20	
Antimony	0.931	0.010	mg/l	1.00	ND	93	75-125	1	20	
Arsenic	0.979	0.010	mg/l	1.00	ND	98	75-125	2	20	
Barium	0.959	0.010	mg/l	1.00	0.016	94	75-125	3	20	
Beryllium	0.962	0.0040	mg/l	1.00	ND	96	75-125	1	20	
Cadmium	0.929	0.0050	mg/l	1.00	ND	93	75-125	4	20	
Calcium	292	0.10	mg/l	5.00	280	240	75-125	3	20	*M-HA*
Chromium	0.954	0.0050	mg/l	1.00	ND	95	75-125	2	20	
Cobalt	0.978	0.010	mg/l	1.00	ND	98	75-125	4	20	
Copper	1.03	0.010	mg/l	1.00	ND	103	75-125	4	20	
Iron	0.993	0.040	mg/l	1.00	0.046	95	75-125	4	20	
Lead	0.924	0.0050	mg/l	1.00	ND	92	75-125	2	20	
Magnesium	191	0.020	mg/l	5.00	180	220	75-125	2	20	*M-HA*
Manganese	1.18	0.040	mg/l	1.00	0.094	109	75-125	3	20	
Molybdenum	0.945	0.020	mg/l	1.00	0.0055	94	75-125	2	20	
Nickel	0.922	0.010	mg/l	1.00	0.0024	92	75-125	2	20	
Potassium	18.4	0.50	mg/l	10.0	7.1	113	75-125	2	20	
Selenium	0.910	0.010	mg/l	1.00	ND	91	75-125	2	20	
Silver	0.544	0.010	mg/l	0.500	ND	109	75-125	3	20	
Sodium	698	1.0	mg/l	10.0	610	880	75-125	6	20	*M-HA*
Thallium	0.879	0.010	mg/l	1.00	ND	88	75-125	2	20	
Vanadium	1.01	0.010	mg/l	1.00	ND	101	75-125	4	20	
Zinc	0.970	0.020	mg/l	1.00	ND	97	75-125	5	20	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

METALS

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J24066 Extracted: 10/24/06										
Blank Analyzed: 10/24/2006 (6J24066-BLK1)										
Mercury	ND	0.00020	mg/l							
LCS Analyzed: 10/24/2006 (6J24066-BS1)										
Mercury	0.00825	0.00020	mg/l	0.00800		103	90-115			
Matrix Spike Analyzed: 10/24/2006 (6J24066-MS1)					**Source: IPJ2245-01**					
Mercury	0.00782	0.00020	mg/l	0.00800	ND	98	75-120			
Matrix Spike Dup Analyzed: 10/24/2006 (6J24066-MSD1)					**Source: IPJ2245-01**					
Mercury	0.00779	0.00020	mg/l	0.00800	ND	97	75-120	0	20	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

DISSOLVED METALS

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J18110 Extracted: 10/18/06										
Blank Analyzed: 10/21/2006-10/23/2006 (6J18110-BLK1)										
Iron	ND	0.040	mg/l							
Manganese	ND	0.020	mg/l							
LCS Analyzed: 10/21/2006-10/22/2006 (6J18110-BS1)										
Iron	1.01	0.040	mg/l	1.00		101	80-120			
Manganese	0.980	0.020	mg/l	1.00		98	80-120			
Matrix Spike Analyzed: 10/21/2006 (6J18110-MS1)					**Source: IPJ1647-01**					
Iron	1.02	0.040	mg/l	1.00	ND	102	75-125			
Manganese	0.986	0.020	mg/l	1.00	ND	99	75-125			
Matrix Spike Dup Analyzed: 10/21/2006 (6J18110-MSD1)					**Source: IPJ1647-01**					
Iron	1.01	0.040	mg/l	1.00	ND	101	75-125	1	20	
Manganese	0.994	0.020	mg/l	1.00	ND	99	75-125	1	20	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J17032 Extracted: 10/17/06										
Blank Analyzed: 10/17/2006 (6J17032-BLK1)										
Fluoride	ND	0.50	mg/l							
Nitrite-N	ND	0.15	mg/l							
LCS Analyzed: 10/17/2006 (6J17032-BS1)										
Fluoride	4.99	0.50	mg/l	5.00		100	90-110			
Nitrite-N	1.52	0.15	mg/l	1.52		100	90-110			
Matrix Spike Analyzed: 10/17/2006 (6J17032-MS1)					**Source: IPJ1526-02**					
Fluoride	5.49	0.50	mg/l	5.00	0.53	99	80-120			
Nitrite-N	2.31	0.15	mg/l	1.52	0.76	102	80-120			
Matrix Spike Dup Analyzed: 10/17/2006 (6J17032-MSD1)					**Source: IPJ1526-02**					
Fluoride	5.46	0.50	mg/l	5.00	0.53	99	80-120	1	20	
Nitrite-N	2.28	0.15	mg/l	1.52	0.76	100	80-120	1	20	
Batch: 6J17033 Extracted: 10/17/06										
Blank Analyzed: 10/17/2006 (6J17033-BLK1)										
Chloride	ND	0.50	mg/l							
Nitrate-N	ND	0.11	mg/l							
Sulfate	ND	0.50	mg/l							
LCS Analyzed: 10/17/2006 (6J17033-BS1)										
Chloride	4.57	0.50	mg/l	5.00		91	90-110			*M-3*
Nitrate-N	1.21	0.11	mg/l	1.13		107	90-110			
Sulfate	10.2	0.50	mg/l	10.0		102	90-110			

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J17033 Extracted: 10/17/06										
Matrix Spike Analyzed: 10/17/2006 (6J17033-MS1)					**Source: IPJ1556-01**					
Nitrate-N	1.24	0.11	mg/l	1.13	0.071	103	80-120			
Sulfate	24.2	0.50	mg/l	10.0	14	102	80-120			
Matrix Spike Dup Analyzed: 10/17/2006 (6J17033-MSD1)					**Source: IPJ1556-01**					
Nitrate-N	1.22	0.11	mg/l	1.13	0.071	102	80-120	2	20	
Sulfate	24.2	0.50	mg/l	10.0	14	102	80-120	0	20	
Batch: 6J17132 Extracted: 10/17/06										
Blank Analyzed: 10/17/2006 (6J17132-BLK1)										
Chromium VI	ND	0.0020	mg/l							
LCS Analyzed: 10/17/2006 (6J17132-BS1)										
Chromium VI	0.0473	0.0020	mg/l	0.0500		95	90-110			
Matrix Spike Analyzed: 10/17/2006 (6J17132-MS1)					**Source: IPJ1647-01**					
Chromium VI	0.0532	0.0020	mg/l	0.0500	0.0049	97	80-115			
Matrix Spike Dup Analyzed: 10/17/2006 (6J17132-MSD1)					**Source: IPJ1647-01**					
Chromium VI	0.0533	0.0020	mg/l	0.0500	0.0049	97	80-115	0	15	
Batch: 6J17135 Extracted: 10/17/06										
Duplicate Analyzed: 10/17/2006 (6J17135-DUP1)					**Source: IPJ1647-01**					
Dissolved Oxygen	4.93	1.0	mg/l		5.0			1	20	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

INORGANICS

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: 6J20100 Extracted: 10/20/06										
Blank Analyzed: 10/20/2006 (6J20100-BLK1)										
Perchlorate	ND	4.0	ug/l							
LCS Analyzed: 10/20/2006 (6J20100-BS1)										
Perchlorate	51.9	4.0	ug/l	50.0		104	85-115			
Matrix Spike Analyzed: 10/20/2006 (6J20100-MS1)					**Source: IPJ1336-01**					
Perchlorate	50.4	4.0	ug/l	50.0	0.90	99	80-120			
Matrix Spike Dup Analyzed: 10/20/2006 (6J20100-MSD1)					**Source: IPJ1336-01**					
Perchlorate	53.4	4.0	ug/l	50.0	0.90	105	80-120	6	20	
Batch: 6J24096 Extracted: 10/24/06										
Blank Analyzed: 10/24/2006 (6J24096-BLK1)										
Sulfide	ND	1.0	mg/l							
LCS Analyzed: 10/24/2006 (6J24096-BS1)										
Sulfide	9.63	1.0	mg/l	10.4		93	80-120			
Matrix Spike Analyzed: 10/24/2006 (6J24096-MS1)					**Source: IPJ1647-03**					
Sulfide	7.23	1.0	mg/l	10.4	ND	70	70-130			
Matrix Spike Dup Analyzed: 10/24/2006 (6J24096-MSD1)					**Source: IPJ1647-03**					
Sulfide	7.23	1.0	mg/l	10.4	ND	70	70-130	0	30	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

METHOD BLANK/QC DATA

CDHS SRL 524 MODIFIED METHOD FOR 1,2,3-TRICHLOROPROPANE PT/GCMS

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Data Qualifiers
Batch: C6J1909 Extracted: 10/19/06										
Blank Analyzed: 10/19/2006 (C6J1909-BLK1)										
1,2,3-Trichloropropane (123-TCP)	ND	5.0	ng/L							
LCS Analyzed: 10/19/2006 (C6J1909-BS1)										
1,2,3-Trichloropropane (123-TCP)	5.65	5.0	ng/L	5.00		113	80-120			
Matrix Spike Analyzed: 10/19/2006 (C6J1909-MS1)					**Source: CPJ0616-02**					
1,2,3-Trichloropropane (123-TCP)	54.0	5.0	ng/L	50.0	ND	108	80-120			
Matrix Spike Dup Analyzed: 10/19/2006 (C6J1909-MSD1)					**Source: CPJ0616-02**					
1,2,3-Trichloropropane (123-TCP)	49.7	5.0	ng/L	50.0	ND	99	80-120	8	20	

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

DATA QUALIFIERS AND DEFINITIONS

C	Calibration Verification recovery was above the method control limit for this analyte. Analyte not detected, data not impacted.
L	Laboratory Control Sample recovery was above the method control limits. Analyte not detected, data not impacted.
M-3	Results exceeded the linear range in the MS/MSD and therefore are not available for reporting. The batch was accepted based on acceptable recovery in the Blank Spike (LCS).
M7	The MS and/or MSD were above the acceptance limits. See Blank Spike (LCS).
M-HA	Due to high levels of analyte in the sample, the MS/MSD calculation does not provide useful spike recovery information. See Blank Spike (LCS).
M-NR1	There was no MS/MSD analyzed with this batch due to insufficient sample volume. See Blank Spike/Blank Spike Duplicate.
ND	Analyte NOT DETECTED at or above the reporting limit or MDL, if MDL is specified.
RPD	Relative Percent Difference

ADDITIONAL COMMENTS

For 8260 analyses:

Due to the high water solubility of alcohols and ketones, the calibration criteria for these compounds is <30% RSD. The average % RSD of all compounds in the calibration is 15%, in accordance with EPA methods.

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

TestAmerica
ANALYTICAL TESTING CORPORATION

Burns & McDonnell, San Francisco	Project ID:	Weber Waters	
393 E. Grand Avenue, Suite J		40641	Sampled: 10/17/06
San Francisco, CA 94080	Report Number:	IPJ1647	Received: 10/17/06
Attention: Gary Messerotes			

Certification Summary

TestAmerica - Irvine, CA

Method	Matrix	Nelac	California
EPA 1625C-CI Mod	Water		
EPA 300.0	Water	X	X
EPA 314.0	Water	N/A	X
EPA 360.1	Water	X	X
EPA 524.2 Mod	Water		
EPA 6010B-Diss	Water	X	X
EPA 6010B	Water	X	X
EPA 7199	Water	X	X
EPA 7470A	Water	X	X
EPA 8260B	Water	X	X
EPA 8270C	Water	X	X
EPA 9034	Water		X

Nevada and NELAP provide analyte specific accreditations. Analyte specific information for TestAmerica may be obtained by contacting the laboratory or visiting our website at www.testamericainc.com

Subcontracted Laboratories

TestAmerica - Ontario, CA *California Cert #1169, Arizona Cert #AZ0062, Nevada Cert #CA-242*

1014 E. Cooley Drive, Suite AB - Colton, CA 92324

Method Performed: SRL 524M-TCP

Samples: IPJ1647-01, IPJ1647-02, IPJ1647-03, IPJ1647-04

TestAmerica - Irvine, CA
Lisa Reightley For Patty Mata
Project Manager

The results pertain only to the samples tested in the laboratory. This report shall not be reproduced, except in full, without written permission from TestAmerica.

Del Mar Analytical

A TEST AMERICA COMPANY

CHAIN OF CUSTODY FORM

IPD1647

Page 1 of 4

Client Name/Address: [illegible]

Project/PO Number: [illegible]

Project Manager: [illegible]

Phone Number: (650) [illegible]

Sampler: Patrick Button

Fax Number: (650) 871-2653

Analysis Required

Sample Description	Sample Matrix	Container Type	# of Cont.	Sampling Date	Sampling Time	Preservatives	[illegible]	EPA 601/602 NDMA	Cadmium CAM [illegible]	Perchlorate (EPA 314.0) Ammonia Nitrogen Chromium VI (7196)	Total [illegible]	Dissolved Oxygen	[illegible]	[illegible]		Special Instructions
SW-1	L	40mL VOA	3	10/17/06	0900	HCl	x									
	~~X~~L	1L Amber	4		0900	—		x								
	~~X~~L	500mL Poly	1		0900	HNO_3			x							
	L	500mL Poly	1		0900	—				x	x					
	~~X~~	Amber VOA	2		0900	—						x				
	L	500mL Poly	1		0900	NaOH + zinc							x			
↓	~~X~~L	40mL V	3		0900	Ascorbic Acid								x		
SW-3	L	40mL VOA	3		1140	HCl	x									
	L	1L Amber	4		1140	—		x								
	L	500mL Poly	1		1140	HNO_3			x							
	L	~~500mL Poly~~ ~~Amber VOA~~	1		1140	—				x	x					
	L	Amber VOA ~~500mL Poly~~	2		1140	—						x				ER 2145
	L	500mL Poly	1		1140	NaOH + zinc							x			
↓	L	40mL VOA	3	↓	1140	Ascorbic Acid								x		

Relinquished By:	Date/Time:	Received By:	Date/Time:	Turnaround Time: (Check)
Patrick Button	10-17-06 1550	[illegible]	10/17/06 1550	same day ____ 72 hours ____
Relinquished By: [illegible]	Date/Time: 10/17/06 1830	Received By:	Date/Time:	24 hours ____ 5 days ____
				48 hours ____ normal ✓
Relinquished By:	Date/Time:	Received in Lab By: [illegible]	Date/Time: 10/17/06 1830	Sample Integrity: (Check) intact ✓ on ice 3

Note: By relinquishing samples to Del Mar Analytical, client agrees to pay for the services requested on this chain of custody form and any additional analyses performed on this project. Payment for services is due within 30 days from the date of invoice. Sample(s) will be disposed of after 30 days.

037

Del Mar Analytical

A TEST AMERICA COMPANY

CHAIN OF CUSTODY FORM

[illegible] Corona Del Mar, Suite A, Costa Mesa, CA 92626 (949) 270-4067 FAX (949) 270-1040
[illegible] (800) 795-0651 FAX (800) 795-0651

Client Name/Address: [illegible]

Phone/FAX Number: [illegible]

Project Manager: [illegible]

Phone Number: [illegible]

Sampler: [illegible]

Fax Number: (650) 571-2453

Analysis Required

Sample Description	Sample Matrix	Container Type	# of Cont.	Sampling Date	Sampling Time	Preservatives	[illegible]	[illegible] MDMA-1625 [illegible]	[illegible] CAM Metals	Pesticides [illegible] Aroclors by [illegible]	[illegible]	Pesticides [illegible]	Sulfide [illegible]	[illegible]		Special Instructions
SW-2	L	VOA	3	10-17-06	1345	HCl	X									
	L	1L Amber	4		1345	—		X								
	L	500mL Poly	1		1345	HNO_3			X							
	L	500mL Poly	1		1345	—				X	X					
	L	Amb. VOA	2		1345	—						X				
	L	500mL Poly	1		1345	NaOH + Zinc							X			
	L	VOA	3		1345	Ascorbic Acid								X		
Trip Blank	L	VOA	3		-	HCl/Ascorbic	X							X		TM

Relinquished By:	Date/Time:	Received By:	Date/Time:	Turnaround Time: (Check)	
[illegible]	10-17-06 1550	[illegible]	10/17/06 1550	same day ____	72 hours ____
Relinquished By: [illegible]	Date/Time: 10/17/06 1830	Received By:	Date/Time:	24 hours ____	5 days ____
				48 hours ____	normal ✓
Relinquished By:	Date/Time:	Received in Lab By: [illegible]	Date/Time: 10/17/06 1830	Sample Integrity: (Check) intact ✓	on ice 3

Note: By relinquishing samples to Del Mar Analytical, client agrees to pay for the services requested on this chain of custody form and any additional analyses performed on this project. Payment for services is due within 30 days from the date of invoice. Sample(s) will be disposed of after 30 days.

037