Quality Assured
100%
Laboratory Tested

All Office Depot® Brand products endure laboratory testing to assure the quality, durability and safety of our products.

1-800-GO-DEPOT
(1-800-463-3768)
www.officedepot.com
www.officedepot.ca

Distributed by: Office Depot, Inc.
2200 Old Germantown Road
Delray Beach, FL 33445, USA
©2007, Office Depot, Inc.

Made in India / Fabriqué en Inde

MULTIRAE PRO

EPA# SB6192

4/3/15 ops check / Fresh Air cal / bump check ok R
4/10/15 calib. w/4 gas lot # T082-506 calib. Pid #1 w/10 ppm Iso lot # T083-501 calib. Pid #2 w/100 ppm Iso lot # T076-501 passed. af
4/17/15 ops ✓ good. af
4/21/15 ops ✓ bump ✓ EMC
5/6/15 Tried to calib. LEL Failed will try and fix & Re calib. af
5/7/15 LEL won't go above 20ppm ordering New LEL af

5/15/15 RePL. LEL sensor & Recalib. w/4 gas lot # T082-501 good Now. af
5/22/15 ops/bump check good. af
6/1/15 Cal. all sensors passed Mw
6/12/15 Returned from John Day. All sensors passed calibration. Mw
6/15/15 Calib. unit w/4 gas lot # T082-503 calib. pid #1 w/10 ppm Iso lot # T083-503 calib. pid #2 w/100 ppm Iso lot # T076-501 passed. af
6/18/15 OPS CHECK / FRESH AIR CAL / BUMP CHECK EMC

Quality Assured
100%
Laboratory Tested

All Office Depot® Brand products endure laboratory testing to assure the quality, durability and safety of our products.

1-800-GO-DEPOT
(1-800-463-3768)
www.officedepot.com
www.officedepot.ca

Distributed by: Office Depot, Inc.
2200 Old Germantown Road
Delray Beach, FL 33445, USA
©2007, Office Depot, Inc.

Made in India / Fabriqué en Inde

MULTIRAE PRO
composition book
SB7157
80

3/11/15 OPS TEST OK, FRESH AIR CAL OK, BATT OK, BUMP TEST W/ MULTI-GAS T082-506 AND 100ppm ISO T076-503 YIELDS

	O_2	LEL	CO	H_2S	VOC
UPRT	18.1%	48%	48ppm	10.0ppm	96.34
BUMP	18%	50%	50ppm	10ppm	100ppm

OK CRW

3/16/15 calib. w/4 gas lot# T082-501 calib. Pid #1 w/10ppm ISO lot# T083-501 calib. Pid #2 w/100ppm ISO lot# T076-503 passed gf

4/3/15 bump check & Fresh Air cal OK

4/10/15 Calib. w/4 gas lot# T082-506. Calib. Pid #1 w/10ppm Iso lot# T083-501. Calib. #2 Pid w/100 ppm ISO lot# T076-501 passed, gf

5/7/15 Calib. w/4gas lot# T082-502 Calib. Pid #1 w/10ppm lot# T083-501 Calib. Pid #2 w/100ppm lot# T076-501 passed gf

6/12/15 Returned from John Day VI. Replaced PID/VOC w/new sensor. Calibrated all sensors passed MW

6/15/15 calib. w/4 gas lot# T082-503 calib. Pid #1 w/10ppm ISO lot# T083-503 calib.

Quality Assured 100% Laboratory Tested

All Office Depot® Brand products endure laboratory testing to assure the quality, durability and safety of our products.

1-800-GO-DEPOT
(1-800-463-3768)
www.officedepot.com
www.officedepot.ca

Distributed by: Office Depot, Inc.
2200 Old Germantown Road
Delray Beach, FL 33445, USA
©2007, Office Depot, Inc.

Made in India / Fabriqué en Inde

composition book

narrow ruled
80 sheets • 3 1/4 in x 4 1/2 in
(8.3 cm x 11.4 cm)

MULTI-RAE PRO

EPA# SB7022

	O_2	LEL	CO	H_2S	VOC
UNIT	18.0%	50%	9ppm	9.5ppm	8.7ppm
BUMP	18%	50%	50ppm	10ppm	10ppm
				OK	OK

3/23 ops test ✓ bump ✓ MW

4/1/15 OPS ✓, Battery Good. Zeroed Bump test Good. Within parameters

4/10/15 Calib. w/4 gas lot # T082-506. Calib pid #1 w/10ppm Iso. lot # T083-501 calib. pid #2 w/100 ppm Iso. lot # T076-501. passed gf

4/27/15 OPS CK. Fresh air Calibration Pass. Bump test each sensor

5/6/15 Calib. w/4 gas lot # T082-502. Calib. Pid #1 w/10 ppm lot # T083-501 Calib. Pid #2 w/100 ppm lot # T076-501. passed. gf

5/15/15 ops test, bump ✓ MW

5/29/15 OPS CK. Fresh Air CK. Bump test c̄ multigas - passed. ~~[illegible]~~ H_2S + CO sensors not responding to Bump test.

6/1/15 CO sensor drifting negative. Calibrated all sensors passed ✓ MW

6/12/15 Returned from ~~Field~~ John Pay LEL Failed cal all others passed. MW

6/15/15 calib. w/4 gas lot# TO 82-503 calib. pid #1 w/10ppm Iso lot # TO83-503 calib. Pid #2 w/100ppm Iso lot# TO76-501 passed. gf

6/24/15 Recalib to check sensor op calib. w/4 gas lot # TO82-503 all sensor's passed. gf

6-26-15 OPS ✓ BUCLEAN AIR CAL ✓ EMA

7/1/15 calib. w/4 gas lot # TO 82-502 calib. Pid #1 w/10ppm Iso lot # TO83-502 calib. Pid #2 w/100ppm Iso lot# TO 76-505 unit passed. gf

7/15/15 Calib. w/4 gas lot# TO 82-504 calib. Pid #1 w/10ppm Iso lot# TO83-503 calib. Pid #2 w/100ppm Iso lot # TO76-501 passed. gf

7/20 OPS ✓ Bump Iso 10 ppm lot TØ83-502 pass (SP)

7/27/15 OPS TEST OK ~~[illegible]~~ AIR CAL; BUMP TEST W/ MULTI GAS TO82-502 AND 10ppm ISO TO83-502 YIELDS NO SIGNIFICANT DEVIATION FROM BACKGROUND EXCEPT LEL SENSOR WHICH EXCEEDED 19%. UNIT CORRECTED AFTER RESETTING OK CW

Quality Assured
100%
Laboratory Tested

All Office Depot® Brand products endure laboratory testing to assure the quality, durability and safety of our products.

1-800-GO-DEPOT
(1-800-463-3768)
www.officedepot.com
www.officedepot.ca

Distributed by: Office Depot, Inc.
2200 Old Germantown Road
Delray Beach, FL 33445, USA
©2007, Office Depot, Inc.

Made in India / Fabriqué en Inde

composition book

narrow ruled
80 sheets • 3 ¼ in x 4 ½ in.
(8.3cm x 11.4 cm)

MULTI-RAE PRO

EPA# SB7023

2/6 Calibrated all sensors passed. PID reading negative on 100 ppm. Reported for lamp cleaning MW

2/27 Replaced CO/H2S + PID sensors Calib all sensors w/ Multi Gas SOS, 10 ppm Iso SO2 + 100 ppm Iso SOS. All passed ✓ MW

3/9/15 calib w/ 4 gas lot# T082-506 calib Pid #1 w/ 10 ppm ISO lot # T083-50 calib. Pid #2 w/ 100 ppm Iso lot# T076-503 passed af

3/16/15 Calib w/ 4 gas lot# T082-506 Calib Pid #1 w/ 10 ppm Iso lot# T083-501 calib. Pid #2 w/ 100 ppm Iso lot# T076-503 passed af

4/10/15 calib. w/ 4 gas lot# T082-506 calib. Pid #1 w/ 10 ppm ISO lot # T083-501 Calib Pid #2 w/ 100 ppm Iso lot# T076-501 passed af

5/7/15 Calib. w/ 4 gas lot # T082-502 calib. Pid #1 w/ 10 ppm lot # T083-501 Calib. Pid #2 w/ 100 ppm Iso lot # T076-501 passed. af

5/15 ops ✓ bump ✓ need VOC alarm limits reset to defaults (it alarms @ 10 ppm) MW

5/29/2015 Not on the ERT - in the Field

6/1/15 Cal. all sensors passed MW af

6/12/15 Returned From John Day. LEL sensor failed calibration all others passed. gf
6/16/15 LEL sensor is on order. gf
6/19/15 Repl. LEL sensor Calib. w/4gas lot # T082-502 passed, gf

7/15/15 Calib. w/4gas lot# T082-504 Calib. Pid #1 w/10 ppm ISO lot# T083-503 calib Pid #2 w/100 ppm ISO lot# T076-501. passed gf.

Quality Assured 100% Laboratory Tested

All Office Depot® Brand products endure laboratory testing to assure the quality, durability and safety of our products.

1-800-GO-DEPOT
(1-800-463-3768)
www.officedepot.com
www.officedepot.ca

Distributed by: Office Depot, Inc.
2200 Old Germantown Road
Delray Beach, FL 33445, USA
©2007, Office Depot, Inc.

Made in India / Fabriqué en Inde

MULTIRAE PRO
composition book
SB7058
narrow ruled
80 sheets • 3 ¼ in x 4 ½ in
(8.3 cm x 11.4 cm)

5/6/15 Calib. w/ 4 gas lot # T082-502 calib. Pid #1 w/10 ppm lot # T083-501 Calib. Pid #2 w/100ppm lot # T076-501. gf

5-15-15 Checkout. Fresh air cal. Bump iso T083-502 10 ppm. Read 21900 ppb

6/12/15 Returned from John Day. LEL sensor failed Calibration all other sensors passed. MW

6/15/15 Calib. w/4 gas lot # T082-503 calib. Pid #1 w/10ppm Iso lot # T083-503 calib. Pid #2 w/100 ppm Iso lot # 7076-501 passed. gf

6-26-15 OPS ✓ CLEAN AIR CAL ✓ EMO

7/1/15 calib. w/4 gas lot # T082-502 calib. Pid #1 w/10ppm Iso lot # T083-502 calib. Pid #2 w/100ppm Iso lot # T076-505 unit passed. gf

7/15/15 calib. w/4 gas lot # T082-504 calib. Pid #1 w/10 ppm Iso lot # T083-503 calib. Pid #2 w/100ppm Iso lot # T076-501 passed. gf

7/20/15 ops ✓ Bump ISO 10 ppm lot T083-502 pass JM

Quality Assured 100% Laboratory Tested

All Office Depot® Brand products endure laboratory testing to assure the quality, durability and safety of our products.

1-800-GO-DEPOT
(1-800-463-3768)
www.officedepot.com
www.officedepot.ca

Distributed by: Office Depot, Inc.
2200 Old Germantown Road
Delray Beach, FL 33445, USA
©2007, Office Depot, Inc.

Made in India / Fabriqué en Inde

MULTIRAE PRO
composition book
SB7069
80

	O_2	LEL	CO	H_2S	VOC
UNIT	15.8%	48%	48ppm	9.5ppm	9.3ppm
BUMP	18%	50%	50ppm	10ppm	10ppm

OK cal

3/23 ops test ✓ bump ✓ MW

4/1/2015 OPS CK. Battery Good. Bump test & multigas Gas. Good. LEL read a little low to bump Gas

4/10/15 Calib w/4 gas lot# T082-506. Calib PID w/10ppm Iso. lot# T083-501 Calib. pid #2 w/100ppm Isg. lot# T076-501 passed. gl

4/27/2015 OPS CK. Fresh air Calibrate Passed Bump test each sensor.

5/6/15 Calib. w/4 gas lot# T082-502 Calib. PID #1 w/10ppm lot# T083-501 calib. PID #2 w/100ppm lot# T076-501 passed. gl

5/15 ops test, bump test MW

5/29/2015 OPS CK. Fresh Air CK Good ~~CO & H_2S sensors alarming.~~ Bump test to multigas - Good/Passed

6/1/15 LEL sensor will not calibrate. all others passed MW

6/1/15 Allowed unit to run for 4+ hrs + LEL sensor still failed calibration MW

6/12/15 LEL Failed calibration
all others passed. Still
awaiting LEL sensor MW
6/15/15 Calib. meter LEL
Calib. meter passed Calib. gf
6/18/15 OPS CHECK/FRESH AIR CAL/
BUMP CHECK EMM
6/25/15 OPS/bump check OK, Fresh air
cal
7/21/15 Repl. LEL Sensor &
Recalib. w/4gas lot # T082
-502 passed. gf
7/24/15 OPS/cal
Woods Calibration

7/30/2015 Calib. w/4 gas lot
T082-502 Calib. PID #1 w/
10 ppm Iso lot # T083-501
Calib. PID # 2 w/ 100 ppm Iso
lot # T076-501 passed. gf
8/7/2015 OPS/ck - Fresh Cal.
EOlalan

25

ecology and environment, inc.

International Specialists in the Environment

Job Number [redacted]

[redacted]

[redacted]

AREA RAE

METERS 1,2,3

Bump tested c̄ 4-gas & Iso. Connected all units to Computer monitor software
[illegible]

2-25-2015 OPS CHECK / FRESH AIR CALIBRATION

3/2/15 Fresh air cal, ops check, batt ok

	CO	LEL	VOC	H_2S	O_2
1					
2			11.1		
3			11.5		

3/11/15 OPS CHECK OK FRESH AIR CAL OK BUMP TEST W/ MULTI-GAS T082-506 AND ISO T076-505 YIELDS

	CO	H_2S	LEL	O_2	VOC
1	47	10	49	17.9	90.2
2	47	10	47	17.8	82.7
3	47	10	51	18.0	102
STD	50	10	50%	18%	100 ppm

4/1/15 OPS ✓ BATT ✓ BMP ✓ [illegible]
/3/15 All units Fresh Air cal'd + Bump checked [illegible]
/10/15 Calib. units 1,2,3 w/4 gas t # T082-506. Calib. PID w/100ppm Iso t # T076-501 passed. [illegible]
/17/15 ops ✓ + bump test pass [illegible]
27-15 OPS ✓, BUMP TEST PASS, ISO + 4 GAS [illegible]

5/3/14 Checked Raw readings on units 1,2,3 unit #2 VOC didn't pass Repl. PID bulb + Re calib. [illegible]

5-15-15 Checkout.
- Fresh air calibration all 3 units
- Bump check iso T083-503 10ppm
  - unit 1 read 33.2 ppm
  - unit 2 read 18.8 ppm
  - unit 3 read 14.9 ppm
- Unit 1 alarmed with elevated O_2 reading of 23.7, unit re-zeroed (by [illegible])
- Bumped unit 1 w/ iso 10ppm (read 32.4) which also triggered/coincided with increased O_2 reading 23.2 %

5/15/15 Repl. O_2 sensor + Recalib w/4 gas lot # T082-501, Re calib. PID w/100ppm Iso lot # T076-505 passed [illegible]
5/22/15 ops/bump check batts good a 3 passed, [illegible]
6/12/15 Returned from John Day. All sensors, all units passed calibration [illegible]

6/1/15 Cleaned all Pid lamps & Recalib. w/4gas lot# T082-501 calib. Pids w/100ppm ISO lot# T076-505 all passed. gf

6-26-15 OPS ✓ ALL UNITS BUMP TESTED ✓ EMC

7/1/15 Calib. units 1,2,3 w/4 gas lot# T082-502, Calib. pid w/ 100ppm ISO lot# T076-505 all passed gf

7/15/15 Calib. units 1,2,3 w/4gas lot# T082-503 calib. Pid w/100ppm ISO lot# T076-501 all passed, gf

7/20 Ops ✓ Bump ISO 10ppm lot T082-502. All passed

7/27/15 OPS TEST OK ALL UNITS FRESH AIR CAL AND BUMP W/ MULTI-GAS T082-502; NO DEVIATIONS FROM NORM

ALSO 10 ppm ISO T083-502 OK

7/31/15 ops ✓ bump tested w/4gas Tested Pids w/100ppm ISO. all passed gf

ecology and environment, inc.

International Specialists in the Environment

Job Number

ER AREA Rues
units 4,5,6

3/16/15 Calib. unit 4,5,6 w/4 gas lot # T082-501 calib. PID w/100ppm ISO lot # T076-503 all passed. gf

3/18/15 OPS CHECK OK, BATT OK FRESH AIR CAL OK, BUMP W/ MULTI-GAS T082-501

	O_2	LEL	CO	H_2S	VOC
UNIT 4	17.8%	48%	53 ppm	10 ppm	7.5 ppm
UNIT 5	17.9%	50%	46 ppm	9 ppm	6.1 ppm
UNIT 6	18.1%	52%	51 ppm	10 ppm	7.0 ppm
BUMP	18%	50%	50 ppm	10 ppm	10 ppm

MULTI-GAS T082-501 DEPLETED DURING BUMP OF UNIT 6; SWITCHED TO T082-506 FOR UNIT 6 BUMP ONLY. ALSO USED 10 ppm ISO T083-501 FOR ALL UNITS OK

3/23/15 Unit 4 left on so cleared memory. Unit 5 lamp took a longer time to light. All units bump test passed. lw

4/1/2015 OPS CK. Battery Good, Zeroed. Bump test c̄ multi-span gas Good.

1/10/15 calib units 4,5,6 w/4 gas lot # T082-506. Calib. PID w/100 ppm ISO, lot # T076-501 all passed gf

2-15 OPS TEST, BUMP TEST, CALIBRATED W/ 4 GAS + ISO EML

1/29/15 Rotated Area Rae battery w/ back Kup Batt. gf

5/15/15 ops test + bump all units Unit 4 need to be re-cal. VOC was calibrated to 10 ppm as 100 ppm. Bump w/10 ppm Iso read 88 ppm. lw

7/15/15 Recalib. w/4 gas lot # T082-501 calib. PID w/100 ppm ISO lot # T076-505 passed. gf

5/29/2015 OPS CK - Batteries Good. Fresh air ~~calibration~~ CK all passed. Bump test c̄ multigas OK.

6/26/15 ops/bump check OK, Fresh Air cal.

7/15/15 Calib. units 4,5,6 w/4 gas lot # T082-504. Calib. PID w/100 ppm ISO lot # T076-501 all passed. gf

7/20/15 OPS TEST OK, FRESH AIR CAL ALL UNITS OK; BUMP TEST W/ MULTI-GAS T082-504 AND 10 ppm ISO T083-503 ALSO

	CO	VOC	H_2S	LEL	O_2
BUMP	50 ppm	10 ppm	10 ppm	50%	18%
UNIT 4	48 ppm	14.2 ppm	9 ppm	49%	17.6%
UNIT 5	48 ppm	9.4 ppm	9 ppm	50%	17.6%
UNIT 6	49 ppm	10.8 ppm	9 ppm	46%	17.10%

Quality Assured
100%
Laboratory Tested

All Office Depot® Brand products endure laboratory testing to assure the quality, durability and safety of our products.

1-800-GO-DEPOT
(1-800-463-3768)
www.officedepot.com
www.officedepot.ca

Distributed by: Office Depot, Inc.
2200 Old Germantown Road
Delray Beach, FL 33445, USA
©2007, Office Depot, Inc.

Made in India / Fabriqué en Inde

composition book

narrow ruled
80 sheets • 3 1/4 in x 4 1/2 in
(8.3 cm x 11.4 cm)

ULTRA RAE
3000
EPA #
SB7016

4/30/15 OPS TEST OK, BATT OK, BUMP TEST W/ T083-502 10 ppm ISO READS 20 ppm GOOD ENOUGH OK [illegible]

5/7/15 Calib. w/10 ppm ISO 10t # T083-501 passed. gf

5-15-15 checkout iso 10 ppm T083-502 10 ppm read 24.70 ppm.

6/19/15 cleaned Pid. Calib. w/ 10 ppm ISO 10t # T083-502 passed. gf

6/25/15 OPS/bump check ok Fresh Air cal

7/17/15 Calib. Pid w/10 ppm ISO 10t # T083-503 passed. gf

7/20/15 OPS TEST OK, FRESH AIR CAL OK, BUMP TEST W/ 10 ppm ISO T083-503 YIELDS 10.40 ppm [illegible]

7/24/15 OPS [illegible]

8/7/2014 OPS CK. Zeroed. No Bump test Calibrated 7/20/14. [illegible]