National® Brand CHEMISTRY NOTEBOOKS

Blue Cover			
Item No.	Numbered Pages	Ruling	Size
Item No. 43-571	120	Record	9 ¼" x 7 ¼"
Item No. 43-581	120	Record	11" x 8 ½"

Product Guarantee

Rediform is committed to providing you with quality products, and will gladly replace any product which does not provide complete satisfaction. We also welcome your comments and suggestions. Please send your correspondence with product code to:

Rediform Inc.
555 Airline Drive • Coppell, TX 75019

FSC www.fsc.org MIX Paper from responsible sources FSC® C102682

RECYCLED CONTENT Paper: Contains a minimum of 50% post-consumer fiber.

ACID-FREE PAPER

Portage - Hwy 12 hi-vol
3/17/2021 10:35 am
Katie Healy, Kate Haile, Scott Hamilton (EPA)
Scott Keller (IDEM)

Calibrating TE-PUFPLUS PUF sampler
using EPA's Graseby orifice (SN: 584)
EPA's DeltaCal T/P standard (SN: 178722)
IDEM's digital manometer (SN: 49529521)

Calibration results:
*need to do a 3-point Tamb calibration

R = 0.9986
m = 35.7936
b = -0.9826

Took USB back to Chicago
Left at 11:48 am

KH