8/27/21 SK Setup PTG-A-25
run day 8/28/21

8/30/21 SK Pickup PTG-A-25

9/1/2021 Katie Healy & Kate Haile
arrived at 11:28 CST
warmed up motor for 15 minutes
performed one-point flow check
ΔT: 0.7°C
ΔP: 0.5 mmHg
Q % difference: 1.2%
all checks pass
deployed PTG-A-26 at 11:55 CST
with run date 9/2/2021
flow rate still set at 217 lpm
and duration of 24hr 30min
~~left at~~ KMH 9/1/21
downloaded data onto USB
left at 12:00 CST

KMH

9/3/21 SK pick up PTG-A-26

9/8/21 SK Setup PTG-A-27
run date 9/9/21

9/10/21 SK pickup PTG-A-27

9/13/21 Katie Healy arrived at 10:28 CST
downloaded data onto USB
left at 10:32 CST

9/14/21 SK Setup PTG-A-28
run date 9/15/21

9/16/21 SK pickup PTG-A-28 &
blank FBA-PTG-A-28

9/20/21 SK Setup PTG-A-29
run date 9/21/21

9/22/21 SK pickup PTG-A-29

9/24/21 Katie Healy arrived at 11:09 CDT
downloaded data onto USB
set up PTG-A-30 to run on 9/27/21
left at 11:35 ~~EST~~ KMH 9/24/21 CDT

9/28/21 SK pickup PTG-A-30

9/29/21 IDEM QA AUDIT SAMPLER OFF ON ARRIVAL
MBURKS TAMB +0.4°C BP: -1.3mmHg Q: +1.1%
DATE/TIME OK, -1.0 MIN FROM REF
ALL V's PASS ON CLOSEOUT AUDIT

9/29/21 Kate Haile & Katie Healy arrived at 1:35 pm CDT
downloaded data onto USB
took apart sampler & brought sampler
back to Chicago - Region 5 EPA