ANALYTICAL REPORT

Job Number: 500-4479-1

Job Description: SECO

For:
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Attention: Mr. Alan Cork

Rich Mannz
Project Manager II
rmannz@stl-inc.com
06/15/2007

Project Manager: Rich Mannz

These test results meet all the requirements of NELAC for accredited parameters.

The Lab Certification ID# is 100201.

All questions regarding this test report should be directed to the STL Project Manager whose signature appears on this report. All pages of this report are integral parts of the analytical data. Therefore, this report should be reproduced only in its entirety.

Reporting limits are adjusted for sample size used, dilutions and moisture content if applicable.

Job Narrative
500-J4479-1

I. Comments
No additional comments.

II. Receipt
All samples were received in good condition within temperature requirements.

III. GC/MS VOA
Samples 8, 9, 20, 22, and 23 were diluted due to the abundance of target analytes. Elevated reporting limits (RLs) are provided.

No other analytical or quality issues were noted.

IV. Metals
No analytical or quality issues were noted.

V. General Chemistry
No analytical or quality issues were noted.

EXECUTIVE SUMMARY - Detections

Client: ERM Inc. Job Number: 500-4479-1

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier	Reporting Limit	Units	Method
500-4479-3	**MW-3-2**				
cis-1,2-Dichloroethene		8.6	1.0	ug/L	8260B
Trichloroethene		17	1.0	ug/L	8260B
500-4479-4	**MW-3-3**				
cis-1,2-Dichloroethene		91	1.0	ug/L	8260B
Tetrachloroethene		1.8	1.0	ug/L	8260B
trans-1,2-Dichloroethene		3.2	1.0	ug/L	8260B
Trichloroethene		180	10	ug/L	8260B
500-4479-6	**MW-4-2**				
Trichloroethene		1.6	1.0	ug/L	8260B
500-4479-7	**MW-4-3**				
Trichloroethene		1.5	1.0	ug/L	8260B
500-4479-8	**MS-2-1**				
1,1-Dichloroethane		11	10	ug/L	8260B
1,1-Dichloroethene		25	10	ug/L	8260B
cis-1,2-Dichloroethene		7300	100	ug/L	8260B
Tetrachloroethene		21	10	ug/L	8260B
trans-1,2-Dichloroethene		41	10	ug/L	8260B
Trichloroethene		7800	100	ug/L	8260B
Vinyl chloride		19	10	ug/L	8260B
500-4479-9	**MS-2-2**				
cis-1,2-Dichloroethene		1600	50	ug/L	8260B
trans-1,2-Dichloroethene		7.5	5.0	ug/L	8260B
Trichloroethene		95	5.0	ug/L	8260B
Vinyl chloride		52	5.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: ERM Inc. Job Number: 500-4479-1

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier	Reporting Limit	Units	Method
500-4479-10	**RW-3**				
1,1-Dichloroethene		1.4	1.0	ug/L	8260B
cis-1,2-Dichloroethene		450	10	ug/L	8260B
trans-1,2-Dichloroethene		2.3	1.0	ug/L	8260B
Trichloroethene		4.5	1.0	ug/L	8260B
Vinyl chloride		140	10	ug/L	8260B
Sulfate		42	10	mg/L	375.4
Total Recoverable					
Sodium		9.8	0.50	mg/L	200.7 Rev 4.4
500-4479-11	**SWGW/AS-1**				
cis-1,2-Dichloroethene		1.5	1.0	ug/L	8260B
Vinyl chloride		15	1.0	ug/L	8260B
500-4479-12	**SWGW/AS-2**				
cis-1,2-Dichloroethene		1.7	1.0	ug/L	8260B
Vinyl chloride		15	1.0	ug/L	8260B
500-4479-13	**SWGW/A-1**				
1,1-Dichloroethane		1.3	1.0	ug/L	8260B
1,1-Dichloroethene		2.2	1.0	ug/L	8260B
Benzene		2.5	1.0	ug/L	8260B
cis-1,2-Dichloroethene		1000	20	ug/L	8260B
trans-1,2-Dichloroethene		4.0	1.0	ug/L	8260B
Vinyl chloride		320	10	ug/L	8260B
500-4479-14	**SWGW/A-2**				
1,1-Dichloroethane		1.2	1.0	ug/L	8260B
1,1-Dichloroethene		2.3	1.0	ug/L	8260B
Benzene		2.7	1.0	ug/L	8260B
cis-1,2-Dichloroethene		1100	20	ug/L	8260B
trans-1,2-Dichloroethene		4.9	1.0	ug/L	8260B
Vinyl chloride		370	10	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: ERM Inc. Job Number: 500-4479-1

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier	Reporting Limit	Units	Method
500-4479-15	**RW-2**				
1,1-Dichloroethane		1.7	1.0	ug/L	8260B
1,1-Dichloroethene		7.7	1.0	ug/L	8260B
Benzene		4.1	1.0	ug/L	8260B
cis-1,2-Dichloroethene		2600	50	ug/L	8260B
Tetrachloroethene		1.3	1.0	ug/L	8260B
trans-1,2-Dichloroethene		30	1.0	ug/L	8260B
Trichloroethene		670	10	ug/L	8260B
Vinyl chloride		250	10	ug/L	8260B
Sulfate		21	10	mg/L	375.4
Total Recoverable					
Sodium		13	0.50	mg/L	200.7 Rev 4.4
500-4479-16	**SWGW/B-1**				
cis-1,2-Dichloroethene		410	10	ug/L	8260B
trans-1,2-Dichloroethene		3.4	1.0	ug/L	8260B
Vinyl chloride		320	10	ug/L	8260B
500-4479-17	**SWGW/B-2**				
cis-1,2-Dichloroethene		400	10	ug/L	8260B
trans-1,2-Dichloroethene		3.2	1.0	ug/L	8260B
Vinyl chloride		340	10	ug/L	8260B
500-4479-18	**MW-5-2**				
1,1-Dichloroethene		3.1	1.0	ug/L	8260B
cis-1,2-Dichloroethene		190	10	ug/L	8260B
Trichloroethene		250	10	ug/L	8260B
500-4479-19	**MW-5-3**				
1,1-Dichloroethene		2.8	1.0	ug/L	8260B
cis-1,2-Dichloroethene		180	10	ug/L	8260B
trans-1,2-Dichloroethene		1.8	1.0	ug/L	8260B
Trichloroethene		250	10	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: ERM Inc. Job Number: 500-4479-1

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier	Reporting Limit	Units	Method
500-4479-20	**RW-8**				
1,1-Dichloroethene		7.5	5.0	ug/L	8260B
cis-1,2-Dichloroethene		2000	50	ug/L	8260B
trans-1,2-Dichloroethene		29	5.0	ug/L	8260B
Trichloroethene		720	50	ug/L	8260B
Vinyl chloride		150	5.0	ug/L	8260B
Sulfate		69	10	mg/L	375.4
Total Recoverable					
Sodium		15	0.50	mg/L	200.7 Rev 4.4
500-4479-21	**RW-1**				
1,1-Dichloroethene		1.6	1.0	ug/L	8260B
Benzene		3.0	1.0	ug/L	8260B
cis-1,2-Dichloroethene		1300	20	ug/L	8260B
trans-1,2-Dichloroethene		52	1.0	ug/L	8260B
Vinyl chloride		330	10	ug/L	8260B
Sulfate		57	10	mg/L	375.4
Total Recoverable					
Sodium		11	0.50	mg/L	200.7 Rev 4.4
500-4479-22	**SWGW/C-1**				
1,1-Dichloroethene		20	10	ug/L	8260B
cis-1,2-Dichloroethene		12000	500	ug/L	8260B
trans-1,2-Dichloroethene		2000	100	ug/L	8260B
Vinyl chloride		1500	100	ug/L	8260B
500-4479-23	**SWGW/C-2**				
1,1-Dichloroethene		13	10	ug/L	8260B
Benzene		12	10	ug/L	8260B
cis-1,2-Dichloroethene		9500	100	ug/L	8260B
trans-1,2-Dichloroethene		910	10	ug/L	8260B
Vinyl chloride		1600	100	ug/L	8260B
500-4479-24	**SWGW/CS-1**				
cis-1,2-Dichloroethene		820	10	ug/L	8260B
trans-1,2-Dichloroethene		4.5	1.0	ug/L	8260B

METHOD SUMMARY

Client: ERM Inc. Job Number: 500-4479-1

Description	Lab Location	Method	Preparation Method
Matrix: Water			
Volatile Organic Compounds by GC/MS	STL CHI	SW846 8260B	
Purge-and-Trap	STL CHI		SW846 5030B
ICP Metals by 200.7	STL CHI	EPA 200.7 Rev 4.4	
Total Recoverable Metals Concentration for 200.7	STL CHI		EPA 200.7
Sulfate (Turbidimetric)	STL CHI	MCAWW 375.4	

LAB REFERENCES:

STL CHI = STL Chicago

METHOD REFERENCES:

EPA - US Environmental Protection Agency

MCAWW - "Methods For Chemical Analysis Of Water And Wastes", EPA-600/4-79-020, March 1983 And Subsequent Revisions.

SW846 - "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

METHOD / ANALYST SUMMARY

Client: ERM Inc.

Job Number: 500-4479-1

Method	Analyst	Analyst ID
SW846 8260B	Kras, Michael J	MJK
EPA 200.7 Rev 4.4	Smith, Todd D	TDS
MCAWW 375.4	Muno, Rebecca N	RNM

SAMPLE SUMMARY

Client: ERM Inc. Job Number: 500-4479-1

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
500-4479-1	MW-2-1	Water	05/29/2007 1115	06/01/2007 1107
500-4479-2	MW-2-2	Water	05/29/2007 1120	06/01/2007 1107
500-4479-3	MW-3-2	Water	05/29/2007 1200	06/01/2007 1107
500-4479-4	MW-3-3	Water	05/29/2007 1210	06/01/2007 1107
500-4479-5	MW-4-1	Water	05/29/2007 1230	06/01/2007 1107
500-4479-6	MW-4-2	Water	05/29/2007 1240	06/01/2007 1107
500-4479-7	MW-4-3	Water	05/29/2007 1250	06/01/2007 1107
500-4479-8	MS-2-1	Water	05/29/2007 1400	06/01/2007 1107
500-4479-9	MS-2-2	Water	05/29/2007 1410	06/01/2007 1107
500-4479-10	RW-3	Water	05/29/2007 1420	06/01/2007 1107
500-4479-11	SWGW/AS-1	Water	05/29/2007 1450	06/01/2007 1107
500-4479-12	SWGW/AS-2	Water	05/29/2007 1500	06/01/2007 1107
500-4479-13	SWGW/A-1	Water	05/29/2007 1505	06/01/2007 1107
500-4479-14	SWGW/A-2	Water	05/29/2007 1515	06/01/2007 1107
500-4479-15	RW-2	Water	05/29/2007 1525	06/01/2007 1107
500-4479-16	SWGW/B-1	Water	05/29/2007 1550	06/01/2007 1107
500-4479-17	SWGW/B-2	Water	05/29/2007 1600	06/01/2007 1107
500-4479-18	MW-5-2	Water	05/29/2007 1610	06/01/2007 1107
500-4479-19	MW-5-3	Water	05/29/2007 1620	06/01/2007 1107
500-4479-20	RW-8	Water	05/29/2007 1625	06/01/2007 1107
500-4479-21	RW-1	Water	05/29/2007 1630	06/01/2007 1107
500-4479-22	SWGW/C-1	Water	05/29/2007 1650	06/01/2007 1107
500-4479-23	SWGW/C-2	Water	05/29/2007 1655	06/01/2007 1107
500-4479-24	SWGW/CS-1	Water	05/29/2007 1710	06/01/2007 1107

SAMPLE RESULTS

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-2-1
Lab Sample ID: 500-4479-1

Date Sampled: 05/29/2007 1115
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 1713			
Prep Method: 5030B	Date Prepared:	06/05/2007 1713			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	ND	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,2-Dichloroethene	ND	ug/L	0.20	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-2-1
Lab Sample ID: 500-4479-1

Date Sampled: 05/29/2007 1115
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 1713			
Prep Method: 5030B	Date Prepared:	06/05/2007 1713			
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	ND	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	ND	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0
Vinyl chloride	ND	ug/L	0.16	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103	%	70 - 125
Toluene-d8 (Surr)	102	%	75 - 120
4-Bromofluorobenzene (Surr)	96	%	75 - 120
Dibromofluoromethane	101	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-2-2
Lab Sample ID: 500-4479-2

Date Sampled: 05/29/2007 1120
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 1736			
Prep Method: 5030B	Date Prepared:	06/05/2007 1736			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	ND	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,2-Dichloroethene	ND	ug/L	0.20	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-2-2
Lab Sample ID: 500-4479-2

Date Sampled: 05/29/2007 1120
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 1736			
Prep Method: 5030B	Date Prepared:	06/05/2007 1736			
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	ND	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	ND	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0
Vinyl chloride	ND	ug/L	0.16	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101	%	70 - 125
Toluene-d8 (Surr)	99	%	75 - 120
4-Bromofluorobenzene (Surr)	98	%	75 - 120
Dibromofluoromethane	102	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-3-2
Lab Sample ID: 500-4479-3

Date Sampled: 05/29/2007 1200
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 1758			
Prep Method: 5030B	Date Prepared:	06/05/2007 1758			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	ND	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,2-Dichloroethene	8.6	ug/L	0.20	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-3-2
Lab Sample ID: 500-4479-3

Date Sampled: 05/29/2007 1200
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 1758			
Prep Method: 5030B	Date Prepared:	06/05/2007 1758			
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	ND	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	17	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0
Vinyl chloride	ND	ug/L	0.16	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	95	%	75 - 120
Dibromofluoromethane	103	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-3-3
Lab Sample ID: 500-4479-4

Date Sampled: 05/29/2007 1210
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 1844			
Prep Method: 5030B	Date Prepared:	06/05/2007 1844			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	ND	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,2-Dichloroethene	91	ug/L	0.20	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-3-3
Lab Sample ID: 500-4479-4

Date Sampled: 05/29/2007 1210
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 1844			
Prep Method: 5030B	Date Prepared:	06/05/2007 1844			
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	1.8	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	3.2	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0
Vinyl chloride	ND	ug/L	0.16	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	98	%	75 - 120
Dibromofluoromethane	102	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/05/2007 1906			
Prep Method: 5030B	Date Prepared:	06/05/2007 1906			
Trichloroethene	180	ug/L	1.3	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	98	%	75 - 120
Dibromofluoromethane	104	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-3-3
Lab Sample ID: 500-4479-4

Date Sampled: 05/29/2007 1210
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL	Date Analyzed:	06/05/2007 1906			
Prep Method: 5030B	Date Prepared:	06/05/2007 1906			

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-4-1
Lab Sample ID: 500-4479-5

Date Sampled: 05/29/2007 1230
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 1929			
Prep Method: 5030B	Date Prepared:	06/05/2007 1929			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	ND	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,2-Dichloroethene	ND	ug/L	0.20	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-4-1
Lab Sample ID: 500-4479-5

Date Sampled: 05/29/2007 1230
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 1929			
Prep Method: 5030B	Date Prepared:	06/05/2007 1929			
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	ND	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	ND	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0
Vinyl chloride	ND	ug/L	0.16	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	105	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	107	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-4-2
Lab Sample ID: 500-4479-6

Date Sampled: 05/29/2007 1240
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 1952			
Prep Method: 5030B	Date Prepared:	06/05/2007 1952			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	ND	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,2-Dichloroethene	ND	ug/L	0.20	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-4-2
Lab Sample ID: 500-4479-6

Date Sampled: 05/29/2007 1240
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 1952			
Prep Method: 5030B	Date Prepared:	06/05/2007 1952			
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	ND	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	1.6	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0
Vinyl chloride	ND	ug/L	0.16	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	105	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	98	%	75 - 120
Dibromofluoromethane	108	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-4-3
Lab Sample ID: 500-4479-7

Date Sampled: 05/29/2007 1250
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 2015			
Prep Method: 5030B	Date Prepared:	06/05/2007 2015			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	ND	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,2-Dichloroethene	ND	ug/L	0.20	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-4-3
Lab Sample ID: 500-4479-7

Date Sampled: 05/29/2007 1250
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 2015			
Prep Method: 5030B	Date Prepared:	06/05/2007 2015			
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	ND	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	1.5	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0
Vinyl chloride	ND	ug/L	0.16	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	106	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MS-2-1
Lab Sample ID: 500-4479-8

Date Sampled: 05/29/2007 1400
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 2038			
Prep Method: 5030B	Date Prepared:	06/05/2007 2038			
1,1,1,2-Tetrachloroethane	ND	ug/L	3.3	10	10
1,1,1-Trichloroethane	ND	ug/L	1.7	10	10
1,1,2,2-Tetrachloroethane	ND	ug/L	3.4	10	10
1,1,2-Trichloroethane	ND	ug/L	2.4	10	10
1,1-Dichloroethane	11	ug/L	1.5	10	10
1,1-Dichloroethene	25	ug/L	2.5	10	10
1,1-Dichloropropene	ND	ug/L	3.8	10	10
1,2,3-Trichlorobenzene	ND	ug/L	4.3	10	10
1,2,3-Trichloropropane	ND	ug/L	3.5	10	10
1,2,4-Trichlorobenzene	ND	ug/L	3.6	10	10
1,2,4-Trimethylbenzene	ND	ug/L	2.6	10	10
1,2-Dichloroethane	ND	ug/L	2.5	10	10
1,2-Dichlorobenzene	ND	ug/L	2.9	10	10
1,2-Dichloropropane	ND	ug/L	1.9	10	10
1,3,5-Trimethylbenzene	ND	ug/L	1.8	10	10
1,3-Dichlorobenzene	ND	ug/L	2.1	10	10
1,3-Dichloropropane	ND	ug/L	2.2	10	10
1,4-Dichlorobenzene	ND	ug/L	2.5	10	10
2,2-Dichloropropane	ND	ug/L	1.7	10	10
2-Hexanone	ND	ug/L	9.9	50	10
4-Methyl-2-pentanone (MIBK)	ND	ug/L	9.2	50	10
Acetone	ND	ug/L	14	50	10
Benzene	ND	ug/L	2.3	10	10
Bromobenzene	ND	ug/L	2.2	10	10
Bromochloromethane	ND	ug/L	2.7	10	10
Bromodichloromethane	ND	ug/L	2.2	10	10
Bromomethane	ND	ug/L	5.9	10	10
Bromoform	ND	ug/L	3.2	10	10
n-Butylbenzene	ND	ug/L	3.5	10	10
cis-1,3-Dichloropropene	ND	ug/L	1.5	10	10
Carbon disulfide	ND	ug/L	1.5	50	10
Carbon tetrachloride	ND	ug/L	3.4	10	10
Chlorobenzene	ND	ug/L	1.5	10	10
Chloroethane	ND	ug/L	3.2	10	10
Chloroform	ND	ug/L	1.4	10	10
Chloromethane	ND	ug/L	2.0	10	10
1,2-Dibromo-3-Chloropropane	ND	ug/L	4.1	10	10
Dibromochloromethane	ND	ug/L	2.2	10	10
Dibromomethane	ND	ug/L	2.1	10	10
Dichlorodifluoromethane	ND	ug/L	1.2	10	10

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MS-2-1
Lab Sample ID: 500-4479-8

Date Sampled: 05/29/2007 1400
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/05/2007 2038			
Prep Method: 5030B	Date Prepared:	06/05/2007 2038			
1,2-Dibromoethane	ND	ug/L	3.3	10	10
Ethylbenzene	ND	ug/L	2.1	10	10
Hexachlorobutadiene	ND	ug/L	3.6	10	10
Isopropylbenzene	ND	ug/L	2.0	10	10
p-Isopropyltoluene	ND	ug/L	2.9	10	10
2-Butanone (MEK)	ND	ug/L	10	50	10
Methylene Chloride	ND	ug/L	2.4	10	10
m&p-Xylene	ND	ug/L	3.6	20	10
Methyl tert-butyl ether	ND	ug/L	2.1	10	10
Naphthalene	ND	ug/L	3.7	10	10
2-Chlorotoluene	ND	ug/L	1.6	10	10
o-Xylene	ND	ug/L	1.9	10	10
Tetrachloroethene	21	ug/L	1.8	10	10
4-Chlorotoluene	ND	ug/L	1.8	10	10
N-Propylbenzene	ND	ug/L	1.6	10	10
sec-Butylbenzene	ND	ug/L	1.9	10	10
Styrene	ND	ug/L	1.8	10	10
trans-1,2-Dichloroethene	41	ug/L	2.9	10	10
trans-1,3-Dichloropropene	ND	ug/L	1.6	10	10
tert-Butylbenzene	ND	ug/L	1.6	10	10
Trichlorofluoromethane	ND	ug/L	1.4	10	10
Toluene	ND	ug/L	1.8	10	10
Vinyl chloride	19	ug/L	1.6	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104	%	70 - 125
Toluene-d8 (Surr)	102	%	75 - 120
4-Bromofluorobenzene (Surr)	98	%	75 - 120
Dibromofluoromethane	105	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/06/2007 1819			
Prep Method: 5030B	Date Prepared:	06/06/2007 1819			
cis-1,2-Dichloroethene	7300	ug/L	20	100	100
Trichloroethene	7800	ug/L	13	100	100

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	99	%	75 - 120
Dibromofluoromethane	111	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MS-2-1
Lab Sample ID: 500-4479-8

Date Sampled: 05/29/2007 1400
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL	Date Analyzed:	06/06/2007 1819			
Prep Method: 5030B	Date Prepared:	06/06/2007 1819			

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MS-2-2
Lab Sample ID: 500-4479-9

Date Sampled: 05/29/2007 1410
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 1842			
Prep Method: 5030B	Date Prepared:	06/06/2007 1842			
1,1,1,2-Tetrachloroethane	ND	ug/L	1.7	5.0	5.0
1,1,1-Trichloroethane	ND	ug/L	0.85	5.0	5.0
1,1,2,2-Tetrachloroethane	ND	ug/L	1.7	5.0	5.0
1,1,2-Trichloroethane	ND	ug/L	1.2	5.0	5.0
1,1-Dichloroethane	ND	ug/L	0.75	5.0	5.0
1,1-Dichloroethene	ND	ug/L	1.3	5.0	5.0
1,1-Dichloropropene	ND	ug/L	1.9	5.0	5.0
1,2,3-Trichlorobenzene	ND	ug/L	2.2	5.0	5.0
1,2,3-Trichloropropane	ND	ug/L	1.8	5.0	5.0
1,2,4-Trichlorobenzene	ND	ug/L	1.8	5.0	5.0
1,2,4-Trimethylbenzene	ND	ug/L	1.3	5.0	5.0
1,2-Dichloroethane	ND	ug/L	1.3	5.0	5.0
1,2-Dichlorobenzene	ND	ug/L	1.5	5.0	5.0
1,2-Dichloropropane	ND	ug/L	0.95	5.0	5.0
1,3,5-Trimethylbenzene	ND	ug/L	0.90	5.0	5.0
1,3-Dichlorobenzene	ND	ug/L	1.1	5.0	5.0
1,3-Dichloropropane	ND	ug/L	1.1	5.0	5.0
1,4-Dichlorobenzene	ND	ug/L	1.3	5.0	5.0
2,2-Dichloropropane	ND	ug/L	0.85	5.0	5.0
2-Hexanone	ND	ug/L	5.0	25	5.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	4.6	25	5.0
Acetone	ND	ug/L	7.0	25	5.0
Benzene	ND	ug/L	1.2	5.0	5.0
Bromobenzene	ND	ug/L	1.1	5.0	5.0
Bromochloromethane	ND	ug/L	1.4	5.0	5.0
Bromodichloromethane	ND	ug/L	1.1	5.0	5.0
Bromomethane	ND	ug/L	3.0	5.0	5.0
Bromoform	ND	ug/L	1.6	5.0	5.0
n-Butylbenzene	ND	ug/L	1.8	5.0	5.0
cis-1,3-Dichloropropene	ND	ug/L	0.75	5.0	5.0
Carbon disulfide	ND	ug/L	0.75	25	5.0
Carbon tetrachloride	ND	ug/L	1.7	5.0	5.0
Chlorobenzene	ND	ug/L	0.75	5.0	5.0
Chloroethane	ND	ug/L	1.6	5.0	5.0
Chloroform	ND	ug/L	0.70	5.0	5.0
Chloromethane	ND	ug/L	1.0	5.0	5.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	2.1	5.0	5.0
Dibromochloromethane	ND	ug/L	1.1	5.0	5.0
Dibromomethane	ND	ug/L	1.1	5.0	5.0
Dichlorodifluoromethane	ND	ug/L	0.60	5.0	5.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MS-2-2
Lab Sample ID: 500-4479-9

Date Sampled: 05/29/2007 1410
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 1842			
Prep Method: 5030B	Date Prepared:	06/06/2007 1842			
1,2-Dibromoethane	ND	ug/L	1.7	5.0	5.0
Ethylbenzene	ND	ug/L	1.1	5.0	5.0
Hexachlorobutadiene	ND	ug/L	1.8	5.0	5.0
Isopropylbenzene	ND	ug/L	1.0	5.0	5.0
p-Isopropyltoluene	ND	ug/L	1.5	5.0	5.0
2-Butanone (MEK)	ND	ug/L	5.0	25	5.0
Methylene Chloride	ND	ug/L	1.2	5.0	5.0
m&p-Xylene	ND	ug/L	1.8	10	5.0
Methyl tert-butyl ether	ND	ug/L	1.1	5.0	5.0
Naphthalene	ND	ug/L	1.9	5.0	5.0
2-Chlorotoluene	ND	ug/L	0.80	5.0	5.0
o-Xylene	ND	ug/L	0.95	5.0	5.0
Tetrachloroethene	ND	ug/L	0.90	5.0	5.0
4-Chlorotoluene	ND	ug/L	0.90	5.0	5.0
N-Propylbenzene	ND	ug/L	0.80	5.0	5.0
sec-Butylbenzene	ND	ug/L	0.95	5.0	5.0
Styrene	ND	ug/L	0.90	5.0	5.0
trans-1,2-Dichloroethene	7.5	ug/L	1.5	5.0	5.0
trans-1,3-Dichloropropene	ND	ug/L	0.80	5.0	5.0
tert-Butylbenzene	ND	ug/L	0.80	5.0	5.0
Trichloroethene	95	ug/L	0.65	5.0	5.0
Trichlorofluoromethane	ND	ug/L	0.70	5.0	5.0
Toluene	ND	ug/L	0.90	5.0	5.0
Vinyl chloride	52	ug/L	0.80	5.0	5.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	106	%	70 - 125
Toluene-d8 (Surr)	102	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	115	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/06/2007 1905			
Prep Method: 5030B	Date Prepared:	06/06/2007 1905			
cis-1,2-Dichloroethene	1600	ug/L	10	50	50

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	114	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MS-2-2
Lab Sample ID: 500-4479-9

Date Sampled: 05/29/2007 1410
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL	Date Analyzed:	06/06/2007 1905			
Prep Method: 5030B	Date Prepared:	06/06/2007 1905			

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: RW-3
Lab Sample ID: 500-4479-10

Date Sampled: 05/29/2007 1420
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 1928			
Prep Method: 5030B	Date Prepared:	06/06/2007 1928			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	1.4	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: RW-3
Lab Sample ID: 500-4479-10

Date Sampled: 05/29/2007 1420
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 1928			
Prep Method: 5030B	Date Prepared:	06/06/2007 1928			
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	2.3	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	4.5	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	99	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	98	%	75 - 120
Dibromofluoromethane	97	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/06/2007 1950			
Prep Method: 5030B	Date Prepared:	06/06/2007 1950			
cis-1,2-Dichloroethene	450	ug/L	2.0	10	10
Vinyl chloride	140	ug/L	1.6	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	100	%	75 - 120
Dibromofluoromethane	98	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: RW-3
Lab Sample ID: 500-4479-10

Date Sampled: 05/29/2007 1420
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL	Date Analyzed:	06/06/2007 1950			
Prep Method: 5030B	Date Prepared:	06/06/2007 1950			
Method: TOT-REC-200.7 Rev 4.4	Date Analyzed:	06/06/2007 1823			
Prep Method: 200.7	Date Prepared:	06/05/2007 0940			
Sodium	9.8	mg/L	0.17	0.50	1.0
Method: 375.4	Date Analyzed:	06/02/2007 1228			
Sulfate	42	mg/L	1.7	10	2.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/AS-1
Lab Sample ID: 500-4479-11

Date Sampled: 05/29/2007 1450
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 2013			
Prep Method: 5030B	Date Prepared:	06/06/2007 2013			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	ND	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,2-Dichloroethene	1.5	ug/L	0.20	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/AS-1
Lab Sample ID: 500-4479-11

Date Sampled: 05/29/2007 1450
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 2013			
Prep Method: 5030B	Date Prepared:	06/06/2007 2013			
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	ND	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	ND	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0
Vinyl chloride	15	ug/L	0.16	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	100	%	75 - 120
Dibromofluoromethane	100	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/AS-2
Lab Sample ID: 500-4479-12

Date Sampled: 05/29/2007 1500
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 2036			
Prep Method: 5030B	Date Prepared:	06/06/2007 2036			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	ND	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,2-Dichloroethene	1.7	ug/L	0.20	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/AS-2
Lab Sample ID: 500-4479-12

Date Sampled: 05/29/2007 1500
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 2036			
Prep Method: 5030B	Date Prepared:	06/06/2007 2036			
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	ND	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	ND	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0
Vinyl chloride	15	ug/L	0.16	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	102	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	102	%	75 - 120
Dibromofluoromethane	100	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/A-1
Lab Sample ID: 500-4479-13

Date Sampled: 05/29/2007 1505
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 2059			
Prep Method: 5030B	Date Prepared:	06/06/2007 2059			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	1.3	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	2.2	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	2.5	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/A-1
Lab Sample ID: 500-4479-13

Date Sampled: 05/29/2007 1505
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 2059			
Prep Method: 5030B	Date Prepared:	06/06/2007 2059			
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	4.0	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	ND	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	98	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/06/2007 2121			
Prep Method: 5030B	Date Prepared:	06/06/2007 2121			
Vinyl chloride	320	ug/L	1.6	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	102	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/A-1
Lab Sample ID: 500-4479-13

Date Sampled: 05/29/2007 1505
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL2	Date Analyzed:	06/07/2007 1120			
Prep Method: 5030B	Date Prepared:	06/07/2007 1120			
cis-1,2-Dichloroethene	1000	ug/L	4.0	20	20

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	99	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	101	%	75 - 120
Dibromofluoromethane	95	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/A-2
Lab Sample ID: 500-4479-14

Date Sampled: 05/29/2007 1515
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 2144			
Prep Method: 5030B	Date Prepared:	06/06/2007 2144			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	1.2	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	2.3	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	2.7	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/A-2
Lab Sample ID: 500-4479-14

Date Sampled: 05/29/2007 1515
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 2144			
Prep Method: 5030B	Date Prepared:	06/06/2007 2144			
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	4.9	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	ND	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	102	%	75 - 120
Dibromofluoromethane	102	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/06/2007 2207			
Prep Method: 5030B	Date Prepared:	06/06/2007 2207			
Vinyl chloride	370	ug/L	1.6	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	102	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	98	%	75 - 120
Dibromofluoromethane	102	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/A-2
Lab Sample ID: 500-4479-14

Date Sampled: 05/29/2007 1515
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL2	Date Analyzed:	06/07/2007 1142			
Prep Method: 5030B	Date Prepared:	06/07/2007 1142			
cis-1,2-Dichloroethene	1100	ug/L	4.0	20	20

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	99	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	97	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: RW-2
Lab Sample ID: 500-4479-15

Date Sampled: 05/29/2007 1525
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 2230			
Prep Method: 5030B	Date Prepared:	06/06/2007 2230			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	1.7	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	7.7	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	4.1	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: RW-2
Lab Sample ID: 500-4479-15

Date Sampled: 05/29/2007 1525
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 2230			
Prep Method: 5030B	Date Prepared:	06/06/2007 2230			
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	1.3	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	30	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104	%	70 - 125
Toluene-d8 (Surr)	103	%	75 - 120
4-Bromofluorobenzene (Surr)	96	%	75 - 120
Dibromofluoromethane	100	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/06/2007 2252			
Prep Method: 5030B	Date Prepared:	06/06/2007 2252			
Trichloroethene	670	ug/L	1.3	10	10
Vinyl chloride	250	ug/L	1.6	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	99	%	75 - 120
Dibromofluoromethane	105	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: RW-2
Lab Sample ID: 500-4479-15

Date Sampled: 05/29/2007 1525
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL2	Date Analyzed:	06/07/2007 1205			
Prep Method: 5030B	Date Prepared:	06/07/2007 1205			
cis-1,2-Dichloroethene	2600	ug/L	10	50	50
Surrogate				Acceptance Limits	
1,2-Dichloroethane-d4 (Surr)	100	%		70 - 125	
Toluene-d8 (Surr)	100	%		75 - 120	
4-Bromofluorobenzene (Surr)	99	%		75 - 120	
Dibromofluoromethane	97	%		75 - 120	
Method: TOT-REC-200.7 Rev 4.4	Date Analyzed:	06/06/2007 1829			
Prep Method: 200.7	Date Prepared:	06/05/2007 0940			
Sodium	13	mg/L	0.17	0.50	1.0
Method: 375.4	Date Analyzed:	06/02/2007 1229			
Sulfate	21	mg/L	1.7	10	2.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/B-1
Lab Sample ID: 500-4479-16

Date Sampled: 05/29/2007 1550
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 2315			
Prep Method: 5030B	Date Prepared:	06/06/2007 2315			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	ND	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/B-1
Lab Sample ID: 500-4479-16

Date Sampled: 05/29/2007 1550
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/06/2007 2315			
Prep Method: 5030B	Date Prepared:	06/06/2007 2315			
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	3.4	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	ND	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	96	%	75 - 120
Dibromofluoromethane	104	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/06/2007 2338			
Prep Method: 5030B	Date Prepared:	06/06/2007 2338			
cis-1,2-Dichloroethene	410	ug/L	2.0	10	10
Vinyl chloride	320	ug/L	1.6	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	105	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	98	%	75 - 120
Dibromofluoromethane	103	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/B-1
Lab Sample ID: 500-4479-16

Date Sampled: 05/29/2007 1550
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL	Date Analyzed:	06/06/2007 2338			
Prep Method: 5030B	Date Prepared:	06/06/2007 2338			

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/B-2
Lab Sample ID: 500-4479-17

Date Sampled: 05/29/2007 1600
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 0001			
Prep Method: 5030B	Date Prepared:	06/07/2007 0001			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	ND	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/B-2
Lab Sample ID: 500-4479-17

Date Sampled: 05/29/2007 1600
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 0001			
Prep Method: 5030B	Date Prepared:	06/07/2007 0001			
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	3.2	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	ND	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103	%	70 - 125
Toluene-d8 (Surr)	102	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	102	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 0024			
Prep Method: 5030B	Date Prepared:	06/07/2007 0024			
cis-1,2-Dichloroethene	400	ug/L	2.0	10	10
Vinyl chloride	340	ug/L	1.6	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	96	%	75 - 120
Dibromofluoromethane	102	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/B-2
Lab Sample ID: 500-4479-17

Date Sampled: 05/29/2007 1600
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 0024			
Prep Method: 5030B	Date Prepared:	06/07/2007 0024			

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-5-2
Lab Sample ID: 500-4479-18

Date Sampled: 05/29/2007 1610
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 0046			
Prep Method: 5030B	Date Prepared:	06/07/2007 0046			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	3.1	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-5-2
Lab Sample ID: 500-4479-18

Date Sampled: 05/29/2007 1610
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 0046			
Prep Method: 5030B	Date Prepared:	06/07/2007 0046			
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	ND	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0
Vinyl chloride	ND	ug/L	0.16	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104	%	70 - 125
Toluene-d8 (Surr)	102	%	75 - 120
4-Bromofluorobenzene (Surr)	95	%	75 - 120
Dibromofluoromethane	100	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 0109			
Prep Method: 5030B	Date Prepared:	06/07/2007 0109			
cis-1,2-Dichloroethene	190	ug/L	2.0	10	10
Trichloroethene	250	ug/L	1.3	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	106	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-5-2
Lab Sample ID: 500-4479-18

Date Sampled: 05/29/2007 1610
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 0109			
Prep Method: 5030B	Date Prepared:	06/07/2007 0109			

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-5-3
Lab Sample ID: 500-4479-19

Date Sampled: 05/29/2007 1620
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 0132			
Prep Method: 5030B	Date Prepared:	06/07/2007 0132			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	2.8	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-5-3
Lab Sample ID: 500-4479-19

Date Sampled: 05/29/2007 1620
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 0132			
Prep Method: 5030B	Date Prepared:	06/07/2007 0132			
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	1.8	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0
Vinyl chloride	ND	ug/L	0.16	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	105	%	70 - 125
Toluene-d8 (Surr)	102	%	75 - 120
4-Bromofluorobenzene (Surr)	100	%	75 - 120
Dibromofluoromethane	107	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 0155			
Prep Method: 5030B	Date Prepared:	06/07/2007 0155			
cis-1,2-Dichloroethene	180	ug/L	2.0	10	10
Trichloroethene	250	ug/L	1.3	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	107	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	98	%	75 - 120
Dibromofluoromethane	107	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: MW-5-3
Lab Sample ID: 500-4479-19

Date Sampled: 05/29/2007 1620
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 0155			
Prep Method: 5030B	Date Prepared:	06/07/2007 0155			

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: RW-8
Lab Sample ID: 500-4479-20

Date Sampled: 05/29/2007 1625
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 0217			
Prep Method: 5030B	Date Prepared:	06/07/2007 0217			
1,1,1,2-Tetrachloroethane	ND	ug/L	1.7	5.0	5.0
1,1,1-Trichloroethane	ND	ug/L	0.85	5.0	5.0
1,1,2,2-Tetrachloroethane	ND	ug/L	1.7	5.0	5.0
1,1,2-Trichloroethane	ND	ug/L	1.2	5.0	5.0
1,1-Dichloroethane	ND	ug/L	0.75	5.0	5.0
1,1-Dichloroethene	7.5	ug/L	1.3	5.0	5.0
1,1-Dichloropropene	ND	ug/L	1.9	5.0	5.0
1,2,3-Trichlorobenzene	ND	ug/L	2.2	5.0	5.0
1,2,3-Trichloropropane	ND	ug/L	1.8	5.0	5.0
1,2,4-Trichlorobenzene	ND	ug/L	1.8	5.0	5.0
1,2,4-Trimethylbenzene	ND	ug/L	1.3	5.0	5.0
1,2-Dichloroethane	ND	ug/L	1.3	5.0	5.0
1,2-Dichlorobenzene	ND	ug/L	1.5	5.0	5.0
1,2-Dichloropropane	ND	ug/L	0.95	5.0	5.0
1,3,5-Trimethylbenzene	ND	ug/L	0.90	5.0	5.0
1,3-Dichlorobenzene	ND	ug/L	1.1	5.0	5.0
1,3-Dichloropropane	ND	ug/L	1.1	5.0	5.0
1,4-Dichlorobenzene	ND	ug/L	1.3	5.0	5.0
2,2-Dichloropropane	ND	ug/L	0.85	5.0	5.0
2-Hexanone	ND	ug/L	5.0	25	5.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	4.6	25	5.0
Acetone	ND	ug/L	7.0	25	5.0
Benzene	ND	ug/L	1.2	5.0	5.0
Bromobenzene	ND	ug/L	1.1	5.0	5.0
Bromochloromethane	ND	ug/L	1.4	5.0	5.0
Bromodichloromethane	ND	ug/L	1.1	5.0	5.0
Bromomethane	ND	ug/L	3.0	5.0	5.0
Bromoform	ND	ug/L	1.6	5.0	5.0
n-Butylbenzene	ND	ug/L	1.8	5.0	5.0
cis-1,3-Dichloropropene	ND	ug/L	0.75	5.0	5.0
Carbon disulfide	ND	ug/L	0.75	25	5.0
Carbon tetrachloride	ND	ug/L	1.7	5.0	5.0
Chlorobenzene	ND	ug/L	0.75	5.0	5.0
Chloroethane	ND	ug/L	1.6	5.0	5.0
Chloroform	ND	ug/L	0.70	5.0	5.0
Chloromethane	ND	ug/L	1.0	5.0	5.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	2.1	5.0	5.0
Dibromochloromethane	ND	ug/L	1.1	5.0	5.0
Dibromomethane	ND	ug/L	1.1	5.0	5.0
Dichlorodifluoromethane	ND	ug/L	0.60	5.0	5.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: RW-8
Lab Sample ID: 500-4479-20

Date Sampled: 05/29/2007 1625
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 0217			
Prep Method: 5030B	Date Prepared:	06/07/2007 0217			
1,2-Dibromoethane	ND	ug/L	1.7	5.0	5.0
Ethylbenzene	ND	ug/L	1.1	5.0	5.0
Hexachlorobutadiene	ND	ug/L	1.8	5.0	5.0
Isopropylbenzene	ND	ug/L	1.0	5.0	5.0
p-Isopropyltoluene	ND	ug/L	1.5	5.0	5.0
2-Butanone (MEK)	ND	ug/L	5.0	25	5.0
Methylene Chloride	ND	ug/L	1.2	5.0	5.0
m&p-Xylene	ND	ug/L	1.8	10	5.0
Methyl tert-butyl ether	ND	ug/L	1.1	5.0	5.0
Naphthalene	ND	ug/L	1.9	5.0	5.0
2-Chlorotoluene	ND	ug/L	0.80	5.0	5.0
o-Xylene	ND	ug/L	0.95	5.0	5.0
Tetrachloroethene	ND	ug/L	0.90	5.0	5.0
4-Chlorotoluene	ND	ug/L	0.90	5.0	5.0
N-Propylbenzene	ND	ug/L	0.80	5.0	5.0
sec-Butylbenzene	ND	ug/L	0.95	5.0	5.0
Styrene	ND	ug/L	0.90	5.0	5.0
trans-1,2-Dichloroethene	29	ug/L	1.5	5.0	5.0
trans-1,3-Dichloropropene	ND	ug/L	0.80	5.0	5.0
tert-Butylbenzene	ND	ug/L	0.80	5.0	5.0
Trichlorofluoromethane	ND	ug/L	0.70	5.0	5.0
Toluene	ND	ug/L	0.90	5.0	5.0
Vinyl chloride	150	ug/L	0.80	5.0	5.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	105	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	109	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 0240			
Prep Method: 5030B	Date Prepared:	06/07/2007 0240			
cis-1,2-Dichloroethene	2000	ug/L	10	50	50
Trichloroethene	720	ug/L	6.5	50	50

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	105	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	107	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: RW-8
Lab Sample ID: 500-4479-20

Date Sampled: 05/29/2007 1625
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 0240			
Prep Method: 5030B	Date Prepared:	06/07/2007 0240			
Method: TOT-REC-200.7 Rev 4.4	Date Analyzed:	06/06/2007 1835			
Prep Method: 200.7	Date Prepared:	06/05/2007 0940			
Sodium	15	mg/L	0.17	0.50	1.0
Method: 375.4	Date Analyzed:	06/02/2007 1230			
Sulfate	69	mg/L	1.7	10	2.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: RW-1
Lab Sample ID: 500-4479-21

Date Sampled: 05/29/2007 1630
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 0303			
Prep Method: 5030B	Date Prepared:	06/07/2007 0303			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	1.6	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	3.0	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: RW-1
Lab Sample ID: 500-4479-21

Date Sampled: 05/29/2007 1630
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 0303			
Prep Method: 5030B	Date Prepared:	06/07/2007 0303			
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	52	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	ND	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103	%	70 - 125
Toluene-d8 (Surr)	99	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	104	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 0326			
Prep Method: 5030B	Date Prepared:	06/07/2007 0326			
Vinyl chloride	330	ug/L	1.6	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	107	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	100	%	75 - 120
Dibromofluoromethane	109	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: RW-1
Lab Sample ID: 500-4479-21

Date Sampled: 05/29/2007 1630
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL2	Date Analyzed:	06/07/2007 1227			
Prep Method: 5030B	Date Prepared:	06/07/2007 1227			
cis-1,2-Dichloroethene	1300	ug/L	4.0	20	20
Surrogate				Acceptance Limits	
1,2-Dichloroethane-d4 (Surr)	103	%		70 - 125	
Toluene-d8 (Surr)	102	%		75 - 120	
4-Bromofluorobenzene (Surr)	99	%		75 - 120	
Dibromofluoromethane	99	%		75 - 120	
Method: TOT-REC-200.7 Rev 4.4	Date Analyzed:	06/06/2007 1908			
Prep Method: 200.7	Date Prepared:	06/05/2007 0940			
Sodium	11	mg/L	0.17	0.50	1.0
Method: 375.4	Date Analyzed:	06/02/2007 1231			
Sulfate	57	mg/L	1.7	10	2.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/C-1
Lab Sample ID: 500-4479-22

Date Sampled: 05/29/2007 1650
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 0348			
Prep Method: 5030B	Date Prepared:	06/07/2007 0348			
1,1,1,2-Tetrachloroethane	ND	ug/L	3.3	10	10
1,1,1-Trichloroethane	ND	ug/L	1.7	10	10
1,1,2,2-Tetrachloroethane	ND	ug/L	3.4	10	10
1,1,2-Trichloroethane	ND	ug/L	2.4	10	10
1,1-Dichloroethane	ND	ug/L	1.5	10	10
1,1-Dichloroethene	20	ug/L	2.5	10	10
1,1-Dichloropropene	ND	ug/L	3.8	10	10
1,2,3-Trichlorobenzene	ND	ug/L	4.3	10	10
1,2,3-Trichloropropane	ND	ug/L	3.5	10	10
1,2,4-Trichlorobenzene	ND	ug/L	3.6	10	10
1,2,4-Trimethylbenzene	ND	ug/L	2.6	10	10
1,2-Dichloroethane	ND	ug/L	2.5	10	10
1,2-Dichlorobenzene	ND	ug/L	2.9	10	10
1,2-Dichloropropane	ND	ug/L	1.9	10	10
1,3,5-Trimethylbenzene	ND	ug/L	1.8	10	10
1,3-Dichlorobenzene	ND	ug/L	2.1	10	10
1,3-Dichloropropane	ND	ug/L	2.2	10	10
1,4-Dichlorobenzene	ND	ug/L	2.5	10	10
2,2-Dichloropropane	ND	ug/L	1.7	10	10
2-Hexanone	ND	ug/L	9.9	50	10
4-Methyl-2-pentanone (MIBK)	ND	ug/L	9.2	50	10
Acetone	ND	ug/L	14	50	10
Benzene	ND	ug/L	2.3	10	10
Bromobenzene	ND	ug/L	2.2	10	10
Bromochloromethane	ND	ug/L	2.7	10	10
Bromodichloromethane	ND	ug/L	2.2	10	10
Bromomethane	ND	ug/L	5.9	10	10
Bromoform	ND	ug/L	3.2	10	10
n-Butylbenzene	ND	ug/L	3.5	10	10
cis-1,3-Dichloropropene	ND	ug/L	1.5	10	10
Carbon disulfide	ND	ug/L	1.5	50	10
Carbon tetrachloride	ND	ug/L	3.4	10	10
Chlorobenzene	ND	ug/L	1.5	10	10
Chloroethane	ND	ug/L	3.2	10	10
Chloroform	ND	ug/L	1.4	10	10
Chloromethane	ND	ug/L	2.0	10	10
1,2-Dibromo-3-Chloropropane	ND	ug/L	4.1	10	10
Dibromochloromethane	ND	ug/L	2.2	10	10
Dibromomethane	ND	ug/L	2.1	10	10
Dichlorodifluoromethane	ND	ug/L	1.2	10	10

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/C-1
Lab Sample ID: 500-4479-22

Date Sampled: 05/29/2007 1650
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 0348			
Prep Method: 5030B	Date Prepared:	06/07/2007 0348			
1,2-Dibromoethane	ND	ug/L	3.3	10	10
Ethylbenzene	ND	ug/L	2.1	10	10
Hexachlorobutadiene	ND	ug/L	3.6	10	10
Isopropylbenzene	ND	ug/L	2.0	10	10
p-Isopropyltoluene	ND	ug/L	2.9	10	10
2-Butanone (MEK)	ND	ug/L	10	50	10
Methylene Chloride	ND	ug/L	2.4	10	10
m&p-Xylene	ND	ug/L	3.6	20	10
Methyl tert-butyl ether	ND	ug/L	2.1	10	10
Naphthalene	ND	ug/L	3.7	10	10
2-Chlorotoluene	ND	ug/L	1.6	10	10
o-Xylene	ND	ug/L	1.9	10	10
Tetrachloroethene	ND	ug/L	1.8	10	10
4-Chlorotoluene	ND	ug/L	1.8	10	10
N-Propylbenzene	ND	ug/L	1.6	10	10
sec-Butylbenzene	ND	ug/L	1.9	10	10
Styrene	ND	ug/L	1.8	10	10
trans-1,3-Dichloropropene	ND	ug/L	1.6	10	10
tert-Butylbenzene	ND	ug/L	1.6	10	10
Trichloroethene	ND	ug/L	1.3	10	10
Trichlorofluoromethane	ND	ug/L	1.4	10	10
Toluene	ND	ug/L	1.8	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	105	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	99	%	75 - 120
Dibromofluoromethane	109	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 0411			
Prep Method: 5030B	Date Prepared:	06/07/2007 0411			
trans-1,2-Dichloroethene	2000	ug/L	29	100	100
Vinyl chloride	1500	ug/L	16	100	100

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	105	%	70 - 125
Toluene-d8 (Surr)	100	%	75 - 120
4-Bromofluorobenzene (Surr)	94	%	75 - 120
Dibromofluoromethane	108	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/C-1
Lab Sample ID: 500-4479-22

Date Sampled: 05/29/2007 1650
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL2	Date Analyzed:	06/07/2007 1250			
Prep Method: 5030B	Date Prepared:	06/07/2007 1250			
cis-1,2-Dichloroethene	12000	ug/L	100	500	500

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	102	%	70 - 125
Toluene-d8 (Surr)	99	%	75 - 120
4-Bromofluorobenzene (Surr)	100	%	75 - 120
Dibromofluoromethane	101	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/C-2
Lab Sample ID: 500-4479-23

Date Sampled: 05/29/2007 1655
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 1313			
Prep Method: 5030B	Date Prepared:	06/07/2007 1313			
1,1,1,2-Tetrachloroethane	ND	ug/L	3.3	10	10
1,1,1-Trichloroethane	ND	ug/L	1.7	10	10
1,1,2,2-Tetrachloroethane	ND	ug/L	3.4	10	10
1,1,2-Trichloroethane	ND	ug/L	2.4	10	10
1,1-Dichloroethane	ND	ug/L	1.5	10	10
1,1-Dichloroethene	13	ug/L	2.5	10	10
1,1-Dichloropropene	ND	ug/L	3.8	10	10
1,2,3-Trichlorobenzene	ND	ug/L	4.3	10	10
1,2,3-Trichloropropane	ND	ug/L	3.5	10	10
1,2,4-Trichlorobenzene	ND	ug/L	3.6	10	10
1,2,4-Trimethylbenzene	ND	ug/L	2.6	10	10
1,2-Dichloroethane	ND	ug/L	2.5	10	10
1,2-Dichlorobenzene	ND	ug/L	2.9	10	10
1,2-Dichloropropane	ND	ug/L	1.9	10	10
1,3,5-Trimethylbenzene	ND	ug/L	1.8	10	10
1,3-Dichlorobenzene	ND	ug/L	2.1	10	10
1,3-Dichloropropane	ND	ug/L	2.2	10	10
1,4-Dichlorobenzene	ND	ug/L	2.5	10	10
2,2-Dichloropropane	ND	ug/L	1.7	10	10
2-Hexanone	ND	ug/L	9.9	50	10
4-Methyl-2-pentanone (MIBK)	ND	ug/L	9.2	50	10
Acetone	ND	ug/L	14	50	10
Benzene	12	ug/L	2.3	10	10
Bromobenzene	ND	ug/L	2.2	10	10
Bromochloromethane	ND	ug/L	2.7	10	10
Bromodichloromethane	ND	ug/L	2.2	10	10
Bromomethane	ND	ug/L	5.9	10	10
Bromoform	ND	ug/L	3.2	10	10
n-Butylbenzene	ND	ug/L	3.5	10	10
cis-1,3-Dichloropropene	ND	ug/L	1.5	10	10
Carbon disulfide	ND	ug/L	1.5	50	10
Carbon tetrachloride	ND	ug/L	3.4	10	10
Chlorobenzene	ND	ug/L	1.5	10	10
Chloroethane	ND	ug/L	3.2	10	10
Chloroform	ND	ug/L	1.4	10	10
Chloromethane	ND	ug/L	2.0	10	10
1,2-Dibromo-3-Chloropropane	ND	ug/L	4.1	10	10
Dibromochloromethane	ND	ug/L	2.2	10	10
Dibromomethane	ND	ug/L	2.1	10	10
Dichlorodifluoromethane	ND	ug/L	1.2	10	10

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/C-2
Lab Sample ID: 500-4479-23

Date Sampled: 05/29/2007 1655
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 1313			
Prep Method: 5030B	Date Prepared:	06/07/2007 1313			
1,2-Dibromoethane	ND	ug/L	3.3	10	10
Ethylbenzene	ND	ug/L	2.1	10	10
Hexachlorobutadiene	ND	ug/L	3.6	10	10
Isopropylbenzene	ND	ug/L	2.0	10	10
p-Isopropyltoluene	ND	ug/L	2.9	10	10
2-Butanone (MEK)	ND	ug/L	10	50	10
Methylene Chloride	ND	ug/L	2.4	10	10
m&p-Xylene	ND	ug/L	3.6	20	10
Methyl tert-butyl ether	ND	ug/L	2.1	10	10
Naphthalene	ND	ug/L	3.7	10	10
2-Chlorotoluene	ND	ug/L	1.6	10	10
o-Xylene	ND	ug/L	1.9	10	10
Tetrachloroethene	ND	ug/L	1.8	10	10
4-Chlorotoluene	ND	ug/L	1.8	10	10
N-Propylbenzene	ND	ug/L	1.6	10	10
sec-Butylbenzene	ND	ug/L	1.9	10	10
Styrene	ND	ug/L	1.8	10	10
trans-1,2-Dichloroethene	910	ug/L	2.9	10	10
trans-1,3-Dichloropropene	ND	ug/L	1.6	10	10
tert-Butylbenzene	ND	ug/L	1.6	10	10
Trichloroethene	ND	ug/L	1.3	10	10
Trichlorofluoromethane	ND	ug/L	1.4	10	10
Toluene	ND	ug/L	1.8	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	98	%	75 - 120
Dibromofluoromethane	99	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 1335			
Prep Method: 5030B	Date Prepared:	06/07/2007 1335			
cis-1,2-Dichloroethene	9500	ug/L	20	100	100
Vinyl chloride	1600	ug/L	16	100	100

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	96	%	75 - 120
Dibromofluoromethane	99	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/C-2
Lab Sample ID: 500-4479-23

Date Sampled: 05/29/2007 1655
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 1335			
Prep Method: 5030B	Date Prepared:	06/07/2007 1335			

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/CS-1
Lab Sample ID: 500-4479-24

Date Sampled: 05/29/2007 1710
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 1358			
Prep Method: 5030B	Date Prepared:	06/07/2007 1358			
1,1,1,2-Tetrachloroethane	ND	ug/L	0.33	1.0	1.0
1,1,1-Trichloroethane	ND	ug/L	0.17	1.0	1.0
1,1,2,2-Tetrachloroethane	ND	ug/L	0.34	1.0	1.0
1,1,2-Trichloroethane	ND	ug/L	0.24	1.0	1.0
1,1-Dichloroethane	ND	ug/L	0.15	1.0	1.0
1,1-Dichloroethene	ND	ug/L	0.25	1.0	1.0
1,1-Dichloropropene	ND	ug/L	0.38	1.0	1.0
1,2,3-Trichlorobenzene	ND	ug/L	0.43	1.0	1.0
1,2,3-Trichloropropane	ND	ug/L	0.35	1.0	1.0
1,2,4-Trichlorobenzene	ND	ug/L	0.36	1.0	1.0
1,2,4-Trimethylbenzene	ND	ug/L	0.26	1.0	1.0
1,2-Dichloroethane	ND	ug/L	0.25	1.0	1.0
1,2-Dichlorobenzene	ND	ug/L	0.29	1.0	1.0
1,2-Dichloropropane	ND	ug/L	0.19	1.0	1.0
1,3,5-Trimethylbenzene	ND	ug/L	0.18	1.0	1.0
1,3-Dichlorobenzene	ND	ug/L	0.21	1.0	1.0
1,3-Dichloropropane	ND	ug/L	0.22	1.0	1.0
1,4-Dichlorobenzene	ND	ug/L	0.25	1.0	1.0
2,2-Dichloropropane	ND	ug/L	0.17	1.0	1.0
2-Hexanone	ND	ug/L	0.99	5.0	1.0
4-Methyl-2-pentanone (MIBK)	ND	ug/L	0.92	5.0	1.0
Acetone	ND	ug/L	1.4	5.0	1.0
Benzene	ND	ug/L	0.23	1.0	1.0
Bromobenzene	ND	ug/L	0.22	1.0	1.0
Bromochloromethane	ND	ug/L	0.27	1.0	1.0
Bromodichloromethane	ND	ug/L	0.22	1.0	1.0
Bromomethane	ND	ug/L	0.59	1.0	1.0
Bromoform	ND	ug/L	0.32	1.0	1.0
n-Butylbenzene	ND	ug/L	0.35	1.0	1.0
cis-1,3-Dichloropropene	ND	ug/L	0.15	1.0	1.0
Carbon disulfide	ND	ug/L	0.15	5.0	1.0
Carbon tetrachloride	ND	ug/L	0.34	1.0	1.0
Chlorobenzene	ND	ug/L	0.15	1.0	1.0
Chloroethane	ND	ug/L	0.32	1.0	1.0
Chloroform	ND	ug/L	0.14	1.0	1.0
Chloromethane	ND	ug/L	0.20	1.0	1.0
1,2-Dibromo-3-Chloropropane	ND	ug/L	0.41	1.0	1.0
Dibromochloromethane	ND	ug/L	0.22	1.0	1.0
Dibromomethane	ND	ug/L	0.21	1.0	1.0
Dichlorodifluoromethane	ND	ug/L	0.12	1.0	1.0

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/CS-1
Lab Sample ID: 500-4479-24

Date Sampled: 05/29/2007 1710
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B	Date Analyzed:	06/07/2007 1358			
Prep Method: 5030B	Date Prepared:	06/07/2007 1358			
1,2-Dibromoethane	ND	ug/L	0.33	1.0	1.0
Ethylbenzene	ND	ug/L	0.21	1.0	1.0
Hexachlorobutadiene	ND	ug/L	0.36	1.0	1.0
Isopropylbenzene	ND	ug/L	0.20	1.0	1.0
p-Isopropyltoluene	ND	ug/L	0.29	1.0	1.0
2-Butanone (MEK)	ND	ug/L	1.0	5.0	1.0
Methylene Chloride	ND	ug/L	0.24	1.0	1.0
m&p-Xylene	ND	ug/L	0.36	2.0	1.0
Methyl tert-butyl ether	ND	ug/L	0.21	1.0	1.0
Naphthalene	ND	ug/L	0.37	1.0	1.0
2-Chlorotoluene	ND	ug/L	0.16	1.0	1.0
o-Xylene	ND	ug/L	0.19	1.0	1.0
Tetrachloroethene	ND	ug/L	0.18	1.0	1.0
4-Chlorotoluene	ND	ug/L	0.18	1.0	1.0
N-Propylbenzene	ND	ug/L	0.16	1.0	1.0
sec-Butylbenzene	ND	ug/L	0.19	1.0	1.0
Styrene	ND	ug/L	0.18	1.0	1.0
trans-1,2-Dichloroethene	4.5	ug/L	0.29	1.0	1.0
trans-1,3-Dichloropropene	ND	ug/L	0.16	1.0	1.0
tert-Butylbenzene	ND	ug/L	0.16	1.0	1.0
Trichloroethene	ND	ug/L	0.13	1.0	1.0
Trichlorofluoromethane	ND	ug/L	0.14	1.0	1.0
Toluene	ND	ug/L	0.18	1.0	1.0
Vinyl chloride	ND	ug/L	0.16	1.0	1.0

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103	%	70 - 125
Toluene-d8 (Surr)	101	%	75 - 120
4-Bromofluorobenzene (Surr)	97	%	75 - 120
Dibromofluoromethane	99	%	75 - 120

Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 1421			
Prep Method: 5030B	Date Prepared:	06/07/2007 1421			
cis-1,2-Dichloroethene	820	ug/L	2.0	10	10

Surrogate			Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101	%	70 - 125
Toluene-d8 (Surr)	99	%	75 - 120
4-Bromofluorobenzene (Surr)	96	%	75 - 120
Dibromofluoromethane	102	%	75 - 120

Mr. Alan Cork
ERM Inc.
12755 Olive Blvd
Suite 110
St. Louis, MO 63141

Job Number: 500-4479-1

Client Sample ID: SWGW/CS-1
Lab Sample ID: 500-4479-24

Date Sampled: 05/29/2007 1710
Date Received: 06/01/2007 1107
Client Matrix: Water

Analyte	Result/Qualifier	Unit	MDL	RL	Dilution
Method: 8260B Run Type: DL	Date Analyzed:	06/07/2007 1421			
Prep Method: 5030B	Date Prepared:	06/07/2007 1421			

DATA REPORTING QUALIFIERS

Lab Section	Qualifier	Description

QUALITY CONTROL RESULTS

Client: ERM Inc. Job Number: 500-4479-1

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:500-16410					
LCS 500-16410/4	Lab Control Spike	T	Water	8260B	
MB 500-16410/3	Method Blank	T	Water	8260B	
500-4479-1	MW-2-1	T	Water	8260B	
500-4479-2	MW-2-2	T	Water	8260B	
500-4479-3	MW-3-2	T	Water	8260B	
500-4479-4	MW-3-3	T	Water	8260B	
500-4479-4DL	MW-3-3	T	Water	8260B	
500-4479-5	MW-4-1	T	Water	8260B	
500-4479-6	MW-4-2	T	Water	8260B	
500-4479-7	MW-4-3	T	Water	8260B	
500-4479-8	MS-2-1	T	Water	8260B	
Analysis Batch:500-16448					
LCS 500-16448/4	Lab Control Spike	T	Water	8260B	
LCSD 500-16448/5	Lab Control Spike Duplicate	T	Water	8260B	
MB 500-16448/3	Method Blank	T	Water	8260B	
500-4479-8DL	MS-2-1	T	Water	8260B	
500-4479-9	MS-2-2	T	Water	8260B	
500-4479-9DL	MS-2-2	T	Water	8260B	
500-4479-10	RW-3	T	Water	8260B	
500-4479-10DL	RW-3	T	Water	8260B	
500-4479-11	SWGW/AS-1	T	Water	8260B	
500-4479-12	SWGW/AS-2	T	Water	8260B	
500-4479-13	SWGW/A-1	T	Water	8260B	
500-4479-13DL	SWGW/A-1	T	Water	8260B	
500-4479-14	SWGW/A-2	T	Water	8260B	
500-4479-14DL	SWGW/A-2	T	Water	8260B	
500-4479-15	RW-2	T	Water	8260B	
500-4479-15DL	RW-2	T	Water	8260B	
500-4479-16	SWGW/B-1	T	Water	8260B	
500-4479-16DL	SWGW/B-1	T	Water	8260B	
500-4479-17	SWGW/B-2	T	Water	8260B	
500-4479-17DL	SWGW/B-2	T	Water	8260B	
500-4479-18	MW-5-2	T	Water	8260B	
500-4479-18DL	MW-5-2	T	Water	8260B	
500-4479-19	MW-5-3	T	Water	8260B	
500-4479-19DL	MW-5-3	T	Water	8260B	
500-4479-20	RW-8	T	Water	8260B	
500-4479-20DL	RW-8	T	Water	8260B	
500-4479-21	RW-1	T	Water	8260B	
500-4479-21DL	RW-1	T	Water	8260B	
500-4479-22	SWGW/C-1	T	Water	8260B	
500-4479-22DL	SWGW/C-1	T	Water	8260B	

Client: ERM Inc. Job Number: 500-4479-1

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:500-16528					
LCS 500-16528/4	Lab Control Spike	T	Water	8260B	
MB 500-16528/3	Method Blank	T	Water	8260B	
500-4479-13DL2	SWGW/A-1	T	Water	8260B	
500-4479-14DL2	SWGW/A-2	T	Water	8260B	
500-4479-15DL2	RW-2	T	Water	8260B	
500-4479-21DL2	RW-1	T	Water	8260B	
500-4479-22DL2	SWGW/C-1	T	Water	8260B	
500-4479-23	SWGW/C-2	T	Water	8260B	
500-4479-23DL	SWGW/C-2	T	Water	8260B	
500-4479-24	SWGW/CS-1	T	Water	8260B	
500-4479-24DL	SWGW/CS-1	T	Water	8260B	

Report Basis
T = Total

Metals					
Prep Batch: 500-16296					
LCS 500-16296/2-AA	Lab Control Spike	R	Water	200.7	
MB 500-16296/1-AA	Method Blank	R	Water	200.7	
500-4479-10	RW-3	R	Water	200.7	
500-4479-15	RW-2	R	Water	200.7	
500-4479-20	RW-8	R	Water	200.7	
500-4479-21	RW-1	R	Water	200.7	
Analysis Batch:500-16428					
LCS 500-16296/2-AA	Lab Control Spike	R	Water	200.7 Rev 4.4	500-16296
MB 500-16296/1-AA	Method Blank	R	Water	200.7 Rev 4.4	500-16296
500-4479-10	RW-3	R	Water	200.7 Rev 4.4	500-16296
500-4479-15	RW-2	R	Water	200.7 Rev 4.4	500-16296
500-4479-20	RW-8	R	Water	200.7 Rev 4.4	500-16296
500-4479-21	RW-1	R	Water	200.7 Rev 4.4	500-16296

Report Basis
R = Total Recoverable

Client: ERM Inc. Job Number: 500-4479-1

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
General Chemistry					
Analysis Batch:500-16159					
LCS 500-16159/4	Lab Control Spike	T	Water	375.4	
MB 500-16159/3	Method Blank	T	Water	375.4	
500-4479-10	RW-3	T	Water	375.4	
500-4479-15	RW-2	T	Water	375.4	
500-4479-20	RW-8	T	Water	375.4	
500-4479-21	RW-1	T	Water	375.4	

Report Basis
T = Total

Client: ERM Inc.

Job Number: 500-4479-1

Surrogate Recovery Report

8260B Volatile Organic Compounds by GC/MS

Client Matrix: Water

Lab Sample ID	Client Sample ID	(BFB) (%Rec)	(DCE) (%Rec)	(DFM) (%Rec)	(TOL) (%Rec)
500-4479-1	MW-2-1	96	103	101	102
500-4479-2	MW-2-2	98	101	102	99
500-4479-3	MW-3-2	95	100	103	101
500-4479-4	MW-3-3	98	104	102	101
500-4479-4 DL	MW-3-3	98	103	104	101
500-4479-5	MW-4-1	97	105	107	101
500-4479-6	MW-4-2	98	105	108	100
500-4479-7	MW-4-3	97	104	106	101
500-4479-8	MS-2-1	98	104	105	102
500-4479-8 DL	MS-2-1	99	103	111	100
500-4479-9	MS-2-2	97	106	115	102
500-4479-9 DL	MS-2-2	97	104	114	100
500-4479-10	RW-3	98	99	97	101
500-4479-10 DL	RW-3	100	101	98	100
500-4479-11	SWGW/AS-1	100	101	100	100
500-4479-12	SWGW/AS-2	102	102	100	100
500-4479-13	SWGW/A-1	97	101	98	100
500-4479-13 DL	SWGW/A-1	97	101	102	101

Client: ERM Inc.

Job Number: 500-4479-1

Surrogate Recovery Report

8260B Volatile Organic Compounds by GC/MS

Client Matrix: Water

		(BFB) (%Rec)	(DCE) (%Rec)	(DFM) (%Rec)	(TOL) (%Rec)
500-4479-13 DL2	SWGW/A-1	101	99	95	100
500-4479-14	SWGW/A-2	102	103	102	101
500-4479-14 DL	SWGW/A-2	98	102	102	100
500-4479-14 DL2	SWGW/A-2	97	99	97	100
500-4479-15	RW-2	96	104	100	103
500-4479-15 DL	RW-2	99	104	105	101
500-4479-15 DL2	RW-2	99	100	97	100
500-4479-16	SWGW/B-1	96	103	104	101
500-4479-16 DL	SWGW/B-1	98	105	103	100
500-4479-17	SWGW/B-2	97	103	102	102
500-4479-17 DL	SWGW/B-2	96	103	102	101
500-4479-18	MW-5-2	95	104	100	102
500-4479-18 DL	MW-5-2	97	104	106	100
500-4479-19	MW-5-3	100	105	107	102
500-4479-19 DL	MW-5-3	98	107	107	101
500-4479-20	RW-8	97	105	109	101
500-4479-20 DL	RW-8	97	105	107	101
500-4479-21	RW-1	97	103	104	99

Client: ERM Inc.

Job Number: 500-4479-1

Surrogate Recovery Report

8260B Volatile Organic Compounds by GC/MS

Client Matrix: Water

		(BFB) (%Rec)	(DCE) (%Rec)	(DFM) (%Rec)	(TOL) (%Rec)
500-4479-21 DL	RW-1	100	107	109	100
500-4479-21 DL2	RW-1	99	103	99	102
500-4479-22	SWGW/C-1	99	105	109	100
500-4479-22 DL	SWGW/C-1	94	105	108	100
500-4479-22 DL2	SWGW/C-1	100	102	101	99
500-4479-23	SWGW/C-2	98	104	99	101
500-4479-23 DL	SWGW/C-2	96	101	99	101
500-4479-24	SWGW/CS-1	97	103	99	101
500-4479-24 DL	SWGW/CS-1	96	101	102	99
LCS 500-16410/4		106	98	98	101
LCS 500-16448/4		106	98	99	102
LCS 500-16528/4		104	99	101	100
LCSD 500-16448/5		103	103	105	104
MB 500-16410/3		98	98	94	101
MB 500-16448/3		102	101	99	100
MB 500-16528/3		98	102	101	100

Surrogate		Acceptance Limits
(BFB)	4-Bromofluorobenzene (Surr)	75 - 120
(DCE)	1,2-Dichloroethane-d4 (Surr)	70 - 125
(DFM)	Dibromofluoromethane	75 - 120

Client: ERM Inc.

Job Number: 500-4479-1

Surrogate Recovery Report

8260B Volatile Organic Compounds by GC/MS

Client Matrix: Water

(TOL)	Toluene-d8 (Surr)	75 - 120

Client: ERM Inc.

Job Number: 500-4479-1

Method Blank - Batch: 500-16410

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 500-16410/3
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/05/2007 1111
Date Prepared: 06/05/2007 1111

Analysis Batch: 500-16410
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18M0605.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Result	Qual	MDL	RL
1,1,1,2-Tetrachloroethane	ND		0.33	1.0
1,1,1-Trichloroethane	ND		0.17	1.0
1,1,2,2-Tetrachloroethane	ND		0.34	1.0
1,1,2-Trichloroethane	ND		0.24	1.0
1,1-Dichloroethane	ND		0.15	1.0
1,1-Dichloroethene	ND		0.25	1.0
1,1-Dichloropropene	ND		0.38	1.0
1,2,3-Trichlorobenzene	ND		0.43	1.0
1,2,3-Trichloropropane	ND		0.35	1.0
1,2,4-Trichlorobenzene	ND		0.36	1.0
1,2,4-Trimethylbenzene	ND		0.26	1.0
1,2-Dichloroethane	ND		0.25	1.0
1,2-Dichlorobenzene	ND		0.29	1.0
1,2-Dichloropropane	ND		0.19	1.0
1,3,5-Trimethylbenzene	ND		0.18	1.0
1,3-Dichlorobenzene	ND		0.21	1.0
1,3-Dichloropropane	ND		0.22	1.0
1,4-Dichlorobenzene	ND		0.25	1.0
2,2-Dichloropropane	ND		0.17	1.0
2-Hexanone	ND		0.99	5.0
4-Methyl-2-pentanone (MIBK)	ND		0.92	5.0
Acetone	ND		1.4	5.0
Benzene	ND		0.23	1.0
Bromobenzene	ND		0.22	1.0
Bromochloromethane	ND		0.27	1.0
Bromodichloromethane	ND		0.22	1.0
Bromomethane	ND		0.59	1.0
Bromoform	ND		0.32	1.0
n-Butylbenzene	ND		0.35	1.0
cis-1,2-Dichloroethene	ND		0.20	1.0
cis-1,3-Dichloropropene	ND		0.15	1.0
Carbon disulfide	ND		0.15	5.0
Carbon tetrachloride	ND		0.34	1.0
Chlorobenzene	ND		0.15	1.0
Chloroethane	ND		0.32	1.0
Chloroform	ND		0.14	1.0
Chloromethane	ND		0.20	1.0
1,2-Dibromo-3-Chloropropane	ND		0.41	1.0
Dibromochloromethane	ND		0.22	1.0
Dibromomethane	ND		0.21	1.0
Dichlorodifluoromethane	ND		0.12	1.0

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: ERM Inc.

Job Number: 500-4479-1

Method Blank - Batch: 500-16410

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 500-16410/3
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/05/2007 1111
Date Prepared: 06/05/2007 1111

Analysis Batch: 500-16410
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18M0605.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Result	Qual	MDL	RL
1,2-Dibromoethane	ND		0.33	1.0
Ethylbenzene	ND		0.21	1.0
Hexachlorobutadiene	ND		0.36	1.0
Isopropylbenzene	ND		0.20	1.0
p-Isopropyltoluene	ND		0.29	1.0
2-Butanone (MEK)	ND		1.0	5.0
Methylene Chloride	ND		0.24	1.0
m&p-Xylene	ND		0.36	2.0
Methyl tert-butyl ether	ND		0.21	1.0
Naphthalene	ND		0.37	1.0
2-Chlorotoluene	ND		0.16	1.0
o-Xylene	ND		0.19	1.0
Tetrachloroethene	ND		0.18	1.0
4-Chlorotoluene	ND		0.18	1.0
N-Propylbenzene	ND		0.16	1.0
sec-Butylbenzene	ND		0.19	1.0
Styrene	ND		0.18	1.0
trans-1,2-Dichloroethene	ND		0.29	1.0
trans-1,3-Dichloropropene	ND		0.16	1.0
tert-Butylbenzene	ND		0.16	1.0
Trichloroethene	ND		0.13	1.0
Trichlorofluoromethane	ND		0.14	1.0
Toluene	ND		0.18	1.0
Vinyl chloride	ND		0.16	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98	70 - 125
Toluene-d8 (Surr)	101	75 - 120
4-Bromofluorobenzene (Surr)	98	75 - 120
Dibromofluoromethane	94	75 - 120

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: ERM Inc.

Job Number: 500-4479-1

Lab Control Spike - Batch: 500-16410

Method: 8260B
Preparation: 5030B

Lab Sample ID: LCS 500-16410/4
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/05/2007 1133
Date Prepared: 06/05/2007 1133

Analysis Batch: 500-16410
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18S0605.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
1,1,1,2-Tetrachloroethane	25.0	23.3	93	72 - 120	
1,1,1-Trichloroethane	25.0	22.6	90	65 - 129	
1,1,2,2-Tetrachloroethane	25.0	22.2	89	68 - 120	
1,1,2-Trichloroethane	25.0	22.4	90	59 - 135	
1,1-Dichloroethane	25.0	22.0	88	63 - 121	
1,1-Dichloroethene	25.0	21.6	86	50 - 121	
1,1-Dichloropropene	25.0	22.5	90	62 - 122	
1,2,3-Trichlorobenzene	25.0	22.2	89	62 - 120	
1,2,3-Trichloropropane	25.0	22.1	88	70 - 120	
1,2,4-Trichlorobenzene	25.0	21.3	85	63 - 120	
1,2,4-Trimethylbenzene	25.0	23.6	94	76 - 120	
1,2-Dichloroethane	25.0	21.9	88	68 - 120	
1,2-Dichlorobenzene	25.0	22.5	90	74 - 120	
1,2-Dichloropropane	25.0	22.5	90	72 - 120	
1,3,5-Trimethylbenzene	25.0	23.3	93	76 - 120	
1,3-Dichlorobenzene	25.0	22.0	88	76 - 120	
1,3-Dichloropropane	25.0	23.2	93	73 - 120	
1,4-Dichlorobenzene	25.0	22.0	88	74 - 120	
2,2-Dichloropropane	25.0	22.6	91	56 - 134	
2-Hexanone	25.0	22.6	90	54 - 139	
4-Methyl-2-pentanone (MIBK)	25.0	23.0	92	65 - 128	
Acetone	25.0	19.6	78	22 - 175	
Benzene	25.0	23.0	92	68 - 120	
Bromobenzene	25.0	22.2	89	76 - 120	
Bromochloromethane	25.0	21.1	85	61 - 125	
Bromodichloromethane	25.0	25.3	101	71 - 131	
Bromomethane	25.0	29.3	117	61 - 172	
Bromoform	25.0	20.4	82	55 - 120	
n-Butylbenzene	25.0	24.8	99	68 - 120	
cis-1,2-Dichloroethene	25.0	23.5	94	62 - 127	
cis-1,3-Dichloropropene	26.9	21.9	82	60 - 120	
Carbon disulfide	25.0	19.1	77	33 - 120	
Carbon tetrachloride	25.0	23.0	92	67 - 121	
Chlorobenzene	25.0	22.5	90	75 - 120	
Chloroethane	25.0	24.1	96	56 - 152	
Chloroform	25.0	21.7	87	65 - 127	
Chloromethane	25.0	27.5	110	50 - 140	
1,2-Dibromo-3-Chloropropane	25.0	21.7	87	60 - 120	
Dibromochloromethane	25.0	21.9	88	57 - 132	
Dibromomethane	25.0	22.9	92	71 - 120	
Dichlorodifluoromethane	25.0	28.5	114	21 - 178	

Calculations are performed before rounding to avoid round-off errors in calculated results.

Quality Control Results

Client: ERM Inc.

Job Number: 500-4479-1

Lab Control Spike - Batch: 500-16410

Method: 8260B
Preparation: 5030B

Lab Sample ID: LCS 500-16410/4
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/05/2007 1133
Date Prepared: 06/05/2007 1133

Analysis Batch: 500-16410
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18S0605.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
1,2-Dibromoethane	25.0	23.7	95	68 - 125	
Ethylbenzene	25.0	23.2	93	75 - 120	
Hexachlorobutadiene	25.0	22.7	91	54 - 131	
Isopropylbenzene	25.0	20.4	82	68 - 120	
p-Isopropyltoluene	25.0	23.1	92	71 - 120	
2-Butanone (MEK)	25.0	20.9	84	36 - 157	
Methylene Chloride	25.0	19.1	76	52 - 126	
m&p-Xylene	50.0	46.6	93	75 - 120	
Methyl tert-butyl ether	25.0	22.3	89	54 - 140	
Naphthalene	25.0	21.7	87	50 - 120	
2-Chlorotoluene	25.0	22.6	90	74 - 120	
o-Xylene	25.0	23.7	95	75 - 120	
Tetrachloroethene	25.0	21.5	86	65 - 120	
4-Chlorotoluene	25.0	22.7	91	75 - 120	
N-Propylbenzene	25.0	22.8	91	74 - 120	
sec-Butylbenzene	25.0	24.1	96	73 - 120	
Styrene	25.0	25.0	100	77 - 120	
trans-1,2-Dichloroethene	25.0	22.7	91	57 - 122	
trans-1,3-Dichloropropene	24.3	19.8	81	61 - 120	
tert-Butylbenzene	25.0	23.5	94	75 - 120	
Trichloroethene	25.0	22.8	91	73 - 120	
Trichlorofluoromethane	25.0	20.8	83	58 - 147	
Toluene	25.0	23.2	93	75 - 120	
Vinyl chloride	25.0	25.9	104	57 - 135	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98	70 - 125
Toluene-d8 (Surr)	101	75 - 120
4-Bromofluorobenzene (Surr)	106	75 - 120
Dibromofluoromethane	98	75 - 120

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: ERM Inc.

Job Number: 500-4479-1

Method Blank - Batch: 500-16448

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 500-16448/3
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/06/2007 1733
Date Prepared: 06/06/2007 1733

Analysis Batch: 500-16448
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18M0606.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Result	Qual	MDL	RL
1,1,1,2-Tetrachloroethane	ND		0.33	1.0
1,1,1-Trichloroethane	ND		0.17	1.0
1,1,2,2-Tetrachloroethane	ND		0.34	1.0
1,1,2-Trichloroethane	ND		0.24	1.0
1,1-Dichloroethane	ND		0.15	1.0
1,1-Dichloroethene	ND		0.25	1.0
1,1-Dichloropropene	ND		0.38	1.0
1,2,3-Trichlorobenzene	ND		0.43	1.0
1,2,3-Trichloropropane	ND		0.35	1.0
1,2,4-Trichlorobenzene	ND		0.36	1.0
1,2,4-Trimethylbenzene	ND		0.26	1.0
1,2-Dichloroethane	ND		0.25	1.0
1,2-Dichlorobenzene	ND		0.29	1.0
1,2-Dichloropropane	ND		0.19	1.0
1,3,5-Trimethylbenzene	ND		0.18	1.0
1,3-Dichlorobenzene	ND		0.21	1.0
1,3-Dichloropropane	ND		0.22	1.0
1,4-Dichlorobenzene	ND		0.25	1.0
2,2-Dichloropropane	ND		0.17	1.0
2-Hexanone	ND		0.99	5.0
4-Methyl-2-pentanone (MIBK)	ND		0.92	5.0
Acetone	ND		1.4	5.0
Benzene	ND		0.23	1.0
Bromobenzene	ND		0.22	1.0
Bromochloromethane	ND		0.27	1.0
Bromodichloromethane	ND		0.22	1.0
Bromomethane	ND		0.59	1.0
Bromoform	ND		0.32	1.0
n-Butylbenzene	ND		0.35	1.0
cis-1,2-Dichloroethene	ND		0.20	1.0
cis-1,3-Dichloropropene	ND		0.15	1.0
Carbon disulfide	ND		0.15	5.0
Carbon tetrachloride	ND		0.34	1.0
Chlorobenzene	ND		0.15	1.0
Chloroethane	ND		0.32	1.0
Chloroform	ND		0.14	1.0
Chloromethane	ND		0.20	1.0
1,2-Dibromo-3-Chloropropane	ND		0.41	1.0
Dibromochloromethane	ND		0.22	1.0
Dibromomethane	ND		0.21	1.0
Dichlorodifluoromethane	ND		0.12	1.0

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: ERM Inc.

Job Number: 500-4479-1

Method Blank - Batch: 500-16448

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 500-16448/3
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/06/2007 1733
Date Prepared: 06/06/2007 1733

Analysis Batch: 500-16448
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18M0606.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Result	Qual	MDL	RL
1,2-Dibromoethane	ND		0.33	1.0
Ethylbenzene	ND		0.21	1.0
Hexachlorobutadiene	ND		0.36	1.0
Isopropylbenzene	ND		0.20	1.0
p-Isopropyltoluene	ND		0.29	1.0
2-Butanone (MEK)	ND		1.0	5.0
Methylene Chloride	ND		0.24	1.0
m&p-Xylene	ND		0.36	2.0
Methyl tert-butyl ether	ND		0.21	1.0
Naphthalene	ND		0.37	1.0
2-Chlorotoluene	ND		0.16	1.0
o-Xylene	ND		0.19	1.0
Tetrachloroethene	ND		0.18	1.0
4-Chlorotoluene	ND		0.18	1.0
N-Propylbenzene	ND		0.16	1.0
sec-Butylbenzene	ND		0.19	1.0
Styrene	ND		0.18	1.0
trans-1,2-Dichloroethene	ND		0.29	1.0
trans-1,3-Dichloropropene	ND		0.16	1.0
tert-Butylbenzene	ND		0.16	1.0
Trichloroethene	ND		0.13	1.0
Trichlorofluoromethane	ND		0.14	1.0
Toluene	ND		0.18	1.0
Vinyl chloride	ND		0.16	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101	70 - 125
Toluene-d8 (Surr)	100	75 - 120
4-Bromofluorobenzene (Surr)	102	75 - 120
Dibromofluoromethane	99	75 - 120

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: ERM Inc.

Job Number: 500-4479-1

Lab Control Spike/
Lab Control Spike Duplicate Recovery Report - Batch: 500-16448

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID: LCS 500-16448/4
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/06/2007 1756
Date Prepared: 06/06/2007 1756

Analysis Batch: 500-16448
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18S0606.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

LCSD Lab Sample ID: LCSD 500-16448/5
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/07/2007 0434
Date Prepared: 06/07/2007 0434

Analysis Batch: 500-16448
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975
Lab File ID: 18T0606.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
1,1,1,2-Tetrachloroethane	97	98	72 - 120	1	20		
1,1,1-Trichloroethane	96	99	65 - 129	4	20		
1,1,2,2-Tetrachloroethane	89	95	68 - 120	7	20		
1,1,2-Trichloroethane	93	102	59 - 135	9	20		
1,1-Dichloroethane	92	96	63 - 121	5	20		
1,1-Dichloroethene	97	93	50 - 121	4	20		
1,1-Dichloropropene	94	94	62 - 122	0	20		
1,2,3-Trichlorobenzene	94	97	62 - 120	3	20		
1,2,3-Trichloropropane	87	91	70 - 120	6	20		
1,2,4-Trichlorobenzene	87	86	63 - 120	1	20		
1,2,4-Trimethylbenzene	99	100	76 - 120	1	20		
1,2-Dichloroethane	94	98	68 - 120	4	20		
1,2-Dichlorobenzene	92	98	74 - 120	7	20		
1,2-Dichloropropane	94	94	72 - 120	0	20		
1,3,5-Trimethylbenzene	98	99	76 - 120	1	20		
1,3-Dichlorobenzene	91	93	76 - 120	3	20		
1,3-Dichloropropane	95	97	73 - 120	1	20		
1,4-Dichlorobenzene	89	93	74 - 120	4	20		
2,2-Dichloropropane	95	88	56 - 134	7	20		
2-Hexanone	84	85	54 - 139	1	20		
4-Methyl-2-pentanone (MIBK)	92	91	65 - 128	1	20		
Acetone	79	84	22 - 175	6	20		
Benzene	96	96	68 - 120	0	20		
Bromobenzene	91	95	76 - 120	4	20		
Bromochloromethane	88	98	61 - 125	11	20		
Bromodichloromethane	108	112	71 - 131	4	20		
Bromomethane	90	91	61 - 172	2	20		
Bromoform	87	86	55 - 120	1	20		
n-Butylbenzene	104	102	68 - 120	2	20		
cis-1,2-Dichloroethene	97	105	62 - 127	8	20		
cis-1,3-Dichloropropene	83	85	60 - 120	2	20		
Carbon disulfide	86	80	33 - 120	7	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: ERM Inc.

Job Number: 500-4479-1

Lab Control Spike/
Lab Control Spike Duplicate Recovery Report - Batch: 500-16448

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID: LCS 500-16448/4
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/06/2007 1756
Date Prepared: 06/06/2007 1756

Analysis Batch: 500-16448
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18S0606.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

LCSD Lab Sample ID: LCSD 500-16448/5
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/07/2007 0434
Date Prepared: 06/07/2007 0434

Analysis Batch: 500-16448
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975
Lab File ID: 18T0606.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Carbon tetrachloride	102	97	67 - 121	5	20		
Chlorobenzene	93	93	75 - 120	0	20		
Chloroethane	87	91	56 - 152	5	20		
Chloroform	92	98	65 - 127	7	20		
Chloromethane	105	98	50 - 140	7	20		
1,2-Dibromo-3-Chloropropane	82	86	60 - 120	4	20		
Dibromochloromethane	91	92	57 - 132	1	20		
Dibromomethane	96	99	71 - 120	3	20		
Dichlorodifluoromethane	110	114	21 - 178	3	20		
1,2-Dibromoethane	98	101	68 - 125	4	20		
Ethylbenzene	98	94	75 - 120	4	20		
Hexachlorobutadiene	94	96	54 - 131	2	20		
Isopropylbenzene	86	84	68 - 120	2	20		
p-Isopropyltoluene	97	99	71 - 120	2	20		
2-Butanone (MEK)	99	104	36 - 157	5	20		
Methylene Chloride	86	93	52 - 126	8	20		
m&p-Xylene	98	95	75 - 120	3	20		
Methyl tert-butyl ether	93	102	54 - 140	8	20		
Naphthalene	87	91	50 - 120	4	20		
2-Chlorotoluene	95	95	74 - 120	1	20		
o-Xylene	99	98	75 - 120	0	20		
Tetrachloroethene	91	85	65 - 120	7	20		
4-Chlorotoluene	95	95	75 - 120	1	20		
N-Propylbenzene	96	94	74 - 120	3	20		
sec-Butylbenzene	101	101	73 - 120	0	20		
Styrene	104	103	77 - 120	1	20		
trans-1,2-Dichloroethene	96	100	57 - 122	4	20		
trans-1,3-Dichloropropene	86	85	61 - 120	1	20		
tert-Butylbenzene	99	102	75 - 120	3	20		
Trichloroethene	95	95	73 - 120	0	20		
Trichlorofluoromethane	91	94	58 - 147	3	20		
Toluene	98	100	75 - 120	2	20		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: ERM Inc.

Job Number: 500-4479-1

Lab Control Spike/
Lab Control Spike Duplicate Recovery Report - Batch: 500-16448

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID: LCS 500-16448/4
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/06/2007 1756
Date Prepared: 06/06/2007 1756

Analysis Batch: 500-16448
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18S0606.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

LCSD Lab Sample ID: LCSD 500-16448/5
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/07/2007 0434
Date Prepared: 06/07/2007 0434

Analysis Batch: 500-16448
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975
Lab File ID: 18T0606.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Vinyl chloride	103	102	57 - 135	1	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98	103	70 - 125
Toluene-d8 (Surr)	102	104	75 - 120
4-Bromofluorobenzene (Surr)	106	103	75 - 120
Dibromofluoromethane	99	105	75 - 120

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: ERM Inc.

Job Number: 500-4479-1

Method Blank - Batch: 500-16528

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 500-16528/3
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/07/2007 1034
Date Prepared: 06/07/2007 1034

Analysis Batch: 500-16528
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18M0607.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Result	Qual	MDL	RL
1,1,1,2-Tetrachloroethane	ND		0.33	1.0
1,1,1-Trichloroethane	ND		0.17	1.0
1,1,2,2-Tetrachloroethane	ND		0.34	1.0
1,1,2-Trichloroethane	ND		0.24	1.0
1,1-Dichloroethane	ND		0.15	1.0
1,1-Dichloroethene	ND		0.25	1.0
1,1-Dichloropropene	ND		0.38	1.0
1,2,3-Trichlorobenzene	ND		0.43	1.0
1,2,3-Trichloropropane	ND		0.35	1.0
1,2,4-Trichlorobenzene	ND		0.36	1.0
1,2,4-Trimethylbenzene	ND		0.26	1.0
1,2-Dichloroethane	ND		0.25	1.0
1,2-Dichlorobenzene	ND		0.29	1.0
1,2-Dichloropropane	ND		0.19	1.0
1,3,5-Trimethylbenzene	ND		0.18	1.0
1,3-Dichlorobenzene	ND		0.21	1.0
1,3-Dichloropropane	ND		0.22	1.0
1,4-Dichlorobenzene	ND		0.25	1.0
2,2-Dichloropropane	ND		0.17	1.0
2-Hexanone	ND		0.99	5.0
4-Methyl-2-pentanone (MIBK)	ND		0.92	5.0
Acetone	ND		1.4	5.0
Benzene	ND		0.23	1.0
Bromobenzene	ND		0.22	1.0
Bromochloromethane	ND		0.27	1.0
Bromodichloromethane	ND		0.22	1.0
Bromomethane	ND		0.59	1.0
Bromoform	ND		0.32	1.0
n-Butylbenzene	ND		0.35	1.0
cis-1,2-Dichloroethene	ND		0.20	1.0
cis-1,3-Dichloropropene	ND		0.15	1.0
Carbon disulfide	ND		0.15	5.0
Carbon tetrachloride	ND		0.34	1.0
Chlorobenzene	ND		0.15	1.0
Chloroethane	ND		0.32	1.0
Chloroform	ND		0.14	1.0
Chloromethane	ND		0.20	1.0
1,2-Dibromo-3-Chloropropane	ND		0.41	1.0
Dibromochloromethane	ND		0.22	1.0
Dibromomethane	ND		0.21	1.0
Dichlorodifluoromethane	ND		0.12	1.0

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: ERM Inc.

Job Number: 500-4479-1

Method Blank - Batch: 500-16528

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 500-16528/3
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/07/2007 1034
Date Prepared: 06/07/2007 1034

Analysis Batch: 500-16528
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18M0607.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Result	Qual	MDL	RL
1,2-Dibromoethane	ND		0.33	1.0
Ethylbenzene	ND		0.21	1.0
Hexachlorobutadiene	ND		0.36	1.0
Isopropylbenzene	ND		0.20	1.0
p-Isopropyltoluene	ND		0.29	1.0
2-Butanone (MEK)	ND		1.0	5.0
Methylene Chloride	ND		0.24	1.0
m&p-Xylene	ND		0.36	2.0
Methyl tert-butyl ether	ND		0.21	1.0
Naphthalene	ND		0.37	1.0
2-Chlorotoluene	ND		0.16	1.0
o-Xylene	ND		0.19	1.0
Tetrachloroethene	ND		0.18	1.0
4-Chlorotoluene	ND		0.18	1.0
N-Propylbenzene	ND		0.16	1.0
sec-Butylbenzene	ND		0.19	1.0
Styrene	ND		0.18	1.0
trans-1,2-Dichloroethene	ND		0.29	1.0
trans-1,3-Dichloropropene	ND		0.16	1.0
tert-Butylbenzene	ND		0.16	1.0
Trichloroethene	ND		0.13	1.0
Trichlorofluoromethane	ND		0.14	1.0
Toluene	ND		0.18	1.0
Vinyl chloride	ND		0.16	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	102	70 - 125
Toluene-d8 (Surr)	100	75 - 120
4-Bromofluorobenzene (Surr)	98	75 - 120
Dibromofluoromethane	101	75 - 120

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: ERM Inc.

Job Number: 500-4479-1

Lab Control Spike - Batch: 500-16528

Method: 8260B
Preparation: 5030B

Lab Sample ID: LCS 500-16528/4
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/07/2007 1057
Date Prepared: 06/07/2007 1057

Analysis Batch: 500-16528
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18S0607.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
1,1,1,2-Tetrachloroethane	25.0	24.2	97	72 - 120	
1,1,1-Trichloroethane	25.0	24.6	98	65 - 129	
1,1,2,2-Tetrachloroethane	25.0	22.4	89	68 - 120	
1,1,2-Trichloroethane	25.0	23.6	95	59 - 135	
1,1-Dichloroethane	25.0	23.4	94	63 - 121	
1,1-Dichloroethene	25.0	25.6	102	50 - 121	
1,1-Dichloropropene	25.0	24.1	96	62 - 122	
1,2,3-Trichlorobenzene	25.0	23.1	93	62 - 120	
1,2,3-Trichloropropane	25.0	21.8	87	70 - 120	
1,2,4-Trichlorobenzene	25.0	21.1	84	63 - 120	
1,2,4-Trimethylbenzene	25.0	24.6	98	76 - 120	
1,2-Dichloroethane	25.0	23.4	93	68 - 120	
1,2-Dichlorobenzene	25.0	23.2	93	74 - 120	
1,2-Dichloropropane	25.0	23.0	92	72 - 120	
1,3,5-Trimethylbenzene	25.0	24.5	98	76 - 120	
1,3-Dichlorobenzene	25.0	23.1	92	76 - 120	
1,3-Dichloropropane	25.0	23.0	92	73 - 120	
1,4-Dichlorobenzene	25.0	22.1	89	74 - 120	
2,2-Dichloropropane	25.0	24.8	99	56 - 134	
2-Hexanone	25.0	21.1	84	54 - 139	
4-Methyl-2-pentanone (MIBK)	25.0	23.2	93	65 - 128	
Acetone	25.0	22.1	88	22 - 175	
Benzene	25.0	24.1	96	68 - 120	
Bromobenzene	25.0	22.4	90	76 - 120	
Bromochloromethane	25.0	22.7	91	61 - 125	
Bromodichloromethane	25.0	26.6	106	71 - 131	
Bromomethane	25.0	26.7	107	61 - 172	
Bromoform	25.0	20.5	82	55 - 120	
n-Butylbenzene	25.0	26.3	105	68 - 120	
cis-1,2-Dichloroethene	25.0	25.1	100	62 - 127	
cis-1,3-Dichloropropene	26.9	22.0	82	60 - 120	
Carbon disulfide	25.0	22.7	91	33 - 120	
Carbon tetrachloride	25.0	25.4	102	67 - 121	
Chlorobenzene	25.0	23.0	92	75 - 120	
Chloroethane	25.0	24.6	98	56 - 152	
Chloroform	25.0	23.7	95	65 - 127	
Chloromethane	25.0	27.9	112	50 - 140	
1,2-Dibromo-3-Chloropropane	25.0	20.6	82	60 - 120	
Dibromochloromethane	25.0	22.3	89	57 - 132	
Dibromomethane	25.0	23.8	95	71 - 120	
Dichlorodifluoromethane	25.0	31.9	128	21 - 178	

Calculations are performed before rounding to avoid round-off errors in calculated results.

Quality Control Results

Client: ERM Inc.

Job Number: 500-4479-1

Lab Control Spike - Batch: 500-16528

Method: 8260B
Preparation: 5030B

Lab Sample ID: LCS 500-16528/4
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/07/2007 1057
Date Prepared: 06/07/2007 1057

Analysis Batch: 500-16528
Prep Batch: N/A
Units: ug/L

Instrument ID: Agilent 6890N GC - 5975N
Lab File ID: 18S0607.D
Initial Weight/Volume: 10 mL
Final Weight/Volume: 10 mL

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
1,2-Dibromoethane	25.0	24.0	96	68 - 125	
Ethylbenzene	25.0	24.0	96	75 - 120	
Hexachlorobutadiene	25.0	24.0	96	54 - 131	
Isopropylbenzene	25.0	21.2	85	68 - 120	
p-Isopropyltoluene	25.0	24.4	98	71 - 120	
2-Butanone (MEK)	25.0	25.7	103	36 - 157	
Methylene Chloride	25.0	22.8	91	52 - 126	
m&p-Xylene	50.0	48.0	96	75 - 120	
Methyl tert-butyl ether	25.0	23.7	95	54 - 140	
Naphthalene	25.0	20.9	83	50 - 120	
2-Chlorotoluene	25.0	23.4	94	74 - 120	
o-Xylene	25.0	24.8	99	75 - 120	
Tetrachloroethene	25.0	22.2	89	65 - 120	
4-Chlorotoluene	25.0	23.4	93	75 - 120	
N-Propylbenzene	25.0	23.6	94	74 - 120	
sec-Butylbenzene	25.0	25.5	102	73 - 120	
Styrene	25.0	25.6	102	77 - 120	
trans-1,2-Dichloroethene	25.0	24.4	97	57 - 122	
trans-1,3-Dichloropropene	24.3	20.1	82	61 - 120	
tert-Butylbenzene	25.0	24.9	99	75 - 120	
Trichloroethene	25.0	24.1	97	73 - 120	
Trichlorofluoromethane	25.0	26.1	104	58 - 147	
Toluene	25.0	24.3	97	75 - 120	
Vinyl chloride	25.0	28.3	113	57 - 135	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	99	70 - 125
Toluene-d8 (Surr)	100	75 - 120
4-Bromofluorobenzene (Surr)	104	75 - 120
Dibromofluoromethane	101	75 - 120

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: ERM Inc.

Job Number: 500-4479-1

Method Blank - Batch: 500-16296

Method: 200.7 Rev 4.4
Preparation: 200.7
Total Recoverable

Lab Sample ID: MB 500-16296/1-AA
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/06/2007 1733
Date Prepared: 06/05/2007 0940

Analysis Batch: 500-16428
Prep Batch: 500-16296
Units: mg/L

Instrument ID: TJA ICAP 61E Trace Analy
Lab File ID: P40606A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 25 mL

Analyte	Result	Qual	MDL	RL
Sodium	ND		0.17	0.50

Lab Control Spike - Batch: 500-16296

Method: 200.7 Rev 4.4
Preparation: 200.7
Total Recoverable

Lab Sample ID: LCS 500-16296/2-AA
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/06/2007 1739
Date Prepared: 06/05/2007 0940

Analysis Batch: 500-16428
Prep Batch: 500-16296
Units: mg/L

Instrument ID: TJA ICAP 61E Trace Analy
Lab File ID: P40606A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 25 mL

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Sodium	5.00	4.79	96	85 - 115	

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: ERM Inc.

Job Number: 500-4479-1

Method Blank - Batch: 500-16159

Method: 375.4
Preparation: N/A

Lab Sample ID: MB 500-16159/3
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/02/2007 1202
Date Prepared: N/A

Analysis Batch: 500-16159
Prep Batch: N/A
Units: mg/L

Instrument ID: ThermoSpectronic Genesys
Lab File ID: N/A
Initial Weight/Volume: mL
Final Weight/Volume: mL

Analyte	Result	Qual	MDL	RL
Sulfate	ND		0.84	5.0

Lab Control Spike - Batch: 500-16159

Method: 375.4
Preparation: N/A

Lab Sample ID: LCS 500-16159/4
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 06/02/2007 1203
Date Prepared: N/A

Analysis Batch: 500-16159
Prep Batch: N/A
Units: mg/L

Instrument ID: ThermoSpectronic Genesys
Lab File ID: N/A
Initial Weight/Volume: 100 mL
Final Weight/Volume: 100 mL

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Sulfate	20.0	20.4	102	80 - 120	

Calculations are performed before rounding to avoid round-off errors in calculated results.

SEVERN TRENT **STL**

STL Chicago
2417 Bond Street
University Park, IL 60466
Phone: 708-534-5200
Fax: 708-534-5211

Report To:
Contact: ALAN J. CORK
Company: ERM
Address: 12755 OLIVE BLVD SUITE 110
Phone: 314-682-3880
Fax: 314-682-3970
E-Mail: alan.cork@erm.com

Bill To:
Contact: SAME AS REPORT TO:
Company:
Address:
Phone:
Fax:
PO#: 0060492 Quote: 50000175

Shaded Areas For Internal Use Only 2 of 2

Lab Lot# 500-479

Package Sealed: Yes (No)	Samples Sealed: Yes (No)
Received on Ice: (Yes) No	Samples Intact: (Yes) No
Temperature °C of Cooler: 2.0 2.6	
Within Hold Time: (Yes) No	Preserv. Indicated: (Yes) No NA
pH Check OK: (Yes) No NA	Res Cl_2 Check OK: Yes No (NA)
Sample Labels and COC Agree: (Yes) No COC not present	

Sampler Name: ERM
Signature: [signature]
Project Name: SECO
Project Number: 0060492
Project Location: WASHINGTON, MO
Date Required: STD TAT
Hard Copy: ___/___/___
Fax: ___/___/___
Lab PM: R. MANNZ

Refrg #				
# / Cont.	3	1	1	1
Volume	40 ML	500 ML	500 ML	500 ML
Preserv	HCL	HNO_3	HNO_3	–

Laboratory ID	MS-MSD	Client Sample ID	Sampling Date	Sampling Time	Matrix	Comp/Grab	VOCs SW8260B	RCRA METALS SW6010B/7471	SODIUM EPA 200.7	SULFATE EPA 375.4	Additional Analyses / Remarks
13		SW6W/A-1	5/29/07	1505	W	G	X				
14		SW6W/A-2		1515	W	G	X				
15		RW-2		1525	W	G	X		X	X	
16		SW6W/B-1		1550	W	G	X				
17		SW6W/B-2		1600	W	G	X				
18		MW-5-2		1610	W	G	X				
19		MW-5-3		1620	W	G	X				
20		RW-8		1625	W	G	X		X	X	
21		RW-1		1630	W	G	X		X	X	
22		SW6W/C-1		1650	W	G	X				
23		SW6W/C-2		1655	W	G	X				
24		SW6W/CS-1		1710	W	G	X				

RELINQUISHED BY	COMPANY	DATE	TIME	RECEIVED BY	COMPANY	DATE	TIME
[signature]	ERM	5/31/07	1615	[signature]	STL	5.31.07	1615
[signature]	STL	5.31.07	1700	[signature]	STL	6-1-07	0910

Matrix Key
WW = Wastewater
W = Water
S = Soil
SL = Sludge
MS = Miscellaneous
OL = Oil
A = Air
SE = Sediment
SO = Solid
DS = Drum Solid
DL = Drum Liquid
L = Leachate
WI = Wipe
O = _____

Container Key
1. Plastic
2. VOA Vial
3. Sterile Plastic
4. Amber Glass
5. Widemouth Glass
6. Other _____

Preservative Key
1. HCl, Cool to 4°
2. H2SO4, Cool to 4°
3. HNO3, Cool to 4°
4. NaOH, Cool to 4°
5. NaOH/Zn, Cool to 4°
6. Cool to 4°
7. None

COMMENTS
SHIP TO RICH MANNZ @ STL CHICAGO

Date Received ___/___/___
Courier FX
Hand Delivered ☐
Bill of Lading

STL Chicago is a part of Severn Trent Laboratories, Inc.

STL-8208 (0600)

SEVERN TRENT **STL**

STL Chicago
2417 Bond Street
University Park, IL 60466
Phone: 708-534-5200
Fax: 708-534-5211

Report To:
Contact: ALAN J. CORK
Company: ERM
Address: 12755 OLIVE BLVD SUITE 100
Phone: 314-682-3900
Fax: 314-682-3970
E-Mail: alan.cork@erm.com

Bill To:
Contact: SAME AS REPORT TO:
Company:
Address:
Phone:
Fax:
PO#: 0060492 Quote: 50000175

Shaded Areas For Internal Use Only 1 of 6

Lab Lot# 500-4479

Package Sealed	Samples Sealed
Yes No	Yes (No)
Received on Ice	**Samples Intact**
(Yes) No	(Yes) No

Temperature °C of Cooler 2.0 / 2.6

Within Hold Time	Preserv. Indicated
(Yes) No	(Yes) No NA
pH Check OK	**Res Cl_2 Check OK**
(Yes) No NA	Yes No (NA)

Sample Labels and COC Agree: (Yes) No COC not present

Sampler Name: ERM
Signature: [signature]
Project Name: SECO
Project Number: 0060492
Project Location: WASHINGTON, MO
Date Required: STD TAT
Hard Copy: __/__/__
Fax: __/__/__
Lab PM: R. MANNZ

Laboratory ID	MS-MSD	Client Sample ID	Sampling Date	Sampling Time	Matrix	Comp/Grab	VOCs SW8260B	RCRA Metals SW6010B/7471	Sodium EPA 200.7	Sulfate EPA 375.4	Additional Analyses / Remarks
Refrig #											
# / Cont.							3	1	1	1	
Volume							40ML	500ML	500ML	500ML	
Preserv							HCL	HNO_3	HNO_3	-	
1		MW-2-1	5/29/07	1115	W	G	X				
2		MW-2-2		1120	W	G	X				
3		MW-3-2		1200	W	G	X				
4		MW-3-3		1210	W	G	X				
5		MW-4-1		1230	W	G	X				
6		MW-4-2		1240	W	G	X				
7		MW-4-3		1250	W	G	X				
8		MS-2-1		1400	W	G	X				
9		MS-2-2		1410	W	G	X				
10		RW-3		1420	W	G	X		X	X	
11		SWGW/AS-1		1450	W	G	X				
12		SWGW/AS-2		1500	W	G	X				

RELINQUISHED BY	COMPANY	DATE	TIME	RECEIVED BY	COMPANY	DATE	TIME
[signature]	ERM	5/31/07	1615	[signature]	STL	5.31.07	1615
[signature]	STL	5.31.07	1700	[signature]	STL	6-1-07	0910

Matrix Key
WW = Wastewater
W = Water
S = Soil
SL = Sludge
MS = Miscellaneous
OL = Oil
A = Air
SE = Sediment
SO = Solid
DS = Drum Solid
DL = Drum Liquid
L = Leachate
WI = Wipe
O = _____

Container Key
1. Plastic
2. VOA Vial
3. Sterile Plastic
4. Amber Glass
5. Widemouth Glass
6. Other _____

Preservative Key
1. HCl, Cool to 4°
2. H2SO4, Cool to 4°
3. HNO3, Cool to 4°
4. NaOH, Cool to 4°
5. NaOH/Zn, Cool to 4°
6. Cool to 4°
7. None

COMMENTS: SHIP TO RICH MANNZ @ STL CHICAGO

Date Received __/__/__
Courier:
Hand Delivered ☐
Bill of Lading

STL Chicago is a part of Severn Trent Laboratories, Inc.

STL-8208 (0600)

LOGIN SAMPLE RECEIPT CHECK LIST

Client: ERM Inc. Job Number: 500-4479-1

Login Number: 4479

Question	T/F/NA	Comment
Radioactivity either was not measured or, if measured, is at or below background	True	
The cooler's custody seal, if present, is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	2.0,2.6
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	True	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	