Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No.	B4760 S51 A1	Volume (L)	1209.75
Client Sample No.	SVM-H2-1FD-100204	No. of Grid Openings	22
Description		Filter Area (mm2)	385
Analysis Date	11/8/2004	Area Analyzed (mm2)	0.319
Analyst	DW	Analytical Sens. (struc/cc)	0.000998
		Dectection Limit. (struc/cc)	0.00299

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	6.3	0.00200	0.00 - 0.00629	2
Total Asbestos Structures	6.3	0.00200	0.00 - 0.00629	2
Asbestos Structures > 5um	3.1	0.000998	0.00 - 0.00473	1
Asbestos Fibers and Bundles > 5um	3.1	0.000998	0.00 - 0.00473	1
PCM Equivalent Fibers-US	3.1	0.000998	0.00 - 0.00473	1
PCM Equivalent Structures-US	3.1	0.000998	0.00 - 0.00473	1
PROTOCOL ASB STRUCS 5-10	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL ASB STRUCS >10	3.1	0.000998	0.00 - 0.00473	1
PROTOCOL ASB STRUCS TOTAL	3.1	0.000998	0.00 - 0.00473	1
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL CHRYS STRUCS >10	3.1	0.000998	0.00 - 0.00473	1
PROTOCOL CHRYS STRUCS TOTAL	3.1	0.000998	0.00 - 0.00473	1
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00299	0.00 - 0.00299	0
AHERA-like Total Structures 3:1	6.3	0.00200	0.00 - 0.00629	2
AHERA-like Asb Strucs >5 and 3:1	3.1	0.000998	0.00 - 0.00473	1
AHERA-like Asb Strucs 5 - 10 and 3:1	0.0	<0.00299	0.00 - 0.00299	0
AHERA-like Asb Strucs >10 and 3:1	3.1	0.000998	0.00 - 0.00473	1
Total Other Amphibole Strucs 3:1	0.0	<0.00299	0.00 - 0.00299	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00299	0.00 - 0.00299	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00299	0.00 - 0.00299	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00299	0.00 - 0.00299	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041176R1

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No.	B4763 S6 A1	Volume (L)	1209.7
Client Sample No.	SVM-H2-1FD-100204	No. of Grid Openings	22
Description	Duplicate Analysis of 041172-51 - Different Grid Opening	Filter Area (mm2)	385
Analysis Date	12/2/2004	Area Analyzed (mm2)	0.319
Analyst	KM	Analytical Sens. (struc/cc)	0.000998
		Dectection Limit. (struc/cc)	0.00299

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	6.3	0.00200	0.00 - 0.00629	2

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No.	B4760 S52 A1	Volume (L)	1192.92
Client Sample No.	SVM-H2-2FD-100204	No. of Grid Openings	23
Description		Filter Area (mm2)	385
Analysis Date	11/8/2004	Area Analyzed (mm2)	0.333
Analyst	DW	Analytical Sens. (struc/cc)	0.000968
		Dectection Limit. (struc/cc)	0.00290

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	9.0	0.00291	0.00 - 0.00751	3
Total Asbestos Structures	9.0	0.00291	0.00 - 0.00751	3
Asbestos Structures > 5um	6.0	0.00194	0.00 - 0.00610	2
Asbestos Fibers and Bundles > 5um	6.0	0.00194	0.00 - 0.00610	2
PCM Equivalent Fibers-US	6.0	0.00194	0.00 - 0.00610	2
PCM Equivalent Structures-US	6.0	0.00194	0.00 - 0.00610	2
PROTOCOL ASB STRUCS 5-10	0.0	<0.00290	0.00 - 0.00290	0
PROTOCOL ASB STRUCS >10	0.0	<0.00290	0.00 - 0.00290	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00290	0.00 - 0.00290	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00290	0.00 - 0.00290	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00290	0.00 - 0.00290	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00290	0.00 - 0.00290	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00290	0.00 - 0.00290	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00290	0.00 - 0.00290	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00290	0.00 - 0.00290	0
AHERA-like Total Structures 3:1	9.0	0.00291	0.00 - 0.00751	3
AHERA-like Asb Strucs >5 and 3:1	6.0	0.00194	0.00 - 0.00610	2
AHERA-like Asb Strucs 5 - 10 and 3:1	6.0	0.00194	0.00 - 0.00610	2
AHERA-like Asb Strucs >10 and 3:1	0.0	<0.00290	0.00 - 0.00290	0
Total Other Amphibole Strucs 3:1	0.0	<0.00290	0.00 - 0.00290	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00290	0.00 - 0.00290	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00290	0.00 - 0.00290	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00290	0.00 - 0.00290	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041176R1

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No.	B4763 S7 A1	Volume (L)	1192.92
Client Sample No.	SVM-H2-2FD-100204	No. of Grid Openings	24
Description	Replicate Analysis of 041172-52 - Reprep	Filter Area (mm2)	385
Analysis Date	12/4/2004	Area Analyzed (mm2)	0.348
Analyst	DW	Analytical Sens. (struc/cc)	0.000928
		Dectection Limit. (struc/cc)	0.00277

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	5.8	0.00186	0.00 - 0.00585	2

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041176R1

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No. B4763 S8 A1
Client Sample No. SVM-H2-2FD-100204
Description Duplicate Analysis of 041172-52 - Same Grid Openings
Analysis Date 11/18/2004
Analyst KM

Volume (L) 1192.92
No. of Grid Openings 22
Filter Area (mm2) 385
Area Analyzed (mm2) 0.319
Analytical Sens. (struc/cc) 0.00101
Dectection Limit. (struc/cc) 0.00303

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	9.4	0.00304	0.00 - 0.00785	3

Lab/Cor Sample No. B4763 S9 A1
Client Sample No. SVM-H2-2FD-100204
Description Duplicate Analysis of 041172-52 - Different Grid Opening
Analysis Date 12/6/2004
Analyst KM

Volume (L) 1192.92
No. of Grid Openings 22
Filter Area (mm2) 385
Area Analyzed (mm2) 0.319
Analytical Sens. (struc/cc) 0.00101
Dectection Limit. (struc/cc) 0.00303

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	9.4	0.00304	0.00 - 0.00785	3

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No.	B4760 S53 A1	Volume (L)	1215.12
Client Sample No.	SVM-H2-3FD-100204	No. of Grid Openings	22
Description		Filter Area (mm2)	385
Analysis Date	11/8/2004	Area Analyzed (mm2)	0.319
Analyst	DW	Analytical Sens. (struc/cc)	0.000994
		Dectection Limit. (struc/cc)	0.00297

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	3.1	0.000994	0.00 - 0.00471	1
Total Asbestos Structures	3.1	0.000994	0.00 - 0.00471	1
Asbestos Structures > 5um	3.1	0.000994	0.00 - 0.00471	1
Asbestos Fibers and Bundles > 5um	3.1	0.000994	0.00 - 0.00471	1
PCM Equivalent Fibers-US	3.1	0.000994	0.00 - 0.00471	1
PCM Equivalent Structures-US	3.1	0.000994	0.00 - 0.00471	1
PROTOCOL ASB STRUCS 5-10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL ASB STRUCS >10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00297	0.00 - 0.00297	0
AHERA-like Total Structures 3:1	3.1	0.000994	0.00 - 0.00471	1
AHERA-like Asb Strucs >5 and 3:1	3.1	0.000994	0.00 - 0.00471	1
AHERA-like Asb Strucs 5 - 10 and 3:1	0.0	<0.00297	0.00 - 0.00297	0
AHERA-like Asb Strucs >10 and 3:1	3.1	0.000994	0.00 - 0.00471	1
Total Other Amphibole Strucs 3:1	0.0	<0.00297	0.00 - 0.00297	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00297	0.00 - 0.00297	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00297	0.00 - 0.00297	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00297	0.00 - 0.00297	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No. B4760 S86 A1
Client Sample No. SVM-H2-4FD-100204
Description
Analysis Date 12/18/2004
Analyst KM

Volume (L) 1194.88
No. of Grid Openings 23
Filter Area (mm2) 385
Area Analyzed (mm2) 0.333
Analytical Sens. (struc/cc) 0.000967
Dectection Limit. (struc/cc) 0.00289

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	0.0	<0.00289	0.00 - 0.00289	0
Total Asbestos Structures	0.0	<0.00289	0.00 - 0.00289	0
Asbestos Structures > 5um	0.0	<0.00289	0.00 - 0.00289	0
Asbestos Fibers and Bundles > 5um	0.0	<0.00289	0.00 - 0.00289	0
PCM Equivalent Fibers-US	0.0	<0.00289	0.00 - 0.00289	0
PCM Equivalent Structures-US	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL ASB STRUCS 5-10	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL ASB STRUCS >10	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00289	0.00 - 0.00289	0
AHERA-like Total Structures 3:1	0.0	<0.00289	0.00 - 0.00289	0
AHERA-like Asb Strucs >5 and 3:1	0.0	<0.00289	0.00 - 0.00289	0
AHERA-like Asb Strucs 5 - 10 and 3:1	0.0	<0.00289	0.00 - 0.00289	0
AHERA-like Asb Strucs >10 and 3:1	0.0	<0.00289	0.00 - 0.00289	0
Total Other Amphibole Strucs 3:1	0.0	<0.00289	0.00 - 0.00289	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00289	0.00 - 0.00289	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00289	0.00 - 0.00289	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00289	0.00 - 0.00289	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No. B4760 S87 A1
Client Sample No. SVM-H2-5FD-100204
Description
Analysis Date 12/18/2004
Analyst KM

Volume (L) 1194.26
No. of Grid Openings 23
Filter Area (mm2) 385
Area Analyzed (mm2) 0.333
Analytical Sens. (struc/cc) 0.000967
Dectection Limit. (struc/cc) 0.00289

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	18.0	0.00580	0.00213 - 0.0126	6
Total Asbestos Structures	18.0	0.00580	0.00213 - 0.0126	6
Asbestos Structures > 5um	12.0	0.00387	0.00105 - 0.00991	4
Asbestos Fibers and Bundles > 5um	12.0	0.00387	0.00105 - 0.00991	4
PCM Equivalent Fibers-US	12.0	0.00387	0.00105 - 0.00991	4
PCM Equivalent Structures-US	12.0	0.00387	0.00105 - 0.00991	4
PROTOCOL ASB STRUCS 5-10	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL ASB STRUCS >10	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00289	0.00 - 0.00289	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00289	0.00 - 0.00289	0
AHERA-like Total Structures 3:1	18.0	0.00580	0.00213 - 0.0126	6
AHERA-like Asb Strucs >5 and 3:1	12.0	0.00387	0.00105 - 0.00991	4
AHERA-like Asb Strucs 5 - 10 and 3:1	3.0	0.000967	0.00 - 0.00459	1
AHERA-like Asb Strucs >10 and 3:1	9.0	0.00290	0.00 - 0.00750	3
Total Other Amphibole Strucs 3:1	0.0	<0.00289	0.00 - 0.00289	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00289	0.00 - 0.00289	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00289	0.00 - 0.00289	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00289	0.00 - 0.00289	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No. B4760 S89 A1
Client Sample No. SVM-L2-1AD-100204
Description
Analysis Date 12/19/2004
Analyst KM

Volume (L) 305.61
No. of Grid Openings 87
Filter Area (mm2) 385
Area Analyzed (mm2) 1.26
Analytical Sens. (struc/cc) 0.000999
Dectection Limit. (struc/cc) 0.00299

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	4.0	0.00500	0.00162 - 0.0117	5
Total Asbestos Structures	4.0	0.00500	0.00162 - 0.0117	5
Asbestos Structures > 5um	0.8	0.000999	0.00 - 0.00474	1
Asbestos Fibers and Bundles > 5um	0.8	0.000999	0.00 - 0.00474	1
PCM Equivalent Fibers-US	0.0	<0.00299	0.00 - 0.00299	0
PCM Equivalent Structures-US	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL ASB STRUCS 5-10	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL ASB STRUCS >10	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00299	0.00 - 0.00299	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00299	0.00 - 0.00299	0
AHERA-like Total Structures 3:1	4.0	0.00500	0.00162 - 0.0117	5
AHERA-like Asb Strucs >5 and 3:1	0.8	0.000999	0.00 - 0.00474	1
AHERA-like Asb Strucs 5 - 10 and 3:1	0.0	<0.00299	0.00 - 0.00299	0
AHERA-like Asb Strucs >10 and 3:1	0.8	0.000999	0.00 - 0.00474	1
Total Other Amphibole Strucs 3:1	0.0	<0.00299	0.00 - 0.00299	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00299	0.00 - 0.00299	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00299	0.00 - 0.00299	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00299	0.00 - 0.00299	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No. B4760 S88 A1
Client Sample No. SVM-L2-15AD-100204
Description
Analysis Date 12/18/2004
Analyst TM

Volume (L) 304.76
No. of Grid Openings 87
Filter Area (mm2) 385
Area Analyzed (mm2) 1.26
Analytical Sens. (struc/cc) 0.00100
Dectection Limit. (struc/cc) 0.00300

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	0.8	0.00100	0.00 - 0.00475	1
Total Asbestos Structures	0.8	0.00100	0.00 - 0.00475	1
Asbestos Structures > 5um	0.8	0.00100	0.00 - 0.00475	1
Asbestos Fibers and Bundles > 5um	0.0	<0.00300	0.00 - 0.00300	0
PCM Equivalent Fibers-US	0.0	<0.00300	0.00 - 0.00300	0
PCM Equivalent Structures-US	0.0	<0.00300	0.00 - 0.00300	0
PROTOCOL ASB STRUCS 5-10	0.0	<0.00300	0.00 - 0.00300	0
PROTOCOL ASB STRUCS >10	0.0	<0.00300	0.00 - 0.00300	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00300	0.00 - 0.00300	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00300	0.00 - 0.00300	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00300	0.00 - 0.00300	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00300	0.00 - 0.00300	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00300	0.00 - 0.00300	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00300	0.00 - 0.00300	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00300	0.00 - 0.00300	0
AHERA-like Total Structures 3:1	0.8	0.00100	0.00 - 0.00475	1
AHERA-like Asb Strucs >5 and 3:1	0.8	0.00100	0.00 - 0.00475	1
AHERA-like Asb Strucs 5 - 10 and 3:1	0.8	0.00100	0.00 - 0.00475	1
AHERA-like Asb Strucs >10 and 3:1	0.0	<0.00300	0.00 - 0.00300	0
Total Other Amphibole Strucs 3:1	0.0	<0.00300	0.00 - 0.00300	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00300	0.00 - 0.00300	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00300	0.00 - 0.00300	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00300	0.00 - 0.00300	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No.	B4760 S90 A1	Volume (L)	306.83
Client Sample No.	SVM-L2-2AD-100204	No. of Grid Openings	87
Description		Filter Area (mm2)	385
Analysis Date	12/23/2004	Area Analyzed (mm2)	1.26
Analyst	JH	Analytical Sens. (struc/cc)	0.000995
		Dectection Limit. (struc/cc)	0.00298

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	4.0	0.00498	0.00162 - 0.0116	5
Total Asbestos Structures	4.0	0.00498	0.00162 - 0.0116	5
Asbestos Structures > 5um	3.2	0.00398	0.00108 - 0.0102	4
Asbestos Fibers and Bundles > 5um	3.2	0.00398	0.00108 - 0.0102	4
PCM Equivalent Fibers-US	2.4	0.00299	0.00 - 0.00771	3
PCM Equivalent Structures-US	2.4	0.00299	0.00 - 0.00771	3
PROTOCOL ASB STRUCS 5-10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL ASB STRUCS >10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00298	0.00 - 0.00298	0
AHERA-like Total Structures 3:1	4.0	0.00498	0.00162 - 0.0116	5
AHERA-like Asb Strucs >5 and 3:1	3.2	0.00398	0.00108 - 0.0102	4
AHERA-like Asb Strucs 5 - 10 and 3:1	0.8	0.000995	0.00 - 0.00472	1
AHERA-like Asb Strucs >10 and 3:1	2.4	0.00299	0.00 - 0.00771	3
Total Other Amphibole Strucs 3:1	0.0	<0.00298	0.00 - 0.00298	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00298	0.00 - 0.00298	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00298	0.00 - 0.00298	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00298	0.00 - 0.00298	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No. B4760 S91 A1
Client Sample No. SVM-L2-3AD-100204
Description
Analysis Date 12/23/2004
Analyst JH

Volume (L) 305.98
No. of Grid Openings 87
Filter Area (mm2) 385
Area Analyzed (mm2) 1.26
Analytical Sens. (struc/cc) 0.000998
Dectection Limit. (struc/cc) 0.00298

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	5.6	0.00699	0.00281 - 0.0144	7
Total Asbestos Structures	5.6	0.00699	0.00281 - 0.0144	7
Asbestos Structures > 5um	4.8	0.00599	0.00220 - 0.0130	6
Asbestos Fibers and Bundles > 5um	4.8	0.00599	0.00220 - 0.0130	6
PCM Equivalent Fibers-US	4.0	0.00499	0.00162 - 0.0116	5
PCM Equivalent Structures-US	4.0	0.00499	0.00162 - 0.0116	5
PROTOCOL ASB STRUCS 5-10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL ASB STRUCS >10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00298	0.00 - 0.00298	0
AHERA-like Total Structures 3:1	5.6	0.00699	0.00281 - 0.0144	7
AHERA-like Asb Strucs >5 and 3:1	4.8	0.00599	0.00220 - 0.0130	6
AHERA-like Asb Strucs 5 - 10 and 3:1	3.2	0.00399	0.00109 - 0.0102	4
AHERA-like Asb Strucs >10 and 3:1	1.6	0.00200	0.00 - 0.00629	2
Total Other Amphibole Strucs 3:1	0.0	<0.00298	0.00 - 0.00298	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00298	0.00 - 0.00298	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00298	0.00 - 0.00298	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00298	0.00 - 0.00298	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No. B4760 S92 A1
Client Sample No. SVM-L2-4AD-100204
Description
Analysis Date 12/27/2004
Analyst JH

Volume (L) 307.93
No. of Grid Openings 87
Filter Area (mm2) 385
Area Analyzed (mm2) 1.26
Analytical Sens. (struc/cc) 0.000992
Dectection Limit. (struc/cc) 0.00297

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	2.4	0.00298	0.00 - 0.00769	3
Total Asbestos Structures	2.4	0.00298	0.00 - 0.00769	3
Asbestos Structures > 5um	0.8	0.000992	0.00 - 0.00470	1
Asbestos Fibers and Bundles > 5um	0.8	0.000992	0.00 - 0.00470	1
PCM Equivalent Fibers-US	0.8	0.000992	0.00 - 0.00470	1
PCM Equivalent Structures-US	0.8	0.000992	0.00 - 0.00470	1
PROTOCOL ASB STRUCS 5-10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL ASB STRUCS >10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00297	0.00 - 0.00297	0
AHERA-like Total Structures 3:1	2.4	0.00298	0.00 - 0.00769	3
AHERA-like Asb Strucs >5 and 3:1	0.8	0.000992	0.00 - 0.00470	1
AHERA-like Asb Strucs 5 - 10 and 3:1	0.8	0.000992	0.00 - 0.00470	1
AHERA-like Asb Strucs >10 and 3:1	0.0	<0.00297	0.00 - 0.00297	0
Total Other Amphibole Strucs 3:1	0.0	<0.00297	0.00 - 0.00297	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00297	0.00 - 0.00297	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00297	0.00 - 0.00297	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00297	0.00 - 0.00297	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No. B4760 S93 A1
Client Sample No. SVM-L2-5AD-100204
Description
Analysis Date 12/26/2004
Analyst TM

Volume (L) 306.46
No. of Grid Openings 87
Filter Area (mm2) 385
Area Analyzed (mm2) 1.26
Analytical Sens. (struc/cc) 0.000997
Dectection Limit. (struc/cc) 0.00298

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	4.0	0.00498	0.00162 - 0.0116	5
Total Asbestos Structures	4.0	0.00498	0.00162 - 0.0116	5
Asbestos Structures > 5um	3.2	0.00399	0.00109 - 0.0102	4
Asbestos Fibers and Bundles > 5um	2.4	0.00299	0.00 - 0.00772	3
PCM Equivalent Fibers-US	2.4	0.00299	0.00 - 0.00772	3
PCM Equivalent Structures-US	1.6	0.00199	0.00 - 0.00628	2
PROTOCOL ASB STRUCS 5-10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL ASB STRUCS >10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00298	0.00 - 0.00298	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00298	0.00 - 0.00298	0
AHERA-like Total Structures 3:1	4.0	0.00498	0.00162 - 0.0116	5
AHERA-like Asb Strucs >5 and 3:1	3.2	0.00399	0.00109 - 0.0102	4
AHERA-like Asb Strucs 5 - 10 and 3:1	3.2	0.00399	0.00109 - 0.0102	4
AHERA-like Asb Strucs >10 and 3:1	0.0	<0.00298	0.00 - 0.00298	0
Total Other Amphibole Strucs 3:1	0.0	<0.00298	0.00 - 0.00298	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00298	0.00 - 0.00298	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00298	0.00 - 0.00298	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00298	0.00 - 0.00298	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No. B4760 S94 A1
Client Sample No. SVM-L2-6AD-100204
Description
Analysis Date 12/28/2004
Analyst JH

Volume (L) 311.83
No. of Grid Openings 86
Filter Area (mm2) 385
Area Analyzed (mm2) 1.25
Analytical Sens. (struc/cc) 0.000991
Dectection Limit. (struc/cc) 0.00296

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	2.4	0.00297	0.00 - 0.00768	3
Total Asbestos Structures	2.4	0.00297	0.00 - 0.00768	3
Asbestos Structures > 5um	1.6	0.00198	0.00 - 0.00624	2
Asbestos Fibers and Bundles > 5um	1.6	0.00198	0.00 - 0.00624	2
PCM Equivalent Fibers-US	1.6	0.00198	0.00 - 0.00624	2
PCM Equivalent Structures-US	0.8	0.000991	0.00 - 0.00470	1
PROTOCOL ASB STRUCS 5-10	1.6	0.00198	0.00 - 0.00624	2
PROTOCOL ASB STRUCS >10	0.0	<0.00296	0.00 - 0.00296	0
PROTOCOL ASB STRUCS TOTAL	1.6	0.00198	0.00 - 0.00624	2
PROTOCOL CHRYS STRUCS 5-10	0.8	0.000991	0.00 - 0.00470	1
PROTOCOL CHRYS STRUCS >10	0.0	<0.00296	0.00 - 0.00296	0
PROTOCOL CHRYS STRUCS TOTAL	0.8	0.000991	0.00 - 0.00470	1
PROTOCOL AMPH STRUCS 5-10	0.8	0.000991	0.00 - 0.00470	1
PROTOCOL AMPH STRUCS >10	0.0	<0.00296	0.00 - 0.00296	0
PROTOCOL AMPH STRUCS TOTAL	0.8	0.000991	0.00 - 0.00470	1
AHERA-like Total Structures 3:1	2.4	0.00297	0.00 - 0.00768	3
AHERA-like Asb Strucs >5 and 3:1	1.6	0.00198	0.00 - 0.00624	2
AHERA-like Asb Strucs 5 - 10 and 3:1	1.6	0.00198	0.00 - 0.00624	2
AHERA-like Asb Strucs >10 and 3:1	0.0	<0.00296	0.00 - 0.00296	0
Total Other Amphibole Strucs 3:1	0.0	<0.00296	0.00 - 0.00296	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00296	0.00 - 0.00296	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00296	0.00 - 0.00296	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00296	0.00 - 0.00296	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No.	B4760 S8 A1	**Volume (L)**	569.77
Client Sample No.	CC1-L6-1CA-100204	**No. of Grid Openings**	47
Description		**Filter Area (mm2)**	385
Analysis Date	10/8/2004	**Area Analyzed (mm2)**	0.681
Analyst	TM	**Analytical Sens. (struc/cc)**	0.000992
		Dectection Limit. (struc/cc)	0.00297

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	8.8	0.00595	0.00218 - 0.0130	6
Total Asbestos Structures	8.8	0.00595	0.00218 - 0.0130	6
Asbestos Structures > 5um	1.5	0.000992	0.00 - 0.00470	1
Asbestos Fibers and Bundles > 5um	1.5	0.000992	0.00 - 0.00470	1
PCM Equivalent Fibers-US	1.5	0.000992	0.00 - 0.00470	1
PCM Equivalent Structures-US	1.5	0.000992	0.00 - 0.00470	1
PROTOCOL ASB STRUCS 5-10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL ASB STRUCS >10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00297	0.00 - 0.00297	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00297	0.00 - 0.00297	0
AHERA-like Total Structures 3:1	8.8	0.00595	0.00218 - 0.0130	6
AHERA-like Asb Strucs >5 and 3:1	1.5	0.000992	0.00 - 0.00470	1
AHERA-like Asb Strucs 5 - 10 and 3:1	1.5	0.000992	0.00 - 0.00470	1
AHERA-like Asb Strucs >10 and 3:1	0.0	<0.00297	0.00 - 0.00297	0
Total Other Amphibole Strucs 3:1	0.0	<0.00297	0.00 - 0.00297	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00297	0.00 - 0.00297	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00297	0.00 - 0.00297	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00297	0.00 - 0.00297	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No. B4760 S9 A1
Client Sample No. CC1-L6-1CB-100204
Description
Analysis Date 10/8/2004
Analyst JH

Volume (L) 607.18
No. of Grid Openings 45
Filter Area (mm2) 385
Area Analyzed (mm2) 0.652
Analytical Sens. (struc/cc) 0.000972
Dectection Limit. (struc/cc) 0.00291

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	7.7	0.00486	0.00158 - 0.0113	5
Total Asbestos Structures	6.1	0.00389	0.00106 - 0.00996	4
Asbestos Structures > 5um	1.5	0.000972	0.00 - 0.00461	1
Asbestos Fibers and Bundles > 5um	1.5	0.000972	0.00 - 0.00461	1
PCM Equivalent Fibers-US	1.5	0.000972	0.00 - 0.00461	1
PCM Equivalent Structures-US	1.5	0.000972	0.00 - 0.00461	1
PROTOCOL ASB STRUCS 5-10	0.0	<0.00291	0.00 - 0.00291	0
PROTOCOL ASB STRUCS >10	0.0	<0.00291	0.00 - 0.00291	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00291	0.00 - 0.00291	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00291	0.00 - 0.00291	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00291	0.00 - 0.00291	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00291	0.00 - 0.00291	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00291	0.00 - 0.00291	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00291	0.00 - 0.00291	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00291	0.00 - 0.00291	0
AHERA-like Total Structures 3:1	6.1	0.00389	0.00106 - 0.00996	4
AHERA-like Asb Strucs >5 and 3:1	1.5	0.000972	0.00 - 0.00461	1
AHERA-like Asb Strucs 5 - 10 and 3:1	0.0	<0.00291	0.00 - 0.00291	0
AHERA-like Asb Strucs >10 and 3:1	1.5	0.000972	0.00 - 0.00461	1
Total Other Amphibole Strucs 3:1	1.5	0.000972	0.00 - 0.00461	1
Other Amphibole Strucs >5 and 3:1	0.0	<0.00291	0.00 - 0.00291	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00291	0.00 - 0.00291	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00291	0.00 - 0.00291	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

ANALYSIS DETAIL

Lab/Cor Sample No.	B4760 S10 A1	**Volume (L)**	597.98
Client Sample No.	CC1-L6-2CB-100204	**No. of Grid Openings**	44
Description		**Filter Area (mm2)**	385
Analysis Date	10/8/2004	**Area Analyzed (mm2)**	0.638
Analyst	DW	**Analytical Sens. (struc/cc)**	0.00101
		Dectection Limit. (struc/cc)	0.00302

Structure Type	Filter Density (s/mm2)	Concen-tration (struc/cc)	95% Confidence Interval (struc/cc)	Struc. Count
Primary Asbestos Structures	12.5	0.00808	0.00349 - 0.0159	8
Total Asbestos Structures	12.5	0.00808	0.00349 - 0.0159	8
Asbestos Structures > 5um	4.7	0.00303	0.00 - 0.00783	3
Asbestos Fibers and Bundles > 5um	4.7	0.00303	0.00 - 0.00783	3
PCM Equivalent Fibers-US	3.1	0.00202	0.00 - 0.00636	2
PCM Equivalent Structures-US	3.1	0.00202	0.00 - 0.00636	2
PROTOCOL ASB STRUCS 5-10	0.0	<0.00302	0.00 - 0.00302	0
PROTOCOL ASB STRUCS >10	0.0	<0.00302	0.00 - 0.00302	0
PROTOCOL ASB STRUCS TOTAL	0.0	<0.00302	0.00 - 0.00302	0
PROTOCOL CHRYS STRUCS 5-10	0.0	<0.00302	0.00 - 0.00302	0
PROTOCOL CHRYS STRUCS >10	0.0	<0.00302	0.00 - 0.00302	0
PROTOCOL CHRYS STRUCS TOTAL	0.0	<0.00302	0.00 - 0.00302	0
PROTOCOL AMPH STRUCS 5-10	0.0	<0.00302	0.00 - 0.00302	0
PROTOCOL AMPH STRUCS >10	0.0	<0.00302	0.00 - 0.00302	0
PROTOCOL AMPH STRUCS TOTAL	0.0	<0.00302	0.00 - 0.00302	0
AHERA-like Total Structures 3:1	12.5	0.00808	0.00349 - 0.0159	8
AHERA-like Asb Strucs >5 and 3:1	4.7	0.00303	0.00 - 0.00783	3
AHERA-like Asb Strucs 5 - 10 and 3:1	3.1	0.00202	0.00 - 0.00636	2
AHERA-like Asb Strucs >10 and 3:1	1.6	0.00101	0.00 - 0.00479	1
Total Other Amphibole Strucs 3:1	0.0	<0.00302	0.00 - 0.00302	0
Other Amphibole Strucs >5 and 3:1	0.0	<0.00302	0.00 - 0.00302	0
Other Amphibole Strucs 5 - 10 and 3:1	0.0	<0.00302	0.00 - 0.00302	0
Other Amphibole Strucs >10 and 3:1	0.0	<0.00302	0.00 - 0.00302	0

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-H2-1FD-100204 *Lab/Cor Sample No.:* B4760 S51 A1

Description:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	B41		NSD									
A	2	B31		NSD									
A	3	B21		NSD									
A	4	B11		NSD									
A	5	B1		NSD									
A	6	C11		NSD									
A	7	C31	CMQ	1	1	B	16	0.5	32	999	734	Mg, Si Chrysotile No Diffraction	AS>5, AFB>5, PCMEF-US, PCMES-US, PSAS >10, PSAS TOT, PCAS >10, PCAS TOT, TAS_AHRA, AS>5_AHRA, AS>10_AHRA
A	8	C42		NSD									
A	9	C22		NSD									
A	10	C2	CMQ	2	2	B	2.5	0.4	6.2			Mg, Si Chrysotile	TAS_AHRA
A	11	B12		NSD									
B	12	A41		NSD									
B	13	A21		NSD									
B	14	A11		NSD									
B	15	D1		NSD									
B	16	D11		NSD									
B	17	D31		NSD									
B	18	D42		NSD									
B	19	D22		NSD									
B	20	D2		NSD									
B	21	A12		NSD									
B	22	A32		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041176R1

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-H2-1FD-100204 *Lab/Cor Sample No.:* B4763 S6 A1

Descripiton: Duplicate Analysis of 041172-51 - Different Grid Openings

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	A4		NSD									
A	2	A14		NSD									
A	3	A3		NSD									
A	4	A13		NSD									
A	5	A23		NSD									
A	6	D2		NSD									
A	7	A2	CD	1		MD1-0	4.5	3.8	1.2			Chrysotile	
A	7	A2	CD		1	MF	1.2	0.15	8.0			Mg, Si Chrysotile	
A	8	A12		NSD									
A	9	A22		NSD									
A	10	D1		NSD									
A	11	A11		NSD									
B	12	D14		NSD									
B	13	D4		NSD									
B	14	A14		NSD									
B	15	A24		NSD									
B	16	D13		NSD									
B	17	D3		NSD									
B	18	A13		NSD									
B	19	A23		NSD									
B	20	A1	CD	2	2	B	8	0.5	16			Mg, Si Chrysotile	
B	21	A11		NSD									
B	22	A21		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-H2-2FD-100204 *Lab/Cor Sample No.:* B4760 S52 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	B34		NSD									
A	2	B24		NSD									
A	3	B14		NSD									
A	4	C4		NSD									
A	5	C14		NSD									
A	6	C34		NSD									
A	7	C43		NSD									
A	8	C33		NSD									
A	9	C13		NSD									
A	10	C3		NSD									
A	11	B13		NSD									
A	12	B23		NSD									
A	13	A41		NSD									
A	14	A21		NSD									
B	15	B44		NSD									
B	16	B34		NSD									
B	17	B24		NSD									
B	18	C4		NSD									
B	19	C14		NSD									
B	20	C24		NSD									
B	21	C34	CDQ	1	1	B	10	2	5.0	1010	15331	Mg, Si Chrysotile Verified - KM	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
B	22	C44		NSD									
B	23	C23	AZQ	2	2	F	10	2	5.0	5435	15330	Mg, Al, Si, Ca, Fe Actinolite Zone Axis [5 3 2] - KM	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
B	23	C23	AQ	3	3	F	5	0.85	5.9			Mg, Al, Si, Ca, Fe Actinolite	TAS_AHRA

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041176R1

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-H2-2FD-100204 *Lab/Cor Sample No.:* B4763 S7 A1

Descripiton: Replicate Analysis of 041172-52 - Reprep

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	B44		NSD									
A	2	B24		NSD									
A	3	B4		NSD									
A	4	C14		NSD									
A	5	C34		NSD									
A	6	C43		NSD									
A	7	C32		NSD									
A	8	C12		NSD									
A	9	B2		NSD									
A	10	B22		NSD									
A	11	B42	AQ	1		MD2-1	17	15	1.1			Actinolite	
A	11	B42	AQ		1	MF	6	1.2	5.0			Mg, Al, Si, Ca, Fe Actinolite	
A	11	B42	AQ		2	MF	2.75	0.65	4.2			Mg, Al, Si, Ca, Fe Actinolite	
A	12	B31		NSD									
A	13	B11		NSD									
A	14	A10		NSD									
B	15	A41		NSD									
B	16	A21		NSD									
B	17	A1		NSD									
B	18	D11		NSD									
B	19	D31		NSD									
B	20	D40		NSD									
B	21	D20		NSD									
B	22	C10	CDQ	2	3	B	5	1.2	4.2			Mg, Si Chrysotile	
B	23	C30		NSD									
B	24	C41		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041176R1

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-H2-2FD-100204 *Lab/Cor Sample No.:* B4763 S8 A1

Descripiton: Duplicate Analysis of 041172-52 - Same Grid Openings

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	A41		NSD									
A	2	A21		NSD									
A	3	B34		NSD									
A	4	B24		NSD									
A	5	B14		NSD									
A	6	B23		NSD									
A	7	B13		NSD									
A	8	C4		NSD									
A	9	C14		NSD									
A	10	C34		NSD									
A	11	C43		NSD									
A	12	C3		NSD									
A	13	C33		NSD									
A	14	C13		NSD									
B	15	B44		NSD									
B	16	B34		NSD									
B	17	B24		NSD									
B	18	C4		NSD									
B	19	C14		NSD									
B	20	C24		NSD									
B	21	C34	CDQ	1	1	B	16	1	16			Mg, Si Chrysotile	
B	22	C23	AQ	2	2	F	10	1.8	5.6			Mg, Al, Si, Ca, Fe Actinolite	
B	22	C23	AQ	3	3	F	4.3	0.9	4.8			Mg, Al, Si, Ca, Fe Actinolite	

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041176R1

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-H2-2FD-100204 *Lab/Cor Sample No.:* B4763 S9 A1

Descripiton: Duplicate Analysis of 041172-52 - Different Grid Openings

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
C	1	A31		NSD									
C	2	A11		NSD									
C	3	D1		NSD									
C	4	D21	CDQ	1	1	F	1.2	0.1	12			Mg, Si Chrysotile	
C	5	D41	ADQ	2	2	F	2.5	0.2	12			Mg, Si, Ca, Fe Actinolite	
C	6	B32		NSD									
C	7	B12		NSD									
C	8	C2		NSD									
C	9	C22		NSD									
C	10	B42		NSD									
C	11	B10		NSD									
D	12	A31		NSD									
D	13	A11		NSD									
D	14	D1		NSD									
D	15	D21		NSD									
D	16	D41		NSD									
D	17	A32		NSD									
D	18	A12		NSD									
D	19	D2		NSD									
D	20	D22	ADQ	3	3	F	8.5	0.9	9.4			Mg, Si, Ca, Fe Actinolite	
D	21	D42		NSD									
D	22	B10		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-H2-3FD-100204 *Lab/Cor Sample No.:* B4760 S53 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	B34		NSD									
A	2	B24		NSD									
A	3	B14		NSD									
A	4	B4		NSD									
A	5	C4		NSD									
A	6	C14		NSD									
A	7	C34		NSD									
A	8	B12		NSD									
A	9	C32		NSD									
A	10	B41		NSD									
A	11	B31		NSD									
A	12	B1		NSD									
B	13	B41		NSD									
B	14	B31		NSD									
B	15	B21		NSD									
B	16	B11		NSD									
B	17	B1		NSD									
B	18	C11		NSD									
B	19	C31		NSD									
B	20	C42		NSD									
B	21	C33	AZQ	1	1	F	11	1.5	7.3	5437	15333	Mg, Al, Si, Ca, Fe Actinolite Zone Axis [5 -1 2] - KM	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS>10_AHRA
B	22	C23		NSD									

NSD = No Structures Detected
PAS = Primary Asbestos Structures
TAS = Total Asbestos Structures
AS>5 = Asbestos Structures > 5um
AFB>5 = Asbestos Fibers and Bundles > 5um
PCMEF-US = PCM Equivalent Fibers-US
PCMES-US = PCM Equivalent Structures-US
PCMEF-ISO = PCM Equivalent Fibers-ISO
PCMES-ISO = PCM Equivalent Structures-ISO

CF = Cleavage Fragments
TS = Transitional Structures

PSAS 5-10 = PROTOCOL ASB STRUCS 5-10
PSAS >10 = PROTOCOL ASB STRUCS >10
PSAS TOT = PROTOCOL ASB STRUCS TOTAL
PSCH 5-10 = PROTOCOL CHRYS STRUCS 5-10
PSCH >10 = PROTOCOL CHRYS STRUCS >10
PSCH TOT = PROTOCOL CHRYS STRUCS TOTAL
PSAM 5-10 = PROTOCOL AMPH STRUCS 5-10
PSAM >10 = PROTOCOL AMPH STRUCS >10
PSAM TOT = PROTOCOL AMPH STRUCS TOTAL

PChS = Primary Chrysotile Structures
PAmS = Primary Amphibole Structures

TAS_AHRA = AHERA-like Total Strucs 3:1
AS>5_AHRA = AHERA-like Asb Strucs >5 and 3:1
AS5-10_AHRA = AHERA-like Asb Strucs 5 - 10 and 3:1
AS>10_AHRA = AHERA-like Asb Strucs >10 and 3:1
TOS_AHRA = Total Other Amphibole Strucs 3:1
OS>5_AHRA = Other Amphibole Struc >5 and 3:1
OS5-10_AHRA = Other Amphibole Struc 5 - 10 and 3:1
OS>10_AHRA = Other Amphibole Strucs >10 and 3:1

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-H2-4FD-100204 *Lab/Cor Sample No.:* B4760 S86 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	B31		NSD									
A	2	B11		NSD									
A	3	C1		NSD									
A	4	C21		NSD									
A	5	C41		NSD									
A	6	A41		NSD									
A	7	A21		NSD									
A	8	A1		NSD									
A	9	D11		NSD									
A	10	D31		NSD									
A	11	C10		NSD									
A	12	C30		NSD									
B	13	B42		NSD									
B	14	B22		NSD									
B	15	B2		NSD									
B	16	C12		NSD									
B	17	B41		NSD									
B	18	B21		NSD									
B	19	B11		NSD									
B	20	C1		NSD									
B	21	C21		NSD									
B	22	A41		NSD									
B	23	A21		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-H2-5FD-100204 *Lab/Cor Sample No.:* B4760 S87 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	2	B12	AQ	NSD		F	13	2.5	5.2			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS>10_AHRA
A	1	B32		NSD									
A	2	B12	AZQ	1		MD1-0	4.8	1.5	3.2			Actinolite	TAS_AHRA
A	2	B12	AZQ		1	MF	4.8	0.5	9.6	1225	906	Mg, Al, Si, Ca, Fe Actinolite Zone Axis [5 1 2] - JH	
A	3	C2	AQ	2		MD1-1	9	1.5	6.0			Actinolite	AS>5, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
A	3	C2	AQ		2	MF	6.8	0.8	8.5			Mg, Al, Si, Ca, Fe Actinolite	AFB>5, PCMEF-US
A	4	C22		NSD									
A	5	C42		NSD									
A	6	B41		NSD									
A	7	B21		NSD									
A	8	B1	AQ	3	3	F	11	1	11			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS>10_AHRA
A	9	C11		NSD									
A	10	C31		NSD									
A	11	A20		NSD									
A	12	A10		NSD									
B	13	B2	AQ	4	4	F	5	1.3	3.8			Mg, Al, Si, Ca, Fe Actinolite	TAS_AHRA
B	14	B42		NSD									
B	15	B22		NSD									
B	16	C12	AQ	5	5	F	13.5	2.5	5.4			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS>10_AHRA
B	17	C32		NSD									
B	18	A41		NSD									
B	19	A21		NSD									
B	20	A1		NSD									
B	21	D11		NSD									
B	22	D31		NSD									
B	23	D20		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-1AD-100204 *Lab/Cor Sample No.:* B4760 S89 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	B44		NSD									
A	2	B24		NSD									
A	3	B4		NSD									
A	4	C14		NSD									
A	5	C34		NSD									
A	6	B33		NSD									
A	7	B13		NSD									
A	8	C3		NSD									
A	9	C23		NSD									
A	10	C43		NSD									
A	11	B42		NSD									
A	12	B22		NSD									
A	13	B2		NSD									
A	14	C12		NSD									
A	15	C32		NSD									
A	16	B31		NSD									
A	17	B11		NSD									
A	18	C1		NSD									
A	19	C21		NSD									
A	20	C31		NSD									
A	21	A41		NSD									
A	22	A21		NSD									
A	23	A1		NSD									
A	24	D11		NSD									
A	25	D31		NSD									
A	26	A32		NSD									
A	27	A12		NSD									
A	28	D2		NSD									
A	29	D22		NSD									
A	30	D42		NSD									
B	31	B44		NSD									
B	32	B24		NSD									
B	33	B4		NSD									
B	34	C14		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-1AD-100204 *Lab/Cor Sample No.:* B4760 S89 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
B	35	C34		NSD									
B	36	B33		NSD									
B	37	B13		NSD									
B	38	C3		NSD									
B	39	C23		NSD									
B	40	C43		NSD									
B	41	B42		NSD									
B	42	B22		NSD									
B	43	B2		NSD									
B	44	C12	AZQ	1	1	F	3.8	0.5	7.6	1196	882	Mg, Al, Si, Ca, Fe Actinolite Zone Axis - [3 1 0] KM	TAS_AHRA
B	45	C32		NSD									
B	46	B31		NSD									
B	47	B11		NSD									
B	48	C1		NSD									
B	49	C21		NSD									
B	50	C41		NSD									
B	51	A31		NSD									
B	52	A11		NSD									
B	53	D1	AQ	2	2	F	23	4.3	5.3			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, TAS_AHRA, AS>5_AHRA, AS>10_AHRA
B	54	D21		NSD									
B	55	D41		NSD									
B	56	A22		NSD									
B	57	A2		NSD									
B	58	D12		NSD									
B	59	D32		NSD									
C	60	C34	AQ	3	3	F	2	0.6	3.3			Mg, Al, Si, Ca, Fe Actinolite	TAS_AHRA
C	61	B44		NSD									
C	62	B24		NSD									
C	63	B4		NSD									
C	64	C14		NSD									
C	65	B33		NSD									
C	66	B13		NSD									
C	67	C3		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-1AD-100204 *Lab/Cor Sample No.:* B4760 S89 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
C	68	C23		NSD									
C	69	C43		NSD									
C	70	B42	AQ	4	4	F	5	0.6	8.3			Mg, Al, Si, Ca, Fe Actinolite	TAS_AHRA
C	71	B22		NSD									
C	72	B2		NSD									
C	73	C12		NSD									
C	74	C32		NSD									
C	75	B31		NSD									
C	76	B11		NSD									
C	77	C1		NSD									
C	78	C21		NSD									
C	79	C41		NSD									
C	80	A41	AQ	5	5	F	4	0.5	8.0			Mg, Al, Si, Ca, Fe Actinolite	TAS_AHRA
C	81	A21		NSD									
C	82	A1		NSD									
C	83	D11		NSD									
C	84	D31		NSD									
C	85	A22		NSD									
C	86	A2		NSD									
C	87	D2		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-15AD-100204 *Lab/Cor Sample No.:* B4760 S88 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	C44		NSD									
A	2	C42		NSD									
A	3	C40		NSD									
A	4	D41		NSD									
A	5	D43		NSD									
A	6	D34		NSD									
A	7	D32		NSD									
A	8	D30		NSD									
A	9	C31		NSD									
A	10	C33		NSD									
A	11	C24		NSD									
A	12	C22		NSD									
A	13	C20		NSD									
A	14	D21		NSD									
A	15	D23		NSD									
A	16	D14		NSD									
A	17	D12		NSD									
A	18	D10		NSD									
A	19	C11		NSD									
A	20	C13		NSD									
A	21	D1		NSD									
A	22	D3		NSD									
A	23	A4		NSD									
A	24	A2		NSD									
A	25	B1		NSD									
A	26	B3		NSD									
A	27	B14		NSD									
A	28	B12	AZQ	1		MD1-0	5.25	4.75	1.1			Actinolite	AS>5, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
A	28	B12	AZQ		1	MF	4.1	1.2	3.4	1226	15603	Mg, Al, Si, Ca, Fe Actinolite Zone Axis [0 1 2] - JH	
A	29	B10		NSD									
A	30	A13		NSD									
B	31	C42		NSD									
B	32	C40		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-15AD-100204 *Lab/Cor Sample No.:* B4760 S88 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
B	33	D41		NSD									
B	34	D43		NSD									
B	35	D34		NSD									
B	36	D32		NSD									
B	37	D30		NSD									
B	38	C31		NSD									
B	39	C33		NSD									
B	40	C22		NSD									
B	41	C20		NSD									
B	42	D21		NSD									
B	43	D23		NSD									
B	44	D14		NSD									
B	45	D12		NSD									
B	46	D10		NSD									
B	47	C11		NSD									
B	48	C13		NSD									
B	49	C4		NSD									
B	50	C2		NSD									
B	51	D1		NSD									
B	52	D3		NSD									
B	53	A4		NSD									
B	54	A2		NSD									
B	55	B1		NSD									
B	56	B14		NSD									
B	57	B12		NSD									
B	58	B10		NSD									
B	59	A11		NSD									
C	60	B42		NSD									
C	61	B22		NSD									
C	62	B2		NSD									
C	63	C12		NSD									
C	64	C32		NSD									
C	65	B31		NSD									
C	66	B11		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-15AD-100204 *Lab/Cor Sample No.:* B4760 S88 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
C	67	C1		NSD									
C	68	C21		NSD									
C	69	C41		NSD									
C	70	A41		NSD									
C	71	A21		NSD									
C	72	A1		NSD									
C	73	D11		NSD									
C	74	D31		NSD									
C	75	A32		NSD									
C	76	A12		NSD									
C	77	D2		NSD									
C	78	D22		NSD									
C	79	D42		NSD									
C	80	A43		NSD									
C	81	A23		NSD									
C	82	A3		NSD									
C	83	D13		NSD									
C	84	D33		NSD									
C	85	D20		NSD									
C	86	D30		NSD									
C	87	D40		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-2AD-100204 *Lab/Cor Sample No.:* B4760 S90 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	A33	AQ	1	1	F	23	5	4.6			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, TAS_AHRA, AS>5_AHRA, AS>10_AHRA
A	2	A13		NSD									
A	3	A3	AZQ	2	2	F	2.7	0.4	6.8	1227	895	Mg, Al, Si, Ca, Fe Actinolite Zone Axis [4 0 3] - JH	TAS_AHRA
A	4	D11		NSD									
A	5	A1		NSD									
A	6	A21		NSD									
A	7	A41		NSD									
A	8	B40		NSD									
A	9	B20		NSD									
A	10	B10		NSD									
A	11	C10		NSD									
A	12	C30		NSD									
A	13	C42		NSD									
A	14	C22		NSD									
A	15	C2		NSD									
A	16	B12		NSD									
A	17	B32		NSD									
A	18	B42		NSD									
A	19	B44		NSD									
A	20	B24		NSD									
A	21	B4		NSD									
A	22	C14		NSD									
A	23	C34		NSD									
A	24	C44		NSD									
A	25	C33		NSD									
A	26	C13		NSD									
A	27	B3		NSD									
A	28	B23		NSD									
A	29	B43		NSD									
B	30	A34		NSD									
B	31	A4		NSD									
B	32	D14	AQ	3	3	F	14	2	7.0			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS>10_AHRA

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-2AD-100204 *Lab/Cor Sample No.:* B4760 S90 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
B	33	D34		NSD									
B	34	D44		NSD									
B	35	D42		NSD									
B	36	D22		NSD									
B	37	D2		NSD									
B	38	A12		NSD									
B	39	A32		NSD									
B	40	A42		NSD									
B	41	A40		NSD									
B	42	A20		NSD									
B	43	A10		NSD									
B	44	D10		NSD									
B	45	D30		NSD									
B	46	D40		NSD									
B	47	C41		NSD									
B	48	C21		NSD									
B	49	C1		NSD									
B	50	B1		NSD									
B	51	B21		NSD									
B	52	B41		NSD									
B	53	B43		NSD									
B	54	B23		NSD									
B	55	B3		NSD									
B	56	C13	AQ	4	4	F	15	3	5.0			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS>10_AHRA
B	57	C33		NSD									
B	58	C4		NSD									
B	59	B14		NSD									
B	60	B34	AQ	5	5	F	8	1.6	5.0			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
C	61	A33		NSD									
C	62	A13		NSD									
C	63	D3		NSD									
C	64	D23		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-2AD-100204 *Lab/Cor Sample No.:* B4760 S90 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
C	65	D43		NSD									
C	66	D41		NSD									
C	67	D21		NSD									
C	68	D1		NSD									
C	69	A11		NSD									
C	70	A31		NSD									
C	71	A41		NSD									
C	72	B40		NSD									
C	73	B20		NSD									
C	74	B10		NSD									
C	75	C10		NSD									
C	76	C30		NSD									
C	77	C40		NSD									
C	78	C42		NSD									
C	79	C22		NSD									
C	80	C2		NSD									
C	81	B12		NSD									
C	82	B32		NSD									
C	83	B42		NSD									
C	84	B44		NSD									
C	85	B24		NSD									
C	86	B4		NSD									
C	87	C14		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-3AD-100204 *Lab/Cor Sample No.:* B4760 S91 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	A33		NSD									
A	2	A13	CDQ	1		MD1-0	4	3	1.3			Chrysotile	TAS_AHRA
A	2	A13	CDQ		1	MF	4	0.05	80	1212	896	Mg, Al, Si Chrysotile Verified - JH	
A	3	D3		NSD									
A	4	D23		NSD									
A	5	D43		NSD									
A	6	D41		NSD									
A	7	D21		NSD									
A	8	D1		NSD									
A	9	A11		NSD									
A	10	A31		NSD									
A	11	A41		NSD									
A	12	B40		NSD									
A	13	B20		NSD									
A	14	B10		NSD									
A	15	C10	AZQ	2	2	F	13	2.5	5.2	1213	897	Mg, Al, Si, Ca, Fe Actinolite Zone Axis - [7 1 0] KM	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS>10_AHRA
A	16	C30		NSD									
A	17	C40		NSD									
A	18	C42		NSD									
A	19	C22		NSD									
A	20	C2	AQ	3	3	B	7.5	1.1	6.8			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
A	21	B12		NSD									
A	22	B32	AQ	4	4	F	7.5	1.3	5.8			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
A	23	B42		NSD									
A	24	B44		NSD									
A	25	B34		NSD									
A	26	B14	AQ	5	5	F	9	2.5	3.6			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
A	27	C4		NSD									
A	28	C24		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-3AD-100204 *Lab/Cor Sample No.:* B4760 S91 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	29	C44		NSD									
B	30	A33		NSD									
B	31	A13		NSD									
B	32	D3		NSD									
B	33	D23		NSD									
B	34	D43		NSD									
B	35	D31		NSD									
B	36	D11		NSD									
B	37	A1		NSD									
B	38	A21		NSD									
B	39	A41		NSD									
B	40	B40		NSD									
B	41	B20		NSD									
B	42	B10		NSD									
B	43	C10		NSD									
B	44	C30		NSD									
B	45	C40		NSD									
B	46	C42		NSD									
B	47	C22		NSD									
B	48	C2		NSD									
B	49	B12		NSD									
B	50	B32		NSD									
B	51	B42		NSD									
B	52	B33		NSD									
B	53	B13		NSD									
B	54	C3		NSD									
B	55	C23		NSD									
B	56	C43		NSD									
B	57	C34		NSD									
B	58	C14		NSD									
C	59	A34		NSD									
C	60	A14		NSD									
C	61	D4		NSD									
C	62	D24		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-3AD-100204 *Lab/Cor Sample No.:* B4760 S91 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
C	63	D44		NSD									
C	64	D42		NSD									
C	65	D22		NSD									
C	66	D2		NSD									
C	67	A12		NSD									
C	68	A32	AQ	6	6	F	7	2.2	3.2			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
C	69	A42		NSD									
C	70	A40		NSD									
C	71	A20		NSD									
C	72	D10		NSD									
C	73	D30		NSD									
C	74	D40		NSD									
C	75	C41		NSD									
C	76	C21		NSD									
C	77	C1		NSD									
C	78	B11	AQ	7	7	F	40	4.5	8.9			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, TAS_AHRA, AS>5_AHRA, AS>10_AHRA
C	79	B31		NSD									
C	80	B41		NSD									
C	81	B43		NSD									
C	82	B23		NSD									
C	83	B3		NSD									
C	84	C13		NSD									
C	85	C33		NSD									
C	86	C43		NSD									
C	87	C34		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-4AD-100204 *Lab/Cor Sample No.:* B4760 S92 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	A44		NSD									
A	2	A24		NSD									
A	3	A4		NSD									
A	4	D14		NSD									
A	5	D34		NSD									
A	6	D42		NSD									
A	7	D22		NSD									
A	8	D2		NSD									
A	9	A12		NSD									
A	10	A42		NSD									
A	11	A40		NSD									
A	12	A20		NSD									
A	13	A10		NSD									
A	14	D10		NSD									
A	15	D30		NSD									
A	16	D40	CDQ	1	1	F	2.2	0.04	55	1216	900	Mg, Si Chrysotile Verified - JH TAS_AHRA	
A	17	C41		NSD									
A	18	C21		NSD									
A	19	C1		NSD									
A	20	B11		NSD									
A	21	B31		NSD									
A	22	B41		NSD									
A	23	B43		NSD									
A	24	B23		NSD									
A	25	B3		NSD									
A	26	C13		NSD									
A	27	C33		NSD									
A	28	C43		NSD									
A	29	C34		NSD									
B	30	A24		NSD									
B	31	A4		NSD									
B	32	D14		NSD									
B	33	D34		NSD									
B	34	D44		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-4AD-100204 *Lab/Cor Sample No.:* B4760 S92 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
B	35	D42		NSD									
B	36	D22		NSD									
B	37	D2		NSD									
B	38	A12		NSD									
B	39	A22		NSD									
B	40	A42		NSD									
B	41	A40		NSD									
B	42	A20		NSD									
B	43	A10		NSD									
B	44	D10		NSD									
B	45	D30		NSD									
B	46	D40		NSD									
B	47	C41		NSD									
B	48	C21		NSD									
B	49	C1		NSD									
B	50	B11		NSD									
B	51	B31		NSD									
B	52	B41		NSD									
B	53	B43		NSD									
B	54	B23		NSD									
B	55	B3	CDQ	2	2	F	0.7	0.08	8.8	1217	901	Mg, Si Chrysotile Verified - JH TAS_AHRA	
B	56	C13		NSD									
B	57	C33		NSD									
B	58	C43		NSD									
C	59	A34		NSD									
C	60	A14		NSD									
C	61	D4		NSD									
C	62	D24		NSD									
C	63	D44		NSD									
C	64	D42		NSD									
C	65	D2		NSD									
C	66	A12		NSD									
C	67	A32		NSD									
C	68	A41		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-4AD-100204 *Lab/Cor Sample No.:* B4760 S92 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
C	69	A40		NSD									
C	70	A20		NSD									
C	71	A10		NSD									
C	72	D10		NSD									
C	73	D30		NSD									
C	74	D40		NSD									
C	75	C41		NSD									
C	76	C21		NSD									
C	77	C1		NSD									
C	78	B11		NSD									
C	79	B21		NSD									
C	80	B41		NSD									
C	81	B43		NSD									
C	82	B23		NSD									
C	83	B3	AZQ	3	3	F	6	1.5	4.0	1218	902	Mg, Al, Si, Ca, Fe Actinolite Zone Axis [3 2 3] - JH	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
C	84	C13		NSD									
C	85	C33		NSD									
C	86	C43		NSD									
C	87	C24		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-5AD-100204 *Lab/Cor Sample No.:* B4760 S93 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	B44		NSD									
A	2	B34		NSD									
A	3	B24		NSD									
A	4	B14		NSD									
A	5	B4		NSD									
A	6	C4		NSD									
A	7	C14		NSD									
A	8	C24		NSD									
A	9	C34		NSD									
A	10	C44		NSD									
A	11	B42		NSD									
A	12	B32		NSD									
A	13	B22		NSD									
A	14	B12		NSD									
A	15	B2	AZQ	1	1	F	7.5	1.2	6.2	5770	15617	Mg,Si,Ca,Fe Actinolite Zone Axis - [1 0 1] KM	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
A	16	C2		NSD									
A	17	C12		NSD									
A	18	C22		NSD									
A	19	C32		NSD									
A	20	C42		NSD									
A	21	A41		NSD									
A	22	A31		NSD									
A	23	A21		NSD									
A	24	A11		NSD									
A	25	A1		NSD									
A	26	D1		NSD									
A	27	D11		NSD									
A	28	D21		NSD									
A	29	D31		NSD									
A	30	D41		NSD									
B	31	C44		NSD									
B	32	C24		NSD									
B	33	C4		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-5AD-100204 *Lab/Cor Sample No.:* B4760 S93 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
B	34	B14		NSD									
B	35	B23		NSD									
B	36	B3	AQ	2		MD1-1	8	3	2.7			Actinolite	AS>5, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
B	36	B3	AQ		2	MF	8	1.25	6.4			Mg, Al, Si, Ca, Fe Actinolite	AFB>5, PCMEF-US
B	37	C13		NSD									
B	38	C33		NSD									
B	39	C22		NSD									
B	40	C2		NSD									
B	41	B12		NSD									
B	42	B21		NSD									
B	43	B1		NSD									
B	44	C11		NSD									
B	45	C31		NSD									
B	46	C20		NSD									
B	47	B10		NSD									
B	48	A20		NSD									
B	49	D10		NSD									
B	50	D30		NSD									
B	51	D1		NSD									
B	52	A11		NSD									
B	53	A22		NSD									
B	54	A2		NSD									
B	55	D12		NSD									
B	56	D32		NSD									
B	57	D23		NSD									
B	58	D3		NSD									
B	59	A13		NSD									
B	60	A24		NSD									
C	61	C44		NSD									
C	62	C24		NSD									
C	63	C4		NSD									
C	64	B14		NSD									
C	65	B23		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-5AD-100204 *Lab/Cor Sample No.:* B4760 S93 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
C	66	B3		NSD									
C	67	C13		NSD									
C	68	C33		NSD									
C	69	C42		NSD									
C	70	C22		NSD									
C	71	C2		NSD									
C	72	B12		NSD									
C	73	B21		NSD									
C	74	B1		NSD									
C	75	C11		NSD									
C	76	C31		NSD									
C	77	C40		NSD									
C	78	C20		NSD									
C	79	B10		NSD									
C	80	A20		NSD									
C	81	D10	AQ	3	3	F	10	1.5	6.7			Mg, Si, Ca, Fe Actinolite	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
C	82	D30	AQ	4		MD1-0	8	6	1.3			Actinolite	AS>5, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
C	82	D30	AQ		4	MF	4	0.1	40			Mg, Si, Ca, Fe Actinolite	
C	83	D41		NSD									
C	84	D21	AQ	5	5	F	4.5	0.45	10			Mg, Si, Ca, Fe Actinolite	TAS_AHRA
C	85	D1		NSD									
C	86	A11		NSD									
C	87	A22		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-6AD-100204 *Lab/Cor Sample No.:* B4760 S94 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	A44		NSD									
A	2	A24		NSD									
A	3	A4		NSD									
A	4	D14		NSD									
A	5	D34		NSD									
A	6	D44		NSD									
A	7	D42		NSD									
A	8	D22		NSD									
A	9	D2		NSD									
A	10	A12		NSD									
A	11	A32		NSD									
A	12	A42		NSD									
A	13	A40		NSD									
A	14	A20		NSD									
A	15	A10		NSD									
A	16	D10		NSD									
A	17	D30		NSD									
A	18	D40		NSD									
A	19	C41		NSD									
A	20	C21		NSD									
A	21	C1		NSD									
A	22	B11		NSD									
A	23	B31		NSD									
A	24	B41		NSD									
A	25	B43		NSD									
A	26	B23		NSD									
A	27	B3		NSD									
A	28	C13		NSD									
A	29	C33		NSD									
B	30	B44		NSD									
B	31	B24		NSD									
B	32	B4		NSD									
B	33	C14		NSD									
B	34	C34		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-6AD-100204 *Lab/Cor Sample No.:* B4760 S94 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
B	35	C42		NSD									
B	36	C22		NSD									
B	37	C2		NSD									
B	38	B12		NSD									
B	39	B32		NSD									
B	40	B42		NSD									
B	41	B40		NSD									
B	42	B20		NSD									
B	43	B10	AZQ	1	1	F	6	0.5	12	1220	904	Mg, Al, Si, Ca, Fe Actinolite Zone Axis [4 1 3] - JH	AS>5, AFB>5, PCMEF-US, PCMES-US, PSAS 5-10, PSAS TOT, PSAM 5-10, PSAM TOT, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
B	44	C10		NSD									
B	45	C30		NSD									
B	46	C40	CDQ	2		MD1-1	8	3	2.7			Chrysotile	AS>5, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
B	46	C40	CDQ		2	MB	6	0.25	24	1221	905	Mg, Si Chrysotile Verified - JH	AFB>5, PCMEF-US, PSAS 5-10, PSAS TOT, PCAS 5-10, PCAS TOT
B	47	D41		NSD									
B	48	D21		NSD									
B	49	D1		NSD									
B	50	A11		NSD									
B	51	A31		NSD									
B	52	A41		NSD									
B	53	A43		NSD									
B	54	A23		NSD									
B	55	A3		NSD									
B	56	D13		NSD									
B	57	D33		NSD									
B	58	D43		NSD									
C	59	A43		NSD									
C	60	A23		NSD									
C	61	A3		NSD									
C	62	D13		NSD									
C	63	D33		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R6

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: SVM-L2-6AD-100204 *Lab/Cor Sample No.:* B4760 S94 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
C	64	D43		NSD									
C	65	D41		NSD									
C	66	D21		NSD									
C	67	D1		NSD									
C	68	A11		NSD									
C	69	A31		NSD									
C	70	A41		NSD									
C	71	B40		NSD									
C	72	B20		NSD									
C	73	B10		NSD									
C	74	C10		NSD									
C	75	C30		NSD									
C	76	C40		NSD									
C	77	C42		NSD									
C	78	C22		NSD									
C	79	C2		NSD									
C	80	B12		NSD									
C	81	B32	AQ	3	3	F	4.5	1.2	3.7			Mg, Al, Si, Ca, Fe Actinolite	TAS_AHRA
C	82	B42		NSD									
C	83	B44		NSD									
C	84	B24		NSD									
C	85	B4		NSD									
C	86	C14		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: CC1-L6-1CA-100204 *Lab/Cor Sample No.:* B4760 S8 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	A1		NSD									
A	2	A11	AZQ	1	1	F	3.5	0.35	10	844	15152	Mg, Al, Si, Ca, Fe Actinolite Zone Axis [3 0 2] - KM	TAS_AHRA
A	3	A10		NSD									
A	4	B10		NSD									
A	5	B11		NSD									
A	6	B1		NSD									
A	7	C1		NSD									
A	8	C11		NSD									
A	9	C10		NSD									
A	10	D10		NSD									
A	11	D11		NSD									
A	12	D1		NSD									
A	13	A21		NSD									
A	14	A20		NSD									
A	15	B20		NSD									
A	16	B21		NSD									
B	17	B10		NSD									
B	18	B11		NSD									
B	19	B1		NSD									
B	20	C1		NSD									
B	21	C11		NSD									
B	22	C10	CDQ	2	2	F	0.6	0.05	12	5233	15153	Mg, Si, Fe Chrysotile Confirmed - KM	TAS_AHRA
B	22	C10	AZQ	3	3	F	3.5	0.5	7.0	5234	15154	Mg, Al, Si, Ca, Fe Actinolite Zone Axis [1 0 0] - KM	TAS_AHRA
B	23	D10	AZQ	4	4	F	9.5	2	4.8	5235	15155	Mg, Al, Si, Ca, Fe Actinolite Zone Axis [3 1 0] - KM	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
B	24.	D11		NSD									
B	25	D1		NSD									
B	26	A1		NSD									
B	27	A11		NSD									
B	28	A32		NSD									
B	29	A31		NSD									
B	30	A30	ADQ	5	5	F	4.25	1.1	3.9			Mg, Si, Ca, Fe Actinolite	TAS_AHRA

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001-FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: CC1-L6-1CA-100204 *Lab/Cor Sample No.:* B4760 S8 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
B	31	B30		NSD									
B	32	B31		NSD									
B	33	B32		NSD									
B	34	B33		NSD									
B	35	B34		NSD									
B	36	B14		NSD									
B	37	B13		NSD									
B	38	B12		NSD									
B	39	A12		NSD									
B	40	D2		NSD									
B	41	C2		NSD									
B	42	C3		NSD									
B	43	C4		NSD									
B	44	C24		NSD									
B	45	C23		NSD									
B	46	C22		NSD									
B	47	C21	AZQ	6	6	F1-0	2.2	0.4	5.5	5236	15156	Mg, Si, Ca, Fe Tremolite Zone Axis [3 1 0] - KM	TAS_AHRA

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: CC1-L6-1CB-100204 *Lab/Cor Sample No.:* B4760 S9 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	D1		NSD									
A	2	D3		NSD									
A	3	D24	CDQ	1		MD1-0	4.9	1.5	3.3			Chrysotile	TAS_AHRA
A	3	D24	CDQ		1	MB	4.9	0.4	12	811	558	Mg, Si Chrysotile Confirmed - KM	
A	4	D22		NSD									
A	5	D20		NSD									
A	6	C21	AZQ	2	2	F	11	1.5	7.3	812	559	Mg, Si, Ca, Fe Actinolite Zone Axis [2 0 -3] - KM	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS>10_AHRA
A	7	C23	OZQ	3	3	F	4.8	0.6	8.0	813	560	Na, Mg, Al, Si, Ca, Fe Edenite Zone Axis [3 2 1] - KM	TOS_AHRA
A	8	C43		NSD									
A	9	C41		NSD									
A	10	D40		NSD									
A	11	D42		NSD									
A	12	A2		NSD									
A	13	A4		NSD									
A	14	A24		NSD									
A	15	A22		NSD									
A	16	A20		NSD									
A	17	B21		NSD									
A	18	B23		NSD									
A	19	B24		NSD									
A	20	B34		NSD									
A	21	B44		NSD									
A	22	B33		NSD									
A	23	B13		NSD									
A	24	C3		NSD									
B	25	B24		NSD									
B	26	B4		NSD									
B	27	C14		NSD									
B	28	C34		NSD									
B	29	C44		NSD									
B	30	C43		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: CC1-L6-1CB-100204 *Lab/Cor Sample No.:* B4760 S9 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
B	31	C23		NSD									
B	32	C3		NSD									
B	33	B13		NSD									
B	34	B33		NSD									
B	35	B22		NSD									
B	36	B2		NSD									
B	37	C12		NSD									
B	38	C32		NSD									
B	39	C42		NSD									
B	40	C41		NSD									
B	41	C21		NSD									
B	42	C1		NSD									
B	43	B11	CDQ	4	4	F	1.2	0.05	24			Chrysotile	TAS_AHRA
B	44	B20	ADQ	5	5	F	1.25	0.4	3.1			Mg, Al, Si, Ca, Fe Actinolite	TAS_AHRA
B	45	A20		NSD									

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: CC1-L6-2CB-100204 *Lab/Cor Sample No.:* B4760 S10 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	1	B44		NSD									
A	2	B24		NSD									
A	3	B4		NSD									
A	4	C14		NSD									
A	5	C24		NSD									
A	6	C44		NSD									
A	7	C43		NSD									
A	8	C23	ADQ	1	1	F	2.5	0.4	6.2			Mg, Al, Si, Ca, Fe Actinolite	TAS_AHRA
A	8	C3		NSD									
A	9	B13	ADQ	2	2	F	6	1.3	4.6			Mg, Al, Si, Ca, Fe Actinolite	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
A	10	B33		NSD									
A	11	B43		NSD									
A	12	B32		NSD									
A	13	B12		NSD									
A	14	C2		NSD									
A	15	C22		NSD									
A	16	C42		NSD									
A	17	C41		NSD									
A	18	C21		NSD									
A	19	C1		NSD									
A	20	B11		NSD									
A	21	B31		NSD									
A	22	B40	ADQ	3		MD1-0	5	5	1.0			Actinolite	TAS_AHRA
A	22	B40	ADQ		3	MF	2.5	0.3	8.3			Mg, Al, Si, Ca, Fe Actinolite	
A	23	B20		NSD									
A	24	A10		NSD									
A	25	A30		NSD									
A	26	A41		NSD									
A	27	A21		NSD									
A	28	A1	ADQ	4	4	F	5	0.5	10			Mg, Al, Si, Ca, Fe Actinolite	TAS_AHRA
A	29	D11	CDQ	5	5	F	5	0.15	33	814	561	Mg, Si Chrysotile Confirmed - KM	TAS_AHRA

Lab/Cor, Inc.

A Professional Service Corporation in the Northwest

Report # 041172R3

Client: Ecology and Environment, Inc.

Project Name: Site# 0440.01CP, 0440.01CP-0001- FINAL RESULTS

TEM ASBESTOS STRUCTURE COUNT - RAW DATA

Sample No.: CC1-L6-2CB-100204 *Lab/Cor Sample No.:* B4760 S10 A1

Descripiton:

Gr	No.	Loc.	ID	Prim	Tot	Class	Len	Wid	Asp	Neg#	EDS#	Comment	Count Categories
A	29	D11	AZQ	6	6	F	10	0.7	14	815	562	Mg, Al, Si, Ca, Fe Actinolite Zone Axis [3 0 4] - KM	AS>5, AFB>5, PCMEF-US, PCMES-US, TAS_AHRA, AS>5_AHRA, AS5-10_AHRA
B	30	B44		NSD									
B	31	B24		NSD									
B	32	B4		NSD									
B	33	C14		NSD									
B	34	C34		NSD									
B	35	C43		NSD									
B	36	C23		NSD									
B	37	C3		NSD									
B	38	B13		NSD									
B	39	B33	ADQ	7	7	F	15	3.5	4.3			Mg, Si, Ca, Fe Actinolite	AS>5, AFB>5, TAS_AHRA, AS>5_AHRA, AS>10_AHRA
B	40	B42		NSD									
B	41	B22		NSD									
B	42	B12		NSD									
B	43	B2	ADQ	8	8	F	4	1	4.0			Mg, Al, Si, Ca, Fe Actinolite	TAS_AHRA
B	44	C12		NSD									