LaPlace Chloroprene Air Sampling

Date: 8/28 - 8/29/16			Set Up			Timer Setting				Collection			
Location	Canister ID	Flow Controller ID	Date	Time	Pres. ("Hg)	Start Date	Start Time	Finish Date	Finish Time	Date	Time	Pres. ("Hg)	Comments
E. St. John HS	A21080	PR-10	8/28/16	1150	L-30					8/29/16	1151	-7	PASSED LEAK CHECK (LC)
Ochsners	SAT082	PR-9	8/28/16	1154	-30					8/29/16	1155	-8	PASSED LC W/IN-LINE FILTER
Acorn and Hwy 44	SAT 114	PR-22	8/28/16	1202	L-30					8/29/16	1204	-9	PASSED LC W/IN-LINE FILTER
Levee	SAT050	PR-7	8/28/16	1205	L-30					8/29/16	1208	-11	PASSED LC *
5th Ward Elementary	5096	PR-4	8/28/16	1213	L-30					8/29/16	1214	-11	PASSED LC CO-LOCATED
238 Chad Baker	183	PR-23	8/28/16	1217	L-30					8/29/16	1218	-7.8	PASSED LC W/IN-LINE FILTER
5TH WARD ELEMENTARY	5062	PR-18	8/28/16 ~~8/29/16~~ EW	1213	L-30 ~~L-70~~ EW					8/29/16	1214	-8	PASSED LC CO-LOCATED
				LAST LINE EW 8/29/16									

Notes: WEATHER: WIND PREDICTED FROM E @ 10-15 mph DURING DAY of 8/28, THEN BECOME LIGHT AND VARIABLE DURING EVENING of 8/28 AND THEN FROM ENE @ 5-10 mph DURING DAY of 8/29. BASED ON PREDICTED WIND, SELECTED 5TH WARD ELEMENTARY AS CO-LOCATED SAMPLER LOCATION.

* - GAUGE READS -12"Hg, BUT NEEDLE RESTS @ -1"Hg.