LaPlace Chloroprene Air Sampling

Date: 8/22/16 - 8/23/16

Location	Canister ID	Flow Controller ID	Set Up Date	Set Up Time	Set Up Pres.	Timer Setting Start Date	Timer Setting Start Time	Timer Setting Finish Date	Timer Setting Finish Time	Collection Date	Collection Time	Collection Pres.	Comments
E. St. John HS	5126 ~~5106~~	PR-9	8/22/16	1238	-30" Hg					8/23/16	1333	-7" Hg	PASSED LEAK CHECK (LC)
Ochsners	5015	PR-12	8/22/16	1252	-30" Hg					8/23/16	1340	-8" Hg	PASSED LC CO-LOCATED
Acorn and Hwy 44	A21073	PR-23	8/22/16	1313	-30" Hg					8/23/16	1350	~~-8"~~ DB -7" Hg	PASSED LC
Levee	SAT 157	PR-19	8/22/16	1322	-30" Hg					8/23/16	1353	~~-4" Hg~~ -4" Hg	PASSED LC
5th Ward Elementary	5106	PR-22	8/22/16	1329	-30" Hg					8/23/16	1357	-30" Hg	PASSED LC VOIDED, VALVE NOT OPENED
238 Chad Baker	A21094	PR-4	8/22/16	1336	-30" Hg					8/23/16	1400	-9" Hg	PASSED LC
Ochsner	5051	PR-10	8/24/16	1254	-30" Hg					8/23/16	1340	-8" Hg	PASSED LC CO-LOCATED

Notes: Weather forecasted winds SSW - WEST 5-10 mph over next 24 hrs. Colocated at Ochsner due to downwind location. FC PR9 NEEDED STICKING. TAP GLASS TO GET READING. VALVE NOT OPENED ENOUGH AT 5th WARD