LaPlace Chloroprene Air Sampling

Date: 12/14-15/16			Set Up			Collection			
Location	Canister ID	Flow Controller ID	Date	Time (Hrs)	Pres. ("Hg)	Date	Time (Hrs)	Pres. ("Hg)	Comments
E. St. John HS	A21076	PR-18	12/14/16	1254	L-30	12/15/16	1255	-7.5	
Ochsners	19280	PR-4	12/14/16	1259	L-30	12/15/16	1300	-8	
Acorn and Hwy 44	SAT156	PR-22	12/14/16	1306	L-30	12/15/16	1307	-8.5	w/ in-line filter
Levee	SAT099	PR-24	12/14/16	1310	L-30	12/15/16	1311	-8	
5th Ward Elementary	SAT174	PR-23	12/14/16	1315	L-30	12/15/16	1316	-8	w/ in-line filter
238 Chad Baker	SAT125	PR-10	12/14/16	1319	L-30	12/15/16	1320	-7	
5TH WARD ELEMENTARY	SAT077	PR-9	12/14/16	1315	L-30	12/15/16	1316	-7.5	CO-LOCATED w/ in-line filter
	LAST LINE EW 12/15/16								

Notes: 1) SAMPLER = ~~ERIK~~ HADWIN = EH =

2) All canisters/flow controllers passed leak check prior to deployment.

3). WEATHER = CURRENT (12/14) = ~~WNW~~ WNW @ 3mph, PREDICTED - LIGHT & VARIABLE DURING DAY of 12/14, THEN N WINDS @ 5-10 mph DURING EVENING of 12/14, AND THEN NNE WINDS @ 10-15 mph DURING DAY of 12/15. BASED ON PREDICTED WINDS, SELECTED 5TH WARD ELEMENTARY AS CO-LOCATED SAMPLE LOCATION.