LaPlace Chloroprene Air Sampling

Date: 7/23 - 7/24/16			Set Up			Timer Setting				Collection			
Location	Canister ID	Flow Controller ID	Date	Time	Pres. ("Hg)	Start Date	Start Time	Finish Date	Finish Time	Date	Time	Pres. ("Hg)	Comments
E. St. John HS	406	PR-2	7/23/16	1222	.30					7/24/16	1223	-6	PASSED LEAK CHECK (LC) w/ IN-LINE FILTER (*)
Ochsners	5134	PR-9	7/23/16	1226	.30					7/24/16	1227	-8	PASSED LC w/ IN-LINE FILTER
Acorn and Hwy 44	SAT120	PR-19	7/23/16	1233	-30					7/24/16	1234	(2) -4	PASSED L w/ IN-LINE FILTER (*)
Levee	19284	PR-7	7/23/16	1238	<-30					7/24/16	1239	(1) -9	PASSED LC CO-LOCATED
5th Ward Elementary	036	PR-5	7/23/16	1242	<-30					7/24/16	1243	-8.5	PASSED LC
238 Chad Baker	AZ1015	PR-18	7/23/16	1246	<-30					7/24/16	1247	-7	PASSED LC
LEVEE	SAT005	PR-10	7/23/16	1238	<-30					7/24/16	1239	-8	PASSED LC CO-LOCATED
				LAST LINE									

N/A

Notes: 1) WEATHER - 7/23 - 93°F (75°F/Lo / 93°F/Hi), MOSTLY CLOUDY; 30% CHANCE OF RAIN; WIND NW @ 3-5 mph. WIND PREDICTED TO BE LIGHT; VARIABLE DURING DAY OF ~~7/23~~ 7/23; THEN ~~REMAIN~~ REMAIN LIGHT AND VARIABLE IN EVENING OF 7/23; AND THEN REMAIN LIGHT AND VARIABLE DURING DAY OF 7/24/16. BASED ON ~~PRD~~ WIND (PREDICTED AND CURRENT), SELECTED LEVEE AS CO-LOCATED SAMPLE LOCATION.

2) SAMPLER = ERIK HADWIN = EY =

(*) MOISTURE OBSERVED IN FLOW CONTROLLER GAUGE @ TIME OF COLLECTION

(1) FLOW CONTROLLER GAUGE READ -10" Hg; HOWEVER GAUGE NEEDLE REST @ -1" Hg

(2) PASSED ~~POST CF~~ POST-COLLECTION LEAK TEST