LaPlace Chloroprene Air Sampling

Date: 7/26 - 7/27/16

Location	Canister ID	Flow Controller ID	Set Up Date	Set Up Time	Set Up Pres. (" Hg)	Timer Setting Start Date	Timer Setting Start Time	Timer Setting Finish Date	Timer Setting Finish Time	Collection Date	Collection Time	Collection Pres. (" Hg)	Comments
E. St. John HS	5045	PR-19	7/26/16	1202	-30					7/27/16	1210	-6	PASSED LEAK CHECK (LC) w/ IN-LINE FILTER
Ochsners	A21053	PR-18	7/26/16	1210	<-30					7/27/16	1214	-7	PASSED LC CO-LOCATED
Acorn and Hwy 44	A21053	PR-5	7/26/16	1218	<-30					7/27/16	1223	-8.5	PASSED LC
Levee	A21000	PR-7	7/26/16	~~1223~~ 1223 GH	<-30					7/27/16	1227	-9.5	PASSED LC
5th Ward Elementary	5074	PR-2	7/26/16	1227	-30					7/27/16	1230	-7.5	PASSED LC (*) w/ IN-LINE FILTER
238 Chad Baker	A21032	PR-10	7/26/16	1232	<-30					7/27/16	1234	-7.5	PASSED LC
OCHSNERS	5024	PR-9	7/26/16	1210	-30					7/27/16	1214	-9	PASSED LC CO-LOCATED w/ IN-LINE FILTER
			LAST LINE GH 7/27/16										

Notes: 1) WEATHER - 7/26 - CURRENTLY ~~CURRENTLY~~ WIND 0-2 mpH FROM SE, WIND PREDICTED TO BE FROM SSW @ 5-10 mpH THROUGH EVENING OF 7/26/16 AND THEN PREDICTED TO COME FROM SW @ 10-15 mpH DURING DAY OF 7/27/16. BASED ON PREDICTED WIND DIRECTION, SELECTED OCHSNERS AS CO-LOCATED SAMPLE LOCATION.

2) SAMPLER = ERIK HADWIN = GH

3) (*) - PRESSURE GAUGE ON FLOW CONTROLLER PR-2 w/ CANISTER 5074 HAD CONDENSATION INSIDE GAUGE AT TIME OF COLLECTION.