LaPlace Chloroprene Air Sampling

Date: 5/25-26/17

Location	Canister ID	Flow Controller ID	Set Up Date	Set Up Time (HRS)	Set Up Pres. ("Hg)	Collection Date	Collection Time (HRS)	Collection Pres. ("Hg)	Comments
E. St. John HS	SAT061	PR-5	5/25/17	1246	2.30	5/26/17	1247	-8.5	CO-LOCATED
Ochsners	~~SAT105~~	PR-15		1250	2.30		1252 ~~1251~~ EH 5/26/17	-7.25	w/in-line filter
Acorn and Hwy 44	SAT170 (4) ~~5116~~ EH 5/25/17	PR-20		1300	-29.5		1301	-6.1	w/in-line filter
Levee	3225	PR-23		1304	2.30		1305	-9	w/in-line filter
5th Ward Elementary	A21101	PR-10		1308	2.30		1309	-9	
238 Chad Baker	SAT098	PR-8	EH 5/25/17	1312	2.30	EH 5/26/17	1313	-7.5	w/in-line filter
E. ST. JOHN HS	A21060	PR-24	5/25/17	1246	2.30	5/26/17	1247	-7.5	CO-LOCATED w/in-line filter
			LAST LINE EH			5/26/17			

Notes: 1) SAMPLER = ERIK HADWIN = EH =

2) ALL CANISTERS / FLOW CONTROLLERS PASSED LEAK CHECK PRIOR TO DEPLOYMENT.

3) WEATHER - CURRENT (5/25) - WINDS SSW @ 8 mph, PREDICTED - WINDS SW @ 5-10 mph DURING DAY OF 5/25, AND THEN WINDS S 5-20 mph THROUGH DAY OF 5/26. BASED ON WINDS, SAMPLER SELECTED EAST ST. JOHN HIGH SCHOOL AS CO-LOCATED SAMPLE LOCATION.

(4) ~~CANIS~~ CANISTER LOST 1 TO 2 "Hg FROM TIME OF SET-UP ~~AND~~ DEPLOYMENT (~3 hrs). DUE TO LOSS OF ~~VA~~ VACUUM, CANISTER EXCHANGED. FAULTY CANISTER RETURNED TO LAB AS INVALID.
on 5/25/17
5/25/17