LaPlace Chloroprene Air Sampling

Date: 7/30-31/17			Set Up			Collection			
Location	Canister ID	Flow Controller ID	Date	Time (Hrs)	Pres. ("Hg)	Date	Time (Hrs)	Pres. ("Hg)	Comments
E. St. John HS	5149	PR-10	7/30/17	1334	-30	7/31/17	1335	-8	
Ochsners	A21074	PR-15		1338	-30		1339	-7.75	
Acorn and Hwy 44	19288	PR-2		1345	-30		1346	-7.75	
Levee	SAT032	PR-23		1349	-30		1350	-8.6	
5th Ward Elementary	SAT082	PR-24		1353	-30		1354	-8.5	CO-LOCATED
238 Chad Baker	SAT087	PR-8	EW 7/30/17	1356	-30	EW 7/31/17	1357	-8	
5TH WARD ELEMENTARY	A21109	PR-18	7/30/17	1353	-30	7/31/17	1354	-7.75	CO-LOCATED
		LAST	LINE	EW		7/31/17			

Notes: 1) SAMPLER = ERIK LADWIN = EW

2) ALL CANISTERS / FLOW CONTROLLERS PASSED LEAK CHECK PRIOR TO DEPLOYMENT. ALL FLOW CONTROLLERS EQUIPPED W/ IN-LINE FILTERS EXCEPT PR-10.

3) WEATHER = CURRENT (7/30) - WINDS NNW @ 7mph, PREDICTED - WINDS NNE @ 5-10mph. DURING DAY OF 7/30, THEN LIGHT & VARIABLE DURING EVENING OF 7/30, AND THEN WINDS E @ 10-15 DURING DAY OF 7/31. BASED ON WINDS, SAMPLER SELECTED 5TH WARD ~~ELEMENTARY~~ ELEMENTARY/CO 7/30 AS CO-LOCATED SAMPLE LOCATION.

SHIPPED 7/31/17