LaPlace Chloroprene Air Sampling

Date: 7/6-7/17

Location	Canister ID	Flow Controller ID	Set Up Date	Set Up Time (Hrs)	Set Up Pres. ("Hg)	Collection Date	Collection Time (Hrs)	Collection Pres. ("Hg)	Comments
E. St. John HS	A21046	PR-10	7/6/17	1339	-30	7/7/17	1340	-9	CO-LOCATED
Ochsners	5004	PR-15		1344	-30		1345	-8	
Acorn and Hwy 44	A21106	PR-2		1351	-30		1352	-8.25	
Levee	5035	PR-23		1355	-30		1356	-9.5	
5th Ward Elementary	SAT089	PR-20		1359	-29.75		1400	-6.5	
238 Chad Baker	SAT140	PR-8	7/6/17	1403	-30	7/7/17	1404	-8	
EAST ST. JOHN HS	5059	PR-24	7/6/17	1339	-30	7/7/17	1340	-9.5	CO-LOCATED
		LAST	LINE	ON	7/7/17				

Notes: 1) SAMPLER = ERIK HADWIN = EH =

2) All canisters / flow controllers passed leak check prior to deployment. All flow controllers equipped w/ in-line filters except PR-10.

3) Weather - current (7/6) - winds S @ 2-5 mph; predicted - winds SSE @ 5-10 mph during day of 7/6, then winds SSW @ 0-5 mph during evening of 7/6, and then winds WSW @ 5-10 mph during day of 7/7. Based on 7/6/17 winds, sampler selected East St. John High School as co-located sampler location.

Shipped 7/7/17