LaPlace Chloroprene Air Sampling

Date: 5/31 - 6/1/17			Set Up			Collection			
Location	Canister ID	Flow Controller ID	Date	Time (HRS)	Pres. ("Hg)	Date	Time (HRS)	Pres. ("Hg)	Comments
E. St. John HS	SAT042	PR-4	5/31/17	1156	-30	6/1/17	1249	-12	
Ochsners	SAT016	PR-15		1201	-30		1254	-7	w/ in-line filter
Acorn and Hwy 44	SAT112	PR-20		1208	-29.5		1302	-5.75	w/ in-line filter
Levee	SAT017	PR-23		1214	-30		1306	-8.25	w/ in-line filter
5th Ward Elementary	5065	PR-10		1218	-30		1310	-10	
238 Chad Baker	SAT032	PR-8	5/31/17	1223	-30	6/1/17	1314	-7.5	co-located w/ in-line filter
238 CHADBAKER	5016	PR-24	5/31/17	1223	-30	6/1/17	1314	-8	co-located w/ in-line filter
			LAST LINE			6/1/17			

Notes: 1) Sampler = Eric Hadwin = EH =

2) All canisters / flow controllers passed leak check prior to deployment.

3) Weather - current (5/31) - winds SE 7 mph; predicted - winds ESE @ 5-10 mph during day of 5/31, light and variable during evening of 5/31, and then winds ESE @ 5-10 mph during day of 6/1/17. Based on winds, sampler selected 238 Chad Baker as co-located sample location.